TITLE:

A target product profile for electronic clinical decision support algorithms combined with point-of-

care diagnostic test results to support evidence-based decisions during patient consultations by

health workers

AUTHORS:

Karell G. Pellé1*, Clotilde Rambaud-Althaus2, Valérie D'Acremont3,4, Gretchen Moran5, Rangarajan

Sampath1, Zach Katz1, Francis Moussy6, Garrett L. Mehl7, Sabine Dittrich1

AUTHORS INSTITUTIONS:

1. Foundation for Innovative New Diagnostics, Chemin des Mines 9, 1202 Geneva Switzerland

2. Doctors Without Borders (MSF), Operational Centre Geneva, Rue de Lausanne 78, 1202 Geneva

Switzerland

3. Swiss Tropical and Public Health Institute, University of Basel, 4002 Basel, Switzerland

4. Centre for Primary Care and Public Health, University of Lausanne, Switzerland

5. eHealth Africa, 4-6 Independent Road, Kano Nigeria

6. Department of Essential Medicines and Health Products, World Health Organization, 20 Avenue

Appia, 1211 Geneva Switzerland

7. Department of Reproductive Health and Research, World Health Organization, 20 Avenue Appia,

1211 Geneva Switzerland

***CORRESPONDING AUTHOR:**

Karell G. Pellé, PhD

Foundation for Innovative New Diagnostics

Chemin des Mines 9, 1202 Geneva Switzerland

Email: Karell.Pelle@finddx.org

WORD COUNT: 3498

KEYWORDS

Target Product Profile, Clinical Guideline, Algorithm, Clinical Decision Support System, Diagnostics,

Point of Care

ABSTRACT

Health workers in low-resource settings often lack the support and tools to follow evidence-based clinical recommendations for diagnosing, treating and managing sick patients. Digital technologies, by combining patient health information and point of care diagnostics with evidence-based clinical protocols, can help improve the quality of care, the rational use of resources (humans, diagnostics and medicines) and save patient lives. The development of a target product profile for electronic clinical decision support algorithms (CDSAs) aimed at guiding preventive or curative consultations, and that integrate diagnostic test results will help align developer and implementer processes and specifications with the needs of end-users, in terms of quality, safety, performance and operational functionality. To identify characteristics for a CDSA, experts from academia, research institutions, and industry as well as policy makers with expertise in diagnostic and CDSA development, and implementation in LMICs were convened. Experts discussed the critical characteristics of a draft TPP which was revised and finalised through a Delphi process to facilitate consensus building. Experts were in overwhelming agreement that, given that CDSAs provide patients' management recommendations, the underlying clinical algorithms should be available in human readable format and evidence-based. Whenever possible, the algorithm output should take into account pre-test disease probabilities, diagnostic likelihood ratios of clinical or laboratory predictors and disease probability thresholds to test and to treat. Validation processes should at a minimum ensure the CDSA are implementing faithfully the evidence-based algorithm they are based on (internal validation through clinical association and analytical validation). Additionally, clinical validation, bringing practice evidence about the impact of the CDSA use on health outcomes, was recognized as a good to have. The CDSAs should be designed to fit within clinic workflows, connectivity challenges and high volume settings. Data collected through the tool should conform to local patient privacy regulations and international data standards.

INTRODUCTION

Health workers at primary health care (PHC) have the greatest challenge in ensuring appropriate care for their communities. One way to support health workers is through the provision of clinical guidance in the form of clinical practice guidelines or clinical decision support algorithms (CDSA). Clinical practice guidelines are a set of recommendations on diagnostic and treatment modalities based on systematic review of evidence and assessment of the benefits and harms for the patients. The World Health Organisation (WHO) developed several simple clinical decision guidelines to support health workers in low- and middle-income countries (LMICs) in the evaluation and management of clinical problems, with the ultimate goal of improving quality of care[1]. This includes the Integrated Management of Childhood Illness (IMCI) designed in the 1990s for common diseases affecting children younger than 5 years of age[2], which was later adapted for use at the community as the integrated Community Case Management guideline (iCCM)[3]. Other guidelines have been developed and either take an integrated management approach similar to IMCI, such as the Integrated Management of Adolescent and Adult Illness (IMAI[4]), or a disease-specific approach (i.e. WHO's dengue guidelines for diagnosis, treatment, prevention and control[5]).

While the global health community awaits updated and more comprehensive clinical algorithms, the Fourth Industrial Revolution[6] is seeing developing economies "leapfrogging"[7] mobile technologies across multiple sectors and including for health. As a result, a plethora of digital clinical decision support systems (CDSS) have been implemented in LMICs. CDSS are defined as "digitised job aids that combine individual's health information with the health worker's knowledge and clinical protocols to assists health workers in making diagnosis and treatment decisions"[8]. It is well recognized that adherence to standardized clinical practice guidelines is a measurement of quality of care[9,10], and that improving adherence to evidence-based guidelines is associated with better health outcomes[11]. By digitalizing clinical guidelines such as IMCI and iCCM, into electronic CDSAs

in its simplest form, a decision tree, several groups have shown significant increase in health worker

adherence to guidelines compared to routine care (paper-based IMCI)[12, 13,14].

However, based on IMCI/iCCM, today these CDSAs carry the same limitations than the main

guidelines, and still lack point of care (POC) diagnostic tests that could improve the accuracy of

diagnoses and treatment recommendations. The only diagnostic test widely used with IMCI/iCCM

guidelines is a malaria rapid diagnostic test (mRDT) and while this inclusion was important in the

context of "test-and-treat"[15], it is limited to ruling in or out one pathogen alone. Acute febrile

illness can be caused by bacterial or viral respiratory infections, or seasonal, geographically localised

pathogens, like dengue. Hence, the inclusion of additional simple wet and digital diagnostics into

adapted algorithms can reduce antimicrobial over-prescriptions as well as patient outcome[16, 17].

With the growing development of CDSAs, especially related to the management of sick children in

LMICs (**Supplemental File 1**), it is important to ensure that new tools meet the needs of the end-user

and take into consideration the emerging landscape of diagnostic tools that might be included in

new algorithms forming a toolkit consisting of CDSA and diagnostics.

The aim of this work was to define a target product profile (TPP) to inform developers and

implementers alike of the requirements for CDSAs aimed at guiding health workers during a

preventive or curative consultation with any type of patient, and that have the ability to incorporate

diagnostic test results, especially when diagnostics tests have been implemented in pre-existing

guidelines.

1. METHODS

1.1. Draft Target Product Profile

A draft TPP (version 0) for CDSAs combined with POC test results was developed (by: KP, SD, RS, ZK, GLM, FM) based on standard procedure at FIND and WHO, where characteristics are defined as "minimal" and "optimal" criteria (**Supplemental File 2**). "Minimal" is used to refer to the lowest acceptable output for a characteristic and "optimal" for the ideal target of a characteristic. Comments are used to explain or provide examples for the characteristics. As this TPP is for a multi-component toolkit, CDSA and POC diagnostics, the structure of the TPP was adapted to cover the different components and for the purpose of clarity. These sections included: Scope General, Scope Toolkit, Electronic Clinical Decision Algorithm, Point of Care Tools, App, Data and Procurement.

1.2. Expert Meeting

To discuss selected aspects of the draft TPP (algorithm validation, performance, machine learning, diagnostic data, disease prediction, clinical workflow and application functionality) (**Supplemental File 2**), a meeting on this subject was convened by FIND and WHO involving 39 experts from academic institutions, industry, private and public sectors from 11 countries[18]. Experts were selected based on their experience in relevant areas of work directly related to digital health (clinical decision support or other digital tools targeting resource-challenges settings) and/or diagnostics (wet and digital). Following the expert meeting the TPP was updated based on expert feedback (version 1.0) (**Supplemental File 3**).

1.3. Delphi Process

To facilitate consensus building for the TPP, we followed a Delphi-like process where the TPP version

1.0 was reviewed. For this, a three- part online survey was developed. The first part of the survey

collected professional information for survey respondents as well as experience with clinical

practice/research, diagnostic development, software development, implementation of healthcare

programs in LMICs, and health information systems. The second part of the survey allowed for

rating of both minimal and optimal criteria for all TPP characteristics. Agreement was scored on a

Likert scale from 1- 5, where 1 corresponds to fully disagree, 2-mostly disagree, 3-neither agree nor

disagree, 4-mostly agree and 5- fully agree. Disagreement with a criterion was based on a rating

from 1 to 3 and required a comment/suggestions by the survey respondent. To reach consensus,

more than 75% of survey respondents should provide a rating of at least 4 (agreement) on the Likert

scale. The third part of the survey provided an opportunity for respondents to share

comments/feedback/ideas on the TPP if not covered in the characteristics section. The survey was

developed with the online software SurveyGizmo (SurveyGizmo LLC, Boulder CO, USA). Survey

analyses were performed using Microsoft Excel 2013 (Microsoft Corporation, Redmond, USA).

Patient and public involvement statement

At no point in this work were patients or the public involved in the design of this study or reporting

and dissemination of the results.

RESULTS

1. Expert Consensus on the Target Product Profile through a Delphi Process

In the first round, the online survey was sent to 77 experts (including workshop participants) and 28 responded to the survey (response rate, 36.4%), with some respondents submitting a coordinated response for their institution. Twenty-five percent of respondents were from LMICs (7/28), however the majority of respondents, (22/28, 78%), had experience working in Africa and 57% (16/28) in South East Asia. More than 85% of respondents (24/28) had more than 5 years of experience in clinical practice and research, 50% (14/28) in health programme implementation in LMICs, 39% (11/28) in health information systems, 21% (6/28) in software development and 14% (4/28) in diagnostic product development **(Supplemental File 4)**. Most participants work in international organizations (7/28, 25%) and academic institutions (10/28, 35.7%). In the first round, at least 50% agreement was reached for all criteria for all characteristics and feedback was taken into consideration to develop TPP version 1.1 for the second round of the Delphi survey (**Supplemental File 5**). In this new version, the characteristic "Encounter" was removed as several participants expressed it was redundant with the characteristic "Workflow". Therefore version 1.1 had 40 characteristics.

A second online survey was sent to the 28 respondents from round 1. In this round, 23 participants responded to the survey (response rate, 82%). Most participants were from academia (7/23, 30%), non-profit international organizations (7/23, 30%) and had a similar profile than respondents from round 1 (**Supplemental File 4**). In this round of review, at least 75% of participants agreed with the minimal and optimal criteria for all TPP characteristics. On average, 91% of survey respondents agreed with the minimal and optimal criteria (**Figure 1**) and the TPP was finalized (version 2).

2. Final Target Product Profile

Scope General and Scope Toolkit Components. This TPP defines a toolkit composed of CDSAs and

POC diagnostics intended to support evidence-based clinical decisions taken by health workers

during a preventive or curative consultation, by capturing patient clinical and context-specific data,

as well as diagnostic test results to provide recommendations on diagnosis and patient management

(including preventive & curative treatment, counseling, etc.). The clinical algorithm integrates POC

diagnostic results and is embedded in an App. POC diagnostics in this toolkit should be regulatory

approved for use in the setting of implementation.

The experts agreed that the CDSA should first define its target i.e. i) patients (e.g. age group), ii)

medical problems (e.g. syndrome or disease), iii) end users (e.g. minimally trained health worker or

trained clinician) and iv) level of care (e.g. community or health facility) for which the clinical

recommendations are meant and relevant. These specifications should be clearly communicated to

the end-user at the beginning of the algorithm. The health worker should immediately be warned if

information she or he enters in the system does not fit with these specifications. In cases where an

App proposes different algorithms for different use cases (i.e. antenatal care, management of

childhood illness), the specifications shall be clearly available for each individual algorithm (**Table 1**).

Table 1: Minimal and optimal target product profile characteristics focused on the scope of the toolkit and toolkit component characteristics, as defined

by expert consensus process.

Scope General			
CHARACTERISTIC	MINIMAL REQUIREMENT	OPTIMAL REQUIREMENT	COMMENT
Intended Use	The toolkit, composed of an electronic clinical decision support algorithm(s) and point of care diagnostic tests, is intended to increase **evidence-based** treatment decisions by capturing diagnostic test results, patient clinical data (e.g. exposures, signs including vital signs) and context specific data (e.g. disease incidence, seasonality) to provide treatment and care recommendations		
Target Population	The algorithm shall define the target population Inclusion and exclusion criteria are used at the encounter to enrol the patient		The system can be modular i.e. composed of discrete algorithms such that one can be used for a specific population based on pre-defined criteria
Setting (level of implementation in the healthcare system)	Defined by the algorithm		The system can be modular i.e. composed of discrete algorithms such that one can be used for a specific healthcare setting based on the infrastructure, workforce knowledge and skills, point of care tools available
Targeted End User	Defined by the algorithm		The end user shall have the required training/skills to use the App appropriately
Scope Toolkit Components			
Algorithm Access	The electronic clinical decision-support algorithm is accessible through an App downloaded on compatible target devices		The App can be a web app, a native app or a hybrid app

Algorithm Content	Built on: — well-established clinical evidence based on WHO/international/local clinical care guidelines, peer-reviewed articles (systematic reviews, original clinical research), clinical experience/practice; **and/or** — appropriate clinical validation research*		* See FDA's SaMD Clinical Evaluation for guidance[20]
Algorithm Treatment Recommendations	Therapeutic recommendations shall be compliant with national treatment guidelines and national EML and support antimicrobial stewardship	Same and evidence based medicine to support optimal treatment recommendations	Recommendations support the appropriate selection, dosage and duration of antimicrobial and any other kind of treatment and management, causing the least harm to the patient
Compatible Point of Care Tools	POCs or other relevant medical devices prompted for use by the App shall be locally relevant, i.e. recommended by EDL or relevant national equivalent, or country program	Same, plus emerging diagnostic tools and medical devices relevant to the algorithm and implementation setting	
Regulated Toolkit Components	POC diagnostic tests and medical devices are regulatory approved and compliant with local regulations	Same, and if the software is a medical device, the App shall also have regulatory approval	
Compatible Devices	The App is compatible with any device including: — Smartphones — Tablets — Computers		Computers are included as it may be necessary for health facility supervisors to access data collected at the facility level to make informed decisions (i.e. restocking medical supplies)
Compatible Operating Systems	OS agnostic	Same as Minimal	

Clinical Decision Support Algorithms. Regarding the algorithm medical content, the working group agreed that it must be evidence-based. Whether newly developed or adapted from existing guidelines, the algorithm medical content should be developed following international standards for the development or adaptation of evidence-based clinical guidelines, to ensure it implements the best available evidence. This evidence can originate from well-established clinical guidelines such as WHO/international/local clinical care guidelines, or from peer-reviewed articles, clinical practice and/or validation research and the level of evidence and the strength of recommendations in the target context should ideally be communicated. In line with this, recommendations should cause the least harm to the patient and to the community. For example, algorithms related to treatment and management of patient suffering from infectious diseases should also support antimicrobial stewardship to avoid unnecessary side effects for the patient and the development of resistance in the community and at global level. In addition, the algorithm should be validated both analytically and semantically to ensure that the algorithm output is accurate and reproducible, doesn't deviate from expert content/evidence and that there are no interactions or conflicts in the logic.

Further, important characteristics were content quality, machine learning (ML). With the rise of ML also in this domain of CDSAs to analyse the data captured through algorithms and then for example augment algorithm diagnostic or prognostic accuracy, workshop experts thought it important to address this in the TPP. It was agreed that an algorithm that recommends treatment decisions and management of patients should be at a minimum human interpretable, meaning one should understand how input data are processed into output data. Changes to the algorithm logic based on ML analyses should also be validated (**Table 2**).

207   **Table 2: Minimal and optimal target product profile characteristics focused on the electronic clinical decision-support algorithm and point of care tools,**

208   **as defined by expert consensus process.**

Clinical Decision-Support Algorithm			
CHARACTERISTIC	**MINIMAL REQUIREMENT**	**OPTIMAL REQUIREMENT**	**COMMENT**
Content Transparency	The algorithm is "human interpretable". The healthcare programme and end user can comprehend the algorithm decision-making processes	The healthcare programme and end user have access to underlying evidence and methodology used to develop the algorithm	Human interpretable: a human can understand the choices taken by the model in the decision-making process, i.e. how output variables are generated based on input variables. Visual representations (e.g.. decision trees, Principle Component Analyses, protocol charts, etc.) and performance metrics can be used to support content transparency
Quality control	The algorithm has been A) analytically and B) semantically tested: A) Analytical verification: the algorithm output is accurate and reproducible B) Semantical verification: the algorithm doesn't deviate from expert content/evidence and there are no interactions or conflicts in the logic		A) and B) answer the question "did I build the model right?" (See FDA's SaMD Clinical Evaluation for current guidance[20])
Algorithm Validation	The algorithm has been validated. The level of validation will depend on the CDSS status as a Software as a Medical Device (SaMD) Refer to upcoming rulings from regulatory bodies such as the FDA or the European Commission		Answers the question "did I build the right model"? (See FDA's SaMD Clinical Evaluation for current guidance[20])

Machine Learning	None. The algorithm is static	ML is applied to generate data on the algorithm performance, improve content, inform healthcare system processes, etc. Changes in the algorithm based on ML shall be validated	
Point Of Care Tool			
POC Data Inputs	Any kind of data (qualitative data such as positive/negative/invalid lateral flow test results and quantitative data such as data provided by haemoglobinometers and glucometers)		
Disease Likelihood	Based on: — pre-test probability (in the absence of POC or POC performance data); **or** — POC positive/negative likelihood ratio	Based on: — pre-test probability; **and** — POC positive/negative likelihood ratio	The test performance (eg. likelihood ratio) is known and performance data ideally previously determined through independent studies in relevant settings The test brand should ideally also be considered so as to account for changes between manufacturers
POC Training	On-site training performed by local authority or implementer		

Point of Care Tools. Experts agreed that the post-test probability or the likelihood of a disease being

present, should be assessed based on the pre-test probability—the prevalence of the disease in the

corresponding patient population—and the diagnostic positive/negative likelihood ratio—how good

a test predicts the presence or absence of disease—of all clinical information i.e. signs, symptoms,

and POC diagnostic test results (**Table 2**). Clinical diagnosis often lacks accuracy and therefore

diagnostic tests that have the proven added value for patient management should be integrated

within algorithms. These tests can be wet diagnostics such as pathogen-specific RDTs or hand-held

devices such as haemoglobinometers, and do not exclude digital diagnostics such as digital

thermometers.

App. An important characteristic in this section is System Validation. Today, there is no standard

process to validate and assess performance accuracy of CDSAs. There is also a severe lack of

evidence and peer-reviewed publications on both the development and the performance of these

tools in LMICs. While some groups have performed randomised controlled trials (RCTs) to measure

patient outcome and other parameters that impact patient health, such as the reduction of

antibiotic over-prescription[16, 17], others have assessed the performance of their tool by

measuring health worker adherence to guidelines[12,13,14]. The level of validation ultimately

depends on the type of algorithm. The USFDA has provided some guidance to developers and

implementers on the kind of evidence that needs to be provided[19]. The TPP criteria are inspired by

this guidance. The minimal criteria guarantee that the App has a CDSA that is based on evidence

(clinical association) and that input data is processed correctly (analytical validation). The optimal

criteria include clinical validation in addition to clinical association and analytical validation. Through

clinical validation, the App has been assessed for the intended outcome (i.e. improved clinical

outcome, better quality of care, rational use of resources and reduction of antibiotic over-

prescription, etc.) (**Table 3**).

Table 3: Minimal and optimal target product profile characteristics focused on App characteristics, as defined by expert consensus process.

App			
CHARACTERISTIC	**MINIMAL REQUIREMENT**	**OPTIMAL REQUIREMENT**	**COMMENT**
System Validation	The App has been validated in the intended implementation setting. There is evidence demonstrating: — Valid Clinical Association (clinical output based on input data is supported by well-established or novel evidence) **and** — Analytical Validation (input data is processed correctly into expected output data)	The App has been validated in the intended setting. There is evidence demonstrating: — Valid Clinical Association (clinical output based on input data is supported by well-established or novel evidence), **and** — Analytical Validation (input data is processed correctly into expected output data), **and** — Clinical Validation (clinical safety or other meaningful outcome relevant to the intended use)	There is evidence that the system is based on evidence and working to achieve the intended use for the intended setting See FDA's SaMD Clinical Evaluation for current guidance[20]
System Access (public API)	Publicly available application programming interface (API) for data access protected by authentication and authorization. At a minimum, technical standards are adhered to	Optimally, HIE/HL7 standards are adhered to	
Context Configuration	— Language: UN official languages — Local time — Local weights and measures	Same as minimal and — Option to customise to local official language — Other country preferences	Language can be a major barrier for the proper use of the tool for patient management and can lead to errors and misinterpretation
Customization	Changes to the App, such as updates to the list of medicines and POCs available in the setting, can be made by the healthcare programme. Validation of this change shall be provided.	Same as Minimal	

User Access Rights	Appropriate data access is provided based on specific roles		Roles may include data manager (facility supervisor) or data entry person (nurse)
Expert Support	None	Built-in access to online/remote expert advice to assist in patient consultation via SMS, audio call, video conferencing	
App Training	On-site training	Same and remote training and/or remote "Train the Trainer"	
Internet Availability	— Functions offline (allows for service delivery and key workflows) — Allows automatic resynchronisation	Same and trigger alerts on user device when data has not been synchronised for a long time	Internet connection can be very unstable therefore the tool should work in offline mode so as to not disrupt the workflow of the user
Clinical Data Entry	Manual entry by the operator	Same, plus automatic upload of digital data (e.g. from biosensors, medical devices)	Optimal: This allows external integration of other App modules, built-in and third party Apps and devices
Patient Management Recommendation	Consultation data is summarised and actionable recommendations provided (e.g. treatment, referral, home care or follow up)	Same **and** recommendations are integrated in EMRs and HIS	
Navigation	Sequential: the user follows a strict sequence of data input to reach a final recommendation	Non-sequential: the user can move in any direction through an assessment and change input data to reach a final recommendation	

Workflow requirements to enable time-delayed POC data input	Ability for a user to perform multiple, simultaneous consultations, with pause and resume capability, to allow clinical and laboratory data entry	Same as minimal plus ability to disable simultaneous workflow feature in settings with minimally skilled workers	This is particularly relevant for implementation in settings where testing and clinical consultations are performed in different locations
Task Management	Multiple algorithms can be supported simultaneously in one application against a common dataset		Can accommodate task-shifting capability i.e. multiple consultations can be opened at a time and patient profiles can be accessed using pre-set user access rights
Follow up	None	Ability to retrieve patient information using patient registration information	The optimal implies data recoverability, also covered below in Data Characteristics section
System Malfunction Protection	System malfunctions are made clear to the user		
Scalability	The App shall allow high transaction volumes with complex workflows to cover primary care workforce at a national scale		
Updates and Versioning	Processes are in place to control any App change (including algorithm version updates) and provide the appropriate and correct update to the user		

In May 2017, a new regulation (EU 2017/745) on medical devices, to which CDSA pertain by definition[20] was adopted by the European Commission, that will be applied starting spring 2020[21]. This regulation includes a reinforcement of the rules on the clinical evidence that should be provided by developers and manufacturers, at least if based in Europe. While waiting for LMICs to have their own regulations, this new regulation will certainly impact development and validation by European developers.

Experts also called for Application Programming Interface (API) to allow other software to 'talk to' and interact with the tool through commonly known programming languages. The App should be protected by authorisation and authentication that as a minimum requirement adhere to usual technical standards, and optimally would adhere to HIE (Health Information Exchange) and/or Health Level 7 International[22] medical standards. It was noted that adherence to HL7 can be expensive, hence this was proposed as an optimal rather than a minimal requirement. The App should also be designed to integrate with patient care pathways and health worker workflows with limited disruption. The CDSA accordingly has pause and resume capability to allow time for diagnostic tests to be performed and results entered in the algorithm. POC diagnostic tests such as mRDTs can require up to 20 min for results to be read, depending on the manufacturer.

Data. Digital tools have the potential to amass enormous amount of data. With CDSA, patient personal data is captured, stored and potentially used for other operational (i.e. supply chain management) and algorithm development purposes. Many countries have personal data privacy and protection laws in place to protect citizens from the abuse of their personal information. In the European Union this is known as the General Data Protection Regulation[23]. However, there are still many countries around the world that do not have a national eHealth strategy or personal data legislation. This called for the need to address personal data handling with CDSA for security and

accountability purposes. Experts agreed that the implementer, who would be considered as the data

controller, is responsible for the data and should comply with local data policies and legislations. No

specific legislation was mentioned as these differ by country and region. In addition, the historic of

the processes that are applied to data as well as the origin of data should be documented (Data

Provenance) and the App should function under secure connectivity to avoid loss and corruption of

sensitive data, and mitigate cyber-attacks, whether data is at rest or in transmission **(Table 4).**

Table 4: Minimal and optimal target product profile characteristics focused on data characteristics, as defined by expert consensus process.

Data			
CHARACTERISTIC	**MINIMAL REQUIREMENT**	**OPTIMAL REQUIREMENT**	**COMMENT**
Data Capture	Text, numeric, image, audio, video	Same, and GPS, barcode, biometric	
Data Validation	The system validates data entry to prevent errors that diminish value of the data or the outcome		
Data Ownership	Ownership shall be determined by the healthcare programme		The healthcare programme is responsible for compliance with any country law, policy and regulation
Data Storage	The healthcare programme shall be able to choose the destination of the App's data	Same as Minimal	
Data Recovery	Data can be recovered or the system can be re-established to the desired state in the event of interruption or failure		
Data Flow	The flow of data shall be determined by the healthcare programme	Same as Minimal	
Data Reporting	Data export available from all target devices	Pre-built data reporting, analytics and dashboards are available with the App	The level of data manipulation, aggregation and reporting should be sensitive to the device the App is running on i.e. the computer App can be rich in functionality, the mobile App is optimised for data collection and exchange only

Data Provenance	Included	Same	Provides origin and processes applied to output data. When data is downloaded or shared, the version of the model is tagged so it is always clear how the data was obtained
Data Dictionary	Available, referencing standards used (eg. International Classification of Diseases, Systematized Nomenclature of Medicine)		Ensures indicators reported are uniform across different health programs
Data Security & Privacy	The App operates under secure connectivity, which meet data protection and regulations of individual countries to avoid loss and corruption of sensitive data, and mitigate cyber-attacks, whether data is at rest or in transmission. Conforms to national privacy laws. Includes processes such as: — Two factor authentication — Authorization/Access Control — De-identified data — Data encryption		Encourages GDPR (should no national data security policies exist) to ensure a system that: — preserves data integrity — identifies & mitigates risks — provides relevant parties security processes

DISCUSSION

In this work, our aim was to develop a TPP to enable the development of a toolkit that includes

CDSAs that are based on evidence and designed to integrate results of diagnostics performed at POC

to guide rational clinical decisions. The use of such toolkit would aid in strengthening healthcare

particularly at the community and primary care echelons, and help address health workers

difficulties to differentiate diseases and syndromes without diagnostic tests and the aid of

formalised diagnostic processes that can be provided through algorithms.

In terms of TPP characteristics, "Data" and "Validation" emerged as key topics of discussion. Data

generated from CDSA require storage and connectivity and are therefore vulnerable to abuse, loss or

unlawful purposes. Although many countries have enacted personal data protection legislation,

many fall short on enforcing these and others are still drafting bills. For example, in the African

region, in 2017, 17 (out of 54) African countries had put in place comprehensive personal data

protection legislation[24]. Several regimes in the region are starting to draft legislation regarding

processing and movement of personal data[25], however funding attrition and the lack of regulators

cripple the process. That said, there are continent-wide initiatives that promise to set data

protection framework such as EU's GDPR. The African Network of Personal Data Protection

Authorities (RAPDP), a cooperation of 8 African countries seeks to draft data protection laws,

formulate opinions on specific issues and establish a consultative framework on data protection.

There are also regional conventions for data protection and privacy that have been enacted such as

the Southern African Development Community (SADC) Model Law[26] for the south and the Act on

Personal Data Protection within ECOWAS (Economic Community of West African States) for the

west[27]. Unlike their southern and western counterparts, east African nations have not adopted

similar regional frameworks, which can set cross-border and data portability limitations. In addition,

some countries national laws fall short in setting the required safeguards for data privacy breach and

data portability and lack bodies such as Data Protection Authorities to enforce these legislations[24].

This is just the African scenario, but every country and region will have their own set of policies.

Therefore, the healthcare programme and data controllers have an important responsibility to

manage the programme in compliance with local laws or, where lacking, consult with government

authorities on which regulation to abide by to foremost protect patient rights. Regulation of CDSS is

also very contentious topic. Groups will have to watch this space, FDA and European Commission

regulations, closely to determine whether their tool fits the definition of a Software as a Medical

Device (SaMD) and what data they will need to generate for regulatory approval.

In terms of "Validation", the big question was how to ensure CDSAs are safe for use. Developers and

implementers should aim for clinical association, analytical validation and clinical validation.

However clinical validation can require significant amount of financial resources and time. As an

example, e-POCT was developed in 2014 to incorporate the latest scientific evidence to expand the

medical content of IMCI. Its safety and efficacy was determined in a RCT in Tanzanian outpatient

clinics[17]. However, in 2019 the tool has yet been implemented at scale or evaluated in non-

controlled settings. Efforts like e-POCT to optimise CDSA content and assess its impact on patient

health outcomes are needed to grow the evidence for CDSA's role in improving patient safety. But

providing these tools to patient also demands that we put our efforts in studies to measure the

effectiveness of CDSA while in the hands of the end-user and in the setting of intended use. Indeed,

there may be factors when CDSAs are implemented as part of routine practice that affect the overall

effectiveness and impact of the tool. This includes a lack of or poor health worker training, lack of

connectivity, software malfunctions, poor design leading to low usability and acceptability of the

tool by health workers. Cycles of evaluation and improvement can be sought to perfect these tools

in intended settings.

In addition, digital data obtained from CDSA validation can be used in turn to dynamically improve

the tool itself faster than development improvements in classic wet diagnostics. CDSA also have the

potential to serve as platforms to support the evaluation or development of new content or

guidelines. These can be faster adapted to geographies to conform to local policies.

The call for the development of PHC during the declaration of Astana revived the global commitment

to provide quality health services to all[28]. Achieving universal health coverage will therefore

require that digital health interventions are quality assured and evidence-based. Indeed, the WHO

has begun developing frameworks to create a common digital language and synthetizing the

evidence around emerging mobile-based digital health technologies through two recent

publications[29, 88]. These resources are meant to guide the global health community in assessing

digital interventions that will improve quality of care, meaning interventions that are safe, effective,

timely, efficient, equitable and people-centred, and this include CDSAs[30].

The management of patients in low-resource settings with complex epidemiology is extremely

challenging without clinical algorithms and accurate and portable rapid diagnostic tests.

Consequences of suboptimal quality of care can go even beyond the patient's health to include

public health considerations such as the development of antimicrobial resistance due to the over-

prescribing of medicines. The aim of the proposed TPP is to support developers and implementers of

these toolkits aimed at guiding health workers throughout clinical assessment, to give them

confidence in clinical decisions and actions, whether it is sending a patient to a hospital or not,

prescribing or not prescribing antibiotics or recommending rehydration at home.

ACKNOWLEDGEMENTS

We would like to thank all the experts who participated in the FIND/WHO workshop to discuss the

first draft of the TPP, as well as Rigveda Kadam, Wallace White and Mo Tobin (FIND) for discussions

on TPP development. We would also like to thank the online reviewers who provided feedback to

develop the final consensus-driven TPP characteristics.

FUNDING

This work was funded by the Fondation Botnar and supported by the Global Antimicrobial Resistance

Innovation Fund (GAMRIF), a UK aid programme. The funders had no role in the study design, data

collection and analysis, or preparation of the manuscript.

Tunçalp Ӧ, Were WM, MacLennan C, et al. Quality of care for pregnant women and newborns—the WHO vision. BJOG 122(8):1045-9. doi: 10.1111/1471-0528.13451. Epub 2015 May 1. http://onlinelibrary.wiley.com/doi/10.1111/1471-0528.13451/full

World Health Organisation. Integrated Management of Childhood Illness (WHO) 2014 [May 2019]. Retrieved from http://apps. who.int/iris/bitstream/10665/104772/16/9789241506823_Chartbook_eng.pdf.

World Health Organisation. Integrated management of childhood illness: caring for newborns and children in the community. 2011. World Health Organization. [May 2019]. Retrieved from http://www.who.int/iris/handle/10665/44398

World Health Organisation. Integrated Management of Adolescent and Adult Illness (IMAI). [May 2019]. Retrieved from http://www.who.int/3by5/publications/documents/imai/en/.

WHO and TDR. Dengue: Guidelines for diagnosis, treatment, prevention and control: new edition. 2009 (WHO/HTM/NTD/DEN/2009.1). [May 2019]. Retrieved from https://www.who.int/tdr/publications/training-guideline-publications/dengue-diagnosis-treatment/en/

Klaus Schwab (2015). The Fourth Industrial Revolution: what it means, how to respond. [May 2019]. Retrieved from https://www.foreignaffairs.com/articles/2015-12-12/fourth-industrial-revolution.

Goldemberg José. Leapfrog energy technologies. Energy Policy1998; 26(10):729–741.

World Health Organisation guideline: recommendations on digital interventions for health system strengthening. Executive summary. Geneva: World Health Organization; 2019. (WHO/RHR/19.8). Licence: CC BY-NC-SA 3.0 IGO.

Darling G. The impact of clinical practice guideline and clinical trial on treatment decisions. Surg Oncol. 2002 Dec; 11.

Armstrong Schellenberg J, Bryce J, de Savigny D, et al. The effect of Integrated Management of Childhood Illness on observed quality of care of under-fives in rural Tanzania. Health Policy Plan. 2004;19

Rowe AK, Labadie G, Jackson D, et al. Improving health worker performance: an ongoing challenge for meeting the sustainable development goals. BMJ 2018;362:k2813.

Mitchell M, Hedt-Gauthier BL, Msellemu D, et al. Using electronic technology to improve clinical care - results from a before-after cluster trial to evaluate assessment and classification of sick

children according to Integrated Management of Childhood Illness (IMCI) protocol in Tanzania. BMC Med Inform Decis Mak. 2013;13:95. Published 2013 Aug 27. doi:10.1186/1472-6947-13-95

Bernasconi A, Crabbé F, Raab M, et al. Can the use of digital algorithms improve quality care? An example from Afghanistan. PLOS ONE 13(11): e0207233.https://doi.org/10.1371/journal.pone.0207233

Sarrassat S, Blanchet K, Lewis J et al. Integrated eDiagnosis Approach (IeDA) for the management of illness in under-five children at the primary health care level in Burkina Faso: Findings from a stepped-wedge cluster randomised trial.

World Health Organization (2010). Guidelines for the Treatment of Malaria. 2nd edition. Geneva, Switzerland: World Health Organization.

Shao AF, Rambaud-Althaus C, Samaka J, et al. New Algorithm for Managing Childhood Illness Using Mobile Technology (ALMANACH): A Controlled Non-Inferiority Study on Clinical Outcome and Antibiotic Use in Tanzania. PLoS One. 2015 Jul 10;10(7):e0132316. doi: 10.1371/journal.pone.0132316.

Keitel K, Kagoro F, Samaka J, et al. A novel electronic algorithm using host biomarker point-of-care tests for the management of febrile illnesses in Tanzanian children (e-POCT): A randomized, controlled non-inferiority trial. PLoS Med. 2017 Oct 23;14(10):e1002411. doi: 10.1371/journal.pmed.1002411. eCollection 2017 Oct.

FIND and WHO. Electronic clinical decision support algorithms_ Workshop report.2018. Available from: https://www.finddx.org/mal-fev/ecdm-toolbox/

U.S. Food & Drug Administration. "Software as a Medical Device (SAMD): Clinical Evaluation. Guidance for Industry and Food and Drug Administration Staff". December 8, 2017. [October 2018]. Retrieved from https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM524904.pdf

European Commission. Guidance document Medical Devices – Scope, field of application, definition – Qualification and Classification of standalone software – MEDDEV 2.1/6. [May 2019]. Retrieved from http://ec.europa.eu/DocsRoom/documents/17921/attachments/1/translations

European Commission. (2017) Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices, amending Directive 2001/83/EC, Regulation (EC) No 178/2002 and Regulation (EC) No 1223/2009 and repealing Council Directives 90/385/EEC and 93/42/EEC (Text with EEA relevance)Text with EEA relevance. [July 2019] Retrieved from http://data.europa.eu/eli/reg/2017/745/2017-05-05

HL7 International. Retrieved from https://www.hl7.org/

General Data Protection Regulation (GDPR). [May 2019]. Retrieved from https://gdpr-info.eu/

Cynthia Rich. A look at New Trends in 2017: Privacy laws in Africa and the Near East (16) 6 Blomberg BNA. [May 2019]. Retrieved from https://media2.mofo.com/documents/170911-privacy-africa.pdf

UNCTAD. Data Protection and Privacy Legislation Worldwide. [May 2019]. Retrieved from https://unctad.org/en/Pages/DTL/STI_and_ICTs/ICT4D-Legislation/eCom-Data-Protection-Laws.aspx. Accessed May 14 2019.

ITU. (2013) Establishment of harmonized policies for the ICT market in the ACP Countries. Data protection: Southern African Development Community (SADC) Model Law. [May 2019]. Retrieved from https://www.itu.int/en/ITU-D/Projects/ITU-EC-ACP/HIPSSA/Documents/FINAL%20DOCUMENTS/FINAL%20DOCS%20ENGLISH/sadc_model_law_data_protection.pdf

ECOWAS (2010). [May 2019]. Supplementary Act A/SA.1/01/10 on Personal Data Protection within ECOWAS. https://www.statewatch.org/news/2013/mar/ecowas-dp-act.pdf

The Lancet. The Astana Declaration: the future of primary health care? Lancet 2018;392(10156):1369. doi.org/10.1016/S0140-6736(18)32478-

Classification of digital health interventions. Geneva: World Health Organization; 2018(WHO/RHR/18.06). Licence: CC BY-NC-SA 3.0 IGO

World Health Organisation. Quality of Care: A Process for Making Strategic Choices in Health Systems. Geneva: WHO, 2006. [July 2019]. Retrieved from https://apps.who.int/iris/handle/10665/43470

Figure 1. Target product profile second and final survey results

Section	Characteristic	Agreement Min	Agreement Opt	Agreement with Min	Agreement with Opt
Scope General	Intended use	96%	96%	96%	96%
	Target Population	96%	96%	96%	96%
	Setting (level of implementation in the hea	96%	96%	96%	96%
	Targeted End User	100%	100%	100%	100%
	Algorithm Access	96%	96%	96%	96%
Scope Toolkit	Algorithm Content	96%	96%	96%	96%
	Algorithm Treatment Recommendations	96%	91%	96%	91%
	Compatible Point of Care Tools	91%	91%	91%	91%
	Regulated Toolkit Components	91%	87%	91%	87%
	Compatible Devices	83%	87%	83%	87%
	Compatible Operating Systems	87%	87%	87%	87%
Electronic CDS	Content Transparency	87%	91%	87%	91%
	Quality control	91%	91%	91%	91%
	Algorithm Validation	91%	91%	91%	91%
	Machine Learning	96%	87%	96%	87%
Point of Care Tools	POC Data Inputs	87%	87%	87%	87%
	Disease Likelihood	83%	87%	83%	87%
	POC Training	87%	87%	87%	87%
App	System Validation	83%	83%	83%	83%
	System Access (public API)	91%	96%	91%	96%
	Context Configuration	87%	91%	87%	91%
	Customization	100%	100%	100%	100%
	User Access Rights	100%	96%	100%	96%
	Expert Support	87%	87%	87%	87%
	App Training	91%	91%	91%	91%
	Internet Availability	96%	96%	96%	96%
	Clinical Data Entry	87%	96%	87%	96%
	Patient Management Recommendation	96%	100%	96%	100%
	Navigation	96%	83%	96%	83%
	Workflow	83%	83%	83%	83%
	Task Management	87%	91%	87%	91%
	Follow up	100%	96%	100%	96%
	System Malfunction Protection	96%	96%	96%	96%
	Scalability	91%	91%	91%	91%
	Updates and Versioning	100%	100%	100%	100%
Data	Data Capture	83%	91%	83%	91%
	Data Validation	96%	96%	96%	96%
	Data Ownership	96%	96%	96%	96%
	Data Storage	91%	91%	91%	91%
	Data Recovery	100%	100%	100%	100%
	Data Flow	87%	87%	87%	87%
	Data Reporting	100%	100%	100%	100%
	Data Provenance	96%	96%	96%	96%
	Data Dictionary	91%	96%	91%	96%
	Data Security & Privacy	87%	91%	87%	91%
Procurement	Procurement Models	91%	87%	91%	87%
	Average	91%	91%	91%	91%

Legend:
≥50% ≥75%

- Fully agree
- Mostly agree
- Neither agree nor disagree
- Mostly disagree
- Fully disagree