A deep learning algorithm using CT images to screen for Corona Virus

Disease (COVID-19)

Shuai Wang[1,*], Bo Kang[2,3,*], Jinlu Ma[4,*], Xianjun Zeng[5,*], Mingming Xiao[1,*], Jia

Guo[3], Mengjiao Cai[4], Jingyi Yang[4], Yaodong Li[6], Xiangfei Meng[2,#], Bo Xu[1,#]

[1] Department of Molecular Radiation Oncology, National Clinical Research Center

for Cancer, Key Laboratory of Cancer Prevention and Therapy, Key Laboratory of

Breast Cancer Prevention and Therapy, Ministry of Education, Tianjin Clinical

Research Center for Cancer, Tianjin Medical University Cancer Institute and

Hospital, Tianjin 300060, China

[2] College of Intelligence and Computing, Tianjin University, Tianjin 300350, China

[3] National Supercomputer Center in Tianjin, Tianjin 300457, China

[4] Department of Radiation Oncology, First Affiliated Hospital, Xi'an Jiaotong

University, Xi'an, China

[5] Department of Radiology, Nanchang University First Hospital, Nanchang, China

[6] Department of Radiology, No.8 Hospital, Xi'an Medical College, Xi'an, China

* Equal contribution

Corresponding authors:

Bo Xu, MD, PhD
Tianjin Medical University Cancer Institute and Hospital
Tianjin 300060, China
Email: xubo@tmu.edu.cn

Xiangfei Meng, PhD
National Supercomputer Center in Tianjin
Tianjin 300457, China
Email: mengxf@nscc-tj.cn

Abstract

Background: To control the spread of Corona Virus Disease (COVID-19),

screening large numbers of suspected cases for appropriate quarantine and

treatment is a priority. Pathogenic laboratory testing is the diagnostic gold standard

but it is time consuming with significant false negative results. Fast and accurate

diagnostic methods are urgently needed to combat the disease. Based on

COVID-19 radiographical changes in CT images, we aimed to develop a deep

learning method that could extract COVID-19's graphical features in order to provide

a clinical diagnosis ahead of the pathogenic test, thus saving critical time for

disease control.

Methods and Findings: We collected 1,065CT images of pathogen-confirmed

COVID-19 cases (325 images) along with those previously diagnosed with typical

viral pneumonia (740 images). We modified the Inception transfer-learning model

to establish the algorithm, followed by internal and external validation. The internal

validation achieved a total accuracy of 89.5% with specificity of 0.88 and sensitivity

of 0.87. The external testing dataset showed a total accuracy of 79.3% with

specificity of 0.83 and sensitivity of 0.67. In addition, in 54 COVID-19 images that

first two nucleic acid test results were negative, 46 were predicted as COVID-19

positive by the algorithm, with the accuracy of 85.2%.

Conclusion: These results demonstrate the proof-of-principle for using

artificial intelligence to extract radiological features for timely and accurate

COVID-19 diagnosis.

Keywords: COVID-19, Computed Tomography, Artificial Intelligence, Deep

Learning, Diagnosis

Short Title: AI to screen for COVID-19

Introduction

The outbreak of atypical and person-to-person transmissible pneumonia caused by the severe acute respiratory syndrome coronavirus 2 (SARS-COV-2, also known as 2019-nCov) has caused a global alarm. There have been more than 500,000 confirmed cases of the Corona Virus Disease (COVID-19) in the world, as of March 26, 2020. According to the WHO, 16-21% of people with the virus in China have become severely ill with a 2-3% mortality rate. With the most recent estimated viral reproduction number (R0), the average number of other people that an infected individual will transmit the virus to in a completely non-immune population, stands at about 3.77 [1] , indicating that a rapid spread of the disease is imminent. It is crucial to identify infected individuals as early as possible for quarantine and treatment procedures.

The diagnosis of COVID-19 relies on the following criteria: clinical symptoms, epidemiological history and positive CT images, as well as positive pathogenic testing. The clinical characteristics of COVID-19 include respiratory symptoms, fever, cough, dyspna, and pneumonia [3-6]. However, these symptoms are nonspecific, as there are isolated cases where, for example, in an asymptomatic infected family a chest CT scan revealed pneumonia and the pathogenic test for the virus came back positive. Once someone is identified as a PUI (person under investigation), lower respiratory specimens, such as bronchoalveolar lavage, tracheal aspirate or sputum, will be collected for pathogenic testing. This laboratory technology is based on real-time RT-PCR

and sequencing of nucleic acid from the virus [7,8]. Since the beginning of the

outbreak, the efficiency of nucleic acid testing has been dependent on several

rate-limiting factors, including availability and quantity of the testing kits in the

affected area. More importantly, the quality, stability and reproducibility of the

detection kits are questionable. The impact of methodology, disease stages,

specimen collection methods, nucleic acid extraction methods, and the

amplification system are all determinant factors for the accuracy of test results.

Conservative estimates of the detection rate of nucleic acid are low (between

30-50%) [7,8,9], and tests need to be repeated several times in many cases

before they can be confirmed.

Radiological imaging is also a major diagnostic tool for COVID-19. The

majority of COVID-19 cases have similar features on CT images including

ground-glass opacities in the early stage and pulmonary consolidation in the

late stage. There is also sometimes a rounded morphology and a peripheral

lung distribution [6,10]. Although typical CT images may help early screening

of suspected cases, the images of various viral pneumonias are similar and

they overlap with other infectious and inflammatory lung diseases. Therefore, it

is difficult for radiologists to distinguish COVID-19 from other viral pneumonias.

Artificial Intelligence involving medical imaging deep-learning systems has

been developed in image feature extraction, including shape and spatial

relation features. Specifically, Convolutional Neural Network (CNN) has been

proven in feature extraction and learning. CNN was used to enhance low-light

images from high-speed video endoscopy with the limited training data being

just 55 videos [11]. Also, CNN has been applied to identify the nature of

pulmonary nodules via CT images, the diagnosis of pediatric pneumonia via

chest X-ray images, automated precising and labeling of polyps during

colonoscopic videos, cystoscopic image recognition extraction from videos

[12-15].

There are a number of features for identifying viral pathogens on the basis

of imaging patterns, which are associated with their specific pathogenesis [16].

The hallmarks of COVID-19 are bilateral distribution of patchy shadows and

ground glass opacity in early stages。 As the disease progresses, multiple

ground glass and infiltrates in both lungs will appear [3]. Theses features are

quite similar to typical viral pneumonia with only slight differences, which are

difficult to be distinguished by radiologists. Based on this, we believed that

CNN might help us identify unique features that might be difficult for visual

recognition.

Hence, the purpose of our study was to evaluate the diagnostic

performance of a deep learning algorithm using CT images to screen for

COVID-19 during the influenza season. To test this notion, we retrospectively

enrolled 1,065 CT images of pathogen-confirmed COVID-19 cases along with

previously diagnosed typical viral pneumonia. Our results reported below

demonstrate the proof-of-principle using the deep learning method to extract

radiological graphical features for COVID-19 diagnosis.

Methods and Materials

Retrospective collection of datasets.

We retrospectively collected CT images from 259 patients, in which the cohort includes 180 cases of typical viral pneumonia and the other 79 cases from three hospitals with confirmed nucleic acid testing of SARS-COV-2. In addition, we enrolled additional 15 COVID cases, in which first two nucleic acid tests were negative at initial diagnoses. Hospitals providing the images were Xi'an Jiaotong University First Affiliated Hospital (Center 1), Nanchang University First Hospital (Center 2) and Xi'an No.8 Hospital of Xi'an Medical College (Center 3). All CT images were de-identified before sending for analysis. This study is in compliance with the Institutional Review Board of each participating institutes. Informed consent was exempted by the IRB because of the retrospective nature of this study.

Delineation of ROIs

To establish a binary model for distinguishing COVID-19 and typical pneumonia, we drew the Region Of Interest (ROI) as input images for the training cohort and validation cohorts. We sketched the ROI from CT images based on features of COVID-19, such as small patchy shadows and interstitial changes in the early stage, multiple ground glass and infiltrates in both lungs in the progression stage, and delineated the ROIs on the CT images of other typical viral pneumonia such as pseudocavity, enlarged lymphnodes and multifocal GGO as the control. The ROIs were divided into three cohorts: one

training cohort (n=320 from Center 1), one internal validation cohort (n=455

from Center 1) and one external validation cohort (n=290 from Center 2 and 3).

For a ROI, it is sized approximately from 395*223 to 636*533 pixels.

Overview of the proposed architecture

Our systematic pipeline for the prediction architecture is depicted in **Figure**

1. The architecture consists of three main processes: 1) Pre-processing of

input images; 2) Feature extraction of ROI images and training; and 3)

Classification with fully connected network and prediction of multiple classifiers.

We built a transfer learning neural network based on the Inception network.

The entire neural network can be roughly divided into two parts: the first part

used a pre-trained inception network to convert image data into

one-dimensional feature vectors, and the second part used a fully connected

network and the main role is for classification prediction. ROI images from

each case were preprocessed and inputted into the model for training. The

number of various types of pictures in the training set is equal, with a total

number of 320. The remaining CT pictures of each case were used for internal

validation. The model training was iterated 15,000 times with a step size of

0.01.

Figure 1. ROI images extraction and Deep Learning (DL) algorithm

framework. ROI images were extracted by the CV model and then trained

using a modified Inception network to extract features. The full connection

layer then makes a classification and prediction.

Image preprocessing and feature extraction.

Based on the signs of characteristic of pneumonia, ROI images were

defined inflammatory lesions and extracted by our computer vision (CV) model

following the steps. 1) Convert the image to grayscale. 2) Binarize grayscale.

Because using the OSTU's method directly may cause the threshold selection

failure in the case of multi-peaks, the selection of the binarization threshold in

this paper was based on the statistics of all pixel frequency histograms of the

gray color values Vmin (80) and Vmax (200). The minimum frequency in the

selection interval is threshold, and the interval of frequency statistics is five. 3)

Background area filling. Using the flood filling method to expand the image by

1 black pixel, and fill the black pixels near the border with white. 4) Reverse

color, find all the contour areas of the image, and keep the two largest contour

areas as the two lung areas. 5) Take the smallest bounding rectangle of the

lung area as the ROI frame and crop the original image to obtain the ROI

images. The delineated ROIs were obtained for classification model building.

We modified the typical Inception network, and fine-tuned the modified

Inception (M-Inception) model with pre-trained weights. During the training

phase, the original Inception part was not trained, and we only trained the

modified part. The architecture of M-Inception is shown in **Table 1**. The

difference between Inception and M-Inception in classification lies in the last

fully-connected layers. We reduced the dimension of features before it was

sent to the final classification layer. The training dataset made up of all those

patches aforementioned. The Inception network is shown in **Table 1**.

Table 1. The architecture of M-Inception

	Layer	Patch size/stride or remarks
Inception part	conv	3×3/2
	conv	3×3/1
	conv padded	3×3/1
	pool	3×3/2
	onv	3×3/1
	conv	3×3/2
	conv	3×3/1
	Inception	3x, 5x, 2x
	pool	8x8
	linear	logits

	softmax	classifier
Modified part	Fc1	$\begin{bmatrix} \text{batchnorm} \\ \text{dropout(0.5)} \\ \text{512d Linear} \end{bmatrix}$
	Fc2	$\begin{bmatrix} \text{batchnorm} \\ \text{dropout(0.5)} \\ \text{2d Linear} \end{bmatrix}$

Prediction.

After generating the features, the final step was to classify the pneumonia

based on those features. Ensembling of classifiers was used to improve the

classification accuracy. In this study, we adopted end-to-end learning to make

the model convergence.

Performance evaluation metrics.

We compared the classification performance using Accuracy, Sensitivity,

Specificity, Area Under Curve (AUC), Positive predictive value (PPV), Negative

predictive value (NPV), F1 score and Youden Index. TP and TN represent the

number of true positive or true negative samples. FP and FN mean the number

of false positive or false negative samples. Sensitivity measures the ratio of

positives that are correctly discriminated. Specificity measures the ratio of

negatives that are correctly discriminated. AUC is an index to measure the

performance of the classifier. NPV was used to evaluate the algorithm for

screening, and PPV was the probability of getting a disease when the

diagnostic index is positive. Youden Index was the determining exponent of the

optimal bound. F1 score was a measure of the accuracy of a binary model.

208 Additionally, performance was evaluated with F-measure (F1) to compare the

209 similarity and diversity of performance. Kappa value measures the agreement

210 between the CNN model prediction and the clinical report.

 Results

Algorithm development.

In order to develop a deep learning algorithm for the identification of viral

pneumonia images, we initially enrolled 259 patients, in which the cohort

includes 180 cases of typical viral pneumonia that were diagnosed previously

before the COVID-19 outbreak. These patients are termed COVID-19 negative

in the cohort. The other 79 cases were from the three hospitals with confirmed

nucleic acid testing of SARS-COV-2, therefore termed COVID-19 positive. Two

radiologists were asked to review the images and sketched 1,065

representative images (740 for COVID-19 negative and 325 for COVID-19

positive) for analysis (**Figure 2** is shown as an example). These images were

randomly divided into a training set and a validation set. The model training

was iterated for 15,000 times with a step size of 0.01. The training loss curve is

shown in **Figure 3A**. 320 images (160 images from COVID-19 negative and

160 images from COVID-19 positive) were obtained to construct the model. To

test the stability and generalization of the model,455 images (COVID-19

ngegative 360 images and COVID-19 positive 95 images) were obtained for

internal validation from Center 1 and 290 images (COVID-19 negative 220

images and COVID-19 positive 70 images) were obtained from Center 2 and 3

for external validation. The model training was also iterated for 15,000 times

with a step size of 0.01. The training loss curve is shown in **Figure 3B**.

Figure 2. An example of COVID-19 pneumonia features. The blue arrow

points to ground-glass opacity, and the yellow arrow points to the pleural

indentation sign.

Figure 3. Training loss curves of the models on internal (**A**) and external (**B**)

validation. The loss curve tends to be stable after descending, indicating that

the training process converges

Deep learning performance.

The deep learning algorithm yielded an AUC of 0.93 (95% CI, 0.90 to 0.96)

on the internal validation and 0.81 (95% CI, 0.71 to 0.84) on the external

validation based on the number of CT images (**Figure 4)**. Using the maximized

Youden index threshold probability, the sensitivity was 0.88 and 0.83,

specificity 0.87 and 0.67, the accuracy 89.5% and 79.3%, the negative

prediction values 0.95 and 0.90, the Youden indexes 0.75 and 0.48, and the F1

scores were 0.77 and 0.63 for the internal and external datasets, respectively

(**Table 2**). The kappa values were 0.69 and 0.48 for internal and external

validation, indicating that prediction of COVID-19 from the CNN model is a

highly consistent with pathogenic testing results. We also performed an

external validation based on each patient's multiple images. The accuracy was

82.5%, with the sensitivity 0.75, the specificity 0.86, the PPV 0.69, the NPV

0.89, and the kappa value 0.59.

Figure 4. Receiver operating characteristic plots for COVID-19 identification

for the deep learning (Inception) algorithm. A) Internal Validation. B) External

Validation.

261 **Table 2. Deep learning Algorithm Performance**

Performance Metric	Internal	External
AUC（95%CI）	0.93(0.86 to 0.94)	0.81(0.71 to 0.84)
Accuracy, %	89.5	79.3
Sensitivity	0.88	0.83
Specificity	0.87	0.67
PPV	0.71	0.55
NPV	0.95	0.90
Kappa*	0.69	0.48
Yoden index	0.75	0.50
F1 score‡	0.77	0.63

* Measures the agreement between the CNN model prediction and the clinical

diagnosis. ‡Measures the accuracy of the CNN model.

Comparison of AI with radiologist prediction.

At the same time, we asked two skilled radiologists to assess the 745

images for a prediction. Radiologist 1 achieved the accuracy of 55.8% with

sensitivity of 0.71 and specificity of 0.51, and Radiologist 2 achieved a similar

accuracy of 55.4% with sensitivity of 0.73 and specificity of 0.50 (**Table 3**).

These results indicates that it is difficult for radiologists to make prediction of

COVID-19 with eye recognition, further showing the advantage of the

algorithm we developed.

Table 3. Performance metrics for the CNN model versus skilled radiologists.

Performance Metric	Internal	External (Based on ROI)	External (Based on patients)	R1	R2
Accuracy, %	89.5	79.3	82.5	55.8	55.4
Sensitivity	0.88	0.83	0.75	0.71	0.73
Specificity	0.87	0.67	0.86	0.51	0.5
PPV	0.71	0.55	0.69	0.29	0.29
NPV	0.95	0.90	0.89	0.86	0.86
F1 score	0.77	0.63	0.72	0.41	0.42
Kappa	0.69	0.48	0.59	0.15	0.15
Yoden index	0.75	0.50	0.61	0.22	0.23

Prediction of COVID-19 on CT images from pathogenic negative patients.

Because high false negative results were frequently reported from nucleic acid testing, we aimed to test whether the algorithm could detect COVID-19 when the pathogenic test came negative. To achieve this goal, we enrolled additional 15 COVID-19 cases, in which initial two nucleic acid tests came negative and for the third test they became positive. These CT results were taken on the same day of the nucleic acid tests **(Figure 5)**. Interestingly, we found that, 46 out of the 54 images when nucleic acid test results were negative were predicted as COVID-19 positive by the algorithm, with the accuracy of 85.2%. These results indicate that the algorithm has high value serving as a screening method for COVID-19.

Figure 5. Representative images from a COVID-19 patient with two negatively

reported nucleic acid tests at earlier stages and one final positively reported

test at a later stage. On the left, only one inflammatory lesion (blue arrow) can

be seen near diaphragm. In the middle, lesions (yellow arrows) were found in

two levels of images. On the right, the images were taken on the ninth day

after admission. The inflammation continued to progress, extending to both

lungs (red arrows), and the nucleic acid test became positive.

Discussion

Timely diagnosis and triaging of PUIs are crucial for the control of

emerging infectious diseases such as the current COVID-19. Due to the

limitation of nucleic acid -based laboratory testing, there is an urgent need to

look for fast alternative methods that can be used by front-line health care

personals for quickly and accurately diagnosing the disease {Rao, 2020 #67}.

In the present study, we have developed an AI program by analyzing

representative CT images using a deep learning method. This is a

retrospective, multicohort, diagnostic study using our modified Inception

migration neuro network, which has achieved an overall 89.5% accuracy.

Moreover, the high performance of the deep learning model we developed in

this study was tested using external samples with 79.3% accuracy. More

importantly, as a screening method, our model achieved a relative high

sensitivity, 0.88 and 0.83 on internal and external datasets, respectively.

Furthermore, the model achieved a performance on the patient level, with the

accuracy of 82.5%. During current COVID-19 global pandemics, this CNN

model can potentially serve as a powerful tool for COVID-19 screening.

It is important to note that our model aims to distinguish between

COVID-19 and other typical viral pneumonia, both of which have similar

radiologic characteristics. We compared the performance of our model with

that of two skilled radiologists, and our model has shown much higher

accuracy and sensitivity. These findings have demonstrated the

proof-of-principle that deep learning can extract CT image features of

COVID-19 for diagnostic purposes. Using the supercomputer system, the time

for each case takes only about 10 seconds, and it can be performed remotely

via a shared public platform. Therefore, further developing this system can

significantly shorten the diagnosis time for disease screening, especially at the

time when many places have nucleic acid test shortage. Our study represents

the first study to apply artificial intelligence technologies to CT images for

effectively screening for COVID-19.

The gold standard for COVID-19 diagnosis has been nucleic acid based

detection for the existence of specific sequences of the SARS-COV-2 gene.

While we still value the importance of nucleic acid detection in the diagnosis of

SARS-COV-2 infection, we must also note that the high number of false

negatives due to several factors such as methodological disadvantages,

disease stages, and methods for specimen collection might delay diagnosis

and disease control. Recent data have suggested that the accuracy of nucleic

acid testing is only about 30-50% [6,7,8]. Using CT imaging feature extraction,

we are able to achieve above 89.5% accuracy, significantly outplaying nucleic

acid testing. More interestingly, testing CT images from COVID-19 patients

when initial pathogenic testing came negative, our model has achieved the

accuracy of 85.2% for correctly predicting COVID-19. According to a study

authored by Xia L et al, 75% patients with negative RT-PCR results had

positive CT findings [17]. The study also recommended that chest CT as a

primary tool for the current COVID-19 detection in affected areas.

Deep learning methods have been used to solve data-rich biology and

medicine. A large number of labeled data is required for training [18]. Although

we are satisfied with the initial results, we believe that with more CT images

included in the training, we will achieve higher accuracy. Therefore, further

optimizing and testing this system is warranted. To achieve this, we have

generated a webpage that licensed healthcare personnel can access to upload

CT images for testing and validation. The webpage information is as following:

https://ai.nscc-tj.cn/thai/deploy/public/pneumonia_ct.

There are some limitations to our study. Although DL has been used to

represent and learn predictable relationships in many diverse forms of data,

and it holds promise for applications in precision medicine, many factors such

as low signal to noise and complex data integration have challenged the DL

efficacy [19]. CT images represent a difficult classification task due to the

relatively large number of variable objects, specifically the imaged areas

outside the lungs that are irrelevant to the diagnosis of pneumonia [12]. In

addition, the training data set is relatively small. The performance of this

system is expected to increase when the training volume is increased. It

should also be noted that, the features of the CT images we analyzed were

from patients with severe lung lesions at later stages of disease development.

Although we have enrolled 15 cases of COVID patients for assessing the value

of the algorithm for early diagnosis, larger numbers of database to associate

this with the disease progress and all pathologic stages of COVID-19 is

necessary to optimize the diagnostic system.

In future, we intend to link hierarchical features of CT images to features of

other factors such as genetic, epidemiological and clinical information for

multi-modeling analysis for an enhanced diagnosis. The artificial intelligence

system developed in our study should significantly contribute to COVID-19

disease control by reducing the number of PUIs for timely quarantine and

treatment.

Ethics Committee Approval and Patient Consent: This study complies with

the Institutional Review Board of each participating institutes. Informed

consent was exempted by the IRB because of the retrospective nature of this

study.

Funding Source: None

Competing interests: The authors have declared that no competing interest

exists

Abbreviations:

COVID-19, Corona Virus Disease

CT, Computed Tomography

SARS-COV-2, severe acute respiratory syndrome coronavirus 2Convolutional

CNN, Neural Network

ROI, region of interest

M-Inception, modified Inception

AUC, Area Under Curve

PPV, Positive predictive value

NPV, Negative predictive value

CV, computer vision

394
395

Reference

1. Yang Y, Lu Q, Liu M, Wang Y, Zhang A, Jalali N, et al. Epidemiological and clinical features of the 2019 novel coronavirus outbreak in China. medRxiv [Preprint]. 2020 medRxiv [posted February 21], Available from: https://www.medrxiv.org/content/10.1101/2020.02.10.20021675v2 doi:10.1101/2020.02.10.20021675

2. Rao ASRS, Vazquez JA. Identification of COVID-19 Can be Quicker through Artificial Intelligence framework using a Mobile Phone-Based Survey in the Populations when Cities/Towns Are Under. Infect Control Hosp Epidemiol. 2020;1–18.

3. Wang D, Hu B, Hu C, Zhu F, Liu X, Zhang J, et al. Clinical Characteristics of 138 Hospitalized Patients With 2019 Novel Coronavirus-Infected Pneumonia in Wuhan, China. Jama. 2020

4. Chen N, Zhou M, Dong X, Qu J, Gong F, Han Y, et al. Epidemiological and clinical characteristics of 99 cases of 2019 novel coronavirus pneumonia in Wuhan, China: a descriptive study. Lancet. 2020.

5. Li Q, Guan X, Wu P, Wang X, Zhou L, Tong Y, et al. Early Transmission Dynamics in Wuhan, China, of Novel Coronavirus-Infected Pneumonia. The New England journal of medicine . 2020.

6. Huang C, Wang Y, Li X, Ren L, Zhao J, Hu Y, et al. Clinical features of patients infected with 2019 novel coronavirus in Wuhan, China. Lancet. 2020.

7. Corman VM, Landt O, Kaiser M, Molenkamp R, Meijer A, Chu DK, et al.

Detection of 2019 novel coronavirus (2019-nCoV) by real-time RT-PCR. Euro surveillance : bulletin Europeen sur les maladies transmissibles = European communicable disease bulletin. 2020; 25(3).

8. Chu DKW, Pan Y, Cheng SMS, Hui KPY, Krishnan P, Liu Y, et al. Molecular Diagnosis of a Novel Coronavirus (2019-nCoV) Causing an Outbreak of Pneumonia. Clinical chemistry. 2020.

9. Zhang N, Wang L, Deng X, Liang R, Su M, He C, et al. Recent advances in the detection of respiratory virus infection in humans. J Med Virol. 2020.

10. Chung M, Bernheim A, Mei X, Zhang N, Huang M, Zeng X, et al. CT Imaging Features of 2019 Novel Coronavirus (2019-nCoV). Radiology. 2020: 200230.

11. Gomez P, Semmler M, Schutzenberger A, Bohr C, Dollinger M. Low-light image enhancement of high-speed endoscopic videos using a convolutional neural network. Med Biol Eng Comput. 2019; 57(7): 1451-63.

12. Choe J, Lee SM, Do KH, Lee G, Lee JG, Lee SM, et al. Deep Learning-based Image Conversion of CT Reconstruction Kernels Improves Radiomics Reproducibility for Pulmonary Nodules or Masses. Radiology. 2019; 292(2): 365-73.

13. Kermany DS, Goldbaum M, Cai W, Valentim CCS, Liang H, Baxter SL, et al. Identifying Medical Diagnoses and Treatable Diseases by Image-Based Deep Learning. Cell 2018; 172(5): 1122-31.

14. Negassi M, Suarez-Ibarrola R, Hein S, Miernik A, Reiterer A. Application of

artificial neural networks for automated analysis of cystoscopic images: a review of the current status and future prospects. World J Urol. 2020.

15. Wang P, Xiao X, Glissen Brown JR, Berzin TM, Tu M, Xiong F, et al. Development and validation of a deep-learning algorithm for the detection of polyps during colonoscopy. Nat Biomed Eng. 2018; 2(10): 741-8.

16. Koo HJ, Lim S, Choe J, Choi SH, Sung H, Do KH. Radiographic and CT Features of Viral Pneumonia. Radiographics. 2018; 38(3): 719-39.

17. Ai T, Yang Z, Hou H, Zhan C, Chen C, Lv W, et al. Correlation of Chest CT and RT-PCR Testing in Coronavirus Disease 2019 (COVID-19) in China: A Report of 1014 Cases [published online ahead of print, 2020 Feb 26]. Radiology. 2020;200642.

18. Ching T, Himmelstein DS, Beaulieu-Jones BK, Kalinin AA, Do BT, Way GP, et al. Opportunities and obstacles for deep learning in biology and medicine. J R Soc Interface. 2018;15(141):20170387. doi:10.1098/rsif.2017.0387

19. Grapov D, Fahrmann J, Wanichthanarak K, Khoomrung S. Rise of Deep Learning for Genomic, Proteomic, and Metabolomic Data Integration in Precision Medicine. OMICS. 2018;22(10):630–636.