Auditory tests for characterizing hearing deficits in listeners with various hearing abilities: The BEAR test battery

Raul Sanchez-Lopez[1,5], **Silje Grini Nielsen**[1], **Mouhamad El-Haj-Ali**[2], **Federica Bianchi**[1,6], **Michal Fereczkowski**[1,2,3], **Oscar M Cañete**[1], **Mengfan Wu**[2,3], **Tobias Neher**[2,3], **Torsten Dau**[1†], **and Sébastien Santurette**[1,4†]

[1] *Hearing Systems Section, Department of Health Technology, Technical University of Denmark, Kgs. Lyngby, Denmark.*
[2] *Institute of Clinical Research, University of Southern Denmark, Odense, Denmark.*
[3] *Research Unit for ORL - Head Neck Surgery and Audiology, Odense University Hospital, Odense, Denmark; University of Southern Denmark, Odense, Denmark*
[4] *Centre for Applied Audiology Research, Oticon A/S, Smørum, Denmark*
[5] *Interacoustics Research Unit, Interacoustics A/S, Kgs. Lyngby, Denmark.*
[6] *Oticon Medical, Kongebakken 9, 2765 Smørum, Denmark*
[†] *These authors have contributed equally to this work and share last authorship*

Correspondence*:
Raul Sanchez-Lopez, Ørsteds Plads 352, 2800 Kgs. Lyngby, Denmark,rsalo@dtu.dk

Torsten Dau, Ørsteds Plads 352, 2800 Kgs. Lyngby, Denmark, tdau@dtu.dk

Sébastien Santurette, Kongebakken 9, 2765 Smørum, Denmark, sesn@oticon.com

ABSTRACT

The Better hEAring Rehabilitation (BEAR) project aims to provide a new clinical profiling tool
– a test battery – for hearing loss characterization. Whereas the loss of sensitivity can be
efficiently measured using pure-tone audiometry, the assessment of supra-threshold hearing
deficits remains a challenge. In contrast to the classical 'attenuation-distortion' model, the
proposed BEAR approach is based on the hypothesis that the hearing abilities of a given listener
can be characterized along two dimensions, reflecting independent types of perceptual deficits
(distortions). A data-driven approach provided evidence for the existence of different auditory
profiles with different degrees of distortions. Ten tests were included in a test battery, based on
their clinical feasibility, time efficiency and related evidence from the literature. The tests were
divided into six categories: audibility, speech perception, binaural processing abilities, loudness
perception, spectro-temporal modulation sensitivity and spectro-temporal resolution. Seventy-five
listeners with symmetric, mild-to-severe sensorineural hearing loss were selected from a clinical
population. The analysis of the results showed interrelations among outcomes related to high-
frequency processing and outcome measures related to low-frequency processing abilities. The
results showed the ability of the tests to reveal differences among individuals and their potential
use in clinical settings.

Keywords: audiology, hearing loss, loudness, binaural processing, speech perception, spectro-temporal resolution, aging

1 INTRODUCTION

21 In current clinical practice, hearing loss is diagnosed mainly on the basis of pure-tone audiometry (ISO
22 8253-1, 2010). The audiogram helps differentiate between conductive and sensorineural hearing losses and
23 can characterize the severity of the hearing loss from mild to profound. However, the pure-tone audiogram
24 only assesses the sensitivity to simple sounds, which is not necessarily related to listening abilities at
25 supra-threshold sound pressure levels (e.g. a person's ability to discriminate speech in noise)
26

Pure-tone audiometry is often complemented by speech audiometry (ISO 8253-3, 2012), which is a
test typically performed in the form of word recognition performance in quiet (Anderson et al., 2018).
Although this test can provide information about supra-threshold deficits (Gelfand, 2009), measurements
of speech understanding in noise have been found more informative (Killion et al., 2004; Nilsson et al.,
1994). Since improving speech intelligibility is usually the main goal of successful hearing rehabilitation,
several auditory factors affecting speech intelligibility in noise have been investigated (e.g. Glasberg
and Moore, 1989; Houtgast and Festen, 2008; Strelcyk and Dau, 2009). Audibility (in conditions with
fluctuating maskers), frequency selectivity (in conditions with stationary noise), and temporal processing
acuity (in conditions with speech interferers), have been identified as important factors affecting speech
reception thresholds in noise when using meaningful sentences as speech material (e.g. Desloge et al.,
2017; Johannesen et al., 2016; Oxenham and Simonson, 2009; Rhebergen et al., 2006)[2] Thus, a hearing
evaluation that goes beyond pure-tone sensitivity and speech intelligibility in quiet would be expected to
provide a more accurate characterization of a listener's hearing deficits.
40

In Denmark, the Better hEAring Rehabilitation (BEAR) project was initiated with the aim of developing
new diagnostic tests and hearing-aid compensation strategies for audiological practice. Although the
assessment of individual hearing deficits can be complex, new evidence suggests that the perceptual
consequences of a hearing loss can be characterized effectively by two types of hearing deficits, defined as
"auditory distortions" (Sanchez-Lopez et al., 2018). By analysing the outcomes of two previous studies
(Johannesen et al., 2016; Thorup et al., 2016) with a data-driven approach, Sanchez-Lopez et al. (2018)
identified high-frequency hearing loss as the main predictor of one of the distortions, whereas the definition
of the second type of distortion was inconclusive. The inconclusiveness in the prediction of the second
distortion was most likely due to differences between the two studies in terms of hearing loss profiles and
outcome measures. Here, a new dataset was therefore collected based on a heterogeneous group of listeners
with audiometric hearing losses ranging from very mild to severe and with a large range of audiometric
profiles. To that end, the most informative tests resulting from the analysis of Sanchez-Lopez et al. (2018)
were included, together with additional auditory tests that had shown potential for hearing profiling in other
previous studies. The tests included in the current study are referred to as the BEAR test battery.
55

56 The characterization of hearing deficits beyond the audiogram was considered in several earlier studies
57 (e.g., Lecluyse et al., 2013; Brungart et al., 2014; Rönnberg et al., 2016; Santurette and Dau, 2012; Saunders
58 et al., 1992; Vlaming et al., 2011). Among them, the HEARCOM project (Vlaming et al., 2011) proposed
59 an extended hearing profile formed by the results of several behavioural tests. These tests targeted various

---

[2] The factors identified correspond to the authors' conclusions based on cited references. For example, Johannesen et al. (2016) identified the basilar membrane compession as a predictor of speech intelligibility in stationary noise and temporal processing as a predictor of speech-in-speech intelligibility. Desloge et al. (2017); Oxenham and Simonson (2009); Rhebergen et al. (2006) identified the audibility of the soft speech sounds in the presence of fluctuating maskers as a crucial factor for speech intelligibility.

auditory domains, such as audibility, loudness perception, speech perception, binaural processing, and spectro-temporal resolution, as well as a test of cognitive abilities. Importantly, while the auditory domains considered in the BEAR test battery are similar to the ones considered in the HEARCOM project, the BEAR project aims to additionally classify the patients in subcategories and to create a link between hearing capacities and hearing-aid parameter settings.

The tests included in the BEAR test battery were chosen based on the following criteria: 1) There is evidence from the hearing research literature that the considered test is informative (i.e., it provides information about the individual hearing deficits) and reliable (i.e., the result of the test does not vary over time); 2) The outcomes of the test may be linked to a hearing-aid fitting strategy; 3) The outcome measures are easy to interpret and to explain to the patient; 4) The task is reasonably time-efficient or can be suitably modified to meet this requirement (e.g., by changing the test paradigm or developing an out-of-clinic solution); 5) The test implementation can be done with equipment available in clinics; 6) The tasks are not too demanding for patients and clinicians; 7) Tests with several outcome measures are prioritized, and 8) The language independent tests are also prioritized.

The selected test battery included measures of audibility, loudness perception, speech perception, binaural processing abilities, spectro-temporal modulation (STM) sensitivity and spectro-temporal resolution. It was implemented and tested in older listeners with different hearing abilities (from mild to severe hearing losses). The goals of the study were: 1) To collect reference data from a representative sample of HI listeners for each of the selected tests, 2) to analyse the test-retest reliability of these tests, 3) to analyse the relationships between the different outcome measures, and 4) to propose a version of the test battery that can be implemented in hearing clinics.

2 OVERVIEW OF THE TEST BATTERY

The test battery consisted of ten tests (9 tests besides the pure-tone audiometry). The outcomes of the proposed tests are divided into six categories. Table 1 shows the tests and the corresponding auditory domains or categories. By convenience, the domains spectro-temporal modulation sensitivity and spectro-temporal resolution are together in the category spectro-temporal processing. The following sections introduce the experimental methods and present all tests individually. The dataset is publicly available in a Zenodo repository (Sanchez-Lopez et al., 2019). More details about the method can be found in the supplementary material in the data repository.

2.1 Reference data from younger normal-hearing listeners

Although many of the tests included in the test battery are based on previous studies with normative data, a group of 11 young normal-hearing listeners were tested in the facilities of the Technical University of Denmark (DTU) and the University of Southern Denmark (SDU) to obtain reference data for this specific implementation of each of the tests. The summary statistics of the outcome variables from Table 1 are shown in Table 2.

2.2 Time efficiency of the test battery

The examiners kept track of the time used by each of the participants in completing the test battery. In the case of unexpected events (e.g., unexpected or incongruent results), these events were cautiously annotated for later investigation. Regarding the test procedure, additional repetitions of the threshold estimations were needed if: 1) a repetition was considered as an outlier if a given threshold was greater than three

Table 1. List of the tests included in the BEAR test battery and their corresponding auditory domains.

Test Name	Category	Variables
Pure-tone audiometry[1]	Audibility	AUD_{LF}, AUD_{HF}
Extended audiometry at high frequencies [2]		FLFT
Adaptive categorical loudness scaling[3]		HTL_{LF}, HTL_{HF}
	Loudness perception	MCL_{LF}, MCL_{HF} $DynR_{LF}$, $DynR_{HF}$ $Slope_{LF}$, $Slope_{HF}$
Word recognition scores[4]	Speech Perception	SRT_Q, maxDS
Hearing in noise test[5]		SRT_N, SS^{4dB}
Spectro-temporal modulation test[9]	Spectro-temporal processing	$sSTM_8$, $sSTM_{4k}$ $fSTM_8$, $fSTM_{4k}$
Extended audiometry in noise [10,11,12]		TiN_{LF}, TiN_{HF} SMR_{LF}, SMR_{HF} , TMR_{LF}, TMR_{HF}
Maximum frequency for IPD detection[6]	Binaural processing abilities	IPD_{fmax}
Binaural pitch[7]		BP_{20}
Extended binaural audiometry in noise[8]		BMR

LF: Lower-frequencies, HF: high-frequencies

AUDx: Pure-tone average at low (x=LF; f $\leq$ 1kHz) or high (x=HF; f > 1kHz) frequencies. // FLFT: Fixed-level frequency threshold at 80 dB SPL ACALOS outcome variabless are averaged for low (x=LF; f $\leq$ 1kHz) and high (x=HF; f > 1kHz) frequencies. // HTL: Hearing threshold level MCL: Most comfortable level / Slope: Slope of the loudness function / DynR: Dynamic range // SRTQ: Speech reception threshold in quiet / Max DS: Maximum speech discrimination score. // SRTN: Speech reception threshold in noise / Score +4: Sentence recognition score at +4 dB SNR // sSTM: Sensitivity for detecting a spectro-temporally modulated noise at 20log(m) = -3 dB, where m is the modulation depth / fSTM: Fast version of the STM test (Bernstein et al., 2016) // Extended audiometry outcome measures were measured at 0.5 kHz (x=LF) and at 2 kHz (x=HF) // eAUD-N: Tone detection in TEN noise // TMR: Temporal masking release // SMR: Spectral masking release // IPD fmax: Frequency threshold for detecting an interaural phase difference of 180°. // BP20: Binaural pitch detection scores for 20 presentations // BMR: Binaural masking release.

[1]ISO 8253-1 (2010); [2] Rieke et al. (2017); [3]Brand and Hohmann (2002), [4]ISO 8253-3 (2012) ; [5]Nielsen and Dau (2011); [6]Füllgrabe and Moore (2017); [7]Santurette and Dau (2012); [8]Durlach (1963); [9]Bernstein et al. (2016); [10]Moore et al. (2000) ; [11]Schorn and Zwicker (1990) ; [12]van Esch and Dreschler (2011)

scaled median absolute deviations of the two repetitions; or 2) the responses of the listeners during the
tracking procedure were inconsistent or reached the maximum or minimum possible values. In that case,
the measurement was considered an invalid or "missing" data point.

The timing annotated of the individual tests are shown in Figure 1. Besides, the probability of needing
an additional measurement and mean number of extra repetitions per listener are shown in Table 3.
The repetitions were only suggested when the test was done using the alternaltive-forced choice (AFC)
framework (i.e., the IPD test, the STM test and the eAUD test in all the conditions). The total testing time
was approximately 2'5 hours excluding the initial interview, information about the study and preparations.
110

Table 2. Reference data of the young normal-hearing group. The data is shown as the 5th, 25th, 50th, 75th and 95th percentiles. In tests performed monoaurally, the summary corresponds to the data of both ears merged together (i.e., 22 ears).

Outcome	5th	25th	50th	75th	95th	Unit
AUD_{LF}	-5	0	0	5	10	dB HL
AUD_{HF}	-10	0	5	10	10	dB HL
FLFT	14.42	15.60	17.88	18.02	19.31	kHz
HTL_{LF}	-7.5	-2.5	1.25	2.5	7.5	dB HL
HTL_{HF}	-5.5	2.5	8.75	15	23.5	dB HL
MCL_{LF}	55	62.5	70	72.5	78	dB HL
MCL_{HF}	62.5	70	75	80	87.5	dB HL
$DynR_{LF}$	87.5	92.5	97.5	105	117.5	dB
$DynR_{HF}$	74.5	85	92.5	100	105	dB
$SLope_{LF}$	0.27	0.30	0.33	0.37	0.42	CU/dB
$Slope_{HF}$	0.27	0.30	0.33	0.37	0.44	CU/dB
SRT_Q	6	10	14	18	21	dB SPL
maxDS	96.7	100	100	100	100	% Corr
SRT_N	-3.48	-1.84	-0.85	-0.19	1	dB SNR
SS_{+4dB}	80	90	95	100	100	% Corr
$sSTM_8$	2.40	3.07	3.07	3.07	3.07	d'
$sSTM_{4k}$	0.30	2.40	3.07	3.07	3.07	d'
$fSTM_8$	-13.5	-12.7	-12.3	-10.5	-6.1	dB ML
$fSTM_{4k}$	-9.1	-6.5	-4.375	-3.375	-2.7	dB ML
TiN_{LF}	64.25	66.5	68.25	69.5	73.6	dB HL
TiN_{HF}	64.1	68	69.375	70	73.85	dB HL
TMR_{LF}	3.55	7.4167	9.375	13	15.95	dB
TMR_{HF}	3.1	8.25	10.875	13	18.15	dB
SMR_{LF}	16.55	19.75	21.75	23.25	26.15	dB
SMR_{HF}	19.6	26	28.5833	30.75	31.25	dB
IPD	0.86	1.20	1.22	1.29	1.40	kHz
BP_{20}	50	98.75	100	100	100	% Corr
BMR	14.3	15.875	17	18.4375	22.1	dB

Table 3. Table with the probability of needing repetitions (PR), and the probability of having missing values (PM). The total probability of repetitions (PT). The mean number of extra repetitions (E. Rep).

Test	PR (%)	PM (%)	PT (%)	E.Rep.
STM	42.86	90.79	88.16	4.32
IPD	10.77	10.97	20.55	1.87
eAUD-HF	5.63	4.05	9.46	1.85
eAUD-N	66.67	46.58	82.19	3.00
eAUD-S	48.57	52.70	75.68	3.07
eAUD-T	53.85	46.58	75.34	3.27
S_0N_0	42.59	27.03	58.11	2.00
$S_\pi N_0$	20.59	9.11	27.03	1.85

3 GENERAL METHODS

3.1 Participants and general setup

111   Seventy-five listeners (38 of them females) participated in the study, who were aged between 59 and 82
112 years (median: 71 years). Five participants were considered older normal hearing (ONH) with thresholds

Figure 1. The overall time of the different tests in the test battery including the instructions. The data corresponds to the annotations of the examiners. The basic examination with the audiometry and the tympanometry (Tymp) are included. The numbers represent the rounded median in minutes.

below 25 dB Hearing Level (HL) in the frequency range between 0.25 and 4 kHz in both ears and no larger
than 40 dB HL at 8kHz (PTA $\leq$ 22 dB HL)[3]. PTA was defined as the pure-tone average between 0.5, 1
and 2 kHz as it is typically reported (Vermiglio et al., 2020). Two of these participants were not usual
hearing-aid users. The hearing-impaired listeners (HI) group consisted of 70 participants with symmetric
sensorineural hearing losses. Symmetric sensorineural hearing loss was defined as an interaural difference
(ID) $\leq$ 15 dB HL at frequencies below 8 kHz and ID $\leq$ 25 dB HL at 8 kHz and air-bone gap $\leq$ 10 dB
HL. The pure-tone audiograms of the participants are shown in Figure 2. The participants eligible for the
present study had audiometric thresholds $\leq$ 55 dB HL (pure-tone audiometry not older than 1 year) in
the range between 125 and 1000 Hz. Participants with a pure tone threshold $\geq$ 75 dB HL at 2 kHz were
excluded from the study as it was unlikely that it would be feasible to perform all the tests due to audibility
issues.

The participants were recruited from the BEAR database (Houmoller et al., 2021) at Odense University
Hospital (OUH), from the patient database at Bispebjerg Hospital (BBH), and from the database at the
Hearing Systems Section at the Technical University of Denmark (DTU). The study was approved by the
Science-Ethics Committee for the Capital Region of Denmark, H-16036391. All participants gave written
informed consent and some of them received economical compensation for their participation depending
on the test site and whether the participant was willing to particpate without compensation.
131

132   ## 3.2   Equipment
133   The basic audiological assessment consisted of pure-tone audiometry, wideband tympanometry (Rosowski
134   et al., 2013) and middle ear muscle reflex, and was conducted in the facilities of OUH, BBH and DTU.
135   The rest of the tests were performed via PC in a double-walled sound-insulated booth (BBH and DTU)
136   or in a small anechoic chamber (OUH). The tests were implemented in Matlab with a graphical user

---

3   Despite other listeners presented PTA $\leq$ 22 dB, the individual thresholds did not fulfil this criteria.

Figure 2. Audiograms of the 75 participants of the study together with the average for each ear (dark solid lines) and interquartile ranges (grey areas). The grey dashed lines correspond to the standard audiograms N1 and N4 from Bisgaard et al. (2010).

interface (GUI) that the examiner could operate without programming experience. Most of the tests were implemented using a modular framework for psychoacoustic experiments (AFC; Ewert, 2013), except for HINT, provided by Jens Bo Nielsen and Binaural Pitch test which was a reimplementation of the Binaural Pitch Test v1.0, Bispebjerg hospital, 2008. The participants were seated in the room and the stimuli were presented through headphones (Sennheiser HDA200) connected to a headphone-amplifier (SPL phonic) and an audio interface (RME Surface 24-bit). The equipment was calibrated using an artificial ear according to IEC 60318-1:2009. The tests consisting of threshold estimation using the AFC framework were repeated at least two times and the mean of the two measurements was considered as the final value. To ensure the quality of the data collected, a repetition was considered as an outlier if it was greater than three scaled median absolute deviations and additional repetitions were suggested by the framework until a certain standard deviation across repetitions was achieved.

3.3 Analysis of test reliability

The test-retest reliability of the test battery was assessed using intraclass correlation coefficients (ICC; Koo and Li, 2016), and the standard error of measurement (SEM; Stratford and Goldsmith, 1997). It was of special interest to test the reliability in older listeners with different hearing abilities. Therefore, test-retest measurements were performed with a subgroup consisting of 11 participants for all tests of the test battery. The seven listeners had bilateral hearing loss with a mean PTA of 31 dB HL. The participants were aged between 59 and 82 years (median 69 years). The retest session was conducted within four months after the first visit.

4   HIGH-FREQUENCY AUDIBILITY

Recently, elevated thresholds at high frequencies (> 8 kHz) have been linked to the concept of "hidden
hearing loss" and synaptopathy (Liberman et al., 2016). However, the measurement of audiometric
thresholds above 8 kHz is not part of the current clinical practice. The fixed-level frequency threshold
(FLFT) has been proposed as a quick and efficient alternative to high-frequency audiometry (Rieke et al.,
2017; Prendergast et al., 2020). The test is based on the detection of a tone presented at a fixed level. The
frequency of the tone is varied towards high frequencies and the maximum audible frequency at the given
level is estimated in an adaptive procedure. Here, a modified version of FLFT, using warble tones presented
at 80 dB SPL, was used as the extended audiometry at high frequencies (eAUD-HF).
164

4.1   Method

The procedure used here was a yes/no task using a SIAM procedure (Kaernbach, 1990). As in traditional
up-down procedures, the target can be presented in a given trial or not. If the target was detected, the
frequency of the warble tone was increased according to a given step size; if it was not detected, the
frequency was decreased. For each run, the first two reversals were discarded, and the threshold of each
trial was calculated as the average of the four subsequent reversals. This procedure includes catch trials
where no sound was presented.
172

4.2   Results and discussion

174 The results of the FLFT measured at 80 dB SPL are shown in Table 4.
175

Table 4. Summary of the results of the Extended Audiometry for high frequencies eAUD-HF. The results
are shown as the mean, standard deviation, first and third quantiles (25th and 75th percentiles) for each ear.

			ONH			HI		
Outcome measure	*Level*	*Ear*	*Mean (SD)*	*Q1*	*Q3*	*Mean (SD)*	*Q1*	*Q3*
eAUD-HF	80	*LE*	*10.9 (1.2)*	*10.2*	*11.9*	*7.57 (2.7)*	*5.3*	*10*
FLFT (kHz)	dB SPL	*RE*	*11.7 (1.1)*	*10.9*	*12.5*	*8.12 (2.3)*	*6.7*	*10.2*

The maximum frequency threshold for a tone presented at 80 dB SPL (eAUD-HF) was 11 kHz for the
ONH listeners and 8 kHz for the HI listeners. The HI group showed larger variability compared to the
NH group (interquartile range: 6 kHz vs. 10 kHz). The eAUD-HF test showed very good reliability (ICC
= 0.89; SEM = 495 Hz). These results suggest that the FLFT paradigm might be a good time-efficient
alternative to the traditional audiometry for measuring high-frequency sensitivity. A recent study pointed
out the importance of off-frequency listening and the role of the excitation of the basal cochlea when
presenting narrow-band stimuli in high levels (Encina-Llamas et al., 2019). Knowing the hearing sensitivity
at high-frequencies of a given patient might be crucial for better understanding of their supra-threshold
deficits. Moreover, The eAUD-HF can include different levels and be a useful not only for ototoxicity
monitoring but also in association with other supra-threshold measues.

5   LOUDNESS PERCEPTION

Loudness perception can substantially differ between NH and HI listeners and has been connected to the
peripheral non-linearity (e.g. Jürgens et al., 2011). While the growth of loudness shows a non-linear
behaviour in a healthy ear, the results from HI listeners suggest that loudness perception becomes

linear when outer-hair cell (OHC) function is affected (e.g Moore, 2007). Besides, the possibilities
of characterizing hearing deficits, loudness function can be used for fitting hearing aids (e.g. Oetting et al.,
2018). Adaptive categorical loudness scaling (ACALOS; Brand and Hohmann, 2002) is the reference
method for the current standard (ISO 16832, 2006) for loudness measurements.

5.1 Methods

According to the ACALOS method, 1/3-octave band noise were presented sequentially, and the participant
had to judge the perceived loudness using a 11-category scale ranging from "not heard" to "extremely
loud". The presentation level of the next stimulus was calculated based on the previous trials. The raw
results, which correspond to categorical units (CU) spanned between 0 and 50, were fitted to a model of
loudness as described in (Oetting et al., 2014). The outcome measures of the ACALOS presented here
are the most comfortable level (MCL), the slope of the loudness function (Slope), and the dynamic range
(DynR) defined as the difference between the uncomfortable level (50 CU) and the hearing threshold (0.5
CU). Low-frequency (LF) average corresponds to frequencies below 1.5 kHz, high-frequency (HF) average
correspond to frequencies above 1.5 kHz.

5.2 Results and discussion

The results of the ACALOS outcome measures are shown in Table 5. The hearing thresholds (HTL)
estimated by ACALOS were significantly correlated with the pure-tone audiometric thresholds ($\rho =$
$0.88; p << 0.0001$) even when looking at the HI group alone ($\rho = 0.83; p << 0.0001$) despite the use of
different stimuli and procedure.

Table 5. Summary of the results of the adaptive categorial loudness scaling test (ACALOS). The results of
the most comfortable level (MCL), slope of the growth of loudness (Slope) and dynamic range (DynR) are
shown for low (LF) and high frequencies (HF) as the mean, standard deviation, first and third quantiles
(25th and 75th percentiles) for each ear.

Outcome measure	Freq. Range	Ear	ONH Mean (SD)	Q1	Q3	HI Mean (SD)	Q1	Q3
MCL (dB HL)	LF	LE	81.5 (14.8)	73.3	84.1	80.6 (8.4)	76.4	85.8
		RE	76.5 (13.2)	70	80	79.1 (7.9)	74.7	84.1
	HF	LE	79.0 (17.6)	66.6	90.8	82.7 (12.3)	75.8	90
		RE	73.8 (17.2)	65	80	80.3 (9.9)	74.7	87.5
Slope (CU/dB)	LF	LE	0.35 (0.1)	0.3	0.4	0.45 (0.1)	0.3	0.5
		RE	0.36 (0.1)	0.3	0.4	0.48 (0.2)	0.3	0.5
	HF	LE	0.45 (0.1)	0.3	0.4	0.84 (0.5)	0.5	0.9
		RE	0.41 (0.1)	0.3	0.4	0.81 (0.4)	0.5	0.9
DynR (dB HL)	LF	LE	91.5 (16.8)	78.3	97.5	76.7 (15.8)	64.5	88.3
		RE	91.1 (18.8)	79.1	100	73.9 (16.0)	61.6	86.8
	HF	LE	77.6 (18.2)	72.5	85.8	50.8 (15.1)	40.6	60.2
		RE	78.6 (17.9)	67.5	90.8	50.7 (15.5)	38.9	60.4

The average MCL estimate ranged between 73 and 82 dB HL in both groups and for both frequency ranges.
The average slope of the loudness growth was slightly steeper for the HI listeners in the low-frequency
range (0.45 CU/dB for HI vs. 0.35 CU/dB for ONH) and substantially steeper in the high-frequency range
(0.8 CU/dB for HI vs 0.45 CU/dB for NH). The average dynamic range was between 80 and 90 dB HL for
the ONH listeners, and smaller for the HI listeners, especially at high frequencies (50.8 dB). Regarding the
test-retest reliability, ACALOS showed an excellent reliability for estimating the hearing thresholds (ICC =

0.94; SEM = 4.5 dB), good reliability for estimating the MCL (ICC = 0.68, SEM = 6.5 dB) and very good
reliability for estimating the slope (ICC = 0.82; SEM = 0.07 CU/dB). Overall, these results supported the
inclusion of ACALOS in a clinical test battery, as it provides several outcomes (hearing thresholds, growth
of loudness, MCL and dynamic range). ACALOS also showed a high time efficiency (around 10 min. per
ear).

6 SPEECH PERCEPTION IN QUIET

6.1 Method

The word recognition score with four unforced choices (WRS-4UFC) test was proposed as a systematic
and self-administered procedure that allows the estimation of supra-threshold deficits in speech perception
in quiet. The speech material was the same as the one used for standard speech audiometry (Dantale I;
Elberling et al., 1989) in Danish. The self-administered procedure consisted of the presentation of one word
were the patient has to answer in a 4-unforced-choice paradigm (4UFC). After the acoustical presentation
of each word, the target written word was assigned randomly to one of four buttons shown to the patient.
The other three buttons contained words that were also taken from the Dantale-I corpus. They were chosen
based on the lowest Levenshtein phonetic distance (Sanders and Chin, 2009) from the target. The term
"unforced" corresponds to an additional choice, a question mark, that the listener can press if none of the
four options are considered the right answer. Four lists of 25 words were presented at 40, 30, 20 and 10 dB
above the individual PTA, in this order. A logistic function was fitted to the results from each individual ear
and the speech reception threshold ($\mathrm{SRT_Q}$), at which 50% of the words were recognized, and maximum
speech discrimination score (maxDS), which was the maximum value of the function. Both outcomes were
estimated using psignifit 4 software (Schütt et al., 2016).

6.2 Results and discussion

236 The results of the WRS-4UAFC outcome measures are shown in Table 6.

Table 6. Summary of the results of the word recognition scores (WRS-4UFC) test. The results of the speech recognition threshold ($\mathrm{SRT_Q}$), and maximum discrimination score (maxDS) are shown as the mean, standard deviation, first and third quantiles (25th and 75th percentiles) for each ear.

Outcome measure	Ear	ONH			HI		
		Mean (SD)	*Q1*	*Q3*	*Mean (SD)*	*Q1*	*Q3*
SRTQ (dB HL)	*LE*	*19.9 (7.1)*	*16.5*	*19.2*	*41.5 (13.5)*	*31.8*	*50.6*
	RE	*23.3 (8.9)*	*17.2*	*29.0*	*42.7 (12.6)*	*33.9*	*51.1*
Max DS (%)	*LE*	*99.2 (1.6)*	*100*	*100*	*97.2 (4.1)*	*95.3*	*100*
	RE	*97.2 (1.8)*	*95.5*	*97.6*	*93.9 (6.4)*	*92.1*	*98.4*

The HI listeners' $\mathrm{SRT_Q}$ were, on average, 20 dB higher than the ones of the ONH group. The interquartile
range for the HI group was about 19 dB whereas for the ONH group it was 3 dB for the left ear (LE)
and 11.8 dB for the right ear (RE). The maxDS for both groups was close to 100%. However, the HI
listeners showed larger variability, especially in the right ear (SD= 6.42%). In the analysis of the test-retest
variability, the WRS-4UFC test showed poor to moderate reliability, especially at low levels (PTA + 10
dB; ICC = 0.25). However, at the higher presentation levels (i.e., individual PTA + 40 dB) the standard
error of the measurement was only 4% (1 word). Regarding clinical applicability, the WRS-4UFC needs to
be compared to traditional speech audiometry to explore the influence of using closed- vs. open-set and
forced- vs. unforced-choice test procedures on the results.

7 SPEECH PERCEPTION IN NOISE

The Hearing in Noise Test (HINT; Nilsson et al., 1994) is an adaptive sentence recognition test carried out
with speech-shaped noise. The following assumptions are considered in HINT (based on Plomp, 1978): 1)
Speech materials made of meaningful sentences yield a steep psychometric function; 2) Stationary noise
with the same spectral shape as the average spectrum of the speech material makes the speech reception
threshold in noise (SRT_N) less dependent of the spectral characteristics of the speaker's voice. Furthermore,
the signal-to-noise ratio (SNR) between the target and masker is better defined across the frequency range;
3) The (SRT_N) is independent of the absolute noise level as long as the noise level is above the "internal
noise" level. Therefore, it is recommended to present the noise at least 30 dB above the "internal noise".
The internal noise is defined as the sum of the SRT in quiet of the tested listener and the SRT in noise for
NH listeners, for a given speech material.

7.1 Methods

The Danish HINT was used as in Nielsen and Dau (2011) to obtain the SRT_N but in a monaural
presentation. Additionally, a 20-sentence list was presented at a fixed signal-to-noise ratio of +4 dB and
scored to obtain a sentence recognition score ($SScore^{+4dB}$). The presentation level of the noise was set
between 65 and 85 dB SPL to ensure that the noise was always presented 30 dB above the individual PTA.
Each ear was tested individually. All participants were tested using the same list with the same ear. Since
small differences across lists were found in Nielsen and Dau (2011), this was done to ensure that all the
listeners were tested with an equally difficult list. However, for the test-retest reliability study, the list and
ear presented were randomized, only using lists 6-10. The listeners did not report recalling sentences from
the test.

7.2 Results and discussion

267 The results of the HINT outcome measures are shown in Table 7.

Table 7. Summary of the results of the hearing in noise (HINT) test. The results of the speech recognition
threshold in noise (SRT_N), and sentence recognition score at +4 dB SNR ($SScore^{+4dB}$) are shown as the
mean, standard deviation, first and third quantiles (25th and 75th percentiles) for each ear.

| *Outcome measure* | *Ear* | *ONH* | | | *HI* | | |
		Mean (SD)	*Q1*	*Q3*	*Mean (SD)*	*Q1*	*Q3*
SRT_N (dB SNR)	*LE*	*1.0 (0.7)*	*0.4*	*1.5*	*4.1 (3.4)*	*1.4*	*6.7*
	RE	*-0.5 (1.1)*	*-1.0*	*0.0*	*2.6 (3.8)*	*0.0*	*4.2*
$SScore^{+4dB}$ (%)	*LE*	*85.0 (11.7)*	*85*	*90*	*60.0 (26.6)*	*40*	*85*
	RE	*91.0 (9.6)*	*90*	*95*	*62.3 (24.0)*	*48.7*	*80*

The SRT_N for ONH listeners were, on average, 2 dB higher than the ones reported (Nielsen and Dau,
2011). This bias was also observed in the YNH listeners. However, this might be explained by the fact
that they used diotic presentation which can lead to a 1.5 dB improvement as reported by Plomp and
Mimpen (1979). The results also showed a lower SRT_N (1.5 dB) and higher $SScore^{+4dB}$ (4%) for the
right ear in both groups of listeners. According to (Nielsen and Dau, 2011), there was a significant main
effect of test list. Such differences are seen mainly for lists 1-4, which were the lists used here. Therefore,
the observed interaural difference can be ascribed to a list effect, however, it might be ascribed to other
factors as, for example, a right-ear advantage as the one observed in NH listeners with tinnitus (Tai and
Husain, 2018). The ICC values (SRT_N: ICC = 0.61; $SScore^{+4dB}$: ICC = 0.57) indicated only moderate
reliability of the HINT. The SRT_N showed an SEM = 1.02 dB, which is below the step size of the test

(2 dB). The $\text{SScore}^{+4\text{dB}}$ showed an SEM value of 7.94%, which corresponds to an error in one of the
sentences. However, the reliability of the test can be improved by using an adaptive method as the one
described in Wagener et al. (2003); Rønne et al. (2017) where the SRT_N estimation was optimized using a
combination of word scoring and a maximum likelihood procedure.

8 SPECTRO-TEMPORAL MODULATION SENSITIVITY

A speech signal can be decomposed into spectral and temporal modulations. While speech-in-noise
perception assessment leads to some confounds due to the variety of speech corpora, noise maskers,
and test procedures that can all affect the results, the assessment of the contrast sensitivity of simpler
sounds might be of interest for characterizing a listener's spectro-temporal processing abilities. Bernstein
et al. (2013) showed significant differences between NH and HI listeners for detecting STM in random
noise. These differences corresponded to specific conditions that were also useful for the prediction of
speech-in-noise performance in the same listeners. Lately, the assessment of STM sensitivity in these
specific conditions gained an increasing interest due to its potential for predicting speech intelligibility
(Bernstein et al., 2016; Zaar et al., 2020) and for assessing cochlear-implant candidacy (Choi et al., 2016).
Here, STM sensitivity was assessed using a new test paradigm that may be more suitable for a clinical
implementation. The test was performed in two conditions: a low-frequency condition (similar to the one
previously used in Bernstein et al., 2016) and a high-frequency condition (Mehraei et al., 2014).

8.1 Methods

The stimuli were similar to those of Bernstein et al. (2016) and Mehraei et al. (2014), but a different
presentation paradigm was employed. A sequence of four noises was presented in each trial. The first and
third stimulus always contained unmodulated noise, whereas the second and fourth stimuli could be either
modulated or unmodulated. The test was performed with a low-frequency 3-octaves wide stimulus centred
at 800Hz (sSTM_8 and fSTM_8), and a 1-octave wide stimulus centred at 4 kHz (sSTM_{4k} and fSTM_{4k}.
The stimuli were presented at 75 dB sound pressure level (SPL). After the sequence was presented, the
listener had to respond whether the four sounds were different ('yes') or the same ('no'). Two procedures
involving catch trials were evaluated. The first test (sSTM -3 dB) was a screening test consisting of
10 stimuli modulated at 20log(m) = -3 dB modulation level (ML), where m is the modulation depth,
and five unmodulated ones presented in random order. The outcome measure was the listener's contrast
sensitivity (d')[3] in the task. The second test (fSTM) tracked the 80% threshold using a yes/no task and the
single-interval adjusted matrix (SIAM; Kaernbach, 1990) paradigm. For the sSTM test, the stimulus was
presented diotically whereas for the fSTM the test was presented in each ear individually in a monoaural
presentation.

8.2 Results and discussion

The results of the spectro-temporal modulation sensitivity tests outcomes are shown in Table 8.

The screening STM test (sSTM) shows the sensitivity in terms of d', where the maximum value is d' = 3,
(i.e., 10 modulated and 5 unmodulated stimuli correctly detected). In the hypothetical case when all the
catch trials are detected, the lowest d' value can be -0.3. The ONH listeners showed a high sensitivity in
the low-frequency condition (d' = 2.6) and a somewhat lower sensitivity in the high-frequency condition
(d' = 1.63) corresponding to 65% correct responses. The HI listeners showed a higher variability and a
lower sensitivity in the low-frequency condition ($\approx$ 70% correct) and substantially lower sensitivity in the
high-frequency condition (0-50% correct responses). The threshold-tracking procedure (fSTM) showed

---

[3] d' was defined as $Z(\frac{N_H+0.5}{H+1}) - Z(\frac{N_{FA}+0.5}{FA+1})$ where Z refers to the z-score transformation, H is the total number of target presentations and FA the total number of catch trials.

Table 8. Summary of the results of the spectro-temporal modulation sensitivity tests. The results of the screening STM sensitivity test (sSTM), and the threshold of the fast STM sensitivity test (fSTM) are shown for the low-frequency stimulus (8, because it is centered at 800 Hz) and high-frequency (4k, centered at 4kHz) as the mean, standard deviation, first and third quantiles (25th and 75th percentiles) for each ear.

Outcome measure	Freq. Range	Ear	ONH Mean (SD)	Q1	Q3	HI Mean (SD)	Q1	Q3
sSTM -3dB (d')	LF	Bin	2.6 (0.6)	2.4	3	1.7 (1.3)	0.4	3
	HF		1.6 (0.8)	1.1	2.4	0.6 (1.1)	-0.3	1.4
fSTM (dB)	LF	LE	-7.7 (1.8)	-9	-7.6	-2.8 (2.1)	-3.5	-0.8
		RE	-5.1 (3.1)	-7.2	-1.6	-1.6 (1.3)	-2	-0.6
	HF	LE	-8.0 (2.0)	-8.6	-6.2	-2.6 (2.4)	-3.8	-0.6
		RE	-5.6 (3.6)	-8.6	-2.1	-1.9 (1.5)	-2	-1

results between -9 and -1.6 dB ML in the ONH group, whereas the HI listeners showed thresholds between
-3.50 and -0.6 dB ML in the low-frequency condition. Although the results of the fSTM low-frequency
condition were consistent with Bernstein et al. (2016), the results in the high-frequency condition showed
higher thresholds than the ones in Mehraei et al. (2014). This can be ascribed to the higher presentation
level used in Mehraei et al. (2014) than in the current test procedure. The fSTM showed an excellent
reliability (ICC = 0.91; SEM = 0.93 dB ML) in the LF condition. However, several HI listeners were not
able to complete the procedure for the HF condition. Overall, the use of the SIAM tracking procedure
allowed us to obtain accurate thresholds, although additional repetitions were required, especially in
the HF condition. This might be because the psychometric function for detecting the stimulus can be
shallower in this condition, or because the 100% detection could not be reached even in the fully-modulated
trials. Therefore, a Bayesian procedure being able to estimate the threshold and slope of the psychometric
function, such as the Bayes Fisher information gain (FIG: Remus and Collins, 2008), might be more
suitable for this type of test. Another reason explaining the inability of the listeners to perform the test can
be ascribed to the stimulus.Zaar et al. (2018) used a longer stimulus (1s) a diotic presentation and a hearing
loss compensation that ensured the audibility of the stimulus in all its frequency range. In their study, all
the listeners were able to perform the tests and their sensitivity thresholds were well below the maximum
value.

9 EXTENDED AUDIOMETRY IN NOISE (EAUD)

335 The extended audiometry in noise (eAUD) is a tone detection test intended to assess different aspects
336 of auditory processing by means of a task similar to pure-tone audiometry. The tone is presented in the
337 presence of noise and the listener has to indicate whether the tone was perceived or not. The aspects of
338 auditory processing assessed here are 1) tone-in-noise detection, and 2) spectral and temporal resolution.
339

9.1 Tone-in-noise detection

In the pure-tone audiometry, a given patient has to detect the simple stimulus (e.g., sinusioids) in quiet
aiming at estimating the hearing thresholds of the listener. A simple way to explore the supra-threshold
performance is to perform a tone-in-noise detection test by presenting noise at supra-threshold levels and
obtaining the masked thresholds. However, the characteristics of the noise such as bandwidth, level or
inherit modulations can affect the results. Moore et al. (2000) proposed a test paradigm using a special type
of noise which is able to provide the same masking in the entire frequency range, so the hearing thresholds
of a NH listener would raise according to the level of the noise. This is the so-called threshold-equalizing

noise (TEN). The advantage of the TEN test is that the expected masked thresholds are similar to the
level of the noise (i.e., as TEN is played at 70dB per equal rectangular bandwidth (ERB), the masked
threshold is expected to be at 70 dB SPL). Although this test was originally design to detect dead cochlear
regions, recent evidence suggests that tone-in-noise detection can be representative of supra-threshold
deficits beyond the audiogram (Schädler et al., 2020).

9.2   Spectro-temporal resolution

Frequency and temporal resolution are aspects of hearing that are fundamental for the analysis of
perceived sounds. While NH listeners exhibit a frequency selectivity on the order of one third of an octave
when using isoinput levels (from Glasberg and Moore, 1990; Eustaquio-Martín and Lopez-Poveda, 2011),
HI listeners have typically broader auditory filters, leading to impaired frequency selectivity (Moore, 2007).
Temporal resolution can be characterized by the ability to "listen in the dips" when the background noise is
fluctuating based on the so-called masking release (Festen and Plomp, 1990). Schorn and Zwicker (1990)
proposed an elaborated technique for assessing both spectral and temporal resolution using two tests:
1) Psychoacoustical tuning curves and 2) temporal resolution curves. In both cases, the task consists of
detecting a pure tone that is masked by noise or another tone while the spectral or temporal characteristics
of the masker are varied. Later, Larsby and Arlinger (1998) proposed a similar paradigm, the F-T test,
which was successfully tested in HI listeners (van Esch and Dreschler, 2011). Here, the spectro-temporal
resolution was assessed using a new test. This test is a tone-in-noise detection task consisting of three
conditions as sketched in Figure 3.

Figure 3. Sketch of the conditions of the spectro-temporal resolution measures of the extended audiometry in noise (eAUD). The top panel shows the spectrum of the noise and target pure-tone (delta), the bottom panel shows both signals in the time domain. Left panel: Tone in noise condition (eAUD-N). Middle panel: Spectral condition (eAUD-S). Right panel: Temporal condition (eAUD-T).

1. eAUD-N: The tone is embedded in a 1-octave-wide threshold equalizing noise (TEN-HL; Moore,
2001). Because of the properties of the TEN-HL, the tone detection threshold is comparable to the
level of the noise in dB HL.

2. eAUD-S: The tone is embedded in a TEN that has been shifted up in frequency. In the spectral domain,
this yields spectral unmasking of the tone, so the detection threshold is lower than in eAUD-N.

3. eAUD-T: The tone is embedded in a temporally-modulated noise with the same spectral properties as
the one in eAUD-N. In the temporal domain, the modulations of the noise yield temporal unmasking,
so the tone can be detected in the dips.

The outcome measures were focused on the temporal and spectral benefits expected in the eAUD-S
and eAUD-T conditions compared to the eAUD-N condition. While in the noise condition (eAUD-N) the
threshold is expected to be approximately at the level of the noise, in the temporal and spectral conditions
the thresholds should be obtained at a lower level, showing temporal masking release (TMR) and spectral
masking release (SMR)

380
9.3   Methods

The procedure used here was a yes/no task using a SIAM procedure (Kaernbach, 1990) similar to the one
used in the eAUD-HF. Here, a TEN was presented together with a warble tone. If the target was detected,
the target-presentation level is decreased according to a given step size; if it was not detected, the level
is increased. If the stimulus was not presented (catch trial) but the listener provided a positive response,
the level is decreased compared to the previous trial. As in the eAUD-HF, for each run, the threshold of
each trial was calculated as the average of the last four reversals. The noise was presented at 70 dB HL.
The low-frequency condition (LF) corresponds to the detection of a 0.5-kHz warble tone, whereas the
high-frequency (HF) condition corresponded to a 2-kHz warble tone. The final threshold was calculated
as the mean threshold of two repetitions. In the eAUD-S condition, the center frequency of the noise was
$f_{c,\text{noise}} = 1.1 f_{\text{tone}}$. In the eAUD-T condition, the modulation frequency of the noise was set to, $f_m = 4$
Hz. The outcome measures of the eAUD are 1) the tone-in-noise threshold (TiN), 2) the temporal masking
release (TMR), and 3) the spectral masking release (SMR).

394
9.4   Results and discussion

The results of the extended audiometry in noise outcomes are shown in Table 9.

The TiN showed a larger variance for the ONH group (SD = 4.5 dB HL) at low frequencies. The detection
thresholds were in line with previous work with thresholds close to the noise presentation level (70 dB HL)
(Vinay et al., 2017). The TMR shown by the NH group was larger at high frequencies (10 dB) than at low
frequencies (7 dB). The HI group showed, on average, similar TMR only at low frequencies. The SMR
shown by the ONH listeners was 19 dB for low frequencies and 26 dB for high frequencies. In contrast,
for the HI listeners, the SMR was 7 dB lower only in the high-frequency condition. The reliability of the
eAUD was moderate for most of the conditions (ICC ≤ 0.75). The eAUD-S at low frequencies showed
good reliability (ICC = 0.85; SEM = 1.78 dB). The masking release estimates showed good reliability
only for the high-frequency condition. The reason for this might be that masking release is a differential
measure, and the cumulative error is, therefore, higher than that of each individual measure. The reduced
reliability can be explained to some extent by the method used. To have a similar procedure as in pure-tone
audiometry, the parameters of the SIAM tracking procedure were set accordingly. However, this made the
test challenging and the listeners consistently missed several catch trials. Thus, extra trials were required to

Table 9. Summary of the results of the extended audiometry in noise (eAUD). The results of the tone-in-noise (TiN), temporal masking release (TMR) and spectral masking release (SMR) are shown for the low-frequency condition (LF; 500Hz) and high-frequency condition (HF; 2kHz) as the mean, standard deviation, first and third quantiles (25th and 75th percentiles) for each ear.

Outcome measure	Freq. Range	Ear	ONH Mean (SD)	Q1	Q3	HI Mean (SD)	Q1	Q3
TiN (dB HL) eAUD-N	LF	LE	70.4 (4.5)	68	71.5	71.8 (2.6)	70.2	73.2
		RE	69.2 (4.6)	65.2	72.5	72.0 (2.8)	69.6	74.3
	HF	LE	71.1 (2.5)	69.7	72.7	74.7 (3.4)	72.5	76.1
		RE	70.8 (3.6)	70.5	71.7	74.2 (3.1)	72	76.2
TMR (dB) eAUD (N -T)	LF	LE	7.5 (3.4)	6	7.5	7.7 (4.0)	6.1	10.1
		RE	5.2 (3.3)	4	7.6	8.3 (2.7)	6.5	10.3
	HF	LE	13.0 (0.6)	12.7	13.2	7.9 (5.0)	5	11.6
		RE	10.7 (3.1)	9.1	10.2	8.1 (5.2)	5.1	10.7
SMR (dB) eAUD (N -S)	LF	LE	19.3 (3.6)	16.5	21.7	19.6 (17.7)	17.7	23.2
		RE	18.8 (4.6)	17	21.2	20.0 (5.2)	16.5	23.8
	HF	LE	26.8 (4.5)	27.5	29	19.3 (9.5)	12.1	26.3
		RE	27.2 (3.7)	26.2	29.5	19.5 (9.9)	12	26.8

410 improve measurement accuracy, especially in the eAUD-N condition. Furthermore, the standard error of
411 the measurement was in most cases larger than the final step size (2 dB). As in the case of the fSTM, a
412 different procedure, such as Bayesian adaptive methods, might increase measurement reliability.

10 BINAURAL PROCESSING ABILITIES

Binaural hearing is useful for sound localization and the segregation of complex sounds (Darwin, 1997).
Interaural differences in level or timing are processed for spatial hearing purposes in the auditory system.
In the case of hearing loss, the neural signal at the output of the cochlea can be degraded which may lead
to reduced binaural abilities typically connected to temporal fine structure (TFS) processing. Based on
a method of estimating the upper-frequency limit for detecting an interaural phase difference (IPD) of
180° (IPD$_{fmax}$ Ross et al., 2007; Neher et al., 2011; Santurette and Dau, 2012), Füllgrabe and Moore
(2017) recently proposed a refined test as a feasible way to evaluate TFS sensitivity. This paradigm was
used in recent research that suggested that IPD$_{fmax}$ might be related to non-auditory factors (Strelcyk
et al., 2019) and affected by factors beyond hearing loss, such as musical training (Bianchi et al., 2019).
Therefore, the IPD$_{fmax}$ might be a task that requires auditory and non-auditory processing abilities beyond
TFS sensitivity.
In contrast, binaural pitch detection assesses binaural processing abilities in a different manner. This test
requires the detection of pitch contours embedded in noise, which are diotically or dichotically evoked.
While the diotic condition can be resolved monoaurally, the dichotic condition requires the binaural
processing abilities to be sufficiently intact to detect the contour. Previous studies showed that some
listeners were unable to detect binaural pitch, regardless of the audiometric configuration (Sanchez-Lopez
et al., 2018; Santurette and Dau, 2012). Therefore, it was of interest to compare the results of these two
binaural processing tests.
Another approach for evaluating the binaural processing abilities is assessing binaural masking release
(Durlach, 1963), which has been used in several studies (e.g. Neher, 2017; Strelcyk and Dau, 2009) and
implemented in some commercial audiometers (Brown and Musiek, 2013). In this paradigm, a tone-in-noise
stimulus is presented in two conditions: (1) a diotic condition where the tone is in phase in the two ears,
and (2) a dichotic condition where the tone is in antiphase in the two ears. The difference between the

two yields the benefit for tone detection due to binaural processing, the so-called binaural masking release
(BMR).
438

10.1 Methods

The maximum frequency for detecting an IPD of 180° with pure-tones was obtained using a 2-AFC
tracking procedure similar to the one used in Füllgrabe and Moore (2017). The stimuli were presented
bilaterally in both ears as two sequencies of four tones. One sequence contained an ABAB sequence,
where A means a dichotic presentation and B an IPD of 180° between the tones presented to each ear,
and the other an AAAA sequence. A positive response (detection) increased the frequency of the tone,
and a negative response a decrease of the frequency. Although the stimuli duration and procedure were
similar, the step size used here was slightly different, starting with steps of 2/3 octave and decreasing to a
final step size of 1/6 octave in each reversal. The last six reversals were used for estimating the threshold.
The frequency threshold (IPD_{fmax}) was obtained from the average of two runs. Binaural pitch detection
scores were obtained using a clinical implementation of the test proposed by Santurette and Dau (2012). A
3-minute sequence of noise was presented bilaterally. Ten diotic and ten dichotic pitch contours, embedded
in the noise, had to be detected by the listener. The tones forming the pitch contours were generated by
adding frequency-specific IPDs to the presented noise (Cramer and Huggins, 1958). The outcome measure
of the binaural pitch test was the percentage score of detecting the dichotic pitch contours only, averaged
across two repetitions (BP20). The BMR was assessed using the same method as the extended audiometry.
Two measurements were required: 1) tone-in-noise detection presented diotically (S_0N_0) and tone-in-noise
detection presented dichotically, i.e., with the tone in anti-phase across the two ears ($S_\pi N_0$).

10.2 Results and discussion

458   The results of the tests assessing binaural processing abilities are shown in Table 10.
459

Table 10. Summary of the results of the binaural processing abilities tests. The results of the maximum frequency for IPD detection (IPD_{fmax}), Binaural Pitch detection scores (BP_{20}) and the binaural masking release (BMR) are shown as the mean, standard deviation, first and third quantiles (25th and 75th percentiles) for each ear.

Outcome measure	Ear	ONH			HI		
		Mean (SD)	Q1	Q3	Mean (SD)	Q1	Q3
IPD fmax (kHz)	Bin	0.76 (0.26)	0.59	0.98	0.69 (0.27)	0.52	0.88
Bin Pitch 20 (%)	Bin	87.5 (25.0)	87.5	100	80.7 (30.9)	70	100
BMR (dB) $(S_0N_0 - S_\pi N_0)$	Bin	16.5 (4.7)	13.5	17.5	14.7 (4.6)	12.2	17.5

The listeners in the ONH and HI groups showed IPD_{fmax} thresholds around 700 Hz with a standard
deviation (≈ 270 Hz) and interquartile range (≈ 370 Hz) similarly in both groups. These results are in line
with the ones reported in Füllgrabe and Moore (2017). The IPD_{fmax} test showed excellent reliability (ICC
= 0.95; SEM = 65.4 Hz), and the median time needed for two repetitions was 10 minutes. This suggests
that IPD_{fmax} is a reliable measure of binaural processing abilities that can reveal substantial variability
among both NH and HI listeners, which is valuable for highlighting individual differences among patients.
The overall results from the binaural pitch test for the NH listeners showed $> 87.5\%$ correct detection,
whereas the HI listeners' results showed a higher variability with an interquartile range from 70-100%.
The test showed excellent reliability (ICC = 0.98; SEM = 4%). Listeners reported a positive experience

due to the test being short and easy to understand. The BMR shown by both groups was around 15 dB, as
expected from previous studies (Durlach, 1963).

11 EXPLORATORY ANALYSIS

The collection of tests included in the test battery was intended to explore different and potentially
independent aspects of hearing to obtain an auditory profile with controlled interrelations among the
tests. A factor analysis performed in the HEARCOM study (Vlaming et al., 2011) based on data from 72
HI subjects revealed auditory dimensions: 1) high-frequency processing, 2) audibility, 3) low-frequency
processing and 4) recruitment. In the current study, the results of the behavioural tests were analysed further
in order to explore possible interrelations between the various outcome measures.

11.1 Methods

477
First, the data were pre-processed as in Sanchez-Lopez et al. (2018) to reduce the number of variables.
The outcome variables of the frequency-specific tests were divided into LF ($\leq$1 kHz) and HF ($\i$1 kHz)
variables. This decision was supported by a correlation analysis performed on the complete set of outcome
variables, where the outcomes corresponding to 2, 4 and 6 kHz as well as the ones corresponding to
0.25, 0.5 and 1 kHz were highly intercorrelated. For the tests performed monaurally, the mean of the
two ears was taken as the resulting outcome variable. The resulting dataset (BEAR3 dataset[1]) contained
26 variables, divided into six groups corresponding to the six aspects of auditory processing considered
here. The exploratory analysis consisted of a correlation analysis using Spearman correlations and factor
analysis. The factor analysis was performed using an orthogonal rotation ("varimax") and the method of
maximum likelihood. The number of components was chosen to use parallel analysis, the resulting number
of components was four.

11.2 Results

Figure 4 shows the results from the correlation analysis performed on the BEAR3 dataset. For convenience,
the absolute value of the correlation was used when visualizing the data to show the strength of the
correlation. The circles on the left-hand side of the figure depict significant correlations ($p < 0.00001$), and
the correlation values are presented on the left-hand side of the figure. Two groups of correlated variables
can be observed. The upper-left corner shows variables related to LF processing (dynamic range, the slope
of the loudness function, and hearing thresholds) and speech intelligibility in quiet. The bottom-right corner
shows a larger group of correlated variables including HF processing, speech intelligibility in noise, and
spectro-temporal resolution at high frequencies. The variables that are not significantly interrelated are
shown in the middle part of Figure 3, including the three variables related to binaural processing abilities
(IPD_{fmax}, BP20 and BMR) which were not significantly correlated to each other. The speech reception
threshold in quiet (SRT_Q) and the STM detection were correlated to various variables such as tone-in-noise
detection, HF spectro-temporal resolution, LF hearing thresholds and speech-in-noise perception.
502

The four factors resulting from the factor analysis showed 63% of explained cumulative variance. The
variables with higher loadings (> 0.65) for each of the factors are shown in Table 11. The first factor, in
terms of the amount of variance explained (19%), was associated with LF loudness perception and speech
intelligibility in quiet, whereas the second factor (18% of variance explained) was associated with HF
loudness perception. Despite loudness perception being associated with the first and second factor, the
MCL was associated, both at high and low frequencies, with the third factor, while the fourth factor was

---

[1] The BEAR3 available at Zenodo contains an observation labelled '0', which corresponds to the results of one of the examiners and it is not used in the present analysis

Figure 4. Correlation plot of the data set BEAR3. The upper part shows the significantly correlated variables as coloured circles. The lower panel shows the numeric correlation value.

Table 11. Variables correlated to the four latent orthogonal factors resulting from the factor analysis with the method of maximum likelihood (ML). Columns are sorted in terms of the variance explained by each factor.

	ML2(19%)	*ML1(18%)*	*ML3(14%)*	*ML4(12%)*
HTL_LF	*0.93*			
DynR_LF	*-0.9*			
AUD_LF	*0.82*			
Slope_LF	*0.81*			
SRTQ	*0.67*			
DynR_HF		*-0.93*		
Slope_HF		*0.82*		
HTL_HF		*0.79*		
AUD_HF		*0.73*		
MCL_HF			*0.92*	
MCL_LF			*0.85*	
SRT_N				*0.77*
SScore_4dB				*-0.78*

509 associated with speech intelligibility in noise.

12 GENERAL DISCUSSION

The first goal of the present study was to collect data of a heterogeneous population of HI listeners, reflecting
their hearing abilities in different aspects of auditory processing. The current study was motivated by the
need for a new dataset to refine the data-driven approach for auditory profiling. The dataset should contain
a representative population of listeners and outcome measures (Sanchez-Lopez et al., 2018) to allow a
refined definition of the two types of auditory distortions and to identify subgroups of listeners with clinical
relevance. To refine the data-driven auditory profiling, the BEAR3 dataset fulfils all the requirements
discussed in Sanchez-Lopez et al. (2018). Other datasets containing a large number of listeners (e.g.,
Gieseler et al., 2017; Rönnberg et al., 2016) or physiological measures (e.g., Kamerer et al., 2019) could
also be interesting for complementing the auditory profiling beyond auditory perceptual measures.

12.1 Relationships across different aspects of auditory processing

The proposed test battery considers outcomes divided into six dimensions of auditory processing. One
of the objectives of the study was to investigate the interrelations of different dimensions and measures.
The present analysis showed two interesting findings. First, the correlation analysis shows two clusters
of variables related to either low- or high-frequency audiometric thresholds. Speech-in-noise perception
was associated with high-frequency sensitivity loss, temporal, and spectral masking release, whereas
speech-in-quiet was correlated with both low- and high-frequency hearing loss. Several outcomes were not
interrelated, especially the outcomes associated with binaural processing abilities. Second, factor analysis
yielded latent factors related to low- and high-frequency processing, most comfortable level and speech in
noise. Vlaming et al. (2011) showed four dimensions in the factor analysis of the HEARCOM project data
corresponding to high-and low-frequency spectro-temporal processing, MCL and recruitment. In contrast,
the current study showed that the slopes of the loudness growth, both at low and high frequencies, were
not interrelated and contributed to the first and second latent factors. Additionally, the speech-in-noise
test performed in HEARCOM was associated with the low-frequency processing, whereas, in the present
study, speech-in-noise dominates the fourth factor and is significantly correlated with high frequencies. The
reason for this discrepancy might be the use of different types of noise and test procedures in the two studies.

Overall, the data of the present study seem to be dominated by the audiometric profiles, with low-
and high-frequency processing reflecting the main sources of variability in the data. However, binaural
processing abilities, loudness perception and speech-in-noise outcomes showed a greater contribution to
the variability of the supra-threshold measures than spectro-temporal processing outcomes.

12.2 Effects of the participant's cognitive abilities

In this study, only auditory tests were considered. Indeed, some of the tests were quite demanding, which
might have affected the results of listeners with reduced cognitive abilities. Here, only the age could be
indicative of a likely cognitive decline, but it can not confirm or deny this. In terms of age, we observed a
significant effect on the results of the IPD and tone-in-noise tests. However, a thorough analysis on the
effect of age was not carried out with the existing data. A more interesting approach would be to replicate
this study including cognitive tests assessing executive functions, working memory or attention span with
the aim of including a heterogeneous group of listeners with various cognitive abilities. Such a study could
potentially assess both hearing and cognitive abilities towards a feasible test battery that can be included in
the audiological assessment, either in or out of the clinic.

12.3 Towards clinical feasibility of the tests

The test-retest reliability of the test battery was investigated based on the results of a subset of listeners who participated 2-5 months after the first visit. The analysis was based on the ICC and the SEM. Some of the tests, such as IPD_{fmax}, binaural pitch and eAUD-HF (FLFT) showed good to excellent test-retest reliability with all ICC values above 0.9, while other tests, such as the extended audiometry in noise and speech intelligibility in quiet, showed poor reliability. The selected tests were conducted in two sessions and the total time was, on average, three hours including the instructions and interview. In realistic clinical setups, a subset of tests with high reliability and a reasonably low difficulty would need to be prioritized. For a clinical version of the test battery, other tracking procedures such as Bayesian Functional information (Remus and Collins, 2008) might be adopted to improve the reliability and time-efficiency in some tasks such as STM and tone detection in noise. Moreover, if time-efficiency is crucial, testing some aspects of auditory processing out of the clinic, as other proposed test batteries for auditory research (Lelo De Larrea-Mancera et al., 2020), might be a solution for completing the patient's hearing profile. The use of speech-in-noise tests can be a useful tool for the characterization of the listener's hearing deficits that can be performed under different conditions, including monaural, binaural, unaided and aided stimuli presentations. While here the tests were performed monaurally and unaided, a binaural condition as well as at least one aided measure (i.e., with hearing aids), could also be included in clinical practice. A clinical test battery with the subset of tests that showed a good or excellent test-retest reliability should be evaluated in a large scale study. In this study, we explored the use of an extended audiometry using the same test procedure for assessing high-frequency audibility (eAUD-HF), tone-in-noise detection (eAUD-N), spectro-temporal resolution (eAUD-S and eAUD-T), and binaural processing abilities. This procedure can be further explored and be performed by a hearing-care professional rather than in the current experimental setup. However, if the goal is to accurately estimate the hearing deficits of the patient, the test battery should include several aspects of auditory processing and provide detailed information on the supra-threshold deficits of the patient. The tests that showed potential for the clinical implementation were ACALOS, HINT, fSTM (only the LF condition), Binaural Pitch and IPD_{fmax}. Such a test battery could serve to identify a clinically relevant subset of patients (auditory profiles) that may benefit from specific types of hearing rehabilitation towards a "stratified approach" (Lonergan et al., 2017) for audiology practice.

13 CONCLUSION

The current study has shown the rationale behind the BEAR test battery and the selected tests for characterizing hearing deficits in listeners with various hearing abilities. The analysis of the data showed that a reduced BEAR test battery has the potential for clinical implementation, providing relevant and reliable information reflecting several auditory domains. The proposed test battery showed good reliability, was reasonably time-efficient and easy to perform. The implementation of a clinical version of the test battery is to be evaluated in future research, e.g., in a larger field study to further refine the auditory profiling approach. Moreover, the current data has been already analyzed for the purpose of auditory profiling (Sanchez Lopez et al., 2020), showing the potential of this test battery for hearing rehabilitation.

14 ETHICS STATEMENT

All the patients received written informed consent. The study was approved by the Science-Ethics Committee for the Capital Region of Denmark H-16036391.

CONFLICT OF INTEREST STATEMENT

590 The authors declare that the research was conducted in the absence of any commercial or financial
591 relationships that could be construed as a potential conflict of interest.

AUTHOR CONTRIBUTIONS

Author contributions according to CRediT (Contributor Roles Taxonomy). RSL: conceptualization,
methodology, software, validation, formal analysis, investigation, data curation, visualization and writing-
original draft and editing. SGN: investigation, validation, resources, formal analysis, visualization, writing-
Original draft. ME: methodology, investigation, resources, writing-review. MF and FB: conceptualization,
methodology, supervision and writing-editing. MW and OC: investigation, resources and writing-review.
TN: methodology, supervision, project administration and writing-review. SS and TD: conceptualization,
methodology, supervision, project administration and funding acquisition. All authors contributed to the
article and approved the submitted version.

FUNDING

600 This work was supported by Innovation Fund Denmark Grand Solutions 5164-00011B (Better hEAring
601 Rehabilitation project), Oticon, GN Resound, Widex and other partners (Aalborg University, University
602 of Southern Denmark, the Technical University of Denmark, Force, Aalborg, Odense and Copenhagen
603 University Hospitals).

ACKNOWLEDGMENTS

We thank the staff from OUH, BBH and HEA, especially JH Schmidt, SS Houmøller, E Kjærbøl, RS
Sørensen, and the student helpers from the MSc of Audiology at SDU. We also want to show our gratitude
to all the participants in the study. We want to thank I. Holube and two annonymous reviewers for their
comments in an earlier version of this manuscript, as well as SD Ewert for the AFC framework. This work
was supported by Innovation Fund Denmark Grand Solutions 5164-00011B (Better hEAring Rehabilitation
project), Oticon, GN Resound, Widex and other partners (Aalborg University, University of Southern
Denmark, the Technical University of Denmark, Force technology, Aalborg, Odense and Copenhagen
University Hospitals).The funding and collaboration of all BEAR partners are sincerely acknowledged.

DATA AVAILABILITY STATEMENT

612 The datasets generated from this study can be found in the Zenodo repository: `https://www.doi.org/`
613 `10.5281/zenodo.3459579` A working version of the Test Battery can be shared upon a reasonable
614 request to the corresponding author. This study is part of the PhD thesis entitled "Clinical auditory profiling
615 and profile-based hearing-aid fitting" and available as the volume 43 of the Contributions to Hearing
616 Research at the Technical University of Denmark.

REFERENCES

Anderson, M. C., Arehart, K. H., and Souza, P. E. (2018). engSurvey of Current Practice in the Fitting
and Fine-Tuning of Common Signal-Processing Features in Hearing Aids for Adults. *Journal of the*
American Academy of Audiology 29, 118–124. doi:10.3766/jaaa.16107
Bernstein, J. G., Mehraei, G., Shamma, S., Gallun, F. J., Theodoroff, S. M., and Leek, M. R. (2013).
Spectrotemporal modulation sensitivity as a predictor of speech intelligibility for hearing-impaired
listeners. *Journal of the American Academy of Audiology* 24, 293–306. doi:10.3766/jaaa.24.4.5

Bernstein, J. G. W., Danielsson, H., Hällgren, M., Stenfelt, S., Rönnberg, J., and Lunner, T. (2016). Spectrotemporal modulation sensitivity as a predictor of speech-reception performance in noise with hearing aids. *Trends in hearing* 20, 1–17. doi:10.1177/2331216516670387

Bianchi, F., Carney, L. H., Dau, T., and Santurette, S. (2019). Effects of musical training and hearing loss on fundamental frequency discrimination and temporal fine structure processing: Psychophysics and modeling. *Journal of the Association for Research in Otolaryngology* 20, 263–277. doi:10.1007/s10162-018-00710-2

Bisgaard, N., Vlaming, M. S. M. G., and Dahlquist, M. (2010). Standard audiograms for the iec 60118-15 measurement procedure. *Trends in Amplification* 14, 113–120. doi:10.1177/1084713810379609

Brand, T. and Hohmann, V. (2002). An adaptive procedure for categorical loudness scaling. *The Journal of the Acoustical Society of America* 112, 1597–1604. doi:10.1121/1.1502902

Brown, M. and Musiek, F. (2013). Pathways: The fundamentals of masking level differences for assessing auditory function. *The Hearing Journal* 66, 16. doi:10.1097/01.HJ.0000425772.41884.1d

Brungart, D. S., Sheffield, B. M., and Kubli, L. R. (2014). Development of a test battery for evaluating speech perception in complex listening environments. *The Journal of the Acoustical Society of America* 136, 777–790. doi:10.1121/1.4887440

Choi, J. E., Hong, S. H., Won, J. H., Park, H.-S., Cho, Y. S., Chung, W.-H., et al. (2016). Evaluation of cochlear implant candidates using a non-linguistic spectrotemporal modulation detection test. *Scientific Reports* 6, 35235. doi:10.1038/srep35235

Cramer, E. M. and Huggins, W. H. (1958). Creation of pitch through binaural interaction. *The Journal of the Acoustical Society of America* 30, 413–417. doi:10.1121/1.1909628

Darwin, C. J. (1997). Auditory grouping. *Trends in Cognitive Sciences* 1, 327–333. doi:10.1016/S1364-6613(97)01097-8

Desloge, J. G., Reed, C. M., Braida, L. D., Perez, Z. D., and D'Aquila, L. A. (2017). Masking release for hearing-impaired listeners: The effect of increased audibility through reduction of amplitude variability. *The Journal of the Acoustical Society of America* 141, 4452–4465. doi:10.1121/1.4985186

Durlach, N. I. (1963). Equalization and cancellation theory of binaural masking-level differences. *The Journal of the Acoustical Society of America* 35, 1206–1218. doi:10.1121/1.1918675

Elberling, C., Ludvigsen, C., and Lyregaard, P. E. (1989). Dantale: A new danish speech material. *Scandinavian Audiology* 18, 169–175. doi:10.3109/01050398909070742

Encina-Llamas, G., Harte, J., Dau, T., Shinn-Cunningham, B., and Epp, B. (2019). engInvestigating the effect of cochlear synaptopathy on envelope following responses using a model of the auditory nerve. *J a R O* 20, 363–382. doi:10.1007/s10162-019-00721-7

Eustaquio-Martín, A. and Lopez-Poveda, E. A. (2011). engIsoresponse Versus Isoinput Estimates of Cochlear Filter Tuning. *Journal of the Association for Research in Otolaryngology* 12, 281–299. doi:10.1007/s10162-010-0252-1

Ewert, S. (2013). Afc - a modular framework for running psychoacoustic experiments and computational perception models. In *Proceedings of the International Conference on Acoustics AIA-DAGA 2013* (Merano, Italy), 1326–1329

Festen, J. M. and Plomp, R. (1990). Effects of fluctuating noise and interfering speech on the speech-reception threshold for impaired and normal hearing. *The Journal of the Acoustical Society of America* 88, 1725–1736. doi:10.1121/1.400247

Füllgrabe, C. and Moore, B. C. J. (2017). Evaluation of a method for determining binaural sensitivity to temporal fine structure (tfs-af test) for older listeners with normal and impaired low-frequency hearing. *Trends in Hearing* 21, 233121651773723. doi:10.1177/2331216517737230

[Dataset] Gelfand, S. A. (2009). Essentials of audiology. doi:10.1097/MAO.0b013e3181c99550

Gieseler, A., Tahden, M. A. S., Thiel, C. M., Wagener, K. C., Meis, M., and Colonius, H. (2017). Auditory
and non-auditory contributions for unaided speech recognition in noise as a function of hearing aid use.
Frontiers in Psychology 8. doi:10.3389/fpsyg.2017.00219

Glasberg, B. R. and Moore, B. C. (1990). Derivation of auditory filter shapes from notched-noise data.
Hearing Research 47, 103–138. doi:10.1016/0378-5955(90)90170-T

Glasberg, B. R. and Moore, B. J. (1989). Psychoacoustic abilities of subjects with unilateral and bilateral
cochlear hearing impairments and their relationship to the ability to understand speech. *Scandinavian*
audiology. Supplementum 32, 1–25

Houmoller, S. S., Wolff, A., Moller, S., Narne, V. K., Narayanan, S. K., Godballe, C., et al. (2021).
engPrediction of successful hearing aid treatment in first-time and experienced hearing aid users:
Using the international outcome inventory for hearing aids. *International Journal of Audiology* ,
1–11doi:10.1080/14992027.2021.1916632

Houtgast, T. and Festen, J. M. (2008). On the auditory and cognitive functions that may explain an
individual's elevation of the speech reception threshold in noise. *International Journal of Audiology* 47,
287–295. doi:10.1080/14992020802127109

ISO 16832 (2006). Acoustics - loudness scaling by means of categories. *International Organization for*
Standardization 2006, 1–12

ISO 8253-1 (2010). Acoustics - audiometric test methods - part 1: Pure-tone air and bone conduction
audiometry. *International Organization for Standardization*

ISO 8253-3 (2012). Acoustics. audiometric test methods - part 3: Speech audiometry. *International*
Organization for Standardization

Johannesen, P. T., Pérez-González, P., Kalluri, S., Blanco, J. L., and Lopez-Poveda, E. A. (2016). The
influence of cochlear mechanical dysfunction, temporal processing deficits, and age on the intelligibility
of audible speech in noise for hearing-impaired listeners. *Trends in Hearing* 20, 233121651664105.
doi:10.1177/2331216516641055

Jürgens, T., Kollmeier, B., Brand, T., and Ewert, S. D. (2011). Assessment of auditory nonlinearity for
listeners with different hearing losses using temporal masking and categorical loudness scaling. *Hearing*
Research 280, 177–191. doi:10.1016/j.heares.2011.05.016

Kaernbach, C. (1990). A single-interval adjustment-matrix (siam) procedure for unbiased adaptive testing.
The Journal of the Acoustical Society of America 88, 2645–2655. doi:10.1121/1.399985

Kamerer, A. M., Kopun, J. G., Fultz, S. E., Neely, S. T., and Rasetshwane, D. M. (2019). Reliability of
measures intended to assess threshold-independent hearing disorders. *Ear and Hearing* 40, 1267–1279.
doi:10.1097/AUD.0000000000000711

Killion, M. C., Niquette, P. A., Gudmundsen, G. I., Revit, L. J., and Banerjee, S. (2004). Development
of a quick speech-in-noise test for measuring signal-to-noise ratio loss in normal-hearing and hearing-
impaired listeners. *The Journal of the Acoustical Society of America* 116, 2395–2405. doi:10.1121/1.
1784440

Koo, T. K. and Li, M. Y. (2016). A guideline of selecting and reporting intraclass correlation coefficients
for reliability research. *Journal of chiropractic medicine* doi:10.1016/j.jcm.2016.02.012

Larsby, B. and Arlinger, S. (1998). A method for evaluating temporal, spectral and combined temporal-
spectral resolution of hearing. *Scandinavian audiology* 27, 3–12. doi:10.1080/010503998419641

Lecluyse, W., Tan, C. M., McFerran, D., and Meddis, R. (2013). Acquisition of auditory profiles for good
and impaired hearing. *International Journal of Audiology* 52, 596–605. doi:10.3109/14992027.2013.
796530

Lelo De Larrea-Mancera, E. S., Stavropoulos, T., Hoover, E. C., Eddins, D. A., Gallun, F. J., and Seitz,
A. R. (2020). engPortable automated rapid testing (part) for auditory assessment: Validation in a
young adult normal-hearing population. *Journal of the Acoustical Society of America* 148, 1831–1851.
doi:10.1121/10.0002108

Liberman, M. C., Epstein, M. J., Cleveland, S. S., Wang, H., and Maison, S. F. (2016). Toward a differential
diagnosis of hidden hearing loss in humans. *PLOS ONE* 11, e0162726. doi:10.1371/journal.pone.
0162726

[Dataset] Lonergan, M., Senn, S. J., McNamee, C., Daly, A. K., Sutton, R., Hattersley, A., et al. (2017).
Defining drug response for stratified medicine. doi:10.1016/j.drudis.2016.10.016

Mehraei, G., Gallun, F. J., Leek, M. R., and Bernstein, J. G. W. (2014). Spectrotemporal modulation
sensitivity for hearing-impaired listeners: Dependence on carrier center frequency and the relationship to
speech intelligibility. *The Journal of the Acoustical Society of America* 136, 301–316. doi:10.1121/1.
4881918

Moore, B. C., Huss, M., Vickers, D. A., Glasberg, B. R., and Alcantara, J. I. (2000). engA test for the
diagnosis of dead regions in the cochlea. *British Journal of Audiology* 34, 205–224. doi:10.3109/
03005364000000131

Moore, B. C. J. (2001). Dead regions in the cochlea: Diagnosis, perceptual consequences, and implications
for the fitting of hearing aids. *Trends in Amplification* 5, 1–34. doi:10.1177/108471380100500102

Moore, B. C. J. (2007). *Cochlear Hearing Loss* (West Sussex, England: John Wiley & Sons, Ltd).
doi:10.1002/9780470987889

Neher, T. (2017). Characterizing the binaural contribution to speech-in-noise reception in elderly hearing-
impaired listeners. *The Journal of the Acoustical Society of America* doi:10.1121/1.4976327

Neher, T., Laugesen, S., Søgaard Jensen, N., and Kragelund, L. (2011). Can basic auditory and cognitive
measures predict hearing-impaired listeners' localization and spatial speech recognition abilities? *The
Journal of the Acoustical Society of America* 130, 1542–1558. doi:10.1121/1.3608122

Nielsen, J. B. and Dau, T. (2011). The danish hearing in noise test. *International Journal of Audiology* 50,
202–208. doi:10.3109/14992027.2010.524254

Nilsson, M., Soli, S. D., and Sullivan, J. a. (1994). Development of the hearing in noise test for the
measurement of speech reception thresholds in quiet and in noise. *The Journal of the Acoustical Society
of America* 95, 1085–1099. doi:10.1121/1.408469

Oetting, D., Brand, T., and Ewert, S. D. (2014). Optimized loudness-function estimation for categorical
loudness scaling data. *Hearing Research* 316, 16–27. doi:10.1016/j.heares.2014.07.003

Oetting, D., Hohmann, V., Appell, J. E., Kollmeier, B., and Ewert, S. D. (2018). Restoring perceived
loudness for listeners with hearing loss. *Ear and hearing* doi:10.1097/AUD.0000000000000521

Oxenham, A. J. and Simonson, A. M. (2009). Masking release for low- and high-pass-filtered speech in
the presence of noise and single-talker interference. *The Journal of the Acoustical Society of America*
125, 457–468. doi:10.1121/1.3021299

Plomp, R. (1978). Auditory handicap of hearing impairment and the limited benefit of hearing aids. *The
Journal of the Acoustical Society of America* 63, 533–549. doi:10.1121/1.381753

Plomp, R. and Mimpen, A. M. (1979). Speech-reception threshold for sentences as a function of age and
noise level. *The Journal of the Acoustical Society of America* 66, 1333–42. doi:10.1121/1.383554

Prendergast, G., Hymers, M., and Lee, A. (2020). engA quick and reliable estimate of extended high-
frequency hearing. *International Journal of Audiology* 59, 823–827. doi:10.1080/14992027.2020.
1767810

Remus, J. J. and Collins, L. M. (2008). Comparison of adaptive psychometric procedures motivated by the theory of optimal experiments: Simulated and experimental results. *The Journal of the Acoustical Society of America* 123, 315–326. doi:10.1121/1.2816567

Rhebergen, K. S., Versfeld, N. J., and Dreschler, W. a. (2006). Extended speech intelligibility index for the prediction of the speech reception threshold in fluctuating noise. *The Journal of the Acoustical Society of America* 120, 3988–3997. doi:10.1121/1.2358008

Rieke, C. C., Clavier, O. H., Allen, L. V., Anderson, A. P., Brooks, C. A., Fellows, A. M., et al. (2017). Fixed-Level Frequency Threshold Testing for Ototoxicity Monitoring. *Ear and Hearing* 38, e369–e375. doi:10.1097/AUD.0000000000000433

Rönnberg, J., Lunner, T., Ng, E. H. N., Lidestam, B., Zekveld, A. A., Sörqvist, P., et al. (2016). Hearing impairment, cognition and speech understanding: exploratory factor analyses of a comprehensive test battery for a group of hearing aid users, the n200 study. *International Journal of Audiology* 55, 623–642. doi:10.1080/14992027.2016.1219775

Rosowski, J. J., Stenfelt, S., and Lilly, D. (2013). An overview of wideband immittance measurements techniques and terminology. *Ear and Hearing* 34, 9s–16s. doi:10.1097/AUD.0b013e31829d5a14

Ross, B., Tremblay, K. L., and Picton, T. W. (2007). Physiological detection of interaural phase differences. *The Journal of the Acoustical Society of America* 121, 1017–1027. doi:10.1121/1.2404915

Rønne, F. M., Laugesen, S., and Jensen, N. S. (2017). engSelection of test-setup parameters to target specific signal-to-noise regions in speech-on-speech intelligibility testing. *International Journal of Audiology* 56, 559–567. doi:10.1080/14992027.2017.1300349

Sanchez-Lopez, R., Bianchi, F., Fereczkowski, M., Santurette, S., and Dau, T. (2018). Data-driven approach for auditory profiling and characterization of individual hearing loss. *Trends in Hearing* 22, 233121651880740. doi:10.1177/2331216518807400

Sanchez Lopez, R., Fereczkowski, M., Neher, T., Santurette, S., and Dau, T. (2020). engRobust data-driven auditory profiling towards precision audiology. *Trends in Hearing* 24, 2331216520973539. doi:10.1177/2331216520973539

[Dataset] Sanchez-Lopez, R., Nielsen, S., El-Haj-Ali, M., Bianchi, F., Fereckzowski, M., Cañete, O., et al. (2019). Data from "auditory tests for characterizing hearing deficits: The bear test battery". doi:10.5281/zenodo.3459579

Sanders, N. C. and Chin, S. B. (2009). Phonological distance measures. *Journal of Quantitative Linguistics* 16, 96–114. doi:10.1080/09296170802514138

Santurette, S. and Dau, T. (2012). Relating binaural pitch perception to the individual listener's auditory profile. *The Journal of the Acoustical Society of America* 131, 2968–2986. doi:10.1121/1.3689554

Saunders, G. H., Field, D. L., and Haggard, M. P. (1992). A clinical test battery for obscure auditory dysfunction (oad): Development, selection and use of tests. *British Journal of Audiology* 26, 33–42. doi:10.3109/03005369209077869

Schorn, K. and Zwicker, E. (1990). Frequency selectivity and temporal resolution in patients with various inner ear disorders. *Audiology : official organ of the International Society of Audiology* 29, 8–20. doi:10.3109/00206099009081641

Schütt, H. H., Harmeling, S., Macke, J. H., and Wichmann, F. A. (2016). Painfree and accurate bayesian estimation of psychometric functions for (potentially) overdispersed data. *Vision Research* 122, 105–123. doi:10.1016/j.visres.2016.02.002

Schädler, M. R., Hülsmeier, D., Warzybok, A., and Kollmeier, B. (2020). engIndividual aided speech-recognition performance and predictions of benefit for listeners with impaired hearing employing fade. *Trends in Hearing* 24, 2331216520938929. doi:10.1177/2331216520938929

Stratford, P. W. and Goldsmith, C. H. (1997). Use of the standard error as a reliability index of interest :
An applied example using elbow flexor strength data. *Physical Therapy* 77, 745–750. doi:10.1093/ptj/
77.7.745

Strelcyk, O. and Dau, T. (2009). Relations between frequency selectivity, temporal fine-structure processing,
and speech reception in impaired hearing. *The Journal of the Acoustical Society of America* 125, 3328.
doi:10.1121/1.3097469

Strelcyk, O., Zahorik, P., Shehorn, J., Patro, C., and Derleth, R. P. (2019). Sensitivity to interaural phase
in older hearing-impaired listeners correlates with nonauditory trail making scores and with a spatial
auditory task of unrelated peripheral origin. *Trends in Hearing* 23, 233121651986449. doi:10.1177/
2331216519864499

Tai, Y. and Husain, F. T. (2018). engRight-ear advantage for speech-in-noise recognition in patients with
nonlateralized tinnitus and normal hearing sensitivity. *Jaro - Journal of the Association for Research in
Otolaryngology* 19, 211–221. doi:10.1007/s10162-017-0647-3

Thorup, N., Santurette, S., Jørgensen, S., Kjærbøl, E., Dau, T., and Friis, M. (2016). Auditory profiling and
hearing-aid satisfaction in hearing-aid candidates. *Danish medical journal* 63, 1–5

van Esch, T. E. M. and Dreschler, W. A. (2011). Measuring spectral and temporal resolution simultaneously:
a comparison between two tests. *International journal of audiology* 50, 477–90. doi:10.3109/14992027.
2011.572083

Vermiglio, A. J., Soli, S. D., Freed, D. J., and Fang, X. (2020). engThe effect of stimulus audibility on the
relationship between pure-tone average and speech recognition in noise ability. *Journal of the American
Academy of Audiology* 31, 224–232. doi:10.3766/jaaa.19031

Vinay, Hansen, A. S., Raen, Ø., and Moore, B. C. J. (2017). Reference thresholds for the ten(hl) test for
people with normal hearing. *International Journal of Audiology* 56, 672–676. doi:10.1080/14992027.
2017.1307531

Vlaming, M. S. M. G., Kollmeier, B., Dreschler, W. A., Martin, R., Wouters, J., Grover, B., et al. (2011).
Hearcom: Hearing in the communication society. *Acta Acustica united with Acustica* 97, 175–192.
doi:10.3813/AAA.918397

Wagener, K., Josvassen, J. L., and Ardenkjær, R. (2003). engDesign, optimization and evaluation of a danish
sentence test in noise. *International Journal of Audiology* 42, 10–17. doi:10.3109/14992020309056080

Zaar, J., Simonsen, L., Behrens, T., and Laugesen, S. (2018). Towards a clinically viable spectro-
temporal modulation test. International Hearing Aid Conference 2018, IHCON 2018 ; Conference date:
15-08-2018 Through 19-08-2018

Zaar, J., Simonsen, L. B., Behrens, T., Dau, T., and Laugesen, S. (2020). Investigating the relationship
between spectro-temporal modulation detection, aided speech perception, and directional noise reduction
preference in hearing-impaired listeners. *Proceedings of the International Symposium on Auditory and
Audiological Research* 7, 181–188