**Rapid Detection of 2019 Novel Coronavirus SARS-CoV-2 Using a CRISPR-based DETECTR Lateral Flow Assay**

James P. Broughton[1,&], Xianding Deng[2,3,&], Guixia Yu[2,3], Clare L. Fasching[1], Jasmeet Singh[1], Jessica Streithorst[2], Andrea Granados[2,3], Alicia Sotomayor-Gonzalez[2,3], Kelsey Zorn[4], Allan Gopez[2], Elaine Hsu[2], Wei Gu[2], Steve Miller[2], Chao-Yang Pan[5], Hugo Guevara[5], Debra A. Wadford[5], Janice S. Chen[1,*], and Charles Y. Chiu[2,3,6,*]

[1]Mammoth Biosciences, Inc., San Francisco, California, USA

[2]Department of Laboratory Medicine, University of California, San Francisco, California, USA

[3]UCSF-Abbott Viral Diagnostics and Discovery Center, San Francisco, California, USA

[4]Department of Biochemistry and Biophysics, San Francisco, California, USA

[5]Viral and Rickettsial Disease Laboratory, California Department of Public Health, Richmond, CA, USA

[6]Department of Medicine, Division of Infectious Diseases, University of California, San Francisco, California, USA

[&]These authors contributed equally.

*Co-corresponding authors:

Charles Chiu, 185 Berry Street, Box #0134, San Francisco, CA 94107, Office (415) 514-8219, Fax (415) 889-6441, e-mail: charles.chiu@ucsf.edu

Janice Chen, Mammoth Biosciences, Inc., 279 E. Grand Ave. Suite #400, South San

Francisco, CA 94080, Phone (650) 294-8583, e-mail: janice@mammoth.bio

Short title: CRISPR-Cas12 detection of SARS-CoV-2

**KEYWORDS:** coronavirus, COVID-19, 2019-nCoV, SARS-CoV-2, betacoronavirus,

outbreak, Wuhan, CRISPR, CRISPR-Cas12, DETECTR, loop-mediated isothermal

amplification (LAMP), isothermal amplification, lateral flow, zoonotic, outbreak,

epidemic, pandemic, diagnostic testing, molecular testing

**ABSTRACT**

An outbreak of novel betacoronavirus, SARS-CoV-2 (formerly named 2019-

nCoV), began in Wuhan, China in December 2019 and the COVID-19 disease

associated with infection has since spread rapidly to multiple countries. Here we report

the development of SARS-CoV-2 DETECTR, a rapid (~30 min), low-cost, and accurate

CRISPR-Cas12 based lateral flow assay for detection of SARS-CoV-2. We validated

this method using contrived reference samples and clinical samples from infected US

patients and demonstrated comparable performance to the US CDC SARS-CoV-2 real-

time RT-PCR assay.

Over the past 40 years, there have been recurrent large-scale epidemics from

novel emerging viruses, including human immunodeficiency virus (HIV), SARS and

MERS coronaviruses, 2009 pandemic influenza H1N1 virus, Ebola virus (EBOV), Zika

virus (ZIKV), and most recently SARS-CoV-2[1,2]. All of these epidemics presumably

resulted from an initial zoonotic animal-to-human transmission event, with either

clinically apparent or occult spread into vulnerable human populations. Each time, a

lack of rapid, accessible, and accurate molecular diagnostic testing has hindered the

public health response to the emerging viral threat.

In early January 2020, a cluster of cases of pneumonia from a novel coronavirus,

SARS-CoV-2 (with the disease referred to as COVID-19), was reported in Wuhan,

China[1,2]. This outbreak has spread rapidly, with over 90,000 reported cases and 3,000

deaths as of March 4th, 2020[3]. Person-to-person transmission from infected individuals

with no or mild symptoms has been reported[4,5]. Assays using quantitative reverse

transcription-polymerase chain reaction (qRT-PCR) approaches for detection of the

virus in 4-6 hours have been developed by several laboratories, including an

Emergency Use Authorization (EUA)-approved assay developed by the US CDC[6].

However, the typical turnaround time for screening and diagnosing patients with

suspected SARS-CoV-2 has been >24 hours given the need to ship samples overnight

to reference laboratories. To accelerate clinical diagnostic testing for COVID-19 in the

United States, the FDA on February 28th, 2020 permitted individual clinically licensed

laboratories to report the results of in-house developed SARS-CoV-2 diagnostic assays

while awaiting results of an EUA submission for approval[7].

Here we report the development and initial validation of a CRISPR (clustered

regularly interspaced short palindromic repeats)-Cas12 based assay[8-11] for detection of

SARS-CoV-2 from extracted patient sample RNA in approximately 30 minutes, called

SARS-CoV-2 DETECTR. This assay performs simultaneous reverse transcription and

isothermal amplification using loop-mediated amplification (RT-LAMP)[12] from RNA

extracted from nasopharyngeal or oropharyngeal swabs in universal transport media

(UTM), followed by Cas12 detection of predefined coronavirus sequences, after which

cleavage of a reporter molecule confirms detection of the virus. We first designed

primers targeting the E (envelope) and N (nucleoprotein) genes of SARS-CoV-2 (**Fig.**

**1a**). The primers amplify regions that overlap the WHO assay (E gene region) and US

CDC assay (N2 region in the N gene)[6,13], but are modified to meet design requirements

for LAMP. We did not target the N1 and N3 regions used by the US CDC assay, as

these regions lacked suitable protospacer adjacent motif (PAM) sites for the Cas12

guide RNAs (gRNAs). Next, we designed Cas12 gRNAs to detect three SARS-like

coronaviruses (SARS-CoV-2 accession NC_045512, bat SARS-like coronavirus (bat-

SL-CoVZC45, accession MG772933), and SARS-CoV, accession NC_004718)) in the E

gene and specifically detect SARS-CoV-2 only in the N gene **(Supplementary Fig. 1).**

Using synthetic, *in vitro* transcribed (IVT) SARS-CoV-2 RNA gene targets in

nuclease-free water, we demonstrated that the CRISPR-Cas12 based detection can

distinguish SARS-CoV-2 with no cross-reactivity for related coronavirus strains **(Fig. 1b,**

**Supplementary Fig. 2)**. We then optimized the conditions for the SARS-CoV-2

DETECTR assay on the E gene, N gene and the human RNase P gene as a control,

which consists of an RT-LAMP reaction at 62°C for 20 min and Cas12 detection

reaction at 37°C for 10 min. The DETECTR assay can be run in approximately 30

minutes and visualized on a lateral flow strip **(Fig. 1c, d).** The SARS-CoV-2 DETECTR

assay requires detection of both the E and N genes to confirm a positive test **(Fig. 1e),**

and interpretation is consistent with that for the CDC assay N1 and N2 genes (the N3

gene target region for the CDC assay is no longer being used due to concerns

regarding flaws in manufacturing reagents and potential decreased sensitivity)[14].

We next compared the analytic limits of detection (LoD) of the RT-LAMP/Cas12

DETECTR assay relative to the US FDA Emergency Use Authorization (EUA)-approved

CDC assay for detection of SARS-CoV-2 **(Table 1; Fig. 2d).** A standard curve for

quantitation was constructed using 7 dilutions of a control IVT viral nucleoprotein RNA

("CDC VTC nCoV Transcript")[6], with 3 replicates at each dilution (**Fig. 2d, left;**

**Extended Data 1).** Ten two-fold serial dilutions of the same control nucleoprotein RNA

were then used to run the DETECTR assay, with 6 replicates at each dilution (**Fig. 2d,**

**right; Supplementary Fig. 3).** The estimated LoD for the CDC assay tested by

California Department of Public Health was 1 copy/µL reaction, consistent with the

analytic performance in the FDA package insert, versus 10 copies/µL reaction for the

DETECTR assay.

We then assessed the capability of the RT-LAMP assay to amplify SARS-CoV-2

nucleic acid directly from raw sample matrix consisting of nasopharyngeal swabs from

asymptomatic donors placed in universal transport medium (UTM) or phosphate

buffered saline (PBS) and spiked with SARS-CoV-2 IVT target RNA. Assay

performance was degraded at reaction concentration of ≥10% UTM and ≥20% PBS by

volume, with estimated limits of detection decreasing to 500 and 1,500 copies/µL,

respectively (**Supplementary Fig. 4).**

Finally, we tested extracted RNA from 11 respiratory swab samples collected

from 6  PCR-positive COVID-19 patients (COVID19-1A/B to COVID19-5A/B, where

A=nasopharyngeal swab and B=oropharyngeal swab and COVID19-6, a single

nasopharyngeal swab) and 12 nasopharyngeal swab samples from patients with

influenza (n=4), common human seasonal coronavirus infections (n=3, representing

OC43, HKU1, NL63), and healthy donors (n=5) **(Fig. 2e, f; Supplementary Fig. 5).**

Relative to the CDC qRT-PCR, SARS-CoV-2 DETECTR was 90% sensitive and 100%

specific for detection of the coronavirus in respiratory swab samples, corresponding to

positive and negative predictive values of 100% and 91.7%, respectively **(Fig. 2g).**

Here we combined isothermal amplification with CRISPR-Cas12 DETECTR

technology to develop a rapid (~30 min) and low-cost test for detection of SARS-CoV-2

in clinical samples. The use of existing qRT-PCR based assays is hindered by the need

for expensive lab instrumentation, and availability is currently restricted to public health

laboratories. Importantly, the DETECTR assays developed here have comparable

accuracy to qRT-PCR and are broadly accessible, as they use routine protocols and

commercially available, "off-the-shelf" reagents. Key advantages of our approach over

existing methods such as qRT-PCR include (1) isothermal signal amplification for rapid

target detection obviating the need for thermocycling, (2) single nucleotide target

specificity (guide RNAs at the N2 site can distinguish SARS-CoV-2 from SARS-CoV

and MERS-CoV), (3) integration with portable, low-cost reporting formats such as lateral

flow strips, and (4) quick development cycle to address emerging threats from novel

zoonotic viruses (<2 weeks for SARS-CoV-2, **Supplementary Fig. 6**).

Although most of the cases of COVID-19 infection during the first month of the

epidemic were traced to the city of Wuhan and Hubei province in China, the ongoing

rise in cases now appears to be driven to local community transmission[15,16]. For a

number of reasons, there is an urgent public health need for rapid diagnostic tests for

SARS-CoV-2 infection. The documented cases of asymptomatic infection and

transmission in COVID-19 patients[4,5] greatly expand the pool of individuals who need to

be screened. Viral titers in hospitalized patients can fluctuate day-to-day with lack of

correlation to disease severity[17-19], and thus a single negative qRT-PCR test for SARS-

CoV-2 does not exclude infection. The virus has also been shown to be shed in stool[20],

raising the possibility of environmental contamination contributing to local disease

outbreaks. Low testing platforms such as the DETECTR CRISPR-Cas12 assay

developed here may be useful for periodic repeat testing of patient samples. Clinical

validation of this assay in response to recent draft guidance from the US FDA[7] is

currently ongoing in a CLIA (Clinical Laboratory Improvement Amendments)-certified

microbiology laboratory.

The major pandemics and large-scale epidemics of the past half century have all

been caused by zoonotic viruses. Despite these recurrent outbreaks, we still do not

have a programmable point of care (POC) diagnostic platform that can be used to

promptly address any emerging viral threat. The CRISPR-based DETECTR technology

provides such a platform, which we have reconfigured within days to detect SARS-CoV-

2 (**Supplementary Fig. 6**). Here we developed a SARS-CoV-2 DETECTR assay,

described its performance characteristics and demonstrated compatibility with lateral

flow strips. The future development of portable microfluidic-based cartridges to run the

assay and use of lyophilized reagents will enable POC testing outside of the clinical

diagnostic laboratory, such as airports, local emergency departments and clinics, and

other decentralized locations.

**(~1343 words, max 2000 words)**

 **FIGURE LEGENDS**

**Figure 1. A CRISPR-Cas12 based assay for detection of SARS-CoV-2. (a) Genome**

**map showing primers, probes and gRNAs.** Visualization of primers and probes on

the SARS-CoV-2 genome **(b) gRNA specificity.** Cas12 gRNAs are programmed to

specifically target SARS-CoV-2 or broadly detect related coronavirus strains. The N

gene gRNA used in the assay (left) is specific for SARS-CoV-2, whereas the E gene

gRNA is able to detect 3 SARS-like coronavirus (right). A separate N gene gRNA

targeting SARS-CoV and a bat coronavirus and differing by a single nucleotide from the

N gene gRNA used in the assay fails to detect SARS-CoV-2 (middle). **(c) Minimum**

**equipment needed to run protocol.** With appropriate biosafety level 2 requirements,

the minimum equipment required to run the protocol includes Eppendorf tubes with

reagents, heat blocks or water bath (37°C and 62°C), nuclease-free water, pipettes and

tips, and lateral flow strips. **(d) Schematic of SARS-CoV-2 DETECTR workflow.**

Conventional RNA extraction or sample matrix can be used as an input to DETECTR

(LAMP pre-amplification and Cas12-based detection for E gene, N gene and RNase P),

which is visualized by a fluorescent reader or lateral flow strip. **(e) Lateral flow strip**

**assay readout.** A positive result requires detection of at least the two SARS-CoV-2 viral

gene targets (N gene and E gene).

**Figure 2. Detection of SARS-CoV-2 in contrived and clinical nasopharyngeal or oropharyngeal swab samples. (a) Schematic of DETECTR coupled with lateral flow readout.** The intact FAM-biotinylated reporter molecule flows to the control capture line. Upon recognition of the matching target, the Cas-gRNA complex cleaves the reporter molecule, which flows to the target capture line. **(b-c) Comparison of fluorescence to lateral flow. (b)** Fluorescence signal of LbCas12a detection assay on RT-LAMP amplicon for SARS-CoV-2 N-gene saturates within 10 minutes. RT-LAMP amplicon generated from 2 µL of 5 fM or 0 fM SARS-CoV-2 N-gene IVT RNA by amplifying at 62°C for 20 minutes. **(c)** LbCas12a on the same RT-LAMP amplicon produces visible signal through lateral flow assay within 5 minutes. **(d) Limit of detection for CDC qPCR and DETECTR.** Ct values using the CDC qPCR assay (n=3) and fluorescence values using SARS-CoV-2 DETECTR (n=6) using SARS-CoV-2 N2 gene IVT-RNA. **(e) Patient sample DETECTR data.** DETECTR fluorescence values were normalized to the highest value within the N gene, E gene or RNase P set, with a positive threshold at five standard deviations above background. Final determination of the SARS-CoV-2 test was based on the interpretation matrix in Fig. 1e, with results indicated above the heat map. **(f) SARS-CoV-2 DETECTR assay identifies presence of SARS-CoV-2 viral RNA from clinical sample.** Two replicate assays were performed using 2 µL of extracted RNA for each reaction (titer 12 copies/µL). Positive controls used IVT RNA for SARS-CoV-2 targets and total human RNA for RNase P. LbCas12a detection assays were run on lateral flow strips (TwistDx) and imaged after 3 minutes. **(g) Performance characteristics of the SARS-CoV-2 DETECTR assay.**

204 Abbreviations: fM, femtomolar; NTC, no-template control; PPV, positive predictive

205 value; NPV, negative predictive value.

206 **Table 1. Comparison of the DETECTR (RT-LAMP/Cas12) assay with the CDC qRT-**

207 **PCR assay for detection of SARS-CoV-2**

| | SARS-CoV-2 DETECTR, RT-LAMP/Cas12 | CDC SARS-CoV-2 qRT-PCR |
|---|---|---|
| Target | E gene & N gene* | N-gene (3 amplicons, N1, N2, and N3) |
| Sample control | RNase P | RNase P |
| Limit of Detection | 10 copies/µL input | 3.2 copies/µL input |
| Assay reaction time (approximate) | 40 min | 120 min |
| Assay sample-to-result time (approximate) | 45 min (with manual RNA extraction) | 4 hr (including RNA extraction) |
| Assay components | RT-LAMP (62°C, 30 min)<br>Cas12 (37°C, 10 min)<br>Lateral flow (RT, 2 min; no additional time if using fluorescence readout) | UDG digestion (25°C, 2 min), reverse transcription (50°C, 15 min), denature (95°C, 2 min) amplification, (95°C, 3 sec; 55°C 30 sec; 45 cycles) |
| Bulky instrumentation required | No | Yes |
| FDA EUA approval | No | Yes |

210 *E gene primers target same amplicon region as in the WHO protocol; N gene primers

211 target same N2 amplicon region as in the CDC protocol

## METHODS

*Nucleic acid preparation*

SARS-CoV-2 target sequences were designed using all available genomes

available from GISAID[21] as of January 27, 2020. Briefly, viral genomes were aligned

using Clustal Omega. Next, LbCas12a target sites on the SARS-CoV-2 genome were

filtered against SARS-CoV, two bat-SARS-like-CoV genomes and common human

coronavirus genomes. Compatible target sites were finally compared to those used in

current protocols from the CDC and WHO. LAMP primers for SARS-CoV-2 were

designed against regions of the N-gene and E-gene using PrimerExplorer v5

(https://primerexplorer.jp/e/). RNase P POP7 primers were originally published by

Curtis, et al. (2018) and a compatible gRNA was designed to function with these primer

sets.

Target RNAs were generated from synthetic gene fragments of the viral genes of

interest. First a PCR step was performed on the synthetic gene fragment with a forward

primer that contained a T7 promoter. Next, the PCR product was used as the template

for an in-vitro transcription (IVT) reaction at 37°C for 2 hours. The IVT reaction was then

treated with TURBO DNase (Thermo) for 30 minutes at 37°C, followed by a heat-

denaturation step at 75°C for 15 minutes. RNA was purified using RNA Clean and

Concentrator 5 columns (Zymo Research). RNA was quantified by Nanodrop and Qubit

and diluted in nuclease-free water to working concentrations.

*DETECTR assays*

DETECTR assays were performed using RT-LAMP for pre-amplification of viral

or control RNA targets and LbCas12a for the *trans*-cleavage assay. RT-LAMP was

prepared as suggested by New England Biolabs

(https://www.neb.com/protocols/2014/10/09/typical-rt-lamp-protocol) with a $MgSO_4$

concentration of 6.5 mM and a final volume of 10 µL. LAMP primers were added at a

final concentration of 0.2 µM for F3 and B3, 1.6 µM for FIP and BIP, and 0.8 µM for LF

and LB. Reactions were performed independently for N-gene, E-gene, and RNase P

using 2 µL of input RNA at 62°C for 20 minutes.

LbCas12a *trans*-cleavage assays were performed similar to those described in

Chen, et al. (2018). 50 nM LbCas12a (available from NEB) was pre-incubated with 62.5

nM gRNA in 1X NEBuffer 2.1 for 30 minutes at 37°C. After formation of the RNA-protein

complex, the lateral flow cleavage reporter (/56-FAM/TTATTATT/3Bio/, IDT) was added

to the reaction at a final concentration of 500 nM. RNA-protein complexes were used

immediately or stored at 4°C for up to 24 hours before use.

*Lateral flow readout*

After completion of the pre-amplification step, 2 µL of amplicon was combined

with 18 µL of LbCas12a-gRNA complex and 80 µL of 1X NEBuffer 2.1. The 100 µL

LbCas12a trans-cleavage assay was allowed to proceed for 10 minutes at 37°C.

A lateral flow strip (Milenia HybriDetect 1, TwistDx) was then added to the reaction tube

and a result was visualized after approximately 2-3 minutes. A single band, close to the

sample application pad indicated a negative result, whereas a single band close to the

top of the strip or two bands indicated a positive result.

*Optimized DETECTR method for patient samples*

264           The patient optimized DETECTR assays were performed using RT-LAMP

265    method as described above with the following modifications: A DNA binding dye,

266    SYTO9 (Thermo Fisher Scientific), was included in the reaction to monitor the

267    amplification reaction and the incubation time was extended to 30 minutes to capture

268    data from lower titre samples.

269           The fluorescence based patient optimized LbCas12a *trans*-cleavage assays

were performed as described above with modifications; 40nM LbCas12a was pre-

incubated with 40nM gRNA, after which 100nM of a fluorescent reporter molecule

compatible with detection in the presence of the SYTO9 dye

(/5Alex594N/TTATTATT/3IAbRQSp/) was added to the complex. 2 µL of amplicon was

combined with 18 µL of LbCas12a-gRNA complex in a black 384-well assay plate and

monitored for fluorescence using a Tecan plate reader.

*Contrived sample preparation.*

278           In-vitro transcribed RNA (gift from California Department of Public Health

279    (CDPH)), with a titer of 10,000 copies/µL (Ct value of 21) was diluted into 2,500

280    copies/µL first, then serially diluted in water to concentration of 1 250, 625, 312, 156, 78,

281    39, 10 and 2.5 copies per microliter.

*Human clinical sample collection and preparation*

Negative nasopharyngeal swabs were acquired from healthy donors in Chiu lab

with the approval of the University of California, San Francisco (UCSF) IRB. Clinical

nasopharyngeal and oropharyngeal swab samples of SARS-CoV-2 patients were

collected in UTM and transported to the CDPH or UCSF lab. Sample RNA of SARS-

CoV-2 was extracted following instructions as described in the CDC EUA-approved

protocol[6] (input 120 µL, elution of 120 µL) using Qiagen DSP Viral RNA Mini Kit

(Qiagen) at CDPH and the MagNA Pure 24 instrument (Roche Life Science) at UCSF.

Nasopharyngeal swab samples of influenza and common coronavirus were extracted at

UCSF using the MagNA Pure 24 instrument.

*CDC real-time qRT-PCR assay*

The CDC assay was performed using the ABI 7500 Fast DX instrument (Applied

Biosystems) according the CDC EUA-approved protocol[6].

**SUPPLEMENTARY FIGURES**

**Supplementary Figure 1.** Comparison of sequences between SARS-CoV-2, SARS-

CoV, and bat-SL-CoVZC45 at the sites targeted by the gRNAs used in this study. **(a)**

The N-gene gRNA is compatible with the CDC-N2 amplicon, and **(b)** the E-gene gRNA

is compatible with the WHO E-Sarbeco amplicon. **(c-d)** DETECTR fluorescence values

using **(c)** N gene gRNAs and **(d)** E gene gRNAs.

**Supplementary Figure 2.** Cross-reactivity of DETECTR to common human

coronaviruses. SARS-CoV-2 DETECTR assay (RT-LAMP + Cas12a) was evaluated on

IVT RNA products from SARS-CoV-2, SARS-CoV, bast-SL-CoVZC45, and clinical

samples from common human coronaviruses. As expected, the N-gene is only detected

in SARS-CoV-2, whereas the E-gene is detected only in SARS-CoV-2 and bat-SL-

CoVZC45. SARS-CoV E-gene was not detected as the RT-LAMP primer set is not

capable of amplifying the SARS-CoV E-gene, even though the E-gene gRNA is capable

of detecting the SARS-CoV E-gene target site. RNase P is detected in common human

coronaviruses because these samples are RNA extracted from clinical samples. Result

shown at 15 minutes of LbCas12a detection assay signal on fluorescent plate reader.

**Supplementary Figure 3.** DETECTR analysis of SARS-CoV-2 identifies down to 10 viral genomes in approximately 30 minutes. Duplicate LAMP reactions were amplified for twenty minutes followed by LbCas12a DETECTR analysis. **(a)** Raw fluorescence curves generated by LbCas12a detection of SARS-CoV-2 N-gene (n=6) show saturation in less than 20 minutes. **(b)** Further analysis reveals the limit of detection of the SARS-CoV-2 N-gene to be 10 viral genomes per reaction (n=6). **(c)** Evaluation of the time to result of these reactions highlights detection of 10 viral genomes of SARS-CoV-2 in under 5 minutes (n=6).

333
**Supplementary Figure 4.** Impact of sample buffers on performance of RT-LAMP pre-

amplification. Time-to-result for RT-LAMP amplification (lower value indicates faster

amplification) with 10% universal transport medium (UTM), 10% PBS, or 10% water

final volume for the SARS-CoV-2 N-gene on a standard curve of the 2019-nCoV

positive control plasmid (IDT) in 10% reaction volume. Results indicate that 10% PBS

inhibits the assay less than 10% UTM.

**Supplementary Figure 5.** DETECTR kinetic curves on COVID-19 infected patient

samples. Ten nasopharyngeal/oropharyngeal swab samples from 5 patients (COVID19-

1 to COVID19-5) were tested for SARS-CoV-2 using two different genes, N2 and E as

well as a sample input control, RNase P. **(a)** Using the standard amplification and

detection conditions, 9 of the 10 patient samples resulted in robust fluorescence curves

indicating presence of the SARS-CoV-2 E-gene (20-minute amplification, signal within

10 minutes). **(b)** The SARS-CoV-2 N-gene required extended amplification time to

produce strong fluorescence curves (30-minute amplification, signal within 10 minutes)

for 8 of the 10 patient samples. **(c)** As a sample input control, RNase P was positive for

17 of the 22 total samples tested (20-minute amplification, signal within 10 minutes).

355
**Supplementary Figure 6.** Timeline showing major events in the progression of COVID-

19 detection and assay development.

**Extended Data 1 ("Extended_Data_1.xlsx").** Standard curve generated by running

seven 5- or 10-fold dilutions of the CDC N2 qRT-PCR assay, with 3 replicates each

dilution. The R-squared measure corresponding to the regression line is 0.9981.

**Extended Data 2 ("Extended_Data_2.xlsx").** Primer, reporter molecules, target gene

fragments, and guide RNAs used in this study.

**ACKNOWLEDGEMENTS**

This work was funded by NIH grants R33-AI129455 (CYC) from the National Institute of

Allergy and Infectious Diseases and R01-HL105704 (CYC) from the National Heart,

Lung, and Blood Institute. We thank Jill Hakim and Dustin Glasner for assisting with

sample collection, and Vikram Joshi, Maria-Nefeli Tsaloglou and Xin Miao for helpful

discussions in the preparation of this manuscript.

**AUTHOR CONTRIBUTIONS**

CYC and JSC conceived the study. JPB conceived, designed and validated DETECTR

reagents and protocols. XD and GY validated RT-PCR and LAMP on patient samples.

JPB, XD, CLF performed experiments and analyzed data. GY, JS, AG and AS

performed experiments. KZ, SM, EH and WG coordinated the study, consented UCSF

patients and collected samples. JS, CYP, HG and DW collected samples from patients

and extracted the viral RNA. CYC, JPB, XD, and JSC wrote and edited the manuscript.

All authors read the manuscript and agree to its contents.

**COMPETING INTERESTS**

CYC is the director of the UCSF-Abbott Viral Diagnostics and Discovery Center

(VDDC), receives research support funding from Abbott Laboratories, and is on the

Scientific Advisory Board of Mammoth Biosciences, Inc. JSC is a co-founder of

Mammoth Biosciences, Inc. JPB, CLF, and JS are employees of Mammoth

Biosciences, Inc. CYC, JPB, XD, CLF, JS and JSC are co-inventors of CRISPR-related

technologies.

**REFERENCES**

1    Wang, C., Horby, P. W., Hayden, F. G. & Gao, G. F. A novel coronavirus
outbreak of global health concern. *Lancet* **395**, 470-473, doi:10.1016/S0140-
6736(20)30185-9 (2020).
2    Zhu, N. *et al.* A Novel Coronavirus from Patients with Pneumonia in China, 2019.
*N Engl J Med* **382**, 727-733, doi:10.1056/NEJMoa2001017 (2020).
3    World Health Organization. Coronavirus disease 2019 (COVID-19) Situation
Report - 44. 8 (World Health Organization, Geneva, Switzerland, 2020).
4    Bai, Y. *et al.* Presumed Asymptomatic Carrier Transmission of COVID-19. *JAMA*,
doi:10.1001/jama.2020.2565 (2020).
5    Rothe, C. *et al.* Transmission of 2019-nCoV Infection from an Asymptomatic
Contact in Germany. *N Engl J Med* **382**, 970-971, doi:10.1056/NEJMc2001468
(2020).
6    US Centers for Disease Control and Prevention. *Real-Time RT-PCR Panel for*
*Detection 2019-nCoV*, <https://www.cdc.gov/coronavirus/2019-ncov/lab/rt-pcr-
detection-instructions.html> (2020).
7    US Food and Drug Administration. Policy for Diagnostics Testing in Laboratories
Certified to Perform High Complexity Testing under CLIA prior to Emergency Use
Authorization for Coronavirus Disease-2019 during the Public Health Emergency.
(US Food and Drug Administration, 2020).
8    Chen, J. S. *et al.* CRISPR-Cas12a target binding unleashes indiscriminate
single-stranded DNase activity. *Science* **360**, 436-439,
doi:10.1126/science.aar6245 (2018).
9    Chiu, C. Cutting-Edge Infectious Disease Diagnostics with CRISPR. *Cell Host*
*Microbe* **23**, 702-704, doi:10.1016/j.chom.2018.05.016 (2018).
10    Gootenberg, J. S. *et al.* Nucleic acid detection with CRISPR-Cas13a/C2c2.
*Science* **356**, 438-442, doi:10.1126/science.aam9321 (2017).
11    Myhrvold, C. *et al.* Field-deployable viral diagnostics using CRISPR-Cas13.
*Science* **360**, 444-448, doi:10.1126/science.aas8836 (2018).
12    Notomi, T. *et al.* Loop-mediated isothermal amplification of DNA. *Nucleic Acids*
*Res* **28**, E63, doi:10.1093/nar/28.12.e63 (2000).
13    World Health Organization. *Diagnostic detection of Wuhan coronavirus 2019 by*
*real-time RT-PCR,* < https://www.who.int/docs/default-
source/coronaviruse/wuhan-virus-assay-
v1991527e5122341d99287a1b17c111902.pdf>. (World Health Organization,
Geneva, Switzerland, 2020).
14    GenomeWeb. *CDC Revises SARS-CoV-2 Assay Protocol*,
<https://www.genomeweb.com/pcr/cdc-revises-sars-cov-2-assay-protocol-
surveillance-testing-track-start-next-week#.XmFyBJNKjxs> (2020).
15    Liu, J. *et al.* Community Transmission of Severe Acute Respiratory Syndrome
Coronavirus 2, Shenzhen, China, 2020. *Emerg Infect Dis* **26**,
doi:10.3201/eid2606.200239 (2020).
16    California Department of Public Health. *CDC confirms possible first instance of*
*COVID-19 community transmission in California*,
<https://www.cdph.ca.gov/Programs/OPA/Pages/NR20-006.aspx> (2020).

17   Holshue, M. L. *et al.* First Case of 2019 Novel Coronavirus in the United States. *N Engl J Med* **382**, 929-936, doi:10.1056/NEJMoa2001191 (2020).
18   Pan, Y., Zhang, D., Yang, P., Poon, L. L. M. & Wang, Q. Viral load of SARS-CoV-2 in clinical samples. *Lancet Infect Dis*, doi:10.1016/S1473-3099(20)30113-4 (2020).
19   Zou, L. *et al.* SARS-CoV-2 Viral Load in Upper Respiratory Specimens of Infected Patients. *N Engl J Med*, doi:10.1056/NEJMc2001737 (2020).
20   Young, B. E. *et al.* Epidemiologic Features and Clinical Course of Patients Infected With SARS-CoV-2 in Singapore. *JAMA*, doi:10.1001/jama.2020.3204 (2020).
21   Elbe, S. & Buckland-Merrett, G. Data, disease and diplomacy: GISAID's innovative contribution to global health. *Glob Chall* **1**, 33-46, doi:10.1002/gch2.1018 (2017).
450