Deep Learning-Based Universal Expert-Level Recognizing

Pathological Images of Hepatocellular Carcinoma and beyond

Running title: Universal Recognizing Pathological Images

Wei-Ming Chen[1], Min Fu[2], Cheng-Ju Zhang[3], Qing-Qing Xing[1], Fei Zhou[1], Meng-Jie

Lin[4], Xuan Dong[1], Qi-Zhong Zheng[5*], Mei-Zhu Hong[6*], and Jin-Shui Pan[1*]

[1]Department of Gastroenterology, Zhongshan Hospital Affiliated to Xiamen

University, Xiamen, Fujian, China;

[2]School of Aerospace Engineering, Xiamen University, Xiamen, Fujian, China;

[3]Department of Anesthesiology, Zhongshan Hospital Affiliated to Xiamen University,

Xiamen, Fujian, China;

[4]Department of Pathology, Zhongshan Hospital Affiliated to Xiamen University,

Xiamen, Fujian, China;

[5]Department of Pathology, Xiamen Hospital of Traditional Chinese Medicine, Xiamen,

Fujian, China;

[6]Department of Traditional Chinese Medicine, Zhongshan Hospital Affiliated to

Xiamen University, Xiamen, Fujian, China

Emails:

Wei-Ming Chen   345173327@qq.com

Min Fu   121820447@qq.com

Cheng-Ju Zhang   1315363531@qq.com

Qing-Qing Xing   582398903@qq.com

Fei Zhou   feiflyfei888@163.com

Meng-Jie Lin   linmengjie1016@163.com

Xuan Dong   362485659@qq.com

Qi-Zhong Zheng   188192127@qq.com

Mei-Zhu Hong   546777397@qq.com

Jin-Shui Pan   j.s.pan76@gmail.com

*Correspondence:**

Jin-Shui Pan, MD, PhD.

Department of Gastroenterology, Zhongshan Hospital Affiliated to Xiamen University

No. 201-209, Hubin Nan Road

Xiamen 361004, Fujian, China

E-mail: j.s.pan76@gmail.com

Fax: +86 592 2590 150

Tel.: +86 592 2590 150.

Mei-Zhu Hong, M.D.

Department of Traditional Chinese Medicine

Zhongshan Hospital Affiliated to Xiamen University,

201 Hu-Bin Nan Road

Xiamen, Fujian 361004, China

E-mail: 546777397@qq.com

Tel: +86-592-229-2015

Qi-Zhong Zheng, M.D.

Department of Pathology

Xiamen Hospital of Traditional Chinese Medicine

1739 Xian-Yue Road

Xiamen, Fujian 361004, China

E-mail: 188192127@qq.com

Tel: +86-592-557-9529

List of Abbreviations:

HCC, Hepatocellular carcinoma; CRC, colorectal cancer; BIDC, breast invasive

ductal carcinoma; AI, artificial intelligence.

Word Count: 2811

ABSTRACT

Purpose: Human–based medical–image interpretation always falls into the predicament between specialized practitioners and expanding medical imaging. We aim at developing a diagnostic tool for pathological–image classification by using transfer learning that can be applied to diverse tumor types.

Experimental Design: In this study, images were retrospectively collected and prospectively analyzed using machine learning. Microscopic images of liver tissue that show or do not show hepatocellular carcinoma were used to train and validate a classification framework based on convolutional neural network. To evaluate the universal classification performance of the artificial–intelligence (AI) framework, histological images from colorectal tissue and breast were also collected. Training and validation set of images were collected from Xiamen Hospital of Traditional Chinese Medicine whereas test set of images were collected from Zhongshan Hospital Xiamen University.

Results: Accuracy, sensitivity, and specificity were reported and compared to human image interpretation and other AI image classification systems such as AlexNet and GoogLeNet. For the test dataset, sensitivity, specificity, and area under the curve of the AI framework were 99.1%, 98.0%, and 0.960, respectively. In human–machine comparisons, the accuracy of the AI framework was 98.5%, while the accuracy of human experts fluctuated between 93.0% and 95.0%. Based on transfer learning, the AI framework accuracy for colorectal carcinoma, breast invasive ductal carcinoma,

were 96.8%, and 96.0%, respectively.

Conclusions: The performance of the proposed AI framework in classifying

histological images with hepatocellular carcinoma is comparable to the classification

by human experts. With limited training, the proposed AI framework has potential

universality in histological image classification.

Key Words: machine learning, pathology, transfer learning, diagnostic imaging,

hepatocellular carcinoma

Study Highlights

WHAT IS KNOWN

✓ Accurate recognition of medical images is the basis for clinical

decision-making.

✓ Unresolved challenge exists between specialized practitioners and

expanding medical imaging output.

WHAT IS NEW HERE

✓ Proposed AI framework has excellent performance in classifying

hepatocellular carcinoma.

✓ The AI framework has universal feature in classifying other types of

medical images.

✓ The AI framework may help to interpret emerging imaging technology.

Introduction

Hepatocellular carcinoma (HCC) ranks as the fifth most common cancer worldwide and the second most common cause of cancer-related deaths.(1) In the United States and China, HCC is estimated to be the fourth and third most common cause of cancer-related deaths, respectively.(2, 3) This type of liver cancer usually develops in patients with liver cirrhosis, especially in chronic hepatitis B (CHB) or chronic hepatitis C (CHC)-related liver cirrhosis.(4-6) In cirrhosis cases, numerous nodules of varying sizes are found in the liver. Benign and malignant intrahepatic nodule identification is often a huge challenge for computed tomography (CT) or magnetic resonance imaging (MRI)-based diagnosis. Definitive diagnosis of intrahepatic nodules usually depends on liver biopsy. Histopathology diagnosis is the gold standard for determining the nature of hepatic space occupying lesions. However, diagnosing a large number of pathology slide images is laborious. Additionally, substantial observer-to-observer variation in liver biopsy assessments cannot be neglected.(7) Another challenge faced during medical-image diagnostics is the patient-to-patient variability in the pathology of disease manifestation. Even experienced pathologists differ significantly in the interpretation of histopathology of the same disease. Therefore, it is necessary to develop novel auxiliary diagnostic facilities.

Diagnostic approaches to HCC include ultrasound, CT, and MRI, with histopathology

diagnosis as the "gold standard".(8) In addition to HCC, the colorectal cancer (CRC)

and breast invasive ductal carcinoma (BIDC), described later, are some of the most

common tumors. In 2015, CRC was estimated to be the fifth most common cause of

cancer-related deaths in China.(3) Similarly, according to the 2018 United States

cancer statistics published by Siegel et al.,(2) CRC ranks as the third most common

cause of cancer-related deaths in both men and women. Moreover, adenocarcinoma is

the most common type of CRC, making a confirmed diagnosis of CRC dependent on

pathology imaging and interpretation. Diagnostic methods of CRC include CT,

colonoscopy and subsequent tissue examination,(9) while breast cancer diagnostics

include ultrasound and mammography.(10) In 2015, breast cancer was estimated to

have contributed toward most new cancer cases in China.(3) The same was true in the

United States, according to cancer statistics published by Siegel et al.(2) Invasive

ductal carcinoma is the most common type of breast cancer that is diagnosed

histologically.(11) Similar to CRC, histopathology examination is also the diagnostic

gold standard for BIDC. To summarize, there are numerous imaging diagnostic means

for almost all diseases. More than that, almost all diagnostic imaging can produce a

large number of medical images. For example, a plain scan combined with

contrast-enhanced CT or MRI can produce more than 1000 images per examination,

while capsule endoscopy can produce more than 40000 medical images per

examination. Thus, interpretation of these images can be time-consuming.

With the continuous emergence of new technologies, the number and types of medical

images have expanded at an unprecedented rate. However, an unresolved challenge

exists due to the imbalance between the ability and number of specialized

practitioners and expanding medical imaging output. Usually, a physician is familiar

with only a few or even just one type of diagnostic imaging, whereas the

interpretation of countless medical images requires human expertise and judgement to

correctly understand and triage. Therefore, it is undoubtedly attractive to develop an

artificial intelligence (AI) system of high classification accuracy with universal

recognition capability.

In recent years, AI has been widely used in various fields.(12-14) For classification

that is difficult for human experts or where rapid review of an immense number of

images is needed, AI has outstanding advantages and has a revolutionary role in

disease diagnosis. In the present study, we developed an effective convolutional

neural network (CNN) based on a deep-learning algorithm to classify medical images.

We also evaluated the generalizability performance of the proposed AI system in

interpreting histological images of several common types of tumors through transfer

learning.

Methods

Patients

Patients who underwent biopsy or surgical resection due to diseases of the liver,

colorectum, or breast in Xiamen Hospital of Traditional Chinese Medicine, or

Zhongshan Hospital Xiamen University between June 1, 2010 and December 31,

2017. Adult patients who aged between 18 and 75 were enrolled. The enrolled

participants had biopsy or surgical resection specimens with completed structure, and

one of the following conditions: 1) chronic hepatitis-related to hepatis B virus (HBV)

or hepatis C virus (HCV), without HCC, with or without liver cirrhosis; 2) HCC

companied by HBV-related or HCV-related chronic hepatitis, with or without liver

cirrhosis; 3) CRC; 4) BIDC. All of HCC, CRC, and BIDC were further confirmed on

the basis of surgically resected specimens. Necroinflammatory activity and fibrosis or

cirrhosis related to chronic hepatitis were recorded using the Scheuer system.(15)

Histological diagnosis of HCC or CRC was made according to the digestive

tumor-classification system formulated by the World Health Organization (WHO) in

2010 (16) while the histological diagnosis of BIDC was made according to the

breast-tumor-classification system formulated by WHO in 2012.(17) There were no

exclusion criteria regarding gender, or race. This study was approved by the Ethics

Committees of Zhongshan Hospital, Xiamen University. Written informed consent

was waived by the Ethics Commission of the designated hospital because of

non-interventional study and no identifiable personal information was recorded.

Images

Collected tissue samples were fixed in the wax, followed by slicing at the thickness of

3 μm, and then staining by hematoxylin-eosin. After the process, histological images

were collected with a 200-fold magnification. Two to 6 images were collected for

each patient, and there was no overlap between the images. For a case with tumor, an

image of tumor was collected companied by corresponding non-tumor image, which

was taken at 2 cm away from tumor focus. Strategy of "full field" was adopted. An

image of tumor was taken in the focus of tumor whereas an image of non-tumor was

taken in the focus non-tumor. In other words, the entire image of tumor was consisted

of tumor tissue while the whole image of non-tumor was consisted of non-tumor

tissue. Each image was examined by a panel of two independent pathologists, each

with over 15 years of pathology experience. If there was a disagreement in clinical

labeling, the image was further arbitrated by a panel of senior pathology specialists.

Before training the AI framework, a whole image was divided into partial regions of

1920 × 1280 pixels. Identifiable personal information, such as name of the enrolled

patients, name of hospital, etc, was removed.

Datasets

Histological images collected from Department of Pathology, Xiamen Hospital of

Traditional Chinese Medicine acted as training and validation sets while histological

images collected from Department of Pathology, Zhongshan Hospital Xiamen

University acted as test sets to further verify the performance in classifying.

Histological images of HCC, non-HCC, CRC, non-CRC, BIDC, and non-BIDC were

collected independently from these two hospitals in the same way.

Training and validation the AI algorithm

Each divided image of 1920 × 1280 pixels was imported into the database with

multiple layers of classification. For an entire image, it would be classified as "tumor"

if tumor was identified even in only one divided image. However, only when all of

divided images were recognized as "non-tumor" would the entire image be classified

into the group of "non-tumor". Collected liver pathology images were randomly

divided into training set and validation set at a ratio of 3:1. The training set was used

to train the AI algorithm whereas the validation set was employed to evaluate the

classification performance of the trained AI algorithm. This process was repeated five

times.

Based on deep learning, we developed an AI algorithm, and used the PyTorch

platform to adopt the ResNet-34 architecture pretrained using the ImageNet

dataset.(18) The retraining consisted of initialization of the convolutional layers with

loaded pretrained weights and updating of the neural network to recognize our classes

such as HCC or non-HCC. In the training process, the network structure was kept

unchanged. However, the network learning rates, except the last fully connected layer,

were tuned to be 0.001. The learning rate of the last fully connected layer was tuned

to be 0.02 (0.001 × 20), and weights were updated using backpropagation. This

strategy tended to update the first several layers slowly, while updating the output

layer more efficiently. Layer training was performed by stochastic gradient descent in

batches of 64 images per step by using stochastic gradient descent optimizer. The

training procedure was run for 25 epochs with a dropout ratio equal to 0.5. The

modified ResNet-34 was trained on an 14.04.1-Ubuntu computer with Intel(R)

i7-5930K CPU @ 3.50 GHz. An NVIDIA GTX 1080Ti 11 GB GPU was utilized to

accelerate training.

Testing of the AI algorithm

232   The classification performance was evaluated independently in the images collected

233   from the two hospitals. After the training process was finished, histological images

234   collected from Department of Pathology, Zhongshan Hospital Xiamen University was

235   used as test set to monitor the classifying decisions of the trained algorithm.

Comparison between the AI algorithm and human experts

Histological images collected from Department of Pathology, Zhongshan Hospital

Xiamen University were also sent to expert pathologist to make a diagnosis.

Classification performance was compared with that of the AI algorithm. Expert

pathologists were senior staffs of Department of Pathology, Zhongshan Hospital

Xiamen University, with clinical experience about 15 years. Diagnosis was made

independently. The error rates were determined for the AI algorithm and for each of

the human experts. Comparison was also conducted between the proposed AI

algorithm and other framework such as AlexNet and GoogLeNet.(19, 20)

Transfer Learning of the AI algorithm

Transfer learning was developed by Donahue et al.(21) To evaluate the

transfer-learning performance, the trained AI algorithm was further tested by two

other kinds of tumors: CRC and BIDC. The classifying performance of the trained AI

algorithm was determined independently in each kind of tumor.

The study design was shown in Figure 1 while the CNN schematic for HCC

classification was shown in Supplementary Figure 1.

Statistical analysis

257   To evaluate the classification performance of the AI algorithm on histological images,

258   three indies including accuracy, sensitivity, and specificity were calculated. The

259   receiver operating characteristics (ROC) curves plot the true positive rate (sensitivity)

260   versus the false positive rate (1-specificity). $P < 0.05$ was set as the level for statistical

261   significance for two-tailed paired test.

Results

Patient and image characteristics

We obtained 7000 liver pathology slide images generated from 2745 patients enrolled

from Xiamen Hospital of Traditional Chinese Medicine, where 4000 images showed

confirmed HCC while the other 3000 images confirmed other diseases, such as

CHB/CHC with or without cirrhosis. All these images passed an initial image-quality

check, and were randomly divided into training set and validation set at a ratio of 3:1

to train and validate the classifying performance of the AI algorithm. This process was

repeated five times.

AI algorithm performance during process of training and validation

During training and validation process, accuracy and cross-entropy were plotted

against the iteration step, which were shown in Supplementary Figure 2. By using the

validation set as monitor, the mean sensitivity, specificity, and accuracy of the AI

algorithm were calculated as 98.6%, 98.5%, and 98.5%, respectively.

AI algorithm performance evaluated by test set

We also obtained 2400 images showing and not showing HCC, generated from 873

patients enrolled from Zhongshan Hospital Xiamen University, which were used to

further evaluate the performance of the AI algorithm. In these 2400 images, 1324 of

282 them showed HCC while 1076 of them showed non-HCC. By using the test set as

283 monitor, the sensitivity, specificity, and area under ROC curve of the AI algorithm

284 were calculated as 99.1%, 98.0%, and 0.960, respectively.

Comparison between the results of the AI algorithm and human experts

288 To compare the performances of the AI algorithm and human experts, we chose

289 another randomly selected set of 200 images consisting of 100 images with HCC and

290 100 images without HCC. All 200 images were sent to both the AI algorithm and

291 human experts to make clinical decisions. The accuracy, sensitivity, and specificity for

292 the AI algorithm were 98.5%, 98.0%, and 99.0%. For human experts, sensitivity

293 ranged from 86.0% to 97.0% while specificity ranged from 91.0% to 100%.

294 Compared with human experts, the AI algorithm tended toward a more balanced

295 performance between sensitivity and specificity. However, a remarkable variation was

296 observed between sensitivity and specificity distinguishing it from the results by the

297 human experts. The performances of the AI algorithm and human experts were

298 presented in Figure 2.

Comparison between the AI algorithm and other architectures

301 Two hundred images used for the human–machine comparison was also employed to

302   compare the performance of the AI algorithm with those of the other architectures.

303   The HCC image-recognition sensitivity, specificity, and accuracy of the proposed AI

304   system was superior to those of AlexNet and GoogLeNet, as reported in

305   Supplementary Table 1 and Supplementary Figure 3.

Transfer learning of the AI algorithm to CRC

To evaluate the proposed transfer-learning performance of the AI system, 3600

colorectal-tissue microscope slide images from Xiamen Hospital of Traditional

Chinese Medicine were collected to train and validate the AI algorithm. These 3600

images consisted of 1800 images showing CRC and another 1800 images not showing

CRC. Another 600 colorectal-tissue microscope images from Zhongshan Hospital

Xiamen University were also collected as a test set. As is shown in Figure 3, after

limited training, the proposed AI algorithm showed excellent accuracy in CRC and

non-CRC image classification based on transfer learning. An accuracy of 96.8% was

achieved, with a sensitivity of 97.0% and a specificity of 96.7%.

Transfer learning of the AI algorithm to BIDC

Microscope slide images from breast tissue were collected to further evaluate the

transfer-learning performance of the proposed AI system. A total of 3600 histologic

images of breast obtained from Xiamen Hospital of Traditional Chinese Medicine

were employed to train and validate the AI algorithm. In these 3600 images, 1800

images showed BIDC while another 1800 images showed non-BIDC. Another 600

breast microscope images from Zhongshan Hospital Xiamen University were also

obtained as a test set. As shown in Figure 4, after training, the proposed AI system

showed an accuracy of 96.0%, with a sensitivity of 95.7% and a specificity of 96.3%,

in classifying images into BIDC or non-BIDC.

Discussion

In this paper, we described a general AI algorithm for the interpretation of histological

images from the liver, colorectum, and breast. Although medical imaging, such as CT

and MRI, are widely used for HCC diagnosis, CT and MRI detection is poor for

HCCs < 1.0 cm, especially for patients with cirrhosis.(22) For those patients without

definitive findings on either CT or MRI, a biopsy may be a the only detection

method.(1) Owing to potential interobserver bias that may be present when reviewing

histological images generated from biopsy, AI may be a useful ancillary tool for HCC

identification. In addition, the proposed model demonstrated competitive performance

when analyzing liver histological images. This was accomplished without the need for

a highly specialized, deep learning machine and without a very large training database.

When the model was trained with 7000 images (3500 images for each class), high

performance accuracy, sensitivity, and specificity were achieved for the correct

diagnosis. Moreover, the model's performance in diagnosing HCC was comparable to

the diagnosis by experts with significant clinical experience in liver pathology.

By employing 2400 images (100 for each class) as test set, the proposed AI model

showed a more balanced performance between sensitivity and specificity in

recognizing HCC compared with that of human experts. The proposed AI model

accuracy was superior to that of experts, which also showed a remarkable variation

between sensitivity and specificity. The test set as aforementioned was used for

comparison between the proposed ResNet-34-architecture-based AI model and other

AI architectures including AlexNet and GoogLeNet. The proposed AI model achieved

superior accuracy, sensitivity, and specificity, thus demonstrating the robustness of the

model.

In the study by Kermany et al, over 100000 OCT images were employed to train the

AI framework.(14) However, only 3600 images were used to train our AI system in

our study based on transfer learning, yet excellent diagnostic performance was

achieved. Thus, even with a limited training dataset, the transfer-learning system

demonstrated highly effective classifications. Transfer-learning techniques for image

analysis could potentially be employed for a wide range of medical images across

multiple disciplines. In fact, a direct illustration of its wide applicability in the

analysis of two similar histological image types (CRC and BIDC) was shown. After a

much smaller amount of training, the proposed AI model reported an accuracy of

96.8% and 96.0% for CRC and BIDC, respectively. Importantly, the proposed AI

model showed a balance between sensitivity and specificity. Thus, the proposed AI

system has potential universality in the classification of histological images.

An AI model trained from an extremely large training dataset would have superior

performance to that of a transfer-learning-based model trained from a relatively small

training dataset. However, in practice, *de novo* training of a convolutional neural

network usually needs an unlimited supply of training data and would require weeks

to achieve a good accuracy. By using the retraining layers from other medical

classifications, a transfer-learning-based model would likely yield a highly accurate

model in much less time. Thus, for difficult-to-collect medical images,

transfer-learning-based image recognition is more practical. Given that imaging-based

diagnosis has played a crucial role in guiding treatment, extending the proposed AI's

application to diagnoses and treatment recommendations is a promising area for

future investigation.

CONFLICTS OF INTEREST

Guarantor of the article: Jin-Shui Pan, MD, PhD.

Specific author contributions: Conception or design of the work: J.S.P., M.Z.H. and Q.Z.Z. Acquisition and analysis of data: Q.Z.Z., W.M.C., C.J.Z., Q.Q.X., F.Z., M.J.L., X.D., Q.Z.Z., and M.Z.H. Interpretation of data: M.F and J.S.P. Drafting of the manuscript: J.S.P. Revision for critically important intellectual content and final approval of the version to be published: J.S.P., M.Z.H. and Q.Z.Z.

Financial support: This work was supported by the National Natural Science Foundation of China No.81871645 (J.S.P.). The funding sources did not have any role in the design and conduct of the study; collection, management, analysis, and interpretation of the data; preparation, review, or approval of the manuscript; and decision to submit the manuscript for publication.

Potential competing interests: None.

Acknowledgements: We are very grateful to the enrolled patients who provides precious histological images.

References

1.   Heimbach JK, Kulik LM, Finn RS, et al. AASLD guidelines for the treatment of

hepatocellular carcinoma. Hepatology 2018;67:358-80.

2.   Siegel RL, Miller KD, Jemal A. Cancer statistics, 2018. CA Cancer J Clin 2018;68:7-30.

3.   Chen W, Zheng R, Baade PD, et al. Cancer statistics in China, 2015. CA Cancer J Clin

2016;66:115-32.

4.   Terrault NA, Lok ASF, McMahon BJ, et al. Update on prevention, diagnosis, and treatment

of chronic hepatitis B: AASLD 2018 hepatitis B guidance. Hepatology 2018;67:1560-99.

5.   European Association for the Study of the Liver. EASL Recommendations on Treatment of

Hepatitis C 2018. J Hepatol 2018;69:461-511.

6.   European Association for the Study of the Liver. EASL 2017 Clinical Practice Guidelines on

the management of hepatitis B virus infection. J Hepatol 2017;67:370-98.

7.   Theodossi A, Skene AM, Portmann B, et al. Observer variation in assessment of liver

biopsies including analysis by kappa statistics. Gastroenterology 1980;79:232-41.

8.   European Association for the Study of the Liver. EASL Clinical Practice Guidelines:

Management of hepatocellular carcinoma. J Hepatol 2018;69:182-236.

9.   Wolf AMD, Fontham ETH, Church TR, et al. Colorectal cancer screening for average-risk

adults: 2018 guideline update from the American Cancer Society. CA Cancer J Clin

2018;68:250-81.

10.   Cardoso F, Kyriakides S, Ohno S, et al. Early breast cancer: ESMO Clinical Practice

Guidelines for diagnosis, treatment and follow-up. Ann Oncol 2019.

11. Li CI, Anderson BO, Daling JR, et al. Trends in incidence rates of invasive lobular and ductal

breast carcinoma. JAMA 2003;289:1421-4.

12. Esteva A, Kuprel B, Novoa RA, et al. Dermatologist-level classification of skin cancer with

deep neural networks. Nature 2017;542:115-18.

13. De Fauw J, Ledsam JR, Romera-Paredes B, et al. Clinically applicable deep learning for

diagnosis and referral in retinal disease. Nat Med 2018;24:1342-50.

14. Kermany DS, Goldbaum M, Cai W, et al. Identifying Medical Diagnoses and Treatable

Diseases by Image-Based Deep Learning. Cell 2018;172:1122-31 e9.

15. Scheuer PJ. Classification of chronic viral hepatitis: a need for reassessment. J Hepatol

1991;13:372-4.

16. WHO Classification of Tumours of the Digestive System. 4 ed: World Health Organization.;

2010.

17. WHO Classification of Tumours of the Breast. 4 ed: World Health Organization.; 2012.

18. He K, Zhang X, Ren S, et al. Deep residual learning for image recognition. In: Computer

Vision and Pattern Recognition 2016.

19. Krizhevsky A, Sutskever I, Hinton G. ImageNet classification with deep convolutional neural

networks. In: Proc. Advances in Neural Information Processing Systems 25; 2012; 2012. p.

1090.

20. Szegedy C, Liu W, Jia Y, et al. Going deeper with convolutions. Preprint at

http://arxiv.org/abs/1409.4842 2014.

21. Donahue J, Jia Y, Vinyals O, et al. DeCAF: A Deep Convolutional Activation Feature for

438   Generic Visual Recognition. In: the 31 st International Conference on Machine Learning;

439   2014; Beijing, China; 2014.

440  22. Roberts LR, Sirlin CB, Zaiem F, et al. Imaging for the diagnosis of hepatocellular carcinoma:

441   A systematic review and meta-analysis. Hepatology 2018;67:401-21.

Figure

Figure 1. Study Design

Figure 2. Performances of the proposed AI model and human experts during the

human–machine comparison.

A, Confusion matrix of the proposed AI model for HCC diagnosis;

B, Confusion matrix of human experts for HCC diagnosis;

C, Comparison of between the performance of the proposed AI model and that of

human experts for HCC diagnosis.

Figure 3. Transfer-learning performance of CRC diagnosis using colorectal tissue

microscope slide images.

In panel (A) and (B), the training dataset is shown in blue and the test dataset is

shown in red. Accuracy is plotted against the iteration step (A), and cross-entropy loss

is plotted against the iteration step (B) during the length of the training of the

binary-class classifier over the course of 8000 steps. The curve is smoothed. The test

accuracy and loss show better performance. Panel (C) shows the confusion matrix of

best test image model classification. The model successfully classifies CRC

separately from non-CRC.

Figure 4. BIDC diagnosis transfer-learning performance using breast tissue

microscope slide images.

In panel (A) and (B), the training and test datasets are shown in blue and red,

respectively. Classification accuracy is plotted against training epochs, and in (B), the

categorical cross-entropy loss is shown as a function of training epochs for the binary

classification problem. The curve is smoothed. Panel (C) shows the

model-classification confusion matrix for test image classification. As shown, the

proposed model successfully classifies BIDC from non-BIDC images.

Supplementary Tables

Supplementary Table 1. HCC image-recognition performance of the AI model and

other architectures

	Sensitivity (%)	Specificity (%)	Accuracy (%)
AlexNet	97.0	95.0	96.0
GoogLeNet	98.0	96.0	97.0
Our AI Model	99.0	98.0	98.5

476   Supplementary Table 2. Proposed AI framework performance in classifying other

477   types of medical images by using transfer learning

Images	Accuracy (%)	Sensitivity (%)	Specificity (%)
Colorectal cancer	96.8	97.0	96.7
Breast cancer	96.0	95.7	96.3

Supplementary Figures

Supplementary Figure 1 Algorithm Architecture

Whole slide images were divided into n images, which were sent to the CNN feature

extractor. Finally, the image features were determined by the constructed classifier.

Supplementary Figure 2 Performance of HCC recognition during validation.

(A) Classification accuracy is plotted against training epochs, and in (B), the

categorical cross-entropy loss is shown as a function of training epochs for the binary

classification problem. (C) Classification accuracy is plotted against validation

epochs, and in (D), the categorical cross-entropy loss is shown as a function of

validation epochs for the binary classification problem. The curve is smoothed.

492 Supplementary Figure 3. Performances of the proposed AI model and other

493 architectures for HCC diagnosis.