Portable and accurate diagnostics for COVID-19: Combined use of the miniPCR thermocycler and a well-plate reader for SARS-CoV-2 virus detection

Everardo González-González[1,2], Grissel Trujillo-de Santiago[1,3], Itzel Montserrat Lara-Mayorga[1,2], Sergio Omar Martínez-Chapa[3], Mario Moisés Alvarez[1,2*]

[1] Centro de Biotecnología-FEMSA, Escuela de Ingeniería y Ciencias, Tecnologico de Monterrey, Monterrey, Nuevo León, México

[2] Departamento de Bioingeniería, Escuela de Ingeniería y Ciencias, Tecnologico de Monterrey, Monterrey, Nuevo León, México

[3] Departamento de Ingeniería Mecátrónica y Eléctrica, Escuela de Ingeniería y Ciencias, Tecnologico de Monterrey, Monterrey, Nuevo León, México

*Corresponding author

E-mail: mario.alvarez@tec.mx (MMA)

Abstract

The coronavirus disease 2019 (COVID-19) pandemic has crudely demonstrated the need for massive and rapid diagnostics. By the first week of July, more than 10,000,000 positive cases of COVID-19 have been reported worldwide, although this number could be greatly underestimated. In the case of an epidemic emergency, the first line of response should be based on commercially available and validated resources. Here, we demonstrate the use of the miniPCR, a commercial compact and portable PCR device recently available on the market, in combination with a commercial well-plate reader as a diagnostic system for detecting genetic material of the severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2), the causal agent of COVID-19. We used the miniPCR to detect and amplify SARS-CoV-2 DNA sequences using the sets of initiators recommended by the World Health Organization (WHO) for targeting three different regions that encode for the N protein. Prior to amplification, samples were combined with a DNA intercalating reagent (i.e., EvaGreen Dye). Sample fluorescence after amplification was then read using a commercial 96-well plate reader. This straightforward method allows the detection and amplification of SARS-CoV-2 nucleic acids in the range of ~625 to 2×10^5 DNA copies. The accuracy and simplicity of this diagnostics strategy may provide a cost-efficient and reliable alternative for COVID-19 pandemic testing, particularly in underdeveloped regions where RT-QPCR instrument availability may be limited. The portability, ease of use, and reproducibility of the miniPCR makes it a reliable alternative for deployment in point-of-care SARS-CoV-2 detection efforts during pandemics.

Introduction

Recent epidemic events (i.e., Zika in Southeast Asia and Latin-America in 2016 [1,2], Ebola in West Africa in 2013-2015 [3], and pandemic Influenza A/H1N1/2009 [4]) have clearly evidenced the urgent need for low-cost, portable, and easy-to-use diagnostic systems that can be effectively deployable to address epidemic episodes [5–8]. However, these portable diagnostic systems have been mainly viewed as solutions for underprivileged or remote places and/or for catastrophic scenarios. Nevertheless, the coronavirus disease 2019 (COVID-19) pandemic caused by the severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) [9] has broadsided most well developed and developing countries with only a few (i.e., South Korea [10], China, Singapore [11], and Taiwan [12]) showing an ability to deploy massive efforts for rapid and accurate detection of positive infection cases. The swift and massive testing of thousands of possibly infected subjects has been an important component of the strategy of these countries that has helped to effectively mitigate the spreading of COVID-19 among their populations [10,12–14]. And yet, most nations are still struggling to implement massive testing[15–17]. Current testing methods have exhibited important limitations in widespread reach, flexibility, cost-effectiveness, and scalability during this pandemic.

Through the last two pandemic events involving influenza A/H1N1/2009 and COVID-19 [18], the Centers for Disease Control (CDC) and the World Health Organization (WHO)

recommended the reverse transcription quantitative polymerase chain reaction (RT-qPCR) as the gold standard for official detection of positive cases. Without any doubt, nucleic acid amplification, and particularly RT-qPCR, is the most reliable technique for the early and accurate detection of viral diseases [19,20]. Unfortunately, conducting RT-qPCR diagnostics often depends on access to centralized laboratory facilities for testing [21–23]. To resolve this limitation, several different versions of compact PCR platforms (some of them q-PCR systems) have been described recently in the scientific literature [24–27]. Unfortunately, most of these devices have not yet reached the market. During epidemic emergencies, resourcing of incompletely developed technologies is impractical, and the use of commercially available diagnostic platforms becomes the first and arguably the most cost-efficient line of defense.

Only recently, several miniaturized PCR machines become commercially available [28,29]. One of them, the miniPCR [30,31], reached the international market in 2015. The most recent version of this compact PCR machine has an approximate cost of ~$800 USD (www.minipcr.com) as compared to $3000 USD for a conventional PCR thermocycler [28]. Several papers have documented the value of the miniPCR® system as a portable and robust diagnostic tool [32–36]. We recently published a comparison of the performance of the miniPCR and a commercial thermal cycler for the identification of artificial Zika and Ebola genetic sequences. Our experiments using a wide variety of primers sets and template concentrations revealed no differences in performance between either thermal cycler type [37]. The commercial availability, low price (as compared to conventional thermocyclers), portability, and user friendliness of the miniPCR makes it an attractive and tangible solution that effectively brings PCR analysis to the POC. In the present study, we

90   demonstrate the convenience of using the miniPCR for the detection and amplification of

91   synthetic samples of SARS-CoV-2 [18], the causal viral agent of the current COVID-19

92   pandemic.

94 **Materials and Methods**

Equipment specifications: We ran several sets of amplification experiments in a miniPCR

from Amplyus (MA, USA). The unit has dimensions of $20 \times 5 \times 15$ cm, weighs 0.7 kg,

and requires 120V (AC) and 3.5 A to operate. The miniPCR can run 8-16 amplifications

in parallel (depending on the model employed).

A commercial power supply (PowerPac from Bio-Rad, CA, USA) was used to operate the

electrophoresis unit used to run the agarose gels to reveal the amplification products

obtained by the miniPCR thermocycler. A Bio-Rad ChemiDoc XRS imaging system was

used for end-point PCR detection. Alternatively, the miniPCR unit has its own blueGel

electrophoresis unit (Figure 1 A, B), a compact electrophoresis unit ($23 \times 10 \times 7$ cm) that

weighs 350 g, that is powered by a built-in power supply (AC 100–240 V, 50–60 Hz).

We also used a Synergy HT microplate reader (BioTek Instruments, VT, USA) to detect

the fluorescence induced by an intercalating reagent in positive samples from the PCR

reactions.

Controls for validation: We used a plasmid containing the complete N gene from 2019-

nCoV, SARS, and MERS as positive controls at a concentration of 200,000 copies/µL

(Integrated DNA Technologies, IA, USA). Samples containing different concentrations of

synthetic nucleic acids of SARS-CoV-2 were prepared by successive dilutions from stocks

containing 200,000 copies mL^{-1} ng/L of viral nucleic acids. We used a plasmid containing

the GP gene from Ebola Virus (EBOV) as a negative control. The production of this EBOV

genetic material has been documented previously by our group [37].

Amplification mix: We used REDTaq Ready Mix from Sigma-Aldrich (USA), and followed

the recommended protocol: 10 μL Readymix, 0.5 μM of forward primer, 0.5 μM of reverse

primer,1μL of DNA template (~ 625 to $2x10^5$ DNA copies), 1μl of EvaGreen Dye, and

nuclease free water to final volume of reaction 20 μL.

120   **Figure 1.**

Primers used:

Three different sets of primers were used to target three different regions of the SARS-

CoV-2 N gene sequence. These primer sets are identical to those recommended by the

Center of Disease Control (CDC) for the standard diagnostics of COVID-19 (i.e., N1, N2,

and N3 assays) using quantitative real time PCR. Sequences of all these primers and their

corresponding amplicons are presented in Table 1 and 2.

Table 1. Primer sequences used in PCR amplification experiments

Name	Description	Primers Sequence (5'>3')
2019-nCoV_N1-F	2019-nCoV_N1 Forward Primer	5'-GAC CCC AAA ATC AGC GAA AT-3'
2019-nCoV_N1-R	2019-nCoV_N1 Reverse Primer	5'-TCT GGT TAC TGC CAG TTG AAT CTG-3'
2019-nCoV_N2-F	2019-nCoV_N2 Forward Primer	5'-TTA CAA ACA TTG GCC GCA AA-3'
2019-nCoV_N2-R	2019-nCoV_N2 Reverse Primer	5'-GCG CGA CAT TCC GAA GAA-3'

2019-nCoV_N3-F	2019-nCoV_N3 Forward Primer	5'-GGG AGC CTT GAA TAC ACC AAA A-3'
2019-nCoV_N3-R	2019-nCoV_N3 Reverse Primer	5'-TGT AGC ACG ATT GCA GCA TTG-3'

131 **Table 2. Amplicon sequences generated (and their corresponding lengths) by each of the**
132 **primer pairs used in the PCR amplification experiments (see also Fig XX).**
133

Primer pair	Amplicon sequence	Amplicon Length (nt)
N1	GACCCCAAAATCAGCGAAATGCACCCCGCATTACGTTTG GTGGACCCTCAGATTCAACTGGCAGTAACCAGA	72
N2	TTACAAACATTGGCCGCAAATTGCACAATTTGCCCCCAGC GCTTCAGCGTTCTTCGGAATGTCGCGC	67
N3	GGGAGCCTTGAATACACCAAAAGATCACATTGGCACCCG CAATCCTGCTAACAATGCTGCAATCGTGCTACA	72

Amplification protocols: For all PCR experiments, we used the same three-stage protocol

(see Fig 1D) consisting of a denaturation stage at 94 □C for 5 min, followed by 25 cycles

of 94 °C for 20s, 60 °C for 30s, and 72 °C for 20s, and then a final stage at 72 °C for 5 min,

for a total duration of 60 minutes in the miniPCR® thermocycler.

Documentation of PCR products: We analyzed 10 µL of each PCR product using 2%

agarose electrophoresis in Tris-acetic acid-EDTA (TAE) buffer (Sigma-Aldrich, MO,

USA). Gels were dyed with GelGreen (Biotium, CA, USA) using a 1:10,000 dilution and a

current of 110 V supplied by a Bio-Rad PowerPac HC power supply (Bio-Rad, CA, USA)

for 40 min. We used the Quick-Load Purple 2-Log DNA ladder (NEB, MA, USA) as a

molecular weight marker. We analyzed the gels by UV transillumination using a Bio-Rad

ChemiDoc XRS imaging system. In some of our experiments, we also used the blueGel

unit, a portable electrophoresis unit sold by MiniPCR from Amplyus (MA, USA). In these

experiments, we analyzed 10 µL of PCR product using 2% agarose electrophoresis tris-

borate-EDTA buffer (TBE). Gels were dyed with Gel- Green (CA, USA) using a 1:10,000

dilution, and a current of 48 V was supplied. Photo-documentation was done using a

smartphone camera. As a third method of detection and to read the amplification product,

we evaluated the amplification products by detecting the fluorescence emitted by a DNA

intercalating agent, the EvaGreen® Dye, in the Synergy HT microplate reader (BioTek

Instruments, VT, USA). Briefly, 20 μL of the PCR reaction mix was placed in distinct

wells of a 96-well plate, after completion of the PCR program. Each well was made to a

final volume of 150 μL by adding 130 μL of nuclease free water and the samples were well

mixed by pipetting. These experiments were run in triplicate. The following conditions

were used in the microplate reader: excitation of 485/20, emission of 528/20, gain of 75.

Fluorescence readings were made from the above at room temperature.

Statistical Analysis: Determination of mean values and standard deviations were conducted

using Excel tools. All experiments were run by triplicate. Regression analysis was

conducted in Excel.

Results and Discussion

Time is the most limiting factor in epidemic emergencies. Therefore, the integration of

well-developed and commercially available technologies [31,37,38] becomes an obvious,

expedient, and cost-effective first line of defense in the context of COVID-19 pandemics.

Here, we demonstrate that the combined use of a commercial and portable PCR unit (the

miniPCR) and a 96-well plate reader is potentially adequate for the fast deployment of

diagnostic efforts. We show the combined ability of both units to amplify and identify

different synthetic genetic sequences of SARS-CoV-2 (see Materials and Methods).

Analysis of sensitivity

We conducted a series of experiments to assess the sensitivity of the PCR reactions

conducted in the miniPCR thermocycler using the three sets of primers recommended by

CDC to diagnose infection by SARS-CoV-2. Table 1 shows the sets of primers used to

target genetic sequences that code for the expression of the SARS-CoV-2 N protein. Table

2 shows the sequence of the DNA products (amplicons) generated by successful targeting

of these regions with the N1, N2, and N3 primer pairs.

Figures 2A-C show the PCR products of the amplification reactions conducted using three

different primer pairs. In all cases, different concentrations of SARS-CoV-2 genetic

material, in the range of 2.0×10^5 to 625 DNA copies, were used as reaction templates. If

we put this range in the proper clinical context, the actual viral load of COVID-19 in nasal

swabs from patients has been estimated to fall within the range of 10^5 to 10^6 viral copies

per mL [18]. The amplification proceeds with sufficient quality to allow proper

visualization of the amplification products in electrophoresis gels, even at low nucleic acid

concentrations. Figure 2A-C shows agarose gels containing the amplification products of

each one of three experiments, where the three different sets of primers (namely N1, N2,

and N3) were used to amplify the same range of concentrations of template. The

miniPCR® was able to generate a visible band of amplification products for all three

primer sets and across the whole range of synthetic viral loads.

In general, the products of amplification in final point PCR are primarily detected on

agarose gels using conventional electrophoresis techniques conducted with conventional

lab equipment. The miniPCR® system is commercialized with its own electrophoretic unit

("blueGel®"; Figure 1B-C). The blueGel® has several important advantages and represents

a valid and portable solution for detecting PCR amplification products. Nevertheless,

running an experiment aimed at visualizing amplification products, as with any standard gel

electrophoresis procedure requires time. A good separation of bands typically involves a

processing time of 35 to 60 minutes from the loading of the amplification product to the

final documentation through photography.

As an alternative, we show here that the amount of amplification product can be

quantitatively evaluated using a commercial 96-well plate reader. To do this, we used an

intercalating agent during amplification in the miniPCR apparatus. Figures 2D-F shows the

fluorescence readings associated with the analysis of the different dilutions of synthetic

SARS-CoV-2 samples previously revealed by gel electrophoresis (S1 Fig., S2 Fig., and S3

Fig). We ran triplicate reactions for each dilution and for each primer data set. The

fluorescence readings were capable of clearly discriminating between positive and negative

samples across the whole range of dilutions tested (from 2×10^5 to 625 copies). This

observation holds true for each of the three primer sets tested. Note that the use of a plate

reader, instead of a conventional gel electrophoresis unit, presupposes a significant savings

in time. Up to 96 PCR reactions can be read in a matter of 5 to 10 minutes. This implies

that an array of 12 miniPCR units and a plate reader could equal the throughput of a

traditional RT-QPCR platform, but at one third of the capital cost. In addition, during

emergencies and particularly in developing countries, attaining or buying regular thermal

cyclers and plate readers is much easier than purchasing or accessing RT-qPCR systems.

In addition, our results suggest that fluorescence readings using a plate reader exhibit high

reproducibility and robustness. Overall, we obtained small standard deviations (in the range

of 6 to 40 arbitrary fluorescence units [a.f.u.]) and a small average variance coefficient

(2.6%) in fluorescence readings across the whole range of values of viral copies tested. We

observed similar variability indicators in experiments using different primer pairs. For

instance, we observed variance coefficients of 2.31%, 2.15%, and 3.34% when using

primer sets N1, N2, and N3, respectively. If we considered only fluorescence readings from

positive samples, we observed variance coefficients of 2.23%, 2.34%, and 1.31% when

using primer sets N1, N2, and N3, respectively.

Figure 2.

Figure 2G consolidates the fluorescence readings obtained from miniPCR amplifications

using synthetic SARS-CoV-2 samples and the primer sets N1 (blue bars), N2 (red bars),

and N3 (yellow bars). Overall, this data set is consistent. These results suggest that any of

the primer sets tested (N1, N2, or N3) may be used to amplify SARS-CoV-2 genetic

material in the miniPCR. However, for the experimental conditions tested (i.e., the nature

and concentration of the intercalating agent, the concentrations of primers, and the

concentration of enzyme, among others), we observe differences in the performance of each

primer pair. For example, primer sets N1 and N3 appear to promote amplifications in which

the observed fluorescence is proportional to the initial concentration of DNA template (i.e.,

the viral load). By contrast, primer pair N2 appears to generate amplification product with

high fluorescence emissions even at low values of the initial final copy numbers. Note that

all fluorescence readings for positive samples shown in Figure 2E exhibit a fluorescence

reading between 1300 and 1400 a.f.u.

Furthermore, measuring the fluorescence with the plate reader may add a quantitative

element to the analysis of positive COVID-19 samples. In principle, samples with higher

viral loads will exhibit higher fluorescence if processed through the same PCR program

(i.e., exposed to the same number of cycles). For example, for amplifications using primer

set N3, we observe a linear relationship between the natural logarithm of the number of

viral copies and the natural logarithm of fluorescence signal for the range of 625 to 40,000

viral copies:

$$\text{Ln (viral load)} = \alpha * \text{Ln}(F_{sample} - F_o) \quad (1)$$

where F_o is the fluorescence reading exhibited by a blank (i.e., a negative sample prepared

and processed in the same way than the positive samples) and $\alpha = 8.897$ (as determined by

fitting of the data presented in Figure 3A). For instance, we believe we can adjust the

concentration of intercalating reagent to assure linearity of the fluorescence signal with

respect to the viral load for experiments with different primer sets. This simple strategy will

result in a fully quantitative, reliable, and easily implemented quantitative version of a

straightforward final-point PCR protocol.

260   **Figure 3.**

Using the primers and methods described here, we were able to consistently detect the

presence of SARS-CoV-2 synthetic DNA using a miniPCR and a simple plate reader. In the

current context of the COVID-19 pandemics, the importance of communicating this result

does not reside in its novelty but in its practicality. In our experiments, we have used the

three sets of primers designed and recommended by the CDC to identify the presence of

SARS-CoV-2, the causal agent of COVID-19. These primer pairs, aimed at identifying

three different regions encoding for the N protein of SARS-CoV-2, have been widely

validated and used for diagnostic purposes in actual COVID-19 patients, Here we simply

translated widely tested protocols from the framework of an RT-qPCR apparatus (the gold

standard platform recommended for analyzing and confirming positive cases) to execution

in a miniaturized and already commercial POC thermal cycler. While the cost of a

commercial RT-qPCR apparatus falls in the range of $10,000 to $40,000 USD, the

commercial value of the miniPCR is under $800 USD. This difference is significant,

especially when considering the need for rational investment of resources during an

epidemic crisis.

While the quantitative capabilities of testing in a RT-QPCR platform are undisputable, the

capacity of many countries for rapid, effective, and massive establishment of diagnostic

centers based on RT-qPCR is questionable. The current pandemic scenarios experienced in

the USA, Italy, France, and Spain, among others, have crudely demonstrated that

centralized labs are not an ideal solution during emergencies. Portable diagnostic systems

may provide the required flexibility and speed of response that RT-qPCR platforms cannot

deliver.

To further illustrate the deterministic and quantitative dependence between the

concentration of amplification product and the fluorescence signal, as measured in a plate

reader, we simulated some real-time amplification experiments. To that end, we conducted

amplification reactions using initial amounts of 4×10^4 copies of synthetic SARS-CoV-2 in

the miniPCR cycler. We added the intercalating agent, EvaGreen Dye, to the reaction mix

at the initial time and extracted samples after 1, 5, 10, 15, 20, 25, and 30 PCR cycles. The

fluorescence from these samples was then measured in a plate reader. We observed a linear

increase in fluorescence as more PCR cycles were performed (Figure 3B); this highlights

the quantitative nature of the intercalating reaction.

Our results suggest that using a commercial plate reader to determine the extent of advance

of PCR amplifications is a practical, reliable, reproducible, and robust alternative to the use

of gel electrophoresis. Moreover, fluorescence reading of PCR products may lead to precise

quantification of viral loads.

298   Conclusions

The current COVID-19 pandemic has crudely demonstrated that our available methods of

detection have severe limitations in terms of cost-efficiency, scalability, and amenability

for rapid implementation. Developing and well-developed countries have experienced

severe difficulties in intensifying diagnostics, a required condition to stop the pandemic

advance in densely populated cities. Since time is the most limiting factor in emergencies,

the integration of well-developed and commercially available technologies becomes an

obvious, expedient, and cost-effective first line of defense during epidemic events. Our

research extends the validation of the miniPCR technology to the as-yet-unexplored topic

of detection of COVID-19. Furthermore, we suggest the combined use of the miniPCR and

a conventional well-plate reader as a reliable strategy that can expand the testing

capabilities of RT-qPCR.

We used the set of primers developed by the CDC and recommended by the WHO for

conducting the standard PCR diagnostics of COVID-19. These primers target three

different regions of the viral nucleic acids encoding for the N protein. In our experiments,

we corroborate that the miniPCR apparatus is capable of amplifying small amounts of

SARS-CoV-2 synthetic nucleic acids. We were able to detect and amplify 64 copies of genes encoding for the N protein of SARS-CoV-2. In the context of the COVID-19 pandemics, the use of the miniPCR thermocycler may be a valuable tool to intensify diagnostics by providing relevant advantages of higher portability, lower capital cost, and easier operation than can be achieved with other RT-qPCR platforms. We found the miniPCR® to be simple and intuitive to use; these are important attributes that would facilitate the widespread adoption of any diagnostic technology.

Moreover, the combined use of the miniPCR thermocycler and a 96-well plate reader enables the possibility of obtaining immediate readings of the amplification products, thereby providing faster (and potentially quantitative) diagnostic results in shorter times than when gel electrophoresis techniques are used. Therefore, the integration of these two already commercially available devices—a miniPCR thermocycler and a 96-well plate reader—has great potential for use during epidemic emergencies.

Acknowledgments

The authors aknowledge the funding provided by the Federico Baur Endowed Chair in Nanotechnology (0020240I03). EGG acknowledges funding from a doctoral scholarship provided by CONACyT (Consejo Nacional de Ciencia y Tecnología, México). GTdS and MMA acknowledge the institutional funding received from Tecnológico de Monterrey (Grant 002EICIS01). MMA, GTdS, SOMC and IMLM acknowledge funding provided by CONACyT (Consejo Nacional de Ciencia y Tecnología, México) through grants (SNI 26048, SNI 256730, SNI 31803, SNI 228923).

References

1. Petersen LR, Jamieson DJ, Powers AM, Honein MA. Zika Virus. Baden LR, editor. N Engl J Med. Massachussetts Medical Society; 2016;374: 1552–1563. doi:10.1056/NEJMra1602113

2. Fauci AS, Morens DM. Zika Virus in the Americas — Yet Another Arbovirus Threat. N Engl J Med. Massachussetts Medical Society; 2016;374: 601–604. doi:10.1056/NEJMp1600297

3. Kaushik A, Tiwari S, Dev Jayant R, Marty A, Nair M. Towards detection and diagnosis of Ebola virus disease at point-of-care. Biosensors and Bioelectronics. Elsevier Ltd; 2016. pp. 254–272. doi:10.1016/j.bios.2015.08.040

4. Dawood FS, Jain S, Finelli L, Shaw MW, Lindstrom S, Garten RJ, et al. Emergence of a Novel Swine-Origin Influenza A (H1N1) Virus in Humans. N Engl J Med. Massachussetts Medical Society; 2009;360: 2605–2615. doi:10.1056/NEJMoa0903810

5. Kaushik A, Tiwari S, Dev Jayant R, Marty A, Nair M. Towards detection and diagnosis of Ebola virus disease at point-of-care. Biosens Bioelectron. NIH Public Access; 2016;75: 254–72. doi:10.1016/j.bios.2015.08.040

6. Jansen van Vuren P, Grobbelaar A, Storm N, Conteh O, Konneh K, Kamara A, et al. Comparative Evaluation of the Diagnostic Performance of the Prototype Cepheid GeneXpert Ebola Assay. J Clin Microbiol. American Society for Microbiology; 2016;54: 359–67. doi:10.1128/JCM.02724-15

7. Mauk MG, Song J, Bau HH, Liu C. Point-of-Care Molecular Test for Zika Infection. Clin Lab Int. NIH Public Access; 2017;41: 25–27. Available:

http://www.ncbi.nlm.nih.gov/pubmed/28819345

8.   Duchesne L, Lacombe K. Innovative technologies for point-of-care testing of viral

hepatitis in low-resource and decentralized settings. J Viral Hepat. John Wiley &

Sons, Ltd (10.1111); 2018;25: 108–117. doi:10.1111/jvh.12827

9.   Gorbalenya AE. Severe acute respiratory syndrome-related coronavirus – The

species and its viruses, a statement of the Coronavirus Study Group. bioRxiv. Cold

Spring Harbor Laboratory; 2020; 2020.02.07.937862.

doi:10.1101/2020.02.07.937862

10.   Cohen J, Kupferschmidt K. Countries test tactics in "war" against COVID-19.

Science. American Association for the Advancement of Science; 2020;367: 1287–

1288. doi:10.1126/science.367.6484.1287

11.   Pung R, Chiew CJ, Young BE, Chin S, I-C Chen M, Clapham HE, et al. Articles

Investigation of three clusters of COVID-19 in Singapore: implications for

surveillance and response measures. Lancet. Elsevier; 2020;19: 1–8.

doi:10.1016/S0140-6736(20)30528-6

12.   Wang CJ, Ng CY, Brook RH. Response to COVID-19 in Taiwan: Big Data

Analytics, New Technology, and Proactive Testing. JAMA - Journal of the

American Medical Association. American Medical Association; 2020.

doi:10.1001/jama.2020.3151

13.   Salathé M, Althaus CL, Neher R, Stringhini S, Hodcroft E, Fellay J, et al. COVID-

19 epidemic in Switzerland: on the importance of testing, contact tracing and

isolation. Swiss Med Wkly. EMH Media; 2020;150: w20225.

doi:10.4414/smw.2020.20225

14.   Moisés Alvarez M, Trujillo-De Santiago G. Only a combination of social distancing

and massive testing can [Internet]. medRxiv. Cold Spring Harbor Laboratory Press;

2020 Jun. doi:10.1101/2020.06.23.20138743

15. Moon MJ. Fighting <scp>COVID</scp> □19 with Agility, Transparency, and

Participation: Wicked Policy Problems and New Governance Challenges. Public

Adm Rev. Blackwell Publishing Ltd; 2020;80: 651–656. doi:10.1111/puar.13214

16. Chakraborty I, Maity P. COVID-19 outbreak: Migration, effects on society, global

environment and prevention. Sci Total Environ. Elsevier B.V.; 2020;728: 138882.

doi:10.1016/j.scitotenv.2020.138882

17. Humanity tested [Internet]. Nature Biomedical Engineering. Nature Research; 2020.

pp. 355–356. doi:10.1038/s41551-020-0553-6

18. Wang W, Xu Y, Gao R, Lu R, Han K, Wu G, et al. Detection of SARS-CoV-2 in

Different Types of Clinical Specimens. JAMA - Journal of the American Medical

Association. American Medical Association; 2020. doi:10.1001/jama.2020.3786

19. Nguyen Van JC, Caméléna F, Dahoun M, Pilmis B, Mizrahi A, Lourtet J, et al.

Prospective evaluation of the Alere i Influenza A&B nucleic acid amplification

versus Xpert Flu/RSV. Diagn Microbiol Infect Dis. Elsevier Inc.; 2016;85: 19–22.

doi:10.1016/j.diagmicrobio.2015.11.012

20. Murphy J, Bustin SA. Reliability of real-time reverse-transcription PCR in clinical

diagnostics: gold standard or substandard? Expert Rev Mol Diagn. Taylor & Francis;

2009;9: 187–197. doi:10.1586/14737159.9.2.187

21. Kozel TR, Burnham-Marusich AR. Point-of-Care Testing for Infectious Diseases:

Past, Present, and Future. J Clin Microbiol. American Society for Microbiology;

2017;55: 2313–2320. doi:10.1128/JCM.00476-17

22. Su W, Gao X, Jiang L, Qin J. Microfluidic platform towards point-of-care

diagnostics in infectious diseases. J Chromatogr A. Elsevier; 2015;1377: 13–26.

doi:10.1016/J.CHROMA.2014.12.041

23. Drancourt M, Michel-Lepage A, Boyer S, Raoult D. The Point-of-Care Laboratory

in Clinical Microbiology. Clin Microbiol Rev. American Society for Microbiology;

2016;29: 429–47. doi:10.1128/CMR.00090-15

24. Qiu X, Ge S, Gao P, Li K, Yang S, Zhang S, et al. A smartphone-based point-of-care

diagnosis of H1N1 with microfluidic convection PCR. Microsyst Technol. Springer

Berlin Heidelberg; 2017;23: 2951–2956. doi:10.1007/s00542-016-2979-z

25. Petralia S, Conoci S. PCR technologies for point of care testing: Progress and

perspectives. ACS Sensors. American Chemical Society; 2017. pp. 876–891.

doi:10.1021/acssensors.7b00299

26. Ahrberg CD, Ilic BR, Manz A, Neužil P. Handheld real-time PCR device. Lab Chip.

Royal Society of Chemistry; 2016;16: 586–592. doi:10.1039/c5lc01415h

27. Bissonnette L, Bergeron M. The GenePOC Platform, a Rational Solution for

Extreme Point-of-Care Testing. Micromachines. MDPI AG; 2016;7: 94.

doi:10.3390/mi7060094

28. Marx V. PCR heads into the field. Nat Methods. 2015;12: 393–397.

doi:10.1038/nmeth.3369

29. Wolters F, van de Bovenkamp J, van den Bosch B, van den Brink S, Broeders M,

Chung NH, et al. Multi-center evaluation of cepheid xpert® xpress SARS-CoV-2

point-of-care test during the SARS-CoV-2 pandemic. J Clin Virol. Elsevier B.V.;

2020;128: 104426. doi:10.1016/j.jcv.2020.104426

30. Kwon H-S, Park H-C, Lee K, An S, Oh Y-L, Ahn E-R, et al. Performance of

MiniPCR [TM] mini8, a portable thermal cycler. Anal Sci Technol. The Korean Society

of Analytical Science; 2016;29: 79–84. doi:10.5806/AST.2016.29.2.79

31.   González-González E, Mendoza-Ramos JL, Pedroza SC, Cuellar-Monterrubio AA,

434   Márquez-Ipiña AR, Lira-Serhan D, et al. Validation of use of the miniPCR

435   thermocycler for Ebola and Zika virus detection. Ansumana R, editor. PLoS One.

436   Public Library of Science; 2019;14: e0215642. doi:10.1371/journal.pone.0215642

437   32.   Guevara EE, Frankel DC, Ranaivonasy J, Richard AF, Ratsirarson J, Lawler RR, et

438   al. A simple, economical protocol for DNA extraction and amplification where there

439   is no lab. Conserv Genet Resour. Springer Netherlands; 2018;10: 119–125.

440   doi:10.1007/s12686-017-0758-5

441   33.   Pomerantz A, Peñafiel N, Arteaga A, Bustamante L, Pichardo F, Coloma LA, et al.

442   Real-time DNA barcoding in a rainforest using nanopore sequencing: opportunities

443   for rapid biodiversity assessments and local capacity building. Gigascience. Oxford

444   University Press; 2018;7. doi:10.1093/gigascience/giy033

445   34.   Boguraev A-S, Christensen HC, Bonneau AR, Pezza JA, Nichols NM, Giraldez AJ,

446   et al. Successful amplification of DNA aboard the International Space Station. npj

447   Microgravity. 2017;3. doi:10.1038/s41526-017-0033-9

448   35.   Montague TG, Almansoori A, Gleason EJ, Copeland DS, Foley K, Kraves S, et al.

449   Gene expression studies using a miniaturized thermal cycler system on board the

450   International Space Station. Reddy S V., editor. PLoS One. Public Library of

451   Science; 2018;13: e0205852. doi:10.1371/journal.pone.0205852

452   36.   Zaky WI, Tomaino FR, Pilotte N, Laney SJ, Williams SA. Backpack PCR: A point-

453   of-collection diagnostic platform for the rapid detection of Brugia parasites in

454   mosquitoes. McCarthy JS, editor. PLoS Negl Trop Dis. Public Library of Science;

455   2018;12: e0006962. doi:10.1371/journal.pntd.0006962

37.   Gonzá Lez-Gonzá Lez E, Mendoza-Ramos JL, Pedrozaid SC, Cuellar-Monterrubio

A, Má Rquez-Ipiña AR, Lira-Serhanid D, et al. Validation of use of the miniPCR

thermocycler for Ebola and Zika virus detection. 2019;

doi:10.1371/journal.pone.0215642

38.   Chin CD, Linder V, Sia SK. Commercialization of microfluidic point-of-care

diagnostic devices{. doi:10.1039/c2lc21204h

Supporting Information

S1 File. Excel file that contains raw data

S1 Figure. Uncropped and unadjusted images of gels presented in Figure 2A: Images of
agarose gel electrophoresis of the DNA amplification product generated by targeting a
region of the sequence coding for SARS-CoV-2 N protein. PCR was performed using a
miniPCR® thermocycler. The primer set N1 was used. The initial template amount was
gradually increased from left to right: negative control (lane 1), 625 copies (lane 2), $2.5 \times$
10^3 (lane 3), 1.0×10^4 (lane 4), repetition of 1.0×10^4 (lane 5), 4.0×10^4 (lane 6), 2.0×10^5
DNA copies (lane 7), and molecular weight ladder (lane 8).

S2 Figure. Uncropped and unadjusted images of gels presented in Figure 2B: Images of
agarose gel electrophoresis of the DNA amplification product generated by targeting a
region of the sequence coding for SARS-CoV-2 N protein. PCR was performed using a
miniPCR® thermocycler. The primer set N2 was used. The initial template amount was
gradually increased from left to right: negative control (lane 1), 625 copies (lane 2), $2.5 \times$
10^3 (lane 3), 1.0×10^4 (lane 4), repetition of 1.0×10^4 (lane 5), 4.0×10^4 (lane 6), 2.0×10^5
DNA copies (lane 7), and molecular weight ladder (lane 8).

S3 Figure. Uncropped and unadjusted images of gels presented in Figure 2C: Images of
agarose gel electrophoresis of the DNA amplification product generated by targeting a
region of the sequence coding for SARS-CoV-2 N protein. PCR was performed using a
miniPCR® thermocycler. The primer set N3 was used. The initial template amount was
gradually increased from left to right: negative control (lane 1), 625 copies (lane 2), $2.5 \times$
10^3 (lane 3), 1.0×10^4 (lane 4), repetition of 1.0×10^4 (lane 5), 4.0×10^4 (lane 6), 2.0×10^5
DNA copies (lane 7), and molecular weight ladder (lane 8).

Figure and captions:

Figure 1. **Equipment and initiators for SARS-CoV-2 detection using a miniPCR**. A) The miniPCR® thermocycler. (B) The blueGel® electrophoresis chamber: blueGel® allows visualization of a 15 ml agarose gel using an integrated blue LED array. (C) Agarose gel electrophoresis of the SARS-CoV-2 amplification products. (D) Commercial 96-well plate with COVID-19 controls (artificial DNA samples). (E) Temperature cycling scheme used in our PCR protocol. (F) Three different sets of primers were used to target a gene sequence encoding the SARS-CoV-2 N protein.

Figure 2. Evaluation of the sensitivity of the combined use of a miniPCR® thermal cycler (for amplification) and a plate reader (for determination of the amplification extent). (A-C) Sensitivity trials using different concentrations of the template (positive control) and three different primers sets (A) N1, indicated in blue; (B) N2, indicated in red; and (C) N3, indicated in yellow. Images of agarose gel electrophoresis of the DNA amplification product generated by targeting three different regions of the sequence coding for SARS-CoV-2 N protein. PCR was performed using a miniPCR® thermocycler. Three different primer sets were used (N1, N2, and N3). The initial template amount was gradually

increased from left to right: negative control (lane 1), 625 copies (lane 2), 2.5×10^3 (lane

3), 1.0×10^4 (lane 4), repetition of 1.0×10^4 (lane 5), 4.0×10^4 (lane 6), 2.0×10^5 DNA

copies (lane 7), and molecular weight ladder (lane 8). (D-F) Determination of fluorescence,

as measured in a commercial plate reader, for different dilutions of SARS-CoV-2 synthetic

DNA templates. Results using three different primer sets are shown: (A) N1, indicated in

blue; (B) N2, indicated in red; and (C) N3, indicated in yellow. (G) Summary and

comparison of fluorescence readings form synthetic samples of SARS-CoV-2 in a wide

span of dilutions. Results using three different primer sets are shown: (A) N1, indicated in

blue; (B) N2, indicated in red; and (C) N3, indicated in yellow.

Figure 3. Potential use of a plate reader for quantitation of the initial viral load in a sample

and the extent of amplification. (A) Linear relationship between the natural logarithm of the

fluorescence reading and the natural logarithm of the viral load. (B) Results of the

simulation of real time PCR in a microplate reader.