Scaling diagnostics in times of COVID-19: Colorimetric Loop-mediated Isothermal Amplification (LAMP) assisted by a 3D-printed incubator for cost-effective and scalable detection of SARS-CoV-2

Everardo González-González[1,2], Itzel Montserrat Lara-Mayorga[1,3], Iram Pablo Rodríguez-Sánchez[4,5], Felipe Yee-de León[6], Andrés García-Rubio[1,3], Carlos Ezio Garciaméndez-Mijares[1,3,7], Gilberto Emilio Guerra-Alvarez[1,3], Germán García-Martínez[1,3,7], Juan Andrés Aguayo-Hernández[1,3], Eduardo Márquez-García[8], Yu Shrike Zhang[7], Sergio O. Martínez-Chapa[3], Joaquín Zúñiga[9,10], Grissel Trujillo-de Santiago[1,3,*], Mario Moisés Alvarez[1,2*]

[1] Centro de Biotecnología-FEMSA, Tecnologico de Monterrey, CP 64849, Monterrey, NL, México

[2] Departamento de Bioingeniería, Tecnologico de Monterrey, CP 64849, Monterrey, NL, México

[3] Departamento de Ingeniería Mecátrónica y Eléctrica, Tecnologico de Monterrey, CP 64849, Monterrey, NL, México

[4] Universidad Autónoma de Nuevo León, Facultad de Ciencias Biológicas, Laboratorio de Fisiología Molecular y Estructural. 66455, San Nicolás de los Garza, NL, México.

[5] Alfa Medical Center, Guadalupe, CP 67100, NL, México.

[6] Delee Corp. Mountain View 94041, CA, USA

[7] Division of Engineering in Medicine, Department of Medicine, Brigham and Women's Hospital, Harvard Medical School, Cambridge 02139, MA, USA

[8] Laboratorio de Inmunobiología y Genética, Instituto Nacional de Enfermedades Respiratorias Ismael Cosío Villegas, Cd. de México, México.

[9] Unidad de Biología Molecular, Instituto Nacional de Enfermedades Respiratorias Ismael Cosío Villegas, Cd. de México, México.

[10] Escuela de Medicina y Ciencias de la Salud, Tecnologico de Monterrey, CP 14380, Cd. de México, México

*Corresponding authors. E-mails: mario.alvarez@tec.mx; grissel@tec.mx;

Abstract

By the third week of June 2020, more than 8,500,000 positive cases of COVID-19 and
more than 450,000 deaths had been officially reported worldwide. The COVID-19
pandemic arrived in Latin America, India, and Africa—territories in which the mounted
infrastructure for diagnosis is greatly underdeveloped. Here, we demonstrate the combined
use of a three-dimensional (3D)-printed incubation chamber for commercial Eppendorf
PCR tubes, and a colorimetric embodiment of a loop-mediated isothermal amplification
(LAMP) reaction scheme for the detection of SARS-CoV-2 nucleic acids. We used this
strategy to detect and amplify SARS-CoV-2 genetic sequences using a set of in-house
designed initiators that target regions encoding the N protein. We were able to detect and
amplify SARS-CoV-2 nucleic acids in the range of 62 to 2×10^5 DNA copies by this
straightforward method. Using synthetic SARS-CoV-2 samples and a limited number of
RNA extracts from patients, we also demonstrate that colorimetric LAMP is a quantitative
method comparable in diagnostic performance to RT-qPCR. We envision that LAMP may
greatly enhance the capabilities for COVID-19 testing in situations where RT-qPCR is not
feasible or is unavailable. Moreover, the portability, ease of use, and reproducibility of this
strategy make it a reliable alternative for deployment of point-of-care SARS-CoV-2
detection efforts during the pandemics.

Introduction

By the end of the third week of June 2020, more than 8.5 million positive cases of COVID-

19 were officially reported across the globe[1]. Even developed countries, such as the USA,

England, France, and Germany, are still struggling to mitigate the propagation of SARS-

CoV-2 by implementing social distancing and widespread testing. Less developed regions,

such as Latin America, India, and Africa, are now experiencing the arrival of COVID-19,

but these—territories are woefully lacking in the finances or the mounted infrastructure for

diagnosis of this pandemic infection. Rapid and massive testing of thousands of possibly

infected subjects has been an important component of the strategy of the countries that are

effectively mitigating the spreading of COVID-19 among their populations (i.e., China[2],

South Korea [3], and Singapore [4]). By comparison, developing countries with high

demographic densities, such as México [5], India [6], or Brazil [7], may not be able to

implement a sufficient number of centralized laboratories for rapid large-scale testing for

COVID-19.

Many methodologies have been proposed recently to deliver cost-effective diagnosis (i.e.,

those based on immunoassays [8–11] or specific gene hybridization assisted by CRISPR-

Cas systems [12–14]). While immunoassays are an accurate and efficacious tool for

assessing the extent of the infection for epidemiological studies [15], their usefulness is

limited to the identification of infected subjects during early phases of infection [11,16], a

critical period for infectiveness. For instance, experimental evidence collected from a small

number of COVID-19 patients (9 subjects) showed that 100% of them produced specific

immunoglobulins G (IgGs) for SARS-CoV-2 within two weeks of infection, but only 50%

of them did during the first week post infection [17].

Nucleic acid amplification continues to be the gold standard for the detection of viral

diseases in the early stages [18–22], and very small viral loads present in symptomatic or

asymptomatic patients can be reliably detected using amplification based technics, such as

polymerase chain reaction (PCR) [23–25], recombinase polymerase amplification

(RPA)[26], and loop-mediated isothermal amplification (LAMP) [27–29].

During the last two pandemic events with influenza A/H1N1/2009 and COVID-19, the

Centers for Disease Control and Prevention (CDC) and the World Health Organization

(WHO) recommended real-time quantitative PCR (RT-qPCR) methods as the gold standard

for official detection of positive cases[16,30]. However, the reliance on RT-qPCR often

leads to dependence on centralized laboratory facilities for testing [16,30–33]. To resolve

this drawback, isothermal amplification reaction schemes (i.e., LAMP and RPA) have been

proposed as alternatives to PCR-based methods and devices for point-of-care settings

[32,34,35]. The urgency of using reliable molecular-based point of care (POC) methods for

massive diagnostic during epidemiological emergencies has become even more evident

during the current COVID-19 pandemics [30,36,37].

In these times of COVID-19 [38], scientists and philanthropists around the globe have

worked expeditiously on the development of rapid and portable diagnostics for SARS-

CoV-2. Several reports have demonstrated the use of colorimetric LAMP-based methods

for diagnosis of pandemic COVID-19 [39–44]. Some of these reports (currently available

as preprints) use phenol red, a well-known pH indicator, to assist in the visual

discrimination between positive and negative samples [39,40,45].

In this study, we demonstrate the use of a simple embodiment of a colorimetric LAMP

protocol for the detection and amplification of synthetic samples and actual RNA samples

from patients of SARS-CoV-2, the causal viral agent of COVID-D. In this LAMP-based

strategy, also assisted by the use of phenol red, sample incubation is greatly facilitated by

the use of a three-dimensional (3D)-printed incubator connected to a conventional water

circulator, while discrimination between positive and negative samples is achieved by

visual inspection. We quantitatively analyze differences in color between positive and

negative samples using color decomposition and analysis in the color CIELab space[46].

Moreover, we compare the sensibility of this LAMP colorimetric method versus PCR

protocols. This simple strategy is potentially adequate for the fast deployment of diagnostic

efforts in the context of COVID-19 pandemics.

Rationale

We have developed a simple diagnostic method for the detection of SARS-CoV-2, the

causal agent of COVID-19. The rationale underlying this strategy is centered on achieving

the simplest possible integration of easily available reagents, materials, and fabrication

techniques to facilitate fast and massive implementation during the current COVID-19

pandemics in low- or middle-income regions. This method is based on the amplification of

the genetic material of SARS-CoV-2 using LAMP. The amplification is conducted using a

commercial reaction mix in commercial and widely available 200 µL Eppendorf PCR

tubes. Moreover, we have designed and fabricated a simple 3D-printed chamber (Figure 1)

for incubation of the Eppendorf tubes and to enable LAMP at high temperatures (50–65 ℃)

and extended times (up to 1 h). We show that this incubation chamber, when connected to a

conventional water recirculator, enables the successful amplification of positive samples

(i.e., samples containing SARS-CoV-2 nucleic acids).

This incubation chamber is one of the key elements that enable rapid and widespread

implementation of this diagnostic method at low cost. This 3D-printed incubator can be

rapidly printed using standard SLA printers widely available in markets worldwide.

Standard 3D-printing resins can be used. The availability of the original AutoCAD files

(included here as supplemental material) enables fast modification/optimization of the

design for accommodation of a larger number of samples or larger or smaller tubes,

adaptation to any available hoses (tubing), and possible incorporation of an on-line color-

reading system.

Figure 1. Experimental setup. (A) Commercial 200 microliter Eppendorf PCR tubes, and (B) a
3D-printed incubator was used in amplification experiments of samples containing the synthetic
SARS-CoV-2 nucleic acid material. (C) 3D CAD model of the LAMP reaction incubator. (D)
Actual image of the Eppendorf tube incubator connected to a conventional water circulator.

Indeed, all this is consistent with the main rationale of our proposed diagnostic strategy for

pandemic COVID-19: To enabling a fast and feasible response using widespread,

distributed, and scalable diagnostics fabricated with widely available resources.

In the following section, we briefly discuss the mechanisms of amplification and visual

discrimination between positive and negative samples.

Colorimetric LAMP amplification

The presence of phenol red within the LAMP reaction mix allows for naked-eye

discrimination between positive and negative samples (Figure 2). The reaction mix is

coupled with the pH color transition of phenol red, a widely used pH indicator, which shifts

in color from red to yellow at pH 6.8. During LAMP amplification, the pH of the reaction

mix continuously evolves from neutrality to acidic values as protons are produced [27,47].

The mechanism of production of hydrogen ions (H^+) during amplification in weakly

buffered solutions has been described [47]. DNA polymerases incorporate a

deoxynucleoside triphosphate into the nascent DNA chain. During this chemical event, a

pyrophosphate moiety and a hydrogen ion are released as byproducts (Figure 2A). This

release of hydrogen ions is quantitative, according to the reaction scheme illustrated in

Figure 2. The caudal of H^+ is high, since it is quantitatively proportional to the number of

newly integrated dNTPs. In fact, the quantitative production of H^+ is the basis of previously

reported detection methods, such as the semiconductor sequencing technology operating in

Ion Torrent sequencers[48]. In the initially neutral and weakly buffered reaction mixes, the

production of H^+ during LAMP amplification progressively and rapidly shifts the pH across

the threshold of phenol red (Figure 2B).

Figure 2. **Initiators and pH indicator for SARS-Co2 detection using a colorimetric LAMP
method**. (A) The LAMP reaction scheme. (B) Chemical structure of phenol red. (C) Two different
sets of LAMP primers were used for successfully targeting a gene sequence encoding the SARS-
Co2 N protein. Successful targeting and amplification are clearly evident to the naked eye: positive
samples shift from red to yellow.

Moreover, the pH shift is clearly evident to the naked eye, thereby freeing the user from

reliance on spectrophotometric instruments and facilitating simple implementation during

emergencies (Figure 2C). Images in Figure 2C show representative colors of the

amplification reaction mixes contained in Eppendorf PCR tubes after incubation for 30

174   min. Three different incubation temperatures were tested (50, 60, and 65 ºC) and two

175   different sets of LAMP-primers (α and β) were used (Table 1).

177   **Table 1. Primer sequences used in LAMP amplification experiments.** Two different sets of
178   primers were used, directed at the RNA sequence encoding the N sequence of the SARS-CoV-2.

Set	Description	Primers Sequence (5'>3')
Primer set α	2019-nCoV 1-F3	TGGACCCCAAAATCAGCG
	2019-nCoV 1-B3	GCCTTGTCCTCGAGGGAAT
	2019-nCoV 1-FIP	CCACTGCGTTCTCCATTCTGGTAAATGCACCCCGCATTACG
	2019-nCoV 1-BIP	CGCGATCAAAACAACGTCGGCCCTTGCCATGTTGAGTGAGA
	2019-nCoV 1-LF	TGAATCTGAGGGTCCACCAA
	2019-nCoV 1-LB	TTACCCAATAATACTGCGTCTTGGT
Primer set β	2019-nCoV 2-F3	CCAGAATGGAGAACGCAGTG
	2019-nCoV 2-B3	CCGTCACCACCACGAATT
	2019-nCoV 2-FIP	AGCGGTGAACCAAGACGCAGGGCGCGATCAAAACAACG
	2019-nCoV 2-BIP	AATTCCCTCGAGGACAAGGCGAGCTCTTCGGTAGTAGCCAA
	2019-nCoV 2-LF	TTATTGGGTAAACCTTGGGGC
	2019-nCoV 2-LB	TAACACCAATAGCAGTCCAGATGA

Both sets of primers performed equivalently, at least based on visual inspection, in the three

temperature conditions tested. Discrimination between positive and negative controls is

possible using only the naked eye to discern the reaction products from amplifications

conducted at 60 and 65 ºC. No or negligible amplification was evident at 50 ºC or in the

control group.

Furthermore, we were able to successfully discriminate between positive and negative

samples using LAMP reaction mix already added with primers and kept at 20 °C or 4 °C

for 24, 48, 72, and 96 h (Figure S5). The stability of the reaction, the isothermal nature of

the amplification process, and its independence from specialized equipment greatly

simplifies the logistic of implementation of this diagnostic method outside centralized labs.

Analysis of sensitivity

We conducted a series of experiments to assess the sensitivity of the LAMP reactions in the

3D-printed incubation chamber using the two sets of primers (α and β; Table 1). The

amplification proceeds with sufficient quality to also allow proper visualization of the

amplification products in electrophoresis gels, even at low nucleic acid concentrations. We

observed that amplification proceeded successfully in a wide range of viral loads, from 625

to 5×10^5 copies in experiments using synthetic SARS-CoV-2 nucleic acid material (Figure

3A). We clearly observed amplification in samples containing as few as 625 viral copies

after incubation times of 50 min at 65 ℃. If we put this range into a proper clinical context,

the actual viral load of COVID-19 in nasal swabs from patients has been estimated to fall

within the range of 10^5 to 10^6 viral copies per mL [49]. Discrimination between positive

and negative samples (controls) can be clearly established by the naked eye in all reactions

incubated for 50 min, regardless of the number of viral copies present. In addition, we did

not observe any non-specific amplification in negative samples (i.e., containing synthetic

genetic material form EBOV) incubated for 50 min at 65 ℃. Indeed, the identification and

amplification of SARS-CoV-2 synthetic material is feasible in samples that contained ~62.5

viral copies using this LAMP strategy (Figure S3) and incubation times of 50–60 min.

Figure 3. **Two different sets of LAMP-primers were used for successfully targeting of a gene sequence encoding the SARS-Co2 N protein**. (A) LAMP primer sets α and β both enable the amplification of synthetic samples of SARS-CoV-2 nucleic acids in a wide range of template concentrations, from 625 to 2.0×10^5 DNA copies of SARS-CoV-2 when incubated for 50 minutes at a temperature range from 60 to 65 ºC. (B, C) Agarose gel electrophoresis of DNA amplification products generated by targeting two different regions of the sequence coding for SARS-Co2 N protein. Two different primer sets were used: (B) primer set α, and (C) primer set β. The initial template amount was gradually decreased from left to right: 2.0×10^5 DNA copies (lane 1), 4.0×10^4 copies, (lane 2), 1.0×10^4 copies (lane 3), 2.5×10^3 copies (lane 4), 625 copies (lane 5), negative control (lane 6), and molecular weight ladder (lane 7). Panel B and C corresponds to portions of the full-length gels presented in supplemental figures S8A and S8B, respectively.

We corroborated the amplification by visualizing LAMP products with gel electrophoresis for the different viral loads tested. Figures 3B, C show agarose gels of the amplification products of each one of the LAMP experiments, where two different sets of primers (α and β) were used to amplify the same range of concentrations of template (from 625 to 2×10^5

synthetic viral copies). We were able to generate a visible array of bands of amplification

products, a typical signature of LAMP, for both LAMP primer sets and across the whole

range of synthetic viral loads. Indeed, both primer sets rendered similar amplification

profiles.

In summary, using the primers and methods described here, we were able to consistently

detect the presence of SARS-CoV-2 synthetic nucleic acids. We have used a simple 3D-

printed incubator, connected to a water circulator, to conduct LAMP. We show that, after

only 30 min of incubation, samples containing a viral load in the range of 10^4 to 10^5 copies

could be clearly discriminated from negative samples by visual inspection with the naked

eye (Figure 2C). Samples with a lower viral load were clearly discriminated when the

LAMP reaction was incubated for 50 min. Incubation periods of up to 1 h at 68 ºC did not

induced false positives and were able to amplify as few as ~62 copies of SARS-CoV-2

synthetic genetic material. These results are consistent with those of other reports in which

colorimetric LAMP, assisted by phenol red, has been used to amplify SARS-COV-2

genetic material [39,40]. We observe 0 false positive cases in experiments where synthetic

samples containing EBOV genetic material were incubated at 65 ºC for 1 h.

In the current context of the COVID-19 pandemics, the importance of communicating this

result does not reside in its novelty but in its practicality. Some cost considerations follow.

While the market value of a traditional RT-qPCR apparatus (the current gold standard for

COVID-19 diagnostics) is in the range of 10,000 to 40,000 USD, a 3D-printed incubator,

such as the one described here (Figure S1,S2; Supplementary file S1), could be fabricated

for under 200 USD at any 3D printer shop. This difference is significant, especially during

an epidemic or pandemic crisis when rational investment of resources is critical. While the

quantitative capabilities of testing using an RT-qPCR platform are undisputable, the

capacity of many countries to rapidly, effectively, and massively establish diagnostic

centers based on RT-qPCR is questionable. The current pandemic scenarios experienced in

the USA, Italy, France, and Spain, among others, have crudely demonstrated that

centralized labs are not an ideal solution during emergencies. Portable diagnostic systems

may provide a vital flexibility and speed of response that RT-qPCR platforms cannot

deliver.

Feasibility of real-time quantification

Here, we further illustrate the deterministic and quantitative dependence between the

concentration of the amplification product and the color signal produced during this

colorimetric LAMP reaction. For this purpose, we simulated real-time amplification

experiments by conducting a series of amplification reactions using initial amounts of 625,

1×10^4, and 2×10^5 copies of synthetic SARS-CoV-2 genetic material in our 3D-printed

incubator.

We extracted samples from the incubator after 0, 10, 20, 30, 40, and 50 min of incubation at

65 ºC. The color of these samples was documented as images captured using a smart phone

(iPhone 7) against a white background (Figure 4A). The images were analyzed using the

free access application Color Companion® for the iPhone or iPad. Briefly, color images

were decomposed into their CIELab space components. In the CIELab color space, each

color can be represented as a point in a 3D-space, defined by the values **L**, **a**, and **b** [46]. In

this coordinate system, **L** is the luminosity (which ranges from 0 to +100), **a** is the blue-

yellow axis (which ranges from -50 to 50), and **b** is the green-red axis (which ranges from -

50 to 50) (Figure S4).

Figure 4. Evaluation of the sensitivity of the combined use of a colorimetric LAMP method assisted by the use of phenol red. (A) Sensitivity trials using different concentrations of the template (positive control) and two different primers sets: α (indicated in blue) and β (indicated in red). Photographs of the Eppendorf PCR tubes containing positive samples and negative controls were acquired using a smartphone. (C, D) Distance in the color CIELab space between negative controls (red) and samples containing different concentrations of SARS-CoV-2 nucleic acid material (i.e., 625, 10000, and 200000 synthetic copies) analyzed after different times of incubation (i.e., 10, 20, 30, 40, and 50 minutes) at 65 °C. The analysis of color distances is presented for amplifications conducted using primer set (B) α and (C) β.

The difference between two colors can be quantitatively represented as the distance between the two points that those colors represent in the CIELab coordinate system. For the colorimetric LAMP reaction mixture used in our experiments, the spectrum of possible colors evolves from red (for negative controls and negative samples) to yellow (for positive samples). Conveniently, the full range of colors for samples and controls can be represented

in the red and yellow quadrant defined by L [0,100], a [0,50], and b [0,50]. For instance, the

difference between the color of a sample (at any time of the reaction) and the color of the

negative control (red; L=53.72 ± 0.581, a=38.86 ± 2.916, and b=11.86 ± 0.961) can be

calculated in the CIELab space. We determined the distance in the CIELab space between

the color of samples taken at different incubation times that contained SARS-CoV-2

genetic material and negative controls (Figure 4B, C). We repeated this calculation for each

of the LAMP primer sets that we used, namely primer set α (Figure 4B) and β (Figure 4C).

These results suggest that the color difference between the samples and negative controls is

quantifiable. Therefore, color analysis may be implemented to assist the discrimination

between positives and negatives. Furthermore, imaging and color analysis techniques may

be implemented in this simple colorimetric LAMP diagnostic strategy to render a real-time

quantitative Lamp (RT-qLAMP).

Alternatively, the progression of the amplification at different times can be monitored by

adding an intercalating DNA agent (i.e., EvaGreen Dye), and measuring fluorescence on

time (Figure S5).

Note that the variance coefficients for the control are 1.08, 7.50, and 8.10% for L, a, and b,

respectively. These small values suggest robustness and reproducibility in the location of

the coordinates of the control point (reference point). Similarly, the variation in color

between negative controls and positive samples incubated for 50 min was reproducible and

robust (average of 46.60 +/- 4.02 d.u.; variance coefficient of 8.62%).

Figure 5. Time progression of the distance in color with respect to negative controls (red color) in the CIELab space for positive SARS-CoV-2 samples containing 625 (light blue, ▪), 1×10^4 (medium blue, ▪), and 2.5×10^6 (dark blue, ▪) copies of synthetic of SARS-CoV-2 nucleic acids. Results obtained in experiments using (A) primer set α, and (B) primer set β. (C) Comparison between the performance of PCR and LAMP in a simulated real-time experiment. Progression of the fluorescence signal, as measured in a plate reader, in PCR (black circles) and LAMP (red squares) experiments. The inset shows a zoom at the exponential stage of the amplification process.

Interestingly, we observed significant differences in the performance of the two LAMP

primer sets used in the experiments reported here (Figures 4B and 5).

Our results suggest that primer set α enabled faster amplification in samples with fewer

viral copies. Consistently, this primer set yielded positive discrimination in samples with

625 viral copies in 30 min (Figure 4B). The use of primer set β enabled similar differences

in color, measured as distances in the CIELab 3D-space, in 40 min (Figure 4C).

These findings suggest that primer set α should be preferred for final-point implementations

of this colorimetric LAMP method. Interestingly, primer set β may better serve the purpose

of a real-time implementation. While primer set α produced similar trajectories of evolution

of color in samples that contained 1.0×10^4 and 2.0×10^5 viral copies (Figure 5A), primer

set β was better at discriminating between amplifications produced from different initial

viral loads (Figure 5B).

Comparison of LAMP versus PCR

LAMP has been regarded before as a more efficient amplification reaction than PCR, since

more DNA is produced per unit of time due to the use of a higher number of primers[50]

(in this case 6 versus 2). To finalize our analysis, we simulated some real-time

amplification experiments to compare the performance of LAMP and PCR in similar

conditions (Figure 6A). To that end, we conducted amplification reactions using initial

amounts of 4×10^4 copies of synthetic SARS-CoV2 in a commercial miniPCR

cycler[24,51] (using primer set N1) and in our LAMP 3D-printed incubator (using primer

set β). We added the intercalating agent, EvaGreen® Dye, to the reaction mix at the initial

time and extracted samples after 0, 7, 14, 21, 28, 35, 42, and 51 minutes. These samples

form PCR and LAMP experiments were dispensed in 96-weel plates. The fluorescence

from these samples was then measured in a commercial plate reader[24] (Figure 5C). We

observed an exponential increase in fluorescence as more LAMP or PCR cycles were

performed, which highlights the quantitative nature of the intercalating reaction. The

LAMP reaction produces significantly higher fluorescence signals that the PCR reaction

throughout the entire reaction time. The difference between the fluorescence emissions of

both amplifications is more evident after the first 20 minutes of amplification. These

results also suggest that using a commercial plate reader to determine the extent of advance

of LAMP amplifications is a practical and reliable alternative to the use of colorimetric

evaluations. Moreover, fluorescence reading of LAMP products may lead to precise

quantification of SARS-CoV-2 viral loads.

We also compared the performance of RT-qPCR and colorimetric LAMP using actual RNA

extracts isolated from human volunteers. For this purpose, first we used colorimetric

LAMP for diagnosis of one RNA sample confirmed as positive for COVID-19 and one

confirmed as negative according to RT-qPCR results. RNA extracts from the COVID-19

(+) patient were clearly discriminated from the COVID-19 (-) patient extracts by our

colorimetric LAMP amplifications (Figure 6A).

Similar results were obtained regardless of the LAMP primer set used (i.e., α and β). We

corroborated our LAMP amplification results using standard gel electrophoresis (Figure

6B). In addition, samples were serially diluted to challenge the sensitivity of colorimetric

LAMP. We were able to discriminate between positive and negative samples in the entire

concentration range tested (300 ng of total RNA, as determined by nanoDrop assays). The

color shift (red to yellow) was clearly perceived after 30 minutes of amplification in

samples containing 300 ng of total RNA from COVID (+) patients. Samples containing 30

and 3 ng of total RNA required longer times (Figure 6C).

Figure 6. Progression of color changes during amplification in actual RNA extracts from patients.
(A) RNA extracts from COVID-19(+) and COVID-19(-) samples, amplified by colorimetric
LAMP, can be easily discriminated by visual inspection. (B) LAMP amplification products from
RNA (lane 1 and 2) and RNA extracts (lane 3 and 4) from a COVID (+) patient, and a COVID (-)
volunteer (lane 5), as revealed by gel electrophoresis experiments. A molecular weight ladder is
shown in lane 6. Panel B corresponds to a portion of the full-length gel presented in Supplemental
figures (Figure S9). (C) Time progression of color changes in LAMP reaction mixes containing 300
ng of RNA extract from a COVID(-) volunteer (as diagnosed by RT-qPCR), and 3, 30, and 300 ng
of RNA extract from a COVID(+) patient (as diagnosed by RT-qPCR). (D) Distance in color with
respect to negative controls (red color) in the CIELab space for RNA extracts from a COVID(-)
volunteer (as diagnosed by RT-qPCR) containing 300 ng of nucleic acids, and a COVID(+) patient
(as diagnosed by RT-qPCR) containing 3, 30, and 35 (dark blue, ■) ng of nucleic acids. Readings at
0, 30, and 60 minutes are shown. A suggested positive–negative threshold value is indicated with a
red line. (E) Time progression of the distance in color with respect to negative controls (red color)
in the CIELab space for RNA extracts from a COVID(-) volunteer (as diagnosed by RT-qPCR)

containing 300 ng of nucleic acids (red, ■), and from a COVID(+) patient (as diagnosed by RT-

qPCR) containing 3 (light blue, ■), 30 (medium blue, ■), and 35 (dark blue, ■) ng of nucleic acids.

A suggested positive–negative threshold value is indicated with a red line.

Positive samples exhibited a shift in color after 60 minutes of amplification, while negative

samples remained unchanged. We quantified the change in color in positive and negative

samples using color image analysis and by calculating color distances in the CieLab color

space (Figure 6D,E).

Our experiments show that the distance in color between positive and negative RNA

samples from human volunteers is proportional to the number of viral copies. These results

suggest that the change in color can be quantitatively related to the viral load of SARS-

CoV-2 in actual RNA extracts, similarly to synthetic samples.

In a final set of experiments, we blind tested a set of 8 extracts of human RNA from

nasopharyngeal samples corresponding to 2 patients that were diagnosed as COVID-19 (-)

and 6 patients diagnosed as COVID-19 (+) by RT-qPCR. We adjust the RNA content in all

samples to 300 ng/µL of RNA and then diluted them to derive samples containing 30

ng/µL. All samples, undiluted and diluted, were added with the LAMP reactive mix and

incubated at 65 ° C by 50 minutes. All samples exhibit a red color before incubation

(Figure 7A), and only positive samples shifted to yellow during incubation (Figure 7B). We

confirmed results by gel electrophoresis of the amplification products. Only positive

samples exhibited the characteristic DNA profile associated with LAMP products (Figure

7C,D). Colorimetric LAMP was also able to discriminate positive samples correctly even in

diluted extracts containing one order of magnitude less RNA than original extracts.

COVID-19 positive RNA samples, original or diluted, showed similar values of distance in

color with respect to negative samples, although standard deviations were higher in samples

that contained 30 ng/µL than in samples that contained 30 ng/µL (Figure 7E).

In this reduced set of extracts from nasopharyngeal patient samples, diagnostic results from

colorimetric LAMP were completely consistent with RT-qPCR results.

Figure 7. Discrimination of actual RNA extracts from COVID-19 positive and negative samples.
Color of RNA extracts from 6 COVID-19(+) and 2 COVID-19(-) samples, at two different
concentrations (300 and 30 ng/µL) (A) before, and (B) after colorimetric LAMP reaction. COVID-

19 ositive samples (S1, S2, S3. S4. S6, and S8) can be easily discriminated by visual inspection. (C)

Distance in color of samples of RNA extracts with respect to negative samples (S5 and S7) in the

CIELab space. Distances in color of samples containing 300 ng of nucleic acids (orange bars), or 30

ng of nucleic acids (yellow bars) are presented. (D-E) LAMP amplification products from RNA

extracts containing (D) 300 ng/μL and (E) 30 ng/ μL from the same set of COVID (+) patients (S1-

S4, S6 and S8), and COVID (-) volunteers (S5 and S7), as revealed by gel electrophoresis

experiments. Lanes 1 to 8 contained amplification products from samples S1 to S8. Lane 9 was

reserved for the molecular weight ladder. Panel D and E corresponds to portions of the full-length

gels presented in supplemental figures S10A and S10B, respectively.

Moreover, discrimination of positive samples even in diluted samples suggests that this

colorimetric technique may be useful even in situations where the amount of RNA

extracted is low due to improper sampling/extraction or degradation during transportation.

Concluding remarks

The challenge of point-of-care detection of viral threats is of paramount importance,

particularly in underdeveloped regions and in emergency situations (i.e., epidemic

outbreaks). In the context of the current COVID-19 pandemic, the availability of testing

infrastructure based on RT-qPCR is recognized as a serious challenge around the world. In

developing economies (i.e. Latin America, India, and most African countries), the currently

available resources for massive COVID-19 testing by RT-qPCR will clearly be insufficient.

Even in developed countries, the time to get diagnostic RT-qPCR results from a COVID-19

RT-qPCR test currently ranges from 1 to 5 days.

Clearly, the available PCR labs are overburdened with samples, have too few personnel to

conduct the tests, are struggling with backlogs on the instrumentation, and face complicated

logistics to transport delicate and infective samples while preserving the cold chain.

Here, we have demonstrated that a simple embodiment of a LAMP reaction, assisted by the

use of phenol red as a pH indicator and the use of a simple 3D-printed chamber connected

to a water circulator can enable the rapid and highly accurate identification of samples that

contain artificial SARS-CoV-2 genetic sequences. We also showed, using synthetic SARS-

CoV-2 and a limited number of RNA extracts from patients, that colorimetric LAMP is a

quantitative method, comparable to RT-qPCR. Amplification is visually evident, without

the need for any additional instrumentation, even at low viral copy numbers. In our

experiments with synthetic samples, we observed 100% accuracy in samples containing as

few as 625 copies of SARS-CoV-2 genetic material.

Validation of these results using a larger number of real human samples from positive and

negative COVID-19 subjects is obviously needed to obtain a full assessment of the

potential of this strategy as an alternative to RT-qPCR platforms. However, our results with

synthetic samples and with a reduced number of samples containing RNA from human

volunteers (8 samples) suggest that this simple strategy may greatly enhance the

capabilities for COVID-19 testing in situations where RT-qPCR is not feasible or is

unavailable. Recently, other groups have shown that accurate discrimination between

COVID-19 positive and negative samples is positive in extraction-free RT-qPCR[52] and

colorimetric LAMP implementations[45]. Smyrlaki et al. extensively investigated the

performance of an extraction-free RT-qPCR protocol in which saliva or nasopharyngeal

samples were heated at 95 °C for 5 minutes for virus inactivation and then directly

amplified. Lalli et al. showed successful amplification using a similar colorimetric LAMP

protein in which they heated saliva samples at 50 or 64 °C for 5 minutes optionally adding

protease K. These results suggest that extraction-free implementations of amplification

methods, including colorimetric LAMP, can successfully identify COVID-19 positive

patients from nasopharyngeal and saliva samples. Altogether, this body of evidence

suggests that extraction-free colorimetric LAMP provides means for cost-effective massive

diagnostics of SARS-CoV-2 and is a promising tool for pandemic contention that deserves

further exploration.

Materials and Methods

Equipment specifications: We ran several hundred amplification experiments using a

colorimetric LAMP method in a 3D-printed incubation chamber designed in house and

connected to a conventional water circulator (Figure 1). The design and all dimensional

specifications of this chamber have been made available in Supplementary Information

(Figures S1, S2; Supplementary File S1). In the experiments reported here, we used a

chamber with dimensions of $20 \times 5 \times 15$ cm^3 and a weight of 0.4 kg (without water). A

conventional water circulator (WVR, PA, USA), was used to circulate hot water (set point

value at 76 ºC) through the 3D-printed chamber for incubation of the Eppendorf PCR tubes

(0.2 mL). In this first chamber prototype, twelve amplification reactions can be run in

parallel. This concept design is amenable for fabrication in any STL-3D printing platform

and may be scaled up to accommodate a larger number of tubes.

We used a blueGel electrophoresis unit, powered by 120 AC volts, to validate the LAMP

amplification using gel electrophoresis. Photo-documentation was done using a

smartphone camera. We also used a Synergy HT microplate reader (BioTek Instruments,

VT, USA) to detect the fluorescence induced by an intercalating reagent in positive

samples from the PCR reactions.

Validation DNA templates: We used plasmids containing the complete N gene from 2019-

nCoV, SARS, and MERS as positive controls, with a concentration of 200,000 copies/μL

(Integrated DNA Technologies, IA, USA). Samples containing different concentrations of

synthetic nucleic acids of SARS-CoV-2 were prepared by successive dilutions from stocks

(from 2×10^5 copies to 65 copies). We used a plasmid that contained the gene GP from

EBOV as a negative control. The production of this EBOV genetic material has been

documented elsewhere by our group [23].

RNA extracts from human volunteers. In addition, we used 8 samples of RNA extracts from

6 COVID-19 positive and 2 negative subjects, as determined by RT-PCR analysis. Samples

were kindly donated by Hospital Alfa, Medical Center, in Guadalupe, Nuevo León,

México. Nasopharyngeal samples were collected from two patients after obtaining

informed and signed written consent and in complete observance of good clinical practices,

the principles stated in the Helsinki Declaration, and applicable lab operating procedures at

Hospital Alfa. Every precaution was taken to protect the privacy of sample donors and the

confidentiality of their personal information. The experimental protocol was approved on

May 20th, 2020 by a named institutional committee (Alfa Medical Center, Research

Comitte; resolution AMCCI-TECCOVID-001). RNA extraction and purification was

conducted at the molecular biology laboratory at Hospital Alfa. The Qiagen QIAamp DSP

Viral RNA Mini kit was used for RNA extraction and purification by closely following the

directions of the manufacturer.

Amplification mix: We used WarmStart® Colorimetric LAMP 2× Master Mix (DNA &

RNA) from New England Biolabs (MA, USA), and followed the recommended protocol:

12.5 μL Readymix, 1.6 μM FIP primer, 1.6 μM BIP primer, 0.2 μM F3 primer, 0.2 μM B3

primer, 0.4 μM LF primer, 0.4 μM LB primer, 1μL DNA template (~ 625 to 2×10^5 DNA

copies), 1.25 μL EvaGreen® Dye from Biotium (CA, USA), and nuclease-free water to a

final volume of reaction 25 μL. This commercial mix contains phenol red as a pH indicator

for revealing the shift of pH during LAMP amplification across the threshold of pH=6.8.

Primers used: Two different sets of LAMP primers, referred to here as α and β, were

designed in house using the LAMP primer design software Primer Explorer V5

(http://primerexplorer.jp/lampv5e/index.html). These primers were based on the analysis of

alignments of the SARS-Co2 N gene sequences using the software Geneious (Auckland,

New Zealand), downloaded from https://www.ncbi.nlm.nih.gov/genbank/sars-cov-2-

seqs/#nucleotide-sequences.

Each set, containing six LAMP primers, were used to target two different regions of the

sequence of the SARS-Co2 N gene. In addition, for comparison purposes, we conducted

PCR amplification experiments using one of the primer sets recommended by the CDC for

the standard diagnostics of COVID-19 (i.e., N1 assay) using RT-qPCR. The sequences of

our LAMP primers are presented in Table 1. The sequences of the PCR primers (N1) have

been reported elsewhere[24,53].

Amplification protocols: For all LAMP experiments, we performed isothermal heating for

30 or 60 min. In our experiments, we tested three different temperatures: 50, 60, and 65 ºC.

For PCR experiments, we used a three-stage protocol consisting of a denaturation stage at

94℃ for 5 min, followed by 25 cycles of 94 °C for 20 s, 60 °C for 30 s, and 72 °C for 20 s,

and then a final stage at 72 °C for 5 min, for a total duration of 60 min in the miniPCR®

thermocycler from Amplyus (MA, USA).

Documentation of LAMP products: We analyzed 10 μL of each LAMP reaction in a

blueGel unit, a portable electrophoresis unit sold by MiniPCR from Amplyus (MA, USA).

This is a compact electrophoresis unit ($23 \times 10 \times 7$ cm^3) that weighs 350 g. In these

experiments, we analyzed 10 μL of the LAMP product using 1.2% agarose electrophoresis

tris-borate-EDTA buffer (TBE). We used the Quick-Load Purple 2-Log DNA Ladder

(NEB, MA, USA) as a molecular weight marker. Gels were dyed with Gel-Green from

Biotium (CA, USA) using a 1:10,000 dilution, and a current of 48 V was supplied by the

blueGel built-in power supply (AC 100–240V, 50–60Hz).

As an alternative method for detection and reading of the amplification product, we

evaluated the amplification products by detecting the fluorescence emitted by a DNA-

intercalating agent, the EvaGreen® Dye from Biotium (CA, USA), in a Synergy HT

microplate reader (BioTek Instruments, VT, USA). Briefly, 25 μL of the LAMP reaction

was placed in separate wells of a 96-well plate following completion of the LAMP

incubation. A 125 μL volume of nuclease-free water was added to each well for a final

sample volume of 150 μL and the samples were well-mixed by pipetting. These

experiments were run in triplicate. The following conditions were used in the microplate

reader: excitation of 485/20, emission of 528/20, gain of 75. Fluorescence readings were

done from the top at room temperature.

Color determination by image analysis: We also photographically documented and

analyzed the progression of color changes in the positive and negative SARS-CoV-2

synthetic samples during the LAMP reaction time (i.e., from 0 to 50 min). For that purpose,

Eppendorf PCR tubes containing LAMP samples were photographed using a smartphone

(iPhone, from Apple, USA). We used an application for IOS (Color Companion, freely

available at Apple store) to determine the components of color of each LAMP sample in the

CIELab color space. Color differences between the positive samples and negative controls

were calculated as distances in the CIELab coordinate system according to the following

formula:

$\text{Color Distance}_{\text{sample-negative}} = \text{SQRT} [(L_{\text{sample}} - L_{\text{negative}})^2 + (a_{\text{sample}} - a_{\text{negative}})^2 + (b_{\text{sample}} - b_{\text{negative}})^2]$

Here L, a, and b are the color components of the sample or the negative control in the

CIELab color space (Figure S4).

573 Acknowledgments

The authors aknowledge the funding provided by the Federico Baur Endowed Chair in

Nanotechnology (0020240I03). EGG acknowledges funding from a doctoral scholarship

provided by CONACyT (Consejo Nacional de Ciencia y Tecnología, México). GTdS and

MMA acknowledge the institutional funding received from Tecnológico de Monterrey

(Grant 002EICIS01). MMA, GTdS, SOMC and IMLM acknowledge funding provided by

CONACyT (Consejo Nacional de Ciencia y Tecnología, México) through grants SNI

26048, SNI 256730, SNI 31803, SNI 1056909, respectively. YSZ acknowdges the suppport

by the Brigham Research Institute.

Author participation

E.G. and I.L. conducted most of the the amplification experiments; G.T. and M.A. wrote

the manuscript; E.G. I.L. and M.A. prepared the figures; C.G., G.E.G., G.G, and J.A.

designed the incubation chamber; A.G. and F.Y. 3D-printed the incubation chamber; I.R.

and E.M. conducted the experiments of RNA extraction and purification form human

samples at Alfa Medical and INER; Y-S.Z., S.M., G.T., and J.Z. edited the manuscript; J.Z.

coordinated the collection of samples and the execution of RT-qPCR experiments at INER,

and G.T., S.M., Y-S.Z.; and M.A. designed the study.

Supporting Information

Figure S1. (A) Photograph and (B) rendering of the 3D-printed incubation chamber used in

LAMP experiments.

Figure S2. Schematic representation of the chamber (different views) showing its relevant

dimensions.

Figure S3. Commercial plasmid that contains the plasmids containing the complete N gene

from 2019-nCoV, SARS, and MERS. We use this plasmid as a SARS-CoV-2 synthetic

nucleic acid material in our amplification experiments.

Figure S4. *In house* designed plasmid containing the gene that codes for the expression of

protein GP from EBOV. This plasmid was added as nucleic acid material in negative

controls in our amplification experiments.

Figure S5. (A) The colorimetric LAMP method described here was able to identify and
amplify synthetic SARS-CoV-2 genetic material in samples containing as few as ~62 viral
copies. (B) Evaluation of the stability and functionality of the LAMP reaction mix at
different storage times and temperatures. The reaction mix, which is formulated with
LAMP primers and ready for the addition of nucleic acid extracts, is functional and
discriminates between positive and negative samples when stored (i) at room temperature
for 48 h or (ii) at 4 °C for 72 h.

Figure S6. (A) Color analysis conducted on positive and negative SARS-CoV-2 samples
contained in Eppendorf PCR tubes (yellow inset) using Color Companion, an app from
Apple (downloadable at Apple Store, USA). This app identifies the components of color in
a specific location of an image (black circle in the yellow inset) in the CIELab, RGB, HSB,
or CMYK spaces. The image can be uploaded using e-mail, airdrop, or Whatsapp. (B)
Schematic representation of the CIELab space, a color system where any color can be
represented in terms of a point and its coordinates in a 3D space, where L is luminosity, a is
the axis between green and red, and b is the axis between yellow and red.

Figure S7. (A) The amount of amplification product in LAMP experiments was evaluated
by measuring the fluorescence emitted by the amplification product in reactions with an
added intercalating agent. Fluorescence readings were conducted in standard 96-well plates
using a conventional plate reader. (A) Fluorescence readings, as measured in a commercial
plate reader, for different dilutions of SARS-CoV-2 synthetic DNA templates. Results
using two different LAMP primer sets are shown: set α (indicated in blue), and set β
(indicated in red).

Figure S8. Images of the full-length gels from which Figure 3B and 3C were obtained. (A,
B) Agarose gel electrophoresis of DNA amplification products generated by targeting two
different regions of the sequence coding for SARS-Co2 N protein. Two different primer
sets were used: (B) primer set α, and (C) β. The initial template amount was gradually
decreased from left to right: 2.0×10^5 DNA copies (lane 1), 4.0×10^4 copies, (lane 2), 1.0
$\times 10^4$ copies (lane 3), 2.5×10^3 copies (lane 4), 625 copies (lane 5), negative control (lane

6), and molecular weight ladder (lane 7).

Figure S9. Image of the full-length gel from which Figure 6B was obtained. LAMP

amplification products from RNA extracts (lane 1 and 2) and cDNA (lane 3 and 4) from a

COVID (+) patient, and a COVID (-) volunteer (lane 5), as revealed by gel electrophoresis

experiments. A molecular weight ladder is shown in lane 6.

Figure S10. Image of the full-length gels from which Figure 7D and 7E were obtained. (A-

B) LAMP amplification products from RNA extracts containing (D) 300 ng/µL and (E) 30

ng/ µL from the same set of COVID (+) patients (S1-S4, S6 and S8), and COVID (-)

volunteers (S5 and S7), as revealed by gel electrophoresis experiments. Lanes 1 to 8

contained amplification products from samples S1 to S8. Lane 9 was reserved for the

molecular weight ladder.

References

1.　Coronavirus Disease (COVID-19) – Statistics and Research - Our World in Data

[Internet]. [cited 8 Apr 2020]. Available:

https://ourworldindata.org/coronavirus?fbclid=IwAR28tcRVA1rmXsVCrYHcxuHp

RXeyO9-uxFJFSG5-lv5gsJgzDxK7eN08i_Y

2.　Bedford J, Enria D, Giesecke J, Heymann DL, Ihekweazu C, Kobinger G, et al.

COVID-19: towards controlling of a pandemic. The Lancet. Lancet Publishing

Group; 2020. pp. 1015–1018. doi:10.1016/S0140-6736(20)30673-5

3.　Cohen J, Kupferschmidt K. Countries test tactics in "war" against COVID-19.

Science. American Association for the Advancement of Science; 2020;367: 1287–

1288. doi:10.1126/science.367.6484.1287

4.　Pung R, Chiew CJ, Young BE, Chin S, I-C Chen M, Clapham HE, et al. Articles

Investigation of three clusters of COVID-19 in Singapore: implications for

surveillance and response measures. Lancet. Elsevier; 2020;19: 1–8.

doi:10.1016/S0140-6736(20)30528-6

5.   Alvarez MM, Gonzalez-Gonzalez E, Santiago GT. Modeling COVID-19 epidemics

in an Excel spreadsheet: Democratizing the access to first-hand accurate predictions

of epidemic outbreaks. medRxiv. Cold Spring Harbor Laboratory Press; 2020;

2020.03.23.20041590. doi:10.1101/2020.03.23.20041590

6.   Singh R, Adhikari R. Age-structured impact of social distancing on the COVID-19

epidemic in India. 2020; Available: http://arxiv.org/abs/2003.12055

7.   Bastos SB, Cajueiro DO. Modeling and forecasting the Covid-19 pandemic in

Brazil. 2020; Available: http://arxiv.org/abs/2003.14288

8.   Yen C-W, de Puig H, Tam JO, Gómez-Márquez J, Bosch I, Hamad-Schifferli K, et

al. Multicolored silver nanoparticles for multiplexed disease diagnostics:

distinguishing dengue, yellow fever, and Ebola viruses. Lab Chip. The Royal Society

of Chemistry; 2015;15: 1638–1641. doi:10.1039/C5LC00055F

9.   Mou L, Jiang X. Materials for Microfluidic Immunoassays: A Review. Adv Healthc

Mater. Wiley-VCH Verlag; 2017;6: 1601403. doi:10.1002/adhm.201601403

10. Alvarez MM, López-Pacheco F, Aguilar-Yañez JM, Portillo-Lara R, Mendoza-

Ochoa GI, García-Echauri S, et al. Specific Recognition of Influenza A/H1N1/2009

Antibodies in Human Serum: A Simple Virus-Free ELISA Method. Jeyaseelan S,

editor. PLoS One. Public Library of Science; 2010;5: e10176.

doi:10.1371/journal.pone.0010176

11. Zhong L, Chuan J, Gong B, Shuai P, Zhou Y, Zhang Y, et al. Detection of serum

IgM and IgG for COVID-19 diagnosis. Sci CHINA Life Sci. Science China Press;

2020; doi:10.1007/S11427-020-1688-9

12.   Pardee K, Green AA, Takahashi MK, Braff D, Lambert G, Lee JW, et al. Rapid,

690   Low-Cost Detection of Zika Virus Using Programmable Biomolecular Components.

691   Cell. Cell Press; 2016;165: 1255–1266. doi:10.1016/J.CELL.2016.04.059

692   13.   Broughton JP, Deng X, Yu G, Fasching CL, Streithorst J, Granados A, et al. Rapid

693   Detection of 2019 Novel Coronavirus SARS-CoV-2 Using a CRISPR-based 1

694   DETECTR Lateral Flow Assay 2 3. doi:10.1101/2020.03.06.20032334

695   14.   Chen JS, Ma E, Harrington LB, Da Costa M, Tian X, Palefsky JM, et al. CRISPR-

696   Cas12a target binding unleashes indiscriminate single-stranded DNase activity.

697   Science (80-). American Association for the Advancement of Science; 2018;360:

698   436–439. doi:10.1126/science.aar6245

699   15.   Elizondo-Montemayor L, Alvarez MM, Hernández-Torre M, Ugalde-Casas PA,

700   Lam-Franco L, Bustamante-Careaga H, et al. Seroprevalence of antibodies to

701   influenza A/H1N1/2009 among transmission risk groups after the second wave in

702   Mexico, by a virus-free ELISA method. Int J Infect Dis. Elsevier; 2011;15: e781–

703   e786. doi:10.1016/j.ijid.2011.07.002

704   16.   Tang Y-W, Schmitz JE, Persing DH, Stratton CW. The Laboratory Diagnosis of

705   COVID-19 Infection: Current Issues and Challenges. J Clin Microbiol. American

706   Society for Microbiology Journals; 2020; doi:10.1128/JCM.00512-20

707   17.   Wölfel R, Corman VM, Guggemos W, Seilmaier M, Zange S, Müller MA, et al.

708   Virological assessment of hospitalized patients with COVID-2019. Nature. Nature

709   Publishing Group; 2020; 1–10. doi:10.1038/s41586-020-2196-x

710   18.   Mauk MG, Song J, Bau HH, Liu C. Point-of-Care Molecular Test for Zika Infection.

711   Clin Lab Int. NIH Public Access; 2017;41: 25–27. Available:

712   http://www.ncbi.nlm.nih.gov/pubmed/28819345

19.   Liao S-C, Peng J, Mauk MG, Awasthi S, Song J, Friedman H, et al. Smart cup: A

minimally-instrumented, smartphone-based point-of-care molecular diagnostic

device. Sensors Actuators B Chem. Elsevier; 2016;229: 232–238.

doi:10.1016/J.SNB.2016.01.073

20.   Jansen van Vuren P, Grobbelaar A, Storm N, Conteh O, Konneh K, Kamara A, et al.

Comparative Evaluation of the Diagnostic Performance of the Prototype Cepheid

GeneXpert Ebola Assay. J Clin Microbiol. American Society for Microbiology;

2016;54: 359–67. doi:10.1128/JCM.02724-15

21.   Craw P, Balachandran W. Isothermal nucleic acid amplification technologies for

point-of-care diagnostics: a critical review. Lab Chip. The Royal Society of

Chemistry; 2012;12: 2469. doi:10.1039/c2lc40100b

22.   Hsieh K, Patterson AS, Ferguson BS, Plaxco KW, Soh HT. Rapid, sensitive, and

quantitative detection of pathogenic DNA at the point of care through microfluidic

electrochemical quantitative loop-mediated isothermal amplification. Angew Chem

Int Ed Engl. NIH Public Access; 2012;51: 4896–900. doi:10.1002/anie.201109115

23.   Gonzá Lez-Gonzá Lez E, Mendoza-Ramos JL, Pedrozaid SC, Cuellar-Monterrubio

A, Má Rquez-Ipiña AR, Lira-Serhanid D, et al. Validation of use of the miniPCR

thermocycler for Ebola and Zika virus detection. 2019;

doi:10.1371/journal.pone.0215642

24.   Gonzalez-Gonzalez E, Santiago GT, Lara-Mayorga IM, Martinez-Chapa SO,

Alvarez MM. Portable and accurate diagnostics for COVID-19: Combined use of the

miniPCR thermocycler and a well-plate reader for SARS-Co2 virus detection.

medRxiv. Cold Spring Harbor Laboratory Press; 2020; 2020.04.03.20052860.

doi:10.1101/2020.04.03.20052860

25.   Lan L, Xu D, Ye G, Xia C, Wang S, Li Y, et al. Positive RT-PCR Test Results in

738   Patients Recovered From COVID-19. JAMA. 2020; doi:10.1001/jama.2020.2783

739   26.   El-Tholoth M, Bau HH, Song J. A Single and Two-Stage, Closed-Tube, Molecular

740   Test for the 2019 Novel Coronavirus (COVID-19) at Home, Clinic, and Points of

741   Entry. ChemRxiv; 2020; doi:10.26434/CHEMRXIV.11860137.V1

742   27.   Kaarj K, Akarapipad P, Yoon JY. Simpler, Faster, and Sensitive Zika Virus Assay

743   Using Smartphone Detection of Loop-mediated Isothermal Amplification on Paper

744   Microfluidic Chips. Sci Rep. Nature Publishing Group; 2018;8: 1–11.

745   doi:10.1038/s41598-018-30797-9

746   28.   Tomita N, Mori Y, Kanda H, Notomi T. Loop-mediated isothermal amplification

747   (LAMP) of gene sequences and simple visual detection of products. Nat Protoc.

748   Nature Publishing Group; 2008;3: 877–882. doi:10.1038/nprot.2008.57

749   29.   Goto M, Honda E, Ogura A, Nomoto A, Hanaki KI. Colorimetric detection of loop-

750   mediated isothermal amplification reaction by using hydroxy naphthol blue.

751   Biotechniques. 2009;46: 167–172. doi:10.2144/000113072

752   30.   Yang T, Wang Y-C, Shen C-F, Cheng C-M. Point-of-Care RNA-Based Diagnostic

753   Device for COVID-19. Diagnostics. MDPI AG; 2020;10: 165.

754   doi:10.3390/diagnostics10030165

755   31.   Kozel TR, Burnham-Marusich AR. Point-of-Care Testing for Infectious Diseases:

756   Past, Present, and Future. J Clin Microbiol. American Society for Microbiology;

757   2017;55: 2313–2320. doi:10.1128/JCM.00476-17

758   32.   Su W, Gao X, Jiang L, Qin J. Microfluidic platform towards point-of-care

759   diagnostics in infectious diseases. J Chromatogr A. Elsevier; 2015;1377: 13–26.

760   doi:10.1016/J.CHROMA.2014.12.041

33.   Drancourt M, Michel-Lepage A, Boyer S, Raoult D. The Point-of-Care Laboratory

in Clinical Microbiology. Clin Microbiol Rev. American Society for Microbiology;

2016;29: 429–47. doi:10.1128/CMR.00090-15

34.   Jangam SR, Agarwal AK, Sur K, Kelso DM. A point-of-care PCR test for HIV-1

detection in resource-limitedsettings. Biosens Bioelectron. Elsevier; 2013;42: 69–75.

doi:10.1016/J.BIOS.2012.10.024

35.   Qiu X, Ge S, Gao P, Li K, Yang S, Zhang S, et al. A smartphone-based point-of-care

diagnosis of H1N1 with microfluidic convection PCR. Microsyst Technol. Springer

Berlin Heidelberg; 2017;23: 2951–2956. doi:10.1007/s00542-016-2979-z

36.   Nguyen T, Duong Bang D, Wolff A. 2019 Novel Coronavirus Disease (COVID-19):

Paving the Road for Rapid Detection and Point-of-Care Diagnostics.

Micromachines. MDPI AG; 2020;11: 306. doi:10.3390/mi11030306

37.   Udugama B, Kadhiresan P, Kozlowski HN, Malekjahani A, Osborne M, Li VYC, et

al. Diagnosing COVID-19: The Disease and Tools for Detection. ACS Nano.

American Chemical Society; 2020; doi:10.1021/acsnano.0c02624

38.   Gates B. Responding to Covid-19 — A Once-in-a-Century Pandemic? N Engl J

Med. Massachusetts Medical Society; 2020; doi:10.1056/nejmp2003762

39.   Yu L, Wu S, Hao X, Li X, Liu X, Ye S, et al. Rapid colorimetric detection of

COVID-19 coronavirus using a reverse tran-scriptional loop-mediated isothermal

amplification (RT-LAMP) diagnostic plat-form: iLACO. medRxiv. Cold Spring

Harbor Laboratory Press; 2020; 2020.02.20.20025874.

doi:10.1101/2020.02.20.20025874

40.   Zhang Y, Odiwuor N, Xiong J, Sun L, Nyaruaba RO, Wei H, et al. Rapid Molecular

Detection of SARS-CoV-2 (COVID-19) Virus RNA Using Colorimetric LAMP.

medRxiv. Cold Spring Harbor Laboratory Press; 2020;2: 2020.02.26.20028373.

doi:10.1101/2020.02.26.20028373

41. Park G-S, Ku K, Beak S-H, Kim SJ, Kim S Il, Kim B-T, et al. Development of

Reverse Transcription Loop-mediated Isothermal Amplification (RT-LAMP) Assays

Targeting SARS-CoV-2. bioRxiv. Cold Spring Harbor Laboratory; 2020;

2020.03.09.983064. doi:10.1101/2020.03.09.983064

42. Zhu X, Wang X, Han L, Chen T, Wang L, Li H, et al. Reverse transcription loop-

mediated isothermal amplification combined with nanoparticles-based biosensor for

diagnosis of COVID-19. medRxiv. Cold Spring Harbor Laboratory Press; 2020;

2020.03.17.20037796. doi:10.1101/2020.03.17.20037796

43. Lamb LE, Bartolone SN, Ward E, Chancellor MB. Rapid Detection of Novel

Coronavirus (COVID19) by Reverse Transcription-Loop-Mediated Isothermal

Amplification. SSRN Electron J. Elsevier BV; 2020; doi:10.2139/ssrn.3539654

44. Jiang M, Fang W, Aratehfar A, Li X, ling L, Fang H, et al. Development and

validation of a rapid single-step reverse transcriptase loop-mediated isothermal

amplification (RT-LAMP) system potentially to be used for reliable and high-

throughput screening of COVID-19. medRxiv. Cold Spring Harbor Laboratory

Press; 2020; 2020.03.15.20036376. doi:10.1101/2020.03.15.20036376

45. Lalli MA, Chen X, Langmade SJ, Fronick CC, Sawyer CS, Burcea LC, et al. Rapid

and extraction-free detection of SARS-CoV-2 from saliva with colorimetric LAMP.

medRxiv. Cold Spring Harbor Laboratory Press; 2020; 2020.05.07.20093542.

doi:10.1101/2020.05.07.20093542

46. Santiago GT De, Gante CR De, García-Lara S, Ballescá-Estrada A, Alvarez MM.

Studying mixing in Non-Newtonian blue maize flour suspensions using color

analysis. PLoS One. Public Library of Science; 2014;9.

doi:10.1371/journal.pone.0112954

47. Tanner NA, Zhang Y, Evans TC. Visual detection of isothermal nucleic acid

amplification using pH-sensitive dyes. Biotechniques. Eaton Publishing Company;

2015;58: 59–68. doi:10.2144/000114253

48. Rusk N. Torrents of sequence. Nature Methods. Nature Publishing Group; 2011. p.

44. doi:10.1038/nmeth.f.330

49. Wang W, Xu Y, Gao R, Lu R, Han K, Wu G, et al. Detection of SARS-CoV-2 in

Different Types of Clinical Specimens. JAMA - Journal of the American Medical

Association. American Medical Association; 2020. doi:10.1001/jama.2020.3786

50. Gadkar VJ, Goldfarb DM, Gantt S, Tilley PAG. Real-time Detection and Monitoring

of Loop Mediated Amplification (LAMP) Reaction Using Self-quenching and De-

quenching Fluorogenic Probes. Sci Rep. Nature Publishing Group; 2018;8: 1–10.

doi:10.1038/s41598-018-23930-1

51. González-González E, Mendoza-Ramos JL, Pedroza SC, Cuellar-Monterrubio AA,

Márquez-Ipiña AR, Lira-Serhan D, et al. Validation of use of the miniPCR

thermocycler for Ebola and Zika virus detection. Ansumana R, editor. PLoS One.

Public Library of Science; 2019;14: e0215642. doi:10.1371/journal.pone.0215642

52. Smyrlaki I, Ekman M, Lentini A, Vondracek M, Papanicolaou N, Aarum J, et al.

Massive and rapid COVID-19 testing is feasible by extraction-free SARS-CoV-2

RT-qPCR. doi:10.1101/2020.04.17.20067348

53. OSF Preprints | Landscape Coronavirus Disease 2019 test (COVID-19 test) in vitro -

- A comparison of PCR vs Immunoassay vs Crispr-Based test [Internet]. [cited 8 Apr

2020]. Available: https://osf.io/6eagn

833