Open Development and Clinical Validation
Of Multiple 3D-Printed Sample-Collection Swabs:
Rapid Resolution of a Critical COVID-19 Testing Bottleneck

Cody J Callahan BS[a], Rose Lee MD[bc], Katelyn E. Zulauf PhD[bd], Lauren Tamburello MPH[e], Kenneth P. Smith PhD[bd], Joe Previtera AS[f], Annie Cheng BA[b], Alex Green BA[bd], Ahmed Abdul Azim MD[ci], Amanda Yano MHA[g], Nancy Doraiswami RN BSN[h], James E. Kirby MD PhD[bd], and Ramy A. Arnaout MD DPhil[bdj†]

[a]Department of Radiology, Beth Israel Deaconess Medical Center, Boston, MA, USA 02215

[b]Clinical Microbiology Laboratories, Division of Clinical Pathology, Department of Pathology, Beth Israel Deaconess Medical Center, Boston, MA USA 02215

[c]Division of Infectious Disease, Department of Medicine, Beth Israel Deaconess Medical Center, Boston, MA, USA 02215

[d]Harvard Medical School, Boston, MA, USA 02115

[e]Division of Urologic Surgery, Department of Surgery, Beth Israel Deaconess Medical Center, Boston, MA, USA 02215

[f]Division of Respiratory Therapy, Beth Israel Deaconess Medical Center, Boston, MA, USA 02215

[g]Department of Medicine, Beth Israel Deaconess Medical Center, Boston, MA, USA 02215

[h]Division of Perioperative Services, Department of Central Processing, Beth Israel Deaconess Medical Center, Boston, MA, USA 02215

[i]Division of Infection Control/Hospital Epidemiology, Silverman Institute for Healthcare Quality and Safety, Beth Israel Deaconess Medical Center, Boston, MA, USA 02215

[j]Division of Clinical Informatics, Department of Medicine, Beth Israel Deaconess Medical Center, Boston, MA USA 02215

[†]Corresponding author: +1-617-538-5681, rarnaout@bidmc.harvard.edu.

Abstract

The SARS-CoV-2 pandemic has caused a severe international shortage of the nasopharyngeal

swabs that are required for collection of optimal specimens, creating a critical bottleneck in the

way of high-sensitivity virological testing for COVID-19. To address this crisis, we designed and

executed an innovative, radically cooperative, rapid-response translational-research program

that brought together healthcare workers, manufacturers, and scientists to emergently develop

and clinically validate new swabs for immediate mass production by 3D printing. We performed

a rigorous multi-step preclinical evaluation on 160 swab designs and 48 materials from 24

companies, laboratories, and individuals, and shared results and other feedback via a public

data repository (http://github.com/rarnaout/Covidswab/). We validated four prototypes through

an institutional review board (IRB)-approved clinical trial that involved 276 outpatient volunteers

who presented to our hospital's drive-through testing center with symptoms suspicious for

COVID-19. Each participant was swabbed with a reference swab (the control) and a prototype,

and SARS-CoV-2 reverse-transcriptase polymerase chain reaction (RT-PCR) results were

compared. All prototypes displayed excellent concordance with the control (κ=0.85-0.89). Cycle-

threshold (Ct) values were not significantly different between each prototype and the control,

supporting the new swabs' non-inferiority (Mann-Whitney U [MWU] $p>0.05$). Study staff

preferred one of the prototypes over the others and the control swab overall. The total time

elapsed between identification of the problem and validation of the first prototype was 22 days.

Contact information for ordering can be found at http://printedswabs.org. Our experience holds

lessons for the rapid development, validation, and deployment of new technology for this

pandemic and beyond.

Introduction

Since the emergence of the COVID-19 pandemic, more than 2.5 million cases have been diagnosed worldwide (*1*). These diagnoses were made using material collected from NP swabs, which provide the highest sensitivity for detecting SARS-CoV-2 infection during early infection using commercial RT-PCR-based assays. An NP swab is an FDA Class I exempt medical device roughly 15 cm in length and 2-3mm in diameter designed to collect secretions from the posterior nasopharynx (Figs. 1a, left and 1b, top). The head of the swab is generally coated with short synthetic filaments called flock. The swab is inserted into the nasopharynx, rotated several times to collect material, and then placed into a vial containing a few milliliters of transport media. A breakpoint on the shaft enables detachment and release of the head into the vial, which is then sealed and sent for testing.

The rapid spread of SARS-CoV-2 has resulted in severe shortages of NP swabs, due to both manufacturing stoppages resulting in decreased supply and the spread of the pandemic resulting in unprecedented demand (*2*). To address the swab shortage, hospitals and other testing centers have repurposed other commercially available swabs (e.g. throat, urogenital) to collect nasal epithelial mucosa for testing (Fig. 1a, second from left and 1b, second from top). However, such swabs are suboptimal for swabbing the nasopharynx due to differences in size and flexibility and the possibility they contain PCR-inhibitory materials (*3, 4*). Material from other sites has not yet been shown to be able to substitute for swabbing the nasopharynx.

One solution to the swab crisis is to design and 3D-print swabs. Advantages of 3D printing include simplicity (avoiding the multistep process of applying flock), the widespread availability of 3D printing capacity, and the ability to iterate prototypes rapidly (*5*). To resolve the swab-shortage crisis, we have been coordinating an open collaborative process that has brought together many medical centers, individuals, academic laboratories, and both new and well established manufacturers (*6*). As part of this process, we have been testing and continuously providing feedback on prototype swabs in order to proceed rapidly but safely toward the development of swabs that can be used clinically, at volumes equal to the need. The openness of the process was a conscious decision supported by a substantial body of scientific literature, including the previous experience of the present authors, that demonstrates the advantages of openness over closed or hybrid approaches (*7–9*). At our institution, this process has led to an ongoing clinical trial of several prototype swabs, the first results of which we report here.

Materials and Methods

Process. We created a public repository using GitHub, a free website most often used by
programmers to co-develop computer code (http://www.github.com/rarnaout/Covidswab)[6]. We
provided a clear description of the problem and updated the repository with whatever we
learned and encouraged others to do the same. By tapping our personal and professional
networks, we nucleated an ad hoc network of manufacturers that included companies,
academic groups, and individuals. This grew to include other medical centers interested in
helping develop and test new swabs. These other groups were given the ability to add to the
repository as desired.

We devised a three-phase process consisting of preclinical evaluation (Phase I), production
considerations (Phase II), and field testing (Phase III). We described these processes on the
repository for all to see. We took high-resolution photographs of all prototypes and stored Phase
I results in a Microsoft Excel (Microsoft Corporation, Redmond, WA, USA) spreadsheet that
remains publicly available in the repository. All contributors could see each others' designs and
our feedback and iterated accordingly.

We made our personal contact information freely available to facilitate communication and
speed the delivery of prototypes. We involved representatives of our institution's nursing, legal,
intellectual property, leadership, purchasing, human resources, communications, and
contracting teams and the institutional review board early and often in order to facilitate open
development, reassign idled staff to our process, and minimize lead times during the rapidly
changing situation.

Phase I: Preclinical evaluation. *Design.* An infectious disease physician, clinical pathologist
(clinical microbiologist), and respiratory therapist tested each prototype swab for design and
mechanical properties (Fig. 1c-d). These included size measurements of the head, neck, shaft,
and breakpoint (requirement of ~15cm to reach the posterior nasopharynx; head diameter of 1-
3.2mm to pass into the mid-inferior portion of the inferior turbinate and be able maneuver
appropriately without catching on anatomical variants such as septal spurs or a deviated nasal
septum); surface properties such as smoothness (with roughness leading to an unpleasant feel
and risk of bleeding); flexibility vs. brittleness of the head, neck, shaft, and breakpoint (to avoid
fracture during use); durability (e.g. ability to tolerate 20 rough repeated insertions into a 4-mm-
inner-diameter clear plastic tube curved back on itself with a curve radius of ~3 centimeters;

ability to bend tip and neck 90 degrees without breaking; ability to restore to initial form following

bend of 45 degrees; Fig. 1d); strength (resist breakage under rough but reasonable

manipulation); and other factors as applicable (e.g. stickiness, smell).

Collection sufficiency. We assessed the ability to collect sufficient material for testing using

Gram stain of a swab of the interior cheek smeared onto a standard microscopy slide as a

surrogate for NP swabbing and comparison to Gram stain of a swab of the interior cheek using

Copan Diagnostics, Inc. (Mantua, Italy) model 501CS01 NP swab as the control (Fig. 1c). Slides

were heat fixed and Gram stained according to the BD BBL gram stain test kit protocol (*10*).

Slides were examined at 40x magnification for the presence of both epithelial cells and bacteria.

Prototypes were passed if they collected a comparable quantity of the material as the control.

PCR compatibility. We tested PCR compatibility by incubating the head overnight in 3 mL of

modified CDC VTM (Hank's balanced salt solution containing: 2% heat inactivated FBS,

100µg/mL gentamicin, 0.5µg/mL fungizone, and 10mg/L Phenol red (*11*)) to allow any PCR-

inhibitory material to leach into the medium, spiking 1.5mL with 200 copies/mL of control SARS-

CoV-2 amplicon target (representing 2 times the limit of detection on our system), vortexing, and

testing using the Abbott RealTime SARS-CoV-2 Assay on an Abbott m2000 RealTime System

platform (*12*), following the same protocol as for clinical testing. PCR-positive prototypes

passed.

Phase II: Production considerations. We considered stability to autoclaving by repeating

Phase I testing on post-autoclaved materials; manufacturers' short-term strategies for individual

packaging; and manufacturers' stated ability to produce at least 10,000 swabs per day (at the

time roughly a week's worth of swabs for a mid-sized testing center) within a week's notice. We

considered differences in supply chain to minimize the risk of future crises.

Phase III: Field testing. *Trial design and oversight.* COVIDSwab is an adaptive trial for

evaluating the performance of prototypes compared to the control (see above). Participants

under clinical suspicion for COVID-19 who were scheduled for standard clinical SARS-CoV-2

RT-PCR testing with a control swab were asked also to be swabbed afterward with a single

prototype. Prototypes were collected and tested until at least 10 positive and 10 negative results

on control swabs were obtained (*13*). Sample collection was performed by trained nursing or

respiratory-therapy staff ("study staff") overseen by the respiratory therapy department at

BIDMC. The Clinical Microbiology laboratories oversaw data collection. This study was reviewed

and approved by the institutional review board of Beth Israel Deaconess Medical Center

(protocol number 2020P000323).

Participants. Participants were individuals clinically suspected of COVID-19 who were brought

to the drive-through/walk-up ("drive-through") COVID-19 testing site at BIDMC. Adults over 18

years of age were given a participant information sheet by study staff and asked whether they

would agree to being swabbed with a prototype swab performed by a trained nurse or

respiratory therapist in addition to the control swab required for testing. Individuals with known

thrombocytopenia of <50,000 platelets/μl were excluded from the study to avoid risk of mild

bleeding.

Trial procedures. Prototype swabs were individually packaged and autoclaved at BIDMC for

sterilization according to manufacturer protocols. Swabbing was performed per standard

protocol. Participants were first swabbed with the control swab, then the prototype. Choice of

naris for each swab was left to study staff and the participant. Approximately half of all drive-

through arrivals participated. Control and prototype swabs were placed in separate vials of VTM

and transported to the BIDMC Clinical Microbiology Laboratories where each sample was tested

on the Abbott m2000 SARS-CoV-2 RT-PCR platform as per standard clinical protocol.

Statistical analyses. RT-PCR results are reported categorically as either positive or negative.

We tested categorical concordance using Cohen's kappa (*14*). For each positive test, the Ct

value (the RT-PCR cycle number at which the sample first turns positive) was obtained from the

Clinical Microbiology Laboratories. Higher values reflect lower viral load in the sample.

We tested for systematic bias in Ct values by comparing values for controls vs. prototypes using

MWU (*15*). This tested the null hypothesis that values for controls and prototypes are drawn

from the same underlying distribution; $p>0.05$ was interpreted as no bias. For discordant

(positive control/negative prototype or vice versa) samples, the negative was assigned a Ct

value of 37, the total number of cycles run. As a second test for bias, we compared (again by

MWU) the distribution of differences in Ct values between control and prototype swabs to the

distribution of differences between two control swabs taken within 24 hours (quality-control data

independent of our study). This tested the null hypothesis that the differences between control

and prototype swabs and the differences between two control swabs are drawn from the same

underlying distribution; $p>0.05$ was interpreted as no bias.

To quantify relative preferences among the prototypes, we gave study staff members printouts
of all six possible pairs of swabs (a "round robin"), in randomized order, and for each pair asked
them to circle their preference (A-B testing). We collated the results and assessed preferences.

Results

Open process. In the first days of the development effort GitHub repository (*6*) was established
to serve a public resource and knowledge base. We updated the repository continuously with
design information and test results. These updates included high-resolution images of
prototypes submitted to us for testing (*6*), a public database of results of our Phase I testing,
and periodic updates and guidance based on our experiences. Open communication facilitated
rapid design iteration by providing anyone interested with a way to quickly understand the
required specifications and to learn from each other's experiences.

Phase I testing. To date we have evaluated 48 materials and 160 designs submitted to us for
testing by 4 individuals, 2 laboratories, and 18 companies, for a total of 24 manufacturers.
Seven (4.4%) have passed Phase I testing. Most failures were either for inappropriate materials,
including some that were sticky or brittle, or for inappropriate designs, including those with sharp
heads. Prototypes from 19 manufacturers went through at least two iterations, with a maximum
of 28 prototypes from one manufacturer (Prototype 4 below; Fig. 1). The rate-limiting steps were
receipt of new prototypes, with slow mail delivery during the pandemic being a major
contributor, and PCR-compatibility testing, as testing patient samples took priority over testing
prototypes. Communication with and responsiveness by manufacturers were considered
outstanding.

Phase II and III prototypes. Four prototypes passed Phase II testing, all of which have
completed our Phase III clinical trial: these are prototypes from the 3D-printing manufacturers
Resolution Medical (with technology from Carbon3D), EnvisionTec, Origin.io, and HP Inc.
(Prototypes 1-4, respectively; Fig. 1a). Like control swabs, the prototypes were 15-16cm in
length with 1-3cm length radially symmetric heads 2-3mm in diameter, a thin neck 4-7cm long
and 1-2mm in diameter, and a thicker shaft 2-4mm in diameter, with a breakpoint most often 7-
8cm from the tip of the head. The materials were plastics and resins such as Keysplint Soft.
Head design evolved over many iterations to increase surface area. Designs generally featured
either a polygonal matrix connected to a central, tapered strut with multiple branch points or else
some form of spiral (Fig. 1b). Manufacturers were able to balance sample collection (Fig. 1c),

stiffness, and surface texture. Variations of a longitudinal central strut allowed for varying

degrees of stability, flexibility, and impact cushioning (Fig. 1d).

Sample and data acquisition. We collected and tested control and prototype swab pairs from

276 participants. Approximately half of the patients tested at our drive-through testing center

participated. Because testing runs were batched and the COVID-19 status of participants was

not known prior to testing, the number of control-positives usually exceeded the minimum

requirement of 10 (range, 10-19). Total collection time was 2-3 days per prototype. The

frequency of control-positive tests was 18%, generally increasing by prototype as the pandemic

worsened in and around Boston.

Comparison. All four prototypes exhibited a high degree of concordance with the control swab,

with kappas of 0.88, 0.85, 0.89, and 0.88, respectively (Fig. 2a). For convenience we use the

terminology of true positives, true negatives, false positives, and false negatives, with the

control swab result considered the provisional gold standard. Prototypes exhibited 0-1 false

positives and 1-2 false negatives. However, since control swabs are known to be an imperfect

gold standard (<100% sensitivity) and because PCR positives are more likely to reflect true

infection than error, false positives were interpreted as identifying missed infections; indeed,

false positives were referred to clinical care teams as clinically actionable, as per IRB protocol.

Of note, discordant cases were always associated with high Ct values, reflecting low viral load

(Fig. 2b). For example, for Prototype 4, the control swab for one of the two false negatives had a

Ct of 31.47, just shy of 31.50, our hospital's cutoff for reportability (corresponding approximately

to a single virion per mL of VTM); in addition, testing of this false negative was delayed by 16

hours because of prioritizing patient samples, which can result in decreased signal.

To better assess possible performance differences between control and prototype swabs, we

compared Ct values for control-prototype pairs for which at least one was positive (assigning the

maximum-possible Ct to negatives; see Methods). Specifically, we asked whether the Ct values

for the prototype swabs were systematically different from those of the control swabs.

Systematically higher values for prototype swabs would suggest that they may underperform

control swabs, notwithstanding the high kappa values. A p-value of >0.05 indicates no statistical

difference. Although there were more datapoints below the 1:1 line than above it (Fig. 2b),

statistical testing revealed no evidence for underperformance, with MWU p-values of 0.36, 0.26,

0.42, and 0.31 for Prototypes 1-4, respectively (Fig. 2b). This result supports the conclusion that

the prototypes are non-inferior to the control.

As an additional assessment of non-inferiority, we compared the difference in Ct values
observed between control and prototype swabs to the differences between replicates of control
swabs. Independent of our clinical trial, there were 88 cases in which a patient, in the course of
clinical care, was swabbed twice within 24 hours (mean±stdev, 15±7 hours), during the time
period of our study. In 11 of these cases, at least one of the two swabs was positive for SARS-
CoV-2. There were two disagreements between replicate swab tests, resulting in a kappa of
0.90, similar to what was observed in our study for each prototype (kappa=0.85-0.89). Also as in
our study, the Ct values for the first swab and second swab were not significantly different
(MWU p-value of 0.18). Finally, the differences between Ct values for the first and second
control swabs were comparable to the differences between control and prototype swabs (MWU
p-values of 0.31, 0.26, 0.47, and 0.44 for Prototypes 1-4; Fig. 2b).

Staff and participant preferences. A written staff survey showed a preference for Prototype 4,
then Prototypes 2 and 3, then Prototype 1. There was a slight preference for the control swab
over Prototype 4 (Fig. 3a). In narrative feedback, Prototype 4, which underwent the largest
number of revisions through our process (28), was described as comparable to the control swab
(Fig. 3b).

Availability. Swabs are available to order. Details can be found on the GitHub repository in the
updates at https://github.com/rarnaout/Covidswab/tree/master/BIDMC.

Discussion

The COVID-19 pandemic has forced healthcare providers to seek alternative sources of critical
materials affected by supply-chain disruptions and increases in demand. The situation has
forced providers to innovate under extraordinary time-pressure. Over the course of our study we
received numerous anecdotal reports of swab shortages at hospitals across the United States
and in Europe, necessitating urgent stopgap solutions. Scientific literature on time-sensitive
innovation suggests that open, collaborative, decentralized processes outperform closed or
proprietary ones (*7–9*). Here we report the success of such a process, going from the
identification of the swab crisis to multiple clinically validated prototypes capable of high-volume
manufacture beginning at 22 days. Notably, none of the prototypes tested were flocked, yet their
performance was statistically indistinguishable from the flocked control swab.

The urgency of the situation, the configuration of the manufacturing ecosystem, and human
nature contributed to several observations and shortcomings worth mentioning. First, 3D

printing has important advantages in a crisis, including the ability to iterate designs and output

swabs rapidly. It remains to be seen how complementary manufacturing techniques, each with

advantages and disadvantages relative to 3D printing, will contribute in a more mature market

and less urgent setting. Second, in any cooperative process there is a temptation to "defect," i.e.

taking without giving back. Individuals and manufacturers may well exploit open knowledge for

competitive advantage (*16*). This is a known price of openness that can disincentivize

cooperation, absent social or structural mechanisms to enforce norms; managing this

temptation took considerable effort by all. Third, ideally the study size would have been larger,

and there would have been a better null model than replicates separated by many hours, to

which to compare our results. Possible sources of variance in our study include differences in

secretions or viral burden between nares and the possibility that the first (control) swab left less

material for the second (prototype) when the same naris was used for both swabs. Despite

these potential issues, our statistical tests supported analytical non-inferiority for all four

prototypes. And fourth, we note our "round-robin" A-B testing survey was useful in summarize

preferences, although the narrative comments seemed often to be more positive than the round

robin suggested. A possible explanation is that the control swab was preferred in large part

simply due to its being familiar, and preferred only narrowly (if often).

Like the control swab, the prototype swabs we tested can be improved upon, and manufacturers

are currently doing so. The same is true for other prototypes we may test through our ongoing

clinical trial. Especially in a crisis, perfect is the enemy of good enough. The pandemic

continues to change quickly, and bottlenecks will likely continue to appear unpredictably. The

constant requirement is the ability to respond in a timely fashion under this extraordinary

pressure. We hope our experience, based on past scientific work on cooperation and

innovation, will provide a useful case study for how to iterate and produce a clinically validated

medical manufacture under the pressure of an ongoing pandemic, work on which others will

hopefully improve as we continue to fight COVID-19 together

Acknowledgements

The authors would like to thank Abigail Bakken, Alisa Chau, Monique Cole, Zachary Fitts,

Jonathan Goldie, Lan Huynh, Christina Lexinger-Holahan, Lorinda Longhi, Restituto Miranda,

Pavel Gorelik, Jenny Hu, Ofer Mazor, Goutam Reddy, Karen Robinson, Stefan Riedel, Christina

Yen, Summer Decker, Don Ingber, Pawan Jolly, Kit Parker, Adama Sesay, Craig Broady, John

Burpo, Daniel Davis, Joe DeSimone, Annette Friskopp, Ric Fulop, Grant Michael Gonzalez,

Markus Greiner, Marie Herring, Matthew Hurley, Hardik Kabaria, Shawn Patterson, David

Lakatos, Ben Linville-Engler, Oren Mechanic, Richard Novak, Jifei Ou, Michael Papish, Steve

Pollack, Chris Prucha, Christian Reed, Isabel Sanz, Al Siblani, Lihua Zhao, Stephanie

Dominguez, Stuart Steinbock, Greg Mark, Nic Scarfo, John Bentley, Nira Pollock, Paula

Watnick, Colleen Baker, Carol Daugherty. Clementina DiMonda, Virginia Dolan, Isaac Fombuh,

Kylie Griffin, Amy Guadognoli, Allison Wang, Dan Eiref, Johan Sunryd, Jonathan Ford, Sherry

Yu, and Kelsey Ladt for participation in the open process that made this study possible.

K.E.Z. was supported by a National Institute of Allergy and Infectious Diseases training grant

(T32AI007061).

K.P.S. was supported by the National Institute of Allergy and Infectious Diseases of the National

Institutes of Health under award number F32AI124590.

This work was conducted with support from Harvard Catalyst | The Harvard Clinical and

Translational Science Center (National Center for Advancing Translational Sciences, National

Institutes of Health Award UL 1TR002541) and financial contributions from Harvard University

and its affiliated academic healthcare centers. The content is solely the responsibility of the

authors and does not necessarily represent the official views of Harvard Catalyst, Harvard

University and its affiliated academic healthcare centers, or the National Institutes of Health.

This work was supported by Carbon3D, Envisiontec, Origin, and HP to cover the cost of the IRB

and PCR testing.

The content is solely the responsibility of the authors and does not necessarily represent the

official views of the National Institutes of Health.

REFERENCES

1. COVID-19 Map - Johns Hopkins Coronavirus Resource Center. https://coronavirus.jhu.edu/map.html. Accessed April 21, 2020.

2. Thomas K. Coronavirus Test Obstacles: A Shortage of Face Masks and Swabs. *The New York Times*. https://www.nytimes.com/2020/03/18/health/coronavirus-test-shortages-face-masks-swabs.html. Published March 18, 2020. Accessed April 13, 2020.

3. Kodzius R, Xiao K, Wu J, et al. Inhibitory effect of common microfluidic materials on PCR outcome. *Sensors and Actuators B: Chemical*. 2012;161(1):349-358. doi:10.1016/j.snb.2011.10.044

4. Bessetti J. An Introduction to PCR Inhibitors. 2007. https://www.promega.es/-/media/files/resources/profiles-in-dna/1001/an-introduction-to-pcr-inhibitors.pdf?la=es-es. Accessed April 13, 2020.

5. Tack P, Victor J, Gemmel P, Annemans L. 3D-printing techniques in a medical setting: a systematic literature review. *BioMed Eng OnLine*. 2016;15(1):115. doi:10.1186/s12938-016-0236-4

6. rarnaout. *Rarnaout/Covidswab*.; 2020. https://github.com/rarnaout/Covidswab. Accessed April 13, 2020.

7. Lakhani KR, Boudreau KJ, Loh P-R, et al. Prize-based contests can provide solutions to computational biology problems. *Nat Biotechnol*. 2013;31(2):108-111. doi:10.1038/nbt.2495

8. Lee J, Min J, Lee H. The Effect of Organizational Structure on Open Innovation: A Quadratic Equation. *Procedia Computer Science*. 2016;91:492-501. doi:10.1016/j.procs.2016.07.128

9. Mak RH, Endres MG, Paik JH, et al. Use of Crowd Innovation to Develop an Artificial Intelligence–Based Solution for Radiation Therapy Targeting. *JAMA Oncol*. 2019;5(5):654. doi:10.1001/jamaoncol.2019.0159

10. BD Gram Stain Kits and Reagents. September 2017. https://legacy.bd.com/resource.aspx?IDX=19184. Accessed April 13, 2020.

11. Preparation of Viral Transport Medium SOP# DSR-052-01. March 2020.

https://www.cdc.gov/coronavirus/2019-ncov/downloads/Viral-Transport-Medium.pdf.

12. Abbott RealTime SARS-CoV-2. March 2020. https://www.molecular.abbott/sal/9N77-

095_SARS-CoV-2_US_EUA_Amp_PI.pdf. Accessed April 13, 2020.

13. Miller JM, Campbell S, Loeffelholz M. Changing Swabs: To Validate or Not To Validate? *J*

Clin Microbiol. 2013;51(11):3910-3910. doi:10.1128/JCM.02023-13

14. Kwiecien R, Kopp-Schneider A, Blettner M. Concordance Analysis. *Dtsch Arztebl Int.*

2011;108(30):515-521. doi:10.3238/arztebl.2011.0515

15. Mann HB, Whitney DR. On a Test of Whether one of Two Random Variables is

Stochastically Larger than the Other. *Ann Math Statist.* 1947;18(1):50-60.

doi:10.1214/aoms/1177730491

16.   Fehr E, Fischbacher U. The nature of human altruism. *Nature.* 2003;425(6960):785-791.
doi:10.1038/nature02043

 FIGURES AND FIGURE LEGENDS

Figure 1: Control and prototype swabs. **(a**) From left to right: the control swab (Copan 501CS01), a repurposed urogenital cleaning swab approved for NP testing through our process, Prototype 1 (Resolution Medical), Prototype 2 (EnvisionTec), Prototype 3 (Origin.io), and Prototype 4 (Hewlett Packard). **(b)** From top to bottom, closeups of the heads of the swabs in (a). Scale bars, 1cm. **(c)** Examples of Gram stain of cheek swab using control (top) and prototype swabs. Scale bar, 10µm. **(d)** Examples of materials testing. Clockwise from top left: head flexibility and robustness to fracture, neck flexibility and robustness to fracture, robustness to repeat insertion into and removal from a tortuous canal (diameter 3cm), and breakpoint evaluation.

Figure 2: Concordance results. (a) 2x2 tables giving counts for each prototype vs. the control swab (first three panels) and for control vs. replicate control obtained within 24 hours on the same individual. Discordant results in gray; totals for each swab below and to the right of each box; total number of pairs in bold; K=Cohen's kappa. **(b)** Scatterplot of Ct values for pairs of swabs for which at least one swab was SARS-CoV-2 positive. For discordant pairs, the negative swab was assigned a Ct value of 37 (the maximum number of cycles run).

Figure 3: Subjective feedback. (a) Round-robin A-B testing of net preferences among

Prototypes 1-3 (large bold numbers) and the control ("C"). Each arrow points from the less

preferred to the more preferred swab. Arrow weight indicates strength of relative preference.

Preferences were unanimous except where noted with numbers separated by a slash: the first

number denotes the number of responses for the direction indicated by the arrowhead, while the

second number denotes the number of responses that had the opposite preference. The weight

of the arrow is proportional to the difference (e.g. 7-3=a net preference of 4). Unless noted,

each arrow denotes 12-15 separate responses. **(b)** Number of positive and negative comments

received from study staff who administered the swabs, tabulated by category. In each plot,

negative feedback is to the left of the zero, while positive feedback is to the right. Bars on both

the positive and negative sides of zero reflect differing opinions among study staff. *n*, total

number of comments received about each prototype.