Aerosolized Hydrogen Peroxide Decontamination of N95 Respirators, with Fit-Testing and Viral Inactivation, Demonstrates Feasibility for Re-Use During the COVID-19 Pandemic

T. Hans Derr[1]*^, Melissa A. James[2]^, Chad V. Kuny[4,8]^, Devanshi Patel[6,8], Prem P. Kandel[5], Cassandra Field[6,8], Matthew D. Beckman[7], Kevin L. Hockett[5,8], Mark A. Bates[3], Troy C. Sutton[6,8], Moriah Szpara[4,8]*

^co-first-authors

[1] Environmental Health and Safety; [2] Animal Resource Program, [3] Occupational Medicine Program, and [4] Departments of Biology, Biochemistry and Molecular Biology, [5] Plant Pathology and Environmental Microbiology, [6] Veterinary and Biomedical Sciences, [7] Statistics, [8] Center for Infectious Disease Dynamics, and the Huck Institutes of the Life Sciences, Pennsylvania State University, University Park, Pennsylvania 16802, USA

*Co-Corresponding Authors

Moriah L. Szpara, PhD
Depts. of Biology, and Biochemistry &
Molecular Biology
The Huck Institutes of the Life Sciences
Pennsylvania State University
W-208 Millennium Science Complex
(MSC)
University Park, PA 16802 USA
Phone: 814-867-0008
Email: moriah@psu.edu

T. Hans Derr, CIH
Manager, Health & Environmental
Programs
Environmental Health & Safety
Pennsylvania State University
6B Eisenhower Deck
University Park, PA 16802
Phone: 814-863-3834
Email: thd12@psu.edu

Author emails:

Mark A. Bates: mxb20@psu.edu
Matthew D. Beckman: mdb268@psu.edu
T. Hans Derr: thd12@psu.edu
Cassandra Field: cif5202@psu.edu
Kevin L. Hockett: klh450@psu.edu
Melissa A. James: maj22@psu.edu

Prem Kandel: kandelpremp@gmail.com
Chad V. Kuny: cvk108@psu.edu
Devanshi Patel: drp5323@psu.edu
Troy C. Sutton: tcs38@psu.edu
Moriah L. Szpara: moriah@psu.edu

Abstract

In response to the demand for N95 respirators by healthcare workers during the COVID-19 pandemic, we evaluated decontamination of N95 respirators using an aerosolized hydrogen peroxide (aHP) system. This system is designed to dispense a consistent atomized spray of aerosolized, 7% hydrogen peroxide (H_2O_2) solution over a treatment cycle. Multiple N95 respirator models were subjected to ten or more cycles of respirator decontamination, with a select number periodically assessed for qualitative and quantitative fit testing. In parallel, we assessed the ability of aHP treatment to inactivate multiple viruses absorbed onto respirators, including phi6 bacteriophage, HSV-1, CVB3, and SARS-CoV-2. For pathogens transmitted via respiratory droplets and aerosols, it is critical to address respirator safety for reuse. This study provided experimental validation of an aHP treatment process that decontaminates the respirators while maintaining N95 function. External NIOSH certification verified respirator structural integrity and filtration efficiency after ten rounds of aHP treatment. Virus inactivation by aHP was comparable to the decontamination of commercial spore-based biological indicators. These data demonstrate that the aHP process is effective, with successful fit-testing of respirators after multiple aHP cycles, effective decontamination of multiple virus species including SARS-CoV-2, successful decontamination of bacterial spores, and filtration efficiency maintained at or greater than 95%. While this study did not include extended or clinical use of N95 respirators between aHP cycles, these data provide proof of concept for aHP decontamination of N95 respirators before reuse in a crisis-capacity scenario.

Importance

The COVID-19 pandemic led to unprecedented pressure on healthcare and research facilities to provide personal protective equipment. The respiratory nature of the SARS-CoV2 pathogen makes respirator facepieces a critical protection measure to limit inhalation of this virus. While respirator facepieces were designed for single-use and disposal, the pandemic increased overall demand for N95 respirators, and corresponding manufacturing and supply chain limitations necessitated the safe reuse of respirators when necessary. In this study, we repurposed an aerosolized hydrogen peroxide (aHP) system that is regularly utilized to decontaminate materials

30   in a biosafety level 3 (BSL3) facility, to develop methods for decontamination of N95

31   respirators. Results from virus inactivation, biological indicators, respirator fit testing, and

32   filtration efficiency testing all indicated that the process was effective at rendering N95

33   respirators safe for reuse. This proof-of-concept study establishes baseline data for future testing

34   of aHP in crisis capacity respirator-reuse scenarios.

Keywords:

37   N95 respirators, filtering facepiece (FFP) respirators (FFR), decontamination, aerosolized

38   hydrogen peroxide, COVID-19, SARS-CoV2, virologic testing, virus, fit-testing, disinfection,

39   sterilization, CURIS®

Abbreviations:

aHP, aerosolized hydrogen peroxide (H_2O_2); BSL, biosafety level; COVID-19, coronavirus

disease 2019; CVB3, coxsackievirus B3; H_2O_2, hydrogen peroxide; HPM, hydrogen peroxide

vapor monitor; HSV-1, herpes simplex virus 1; phi6, *Pseudomonas* phi6 bacteriophage; OSHA,

Occupational Safety and Health Administration; QLFT, qualitative fit test; QNFT, quantitative

fit-test; SARS-CoV-2, severe acute respiratory syndrome coronavirus 2; VHP, Vapor-phase

hydrogen peroxide.

Introduction

The ongoing severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) pandemic resulted

in a shortage of personal protective equipment (PPE). In healthcare settings, the need for PPE is

critical to protect frontline healthcare providers from infection, and to reduce cross-

contamination between patients with coronavirus disease 2019 (COVID-19) and other uninfected

patients. In healthcare settings, N95 filtering facepiece (FFP) respirators, including surgical N95

respirators, are used to provide protection from airborne infectious particles. The N95

terminology refers to the ability to block at least 95 percent of the most penetrating particle sizes

(0.1 - 0.3 micron). Proper use of N95 respirators requires qualitative fit-testing (QLFT) or

quantitative fit-testing (QNFT), which are designed to ensure a tight face-to-respirator seal for

each specific wearer's facial characteristics.

The shortage of N95 respirators resulted from limitations on the required raw materials, limited

capacity to manufacture respirators, and the ability of supply and distribution chains to handle

increased global demand. For this reason, researchers sought to demonstrate the potential for

decontamination and reuse of existing N95 respirators. Standardized procedures are well-

established for the decontamination and reuse of medical equipment, such as autoclaving, steam

treatment, and chemical inactivation (e.g. bleach) (1). Decontamination of most medical

equipment is verified using spore-based biological indicators (2). Unlike most hospital

equipment (e.g. steel, metal, plastic) or fabrics (e.g. blankets) for which standardized

decontamination methods exist (1), N95 respirators are generally not intended for re-use (3).

Thus many of the standard decontamination approaches deform, damage, or destroy the integrity

of N95 respirator fabric, nosepiece materials, or elastic straps (4–8). Hydrogen peroxide (H_2O_2)-

based methods have been successfully adapted for use in decontamination of N95 respirators (4,

6–22), with indications that these methods are less damaging than other decontamination

methods (e.g. chemical or steam) and can penetrate the densely-woven fabric of respirator

facepieces (4, 6–8, 14, 22). In addition, the viricidal capability of H_2O_2 decontamination methods

has been previously demonstrated (23–25). Vapor-phase H_2O_2 methods (e.g. "VHP" and other

patented methods) have been used to decontaminate N95 respirators and were granted temporary

U.S. Food and Drug Administration (FDA) "Emergency Use Authorization" (EUA) for

healthcare use during the pandemic (26). However, these methods employ high concentrations of

hydrogen peroxide (30-70%) and elevated temperature to achieve vaporization. Vapor-phase

H_2O_2 (VHP) methods at these concentrations may pose increased health risks to decontamination

personnel, and combined with elevated temperature, can result in notable respirator material

decay (27). Historical aerosolized H_2O_2 (aHP) methods utilize lower peroxide concentrations (5-

6%) with silver ions and other antimicrobial agents, activated aerosolization via plasma, nozzle

pressure, or ultrasound, with similar limitations to VHP methods (28–32). These methods have

not received strong comparative support in U.S. markets due to lower decontamination

effectiveness (30–38). Therefore, we utilized a recently-developed aHP method (i.e. the CURIS®

system), which dispenses a low concentration hydrogen peroxide solution through a precision

adjustable nozzle (39). The unit design and decontamination process characteristics allow

consistent distribution of disinfectant over time and enable effective decontamination of space

and materials. The aHP approach has the potential to scale up for large clinical settings, since the

number of respirators that can be decontaminated simultaneously is limited only by the room

size.

At present, respirator manufacturers have not approved protocols for N95 respirator

decontamination (3). To address imminent pandemic needs, healthcare settings have referenced

U.S. Centers for Disease Control and Prevention (CDC) guidelines on provisional N95 respirator

decontamination and reuse (40, 41). Battelle Memorial Institute of Columbus, Ohio, received

FDA EUA status for an N95 decontamination protocol using a VHP method, based on a prior

study addressing the potential for respirator reuse in emergency scenarios (4, 26). This study

included evaluation of respirator structure, filtration, and manikin fit-testing, and used bacterial

spore-based biological indicators to demonstrate effective decontamination (4); however, viral

inactivation testing and respirator fit on live subjects were not performed. Since the current

pandemic entails a respiratory viral pathogen, multiple decontamination protocols have been

actively investigated at research universities and medical centers (6–21).

Given the reduced personnel health risks of using aHP, we assessed the ability of an aHP

decontamination protocol to achieve viral and microbial decontamination of N95 respirators,

while preserving respirator fit and integrity over multiple treatment cycles. Several respirator

models in use by local healthcare and research personnel were included. Viral decontamination

was tested using multiple species representing a range of pathogen characteristics: *Pseudomonas*

phi6 bacteriophage (phi6), herpes simplex virus 1 (HSV-1), coxsackievirus B3 (CVB3), and

SARS-CoV-2. Commercial *Geobacillus stearothermophilus* spore-based biological indicators

were used in parallel throughout the aHP process to verify effectiveness of decontamination. We

measured the inactivation of viruses by passive drying and by active aHP decontamination.

Fitness of respirators for re-use was assessed by qualitative and quantitative respirator fit-testing

after decontamination, including up to 10 cycles of aHP treatment. Respirator structure and

filtration efficiency testing was also performed. We also acquired real-time and diffusion

sampler analyses of hydrogen peroxide levels throughout the decontamination process to monitor

user safety.

118   Results

This study was intended to rigorously evaluate a protocol for decontamination and re-use of N95

respirators using aerosolized hydrogen peroxide (aHP) treatment (7% H_2O_2; Curoxide®). The

N95 respirator facepiece models examined here represent those frequently used at Penn State or

within the Penn State Health system. Six N95 respirator models were selected, with the greatest

number available being the 3M 8511 model (Table 1, Figure S1). The decontamination process

was performed in the BSL3 enhanced facility, enabling the assessment of viral inactivation

across multiple biosafety levels, including SARS-CoV-2 (Table 2, Figure S2). The BSL3 facility

employs aHP on a routine basis to decontaminate solid equipment, and we adapted this protocol

to account for the absorbent nature of N95 respirators. Standard aHP charge, pulse, and dwell

period parameters were adjusted to optimize cycle times (Table 3) and account for biological

indicator and virology results. Our final aHP process (matched to room size) utilized a 11:43

charge period to establish of aHP, followed by six pulse charges evenly spaced over 30 minutes

to maintain H_2O_2 concentrations, and a 20-minute dwell period (Table 3).

132   Chemical Verification and Biological Validation of Aerosolized H_2O_2 Process

Chemical and bacterial spore-based biological indicators (BIs) were used to verify aHP treatment

and decontamination. All chemical indicators located throughout the treatment area (Prep Room)

during all aHP cycles confirmed exposure to hydrogen peroxide. For cycles in which charge,

pulse, and dwell periods were utilized, all biological indicators (BIs) similarly indicated

successful decontamination, with the exception of aHP cycles 3 and 5 (Table 4, Figure S3). In

cycle 3 the additional dwell period was not yet implemented, and BIs indicated an unsuccessful

decontamination cycle (1 positive, 5 negative). External contamination of one spore coupon after

treatment cycle 5 (via dropping) likely resulted in the single positive indicator for this cycle

(Table 4). Overall chemical and biological indicator results indicated successful aHP treatment

and decontamination of N95 respirators.

143   Real-Time Hydrogen Peroxide Monitoring

Real-time measurement of H_2O_2 concentrations during decontamination were obtained with a

portable, real-time ATI PortaSens II monitor (Table 5). H_2O_2 concentrations in the treatment area

were measured at $\geq$ 120 ppm during the charge and pulse periods (maximum sensor capability).

Hydrogen peroxide concentrations outside the sealed entryway were low or undetected (0 ppm).

On occasions where a broken seal arose in the tape around the door (Figure S2), low

concentrations were detectable outside the door (up to 3 ppm). Once corrected, these levels

immediately dropped to 0 ppm. During aeration after aHP treatment, H_2O_2 levels were monitored

until concentrations measured $\leq$ 2 ppm. At this time research personnel entered the Prep Room to

measure H_2O_2 concentrations at respirator surfaces. Hydrogen peroxide concentrations were

observed to decline rapidly during aeration within 20-30 minutes of door opening (Table 5).

Initial respirator surface concentrations (at fabric) typically exceeded 2 ppm. Subsequent

respirator drying to achieve < 1 ppm required further room aeration (exhaust ventilation)

exceeding two hours, or overnight (Table 5). Once H_2O_2 respirator surface concentrations

measured < 1 ppm, they were either subjected to further cycles of aHP, or collected for fit-testing

or virus inactivation analysis.

Qualitative and Quantitative Fit-Testing After aHP Decontamination

We employed both qualitative (QLFT) and quantitative (QNFT) fit-testing. QLFT allowed

repeated non-destructive testing across multiple rounds of decontamination, while QNFT

provided numerical data on fit factor and ruled out any false negatives due to user fatigue. QLFT

and QNFT was conducted after the first, fifth, and tenth decontamination cycles for both a male

and female subject. These tests included a large number of 3M model 8511 respirators (for which

we had the most available in the starting pool), and a smaller number of 5 other respirator models

(Table 1, Figure 1). All respirators passed all QLFT and QNFT fit tests (Figure 1). One respirator

facepiece (3M model 1870+) experienced a broken strap after the eighth cycle of aHP; this

respirator had passed QLFT successfully in earlier cycles but was excluded from final QNFT. No

other failures occurred, therefore with 95% confidence, at least 95% of 3M model 8511

respirators maintained successful fit seal after ten aHP decontamination cycles. QNFT results

with 3M model 8511 respirators passed the minimum passing fit factor of 100, and the maximum

quantifiable fit factor of 200(+) in 8 of 9 tests. The numerical range of passing fit factors (100 –

200), and available sample quantity, limit any further statistical analysis. The successful fit-

testing results indicate that repeated cycles of aHP decontamination do not interfere with

respirator fit for reuse.

Respirator Filtration Efficiency Testing

N95 respirator filtration efficiency testing was conducted on aHP-treated 3M 8511 respirators

using the NIOSH test protocol under full-load test conditions (Table 6). No visible degradation

was found on inspection (e.g. metal nose guard discoloration, unusual thinning, or wear).

Filtration efficiencies were found to exceed 95% for all eight respirators subjected to ten cycles

of aHP decontamination. Therefore, for at least ten aHP cycles, there were no adverse impacts or

loss of N95 respirator efficiencies. Two additional unused, untreated 3M 8511 respirators were

additionally tested and found with slightly reduced filtration efficiencies (94.4%, 94.6%),

consistent with the respirator manufacturer's 5-year shelf-life limitation. These data indicate that

overall respirator performance was maintained over time and after aHP treatment.

Application of Multiple Viral Species to N95 Respirator Facepieces

Viral inactivation on respirator facepieces was anticipated to occur both by the passive process of

drying or desiccation, and by the active process of aHP decontamination. Multiple virus species

were included to test the decontamination potential of aHP against viruses in general, as well as

against SARS-CoV-2. These included phi6, HSV-1, CVB3, and SARS-CoV-2 (see Table 2 and

Methods for details). Respirators used for virus inactivation testing were those previously

subjected to aHP treatment and fit-testing, to spare overall respirator consumption (see Table 4

and Methods). Initial application of virus to different respirator facepiece types revealed clear

differences in relative absorption vs. fluid repulsion (Figure S1). 3M respirator models 1870+

and 9211+ share a common outer fabric which is listed by the manufacturer as having the highest

fluid resistance of any N95 respirator (Table 1) (42). In our testing, these two respirator models

displayed no apparent absorption of virus inoculum and instead dried with a "coffee ring effect".

All other respirator types (Table 1) experienced a combination of liquid spreading, absorption,

and evaporative drying of the virus inoculum droplet. Viruses were inoculated onto different

areas of each respirator facepiece model, including the outer and inner fabric surfaces, the elastic

strap, and where present, the inner and outer surface of the plastic exhalation valves.

Decontamination of Virus-Inoculated Respirators by Aerosolized H_2O_2 Treatment

We set out to assess the effectiveness of aHP treatment for active decontamination of virus

inoculated onto N95 respirators. Virus-inoculated respirators were subjected to aHP treatment,

using both "Modification 1" and "Final" parameters (Table 3). Viral testing of decontamination

was conducted during five independent aHP cycles (Table 4), using the maximum inoculum titer

available for each viral stock preparation. For phi6 bacteriophage, this included 34 "aHP-treated"

sites spanning two independent rounds of testing (Figure 2). For HSV-1 and CVB3, this included

62 and 60 "aHP-treated" sites respectively, spanning three independent rounds of testing (Figure

3 and Figure 4). Across a total of 204 respirator sites inoculated with one of four virus species

tested (Table 2; see also Table S1), only four sites had any detectable virus remaining. Three of

these rare positive virus plaques were detected in an aHP cycle using the "Modification 1"

parameters (Figure 2A and 3A; Table 4; see also Figure S4) – these were a key motivation to add

the dwell time for the "Final" aHP cycle parameters (Table 3). Overall, aHP treatment produced

a 4-7 $\log_{10}$ reduction in viral load (10^7 reduction for phi6, Figure 2A; 10^5 reduction for HSV-1,

Figure 3; 10^4 reduction for CVB3; Figure 4). There was no observable difference in the

effectiveness of aHP decontamination for inner vs. outer surfaces of respirators (Figures 2-5; see

also Table S1), or in limited testing of alternative inoculation sites such as elastic straps (Figure

3, Figure 4, Table S1). The success of virus inactivation by aHP treatment mirrored the results of

spore-based biological indicators (Table 4).

Inactivation of Multiple Viral Species by Air-Drying on Respirators

Since passive viral inactivation by drying likely occurs during the time involved in

decontamination, we measured the amount of virus remaining on inoculated but untreated N95

respirators. For each cycle of viral testing, the duration of passive viral inactivation was matched

to the duration of the decontamination process (including aHP treatment, subsequent aeration,

and transport). These "drying-only" samples confirmed a partial loss of viral infectiousness,

ranging from 10-100-fold for phi6 and HSV-1, and 100-fold or greater for CVB3 and SARS-

CoV-2 (Figures 2-5; see also Figure S4 and Table S1). For phi6 bacteriophage, this included 26

"drying-only" inoculation sites spanning three independent rounds of testing (Figure 2). For

HSV-1 and CVB3, this included 52 "drying-only" inoculation sites for each virus species,

spanning four independent rounds of testing (Figure 3 and Figure 4, respectively). Viral

inactivation by drying was not markedly different on inner vs. outer surfaces of the respirator

models (Figures 2-5; see also Table S1), or in limited testing of alternative inoculation sites, such

as elastic straps or plastic exhalation valves (Figure 3A-B, Figure 4A-B, Table S1).

Inactivation of SARS-CoV-2 by aHP Treatment

Experimental testing of SARS-CoV-2 by aHP necessitated that all work be completed at BSL3. These studies were conducted with the 3M 8511, 1860, and 1870+ N95 respirators and included two independent rounds of testing (Figure 5). As with the surrogate virus species, SARS-CoV-2 was first inoculated onto respirator facepieces. The "drying only" controls were left in the BSL3 ambient environment, while matched samples were subjected to aHP-treatment. The respirators used for SARS-CoV-2 testing had been subjected to 10 prior rounds of aHP before virus inoculation (Table 4). For SARS-CoV-2, viral testing included 48 "drying-only" inoculation sites and 48 "aHP-treated" sites, spanning two independent rounds of testing (Figure 5). We observed a partial loss of viral infectiousness for SARS-CoV-2 due to drying (e.g. $10^{6.125}$ TCID50/mL input vs ~10^2 TCID50/mL after drying; Figure 5). Importantly, no infectious SARS-CoV-2 remained on any respirator model after aHP decontamination (Figure 5).

Discussion

Based on a series of ten respirator decontamination cycles and multiple rounds of viral inactivation testing, multiple N95 respirator models tested were found to be suitable for aHP decontamination and reuse. We found that respirators successfully passed qualitative respirator fit testing after multiple cycles of the aHP decontamination process, and ultimately passed tests indicating no loss in filtration efficiency. Most studies thus far have focused either on verifying fit-testing after respirator decontamination (6–10, 12, 15, 16, 20), or examined how decontamination approaches inactivate one or more virus species on respirators (6–8, 11, 13, 14, 18–21). A handful of N95 respirator decontamination studies have combined fit-testing with measures of viral inactivation (6–8), but none incorporated viral- or fit-testing after 5 or 10 decontamination cycles as done here, or included the parallel use of biological indicators. Only rarely have studies included extended-use between multiple rounds of decontamination cycles (10), although this is a key aspect of reuse that warrants further study. Our study is unique in including multiple measures of respirator integrity via fit-testing, as well as filtration efficiency, and robust verification of viral inactivation using BIs, multiple surrogate viruses, and SARS-CoV-2 in parallel (22). The breadth of this study aims to extend its usefulness beyond the current pandemic.

Respirator Resilience for Re-Use

Aerosolized H_2O_2 decontamination of the N95 respirators used in this study did not indicate any adverse impact on final respirator filtration efficiency, although extended-use of respirators between aHP cycles was not possible in this study. We documented effective control of H_2O_2 levels, supporting effective pathogen decontamination with no detectable researcher exposure. These data support and reflect those of parallel studies that have tested respirator fit after decontamination by other forms of hydrogen peroxide vapor (VHP) and/or similar methods (6–10, 12, 15, 16, 20). Further studies will be needed to assess the H_2O_2 concentration profile inside containment, using real-time instruments with a wider detection range and automated data logging.

As noted above, most studies that include successful fit-testing and verification of viral inactivation have not pursued this testing across multiple (i.e. 10) cycles of decontamination (6–8, 20). Lab-based conditions such as those used here do not fully reflect clinical use conditions. While respirators were physically stretched between each decontamination cycle as a proxy for donning and doffing, sustained clinical use includes other stressors which may influence fit and performance (e.g. exhaled moisture or perspiration) (10, 16). However as noted by the CDC, respirators with obvious signs of use (e.g. makeup or patient fluids) should be discarded and not used for decontamination (40, 41). Other studies have explored respirator decontamination in these real-world use scenarios, albeit without including the parallel testing of viral species, BIs, and cycle numbers that were included here (10, 15–17, 20). We anticipate that future studies will address the implementation of aHP-based decontamination in a clinical-use setting.

During this study, one 3M 1870+ respirator suffered a broken rubber strap after eight cycles of aHP, during inter-cycle strap stretching. The breakage occurred at a point on the strap that corresponded with a penned hash mark (used to denote decontamination cycle; see Figure S2). There were no other instances of failure for this or any other respirator models. Prior groups have likewise noted strap-based failures after multiple cycles of respirator decontamination (4, 6, 19, 20). We recommend using care when marking respirators during re-use protocols.

Viral and Biological Indicator Inactivation

Viral inactivation by drying depends on multiple factors including surface type, humidity,

temperature, virion size and type, and duration of drying (43–46). All studies of N95 respirator

decontamination include both the passive inactivation of viruses by air drying and their active

decontamination by aHP or comparable treatment (43–46). Clinical exposure of respirators to

SARS-CoV-2 or other viral pathogens would likewise entail ambient drying prior to any

respirator decontamination or re-use (e.g. virus may dry onto an N95 in the course of a work-

shift, or during bagging for decontamination or direct re-use). The "drying-only" time frame

used here was shorter than used in studies modeling re-use in a clinical setting (10, 15–17, 20).

This suggests that crisis-capacity protocols that involve respirator re-use after multiple days of

drying, even without aHP or active decontamination, likely provide substantial levels of viral

inactivation (40, 41).

To model diverse routes of respirator exposure to viral pathogens during use, we inoculated

viruses onto different N95 models and surfaces to thoroughly test for the ability of aHP to

inactivate viruses. This included different regions of the respirator (e.g. outer vs. inner surface) to

discern whether any differences in these fabrics would influence viral inactivation. This work

was inspired by early efforts to verify viral inactivation in the context of N95 respirator surfaces,

such as that of Kenney et al (13). We found scant evidence of viral survival during the aHP

decontamination process. Relative to studies using different decontamination methods to test

viral inactivation on N95 respirators, these data show equivalent or better inactivation of viruses

(6–8, 11, 13, 14, 18–21, 43).

Commercial bioindicator tests have been relied upon for verification of decontamination for

decades (2, 47). Overall we observed parallel outcomes in terms of successful decontamination

of viral species and bacterial spore-based BIs (Table 4), echoing the few other studies that have

used these approaches in parallel (14, 18, 20). During establishment of aHP cycle parameters we

noted a concomitant failure of BIs and viral inactivation in cycle 3, before the dwell period was

added (Tables 3 and 4). The parallels in success or failure of both BIs and viral inactivation

suggest that commercial spore-based BIs provide a useful predictor of success or failure for

decontamination of N95 respirators, particularly in settings where direct viral testing is not

feasible (2, 10, 15–17). The ten cycles of aHP decontamination achieved here are well beyond

the CDC's crisis-capacity plans, which recommend no more than five total rounds of respirator

reuse (40, 41). While clinical use of respirators, by a multitude of healthcare workers, was not

included in the present study, we foresee that such testing will be an important next step.

We used multiple virus species to test the viral inactivation capabilities of aHP decontamination.

These viruses represented multiple characteristics of human viral pathogens, with a range of

virion and genome types and sizes, and previously documented environmental stability (44, 45,

48). Like SARS-CoV-1, SARS-CoV-2 has a high level of environmental stability (49–53).

Coronaviruses have a lipid-enveloped virion of ~120 nm, with no icosahedral capsid core,

containing a single-stranded, linear, positive-sense RNA genome (Table 2). Virions of HSV-1

and phi6 bacteriophage have a lipid envelope, with an underlying icosahedral capsid core. In

contrast, CVB3 has a non-enveloped or naked icosahedral capsid virus. Prior work has shown

that naked-capsid viruses have a higher stability than enveloped viruses, thus demonstrating the

range of aHP decontamination abilities (44, 45, 48). While most pathogens utilized here require

propagation in mammalian cell lines at biosafety level 2 (BSL2), phi6 can be assayed more

flexibly, using rapid bacterial cultures (24-hour turnaround) at biosafety level 1. Phi6 is a natural

pathogen of the bacterial species *Pseudomonas syringae pathovar phaseolicola*, which is itself a

pathogen of green beans. All viral species examined in this study were effectively

decontaminated by aHP (Table 4). While the aerosolization of viruses was not incorporated here,

this approach merits inclusion in future studies. Together, the combination of respirator fit

testing and virus inactivation testing used here indicate that aHP is a viable decontamination

process to enable crisis-capacity reuse of N95 respirators during viral pandemics.

Methods

Decontamination Facility

The decontamination process was carried out in the Eva J. Pell Laboratory for Advanced

Biological Research at The Pennsylvania State University, University Park campus. This facility

is a purpose-built BSL3 enhanced facility, and all required approvals were obtained from the

Institutional Biosafety Committee (IBC) for work involving viruses, as described below.

The primary decontamination process was performed within in an approximate 1,700 cubic feet

sealed Preparation Room (Prep Room), followed by additional virus inactivation testing with

SARS-CoV-2 in a separate, nearby 1,840 cubic feet Prep Room. Procedures and personal

protective equipment (PPE) suitable for the viruses and materials in use were strictly observed in

this biosafety level 3 (BSL-3) facility, which has consistently maintained institutional, CDC, and

USDA approval for work with risk group 3 pathogens since its commissioning in 2014.

Decontamination Preparation

Respirators for decontamination were staged on a portable metal rack located centrally in the

Prep Room (Figure S2). Filtered and conditioned air is supplied to the Prep Room and the air

exhausted from the room is HEPA-filtered. Bubble-tight dampers (Camfil Farr®) were operated

to seal both the supply and exhaust air from associated ductwork during the decontamination

cycle. The CURIS® decontamination unit was programmed, equipment positioned, and the room

doors sealed using polyethylene sheeting and non-porous adhesive tape (Figure S2).

Decontamination Process

The CURIS® decontamination unit programming method utilizes room size to establish the

baseline parameters for charge (initial aHP dispensing) and intermittent aHP pulse periods

(additional aerosol pulses). This is followed by a user-defined dwell period (when no further aHP

is introduced), at closure of the pulse period. Aeration to disperse residual H_2O_2 follows the

dwell period (i.e. room seals are broken and ventilation resumed). Once the user inputs the

room's cubic volume or dimensions, the CURIS unit calculates a suggested duration of charge

and pulse periods. The standard aeration period is 3 hours, unless an auxiliary scavenging system

or other space aeration system is utilized. To account for absorption of aHP into porous materials

in the decontamination space (e.g. N95 respirators), a 30-40% increase in these default settings

was initially used, as recommended by the manufacturer. Additionally, to ensure adequate

contact time of disinfectant to the treated surfaces, a dwell period was added to the continuous

and pulse charge periods. Based on initial results, adjustments in the charge, pulse, and dwell

periods were made to optimize the decontamination process (Tables 3-4).

After completion of the decontamination phase, an aeration or dissipation phase was initiated by

removing fixed room seals and opening exhaust dampers for up to 2-hour periods, with the room

under slight vacuum (0.18-0.19" W.G.). Adjustments were made to room air exchange rate

during aeration to efficiently dissipate detectable H_2O_2 from respirator facepieces. At cycle 5,

final parameters were established to include an aeration exhaust rate greater than or equal to 35

air changes/hour, with make-up air supplied by outdoor air. Following this phase, and after room

H_2O_2 concentration was measured at less than 2 ppm, the respirator-holding rack was either

retained under ventilation or transferred to a separate room (referred to as the "Finishing Room")

with an HVAC air supply curtain to further dry and decompose residual aHP from respirator

facepieces to less than 1 ppm H_2O_2 (24, 25).

Respirator Handling Process

For respirators subjected to repeated rounds of decontamination as part of this study,

decontamination cycles were conducted repetitively from staging through drying. In order to be

considered "dry" or ready for the next cycle, the interior and exterior respirator surfaces were

monitored using a calibrated, hand-held real-time H_2O_2 monitor (ATI PortaSens II). Once H_2O_2

concentrations measured at respirator surfaces at less than 1 ppm, respirators were re-staged for

the next round of decontamination, or packaged and transported for subsequent respirator fit-

testing or virus inactivation analysis. Between each cycle, the treated/dried respirators were

subjected to manual stress, by flexing each respirator bi-directionally and stretching each strap

twice, using a hold position similar that used in respirator donning or doffing. A standard thin-

line VWR Lab Marker was to mark each strap for each round of the aHP process (Figure S2C).

Spore-based Biological Indicators

Commercial biological indicators (BI) were used for verification of decontamination. These

commercially-prepared spore discs or "coupons" (Steris Spordex®) are enclosed in

Tyvek/glassine envelopes (see Figure S3 for image) and contain a mean spore count of 2.4 x 10^5

Geobacillus stearothermophilus (ATCC® 7953;) (47). Between 6-12 BIs per cycle were placed

throughout the room for each decontamination cycle. These were located behind or beneath

equipment and surfaces, on the portable metal rack holding respirators, and either on or nested

within the pairs of respirators to test aHP penetration (Figure S3). After each cycle, each BI

spore disc was transferred from its glassine/Tyvek envelope to tryptic soy broth (Spordex®

Culture Media), incubated at 55°C, and analyzed after 7 days as an indicator of effective

decontamination. There was one instance of a spore disc dropped during transfer, which resulted

in a single positive BI from that cycle (cycle 5; see Table 4).

Chemical Indicators of H_2O_2

410   Chemical indicator strips (Steris Steraffirm® or CURIS® System Hydrogen Peroxide Test Strips)

411   (between 1 and 4 total per cycle) were placed in various locations throughout the Prep Room to

412   indicate the presence of H_2O_2, supporting successful decontamination.

Real-Time Hydrogen Peroxide Monitoring

The portable ATI PortaSens II detector (PortaSens) was used to measure H_2O_2 levels both within

and outside the Prep Room during the decontamination process. Hydrogen peroxide

concentrations were also measured at respirator surfaces, and necessarily reduced to less than 1

ppm prior to handling and sealing for transportation to designated tissue culture rooms, or for

subsequent respirator fit-testing. During the charge and pulse periods of decontamination, real-

time instantaneous sampling was conducted through a sealable wall port (designed for sterilizer

tubing) into the Prep Room. The PortaSens was also used to monitor Prep Room concentrations

at the start of the aeration phase, and to verify concentrations were reduced to less than 2 ppm for

safe re-entry to the room (without respiratory protection); these measurements were made at

breathing zone height (BZH) or 5 feet above floor level. H_2O_2 concentrations were also

monitored outside the Prep Room door seal (see Figure S2), to check for any H_2O_2 leakage.

The U.S. Department of Labor/Occupational Safety and Health Administration (OSHA) and the

American Conference of Governmental Industrial Hygienists (ACGIH) have established or

adopted an eight-hour time-weighted average (TWA) occupational permissible exposure limit

(PEL) to hydrogen peroxide of 1 ppm (54, 55). Though researchers and fit-test subjects were not

anticipated to experience this eight-hour exposure level, 2 ppm H_2O_2 concentration was used as a

safety threshold for room entry, and < 1 ppm for removal of respirators, sealing respirators for

transport, and re-use by study participants. See Supplementary Text (and Table S2) for additional

metrics used to verify that research personnel were not adversely exposed to H_2O_2.

Respirator Selection for the Study

The N95 respirator facepiece models examined in this study include a range of characteristics, including those identified as surgical N95 respirators (no exhalation valve to maintain sterile field), and non-surgical N95 respirators with an exhalation valve (see Table 1 and Figure S1). Respirator models were required to be successfully fitted by test subjects. Since fit test results varied widely by subject during early testing of Alpha ProTech respirators, this model was discontinued from further study. Fit-testing participants included experienced test subjects and/or administrators. Several respirator models were qualitatively fit-tested (QLFT) using Occupational Safety and Health Administration (OSHA) fit test protocols (saccharine challenge) described in the OSHA Respiratory Protection Standard (29 CFR 1910.134) (56). This included standard exercises to challenge respirator fit, over an approximate 8 minute, 30 second period. Additionally, quantitative fit-testing (QNFT) was conducted using OSHA protocol requirements with a T.S.I, Inc. PortaCount Pro+ Model 8038 Fit Tester. This latter method employs condensation nucleus or particle counting technology (CNC or CPC) to measure aerosol concentration outside and inside the facepiece to determine a user fit-factor (57).

Sample Size and Acceptance Criteria for Fit-Test Reliability

Study design for the QLFT endpoint was intended to rigorously evaluate user respirator to facepiece seal using the greatest-available pool of respirators (stockpiled model 3M 8511), while also providing representative feasibility data for the other respirator models (Table 1). The stockpiled 3M 8511 respirators were procured and collected by Penn State from 2006 to 2009. The use of these respirators beyond the manufacturer's recommended shelf-life enabled the study to proceed without consuming respirators that were more urgently needed by frontline workers, and it was supported by CDC crisis-capacity scenarios in force at the time (40, 41). All other respirator models (Table 1) used in this study were within their expiration date. All respirators new and unused at study initiation.

Acceptance criteria for 3M 8511 respirators required a minimum sample of 59 facepieces with no failures during QLFT in order to conclude with 95% confidence that at least 95% of 3M 8511 respirators maintain fit integrity after repeated use and decontamination. A total of seventy-seven 3M 8511 respirators were available for study purposes. A small number of respirators were allocated for QNFT and virology testing after QLFT, since these processes entailed respirator

463   destruction (i.e. grommet insertion for QNFT or mask slicing for viral resuspension) and

464   rendered them inaccessible for subsequent rounds of aHP.

Respirator Fit-Testing

Prior to decontamination, respirators were labeled with a unique identifier. After

decontamination, a hash mark was placed on the lower elastic band of each respirator to identify

the round(s) of decontamination completed (Figure S2). Since the study was conducted during

the COVID-19 pandemic, physical distancing and active clinical operating conditions limited the

use of multiple test subjects. Therefore, two subjects were selected (one male, one female) to

maximize varying size and facial features for fit-testing. Respirators were subjected to QLFT on

the first, fifth, and tenth rounds of decontamination. A small number of respirators per model

were allocated for QNFT during each round of fit testing; due to installation of metal

grommet/probe for QNFT, these could not be reused for subsequent aHP cycles or fit testing.

These were repurposed for subsequent virus inactivation testing in order to conserve overall

respirator use. See Supplementary Text for additional details on QNFT metrics.

NIOSH Filtration Efficiency Testing

To determine whether N95 respirators used in this study experienced filtration media breakdown

as a result of sequential aHP disinfection cycles, several respirators were sent for independent

laboratory analysis. These included eight 3M 8511 respirators subjected to cyclic aHP treatment

and intermittent fit-testing, and two additional unused, untreated 3M 8511 respirators. Respirator

filtration efficiency testing was performed by ICS Laboratories, Inc. (ICS) of Brunswick, Ohio.

ICS is one of two firms in the U.S. authorized by the National Institute for Occupational Safety

and Health (NIOSH) to perform respirator certification or re-certification.

An abbreviated "short-cycle" filtration efficiency verification test was conducted using the

NIOSH Standard Test Procedure TEB-APR-STEP-0059 (58), which is summarized below. The

standard test protocol includes initial respirator conditioning at 85 +/- 5% relative humidity

(%RH) and 38 +/- 2°C for 25 hours prior to testing. This conditioning was intended to reflect

active moisture load created by the respirator user. Following sealing of the respirator exhalation

valve, and placing into the test instrument, full-load testing was performed. This testing included

respirator challenge using 200 milligrams sodium chloride aerosol (mean count particle size

distribution verified as 0.075 +/- 0.020 microns, with geometric standard deviation not exceeding

1.86). Sodium chloride aerosol was neutralized to a "Boltzman equilibrium state" (25 +/- 5°C, 30

+/- 10% relative humidity), and introduced at an airflow rate of 85 +/- 4 liters per minute, with

periodic check and adjustment to maintain this flowrate. Instrumental analysis was conducted

using a T.S.I Automated Filter Tester Model 8130A. Recorded data included flow rate,

resistance, penetration, maximum penetration, and filtration efficiency.

The test protocol establishes the means for ensuring that the particulate filtering efficiency of

N95 series filters used on non-powered respirators submitted for Approval, Extension of

Approval, or examined during Certified Product Audits, meets the minimum certification

standards set forth in 42 CFR, Part 84, Subpart K, §84.181.

Virus Inoculation and Titration

Viruses used here included herpes simplex virus 1 (HSV-1) strain F, coxsackievirus B3 (CVB3),

Pseudomonas phi6 bacteriophage (phi6) and SARS-CoV-2 isolate USA-WA1/2020. The key

characteristics of these viruses are summarized in Table 2. All virus-inoculated materials were

handled in accordance with the biosafety level (BSL) specified for that virus (see Table 2).

Viruses were propagated in the same host cells as used for viral titration (see virus-specific

sections below). In all aHP cycles, the respirators used for virus inactivation testing had been

subjected to preceding total number of aHP cycles (see Table 4 for details). The input inoculum

for each virus was set to the maximum available in each viral stock preparation.

For virus inactivation testing, respirator facepieces were inoculated with a controlled amount of

one or more surrogate virus species (refer to Table 2 for complete list). Each virus was added in

duplicate (or in quadruplicate during "Post 1-3" cycles of testing) droplets of 10 μL each, on one

or more surfaces of each respirator type. The zone of viral inoculation was marked at the corners,

to allow excision of the inoculated area after aHP treatment or air-drying. Droplets were allowed

to air-dry or absorb fully onto each respirator inside a Class II biosafety cabinet (BSC), before

proceeding further. Selected samples of each virus-inoculated respirator were left in ambient

conditions, without aHP decontamination, as a "drying-only" virus control. The remainder of

each virus-inoculated facepiece was subjected to aHP decontamination as described above (i.e.

"decontaminated" samples). For each round of viral testing, the duration of air-drying was

matched to the duration of time needed for aHP treatment of parallel respirators (i.e. including

transport, respirator staging in the Prep Room, aHP decontamination, aeration, and return to the

viral testing location). After aHP treatment, respirators, each virus-inoculated area was cut out of

the "dried" or "decontaminated" respirators using dissection scissors. Each excised virus-spot

encompassed all layers of the respirator, to capture any virus that had absorbed beyond the

surface fabric.

Herpes Simplex Virus 1 and Coxsackievirus B3 Quantification

Excised respirator areas inoculated with HSV-1 or CVB3 were transferred into individual

Eppendorf tubes and resuspended in a 250 μL volume of cell media. Cell media consisted of

Dulbecco's Modified Eagle's Medium (DMEM) supplemented with 10% fetal bovine serum

(FBS) and penicillin-streptomycin (Pen/Strep; 100 U/mL; Thermo Fisher Scientific). The

number of infectious units, or plaque-forming units (PFU), for these viruses was determined by

limiting dilution onto confluent Vero detector cell monolayers (*Cercopithecus aethiops* monkey

kidney cells, ATCC® CCL-81™). Plaque formation was assessed at 72 hours post infection

(hpi), after fixation and visualization of plaques using methylene blue staining. Duplicates of

each virus-inoculated respirator piece were frozen for titration at a subsequent date, separate

from the initial viral quantification. This allowed for experimental redundancy in terms of

detector cell monolayers, and reduced handling time for the large numbers of virus-respirator

inoculation sites.

Bacteriophage phi6 Quantification

Excised respirator areas inoculated with phi6 bacteriophage were transferred into individual

Eppendorf tubes and resuspended in 250 μL of King's medium B (KB). These were then

quantified on lawns of *Pseudomonas syringae pathovar phaseolicola* strain 1448A (*Pph*) using a

previously described soft agar overlay protocol (59). Briefly, 100 μl logarithmic culture of *Pph*

($OD_{600} \sim 0.5$) and 100 μl of the phage preparation were sequentially added to 3 ml of molten soft

agar (0.7%) maintained at 55ºC. The mixture was quickly poured on top of a KB agar plate and

dried in the biosafety cabinet before transferring to a 28ºC incubator. Alternatively, for

enumeration of "inoculum" and "dried" samples, soft agar *Pph* lawns were prepared and 10 μl

dilutions of phage preparation were spotted. Plaque forming units (PFU) were enumerated after

24-48 hours of incubation at 28ºC.

SARS-Coronavirus 2 Quantification

All experiments with SARS-CoV-2 were conducted in the Pell Laboratory under biosafety level 3 (BSL3) enhanced conditions using the USA-WA1/2020 isolate. SARS-CoV-2 was spotted onto masks in quadruplicate and treated with aHP or left in ambient BSL3 conditions for drying-only controls. Respirator areas inoculated with SARS-CoV-2 were then excised and resuspended in 250 µL of DMEM supplemented with sodium pyruvate, non-essential amino acids, antibiotics-antimycotics, and 2% FBS. After resuspension, viral titer was determined by tissue culture infectious dose 50 (TCID50) assay in 96-well plates, using Vero E6 cells. Briefly, 20 µL of resuspended sample was added to 4 wells containing 180 µL of resuspension media. The added samples were then serially diluted 10-fold down the plate, and the plates were incubated at 37 ºC for 96-hours. At this time, cytopathic effect (CPE) was scored, and titer was calculated using the method of Reed and Muench (60).

Acknowledgements

We appreciate the contributions of colleagues and scientists who provided intellectual input and/or materials for this work. These include Joyce Jose, Susan Hafenstein, Anthony Schmitt, Gregory Broussard, Michael Brignati, and Tim Simpson. Paul Turner and his lab graciously provided phi6 bacteriophage. We appreciate the contributions and support of Jim Crandall and the Penn State Environmental Health and Safety staff, as well Theresa Engle and other Occupational Medicine staff, who provided support in study implementation and fit-testing. We also thank Irene Miller and Pell Lab decontamination staff, researchers, clinical staff, as well as Neerav Goyal, Justin Soulier, and the healthcare providers of the Penn State Health Milton S. Hershey Medical Center and College of Medicine.

This work was a largely volunteer-based effort, with support from the Pennsylvania State University and The Penn State Institutes of Energy and the Environment. SARS-CoV-2 research in Dr. Troy Sutton's laboratory was initiated by a seed grant from The Huck Institutes of Life Sciences at Pennsylvania State University, and Dr. Sutton's research is further supported by the USDA National Institute of Food and Agriculture, Hatch project 4605.

579 **References**

1.   Rutala WA, Weber DJ, Healthcare Infection Control Practices Advisory Committee
(HICPAC). 2008. Guideline for Disinfection and Sterilization in Healthcare Facilities,
2008. Cent Dis Control CDC 163.

2.   Centers for Disease Control and Prevention (CDC), National Center for Emerging and
Zoonotic Infectious Diseases (NCEZID), Division of Healthcare Quality Promotion
(DHQP). 2008. Guideline for Disinfection and Sterilization in Healthcare Facilities.
Sterilizing Pract.
https://www.cdc.gov/infectioncontrol/guidelines/disinfection/sterilization/sterilizing-
practices.html.

3.   3M Personal Safety Division. 2020. Disinfection of Filtering Facepiece Respirators. 3M
Tech Bull 4.

4.   Battelle Memorial Institute. 2016. Final Report for the Bioquell Hydrogen Peroxide Vapor
(HPV) Decontamination for Reuse of N95 Respirators. Prep Contract – Study Number
3245 US Food Drug Adm 46.

5.   Liao L, Xiao W, Zhao M, Yu X, Wang H, Wang Q, Chu S, Cui Y. 2020. Can N95
Respirators Be Reused after Disinfection? How Many Times? ACS Nano 14:6348–6356.

6.   Kumar A, Kasloff SB, Leung A, Cutts T, Strong JE, Hills K, Gu FX, Chen P, Vazquez-
Grande G, Rush B, Lother S, Malo K, Zarychanski R, Krishnan J. 2020. Decontamination
of N95 masks for re-use employing 7 widely available sterilization methods. PLOS ONE
15:e0243965.

7.   Smith JS, Hanseler H, Welle J, Rattray R, Campbell M, Brotherton T, Moudgil T, Pack TF,
Wegmann K, Jensen S, Jin J, Bifulco CB, Prahl SA, Fox BA, Stucky NL. 2020. Effect of
various decontamination procedures on disposable N95 mask integrity and SARS-CoV-2
infectivity. J Clin Transl Sci 1–5.

8.   Fischer RJ, Morris DH, van Doremalen N, Sarchette S, Matson MJ, Bushmaker T, Yinda
CK, Seifert SN, Gamble A, Williamson BN, Judson SD, de Wit E, Lloyd-Smith JO,
Munster VJ. 2020. Effectiveness of N95 Respirator Decontamination and Reuse against
SARS-CoV-2 Virus. Emerg Infect Dis 26:2253–2255.

9.   Schwartz A, Stiegel M, Greeson N, Vogel A, Thomann W, Brown M, Sempowski GD,
Alderman TS, Condreay JP, Burch J, Wolfe C, Smith B, Lewis S. 2020. Decontamination
and Reuse of N95 Respirators with Hydrogen Peroxide Vapor to Address Worldwide
Personal Protective Equipment Shortages During the SARS-CoV-2 (COVID-19) Pandemic.
Appl Biosaf 25:67–70.

10. Lieu A, Mah J, Zanichelli V, Exantus RC, Longtin Y. 2020. Impact of extended use and
decontamination with vaporized hydrogen peroxide on N95 respirator fit. Am J Infect
Control 48:1457–1461.

11. Cheng VCC, Wong S-C, Kwan GSW, Hui W-T, Yuen K-Y. 2020. Disinfection of N95
respirators by ionized hydrogen peroxide during pandemic coronavirus disease 2019
(COVID-19) due to SARS-CoV-2. J Hosp Infect 105:358–359.

12. Grillet AM, Nemer MB, Storch S, Sanchez AL, Piekos ES, Leonard J, Hurwitz I, Perkins DJ.
2020. COVID-19 global pandemic planning: Performance and electret charge of N95
respirators after recommended decontamination methods. Exp Biol Med
153537022097638.

13. Kenney PA, Chan BK, Kortright KE, Cintron M, Russi M, Epright J, Lee L, Balcezak TJ,
Havill NL, Martinello RA. 2021. Hydrogen peroxide vapor decontamination of N95
respirators for reuse. Infect Control Hosp Epidemiol 1–3.

14. John AR, Raju S, Cadnum JL, Lee K, McClellan P, Akkus O, Miller SK, Jennings WD,
Buehler JA, Li DF, Redmond SN, Braskie M, Hoyen CK, Donskey CJ. 2021. Scalable in-
hospital decontamination of N95 filtering face-piece respirator with a peracetic acid room
disinfection system. Infect Control Hosp Epidemiol 42:678–687.

15. Jatta M, Kiefer C, Patolia H, Pan J, Harb C, Marr LC, Baffoe-Bonnie A. 2021. N95
reprocessing by low temperature sterilization with 59% vaporized hydrogen peroxide
during the 2020 COVID-19 pandemic. Am J Infect Control 49:8–14.

16. Levine C, Grady C, Block T, Hurley H, Russo R, Peixoto B, Frees A, Ruiz A, Alland D.
2021. Use, re-use or discard? Quantitatively defined variance in the functional integrity of
N95 respirators following vaporized hydrogen peroxide decontamination during the
COVID-19 pandemic. J Hosp Infect 107:50–56.

17. Russo R, Levine C, Grady C, Peixoto B, McCormick-Ell J, Block T, Gresko A, Delmas G,
Chitale P, Frees A, Ruiz A, Alland D. 2021. Decontaminating N95 respirators during the
COVID-19 pandemic: simple and practical approaches to increase decontamination
capacity, speed, safety and ease of use. J Hosp Infect 109:52–57.

18. Christie-Holmes N, Tyli R, Budylowski P, Guvenc F, Weiner A, Poon B, Speck M, Naugler
S, Rainville A, Ghalami A, McCaw S, Hayes S, Mubareka S, Gray-Owen SD, Rotstein OD,
Kandel RA, Scott JA. 2021. Vapourized hydrogen peroxide decontamination in a hospital
setting inactivates SARS-CoV-2 and HCoV-229E without compromising filtration
efficiency of unexpired N95 respirators. Am J Infect Control 49:1227–1231.

19. Ludwig-Begall LF, Wielick C, Jolois O, Dams L, Razafimahefa RM, Nauwynck H,
Demeuldre P-F, Napp A, Laperre J, Thiry E, Haubruge E. 2021. "Don, doff, discard" to
"don, doff, decontaminate"—FFR and mask integrity and inactivation of a SARS-CoV-2
surrogate and a norovirus following multiple vaporised hydrogen peroxide-, ultraviolet
germicidal irradiation-, and dry heat decontaminations. PLOS ONE 16:e0251872.

20. Laatikainen K, Mesilaakso M, Kulmala I, Mäkelä E, Ruutu P, Lyytikäinen O, Tella S,
Humppi T, Salo S, Haataja T, Helminen K, Karppinen H, Kähkönen H, Vainiola T,
Blomqvist K, Laitinen S, Peltonen K, Laaksonen M, Ristimäki T, Koivisto J. 2022. Large-
scale decontamination of disposable FFP2 and FFP3 respirators by hydrogen peroxide
vapour, Finland, April to June 2020. Eurosurveillance 27.

21. Oral E, Wannomae KK, Connolly RL, Gardecki JA, Leung HM, Muratoglu OK, Durkin J,
Jones R, Collins C, Gjore J, Budzilowicz A, Jaber T. 2020. Vaporized H_2O_2
decontamination against surrogate viruses for the reuse of N95 respirators in the COVID-19
emergency. preprint. Occupational and Environmental Health.

22. Berger D, Gundermann G, Sinha A, Moroi M, Goyal N, Tsai A. 2021. Review of aerosolized
hydrogen peroxide, vaporized hydrogen peroxide, and hydrogen peroxide gas plasma in the

decontamination of filtering facepiece respirators. Am J Infect Control
S0196655321004314.

23. Goyal SM, Chander Y, Yezli S, Otter JA. 2014. Evaluating the virucidal efficacy of
hydrogen peroxide vapour. J Hosp Infect 86:255–259.

24. Tuladhar E, Terpstra P, Koopmans M, Duizer E. 2012. Virucidal efficacy of hydrogen
peroxide vapour disinfection. J Hosp Infect 80:110–115.

25. Holmdahl T, Odenholt I, Riesbeck K, Medstrand P, Widell A. 2019. Hydrogen peroxide
vapour treatment inactivates norovirus but has limited effect on post-treatment viral RNA
levels. Infect Dis 51:197–205.

26. Hinton, Chief Scientist, Food and Drug Administration D. 2020. FDA Letter - Emergency
Use Authorization (EUA) for the Battelle Critical Care Decontamination System (CCDS).

27. Hultman C, Hill A, McDonnell G. 2006. The physical chemistry of decontamination with
gaseous hydrogen peroxide. Pharm Eng 27:22–32.

28. Otter JA, Yezli S. 2011. A call for clarity when discussing hydrogen peroxide vapour and
aerosol systems. J Hosp Infect 77:83–84.

29. Boyce JM. 2009. New Approaches to Decontamination of Rooms After Patients Are
Discharged. Infect Control Hosp Epidemiol 30:515–517.

30. Fu TY, Gent P, Kumar V. 2012. Efficacy, efficiency and safety aspects of hydrogen peroxide
vapour and aerosolized hydrogen peroxide room disinfection systems. J Hosp Infect
80:199–205.

31. Holmdahl T, Lanbeck P, Wullt M, Walder MH. 2011. A Head-to-Head Comparison of
Hydrogen Peroxide Vapor and Aerosol Room Decontamination Systems. Infect Control
Hosp Epidemiol 32:831–836.

32. Freyssenet C, Karlen S. 2019. Plasma-Activated Aerosolized Hydrogen Peroxide (aHP) in
Surface Inactivation Procedures. Appl Biosaf 24:10–19.

33. Best EL, Parnell P, Thirkell G, Verity P, Copland M, Else P, Denton M, Hobson RP, Wilcox
MH. 2014. Effectiveness of deep cleaning followed by hydrogen peroxide decontamination
during high Clostridium difficile infection incidence. J Hosp Infect 87:25–33.

34. Barbut F, Menuet D, Verachten M, Girou E. 2009. Comparison of the Efficacy of a
Hydrogen Peroxide Dry-Mist Disinfection System and Sodium Hypochlorite Solution for
Eradication of *Clostridium difficile* Spores. Infect Control Hosp Epidemiol 30:507–514.

35. Shapey S, Machin K, Levi K, Boswell TC. 2008. Activity of a dry mist hydrogen peroxide
system against environmental Clostridium difficile contamination in elderly care wards. J
Hosp Infect 70:136–141.

36. Bartels MD, Kristoffersen K, Slotsbjerg T, Rohde SM, Lundgren B, Westh H. 2008.
Environmental meticillin-resistant Staphylococcus aureus (MRSA) disinfection using dry-
mist-generated hydrogen peroxide. J Hosp Infect 70:35–41.

37. Andersen BM, Syversen G, Thoresen H, Rasch M, Hochlin K, Seljordslia B, Snevold I, Berg
E. 2010. Failure of dry mist of hydrogen peroxide 5% to kill Mycobacterium tuberculosis. J
Hosp Infect 76:80–83.

38. Beswick AJ, Farrant J, Makison C, Gawn J, Frost G, Crook B, Pride J. 2011. Comparison of
Multiple Systems for Laboratory Whole Room Fumigation. Appl Biosaf 16:139–157.

39. Henneman JR, McQuade EA, Sullivan RR, Downard J, Thackrah A, Hislop M. 2022.
Analysis of Range and Use of a Hybrid Hydrogen Peroxide System for Biosafety Level 3
and Animal Biosafety Level 3 Agriculture Laboratory Decontamination. Appl Biosaf 27:7–
14.

40. Centers for Disease Control and Prevention (CDC), National Center for Immunization and
Respiratory Diseases (NCIRD), Division of Viral Diseases. 2020. Implementing Filtering
Facepiece Respirator (FFR) Reuse, Including Reuse after Decontamination, When There
Are Known Shortages of N95 Respirators. Coronavirus Dis 2019 COVID-19 Decontam
Reuse Filter Facepiece Respir. https://www.cdc.gov/coronavirus/2019-ncov/hcp/ppe-
strategy/decontamination-reuse-respirators.html.

41. Centers for Disease Control and Prevention. 2020. Strategies for Optimizing the Supply of
N95 Respirators. Cent Dis Control Prev CDC. https://www.cdc.gov/coronavirus/2019-
ncov/hcp/respirators-strategy/index.html. Retrieved 25 October 2021.

42. 3M Personal Safety Division. 2021. 3M™ Aura™ Health Care Particulate Respirator and
Surgical Mask 1870+, N95. https://multimedia.3m.com/mws/media/890183O/3m-aura-
healthcare-particulate-respirator-and-surgical-mask-1870-brochure.pdf. St. Paul, MN.

43. Rockey N, Arts PJ, Li L, Harrison KR, Langenfeld K, Fitzsimmons WJ, Lauring AS, Love
NG, Kaye KS, Raskin L, Roberts WW, Hegarty B, Wigginton KR. 2020. Humidity and
Deposition Solution Play a Critical Role in Virus Inactivation by Heat Treatment of N95
Respirators. mSphere 5.

44. Vasickova P, Pavlik I, Verani M, Carducci A. 2010. Issues Concerning Survival of Viruses
on Surfaces. Food Environ Virol 2:24–34.

45. Kampf G, Todt D, Pfaender S, Steinmann E. 2020. Persistence of coronaviruses on
inanimate surfaces and their inactivation with biocidal agents. J Hosp Infect 104:246–251.

46. Casanova LM, Jeon S, Rutala WA, Weber DJ, Sobsey MD. 2010. Effects of Air
Temperature and Relative Humidity on Coronavirus Survival on Surfaces. Appl Environ
Microbiol 76:2712–2717.

47. Steris Healthcare. Biological Indicators for Sterilization | STERIS.
https://www.steris.com/healthcare/products/sterility-assurance-and-monitoring/biological-
indicators. Retrieved 14 October 2020.

48. Firquet S, Beaujard S, Lobert P-E, Sané F, Caloone D, Izard D, Hober D. 2015. Survival of
Enveloped and Non-Enveloped Viruses on Inanimate Surfaces. Microbes Environ 30:140–
144.

49. van Doremalen N, Bushmaker T, Morris DH, Holbrook MG, Gamble A, Williamson BN,
Tamin A, Harcourt JL, Thornburg NJ, Gerber SI, Lloyd-Smith JO, de Wit E, Munster VJ.
2020. Aerosol and Surface Stability of SARS-CoV-2 as Compared with SARS-CoV-1. N
Engl J Med https://doi.org/10.1056/NEJMc2004973.

50. Sizun J, Yu MWN, Talbot PJ. 2000. Survival of human coronaviruses 229E and OC43 in
suspension and after drying onsurfaces: a possible source ofhospital-acquired infections. J
Hosp Infect 46:55–60.

51. Duan S-M, Zhao X-S, Wen R-F, Huang J-J, Pi G-H, Zhang S-X, Han J, Bi S-L, Ruan L,
Dong X-P, SARS Research Team. 2003. Stability of SARS coronavirus in human
specimens and environment and its sensitivity to heating and UV irradiation. Biomed
Environ Sci BES 16:246–255.

52. Rabenau HF, Cinatl J, Morgenstern B, Bauer G, Preiser W, Doerr HW. 2005. Stability and
inactivation of SARS coronavirus. Med Microbiol Immunol (Berl) 194:1–6.

53. Geller C, Varbanov M, Duval R. 2012. Human Coronaviruses: Insights into Environmental
Resistance and Its Influence on the Development of New Antiseptic Strategies. Viruses
4:3044–3068.

54. Occupational Safety and Health Administration (OSHA), United States Department of
Labor. 2017. TABLE Z-1 Limits for Air Contaminants. 1910.1000 TABLE Z–1.
https://www.osha.gov/laws-regs/regulations/standardnumber/1910/1910.1000TABLEZ1.

55. American Conference of Governmental Industrial Hygienists (ACGIH). 2019. Hydrogen
Peroxide: Threshold Limit Values (TLV) Chemical Substances, 7th ed. Cincinnati, OH.

56. Occupational Safety and Health Administration (OSHA), United States Department of
Labor. 2004. Fit Testing Procedures (Mandatory). 1910.134 App A.
https://www.osha.gov/laws-regs/regulations/standardnumber/1910/1910.134AppA.

57. Operation and Service Manual. 2015. Portacount Pro 8030 and Portacount Pro+ 8038
Respirator Fit TestersRev. P. T.S.I., Inc., Shoreview, MN.

58. National Institute for Occupational Safety and Health. 2019. Determination of Particulate
Filter Efficiency Level for N95 Series Filters Against Solid Particulates for Non-Powered,
Air-Purifying Respirators Standard Testing Procedure (STP), p. 9. *In* TEB-APR-STEP-
0059Revision 3.2. https://www.cdc.gov/niosh/npptl/stps/apresp.html.

59. Hockett KL, Baltrus DA. 2017. Use of the Soft-agar Overlay Technique to Screen for
Bacterially Produced Inhibitory Compounds. J Vis Exp https://doi.org/10.3791/55064.

60. Reed LJ, Muench H. 1938. A simple method of estimating fifty per cent endpoints. Am J
Epidemiol 27:493–497.

61. Advanced Chemical Sensors. AIHA-LAP ID No. 102047. Longwood, Florida.

62. SGS Galson Laboratories. AIHA-LAP ID No. 100324. E. Syracuse, New York 13057.

Tables

Table 1: N-95 respirator facepiece models included in this study

Brand & model	# in study	Style	Type	Exhalation valve	Notes
3M 8511	77	molded	Non-surgical	yes	
3M 1860	10	molded	Surgical	no	
3M 1870+	11	folded	Surgical	no	Highest fluid resistance*
3M 9211+	12	folded	Non-surgical	yes	Same fabric as 1870+
Honeywell Sperian N11125	5	molded	Non-surgical	yes	
Alpha ProTech	65	folded	Surgical	No	

* Highest level of fluid resistance according to ASTM F1862 at 160 mm Hg (42).

Table 2. Characteristics of virus species used to test inactivation by aerosolized H2O2, as compared to SARS-CoV-2

Virus species (abbreviation, taxonomic family)	Diameter	Capsid / virion shape	Genome type, ~size*	Titer inactivated by aHP^	Biosafety Level (BSL)
Severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2, *Coronaviridae*)	120 nm	Enveloped, no icosahedral capsid	linear (+) ssRNA genome, ~30 kbp	1.6×10^5 PFU/ml	BSL3
Herpes simplex virus 1 (HSV-1, *Herpesviridae*)	200 nm	Enveloped, icosahedral	linear dsDNA genome, ~152 kbp	2.0×10^6 PFU/ml	BSL2
Coxsackie virus B3 (CVB3, *Picornaviridae*)	30 nm	Non-enveloped (naked), icosahedral	linear (+) ssRNA genome, ~7.4 kbp	5.9×10^4 PFU/ml	BSL2
Pseudomonas phi6 bacteriophage (phi6, *Cystoviridae*)	85 nm	Enveloped, icosahedral	segmented, dsRNA genome, ~13.3 kbp	2.4×10^8 PFU/ml	BSL1

*ds, double-stranded; ss, single-stranded; ssRNA genomes are either (+) positive or (-) negative sense.

^Each viral species was tested for decontamination at the maximum available titer.

784   **Table 3: Optimization of aHP treatment parameters**

aHP Cycle #	Room Volume (ft^3)	Charge Period	Pulse Period	Dwell Period	aHP Parameter Set
1a	1700	16:20	40:00	0:00	Initial
1b	1700	16:20	40:00	0:00	Initial
2	1700	11:43	30:00	0:00	Modification 1
3	1700	11:43	30:00	0:00	Modification 1
4	1700	11:43	30:00	0:00	Modification 1
5	1700	11:43	30:00	20:00	Final
6	1700	11:43	30:00	20:00	Final
7	1700	11:43	30:00	20:00	Final
8	1700	11:43	30:00	20:00	Final
9	1700	11:43	30:00	20:00	Final
10	1700	11:43	30:00	20:00	Final
11	1700	11:43	30:00	20:00	Final
12	1700	11:43	30:00	20:00	Final
Post 1	1840	12:41	30:00	20:00	Final*
Post 2	1840	12:41	30:00	20:00	Final*
Post 3	1840	12:41	30:00	20:00	Final*

785   * These cycles were carried out in a larger room, which necessitated adjustment to the charge
786   time to account for the larger room volume.

788   **Table 4: Spore-based Biological indicator (BI) and Virus Inactivation Results**

aHP Cycle #	aHP-treated Spore-based BIs	Control Spore-based BIs	Spore-based BI Results	aHP parameter set	Viruses Tested*	Virus Inactivation Results
1a	12	1	PASS	Initial		
1b	12	1	PASS	Initial		
2	--	--	--	Modification 1		
3	6	1	FAIL (1+/5-)	Modification 1	phi6, HSV1, CVB3	3 of 64 sites positive
4	6	1	PASS	Modification 1		
5	6	1	FAIL^ (1+/5-)	Final		
6	6	1	PASS	Final	HSV1, CVB3	1 of 26 sites positive
7	6	1	PASS	Final		
8	6	1	PASS	Final		
9	6	1	PASS	Final		
10	6	1	PASS	Final		
11	6	1	PASS	Final		
12	6	1	PASS	Final		
Post 1	--	--	--	Final	phi6, HSV1, CVB3	0 of 66 sites positive
Post 2	12	1	PASS	Final	SARS-CoV-2	0 of 24 sites positive
Post 3	12	1	PASS	Final	SARS-CoV-2	0 of 24 sites positive

789   * The N95 respirators used for virus inoculation at each cycle were subjected to all preceding aHP cycles
790   – e.g. respirators used for viral testing in aHP cycle #6 had been through 5 prior aHP cycles. For cycles
791   labeled "Post", the respirators for viral testing had been through 10-12 prior aHP cycles.
792   ^A spore disc dropped during transfer to media was the suspected cause of this failure.
793

794   **Table 5: Portable real-time H₂O₂ measurements, using ATI PortaSens II**

aHP Cycle ID	Measurement Location*	Measurement Condition	H₂O₂ Concentration or range (ppm)**	Time period to decayed concentration (minutes)	aHP Parameter set
1a	IC, wall port	Charge	> 120	--	Initial
1a	IC, wall port	Aeration, to entry	2.5 – 57	52	Initial
1b	IC, wall port	Charge	> 120	--	Initial
2	No data	No data	No data	No data	Modification 1
3	IC, door open	Aeration, to entry	1.5 – 11.6	20	Modification 1
3	IC, door open	Aeration, to respirator measurement	1.5 – 0.8	40	Modification 1
4	IC, door open	Aeration, to entry	2 – 10.6	20	Modification 1
4	respirator surface	After overnight drying	0 – 3	overnight	Modification 1
5	IC, wall port	Charge	> 120	--	Final
6	IC, wall port	Pulse	120	--	Final
6	respirator surface	After overnight drying	0	overnight	Final
7	IC, wall port	Charge	> 120	--	Final
7	OC, seal check(s)	Charge, Pulse, Dwell	0	--	Final
7	IC, door open	Aeration, to entry	2 – 7	27	Final
7	respirator surface	After overnight drying	0	overnight	Final
8	IC, wall port	Pulse	88	--	Final
9	IC, wall port	Charge	> 120	--	Final
9	IC, door open	Aeration, to entry	1.6 – 24	20	Final
9	respirator surface	Aeration, to respirator measurement	< 1	85	Final
10	OC, seal check	Pulse, door seal repair	0 – 1.6	--	Final
10	IC, door open	Aeration, to entry	1.3 – 15	32	Final
10	respirator surface	After overnight drying	< 1	overnight	Final
11	IC, door open	Aeration, to entry	1 – 13.6	35	Final
11	respirator surface	After overnight drying	< 1	overnight	Final
Post	respirator surface	Aeration to 1ppm, respirator data log	1 – 8.1	150	Final

795   ***** Measurement Location: IC – inside containment / treatment area, (wall port) – through sealable wall
796   port, (door open) – door ajar after door seal removed, OC – outside containment / treatment area, (seal
797   check) – Prep Room door tape seal, (respirator surface) – at respirator surface following room entry.
798   ****** Low measurement value represents lowest concentration following concentration decay; high value
799   represents starting concentration (prior to decay).
800

801   **Table 6: Respirator filtration efficiency testing results following 10 cycles of**

802   **aerosolized hydrogen peroxide decontamination**

Full-loading efficiencies – 3M Model 8511 (N95 respirator)*						
Sample ID**	Initial Test Flow Rate (LPM)	Initial Test Resistance (mmH$_2$O)	Initial Test Penetration (%)	Maximum Penetration (%)	Filter Efficiency*** (%)	Result
MS – 1	84	6.5	0.67	2.29	97.71	Pass
MS – 2	85	6.7	0.82	2.38	97.62	Pass
MS – 3	84	6.9	1.46	4.16	95.84	Pass
MS – 4	85	6.8	1.75	4.14	95.86	Pass
FS – 1	85	7.1	2.02	4.58	95.42	Pass
FS – 2	84	6.6	1.47	3.5	96.50	Pass
FS – 3	85	6.6	0.20	2.37	97.63	Pass
FS – 4	86	7.0	1.47	4.12	95.88	Pass
Test specification	81-89			≤ 5.0	≥ 95.0	

803   * Respirator testing analysis performed by ICS Laboratories of Brunswick, OH
804   ** MS, FS – [x]Male subject or female subject sample respirator id
805   ***Filter efficiency percent is based on maximum penetration value.

806   **Figures**

Sequential fit testing (QLFT & QNFT) of N95 respirators subjected to repeated aHP decontamination cycles

Figure 1: Sequential fit testing (QLFT & QNFT) of N95 respirators subjected to repeated

aHP decontamination cycles.

Results demonstrate that all 3M model 8511 respirators successfully pass QLFT and QNFT after

1, 5, and 10 cycles. In particular, 8 of 9 QNFT results for the 3M model 8511 respirator

surpassed a fit factor of 200, the maximum reportable by the test method – providing qualitative,

yet objective, evidence of the safety margin related to fit integrity after 10 aHP decontamination

cycles. All respirator models except Alpha Protech (which had inconsistent fit-test results; see

Methods for details) passed all QLFT and QNFT to which they were subjected. Models Alpha

ProTech (aHP cycle 1) and Kimberly Clark (aHP cycle 2) were only included for a single cycle

and are thus not shown here.

Figure 2: Infectious titer of phi6 bacteriophage inoculated on N95 respirator facepieces was

eliminated after aerosolized H$_2$O$_2$ (aHP) decontamination.

(A) Data are plotted for each aHP cycle in which viral testing was done (see Table 4). Multiple

models of N95 respirator (see Table 1) were inoculated and either treated as "drying only"

controls (red) or subjected to aHP treatment (blue). The respirator surface and model are

indicated by the symbol shape and fill. The median of all points within a given aHP cycle and

treatment is indicated by a solid horizontal line. The dashed horizontal line indicates the limit of

detection (LOD) at 1 viral plaque-forming unit (PFU), in the resuspended but undiluted volume

from the site of viral inoculation. (B) Image depicts Petri dish plating of bacterial lawns exposed

to phi6 from drying-only (left side) or aHP-treated (right side) respirator inoculation sites. These

were applied to the bacterial lawn either as an undiluted resuspension (top row) or 1:10^6 dilution

applied to focal points (bottom row). For the purposes of illustrating the decontaminated sites

where zero plaques were detected, these numbers were replaced with fractional values (0.5), to

allow their visualization on this log-scale plot. See Table S1 for all data values.

Figure 3: Infectious titer of HSV-1 inoculated on N95 respirator facepieces was reduced by

drying and eliminated after aerosolized H$_2$O$_2$ (aHP) decontamination.

Data are plotted for each aHP cycle in which viral testing was done (see Table 4). For HSV-1,

the sole positive plaque after aHP treatment occurred in aHP cycle #3, when the "Modification

1" parameters were in use (Tables 3-4). This failure, in concert with a spore-based biological

indicator and 2 CVB3 plaques (Figure 4) motivated the addition of a "dwell time" in the final

aHP parameters. As in Figure 2, multiple models of N95 respirator (see Table 1) were inoculated

and either treated as "drying only" controls (red) or subjected to aHP treatment (blue). The

respirator surface and model are indicated by the symbol shape and fill. The median of all points

within a given aHP cycle and treatment is indicated by a solid horizontal line. The dashed

horizontal line indicates the limit of detection (LOD) at 1 viral plaque-forming unit (PFU), in the

resuspended but undiluted volume from the site of viral inoculation. For the purposes of

illustrating the decontaminated sites where zero plaques were detected, these numbers were

replaced with fractional values (0.5), to allow their visualization on this log-scale plot. See Table

S1 for all data values.

Figure 4: Infectious titer of CVB3 inoculated on N95 respirator facepieces was reduced by

drying and eliminated after aerosolized H₂O₂ (aHP) decontamination.

Data are plotted for each aHP cycle in which viral testing was done (see Table 4). For CVB3,

two positive plaques after aHP treatment occurred in aHP cycle #3, when the "Modification 1"

parameters were in use (Tables 3-4). This failure, in concert with a spore-based biological

indicator and 1 HSV-1 plaques (Figure 4) motivated the addition of a "dwell time" in the final

aHP parameter. The only other positive CVB3 plaque after aHP treatment occurred in aHP cycle

6 (B), and no plaques were detected in the replicate or in parallel samples. As in Figure 2,

multiple models of N95 respirator (see Table 1) were inoculated and either treated as "drying

only" controls (red) or subjected to aHP treatment (blue). The respirator surface and model are

indicated by the symbol shape and fill. The median of all points within a given aHP cycle and

treatment is indicated by a solid horizontal line. The dashed horizontal line indicates the limit of

detection (LOD) at 1 viral plaque-forming unit (PFU), in the resuspended but undiluted volume

from the site of viral inoculation. For the purposes of illustrating the decontaminated sites where

zero plaques were detected, these numbers were replaced with fractional values (0.5), to allow

their visualization on this log-scale plot. See Table S1 for all data values.

Figure 5: Infectious titer of SARS-CoV-2 inoculated on N95 respirator facepieces was reduced by drying and eliminated after aerosolized H2O2 (aHP) decontamination.

(A-B) Data are plotted separately for each aHP cycle in which viral testing was done (see Table 4). For SARS-CoV-2, no infectious virus was detected by TCID50 assay after aHP treatment. As in Figure 2, multiple models of N95 respirator (see Table 1) were inoculated and either treated as "drying only" controls (red) or subjected to aHP treatment (blue). The respirator surface and model are indicated by the symbol shape and fill. The median of all points within a given aHP cycle and treatment is indicated by a solid horizontal line. Viral titer was determined by tissue culture infectious dose 50 (TCID50) assay in 96-well plates, with a limit of detection (LOD) of 1.2 (see Methods for details). For the purposes of illustrating the decontaminated samples where no virus was detected, these numbers were plotted as a value of 1. See Table S1 for all data values.

Supplementary Text

Methods for hydrogen peroxide diffusion sampling (by H₂O₂ vapor monitors)

Diffusion samplers, also known as hydrogen peroxide vapor monitors (HPMs) were utilized to

monitor long-duration H_2O_2 leakage outside the Prep Room entry (outside containment or OC),

on research personnel during decontamination activities (breathing zone samples, clipped to

lapel-collar), and within sealed respirator transport containers (prior to fit-testing) to verify that

H_2O_2 concentrations remained less than 1 ppm throughout transport and fit-testing. Once the

decontamination process was standardized with specific room ventilation parameters, ongoing

personnel monitoring was discontinued.

HPM sampler analysis was conducted by Advanced Chemical Sensors, Inc. (ACS) of

Longwood, FL utilizing an ACS HP-10 hydrogen peroxide vapor monitor, via modified OSHA

Method VI-6 (colorimetric analysis). Analysis was subcontracted and completed by laboratories

participating in the American Industrial Hygiene Association (AIHA) Laboratory Accreditation

Program. The AIHA is an ISO/IEC 17025 accrediting body (61, 62).

Additional methods on quantitative fit-testing metrics

Based on historical studies by the QNFT device and respirator manufacturers, N95 respirators

have maximum breakthrough for particle sizes of 0.1 – 0.3 micron. Particles smaller or larger

than this size range have increased filtration efficiency (greater than 95%). QNFT fit testing of

N95 respirators therefore focuses on measuring this maximum penetration size range, to ensure

that detectable leakage was due to facepiece seal leakage. OSHA has established a minimum

QNFT passing fit factor of 100 for half-face respirators (including N95s or filtering-facepiece

respirators) (56). The QNFT instrument manufacturer has established a maximum quantifiable fit

factor of 200, as limited by measurement reliability and particle counting factors. QNFT fit

factors equal to or exceeding 200 are therefore reported as 200(+). The reported QNFT data were

utilized to verify successful fit reported by QLFT.

Results for hydrogen peroxide diffusion sampling (by H₂O₂ vapor monitors)

Diffusion samplers, also known as hydrogen peroxide vapor monitors (HPMs) were utilized to

assess researcher exposure levels to H_2O_2, with reference to the OSHA permissible exposure

limit (PEL) and ACGIH Threshold Limit Value (TLV®) of 1 ppm, as an eight-hour time-

weighted average (TWA). HPMs were also collected to monitor H_2O_2 build-up within respirator

containers prior to fit-testing (for test subject safety), and to verify that H_2O_2 concentrations

remained low in areas occupied by researchers (outside containment, OC) during the

decontamination process (Table S2). All personal breathing zone HPM sampler results were

reported below sample detection limits (< 0.08 ppm, <0.1 ppm) for researchers conducting aHP

processes (up to 3 hours). All results for HPMs placed within sealed transport containers were

reported below sample detection limits (< 0.03 ppm), indicating no residual H_2O_2 present

through transport and respirator fit-testing. All results for HPMs collected as OC samples were

reported below sample detection limits (< 0.08 ppm, cycle 1b) and < 0.2 ppm (cycle 11).

Supplementary Tables

Table S1 (Excel): aHP Viral Testing Values

Excel file of data underlying Figures 2-4, and Table 4. These include the number of plaque-

forming units (PFU) for phi6, HSV-1, and CVB3, as well as TCID50(log10)/mL values for

SARS-CoV-2. As noted in Figures 2-4, for samples where no virus was detected, these numbers

were replaced with non-zero values to allow their visualization on the log-scale plots in Figures

2-4. HSV-1, CVB3, and phi6 "no virus detected" values were set to 0.5 (less than 1 plaque

detected), and SARS-CoV-2 values were set to 1 (less than the limit of detection of 1.2

TCID50(log10)/mL).

Table S2: Hydrogen peroxide diffusion sampler (HPM) data

aHP Cycle ID	Sample Type	Location Notes*	H_2O_2 (ppm)**	Sampling Period (minutes)***
1a	Researcher A – personal breathing zone	Sampler at worker lapel; transit in all service areas	< 0.1	143
1b	Researcher B – personal breathing zone	Sampler at worker lapel; transit in all service areas	< 0.08	185
1b	area sample	Door Seal (OC), breathing zone height	< 0.08	183
1b	respirator container	3M 1860 (within brown paper bag)	< 0.03	461
1b	respirator container	3M 9211+ (bag)	< 0.03	518
1b	respirator container	Alpha Protech PFL (within brown paper bag)	< 0.03	497
1b	respirator container	Alpha Protech PFL (within brown paper bag)	< 0.03	510
11	area sample	Door Seal (OC), breathing zone height	< 0.2	110

934 *C (charge period), P (pulse period), D (dwell period), A (aeration phase)
935 ** "< Value" indicates sample analysis was below limit of quantitation (LOQ) based on sampler diffusion
936 rate and air volume collected. Time-weighted average (TWA) not calculated for results reported <LOQ.
937 The OSHA PEL and ACGIH TLV® exposure standard for hydrogen peroxide is 1ppm eight-hour TWA.
938 ***Sampling period varied by sampling objectives.
939

Supplementary Figures

Figure S1: Respirator facepiece models & characteristics

(A) 3M models 1860 and 8511 (left) have a molded facepiece, with 8511 including an exhalation

valve. (B) Virus-inoculation via droplets spotted onto the respirator fabric are shown for two

examples. (C) 3M models 1870+ and 9211+ share a common folded fabric model, with a high

fluid resistance of the fabric. 3M 9211+ includes an exhalation valve. Several of the facepieces

(1860, 1870+, 9211+) display the metal grommet from insertion of the probe used for

quantitative fit testing (QNFT). The inner face of each respirator is shown, which for the models

shown in (C) required a prop to hold them open.

Figure S2: Respirator staging and CURIS® unit.

(A) Respirator staging (partial loading) on metal rack with binder clips and rack centrally

positioned in the Prep Room. (B) Prep Room door sealed with polyethylene sheeting and non-

porous adhesive tape to prevent migration of H_2O_2 outside of the Prep Room. (C) Indelible ink

pen hash marks (arrows) placed to notate completed aHP cycle, at time of cycle completion, or

collection/transport for subsequent fit-testing or virus inactivation testing. (D) CURIS®

decontamination unit with adjustable aHP dispersal nozzle and programmable inputs.

Figure S3: Spore-based biological indicators

(A) Commercial biological indicator (BIs; Steris Spordex®) containing *Geobacillius*

stearothermophilis spores are sold on discs in Tyvek/glassine envelopes, with accompanying

vials of culture media. After treatment, spore discs are transferred from to culture media and

incubated for at 55°C for at least 7 days post decontamination cycle. (B) Media color change to

yellow indicates bacterial growth, as seen in the positive control tube on the far right. In addition

to placement of BIs around the treatment room (see Methods for details), BIs were nested

between two molded respirators (3M 1860) and taped together (C), were enclosed within a flat-

fold type respirator (3M Aura 9211+; not pictured). All BIs positioned in this manner were

rendered inactivated after aHP treatment.

Figure S4: Viral titration demonstrates inactivation and loss of infectious units due to

drying and aerosolized H$_2$O$_2$ (aHP) decontamination

Shown here are representative examples of serial dilutions (titration) of HSV-1 or CVB3 that

were spotted onto 3M 9211+ respirators, and then resuspended and plated onto monolayers of

Vero detector cells. Viral plaques, which are visible as clear foci of infection (plaque-forming

units, or PFU) on the background of methylene blue-stained cells, were visualized at 72 hours

post infection. The plate shown in **(A)** illustrates serial dilution of a high concentration of HSV-1

after drying-only (10^5 PFU; see plot in Figure 3A), and a lower concentration of CVB3 after

drying-only (10^2 PFU; see plot in Figure 4A). The plate shown in **(B)** is from an aHP cycle run

with "Modification 1" parameters, when no dwell time was used (aHP cycle 3; see Tables 3-4)

and commercial spore-based biological indicators indicated a failure of decontamination. Even

on this partial aHP decontamination, this sensitive detection method revealed only three wells (of

which two are shown above) with any viral plaques (indicated by arrowheads; these equate to 25

PFU of HSV-1, and 5 PFU of CVB3; see plots in Figure 3A and 4A).