1 **Fast and easy disinfection of coronavirus-contaminated face masks using**
2 **ozone gas produced by a dielectric barrier discharge plasma generator**
3

4 Jinyeop Lee[1+], Cheolwoo Bong[1+], Pan K. Bae[2], Abdurhaman T. Abafogi[1], Seung H. Baek[1], Yong-Beom
5 Shin[2,3,4], Moon S. Bak[1,5,*], and Sungsu Park[1,5,6,*]
6

[1]School of Mechanical Engineering, Sungkyunkwan University, Suwon 16419, Korea
[2]BioNano Health Guard Research Center (H-GUARD), Daejeon 34141, Korea
[3]Bionanotechnology Research Center, Korea Research Institute of Bioscience and Biotechnology
(KRIBB),
Daejeon 34141, Korea
[4]Department of bioengineering, KRIBB School, University of science and Technology (UST), Daejeon
34141,
Korea
[5]Biomedical Institute for Convergence at SKKU (BICS), Sungkyunkwan University, Suwon 16419,
Korea
[6]Institute of Quantum Biophysics (iQB), Sungkyunkwan University, Suwon 16419, Korea
18
19

20 [+] Each contributed equally to the work.

21 *Correspondence and requests for materials should be addressed to M. S. Bak
22 (moonsoo@skku.edu)/S. Park (nanopark@skku.edu).

Abstract

Face masks are one of the currently available options for preventing the transmission of the

severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2), which has caused the 2019

pandemic. However, with the increasing demand for protection, face masks are becoming

limited in stock, and the concerned individuals and healthcare workers from many countries

are now facing the issue of the reuse of potentially contaminated masks. Although various

technologies already exist for the sterilization of medical equipment, most of them are not

applicable for eliminating virus from face masks. Thus, there is an urgent need to develop a

fast and easy method of disinfecting contaminated face masks. In this study, using a human

coronavirus (HCoV-229E) as a surrogate for SARS-CoV-2 contamination on face masks, we

show that the virus loses its infectivity to a human cell line (MRC-5) when exposed for a

short period of time (1 min) to ozone gas produced by a dielectric barrier discharge plasma

generator. Scanning electron microscopy and particulate filtration efficiency (PFE) tests

revealed that there was no structural or functional deterioration observed in the face masks

even after they underwent excessive exposure to ozone (five 1-minute exposures).

Interestingly, for face masks exposed to ozone gas for 5 min, the amplification of HCoV-229E

RNA by reverse transcription polymerase chain reaction suggested a loss of infectivity under

the effect of ozone, primarily owing to the damage caused to viral envelopes or envelope

proteins. Ozone gas is a strong oxidizing agent with the ability to kill viruses on hard-to-

reach surfaces, including the fabric structure of face masks. These results suggest that it may

be possible to rapidly disinfect contaminated face masks using a plasma generator in a well-

ventilated place.

Keyword: face mask, coronavirus, ozone gas, disinfection.

71 Introduction

Face masks are serving as one of the options for preventing severe acute respiratory

syndrome coronavirus 2 (SARS-CoV-2), which has caused the 2019 international pandemic of

the coronavirus disease.[1] However, with increasing demands for protection, face masks are

becoming limited in stock; concerned individuals and healthcare workers from many

countries are now facing the issue of the reuse of potentially contaminated masks in many

countries. Thus, there is an urgent need to develop a fast and easy method of disinfecting

contaminated face masks. There already exist technologies for the sterilization of medical

equipment, including personal protective equipment (PPE); these technologies include

autoclave treatment, ethylene oxide gassing, ionized hydrogen peroxide fogging and

hydrogen peroxide vaporization.[2] However, most of them are not practical for disinfecting

face masks with SARS-CoV-2.[2] It was reported that ozone gas produced by plasma generators

can inactivate various types of viruses on different surfaces, including porous ones.[3-6] Ozone

is a powerful oxidizing agent, but it does not linger. Its production, involving the use of

electricity and a normal atmosphere, is easy and inexpensive. However, it has not been

determined whether ozone gas can disinfect face masks contaminated with SARS-CoV-2

without compromising the filtration efficiency of the masks.

Here we show that a human coronavirus (HCoV-229E)[7,8] as a surrogate for SARS-CoV-2

on face masks lost its infectivity to a human cell line (MRC-5) when exposed to gaseous

ozone produced by a dielectric barrier discharge (DBD) plasma generator[9] for a short time (1

min) (Figure 1). Neither structural nor functional deterioration of the face masks even with

excessive exposures (5 times, 5 min per each time) to the ozone were observed by scanning

electron microscopy (SEM) and a particulate filtration efficiency (PFE) test. Interestingly, RNA

of HCoV-229E on the face masks by the ozone for 5 min was amplified by reverse

transcription polymerase chain reaction (RT-PCR). This is the first demonstration of the

potential of using ozone gas for disinfecting face masks contaminated with a coronavirus.

Figure 1. A schematic diagram describing the disinfection of a face mask contaminated by a

coronavirus using ozone produced by a DBD plasma generator. It consists of a high-voltage,

high-frequency power supply and two electrodes separated by a 1 mm-thick alumina

dielectric barrier. Plasma was produced on the face of the device with the perforated

electrode, along the rims of the holes.

■ **RESULTS AND DISCUSSION**

Inhibitory effect of ozone gas on virus and bacteria on face masks. Among the numerous

types of plasma generators, the dielectric barrier discharge (DBD) plasma generator is

considered the most energy-efficient and cost-effective plasma generator for ozone

production; it forms ozone through the dissociation of molecular oxygen (O_2) by collisions

with excited electronic nitrogen populated by electron impacts and the ensuing combination

between the atomic oxygen and O_2.[9]

When face masks, experimentally contaminated with a human coronavirus (HCoV-

229E)[7,8] as a surrogate, were exposed to ozone gas (about 120 ppm) produced by the

plasma generator for either 1 or 5 min, no viable HCoV-22E was recovered from the face

masks (Table 1). Corresponding untreated face masks showed the recovery of about 3 log

units of tissue culture infective dose 50% (log $TCID_{50}$) per mL following 15 min of air drying.

To the best of our knowledge, this is the first demonstration of the potential of using ozone

gas for disinfecting face masks contaminated with a coronavirus. Similar results were

obtained for face masks experimentally contaminated with either influenza A virus (H1N1)[10]

(Table S1) or Gram-positive bacteria *Staphylococcus aureus* (Table S2 and Figure S1) when

exposed to ozone gas. These results suggest that virus and bacteria on face masks can be

inactivated by ozone gas at a concentration of about 120 ppm within a short time (1-5 min).

Table 1. HCoV-229E titer recovered from contaminated face masks with and without
exposure to ozone gas.

Treatment of face masks[a] contaminated with HCoV-229E	Recovered virus[b] (log $TCID_{50}$ ± S.D.)
None	3.0 ± 0.2 (n = 4)
Ozone gas (120 ppm[c], 1 min)	0 (n = 3)
Ozone gas (120 ppm, 5 min)	0 (n = 3)

[a] Samples (30 mm × 35 mm in size) were cut from face masks (Kleenguard®; product number

Y2-44015, Kimberly-Clark Worldwide, Inc., Irving, TX, USA) with 3-layers filtering. The front

side was sprayed with about 250 µL of HCoV-229E culture (about 4.5 log $TCID_{50}$ per mL) and

dried at room temperature for 15 min in a biosafety cabinet before exposed to ozone gas.

[b] Virus particles on the samples were collected by washing the sample surface with 5 mL of

PBS (phosphate buffered saline, pH 7.4) and measured using MRC-5 cells (ATCC, Bethesda,

MD, USA).[11]

[c] The ozone concentrations produced by the DBD plasma generator were measured via UV

absorption spectroscopy.[12]

n: sample number.

Partial degradation of viral RNA by ozone gas

To understand the mechanism underlying the viral inactivation of face masks by ozone gas,

the experimentally contaminated face masks (Kleenguard®; product # Y2-44015, Kimberly-

Clark Worldwide, Inc., Irving, TX, USA)—with and without exposure to ozone gas for 5 min—

were washed, and the washing solutions were assayed with the quantitative reverse

transcription polymerase chain reaction (qRT-PCR).[13] Surprisingly, there was no significant

difference ($p > 0.05$; student's t-test) in the amount of amplifiable RNAs between the
unexposed and exposed masks, indicating that the short exposure may not fully degrade the
viral RNA (Table 2). Similarly, the RNA of either H1N1 (Table S3) or *S. aureus* (Table S4) on the
face masks was not totally degraded by the exposure to ozone gas. These results suggest
that the loss of infectivity could be due to the damage to the viral envelope or envelope
proteins, resulting in failure of the virus to attach itself to host cells.[14]

149
150 Table 2: qRT-PCR of HCoV-229E in washing solutions obtained from contaminated face
151 masks with and without exposure to ozone gas.

Treatment of face masks contaminated with HCoV-229E	Ct value[a] (mean ± S.D.)
None	22.7 ± 0.4 (n = 6)
Ozone gas (120 ppm, 5 min)	23.1 ± 0.6 (n = 6)

152 [a] cycle threshold (Ct) is defined as the number of cycles required for the fluorescent signal to
153 exceed the background signal level (threshold). n: sample number.

154
No structural damage on the filter layer of face masks
To test if the exposure of face masks (Kleenguard®) to either plasma or ozone gas causes
any damage to their filter layer, uncontaminated face masks were exposed to ozone gas for
5 min (five 1-minute exposures). We did not see any noticeable damage on the front and
back side of the face masks with eyes and under a light microscope, either (data not shown).
Their inner filter layer composed of polypropylene meltblown non-woven fabric was further
examined under a SEM. As shown in Figure 2, there was no detectable structural damage
caused to the filter layer of the exposed face masks. The result showed that the repeated
exposures (5 times) of face masks to ozone gas did cause structural damage to the face
masks.

Figure 2. Scanning electron microscopy images of the filter layer of uncontaminated face

masks (Kleenguard®) with and without exposures to ozone gas for 5 min (five 1-minute

exposures). The images were taken by a field-emission scanning electron microscope (FE-

SEM-EDS, JSM7500F, JEOL Ltd., Tokyo, Japan).

No functional deterioration on face masks

The electrocharged filter is an essential component of dust masks such as N95 and KF94

masks and there was a concern that the charge on the filter could be lost with its exposure

to ozone gas. The functioning of KF94 masks (registration number: F1-28712011; Kleannara

Co., Seoul, Korea) which are certified to filter out 94% of particulate matter (about 0.4 μm

diameter), after exposed to ozone gas for 5 min (five 1-minute exposures) was thus assessed

using a standard test for measuring PFE with paraffin oil mists.[15] There was no statistical

difference in PFE between the exposed and unexposed face masks (Table 3). Taken together

with the SEM images, it is suggested that the repeated exposures (5 times) of face masks to

ozone gas do not cause structural or functional damage to the face masks.

181
182
183

184   Table 3: PFE of uncontaminated face masks (KF94) with and without exposure to ozone gas.

Certified testing lab[a]	Ozone gas treatment on uncontaminated face masks	PFE (%)[b] (Mean ± S.D.)
FITI Testing & Research Institute	None	98.5 ± 1.3 (n=3)
	Ozone gas (120 ppm, 5 min)	99.3 ± 1 (n=3)
Korea Mask Laboratory	None	98.4 ± 0.5 (n=6)
	Ozone gas (120 ppm, 5 min)	98.6 ± 0.5 (n=6)

[a] The lab is certified and registered as a testing lab by the ministry of food and drug safety
(MFDS) in Korea.

[b] Measured by a standard method using paraffin mist. n: sample number.

■   **CONCLUSIONS**

In this study, using a human coronavirus (HCoV-229E) as a surrogate for SARS-CoV-2

contamination on face masks, the virus is shown to lose its infectivity to a human cell line

(MRC-5) when exposed for a short period of time (1 min) to ozone gas produced by the DBD

plasma generator. SEM and PFE tests revealed that there was no structural or functional

deterioration observed in the face masks even after they underwent excessive exposure to

ozone (five 1-minute exposures).

Ozone gas is a strong oxidizing agent with the ability to kill viruses on hard-to-reach

surfaces, including the fabric structure of face masks. Inexpensive consumer-grade ozone

generators are widely available. Our results suggest that it may be possible to rapidly

disinfect face masks contaminated with SARS-CoV-2 using a plasma generator in a well-

ventilated place.

■   **MATERIALS AND METHODS**
DBD plasma generator

The plasma generator consists of a high-voltage, high-frequency generator (Minipuls 2.2;

GBS Elektronik GmbH, Großerkmannsdorf, Germany) and two electrodes separated by a 1

mm -thick alumina dielectric barrier. Each electrode was made of a perforated stainless-steel

plate and bare aluminum tape. The plasma was produced only at the perforated electrode,

along the rims of the holes. As the dielectric barrier allows electrons and ions to accumulate

on the surface, preventing the transition from a cold plasma to an arc, a sinusoidal voltage

was applied to steadily produce the plasma (Figure S2). The plasma was turned on for 1 min

after every 4 min to avoid any damage resulting from thermal heating.

UV absorption spectroscopy

The ozone concentrations produced by the DBD plasma generator were measured via UV

absorption spectroscopy.[16] Light from a mercury lamp (BHK 90-0005-01, spectral line: 253.65

nm) was collimated using lenses and an optical fiber and sent through a gas medium, 4.3 cm

above the electrode surface where the plasma is produced. The transmitted light intensity

was then measured using a spectrometer (AvaSpec-2048L). As the wavelength-dependent

absorption cross section of ozone is broad near 253.65 nm and known as $1.137 \cdot 10^{-17}$

$cm^2 molecule^{-1}$, the ozone concentration can be evaluated by the Beer-Lambert law

described as follows:

$$n_{O_3} = -\frac{1}{\sigma_{O_3} L} ln\left(\frac{I}{I_0}\right) \qquad (1)$$

where I and I_0 are the transmitted and incident light intensities; n_{O3} is the number density of

ozone; σ_{O_3} is the ozone absorption cross-section near 253.65 nm; L is the optical path

length. The ozone concentrations under the test conditions were determined to be

approximately 120 ppm.

Coronavirus culture

HCoV-229E was cultured using human fetal lung fibroblast cell (MRC-5; ATCC, Bethesda, MD,

USA) in a 96-well plate.[11]

Exposure of face masks experimentally contaminated with HCoV-229E to ozone gas

Samples (n = 6) were cut (30 mm × 35 mm in size) from face masks and sprayed with about

250 µL of HCoV-229E culture (4.5 log $TCID_{50}$ per mL) in a biosafety cabinet at a biosafety

level-2 (BSL-2) laboratory while wearing face masks and gloves. The samples were then

dried for 15 min at 25 °C and were individually exposed to ozone for both 1 and 5 min using

the plasma generator (Figure 1) in a chemical hood.

Determination of $TCID_{50}$ value

Immediately after the exposure, the virus particles on the samples were collected by

washing the samples with 5 mL of PBS (phosphate buffered saline, pH 7.4). $TCID_{50}$ was

determined by adding serial 10-fold dilutions of HCoV-229E collected from each mask into a

human fetal lung fibroblast cell (MRC-5, ATCC, Bethesda, MD, USA) monolayer in a 96-well

plate.[11] The plates were observed for cytopathic effects for 4 days. The viral titer was

calculated via the Reed and Munch endpoint method.[11] Viral titer collected from the face

masks was measured using MRC-5 cells (ATCC, Bethesda, MD, USA).[11]

RT-PCR

Immediately after the exposure, the virus particles on the samples were collected by

washing the samples with 5 mL of PBS. qRT-PCR was performed using StepOne[TM] Real-Time

PCR system (Applied biosystems, CA, USA) and MG 2X One Step RT-PCR SYBR® Green Master

Mix reagents (Cancer Rop Co Ltd., Seoul, Korea). A segment of the *N* gene of HCoV229E was

amplified using a forward primer (CGCAAGAATTCAGAACCAGAG) and a reverse primer

(GGCAGTCAGGTTCTTCAACAA)[13] with an amplicon size of 83 bp (Bioneer, Daejeon, Korea).

The thermocycler conditions were as follows: reverse transcriptase at 50 °C for 30 min and

an initial denaturation at 95 °C for 5 min, followed by 45 cycles of denaturation at 95 °C for

15 s, annealing at 52 °C for 30 s, and an extension at 72 °C for 30 s. To confirm that the target

amplicon was properly formed, a melting curve analysis was conducted. The fluorescence

intensity was measured within the range of 60–95 °C at a rate of 0.2 °C/s.

SEM

A field-emission scanning electron microscope (FE-SEM-EDS, JSM7500F, JEOL Ltd., Tokyo,

Japan) was used to take images of untreated and treated masks (n=4). To obtain the

morphology of the electrostatic melt blown filter layer, which is the middle layer of the face

mask, the melt blown filter was cut to obtain a 5 mm x 5 mm sample that was coated with

iridium (Ir) for 15 minutes via the ion sputtering method. The FE-SEM-EDS was operated at

15 kV, and the working distance (WD) was 8 mm. All the sample images were acquired with

magnification factors of 25, 50, and 200.

Paraffin Oil Test

The standard test for measuring the filtration efficiency with paraffin oil mist was performed

on masks at FITI Testing & Research Institute (Cheongju, Korea) and Korea Mask Laboratory

(KML; Hanam city, Korea), the authorized testing organizations in Korea. In this test, a tester

first produces paraffin oil mist with a particle size ranging from 0.05 to 1.7 µm, 0.4 µm on

average.[15] The flow containing the paraffin oil aerosol at a concentration of 20 ± 5 mg/m^3 is

then blown towards the mask at a flow rate of 95 L/min. The filtration efficiency, given by Eq.
(2), is then evaluated by measuring the concentrations of paraffin oil mist upstream and
downstream of the mask. The efficiency is a value averaged over 30 s and must be measured
within 3 min after the test starts.

$$P = \frac{C_1 - C_2}{C_1} \times 100 \qquad (2)$$

where P is the filtration efficiency; C_1 and C_2 are the concentrations of paraffin oil mist
upstream and downstream of the mask, respectively.

Supplementary Materials: Supplementary materials can be found at
www.mdpi.com/xxx/s1.

Acknowledgements: We thank Ms. J. So for giving an idea for the project.

Funding: S.P. was supported by the BioNano Health-Guard Research Centre as a Global
Frontier Project (H-guard 2018M3A6B2057299) through the National Research Foundation
(NRF) of Ministry of Science and ICT (MSIT) in Korea. M.S.B. was supported by a grant
through a future integration program of Kangbuk Samsung Hospital-Biomedical Institute for
Convergence (BICS) at Sungkyunkwan University.

Conflicts of Interest: The authors declare no conflict of interest.

References

1.   Leung NHL, Chu DKW, Shiu EYC, et al. Respiratory virus shedding in exhaled breath
and efficacy of face masks. *Nat Med.* 2020. doi:10.1038/s41591-020-0843-2.

2.   Kumar A, Kasloff SB, Leung A, et al. N95 Mask Decontamination using Standard
Hospital Sterilization Technologies. Preprint. Posted online April 08, 2020. MedRxiv.
doi:10.1101/2020.04.05.20049346.

3.   Hudson JB, Sharma M, Vimalanathan S. Development of a practical method for using
ozone gas as a virus decontaminating agent. *Ozone Sci Eng.* 2009;31(3):216-223.
doi:10.1080/01919510902747969.

4.   Zimmermann JL, Dumler K, Shimizu T, et al. Effects of cold atmospheric plasmas on
adenoviruses in solution. *J Phys D Appl Phys.* 2011;44(50):505201. doi:10.1088/0022-
3727/44/50/505201.

5.   Mastanaiah N, Johnson JA, Roy S. Effect of dielectric and liquid on plasma sterilization
using   dielectric   barrier   discharge   plasma.   *PLoS   One.*   2013;8(8):e70840
doi:10.1371/journal.pone.0070840.

6.   Morfill GE, Shimizu T, Steffes B, Schmidt HU. Nosocomial infections—a new approach
towards preventive medicine using plasmas. *New J Phys.* 2009;11(11):115019..

7.   Van Der Hoek L, Pyrc K, Berkhout B. Human coronavirus NL63, a new respiratory
virus.   *FEMS   Microbiol   Rev.*   2006;30(5):760-773.   doi:10.1111/j.1574-
6976.2006.00032.x.

8.   Wentworth D, Holmes KV. Molecular determinants of species specificity in the
coronavirus receptor aminopeptidase N (CD13): influence of N-linked glycosylation. *J*
Virol. 2001;75(20):9741-9752. doi:10.1128/JVI.75.20.9741-9752.2001.

9.   Stefanovic I, Bibinov N, Deryugin A, et al. Kinetics of ozone and nitric oxides in
dielectric barrier discharges in O2/NOx and N2/O2/NOx mixtures. *Plasma Sources Sci*
Technol. 2001;10(3):406.

10. Dotis J, Roilides EJH. H1N1 Influenza A infection. *Hippokratia.* 2009;13(3):135-138.

11.   Reed LJ, Muench H. A simple method of estimating fifty per cent endpoints. *Am J*
Epidemiol. 1938;27(3):493-497. doi:10.1093/oxfordjournals.aje.a118408.

12.   Reuter S, Winter J, Iseni S, et al. Detection of ozone in a MHz argon plasma bullet jet.
Plasma Sources Sci Technol. 2012;21(3):034015.

13.   Lu R, Yu X, Wang W, et al. Characterization of human coronavirus etiology in Chinese
adults with acute upper respiratory tract infection by real-time RT-PCR assays. *PLoS*
One. 2012;7(6):e38638 doi:10.1371/journal.pone.0038638.

14.   Tseng C, Li C. Inactivation of surface viruses by gaseous ozone. *J Environ Health.*
2008;70(10):56-63.

15.   Choi HW, Yoon S, Lee JH, *et al.* Comparison of pressure drop and filtration efficiency
of particulate respirators using welding fumes and sodium chloride. *Ann. Occup. Hyg.*
2011;55(6):666-680. doi: 10.1093/annhyg/mer032.

16.   Orphal J, Staehelin J, Tamminen J, et al. Absorption cross-sections of ozone in the
ultraviolet and visible spectral regions: Status report 2015. *J Mol Spectrosc.*
2016;327:105-121. doi: 10.1016/j.jms.2016.07.007.

340