**Performance evaluation of the point-of-care SAMBA II SARS-CoV-2 Test for**

**detection of SARS-CoV-2**

Sonny M Assennato[a*], Allyson V Ritchie[a*], Cesar Nadala[b], Neha Goel[a], Hongyi

Zhang[c], Rawlings Datir[d], Ravindra K Gupta[e,f], Martin D Curran[c**], Helen H Lee[a,b**#]

*Assennato and Ritchie contributed equally

**Curran and Lee contributed equally

[a.] Diagnostics for the Real World EU Ltd., Chesterford Research Park, Great

Chesterford, UK

[b.] Diagnostics for the Real World Ltd., San Jose, USA

[c.] Public Health England, Clinical Microbiology and Public Health Laboratory,

Cambridge University Hospitals NHS Foundation Trust, Cambridge, UK

[d.] Division of Infection and Immunity, University College London, London, UK

[e.] Department of Medicine, University of Cambridge, Cambridge, UK

[f.] Africa Health Research Institute, Durban, South Africa

**Running Title:** SAMBA II SARS-CoV-2 analytical and clinical evaluation

# Address correspondence to Dr Helen H Lee, email: hl207@cam.ac.uk

**Key words:** SAMBA II, SARS-CoV-2, point-of-care, COVID-19

**Abstract**

Nucleic acid amplification for the detection of SARS-CoV-2 RNA in respiratory samples is the standard method for diagnosis. These tests are centralised and therefore turnaround times can be 2-5 days. Point-of-care testing with rapid turnaround times would allow more effective triage in settings where patient management and infection control decisions need to be made rapidly.

Inclusivity and specificity of the SAMBA II SARS-CoV-2 assay was determined by *in silico* analyses of the primers and probes. Analytical and clinical sensitivity and specificity of the SAMBA II SARS-CoV-2 Test was evaluated for analytical sensitivity and specificity. Clinical performance was evaluated in residual clinical samples compared to the Public Health England reference tests.

The limit of detection of the SAMBA II SARS-CoV-2 Test is 250 cp/mL and is specific for detection of 2 regions of the SARS-CoV-2 genome. The clinical sensitivity was evaluated in 172 clinical samples provided by Public Health England, Cambridge, which showed a sensitivity of 98.9% (95% CI 94.03-99.97%), specificity of 100% (95% CI 95.55-100%), PPV of 100% and NPV of 98.78% (92.02-99.82%) compared to testing by Public Health England (PHE). SAMBA detected 3 positive samples that were initially negative by PHE.

The data shows that the SAMBA II SARS-CoV-2 Test performs equivalently to the centralised testing methods with a much quicker turnaround time. Point of care testing, such as SAMBA, should enable rapid patient management and effective implementation of infection control measures.

## Introduction

The SARS-CoV-2 was first reported in Wuhan, China, in early December 2019 and is the causative agent of coronavirus disease 19 (COVID-19) [1]. It has since spread to over 188 countries/regions around the world [2], causing 317,234 deaths [3]. It was declared a pandemic by the World Health Organisation on the 11th of March 2020 [4]. In Europe, the country with the highest number of deaths is UK, which as of the 17th of May has had 243,694 lab-confirmed cases and 34,636 deaths in all settings [5]. Nucleic acid testing is essential for early diagnosis of SARS-CoV-2 infection as antibody response is often not detected until 7-10 post onset of symptoms [6]. Upper respiratory tract (URT) specimens such as nose and throat swabs generally have high SARS-CoV-2 viral loads upon symptom onset [7]. The standard diagnostic test for SARS-CoV-2 in the UK is done by real-time RT-PCR of the RdRp gene [8], from a combined throat and nose swab sample. The UK has dramatically scaled up testing from 5,000 tests per day in March 2020 to 100,000 tests per day by the end of April. Although this test has good accuracy, the samples must be transported to centralized testing laboratories and batched for processing, which leads to turnaround times of around 48 hours or more. This means treatment of severely ill patients may be suboptimal when other causative pathogens are in the differential diagnosis and those requiring admission or triage with possible COVID-19 maybe unnecessarily isolated or inappropriately cohorted in a COVID-19 ward. This causes obvious bottlenecks in addition to the sheer number of samples that require processing; at present hospital and regional laboratories are at full capacity and a rapid point of care (POC) testing is required.

The SAMBA II nucleic acid testing system was originally designed for HIV testing in

POC and resource-limited settings, with CE-marked products for early infant diagnosis

[9,10] and viral load monitoring [11-13].  Since 2017, SAMBA HIV tests have been

implemented in Uganda, Malawi, Zimbabwe and Central African Republic. The

SAMBA II SARS-CoV-2 Test system has now been developed to specifically detect

the presence of the novel coronavirus, SARS-CoV-2, in nose and throat swab samples

run on the SAMBA II instrument. Test results are available in approximately 1.5 hours.

We have here assessed the analytical and clinical performance of the SAMBA II

SARS-CoV-2 Test using panels and clinical samples.

**Materials and Methods**

**SAMBA sample preparation, amplification, and detection**

For the SAMBA II system, sample preparation, amplification, and detection as well as

reading and interpretation of the results are fully automated. The test is carried out in

the SAMBA II Assay Module under the control of the SAMBA II Tablet Module as

previously described for HIV [9-12], according to the manufacturer's instructions for

use. Nose and throat swab samples are preferably resuspended in 2 mL of SAMBA

SCoV buffer but the system can be used with viral transport medium (VTM) if the

sample is diluted 1:2 with SCoV buffer prior to processing. The input volume for the

SAMBA test is 300 $\mu$l of which 250 $\mu$l is used by the SAMBA II machine as input into

the sample preparation. The SAMBA II SARS-CoV-2 Test specifically amplifies and

detects two regions of the SARS-CoV-2 genome in the ORF1ab (region 1) and

nucleocapsid protein (region 2), reported as two distinct lines on the test strip. A third

line on the test strip, the internal control, is present to control for false negatives

caused by instrument/reagent problems or inhibition. The presence of one or both test

lines indicates a positive result. The presence of just the internal control line indicates

a negative result.

**Virus inactivation in SAMBA SCoV buffer**

SAMBA SCoV buffer, containing Triton x100, is provided with the SAMBA II SARS-

CoV-2 Test for sample collection of the nose and throat swabs. Single round of VSV-

G pseudotyped lentiviruses were produced by transfecting HEK-293T cells in a 3-

plasmid transfection system (HIV Gag-pol expresser under a CMV promoter,

luciferase genome reporter and VSV-g envelope) as previously described [14]. Virus

supernatant was harvested after 24 hours and filtered. Fifty (50) µL

of pseudovirus preparation was added to increasing dilutions of SAMBA SCoV Buffer

in duplicates and incubated at room temperature for 5 minutes. 100µL of the virus-

reagent prep was then used to infect $1.0 \times 10^6$ TZMbl target cells per well in a 96-well

plate and incubated at $37^{\circ}C$, 5% $CO_2$. Forty-eight (48) hours post-infection, luciferase

expression was measured using SteadyGlo and a Glomax Luminometer (both

Promega).

*In silico* **inclusivity analysis**

The SAMBA-SARS-CoV-2 primers and probes for Orf1ab and N regions were

individually evaluated using *in-silico* analysis with respect to 157 SARS-CoV-2

sequences in the NCBI database.

*In silico* **specificity analysis**

*In silico* analysis for possible cross-reactions with related human Coronaviruses

(Human coronavirus 229E, Human coronavirus OC43, Human coronavirus HKU1,

Human coronavirus NL63, SARS-coronavirus and MERS-coronavirus) was conducted

by mapping primers in the SAMBA II SARS-CoV-2 Test individually to the sequences

downloaded from the NCBI database.

An *in silico* analysis for possible cross-reactions with other high-priority organisms was

conducted by carrying out a blastn search for each of the SAMBA primers and probes

against NCBI databases and retrieving all sequences with homologies > 80%.

**Panels and samples**

Panel members for determination of the limit of detection (LOD) were prepared by

making serial dilutions of SARS-CoV-2 RNA from strain 2019-nCoV/Italy-INMI1 from

EVAg (code number: 008N-03894) in pooled negative combined nose and throat

swab samples to target concentrations of 750, 500, 250, 200, 150 and 100

copies/mL.

A Coronavirus RNA specificity panel containing hCoV-NL63, hCoV-229E, hCoV-

OC43, MERS-CoV and SARS-CoV HKU339849 were sourced from EVAg (code

number 011N-03868). RNA samples from this panel were tested at >100,000 cp/test

in the SAMBA II test to determine specificity against other human Coronaviruses.

**Contrived clinical samples**

Combined nose and throat swabs were collected from 35 presumed negative

individuals using FLOQSwabs (Copan, Italy) and SAMBA SCoV buffer.  Thirty

contrived positive clinical samples were prepared by spiking known concentrations of

SARS-CoV-2 RNA strain 2019-nCoV/Italy-INMI1 (EVAg, Italy) into individual negative

specimens to produce final concentrations of 1x LoD (n=3), 2x LoD (n=17), 3x LoD

(n=5), 5x LoD (n=3) and 100x LoD (n=2).

**Clinical evaluation**

The clinical performance of the SAMBA II SARS-CoV-2 Test was further evaluated retrospectively with 172 residual blinded combined nose and throat swab samples from Public Health England (PHE), Cambridge laboratory. These residual samples were from symptomatic individuals with suspected Covid-19 from around the East of England region sent for routine laboratory diagnosis and provided as VTM diluted 1:2 with SAMBA SCoV buffer. In total 172 samples were tested by the SAMBA II SARS-CoV-2 Test and results were compared to the Cambridge RdRp gene (Wuhan) assay on the Rotor gene Q real-time PCR assay routinely used by Public Health England Cambridge Laboratory based on the publication by Sridhar et al., 2020 [15]. Results are expressed as positive or negative with a Ct cut off of 36 for positive results. Samples were also tested with the PHE Colindale (Reference Laboratory) assay, based on the publication by Corman et al, 2020 [8], which amplifies a different region of the RdRp gene.

For LOD, we used SARS-CoV-2 RNA quantified in cp/mL but in the clinical samples, SAMBA II results were compared to real-time PCR results expressed as Ct number (number of cycles). By systematically testing serial dilutions of samples calibrated in cp/mLby the PHE test a correspondence curve was drawn showing for instance that 25 Ct corresponded to 100,000 cp/mL and 30 Ct to 1,000 cp/mL. The curve obtained is added as supplementary information.

**Research Ethics**

Surplus samples obtained from patients known to be symptomatic for COVID-19 and submitted to the Public Health England Cambridge Laboratory for routine testing were retrieved before being discarded. These samples were rendered anonymous and

provided blinded for the purpose of test validation. Public Health England and NHS

Research Ethics Committee have permitted the use of residual samples in this

manner, strictly for the purpose of diagnostic assay validation [16]. The evaluation was

carried out in accordance with The Use of Human Organs and Tissue act [17].

**Results**

**Viral inactivation by SAMBA SCoV collection buffer**

Due to the highly infectious nature of SARS-CoV-2, a proprietary buffer which has a

low pH and a strong detergent was developed for the SAMBA SARS-CoV-2 Test for

sample collection. The ability of the SAMBA SCoV Buffer to inactivate pseudoviruses

was assessed by incubating pseudoviruses with increasing dilutions of

SAMBA SCoV Buffer in duplicates at room temperature for 5 minutes. Forty-eight (48)

hours post-infection, luciferase expression was measured. Figure 1 shows the plot of

$log_{10}$ relative light units (RLU) of luciferase expression vs concentration of

SAMBA SCoV Buffer containing Triton X100. Results show that luciferase activity was

insignificant up to 1:128 dilution of the SAMBA SCoV Buffer, indicating that

the pseudovirus was inactivated after 5-minute exposure at room temperature (Figure

1). For the SAMBA II Test, a 10-minute soaking of the swab in the

SAMBA SCoV Buffer is recommended.

**Limit of detection**

The limit of detection (LOD) of the SAMBA II SARS-CoV-2 Test was determined using

serial dilutions of SARS-CoV-2 RNA in pooled negative combined nose and throat

swab samples. The initial LOD was determined by testing 6 levels at target

concentrations of 750, 500, 250, 200, 150 and 100 copies/mL. Each panel member

was tested in replicates of 3 (Table 1). The final LOD was confirmed by testing 250

copies/mL in replicates of 20, of which all were detected. The results are summarized

in Table 1. Therefore, the claimed LOD of the SAMBA II SARS-CoV-2 Test is 250

cp/mL.

**Inclusivity**

The SAMBA II SARS-CoV-2 Test primers and probes for Target 1 (Orf 1ab) had 100%

match to all but one available SARS-CoV-2 sequence for this region in the NCBI

database (n=157). For this one sequence, a single nucleotide mismatch was found

that maps to the capture probe with no predicted impact on the assay performance.

The primers and probes for Target 2 (N) had 100% identity to all available SARS-CoV-

2 sequences for this region in the NCBI database (n=157).

**Specificity analysis**

*In silico* analysis for possible cross-reactions with related human coronaviruses

(Human coronavirus 229E, Human coronavirus OC43, Human coronavirus HKU1,

Human coronavirus NL63, SARS-coronavirus and MERS-coronavirus) concluded that

none of the SAMBA primers had >80% homology to the organisms listed. In addition

to *in silico* analysis, the SAMBA II SARS-CoV-2 Test was evaluated for specificity by

using >100,000 cp/test of hCoV-Nl63, hCoV-229E, hCoV-OC43, MERS-CoV and

SARS-CoV RNA as target in the assay. All samples gave negative results in SAMBA

showing that both the Orf1ab and N primers and probes in the SAMBA II SARS-CoV-

2 were specific.

An *in silico* analysis for possible cross-reactions with other high-priority organisms

showed that only one SAMBA probe (N region) had greater than 80% homology (81%)

to one of the high priority organisms (*Pneumocystis jirovecii* [PJP]). This marginal

homology would not impact the performance of the test because the other primers and

probes have no homology to *P. jirovecii*. This homology would not be able to compete

with 100% homology of the SARS-CoV-2 amplicon to the capture probe.

**Contrived clinical specimens**

Positive (n=30) and negative (n=35) swab samples were contrived from individual

nose and throat swab samples collected from 35 individuals and tested with the

SAMBA II SARS-CoV-2 Test. Spiked positive samples were tested at 1x LOD (n=3),

2x LOD (n=17), 3x LOD (n=5), 5x LOD (n=3) and 100x LOD (n=2) by spiking SARS-

CoV-2 RNA into individual negative swab samples. All 35 negative samples were

negative and all 30 spiked positive samples were positive when tested with the

SAMBA II SARS-CoV-2 Test (Table 2). The overall sensitivity was 100% (95%CI:

88.43-100%) and specificity was 100% (95%CI: 90-100%).

**Clinical evaluation**

The clinical performance of the SAMBA II SARS-CoV-2 Test was further evaluated

with 172 combined nose and throat samples from symptomatic individuals provided

blinded by Public Health England (PHE), Cambridge. The samples were provided in

Viral Transport Medium (VTM) diluted 1:2 with SAMBA SCoV buffer. After initial testing

there were 87 concordant positives, 81 concordant negatives and 4 discrepant results

(3 SAMBA positive and one SAMBA negative) compared with the PHE Colindale

reference laboratory test (Table 3). The three SAMBA positive samples were repeat

positive by SAMBA and on retest by PHE they were found to be borderline positive

with high CT values for at least one of the target genes on the Colindale or Cambridge

(Wuhan) test (Table 4). The one SAMBA negative sample was negative on repeat by

SAMBA but was positive by PHE for RdRp using both the Cambridge (Wuhan) and

Colindale assays, with Ct values of 28.87 and 31.18 respectively (Table 4). Therefore,

there was just 1 discrepant sample after retest, a false negative for SAMBA (Table 5).

From this data set the SAMBA II SARS-CoV-2 Test has a sensitivity of 98.9% (95%

CI 94.03-99.97%), specificity of 100% (95% CI 95.55-100%), PPV of 100% and NPV

of 98.78% (92.02-99.82%) when compared to the PHE reference tests. The one

SAMBA false negative gave a high Ct value on the PHE test (>31), suggesting low

viral load, and it was diluted (1:2) for SAMBA testing, which may explain the false-

negative result.

**Discussion**

POC molecular tests for SARS-CoV-2, such as SAMBA II, are required to quickly

triage patients as centralised testing can take 2-5 days for results. In addition, POC

tests would be extremely useful for non-laboratory residential settings such as prisons,

immigration centres, nursing homes and rehabilitative centres. The people living in

such facilities tend to be vulnerable populations who are at a higher risk for adverse

outcome and for infection due to living in close proximity to others [18] and early

identification and implementation of increased infection control measures would

reduce spread amongst residents and staff. POC would also be extremely useful in

other situations where rapid and accurate results are required e.g organ donation,

hospital admissions and emergency surgery. POC molecular tests with rapid

turnaround, such as SAMBA II, will be essential not only during high rates of infection

but also as the country begins to end the lockdown period and localised outbreaks will

need to be managed quickly and efficiently.

Our data shows that the SAMBA II SARS-CoV-2 Test is equivalent to centralised

testing with excellent sensitivity and specificity. Samples are inactivated in the SAMBA

collection buffer and results are available within 86-101 minutes at the POC. High

sensitivity and specificity are essential for the appropriate triaging and treatment of

incoming patients. The assay has a limit of detection of 250 cp/mL, which is in-line

with that claimed by other commercial SARS-CoV-2 tests [19,20]. The specificity of

the SAMBA SARS-CoV-2 Test in clinical samples was 100% and the sensitivity was

98.9% compared to the centralised molecular testing by PHE. This data includes 3

positive samples detected by SAMBA that were original negative by centralised testing

indicating good sensitivity. Clinical evaluation by Zhen et al [21] comparing the

performance of Xpert® Xpress SARS-CoV-2 (Cepheid), ePlex® SARS-CoV-

2 (GenMark) and ID NOW™ COVID-19 (Abbott) showed limit of detection of 100

cp/mL, 1,000 cp/mL and 10,000 cp/mL and clinical agreement with the reference

standard of 98.3%, 91.4% and 87.7% respectively. In another study the Xpert system

has been shown to have good agreement with the reference test over a wide range of

Ct values, including low positives [22,23] with a positive agreement of 98.9% and

negative agreement of 92.0% compared to Roche cobas [22].  The Abbott ID NOW

has been shown to have a positive agreement of 73.9% (95% CI: 63.2-82.3%) and

negative agreement of 100% (95%CI: 92.9-100%) compared to the lab-based Roche

cobas system, with the majority of false negative samples being low viral load (>30 Ct

cycles by Roche cobas) [22]. Abbott have since modified the instructions for use to

remove the use of swab in transport medium as samples may become too dilute and

affect the sensitivity [24].

Potential limitations of this study include that the virus inactivation study was carried

out using a constructed pseudovirus rather than a SARS-CoV-2 or other coronavirus

due to availability. Also clinical samples were collected in VTM and diluted 1:2 in SCoV

buffer rather than collected directly into SCoV buffer, which may affect the sensitivity.

**Acknowledgements**

RKG is funded by Wellcome Senior Fellowship In Clinical Science award no WT108082AIA

**References**

1. Huang C, Wang Y, Li X, Ren L, Zhao J, Hu Y, Zhang L, Fan G, Xu J, Gu X,

Cheng Z, Yu T, Xia J, Wei Y, Wu W, Xie X, Yin W, Li H, Liu M, Xiao Y, Gao H,

Guo L, Xie J, Wang G, Jiang R, Gao Z, Jin Q, Wang J, Cao B. 2020. Clinical

features of patients infected with 2019 novel coronavirus in Wuhan, China.

Lancet 395(10223):497-506.

2. COVID-19 Dashboard by the Center for Systems Science and Engineering

(CSSE) at John Hopkins University (JHU). ArcGIS. John Hopkins University.

(Accessed 18th May 2020).

3. https://www.worldometers.info/coronavirus/ (Accessed 18th May 2020).

4. World Health Organisation. 2020. WHO Director-General's opening remarks at

the media briefing on COVID-19 - 11 March 2020. Available from:

https://www.who.int/dg/speeches/detail/who-director-general-s-opening-

remarks-at-the-media-briefing-on-covid-19---11-march-2020. (Accessed 10th

May 2020)

5. https://coronavirus.data.gov.uk (Accessed 18th May 2020).

6. Grzelak L, Temman S, Plachais C, Grzelak L, Temmam S, Planchais C,

Demeret C, Huon C, Guivel F, Staropoli I, Chazal M, Dufloo J, Planas

D, Buchrieser J, Rajah MM, Robinot R, Porrot F, Albert M, Chen K,

Crescenzo B, Donati F, Anna F, Souque P, Gransagne M, Bellalou J,

Nowakowski M, Backovic M, Bouadma I, Le Fevre L, Le Hingrat

Q,  Descamps D, Pourbaix A, Yazdanpanah Y, Tondeur L, Besombes C,

Ungeheuer M, Mellon G, Morel P, Rolland S, Rey F, Behillil S, Enouf

V, Lemaitre A, Creach M, Petres S, Escriou N, Charneau P, Fontanet A, Hoen

B, Bruel T, Eloit M, Mouquet H, Schwartz O, van der Werf S. 2020. SARS-

CoV-2 serological analysis of COVID-19 hospitalized patients, pauci-

symptomatic individuals and blood donors. medRxiv

https://doi.org/10.1101/2020.04.21.20068858

7.  Pan Y, Zhang D, Yang P, Poon LIM, Wang Q. 2020. Viral load of SARS-CoV-2

in clinical samples. The Lancet. 20(4):411-412.

8.  Corman, VM, Landt O, Kaiser M, Molenkamp R, Meijer A, Chu DKW, Bleicker

T, Brnink S, Schneider J, Schmidt ML, Mulders DGJC, Haagmans BL, van der

Veer B, van den Brink S, Wijsman L, Goderski G, Romette J-L, Ellis J, Zambon

M, Peirlis M, Goossens H, Reusken C, Koopmans MPG, Drosten C. 2020.

Detection of 2019 novel coronavirus (2019-nCoV) by real-time RT-PCR. Euro

Surveill, 25(3)  https://doi.org/10.2807/1560-7917.ES.2020.25.3.2000045

9.  Ritchie AV, Goel N, Sembongi H, Lehga J, Farleigh LE, Edemaga D,

Wisniewski CA, Lee HH.(2016. Performance evaluation of the point-of-care

SAMBA I and II HIV-1 Qual whole blood tests. J Virol Methods. Nov;237:143-

149.

10. Ondiek J, Namukaya Z, Mtapuri-Zinyowera S, Balkan S, Elbireer A, Ushiro

Lumb I, Kiyaga C, Goel N, Ritchie A, Ncube P, Omuomu K, Ndiege K,

Kekitiinwa A, Mangwanya D, Fowler MG, Nadala L, Lee H. 2017. Multi-

country validation of SAMBA – A novel molecular point-of-care test for HIV-1

detection in resource-limited setting. J Acquir Immune Defic Syndr. Oct

1;76(2):e52-e57.

11. Ritchie AV, Ushiro-Lumb I, Edemaga D, Joshi HA, De Ruiter A, Szumilin E,

Jendrulek I, McGuire M, Goel N, Sharma PI, Allain JP, Lee HH. 2014. SAMBA

HIV semiquantitative test, a new point-of-care viral-load-monitoring assay for

resource-limited settings. J Clin Microbiol. 52(9):3377-83.

12. Goel N, Ritchie AV, Mtapuri-Zinyowera S, Zeh C, Stepchenkova T, Lehga J,

De Ruiter A, Farleigh LE, Edemaga D, So R, Sembongi H, Wisniewski C,

Nadala L, Schito M, Lee H. 2017. Performance of the SAMBA I and II HIV-1

Semi-Q Tests for viral load monitoring at the point-of-care. J Virol Methods.

Jun;224:39-45.

13. Titchmarsh L, Zeh, C, Verpoort T, Allain, J-P and Lee, H.2015. Leukodepletion

as a Point-of-Care Method for Monitoring HIV-1 Viral Load in Whole Blood. J. Clin.

Microbiol.53(4):1080-1086.

14. Gupta RK, Kohli A, McCormick AL, Towers GJ, Pillay D, Parry CM. 2017 Full

length HIV-1 Gag determines protease inhibitor susceptibility within *in*

*vitro* assays. AIDS. Jul17;24(11):1615-1655.

15. Sridhar S, Forrest S, Kean I, Young J, Scott JB, Maes M, Pereira-Dias J,

Parmar S, Routledge M, Rivett L, Dougan G, Weekes M, Curran M,

Goodfellow I, Baker S. 2020. A blueprint for the implementation of a

validated approach for the detection of SARS-Cov2 in clinical samples in

academic facilities. bioRxiv https://doi.org/10.1101/2020.04.14.041319

16. Public   Health   England.   2017.   UK   Standards   for   Microbiology

Investigations - Evaluations, validations and verifications of diagnostic tests.

Issue no. 5, 16 March 2017

17. UK Legislation. Human Tissue Act. 2004.

18. Infection Prevention and Control guidance for Long-Term Care Facilities in the

context of COVID-19: World Health Organisation. [updated 21 March 2020].

Available from: https://apps.who.int/iris/handle/10665/331508. (Accessed 11

May 2020).

19. FDA. 2020. Xpert® Xpress SARS-CoV-2 Instructions for Use, April 2020.

https://www.fda.gov/media/136314/download (Accessed 10 May 2020).

20. FDA. 2020. Biomerieux ARGENE® SARS-CoV-2 R Gene® Instructions for

Use, May 2020. https://www.fda.gov/media/137742/download (Accessed 10

May 2020).

21. Zhen W, Smith E, Manj R, Schron D, Berry GJ. (2020) Clinical Evaluation of

Three Sample-To-Answer Platforms for the Detection of SARS-CoV-2. J. Clin.

Microbiol. Apr 24 DOI: 10.1128/JCM.00783-20

22. Smithgall MC, Scherberkova' I, Whittier S, Green DA (2020) Comparison of

Cepheid Xpert Xpress and Abbott ID Now to Roche cobas for the Rapid

Detection of SARS-CoV-2. bioRxiv https://doi.org/10.1101/2020.04.22.055327.

23. Moran A, Beavis KG, Matushek SM, Ciaglia C, Francios N, Tesic V, Love N.

2020. The Detection of SARS-CoV-2 using the Cepheid Xpert Xpress SARS-

CoV-2 and Roche cobas SARS-CoV-2 Assays. J Clin Microbiol. Apr 17.

doi: 10.1128/JCM.00772-2

24. Abbott. 2020. ID NOW COVID-19 Technical Brief and Sample Type Labelling

Update. April 2020

**Figure 1. Viral inactivation as demonstrated by Luminescence from a VSV-G pseudotyped reporter virus in descending concentrations of SAMBA Reagent (SCoV Buffer).**

**Table 1**. **LOD of the SAMBA II SARS-CoV-2 Test**

| Concentration SARS-CoV-2 RNA (cp/mL) | Number tested | Number positive | % positive |
|---|---|---|---|
| 750 | 3 | 3 | 100 |
| 500 | 3 | 3 | 100 |
| **250** | **20** | **20** | **100** |
| 200 | 3 | 2 | 66.7 |
| 150 | 3 | 2 | 66.7 |
| 100 | 3 | 0 | 0 |

407    **Table 2**. **Clinical performance in 65 contrived clinical samples**

| Concentration SARS-CoV-2 RNA (cp/mL) | Number tested | Number positive | % agreement (95% CI) |
|---|---|---|---|
| 1-2x LOD | 20 | 20 | 100% (86.1-100%) |
| ≥3x LOD | 10 | 10 | 100 % (74.1-100%) |
| Negative | 35 | 0 | 100% (91.5-100%) |

409    **Table 3**. **Clinical performance in 172 clinical samples compared to PHE**

410                              **reference test (initial results)**

| | | PHE reference test | |
|---|---|---|---|
| | | POSITIVE | NEGATIVE |
| SAMBA SARS-CoV-2 | POSITIVE | 87 | 3 |
| Test | NEGATIVE | 1 | 81 |

416 **Table 4**. Discrepant analysis of 4 samples

| Sample ID | SAMBA | | PHE Colindale assay RdRp | | PHE Wuhan Cambridge assay | | Final result |
|---|---|---|---|---|---|---|---|
| | Initial result | Repeat result | Initial (Ct) | Repeat (Ct) | RdRp (Ct) | E (Ct) | |
| 1-25 | NEG | NEG | POS (31.18) | Not tested | POS (28.87) | Not tested | POS |
| 2-17 | POS | POS | NEG | NEG | NEG | POS (34.99) | POS |
| 2-25 | POS | POS | NEG (36.58) | POS (33.62) | NEG | NEG | POS |
| 2-38 | POS | POS | NEG | NEG | POS (35.09) | POS (34.06) | POS |

419 **Table 5**. **Clinical performance in 172 clinical samples compared to PHE**

420 **reference test (after retest)**

| | | PHE reference test | |
|---|---|---|---|
| | | POSITIVE | NEGATIVE |
| SAMBA SARS-CoV-2 Test | POSITIVE | 90 | 0 |
| | NEGATIVE | 1 | 81 |