Article Summary Line: Ultraviolet C at a dose of 1.5 J/cm^2 is effective at decontaminating

SARS-CoV-2 on some types/models of N95 respirators.

Running Title: UVC Decontamination of SARS-CoV-2 N95 Respirators

Key Words: Ultraviolet C; N95; decontamination; SARS-Co-V2; healthcare personnel;

pandemic; reuse; filtering facepiece respirators; ultraviolet germicidal irradiation; COVID-19

Title: The Effect of Ultraviolet C Radiation Against SARS-CoV-2 Inoculated N95

Respirators

Authors:

David M. Ozog, Jonathan Z. Sexton, Shanthi Narla, Carla D. Pretto-Kernahan, Carmen

Mirabelli, Henry W. Lim, Iltefat H. Hamzavi, Robert J. Tibbetts, and Qing-Sheng Mi

Affiliations:

Henry Ford Health System, Detroit, Michigan, USA (D. Ozog, S. Narla, H. Lim, I. Hamzavi, R.

Tibbetts, Q. Mi)

University of Michigan Medical School, Ann Arbor, Michigan, USA (J. Sexton, C. Pretto-

Kernahan, C. Mirabelli)

University of Michigan College of Pharmacy, Ann Arbor, Michigan, USA (J. Sexton)

University of Michigan Center for Drug Repurposing, Ann Arbor, Michigan, USA (J. Sexton)

Abstract

Since March 31st 2020, during the height of the pandemic, we have decontaminated

thousands of 3M 1860 respirators with Ultraviolet C (UVC) for our frontline workers. There is

no published peer-reviewed data regarding the dose required to effectively disinfect SARS-CoV-

2 on N95 filtering facepiece respirators (FFRs). Four different locations (facepiece and strap) on

5 different N95 FFR models (3M 1860, 8210, 8511, 9211; Moldex 1511) were inoculated with a

10 μL drop of SARS-CoV-2 viral stock (8 x 10^7 TCID$_{50}$/mL). The outside-facing and wearer-

facing surfaces of the respirators were each irradiated with a dose of 1.5 J/cm^2 UVC (254 nm).

Viable SARS-CoV-2 was quantified by a median tissue culture infectious dose assay (TCID$_{50}$).

UVC delivered using a dose of 1.5 J/cm^2, to each side, was an effective method of

decontamination for the facepieces of 3M 1860 and Moldex 1511, and for the straps of 3M 8210

and the Moldex 1511. This dose is an appropriate decontamination method to facilitate reuse of

respirators for healthcare personnel when applied to certain models/materials. Increasing the

dose may improve decontamination for the other models and straps; however, UVC radiation can

degrade certain polymers in a dose dependent manner, and the effects may vary greatly between

different models. Therefore, fit-testing of UVC decontaminated respirators must be performed

each time a new model and/or dose is introduced into the healthcare system.

Introduction

The shortage of personal protective equipment (PPE) is affecting healthcare workers worldwide during the coronavirus disease 2019 (COVID-19) pandemic. The ability to decontaminate and reuse N95 filtering facepiece respirators (FFRs) is a partial solution to the current shortage.(1) We recently proposed decontaminating respirators with repurposed dermatology office phototherapy devices, which serve as a platform for ultraviolet C (UVC) germicidal disinfection.(2) On March 31st, during the height of the pandemic, Henry Ford Health System (HFHS) began decontaminating 3M 1860 respirators with UVC and returning them to their original users. Since then, several thousand respirators have been decontaminated.

Previous studies have shown that UVC can inactivate other coronaviruses including Severe Acute Respiratory Syndrome coronavirus (SARS-CoV) and Middle East Respiratory Syndrome coronavirus (MERS-CoV).(3, 4) However, there is no peer-reviewed published data showing the effective disinfection of the causative agent of COVID-19, SARS-CoV-2 by UVC, on intact N95 filtering facepiece respirators (FFRs). Consequently, this was causing significant anxiety in our frontline workers using decontaminated PPE.

The objective of this study was to determine the effect of UVC on decontamination of SARS-CoV-2-innoculated N95 respirators using a variety of FFRs that are available to healthcare employees at Henry Ford Health System (HFHS) in Detroit, MI.

Methods

The study was performed as a collaboration between the HFHS and the University of Michigan. All study procedures were approved and conducted according to the University of Michigan Institutional Biosafety Committee BSL3 (Biosafety Level 3). The appropriate training and medical surveillance for experimental procedures and manipulations performed in the BSL3

70 facility were satisfied by all individuals directly involved in laboratory testing at University of

71 Michigan.

Virus and Preparation of Viral Stocks

The SARS-CoV-2 strain used was USA-WA1/2020 NR-52281. Viral stocks of SARS-

COV-2 were obtained from the Biodefense and Emerging Infections Research Resources

Repository and were propagated in Vero-E6 cells grown in Dulbecco's Modified Eagle Medium

(DMEM) without phenol red, with 2% Fetal Bovine Serum (FBS), L-glutamine,

penicillin/streptomycin, non-essential amino acids, and hydroxyethyl piperazineethanesulfonic

acid (HEPES). The virus stock was purposely produced in a phenol red-free medium to avoid

photodegradation or photooxidation that may affect the results. For stock virus titration, aliquots

of viral stock were applied on confluent Vero-E6 cells in 96-well plates for a 50% tissue culture

infectious dose ($TCID_{50}$) assay. Viral stocks were determined to be 8 x 10^7 $TCID_{50}$/mL.

Test respirators and UVGI device

Respirators were tested 100% intact and included the following models: 3M 1860 (St.

Paul, MN); 3M 8210 (St. Paul, MN); Moldex 1511 (Culver City, CA); 3M 8511 (St. Paul, MN);

and 3M 9211(St. Paul, MN).The ultraviolet germicidal irradiation device (UVGI) was

manufactured by Daavlin (Byron, OH), with custom dimensions (22 in. x 10 in. x 8 in) to fit

under the BSL3 biosafety hood. The irradiance of the device was approximately 22 mW/cm^2 at

the level of the surface where the masks were placed during the irradiation. Other than

dimensions and a slightly higher irradiance, this UVGI device is identical to the devices used by

HFHS to decontaminate respirators for frontline healthcare personnel. Prior to initiating the

experiment, the irradiance of the device was measured, and a built-in dosimeter was calibrated to

adjust the irradiation.

Decontamination studies

Intact FFRs in a donned position were inoculated on the outside-facing surface with a single 10 μL drop of viral stock (8 x 10^7 TCID$_{50}$/mL) on 4 areas to account for differing received doses on complex surfaces: nosepiece, apex, chin-piece, and strap **(Figure 1A)**. Inoculated respirators were dried in a biosafety cabinet at room temperature for 40 minutes. For each N95 respirator model, FFRs were UVC-irradiated or left untreated as positive controls for viral load recovery. The respirators were then placed under the UVGI device and were individually treated with a dose of 1.5 J/cm^2. Then, they were rotated and the wearer-facing side of the N95 was again irradiated with 1.5 J/cm^2. The irradiation time for each side was approximately 60-70 seconds. The device does not generate any heat; as such, all FFRs were exposed to UVC at room temperature. Immediately after the completion of the irradiation, 4 mm circles containing the inoculated surface were obtained with a leather belt eyelet hole punch tool and were placed in 300 μL (microliters) of PBS for 1 hour at room temperature.

Recovered viral loads were determined by TCID$_{50}$ assay of the absorbed samples. Briefly, 25 μL aliquots of serially 10-fold diluted samples were inoculated into 96-well plates with a Vero-E6 cell monolayer in sextuplicate and cultured in DMEM with 2% FBS, L-glutamine, penicillin/streptomycin, non-essential amino acids, and HEPES. The plates were observed for cytopathic effects for 4 days. Viral titer was calculated with the Reed and Müench endpoint method.(5) Viral yields were expressed as total TCID$_{50}$ recovered in 300 μL or TCID$_{50}$/ 4mm punch. TCID$_{50}$ negative controls were cells with media only and were included on each plate assayed. All negative controls had no cytopathic effect (CPE). The limit of detection (LOD) for the TCID$_{50}$ assay was determined to be $10^{1.3}$ TCID$_{50}$ / 4 mm punch. If the amount of viral particles was below the LOD, then a theoretical, yet low content of viruses may be present.(6)

However, an absence of CPE in the Vero E6 cells at 4 days post inoculation indicates a loss of

infectivity and is evidence of inactivation of the SARS-CoV-2 samples **(Figure 1B)**.(6)

Results

Following preliminary testing **(Appendix Figure 1, Appendix Table 1)**, virus

inoculation was performed on all 5 types of respirators. For each type, three were irradiated with

UVC, and one was not irradiated to serve as positive control. Similar results were seen for the

3M 1860 respirators as in the preliminary study. All facepiece locations were below the LOD

with absent CPE. Two straps were above the LOD, and one strap was below the LOD with

absent CPE. Sufficient virus (≤ 1 log reduction) was recovered from the untreated positive

controls on all facepiece locations; however, a lower yield was recovered from the untreated

control strap **(Figure 2, Appendix Table 2)**.

On the 3M 8210 respirators, location 1 had two respirators above the LOD, and one

respirator below the LOD with absent CPE. Location 2 had one FFR at the LOD, and two FFRs

below the LOD with absent CPE. Location 3 and all the straps were below the LOD with absent

CPE. Lower virus yields were recovered from the untreated positive control on all facepiece

locations. In contrast, the strap did not absorb the droplet and a sufficient yield was obtained.

(Figure 2, Appendix Table 2). Of note, the amount of virus recovered from the strap on the

untreated positive control was higher than the untreated control virus stock (10 μL in PBS

control). This could have been due to a loss of viral titer of the stock as the control sat in PBS

solution for an hour during the experiment.

On the Moldex 1511, all facepiece locations and straps were below the LOD with absent

CPE. However, there was a lower virus recovery from certain facepiece locations (1 and 3) on

the untreated positive respirator **(Figure 2, Appendix Table 2)**.

For the 3M 8511 and 3M 9211, locations 1 and 2 had FFRs all below the LOD with

absent CPE. Location 3 had one FFR at the LOD, and two respirators below the LOD with

absent CPE. All the straps were above the LOD. All facepiece locations and the strap on the

untreated control had lower virus recovery as compared to the 10 μL in PBS control **(Figure 2,**

Appendix Table 2).

Discussion

Five N95 respirator models were inoculated with SARS-CoV-2 and tested. UVC

delivered using a dose of 1.5 J/cm^2, to each side, was an effective method of decontamination for

the facepieces of 3M 1860 and Moldex 1511, and for the straps of 3M 8210 and the Moldex

1511. This is consistent with previous results using H1N1 influenza demonstrating that UVC

decontamination is dependent on model and material type. Mills et al[7] and Heimbuch et al[4]

reported 1 J/cm^2 dose may not be adequate to kill H1N1 influenza depending on the N95

respirator used. Mills et al found that only facepieces on 12 of 15 models and straps on 7 of 15

models showed a significant (≥ 3 log) reduction of H1N1 influenza viability. Similarly,

Heimbuch et al found that only facepieces on 11 of 15 models and straps on 4 of 15 models

showed a significant (≥ 3 log) reduction of H1N1 influenza viability.[8]

Some respirator models have materials, such as the straps of the 3M 1860, that

demonstrate hydrophilic characteristics when inoculated. Moreover, these seemingly hydrophilic

surfaces showed consistently lower mean log reduction $< 3 \log_{10} TCID_{50}$.[4] In contrast,

seemingly hydrophobic materials, such as the 3M 1860 facepiece, were found to demonstrate a $>$

$3 \log_{10} TCID_{50}$ reduction.[4] In our study using SARS-CoV-2, we observed similar results with

the 3M 1860 facepiece and strap. Further, the facepieces of the 3M 8210 have hydrophilic

properties which were reflected in the reduced decontamination results, while the straps did not

readily absorb the droplets, and hence were adequately disinfected. The Moldex 1511 facepiece

and straps also appeared to be hydrophobic and did not absorb the droplets.

Some straps are prone to twisting. Consequently, when the respirator is flipped during

the irradiation process, care must be taken to make sure the appropriate surface of the strap is

exposed to UVC. Also, straps should not inadvertently lay on top of the respirator, hence

creating a shadowing effect. Reduced decontamination seen amongst the straps may not only be

a result of material but secondary to receiving a reduced dosage. UVC devices which provide

360 degrees of irradiation may obviate this issue. Possible respirator-based solutions include a

secondary disinfection step (e.g. Environmental Protection Agency recommended cleansers)

applied only to the straps, or manufacturers may consider using, for example, the same material

as the straps of the 3M 8210 for all the other models of FFRs to improve UVC decontamination.

Our dosage for this study was partially based on previous work with Influenza A (H1N1),

Avian influenza A virus (H5N1), Influenza A (H7N9) A/Anhui/1/2013, Influenza A (H7N9)

A/Shanghai/1/2013, MERS-CoV, and SARS-CoV,[4, 7-9] where it was determined that all areas

of a respirator should receive at least 1 J/cm^2. Preliminary unpublished data from the

Photomedicine and Photobiology Unit at HFHS demonstrated through theoretical and measured

models that the curvature and the distance of the 3M 1860 N95 respirator from the light source

affected the dosage delivered in a predictable way. Moreover, extrapolating from this model,

after irradiating one side of the respirator with 1.5 J/cm^2, some of the lateral aspects may only

receive 900 mJ/cm^2 while the apex of the respirator may receive almost 3 J/cm^2. Further, it was

also observed that a certain percentage of the dosage received in an area (~10%) permeates to the

other side (I. Kohli, unpub. data). Therefore, 1.5 J/cm^2 was chosen as the lowest irradiance to

ensure that all areas received at least 1 J/cm^2. Increasing the dosage delivered may improve

decontamination, but UVC radiation can degrade certain polymers in a dose dependent manner,

and the effects may vary greatly between different models.(10) Therefore, fit-testing of UVC

decontaminated respirators must be performed each time a new model and/or dose is introduced

into the healthcare system.(11)

Our study sampled different areas of each respirator to ensure that all ranges of dosages

were accounted for in a real-world setting against SARS-CoV-2. Other strengths included the

testing of different model types. Of note, the hydrophilic surfaces (e.g. 3M 8210 facepiece and

3M 1860 strap) of untreated positive controls demonstrated a lower virus recovery than control.

Additional testing was performed to determine if the droplet was drying larger than the 4 mm

area tested. The results showed that there was limited to no virus in the periphery of the 4 mm

area tested, and that no virus could be detected on the wearer-facing surface. Moreover, the

lower yield reflects a diminished ability to resuspend the virus after drying. Limitations of the

study include that no soiling agents were used. However, at Henry Ford Health System, as in

other healthcare facilities, personnel are instructed not to reuse respirators that are visibly soiled.

Further, it is still unclear what the infectious dose is for SARS-CoV-2; therefore, it is unknown if

a significant reduction in viral load eliminates contagious risk.

In conclusion, UVC at a dose of 1.5 J/cm^2 is effective at decontaminating SARS-CoV-2

on some N95 respirators. This dose may only be an appropriate decontamination method to

facilitate reuse of PPE for healthcare personnel when applied to certain models/materials. In

addition, some straps may require additional disinfection to maximize the safety to the frontline

workers. Implementation of widespread UVC decontamination methods requires a careful

consideration of model, material type, design, and fit-testing following irradiation. It should also

be emphasized that similar cautions should be practiced for all other methods of respirator

decontamination.

References

1.   Centers for Disease Control and Prevention. Decontamination and Reuse of Filtering

Facepiece Respirators. 2020 [cited 9 April 2020]; Available from:

https://www.cdc.gov/coronavirus/2019-ncov/hcp/ppe-strategy/decontamination-reuse-

respirators.html

2.   Hamzavi IH, Lyons AB, Kohli I, Narla S, Parks-Miller A, Gelfand JM, et al. Ultraviolet

germicidal irradiation: possible method for respirator disinfection to facilitate reuse

during COVID-19 pandemic. J Am Acad Dermatol. 2020 Apr 1.

3.   Bedell K, Buchaklian AH, Perlman S. Efficacy of an Automated Multiple Emitter

Whole-Room Ultraviolet-C Disinfection System Against Coronaviruses MHV and

MERS-CoV. Infect Control Hosp Epidemiol. 2016;37(5):598-9.

4.   Heimbuch BK, Harnish, D. Research to Mitigate a Shortage of Respiratory Protection

Devices During Public Health Emergencies. 2019 [cited 9 April 2020]; Available from:

https://www.ara.com/news/ara-research-mitigate-shortage-respiratory-protection-devices-

during-public-health-emergencies

5.   Reed LJ, Muench H. A SIMPLE METHOD OF ESTIMATING FIFTY PER CENT

ENDPOINTS12. American Journal of Epidemiology. 1938;27(3):493-7.

6.   Darnell MER, Taylor DR. Evaluation of inactivation methods for severe acute respiratory

syndrome coronavirus in noncellular blood products. Transfusion. 2006;46(10):1770-7.

7.   Mills D, Harnish DA, Lawrence C, Sandoval-Powers M, Heimbuch BK. Ultraviolet

germicidal irradiation of influenza-contaminated N95 filtering facepiece respirators.

American journal of infection control. 2018 Jul;46(7):e49-e55.

8.   Narla S, Lyons AB, Kohli I, Torres AE, Parks-Miller A, Ozog DM, et al. The importance

of the minimum dosage necessary for UVC decontamination of N95 respirators during

the COVID-19 pandemic. Photodermatol Photoimmunol Photomed. 2020 Apr 14.

9.   N95DECON. Technical Report for UV-C-Based N95 Reuse Risk Management. 2020

[cited 9 April 2020]; Available from:

https://static1.squarespace.com/static/5e8126f89327941b9453eeef/t/5e854176021146762

3b1e4e4/1585791351199/200401_N95DECON_UV_technicalreport_v1.2_final.pdf

10. Torres AE, Lyons AB, Narla S, Kohli I, Parks-Miller A, Ozog D, et al. Ultraviolet-C and

other methods of decontamination of filtering facepiece N-95 respirators during the

COVID-19 pandemic. Photochemical & photobiological sciences : Official journal of the

European Photochemistry Association and the European Society for Photobiology. 2020

May 15.

11. Ozog D, Parks-Miller A, Kohli I, Lyons AB, Narla S, Torres AE, et al. The Importance

of Fit-Testing in Decontamination of N95 Respirators: A Cautionary Note. Journal of the

American Academy of Dermatology.

Address for correspondence: David M. Ozog, MD, Photomedicine and Photobiology Unit,

Department of Dermatology, Henry Ford Medical Center New Center One, 3031 W. Grand Blvd

Suite 800, Detroit, MI 48202, USA; Phone: 313-916-0412; Fax: 313-916-0609; email:

DOZOG1@hfhs.org

Figure 1A. Locations 1-4 (Nosepiece, Apex, Chin, Strap) on models 1860 and 8210

*Similar locations were sampled on each of the five N95 respirators.

286   Figure 1B. Bright-field microscopy of wells with Vero-E6 cells and SARS-CoV-2

287   *CPE = cytopathic effect

Left: Representative field of inactivated virus (NO CPE).

Right: Virus-induced CPE in Vero E6 cells

Figure 2. Recovered SARS-CoV-2 $TCID_{50}$ / 4mm punch

*Wells that were below the limit of detection (LOD) and had no cytopathic effect were

arbitrarily assigned the value of zero to represent this phenomenon in the above graphs.

Appendix Figure 1. Preliminary testing of 3M 1860 N95 respirator

*Wells that were below the limit of detection (LOD) and had no cytopathic effect were

arbitrarily assigned the value of zero to represent this phenomenon in the above graphs. SARS-

CoV-2 = Severe Acute Respiratory Syndrome coronavirus 2; PBS = phosphate-buffered-saline;

TCID$_{50}$ = 50% tissue culture infectious dose

329 **Appendix Table 1. Preliminary testing of 3M 1860 N95 respirator**

330 *SARS-CoV-2 = Severe Acute Respiratory Syndrome coronavirus 2; PBS = phosphate-buffered-

331 saline; $TCID_{50}$ = 50% tissue culture infectious dose; LOD = limit of detection.

Log_{10} TCID$_{50}$/ 4mm punch (300 μL)	Treated respirator A	Treated respirator B	Untreated respirator A	Untreated respirator B	10 uL of SARS-CoV-2 stock	SARS-CoV2-stock
Location 1	< LOD	< LOD	4.579181246	4.579181246		
Location 2	< LOD	< LOD	4.479181246	3.779181246		
Location 3	< LOD	< LOD	4.679181246	4.079181246		
Location 4	1.301029996	< LOD	2.479181246	2.379181246		
10 μL of SARS-CoV-2 stock diluted in 300 uL of PBS					4.679181246	
25 μL of SARS-CoV-2 stock						7.079181246
Limit of detection	1.301029996					

 Appendix Table 2. Recovered SARS-CoV-2 Log$_{10}$ TCID$_{50}$ / 4mm punch

348 *SARS-CoV-2 = Severe Acute Respiratory Syndrome coronavirus 2; PBS = phosphate-buffered-

349 saline; TCID$_{50}$ = 50% tissue culture infectious dose; LOD = limit of detection.

3M 1860						
Log$_{10}$ TCID$_{50}$ / 4mm punch (300 µL)	Treated respirator A	Treated respirator B	Untreated respirator A	Untreated respirator B	10 uL of SARS-CoV-2 stock	SARS-CoV2-stock
Location 1	< LOD	< LOD	< LOD	5.479181246		
Location 2	< LOD	< LOD	< LOD	4.879181246		
Location 3	< LOD	< LOD	< LOD	5.379181246		
Location 4	1.602059991	2.279181246	< LOD	2.979181246		
10 µL of SARS-CoV-2 stock diluted in 300 uL of PBS					4.579181246	
25 µL of SARS-CoV-2 stock						7.379181246
Limit of detection	1.301029996					
3M 8210						
Log$_{10}$ TCID$_{50}$ / 4mm punch (300 µL)	Treated respirator A	Treated respirator B	Untreated respirator A	Untreated respirator B	10 uL of SARS-CoV-2 stock	SARS-CoV2-stock
Location 1	< LOD	1.602059991	2.379181246	3.579181246		
Location 2	< LOD	1.301029996	< LOD	3.079181246		
Location 3	< LOD	< LOD	< LOD	2.379181246		
Location 4	< LOD	< LOD	< LOD	5.479181246		
10 µL of SARS-CoV-2 stock diluted in 300 uL of PBS					4.579181246	
25 µL of SARS-CoV-2 stock						7.379181246
Limit of detection	1.301029996					
Moldex 1511						
Log$_{10}$ TCID$_{50}$ / 4mm punch (300 µL)	Treated respirator A	Treated respirator B	Untreated respirator A	Untreated respirator B	10 uL of SARS-CoV-2 stock	SARS-CoV2-stock
Location 1	<LOD	< LOD	< LOD	2.579181246		
Location 2	< LOD	< LOD	< LOD	4.379181246		
Location 3	< LOD	< LOD	< LOD	3.479181246		
Location 4	< LOD	< LOD	< LOD	4.679181246		
10 µL of SARS-CoV-2 stock diluted in 300 uL of PBS					5.779181246	
25 µL of SARS-CoV-2 stock						7.379181246
Limit of detection	1.301029996					
3M 8511						
Log$_{10}$ TCID$_{50}$ / 4mm punch (300 µL)	Treated respirator A	Treated respirator B	Untreated respirator A	Untreated respirator B	10 uL of SARS-CoV-2 stock	SARS-CoV2-stock
Location 1	<LOD	<LOD	<LOD	4.479181246		
Location 2	<LOD	<LOD	<LOD	3.779181246		

Location 3	1.301029996	<LOD	<LOD	3.379181246		
Location 4	2.279181246	2.779181246	2.679181246	3.379181246		
10 µL of SARS-CoV-2 stock diluted in 300 uL of PBS					5.779181246	
25 µL of SARS-CoV-2 stock						7.379181246
Limit of detection	1.301029996					

3M 9211						
Log_{10} $TCID_{50}$ / 4mm punch (300 µL)	Treated respirator A	Treated respirator B	Untreated respirator A	Untreated respirator B	10 uL of SARS-CoV-2 stock	SARS-CoV2-stock
Location 1	<LOD	<LOD	<LOD	3.879181246		
Location 2	<LOD	<LOD	<LOD	3.479181246		
Location 3	1.301029996	<LOD	<LOD	2.279181246		
Location 4	2.279181246	2.779181246	2.679181246	2.379181246		
10 µL of SARS-CoV-2 stock diluted in 300 uL of PBS					5.779181246	
25 µL of SARS-CoV-2 stock						7.379181246
Limit of detection	1.301029996					

Left: Representative field of inactivated virus (NO CPE).

Right: Virus-induced CPE in Vero E6 cells

3M 1860 Facepieces

3M 1860 Straps