TITLE

Decontamination of face masks and filtering facepiece respirators via ultraviolet germicidal irradiation,

hydrogen peroxide vaporisation, and use of dry heat inactivates an infectious SARS-CoV-2 surrogate

virus.

RUNNING TITLE

Coronavirus mask and FFR decontamination

AUTHORS

Louisa F. Ludwig-Begall[1,*], Constance Wielick[1,*], Lorène Dams[1], Hans Nauwynck[2], Pierre-Francois

Demeuldre[3], Aurore Napp[3], Jan Laperre[4], Eric Haubruge[5], Etienne Thiry[1,#]

AUTHOR AFFILIATIONS

[1]Veterinary Virology and Animal Viral Diseases, Department of Infectious and Parasitic Diseases,

FARAH Research Centre, Faculty of Veterinary Medicine, Liège University, Liège, Belgium

[2]Laboratory of Virology, Faculty of Veterinary Medicine, Ghent University, Merelbeke, Belgium

[3] Department of Hospital Pharmacy, The University Hospital Center, University of Liège, Liège,

Belgium

[4] Centexbel Textile Research Centre, Grace-Hollogne, Belgium

[5] TERRA Research Centre, Gembloux Agro-Bio Tech, University of Liège, Gembloux, Belgium

[*]These authors contributed equally to this work.

[#]Corresponding author: etienne.thiry@uliege.be

Orcid ID Etienne Thiry: 0000-0002-9585-7933

WORD COUNT

3608 (excluding Summary and References) +400 words for two figures

Non-standard abbreviations:
FFR: filtering facepiece respirator
SM: surgical mask
SARS-CoV-2: severe acute respiratory syndrome coronavirus 2
PRCV: porcine respiratory coronavirus

ABSTRACT

Background

In the context of the ongoing severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2)

pandemic, the supply of personal protective equipment remains under severe strain. To address this

issue, re-use of surgical face masks and filtering facepiece respirators has been recommended; prior

decontamination is paramount to their re-use.

Aim

We aim to provide information on the effects of three decontamination procedures on porcine respiratory

coronavirus (PRCV)-contaminated masks and respirators, presenting a stable model for infectious

coronavirus decontamination of these typically single-use-only products.

Methods

Surgical masks and filtering facepiece respirator coupons and straps were inoculated with infectious

PRCV and submitted to three decontamination treatments, UV irradiation, H_2O_2 vaporisation, and dry

heat treatment. Viruses were recovered from sample materials and viral titres were measured in swine

testicle cells.

Findings

UV irradiation, H_2O_2 vaporisation and dry heat reduced infectious PRCV by more than three orders of

magnitude on mask and respirator coupons and rendered it undetectable in all decontamination assays.

Conclusion

This is the first description of stable disinfection of face masks and filtering facepiece respirators

contaminated with an infectious SARS-CoV-2 surrogate using UV irradiation, H_2O_2 vaporisation and

dry heat treatment. The three methods permit demonstration of a loss of infectivity by more than three

orders of magnitude of an infectious coronavirus in line with the FDA policy regarding face masks and

respirators. It presents advantages of uncomplicated manipulation and utilisation in a BSL2 facility,

therefore being easily adaptable to other respirator and mask types.

KEYWORDS

SARS-CoV-2; PRCV; decontamination (UV; H_2O_2; dry heat); respirator; surgical mask

INTRODUCTION

In the context of the ongoing severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) pandemic, the supply of personal protective equipment (PPE) remains under severe strain and both availability and affordability of items can be subject to fluctuations and disruptions within health-, social care and other essential public facilities. Access to PPE for the health workforce in all services (public and private, community and hospital) has been identified as a key factor in strengthening the international health system response to SARS-CoV-2, the causative agent of COVID-19 [1]; use of medical/surgical face masks and filtering facepiece respirators (FFR i.e. N95 or FFP2 or FFP3 standard or equivalent with a minimum filtration efficiency of 95% for 0.3 μm (aerodynamic mass mean diameter) of sodium chloride aerosols) has been recommended in conjunction with other mitigating measures to prevent transmission of this and other respiratory pathogens [2,3].

Since the surging global demand cannot currently be met solely by the limited capacities of expanding PPE production, national and internationally-coordinated efforts are increasingly focused on re-use of various items [4,5]. While prior decontamination is paramount to re-use of surgical masks or respirators, little information exists on effective decontamination of these typically single-use-only products. Without compromising the fit and filtration integrity of the masks themselves, decontaminating procedures must guarantee the complete inactivation of contaminating pathogens; the US Food and Drug Administration Enforcement Policy for Face Masks and Respirators issued in April 2020 recommends a robust proof of infectious bioburden reduction of three orders of magnitude for viral pathogens, specifically coronaviruses, and six orders of magnitude for either mycobacteria or bacterial spores [6].

Amongst a variety of different methods under investigation, vaporisation of the oxidizing agent hydrogen peroxide (H_2O_2), already a standard hospital sterilisation technology due to its broad antimicrobial activity and efficacy in surface decontamination [7], has garnered attention as a cost-effective and practical option for mask decontamination [5,6,8]. Mask shape and fit of unused 3M FFP2 NR D face masks (type 8822) were shown to remain intact subsequent to two cycles of vaporised H_2O_2 low pressure gas sterilisation in a study by the Dutch National Institute for Public Health and the Environment [5]; further studies have since demonstrated integrity of other respirator models after numerous cycles of H_2O_2 treatment and have shown virucidal and bactericidal activity of the method on a number of respiratory pathogens and/or biological indicators [8,9]. To our knowledge, no study has yet reported the effect of H_2O_2 treatment on SARS-CoV-2, although H_2O_2 vaporisation decontamination trials of SARS-CoV-2 inoculated masks are reportedly underway [8].

As alternatives to chemical H_2O_2 vaporisation, two physical decontamination methods, the application of dry heat and ultraviolet (UV) irradiation, show promise for decontamination of SARS-CoV-2 contaminated masks in various settings. Heat treatment acts via denaturation of protein secondary

structures thereby altering conformation of viral proteins involved in attachment to and replication

within host cells and has long been recognised as an efficient method of virus inactivation [10].

Temperatures of over 65°C have previously been shown to inactivate SARS-CoV in suspension [11];

more recently, dry heat treatment of 70°C was identified to inactivate SARS-CoV-2 in solution [12] and

was shown to not significantly alter filtration efficiency of N95 respirators within 20 cycles of

application [13]. Easily scalable, dry heat allows mass treatment of large sample sizes and thus

potentially presents a fast and efficient decontamination alternative to H_2O_2 vaporisation in decentralised

hospital centres or industrial settings. Ultraviolet germicidal irradiation with a highly energetic short-

wave (254 nm) acts via viral disruption of viral DNA or RNA and constitutes a physical surface

treatment for contaminated masks or FFRs. A useful sterilisation technique in a variety of applications,

UV irradiation has been implemented to effectively decontaminate influenzavirus (H5N1) from two

different models of FFRs (3M models 1860s and 1870) [14] and does not degrade respirator performance

even after multiple applications [13,15]. It is a promising option for rapid decontamination of smaller,

individual sample contingents and is as such easily adaptable to point-of-care applications. A possible

concern relates to UV penetration depth, necessitating studies that not only investigate viral inactivation

of decontaminated surfaces but that address a potential worst-case scenario in which viruses penetrate

deeper layers of contaminated face masks and FFRs [13].

Since the utilisation, concentration and cultivation of infectious SARS-CoV-2 necessary for analyses

investigating its inactivation, pose obvious problems in terms of the availability and equipping of BSL3

facilities, the use of conservative surrogates to test decontamination efficacy of various methods is

justified and crucial to gain an as accurate as possible insight into SARS-CoV-2 decontamination. A

recent (non-peer reviewed) publication describes inactivation of murine hepatitis virus (MHV), a SARS-

CoV-2 surrogate of the same *Betacoronavirus* genus, via H_2O_2 vaporisation, heat treatment, UV

exposure and other decontamination methods on FFRs; regrettably, the limited dynamic range of the

MHV model (at most 10^1-10^2 inactivation) fell short of demonstrating a 10^3 reduction [16].

In the present investigation into filtering facepiece respirator and surgical mask decontamination via UV

irradiation, H_2O_2 and application of dry heat, we implemented porcine respiratory coronavirus (PRCV),

a spike gene deletion mutant of transmissible gastroenteritis virus (TGEV) and a member of the

Alphacoronavirus 1 species [17,18], as SARS-CoV-2 surrogate. While PRCV, which infects the

respiratory tract of swine [19], is not in the same genus as SARS-CoV-2, the two members of the

subfamily *Coronavirinae* in the family *Coronaviridae* show sufficient similarities as to genome length

and virion structure (notwithstanding differences in envelope glycoproteins), for them to be expected to

behave similarly outside their hosts. Indeed, TGEV has previously been utilised as a surrogate for

another *Betacoronavirus*, SARS-CoV, in studies investigating the persistence of coronaviruses on

inanimate surfaces and/or their inactivation with biocidal agents [20,21]. The PRCV model thus

combines the advantages of sufficient genetic and structural relatedness to SARS-CoV-2, the fact that PRCV can be readily propagated and assayed *in vitro* to high titres in the swine testicle (ST) cell line, and the absence of human infection risk.

This is the first description of stable disinfection of FFRs and surgical masks contaminated with an infectious SARS-CoV-2 surrogate using UV irradiation, H_2O_2 vaporisation and dry heat treatment. The three methods permit demonstration of a loss of infectivity by more than three orders of magnitude of an infectious coronavirus in line with the FDA policy regarding face masks and respirators. It presents advantages of uncomplicated manipulation and utilisation in a BSL2 facility, therefore being easily adaptable to other respirator and mask types to which any of the three or other decontamination methods may be applied.

METHODS

Virus and cells

PRCV strain 91V44 [22] was passaged three times on confluent cell monolayers of the ST continuous cell line. A virus stock with a titre of $10^{7.8}$ TCID$_{50}$/ml was used.

Surgical masks and filtering facepiece respirators

All FFRs and surgical masks, commonly used by the health care community in Belgium at the time of the study, were supplied by the Department of the Hospital Pharmacy, University Hospital Centre of Liege (Sart-Tilman). Manufacturers (and models): KN95 FFR - Guangzhou Sunjoy Auto Supplies CO. LTD, Guangdong, China (2020 N°26202002240270); surgical mask (Type II) - Hangzhou Sunten Textile Co., LTD, Hangzhou, China (SuninCare™, Protect Plus). Surgical masks and FFRs were verified to be from the same respective manufacturing lot to minimise any lot-to-lot variation and to ensure consistency during future respirability and filtration performance testing.

Inoculation of surgical masks and filtering facepiece respirators with porcine respiratory coronavirus (PRCV)

Efficacy of three different decontamination methods in inactivating an infectious coronavirus was assessed using surgical masks and FFRs experimentally inoculated with PRCV. Per decontamination method and mask type, one negative control mask or respirator (not contaminated but treated), three treated masks or respirators (PRCV-contaminated and treated), and three positive controls (PRCV-contaminated but untreated), i.e. seven masks in total, were utilised.

Prior to inoculation, the masks and FFRs were first marked with a graphite pencil to enable sample

identification and to outline three square areas (34 mm x 34 mm) to the left, right and centre of the masks

and FFRs, corresponding to areas to be cut out post-inoculation (coupons) with a central site of

inoculation (the demarcation of the coupons and inoculation sites are shown in Supplementary Figure

1). Utilising an ultra-fine insulin syringe and needle (BD Medical), 100 µl of undiluted viral suspension

in Minimum Essential Medium (MEM) were injected at the centre of each of the three square coupons

under the first outer layer of mask or FFR (this to simulate a "worst case scenario" for viral inactivation

in which a given decontamination method must reach a virus that has penetrated beyond the surface

layer of a mask owing to respiration pressure gradients) for treated and positive control masks. In

addition to inoculation of the *de facto* masks or respirators themselves, 100 µl of viral suspension was

pipetted onto one elastic strap per contaminated surgical mask or FFR. The masks were allowed to dry

for 20 minutes and were then individually packaged in appropriate containers (autoclaved empty tip

boxes for respirators; sealable plastic bags for masks) ensuring a cold chain and minimal contamination

before decontamination. Throughout, masks were handled only with sterile tweezers and gloves to limit

bacterial or fungal contamination. Gloves were changed between handling of individual masks to avoid

potential carry-over of inoculum.

Decontamination of PRCV-inoculated surgical masks and filtering facepiece respirators

PRCV-inoculated surgical masks and FFRs were decontaminated utilising three different

decontamination methods, UV irradiation, H_2O_2 vaporisation, and dry heat. Three FFRs and three

surgical masks were subjected to each of the three decontamination methods. Three additional

inoculated masks or respirators (positive controls) and the negative control mask or respirator remained

individually packaged and cooled before and after each decontaminating treatment to account for the

effect of time-dependent surface absorption and/or interaction on virus recovery.

UV germicidal irradiation

Surgical masks and FFRs were individually irradiated using a LS-AT-M1 (LASEA Company, Sart

Tilman, Belgium) equipped with 4 UV-C lamps of 5.5W (@UV-C). Hung vertically on a metal frame,

masks and FFRs were inserted into a safety enclosure. A 2 min UV-C treatment (surgical masks) led to

a fluence of 2.6J/cm² per mask (1.3J/cm² per side), 4 min UV-C irradiation (FFRs) led to a fluence of

5.2J/cm² per mask. Power and irradiation time (120 s) were monitored and recorded throughout.

Following irradiation, surgical masks and FFRs were unloaded and placed in individual bags.

H_2O_2 vaporisation

203 Surgical masks, FFRs and a chemical indicator were placed in individual Mylar/Tyvek pouches.
204 Vaporous hydrogen peroxide (VHP) treatment was performed with the V-PRO Max Sterilizer (Steris,
205 Mentor, OH) which uses 59% liquid H_2O_2 to generate hydrogen peroxide vapor. A 28-minute non lumen
206 cycle was used, consisting of 2 min 40 sec conditioning (5 g/min), 19 min 47 sec decontamination (2.2
207 g/min) and 7 min 46 sec aeration. Peak VHP concentration was 750 ppm.

Dry heat

210 Surgical masks and FFRs, hung horizontally on a metal frame, were inserted into an electrically heated
211 vessel (M-Steryl, AMB Ecosteryl Company, Mons, Belgium) for treatment with temperatures of 102°C
212 (± 4°C) for 60 min (± 15 min). Temperatures inside the heated vessel were recorded throughout to ensure
213 correct exposure conditions. After termination of the treatment cycle, masks and FFRs were allowed to
214 cool and then bagged individually.

Elution of PRCV from decontaminated and untreated surgical masks and filtering facepiece respirators

Upon completed decontamination, the three previously determined square coupons with the focal point
of PRCV inoculation at their centres were cut from the masks; the inoculated elastic strap was severed
from the masks in its entirety (Supplementary Figure 2). Thus, per mask, three coupons and one elastic
strap were sampled. The different layers of the coupons were separated to facilitate viral recovery and
the separated layers of each individual coupon were placed together in a 15 mL Falcon tube containing
4 mL elution medium (Eagle's MEM (Sigma)) supplemented with 2 % of an association of penicillin
(5000 SI units/mL) and streptomycin (5 mg/mL) (PS, Sigma)). Pilot experiments (results not reported)
indicated a cytotoxicity of eluate from H_2O_2-treated masks. To combat this effect, presumably due to
residual H_2O_2 in the suspension, the elution medium was supplemented with 20% FCS and 0.1% β-
mercaptoethanol in a total volume of 4 mL for virus elution subsequent to H_2O_2 vaporisation. Sterile
sets of scissors and tweezers were used for each coupon, thus avoiding cross-contamination between
inoculation points. The mask and respirator coupon layers and elution medium were mixed for 20
minutes at maximum speed (2500 rounds per minute (rpm)) using a multitube vortex mixer (VWR VX-
2500 Multi-Tube Vortexer). Supernatant eluates were recovered via pipette and either directly utilised
in downstream applications or stored at -80°C until further analysis.

Quantification of infectious PRCV eluted from decontaminated and untreated surgical masks and filtering facepiece respirators

Titres of infectious PRCV recovered from individual coupons and straps of decontaminated surgical

masks and respirators were determined separately using a $TCID_{50}$ assay in ST cells. Briefly, tenfold

dilutions were made from each sample and 50µl of each dilution were inoculated in each of four wells

of a 96-well plate. After one hour of incubation, 100µl of medium (MEM supplemented with 10% FCS

and antibiotics) were added to each well. Due to toxicity in the undiluted inoculum, the inoculum was

completely removed and 150µl of medium were added. Four days after inoculation, monolayers were

analysed for the presence of cytopathic effect by light microscopy. Virus titres were calculated using

the Reed and Muench method [23]. Back titrations of virus inoculum stocks were performed in parallel

to each series of decontamination experiments.

Data analysis and statistics

Statistical analyses of differences in infectious viral titres were performed using GraphPad Prism 7

(Graph-Pad Software) and P-values were computed by using a two-sided independent sample t-test,

where ****$P<0.0001$, ***$P<0.001$, **$P<0.01$, *$P<0.05$, and ns is $P \geq 0.05$.

RESULTS

Back titrations of virus inoculums performed in parallel to each series of experiments confirmed PRCV

inoculum titres to be within a range of 6.31×10^6 to 2×10^7 $TCID_{50}$/mL for all experiments.

The cell culture limit of detection (LOD) was 6.31×10^0 $TCID_{50}$/mL for all assays. An initially observed

H_2O_2 cytotoxicity and correspondingly elevated LOD of 6.31×10^1 $TCID_{50}$/mL of H_2O_2-treated coupon

eluates was corrected via β-mercaptoethanol and FCS supplementation of elution medium; elevated

cytotoxicity of H_2O_2-treated strap eluates (SM and FFR) could not be neutralised and remained at

6.31×10^1 $TCID_{50}$/mL. Values below the LOD were considered as $\leq 6.31 \times 10^0$ $TCID_{50}$/mL, with the

exception of measurements concerning H_2O_2-treated straps.

Infectious PRCV is recovered at high titres from untreated surgical mask- and filtering facepiece

respirator coupons, at lower titres from surgical mask straps, and remains under the limit of

detection following recovery from filtering facepiece respirator straps.

Recovery of infectious PRCV from inoculated untreated surgical masks and FFRs was analysed in ST

cells. Comparable high levels of infectious virus were recovered from PRCV-inoculated, untreated left,

right and middle coupons of all surgical masks. Mean values were 2.83×10^5 (± 2.0) $TCID_{50}$/mL, 1.69×10^5

(± 1.81) $TCID_{50}$/mL and 4.25×10^5 (± 2.88) $TCID_{50}$/mL for recovery from positive control coupons of the

UV, H_2O_2 and dry heat assays, respectively. Mean strap recovery values were also similar between

experiments, however they were lower by two to three orders of magnitude than surgical mask coupon

recovery values, with mean values of 5.88×10^3 (± 2.80) $TCID_{50}$/mL, 1.21×10^2 (± 0.71) $TCID_{50}$/mL, and

3.92×10^2 (± 4.56) $TCID_{50}$/mL for straps utilised as positive controls in the UV, H_2O_2 and dry heat assays,

respectively (Figure 1). Recovery from FFR coupons yielded mean infectious virus at 6.96×10^4 (± 13.27)

$TCID_{50}$/mL, 1.80×10^5 (± 1.58) $TCID_{50}$/mL and 2.16×10^4 (± 3.28) $TCID_{50}$/mL for positive controls of the

UV, H_2O_2 and dry heat assays, respectively. Recovery values for infectious virus from FFR straps

remained below the LOD for UV and dry heat positive controls; recovery from straps used as positive

control in the H_2O_2 experiments was 1.75×10^3 (± 0.43) $TCID_{50}$/mL (Figure 1).

UV irradiation, H_2O_2 vaporisation and dry heat treatment reduce infectious PRCV by more than

three orders of magnitude on surgical mask and filtering facepiece respirator coupons and render

it undetectable in all decontamination assays.

Following UV irradiation (2 minutes exposure time), H_2O_2 vaporisation and dry heat treatment of

surgical masks, all titres for virus recovered from coupons and straps remained below the respective

LOD of the assay, showing a total loss of infectivity of more than five orders of magnitude for all three

treatments on coupons (2.83×10^5 (± 2.0) $TCID_{50}$/mL, 1.69×10^5 (± 1.81) $TCID_{50}$/mL and 4.25×10^5 (± 2.88)

$TCID_{50}$/mL, respectively); titres of virus recovered from treated surgical mask straps were reduced by

three orders of magnitude post UV irradiation (5.88×10^3 (± 2.80) $TCID_{50}$/mL), by two orders of

magnitude for heat-treated straps (3.85×10^2 (± 4.56) $TCID_{50}$/mL), and by one order of magnitude for

H_2O_2-treated straps (5.78×10^1 (± 7.10) $TCID_{50}$/mL).

Decontamination treatment effects followed a similar pattern of viral inactivation for FFR coupons

decontaminated via H_2O_2 and dry heat, reducing viral titres by over five orders of magnitude (1.80×10^5

(± 1.58) $TCID_{50}$/mL) and four orders of magnitude (2.16×10^4 (± 3.28) $TCID_{50}$/mL), respectively. While

UV irradiation was sufficient to deactivate PRCV on surgical masks with an exposure time of 2 minutes,

it was shown to be insufficient to achieve viral inactivation by more than three orders of magnitude on

a different lot of FFRs in a trial run (results not shown); a 4 minute exposure time was thus tested for

FFRs, reducing viral titres by over four orders of magnitude, from 6.96×10^4 (± 13.27) $TCID_{50}$/mL to

below the LOD (Figure 2). The impact of decontamination could not be measured for UV- or dry heat-

treated FFR straps due to insufficient recovery of infectious virus in positive FFR strap controls.

Hydrogen peroxide treatment of FFR straps resulted in a reduction of infectious virus load by 1.69×10^3

(± 0.43) $TCID_{50}$/mL.

DISCUSSION

This is, to our knowledge, the first description of stable disinfection of filtering facepiece respirators

and surgical masks contaminated with an infectious SARS-CoV-2 surrogate using UV irradiation, H_2O_2

vaporisation and dry heat treatment. While other reports have described efficacy of various

decontamination methods on a range of biological indicators, few studies report on the validated

decontamination of masks or respirators inoculated with SARS-CoV-2 or a conservative surrogate virus,
the former being limited by the availability of BSL3 facilities, the latter by the lack of a stable high-titre
model virus with an adequate dynamic range to fulfil FDA policy requirements of demonstrating a loss
of infectivity by more than three orders of magnitude [8,16].

PRCV, a SARS-CoV-2 surrogate and fellow member of the *Coronavirinae* subfamily, is classed as a
BSL2 pathogen. It can be cultured to high viral titres in permissive ST cells, thus possessing advantages
of uncomplicated manipulation and utilisation in a BSL2 facility and a wide dynamic range. Here we
demonstrate successful recovery of high quantities of infectious PRCV from inoculated, otherwise
untreated surgical masks and FFR coupons, with recovery titres stably averaging over 10^5 $TCID_{50}$/mL
for elution from mask coupons and ranging between values of over 10^3 to 10^5 $TCID_{50}$/mL for elution
from FFR coupons. Slightly lower recovery values of the dry heat assay are probably attributable to
longer delays between inoculation and elution of infectious virus. Three decontamination methods,
chemical H_2O_2 vaporisation, physical inactivation via UV irradiation and dry heat treatment, were tested
for their ability to inactivate infectious PRCV on inoculated surgical masks or FFRs. All three methods
rendered PRCV inoculated under the outer surface layer of mask and respirator coupons undetectable,
successfully reducing the infectious load by more than three orders of magnitude.

Since carrier surfaces likely influence decontamination efficacy, we aimed to examine viral inactivation
not only on the *de facto* respirators or masks themselves but on their elastic straps that may become
equally contaminated. We compared titres of infectious virus recovered from inoculated, untreated
straps and those inoculated and subsequently decontaminated via either UV irradiation, H_2O_2
vaporisation, or dry heat treatment. All decontamination methods rendered PRCV undetectable
following recovery from straps; however, owing to insufficient virus recovery from untreated mask
straps, only UV decontamination of surgical mask straps could be successfully validated as reducing
viral loads by more than three orders of magnitude. Recovery of infectious virus applied to FFR straps
proved impossible with a simple elution medium; however, when 20% FCS and 0.1% β-
mercaptoethanol were added to the elution medium (intended to combat H_2O_2 cytotoxicity), infectious
virus was recoverable from FFR straps. Further studies are planned to elucidate these effects, which may
potentially be associated either to inherent virucidal properties of the elastic materials or attributable to
poor elution from the straps. The fact that supplemented medium enabled recovery of infectious virus
from FFR straps suggests that either of the supplemented constituents could have had a protective effect
shielding infectious virus from potential virucidal impacts of the straps.

CONCLUSIONS
In conclusion, we describe successful validation of three decontamination methods, UV irradiation,
H_2O_2 vaporisation and dry heat treatment, in inactivating an infectious coronavirus in line with the FDA

policy regarding face masks and respirators. Without enough proof of inactivation, we cannot recommend safe decontamination of respirator straps and suggest additionally treating straps separately by exposure to 70% ethanol [20] until further results become available. The PRCV surrogate supplements existing data regarding decontamination of surgical masks and FFRs, and both it and the different decontamination methods tested, are easily adaptable to other respirator and mask types, presenting a useful conservative model for stable validation of coronavirus decontamination.

ACKNOWLEDGEMENTS

The authors express their sincere gratitude to Amélie Matton and Frédéric de Meulemeester (AMB Ecosteryl, Mons, Belgium), Axel Kupisiewicz (LASEA, Sart-Tilman, Belgium), Pierre Leonard (Solwalfin, Belgium) for suggestions and technical and administrative support and thank Chantal Vanmaercke and Carine Boone for their excellent technical support.

CONFLICTS OF INTEREST STATEMENT

The authors have no conflicts of interest to disclose.

FUNDING SOURCE

This work was supported by a grant from the Walloon Region, Belgium (Project 2010053 -2020- "MASK - Decontamination and reuse of surgical masks and filtering facepiece respirators").

363   REFERENCES

365   [1]   WHO. Strengthening the health system response to COVID-19 Recommendations for the
366   WHO European Region Policy brief 2020:8.

367   [2]   World Health Organization (WHO). Rational use of personal protective equipment for
368   coronavirus disease 2019 (COVID-19). Who 2020;2019:1–7.

369   [3]   Leung NHL, Chu DKW, Shiu EYC, Chan K-H, McDevitt JJ, Hau BJP, et al. Respiratory virus
370   shedding in exhaled breath and efficacy of face masks. Nat Med 2020.
371   https://doi.org/10.1038/s41591-020-0843-2.

372   [4]   Lee J, Bong C, Bae PK, Abafogi AT, Baek SH. Fast and easy disinfection of coronavirus-
373   contaminated face masks using ozone gas produced by a dielectric barrier discharge plasma
374   generator. MedRxiv 2020;COVID-19 S:1–13.

375   [5]   Dutch National Institute for Public Health and the Environment (RIVM). Reuse of FFP2 masks
376   2020:1–5.

377   [6]   Center for Devices and Radiological Health. Enforcement Policy for Face Masks and
378   Respirators During the Coronavirus Disease (COVID-19) Public Health Emergency (Revised
379   ) Guidance for Industry and Food and Drug Administration Staff 2020.

380   [7]   Zonta W, Mauroy A, Farnir F, Thiry E. Virucidal Efficacy of a Hydrogen Peroxide
381   Nebulization Against Murine Norovirus and Feline Calicivirus, Two Surrogates of Human
382   Norovirus. Food Environ Virol 2016;8:275–82. https://doi.org/10.1007/s12560-016-9253-5.

383   [8]   Kumar A, Kasloff SB, Leung A, Cutts T, Strong JE, Hills K, et al. N95 Mask Decontamination
384   using Standard Hospital Sterilization Technologies. MedRxiv Prepr 2020:1–9.
385   https://doi.org/https://doi.org/10.1101/2020.04.05.20049346.

386   [9]   Schwartz A, Stiegel M, Greeson N, Vogel A, Thomann W, Brown M, et al. Decontamination
387   and Reuse of N95 Respirators with Hydrogen Peroxide Vapor to Address Worldwide Personal
388   Protective Equipment Shortages During the SARS-CoV-2 (COVID-19) Pandemic. Appl Biosaf
389   2020. https://doi.org/10.1177/1535676020919932.

390   [10]   Lelie PN, Reesink HW, Lucas CJ. Inactivation of 12 viruses by heating steps applied during
391   manufacture of a hepatitis B vaccine. J Med Virol 1987;23:297–301.
392   https://doi.org/10.1002/jmv.1890230313.

393   [11]   Darnell MER, Subbarao K, Feinstone SM, Taylor DR. Inactivation of the coronavirus that
394   induces severe acute respiratory syndrome, SARS-CoV. J Virol Methods 2004;121:85–91.
395   https://doi.org/10.1016/j.jviromet.2004.06.006.

396   [12]   Chin AWH, Chu JTS, Perera MRA, Hui KPY, Yen H-L, Chan MCW, et al. Stability of SARS-
397   CoV-2 in different environmental conditions. The Lancet Microbe 2020;1:e10.
398   https://doi.org/10.1016/s2666-5247(20)30003-3.

[13]   Liao L, Xiao W, Zhao M, Yu X, Wang H, Wang Q, et al. Can N95 respirators be reused after

disinfection? And for how many times? MedRxiv 2020:2020.04.01.20050443.

https://doi.org/10.1101/2020.04.01.20050443.

[14] Lore MB, Heimbuch BK, Brown TL, Wander JD, Hinrichs SH. Effectiveness of three

decontamination treatments against influenza virus applied to filtering facepiece respirators.

Ann Occup Hyg 2012;56:92–101. https://doi.org/10.1093/annhyg/mer054.

[15] Viscusi DJ, Bergman MS, Novak DA, Faulkner KA, Palmiero A, Powell J, et al. Impact of

three biological decontamination methods on filtering facepiece respirator fit, odor, comfort,

and donning ease. J Occup Environ Hyg 2011;8:426–36.

https://doi.org/10.1080/15459624.2011.585927.

[16] Wigginton KR, Arts PJ, Clack H, Fitzsimmons WJ, Gamba M, Katherine R, et al. Validation of

N95 filtering facepiece respirator decontamination methods available at a large university

hospital. MedRxiv Prepr 2020. https://doi.org/https://doi.org/10.1101/2020.04.28.20084038.

[17] Pensaert M, Callebaut P, Vergote J. Isolation of a porcine respiratory, non-enteric coronavirus

related to transmissible gastroenteritis. Vet Q 1986;8:257–61.

https://doi.org/10.1080/01652176.1986.9694050.

[18] Laude H, Van Reeth K, Pensaert M. Porcine respiratory coronavirus: molecular features and

virus-host interactions. Vet Res 1993;24:125–50.

[19] Saif LJ, Wang Q, Vlasova AN, Jung K, Xiao S. Coronaviruses. In: Zimmerman JJ, Karriker

LA, Ramirez A, Schwartz KJ, Stevenson GW, Jianqiang Z, editors. Dis. Swine, Elev. Ed.,

2019, p. 488–523.

[20] Kampf G, Todt D, Pfaender S, Steinmann E. Persistence of coronaviruses on inanimate

surfaces and their inactivation with biocidal agents. J Hosp Infect 2020;104:246–51.

https://doi.org/10.1016/j.jhin.2020.01.022.

[21] Casanova LM, Jeon S, Rutala WA, Weber DJ, Sobsey MD. Effects of air temperature and

relative humidity on coronavirus survival on surfaces. Appl Environ Microbiol 2010;76:2712–

7. https://doi.org/10.1128/AEM.02291-09.

[22] Cox E, Hooyberghs J, Pensaert MB. Sites of replication of a porcine respiratory coronavirus

related to transmissible gastroenteritis virus. Res Vet Sci Sci 1990;48:165–169.

[23] Reed LJ, Muench H. A simple method of estimating fifty per cent endpoints. Am J Hyg

1938;27:493–7. https://doi.org/10.1093/oxfordjournals.aje.a118408.

FIGURES

Figure 1

Figure 2

430   LEGENDS
431

Figure 1. Recovery of virus after elution from inoculated, untreated surgical masks and filtering
facepiece respirators. Recovery of infectious porcine respiratory coronavirus from inoculated untreated
surgical masks (SM) and filtering facepiece respirators (FFR) was analysed in swine testicular cells. The
cell culture limit of detection (LOD) was 0.8 $\log_{10}$ $TCID_{50}$/mL (6.31×10^0). Similar levels of virus
recovery were detected for left, right and middle (L, R, M) coupons of masks and respirators; recovery
efficacy of infectious virus from straps (S) deviated significantly in all analyses from the mean of all
coupons and remained below the LOD for all assays performed on FFR straps. P-values were computed
by using a two-sided independent sample t-test to calculate differences between individual coupon
values and differences between mean values of all coupons and straps, where ****P<0.0001,
***P<0.001, **P<0.01, *P<0.05, and ns is P≥0.05.

Figure 2. Effect of three decontaminating treatments on PRCV-inoculated surgical mask- and filtering
facepiece respirator coupons and straps. The infectivity of porcine respiratory coronavirus (PRCV)
recovered from surgical masks (SM) and filtering facepiece respirators (FFR) decontaminated via
exposure to ultraviolet light (UV), hydrogen peroxide vaporisation (H_2O_2), or dry heat treatment was
analysed in swine testicular cells. The cell culture limit of detection (LOD) was 0.8 $\log_{10}$ $TCID_{50}$/ml for
all analyses except those concerning H_2O_2-treated SM or FFR straps (1.8 $\log_{10}$ $TCID_{50}$/ml). Per
decontamination method, nine PRCV-inoculated, decontaminated coupons (n=9) and three inoculated,
decontaminated straps (n=3) were analysed in parallel to inoculated, untreated, positive control coupons
(n=9) and straps (n=3). Sample size deviated for UV-decontaminated FFR respirators, where n=6
coupons (left and right) were analysed. P-values were computed by using a two-sided independent
sample t-test, where ****P<0.0001, ***P<0.001, **P<0.01, *P<0.05, and ns is P≥0.05.

SUPPLEMENTARY FIGURES

Supplementary Figure 1

Supplementary Figure 2

 SUPPLEMENTARY FIGURE LEGENDS

Supplementary Figure 1. Demarcation of the three coupons (black squares), future PRCV inoculation

sites (red dots), and marked elastic strap are shown on a representative surgical mask (top) and filtering

facepiece respirator (bottom).

Supplementary Figure 2. Three square coupons with a focal point of PRCV inoculation at their centres

are cut from a surgical mask (top); to facilitate viral elution, different layers of the coupons are separated

(middle); an inoculated elastic strap is severed from a surgical mask (bottom).