TITLE

The use of germicidal ultraviolet light, vaporised hydrogen peroxide and dry heat to decontaminate

face masks and filtering respirators contaminated with a SARS-CoV-2 surrogate virus

RUNNING TITLE

Coronavirus mask and FFR decontamination

AUTHORS

Louisa F. Ludwig-Begall[1,*], Constance Wielick[1,*], Lorène Dams[1], Hans Nauwynck[2], Pierre-Francois

Demeuldre[3], Aurore Napp[3], Jan Laperre[4], Eric Haubruge[5], Etienne Thiry[1,#]

AUTHOR AFFILIATIONS

[1]Veterinary Virology and Animal Viral Diseases, Department of Infectious and Parasitic Diseases,

FARAH Research Centre, Faculty of Veterinary Medicine, Liège University, Liège, Belgium

[2]Laboratory of Virology, Faculty of Veterinary Medicine, Ghent University, Merelbeke, Belgium

[3] Department of Hospital Pharmacy, The University Hospital Center, University of Liège, Liège,

Belgium

[4] Centexbel Textile Research Centre, Grace-Hollogne, Belgium

[5] TERRA Research Centre, Gembloux Agro-Bio Tech, University of Liège, Gembloux, Belgium

[*]These authors contributed equally to this work.

[#]Corresponding author: etienne.thiry@uliege.be

Orcid ID Etienne Thiry: 0000-0002-9585-7933

WORD COUNT

3944   (excluding   Summary   and   References)   +   two   figures

Non-standard abbreviations:
FFR: filtering facepiece respirator
SM: surgical mask
SARS-CoV-2: severe acute respiratory syndrome coronavirus 2
PRCV: porcine respiratory coronavirus

ABSTRACT

Background

In the context of the ongoing severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) pandemic, the supply of personal protective equipment remains under severe strain. To address this issue, re-use of surgical face masks and filtering facepiece respirators has been recommended; prior decontamination is paramount to their re-use.

Aim

We aim to provide information on the effects of three decontamination procedures on porcine respiratory coronavirus (PRCV)-contaminated masks and respirators, presenting a stable model for infectious coronavirus decontamination of these typically single-use-only products.

Methods

Surgical masks and filtering facepiece respirator coupons and straps were inoculated with infectious PRCV and submitted to three decontamination treatments, UV irradiation, vaporised H_2O_2, and dry heat treatment. Viruses were recovered from sample materials and viral titres were measured in swine testicle cells.

Findings

UV irradiation, vaporised H_2O_2 and dry heat reduced infectious PRCV by more than three orders of magnitude on mask and respirator coupons and rendered it undetectable in all decontamination assays.

Conclusion

This is the first description of stable disinfection of face masks and filtering facepiece respirators contaminated with an infectious SARS-CoV-2 surrogate using UV irradiation, vaporised H_2O_2 and dry heat treatment. The three methods permit demonstration of a loss of infectivity by more than three orders of magnitude of an infectious coronavirus in line with the FDA policy regarding face masks and respirators. It presents advantages of uncomplicated manipulation and utilisation in a BSL2 facility, therefore being easily adaptable to other respirator and mask types.

KEYWORDS

SARS-CoV-2; PRCV; decontamination (UV; H_2O_2; dry heat); respirator; surgical mask

INTRODUCTION

In the context of the ongoing severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2)

pandemic, the supply of personal protective equipment (PPE) remains under severe strain and both

availability and affordability of items can be subject to fluctuations and disruptions within health-,

social care and other essential public facilities. Access to PPE for the health workforce in all services

(public and private, community and hospital) has been identified as a key factor in strengthening the

international health system response to SARS-CoV-2, the causative agent of COVID-19 [1]; use of

medical/surgical face masks and filtering facepiece respirators (FFR i.e. N95 or FFP2 or FFP3

standard or equivalent with a minimum filtration efficiency of 95% for 0.3 µm (aerodynamic mass

mean diameter) of sodium chloride aerosols) has been recommended in conjunction with other

mitigating measures to prevent transmission of this and other respiratory pathogens [2,3].

Since the surging global demand cannot currently be met solely by the limited capacities of expanding

PPE production, national and internationally-coordinated efforts are increasingly focused on re-use of

various items [4,5]. While prior decontamination is paramount to re-use of surgical masks or

respirators, little information exists on effective decontamination of these typically single-use-only

products. Without compromising the fit and filtration integrity of the masks themselves,

decontaminating procedures must guarantee the complete inactivation of contaminating pathogens; the

US Food and Drug Administration Enforcement Policy for Face Masks and Respirators issued in April

2020 recommends a robust proof of infectious bioburden reduction of three orders of magnitude for

viral pathogens, specifically coronaviruses, and six orders of magnitude for either mycobacteria or

bacterial spores [6].

Amongst a variety of different methods under investigation, vaporisation of the oxidizing agent

hydrogen peroxide (H_2O_2), already a standard hospital sterilisation technology due to its broad

antimicrobial activity and efficacy in surface decontamination [7], has garnered attention as a cost-

effective and practical option for mask decontamination [5,6,8]. Mask shape and fit of unused 3M

FFP2 NR D face masks (type 8822) were shown to remain intact subsequent to two cycles of

vaporised H_2O_2 low pressure gas sterilisation in a study by the Dutch National Institute for Public

Health and the Environment [5]; further studies have since demonstrated integrity of other respirator

models after numerous cycles of H_2O_2 treatment and have shown virucidal and bactericidal activity of

the method on a number of respiratory pathogens and/or biological indicators [8,9]. To our knowledge,

no study has yet reported the effect of H_2O_2 treatment on SARS-CoV-2, although vaporised H_2O_2

decontamination trials of SARS-CoV-2 inoculated masks are reportedly underway [8].

As alternatives to chemical vaporised H_2O_2, two physical decontamination methods, the application of

dry heat and ultraviolet (UV) irradiation, show promise for decontamination of SARS-CoV-2

contaminated masks in various settings. Heat treatment acts via denaturation of protein secondary

structures thereby altering conformation of viral proteins involved in attachment to and replication

within host cells and has long been recognised as an efficient method of virus inactivation [10].

Temperatures of over 65°C have previously been shown to inactivate SARS-CoV in suspension [11];

more recently, dry heat treatment of 70°C was identified to inactivate SARS-CoV-2 in solution [12]

and was shown to not significantly alter filtration efficiency of N95 respirators within 20 cycles of

application [13]. Easily scalable, dry heat allows mass treatment of large sample sizes and thus

potentially presents a fast and efficient decontamination alternative to vaporised H_2O_2 in decentralised

hospital centres or industrial settings. Ultraviolet germicidal irradiation with a highly energetic short-

wave (254 nm) acts via viral disruption of viral DNA or RNA and constitutes a physical surface

treatment for contaminated masks or FFRs. A useful sterilisation technique in a variety of applications,

UV irradiation has been implemented to effectively decontaminate influenzavirus (H5N1) from two

different models of FFRs (3M models 1860s and 1870) [14] and does not degrade respirator

performance even after multiple applications [13,15]. It is a promising option for rapid

decontamination of smaller, individual sample contingents and is as such easily adaptable to point-of-

care applications. A possible concern relates to UV penetration depth, necessitating studies that not

only investigate viral inactivation of decontaminated surfaces but that address a potential worst-case

scenario in which viruses penetrate deeper layers of contaminated face masks and FFRs [13].

Since the utilisation, concentration and cultivation of infectious SARS-CoV-2 necessary for analyses

investigating its inactivation, pose obvious problems in terms of the availability and equipping of

BSL3 facilities, the use of conservative surrogates to test decontamination efficacy of various methods

is justified and crucial to gain an as accurate as possible insight into SARS-CoV-2 decontamination. A

recent (non-peer reviewed) publication describes inactivation of murine hepatitis virus (MHV), a

SARS-CoV-2 surrogate of the same *Betacoronavirus* genus, via vaporised H_2O_2, heat treatment, UV

exposure and other decontamination methods on FFRs; regrettably, the limited dynamic range of the

MHV model (at most 10^1-10^2 inactivation) fell short of demonstrating a 10^3 reduction [16].

In the present investigation into filtering facepiece respirator and surgical mask decontamination via

UV irradiation, H_2O_2 and application of dry heat, we implemented porcine respiratory coronavirus

(PRCV), a spike gene deletion mutant of transmissible gastroenteritis virus (TGEV) and a member of

the *Alphacoronavirus 1* species [17,18], as SARS-CoV-2 surrogate. While PRCV, which infects the

respiratory tract of swine [19], is not in the same genus as SARS-CoV-2, the two members of the

subfamily *Coronavirinae* in the family *Coronaviridae* show sufficient similarities as to genome length

and virion structure (notwithstanding differences in envelope glycoproteins), for them to be expected

to behave similarly outside their hosts. Indeed, TGEV has previously been utilised as a surrogate for

another *Betacoronavirus*, SARS-CoV, in studies investigating the persistence of coronaviruses on

inanimate surfaces and/or their inactivation with biocidal agents [20,21]. The PRCV model thus

combines the advantages of sufficient genetic and structural relatedness to SARS-CoV-2, the fact that

PRCV can be readily propagated and assayed *in vitro* to high titres in the swine testicle (ST) cell line,

and the absence of human infection risk.

This is the first description of stable disinfection of FFRs and surgical masks contaminated with an

infectious SARS-CoV-2 surrogate using UV irradiation, vaporised H_2O_2 and dry heat treatment. The

three methods permit demonstration of a loss of infectivity by more than three orders of magnitude of

an infectious coronavirus in line with the FDA policy regarding face masks and respirators. It presents

advantages of uncomplicated manipulation and utilisation in a BSL2 facility, therefore being easily

adaptable to other respirator and mask types to which any of the three or other decontamination

methods may be applied.

METHODS

Virus and cells

PRCV strain 91V44 [22] was passaged three times on confluent cell monolayers of the ST continuous

cell line. A virus stock with a titre of $10^{7.8}$ $TCID_{50}$/ml was used.

Surgical masks and filtering facepiece respirators

All FFRs and surgical masks, commonly used by the health care community in Belgium at the time of

the study, were supplied by the Department of the Hospital Pharmacy, University Hospital Centre of

Liege (Sart-Tilman). Manufacturers (and models): KN95 FFR - Guangzhou Sunjoy Auto Supplies CO.

LTD, Guangdong, China (2020 N°26202002240270); surgical mask (Type II) - Hangzhou Sunten

Textile Co., LTD, Hangzhou, China (SuninCare™, Protect Plus). Surgical masks and FFRs were

verified to be from the same respective manufacturing lot to minimise any lot-to-lot variation and to

ensure consistency during future respirability and filtration performance testing. FFR and mask

materials were verified via scanning electron microscopy and Fourier-transform infrared spectroscopy

(spectral data obtained by direct µ-ATR- FTIR analysis (Nicolet 6700) and compared for best fit

against a Hummel library of know infrared spectra of polymers as well as a Centexbel-curated spectral

library). FFRs consisted of four layers of polypropylene, specifically two outer spunbound structures,

an intermediate meltblown layer, and an inner spunbound layer (Supplementary Figure 3). Surgical

masks were composed of three layers of polypropylene, with an outer and inner layer of spunbound

polypropylene encasing a meltblown polypropylene barrier (Supplementary Figure 4).

Inoculation of surgical masks and filtering facepiece respirators with porcine respiratory
coronavirus (PRCV)

Efficacy of three different decontamination methods in inactivating an infectious coronavirus was
assessed using surgical masks and FFRs experimentally inoculated with PRCV. Per decontamination
method and mask type, one negative control mask or respirator (not contaminated but treated), three
treated masks or respirators (PRCV-contaminated and treated), and three positive controls (PRCV-
contaminated but untreated), i.e. seven masks in total, were utilised.

Prior to inoculation, the masks and FFRs were first marked with a graphite pencil to enable sample
identification and to outline three square areas (34 mm x 34 mm) to the left, right and centre of the
masks and FFRs, corresponding to areas to be cut out post-inoculation (coupons) with a central site of
inoculation (the demarcation of the coupons and inoculation sites are shown in Supplementary Figure
1). Utilising an ultra-fine insulin syringe and needle (BD Medical), 100 µl of undiluted viral
suspension in Minimum Essential Medium (MEM) were injected at the centre of each of the three
square coupons under the first outer layer of mask or FFR (this to simulate a "worst case scenario" for
viral inactivation in which a given decontamination method must reach a virus that has penetrated
beyond the surface layer of a mask owing to respiration pressure gradients) for treated and positive
control masks. In addition to inoculation of the *de facto* masks or respirators themselves, 100 µl of
viral suspension was pipetted onto one elastic strap per contaminated surgical mask or FFR. The
masks were allowed to dry for 20 minutes and were then individually packaged in appropriate
containers (autoclaved empty tip boxes for respirators; sealable plastic bags for masks) ensuring a cold
chain and minimal contamination before decontamination. Throughout, masks were handled only with
sterile tweezers and gloves to limit bacterial or fungal contamination. Gloves were changed between
handling of individual masks to avoid potential carry-over of inoculum.

Decontamination of PRCV-inoculated surgical masks and filtering facepiece respirators

PRCV-inoculated surgical masks and FFRs were decontaminated utilising three different
decontamination methods, UV irradiation, vaporised H_2O_2, and dry heat. Three FFRs and three surgical
masks were subjected to each of the three decontamination methods. Three additional inoculated
masks or respirators (positive controls) and the negative control mask or respirator remained
individually packaged and cooled before and after each decontaminating treatment to account for the
effect of time-dependent surface absorption and/or interaction on virus recovery.

UV germicidal irradiation

Surgical masks and FFRs were individually irradiated using a LS-AT-M1 (LASEA Company, Sart

Tilman, Belgium) equipped with 4 UV-C lamps of 5.5W (@UV-C). Hung vertically on a metal frame,

masks and FFRs were inserted into a safety enclosure. A 2 min UV-C treatment (surgical masks) led

to a fluence of 2.6J/cm² per mask (1.3J/cm² per side), 4 min UV-C irradiation (FFRs) led to a fluence

of 5.2J/cm² per mask. Power and irradiation time (120 s) were monitored and recorded throughout.

Following irradiation, surgical masks and FFRs were unloaded and placed in individual bags.

209   Vaporised H_2O_2

210   Surgical masks, FFRs and a chemical indicator were placed in individual Mylar/Tyvek pouches.

211   Vaporous hydrogen peroxide (VHP) treatment was performed with the V-PRO Max Sterilizer (Steris,

212   Mentor, OH) which uses 59% liquid H_2O_2 to generate hydrogen peroxide vapor. A 28-minute non

213   lumen cycle was used, consisting of 2 min 40 sec conditioning (5 g/min), 19 min 47 sec

214   decontamination (2.2 g/min) and 7 min 46 sec aeration. Peak VHP concentration was 750 ppm.

216   Dry heat

217   Surgical masks and FFRs, hung horizontally on a metal frame, were inserted into an electrically heated

218   vessel (M-Steryl, AMB Ecosteryl Company, Mons, Belgium) for treatment with temperatures of

219   102°C ($\pm$ 4°C) for 60 min ($\pm$ 15 min). Temperatures inside the heated vessel were recorded throughout

220   to ensure correct exposure conditions. After termination of the treatment cycle, masks and FFRs were

221   allowed to cool and then bagged individually.

223   Elution of PRCV from decontaminated and untreated surgical masks and filtering facepiece
224   respirators

Upon completed decontamination, the three previously determined square coupons with the focal

point of PRCV inoculation at their centres were cut from the masks; the inoculated elastic strap was

severed from the masks in its entirety (Supplementary Figure 2). Thus, per mask, three coupons and

one elastic strap were sampled. The different layers of the coupons were separated to facilitate viral

recovery and the separated layers of each individual coupon were placed together in a 15 mL Falcon

tube containing 4 mL elution medium (Eagle's MEM (Sigma)) supplemented with 2 % of an

association of penicillin (5000 SI units/mL) and streptomycin (5 mg/mL) (PS, Sigma)). Pilot

experiments (results not reported) indicated a cytotoxicity of eluate from H_2O_2-treated masks. To

combat this effect, presumably due to residual H_2O_2 in the suspension, the elution medium was

supplemented with 20% FCS and 0.1% β-mercaptoethanol in a total volume of 4 mL for virus elution

subsequent to vaporised H_2O_2 treatment. Sterile sets of scissors and tweezers were used for each

coupon, thus avoiding cross-contamination between inoculation points. The mask and respirator
coupon layers and elution medium were mixed for 20 minutes at maximum speed (2500 rounds per
minute (rpm)) using a multitube vortex mixer (VWR VX-2500 Multi-Tube Vortexer). Supernatant
eluates were recovered via pipette and either directly utilised in downstream applications or stored at -
80°C until further analysis.

Quantification of infectious PRCV eluted from decontaminated and untreated surgical masks
and filtering facepiece respirators

Titres of infectious PRCV recovered from individual coupons and straps of decontaminated surgical
masks and respirators were determined separately using a $TCID_{50}$ assay in ST cells. Briefly, tenfold
dilutions were made from each sample and 50µl of each dilution were inoculated in each of four wells
of a 96-well plate. After one hour of incubation, 100µl of medium (MEM supplemented with 10%
FCS and antibiotics) were added to each well. Due to toxicity in the undiluted inoculum, the inoculum
was completely removed and 150µl of medium were added. Four days after inoculation, monolayers
were analysed for the presence of cytopathic effect by light microscopy. Virus titres were calculated
using the Reed and Muench method [23]. Back titrations of virus inoculum stocks were performed in
parallel to each series of decontamination experiments.

Data analysis and statistics
Statistical analyses of differences in infectious viral titres were performed using GraphPad Prism 7
(Graph-Pad Software) and P-values were computed by using a two-sided independent sample t-test,
where ****$P<0.0001$, ***$P<0.001$, **$P<0.01$, *$P<0.05$, and ns is $P\geq0.05$.

RESULTS
Back titrations of virus inoculums performed in parallel to each series of experiments confirmed
PRCV inoculum titres to be within a range of 6.31×10^6 to 2×10^7 $TCID_{50}$/mL for all experiments.

The cell culture limit of detection (LOD) was 6.31×10^0 $TCID_{50}$/mL for all assays. An initially
observed H_2O_2 cytotoxicity and correspondingly elevated LOD of 6.31×10^1 $TCID_{50}$/mL of H_2O_2-
treated coupon eluates was corrected via β-mercaptoethanol and FCS supplementation of elution
medium; elevated cytotoxicity of H_2O_2-treated strap eluates (SM and FFR) could not be neutralised
and remained at 6.31×10^1 $TCID_{50}$/mL. Values below the LOD were considered as $\leq6.31\times10^0$
$TCID_{50}$/mL, with the exception of measurements concerning H_2O_2-treated straps.

Infectious PRCV is recovered at high titres from untreated surgical mask- and filtering
facepiece respirator coupons, at lower titres from surgical mask straps, and remains under the
limit of detection following recovery from filtering facepiece respirator straps.

Recovery of infectious PRCV from inoculated untreated surgical masks and FFRs was analysed in ST
cells. Comparable high levels of infectious virus were recovered from PRCV-inoculated, untreated
left, right and middle coupons of all surgical masks. Mean values were 2.83×10^5 (± 2.0) $TCID_{50}$/mL,
1.69×10^5 (± 1.81) $TCID_{50}$/mL and 4.25×10^5 (± 2.88) $TCID_{50}$/mL for recovery from positive control
coupons of the UV, H_2O_2 and dry heat assays, respectively. Mean strap recovery values were also
similar between experiments, however they were lower by two to three orders of magnitude than
surgical mask coupon recovery values, with mean values of 5.88×10^3 (± 2.80) $TCID_{50}$/mL, 1.21×10^2
(± 0.71) $TCID_{50}$/mL, and 3.92×10^2 (± 4.56) $TCID_{50}$/mL for straps utilised as positive controls in the
UV, H_2O_2 and dry heat assays, respectively (Figure 1). Recovery from FFR coupons yielded mean
infectious virus at 6.96×10^4 (± 13.27) $TCID_{50}$/mL, 1.80×10^5 (± 1.58) $TCID_{50}$/mL and 2.16×10^4 (± 3.28)
$TCID_{50}$/mL for positive controls of the UV, H_2O_2 and dry heat assays, respectively. Recovery values
for infectious virus from FFR straps remained below the LOD for UV and dry heat positive controls;
recovery from straps used as positive control in the H_2O_2 experiments was 1.75×10^3 (± 0.43)
$TCID_{50}$/mL (Figure 1).

UV irradiation, vaporised H_2O_2 and dry heat treatment reduce infectious PRCV by more than
three orders of magnitude on surgical mask and filtering facepiece respirator coupons and
render it undetectable in all decontamination assays.

Following UV irradiation (2 minutes exposure time), exposure to vaporised H_2O_2, and dry heat
treatment of surgical masks, all titres for virus recovered from coupons and straps remained below the
respective LOD of the assay, showing a total loss of infectivity of more than five orders of magnitude
for all three treatments on coupons (2.83×10^5 (± 2.0) $TCID_{50}$/mL, 1.69×10^5 (± 1.81) $TCID_{50}$/mL and
4.25×10^5 (± 2.88) $TCID_{50}$/mL, respectively); titres of virus recovered from treated surgical mask straps
were reduced by three orders of magnitude post UV irradiation (5.88×10^3 (± 2.80) $TCID_{50}$/mL), by two
orders of magnitude for heat-treated straps (3.85×10^2 (± 4.56) $TCID_{50}$/mL), and by one order of
magnitude for H_2O_2-treated straps (5.78×10^1 (± 7.10) $TCID_{50}$/mL).

Decontamination treatment effects followed a similar pattern of viral inactivation for FFR coupons
decontaminated via H_2O_2 and dry heat, reducing viral titres by over five orders of magnitude
(1.80×10^5 (± 1.58) $TCID_{50}$/mL) and four orders of magnitude (2.16×10^4 (± 3.28) $TCID_{50}$/mL,
respectively. While UV irradiation was sufficient to deactivate PRCV on surgical masks with an
exposure time of 2 minutes, it was shown to be insufficient to achieve viral inactivation by more than
three orders of magnitude on a different lot of FFRs in a trial run (results not shown); a 4 minute
exposure time was thus tested for FFRs, reducing viral titres by over four orders of magnitude, from

6.96×10^4 (± 13.27) $TCID_{50}$/mL to below the LOD (Figure 2). The impact of decontamination could not

be measured for UV- or dry heat-treated FFR straps due to insufficient recovery of infectious virus in

positive FFR strap controls. Hydrogen peroxide treatment of FFR straps resulted in a reduction of

infectious virus load by 1.69×10^3 (± 0.43) $TCID_{50}$/mL.

DISCUSSION

This is, to our knowledge, the first description of stable disinfection of filtering facepiece respirators

and surgical masks contaminated with an infectious SARS-CoV-2 surrogate using UV irradiation,

vaporised H_2O_2, and dry heat treatment. While other reports have described efficacy of various

decontamination methods on a range of biological indicators, few studies report on the validated

decontamination of masks or respirators inoculated with SARS-CoV-2 or a conservative surrogate

virus, the former being limited by the availability of BSL3 facilities, the latter by the lack of a stable

high-titre model virus with an adequate dynamic range to fulfil FDA policy requirements of

demonstrating a loss of infectivity by more than three orders of magnitude [8,16].

PRCV, a SARS-CoV-2 surrogate and fellow member of the *Coronavirinae* subfamily, is classed as a

BSL2 pathogen. It can be cultured to high viral titres in permissive ST cells, thus possessing

advantages of uncomplicated manipulation and utilisation in a BSL2 facility and a wide dynamic

range. Here we demonstrate successful recovery of high quantities of infectious PRCV from

inoculated, otherwise untreated surgical masks and FFR coupons, with recovery titres stably averaging

over 10^5 $TCID_{50}$/mL for elution from mask coupons and ranging between values of over 10^3 to 10^5

$TCID_{50}$/mL for elution from FFR coupons. Slightly lower recovery values of the dry heat assay are

probably attributable to longer delays between inoculation and elution of infectious virus. Three

decontamination methods, chemical vaporised H_2O_2, physical inactivation via UV irradiation and dry

heat treatment, were tested for their ability to inactivate infectious PRCV on inoculated surgical masks

or FFRs. All three methods rendered PRCV inoculated under the outer surface layer of mask and

respirator coupons undetectable, successfully reducing the infectious load by more than three orders of

magnitude.

Since carrier surfaces likely influence decontamination efficacy, we aimed to examine viral

inactivation not only on the *de facto* respirators or masks themselves but on their elastic straps that

may become equally contaminated. We compared titres of infectious virus recovered from inoculated,

untreated straps and those inoculated and subsequently decontaminated via either UV irradiation,

vaporised H_2O_2, or dry heat treatment. All decontamination methods rendered PRCV undetectable

following recovery from straps; however, owing to insufficient virus recovery from untreated mask

straps, only UV decontamination of surgical mask straps could be successfully validated as reducing

viral loads by more than three orders of magnitude. Recovery of infectious virus applied to FFR straps

proved impossible with a simple elution medium; however, when 20% FCS and 0.1% β-
mercaptoethanol were added to the elution medium (intended to combat H_2O_2 cytotoxicity), infectious
virus was recoverable from FFR straps. Further studies are planned to elucidate these effects, which
may potentially be associated either to inherent virucidal properties of the elastic materials or
attributable to poor elution from the straps. The fact that supplemented medium enabled recovery of
infectious virus from FFR straps suggests that either of the supplemented constituents could have had
a protective effect shielding infectious virus from potential virucidal impacts of the straps. It is worth
noting that, although the simple MEM matrix used to inoculate and elute virus loads has similarities to
the natural biocontamination arising from use (i.e. build-up of inorganic salts), a complex
biocontamination (e.g. polymer chains in sputum) may not be replicated.

Decontaminating treatments, by their very nature, are known to have inherently detrimental side
effects; particularly after multiple cycles, the integrity of decontaminated objects may be
compromised. UV irradiation, vaporised H_2O_2, or dry heat treatment, have previously been shown to
not significantly impact performance of polypropylene-based FFRs and/or masks in a number of
studies [8,9,13,15]; however, others have shown that the maximum number of decontamination cycles
may be limited by the respirator model and treatment conditions required for inactivation [24]. In the
present investigation, masks and FFRs were effectively destroyed at the end of each cycle. Since, the
safe reuse of masks and FFRs is important both in the context of the current Covid-19 pandemic and
beyond, when cost-effectiveness, environmental benefits and logistic considerations will advocate a
continued decontamination of these previously single-use items, further work is planned to investigate
how many decontamination cycles may be safely applied to these previously single use items.

CONCLUSIONS
In conclusion, we describe successful validation of three decontamination methods, UV irradiation,
vaporised H_2O_2 and dry heat treatment, in inactivating an infectious coronavirus in line with the FDA
policy regarding face masks and respirators. Without enough proof of inactivation, we cannot
recommend safe decontamination of respirator straps and suggest additionally treating straps
separately by exposure to 70% ethanol [20] until further results become available. Since H_2O_2 breaks
down to water and oxygen, concerns for toxicity of vaporised H_2O_2 treatment are generally held to be
very low risk [25]; however, to eliminate potential exposure of users, we recommend that a short
aeration time should be respected. The PRCV surrogate supplements existing data regarding
decontamination of surgical masks and FFRs, and both it and the different decontamination methods
tested, are easily adaptable to other respirator and mask types, presenting a useful conservative model
for stable validation of coronavirus decontamination.

ACKNOWLEDGEMENTS

The authors express their sincere gratitude to Amélie Matton and Frédéric de Meulemeester (AMB

Ecosteryl, Mons, Belgium), Axel Kupisiewicz (LASEA, Sart-Tilman, Belgium), Pierre Leonard

(Solwalfin, Belgium) for suggestions and technical and administrative support and thank Chantal

Vanmaercke and Carine Boone for their excellent technical support.

CONFLICTS OF INTEREST STATEMENT

The authors have no conflicts of interest to disclose.

FUNDING SOURCE

This work was supported by a grant from the Walloon Region, Belgium (Project 2010053 -2020-

"MASK - Decontamination and reuse of surgical masks and filtering facepiece respirators").

REFERENCES

[1]   WHO. Strengthening the health system response to COVID-19 Recommendations for the

394   WHO European Region Policy brief 2020:8.

395   [2]   World Health Organization (WHO). Rational use of personal protective equipment for

396   coronavirus disease 2019 (COVID-19). Who 2020;2019:1–7.

397   [3]   Leung NHL, Chu DKW, Shiu EYC, Chan K-H, McDevitt JJ, Hau BJP, et al. Respiratory virus

398   shedding in exhaled breath and efficacy of face masks. Nat Med 2020.

399   https://doi.org/10.1038/s41591-020-0843-2.

400   [4]   Lee J, Bong C, Bae PK, Abafogi AT, Baek SH. Fast and easy disinfection of coronavirus-

401   contaminated face masks using ozone gas produced by a dielectric barrier discharge plasma

402   generator. MedRxiv 2020;COVID-19 S:1–13.

403   [5]   Dutch National Institute for Public Health and the Environment (RIVM). Reuse of FFP2 masks

404   2020:1–5.

405   [6]   Center for Devices and Radiological Health. Enforcement Policy for Face Masks and

406   Respirators During the Coronavirus Disease (COVID-19) Public Health Emergency (Revised)

407   Guidance for Industry and Food and Drug Administration Staff 2020.

408   [7]   Zonta W, Mauroy A, Farnir F, Thiry E. Virucidal Efficacy of a Hydrogen Peroxide

409   Nebulization Against Murine Norovirus and Feline Calicivirus, Two Surrogates of Human

410   Norovirus. Food Environ Virol 2016;8:275–82. https://doi.org/10.1007/s12560-016-9253-5.

411   [8]   Kumar A, Kasloff SB, Leung A, Cutts T, Strong JE, Hills K, et al. N95 Mask Decontamination

412   using Standard Hospital Sterilization Technologies. MedRxiv Prepr 2020:1–9.

413   https://doi.org/https://doi.org/10.1101/2020.04.05.20049346.

414   [9]   Schwartz A, Stiegel M, Greeson N, Vogel A, Thomann W, Brown M, et al. Decontamination

415   and Reuse of N95 Respirators with Hydrogen Peroxide Vapor to Address Worldwide Personal

416   Protective Equipment Shortages During the SARS-CoV-2 (COVID-19) Pandemic. Appl Biosaf

417   2020. https://doi.org/10.1177/1535676020919932.

418   [10]   Lelie PN, Reesink HW, Lucas CJ. Inactivation of 12 viruses by heating steps applied during

419   manufacture of a hepatitis B vaccine. J Med Virol 1987;23:297–301.

420   https://doi.org/10.1002/jmv.1890230313.

421   [11]   Darnell MER, Subbarao K, Feinstone SM, Taylor DR. Inactivation of the coronavirus that

422   induces severe acute respiratory syndrome, SARS-CoV. J Virol Methods 2004;121:85–91.

423   https://doi.org/10.1016/j.jviromet.2004.06.006.

424   [12]   Chin AWH, Chu JTS, Perera MRA, Hui KPY, Yen H-L, Chan MCW, et al. Stability of SARS-

425   CoV-2 in different environmental conditions. The Lancet Microbe 2020;1:e10.

426   https://doi.org/10.1016/s2666-5247(20)30003-3.

[13]   Liao L, Xiao W, Zhao M, Yu X, Wang H, Wang Q, et al. Can N95 respirators be reused after

disinfection? And for how many times? MedRxiv 2020:2020.04.01.20050443.

https://doi.org/10.1101/2020.04.01.20050443.

[14] Lore MB, Heimbuch BK, Brown TL, Wander JD, Hinrichs SH. Effectiveness of three

decontamination treatments against influenza virus applied to filtering facepiece respirators.

Ann Occup Hyg 2012;56:92–101. https://doi.org/10.1093/annhyg/mer054.

[15] Viscusi DJ, Bergman MS, Novak DA, Faulkner KA, Palmiero A, Powell J, et al. Impact of

three biological decontamination methods on filtering facepiece respirator fit, odor, comfort,

and donning ease. J Occup Environ Hyg 2011;8:426–36.

https://doi.org/10.1080/15459624.2011.585927.

[16] Wigginton KR, Arts PJ, Clack H, Fitzsimmons WJ, Gamba M, Katherine R, et al. Validation of

N95 filtering facepiece respirator decontamination methods available at a large university

hospital. MedRxiv Prepr 2020. https://doi.org/https://doi.org/10.1101/2020.04.28.20084038.

[17] Pensaert M, Callebaut P, Vergote J. Isolation of a porcine respiratory, non-enteric coronavirus

related to transmissible gastroenteritis. Vet Q 1986;8:257–61.

https://doi.org/10.1080/01652176.1986.9694050.

[18] Laude H, Van Reeth K, Pensaert M. Porcine respiratory coronavirus: molecular features and

virus-host interactions. Vet Res 1993;24:125–50.

[19] Saif LJ, Wang Q, Vlasova AN, Jung K, Xiao S. Coronaviruses. In: Zimmerman JJ, Karriker

LA, Ramirez A, Schwartz KJ, Stevenson GW, Jianqiang Z, editors. Dis. Swine, Elev. Ed.,

2019, p. 488–523.

[20] Kampf G, Todt D, Pfaender S, Steinmann E. Persistence of coronaviruses on inanimate

surfaces and their inactivation with biocidal agents. J Hosp Infect 2020;104:246–51.

https://doi.org/10.1016/j.jhin.2020.01.022.

[21] Casanova LM, Jeon S, Rutala WA, Weber DJ, Sobsey MD. Effects of air temperature and

relative humidity on coronavirus survival on surfaces. Appl Environ Microbiol 2010;76:2712–

7. https://doi.org/10.1128/AEM.02291-09.

[22] Cox E, Hooyberghs J, Pensaert MB. Sites of replication of a porcine respiratory coronavirus

related to transmissible gastroenteritis virus. Res Vet Sci Sci 1990;48:165–169.

[23] Reed LJ, Muench H. A simple method of estimating fifty per cent endpoints. Am J Hyg

1938;27:493–7. https://doi.org/10.1093/oxfordjournals.aje.a118408.

[24] Lindsley WG, Martin SB, Thewlis RE, Sarkisian K, Nwoko JO, Mead KR, et al. Effects of

Ultraviolet Germicidal Irradiation (UVGI) on N95 Respirator Filtration Performance and

Structural Integrity. J Occup Environ Hyg 2015;12:509–17.

https://doi.org/10.1080/15459624.2015.1018518.

[25] Martin S, Duncan E. Sterilisation considerations for implantable sensor systems. Woodhead

Publishing Limited; 2013. https://doi.org/10.1533/9780857096289.2.253.

FIGURES

Figure 1

Figure 2

LEGENDS

Figure 1. Recovery of virus after elution from inoculated, untreated surgical masks and filtering

facepiece respirators. Recovery of infectious porcine respiratory coronavirus from inoculated untreated

surgical masks (SM) and filtering facepiece respirators (FFR) was analysed in swine testicular cells.

The cell culture limit of detection (LOD) was 0.8 $\log_{10}$ $TCID_{50}$/mL (6.31×10^0). Similar levels of virus

recovery were detected for left, right and middle (L, R, M) coupons of masks and respirators; recovery

efficacy of infectious virus from straps (S) deviated significantly in all analyses from the mean of all

coupons and remained below the LOD for all assays performed on FFR straps. P-values were

computed by using a two-sided independent sample t-test to calculate differences between individual

coupon values and differences between mean values of all coupons and straps, where ****$P<0.0001$,

***$P<0.001$, **$P<0.01$, *$P<0.05$, and ns is $P \geq 0.05$.

Figure 2. Effect of three decontaminating treatments on PRCV-inoculated surgical mask- and filtering

facepiece respirator coupons and straps. The infectivity of porcine respiratory coronavirus (PRCV)

recovered from surgical masks (SM) and filtering facepiece respirators (FFR) decontaminated via

exposure to ultraviolet light (UV), vaporised hydrogen peroxide (H_2O_2), or dry heat treatment was

analysed in swine testicular cells. The cell culture limit of detection (LOD) was 0.8 $\log_{10}$ $TCID_{50}$/ml

for all analyses except those concerning H_2O_2-treated SM or FFR straps (1.8 $\log_{10}$ $TCID_{50}$/ml). Per

decontamination method, nine PRCV-inoculated, decontaminated coupons (n=9) and three inoculated,

decontaminated straps (n=3) were analysed in parallel to inoculated, untreated, positive control

coupons (n=9) and straps (n=3). Sample size deviated for UV-decontaminated FFR respirators, where

n=6 coupons (left and right) were analysed. P-values were computed by using a two-sided independent

sample t-test, where ****$P<0.0001$, ***$P<0.001$, **$P<0.01$, *$P<0.05$, and ns is $P \geq 0.05$.

SUPPLEMENTARY FIGURES

Supplementary Figure 1

Supplementary Figure 2

Supplementary Figure 3

Supplementary Figure 4

SUPPLEMENTARY FIGURE LEGENDS

Supplementary Figure 1. Demarcation of the three coupons (black squares), future PRCV inoculation

sites (red dots), and marked elastic strap are shown on a representative surgical mask (top) and

filtering facepiece respirator (bottom).

Supplementary Figure 2. Three square coupons with a focal point of PRCV inoculation at their centres

are cut from a surgical mask (top); to facilitate viral elution, different layers of the coupons are

separated (middle); an inoculated elastic strap is severed from a surgical mask (bottom).

Supplementary Figure 3

Scanning electron microscopy of a KN95 filtering facepiece respirator (FFR). Far left: Cross section

view of the FFR. The two top layers are polypropylene spunbound structures, the third layer from the

top is a polypropylene meltblown layer, the fourth layer is a polypropylene spunbound layer facing the

wearer (the inner layer). Group of four: Details of all four polypropylene layers of the FFR. The

outermost spunbound layer, the second spunbound layer, the intermediate meltblown layer, and the

inner spunbound layer (61.4, 27.5, 24.4, and 28. 8 grams/m^2 from outer to inner layer) are shown in

clockwise direction. The dots in the spunbound layers show points where the fibres are melted

together to ensure layer integrity. Pictures were taken on a JEOL JSM-7600F Analytical Ultrahigh

resolution thermally assisted field-emission gun source scanning electron microscope at 40 times

magnification. Prior to taking the pictures, samples were coated with a platinum/palladium layer

(suited to non-conductive samples) using a JEOL JFC-2300 High Resolution sputter coater.

Supplementary Figure 4

Scanning electron microscopy of a SuninCare™, Protect Plus surgical mask. Top: Cross section view

of the surgical mask. The outer layer (blue) is a polypropylene spunbound structure, the middle layer

is a polypropylene meltblown structure (26.8 grams/m2) and the inner layer is a polypropylene

spunbound structure. Bottom: The outer, intermediate and inner surgical mask layers are depicted

from left to right. The dots in the outer and inner spunbound layers show points where the

polypropylene fibres are melted together to ensure layer integrity. Pictures were taken on a JEOL

JSM-7600F Analytical Ultrahigh resolution thermally assisted field-emission gun source scanning

electron microscope at 40 times magnification. Prior to taking the pictures, samples were coated with a

platinum/palladium layer (suited to non-conductive samples) using a JEOL JFC-2300 High Resolution

sputter coater.