Serological Assays Estimate Highly Variable SARS-CoV-2 Neutralizing Antibody Activity in Recovered

COVID19 Patients

Larry L. Luchsinger,[a,#] Brett Ransegnola,[a] Daniel Jin,[a] Frauke Muecksch,[b] Yiska Weisblum,[b] Weili Bao,[d]

Parakkal Jovvian George,[e] Marilis Rodriguez,[f] Nancy Tricoche,[e] Fabian Schmidt,[b] Chengjie Gao,[g] Shabnam

Jawahar,[e] Mouli Pal,[d] Emily Schnall,[e] Huan Zhang,[g] Donna Strauss,[h] Karina Yazdanbakhsh,[d] Christopher D.

Hillyer,[a,h] Paul D. Bieniasz,[b,c,#] Theodora Hatziioannou,[b,#]

[a.] Laboratory of Stem Cell Regenerative Research, Lindsley F. Kimball Research Institute, New York
Blood Center, New York, NY 10065, USA.
[b.] Laboratory of Retrovirology, The Rockefeller University, New York, NY 10065, USA
[c.] Howard Hughes Medical Institute, New York, NY 10016, USA.
[d.] Laboratory of Complement Biology, Lindsley F. Kimball Research Institute, New York Blood
Center, New York, NY 10065, USA.
[e.] Laboratory of Molecular Parasitology Lindsley F. Kimball Research Institute, New York Blood
Center, New York, NY 10065, USA.
[f.] Laboratory of Blood-Borne Parasites, Lindsley F. Kimball Research Institute, New York Blood
Center, New York, NY 10065, USA.
[g.] Laboratory of Membrane Biology, Lindsley F. Kimball Research Institute, New York Blood Center,
New York, NY 10065, USA.
[h.] New York Blood Center Enterprises, New York, NY 10065, USA.
[i.] Equal authorship
[#] Co-corresponding authors
24
25 Running Head: Covid19 Serology Result Predict Neutralizing Activity
26
27 # Address correspondence to Larry Luchsinger (Lead): lluchsinger@nybc.org

Abstract

The development of neutralizing antibodies (nAb) against SARS-CoV-2, following infection or vaccination, is likely to be critical for the development of sufficient population immunity to drive cessation of the COVID19 pandemic. A large number of serologic tests, platforms and methodologies are being employed to determine seroprevalence in populations to select convalescent plasmas for therapeutic trials, and to guide policies about reopening. However, tests have substantial variability in sensitivity and specificity, and their ability to quantitatively predict levels of nAb is unknown. We collected 370 unique donors enrolled in the New York Blood Center Convalescent Plasma Program between April and May of 2020. We measured levels of antibodies in convalescent plasma using commercially available SARS-CoV-2 detection tests and in-house ELISA assays and correlated serological measurements with nAb activity measured using pseudotyped virus particles, which offer the most informative assessment of antiviral activity of patient sera against viral infection. Our data show that a large proportion of convalescent plasma samples have modest antibody levels and that commercially available tests have varying degrees of accuracy in predicting nAb activity. We found the Ortho Anti-SARS-CoV-2 Total Ig and IgG high throughput serological assays (HTSAs), as well as the Abbott SARS-CoV-2 IgG assay, quantify levels of antibodies that strongly correlate with nAb assays and are consistent with gold-standard ELISA assay results. These findings provide immediate clinical relevance to serology results that can be equated to nAb activity and could serve as a valuable 'roadmap' to guide the choice and interpretation of serological tests for SARS-CoV-2.

Introduction

In late 2019, a cluster of patients in Wuhan, the capital city of China's Hubei providence, were reported to be afflicted with a severe respiratory illness of unknown origin.(1, 2) Patients presented with symptoms that included high fever, pneumonia, dyspnea, and respiratory failure. The causative agent was identified to be severe acute respiratory syndrome coronavirus variant 2 (SARS-CoV-2), the 7th coronavirus strain to infect humans to date,(3) and the clinical syndrome was designated coronavirus disease of 2019 (COVID19). The pathogenesis of COVID19 is similar to previously documented respiratory distress syndromes caused by related coronaviruses, including the 2005 SARS coronavirus (SARS-CoV) and the middle east respiratory syndrome coronavirus (MERS).(4) However, the greater transmissibility of SARS-CoV-2 has enabled a swift global spread that has resulted in substantial mortality. Detection and tracking SARS-CoV-2 spread has been difficult. Moreover, the spectrum of symptomatology observed in SARS-CoV-2 infection is wide, ranging from asymptomatic and mild, reminiscent of numerous seasonal infections, including influenza and common cold viruses, all the way to life-threatening respiratory failure that requires intensive care and invasive ventilation. Currently, increased age and comorbidities are the factors most highly predictive of severe of COVID19 disease.(5)

The utility of serological tests to identify individuals who have acquired antibodies against SARS-CoV-2 is thus recognized as both an indication of the seroprevalence of SARS-CoV-2 infection and, potentially, of immunity afforded to the seropositive individual.(3, 6-8) Seroconversion is determined by detection of antibodies that recognize SARS-CoV-2 antigens. Coronaviruses have 4 major structural proteins: spike (S) protein (including the S1 protein and receptor binding domain (RBD)), nucleocapsid (N) protein, membrane (M) protein and envelope (E) protein.(9) Previous studies of SARS-CoV and MERS found the most immunogenic antigens are the S- and N-proteins,(10) and development of serological tests for SARS-CoV-2 antibodies has focused heavily on these viral proteins.

Three major platforms of serological testing have been adopted; 1) enzyme linked immunosorbent assays (ELISA), 2) high-throughput serological assays (HTSA), and 3) lateral flow assays (LFA). ELISAs offer wide flexibility for research laboratories to select virtually any antigen of interest and provide highly

sensitive, quantitative results. HTSAs are more suitable for clinical laboratories and offer limited antigen

diversity but allow high-throughput and sensitive, semi-quantitative results. LFAs also offer limited

antigen diversity, but function with small volumes (~20µL) of whole blood, plasma or sera and allow rapid

(≤15 minutes) results at the point of care. The clinical community will undoubtedly employ multiple

SARS-CoV-2 serology platforms but a comparative analysis across platforms has not been undertaken.

Further, it is currently unknown whether the detection of antibodies that bind these proteins predicts

neutralizing activity or protection against infection.(11)

Convalescent plasma (CP) transfusion has been recognized as a potential treatment for critically ill

COVID19 patients and the New York Blood Center (NYBC) has led the first COVID19 CP donation

program in the United States. Using 370 unique CP donor samples deposited in our COVID19 Research

Repository (https://nybc.org/covid19repository), we conducted ELISA, HSTA and LFA assays as well as

SARS-CoV-2 pseudovirus neutralization assays. We find that CP donors have a wide range of antibody

titers measured across multiple COVID19 serological and neutralization assays. Notably, we show that

some HTSA and ELISA assays predict neutralizing activity *in vitro* and may thus serve to predict antiviral

activity against SARS-CoV-2 *in vivo*.

Results

Characteristics of the NYC CP Donor Population

Serological analysis of the CP donors was performed using 370 unique samples collected between

April and May of 2020 from the NYC area. CP donors enrolled in the program were required to have tested

positive for SARS-CoV-2 by PCR diagnostic tests and be symptom free for at least 2 weeks. To profile CP

donors, we cross-referenced donor demographic data to the 2010 U.S. Census database.(12) CP donors had

a median age of 41 years (95% CI: 39-44, range 17-75 years,) and showed a gaussian distribution (n=183,

r^2=0.89) compared to the national median age of 38.2 years in 2018 (**Figure 1A**). The frequency of male

and female CP donors was 45.2% and 54.8%, respectively, and was not statistically different from the

national average of 49.2% and 50.8% (**Figure 1B**). The frequency of ABORh blood group antigens was

also largely consistent with the national frequency, with a slightly higher number of A⁻ and O⁻ donors and

slightly lower number of AB$^+$ and B$^+$ donors than expected (**Figure 1C**). Finally, CP donor ethnicity was

largely consistent with the national ethnic composition, with a slightly higher number of multiracial/other

donors and lower number of Black/African American donors than expected (**Figure 1D**). Overall, the

composition of NYC CP donors analyzed was reflective of the United States population demographic.

Neutralizing Activity of the CP Donor Population

Neutralization assays measure how effectively donor plasma or serum can inhibit virus infection of

target cells and are the gold standard for measuring the antiviral activity of antibodies. In the case of SARS-

CoV-2, such assays require in biosafety level 3 (BSL-3) facilities and highly trained personnel. To

overcome this limitation and expedite testing, we employed pseudotyped virus assays based on either HIV-

1 (human immunodeficiency virus type 1) or VSV (vesicular stomatitis virus). Both viruses were

engineered to lack their own envelope glycoproteins and to express a luciferase reporter gene.

Complementation *in trans* with the SARS-CoV-2 Spike (S) protein results in the generation of pseudotyped

virus particles that are dependent on the interaction between the S protein and its receptor ACE2

(angiotensin-converting enzyme 2) for entry into cells.(13) These reporter viruses were used to measure

infection of human cells engineered to express ACE2 (HIV-S assay) or expressed endogenous ACE2

(VSV-S assay) and to determine the ability of plasma dilutions to inhibit S-dependent virus entry. The NT_{50}

values, reflecting the plasma dilution at which virus infection is reduced by 50%, were calculated for each

sample (**Supplementary Figure 1A**).

The neutralizing activity of CP donor samples was extremely variable and NT_{50} values obtained

ranged from <50 to over 20,000. The median NT_{50} values were 390.1 (95% CI: 278.3-499.7) and 450.6

(95% CI: 367.7-538.4) for the HIV-S or VSV-S assays, respectively (**Figure 2A**) and the two assays

showed a high degree of correlation (**Supplementary Figure 1B-C**). Fresh frozen plasma (FFP) samples

donated in 2019, before the SARS-CoV-2 outbreak, were used as negative controls (n=10). Importantly, the

NT_{50} values of all FFP samples were ≤50, which is the highest concentration of plasma used in the

neutralization assays and is hence designated as the signal cutoff (S/co) value. Overall, 83.1% and 92.7% of

the CP donor samples had detectable neutralization activity using HIV-S and VSV-S assays, respectively

(**Figure 2B**). Notably, 11.2% and 8.7% of CP donors had NT_{50} values at or greater than 2000 (40-fold over

S/co) using HIV-S and VSV-S, assays respectively while 55.8% and 52% of CP donors had NT_{50} values at

or less than 500 (10-fold over S/co) (**Figure 2B**). Thus, the majority of CP donors may have relatively

modest neutralizing activity and a small proportion of donors have high neutralization activity.

NT_{50} values were not statistically different between blood groups (**Figure 2E, Supplementary**

Figure 1G) or age groups (**Figure 2C, Supplementary Figure 1E**) and there was no linear correlation of

NT_{50} values with age (**Supplementary Figure 1D**) in contrast to previous reports.[14] However, in

agreement with recent studies,[15] NT_{50} values of male CP donor samples were ~1.7-fold higher than those

from female CP donors using HIV-S and VSV-S assays (**Figure 2D and Supplementary Figure 1F,** n =

195, p = 0.009 and <0.001, median difference 217 and 197, respectively). For CP donors where symptom

dates were reported, the time between last symptom and the date of donation was calculated. Interestingly,

CP donors 2-3 weeks post symptoms had a statistically significant increase in NT_{50} values compared to CP

donors >3 weeks post-symptom (**Figure 2F and Supplementary Figure 1H,** n=52, p = 0.03 and 0.04,

median difference 426 and 226, respectively). Overall, these data suggest CP donors possess a wide range

of neutralizing antibody levels that are proportionately distributed across demographic categories with the

exception of a small sex-dependent effect.

Serological Test Results of the CP Donor Population

Multiple platforms have been deployed to detect seroconversion against SARS-CoV-2. The

simplest tests are LFAs, which solubilize antibodies from whole blood, plasma or sera in an aqueous

mobile phase which moves across a nitrocellulose membrane coated with anti-human IgG and/or IgM to

distinguish between specific classes of immunoglobins while a control band ensures test function. Binding

of antibodies to antigen-conjugated enzyme, such as horseradish peroxidase, generates a colored band at

the test lines. Analysis of 144 CP donor samples showed that only 79.4% of CP donors tested positive for

SARS-CoV-2-specfic IgG antibodies and 24.8% for IgM antibodies (**Figure 3A, top**). While LFAs are not

designed to perform quantitatively, large discrepancies in band intensity between donors (**Supplementary**

Figure 2A) is often presumed to indicate semi-quantitative results. We performed densitometric analysis of

the test bands from LFA cassettes (**Supplementary Figure 2B, 2C**) and normalized each test to control

band intensity. LFAs showed an intensity range of 0% - 99.2% for IgG bands and 0% - 18.5% for IgM

bands, with a median intensity of 20% for IgG and <1% for IgM (**Figure 3A, bottom**). Thus, LFAs have a

high degree of variation in band intensity within the CP donor population.

HTSA systems offer the advantage of performing semi-quantitative seroconversion assays using

clinical laboratory testing infrastructure at large scale. We performed the Ortho-Clinical Diagnostics

VITROS SARS-CoV-2 total Ig assay, the VITROS SARS-CoV-2 IgG assay and the Abbott Labs Architect

SARS-CoV-2 IgG assay using between 100 and 330 CP donor plasma samples. We found 96.4% and

91.0% of CP donor samples were positive using the Ortho total Ig and IgG assays, respectively, and 91.4%

were positive using the Abbott IgG assay (**Figure 3B**). The median value of CP samples using the Ortho

total Ig assay was 101 arbitrary units (AU) (n=333, 95% CI: 78.5 – 123, S/co = 1, range 0 to 1000 AU)

while that of FFP healthy controls was 0.01 AU (n=8, 95% CI: 0.01 – 0.02). Similarly, the median value of

CP samples using the Ortho IgG assay was 11.7 AU (n=100, 95% CI: 8.3 – 16.07, S/co = 1, range 0 to ~30

AU). For the Abbott assay, the median value of CP samples was 6.04 AU (n=315, 95% CI: 5.48 – 6.44,

S/co = 1.4, range 0 to ~10 AU) while that of FFP healthy controls was 0.02 AU (95% CI: 0.01 – 0.15).

These results clearly show HTSA platforms detect a wide variation in antibody levels in the CP donor

population and offer greater dynamic range than LFA assays.

The gold standard for quantification of antigen-specific antibodies is ELISA assays. Studies of

antibody responses during SARS-CoV and MERS outbreaks identified the S- and N-proteins as the

dominant antigens. Therefore, we designed three indirect ELISA assays using SARS-CoV-2 recombinant,

His-tagged, spike protein S1 domain (S1), spike protein RBD domain (RBD) and nucleocapsid protein (N).

We utilized monoclonal antibodies demonstrated to bind antigen in a dose-dependent manner to generate

standard curves from which antibody concentrations were calculated and FFP from healthy controls to

determine signal cutoffs. Thus, we report our ELISA results as monoclonal antibody (mAb) titers. These

ELISA assays showed that 85.2%, 89.1%, and 96.3% of CP donor samples were positive for antibodies

against S1, RBD and N antigens, respectively (**Figure 3C**). Using the S1 ELISA, the median value for CP

donor samples was 445μg/mL (n=285, 95% CI: 342 – 536μg/mL, S/co = 120μg/mL) and for FFP controls

100.9μg/mL (n=10, 95% CI: 78 – 120μg/mL). In the NP ELISA the median value for CP donor samples

was 6432μg/mL (n=271, 95% CI: 2811 – 13792μg/mL, S/co = 700μg/mL) while in the RBD ELISA the

median value of CP donor samples was 15.6μg/mL (n=43, 95% CI: 12.55 – 25.6μg/mL, S/co = 4μg/mL).

Notably, the range of S1 and NP-binding antibody concentrations observed in the ELISAs was extreme,

constituting a 1,000-fold difference in titers within the CP donor population. Taken together, these data

demonstrate that CP donors have a wide range of concentrations of antibodies specific to immunogenic

SARS-CoV-2 antigens, as measured across multiple serological platforms.

Correlation of Serology Tests with Neutralizing Activity

It is not logistically feasible to implement neutralization assays as a measurement of antiviral

antibodies at a scale of the general population. While quantification of seroconversion is practiced,

controlled studies that determine the relationship between quantitative SARS-CoV-2 serology test results

and neutralizing activity is sparse. We examined the correlation between serology and neutralization assays

in the CP donor samples (**Figure 4A, Supplementary Figure 3A, Supplementary Figure 4C**). As

expected, S1 ELISA titers showed a positive linear regression with NT_{50} values ($r^2 = 0.35$) while the RBD

ELISA titers showed slightly higher linearity ($r^2 = 0.38$), commensurate with the fact that the RBD is a key

target for neutralizing antibodies. Conversely, NP ELISA titers showed a comparatively lower degree of

linear regression with neutralization activity ($r^2 = 0.09$). By comparison, both the Ortho HTSA total Ig

assay and the IgG assay showed higher ($r^2 = 0.45$ for both) while the Abbott HTSA IgG assay showed

lower linear regression with neutralization activity ($r^2 = 0.24$). Although Ortho HTSAs and the Abbott

HTSA IgG platforms quantify antibodies against S1 and NP antigens, respectively, a linear regression of

$r^2 = 0.33$ was calculated between these two HTSAs (**Supplementary Figure 3B**). As expected, linear

regression between the Ortho total Ig and IgG assay was strong ($r^2 = 0.72$) since the two assays measure the

same epitope. LFA IgG densitometry measurements showed the poorest correlation with neutralization

activity ($r^2 = 0.22$).

Correlation between serological results and neutralization activity was also examined using the non-

parametric Spearman test that does not assume linear dependence (**Figure 4B**). As expected, a high

correlation between the HIV-S and VSV-S neutralization assays was obtained (r=0.89). The Ortho and

Abbott HTSA platforms exhibited the highest degree of correlation with neutralization among the serology

assays tested (r = 0.75 and 0.72, respectively for the HIV-S assay; 0.70 and 0.69 for the VSV-S assay). The

S1, RBD, and NP ELISAs also showed a high degree of correlation, particularly with the HIV-S

neutralization assay (r = 0.69, 0.65, and 0.65) while the LFA IgG and IgM assay showed the poorest

correlation (r = 0.56, 0.41). Taken together, the data demonstrate that all quantitative serological assays

correlate to some degree with neutralization activity. However, HTSA and S1 ELISA assays that measure

anti-spike protein antigens have the highest predictive value as a surrogate for pseudovirus neutralization

assays. Importantly, correlation between HTSA scores and NT_{50} values suggest presumptive ranges of

neutralizing activity based on ranges of HTSA values (**Figure 4C, Supplementary Figure 4A**).

While ELISA assays revealed S1 and N antibody titers correlated with each other, these titers were

not always proportional among CP donor samples. To examine the coincidence of S1 and NP antibody

titers and using FFP plasma samples as negative controls, we categorized S1 and N antibody titers that fell

below S/co values as 'negative' and titers greater 10-fold over S/co as 'high' (**Supplementary Figure 4B**).

Using 241 CP donor samples that were assayed with both the S1 and N ELISA assays, we found that 81%

of donors were double positive (DP), while 16% of samples were single positive (14% N and 2% S1,

respectively) (**Figure 4D**). Only 2.5% of CP donors were double negative for S1 and NP antibodies. Within

the double positive population, we found that 23% of samples were DP^{high} while 5% and 30% of samples

were only $S1^{high}$ or N^{high} and the remaining 42% were DP^{low}. We then examined the distribution of NT_{50}

values from the HIV-S neutralization assay within these populations (**Figure 4E**). Notably, DN samples

showed NT_{50} values at the S/co observed for FFP healthy control samples while DP^{low} samples had

relatively low NT_{50} values (median value 327, 95% CI: 186 – 444). Importantly, the DP^{high} donors had

NT_{50} values that were 7-fold higher than DP^{low} donors (median value 2130). Additionally, NT_{50} values in

the N^{high} and $S1^{high}$ groups were 2.5- and 4-fold higher than those of the DP^{low} group.

Finally, we sought to determine if the frequency of peripheral blood immune cells varied as a

function of antibody titer. We stained peripheral blood mononuclear cells (PBMCs) isolated from CP donor

buffy coats for classical surface markers associated with B-cell or T-cell populations (**Supplementary**

Figure 5A, 5B). We examined T cell subsets including T central memory (CD45RO+CD62L+) and T

effector memory (TEM; CD45RO$^+$CD62L^{neg}) while B cell (CD20$^+$) subsets analyzed included memory B

cells (CD27$^+$CD24$^+$), plasmablasts (CD24^{neg}CD38^{hi}CD138^{neg}) and the more mature plasma cells

(CD24^{neg}CD38^{hi}CD138$^+$) (**Supplementary Figure 5C**). We found statistically significant differences in

naïve CD4 and CD8 T-cell populations in donors with high S1 ELISA titers compared to those with low

titer. Decrease in CD24$^+$CD27$^+$ memory B cells was detected in individuals with higher anti-S titers.

Although the cause of this lower frequency is not known, it could raise the possibility that individuals with

reduced memory B cells may develop a less robust antibody response with future infections. Although our

phenotypic analysis of B and T cell compartments was limited, these data suggest phenotypic differences in

canonical B and T cell populations are insufficient to explain the large differences in antibody titers or

neutralization activity observed in CP donors and warrants future studies designed to study B and T cell

function from individual donors.

Discussion

Demographic limitations of the CP donor population

Recent studies have noted a disproportion in COVID19 morbidity and mortality among minority

communities.(16) In this study, of the 370 CP samples analyzed, only 204 donors (55%) elected to identify

ethnicity, representing the least reported demographic category we collected. Nevertheless, we did not

observe a significant difference in nAb or serology results as a function of any demographic metric,

including ethnicity. Although we showed that the CP donor samples analyzed in this study comprised a

relatively normal distribution of demographic indicators, based on the U.S. census data, we acknowledge

that some factors, including ethnicity, are underrepresented in this cohort and limit the interpretation of the

study beyond the population aggregate. The potential explanations of this phenomenon are complex and

extend beyond the scope of this study.(17) The blood banking community is continuously working to

recruit minority donors, who are consistently underrepresented amongst regular blood donors.(18) Efforts

to increase public participation in local blood and CP donor programs would both improve blood product

diversity of transfusion products and strengthen the rigor of epidemiological disease. Thus, studies

designed to characterize serological responses to COVID19 specifically in minority groups are warranted

and necessary to augment our current understanding of the pandemic.

Seroconversion assays of the population

Quantification of antiviral antibodies in recovered individuals is an important metric for

determining population immunity conferred by exposure to SARS-CoV-2. Our study suggests that most

New York City convalescent plasma donors have antibodies against SARS-CoV-2. Indeed, our data

demonstrate that the HTSA, including Ortho and Abbot assays, which have received EUA from the FDA,

are well suited to quantify a wide range of antibody titers and reported that 91 – 96.4% of the CP

population possesses detectable SARS-CoV-2 antibodies. LFAs performed less well, and individuals with

low antibody titers scored weakly positive or negative in LFAs. Such outcomes could be interpreted

incorrectly, thus increasing the rate of false negative results. Ultimately, studies that accurately document

SARS-CoV-2 seroprevalence in diverse populations will require highly sensitive, high quality assays such

as HSTA or ELISA to be reliable.

Correlation between serological assay measurements and neutralizing activity

Since patient recovery often precedes the development of efficacious and safe therapeutics, a

longstanding treatment strategy for infectious diseases is passive antibody transfer. Therefore, refining

strategies to improve CP infusion efficacy benefits both the current treatment options of COVID-19 and

will inform the medical community for future pandemics. Our serological analyses are consistent with

previous publications that show a considerable range in antibody titers in recovered COVID19 patients.(19)

However, this study provides a comprehensive analysis of the correlation of quantitative serological test

values with neutralization activity. Importantly, high dynamic range serological assays, such as the HTSA

and S1 ELISA, had a significant linear correlation with neutralization activity. We show, for the first time,

the extent to which three widely available SARS-CoV-2 HTSAs correlated to nAb activity as well as to

each other, providing the clinical and scientific community with a comprehensive overview of clinical serology test performance. To this end, investigators from the Mayo Clinic's COVID-19 Expanded Access Program (EAP) performed an exploratory analysis on the efficacy of CP as a therapeutic agent using data from over 35,000 transfusions.(20) Although the study showed uncertainty as to the statistical significance of effect, the authors noted that patients transfused with high antibody titer CP units, quantified by the Ortho IgG assay, showed a notable reduction in the odds ratio of mortality at both 7 and 30 days after transfusion. These data support the assertion that antibody quantification of CP units using high dynamic range HSTA assays may further improve therapeutic options for COVID-19 and, perhaps, future pandemic responses. This knowledge will also be necessary for deriving potential serologic correlates of protection,(21) and may aid in predicting immunity at the individual and population levels.(15)

Yet, the levels of plasma neutralizing activity required to prevent SARS-CoV-2 re-infection are currently unknown. Anecdotal results have been reported for seasonal coronavirus experimental infection studies. For example, one study of 229E HCoV found a positive correlation between pre-infection antibody titer and neutralization activity with symptom clinical severity.(22) In another study, 7 of 8 individuals with low neutralizing titers excreted virus upon re-exposure, compared to only 1 of 4 subjects with higher titers.(23) However, the conclusions of these studies are not directly comparable to the current SARS-CoV-2 pandemic. As such, the necessity of human epidemiological or vaccination studies are necessary to determine the minimum threshold of neutralizing activity necessary to prevent SARS-CoV-2 re-infection. Conversely, sub-neutralizing antibody levels have been reported to facilitate, rather than inhibit, viral entry of the some coronaviruses *in vitro*, through antibody dependent enhancement (ADE).(24-26) While ADE dependent replication has not been demonstrated to occur in SARS-CoV, viral uptake into macrophages via antibody association with Fc receptors does induce IL-6 and TNFα cytokines which may promote inflammation and tissue damage.(27) Insights gained from an accurate analysis of antibody levels and neutralization activity in SARS-CoV-2-infected individuals will help address these important questions and the corresponding health consequences.

A key biological question is: what underlies the large variation in antibody titers (neutralizing or

otherwise) observed in CP donors? Numerous variables, including the effectiveness of innate immune

responses, SARS-CoV-2 exposure dose, anatomical site of initial infection, and partial cross-reactive

immunity conferred by prior seasonal coronavirus infection, could all impart variation on the amount and

dissemination of SARS-CoV-2 antigen. Variation in the exposure of the adaptive immune system to SARS-

CoV-2 antigen would, in turn, likely impact the magnitude of immune responses. Our observation that the

levels of antibody to N, as well as S, correlates with S-specific neutralizing titer suggests that quantitative

differences in the overall adaptive immune response to SARS-CoV-2, rather than intrinsic differences in

the ability of individuals to mount neutralizing responses, at least partly explains the large variation in

neutralizing capacity of CP. This notion is consistent with recent findings that all individuals examined,

generated very similar, potent monoclonal SARS-CoV-2 neutralizing antibodies, but at very different

levels.(15)

Future utility for vaccine and CP donor strategies

The development of efficacious vaccines against SARS-CoV-2 may be necessary for ending the

COVID19 pandemic. Clinical trials will undoubtedly include a battery of serological and neutralization

assays in test subjects to assess candidate vaccine efficacy. Surrogate serology tests to neutralizing activity

could help to rapidly inform as to the likely effectiveness, as well as immunogenicity, of vaccines against

SARS-CoV-2. To this end, real-time analyses using scalable HTSA testing platforms is effectuate while

future studies are conducted to more precisely measure *in vivo* neutralization activity.

Finally, the utility of convalescent plasma in the treatment of infection has been recognized since

the turn of the 20th century.(28) CP transfusion is thought to be effective through passive immunization,

specifically the transfer of neutralizing antibodies from a recovered individual to another individual

manifesting life-threatening symptoms.(29, 30) Previously CP therapy has been used to treat both SARS

and MERS,(31) and currently can be rapidly deployed against SARS-CoV-2 while other therapies are

under development.(32) Nevertheless, many questions remain regarding the optimal antibody levels

necessary to treat patients at varying stages of COVID19 disease. Accurate quantification using serological

assays that predict neutralization activity may improve clinical outcomes through refinement of CP unit

selection for patients of varying symptomatology. In summary, we demonstrate that HTSA and S1 ELISA

assays show the strongest correlation with neutralization activity and may serve to predict the degree of

antiviral antibody activity present in recovered patients or vaccine recipients.

Authors' Contributions

LLL conceptualized the study, designed and performed serology experiments and managed the collection

of data, figures and statistical analyses. LLL, PB, TH and CDH co-wrote the manuscript. TH, PB, FM, YW

and FS designed and performed the neutralization assays. BR, DJ, WB, SJ, JP, MR and NT performed most

of the serology experiments. CG, MP, ES and HZ processed and preserved donor plasma and PBMCs. DS

coordinated donor demographic information. KY contributed to PBMC flow cytometry and interpretation.

DJ and BR contributed equal authorship to the manuscript. LLL, PB and TH contributed equal

corresponding authorship.

Acknowledgements

We thank Jill Alberigo, Amanda Brites and Kelly Brightman from Rhode Island Blood Center for their help

with performing the Ortho Anti-SARS-CoV-2 Total Test and the Abbott SARS-CoV-2 IgG test. We thank

Chockalingam Palaniappan and Paul Contestable for their assistance with performing the Ortho Anti-

SARS-CoV-2 IgG Test. We thank Haidee Chen for assistance with editing the manuscript.

Conflicts of Interest

The authors declare no conflicts of interest.

Role of the Funding Source

Funding source for TH, PBD, FM, YW and FS were NIH R01AI78788 and R37AI064003. Funding

sources did not have a role in the writing of the manuscript or the decision to submit for publication.

Methods

Cell lines

Huh7.5 cells were a gift from Charles Rice(33). The 293T/ACE2cl.13 cell clone was generated by transducing 293T cells (ATCC® CRL-3216™) with a CSIB-based ACE2 lentivirus expression vector containing a cDNA encoding a catalytically inactive ACE2 mutant. Single cell clones were isolated by limiting dilution and one clone (293T/ACE2cl.13) was used in these studies.

Collection of CP donor information, isolation of convalescent plasma and PBMCs

Disclosure of demographic information was elective at the time of donation and showed that of the 370 CP donors analyzed, 71.1% indicated age, 95.4% indicated blood type, 95.6% indicated sex and 55.1% indicated ethnicity. To examine the demographic characteristics within the convalescent plasma (CP) donor population, we used the 2010 U.S. Census demographic data as expected frequencies. Plasma was isolated from EDTA-anticoagulated human whole blood samples. Samples were shipped from the NYBC Sample Management Facility overnight at 4C and centrifuged for 5 min at 500 xg to facilitate plasma/cell phase separation. The resulting upper plasma layer was extracted, aliquoted to minimize future freeze-thaw cycles, and stored at -80 C. Samples were cryopreserved and stored in the NYBC COVID19 Research Repository (https://nybc.org/covid19repository).

Plasmid constructs

The *env*-inactivated HIV-1 reporter construct (pHIV-1$_{NL4-3}$ ΔEnv-NanoLuc) was generated from a pNL4-3 infectious molecular clone (obtained through NIH AIDS Reagent Program, Division of AIDS, NIAID, NIH from Dr Malcolm Martin). It contains a NanoLuc Luciferase reporter gene in place of nucleotides 1-100 of the *nef*-gene and a 940 bp deletion 3' to the *vpu* stop-codon. The rVSVΔG/NG/NanoLuc plasmid was generated by insertion of a cassette containing an mNeonGreen/FMDV2A/NanoLuc luciferase cDNA into rVSVΔG (Kerafast) (PMID: 20709108) between the M and L genes. The pSARS-CoV-2 S-protein expression plasmid containing a C-terminally truncated SARS-CoV-2 S protein (pSARS-CoV2$_{Δ19}$) was generated by insertion of a synthetic human-codon optimized cDNA encoding SARS-CoV-2 S1 spike protein lacking the C-terminal 19 codons into pCR3.1. An ACE2 lentiviral expression vector was constructed by

inserting a cDNA encoding a catalytically inactive ACE2 mutant into the lentivirus expression vector CSIB

(PMID: 30084827).

SARS-CoV-2 pseudotype particles

To generate (HIV/NanoLuc)-SARS-CoV-2 pseudotype particles, 293T cells were transfected with pHIV-

1_{NL4-3} ΔEnv-NanoLuc reporter virus plasmid and pSARS-CoV-2-$S_{\Delta19}$ at a molar plasmid ratio of 1:0.55. The

transfected cells were washed twice with PBS the following day, and at 48h after transfection, supernatant

was harvested, clarified by centrifugation, passed through a 0.22 μm filter, aliquoted and frozen at -80°C.

To generate (VSV/NG/NanoLuc)-SARS-CoV-2 pseudotype particles, 293T cells were infected with

recombinant T7-expressing vaccinia virus (vTF7-3) and transfected with rVSVΔG/NG/NanoLuc, pBS-N,

pBS-P, pBS-L, and pBS-G (PMID: 20709108). At ~24h post transfection the supernatant was collected,

filtered and used to infect 293T cells transfected with a VSV-G expression plasmid, for amplification. To

prepare stocks of (VSV/NG/NanoLuc)-SARS-CoV-2 pseudotype particles, 293T cells were transfected with

pSARS-CoV2$_{\Delta19}$ and infected with the VSV-G complemented rVSVΔG/NG/NanoLuc virus. At 16h later the

supernatant was collected, clarified by centrifugation, filtered, pelleted through a 20% sucrose cushion and

stored at -80°C. The viral stock was incubated with 20% I1 hybridoma supernatant (ATCC CRL-2700) for

1h at 37°C before use.

Neutralization assays

To measure neutralizing antibody activity in convalescent plasma, five-fold serial dilutions of plasma were

incubated for 1 hour at 37°C in 96-well plates with an aliquot of HIV-1 or VSV-based SARS-CoV-2

pseudotyped virus containing approximately $1x10^3$ infectious units. Thereafter, 100 μl of the plasma/virus

mixture was added to target cells (293T$_{Ace2}$ cl.13, or Huh7.5) cells in 96-well plates. Cells were cultured for

48h (HIV-1 pseudotype viruses) or 16h (VSV pseudotype viruses). Then, cells were washed twice, lysed and

NanoLuc Luciferase activity in lysates was measured using either the Nano-Glo Luciferase Assay System

(Promega) and a Modulus II Microplate Multimode reader (Turner BioSystem) or a Glowmax Navigator

luminometer (Promega). The half maximal neutralizing titer (NT_{50}) for plasma, was determined using a 4-

parameter nonlinear regression in Prism 8.4 (GraphPad).

Lateral Flow ImmunoAssay (LFA)

Lateral flow immunoassays (LFAs) were provided by external companies. Assay cartridges contained detection bands for IgG and IgM against SARS-CoV2 specific epitopes as well as an internal positive control. For each assay, 20 μL convalescent plasma or serum was applied to the sample pad, followed by two drops of proprietary running buffer. After 30 minutes, high resolution pictures of the detection zone were taken and saved as .JPEG files. All tests were performed at room temperature.

LFA Densitometry Analysis

Relative quantification of anti-SARS-CoV-2 IgG and IgM in convalescent plasma samples was performed using built-in gel analysis macros in FIJI (https://fiji.sc/). A rectangular selection covering the detection zone was analyzed using Analyze>Gels>Plot Lanes. Integrated density values were outlined manually and extracted from the resulting plot. Using MS Excel, IgG and IgM values were normalized against the density of the control band.

The remaining whole blood cellular phase was supplemented with 2 mL of 35 g/L HSA/DPBS and diluted 1:1 with DPBS. Diluted whole blood was layered over 7 mL Ficoll-Paque Premium 1.078 g/mL (GE Healthcare) and centrifuged for 20 minutes at 20C and 400xg without braking. Buffer coats were extracted, counted with AOPI viability stain using the Cellometer Auto2000 (Nexelom Bioscience LLC), and frozen in PBMC freezing media (10% DMSO in Knockout SR).

SARS-CoV-2 Binding-Antibody ELISA

Flat-well, nickel-coated 96 well ELISA plates (Thermo Scientific) were coated with 2μg/mL of recombinant S1 spike protein, nucleocapsid protein, or Receptor Binding Domain (RBD) spike protein specific to SARS-CoV-2 in resuspension buffer (1% Human Serum Albumin in 0.01% PBST) and incubated in a stationary humidified chamber overnight at 4 C. On the day of the assay, plates were blocked for 30 min with ELISA blocking buffer (3% W/V non-fat milk in PBST). Standard curves for both S1 and RBD assays were generated by using mouse anti-SARS-CoV spike protein monoclonal antibody (clone [3A2], ABIN2452119, Antibodies-Online) as the standard. Anti-SARS-CoV-2 Nucleocapsid mouse monoclonal antibody (clone [7E1B], bsm-41414M, Bioss Antibodies) was used as a standard for nucleocapsid binding assays.

Monoclonal antibody standard curves and serial dilutions of convalescent donor plasma were prepared in

assay buffer (1% non-fat milk in PBST) and added to blocked plates in technical duplicate for 1 hr with

orbital shaking at room temperature. Plates were then washed three times with PBST and incubated for 1 hr

with ELISA assay buffer containing Goat anti-Human IgA, IgG, IgM (Heavy & Light Chain) Antibody-HRP

(Cat. No. ABIN100792, Antibodies-Online) and Goat anti-Mouse IgG2b (Heavy Chain) Antibody-HRP (Cat.

No. ABIN376251, Antibodies-Online) at 1:30000 and 1:3000 dilutions, respectively. Plates were then

washed three times, developed with Pierce TMB substrate for 5 min, and quenched with 3 M HCl.

Absorbance readings were collected at 450 nm. Standard curves were constructed in Prism 8.4 (Graphpad

Software Inc.) using a Sigmoidal 4PL Non-Linear Regression (curve fit) model.

High-throughput Serology Assays

Convalescent donor plasma samples were barcoded and dispatched to Rhode Island Blood Center (RIBC).

Samples were analyzed using the Abbott SARS-CoV-2 IgG chemiluminescent microparticle immunoassay

with the Abbott Architect *i*2000SR (Abbott Core Laboratories), as well as the VITROS Immunodiagnostic

Products Anti-SARS-CoV-2 Total Test and the Anti-SARS-CoV-2 IgG Test with the VITROS 5600 (Ortho

Clinical Diagnostics). All assays were performed by trained RIBC employees according to the respective

manufacturer standard procedures.

Flow cytometric analysis of PBMCs

Cryopreserved PBMCs were thawed, filtered and stained with a B-cell or T-cell antibody cocktail for 30

minutes in PBS. Cells were washed with PBS and analyzed with a BD LSR Fortessa 4 laser cytometer.

Cytometric analysis was performed using RUO FCS Express 7 (DeNovo Software).

References

1.   F. Wu *et al.*, A new coronavirus associated with human respiratory disease in China. *Nature* **579**, 265-269 (2020).

2.   P. Zhou *et al.*, A pneumonia outbreak associated with a new coronavirus of probable bat origin. *Nature* **579**, 270-273
(2020).

3.   K. G. Andersen, A. Rambaut, W. I. Lipkin, E. C. Holmes, R. F. Garry, The proximal origin of SARS-CoV-2. *Nat Med*
26, 450-452 (2020).

4.   M. Xie, Q. Chen, Insight into 2019 novel coronavirus - An updated interim review and lessons from SARS-CoV and
MERS-CoV. *Int J Infect Dis* **94**, 119-124 (2020).

5.   Q. Yao *et al.*, Retrospective study of risk factors for severe SARS-Cov-2 infections in hospitalized adult patients. *Pol
Arch Intern Med* 10.20452/pamw.15312 (2020).

6.   A. J. Kucharski *et al.*, Early dynamics of transmission and control of COVID-19: a mathematical modelling study. *Lancet
Infect Dis* **20**, 553-558 (2020).

7.   C. M. Chan *et al.*, Examination of seroprevalence of coronavirus HKU1 infection with S protein-based ELISA and
neutralization assay against viral spike pseudotyped virus. *J Clin Virol* **45**, 54-60 (2009).

8.   C. Y. Lee, R. T. P. Lin, L. Renia, L. F. P. Ng, Serological Approaches for COVID-19: Epidemiologic Perspective on
Surveillance and Control. *Front Immunol* **11**, 879 (2020).

9.   X. Zhong *et al.*, B-cell responses in patients who have recovered from severe acute respiratory syndrome target a dominant
site in the S2 domain of the surface spike glycoprotein. *J Virol* **79**, 3401-3408 (2005).

10.   W. Tai *et al.*, Characterization of the receptor-binding domain (RBD) of 2019 novel coronavirus: implication for
development of RBD protein as a viral attachment inhibitor and vaccine. *Cell Mol Immunol* 10.1038/s41423-020-0400-4
(2020).

11.   L. A. VanBlargan, L. Goo, T. C. Pierson, Deconstructing the Antiviral Neutralizing-Antibody Response: Implications for
Vaccine Development and Immunity. *Microbiol Mol Biol Rev* **80**, 989-1010 (2016).

12.   DHHS, U.S. Census Bureau.

13.   M. Hoffmann *et al.*, SARS-CoV-2 Cell Entry Depends on ACE2 and TMPRSS2 and Is Blocked by a Clinically Proven
Protease Inhibitor. *Cell* **181**, 271-280 e278 (2020).

14.   A. W. Fan Wu, Mei Liu, Qimin Wang, Jun Chen, Shuai Xia, Yun Ling, Yuling Zhang, Jingna Xun, Lu Lu, Shibo Jiang,
Hongzhou Lu, Yumei Wen, Jinghe Huang, Neutralizing antibody responses to SARS-CoV-2 in a COVID-19 recovered
patient cohort and their implications. *medRxiv* 10.1101/2020.03.30.20047365 (2020).

15.   C. G. Davide F. Robbiani, Frauke Muecksch, Julio C. C. Lorenzi1 , Zijun Wang, Alice Cho, Marianna Agudelo,
Christopher O. Barnes, Anna Gazumyan, Shlomo Finkin, Thomas Hagglof, Thiago Y. Oliveira, Charlotte Viant, Arlene
Hurley, HansHeinrich Hoffmann, Katrina G. Millard, Rhonda G. Kost, Melissa Cipolla, Kristie Gordon,, S. T. C. Filippo
Bianchini1, Victor Ramos, Roshni Patel, Juan Dizon, IrinaShimeliovich1, Pilar Mendoza, Harald Hartweger, Lilian
Nogueira, Maggi Pack, Jill Horowitz, Fabian Schmidt, Yiska Weisblum, Eleftherios Michailidis, Alison W. Ashbrook,
Eric Waltari,John E. Pak, Kathryn E. Huey-Tubman, Nicholas Koranda, Pauline R. Hoffman, Anthony P.West, Jr.,
Charles M. Rice, Theodora Hatziioannou, Pamela J. Bjorkman, Paul D. Bieniasz, Marina Caskey, Michel C.
Nussenzweig, Convergent Antibody Responses to SARS-CoV-2 Infection in Convalescent Individuals.
10.1101/2020.05.13.092619. (2020).

16.   C. P. Gross *et al.*, Racial and Ethnic Disparities in Population-Level Covid-19 Mortality. *J Gen Intern Med*
10.1007/s11606-020-06081-w (2020).

17.   B. H. Shaz, A. B. James, K. L. Hillyer, G. B. Schreiber, C. D. Hillyer, Demographic patterns of blood donors and donations
in a large metropolitan area. *J Natl Med Assoc* **103**, 351-357 (2011).

18.   B. H. Shaz, C. D. Hillyer, Minority donation in the United States: challenges and needs. *Curr Opin Hematol* **17**, 544-549
(2010).

19.   A. W. Fan Wu, Mei Liu, Qimin Wang, Jun Chen, Shuai Xia, Yun Ling, Yuling Zhang, Jingna Xun, Lu Lu, Shibo Jiang,
Hongzhou Lu, Yumei Wen, Jinghe Huang, Neutralizing antibody responses to SARS-CoV-2 in a COVID-19 recovered
patient cohort and their implications. 10.1101/2020.03.30.20047365.

20.   M. J. Joyner *et al.*, Effect of Convalescent Plasma on Mortality among Hospitalized Patients with COVID-19: Initial
Three-Month Experience. *medRxiv* 10.1101/2020.08.12.20169359 (2020).

21.   P. Kellam, W. Barclay, The dynamics of humoral immune responses following SARS-CoV-2 infection and the potential
for reinfection. *J Gen Virol* 10.1099/jgv.0.001439 (2020).

22.   K. A. Callow, H. F. Parry, M. Sergeant, D. A. Tyrrell, The time course of the immune response to experimental
coronavirus infection of man. *Epidemiol Infect* **105**, 435-446 (1990).

23.   A. F. Bradburne, M. L. Bynoe, D. A. Tyrrell, Effects of a "new" human respiratory virus in volunteers. *Br Med J* **3**, 767-
769 (1967).

24.   M. Jaume *et al.*, Anti-severe acute respiratory syndrome coronavirus spike antibodies trigger infection of human immune
cells via a pH- and cysteine protease-independent FcgammaR pathway. *J Virol* **85**, 10582-10597 (2011).

25.   S. F. Wang *et al.*, Antibody-dependent SARS coronavirus infection is mediated by antibodies against spike proteins.
Biochem Biophys Res Commun **451**, 208-214 (2014).

26.   Y. Wan *et al.*, Molecular Mechanism for Antibody-Dependent Enhancement of Coronavirus Entry. *J Virol* **94** (2020).

27.   A. Iwasaki, Y. Yang, The potential danger of suboptimal antibody responses in COVID-19. *Nat Rev Immunol* **20**, 339-
525   341 (2020).
526   28.   G. Marano *et al.*, Convalescent plasma: new evidence for an old therapeutic tool? *Blood Transfus* **14**, 152-157 (2016).
527   29.   E. M. Bloch *et al.*, Deployment of convalescent plasma for the prevention and treatment of COVID-19. *J Clin Invest*
528   10.1172/JCI138745 (2020).
529   30.   K. Duan *et al.*, Effectiveness of convalescent plasma therapy in severe COVID-19 patients. *Proceedings of the National*
530   *Academy of Sciences of the United States of America* **117**, 9490-9496 (2020).
531   31.   B. L. Brown, J. McCullough, Treatment for emerging viruses: Convalescent plasma and COVID-19. *Transfus Apher Sci*
532   10.1016/j.transci.2020.102790, 102790 (2020).
533   32.   A. Budhai *et al.*, How did we rapidly implement a convalescent plasma program? *Transfusion* 10.1111/trf.15910 (2020).
534   33.   K. J. Blight, J. A. McKeating, C. M. Rice, Highly permissive cell lines for subgenomic and genomic hepatitis C virus
535   RNA replication. *J Virol* **76**, 13001-13014 (2002).
536
537

538 **Figure 1**

541
542

543 **Figure 2**

545
546
547

Figure 3

Figure 4

Figure 1: Demographics of convalescent plasma donors.

A; Distribution of convalescent plasma donor age (left, blue bars) compared to U.S. population (right, red bars). Dotted line represents Gaussian distribution curve fit. N=263; Pearson's correlation coefficient.

B; Distribution of convalescent plasma donor blood group antigen (left, blue bars) compared to U.S. population (right, red bars). N=370, binomial test for discrepancy versus U.S. population; * p < 0.05.

C; Distribution of convalescent plasma donor sex (blue bars) compared to U.S. population (red bars). N=354, binomial test for discrepancy versus U.S. population.

D; Distribution of convalescent plasma donor ethnicity (blue bars) compared to U.S. population (right, red bars). N=204, binomial test for discrepancy versus U.S. population; * p<0.05.

Figure 2: Neutralizing activity analysis of convalescent plasma donors.

A; Distribution of neutralization IC_{50} values (NT50, reciprocal plasma dilution) of convalescent donor plasma using HIV (red) or VSV pseudovirus (blue) overexpressing the SARS-CoV-2 spike protein (S).

B; Frequency of convalescent plasma donor NT50 values within indicated groups using HIV-S (top) or VSV-S pseudovirus constructs.

C; Frequency distribution of convalescent plasma HIV-S NT50 values versus age groups. Signal to cutoff (S/co, dotted grey line) and 10x S/co (solid grey line) thresholds are indicated. n=5-38, Kruskal-Wallis test; * $p < 0.05$.

D; Frequency of convalescent plasma donor NT50 values versus sex. Signal to cutoff (S/co, dotted grey line) and 10x S/co (solid grey line) thresholds are indicated. n=190, Mann-Whitney test, ** $p < 0.01$.

E; Frequency of convalescent plasma donor NT50 values versus blood group antigen. Signal to cutoff (S/co, dotted grey line) and 10x S/co (solid grey line) thresholds are indicated. n=15-82, Kruskal-Wallis test, * $p < 0.05$.

F; Frequency of convalescent plasma donor NT50 values versus time (days) since last reported symptom. Signal to cutoff (S/co, dotted grey line) and 10x S/co (solid grey line) thresholds are indicated. n=19-33, Mann-Whitney t-test, * $p < 0.05$.

592 **Figure 3: Serological analysis of convalescent plasma donors.**
593

594 **A;** Frequency of densitometric IgG (left) or IgM (right) results from LFA bands relative to control band.
595 Median values (red band) with 1st and 3rd quartiles (thin red lines) are shown.
596

597 **B;** Frequency of HTSA results using the total Ig or IgG assays derived from the Ortho Diagnostics platform
598 (left) or Abbott IgG assay platform (right). Results from fresh frozen plasma (FFP) units collected before
599 COVID19 are shown as healthy controls.
600

601 **C;** Frequency of S1 spike protein (left), Nucleocapsid (NP) protein (center) and RBD spike protein (right)
602 ELISA titer results. Titers reflect concentrations calculated using a mAb standard curve and not absolute
603 plasma concentrations. Median values (red band) with 1st and 3rd quartiles (thin red lines) are shown.
604

Figure 4: Correlation of serology assays versus neutralization activity of convalescent plasma donors.

A; Linear regression of HIV-S NT50 values (abscissa) versus serological assay values (ordinate). N indicated in each graph, r^2 = goodness of fit.

B; Spearman correlation coefficients, *r*, of neutralization and serological assays. N=137 samples.

C; Distribution of CP donor sample HTSA scores within indicated HIV-S NT50 groups using Ortho total Ig (left), Ortho IgG (center) or Abbott IgG (right) assays.

D; Frequency of convalescent donor S1 and NP ELISA values defined in **C**. n=241 samples.

E; Distribution of NT50 values corresponding to populations defined in **C**. n=4-51, Kruskall-Wallis test, * $p < 0.05$, ** $p < 0.01$.