Cochlear implant-related speech processing may diminish the advantage of exposure to

infant-directed speech [a]

Meisam K. Arjmandi [1,b], Derek Houston [2,c], Yuanyuan Wang [2], Laura C. Dilley [1]

[1] Department of Communicative Sciences and Disorders, Michigan State University, 1026 Red

Cedar Road, East Lansing, Michigan 48824, USA

[2] Department of Otolaryngology – Head and Neck Surgery, The Ohio State University, 915

Olentangy River Road, Columbus, Ohio, 43212, USA

Date of manuscript submission: May 14, 2020

[a] Portions of this work were presented at the 177th and 178th Meeting of the Acoustical Society

of America.

[b] Current affiliation: Department of Otolaryngology – Head and Neck Surgery, Massachusetts Eye

and Ear and Harvard Medical School, Boston, Massachusetts, USA. Email:

meisam_khalilarjmandi@meei.harvard.edu

[c] Also at: Nationwide Children's Hospital, Columbus, OH, USA

ABSTRACT

Caregivers modify their speech when talking to infants, a specific type of speech known as infant-

directed speech (IDS). This speaking style facilitates language learning compared to adult-directed

speech (ADS) in infants with normal hearing (NH). While infants with NH and those with cochlear

implants (CIs) prefer listening to IDS over ADS, it is yet unknown how CI speech processing may

affect the acoustic distinctiveness between ADS and IDS, as well as the degree of intelligibility of

these. This study analyzed speech of seven female adult talkers to investigate the effects of

simulated CI speech processing on (1) acoustic distinctiveness between ADS and IDS, (2)

estimates of intelligibility of caregivers' speech in ADS and IDS, and (3) individual differences in

caregivers' ADS-to-IDS modification and speech intelligibility. Results suggest that CI speech

processing is substantially detrimental to the acoustic distinctiveness between ADS and IDS, as

well as to the intelligibility benefit derived from ADS-to-IDS modifications. Moreover, the

observed considerable variability across individual talkers in acoustic implementation of ADS-to-

IDS modification and speech intelligibility was significantly reduced due to CI speech processing.

The findings are discussed in the context of the link between IDS and language learning in infants

with CIs.

Keywords: Infant-directed speech, cochlear implant, acoustic distance, speech intelligibility

I. INTRODUCTION

A. Infant-directed speech and language learning

Caregivers around the world modify their speaking style from adult-directed speech (ADS) to

infant-directed speech (IDS) when talking to infants (Fernald & Kuhl, 1987; Snow, 1977). Greater

exposure to IDS and some aspects of its acoustic and linguistic properties facilitates language

learning in children with normal hearing (NH) (Cristia & Seidl, 2014; Liu, Kuhl, & Tsao, 2003;

Song, Demuth, & Morgan, 2010; Trainor, Austin, & Desjardins, 2000; Weisleder & Fernald,

2013). The supportive role of IDS in infants' language learning occurs through several pathways,

such as directing and holding infants' attention to speech (Fernald, 1985, 1989; Kitamura,

Thanavishuth, Burnham, & Luksaneeyanawin, 2001; Kuhl et al., 1997; Schachner & Hannon,

2011; Wang, Bergeson, & Houston, 2017, 2018; Werker, Pegg, & McLeod, 1994), while making

speech more clear by providing more salient cues for speech discrimination (Karzon, 1985), word

segmentation (Singh & Nestor, 2009), speech segmentation (Thiessen, Hill, & Saffran, 2005), and

word learning (Ma, Golinkoff, Houston, & Hirsh-Pasek, 2011). Taken together, these studies

demonstrate the supportive role of IDS in infants' early language learning. The purpose of the

present study was to investigate how signal processing associated with cochlear implants (CIs)

impacts acoustic cues pertaining to distinguishing and recognizing words in IDS versus ADS.

B. Infants' attention to IDS

In listening to speech, infants with NH and those with CIs demonstrate increased attention to IDS

compared to ADS (The ManyBabies Consortium, 2020; Cooper, Abraham, Berman, & Staska,

1997; Fernald, 1985; Wang et al., 2017; Wang, Bergeson, et al., 2018; Werker & McLeod, 1989;

Werker et al., 1994). This effect of IDS on attention begins very early as observed in newborns

and infants as young as 4 months old (Cooper & Aslin, 1990; Fernald, 1985; Werker & McLeod,

1989), demonstrating the capability of infants in using subtle acoustic information to distinguish

IDS from ADS while having acquired only minimal linguistic knowledge. Multiple acoustic-

phonetic cues have been identified as potentially supporting infants' robust preference for IDS

over ADS, such as higher pitch (Cooper & Aslin, 1990; Fernald, 1989), greater pitch fluctuation

(Fernald & Simon, 1984), slower speaking rate (Leong, Kalashnikova, Burnham, & Goswami,

2014; Narayan & McDermott, 2016; Song et al., 2010), and an expanded vowel space (Burnham

et al., 2015) in IDS compared to ADS. These distinctive acoustic-phonetic cues are manifested in

spectro-temporal representations of caregivers' speech, which are actively incorporated through

infants' highly sensitive auditory systems to recognize IDS as distinct from ADS (Kuhl, 2004;

Maye, Werker, & Gerken, 2002; McMurray & Aslin, 2005; Telkemeyer et al., 2009). Although

IDS over ADS preference was observed in infants with CIs (Wang et al., 2017), it is still not clear

how they are different from their peers with NH in terms of access to spectral information involved

in distinguishing IDS from ADS. Infants with CIs may face difficulties in resolving distinctive

acoustic information toward performing this task, leading potentially to a disruption in the

developmental time course of CI infants' attention to IDS and delayed language development, as

compared to that of NH infants. These considerations suggest a benefit of examining how CI-

related speech processing may affect acoustic distinctiveness and intelligibility of IDS compared

with ADS, and how this acoustic distinctiveness and intelligibility may differ as a function of

individual talkers' speech patterns.

C. CI vocoding and acoustic distinction between IDS and ADS

Infants with CIs have access to a spectro-temporally degraded version of speech, due to limitations

of electric stimulation to faithfully transmit this information to the auditory nerve. Friesen et al.

(2001) showed that speech recognition in listeners with NH improved by increasing the number

of spectral channels in the vocoded speech, confirming that spectral resolutions are limited by the

number of channels in the CI vocoder (Friesen et al., 2001; Fu, Chinchilla, Nogaki, & Galvin,

2005; Fu & Nogaki, 2005; Luo & Poeppel, 2007; Shannon, Zeng, Kamath, Wygonski, & Ekelid,

1995; Svirsky, 2017). Temporal resolution is also negatively affected by low-pass filtering of

amplitude envelope, which disrupts faithful transmission of temporal fine-structure in speech to

auditory nerves (Rubinstein, 2004; Svirsky, 2017). Therefore, the CI vocoding process may

negatively affect infants' perception of segmental and suprasegmental features of IDS.

Transmission of multiple speech cues are affected by the CI vocoding process, including

fundamental frequency (F0) (Mehta & Oxenham, 2017; Qin & Oxenham, 2005; Svirsky, 2017),

timbre (Galvin, Mok, Dowell, & Briggs, 2008), and melody (Mehta & Oxenham, 2017; Zeng,

Tang, & Lu, 2014), which are recognized as major attributes that distinguish IDS from ADS

(Fernald, 1989; Fernald & Simon, 1984; Wang et al., 2017). For example, pitch is an important

perceptual cue for children's robust attention in listening to IDS (Fernald, 1985; Piazza, Iordan, &

Lew-Williams, 2017) and was recognized as the most prominent acoustic determinant in

distinguishing IDS from ADS (Fernald & Kuhl, 1987). It also contributes to lexical segmentation

of speech in listeners with CIs (Spitzer, Liss, Spahr, Dorman, & Lansford, 2009), a task critical for

word learning. Recognition of emotion from speech is another important aspect in processing IDS

(Trainor et al., 2000), which relies heavily on perceiving pitch height and fluctuation. Listeners

with CIs performed worse in emotion recognition as the number of spectral channels in vocoded

speech was decreased (Luo, Fu, & Galvin, 2007). Kong et al. (2004) further found that melodic

pitch perception is negatively affected by reducing the number of spectral channels. Taken

together, these studies demonstrate that children with CIs have partial access to spectro-temporal

cues in speech, particularly ones that significantly contribute to the distinctiveness of IDS and

ADS, which may interfere with the connection between listening to IDS and developing better

language skills. Therefore, it is important to understand how the CI vocoding process may

acoustically impact the distinction between IDS and ADS - a major knowledge gap that the present

study aimed to address.

D. CI vocoding and intelligibility of IDS

Another possible way IDS may foster language learning is by making speech more intelligible.

Prior studies showed evidence that IDS may provide acoustic cues that assist infants for parsing

linguistic units in speech. Karzon (1985) showed that the exaggerated suprasegmental features of

IDS may provide supportive perceptual cues for syllabification of multisyllabic sequences. In

another study, infants performed better long-term word recognition when the words presented in

IDS compared to ADS (Singh & Nestor, 2009). It was further shown that IDS facilitates word

segmentation (Thiessen et al., 2005). Although evidence on intelligibility benefits of IDS over

ADS have somehow been contradictory at the level of segmental speech cues (e.g., Kuhl et al.,

1997; McMurray, Kovack-Lesh, Goodwin, & McEchron, 2013), there might be an intelligibility

benefit associated with other properties of IDS (e.g., prosodic cues), similar to the effects observed

in close analogues of a "clear speech" register (Bradlow, Kraus, & Hayes, 2003; Ferguson &

Kewley-Port, 2007). Caregivers' modification of speaking style from ADS to IDS impacts

suprasegmental aspects of speech such as pitch contour and speech rate, which have been shown

to be strong predictors of speech clarity and intelligibility (Bradlow et al., 2003; Cutler, Dahan, &

van Donselaar, 1997; Ferguson & Poore, 2010; Ferguson & Quené, 2014; Spitzer, Liss, & Mattys,

2007; Watson & Schlauch, 2008). Different choices of speaking style often impact speech rate,

where slower rate is thought to contribute to enhanced intelligibility of clear speech compared to

conversational speech both in typical listeners (Ferguson et al., 2010) and recipients of CIs (Li et

al., 2011; Zanto, Hennigan, Östberg, Clapp, & Gazzaley, 2013).

Limited spectral resolution of the cochlear implant device affects the quality by which

listeners with CIs receive speech cues (Croghan, Duran, & Smith, 2017; Jain & Vipin Ghosh,

2018; Peng, Hess, Saffran, Edwards, & Litovsky, 2019; Qin & Oxenham, 2005), which may

negatively impact CI users' ability to recognize words and phonemes, especially children (Grieco-

calub, Saffran, & Litovsky, 2010; Peng et al., 2019). Therefore, any likely beneficial role of IDS

in improvement of speech intelligibility is prone to change when caregivers' speech is perceived

through a CI device. Prior studies have shown that performance of listeners with CIs is

considerably poorer than listeners with NH in understanding sentences spoken with relatively

higher rate (Zanto et al., 2013), which is an acoustic dimension signaling distinction of IDS from

ADS (i.e., slower speaking rate in IDS compared to ADS). Variation of fundamental frequency

(F_0) is another acoustic dimension of the IDS-vs-ADS distinction that contributes to speech

recognition (Spitzer, Liss, Spahr, Dorman, & Lansford, 2009; Spitzer et al., 2007). Findings on the

effect of these acoustic properties on intelligibility of caregivers' speech is still controversial, and

little is known about the effects of CI speech processing on intelligibility of caregivers' speech in

ADS and IDS conditions. The present study aimed to address whether IDS is beneficial for

intelligibility improvement and how this effect may be compromised depending on listeners'

hearing status (NH vs. CI).

E. Individual differences in acoustic implementation of IDS and speech intelligibility

Listeners' familiarity with the range of variability in voice timbres of individual talkers is

important for robust identification and recognition of individuals' voices (Lavan, Burton, Scott, &

McGettigan, 2019; Souza, Gehani, Wright, & McCloy, 2013). Caregiver-specific acoustic

modification is important for effective caregiver-infant interaction and significantly contributes to

infants' recognition of caregivers' voices both before and after birth (Beauchemin et al., 2011;

Kisilevsky et al., 2003), as well as for neural development of infants' auditory cortices (Webb,

Heller, Benson, & Lahav, 2015). Singh (2008) showed that exposure to IDS with relatively higher

variability facilitates word recognition in infants by forming a relatively wider and more

generalizable lexical category, highlighting the supportive role of exposure to a wide range of

ADS-to-IDS acoustic variation in effective processing of caregivers' speech. The degree of

acoustic variability that caregivers introduce to their infants may vary across individuals. The

acoustic characteristics of IDS reveal finer details about the unique voice timbre of individual

caregivers, which potentially contributes to infant' identification of individual caregivers (Piazza

et al., 2017). It is, however, unclear how CI vocoding impacts the range of acoustic variability

within and across caregivers due to shifts in caregivers' speaking style (i.e., ADS to IDS), a change

which likely eliminates acoustic details and potentially negatively impacts the supportive role of

IDS in recognizing caregivers' voices. CI vocoding is also detrimental to the acoustic distance of

voices across caregivers, which potentially makes distinguishing talkers from each other more

challenging for infants with CIs, compared to those with NH. The presents study aimed to

investigate these questions.

F.   Benefits of investigations of acoustic properties of CI-related speech processing

Studies aimed at understanding speech perception in individuals with CIs can be characterized as

involving one of three general approaches. The first approach examines performance of listeners

with NH in response to vocoded speech to simulate hearing in listeners with CIs (Dorman, Loizou,

& Rainey, 1997a; Jahn, DiNino, & Arenberg, 2019; Mehta & Oxenham, 2017; Qin & Oxenham,

2003, 2005). The second method involves directly study of how listeners with CIs perform in

various speech recognition tasks (Brown & Bacon, 2010; Kong, Stickney, & Zeng, 2005; Peng,

Tomblin, & Turner, 2008; Peng, Hess, Saffran, Edwards, & Litovsky, 2019). In a third method –

the one which we focus on in the present paper – acoustic properties of simulated CI speech and

unprocessed analogs are analyzed comparatively. This comparative analysis can be facilitated

through the use of various quantitative metrics that emulate speech recognition in CI users as

proxies to analyze properties of CI speech and its perception in listeners with CIs (Jain & Vipin

Ghosh, 2018; Qin & Oxenham, 2003; Santos, Cosentino, Hazrati, Loizou, & Falk, 2013). Analysis

of simulated CI speech and quantitative measures of speech quality and intelligibility tailored to

listeners with CIs provide repeatable, automated, inexpensive, and fast tools for gaining

preliminary evidence about speech perception in individuals with CIs. As such, a comparative

analysis method offers benefits for undertaking efficient investigations for further study, and as

such provides benefits over various major challenges that accompany the first two categories of

studies, such as participant recruitment, time, and cost.

G. Current study

In the present study, we analyzed unprocessed and simulated CI speech of seven female talkers

who spoke utterances both in ADS and IDS. This permitted examining how CI speech processing

may affect acoustic distinctiveness between ADS and IDS, as well as intelligibility of caregivers'

speech under modifications of speaking style. A signal processing-based approach was taken here

to analyze ADS and IDS stimuli of the female talkers using noise-excited envelope vocoder

processing, which is one type of CI-simulated speech which has been used in a number of prior

studies (Dorman, Loizou, Fitzke, & Tu, 1998; Dorman, Loizou, & Rainey, 1997b; Friesen et al.,

2001; Fu, Shannon, & Wang, 1998; Loizou, Dorman, & Tu, 1999; Mehta, Lu, & Oxenham, 2020;

Mehta & Oxenham, 2017; Qin & Oxenham, 2003, 2005; Rosen, Faulkner, & Wilkinson, 1999;

Shannon et al., 1995).

Our approach involved calculating three main types of acoustic or statistical measures. First,

we calculated mel-frequency cepstral coefficients (MFCCs), which have been shown in multiple

studies to be effective for characterizing the distinctive qualities of IDS vs. ADS (Inoue,

Nakagawa, Kondou, Koga, & Shinohara, 2011; Piazza et al., 2017; Sulpizio et al., 2018). Modeling

the frequency response of human auditory system via MFCCs captures various acoustic-phonetic

features, that distinguish IDS and ADS, such as shifts in timbre (Piazza et al., 2017) and acoustic-

phonetic distinctiveness between IDS and ADS (Inoue et al., 2011), including in other languages

(e.g., Italian and German) (Sulpizio et al., 2018). MFCCs are also able to effectively reflect

caregivers' emotional state and vocal affect (e.g., happiness vs. sadness, Sato & Obuchi, 2007;

Slaney & McRoberts, 1998), attributes that help explain infants' preference for IDS over ADS

(Fernald, 2018; Horowitz, 1983; Mastropieri & Turkewitz, 1999; Moore, Spence, & Katz, 1997;

Papoušek, Bornstein, Nuzzo, Papoušek, & Symmes, 1990; Singh, Morgan, & Best, 2002; Walker-

Andrews & Grolnick, 1983; Walker-Andrews & Lennon, 1991). Furthermore, MFCCs allowed us

to analyze spectral properties of vocoded speech, which was not feasible to do based on calculation

of common acoustic properties such as fundamental frequency (F_0), as these cues are not present

in CI-simulated speech. Second, to quantify the acoustic distinctiveness between IDS and ADS,

we calculated a *Mahalanobis distance (MD)* measure over MFCCs features. MD is a multivariate

distance metric that has been widely used to measure the distances between vectors in a variety of

multidimensional feature spaces (Arjmandi, Dilley, & Wagner, 2018; Masnan et al., 2015; Xiang,

Nie, & Zhang, 2008). Third, to understand how the vocoding process may influence intelligibility

of caregivers' speech, we calculated the speech-to-reverberation-modulation energy ratio (SRMR)

to model the signal-based intelligibility of signals delivered to listeners with NH, as well as its CI-

tailored version (SRMR-CI) to model intelligibility of signals delivered to listeners with CIs

(Santos et al., 2013). These measures were examined both within and across speakers to model the

impacts of speaking styles (IDS vs. ADS) and listener group (NH vs. CI) on signals' distinctiveness

and (estimated) intelligibility. All these analyses were implemented in Matlab 2019a (The

Mathworks Web-Site [http://www.mathworks.com]).

The following four specific questions were addressed in this study. First, we asked how

simulated CI speech processing affects the acoustic distinctiveness of IDS compared with ADS

stimuli, particularly as a function of degree of spectral degradation (i.e., number of spectral

channels) in CI-simulated speech. We hypothesized that (a) CI-related speech processing

significantly degrades the acoustic-phonetic distinctiveness of IDS compared with ADS, and

further that (b) increasing the number of channels would not compensate for degradation imposed

by CI-related processing. Second, we asked whether there is signal-based evidence that IDS may

be more intelligible than ADS, as gauged by the SRMR metric (to simulate a NH listening

condition) or the SRMR-CI metric (to simulate a CI listening condition). We hypothesized that

IDS is more intelligible than ADS, but that (estimated) intelligibility would vary as a function of

(simulated) listening status (NH vs. CI). Third, we asked to what extent the acoustic distinctiveness

between IDS and ADS varies across individual caregivers, as well as how such individual

differences might be impacted by CI speech processing. We predicted individual differences across

caregivers in acoustic implementation of the differences between ADS and IDS; we further

predicted that CI-related speech processing would decrease the extent of acoustic distinctiveness

as a function of speech style, leading to loss of intra- and inter-subject acoustic variability in terms

of IDS vs. ADS distinctiveness. Fourth, and finally, we asked how (estimated) intelligibility

differences for IDS compared with ADS vary across individual caregivers, and how such

intelligibility differ as a function of hearing status (as gauged by SRMR vs. SRMR-CI differences

to simulate effects of NH vs. CI listening conditions, respectively). We hypothesized that

intelligibility differences for IDS vs. ADS would vary across caregivers, and that these

intelligibility differences would be more similar as estimated for CI listeners, compared with NH

listeners.

II. METHODS

A. Speech stimuli

Seven female adult native talkers of American English ranging in age from 21 to 24 years old

spoke fifteen utterances in IDS and ADS (See Appendix in Supplementary Material for the list of

stimuli). The utterances were elicited for use in separate infant word-learning experiments in which

infants were exposed to a novel target word (i.e., "*modi*") in the context of behavioral measures of

infant word recognition to assess whether infants learned the novel word from IDS better than

ADS. To elicit stimuli, the talkers were instructed to speak utterances as if talking to an infant (IDS

condition) or an adult (ADS condition); this procedure was similar to that used in Wang et al.,

(2017; 2018).

Speech stimuli were recorded using an AKG D 542 ST-S microphone in a sound booth and

digitized at a sampling rate of 44.1 kHz with 16-bit resolution. The distance between talkers'

mouths and the microphone was controlled to assure the quality of the recorded stimuli. Prior to

processing as discussed below, the start and end points of recorded utterances were manually

identified in Praat software (Boersma & Weenink, 2001) to remove preceding and following silent

portions. All participants were fully informed about the purpose and procedure of this study, and

they had given informed consent to participate. This study was approved by the Institutional

Review Boards of the Ohio State University and Michigan State University.

B.   Creation of noise-vocoded speech stimuli

CI-simulated versions of the unprocessed (natural) stimuli were created using noise-excited

envelope vocoder processing at six levels of spectral degradation, corresponding to 4-, 8-, 12-, 16-

, 22- and 32-channel noise-vocoded stimuli. The choice of number of channels was made to cover

both the actual number of channels in FDA-approved devices (12, 16, 22) and to query ranges of

variation ranging from minimal cues used in prior studies (4, 8) out to 32 channels to examine CI

vocoder scenarios with higher spectral resolution. The natural stimuli consisting of spoken IDS or

ADS were processed in AngelSim^{TM} Cochlear Implant and Hearing Loss Simulator (Fu, 2019;

Emily Shannon Fu Foundation, www.tigerspeech.com) using 4-, 8-, 12-, 16-, 22-, and 32-channel

noise-vocoding CI-simulated stimuli; the noise vocoding method followed the procedure in

Shannon et al. (1995). The original stimuli were first band-passed filtered using the Greenwood

function (emulating the Greenwood frequency-place map) into N (N = {4, 8, 12, 16, 22, 32})

adjacent frequency channels ranging from 200 Hz to 8000 Hz. This was implemented in

AngelSim^{TM} by setting the absolute lower- and higher-frequency threshold for analysis and carrier

filters to 200 Hz and 8000 Hz with a filter slope of 24 dB/Oct (Fu, 2019). These frequency ranges

are fairly close to the corner frequencies of the Cochlear Nucleus speech processors in CI listeners

cochlear implant devices (Crew & Galvin, 2012; Winn & Litovsky, 2015), which emulate the

performance of average CI listeners in speech envelope discrimination (Chatterjee & Oberzut,

2011; Chatterjee & Peng, 2008). The same analysis filter and carrier filter of the Greenwood

function was used to analyze white noise as a carrier signal (Greenwood, 1990). The *AngelSim*

software used this setup to extract a time-varying amplitude envelope of speech stimuli under each

frequency band using half-wave rectification and then modulated independent white-noise carriers.

There were 210 stimuli per speech style condition (7 talkers x 15 stimuli x 2 speaking styles).

Including the noise-vocoded versions of these utterances (at 4, 8, 12, 16, 22, and 32-channels), we

analyzed 1470 (210 x 7 levels of speech degradation) utterances in the present study.

C. Using MFCC features to characterize acoustic properties of IDS and ADS

We calculated 12 MFCCs for each speech stimulus to characterize its acoustic information (Hunt,

Lennig, & Mermeletein, 1980; Imai, 1983; Shaneh & Taheri, 2009). Figure 1 illustrates this step

for a sample pair of IDS-ADS stimuli in the process for measuring the acoustic distinctiveness

between each ADS-IDS stimulus pair. To calculate MFCCs, each stimulus was re-sampled at 16

kHz using a Hamming window of 25 ms applied to each frame, with a frame shift of 10 ms,

following the procedure used in Inoue et al. (2011). For each pair of IDS and ADS stimuli ($S_{IDS_{ij}}$

and $S_{ADS_{ij}}$ in Figure 1), MFCCs were calculated for all frames of these stimuli ($MFCCs_{IDS_{ij}}$ and

$MFCCs_{ADS_{ij}}$). Here, i indicates the index of the speech stimulus, $i = \{1,2,3,…,15\}$, and j is the

index for the talker, $j = \{1,2,3,…,7\}$. Unlike previous studies in which a single, time-averaged

MFCC vector was calculated to represent acoustic information in IDS and ADS (e.g., Piazza et al.,

2017), our approach took advantage of all MFCCs derived from all frames of a speech stimuli to

calculate the acoustic distinctiveness between ADS-IDS pairs of stimuli within a multidimensional

feature space, thereby preserving the details about spectro-temporal information in speech at the

level of frame. Since each analyzed pair of IDS and ADS stimuli contained identical word strings,

our analysis was expected to mainly model the acoustic-phonetic effect of caregivers' speaking

styles (IDS vs. ADS).

313   >>>>>>>>**FIG. 1 about here** <<<<<<<<

D. Acoustic distinctiveness quantification using Mahalanobis Distance (MD) measure

After representing acoustic features of pairs of IDS and ADS stimuli by calculating their MFCCs,

the acoustic distinctiveness between each pair was measured by calculating MD on the

corresponding MFCC matrices (i.e., $MFCCs_{IDS_{ij}}$ and $MFCCs_{ADS_{ij}}$) using a 12-dimensional feature

space (i.e., 12 MFCCs) (Maesschalck & Massart, 2000; Masnan et al., 2015; Xiang et al., 2008).

MD is a multivariate statistical approach that evaluates distances between two multidimensional

feature vectors or matrices that belong to two classes (here IDS vs. ADS) (Arjmandi et al., 2018;

Heijden, Ferdinand, Ridder, & Tax., 2005; Maesschalck & Massart, 2000; Masnan et al., 2015).

The MD calculation returns the *distance* between means of two classes (here IDS and ADS)

relative to the average per-class covariance matrix (Maesschalck & Massart, 2000). Here, acoustic

properties of each class were represented by two feature matrices (here, MFCCs). The larger the

distance between pairs of $MFCCs_{IDS_{ij}}$ and $MFCCs_{ADS_{ij}}$ feature vectors, the lower the overlap

between two classes of IDS and ADS; this corresponds in turn to greater acoustic distance (or

distinctiveness) between IDS and ADS stimuli.

Each row of the calculated MFCCs matrix for utterance i from talker j (e.g., $MFCCs_{IDS_{ij}}$ or

$MFCCs_{ADS_{ij}}$) corresponds to a frame of that utterance, and 12 MFCCs were presented on the

columns of this matrix. Therefore, the dimension of each MFCC matrix was N_F x 12, where N_F

refers to the number of frames in that speech stimulus. The acoustic distinctiveness between IDS

and ADS stimuli for each female talker j was computed by averaging the fifteen MD values

obtained from that talker's fifteen IDS-ADS pairs ($MD_j = \frac{1}{i}\sum_{n=1}^{i} MD_{ij}$, $i=\{1,2,...,15\}$). This

process was separately performed on pairs of IDS and ADS stimuli for seven levels of spectral

degradation – *unprocessed*, *32*-, *22*-, *16*-, *12*-, *8*-, and/or *4-channel* noise-vocoded CI-simulated

stimuli – to give average MDs used to examine (1) how the acoustic distinctiveness between IDS

and ADS stimuli changes as a function of CI processing (and number of spectral channels), (2)

how individual caregivers vary in acoustic distinctiveness between IDS and ADS stimuli, and (3)

how CI noise-vocoding affects this pattern of individual differences across talker.

E. Estimation of speech intelligibility using quantitative metrics of speech-to-reverberation-

modulation energy ratio (SRMR)

As shown in Figure 1, we estimated the degree of intelligibility of each speech stimulus as it might

be experienced by listeners with NH or with CIs. To approximate stimulus intelligibility as

associated with NH, we calculated the quantitative metric of *speech-to-reverberation-modulation*

energy ratio (SRMR) for IDS and ADS (unprocessed) stimuli, respectively (Santos et al., 2013).

This metric has previously been validated as an approximation of intelligibility of speech to

listeners with NH in behavioral tasks (Falk et al., 2015; Santos et al., 2013). Further, to

approximate intelligibility as might be experienced by listeners with CIs, we calculated an adapted

version of the SRMR metric, which was also developed by Falk et al. (2013) to tailor the SRMR

metric for CI users; this adapted metric is known as SRMR-CI (Santos et al., 2013).

To calculate SRMR, an (unprocessed) speech stimulus was first filtered by a 23-channel

gammatone filterbank in order to emulate cochlear function. Next, a Hilbert transform was applied

to output signals from each of the 23 filters in the filterbank to obtain their temporal envelopes.

Next, modulation spectral energy for each critical band was calculated by windowing stimulus

temporal envelope and computing the discrete Fourier transform. To emulate frequency selectivity

in the modulation domain, the modulation frequency bins were grouped into eight overlapping

modulation bands with logarithmically-spaced center frequencies between 4 and 128 Hz. Finally,

the SRMR value was obtained by calculating the ratio of the average modulation energy in the

first four modulation bands (~ 3–20 Hz) to the average modulation energy in the last four

modulation bands (~ 20-160 Hz), as explained in Santos et al., (2013).

To calculate SRMR-CI, a 22-channel filterbank was used instead of a 23-channel filterbank

in order to emulate the structure of filterbanks in Nucleus CI devices. In addition, the 4-128 Hz

range of modulation for filterbank center frequencies was replaced by a 4–64 Hz range to better

emulate CI users' performance (Santos et al., 2013). Although these measures have not primarily

been developed to estimate the effect of talkers' speaking style on speech intelligibility, the signal

processing algorithms used in these metrics estimate the distribution of energy in various

frequency bands, as well as other spectral properties that are expected to emulate fairly well

performance of NH and CI users in speech recognition. Note that both SRMR and SRMR-CI

intelligibility metrics were calculated from *unprocessed* speech stimulus inputs, rather than noise-

vocoded, CI-simulated analog signals. In fact, intelligibility of signals as estimated for different

hearing statuses (NH or CI) is inherently modeled through the difference in algorithmic

implementation of these metrics (SRMR vs. SRMR-CI, respectively) (Santos et al., 2013).

As shown in Figure 1, SRMR and SRMR-CI values were separately calculated ($SRMR_{IDS_{ij}}$,

$SRMR_{ADS_{ij}}$, $SRMR\text{-}CI_{IDS_{ij}}$ and $SRMR\text{-}CI_{ADS_{ij}}$) for each pair of stimuli in IDS and ADS conditions

($S_{IDS_{ij}}$ and $S_{ADS_{ij}}$). Intelligibility of speech stimuli produced by talker j in the two conditions (ADS

and IDS) was summarized by averaging the relevant SRMR-related values over the 15 stimuli

spoken by this talker at each of these two conditions. Therefore, for talker j, four values were

calculated: (i) $SRMR_{IDS_j} = \frac{1}{i}\sum_{i=1}^{15} SRMR_{IDS_{ij}}$, (ii) $SRMR_{ADS_j} = \frac{1}{i}\sum_{i=1}^{15} SRMR_{ADS_{ij}}$, (iii)

$SRMR\text{-}CI_{IDS_j} = \frac{1}{i}\sum_{i=1}^{15} SRMR\text{-}CI_{IDS_{ij}}$, (iv) $SRMR\text{-}CI_{ADS_j} = \frac{1}{i}\sum_{i=1}^{15} SRMR\text{-}CI_{ADS_{ij}}$. These four values

for each talker j provided an approximation of the intelligibility of IDS and ADS stimuli spoken

by this talker, as heard by listeners with NH ($SRMR_{IDS_j}$ and $SRMR_{ADS_j}$) and by those with CIs

($SRMR\text{-}CI_{IDS_j}$ and $SRMR\text{-}CI_{ADS_j}$). These values were then used to address (1) how the

intelligibility of speech is affected by the two separate factors of caregivers' speaking styles (ADS

vs. IDS), as well as by (simulated) group hearing status (NH vs. CI), (2) the extent to which the

degree of impact of ADS-to-IDS speaking style modifications on intelligibility varies across

individual caregivers, and (3) how this variation is affected by (simulated) hearing status (NH vs.

CI).

F.   Statistical analysis

We constructed two Generalized Linear Mixed Models (GLMMs) in Matlab (using the *fitglme*

function) (Matuschek, Kliegl, Vasishth, Baayen, & Bates, 2017; Quené & van den Bergh, 2008)

in order to (1) identify whether changes in acoustic distinctiveness (i.e., MD) between IDS and

ADS due to CI noise vocoding were statistically significant, and (2) to examine the effect of

speaking style (IDS vs. ADS) and simulated hearing status (NH vs. CI), and any potential

interaction, on intelligibility of speech (as measured by SRMR and SRMR-CI to simulate NH and

CI hearing statuses, respectively). For the first GLMM, the speech degradation level (i.e., 4, 8, 12,

16, 22, and 32-channel, and unprocessed) was entered into the model as a fixed predictor, and MDs

were defined as the response variable. The individual female talkers and speech stimuli were

defined as random effects (intercepts) in the model to account for quality differences that might

exist due to talker- and stimulus-specific variation. A post hoc test using Tukey's multiple

comparison approach (*multcompare*, MATLAB, Mathworks) was conducted to examine

statistically significant mean differences for all possible pairwise comparisons across seven levels

of spectral degradation (4, 8, 12, 16, 22, 32-channels, and unprocessed). In the second GLMM,

factors of speaking style (IDS or ADS) and simulated hearing status (NH or CI) were entered into

the model as independent fixed variables; the estimated speech intelligibility (cf. SRMR or SRMR-

CI values) corresponded to the response variable in the model. Talkers and speech stimuli were

entered as random-effect intercepts in the model to account for quality and/or information

differences that might exist due to talker- and stimulus-specific variation.

Measures of central tendency and variability (i.e., mean, standard deviation, and range)

were calculated as descriptive statistics to evaluate the degree of variability across caregivers in

acoustic implementation of IDS and ADS. We further calculated coefficients of variance (CVs)

for MD differences between ADS and IDS conditions for both unprocessed and simulated CI

speech within a 22-channel noise-vocoder to examine whether variability across individual

caregivers in ADS-IDS acoustic distinctiveness decreased as a function of speech degradation

(unprocessed vs. 22-channel simulated CI speech). The same analysis was performed to examine

whether changes in speech intelligibility – due to an ADS-to-IDS style shift – were variable across

talkers and/or how variability changed as (simulated) listener hearing status changed (from NH to

CI). We further calculated Pearson correlation coefficients to examine whether changes in acoustic

distinctiveness and speech intelligibility due to changes in speaking style (ADS to IDS) across

degradation levels (unprocessed vs. 22-channel vocoded) and hearing status (NH vs. CI) involved

a linear transformation.

III. RESULTS

A. Effects of CI-related spectral degradation on ADS-vs-IDS acoustic distinctiveness

Fig. 2 shows the mean and distribution of MDs obtained as a function of signal degradation that

ranged from a no-degradation (i.e., unprocessed) condition to CI-simulated speech for which the

number of spectral channels in the noise-vocoder was gradually reduced from 32 to 4. Overall,

MDs between IDS and ADS monotonically decreased with decreasing numbers of spectral

channels in the noise-vocoder, thereby highlighting the detrimental effects of noise vocoding in CI

devices for signal fidelity. As illustrated in Fig. 2, the acoustic distinctiveness between IDS and

ADS is reduced by CI speech processing systems and further reduced through decreasing the

number of spectral channels. Even the highest-fidelity CI-simulated condition (32-channel noise-

vocoded speech) showed a substantial drop in acoustic distinctiveness between IDS and ADS,

relative to the unprocessed condition. The average MD across the six noise-vocoded conditions

showed a decline of approximately 67% relative to MD at the unprocessed condition, indicating

that a large portion of acoustic information involved in conveying the distinction between IDS and

ADS was lost due to CI-related speech processing.

Table I represents the results of the GLMM analysis to test statistical significance of the

effect of CI-related speech processing on MD. As this table suggests, the detrimental effect of CI-

related processing, as captured in the decrease in the number of spectral channels on the

distinctiveness between IDS and ADS was statistically significant ($\beta = -0.886$, $t = -16.54$, $p <$

0.0001). Post hoc testing using a Tukey's multiple comparison approach revealed that the effect

of noise vocoding on the decrease in MDs was statistically significant for comparisons of all pairs

of speech degradation conditions except three (32 vs. 22 channels, 22 vs. 16 channels, and 16 vs.

12 channels).

444   **>>>>>>>>FIG. 2 about here <<<<<<<<**

445   **>>>>>>>>>>>>>>> TABLE I about here <<<<<<<<<<<<<<<**

B. Effects of CI-related speech processing on intelligibility of ADS and IDS

Fig. 3 shows average SRMR or SRMR-CI values estimating intelligibility of ADS and IDS stimuli

for NH and CI listeners. The bar graphs show the group data, corresponding to the means averaged

across talkers. The figure suggests that speech spoken in IDS is more intelligible than matched

utterances in ADS, regardless of whether intelligibility is estimated for listeners with NH or those

with CIs. On average, the stimuli spoken in IDS were ~15% more intelligible than those spoken

in ADS for listeners with normal hearing, as measured by the SRMR metric. This average

improvement in intelligibility of IDS stimuli over ADS stimuli was ~5.7% for listeners with CIs,

as estimated by SRMR-CI, suggesting that filtering speech through a CI speech processor

decreases the supportive effects of IDS for speech intelligibility improvement.

The results from the GLMM analysis for examining the effect of talkers' speaking style

(IDS vs. ADS), listeners' group (NH vs. CI), and their interaction on intelligibility of speech (Table

II) revealed a significant effect of speaking style ($\beta = 3.05$, $SE = 0.74$, $t = 4.14$, $p < 0.0001$). This

analysis provides new evidence that IDS may likely improve intelligibility of caregivers' speech

to both listeners with NH and those with CIs. These results further showed a significant, negative

effect of CI-related processing on intelligibility of caregivers' speech ($\beta = -2.23$, $SE = 0.74$, $t = -$

3.02, $p < 0.0001$). In addition, a significant interactive effect between speaking style and listeners

group was revealed by the GLMM model ($\beta = -1.337$, $SE = 0.466$, $t = -2.865$, $p < 0.001$), reflecting

the fact that hearing speech through CIs resulted in a smaller intelligibility gain from IDS relative

to ADS than was observed in the NH condition.

466   **>>>>>>>>FIG. 3 about here <<<<<<<<**

467   >>>>>>>>>>>>>> **TABLE II about here** <<<<<<<<<<<<<<<

C. Individual differences across caregivers in ADS-vs-IDS acoustic distinctiveness and the

effect of CI noise vocoding

To examine individual differences across talkers in acoustic distinctiveness of IDS from ADS, we

focused on MD data from the individual talkers in two conditions (unprocessed and 22-channel

CI-simulated noise-vocoded speech), as shown in Fig. 4A (left vs. right bars, respectively). Two

general observations are noticeable in this data. First, caregivers vary considerably in terms of the

amount of acoustic variability created by their ADS-to-IDS modification (mean $= 7.47$, $SD = 1.70$,

range $= 5.23$). For example, the ADS-to-IDS shift in MD for talker 2 exposes a child with NH (cf.

natural, unprocessed condition) to a much wider range of acoustic information compared to that

for talker 3. This, in turn, suggests that a child of talker 2 would be more advantaged in learning

acoustic patterns of her caregivers' speech compared to a child of talker 3.

The second observation is that the magnitude of this acoustic variability significantly

reduced for CI-simulated signals, due to the noise vocoding process (mean $= 3.22$, SD $= 0.43$,

range $= 1.4$). An analysis of coefficient of variation (CV) revealed that the standard deviation of

MD for unprocessed (natural) speech stimuli was 22.7% of its mean, whereas this value was 13.3%

for CI-simulated speech stimuli for the 22-channel noise-vocoder. Comparing these two CV values

shows that the magnitude of variation in ADS-IDS acoustic distinctiveness (i.e., MD) reduces

approximately by half due to CI-related speech processing, highlighting a major loss of acoustic

information useful for recognition of caregivers themselves and for distinguishing them from other

caregivers.

If CI-simulated processing involves a *linear* transformation of acoustic distinctiveness

across speech styles for each talker, then we should observe MD differences to be significantly

correlated across natural (i.e., unprocessed) and 22-channel noise-vocoded conditions. The

Pearson's correlation coefficient was $r(5) = 0.07$ ($p = 0.87$), suggesting that CI-simulated noise

vocoding did not impact talkers' speech proportionately in terms of the acoustic distinctiveness

between their IDS and ADS. Instead, CI-related speech processing reduced the IDS-ADS acoustic

distance for some talkers (e.g., talker 6) more than others (e.g., talker 4).

495   >>>>>>>>**FIG. 4 about here** <<<<<<<<

496   **D. Individual differences across caregivers in IDS and ADS intelligibility and the effect of**

497   **hearing status**

498   To examine how individual talkers vary in the effects of their ADS-to-IDS modifications

on intelligibility of speech to listeners with NH and those with CIs, we focused on patterns of

SRMR and SRMR-CI metrics from the individual talkers. For (simulated) listeners with NH,

SRMRs of pairs of IDS and ADS stimuli were subtracted for each talker and averaged over 15

pairs of ADS-IDS stimuli in order to measure the amount of change in her speech intelligibility

due to ADS-to-IDS modification (for talker j, $\Delta SRMR_{(IDS\text{-}ADS)_j} = \frac{1}{i}\sum_{i=1}^{15}(SRMR_{IDS_{ij}} - SRMR_{ADS_{ij}})$).

Likewise, for (simulated) listeners with CI, SRMR-CIs of pairs of IDS and ADS stimuli were

subtracted for each talker and averaged over 15 pairs of ADS-IDS stimuli in order to quantify the

amount of change in her speech intelligibility due to ADS-to-IDS modification (for talker j,

$\Delta SRMR\text{-}CI_{(IDS\text{-}ADS)_j} = \frac{1}{i}\sum_{i=1}^{15}(SRMR\text{-}CI_{IDS_{ij}} - SRMR\text{-}CI_{ADS_{ij}})$). Thus, a positive value of this

difference measure represents IDS being, on average, more intelligible than ADS for a talker,

showing an intelligibility benefit for IDS over ADS for that talker.

510   Figure 4B (right panel) shows the results for changes in intelligibility of speech of talkers

due to modifications to speaking style. The bar plot on the left in Figure 4B (labeled as NH on the

x-axis) presents the difference across seven talkers in IDS-ADS intelligibility as estimated for

listeners with NH. The data in this scatter plot shows how much speech of a talker (e.g., talker 6)

would be heard as more or less intelligible as the talker speaks the utterances or words in IDS

compared to ADS.

Two patterns are observable in Figure 4B. First, the change in intelligibility of talkers'

speech due to the speaking style modification (ADS to IDS) was fairly variable across talkers for

simulated listeners with NH (mean = 1.71, SD = 1.73, range = 4.97), such that this change was

associated with relatively more intelligible speech for some talkers compared to others. As this

plot suggests, when talkers change their speaking style from ADS to IDS, not only does this not

lead to an equal amount of change in intelligibility of speech across talkers, but for one talker the

direction of this effect is even slightly negative (talker 5). For the majority of talkers (6 out of 7

talkers), IDS was more intelligible than ADS (as shown by a positive difference value).

Second, relative to the size of the ADS-to-IDS shift for simulated NH listeners, the size of

this ADS-to-IDS shift was overall considerably reduced for simulated listeners with CIs (mean =

0.38, SD = 0.43, range = 1.10). The bar plot on the right panel of figure 4B (labeled as CI on the

x-axis) presents the dispersion of change in speech intelligibility for each talker due to changing

speaking style from ADS to IDS, as estimated for listeners with CIs. Notably, CI-related

processing decreases the degree of variability in speech intelligibility, compared with that for NH,

as expected. An analysis of CV showed that although the mean of change in speech intelligibility

due to ADS-to-IDS modifications was overall smaller for the simulated CI condition compared to

the simulated NH condition, while the magnitude of variability was almost the same for two groups

of listeners (CV $\cong$ 11%). Highlighting non-linearity of effects of CI processing on speech across

different talkers, there was no significant correlation across the seven talkers for ADS-to-IDS

intelligibility differences in simulated CI vs. NH listening conditions ($r(5) = 0.74$, $p = 0.58$).

IV. DISCUSSIONS

The present study investigated effects of CI-related signal processing on acoustic distinctiveness

and intelligibility of speech signals as a function of IDS and ADS speaking styles. Results from

the present study supported the hypothesis that CI-related speech processing significantly degrades

acoustic information involved in distinguishing IDS from ADS. This significant loss of acoustic

information related to the distinction between IDS and ADS may negatively impact infants'

recognizing IDS as distinct from ADS, leading possibly to less attention to caregivers' speech,

with potential consequences for children with CIs developing relatively poorer language skills

compared to those with NH.

These results are in line with prior findings demonstrating difficulties by listeners with CIs

in distinguishing among various speaking styles (Tamati, Janse, & Başkent, 2019). Distinguishing

caregivers' speaking styles is tied to advances in language development (Karzon, 1985; Singh &

Nestor, 2009; Thiessen et al., 2005; Wang et al., 2017), but is made challenging when speech is

degraded by CI-related speech processing and/or other undesirable sources, e.g., noise and

reverberation (Fetterman & Domico, 2002; Hazrati & Loizou, 2012; Zheng, Koehnke, & Besing,

2011). The substantial elimination of IDS-related acoustic information suggested by the present

study indicates that infants with CIs likely do not have access to as wide a range of spectro-

temporal information in caregivers' speech to foster recognizing when speech is directed to them.

As a result, these children may experience considerable difficulties in recognizing IDS as distinct

from ADS, particularly in complex linguistic environments.

Although Wang et al., (2017) showed that IDS enhanced attention to speech in infants with

CIs compared to ADS, results from the present study suggest that infants' capability to recognize

IDS from ADS and to prefer attending to IDS over ADS is not probably comparable to that of

children with NH and is expected to be largely reduced. Furthermore, the fact that infants with CIs

showed this preference after 12 months of CI experience in Wang et al.'s (2017) study highlights

the possibility that the time course of developing certain capabilities for differentiating IDS and

ADS would be longer than peers with NH and would depend on the amount of experience with

CIs.

Our results also suggest that infants with CIs must develop a different cue-weighting

system for recognition of IDS from ADS compared to infants with NH, where the type and

magnitude of the relevant acoustic cues would be different from what infants with NH develop.

For example, in the absence of prominent acoustic cues of IDS such as F_0 (Fernald & Mazzie,

1991; Kuhl & Meltzoff, 1999; Mehler, 1981), which is poorly perceived through CI devices

(Mehta et al., 2020; Mehta & Oxenham, 2017; Qin & Oxenham, 2003), infants with CIs may rely

more on suprasegmental cues such as speech rate (Peng et al., 2017) and/or periodicity in the

temporal envelope (Fu, Chinchilla, & Galvin, 2004; Kong et al., 2004). However, perception of

pitch through periodicity in the temporal envelope is mostly limited to F_0s below around 300 Hz

(Carlyon, Deeks, & McKay, 2010; Kong & Carlyon, 2010), which is generally below the range of

F_0 variation in IDS. This lack of access to the entire spectro-temporal cue range in IDS puts infants

with CIs at high risk for missing IDS-related communicative events and thus for developing sub-

optimal language skills.

Our results corroborated our hypothesis that caregivers likely expose infants to more

intelligible speech when speaking in IDS compared to ADS, which is a novel contribution of the

present study. This was shown using a novel application of a recently-developed metric of

intelligibility, which models intelligibility of speech for listeners with NH (i.e., SRMR) and those

with CIs (i.e., SRMR-CI) (Falk et al., 2015; Santos et al., 2013). This positive effect of IDS on

speech intelligibility might be because caregivers provide more clear speech by speaking louder,

slower, and/or in a hyperarticulated fashion when using IDS (Hazan et al., 2018; Krause & Braida,

2002; Li et al., 2011; Liu, Del Rio, Bradlow, & Zeng, 2004), likely leading to an intelligibility

benefit compared to ADS (Janse, Nooteboom, & Quené, 2007; Liu et al., 2004). Greater

intelligibility for IDS over ADS suggests that use of IDS during caregiver-infant spoken

communication assists infants in better understanding speech, conceivably supporting infants'

word learning process through a direct link between exposure to IDS and improved speech

intelligibility. However, our results also revealed an interactive effect, showing that the size of this

positive effect of IDS on caregivers' speech intelligibility is smaller for infants with CIs compared

to their peers with NH. This, in turn, suggests that the CI noise vocoding process not only disrupts

the communication of IDS-ADS-specific acoustic information, but also decreases the intelligibility

of caregivers' IDS, indicating that infants with CIs are not expected to benefit from exposure to

IDS as much as their NH peers. These results are consistent with prior findings demonstrating

greater difficulties that adult listeners with CIs in processing speech and accomplishing lexical

access, compared to listeners with NH (McMurray, Farris-Trimble, & Rigler, 2017; Nagels,

Bastiaanse, Başkent, & Wagner, 2020). Although prior studies demonstrated a supportive role of

IDS on infants' language outcomes, to our knowledge, this is the first study that provides evidence

supporting detrimental effects of the noise vocoding process entailed in CI signal processing on

the supportive role of IDS in speech understanding.

Prior studies on speech recognition in listeners with CIs have highlighted the importance

of further examining patterns of individual differences, in addition to group differences (Dilley et

al., 2020; Nagels et al., 2020; Peng et al., 2019; Spencer, 2004; Szagun & Schramm, 2016). Our

investigation of patterns of individual differences across caregivers in acoustic distinctiveness

between IDS and ADS and the simulated effect of CI noise vocoding highlighted multiple

important observations. The first observation relates to the amount of acoustic variability produced

by each talker due to changing speaking style from ADS to IDS. The amount of resultant acoustic

variability varied across talkers, indicating that infants would experience different language

environments in terms of qualities of their caregivers' speech. This is expected to result in some

caregivers' exposing infants to relatively larger ranges of acoustic information, as compared to

others, which would foster their infants' speech processing and language acquisition by assisting

infants in more robust recognition of their caregivers' voices (Beauchemin et al., 2011; Kisilevsky

et al., 2003), as well as better development of auditory cortical processing for language

development (Webb et al., 2015). Similar to the recent findings by Dilley et al. (2020), these results

also suggest that IDS is not always readily distinguishable from ADS, due to the fact that

caregivers vary in implementation of IDS. As such, these results confirm that conditions for

recognition of IDS by infants is not always optimal (Piazza et al., 2017) and might differentially

affect language outcomes for infants with CIs (Dilley et al., 2020).

More importantly, the amount of ADS-IDS acoustic difference was considerably reduced

for each caregiver as her speech passed through a CI simulator, indicating that infants with CIs

probably have access to a much narrower range of acoustic variability in the voices of their

caregivers, as compared to their peers with NH. This large decline in the degree of acoustic

variability in caregivers' voices may negatively impact infants' robust identification of their

caregivers' voices, thereby preventing their gaining the maximal benefit from language input that

may be available in their linguistic environments. In fact, encoding and learning these cues is

crucial as they both significantly contribute to infants' familiarity with ranges of acoustic variation

in talkers' speech and their robust performance in understanding speech, despite of multiple

sources of variability, such as talker variation (Allen, Miller, & DeSteno, 2009; Eskenazi, 1993),

language context (Mattys, 2000; McMurray & Aslin, 2005; Miller, 1994), speech rate (Sommers,

Nygaard, & Pisoni, 1992), and background noise and/or reverberation (Hawkins, 2004). Thus,

infants with CIs are probably at risk for partial learning of subtle voice cues specific to their

caregivers, as reflected in their fine-structure spectro-temporal information.

Acoustic implementation of ADS-to-IDS modification varies across individual caregivers,

which may differentially impact infants' understanding of speech and thus their language

outcomes (Dilley et al., 2020; Hoff, 2006; Weisleder & Fernald, 2013). Results from the present

study showed that individual talkers varied in the effect of their speaking style modification on

speech intelligibility, suggesting that modification of speaking style from ADS to IDS does not

always cause an equal degree of improvement in intelligibility of caregivers' speech. Notably, our

results showed that this variability across caregivers in impact of their speaking style modification

on speech intelligibility was reduced due to cochlear implant-related speech processing. In the

present study, we only studied variability across caregivers in acoustic information and

intelligibility of their IDS compared to ADS, whereas caregivers of infants with CIs may vary in

other aspects of spoken communication such as gestural and proprioceptive behaviors, which very

likely change the degree of intelligibility by which infants eventually perceive caregivers' speech

(Kirk et al., 2007; Kirk & Pisoni, 2002; Lachs, Pisoni, & Kirk, 2001).

These findings can be further interpreted in the context of spectro-temporal information

available to infants' auditory systems, which is very sensitive to subtle changes in speech acoustics

(Jusczyk, Hohne, & Bauman, 1999; Kuhl, 2004). When infants' accessibility to fine-grained

spectro-temporal structures in speech is largely compromised because of a limited number of

vocoder channels in CIs, they may not be able to readily recognize and attend to rich IDS. In the

absence of this fine-structure spectral information at the output of CI electrodes, infants must rely

on course-grained cues in the speech envelope, such as temporal envelope periodicity (Green,

Faulkner, & Rosen, 2002; Moore, 2003) or cues from other sensory modalities (e.g., visual and

tactile, Green, Nip, Wilson, Mefferd, & Yunusova, 2010; Rohlfing, Fritsch, Wrede, & Jungmann,

2006) in order to recognize IDS from ADS. This increases the cognitive load in processing speech

particularly in complex auditory environments and negatively contributes to the observed poor

language outcomes in some children (Davidson, Geers, & Nicholas, 2014; Dunn et al., 2014;

Geers, Strube, Tobey, & Moog, 2011; Geers, Nicholas, Tobey, & Davidsonb, 2015; Houston,

Pisoni, Kirk, Ying, & Miyamoto, 2003; Houston, Stewart, Moberly, Hollich, & Miyamoto, 2012;

Miyamoto, Houston, Kirk, Perdew, & Svirsky, 2003; Niparko et al., 2010; Pisoni et al., 2007;

Pisoni, Kronenberger, Harris, & Moberly, 2018).

The present study used a within-talker manipulation of speech style to investigate how CI-

related speech processing may affect processing of speech style shifts which are tied to language

development, namely IDS vs. ADS signals. While the study involved a sizeable number of

individual utterances with control of phonetic properties within talkers, the amount of speech

collected from each of the seven talkers was relatively small. Further work will be needed to test

how these results extend to a larger sample of talkers with utterances with more varied segmental

and lexical composition. Furthermore, studying natural IDS and ADS would be ideal; however, it

is difficult to control the content and quality of the speech collected from naturalistic environments.

Importantly, studies based on simulated CI speech and computational models of human hearing

provide valuable evidence for understanding speech perception in listeners with normal and

impaired hearing (Litvak, Delgutte, & Eddington, 2001; Mehta et al., 2020; Rubinstein, Wilson,

Finley, & Abbas, 1999; Throckmorton & Collins, 2002). Findings from these studies should be

viewed as evidence for general trends, rather than the actual performance of listeners with CIs,

subject to further investigation. Although we simulated limitation of CIs in spectral resolution by

changing the number of channels in noise-vocoded speech, the effect of other aspects of CI

processing on spectral resolution, such as channel interactions (i.e., steepness of spectral slope)

(Crew & Galvin, 2012; Mehta et al., 2020), were not simulated in this study. Additionally, metrics

of speech intelligibility of SRMR and SRMR-CI have been validated for adult listeners with CIs

and are not direct estimates of speech intelligibility, which is very likely different from what infants

with CIs experience in terms of recognition of utterances. Considering these limitations, the results

should not be taken as the final determination of how children with CIs perform in recognition

between IDS and ADS and how ADS-to-IDS modification impacts the degree of intelligibility of

caregivers' speech to children with CIs.

Despite these limitations, results from the present study suggest that, compared to infants

with NH, infants with CIs are disadvantaged in acquiring spoken language due to multiple factors.

First, partial transmission of spectro-temporal cues via CI speech processing decreases the

supportive role of IDS in infants' language learning by disrupting the link between attending to

IDS and speech comprehension. In fact, degraded representation of attention-related acoustic cues

in IDS (e.g., F_0-related acoustic information, Mehta et al., 2020; Mehta & Oxenham, 2017) through

CIs may have detrimental effects on infants' ability to pay attention to caregivers' speech,

something that is a fundamental cognitive skill for spoken language acquisition (Bergeson, 2014;

Glenn, Cunningham, & Joyce, 1981; Houston et al., 2003; Rottmann & Zobrist, 2004; Wang,

Shafto, & Houston, 2018). It is worth mentioning that infants' language learning involves

incorporating information from multiple interwind communication dimensions (i.e., visual, social,

tactile, and emotional), which creates a very rich channel for learning language through infant-

directed speech (Gogate, Bahrick, & Watson, 2000; Nomikou & Rohlfing, 2011), even in the

absence of major spectro-temporal cues such as talker's F_0 in the output of CI electrodes. Our

findings further suggested that CI speech processing diminishes the benefit of exposure to more

intelligible IDS, as compared to ADS, which may negatively impact infants' abilities to process

caregivers' speech, possibly with further significant downstream consequences for later language

skills. Last but not least, our results showed that, compared to infants with NH, infants with CIs

have access to a relatively narrower range of acoustic information (corresponding to smaller

acoustic variability) in their caregivers' speech, which probably leads to experiencing greater

difficulties in robust identification and recognition of their caregivers' voices (Beauchemin et al.,

2011; Kisilevsky et al., 2003; Lavan et al., 2019), as well as developing poorer word recognition

skills (Singh, 2008) and an impaired auditory system for language processing (Webb et al., 2015).

Despite the reduced ADS-vs-IDS distinctiveness, intelligibility, and variability of vocoded IDS, it

is possible that children with CIs may have developed certain coping/adapting strategies to

mitigate the degraded speech input. For example, it is possible that children with CIs may have

higher sensitivity (lower threshold) to the acoustic cues than children with NH.

In summary, the current study used computational approaches to signal processing in order

to provide new evidence for how CI-related speech processing may impact recognition of IDS

from ADS in children with CIs, as well as how these style differences may affect intelligibility

benefits derived by style shifts from ADS to IDS. These findings provide solid grounding for

developing new perceptual studies to test abilities of infants with CIs to recognize their caregivers'

speaking style (ADS vs. IDS) and to recognize intelligible words in caregivers' speech. Focusing

on computational metrics as undertaken here provides an important complement to costly,

complex, labor-intensive perceptual studies. The results provided preliminary evidence for how

CI-related speech processing may alter the pathway from exposure to IDS to processing speech

and, by extension, the acquisition of language by children with CIs compared to that of their NH

peers. The most direct and immediate implication of these findings is the imperative need to

improve signal processing in CI devices to assure the faithful transmission of acoustic cues

relevant to identification and recognition of IDS from ADS. Until then, the major clinical

implication of these findings is that the maximum benefit from exposure to IDS for language

learning in infants with CIs requires caregivers' active use of multimodal (i.e., gesture, tactile,

visual, social, and emotional) communicative behaviors in order to compensate for the degraded

representation of acoustic information relevant to IDS and to support its robust perception.

ACKNOWLEDGMENTS

Research reported in this publication was supported by the National Institute on Deafness and

other Communicative Disorders of the National Institutes of Health under award number

R01DC008581 to D. Houston and L. Dilley.

REFERENCES

Allen, J. S., Miller, J. L., & DeSteno, D. (2009). Individual talker differences in voice-onset-time:

Contextual influences. *The Journal of the Acoustical Society of America*, *125*(6), 3974–3982.

Arjmandi, M. K., Dilley, L. C., & Wagner, S. E. (2018). Investigation of Acoustic Dimension Use

in Dialect Production : Machine Learning of Sonorant Sounds for Modeling Acoustic Cues

of African American Dialect. *11th International Conference on Voice Physiology and*

Biomechanics, (August), 1–2.

Beauchemin, M., González-Frankenberger, B., Tremblay, J., Vannasing, P., Martínez-Montes, E.,

Belin, P., … Lassonde, M. (2011). Mother and stranger: An electrophysiological study of

voice processing in newborns. *Cerebral Cortex*, *21*(8), 1705–1711.

Bergeson, T. R. (2014). Hearing versus Listening: Attention to Speech and Its Role in Language

Acquisition in Deaf Infants with Cochlear Implants. *Lingua*, 10–25.

Boersma, P., & Weenink, D. (2001). Praat, a system for doing phonetics by computer. *Glot*

International, *5:9/10*, 341–345.

Bradlow, A. R., Kraus, N., & Hayes, E. (2003). Speaking clearly for children with learning

disabilities: Sentence perception in noise. *Journal of Speech, Language, and Hearing*

Research, *46*(1), 80–97.

Brown, C. A., & Bacon, S. P. (2010). Fundamental frequency and speech intelligibility in

background noise. *Hearing Research*, *266*(1–2), 52–59.

Burnham, E. B., Wieland, E. A., Kondaurova, M. V., McAuley, J. D., Bergeson, T. R., & Dilley,

L. C. (2015). Phonetic Modification of Vowel Space in Storybook Speech to Infants up to 2

Years of Age. *Journal of Speech, Language, and Hearing Research*, *58*(2), 241–253.

Carlyon, R. P., Deeks, J. M., & McKay, C. M. (2010). The upper limit of temporal pitch for

cochlear-implant listeners: Stimulus duration, conditioner pulses, and the number of

electrodes stimulated. *The Journal of the Acoustical Society of America*, *127*(3), 1469–1478.

Chatterjee, M., & Oberzut, C. (2011). Detection and rate discrimination of amplitude modulation

in electrical hearing. *The Journal of the Acoustical Society of America*, *130*(3), 1567–1580.

Chatterjee, M., & Peng, S. C. (2008). Processing F0 with cochlear implants: Modulation frequency

discrimination and speech intonation recognition. *Hearing Research*, *235*(1–2), 143–156.

Cooper, R. P., Abraham, J., Berman, S., & Staska, M. (1997). The development of infants'

preference for motherese. *Infant Behavior and Development*, *20*(4), 477–488.

Cooper, R. P., & Aslin, R. N. (1990). Preference for Infant-Directed Speech in the First Month

after Birth. *Child Development*, *61*(5), 1584–1595.

Crew, J. D., & Galvin, J. J. (2012). Channel interaction limits melodic pitch perception in

simulated cochlear implants. *The Journal of the Acoustical Society of America*, *132*(5),

EL429–EL435.

Cristia, A., & Seidl, A. (2014). The hyperarticulation hypothesis of infant-directed speech. *Journal

of Child Language*, *41*(4), 935.

Croghan, N. B. H., Duran, S. I., & Smith, Z. M. (2017). Re-examining the relationship between

number of cochlear implant channels and maximal speech intelligibility. *The Journal of the*

Acoustical Society of America, *142*(6), EL537–EL543.

Cutler, A., Dahan, D., & van Donselaar, W. (1997). Prosody in the Comprehension of Spoken

Language: A Literature Review. *Language and Speech*, *40*, 141–201.

Davidson, L. S., Geers, A. E., & Nicholas, J. G. (2014). The effects of audibility and novel word

learning ability on vocabulary level in children with cochlear implants. *Cochlear Implants*

International, *15*(4), 211–221.

Dilley, L., Lehet, M., Wieland, E. A., Arjmandi, M. K., Houston, D., Kondaurova, M., …

Bergeson, T. (2020). Individual differences in mothers' spontaneous infant-directed speech

predict language attainment in children with cochlear implants. *Journal of Speech, Language,*

and Hearing Research, *In press*.

Dorman, M. F., Loizou, P. C., Fitzke, J., & Tu, Z. (1998). The recognition of sentences in noise

by normal-hearing listeners using simulations of cochlear-implant signal processors with 6–

20 channels. *The Journal of the Acoustical Society of America*, *104*(6), 3583–3585.

Dorman, M. F., Loizou, P. C., & Rainey, D. (1997a). Simulating the effect of cochlear-implant

electrode insertion depth on speech understanding. *The Journal of the Acoustical Society of*

America, *102*(5), 2993–2996.

Dorman, M. F., Loizou, P. C., & Rainey, D. (1997b). Speech intelligibility as a function of the

number of channels of stimulation for signal processors using sine-wave and noise-band

outputs. *The Journal of the Acoustical Society of America*, *102*(4), 2403–2411.

Dunn, C. C., Walker, E. A., Oleson, J., Kenworthy, M., Voorst, T. Van, Tomblin, J. B., … Gantz,

B. J. (2014). Longitudinal speech perception and language performance in pediatric cochlear

implant users: The effect of age at implantation. *Ear and Hearing*, *35*(2), 148–160.

Eskenazi, M. (1993). Trends in speaking styles research. *Third European Conference on Speech*

Communication and Technology, 501–509.

Falk, T. H., Parsa, V., Santos, J. F., Arehart, K., Hazrati, O., Falk, T. H., … Scollie, S. (2015).

Objective Quality Prediction for Users of and Intelligibility Assistive Listening Devices.

IEEE Signal Processing Magazine, *32*(2), 114–124.

Ferguson, S. H., & Kewley-Port, D. (2007). Talker differences in clear and conversational speech:

Acoustic characteristics of vowels. *Journal of Speech, Language, and Hearing Research*,

50(5), 1241–1255.

Ferguson, S. H., Poore, M. A., Shrivastav, R., Kendrick, A., McGinnis, M., & Perigoe, C. (2010).

Acoustic Correlates of Reported Clear Speech Strategies. *Journal of the Academy of*

Rehabilitative Audiology, *43*, 45–64.

Ferguson, S. H., & Quené, H. (2014). Acoustic correlates of vowel intelligibility in clear and

conversational speech for young normal-hearing and elderly hearing-impaired listeners. *The*

Journal of the Acoustical Society of America, *135*(6), 3570–3584.

Fernald, A. (1985). Four-month-old infants prefer to listen to motherese. *Infant Behavior and*

Development, *8*(2), 181–195.

Fernald, A. (1989). Intonation and communicative intent in mothers' speech to infants: Is the

melody the message? *Child Development*, *60*(6), 1497–1510.

Fernald, A. (1993). Approval and Disapproval : Infant Responsiveness to Vocal Affect in Familiar

and Unfamiliar Languages. *Child Development*, *64*(3), 657–674.

Fernald, A., & Mazzie, C. (1991). Prosody and focus in speech to infants and adults.

Developmental Psychology, *27*(2), 209–221.

Fernald, A., & Simon, T. (1984). Expanded intonation contours in mothers' speech to newborns.

Developmental Psychology, *20*(1), 104–113.

Fernald, & Kuhl. (1987). Acoustic determinants of infant preference for motherse speech. *Infant*

Behaviour and Development, *10*, 279–293.

Fetterman, B. L., & Domico, E. H. (2002). Speech recognition in background noise of cochlear

implant patients. *Otolaryngology - Head and Neck Surgery*, *126*(3), 257–263.

Friesen, L. M., Shannon, R. V., Baskent, D., & Wang, X. (2001). Speech recognition in noise as a

function of the number of spectral channels: Comparison of acoustic hearing and cochlear

implants. *The Journal of the Acoustical Society of America*, *110*(2), 1150–1163.

Fu, Q.-J. (2019). *AngelSim: Cochlear implant and hearing loss simulator*. [Computer program],

available at http://www.tigerspeech.com/angelsim/angelsim_about.html

Fu, Q.-J., Chinchilla, S., & Galvin, J. J. (2004). The role of spectral and temporal cues in voice

gender discrimination by normal-hearing listeners and cochlear implant users. *Journal of the*

Association for Research in Otolaryngology, *5*(3), 253–260.

Fu, Q.-J., Chinchilla, S., Nogaki, G., & Galvin, J. J. (2005). Voice gender identification by cochlear

implant users: The role of spectral and temporal resolution. *The Journal of the Acoustical*

Society of America, *118*(3), 1711–1718.

Fu, Q.-J., & Nogaki, G. (2005). Noise susceptibility of cochlear implant users: The role of spectral

resolution and smearing. *Journal of the Association for Research in Otolaryngology*, *6*(1),

19–27.

Fu, Q.-J., Shannon, R. V., & Wang, X. (1998). Effects of noise and spectral resolution on vowel

and consonant recognition: Acoustic and electric hearing. *The Journal of the Acoustical*

Society of America, *104*(6), 3586–3596.

Galvin, K. L., Mok, M., Dowell, R. C., & Briggs, R. J. (2008). Speech detection and localization

results and clinical outcomes for children receiving sequential bilateral cochlear implants

842  before four years of age. *International Journal of Audiology*, *47*(10), 636–646.

843 Geers, A. E., Strube, M. J., Tobey, E. A., & Moog, J. S. (2011). Epilogue: factors contributing to

844  long-term outcomes of cochlear implantation in early childhood. *Ear and Hearing*, *32*(1

845  Suppl), 84S.

846 Geers, Ann E., Nicholas, J., Tobey, E., & Davidsonb, L. (2016). Persistent Language Delay Versus

847  Late Language Emergence in Children With Early Cochlear Implantation. *Journal of Speech,*

848  *Language, and Hearing Research*, *59*(1), 155–170.

849 Glenn, S. M., Cunningham, C. C., & Joyce, P. F. (1981). A Study of Auditory Preferences in

850  Nonhandicapped Infants and Infants with down ' s Syndrome. *Child Development*, 1303–

851  1307.

852 Gogate, L. J., Bahrick, L. E., & Watson, J. D. (2000). A study of multimodal motherese: The role

853  of temporal synchrony between verbal labels and gestures. *Child Development*, *71*(4), 878–

854  894.

855 Green, J. R., Nip, I. S. B., Wilson, E. M., Mefferd, A. S., & Yunusova, Y. (2010). Lip movement

856  exaggerations during infant-directed speech. *Journal of Speech, Language, and Hearing*

857  *Research*, *53*(6), 1529–1542.

858 Green, T., Faulkner, A., & Rosen, S. (2002). Spectral and temporal cues to pitch in noise-excited

859  vocoder simulations of continuous-interleaved-sampling cochlear implants. *The Journal of*

860  *the Acoustical Society of America*, *112*(5), 2155–2164.

861 Greenwood, D. D. (1990). A cochlear frequency-position function for several species—29 years

862  later. *Journal of the Acoustical Society of America*, *87*(6), 2592–2605.

Grieco-calub, T. M., Saffran, J. R., & Litovsky, R. Y. (2010). Spoken Word Recognition in

Toddlers Who Use Cochlear Implants. *Journal of Speech Language and Hearing Research*,

52(6), 1390–1400.

Hawkins, S. (2004). Puzzles and patterns in 50 years of research on speech perception. *From Sound*

to Sense: 50+ Years of Discoveries in Speech Communication, 223–246.

Hazan, V., Tuomainen, O., Kim, J., Davis, C., Sheffield, B., & Brungart, D. (2018). Clear speech

adaptations in spontaneous speech produced by young and older adults. *The Journal of the*

Acoustical Society of America, *144*(3), 1331–1346.

Hazrati, O., & Loizou, P. C. (2012). The combined effects of reverberation and noise on speech

intelligibility by cochlear implant listeners. *International Journal of Audiology*, *51*(6), 437–

443.

Heijden, V. Der, Ferdinand, R. P. D., Ridder, D. De, & Tax., D. M. (2005). *Classification,*

Parameter Estimation and State Estimation An Engineering Approach Using MATLAB. John

Wiley & Sons.

Hoff, E. (2006). How social contexts support and shape language development. *Developmental*

Review, *26*(1), 55–88.

Horowitz, F. D. (1983). The Effects of Intonation on Infant Attention: The Role of the Rising

Intonation Contour. *Journal of Child Language*, *10*(3), 521–534.

Houston, D. M., Pisoni, D. B., Kirk, K. I., Ying, E. A., & Miyamoto, R. T. (2003). Speech

perception skills of deaf infants following cochlear implantation: A first report. *International*

Journal of Pediatric Otorhinolaryngology, *67*(5), 479–495.

Houston, D. M., Stewart, J., Moberly, A., Hollich, G., & Miyamoto, R. T. (2012). Word learning

in deaf children with cochlear implants: Effects of early auditory experience. *Developmental*

Science, *15*(3), 448–461.

Hunt, N. J., Lennig, N., & Mermeletein, P. (1980). Experiments in syllable-based recognition of

continuous speech. *IEEE International Conference on Acoustics, Speech, and Signal*

Processing, (5), 880–883.

Imai, S. (1983). Cepstral analysis synthesis on the mel frequency scale. *IEEE International*

Conference on Acoustics, Speech, and Signal Processing, 93–96.

Inoue, T., Nakagawa, R., Kondou, M., Koga, T., & Shinohara, K. (2011). Discrimination between

mothers' infant- and adult-directed speech using hidden Markov models. *Neuroscience*

Research, *70*(1), 62–70.

Jahn, K. N., DiNino, M., & Arenberg, J. G. (2019). Reducing Simulated Channel Interaction

Reveals Differences in Phoneme Identification Between Children and Adults With Normal

Hearing. *Ear and Hearing*, *40*(2), 295–311.

Jain, S., & Vipin Ghosh, P. G. (2018). Acoustic simulation of cochlear implant hearing: Effect of

manipulating various acoustic parameters on intelligibility of speech. *Cochlear Implants*

International, *19*(1), 46–53.

Janse, E., Nooteboom, S. G., & Quené, H. (2007). Coping with gradient forms of /t/-deletion and

lexical ambiguity in spoken word recognition. In *Language and Cognitive Processes* (Vol.

22).

Jusczyk, P. W., Hohne, E. A., & Bauman, A. (1999). Infants' sensitivity to allophonic cues for

word segmentation. *Perception and Psychophysics*, *61*(8), 1465–1476.

Karzon, R. G. (1985). Discrimination of polysyllabic sequences by one- to four-month-old infants.

Journal of Experimental Child Psychology, *39*(2), 326–342.

Kirk, K. I., Hay-McCutcheon, M. J., Holt, R. F., Gao, S., Qi, R., & Gerlain, B. L. (2007).

Audiovisual spoken word recognition by children with cochlear implants. *Audiological*

Medicine, *5*(4), 250–261.

Kirk, K. I., & Pisoni, D. B. (2002). Audiovisual integration of speech by children and adults with

cochlear implants. *International Conference on Spoken Language Processing*, 1689.

Kisilevsky, B. S., Hains, S. M., Lee, K., Xie, X., Huang, H., Ye, H. H., … Wang, Z. (2003). Effects

of experience on fetal voice recognition. *Psychological Science*, *14*(3), 220–224.

Kitamura, C., Thanavishuth, C., Burnham, D., & Luksaneeyanawin, S. (2001). Universality and

specificity in infant-directed speech: Pitch modifications as a function of infant age and sex

in a tonal and non-tonal language. *Infant Behavior and Development*, *24*(4), 372–392.

Kong, Y.-Y., & Carlyon, R. P. (2010). Temporal pitch perception at high rates in cochlear

implants. *The Journal of the Acoustical Society of America*, *127*(5), 3114–3123.

Kong, Y. Y., Cruz, R., Jones, J. A., & Zeng, F. G. (2004). Music Perception with Temporal Cues

in Acoustic and Electric Hearing. *Ear and Hearing*, *25*(2), 173–185.

Kong, Y. Y., Stickney, G. S., & Zeng, F.-G. (2005). Speech and melody recognition in binaurally

combined acoustic and electric hearing. *The Journal of the Acoustical Society of America*,

117(3), 1351–1361.

Krause, J. C., & Braida, L. D. (2002). Investigating alternative forms of clear speech: The effects

of speaking rate and speaking mode on intelligibility. *The Journal of the Acoustical Society*

of America, *112*(5), 2165–2172.

Kuhl, P. K. (2004). Early language acquisition: cracking the speech code. *Nature Reviews*

Neuroscience, *5*(11), 831–843.

Kuhl, P. K., Andruski, J. E., Chistovich, I. A., Chistovich, L. A., Kozhevnikova, E. V., Ryskina,

V. L., … Lacerda, F. (1997). Cross-language analysis of phonetic units in language addressed

to infants. *Science*, *277*(5326), 684–686.

Kuhl, P. K., & Meltzoff, A. N. (1999). The intermodal representation of speech in newborns.

Developmental Science, *2*(1), 42–46.

Lachs, L., Pisoni, D. B., & Kirk, K. I. (2001). Use of audiovisual information in speech perception

by prelingually deaf children with cochlear implants: A first report. *Ear and Hearing*, *22*(3),

236–251.

Lavan, N., Burton, A. M., Scott, S. K., & McGettigan, C. (2019). Flexible voices: Identity

perception from variable vocal signals. *Psychonomic Bulletin and Review*, *26*(1), 90–102.

Leong, V., Kalashnikova, M., Burnham, D., & Goswami, U. (2014). Infant-directed speech

enhances temporal rhythmic structure in the envelope. *Proceedings of the Annual Conference*

of the International Speech Communication Association, INTERSPEECH, (September),

2563–2567.

Li, Y., Zhang, G., Kang, H., Liu, S., Han, D., & Fu, Q.-J. (2011). Effects of speaking style on

speech intelligibility for Mandarin-speaking cochlear implant users. *The Journal of the*

Acoustical Society of America, *129*(6), EL242–EL247.

Litvak, L., Delgutte, B., & Eddington, D. (2001). Auditory nerve fiber responses to electric

stimulation: Modulated and unmodulated pulse trains. *The Journal of the Acoustical Society*

of America, *110*(1), 368–379.

Liu, H. M., Kuhl, P. K., & Tsao, F. M. (2003). An association between mothers' speech clarity

and infants' speech discrimination skills. *Developmental Science*, *6*(3), 1–10.

Liu, S., Del Rio, E., Bradlow, A. R., & Zeng, F.-G. (2004). Clear speech perception in acoustic

and electric hearing. *The Journal of the Acoustical Society of America*, *116*(4), 2374–2383.

Loizou, P. C., Dorman, M., & Tu, Z. (1999). On the number of channels needed to understand

speech. *The Journal of the Acoustical Society of America*, *106*(4), 2097–2103.

Luo, H., & Poeppel, D. (2007). Phase Patterns of Neuronal Responses Reliably Discriminate

Speech in Human Auditory Cortex. *Neuron*, *54*(6), 1001–1010.

Luo, X., Fu, Q.-J., & Galvin, J. J. (2007). Vocal Emotion Recognition by Normal-Hearing

Listeners and Cochlear Implant Users. *Trends in Amplification*, *11*(4), 301–315.

Ma, W., Golinkoff, R. M., Houston, D. M., & Hirsh-Pasek, K. (2011). Word learning in infant-

and adult-directed speech. *Language Learning and Development*, *7*(3), 185–201.

Maesschalck, R. D., & Massart, D. L. (2000). The Mahalanobis distance. *Chemometrics and*

Intelligent Laboratory Systems, *50*(1), 1–18.

Masnan, M. J., Mahat, N. I., Shakaff, A. Y. M., Abdullah, A. H., Zakaria, N. Z. I., Yusuf, N., …

Aziz, A. H. A. (2015). Understanding Mahalanobis distance criterion for feature selection.

AIP Conference Proceedings, *1660*(February 2015), 050075.

Mastropieri, D., & Turkewitz, G. (1999). Prenatal experience and neonatal responsiveness to vocal

expressions of emotion. *Developmental Psychobiology*, *35*(3), 204–214.

Mattys, S. L. (2000). The perception of primary and secondary stress in English. *Perception and*
Psychophysics, *62*(2), 253–265.

Matuschek, H., Kliegl, R., Vasishth, S., Baayen, H., & Bates, D. (2017). Balancing Type I error
and power in linear mixed models. *Journal of Memory and Language*, *94*, 305–315.

Maye, J., Werker, J. F., & Gerken, L. A. (2002). Infant sensitivity to distributional information can
affect phonetic discrimination. *Cognition*, *82*(3), 101–111.

McMurray, B., & Aslin, R. N. (2005). Infants are sensitive to within-category variation in speech
perception. *Cognition*, *95*(2), B15–B26.

McMurray, B., Farris-Trimble, A., & Rigler, H. (2017). Waiting for lexical access: Cochlear
implants or severely degraded input lead listeners to process speech less incrementally.
Cognition, *169*(August), 147–164.

McMurray, B., Kovack-Lesh, K. A., Goodwin, D., & McEchron, W. (2013). Infant directed speech
and the development of speech perception: Enhancing development or an unintended
consequence? *Cognition*, *85*(0 1), 1–27.

Mehler, J. (1981). The role of syllables in speech processing: infant and adult data. *Philosophical*
Transactions of the Royal Society of London. B, Biological Sciences, *295*(1077), 333–352.

Mehta, A. H., Lu, H., & Oxenham, A. J. (2020). The Perception of Multiple Simultaneous Pitches
as a Function of Number of Spectral Channels and Spectral Spread in a Noise-Excited
Envelope Vocoder. *JARO - Journal of the Association for Research in Otolaryngology*, *21*(1),
61–72.

Mehta, A. H., & Oxenham, A. J. (2017). Vocoder Simulations Explain Complex Pitch Perception

Limitations Experienced by Cochlear Implant Users. *JARO - Journal of the Association for*

Research in Otolaryngology, *18*(6), 789–802.

Miller, J. L. (1994). On the internal structure of phonetic categories: a progress report. *Cognition*,

50(1–3), 271–285.

Miyamoto, R. T., Houston, D. M., Kirk, K. I., Perdew, A. E., & Svirsky, M. A. (2003). Language

development in deaf infants following cochlear implantation. *Acta Oto-Laryngologica*,

123(2), 241–244.

Moore, B. C. J. (2003). Coding of sounds in the auditory system and its relevance to signal

processing and coding in cochlear implants. *Otology and Neurotology*, *24*(2), 243–254.

Moore, D. S., Spence, M. J., & Katz, G. S. (1997). Six-month-olds' categorization of natural

infant-directed utterances. *Developmental Psychology*, *33*(6), 980–989.

Nagels, L., Bastiaanse, R., Başkent, D., & Wagner, A. (2020). Individual differences in lexical

access among cochlear implant users. *Journal of Speech, Language, and Hearing Research*,

63(1), 286–304.

Narayan, C. R., & McDermott, L. C. (2016). Speech rate and pitch characteristics of infant-directed

speech: Longitudinal and cross-linguistic observations. *The Journal of the Acoustical Society*

of America, *139*(3), 1272–1281.

Niparko, J. K., Tobey, E. A., Thal, D. J., Eisenberg, L. S., Wang, N.-Y., Quittner, A. L., & Fink,

N. E. (2010). Spoken language development in children following cochlear implantation.

Jama, *303*(15), 1498–1506.

Nomikou, I., & Rohlfing, K. J. (2011). Language does something: Body action and language in

maternal input to three-month-olds. *IEEE Transactions on Autonomous Mental Development*,

3(2), 113–128.

Papoušek, M., Bornstein, M. H., Nuzzo, C., Papoušek, H., & Symmes, D. (1990). Infant responses

to prototypical melodic contours in parental speech. *Infant Behavior and Development*, *13*(4),

539–545.

Peng, S. C., Lu, H. P., Lu, N., Lin, Y. S., Deroche, M. L. D., & Chatterjee, M. (2017). Processing

of acoustic cues in lexical-tone identification by pediatric cochlear-implant recipients.

Journal of Speech, Language, and Hearing Research, *60*(5), 1223–1235.

Peng, S. C., Tomblin, J. B., & Turner, C. (2008). Production and Perception of Speech Intonation

in Pediatric Cochlear Implant Recipients and Individuals with Normal Hearing. *Ear and*

Hearing, *29*(3), 336–351.

Peng, Z., Hess, C., Saffran, J. R., Edwards, J. R., & Litovsky, R. Y. (2019). Assessing Fine-Grained

Speech Discrimination in Young Children With Bilateral Cochlear Implants. *Otology &*

Neurotology, *40*(3), e191–e197.

Peter, J. W., & Robert S, S. (2008). The effect of fundamental frequency on the intelligibility of

speech with flattened intonation contours. *American Journal of Speech-Language Pathology*,

17(4), 348–355.

Piazza, E. A., Iordan, M. C., & Lew-Williams, C. (2017). Mothers Consistently Alter Their Unique

Vocal Fingerprints When Communicating with Infants. *Current Biology*, *27*(20), 3162-

3167.e3.

Pisoni, D. B., Conway, C. M., Kronenberger, W., Horn, D. L., Karpicke, J., & Henning, S. (2007).

Efficacy and effectiveness of cochlear implants in deaf children. *Research on Spoken*

Language Processing, *28*(28), 3–46.

Pisoni, D. B., Kronenberger, W. G., Harris, M. S., & Moberly, A. C. (2018). Three challenges for

future research on cochlear implants. *World Journal of Otorhinolaryngology - Head and Neck*

Surgery, *3*(4), 240–254.

Qin, M. K., & Oxenham, A. J. (2003). Effects of simulated cochlear-implant processing on speech

reception in fluctuating maskers. *The Journal of the Acoustical Society of America*, *114*(1),

446–454.

Qin, M. K., & Oxenham, A. J. (2005). Effects of Envelope-Vocoder Processing on F0

Discrimination. *Ear & Hearing*, *26*(5), 451–460.

Quené, H., & van den Bergh, H. (2008). Examples of mixed-effects modeling with crossed random

effects and with binomial data. *Journal of Memory and Language*, *59*(4), 413–425.

Rohlfing, K. J., Fritsch, J., Wrede, B., & Jungmann, T. (2006). How can multimodal cues from

child-directed interaction reduce learning complexity in robots? *Advanced Robotics*, *20*(10),

1183–1199.

Rosen, S., Faulkner, A., & Wilkinson, L. (1999). Adaptation by normal listeners to upward spectral

shifts of speech: Implications for cochlear implants. *The Journal of the Acoustical Society of*

America, *106*(6), 3629–3636.

Rottmann, N., & Zobrist, P. (2004). Tuned to the signal: The privileged status of speech for young

infants. *Developmental Science*, *7*(3), 270–276.

Rubinstein, J. T., Wilson, B. S., Finley, C. C., & Abbas, P. J. (1999). Pseudospontaneous activity:

Stochastic independence of auditory nerve fibers with electrical stimulation. *Hearing*

Research, *127*(1–2), 108–118.

Rubinstein, Jay T. (2004). How cochlear implants encode speech. *Current Opinion in*

Otolaryngology and Head and Neck Surgery, *12*(5), 444–448.

Santos, J. F., Cosentino, S., Hazrati, O., Loizou, P. C., & Falk, T. H. (2013). Objective speech

intelligibility measurement for cochlear implant users in complex listening environments.

Speech Communication, *55*(7–8), 815–824.

Sato, N., & Obuchi, Y. (2007). Emotion Recognition using Mel-Frequency Cepstral Coefficients.

Information and Media Technologies, *2*(3), 835–848.

Schachner, A., & Hannon, E. E. (2011). Infant-Directed Speech Drives Social Preferences in 5-

Month-Old Infants. *Developmental Psychology*, *47*(1), 19–25.

Shaneh, M., & Taheri, A. (2009). Voice command recognition system based on MFCC and VQ

algorithms. *World Academy of Science, Engineering and Technology*, *57*, 534–538.

Shannon, R. V., Zeng, F.-G., Kamath, V., Wygonski, J., & Ekelid, M. (1995). Speech recognition

with primarily temporal cues. *Science*, *270*(5234), 303–304.

Singh, L. (2008). Influences of high and low variability on infant word recognition. *Cognition*,

106(2), 833–870.

Singh, L., Morgan, J. L., & Best, C. T. (2002). Infants' listening preferences: Baby talk or happy

talk? *Infancy*, *3*(3), 365–394.

Singh, L., & Nestor, S. (2009). Influences of Infant-Directed Speech on Early Word Recognition.

Infancy, *14*(6), 654–666.

Slaney, M., & Mcroberts, G. (1998). BABY EARS: A recognition system for affective

vocalizations. *IEEE International Conference on Acoustics, Speech and Signal Processing*,

985–988.

Snow, C. E. (1977). Mothers' speech research: From input to interaction. *Talking to Children:*

Language Input and Acquisition, *3149*.

Sommers, M. S., Nygaard, L. C., & Pisoni, D. B. (1992). Stimulus variability and the perception

of spoken words: Effects of variations in speaking rate and overall amplitude. *International*

Conference on Spoken Language Processing, (October), 217–220.

Song, J. Y., Demuth, K., & Morgan, J. (2010). Effects of the acoustic properties of infant-directed

speech on infant word recognition. *The Journal of the Acoustical Society of America*, *128*(1),

389–400.

Souza, P., Gehani, N., Wright, R., & McCloy, D. (2013). The advantage of knowing the talker.

Journal of the American Academy of Audiology, *24*(8), 689–700.

Spencer, P. E. (2004). Individual Differences in Language Performance after Cochlear

Implantation at One to Three Years of Age: Child, Family, and Linguistic Factors. *Journal of*

Deaf Studies and Deaf Education, *9*(4), 395–412.

Spitzer, S. M., Liss, J. M., & Mattys, S. L. (2007). Acoustic cues to lexical segmentation: A study

of resynthesized speech. *The Journal of the Acoustical Society of America*, *122*(6), 3678–

3687.

Spitzer, S. M., Liss, J., Spahr, T., Dorman, M., & Lansford, K. (2009). The use of fundamental

frequency for lexical segmentation in listeners with cochlear implants. *The Journal of the*

Acoustical Society of America, *125*(6), EL236–EL241.

Sulpizio, S., Kuroda, K., Dalsasso, M., Asakawa, T., Bornstein, M. H., Doi, H., … Shinohara, K.

(2018). Discriminating between Mothers' Infant- and Adult-Directed Speech: Cross-

Linguistic Generalizability from Japanese to Italian and German. *Neurosci Res.*, *133*, 21-27.

Svirsky, M. A. (2017). Cochlear implants and electronic hearing. *Physics Today*, *70*(8), 53–58.

Szagun, G., & Schramm, S. A. (2016). Sources of variability in language development of children

with cochlear implants: Age at implantation, parental language, and early features of

children's language construction. *Journal of Child Language*, *43*(3), 505–536.

Tamati, T. N., Janse, E., & Başkent, D. (2019). Perceptual Discrimination of Speaking Style under

Cochlear Implant Simulation. *Ear and Hearing*, *40*(1), 63–76.

Telkemeyer, S., Rossi, S., Koch, S. P., Nierhaus, T., Steinbrink, J., Poeppel, D., … Wartenburger,

I. (2009). Sensitivity of Newborn Auditory Cortex to the Temporal Structure of Sounds.

Journal of Neuroscience, *29*(47), 14726–14733.

The ManyBabies Consortium (2020). Quantifying sources of variability in infancy research using

the infant-directed speech preference. *Advances in Methods and Practices in Psychological*

Science, *3*(1), 24–52.

Thiessen, E. D., Hill, E. A., & Saffran, J. R. (2005). Infant directed speech facilitates word

segmentation. *Infancy*, *7*(1), 53–71.

Throckmorton, C. S., & Collins, L. M. (2002). The effect of channel interactions on speech

recognition in cochlear implant subjects: Predictions from an acoustic model. *The Journal of*

the Acoustical Society of America, *112*(1), 285–296.

Trainor, L. J., Austin, C. M., & Desjardins, N. (2000). Is infant-directed speech prosody a result

of the vocal expression of emotion? *Psychological Science*, *11*(3), 188–195.

Walker-Andrews A. S. & Grolnick, W. (1983). Discrimination of vocal expression by young

infants. *Infant Behavior and Development*, *6*, 491–498.

Walker-Andrews, A. S., & Lennon, E. (1991). Infants' discrimination of vocal expressions:

Contributions of auditory and visual information. *Infant Behavior and Development*, *14*(2),

131–142.

Wang, Y., Bergeson, T. R., & Houston, D. M. (2017). Infant-Directed Speech Enhances Attention

to Speech in Deaf Infants With Cochlear Implants. *Journal of Speech Language and Hearing*

Research, *60*(11), 3321.

Wang, Y., Bergeson, T. R., & Houston, D. M. (2018). Preference for Infant-Directed Speech in

Infants With Hearing Aids: Effects of Early Auditory Experience. *Journal of Speech,*

Language & Hearing Research, *61*(9), 2431–2439.

Wang, Y., Shafto, C. L., & Houston, D. M. (2018). Attention to speech and spoken language

development in deaf children with cochlear implants: a 10-year longitudinal study.

Developmental Science, (July 2017), e12677.

Webb, A. R., Heller, H. T., Benson, C. B., & Lahav, A. (2015). Mother's voice and heartbeat

sounds elicit auditory plasticity in the human brain before full gestation. *Proceedings of the*

National Academy of Sciences of the United States of America, *112*(10), 3152–3157.

Weisleder, A., & Fernald, A. (2013). Talking to Children Matters: Early Language Experience

Strengthens Processing and Builds Vocabulary. *Psychological Science*, *24*(11), 2143–2152.

Werker, J. F., & McLeod, P. J. (1989). Infant preference for both male and female infant-directed

talk: a developmental study of attentional and affective responsiveness. *Canadian Journal of*

Psychology, *43*(2), 230–246.

Werker, J. F., Pegg, J., & McLeod, P. J. (1994). A cross-language investigation of infant preference

for infant-directed communication. *Infant Behavior and Development*, *17*(3), 323–333.

Winn, M. B., & Litovsky, R. Y. (2015). Using speech sounds to test functional spectral resolution

in listeners with cochlear implants. *The Journal of the Acoustical Society of America*, *137*(3),

1430–1442.

Xiang, S., Nie, F., & Zhang, C. (2008). Learning a Mahalanobis distance metric for data clustering

and classification. *Pattern Recognition*, *41*(12), 3600–3612.

Zanto, T. P., Hennigan, K., Östberg, M., Clapp, W. C., & Gazzaley, A. (2013). Effect of speaking

rate on recognition of synthetic and natural speech by normal-hearing and cochlear implant

listeners. *Ear and Hearing*, *34*(3), 313.

Zeng, F. G., Tang, Q., & Lu, T. (2014). Abnormal pitch perception produced by cochlear implant

stimulation. *PLoS ONE*, *9*(2).

Zheng, Y., Koehnke, J., & Besing, J. (2011). Effects of Noise and Reverberation on Virtual Sound

Localization for Listeners With Bilateral Cochlear Implants. *American Journal of Audiology*,

32(5), 569–572.

1155

1156

1157

1158

1159 **TABLE I.** Statistical GLMM for modeling the effect of level of speech degradation on acoustic

1160 distinctiveness between IDS and ADS as quantified by MD.

	β Estimate	St. Error	t	Pr(>|t|)
(intercept)	6.694	0.58	11.45	0.000[a]
Level of degradation	-0.886	0.053	-16.54	0.000[a]

1161 [a] The p-value was statistically significant.

1162

1163

1164

1165

1166

1167

1168

1169

1170

1171

1172 **TABLE II.** Statistical GLMM for the effects of speaking style, level of speech degradation, and

1173 their interaction on intelligibility of speech, as measured by SRMR or SRMR-CI.

	β Estimate	St. Error	t	Pr(>|t|)
(intercept)	10.365	1.339	7.74	<0.00001[a]
Speaking Style	3.054	0.737	4.14	<0.00001 [a]
Level of degradation	-2.23	0.737	-3.024	<0.0001 [a]
Speaking Style:Level of degradation	-1.336	0.466	-2.865	<0.0001 [a]

1174 [a] The p-value was statistically significant.

1175

1176

1177

1178

1179

1180

1181

1182

1183

1184

1185

1186

1187

FIG. 1. (Color online) Schematic diagram of the approach used in the present study for measuring

(1) MD between pairs of MFCCs derived from pairs of IDS-ADS stimuli, and (2) intelligibility of

IDS and ADS stimuli as estimated by SRMR value for listeners with NH and SRMR-CI for those

with CIs. Note that the dashed line denotes the process for creating and analyzing the noise-

vocoded versions of the same pairs of stimuli, while N stands for the number of spectral channels

in the noise-excited envelope vocoder. Blocks and lines with blue (dark gray) color indicate paths

for processing ADS, while those with orange (light gray) color indicate paths for processing IDS.

Note that SRMR and SRMR-CI were calculated only for unprocessed stimuli in order to estimate

intelligibility for NH and CI listeners, respectively. In this figure, the waveforms and their

corresponding MFCCs are from the utterance "See the modi?" spoken by one of the seven talkers

both in IDS and ADS speaking styles. $S_{IDS_{ij}}$ and $S_{ADS_{ij}}$ are the i^{th} pair of IDS and ADS stimuli (i

= {1,2,3,…,15}) for talker j (j ={1,2,3,…,7}). $MFCCs_{IDS_{ij}}$ and $MFCCs_{ADS_{ij}}$ are MFCC features

derived from $S_{IDS_{ij}}$ and $S_{ADS_{ij}}$ speech stimuli, respectively. The middle two panels show MFCCs

obtained from the frames of these IDS and ADS stimuli. MD_{ij} is the MD calculated to measure the

acoustic distance between the two matrices for $MFCCs_{IDS_{ij}}$ and $MFCCs_{ADS_{ij}}$. $SRMR_{IDS_{ij}}$ and

$SRMR_{ADS_{ij}}$ are the estimated intelligibility for $S_{IDS_{ij}}$ and $S_{ADS_{ij}}$ speech stimuli, respectively, as

heard by listeners with NH. $SRMR\text{-}CI_{IDS_{ij}}$ and $SRMR\text{-}CI_{ADS_{ij}}$ are the estimated speech

intelligibility for the same $S_{IDS_{ij}}$ and $S_{ADS_{ij}}$ speech stimuli, respectively, as heard by a listener with

CIs.

FIG. 2. (Color online) The mean (bar) and ±1 standard error (vertical error bar in black) of

Mahalanobis distance (MD) across groups of IDS and ADS stimuli at seven levels of spectro-

temporal degradation, ranging from no-degradation (natural/unprocessed) to 4-channel noise-

vocoded stimuli. Green (dark gray) circles show the mean MDs for each talker derived by

averaging MDs over 15 pairs of IDS-ADS stimuli for that talker.

FIG. 3. (Color online) Estimated intelligibility of speech stimuli spoken in IDS (orange; light gray)

and in ADS (blue; dark gray) styles (simulated) groups of listeners with different hearing statuses:

NH (estimated by SRMR) and those with CIs (estimated by SRMR-CI). The bar graphs represent

average values of SRMR or SRMR-CI over the seven talkers.

FIG. 4. (Color online) Variability across seven female talkers in (A) acoustic distinctiveness

between their IDS and ADS for unprocessed stimuli (Natural) and simulated CI speech within a

22-channel noise vocoder (22CH), and (B) the change in talkers' speech intelligibility (SI) due to

a change in their speaking style (ADS to IDS), as heard for two (simulated) listener groups with

either NH (estimated by SRMR) or CIs (estimated by SRMR-CI). The data points (gray circles)

are laid over a 1.96 standard error of the mean (95% confidence interval) in red (rectangle area

with light gray) and 1 standard deviation shown by blue lines (vertical dark gray lines). The solid

and dotted red lines (horizontal solid and dotted dark gray lines) show the mean and median,

respectively.