Evaluation of efficiency and sensitivity of 1D and 2D sample pooling strategies for diagnostic screening purposes

Jasper Verwilt[1,2,3], Pieter Mestdagh[1,2,3,4], Jo Vandesompele[1,2,3,4]

1 OncoRNALab, Cancer Research Institute Ghent

2 Department of Biomolecular Medicine, Faculty of Medicine and Health Sciences, Ghent University

3 Center for Medical Genetics, Ghent University

4 Biogazelle, Belgium

Abstract

As SARS-CoV-2 continues to spread around the world while the pandemic lasts, testing facilities are forced to massively increment their testing capacities to handle the increasing number of samples. While sample pooling methods have been proposed or are effectively implemented in some labs, no systematic and large-scale simulations have been performed using real-life quantitative data from testing facilities. Here, we use anonymous data from 1632 positive cases to simulate and compare 1D and 2D pooling strategies. We show that the choice of pooling method and pool size is an intricate decision with a prevalence-dependent efficiency-sensitivity trade-off.

Introduction

Massive screening of the population for SARS-CoV-2 infection is proposed as one of the key strategies in the global battle against the COVID-19 pandemic. Due to the immense number of samples that are analyzed during population screening, pooling of samples seems to be a valid strategy in overcoming shortages in reagents and increasing testing capacity.

Several pooling strategies in the frame of SARS-CoV-2 testing have been discussed in recent preprints. The most discussed method is referred to as one-time pooling[1–4]. In this strategy, the samples are pooled, pools are tested and only samples in positive pools are tested individually. A second popular approach is called sequential pooling[1,5] in which the samples are pooled, pools are tested and positive pools are split into two equally-sized sub-pools. These sub-pools are tested again and this process is repeated until the individual sample level is reached. The third method that has been explored is two-dimensional or 2D pooling[6]. This method organizes samples in a 2D matrix and then creates pools along the rows and columns of the matrix. The pools are tested and negative rows and columns are excluded from the matrix. Next, all remaining samples are tested individually. Finally, researchers have explored pooling strategies in which samples are assigned to multiple pools, but require only one round of testing. Given the composition and test results of the positive pools, the positive samples can immediately be identified without the need for further individual testing. P-BEST[7] and Tapestry[8] are examples of such strategies.

While attractive, pooling strategies come with inherent limitations. First, pooling dilutes each individual sample, possibly to such an extent that the viral RNA

becomes undetectable, resulting in false negative observations[9–11]. A second

limitation is that an increase in sample manipulations augments the risk of cross-

contamination and sample mix-ups, which can lead to false negatives and false

positives[6]. Other drawbacks are unique to specific strategies. The P-BEST pooling

protocol is very time consuming, even when using a pipetting robot[7]. The repeated

pooling method, on the other hand, suffers from a complicated re-pooling scheme[1].

Although the number of preprints and peer-reviewed publications on pooling

strategies for COVID-19 PCR-based testing has increased rapidly throughout the

pandemic, some important insights are still lacking. First of all, the proposed optimal

pooling strategy is often based on a binary classification of samples as either positive

or negative. However, this Boolean approach is not in accordance with the real-world

situation and does not allow for investigating the dilution effect of pooling. Second,

when using a quantitative representation of the viral loads, these values need to

reflect real-life data, as patients present a very wide range of viral loads. This is

reflected in the wide spectrum of Cq values reported by PCR-based tests[12]. Finally,

as pooling is most effective for population-wide screening (where a very low

prevalence is expected) it is important to determine the performance of strategies

when encountering a low fraction of positive samples.

Here, we evaluate one-time (or 1D) pooling and two-dimensional (2D) pooling (using

practical microtiter plate format pool sizes) as promising strategies for massive, low-

prevalence population screening using real-life RT-qPCR data from 1632 positive

samples.

Results

Inverse relationship between efficiency and prevalence

In order to evaluate the efficiency-gain of the adopted pooling strategies, we calculated the number of tests that are needed to analyze all samples and divided this number by the total number of samples. We calculated the median, minimum and maximum prevalence-specific efficiency for each pooling strategy (Figure 1). First, we observe an inverse relationship between efficiency and prevalence over the evaluated prevalence range from 0.01% to 10% for all pooling strategies. Second, there is no single most efficient strategy, because this depends on the prevalence (notice crossings in Figure 1). Until a prevalence of 0.36%, 1x24 is the most efficient strategy, from 0.40% to 2.51% 16x24 becomes the most efficient, from 2.82% to 4.47% the most efficient strategy is 12x24 and from 5.01% to 10% 8x12 is the most efficient strategy. However, at high prevalence all strategies show similar efficiency gain. Third, strategies employing a larger pool size display a higher efficiency when the prevalence is low, but as the prevalence increases, there is a tipping point for each strategy at which its smaller pool size variant becomes more efficient. At very low prevalence, the efficiency of a 1 x n pool size becomes n and that of a m x n pool strategy becomes $(m \times n) / (m + n)$. As a general trend, 2D pooling methods are less sensitive to changes in prevalence in comparison with 1D pooling methods. We conclude that the most efficient pool size very much depends on the prevalence, but 2D pooling methods generally are most efficient when prevalence is higher than 0.4%.

Sensitivity decreases with lower prevalence

The true performance of a pooling strategy cannot be evaluated by efficiency only.

Since the number of false negatives due to pooling is one of the main possible

drawbacks of sample pooling, it is necessary to take this into account when choosing

the optimal strategy. In this regard, we calculated the average sensitivity of the

different simulated settings (Figure 2, Supplemental Figure 3), which ranges from

0.636 to 0.968. For all of the pooling strategies—and irrespective of pool size—we

primarily see that sensitivity increases with increasing prevalence. There is a non-

linear relationship between sensitivity and prevalence. Furthermore, since prevalence

is linked with efficiency and as a result indirectly linked with sensitivity, we note that

an increase in efficiency comes with a decrease in sensitivity. We also observe that

at a prevalence lower than 1%, there is an increased variation in sensitivity for

different simulation cohorts. Thus, the increased efficiency at low prevalence comes

with a low and pool size-dependent problematically variable sensitivity.

Sensitivity loss in function of viral load

With the intention of determining the influence of the viral load and, as a proxy, Cq

value of the positive sample on the sensitivity associated with the pooling strategies,

we calculated the probabilities of a true-positive for each non-pooled original Cq

value (Supplemental Figure 2). We investigated how the sensitivity changes when

higher Cq values are progressively being included in the cohort, starting with

samples with highest viral load (lowest Cq value) (Figure 3, Supplemental Figure 4).

In the first place, we note that the Cq value at which sensitivity loss starts to occur

only depends on the 1D pool size or largest dimension of the 2D pool; i.e. Cq value

of 35 for 1x4, 34 for 1x8, 33.4 for 1x12 and 8x12, 33 for 1x16 and 12x16 and 32.4 for

1x24   and   16x24.   These   Cq   values   are   (as   expected)   identical   to

$37 - \log_2(largest\ pool\ size)$. Additionally, when Cq values larger than these cut-off

values are systematically included, the sensitivity drops exponentially. The rate at

which this reduction happens, decreases when prevalence increases. Finally, when

the prevalence is 10%, larger pool sizes result in a smaller reduction in sensitivity,

but for all other visualized prevalence values the sensitivity decreases with larger

pool size. Altogether, the extent to which low viral load samples contribute to the drop

in sensitivity depends on pool size and pooling strategy, although the sensitivity

decrease is most problematic when prevalence is low.

Discussion

Sample pooling strategies form an incredible asset in an attempt to increase

throughput in times when massive testing for COVID-19 would be needed. A plethora

of pooling strategies have been suggested, some more performant or practical than

others. 1D and 2D pooling methods were selected in this simulation study because

they are simple and quick to perform, have straightforward pipetting schemes and do

not require re-accessing the same sample more than twice. Pool sizes were selected

to be easily compatible with 96-well plates. Using a large real-life dataset of 1632

positive samples enabled us to simulate relevant settings and provide more accurate

outcomes in comparison with real wet-lab tests, which are rather limited in the

number of positive samples and may not cover the whole range of viral loads.

Because the original Cq distribution depends on the origin of the sample (hospital,

care center, …) and the stage of pandemic, our observations do as well. Firstly, our

results confirm the widely accepted idea that sample pooling methods show a higher

efficiency when prevalence is low[1–6,13] and that, for 1D and 2D pooling methods, as

prevalence increases, a threshold is reached after which smaller pool sizes become

more efficient[1,6]. However, appraising the performance of a pooling method

exclusively by its efficiency would ignore one of the major drawbacks of pooling: loss

of sensitivity due to dilution of the target. This issue becomes most pertinent when

the viral load is low[9–11,13,14]. Our results confirm that all tested pooling methods suffer

from false negatives, to a variable degree (Figure 2). This loss in sensitivity across all

prevalence conditions generally precludes use of pooling for diagnostic testing of

COVID-19 samples according to the U.S. Food and Drug Administration (FDA),

whereby only 1x16 and 1x24 strategies under high prevalence (≥10%) conditions

meet the minimal 95% sensitivity requirement[15]. When prevalence is high, the loss in

sensitivity for large pools is partly compensated by the fact that low viral load

samples can be 'rescued' by high viral load samples when present in the same pool.

(Figure 3). Intuitively, 2D pooling methods are especially vulnerable for false

negatives, as a high Cq sample would have to be 'rescued' in the corresponding row

and column pools, which is confirmed by our results. The influential role of

prevalence on efficiency as well as sensitivity presents an import challenge,

considering that, in order to make an informed decision on the preferential pooling

strategy, the prevalence has to be known. By nature, we cannot know the exact

prevalence before testing our samples, and as a result, the prevalence has to be

estimated. In general, we show that it is of extreme importance that an optimal

equilibrium between efficiency and sensitivity is achieved when deciding on the

pooling strategy and corresponding pool size.

Materials and Methods

Patient samples

Nasopharyngeal swabs were taken by a healthcare professional as a diagnostic test

for SARS-CoV-2, as part of the Belgian national testing platform. The individuals

were tested at nursing homes or in triage centers, between April 9[th] and June 7[th]. To

mimic low prevalence viral loads as much as possible, only batches of 94 patient

samples with fewer than 10 positives were included in this study. After additional

filtering as described in a further paragraph, this resulted in 113 928 patients in total,

of which 1632 positives (1.43%) with corrected Cq values ranging from 9.85 to 36.94

(median of 28.78) (Supplemental Figure 1).

SARS-CoV-2 RT-qPCR test

RNA extraction was performed using the Total RNA Purification Kit (Norgen Biotek

#24300) according to the manufacturer's instructions using 200 µl transport medium,

200 µl lysis buffer and 200 µl ethanol, with processing using a centrifuge (5810R with

rotor A-4-81, both from Eppendorf). RNA was eluted from the plates using 50 µl

elution buffer (nuclease-free water), resulting in approximately 45 µl eluate. RNA

extractions were simultaneously performed for 94 patient samples and 2 negative

controls (nuclease-free water). After addition of the lysis buffer, 4 µl of a proprietary

700 nucleotides spike-in control RNA (40 000 copies) and carrier RNA (200 ng of

yeast tRNA (Roche #10109517001) was added to all 96 wells from the plate). To the

eluate of one of the negative control wells, 7500 RNA copies of positive control RNA

(Synthetic SARS-CoV-2 RNA Control 2, Twist Biosciences #102024) were added.

Six µl of RNA eluate was used as input for a 20 µl RT-qPCR reaction in a CFX384

qPCR instrument using 10 µl iTaq one-step RT-qPCR mastermix (Bio-Rad

#1725141) according to the manufacturer's instructions, using 250 nM final

concentration of primers and 400 nM of hydrolysis probe. Primers and probes were

synthesized by Integrated DNA Technologies using clean-room GMP production. For

detection of the SARS-CoV-2 virus, the Charité E gene assay was used (FAM)[16]; for

the internal control, a proprietary hydrolysis probe assay (HEX) was used. Prior to

May 25th, 2 singleplex assays were performed; after May 25th, 1 duplex RT-qPCR

was performed (with 1/8[th] of spike-in control RNA). Cq values were generated using

the FastFinder software v3.300.5 (UgenTec). Only batches were approved with a

clean negative control and a positive control in the expected range.

Assembly of positive Cq value distribution

First, a subset of full 96-well RNA plates is selected containing less than 10 positives.

The latter selection criterium is introduced to filter out plates originating from high

COVID-19 positive rate circumstances, such as hospitals, to avoid putative bias

towards high viral loads. Only plates are retained with a positive control value that is

within two standard deviations of the mean positive control value of all filtered plates.

Next, in order to correct the Cq values for inter-plate variation, the difference of the

average positive control Cq values per qPCR plate from the overall mean positive

control Cq is calculated. The Cq values in each qPCR plate are then corrected by the

qPCR plate-specific difference from the mean. Supplemental Figure 1 displays the

distribution of the corrected Cq values of the positive samples. Finally, by inspecting

this histogram, we dismiss all Cq values larger than 37 as noisy data, resulting in a

final set of 1632 positive Cq values.

Simulation of 1D and 2D pooling strategies

Simulations of are run using R 4.0.1. First, several cohorts of 100 000 patients are

repeatedly simulated with varying fractions f, defined by $f = 10^{-x} \ with \ x =$

$4, 3.95, …, 1$, of positive cases (resulting in 61 cohorts of 0.01% to 10% prevalence).

This is done five times, resulting in five replicate cohort per prevalence (302 cohorts).

The Cq values of the positive samples are sampled with replacement from the set of

1632 positive Cq values. Next, the patients are randomly separated into pools

depending on the pooling strategy that is simulated. The pooling strategies that were

simulated are 1x4, 1x8, 1x12, 1x16, 1x24 (all 1D), and 8x12, 12x16 and 16x24. The

Cq value of the pool was calculated as follows:

$$c_{pool} = \log_2 P - \log_2 \sum_{i=1}^{p} 2^{-c_i}$$

With c_{pool} the Cq value of the pool, P the number of samples in the pool, p the

number of positive samples in the pool, $c_1, c_2, …, c_p$ the Cq values of the positive

samples. If the Cq value of the pool is smaller than 37, it is classified as a positive

pool. For 1D pooling, only samples in positive pools are retained and the remaining

individual Cq values were checked to be positive. For 2D pooling, the Cq values of

the differently sized pools are checked simultaneously and the samples in negative

pools are removed, after which all Cq values of the remaining samples are checked

individually. Samples that were retained after the testing of the pools and had an

individual Cq lower than 37 are classified as positive, all other samples are classified

as negative.

The sensitivity is calculated as:

$$sensitivity = \frac{\# \; true \; positive \; samples}{\# \; true \; positive \; samples + \# \; false \; negative \; samples}$$

245   The analytical efficiency gain is calculated as:

$$efficiency\ gain = \frac{\#\ tests\ required\ for\ individual\ testing}{\#\ tests\ required\ for\ pooling\ strategy}$$

248   In all simulations, the number of tests required for individual testing is equal to the

249   number of samples.

Acknowledgements

252   We are grateful for the data from the Belgian federal taskforce for COVID-19 qPCR

253   testing.

Authors' contributions

256   Conceptualization: J.Va., P.M. and J.Ve.; Methodology: J.Va., P.M. and J.Ve.;

257   Software: J.Ve.; Formal Analysis: J.Ve.; Resources: J.Va. and P.M.; Data Curation:

258   J.Ve.; Writing – Original Draft: J.Va. and J.Ve.; Writing – Review & Editing; J.Va.,

259   P.M. and J.Ve.; Visualisation: J.Va. and J.Ve.; Supervision: J.Va and P.M.; Project

260   Administration: J.Va. and P.M.

Data availability

263   The   code   and   Cq   values   are   available   on

264   https://github.com/jasperverwilt/covidpooling. The Cq values are available as an

265   RDML file in the RDMLdb database[17], under ID 2008AA74.

References

1. Shani-Narkiss, H., Gilday, O. D., Yayon, N. & Landau, I. D. Efficient and

Practical Sample Pooling High-Throughput PCR Diagnosis of COVID-19.

medRxiv 2020.04.06.20052159 (2020). doi:10.1101/2020.04.06.20052159

2. Guha, P., Guha, A. & Bandyopadhyay, T. Application of pooled testing in

screening and estimating the prevalence of Covid-19. *medRxiv*

2020.05.26.20113696 (2020). doi:10.1101/2020.05.26.20113696

3. Adams, K. Expanding Covid-19 Testing: Mathematical Guidelines for the

Optimal Sample Pool Size Given Positive Test Rate. *medRxiv*

2020.05.21.20108522 (2020). doi:10.1101/2020.05.21.20108522

4. Millioni, R. & Mortarino, C. Sequential informed pooling approach to detect

SARS-CoV2 infection. *medRxiv* 2020.04.24.20077966 (2020).

doi:10.1101/2020.04.24.20077966

5. Bukhari, S. U. K., Khalid, S. S., Syed, A. & Shah, S. S. H. Smart Pooled

sample Testing for COVID-19: A Possible Solution for Sparsity of Test Kits.

medRxiv 2020.04.21.20044594 (2020). doi:10.1101/2020.04.21.20044594

6. Sinnott-Armstrong, N., Klein, D. & Hickey, B. Evaluation of Group Testing for

SARS-CoV-2 RNA. *medRxiv* 2020.03.27.20043968 (2020).

doi:10.1101/2020.03.27.20043968

7. Shental, N. *et al.* Efficient high throughput SARS-CoV-2 testing to detect

asymptomatic carriers. *medRxiv* 2020.04.14.20064618 (2020).

doi:10.1101/2020.04.14.20064618

8. Ghosh, S. *et al.* Tapestry: A Single-Round Smart Pooling Technique for

COVID-19 Testing. *medRxiv* 2020.04.23.20077727 (2020).

doi:10.1101/2020.04.23.20077727

9. Torres, I., Albert, E. & Navarro, D. Pooling of Nasopharyngeal Swab

Specimens for SARS-CoV-2 detection by RT-PCR. *medRxiv*

2020.04.22.20075598 (2020). doi:10.1101/2020.04.22.20075598

10. Yelin, I. *et al.* Evaluation of COVID-19 RT-qPCR test in multi-sample pools.

Clin. Infect. Dis. (2020). doi:https://doi.org/10.1093/cid/ciaa531

11. Gan, Y. *et al.* Sample Pooling as a Strategy of SARS-COV-2 Nucleic Acid

Screening Increases the False-negative Rate. *medRxiv* 2020.05.18.20106138

(2020). doi:10.1101/2020.05.18.20106138

12. Liu, Y. *et al.* Viral dynamics in mild and severe cases of COVID-19. *The Lancet*

Infectious Diseases **20**, 656–657 (2020).

13. Ben-Ami, R. *et al.* Pooled RNA extraction and PCR assay for efficient SARS-

CoV-2 detection. *medRxiv* 2020.04.17.20069062 (2020).

doi:10.1101/2020.04.17.20069062

14. Nguyen, N. T., Aprahamian, H., Bish, E. K. & Bish, D. R. A methodology for

deriving the sensitivity of pooled testing, based on viral load progression and

pooling dilution. *J. Transl. Med.* **17**, 252 (2019).

15. U.S. Food and Drug Administration. *Molecular Diagnostic Template for*

Laboratories. (2020).

16. Corman, V. M. *et al.* Detection of 2019 novel coronavirus (2019-nCoV) by real-

time RT-PCR. *Eurosurveillance* **25**, 2000045 (2020).

17. Ruijter, J. M. *et al.* RDML-Ninja and RDMLdb for standardized exchange of

qPCR data. *BMC Bioinformatics* **16**, (2015).

Figures

Figure 1: Average efficiency increase (number of samples divided by the number of tests performed)

over a 0.01 to 10% prevalence range for different pooling strategies. The averages are approximated

by the median over five replicate simulations of a cohort size of 100 000 patients each. The ranges

around the line indicate the minimum and maximum value over five replicate simulations.

Figure 2: Sensitivity in function of the prevalence for different strategies. Each point is colored by its

corresponding efficiency. Each data point illustrates the median value of five replicate simulations,

each simulation starting from a set prevalence and pooling strategy for a testing cohort of 100 000

patients.

Figure 3: Cumulative sensitivities calculated for all Cq values up to a cut-off Cq for four prevalences.

All true positives and false negatives over five replicate simulations for each sampled Cq value are

counted and used for calculation of the sensitivity up to the cut-off Cq. The data is filtered for

prevalences of 0.01%, 0.1%, 1% and 10% and grouped by prevalence.

336 **Supplemental Figures**

Supplemental Figure 1: The binned distribution of Cq values of all full 96-well RNA plates with good

controls and less than 10 positives. The Cq values are corrected by calculating the difference between

the mean positive control Cq value per 384-well qPCR plate from the overall mean of the positive

control Cq values. Grey bars indicate the Cq values classified as noise.

Supplemental Figure 2: Cq value distributions for false negatives (FN) and true positives (TP) for

different pooling strategies. Each data point represents a Cq value from a replicate simulation with a

set prevalence.

Supplemental Figure 3: Distributions of the average sensitivity for the simulations in each group.

Each data point is the median sensitivity over five simulations.

Supplemental Figure 4: Cumulative sensitivities calculated for all Cq values up to a cut-off Cq for

each strategy. All true positives and false negatives over five replicate simulations for each sampled

Cq value are counted and used for calculation of the sensitivity up to the cut-off Cq. The data is filtered

for prevalences of 0.01%, 0.1%, 1% and 10% and colored accordingly.

358