Title

Going beyond clinical routine in SARS-CoV-2 antibody testing - A multiplex corona virus

antibody test for the evaluation of cross-reactivity to endemic coronavirus antigens

Authors

Matthias Becker[1,#], Monika Strengert[2,3,#], Daniel Junker[1], Tobias Kerrinnes[4], Philipp D. Kaiser[1],

Bjoern Traenkle[1,16], Heiko Dinter[1,16], Julia Häring[1], Anne Zeck[1], Frank Weise[1], Andreas

Peter[5,21,22], Sebastian Hörber[5,21,22], Simon Fink[1], Felix Ruoff[1], Tamam Bakchoul[9], Armin

Baillot[10], Stefan Lohse[11], Markus Cornberg[12], Thomas Illig[13], Jens Gottlieb[14], Sigrun Smola[11],

André Karch[17], Klaus Berger[17], Hans-Georg Rammensee[6,7,8], Katja Schenke-Layland[1,8,18,19],

Annika Nelde[6,8,20], Melanie Märklin[8,20], Jonas S. Heitmann[8,20], Juliane S. Walz[6,8,15,20], Markus

Templin[1], Thomas O. Joos[1], Ulrich Rothbauer[1,16,#], Gérard Krause[2,3], Nicole Schneiderhan-

Marra[1,*]

Affiliations

[1] NMI Natural and Medical Sciences Institute at the University of Tuebingen, Germany

[2] Department of Epidemiology, Helmholtz Centre for Infection Research, Braunschweig,

Germany

[3] TWINCORE GmbH, Centre for Experimental and Clinical Infection Research, a joint venture

of the Hannover Medical School and the Helmholtz Centre for Infection Research, Hannover,

Germany

[4] Helmholtz-Institute for RNA-based Infection Research (HIRI), Würzburg, Germany

[5] Institute for Clinical Chemistry and Pathobiochemistry, Department for Diagnostic Laboratory

Medicine, University Hospital Tuebingen, Tuebingen, Germany

[6] Institute for Cell Biology, Department of Immunology, University of Tuebingen, Tuebingen,

Germany

[7] German Cancer Consortium (DKTK) and German Cancer Research Center (DKFZ), partner

site Tuebingen, Tuebingen, Germany.

[8] Cluster of Excellence iFIT (EXC2180) "Image-Guided and Functionally Instructed Tumor

Therapies", University of Tuebingen, Tuebingen, Germany

[9] Institute for Clinical and Experimental Transfusion Medicine, University Hospital Tuebingen,

Tuebingen, Germany

[10] Niedersächsisches Landesgesundheitsamt, Department of Virology/Serology, Hannover,

Germany

[11] Institute of Virology, Saarland University Medical Center, Homburg/Saar, Germany

[12] Department of Gastroenterology, Hepatology, Endocrinology, Hannover Medical School,

Hannover, Germany; Centre for Individualized Infection Medicine (CiiM), Hannover, Germany

[13] Hannover Unified Biobank (HUB), Hannover Medical School, Hannover, Germany

[14] Clinic for Pneumonology, Hannover Medical School, Hannover, Germany

[15] Department of Hematology, Oncology, Clinical Immunology and Rheumatology, University

Hospital Tuebingen, Tuebingen, Germany

[16] Pharmaceutical Biotechnology, University of Tuebingen, Germany

[17] Institute of Epidemiology and Social Medicine, University of Münster, Münster, Germany

[18] Department of Women's Health, Research Institute for Womens's Health, Eberhard-Karls-

University, Tuebingen, Germany

[19] Department of Medicine/Cardiology, Cardiovascular Research Laboratories, David Geffen

School of Medicine at UCLA, Los Angeles, CA, USA

[20] Clinical Collaboration Unit Translational Immunology, German Cancer Consortium (DKTK),

Department of Internal Medicine, University Hospital Tuebingen, Tuebingen, Germany

[21] Institute for Diabetes Research and Metabolic Diseases of the Helmholtz Center Munich at

the University of Tuebingen, Tuebingen, Germany

[22] German Center for Diabetes Research (DZD), München-Neuherberg, Germany

[#] These authors contributed equally to this work.

* Corresponding author:

Dr. Nicole Schneiderhan-Marra

Markwiesenstrasse 55, 72770 Reutlingen, Germany

Phone: 0049 7121 51530 815

Fax: 0049 7121 51530 16

E-Mail: Nicole.Schneiderhan@nmi.de

Abstract

Given the importance of the humoral immune response to SARS-CoV-2 as a global benchmark

for immunity, a detailed analysis is needed to monitor seroconversion in the general population,

understand manifestation and progression of COVID-19 disease, and ultimately predict the

outcome of vaccine development. In contrast to currently available serological assays, which

are only able to resolve the SARS-CoV-2 antibody response on an individual antigen level, we

developed a multiplex immunoassay, for which we included spike and nucleocapsid proteins

of SARS-CoV-2 and the endemic human coronaviruses (NL63, OC43, 229E, HKU1) in an

expanded antigen panel. Compared to three commercial *in vitro* diagnostic tests, our

MULTICOV-AB assay achieved the highest sensitivity and specificity when analyzing a well-

characterized sample set of SARS-CoV-2 infected and uninfected individuals. Simultaneously,

high IgG responses against endemic coronaviruses became apparent throughout all samples,

but no consistent cross-reactive IgG response patterns could be defined. In summary, we have

established and validated, a robust, high-content-enabled, and antigen-saving multiplex assay

MULTICOV-AB, which is highly suited to monitor vaccination studies and will facilitate

epidemiologic screenings for the humoral immunity toward pandemic as well as endemic

coronaviruses.

Introduction

Given the importance of the humoral immune response to SARS-CoV-2 as a global benchmark for immunity, a detailed analysis is needed to (i) monitor seroconversion in the general population[1,2], (ii) understand manifestation and progression of the disease[3,4], and (iii) predict the outcome of vaccine development[5,6]. Currently available serological assays utilize single analyte technologies such as ELISA to measure antibodies against SARS-CoV-2 antigens including spike (S) or nucleocapsid (N) protein[1,6-8]. To measure individual antibody (IgG and IgA) responses against SARS-CoV-2 and the endemic human coronaviruses (hCoVs) NL63, 229E, OC43, and HKU1, we developed a multiplexed immunoassay (MultiCoV-Ab), for which we included S and N proteins of these coronaviruses in an expanded antigen panel. Compared to commercial *in vitro* diagnostic (IVD) tests our MultiCoV-Ab assay achieved the highest sensitivity and specificity when analyzing 310 SARS-CoV-2 infected and 866 uninfected individuals. Simultaneously we see high IgG responses against hCoVs throughout all samples, whereas no consistent cross reactive IgG response patterns can be defined. In summary, our MultiCoV-Ab assay is highly suited to monitor vaccination studies and will facilitate epidemiologic screenings for the humoral immunity toward pandemic as well as endemic coronaviruses.

Results

To investigate the antibody response of SARS-CoV-2-infected individuals, we developed and established a high-throughput and automatable bead-based multiplex assay, termed MultiCoV-Ab. We expressed and immobilized six different SARS-CoV-2-specific antigens on Luminex MAGPLEX beads with distinct color codes, specifically the trimeric full-length Spike protein (Spike Trimer), receptor binding domain (RBD), S1 domain (S1), S2 domain (S2), full-length nucleocapsid (N) and the N-terminal domain of nucleocapsid (N-NTD) (**Extended Data Fig. 1**). Immunoglobulins from serum and plasma samples were detected using phycoerythrin-labelled anti-human IgG or IgA antibodies. Data on quality control and assay performance is provided in **Extended Data Fig. 2** and **Extended Data Table 1.**

As key antigens for the classification of SARS-CoV-2-induced seroconversion, we used Spike Trimer and RBD previously described by Amanat *et. al*[2], and initially screened a set of 205 SARS-CoV-2-infected and 72 uninfected individuals with the MultiCoV-Ab assay. Using a combined cut-off of both antigens, we identified all uninfected samples as negative (**Fig. 1a**). Of the 205 infected samples, the MultiCoV-Ab assay identified 24 (11.7%) as IgG antibody-negative. This finding is supported by three other commercially available IVD tests (Roche[9], Siemens Healthineers[10], Euroimmun[11]) widely used in clinical routine SARS-CoV-2 antibody testing. However, the IVD tests missed additional 8 (Roche), 11 (Siemens Healthineers) and 9 (Euroimmun) samples of SARS-CoV-2-infected individuals. Furthermore, the Euroimmun test classified 8 additional samples as "borderline" (**Fig. 1b, Extended Fig. 3a-c**). In accordance with our MultiCoV-Ab assay, no samples were classified as false positives by the Roche and Siemens tests, whereas one sample was classified as false positive and one as "borderline" by the Euroimmun test.

When testing for IgA antibodies in samples of SARS-CoV-2-infected individuals, our MultiCoV-Ab assay classified 47 (22.9 %) as IgA-negative, whereas the Euroimmun test classified 32 (15.6 %) as IgA-negative, and 16 (7.8 %) as borderline (**Fig. 1b and Extended Data Fig. 3d**). For the uninfected samples, the Euroimmun test identified 7 (9.7 %) as false positives and 3

(4.2 %) as "borderline", whereas no samples were classified as false positives by the MultiCoV-

Ab assay.

Next, we used an extended sample set with 310 SARS-CoV-2-infected and 866 uninfected

donors for clinical validation of MultiCoV-Ab assay. A simplified overview of this set is shown

in **Fig. 2a**; a complete breakdown is displayed in **Extended Data Table 2**. A direct comparison

revealed that Spike Trimer and RBD were the best predictors of SARS-CoV-2 infection. Signal

cut-offs were defined for both, IgG and IgA detection, based on ROC analysis with focus on

maximum specificity. Additionally, we defined a cut-off for overall IgG and IgA positivity for

which both individual cut-offs for Spike Trimer and RBD had to be met (**Fig. 2b**). As shown

above, cut-offs based on IgG were shown to be more sensitive and specific than those based

on IgA. With the IgG overall cut-off, we reached a specificity of 100 % (**Fig. 2c**), which would

not have been possible for either of the antigens individually, while still retaining acceptable

sensitivity. To identify samples with an early immune response, we simultaneously measured

IgA response. With the MultiCoV-Ab assay, we identified eight IgA-positive samples that

showed no IgG response (**Fig. 2d**). Two of these were uninfected and falsely classified as

positive. For four of the remaining six infected samples, details regarding the time between the

onset of symptoms and sample drawing were available (2, 6, 7, and 15 days). We hypothesized

that IgA in these samples can be used to measure an early onset of antibody response. Thus,

we classified samples with strong IgA positivity - signal to cut-off (S/CO) > 2 for Spike Trimer

and RBD - as "positive", irrespective of their detected IgG response. With this

combined IgG + IgA classification, we reached an optimal sensitivity of 90 % while retaining a

specificity of 100 %.

Further analyzing the Ig response towards both subdomains of the spike, S1 and S2, we

achieved no additional sensitivity for the classifier (**Fig. 2e**). Interestingly, RBD, as a part of

S1, showed much fewer uninfected samples with increased IgG response compared to S1.

For S2 even more uninfected samples had increased signals, pointing to potential cross-

reactivity in this domain of the spike protein (**Fig. 2e**). We further complemented our assay

with the N and N-NTD proteins. Although these antigens were successfully used in single-

analyte assays[12], we observed a high cross reactivity in uninfected samples for both (**Fig. 2f**).

Across the entire data set, only one sample showed a distinct immune response to N and N-

NTD, but not to all spike derived antigens.

Longitudinal samples from five hospitalized patients were subjected to a small-scale time

course analysis of IgG and IgA immune responses (**Fig. 3a**). Levels of both Ig classes strongly

increased within the first ten days after the onset of symptoms. While IgG levels appeared

constant over roughly two months, IgA levels started to decline between day 10 and 20 after

the onset of symptoms where samples were available. These effects were consistent for the

majority of SARS-CoV-2 antigens. Furthermore, we found that patients´ hospitalization, as a

measure of disease severity (**Fig. 3b**), seemed to correlate with an increased humoral immune

response, especially in IgA. Furthermore, there is indication for a trend for increasing age as

well (**Fig. 3c**). However, it should be considered that patients of higher age also had a higher

rate of hospitalization in our study population.

In order to explore cross-reactivity of hCoVs with SARS-CoV-2, we included S1, N, and N-NTD

antigens from human α- (NL63 and 229E) and β-hCoVs (OC43 and HKU1) in our MultiCoV-

Ab panel (**Extended Data Fig. 1**). The immune response towards all hCoV antigens was more

dependent on coronavirus clade than on N or S1 antigen. However, within the clades of α-

hCoVs and β-hCoVs, types of antigens were more dominant than the virus subtype, as

demonstrated by rank correlation analysis and hierarchical clustering. Interestingly, IgG

response against α-hCoVs clustered more closely to SARS-CoV-2 than to β-hCoVs (**Fig. 4a,**

Extended Data Fig. 4a). Overall, we identified a considerable immune response to hCoV

antigens throughout the whole sample set with no notable differences between samples from

SARS-CoV-2-infected and uninfected donors in IgG or IgA for S1 (**Fig. 4b**), N (**Fig. 4c**), or N-

NTD (**Extended Data Fig. 4b**).

We therefore used the IgG signal relative to the average response per antigen for further

analyses, which allowed comparison among all hCoV antigens on one scale. For those

uninfected samples, which showed an IgG cross reactivity towards Spike Trimer (Spike Trimer

false positives), we observed partially increased responses towards hCoV antigens. Those

samples, which did not show an immune response after SARS-CoV-2 infection (false

negatives, as determined by MultiCoV-Ab assay, combined IgG + IgA) were closer to the

baseline (**Fig. 4d-e, Extended Data Fig. 4c**). This indicates that cross-reactivity with hCoVs

causes some of the observed SARS-CoV-2 immune response in samples taken from

individuals not exposed to SARS-CoV-2.

To investigate the correlation of hCoV and SARS-CoV-2 immune response further, we grouped

samples into high and low responders for α-hCoVs and β-hCoVs. High responders had relative

IgG signals > 0 for N and S1 antigens of both hCoV subtypes within the clade, while low-

responders had < 0, respectively. Samples with SARS-CoV-2 immune response (as

determined by MultiCoV-Ab assay, combined IgG + IgA classification) were significantly

overrepresented within the group of α-hCoV high responders (p = 3.78e-03, Fisher's exact

test, two sided), while being significantly underrepresented within the group of α-hCoV and β-

hCoV low responders (p = 1.14e-03 and p = 1.56e-02, respectively, Fisher's exact test, two

sided) (**Fig. 4f**). These results showed that while there were no discernible global effects for

single antigens, there is a correlation between the SARS-CoV-2 immune response with high

hCoV responses, especially towards α-hCoVs.

Discussion

We demonstrated that our MultiCoV-Ab, a novel multiplex assay, is highly suitable to classify seroconversion in SARS-CoV-2-infected individuals. With a combined cut-off using SARS-CoV-2 trimeric full-length spike protein and RBD, we were able to eliminate false positive responses and achieved a sensitivity of 90% with a specificity of 100% for 310 samples from SARS-CoV-2-infected and for 866 samples from uninfected individuals. We found that detection of IgG more accurately reflected infection compared to IgA, although both were highly specific. However, by simultaneously monitoring IgA, we additionally were able to detect an early immune response in some patients. The MultiCoV-Ab approach allows the easy addition of SARS-CoV-2-specific antigens, here six in total, which provides an additional level of confidence in patient classification. Thus, for example, we noticed that the spike S1 domain showed fewer false positive responses compared to the S2 domain. Interestingly, Ng *et al.*[13] reported reactivity towards SARS-CoV-2 S2 from sera of patients with recent seasonal hCoV infection. These sera prevented infection with SARS-CoV-2 pseudotypes in a neutralization assay. Additionally, we found that spike non-responders also did not show a response to nucleocapsid, which has been described as strongest inducer of antibody responses[12,14]; and not vice versa.

In our comparison to commercially available IVD tests, we classified fewer samples as false negative using our MultiCoV-Ab assay. For 10% of all infected samples, we could not detect a SARS-CoV-2 specific immune response, which is in line with previous findings[3,15,16]. Those non-responders may be able to limit viral replication by innate immune mechanisms or cellular immunity is dominant in mediating viral clearance[17,18].

Expanding our MultiCoV-Ab assay to the endemic hCoVs NL63, 229E, OC43, and HKU1 revealed a clear IgG immune response for all tested samples. Furthermore, we did not observe a difference for the samples from proven hCoV-infected individuals, compared to other samples. Due to the general lack of availability of samples from hCoV-naïve individuals, it was difficult to analyze hCoV-mediated cross-reactivity. Nevertheless, our multiplexed readout indicates a correlation between the SARS-CoV-2 immune response and high hCoV responses.

Currently, we are identifying population groups which were highly exposed and showed

different susceptibility to SARS-CoV-2 infection, e.g. the "Ischgl-study group" (unpublished

data)[19], in order to elucidate potential cross protection derived from immune responses towards

endemic hCoVs in more detail. Alternatively, studies analyzing hCoV signatures in samples

from individuals before and after SARS-CoV-2 infection using the MultiCoV-Ab assay would

help to get insight into a potential cross protection.

A multiplex setup such as MultiCoV-Ab assay is especially suited to vaccination studies, since

the flexibility and broad antigen coverage allows to efficiently map vaccine immune responses

to an immunoglobulin isotype and subtype level for the target pathogen and related species [20].

Interestingly, previous SARS-CoV-1 vaccine studies clearly indicated that a detailed

characterization of vaccine-induced antibody responses is mandatory for efficient coronavirus

vaccine development[21,22].

In summary, we have established and validated the MultiCoV-Ab assay, a robust, high-

content-enabled, and antigen-saving multiplex assay. This assay is suitable for comprehensive

characterization of SARS-CoV-2 infection on the humoral immune response and for

epidemiological screenings to accurately measure SARS-CoV-2 seroprevalence in large

cohort studies. It further provides the unique opportunity to assess and correlate immunity for

both endemic and pathogenic coronaviruses. Finally, the multiplex nature of the MultiCoV-Ab

assay can deliver urgently needed data on the outcome of SARS-CoV-2 vaccination.

Supplementary - Materials and Methods

Generation of expression constructs for production of viral antigens

The cDNAs encoding the full-length nucleocapsid proteins of SARS-CoV-2, hCoV-OC43, hCoV-NL63, hCoV-229E, and hCoV-HKU1 (GenBank accession numbers QHD43423.2; YP_009555245.1; YP_003771.1; NP_073556.1; YP_173242.1) were produced with an N-terminal hexahistidine (His$_6$)-tag by DNA synthesis (ThermoFisher Scientific). The cDNAs were cloned by standard techniques into NdeI/HindIII sites of the bacterial expression vector pRSET2b (ThermoFisher Scientific). To generate N-terminal domains (NTDs) of the respective nucleocapsid proteins (SARS-CoV-2 NTD aa 1-189; hCoV-OC43 NTD aa 1-204; hCoV-NL63 NTD aa 1 - 154; hCoV-229E NTD aa 1-156; hCoV-HKU1 NTD aa 1- 203), a stop codon located N-terminally to the Serine-Arginine (SR)-rich linker site[23] was introduced via PCR mutagenesis of the nucleocapsid encoding plasmids using the forward primer pRSET2b down-for 5' - GGT AAG CTT GAT CCG GCT GCT AA - 3' and respective reverse primers: SARS-CoV2_NTD-rev 5' - GGG AAG CTT ACT CAG CAT AGA AGC CCT TTG G - 3', OC43_NTD-rev 5' - GGG AAG CTT ATT CGA TAT AAT AGC CCT GCG G - 3', NL63_NTD-rev 5' - GGG AAG CTT ATT CAA CAA CGC TCA GTT CCG - 3', 229E_NTD-rev 5' - GGG AAG CTT ATT CAA CAA CGG TAA CAC CAT TC - 3' and HKU1_NTD-rev 5' - GGG AAG CTT ATT CCA CAT AGT AGC CCT GAG GC - 3'.

The pCAGGS plasmids encoding the stabilized trimeric Spike protein and the receptor binding domain (RBD) of SARS-CoV-2 were kindly provided by F. Krammer[2].

The cDNA encoding the S1 domain (aa 1 - 681) of the SARS-CoV-2 Spike protein was obtained by PCR amplification using the forward primer S1_CoV2-for 5´- CTT CTG GCG TGT GAC CGG - 3´ and reverse primer S1_CoV2-rev 5´ - GTT GCG GCC GCT TAG TGG TGG TGG TGG TGG TGG GGG CTG TTT GTC TGT GTC TG - 3´ and the full length SARS-CoV-2 SPIKE cDNA as template and cloned into the XbaI/NotI-digested backbone of the pCAGGS vector, thereby adding a C-terminal His$_6$-Tag.

The cDNAs encoding the S1 domains of hCoV-OC43 (aa 1 -760), hCoV-NL63 (aa 1 - 744), hCoV-229E (aa 1 - 561) and hCoV-HKU1 (aa 1 - 755) (GenBank accession numbers

AVR40344.1; APF29071.1; YP_003771.1; APT69883.1; AGW27881.1) were produced by

DNA synthesis (ThermoFisher Scientific), digested using XbaI/NotI and ligated into the

pCAGGS vector. All expression constructs were verified by sequence analysis.

Protein expression and purification

For expression of the viral nucleocapsid proteins (full-length nucleocapsid and N-NTDs), the

respective expression constructs were used to transform *E.coli* BL21 (DE3) cells. Protein

expression was induced in 1 L TB medium at an optical density (OD_{600}) of 2.5 - 3 by addition

of 0.2 mM isopropyl-β-D-thiogalactopyranoside (IPTG) for 16 h at 20 °C. Cells were harvested

by centrifugation (10 min at 6,000 x g) and the pellets were suspended in binding buffer

(1x PBS, ad 0.5 M NaCl, 50 mM imidazole, 2 mM phenylmethylsulfonyl fluoride, 2 mM $MgCl_2$,

150 µg/mL lysozyme (Merck) and 625 µg/mL DNaseI (Applichem)). Cell suspensions were

sonified for 15 min (Bandelin Sonopuls HD70 - power MS72/D, cycle 50%) on ice, incubated

for 1 h at 4 °C in a rotary shaker followed by a second sonification step for 15 min. After

centrifugation (30 min at 20,000 x g), urea was added to a final concentration of 6 M to the

soluble protein extract. The extract was filtered through a 0.45 µm filter and loaded on a pre-

equilibrated 1-mL HisTrap[FF] column (GE Healthcare). The bound His-tagged nucleocapsid

proteins were eluted by a linear gradient (30 mL) ranging from 50 to 500 mM imidazole in

elution buffer (1x PBS, pH 7.4, 0.5 M NaCl, 6 M urea). Elution fractions (0.5 mL) containing the

His-tagged nucleocapsid proteins were pooled and dialyzed (D-Tube Dialyzer Mega, Novagen)

against PBS.

The viral S1-domains, SARS-CoV-2 RBD, and the stabilized trimeric SARS-CoV-2 Spike

protein were expressed in Expi293 cells following the protocol as described in Stadlbauer *et*

al.[8].

All purified proteins were analyzed via standard SDS-PAGE followed by staining with

InstantBlue Coomassie stain (Expedeon) and immunoblotting using an anti-His antibody

(Penta-His Antibody, #34660, Qiagen) in combination with a donkey-anti-mouse antibody

labeled with AlexaFluor647 (Invitrogen) on a Typhoon Trio (GE-Healthcare, Freiburg,

Germany; excitation 633 nm, emission filter settings 670 nm BP 30) to confirm protein integrity.

To further confirm correct expression, integrity, and purity, proteins were analysed by mass

spectrometry. To control the production reproducibility of the antigens, potential aggregation

and melting temperatures of the proteins were investigated by nano differential scanning

fluorimetry (nanoDSF) using a Prometheus (Nanotemper, Munich, Germany).

Commercial antigens

Two commercial antigens were used to complement the in-house-produced antigen panel.

The S2 ectodomain of the SARS-CoV-2 spike protein (aa 686 – 1213) was purchased from

Sino Biological, Eschborn, Germany (cat # 40590, lot # LC14MC3007). A full-length

nucleocapsid protein of SARS-CoV-2 was purchased from Aalto Bioreagents, Dublin, Ireland

(cat # 6404-b, lot # 4629).

Bead-based serological multiplex assay

All antigens were covalently immobilized on spectrally distinct populations of carboxylated

paramagnetic beads (MagPlex Microspheres, Luminex Corporation, Austin, TX) using 1-ethyl-

3-(3-dimethylaminopropyl)carbodiimide (EDC)/ sulfo-N-hydroxysuccinimide (sNHS) chemistry.

For immobilization, a magnetic particle processor (KingFisher 96, Thermo Scientific, Schwerte,

Germany) was used.

Bead stocks were vortexed thoroughly and sonicated for 15 seconds. Subsequently, 83 µL of

0.065% (v/v) Triton X-100 and 1 mL of bead stock containing 12.5×10^7 beads of one single

bead population were pipetted into each well. The beads were then washed twice with 500 µL

of activation buffer (100 mM Na_2HPO_4, pH 6.2, 0.005% (v/v) Triton X-100) and beads were

activated for 20 min in 300 µL of activation mix containing 5 mg/mL EDC and 5 mg/mL sNHS

in activation buffer. Following activation, the beads were washed twice with 500 µL of coupling

buffer (500 mM MES, pH 5.0, 0.005% (v/v) Triton X-100) and the antigens were added to the

activated beads and incubated for 2 h at 21 °C to immobilize the antigens on the surface.

Antigen-coupled beads were washed twice with 800 µL of wash buffer (1x PBS, 0.005 % (v/v)

Triton X-100) and were finally resuspended in 1,000 µL of storage buffer (1x PBS, 1 % (w/v)

BSA, 0.05% (v/v) ProClin). The beads were stored at 4°C until further use.

To detect human IgG and IgA responses against SARS-CoV-2 and the endemic human

coronaviruses (hCoV-NL63, hCoV-229E, hCoV-OC43 and hCoV-HKU1), the purified trimeric

Spike protein (S), S1-domain, S2-domain (Sino Biological GmbH, Europe), RBD, nucleocapsid

(N) and the N-terminal domain of nucleocapsid (N-NTD) of SARS-CoV-2 as well as the S1-

domain, N, and N-NTD of the endemic hCoVs were immobilized on different bead populations

as described above. The individual bead populations were combined into a bead mix. A bead-

based multiplex assay was performed. Briefly, samples were incubated at a 1:400 dilution for

2 hours at 21 °C. Unbound antibodies were removed and the beads were washed three times

with 100 µL of wash buffer (1x PBS, 0.05% (v/v) Tween20) per well using a microplate washer

(Biotek 405TS, Biotek Instruments GmbH). Bound antibodies were detected with R-

phycoerythrin labeled goat-anti-human IgG or IgA antibodies (incubation for 45 min at 21°C).

Measurements were performed using a Luminex FLEXMAP 3D instrument and the Luminex

xPONENT Software 4.3 (settings: sample size: 80 µL, 50 events, Gate: 7,500 – 15,000,

Reporter Gain: Standard PMT).

Data analysis

Data analysis and visualization was performed with R Studio (Version 1.2.5001, using R

version 3.6.1) using the Median Fluorescent Intensity (MFI). Statistical analysis was performed

using R package "stats" from the base repository. Mann-Whitney U test was used to determine

difference between signal distributions from different sample groups. Spearman's ρ coefficient

was calculated in order to correlate antigens by response from the entire sample set, followed

by hierarchical clustering to group antigens. Fishers' exact test was used to calculate

significance of overlap between sample groups.

Quality control

In order to test the repeatability of the MultiCoV-Ab assay three quality control samples (QCs) were processed in duplicate on each test plate (n = 17) during the sample screening and inter-assay variance was assessed for each antigen in the multiplex. For intra-assay variance, 24 replicates for each of the three QC samples were analyzed on one plate. Achieved results are presented in **Extended Data Table 1 and Extended Data Fig. 2**. A limit of detection (LOD) for each antigen was determined by processing a blank in 24 replicates and LOD was set as mean MFI + 3 standard deviations. Sample parallelism and comparability of paired serum and plasma samples was assessed over eight dilution steps ranging from 1:100 to 1:12,800 (**Extended Data Fig. 2**). A set of samples derived from 205 SARS-CoV-2-infected and 72 uninfected individuals was tested repeatedly with two different kit batches. The samples classification in both runs matched 100 %.

Samples

A total of 1176 sera and plasma samples were used for the MultiCoV-Ab assay development. Ethical approval was granted from the Ethics Committee of Hannover Medical School (#9122_BO_K2020). Only de-identified samples were used for the MultiCoV-Ab assay development. All samples were pre-existing. Cohort age was 5-88 years; age was not known for 161 samples.

310 samples were from COVID-19 patients or convalescents. Samples were classified as SARS-CoV-2 infected, if a positive SARS-CoV-2 RT-PCR was reported and/ or if hospitalization/ quarantine for COVID-19 was indicated as part of the samples metadata. ΔT defined as time between PCR test or symptom onset and blood draw was 0-73 days (median= 38 d; n=258). ΔT was not provided for 52 samples. SARS-CoV-2 infected samples used in this study were collected after ethical review (9001_BO_K, Hannover Medical School; 179/2020/BO2, University Hospital Tübingen; 85/20, Ärztekammer des Saarlandes).

866 control samples were from non-SARS-CoV-2 infected individuals and were classified as non-infected as they were obtained prior to the emergence of SARS-CoV-2 in December 2019

or because they were taken from individuals who had not reported cold symptoms since the

beginning of 2020.

The majority of non-SARS-COV-2 infected samples was randomly selected and consistent of

prepandemic blood donors, commercially available (Central BioHub GmbH, Berlin, Germany

and BBI Solutions, Crumlin, UK) or bio-banked specimens. 365 samples were from the

Memory and Morbidity in Augsburg Elderly (MEMO) study (a subcohort of the MONICA S2

cohort (WHO 1988)) and were included based on available serological titers for HSV-1, HSV-

2, HHV-6 and EBV[24]. 88 samples were obtained from transplanted patients with chronic

respiratory conditions.

Collection of non-SARS-CoV-2 infected control samples had been approved by several ethic

votes: 3232-2016 (Ethics Committee of Hannover Medical School); 62/20 (Ethics Committees

of the Medical Faculty of the Saarland University at the Saarland Ärztekammer); WUM

17.02.1997 (Joint ethics committee of the University of Muenster and the Westphalian

Chamber of Physicians),

Additional sample details can be found in Extended Data Table 2.

Data availability

Primary Data including raw MFI and sample annotation will be made available upon request.

Code availability

R Code for data analysis will be made available upon request.

References

1. Okba, N.M.A., et al. Severe Acute Respiratory Syndrome Coronavirus 2-Specific

Antibody Responses in Coronavirus Disease Patients. Emerg Infect Dis 26, 1478-1488

(2020).

2. Amanat, F., et al. A serological assay to detect SARS-CoV-2 seroconversion in humans.

Nat Med (2020).

3. Tan, W., et al. Viral Kinetics and Antibody Responses in Patients with COVID-19.

medRxiv, 2020.2003.2024.20042382 (2020).

4. Long, Q.X., et al. Antibody responses to SARS-CoV-2 in patients with COVID-19. Nat

Med 26, 845-848 (2020).

5. Amanat, F. & Krammer, F. SARS-CoV-2 Vaccines: Status Report. Immunity 52, 583-589

(2020).

6. Robbiani, D.F., et al. Convergent antibody responses to SARS-CoV-2 in convalescent

individuals. Nature (2020).

7. Lassaunière, R., et al. Evaluation of nine commercial SARS-CoV-2 immunoassays.

(medRxiv, 2020).

8. Stadlbauer, D., et al. SARS-CoV-2 Seroconversion in Humans: A Detailed Protocol for a

Serological Assay, Antigen Production, and Test Setup. Curr Protoc Microbiol 57, e100

(2020).

9. https://diagnostics.roche.com/global/en/products/params/elecsys-anti-sars-cov-2.html.

10. https://www.siemens-healthineers.com/de/laboratory-diagnostics/assays-by-diseases-

conditions/infectious-disease-assays/cov2t-assay.

11. https://www.coronavirus-diagnostik.de/antikoerpertestsysteme-fuer-covid-19.html.

12. Burbelo, P.D., et al. Detection of Nucleocapsid Antibody to SARS-CoV-2 is More

Sensitive than Antibody to Spike Protein in COVID-19 Patients. medRxiv (2020).

13. Ng, K., et al. Pre-existing and de novo humoral immunity to SARS-CoV-2 in humans.

bioRxiv, 2020.2005.2014.095414 (2020).

14. Sun, B., et al. Kinetics of SARS-CoV-2 specific IgM and IgG responses in COVID-19

patients. Emerg Microbes Infect 9, 940-948 (2020).

15. den Hartog, G., et al. SARS-CoV-2-specific antibody detection for sero-epidemiology: a

multiplex analysis approach accounting for accurate seroprevalence. medRxiv,

2020.2006.2018.20133660 (2020).

16. Wang, H., et al. SARS-CoV-2 proteome microarray for mapping COVID-19 antibody

interactions at amino acid resolution. bioRxiv, 2020.2003.2026.994756 (2020).

17. Nelde, A., et al. SARS-CoV-2 T-cell epitopes define heterologous and COVID-19-

induced T-cell recognition. (2020).

18. Liu, L., et al. Anti-spike IgG causes severe acute lung injury by skewing macrophage

responses during acute SARS-CoV infection. JCI Insight 4(2019).

19. https://science.apa.at/rubrik/medizin_und_biotech/Ischgl-

Studie_42_4_Prozent_sind_Antikoerper-

positiv/SCI_20200625_SCI39451352255218286.

20. Okba, N.M.A., et al. Particulate multivalent presentation of the receptor binding domain

induces protective immune responses against MERS-CoV. Emerg Microbes Infect 9,

1080-1091 (2020).

21. Yasui, F., et al. Prior immunization with severe acute respiratory syndrome (SARS)-

associated coronavirus (SARS-CoV) nucleocapsid protein causes severe pneumonia in

mice infected with SARS-CoV. J Immunol 181, 6337-6348 (2008).

22. Bolles, M., et al. A double-inactivated severe acute respiratory syndrome coronavirus

vaccine provides incomplete protection in mice and induces increased eosinophilic

proinflammatory pulmonary response upon challenge. J Virol 85, 12201-12215 (2011).

23. Kang, S., et al. Crystal structure of SARS-CoV-2 nucleocapsid protein RNA binding

domain reveals potential unique drug targeting sites. Acta Pharm Sin B (2020).

24. Zeeb, M., et al. Seropositivity for pathogens associated with chronic infections is a risk

factor for all-cause mortality in the elderly: findings from the Memory and Morbidity in

Augsburg Elderly (MEMO) Study. GeroScience (2020).

Acknowledgements

This work has received funding from the European Union's Horizon 2020 research and innovation programme under grant agreement No 101003480 – CORESMA as well as from the German Federal Ministry for Research and Education/the Helmholtz Association. We thank Florian Krammer providing us with expression plasmids for the Spike Trimer and RBD. We thank Shannon Layland for critical proofreading of the manuscript.

Author Contributions

M.B. designed and performed experiments and data analysis; M.S. designed experiments; D.J., J.H., S.F., F.R., planned and performed experiments; A.Z. performed mass spectrometry analysis; J.H. performed nanoDSF analyses; H.D., B.T., P.K., F.W., U.R. designed, cloned, expressed and purified the antigens; S.H. and A.P. performed sample analysis; T.B., A.B., S.L., S.S., M.C., T.I., H-G.R., A.N., J.W., M.T., T.O.J. arranged sample and data collection; M.T., T.O.J, G.K. supported the study planning; N.S-M. planned the assay development and validation and designed experiments; M.B., M.S., U.R. and N.S-M. wrote the manuscript. All authors reviewed the manuscript.

Competing interests

T.O.J is a scientific advisor for Luminex. N.S-M. was a speaker at Luminex user meetings in the past. The Natural and Medical Sciences Institute at the University of Tuebingen is involved in applied research projects as a fee for services with Luminex.

 Figure Legends

 Fig. 1. | MultiCoV-Ab assay, a sensitive and specific tool to monitor SARS-CoV-2

 antibody responses. a, Control sera (blue, n = 72) and sera from individuals with PCR-

 confirmed SARS-CoV-2 infection (red, n = 205) were screened in a multiplex bead-based

 assay using Luminex technology (MultiCoV-Ab) to quantify IgG or IgA responses to various

 antigens. Reactivity towards trimeric SARS-CoV-2 spike protein (Spike Trimer) or SARS-CoV-

 2 receptor binding domain of spike (RBD) was found to be the best predictor of SARS-CoV-2

 infection. Data are presented as Box-Whisker plots of sample median fluorescence intensity

 (MFI) on a logarithmic scale. Outliers determined by 1.5 times IQR of log-transformed data are

 depicted as circles. **b,** Sample set from **a,** was used to compare assay performance of the

 MultiCoV-Ab assay using combined Spike Trimer and RBD antigens with commercially

 available single analyte SARS-CoV-2 assays which detect total Ig (Elecsys Anti-SARS-CoV-2

 (Roche); ADVIA Centaur SARS-CoV-2 Total (COV2T) (Siemens Healthineers)) or IgG (Anti-

 SARS-CoV-2-ELISA - IgG (Euroimmun)) or IgA (Anti-SARS-CoV-2-ELISA - IgA (Euroimmun)).

 SARS-CoV-2 infection status of samples is indicated as SARS-CoV-2 positive (PCR) or

 negative. Antibody test results were classified as negative (blue), positive (red) or borderline

 (grey) as per the manufacturer's definition. Only samples with divergent antibody test results

 are shown.

Fig. 2. | Combination of 2 Spike protein variants and isotype profiling by multiplex assay increases accuracy to identify SARS-CoV-2 antibody positive individuals.

a, An extended sample set of SARS-CoV-2-uninfected (n = 866) and SARS-CoV-2-infected individuals (n = 310) was used to further validate our MultiCoV-Ab assay. Age, gender, SARS-CoV-2 infection status and hospitalization status of study population are shown. NA: information was not provided. **b,** MultiCoV-Ab assay sensitivity and specificity were calculated for IgA or IgG based on a single analyte or a combined cut-off of Spike Trimer and RBD (IgG or IgA overall). A cut-off combining IgG and IgA was calculated as well. **c-d,** Scatterplot detailing MultiCoV-Ab assay cut-offs. Signal to cut-off (S/CO) values are displayed for Spike Trimer against RBD on a logarithmic scale. For IgG (**c,**), cut-offs are visualized by straight lines and SARS-CoV-2-infected and uninfected samples are separated by color (black circles – SARS-CoV-2-uninfected; red circles – SARS-CoV-2-infected). For IgA (**d,**) cut-offs are visualized as dashed lines and S/CO of 2 used for the combined cut-off is shown as straight lines. SARS-CoV-2-infected samples are split into IgG-positives and -negatives by color as indicated in the plot. **e-f,** Scatterplots display IgG response to additional SARS-CoV-2 antigens contained in the MultiCoV-Ab panel: MFI for Spike subdomains S1 vs S2 (**e,**) or Nucleocapsid antigens N vs N-NTD (**f,**) are displayed on logarithmic scale. SARS-CoV-2-uninfected samples are distinguished from SARS-CoV-2-infected and MultiCoV-Ab classification into positives or negatives as indicated by color.

Fig. 3. | Multiplex-based seroprofiling allows in-depth characterization of SARS-CoV-2 antibody responses

a, Kinetic of SARS-CoV-2 antigen-specific IgA and IgG responses is shown for indicated days after symptom onset for six SARS-CoV-2-specific antigens for five different patients. **b-c**, Samples of SARS-CoV-2-infected individuals were analysed to identify antigen- and isotype-specific antibody responses based on hospitalization indicating disease severity (**b,**) or age (**c,**). Data is presented as Box-Whisker plots of sample MFI on a logarithmic scale. Outliers determined by 1.5 times IQR of log-transformed data are depicted as circles. p-value (Mann-Whitney U test, two-sided) is displayed at the top of the boxes, indicating differences between signal distribution for respective groups.

Fig. 4. | Correlation of endemic and SARS CoV-2 antibody responses

a, Correlation of IgG response for the entire sample set (n = 1176) is visualized as heatmap based on Spearman's ρ coefficient; dendrogram on the right side displays antigens after hierarchical clustering was performed. **b-c,** Immune responses (IgG and IgA) towards hCoV S1 (**b,**) and N (**c,**) proteins are presented as Box-Whisker plots of sample MFI on a logarithmic scale for SARS-CoV-2-infected (red, n = 310) and uninfected (blue, n = 866) individuals. Outliers determined by 1.5 times IQR of log-transformed data are depicted as circles. **d-e,** Relative levels of IgG-specific immune response towards hCoV S1 (**d,**) and N (**e,**) proteins are presented as Box-Whisker plots / stripchart overlays of log-transformed and per-antigen scaled and centred MFI for the sample subsets of Spike Trimer false positives (blue, n =17) and combined IgG + IgA false negatives (red, n = 31). **f,** From the entire study population, groups of α- or β-hCoV high and low responders were built as indicated. High responder were defined as samples with above average MFI values for S1 and N-specific IgGs of the respective hCoV clade. Low responders were defined with below MFI values, correspondingly. Responder groups (i) α-hCoV ↑, red, n = 233, (ii) β-hCoV ↑, green, n = 254, (iii) α-hCoV ↓, blue, n = 172 (iv) β-hCoV ↓, purple, n = 210 are shown as Box-Whisker plots of log-transformed and per-antigen scaled and centred MFI values across hCoV N and S1 antigens. Outliers determined by 1.5 times IQR are depicted as circles. The over- or under-representation of SARS-CoV-2 responders (SARS-CoV-2 +, n = 279, as determined by positive MultiCoV-Ab classification) within the four sample groups is visualized in Venn diagrams, stochastic significance was calculated using Fisher's exact test (two-sided).

	Extended Data Figure Legends

Extended Data Fig.1

SDS-PAGE analysis of the recombinant viral antigens used in this study. To test for purity and

integrity 1 - 2 µg of indicated recombinant proteins were boiled in reducing SDS-sample buffer

and subjected to a gradient (4 - 20%) SDS-PAGE followed by coomassie staining. SARS-CoV-

2_Spike, SARS-CoV-2_RBD and the S1-domains of SARS-CoV-2, hCoV-NL63, hCoV-229E,

hCoV-OC43 and hCoV-HKU1 were produced in ExpiHEK™ cells. Nucleocapsid (N) and N-

terminal domain of nucleocapsid (N-NTD) of SARS-CoV-2, hCoV-NL63, hCoV-229E, hCoV-

OC43 and hCoV-HKU1 were produced in *E.coli*.

Extended Data Fig.2

a, Three quality control (QC) samples, as well as a sample of assay buffer (blank sample) were

processed in duplicates on every plate. Performance across 17 assay runs is depicted and

mean and %CVs are shown on the left side. For plate 14, a processing error lead to exclusion

of one blank sample from this evaluation. **b,** To assess parallelism of signals from different

samples, 6 unique serum samples were processed over a dilution series of 8 steps from 1:100

to 1:12,800. For 3 samples, paired plasma (EDTA and/or Heparin) were available and

processed together. For IgG and IgA detection of Spike Trimer and RBD, MFI are plotted

against sample dilution. Color indicates unique sample and shapes indicate sample type.

Extended Data Fig.3

a-d, Scatterplots of sample set with defined SARS-CoV-2 infection status (infected: red, n=205;

uninfected: black, n=72) to compare performance of the MultiCoV-Ab Spike Trimer vs indicated

antigens of commercial SARS–CoV-2 test kits. Signals are depicted as Signal to cut-off ratios

(S/CO) on a logarithmic scale. Lines indicate the respective cut-off values as defined by the

manufacturer to determine positive and negative test results.

Extended Data Fig.4

a, Correlation of IgA response for the entire sample set (n=1176) is visualized as heatmap based on Spearman's ρ coefficient; dendrogram on the right side displays antigens after hierarchical clustering was performed. **b,** Immune response (IgG and IgA) towards hCoV N-NTD proteins are presented as Box-Whisker plots of sample MFI on a logarithmic scale for SARS-CoV-2-infected (red, n=310) and uninfected (blue, n=866) individuals. Outliers determined by 1.5 times IQR of log-transformed data are depicted as circles. **c,** Relative levels of IgG-specific immune response towards hCoV N-NTD proteins are presented as Box-Whisker plots / stripchart overlays of log-transformed and per-antigen scaled and centred MFI for the sample subsets of Spike Trimer false positives (blue, n=17) and combined IgG + IgA false negatives (red, n=31).

Extended Data Table 1

Intra- and inter-assay variance were determined by repeated measurement of QC samples and blank sample as replicates on one plate and in duplicates over 17 plates, respectively. Standard deviation relative to mean (%CV) is given for each antigen. A limit of detection (LOD) was calculated from 24 blank sample replicates on the same plate as the mean MFI + 3 times standard deviation.

Extended Data Table 2

Complete overview of study sample set. Samples are divided into columns by age groups and gender. NA: Information was not available. Samples from SARS-CoV-2-infected donors are further split up by hospitalization status. Age and gender of patients from which multiple samples were available for time course analyses are indicated. SARS-CoV-2-uninfected samples are further divided into samples drawn during the pandemic, which was defined as all samples taken on 01.01.2020 or later, and pre-pandemic samples. 147 samples with previous hCoV infection were included in the SARS-CoV-2-uninfected group. Detailed diagnosis of hCoV subspecies is indicated where available. Other sample conditions for special groups of uninfected samples are listed.

Figure 1

a

Spike Trimer

RBD

b

Figure 2

a

Age group	≤39		40-59		≥60		NA			Σ
n	299		241		475		161			1,176
Gender	M	F	M	F	M	F	M	F	NA	
n	139	160	144	97	271	204	5	3	153	1,176
SARS-CoV-2 infected	60	51	71	63	42	17	3	3	0	310
hospitalised	6	2	14	6	23	4	0	0	0	55
non-hospitalised	52	43	49	43	13	8	0	0	0	208
hospitalisation NA	2	6	8	14	6	5	3	3	0	47
SARS-CoV-2 uninfected	79	109	73	34	229	187	0	0	153	866

b

	Correctly classified		Sensitivity (%)	Specificity (%)
	infected	uninfected		
IgG Spike Trimer	277	849	89.35	98.04
IgG RBD	276	862	89.03	99.54
IgG overall	275	866	88.71	100.00
IgA Spike Trimer	272	850	87.74	98.15
IgA RBD	255	855	82.26	98.73
IgA overall	254	864	81.94	99.77
Combined IgG + IgA	279	866	90.00	100.00

c

IgG overall cut-off

d

Combined IgG + IgA cut-off

e

Spike subdomains

f

Nucleocapsid antigens

Figure 3

Figure 4

Extended Data - Table 1

			SARS-CoV-2						hCoV NL63			hCoV 229E			hCoV OC43			hCoV HKU1		
			Spike Trimer	RBD	S1	S2	N	N-NTD	S1	N	N-NTD	S1	N	N-NTD	S1	N	N-NTD	S1	N	N-NTD
Inter-assay variance (%CV) n = 34, duplicates, 17 plates	IgG	QC1	3.7	3.3	3.4	3.7	2.8	7.4	3.5	3.2	4.4	3.3	2.8	5.2	3.1	6.0	4.4	3.4	4.7	5.4
		QC2	4.1	4.6	6.9	3.4	5.3	4.8	3.0	2.2	6.3	2.4	2.1	6.7	2.7	4.5	2.3	2.7	5.1	2.8
		QC3	3.4	2.4	2.3	3.6	2.1	4.6	3.1	2.5	3.5	2.9	2.0	4.7	2.9	6.4	4.6	3.2	3.2	3.5
		Blank	5.4	5.6	6.7	6.4	5.6	6.1	6.3	7.1	5.7	9.1	6.1	6.1	5.6	7.3	4.1	4.9	6.1	8.3
	IgA	QC1	3.9	4.6	4.9	4.0	5.1	5.0	4.2	3.6	3.9	4.0	5.3	7.4	4.3	7.4	5.0	4.2	6.0	5.0
		QC2	4.6	5.0	5.1	3.9	3.9	4.2	3.7	2.4	7.6	2.9	2.2	6.0	4.1	16.4	4.2	4.3	5.5	3.9
		QC3	3.9	3.8	4.5	3.4	3.4	4.8	3.9	2.8	4.0	3.0	5.1	4.5	3.6	6.1	4.1	3.7	4.5	5.1
		Blank	6.7	5.3	8.2	6.3	5.3	5.3	3.3	5.0	5.0	6.7	7.0	6.1	5.3	7.1	4.7	6.0	6.8	6.3
Intra-assay variance (%CV) n = 24	IgG	QC1	2.5	1.9	2.0	2.1	1.8	2.1	2.4	1.7	2.8	2.0	2.7	3.2	1.9	2.0	2.2	2.7	2.4	2.2
		QC2	5.9	4.3	4.1	2.8	2.7	3.2	1.9	1.9	2.6	2.0	2.2	2.7	2.2	1.6	2.1	2.5	3.1	2.5
		QC3	1.6	4.3	5.1	1.9	1.9	4.5	4.2	1.7	3.2	3.3	4.1	5.7	3.2	3.1	5.5	5.6	6.0	8.4
		Blank	6.0	5.6	5.2	5.8	5.2	4.2	5.0	4.8	4.8	7.3	6.2	6.3	6.5	6.2	4.4	6.1	6.0	6.2
	IgA	QC1	2.5	3.3	5.2	3.8	3.7	4.2	3.2	2.3	2.2	2.0	4.8	4.7	2.9	4.7	3.4	3.3	4.5	4.3
		QC2	4.8	5.7	5.7	3.2	4.1	4.3	3.4	2.0	5.7	2.1	2.1	6.1	3.0	3.1	1.9	3.9	6.4	3.8
		QC3	3.1	4.7	5.5	3.0	4.1	4.4	3.7	2.7	3.7	2.7	5.4	5.8	2.6	4.1	3.1	2.4	4.5	4.3
		Blank	5.8	5.3	6.3	5.0	5.4	5.5	4.5	5.6	5.3	7.2	6.3	7.0	6.7	9.5	7.3	5.2	8.8	7.0
LOD* (MFI)	IgG		32	26	23	29	38	33	29	28	26	65	35	24	33	30	33	37	33	25
n = 24	IgA		31	26	26	27	37	32	57	28	40	28	36	22	35	28	32	33	39	28

*Blank mean MFI + 3 * sd

Extended Data - Table 2

Age	≤39		40-59		≥60		NA			Σ
n	299		241		475		161			1,176
Gender	M	F	M	F	M	F	M	F	NA*	
n	139	160	144	97	271	204	5	3	153	1,176
SARS-CoV-2-infected (total)	60	51	71	63	42	17	3	3	0	310
Hospitalized (for COVID19)	6	2	14	6	23	4	0	0	0	55
Non-Hospitalized	52	43	49	43	13	8	0	0	0	208
Hospitalisation NA	2	6	8	14	6	5	3	3	0	47
Patients with time series	2	0	0	0	2	1	0	0	0	5
SARS-CoV-2-uninfected (total)	79	109	73	34	229	187	2	0	153	866
Sample during pandemic	10	10	12	14	7	5	1	0	6	65
Sample pre-pandemic	69	99	61	20	222	182	1	0	147	801
Previous hCoV Infection	19	18	45	20	29	16	0	0	0	147
confirmed NL63	2	0	3	1	2	2	0	0	0	10
confirmed 229	5	1	4	1	5	4	0	0	0	20
confirmed OC43	0	1	14	1	6	5	0	0	0	27
confirmed HKU1	3	1	4	2	5	0	0	0	0	15
unknown hCoV	9	15	20	15	11	5	0	0	0	75
Pregnant	0	9	0	1	0	0	0	0	0	10
RF/HAMA samples	0	0	0	0	0	0	0	0	6	6
PCT > 3 ng/mL	0	0	0	0	0	0	0	0	21	21
Neuroinflammatory disease	6	6	1	0	1	1	0	0	0	15

Extended Data - Figure 1

Extended Data - Figure 2

a

QC performance − SARS-CoV-2 RBD

QC performance − SARS-CoV-2 Spike Trimer

b

Parallelism IgG Detection
SARS-CoV-2 Spike Trimer

Parallelism IgA Detection
SARS-CoV-2 Spike Trimer

Parallelism IgG Detection
SARS-CoV-2 RBD

Parallelism IgA Detection
SARS-CoV-2 RBD

Extended Data - Figure 3

Extended Data Figure 4

a

Correlation of IgA Signal

b

c