Evidence for sustained mucosal and systemic antibody responses to SARS-CoV-2 antigens in COVID-19 patients

Authors: Baweleta Isho[1]†, Kento T. Abe[2,3]†, Michelle Zuo[1]†, Alainna J. Jamal[4,5]†, Bhavisha Rathod[2], Jenny H. Wang[2], Zhijie Li[3], Gary Chao[1], Olga L. Rojas[1], Yeo Myong Bang[1], Annie Pu[1], Natasha Christie-Holmes[6], Christian Gervais[7], Derek Ceccarelli[2], Payman Samavarchi-Tehrani[2], Furkan Guvenc[3], Patrick Budylowski[6,8], Angel Li[5], Aimee Paterson[5], Yue Feng Yun[1], Lina G. Marin[9], Lauren Caldwell[2], Jeffrey L. Wrana[2,3], Karen Colwill[2], Frank Sicheri[2,3], Samira Mubareka[10], Scott D. Gray-Owen[3,6], Steven J. Drews[11], Walter L. Siqueira[9], Miriam Barrios-Rodiles[2], Mario Ostrowski[1,12,13], James M. Rini[3,14], Yves Durocher[7], Allison J.McGeer[2,4,5], Jennifer L. Gommerman[1]* and Anne-Claude Gingras[2,3]*

Affiliations:

[1] *Department of Immunology, University of Toronto, Toronto, ON, Canada*

[2] *Lunenfeld-Tanenbaum Research Institute at Mount Sinai Hospital, Sinai Health System, Toronto, ON, Canada*

[3] *Department of Molecular Genetics, University of Toronto, Toronto, ON, Canada*

[4] *Institute of Health Policy, Management and Evaluation, University of Toronto, Toronto, ON, Canada*

[5] *Department of Microbiology, at Mount Sinai Hospital, Sinai Health System, Toronto, ON, Canada*

[6] *Combined Containment Level 3 Unit, University of Toronto, Toronto, ON, Canada*

[7] *Mammalian Cell Expression, Human Health Therapeutics Research Centre, National Research Council Canada, Montréal, QC, Canada*

[8] *Institute of Medical Science, University of Toronto, Toronto, ON, Canada*

[9] *College of Dentistry, University of Saskatchewan, Saskatoon, SK, Canada*

[10] *Department of Laboratory Medicine and Molecular Diagnostics, Division of Microbiology, Sunnybrook Health Sciences Centre; Biological Sciences, Sunnybrook Research Institute; and Division of Infectious Diseases, Sunnybrook Health Sciences Centre, Toronto, ON, Canada; Department of Laboratory Medicine and Pathology, University of Toronto, Toronto, ON, Canada*

[11] *Canadian Blood Services, Edmonton, AB & Department of Laboratory Medicine and Pathology, University of Alberta, Edmonton, AB, Canada*

[12] *St. Michael's Hospital, Toronto, ON, Canada; Li Ka Shing Knowledge Institute*

[13] *Department of Medicine, University of Toronto, Toronto, ON, Canada*

[14] *Department of Biochemistry, University of Toronto, Toronto, ON, Canada*

† These authors contributed equally

*Correspondence to: jen.gommerman@utoronto.ca; gingras@lunenfeld.ca

Abstract: While the antibody response to SARS-CoV-2 has been extensively studied in blood, relatively little is known about the mucosal immune response and its relationship to systemic antibody levels. Since SARS-CoV-2 initially replicates in the upper airway, the antibody response in the oral cavity is likely an important parameter that influences the course of infection. We developed enzyme linked immunosorbent assays to detect IgA and IgG antibodies to the SARS-CoV-2 spike protein (full length trimer) and its receptor binding domain (RBD) in serum (n=496) and saliva (n=90) of acute and convalescent patients with laboratory-diagnosed COVID-19 ranging from 3-115 days post-symptom onset (PSO), compared to negative controls. Anti-CoV-2 antibody responses were readily detected in serum and saliva, with peak IgG levels attained by 16-30 days PSO. Whereas anti-CoV-2 IgA antibodies rapidly decayed, IgG antibodies remained relatively stable up to 115 days PSO in both biofluids. Importantly, IgG responses in saliva and serum were correlated, suggesting that antibodies in the saliva may serve as a surrogate measure of systemic immunity.

One Sentence Summary: In this manuscript, we report evidence for sustained SARS-CoV-2-specific IgG responses both at the site of infection and systemically in COVID-19 patients and suggest that saliva could be used as an alternative biofluid for monitoring IgG to SARS-CoV-2 spike and RBD antigens.

Main Text:

Antibodies play an important role in neutralizing virus and provide protection to the host against viral re-infection. The antibody response to SARS-CoV-2 infection has been extensively studied in the blood (serum, plasma) of COVID-19 patients in order to gain insights into the host immune response. Antibody levels to the spike protein are particularly important since this large trimeric glycoprotein harbours the receptor binding domain (RBD). The RBD facilitates SARS-CoV-2 access to human cells by binding to its counter receptor angiotensin converting enzyme 2 (ACE-2) (*1*), and neutralizing antibodies have been shown to target the RBD (*2*). Most studies agree that the IgG antibodies to SARS-CoV-2 spike and RBD antigens are detected in the blood of greater than 90% of subjects by 10-11 days post-symptom onset (PSO) (*3-7*). However, whether levels of IgG specific for SARS-CoV-2 antigen persist (*8, 9*), or alternatively decay (*10*) remains a debated issue. Examination of different biofluids from multiple cohorts, and attention to the antigens tested, is required to resolve this extremely important issue that has high relevance to vaccine design.

Another gap in our knowledge is that we know very little about the local antibody response at the site of infection. SARS-CoV-2 enters the naso- and oro-pharyngeal tracts where it subsequently replicates (*11*). For this reason, nasopharyngeal and throat swabs are used to test for virus using reverse transcriptase quantitative PCR (RT-qPCR) to detect viral RNA. However, saliva has also been shown to be an effective biofluid for testing for the presence of SARS-CoV-2 mRNA (*12-15*). This makes sense given that pharyngeal SARS-CoV-2 shedding precedes viral replication in the lungs (*11*), and, like cytomegalovirus (*16, 17*), the salivary glands themselves can be a reservoir for the virus (*18*). Yet in spite of the oral cavity being a site for viral replication, only limited studies have examined anti-SARS-CoV-2 antibodies in this compartment (*19, 20*).

In this study, we examined the anti-SARS-CoV-2 antibody response over a 115-day period in the serum and saliva of n=496 (serum) and n=90 (saliva) samples from patients with COVID-19, compared to controls. Antigen-specific IgG in both biofluids were maximally detected by 16–30 days PSO and did not drastically decline in their relative levels as late as 100-115 days PSO. In contrast, antigen-specific IgA was rapidly induced but subsequently declined in both serum and saliva; the IgA response was however weakly correlated to age and disease severity. Importantly, IgG levels against both antigens were strongly correlated across paired serum and saliva samples (n=72), indicating that saliva can be used for monitoring the immune response to SARS-CoV-2 infection. Taken together, the systemic and mucosal IgG response to SARS-CoV-2 is sustained over a 3-month period, while the IgA response occurs early and is transient.

Results

A chemiluminescent fully automated method for detecting antibodies to SARS-CoV2 antigens in the serum of acute and convalescent patients.

To study the antibody response to SARS-CoV-2, we initially focused on class switched antibodies (IgG, IgA) to the spike homotrimer and the RBD, since neutralizing antibodies are directed to the spike protein (*21*). Enzyme-linked immunosorbent assays (ELISAs) for the detection in serum (or plasma) of anti-spike trimer and anti-spike RBD IgG were built as in (*3, 22*) as 96-well colorimetric assays, and implemented as automated 384-well chemiluminescence assays (see Methods). Receiver-Operating Characteristic (ROC) curves were generated on cohorts of true negatives (banked samples collected pre-COVID, n=339 for manual and automated assays) and positives (convalescent patients with confirmed PCR diagnostic, n=402 for manual and automated assays, see Table 1). For manual and automated IgG assays, sensitivities of 95.6% and 95.5% for spike and 93.8% and 91.3% for RBD, respectively, at a false positive rate of ≤1%, were obtained in these cohorts (Fig S1A-B, and Table S1 for ROC statistics). The Areas Under the Curves (AUCs) were ≥0.97 in all cases, indicating excellent assay performance. Automated assays for the detection of IgA (and IgM) were also developed (Fig S1C-D). The results for the automated and manual IgG assays were well correlated (Fig S1E-F).

These automated ELISA assays were used to profile cohorts of confirmed acute and convalescent sera from COVID-19 patients collected as part of COVID-19 surveillance by the Toronto Invasive Bacterial Diseases Network (Table 1). As expected based on the ROC analysis, the convalescent and pre-COVID controls had very different ratio distributions for both antigens (Fig. 1A-B). On the other hand, serum collected from patients less than 21 days post-onset of symptoms (acute serum, n=132) had bimodal distributions in their IgG responses for both antigens (with an overall lower mean), suggesting that antibody concentrations were increasing over time. To compare the relationship between RBD and spike trimer IgG levels, we plotted their values against each other. While there was an overall high correlation between the antigens (Fig. 1E), we noted many more acute specimens with high spike-trimer and low RBD response than vice versa, consistent with the fact that RBD is included within the spike trimer antigen. The concentration of IgA in convalescents serum was also clearly higher than that of the pre-COVID samples, but the acute cases had a higher median than the convalescents (Fig. 1C-D). The IgA levels to RBD and spike were also well correlated (Fig. 1F).

The bimodal distribution of the IgG responses in the acute serum (Fig. 1 A-B), along with the different patterns of response for IgG versus IgA in acute and convalescent specimens (Fig. 1 A-D), prompted us to plot the antibody levels against days PSO. Spearman rank correlation analysis revealed an overall increase in the IgG response versus a decrease in the IgA response to both antigens over time, and the IgG response in particular did not appear to be linear (compare panels A-B to C-D in Fig S2; IgG results were reproduced in the analysis of the manual IgG assays, shown in panels F-G). To look at this response more closely, specimens were binned by days PSO (15-day intervals), and the levels of the different immunoglobulins were plotted (the pre-COVID negative control samples were plotted alongside for comparison; Fig. 2). As was reported in other studies (3, 4, 7), the IgG levels reached peak in the 16–30 days bin, and the levels of IgG against spike trimer were sustained over 115 days (<7.3% change in the median as compared to the maximum; Fig. 2A). However, IgG levels against RBD showed a ~25.3 % decrease by day 105, and ~46.0 % by day 115 (Fig. 2B). The behavior of IgA to both antigens was by contrast much less sustained: after reaching a maximum in the 16–30 days bin, there was a clear and continuous decline throughout the time series such that by 115 days, the anti-spike and anti-RBD IgA levels were ~74.1 % and ~84.2 % of their respective maximal levels (Fig. 2 C-D). Multivariable analyses adjusting for severity of illness, sex, and patient age, did not change conclusions about the aforementioned relationships between time PSO and anti-RBD IgA, anti-spike IgA, and anti-RBD IgG; however, the modest decline in anti-spike IgG after day 35 was statistically significant data not shown).

To dissect these results, we analyzed pairs of serum samples from the hospitalized patients (n=58), collected at admission, and 3-12 weeks later, and performed a non-parametric loess analysis (as in (23)). These results depict a relative stability of the IgG anti-spike trimer levels, a partial decrease in the anti-RBD IgG and anti-spike IgA levels, and a near complete loss in the anti-RBD IgA levels over time (Fig. 3).

Although our focus was on the spike protein, we also examined the antibody response to nucleocapsid (a.k.a. nucleoprotein, NP), since this is the antigen targeted by multiple commercial assays. We developed an assay using bacterially-expressed NP (Fig S3A-B). When we examined the levels of anti-NP antibodies binned for time PSO, we found that their patterns closely resembled those for anti-spike and anti-RBD IgG and IgA responses, namely a relative stability in the IgG and more rapid decline in IgA levels in both the binned time series and the longitudinal series (Fig S3C-F).

In summary, our automated chemiluminescent ELISA represents a high-throughput option with excellent assay performance to measure antibody levels to SARS-CoV-2 antigens in serum (spike, RBD, NP) which mirror those from manual assays. Moreover, in a large cross-sectional survey, IgG, but not IgA, levels persisted for at least 3 months PSO for all antigens measured, with the levels of antibodies to the spike trimer being more stable over time than those to the RBD and NP.

Antibodies to SARS-CoV-2 antigens are detected in the saliva of COVID-19 patients.

While our serum-based assays are scalable and robust, saliva represents a relatively unexplored biofluid for detecting antibodies to SARS-CoV-2 antigens with many practical benefits, including being non-invasive and the capacity for self-collection at home. The disadvantage of saliva as a biofluid is its very low concentration of antibodies (24), making it

necessary to optimize the sensitivity of detection. We explored various detection methods and found that plating biotinylated antigen onto streptavidin coated plates was required to obtain reliable signal-to-noise ratios; the method also required that the saliva be pre-adsorbed to remove any streptavidin-binding protein (data not shown). While heat (65°C for 30 min) prevented detection of antibodies in the saliva, incubation with Triton X-100 was compatible with our assay (Fig S4) and resulted in viral inactivation (Table S2). Bolstered by these findings, we measured antibody levels to SARS-CoV2 antigens in n=54 COVID-19 patients (cohort 1), comparing to unexposed negative controls collected locally during the pandemic (n=42). We first performed a pilot experiment, using expectorated saliva samples acquired during the early phase of the pandemic. Since these samples were diluted to varying degrees, we normalized values to total IgG/IgA (depending on the isotype assay) or to albumin levels as done before by others (25). Saliva samples from COVID-19 patients displayed a significantly higher level of IgG and IgA levels to spike and RBD compared to negative controls when normalized with either method (Fig S4).

Following this pilot experiment, we proceeded with further saliva collections using a standardized collection method without a diluent (cohort 2). In cohort 2, we obtained n=90 samples from 80 patients ranging in time PSO from day 3–104. These were compared to 50 unexposed negative controls for cohort 2, of which 42 were also negative controls for cohort 1. To these negative controls, we also added pre-COVID era saliva samples as an additional comparator (n=27). Total IgG levels were found to be higher in COVID-19 patients compared to controls (Fig. 4A-B). Moreover, cohort 2 exhibited statistically significant differences between the relative levels of IgA and IgG antibodies specific to spike and RBD antigens compared to saliva from negative controls (Fig. 4C-F). The sensitivity of the saliva assays for IgG antibodies to spike and RBD (at a false discovery rate <2%) were 89% and 85%, respectively, while the sensitivity of the assays for IgA antibodies to spike and RBD were 51% and 30%, respectively (Fig S6 and Table S3). The lower sensitivity of the IgA assays is attributed in part to the higher levels of anti-spike and anti-RBD IgA levels in the negative controls (see Discussion). Next, we examined the levels of anti-spike and anti-RBD antibodies in our cross-sectional cohort over time PSO. Similar to the serum data, IgG levels in saliva to the spike and RBD antigens remained stable throughout the 3-month collection period. In contrast, significant decreases were observed for IgA levels to spike and RBD (ρ=-0.307 and ρ=-0.296, respectively); by day 100, anti-spike and anti-RBD IgA levels were barely detectable (Fig. 5A-D). Multivariable analysis adjusting for disease severity, age, and sex did not change conclusions about the aforementioned relationships between time PSO and all antibody levels (data not shown). In summary, infection with SARS-CoV-2 results in detectable IgG and IgA response in saliva against the spike and RBD antigens, with only the IgG response persisting beyond day 60.

Association of age, sex and disease severity with IgG and IgA responses to the spike and RBD antigens in serum and saliva

We next asked if clinical predictors such as age, sex, and disease severity were related to antibody responses in the acute and convalescent periods overall. In serum, older age was associated with higher IgA levels to both antigens (Fig S7 and Table 2) and with higher IgG-RBD but not IgG-spike in bivariate analyses. Consistent with a separate study (26), greater disease severity was significantly associated with higher serum levels of IgA and IgG to both

antigens in bivariate analyses (Fig S8 A-D; Table 2). Serum IgA levels to both antigens were higher in men than women, but IgG levels did not differ (Fig S8 E-H; Table 2).

In multivariable analysis including all serum samples, older age and greater severity of illness were independently associated with higher IgG and IgA levels to both antigens; in addition, men had significantly higher levels of IgA to both antigens (data not shown). There was a significant interaction between age and severity of illness for IgG to both antigens: as demonstrated in Figure 6A, patients with mild disease has significantly higher serum antibody levels in older versus younger adults, while this relationship was not observed in patients with moderate or severe disease.

In bivariate (Table 2) and multivariable analyses of saliva samples, there were no significant associations between clinical predictors and any antibody levels. In analyses limited to convalescent saliva specimens, older age and more severe illness were significantly associated with higher levels of anti-RBD IgA and IgG, and anti-spike IgG (data not shown).

Antibody levels to SARS-CoV-2 antigens in the serum correlate with those in the saliva

As mentioned, saliva has many advantages for biofluid collection over serum. To assess whether saliva might be reliably used in a diagnostic test, we determined whether the IgA and IgG levels to spike and RBD in the saliva correlated with those measured in the serum (Fig. 7A-D). Of the COVID-19 patients analyzed, n=72 had paired saliva and serum samples taken at similar timepoints (i.e. within 4 days). When comparing the saliva AUC values to the ratios in serum, we observed a significant positive correlation between saliva and serum for each antigen-antibody combination, with the strongest correlation observed for anti-RBD IgG ($\rho=0.72$) followed by anti-spike IgG ($\rho=0.56$), On the other hand, the correlation between the levels of serum and saliva anti-spike and anti-RBD IgA was more modest ($\rho=0.55$ and 0.40 respectively). Therefore, at least for anti-spike and anti-RBD IgG measurements, saliva may represent a good alternative for antibody testing.

Discussion

Antibodies are key components in the arsenal of protective immunity against novel viral infections such as SARS-CoV-2. Understanding their durability and their system compartmentalization across a diverse population are critical pieces of information informing our ability to monitor seroprevalence in communities, to select plasma donors for treatment, and to design vaccines against COVID-19. We examined the stability of antibody levels over the first three months after infection in both the serum and the saliva and longitudinal sampling of serum. We observed no drastic decline in levels of anti-spike, anti-RBD or anti-NP IgG levels over a 3-month period. The same was true for the antigen-specific measurements in saliva (anti-spike and anti-RBD IgG). On the other hand, similar to other findings (*26, 27*), IgA responses to SARS-CoV-2 antigens were found to decline in both serum and saliva. In summary, our data show that a durable IgG response against SARS-CoV-2 antigens is generated in both the saliva and serum in most patients with COVID-19, and that there are some unique behaviors of the IgA response that may suggest an independent compartmentalized immune response.

Given the presence of SARS-CoV-2 RNA in saliva, it is reasonable to hypothesize that, like other viruses such as rubella (*24*), 229E alpha-coronavirus (*28*), and MERS beta-coronavirus (*29*), the mucosae and draining lymph nodes of the oro- and nasopharyngeal tracts serve as a site for initiation of an immune response to SARS-CoV-2. If so, then plasmablasts (PB) and plasma cells (PC) that produce antibodies to SARS-CoV-2 will migrate back to the oro- and nasopharyngeal mucosae and produce antibodies that should be detectable in the saliva, a fluid that already has high levels of IgA (*30*). With time, this response will be detected in the systemic circulation, possibly due to migration of PC into new niches as we have previously described in mice (*31*). Indeed, we and two other groups have observed SARS-CoV-2 specific antibodies in saliva (*19, 20*). There are some variations between study protocols that are important to consider: Randad *et al.* applied a brush on the gum line as a means to capture IgG from the blood, heat inactivated this material, and performed Luminex to detect antigen-specific antibody levels (*20*). In contrast, our strategy was to collect saliva in a manner that best approximates the immune response that takes place in the local mucosa. In this way, our study more resembles that of Faustini *et al.*, who used ELISA technology on whole saliva, amplifying the signal with an additional antibody step (*19*). Although the saliva dilutions we used closely match those of Faustini *et al*, unlike our findings, agreement between the serum and saliva for each antibody/antigen ELISA pair was less obvious in that study than in ours (*19*). Whether these discrepancies are methodological (i.e. detection of specific versus total Igs) and/or relate to the higher number of asymptomatic subjects in the Faustini *et al.* study remains to be determined.

While the sensitivity of the saliva assays was very good for anti-spike and anti-RBD IgG responses based on ROC curves, this was less true for IgA, particularly the anti-RBD IgA response. This is because some of our negative controls, irrespective of whether they were collected during the pandemic (unexposed negatives) or prior to the pandemic, exhibit AUC levels of anti-RBD IgA that approach 50% of the pooled control saliva AUC (see Fig. 4D). It is unclear why this would occur for only the IgA/RBD combination. Presumably these are cross-reactive IgA that bind to SARS-CoV-2 RBD. Of interest, thus far SARS-CoV-2 neutralizing antibodies appear to have limited somatic hypermutation (*32, 33*), suggesting that they may originate from a naïve repertoire or from B cells that have been activated in extrafollicular responses where somatic hypermutation is limited. It is tempting to speculate that these pre-existing IgA antibodies may provide some stop-gap protection against SARS-CoV-2 in the oral cavity, and if so, it is essential to ascertain their original antigenic specificity. Future work is required to confirm these results in a greater array of subjects and using different sources of RBD antigen.

Our findings that the IgG response to SARS-CoV-2 antigens is stable over a 3-month period are consistent with two other studies who likewise noted durability in the IgG response to the spike trimer (*8, 9*). These data and ours contrast with those of Long *et al.*, who showed rapid decay of antibody levels when profiling the response to a linear peptide motif of the C-terminal part of the spike protein (*10*) instead of the spike trimer used here, and it is possible that the antigen selection accounts for some of the differences. However, this does not explain discrepant results with respect to the anti-NP response in the serum, which we find also persisted over the 3-month period. One potential difference that could explain these divergent results is that we employed a sensitive and robust chemiluminescence-based ELISA whereas Long *et al*. employed Luminex methods.

One weakness of our study is that we have not looked beyond the day 115 PSO – our collections began in mid-March 2020 – and it is entirely plausible that antigen-specific IgG levels will eventually wane with time. Whether this will translate to a reduction in neutralizing antibodies (nAb), as observed by Long *et al.* (*10*) but not by Wajnberg *et al.* (*9*), is unclear. One possible explanation for the (relatively mild) decline in the nAb response observed by Long *et al.* may be due to the rapid drop in antigen-specific IgA levels. Indeed, IgA is an important mediator of protection against gastrointestinal viruses (*34*), is essential in achieving immunity against avian viruses (*35*), and has been shown to contribute to the nAb response to SARS-CoV-2 as shown by Sterlin *et al.* (*27*). Although Sterlin *et al.* show that the initial IgA PB response quickly declines, IgA-producing PC have been shown to persist for decades in the mucosae of humans (*36*), and these will not be readily measurable in the blood. When combined with the parallel formation of re-activatable memory B cells (*37*), many of which will be tissue-resident (*23*), the host has excellent mechanisms for mounting swift and robust humoral immunity upon pathogen re-exposure that may be missed using blood-based measurements. An epidemiological study that prospectively follows confirmed COVID-19 cases for several months will determine if these immunological principals hold true in the context of SARS-CoV-2 infection.

In conclusion, our study provides evidence that the IgG response to SARS-CoV-2 spike persists in the saliva and the serum, and that this response can be correlated between the two biofluids. Given that SARS-CoV-2 initially replicates in the oro- and nasopharyngeal tracts, in the future it will be critical to characterize the nature and kinetics of salivary antibodies at the earliest time points post-infection in contact-traced individuals in order to determine if there are correlates of protection that impact viral setpoint and COVID-19 disease progression.

References and Notes:

1.   M. Letko, A. Marzi, V. Munster, Functional assessment of cell entry and receptor usage for SARS-CoV-2 and other lineage B betacoronaviruses. *Nat Microbiol* 5, 562-569 (2020).

2.   J. D. Berry *et al.*, Neutralizing epitopes of the SARS-CoV S-protein cluster independent of repertoire, antigen structure or mAb technology. *MAbs* 2, 53-66 (2010).

3.   F. Amanat *et al.*, A serological assay to detect SARS-CoV-2 seroconversion in humans. *Nat Med* 26, 1033-1036 (2020).

4.   Q. X. Long *et al.*, Antibody responses to SARS-CoV-2 in patients with COVID-19. *Nat Med* 26, 845-848 (2020).

5.   L. Premkumar *et al.*, The receptor binding domain of the viral spike protein is an immunodominant and highly specific target of antibodies in SARS-CoV-2 patients. *Sci Immunol* 5, eabc8413 (2020).

6.   J. Zhao *et al.*, Antibody responses to SARS-CoV-2 in patients of novel coronavirus disease 2019. *Clin Infect Dis*, ciaa344 (2020).

7.   N. M. A. Okba *et al.*, Severe Acute Respiratory Syndrome Coronavirus 2-Specific Antibody Responses in Coronavirus Disease Patients. *Emerg Infect Dis* 26, 1478-1488 (2020).

8.   A. S. Iyera *et al.*, Dynamics and significance of the antibody response to SARS-CoV-2 infection. *MedRxiv* https://doi.org/10.1101/2020.07.18.20155374, (2020).

9.   A. Wajnberg *et al.*, SARS-CoV-2 infection induces robust, neutralizing antibody responses that are 1 stable for at least three months. *MedRxiv* https://www.medrxiv.org/content/10.1101/2020.07.14.20151126v1, (2020).

10.   Q. X. Long *et al.*, Clinical and immunological assessment of asymptomatic SARS-CoV-2 infections. *Nat Med*, https://doi.org/10.1038/s41591-020-0965-6 (2020).

11.   R. Wolfel *et al.*, Virological assessment of hospitalized patients with COVID-2019. *Nature* 581, 465-469 (2020).

12.   K. K. To *et al.*, Temporal profiles of viral load in posterior oropharyngeal saliva samples and serum antibody responses during infection by SARS-CoV-2: an observational cohort study. *Lancet Infect Dis* 20, 565-574 (2020).

13.   K. K. To *et al.*, Consistent detection of 2019 novel coronavirus in saliva. *Clin Infect Dis*, https://doi.org/10.1093/cid/ciaa149 (2020).

14.   R. Sabino-Silva, A. C. G. Jardim, W. L. Siqueira, Coronavirus COVID-19 impacts to dentistry and potential salivary diagnosis. *Clin Oral Investig* 24, 1619-1621 (2020).

15.   Z. Khurshid, F. Y. I. Asiri, H. Al Wadaani, Human Saliva: Non-Invasive Fluid for Detecting Novel Coronavirus (2019-nCoV). *Int J Environ Res Public Health* 17, 2225 (2020).

16.   I. R. Humphreys *et al.*, Cytomegalovirus exploits IL-10-mediated immune regulation in the salivary glands. *J Exp Med* 204, 1217-1225 (2007).

17.   A. E. Campbell, V. J. Cavanaugh, J. S. Slater, The salivary glands as a privileged site of cytomegalovirus immune evasion and persistence. *Med Microbiol Immunol* 197, 205-213 (2008).

18.   J. Xu, Y. Li, F. Gan, Y. Du, Y. Yao, Salivary Glands: Potential Reservoirs for COVID-19 Asymptomatic Infection. *J Dent Res* 99, 989 (2020).

19.   S. E. Faustini *et al.*, Detection of antibodies to the SARS-CoV-2 spike glycoprotein in both serum and saliva enhances detection of infection. *medRxiv*, https://www.medrxiv.org/content/10.1101/2020.06.16.20133025v1 (2020).

20.   P. R. Randad *et al.*, COVID-19 serology at population scale: SARS-CoV-2-specific antibody responses in saliva. *medRxiv*, https://www.medrxiv.org/content/10.1101/2020.05.24.20112300v1 (2020).

21.   A. McKie, A. Vyse, C. Maple, Novel methods for the detection of microbial antibodies in oral fluid. *Lancet Infect Dis* 2, 18-24 (2002).

22.   K. T. Abe *et al.*, A simple protein-based SARS-CoV-2 surrogate neutralization assay. *BioRxiv* https://www.biorxiv.org/content/10.1101/2020.07.10.197913v1, (2020).

23.   J. L. Johnson *et al.*, The Transcription Factor T-bet Resolves Memory B Cell Subsets with Distinct Tissue Distributions and Antibody Specificities in Mice and Humans. *Immunity* 52, 842-855 e846 (2020).

24.   J. J. Ceron *et al.*, Use of Saliva for Diagnosis and Monitoring the SARS-CoV-2: A General Perspective. *J Clin Med* 9, 1491 (2020).

25.   A. Aase *et al.*, Salivary IgA from the sublingual compartment as a novel noninvasive proxy for intestinal immune induction. *Mucosal Immunol* 9, 884-893 (2016).

26.   C. Cervia *et al.*, Systemic and mucosal antibody secretion specific to SARS-CoV-2 during mild versus severe COVID-19. *BioRxiv* https://www.biorxiv.org/content/10.1101/2020.05.21.108308v1, (2020).

27.   D. Sterlin *et al.*, IgA dominates the early neutralizing antibody response to SARS-CoV-2. *MedRxiv* https://www.medrxiv.org/content/10.1101/2020.06.10.20126532v1, (2020).

28.   K. A. Callow, H. F. Parry, M. Sergeant, D. A. Tyrrell, The time course of the immune response to experimental coronavirus infection of man. *Epidemiol Infect* 105, 435-446 (1990).

29.   D. Muth *et al.*, Infectious Middle East Respiratory Syndrome Coronavirus Excretion and Serotype Variability Based on Live Virus Isolates from Patients in Saudi Arabia. *J Clin Microbiol* 53, 2951-2955 (2015).

30.   J. V. Parry, K. R. Perry, P. P. Mortimer, Sensitive assays for viral antibodies in saliva: an alternative to tests on serum. *Lancet* 2, 72-75 (1987).

31.   O. L. Rojas *et al.*, Recirculating Intestinal IgA-Producing Cells Regulate Neuroinflammation via IL-10. *Cell* 176, 610-624 e618 (2019).

32.   C. Kreer *et al.*, Longitudinal Isolation of Potent Near-Germline SARS-CoV-2-Neutralizing Antibodies from COVID-19 Patients. *Cell*, https://doi.org/10.1016/j.cell.2020.06.044 (2020).

33.   E. Seydoux *et al.*, Analysis of a SARS-CoV-2-Infected Individual Reveals Development of Potent Neutralizing Antibodies with Limited Somatic Mutation. *Immunity* 53, 98-105 e105 (2020).

34.   S. E. Blutt, M. E. Conner, The gastrointestinal frontier: IgA and viruses. *Front Immunol* 4, 402 (2013).

35.   H. Toro, I. Fernandez, Avian infectious bronchitis: specific lachrymal IgA level and resistance against challenge. *Zentralbl Veterinarmed B* 41, 467-472 (1994).

36.   O. J. Landsverk *et al.*, Antibody-secreting plasma cells persist for decades in human intestine. *J Exp Med* 214, 309-317 (2017).

37.   N. L. Bernasconi, E. Traggiai, A. Lanzavecchia, Maintenance of serological memory by polyclonal activation of human memory B cells. *Science* 298, 2199-2202 (2002).

38.   M. E. Darnell, D. R. Taylor, Evaluation of inactivation methods for severe acute respiratory syndrome coronavirus in noncellular blood products. *Transfusion* 46, 1770-1777 (2006).

39.   D. Wrapp *et al.*, Cryo-EM structure of the 2019-nCoV spike in the prefusion conformation. *Science* 367, 1260-1263 (2020).

40.   A. Poulain, A. Mullick, B. Massie, Y. Durocher, Reducing recombinant protein expression during CHO pool selection enhances frequency of high-producing cells. *J Biotechnol* 296, 32-41 (2019).

41.   A. Poulain *et al.*, Rapid protein production from stable CHO cell pools using plasmid vector and the cumate gene-switch. *J Biotechnol* 255, 16-27 (2017).

42.   D.-Y. Kim *et al.*, A Flexible Genome-Scale Resource of SARS-CoV-2 Coding Sequence Clones. *Preprints.org*, https://www.preprints.org/manuscript/202004.0009/v5 (2020).

43.   S. Miersch *et al.*, Synthetic antibodies neutralize SARS-CoV-2 infection of mammalian cells. *BioRxiv* https://www.biorxiv.org/content/10.1101/2020.06.05.137349v2.full.pdf, (2020).

44.   A. Banerjee *et al.*, Isolation, Sequence, Infectivity, and Replication Kinetics of Severe Acute Respiratory Syndrome Coronavirus 2. *Emerg Infect Dis* 26, doi: 10.3201/eid2609.201495 (2020).

45.   J. Pallesen *et al.*, Immunogenicity and structures of a rationally designed prefusion MERS-CoV spike antigen. *Proc Natl Acad Sci U S A* 114, E7348-E7357 (2017).

46.   Z. Li, I. P. Michael, D. Zhou, A. Nagy, J. M. Rini, Simple piggyBac transposon-based mammalian cell expression system for inducible protein production. *Proc Natl Acad Sci U S A* 110, 5004-5009 (2013).

Acknowledgments: We thank Janet McManus at Canadian Blood Services for her technical and logistical expertise, Dzana Dervovic, Cassandra Wong, Mark Jen and Elizabeth Rubie for help with serum-based ELISAs, Joan Wither for the lupus patient samples, Frederick P. Roth for the *SARS-CoV-2* Nucleocapsid entry vector and Jim Hartley for bacteria expression vectors. We thank research associates of the Toronto Invasive Bacterial Diseases Network and Antonio Estacio, Megan Buchholz, Patti Cheatley, Rebecca Zavalunov and Katerina Pavenski for specimen collection. We thank Jeff Browning (Boston University) for critical input into the saliva experiments. **Funding**: This study was supported by an *Ontario Together* grant and funding from the Canadian Institutes of Health Research (CIHR; #VR1-172711 and VR4–172732). Funding for

the development of the assays in the Gingras lab was provided through generous donations from the Royal Bank of Canada (RBC), Questcap and the Krembil Foundation to the Sinai Health System Foundation. The robotics equipment used is housed in the Network Biology Collaborative Centre at the Lunenfeld-Tanenbaum Research Institute, a facility supported by Canada Foundation for Innovation funding, by the Ontarian Government and by Genome Canada and Ontario Genomics (OGI-139). Indirect support for SARS-CoV-2 work in the Toronto Combined Containment Level 3 laboratory was provided by strategic research funds from the University of Toronto and the Temerity Foundation. MO is funded by: OHTN (Ontario HIV Treatment Network), CIHR and the Juan and Stefania Speck Fund. JG is a Canada Research Chair, Tier 1, in Tissue Specific Immunity and is supported by CIHR FDN15992. ACG is the Canada Research Chair, Tier 1, in Functional Proteomics and is supported by CIHR FDN143301. AJJ is supported by a Vanier Canada graduate studentship. **Author Contributions:** KTA and BR performed serum-based ELISA; BI, MZ and AP performed saliva ELISAs; KTA and LC analyzed serum ELISA data; BI, OR, GC and YB analyzed saliva ELISA data; NCH, FG, PD and SDGO designed and performed the virus neutralization assays; ZL, CG, DC, PST, FS, JMR and YD generated antigens; BI, GC, AL, AP, YFY, LGM, SJD, AM, WLS, MO and AJM provided serum and saliva samples; AJ, AM and SM collected COVID19 patient data; MBR, JHW, ACG and KC designed ELISA assays; AJJ and AJM performed statistical analysis; BI and KA generated figures; JLG and ACG supervised the project and wrote the manuscript. **Conflict of interest statement:** Steven J Drews has acted as a content expert for respiratory viruses for Johnson & Johnson (Janssen). Work in the Gingras lab was partially funded by a contribution from QuestCap through the Sinai Health Foundation. Work in the Gommerman lab unrelated to this project has been funded by EMD-Serono, Roche and Novartis.

List of Supplementary Material

Materials and Methods
Fig S1–S10
Table S1–S3

Main text figures and tables

Figure 1. A cross-sectional analysis of IgG and IgA responses to the spike and RBD antigens of SARS-CoV2 in serum. (**A-D**) Indicated immunoglobulins to spike and RBD were profiled by ELISA in cohorts of pre-COVID samples (n=300), patients with acute COVID infection (n=139) and convalescent patients (n=357). All data, expressed as ratios to a pool of convalescent samples, were plotted using bean plots. Solid bars denote the median and dotted line represents the median across all samples used in the plot. (**E-F**) levels of IgG (E) and IgA (F) to the RBD (y-axis) and spike (x-axis) antigens for the indicated patient groups. Spearman correlation coefficient is indicated. Mann-Whitney U test for significance was performed. *= p ≤ 0.05, **** = p < 0.0001.

Figure 2. Persistence of antibodies in the serum of affected individuals. (A-D) Binned relative ratios (to a pool of positive controls) of spike (A-B) and RBD (C-D) to the indicated antibodies, displayed as bean plots. Days PSO are binned in 15-day increments and are compared to pre-COVID samples (neg). Solid bars denote the median and dotted line represents the median across all samples used in the plot.

Figure 3. A longitudinal analysis of IgG and IgA responses to the spike and RBD antigens of SARS-CoV2 in serum. (A-D) Analysis of the changes in the indicated Ig-antigen levels in patients profiled twice, in comparisons to the relative levels in pre-COVID negative controls (left). Dots represent individual serum samples collected at the indicated times, and the samples from the same patients are connected by the lines. A non-parametric loess function is shown as the blue line, with the grey shade representing the 95% confidence interval.

Figure 4. A cross-sectional analysis of IgG and IgA responses to the spike and RBD antigens of SARS-CoV-2 in saliva. A second cohort of COVID-19 patients was tested for the presence of IgG and IgA antibodies to SARS-CoV-2 spike and RBD antigens in the saliva, comparing with a mixture of unexposed negative controls collected locally and pre-COVID era controls. Unlike cohort 1, for cohort 2 samples we had sufficient material to perform several dilutions and to generate an AUC value for each subject. This value was expressed as a percentage of the AUC derived from a control sample consisting of pooled saliva from COVID-19 patients. Because the saliva was not diluted during collection, we were able to derive the concentration of IgG and IgA in both negative controls and COVID-19 patients. (**A-B**) Total IgG and IgA levels in the saliva. (**C-F**) Saliva data for negative controls versus COVID-19 patients. Solid bars denote the median and dotted line represents the median across all samples used in the plot. Mann-Whitney U test for significance was performed. **** = p < 0.0001, n.s. = not significant.

Figure 5. A cross-sectional analysis of IgG and IgA responses to the spike and RBD antigens of SARS-CoV-2 in saliva correlated with time PSO. A second cohort of COVID-19 patients (n=90) was tested for the presence of IgG and IgA antibodies to SARS-CoV-2 spike and RBD antigens in the saliva, comparing with a mixture of unexposed negative controls collected locally and pre-COVID era negative controls. (**A-D**) Saliva data for all 4 antigen-specific ELISA readouts plotted as time PSO. Spearman correlation coefficients (ρ) and p-value are indicated.

Figure 6. Correlating IgG and IgA responses to the spike and RBD antigens with age and disease severity in serum. ELISA results from the automated platform probing for both IgG and IgA on spike and RBD were plotted against patient age and grouped by symptom severity. Spearman correlation coefficients (ρ) were determined for each group of patients with differing levels of severity.

Figure 7. Correlation of IgG and IgA responses to the spike and RBD antigens in serum and saliva. (A-D) A subset of blood and saliva sample pairs (n=72) collected from the same patient within 4 days were analyzed for correlations in levels of anti-spike and anti-RBD IgG and IgA antibodies. Spearman correlation coefficient (ρ) and p-value are indicated.

Table 1. Cohorts of patients and negative controls.

		SALIVA			Sex			BLOOD			Sex	
		No. patient	No. samples	Median Age	No. M	No. F		No. patients	No. samples	Median Age	No. M	No. F
All samples		247	263	-	141	106		777	835			
Patients with COVID-19	Cohort 1	47	54	61	28	19	Patients with COVID-19	438	496	58	262	234
	Cohort 2	81	90	58	48	33	Pre-COVID Negative Controls	300	300	54.5	150	150
Pre-COVID Negative Controls	Cohort 1	-	-	-	-	-						
	Cohort 2	27	27	43	12	15						
Unexposed Negative Controls Collected in 2020	Cohort 1	42	42	60	24	18						
	Cohort 2	50	50	58	29	21						
Matched saliva-serum samples		72	72	58	33	39						

Table 2. Summary of bivariate analyses of age, disease severity, and sex in relation to antibody levels. Spearman's correlation coefficients (denoted as 'ρ' and if applicable) and P values are indicated for each predictor and antibody.

	PREDICTORS Saliva (N=90)			PREDICTORS Serum (N=496)		
	Disease Severity	Age	Sex	Disease Severity	Age	Sex
	P value[a]	Spearman Correlation Coefficient (ρ) and P Value[b]	P value	P value[a]	Spearman Correlation Coefficient (ρ) and P Value[b]	P value[e]
IgA RBD	0.01	ρ=0.2, p=0.09	0.5	<0.0001	ρ=0.2, p<0.0001	<0.0001
IgA spike	0.2	ρ=0.3, p=0.1	1.0	<0.0001	ρ=0.3, p=0.002	0.0004
IgG RBD	0.1	ρ=0.1, p=0.02	0.1	<0.0001	ρ=0.1, p=0.001	0.2
IgG spike	0.1	ρ=0.2, p=0.04	0.3	0.003	ρ=0.02, p=0.9	0.5

[a]See bean plots of antibody levels versus disease severity for serum (Fig S8A-D) and saliva (Fig S10).
[b]See scatter plots of antibody levels versus age in serum (Fig S7) and saliva (Fig S9).
[e]See bean plots of antibody levels versus sex for serum samples in Fig S8E-H.

Supplementary Materials:

Materials and Methods

Recruitment and participants – COVID19 patients

Acute and convalescent serum and saliva samples were obtained from patients identified by surveillance of COVID-19 (confirmed by PCR; in- and out-patients) by the Toronto Invasive Bacterial Diseases Network in metropolitan Toronto and the regional municipality of Peel in south-central Ontario, Canada (REB studies #20-044 Unity Health Network, #02-0118-U/05-0016-C, Mount Sinai Hospital). Consecutive consenting patients admitted to four TIBDN hospitals were enrolled: these patients had serum and saliva collected at hospital admission, and survivors were asked to submit repeat samples at 4-12 weeks after onset of symptoms. Consecutive out-patients diagnosed at the same 4 hospitals prior to March 15[th] and on a convenience sample of later days were approached for consent to collect serum and saliva at 4-12 weeks post onset of symptoms. Patients were interviewed and patient charts reviewed to determine age, sex, symptom onset date, and disease severity (mild, moderate, and severe). For this study, disease was considered mild if it did not require hospitalization, moderate if it required hospitalization but not intensive care unit (ICU) admission, and severe if it required ICU care. Specimens were considered convalescent if they were collected less than 21 days after onset of symptoms, and convalescent if they were collected 21 or more days after onset of symptoms. From March 10-April 14, patients were asked to provide a 5 ml sample of saliva in a sterile specimen container, and 2.5mls of phosphate buffered saline was added to reduce viscosity for PCR testing. From April 16[th] on, saliva specimens were collected in Salivette® tubes (Sarstedt, Numbrecht, Germany). All specimens were aliquotted and stored frozen at -80ºC prior to analysis.

Additional positive samples for test development were obtained through the Canadian Blood Services. Specimen-only serum donations were collected from individuals with a self-declared SARS-CoV-2-positive nucleic acid test. Collections occurred two weeks or more after cessation of clinical symptoms.

Recruitment and participants – control saliva and serum

Control saliva samples were collected from unexposed, asymptomatic individuals residing in an area of very low COVID19 case numbers (Grey County, Ontario) and throughout the Greater Toronto Area (GTA) (REB study# 23901 University of Toronto).

Control serum samples were from patients enrolled in cancer or birth cohort studies prior to COVID-19 (prior to November 2019; REB studies #01-0138-U and #01-0347-U, Mount Sinai Hospital) and archived frozen in the LTRI Biobank, or from previous studies of the immune system or systemic lupus acquired prior to November 2019 (REB studies #31593 University of Toronto, #05-0869, University Health Network).

Sample collection, handling and viral inactivation – serum

Serum (and in some cases plasma) was collected using standard procedures at the collection sites and transferred to the testing lab on dry ice. Inactivation of potential infectious

viruses in plasma or serum was performed by incubation with Triton X-100 to a final concentration of 1% for 1 h prior to use (*38*).

Antigen production – serum assays

Spike trimer was expressed as follows: the SARS-CoV-2 spike sequence (aa 1-1208 from Genebank accession number MN908947 with the S1/S2 furin site (residues 682–685) mutated [RRAR->GGAS] and K986P / V987P stabilizing mutations was codon-optimized (*Cricetulus griseus* codon bias) and synthesized by Genscript. To stabilize the spike protein in a trimeric form, the cDNA was cloned in-frame with the human resistin cDNA (aa 23-108) containing a C-terminal FLAG-(His)$_6$ tag (*Cricetulus griseus* codon bias, GenScript) into a modified cumate-inducible pTT241 expression plasmid and transfected in CHO[2353] cells (Stuible *et al.*, manuscript in preparation) followed by methionine sulfoximine selection for 14 days to generate a stable CHO pool. This CHO pool allows for cumate-inducible trimeric spike expression from the CR5 promoter as described in Poulain *et al.* (*39-41*). Cell culture was harvested 8-10 days post-cumate induction and secreted spike trimer present in the clarified medium purified by immobilized metal-affinity chromatography (Ni-Excel resin). Purified trimeric spike was buffer exchanged in PBS and store as aliquots at -80°C. The purified spike protein integrity and purity was analyzed by SDS-PAGE and analytical size-exclusion ultra-high performance liquid chromatography (SEC-UPLC). The SEC was run in PBS + 0.02% Tween-20 on an 4.6 x 300 mm Acquity BEH450 column (2.5 μm beads size; Waters Limited, Mississauga, ON) coupled to a MALS detector and the spike trimer eluted as a major (>95% integrated area) symmetrical peak of 490 kDa with less than 3% aggregates (not shown). RBD was expressed as for the saliva assay, but left non-biotinylated, as in (*22*).

Nucleocapsid[1-419] from the pEntry-N (closed) Open Reading Frame (a kind gift from Dr. Frederick P. Roth (*42*)) was cloned into pDEST585 gift of Jim Hartley, internal ID V2097) as a HIS-GST-TEV fusion using LR-clonase. The resulting expression vector was confirmed by restriction digest, expressed in *E. coli* BL21(DE3) Codon+ cells (Agilent Technologies) and induced with 0.25 mM isopropyl 1-thio-β-D-galactopyranoside (IPTG) for 16 hours at 18°C. Harvested cells were resuspended in 20 mM HEPES pH 7.5, 400 mM NaCl, 5 mM imidazole and lysed by passage through a cell homogenizer (Avestin Inc.). Following centrifugation at 30,000g, supernatant was passed through a 0.45 μM PVDF filter and applied to a HiTrap nickel chelating HP column (GE Healthcare). Protein eluted with buffer containing 300 mM imidazole was incubated overnight with Tobacco Etch Virus (TEV) protease. Following cleavage of the His-Tag, protein was dialyzed in 20 mM HEPES pH 7.5, 50 mM NaCl and flowed over a 5 ml HiTrap nickel chelating column to remove His-GST. Nucleocapsid protein was further purified by ion exchange on a mono-S column (GE Healthcare) equilibrated in 20 mM HEPES pH 7.5, 50 mM NaCl, 1 mM DTT and eluted with a gradient to 500 mM NaCl. Purified Nucleocapsid protein was concentrated to 6 mg/mL and stored at -80°C.

Enzyme-linked immunosorbent assays for detecting antigen-specific IgG and IgA in serum or plasma

A manual colorimetric ELISA assay (similar to (*3*)) was first implemented in 96-well plates using the RBD and spike non-biotinylated antigens described here for the detection of IgG

(also see (*22*)). Briefly, concentrations and incubation times were optimized to maximize the separation between anti-RBD or anti-spike trimer levels in convalescent plasma or serum from that of pre-COVID era banked serum while maintaining the required levels of antigens as low as possible. 75 ng and 200 ng of RBD and spike, respectively, were first adsorbed onto 96-well clear Immulon 4 HBX (Thermo Scientific, #3855) plates in PBS overnight at 4 °C, then washed three times with 200 µl PBS+ 0.1% Tween-20 (PBS-T; Sigma). Plates were blocked with 3% w/v milk powder (BioShop Canada Inc., #ALB005.250, lot #9H61718) in PBS for 1–2 hr and washed three times with 200 µl PBS-T. Patient samples (pre-treated with 1% final Triton X-100 for viral inactivation) diluted 1:50 in PBS-T containing 1% w/v milk powder were then added to the plates and incubated for 2 hr at room temperature (50 µl total volume): technical duplicates were performed unless otherwise indicated. Positive and negative control recombinant antibodies and serum samples were added to each plate to enable cross-plate comparisons. Wells were washed three times with 200 µl PBS-T. Goat anti-human anti-IgG (Goat anti-human IgG Fcγ - HRP, Jackson ImmunoResearch, #109-035-098) at a 1:60,000 dilution (0.67 ng/well) in 1% w/v milk powder in PBS-T was added and incubated for 1 hr. Wells were washed three times with 200 µl PBS-T, and 50 µl of 1-Step™Ultra TMB-ELISA Substrate Solution (ThermoFisher, #34029) was added for 15 min at room temperature and the reaction was quenched with 50 µL stop solution containing 0.16N sulfuric acid (ThermoFisher, #N600). The plates were read in a spectrophotometer (BioTek Instruments Inc., Cytation 3) at 450 nm. For all ELISA-based assays, raw OD values had blank values subtracted prior to analysis. All data were normalized to the positive serum control (single point) on each plate and expressed as a ratio to this control. The assay performance was assessed by precision-recall analysis of ratio-expressed values (Fig S1).

The assay was then re-designed to be conducted in a customized robotic platform using a 384-well plate format, first by simply scaling down the volume/amounts used, and then switching to a chemiluminescent substrate for detection, and re-optimizing the amounts per well of antigens and secondary antibodies' dilutions to use. A chemiluminescent substrate is ideally-suited for automated ELISAs, because it offers a higher sensitivity and a better dynamic range than standard colorimetric assays. Furthermore, the reaction does not need to be stopped (e.g. with robotics-incompatible acids) and the luminescence signal is stable for at least 60 min. For all steps, liquid dispensers (Beckman Biomek NXp or ThermoFisher Multidrop Combi) and washer (Biotek 405 TS/LS LHC2) were used on a F7 robotic platform available at the Network Biology Collaborative Centre (nbcc.lunenfeld.ca). Each step of the methods to evaluate the different antigen and antibody class combinations were optimized and routine quality control tests were performed for all dispensing steps.

For automated ELISAs, LUMITRAC 600 high-binding white polystyrene 384-well microplates (Greiner Bio-One, through VWR #82051-268) were pre-coated overnight with 10 µl/well of RBD (25 ng) or spike (50 ng). Next day, the wells were washed 4 times (a BioTek washer is used for all washing steps, and all washes are performed with 100 µl PBST). Wells were blocked for 1 hr at room temperature in 80 µl 5% Blocker™ BLOTTO (Thermo Scientific, #37530), then washed 4 times. 10 µl Triton X-100 inactivated serum (or plasma) samples diluted 1:40 in 1% BLOTTO in PBS-T were added to each well from 96-well sample source plates and incubated for 2 hr at room temperature. Positive and negative controls used on each plate are described below. After washing 4 times, 10 µl of one of the following secondary antibodies (all from Jackson ImmunoResearch) diluted in 1% BLOTTO in PBS-T were added at the indicated concentrations followed by incubation for 2 hr at room temperature: Goat anti-human IgG Fcγ – HRP (#109-035-098; 1:40,000 or 0.2 ng per well), Goat anti-human IgM Fc5u – HRP (#109-

035-129; 1:12,000 or 0.66 ng per well) or Goat anti-human IgA a chain - HRP (#109-035-127; 1:10,000 or 0.8 ng per well). After 4 washes, 10 µl of SuperSignal ELISA pico Chemiluminescent substrate (diluted 1:4 in water) was added, followed by a short mix for 10s at 900 rpm, and incubation at room temperature for 5 min. Luminescence was read on an EnVision (Perkin Elmer) plate reader at 100 ms/well using an ultra-sensitive luminescence detector. All automated assays were performed in biological duplicates, processed on different days. Blank values were subtracted for all raw reads prior to data analysis, and the values were expressed as a ratio of the positive reference serum pool on the same plate (see below).

Quality controls and normalization of the samples in the automated assays were as follows: a standard curve with recombinant antibodies reacting to spike RBD or spike S1 was included on each plate. Antibodies used for the standard curves were: Human anti-spike S1 IgG (A02038, GenScript), anti-spike S1 IgM (A02046, GenScript) and Ab01680 anti-spike IgA (Ab01680-16, Absolute Antibody), all used at 0.5 to 10ng per well. Negative antibody controls were immunoglobulins from human serum (I4506 human IgG, I8260 human IgM, and I4036 human IgA, from Millipore-Sigma). A positive and negative control pool of 4 patient samples each was created and added in each plate at a single point concentration for normalization. For all assays, a standard curve is generated by first plotting the mean of the blank-subtracted recombinant antibodies, plotted against antibody amounts (in ng), and the linearity of the curve and comparison to previous runs is assessed, alongside the confirmation that the positive and negative pool sample fall within the expected range of the standard curve [%CV should be 10-15% or less].

Sample collection and handling - saliva study

With the exception of some samples that were acquired early on in the pandemic (cohort 1), Salivette® tubes were used to collect samples according to manufacturer instructions (Sarstedt, Montreal, Quebec). These tubes include a cotton swab that participants are instructed to chew for set amount of time. The swab is then transferred into an inner tube which is then inserted into an outer tube that catches liquid saliva upon centrifugation at 1000 x g for 3 minutes. Salivary flow was controlled by establishing a fixed amount of collection time (2 minutes) for each subject as previously recommended (*22, 43*). For the early pandemic subjects that were not given salivettes and used in our pilot study (cohort 1), these subjects expectorated directly into a 15 mL conical tube containing 2.5 mL of phosphate-buffered saline (PBS). Prior to saliva collection, healthy subjects confirmed they had fasted, refrained from taking oral medication, and had not brushed their teeth for a minimum of 30 minutes.

Viral inactivation in saliva samples

Following centrifugation, all saliva samples, regardless of their SARS-CoV-2 PCR status, underwent viral inactivation by treating with Triton® X-100 (BioShop, CAT# TRX506.100). 10% Triton X-100 (diluted 1:10 from stock) was added to all samples to a final dilution of 1% Triton X-100 and incubated for 1 hour at room temperature. Inactivated samples were immediately frozen and stored at -80°C. Heat inactivation for 30 minutes at 65°C was found to destroy the IgG and IgA signal against RBD and was therefore not used (Fig S4). The efficiency of virus inactivation in a saliva medium is shown in Table S3. Specifically, we assessed the

treatment of saliva collected from healthy individuals using two different methods (Salivette vs. direct saliva collection into a tube). These samples were spiked with known amounts of SARS-CoV-2 viral stock and then treated with 1% Triton X-100 for 30 minutes, 1 hour or 2 hours. Vero-E6 cells (ATCC$_®$ CRL-1586^{TM}) were used to determine outgrowth of virus. Cells were maintained in Dulbecco's Modified Eagle's Media (DMEM) supplemented with L-glutamine, penicillin/streptomycin and 10% fetal bovine serum (FBS). SARS-CoV-2 virus (isolate SB3) was isolated in-house (44). Briefly, viral stocks were created after isolation of virus from a clinical sample in Toronto, Ontario, Canada. Viral stock was expanded using Vero E6 as previously described such that stored aliquots of stock contain 2% FBS. Initial experiments were done with Triton X-100 (Sigma-Aldrich) serially diluted and applied to Vero-E6 cells in 96-well flat bottom plates to determine the minimum concentration required to prevent toxicity to cells. Furthermore, we have also determined if neat saliva itself could be cytotoxic to Vero-E6 cells by providing healthy donor saliva alone or treated with Triton X-100 ranging from final Triton-X100 concentration of 0.03%, 0.01%, 0.001% and 0.0001% (v/v). Since initial Triton X-100 experiments showed that toxicity is averted at 0.03% (v/v), we proceeded to use this concentration as the point of dilution to prevent any Triton X-100 mediated toxicity.

Antigen production – saliva assay

The expression, purification and biotinylation of the SARS-CoV-2 RBD and spike ectodomain were performed as recently described (22, 43). The human codon optimized cDNA of the SARS-CoV-2 spike protein was purchased from Genscript (MC_0101081). The soluble RBD (residues 328-528, RFPN...CGPK) was expressed as a fusion protein containing a C-terminal 6xHis tag followed by an AviTag. The soluble trimeric spike protein ectodomain (residues 1-1211, MFVF...QYIK) was expressed with a C-terminal phage foldon trimerization motif followed by a 6xHis tag and an AviTag. To help stabilize the spike trimer in its prefusion conformation, residues 682–685 (RRAR) were mutated to SSAS to remove the furin cleavage site and residues 986 and 987 (KV) were each mutated to a proline residue (45). Stably transfected FreeStyle 293-F cells secreting the RBD and soluble spike trimer were generated using a previously reported piggyBac transposon-based mammalian cell expression system (46). Protein production was scaled up in 1L shake flasks containing 300 mL FreeStyle 293 medium. At a cell density of 10^6 cells/mL, 1 µg/mL doxycyline and 1 µg/mL Aprotinin were added. Every other day 150 mL of medium was removed and replaced by fresh medium. The collected medium was centrifuged at 10000 x g to remove the cells and debris and the His-tagged proteins were purified by Ni-NTA chromatography. The eluted protein was stored in PBS containing 300 mM imidazole, 0.1% (v/v) protease inhibitor cocktail (Sigma, P-8849) and 40% glycerol at -12 °C. Shortly before use, the RBD and spike proteins were further purified by size-exclusion chromatography on a Superdex 200 Increase (GE healthcare) or Superose 6 Increase (GE healthcare) column, respectively. Purity was confirmed by SDS-PAGE (not shown). For the spike protein, negative stain electron microscopy was used show evidence of high-quality trimers (not shown). The Avi-tagged proteins, at a concentration of 100 µM or less, were biotinylated in reaction mixtures containing 200 µM biotin, 500 µM ATP, 500 µM $MgCl_2$, 30 µg/mL BirA, 0.1% (v/v) protease inhibitor cocktail. The mixture was incubated at 30 °C for 2 hours followed by size-exclusion chromatography to remove unreacted biotin.

Enzyme-linked immunosorbent assays for detecting total IgA and IgG in saliva

Quantitative total IgA and IgG analyses were performed on the same samples used for detection of anti-RBD and anti-spike Ig described above. Anti-human Ig antibody (Southern Biotech, 2010-01) diluted 1:1000 in PBS was added to 96-well Nunc MaxiSorp™ plates (ThermoFisher, 44-2404-21). PBS alone was added to control wells. Plates were allowed to coat overnight at 4^{o}C. Following coating, plates were blocked using 5% BLOTTO for 2 hours at 37^{o}C. Samples were diluted in 2.5% BLOTTO. Standards (purified IgA and IgG purchased from Sigma-Millipore (IgA, I4036 and IgG, I2511) were prepared in 2.5% BLOTTO ranging from 100 ng/mL 3.125 ng/mL. Upon discarding the blocking solution from the plate, samples and standards were immediately transferred to wells and incubated for 2 hours at 37^{o}C. Following incubation, wells were washed with 200 µL of PBS-T. HRP-conjugated secondary antibodies against IgA and IgG (goat anti-human IgA- and IgG-HRP, Southern Biotech, IgA: 2053-05, IgG: 2044-05) were added to the appropriate wells at 1:1000 in 2.5% BLOTTO and incubated for 1 hour at 37^{o}C. Development of the plates was performed as described above. A four-parameter logistic curve was used to determine the line of best fit for the standard curve, and sample Ig quantities were interpolated accordingly to determine final concentrations in µg/ml. The few samples from patient or control groups that exhibited quality control issues (extremely low to negative IgA levels) were excluded from further analysis.

Enzyme-linked immunosorbent assay for detecting albumin in saliva

Salivary albumin was measured for Cohort 1 using a purchased Human Albumin ELISA Kit (Abcam, ab108788). Assay was performed according to manufacturer instructions included with the kit.

Enzyme-linked immunosorbent assays for detecting antigen-specific IgA and IgG in saliva

96-well plates pre-coated with streptavidin (ThermoFisher, 436014) were used for all assays. Without the biotin-streptavidin system, the anti-S/RBD IgG and IgA signals obtained from COVID-19 patient saliva were undetectable. Based on titrations of antigens using saliva from convalescent COVID-19 patients, 100ng of biotinylated-RBD and 1µg of biotinylated S proteins were applied to the appropriate wells one day prior to starting the assay (see Fig S4 for RBD titration, spike titration not shown). Control wells of sterile PBS rather than biotinylated antigen were reserved for each patient and control sample. A few wells with the biotinylated antigen but with no sample added were reserved as negative internal controls for the reagents on the assay. Plates were incubated overnight at 4^{o}C to allow sufficient coating of the antigen. 200 µL of 5% BLOTTO (5% w/vol skim milk powder (BioShop, CAT# SKI400.500) in sterile PBS) was subsequently added to each well to prevent non-specific interactions, followed by a 2-hour incubation at 37^{o}C. Blocking solution was discarded immediately from plates prior to addition of samples to wells. Newly thawed saliva samples were centrifuged at 8000 rpm for 4 minutes, and appropriately diluted using 2.5% BLOTTO. To reduce anti-streptavidin reactivity in the saliva, diluted samples were applied to streptavidin-coated plates with no antigen and allowed to incubate for 30 minutes at 37^{o}C. Subsequently 50 µL of samples were transferred from the pre-adsorption plate into antigen-coated plates and incubated for 2 hours at 37^{o}C. PBS+0.05% Tween 20 (BioShop, CAT# TWN510) (PBS-T) was used for washing plates between steps.

Horseradish peroxidase (HRP)-conjugated Goat anti human-IgA and anti-IgG secondary antibodies (Southern Biotech, 2053-05 and 2044-05) were added to wells at dilutions of 1:2000 and 1:1000 in 2.5% BLOTTO, respectively, and incubated for 1 hour at 37^{o}C. Development of the plate was done by adding 50 μL of 3,3',5,5'-tetramethylbenzidine (TMB) Substrate Solution (ThermoFisher, 00-4021-56) onto plates. Reaction vas then stopped by adding 50ul/well of 1N H_2SO_4 Optical density (OD) was read at a wavelength of 450 nm on a spectrophotometer (OD_{450}). For cohort 1, because some samples had been collected in cups and were therefore diluted, normalization to a separate variable was performed. The resulting OD from antigen-specific IgA and IgG was subtracted from the OD for the PBS control wells for each sample and subsequently normalized to albumin levels or total IgA and IgG levels, respectively (see below). For cohort 1, raw OD_{450} measurements obtained from PBS-coated wells corresponding to each sample diluted at 1/5 ("background signal") was subtracted from readings obtained from antigen-coated wells at each of three dilutions (1/5, 1/10). Data was normalized to the total IgG, total IgA, or albumin content in each saliva sample. A small number of samples (n=9 from negative controls and n=4 from patients) exhibited high OD values that did not titrate and coincided with high OD levels when plated without antigen (PBS control). These were excluded from the analysis. For cohort 2, raw OD_{450} measurements obtained from PBS-coated wells corresponding to each sample diluted at 1/5 ("background signal") was subtracted from readings obtained from antigen-coated wells at each of three dilutions (1/5, 1/10, 1/20). For each plate, a sample of pooled saliva from COVID-19 acute and convalescent patients was likewise plated at 1/5 with no antigen (PBS control), as well as with antigens at 1/5, 1/10 and 1/20. Background subtracted values from all three dilutions were added for each sample (AUC) and normalized to the AUC for the pooled positive control saliva and expressed as a percentage. By using the same positive control that we ran in every single plate, we determined that intra-assay precision was always greater than 90% between wells. Reproducibility between plates was determined by a coefficient of variation of less than 10% through all the plates. A small number of samples (n=6 from negative controls and n=2 from patients) exhibited high OD values that did not titrate and coincided with high OD levels when plated without antigen (PBS control). These were excluded from the final analysis.

Receiver-Operating Characteristic (ROC) curves

For serum and plasma sample analysis, samples acquired prior to November 2019 (pre-COVID) were labeled true negatives while convalescent samples from patients with PCR-confirmed COVID19 were labelled true positives. For saliva samples, all samples from patients with PCR-confirmed COVID19 collected more than 10 days PSO were considered true positives, and saliva collected before 2020 and from unexposed, asymptomatic individuals in March of 2020 were labeled true negative for ROC analysis. Ratio-converted ELISA reads (colorimetric or chemiluminescent) were used for ROC analysis in the easyROC webtool (v 1.3.1) with default parameters (https://journal.r-project.org/archive/2016/RJ-2016-042/index.html). Non-parametric curve fitting was applied alongside DeLong's method for standard error estimation and confidence interval generation.

Data Analysis, Statistical Analysis and Data Visualization

Saliva: For both antigen-specific and total IgA and IgG readouts, raw OD_{450} measurements obtained from PBS-coated wells corresponding to each sample ("background signal") was subtracted from readings obtained from antigen-coated or anti-human Ig-coated wells. Total IgA and IgG quantification were determined relative to standard wells present on each plate. A four-parameter logistic curve was used to determine the line of best fit for the standard curve, and sample Ig quantities were interpolated accordingly, using Prism Graphpad, Version 8.3. The resulting OD from antigen-specific IgA and IgG were subsequently normalized to total IgA and IgG, respectively.

Serum: Raw OD_{450} measurements for IgG, IgA and IgM on spike, RBD and NP from either the manual or automated platforms were subtracted from wells coated with PBS. A pool of serum samples that previously exhibited high levels of IgGs to all antigens was used as an internal standard across all plates, and a relative ratio between blank-adjusted OD_{450} measurements of patient samples to the OD_{450} measurements of this positive pooled standard are reported. Serum data were analyzed in R using version 4.0.1.

The relationships between saliva and serum antibody levels and metadata such as age, sex, time PSO, and clinical severity were also examined. Spearman's correlation coefficients (ρ) and p values were calculated to assess the relationship between age and antibody levels and between time PSO and antibody levels. Median antibody levels were compared in males vs. females using Mann Whitney U tests and in mild vs. moderate vs. severe illness using Kruskal-Wallis H tests. These analyses were performed in SAS 9.4M6.

The relationship between time PSO and antibody levels in the convalescent period was examined in multivariable linear regression models that adjusted for age, sex, and disease severity. For serum samples, four multivariable linear regression models were constructed (one for each of anti-RBD IgA, anti-S IgA, anti-RBD IgG, anti-S IgG). Generalized estimating equations were used (proc genmod in SAS with exchangeable correlation matrices) to account for patient-level clustering. Antibody levels were transformed as appropriate to achieve heteroscedasticity, and the variance inflation factors for all covariates confirmed to be <5 to verify absence of multicollinearity. For saliva samples, four multivariable linear regression models were similarly constructed; however, only the first convalescent sample for each patient was included in the analysis (proc glm in SAS).

The relationships between clinical predictors (age, sex, disease severity) and serum and saliva antibody levels across the acute and convalescent periods were examined in bivariable and multivariable linear regression models. For serum samples, four multivariable linear regression models were constructed to examine potential independent associations between the three clinical predictors and anti-RBD IgA, anti-S IgA, anti-RBD IgG, and anti-S IgG. These models used generalized estimating equations (proc genmod in SAS with exchangeable correlation matrices) to account for patient-level clustering. In all models, antibody levels were appropriately transformed to ensure homoscedasticity, and the absence of multicollinearity was confirmed with variance inflation factors being <5 for all predictors. Four multivariable linear regression models were similarly constructed for saliva samples. All analysis was performed in SAS 9.4M6 (SAS Institute, Cary, NC). P values of <0.01 were considered statistically significant.

Supplementary figures and tables

Fig S1. Development and validation of manual colorimetric and automated chemiluminescent assays for monitoring RBD and spike trimers antibodies in serum or plasma. (A-D) Receiver operating characteristic (ROC) analysis of serum/plasma assays. ROC curves generated for ELISAs conducted on serum (manual IgG and automated IgG/IgA/IgM platforms) on spike and RBD (note that the IgM assays are not used to profile samples in this study but included here for assay completeness). Samples used for the profiling are listed in Table S1. **(E-F)** Spearman correlation between IgG readouts of manual (colorimetric, 96 wells) and automated (chemiluminescent, 384 wells) assays using spike and RBD.

Fig S2. Correlations between antibody levels and day of symptom onset to sample collection. (**A-D**) Spearman rank correlations between ELISA results from the automated chemiluminescent platform for the indicated antigens (spike trimer left column; RBD right column), and immunoglobulins (A-B; IgG; C-D; IgA). (**E-F**) Validation of the trends observed in the automated platform on the manual colorimetric platform for the IgG response to spike and RBD. The shaded area represents the 95% confidence interval.

Fig S3. IgG and IgA responses to the Nucleocapsid antigen of SARS-CoV2 in serum. (A-B) Receiver operating characteristic (ROC) analysis of serum/plasma assays for NP (as in Fig S1). **(C-D)** Binned relative ratios of IgG and IgA to NP, displayed as bean plots (see Fig. 1). Solid bars denote the median and dotted line represents the median across all samples used in the plot. (E-F) Longitudinal profiling of the antibody response to NP in patients profiled twice, with non-parametric loess function shown as the blue line, with the grey shade representing the 95% confidence interval (also see Fig. 2).

Fig S4. Effect of heat versus detergent inactivation of saliva samples on the detection of anti-RBD antibodies in a manual, colourimetric ELISA. (**A**) Anti-RBD IgG levels expressed as raw OD values in heat versus detergent inactivated samples across a titration of RBD-biotin levels. (**B**) Anti-RBD IgA levels expressed as raw OD values in heat versus detergent inactivated samples across a titration of RBD-biotin levels.

Figure 5. IgG and IgA levels against SARS-CoV-2 antigens in the saliva of cohort 1. A pilot cohort of COVID-19 patients was tested for the presence of IgG and IgA antibodies to SARS-CoV-2 spike and RBD antigens in the saliva, comparing with age- and sex-matched unexposed negative controls collected locally. **(A-D)** Antigen-specific (anti-spike, anti-RBD) IgG levels normalized by total IgG or albumin. (**E-H**) Antigen-specific (anti-spike, anti-RBD) IgA levels normalized by total IgA or albumin. Yellow bars denote saliva samples collected at an unknown dilution. Solid bars denote the median. Mann-Whitney U test for significance was performed. **** = p < 0.0001.

Fig S6. Validation of manual colourimetric assays for monitoring RBD and spike trimers antibodies in saliva. (A-B) Receiver operating characteristic **(**ROC) analysis of ELISA assays conducted on saliva (manually conducted IgG/IgA) on spike and RBD. Samples used for the profiling are listed in Table S3.

Fig S7. Correlations between antibody levels in serum and patient age. (**A-D**) Spearman rank correlations between ELISA results from the automated chemiluminescent platform for the indicated antigens (spike trimer left column; RBD right column), and immunoglobulins (A-B; IgG; C-D; IgA). See Table 2 for statistical tests.

Fig S8. Univariate analyses correlating antigen-specific responses in serum with patient metadata. (**A-D**) Correlation between symptom severity. (E-H) Correlations with sex. Solid bars denote the median and dotted line represents the median across all samples used in the plot. See Table 2 for statistical tests.

Fig S9. Correlations between antibody levels in saliva and patient age.
(A-D) Correlation between age and relative levels of antigen-specific antibody responses in the saliva of COVID-19 patients. See Table 2 for statistical tests.

Fig S10. Association between antibody levels in saliva and patient clinical severity. (A-D) Correlation between disease severity and relative levels of antigen-specific antibody responses in the saliva of COVID-19 patients. Solid bars denote the median and dotted line represents the median across all samples used in the plot. See Table 2 for statistical tests.

Table S1. ROC statistics table for ELISAs conducted on blood-derived samples*

Marker	Manual IgG			Automated IgG			Automated IgA			Automated IgM		
	Spike	RBD	NP	Spike	RBD	NP	Spike	RBD	NP	Spike	RBD	NP
AUC	0.990	0.983	0.975	0.992	0.970	0.953	0.967	0.878	0.802	0.957	0.938	0.764
SE.AUC	0.0033	0.0047	0.0056	0.0025	0.0068	0.0079	0.0068	0.0127	0.0160	0.0138	0.0164	0.0284
Lower Limit	0.984	0.974	0.964	0.987	0.957	0.937	0.954	0.853	0.771	0.929	0.906	0.709
Upper Limit	0.997	0v.993	0.986	0.997	0.984	0.968	0.981	0.903	0.834	0.984	0.970	0.820
z	148.215	102.121	85.284	199.89	69.102	57.202	69.008	29.784	18.865	33.01	26.74	9.327
p-value	0	0	0	0	0	0	0	0	0	0	0	0

ROC Coordinate points at <1% FDR

Marker	Spike	RBD	NP	Spike	RBD	NP	Spike	RBD	NP	Spike	RBD	NP
Cutpoint	0.35	0.19	0.49	0.124	0.244	0.53	0.735	0.616	0.559	0.466	0.421	2.14
FPR	0.0088	0.0088	0.0088	0.0088	0.0088	0.0088	0.0088	0.0088	0.0088	0.0088	0.0088	0.0088
TPR	0.956	0.938	0.828	0.955	0.913	0.754	0.445	0.301	0.223	0.663	0.562	0.180

Samples used to generate ROC curves

Sample Cohort	No. Samples
COVID-19 Patient Samples	353 (64)**
Pre-COVID-19 Negative Controls	339 (339)**

* Sensitivity values reported in the manuscript are based on the coordinate points at a false positive rate of <1%.

** Values in parentheses reflect the smaller sample size used to generate the ROC curve for the automated IgM ELISA assay.

Table S2. Testing effect of Triton-X treatment on saliva samples*

Sample Collection Method	Sample	Mock	CPE + replicates / Total replicates			
			1% Triton X-100 Treatment Time			
			0 minutes	30 minutes	1 hour	2 hours
Salivette	1	0/2	3/3	0/3	0/3	0/3
	2	0/2	3/3	0/3	0/3	0/3
	3	0/2	3/3	0/3	0/3	0/3
	4	0/2	3/3	0/3	0/3	0/3
	5	0/2	3/3	0/3	0/3	0/3
	6	0/2	3/3	0/3	0/3	0/3
Saliva into Tube	1	0/2	3/3	3/3**	0/3	0/3
	2	0/2	3/3	0/3	0/3	0/3
	3	0/2	3/3	0/3	0/3	0/3
	4	0/2	3/3	2/3**	0/3	0/3
	5	0/2	3/3	0/3	0/3	0/3
	6	0/2	3/3	0/3	0/3	0/3

* Cytopathic Effect (CPE) observed in Vero E6 cells 5 days post-inoculation with saliva samples spiked with SARS-CoV-2 and subsequently treated with 1% Triton X-100 for varying times. Mock replicates were inoculated with saliva + media. All other replicates contained saliva + virus + Triton X-100 for the indicated time. Orange cells indicate SARS-CoV-2 non-inactivating conditions. Red cells indicate SARS-CoV-2 inactivating conditions. **CPE+wells did not show the same effect upon passage.

Table S3. ROC statistics table for ELISAs conducted on saliva samples*

	Manual IgG		Manual IgA	
Marker	Spike	RBD	Spike	RBD
AUC	0.970	0.947	0.837	0.764
SE.AUC	0.0139	0.0187	0.0341	0.0412
Lower Limit	0.943	0.911	0.770	0.683
Upper Limit	0.998	0.984	0.903	0.845
z	33.77	23.900	9.872	6.399
p-value	0	0	0	0
ROC Coordinate points at <2% FDR				
Marker	Spike	RBD	Spike	RBD
Cutpoint	5.550	3.341	12.172	42.580
FPR	0.0196	0.0196	0.0196	0.0196
TPR	0.886	0.848	0.506	0.304
ROC Coordinate points at <4% FDR				
Marker	Spike	RBD	Spike	RBD
Cutpoint	2.969	2.283	8.667	27.270
FPR	0.0392	0.0392	0.0392	0.0392
TPR	0.924	0.848	0.608	0.430
Samples used to generate ROC curves				
Sample Cohort	No. Samples			
COVID-19 Patient Samples	79			
Pre-COVID-19 Negative Controls	27			
Unexposed Negative Controls Collected in 2020	24			

* Sensitivity values reported in the manuscript are based on the coordinate points at a false positive rate of <2%.