Cumulative survival profiling: a new PAP-based method for detecting heteroresistance in staphylococcal clinical isolates

Ramzi A. Alsallaq[a#], Tina Dao[a], Jason W. Rosch[a], Elisa Margolis[a]

[a] Department of Infectious Diseases, St. Jude Children's Research Hospital, 262 Danny Thomas Place, MS 320, Memphis, Tennessee 38105-3678, USA

Running Title: New PAP-based heteroresistance detection method

[#]Address correspondence to Ramzi Alsallaq, ralsalla@stjude.org

Author contributions: R.A.A. and E.M. designed the study. T.D. and J.W.R. performed the PAP experiments. T.D. and E.M. performed the Etest experiments. R.A.A. developed the CSP method and conducted related analyses. R.A.A. wrote the first draft, and all authors contributed equally to writing the final manuscript.

Conflict of interest: All authors declare no conflict of interest

Abstract

The modified area under the curve of the population analysis profile (PAP-AUC) developed by Wootton et al. (J Antimicrob Chemother 47(4) 2001) is considered the gold standard for detecting heteroresistance in staphylococci. However, the PAP-AUC method does not explicitly control for the variation in the amount of initial inoculum between tested isolates and the comparator reference heterogeneous vancomycin-intermediate *S. aureus* (hVISA) strain (Mu3), which has been reported to influence

classification results. Further, the method requires batching the isolates and utilizing single Mu3 PAP for
each, which does not capture the heterogeneity of Mu3 and adds to the labor intensity of the method.

We confirmed that the area under the PAP curve for Mu3 is strongly dependent on the initial inoculum.
After interpreting this strong relationship, we derived the cumulative survival: a new metric that lacks the
initial inoculum influence and measures the experience under the vancomycin concentration gradient. We
devised a new method, termed cumulative survival profiling (CSP), to infer a distribution for the
cumulative survival metric for isolates; CSP requires conducting multiple PAPs with a broad range of
initial inoculum values. We used CSP to profile Mu3 and one isolate (CoNSB18) from 67 clinical isolates
of coagulase-negative staphylococci associated with central-line bloodstream infection (CoNS CLABSI).
Mu3 and CoNSB18 were the respective references for hVISA and hVISA that is associated with poor
clinical response. Their CSPs reveal two overlapping distributions of cumulative survival, reflecting the
heterogeneity of hVISA and indicating that the most resistant in the hVISA spectrum are likely to be
associated with a poor clinical response. The 2.5% quantiles of their distributions successfully cross-
validated >80% of their PAPs (accuracy). To detect similar isolates to these two reference strains among
the CoNS CLABSI isolates, rather than conducting CSP for each isolate, we retrospectively estimated the
cumulative survivals using point estimate profiling (PEP) utilizing single PAPs. We estimated that two-
thirds of the CoNS CLABSI isolated from unique subjects are heteroresistant and that about 80% of those
may be associated with a poor clinical response.

Standardization of PAP-based detection of heteroresistance and/or poor clinical response is within reach
by using CSP and PEP. Our results should raise the profile of heteroresistance in CoNS CLABSI.

Introduction

Vancomycin is the last-resort treatment for infections caused by multi-drug–resistant bacteria, such as
methicillin-resistant *Staphylococcus aureus* (MRSA). However, therapeutic failures with vancomycin
have been reported in cases where *in vitro* assays measuring the clinical minimum inhibitory
concentration (MIC) indicated susceptibility to vancomycin. These infections are caused by
heterogeneous vancomycin-intermediate staphylococci. In this context, a single isolate represents a
majority of the susceptible population to vancomycin, whereas a tiny proportion of bacterial cells
demonstrate resistance to vancomycin [1]. There have been several cases of reduced susceptibility to
vancomycin only a few decades after the first use of vancomycin. In 1996, a 64-year-old man, who
acquired MRSA pneumonia after an operation for primary lung cancer, experienced failed treatment after
an extensive course of vancomycin [2]. The causative MRSA strain, Mu3, was isolated from sputum and
was categorized as heterogenous vancomycin-resistant *S. aureus* (hVISA). Since then, various methods of

detection have shown vancomycin heteroresistance in other staphylococcal strains, including coagulase-

negative staphylococci (CoNS) [3–9].

The presence of the resistant subpopulation in multi-resistant or opportunistic pathogens (e.g., *S.*

epidermidis and *S. haemolyticus*) contributes to reduced efficacy of vancomycin, limiting treatment

options, especially for patients undergoing chemotherapy for whom vancomycin is the last resort against

bacterial infection. The resistant subpopulation is very difficult to detect and is often misclassified by the

common clinical diagnostic tests. The Clinical and Laboratory Standards Institute methods, which use an

inoculum of approximately 10^4 CFU per assay, cannot detect the resistant subpopulation, which is present

at a frequency of 10^{-6} or lower [10]. Therefore, detecting heteroresistance in clinical isolates with high

accuracy is critical for clinical decision making.

A widely used method for detecting heteroresistance to an antibiotic in staphylococcal clinical isolates is

the modified area under the curve of the population analysis profile (PAP-AUC) method described by

Wootton et.al. [11], which uses the Mu3 strain as a positive-control comparator. The PAP-AUC method is

considered the gold standard with which all other methods are compared [12–15]. In the PAP-AUC

method, a prepared initial inoculum of the bacterial isolate is plated into different plates forming a

gradient of antibiotic concentrations, then the bacterial density that survived the antibiotic at each

concentration is recorded to construct the population analysis profile (PAP). A similar PAP is prepared

for Mu3 in tandem of a batch of one [16] or several [11] isolates' PAPs and used for comparison. Usually,

the initial inoculum values varies widely, by as much as 1.5log10 [17–22], across the batch and differs

from that of the Mu3. It has been reported that the area under the PAP curve is influenced by the initial

inoculum [9].

It is unclear how much this variation in the initial inoculum influences the PAP-AUC results. But if

confirmed to be significant, then neglecting to account for the variation of the initial inoculum could

diminish the role of the experience of isolates under the antibiotic gradient, which is the basis for PAP-

based methods owing to increased PAP-AUC by exposure to antibiotics in heteroresistant staphylococci

[17]. This is because the seemingly increased PAP-AUC could be just due to the initial inoculum of the

isolate being higher than that of Mu3.

Further, a comparative method such as the PAP-AUC must combine cross-validation and detection by

demonstrating high-accuracy heteroresistance detection in replicas of the positive control (Mu3) in

conjunction with and before detecting heteroresistance in clinical isolates. However, following Wootton

et al's cross-validation of their method on replicas of Mu3 PAPs [11], the PAP-AUC method has been

typically applied in literature without the necessary prior cross-validation by using a single Mu3 PAP to

detect heteroresistance in multiple isolates.

Even if the initial inoculums of Mu3 and the isolates' PAPs are hypothetically identical, using a single

Mu3 PAP represents picking a random single value from a broad spectrum that typically characterizes

heteroresistance [23]. For example, a single Mu3 PAP could reflect susceptibility to vancomycin [5].

Using a single Mu3 PAP could lead to biased classification of tested isolates by increasing the possibility

of false-negatives and clouds the association between clinical outcomes and heteroresistance. Therefore, it

is not surprising to see conflicting results on the prevalence of heteroresistance in clinical isolates and on

the association between poor clinical response and heteroresistance.

Here we introduced a PAP-based method that is free from the influence of the initial inoculum,

characterizes the broad spectrum of heteroresistance, and allows cross-validation and detection of

heteroresistance and of heteroresistance that is associated with poor clinical response based on this broad

spectrum. We believe that this method would pave the way to standardize PAP-based detection of

heteroresistance in staphylococci and offers an adequate assessment of the association of heteroresistance

with poor clinical response.

Methods

Bacterial isolates

We retrospectively studied 74 isolates obtained with permission from all patients in whom coagulase-

negative staphylococcal central-line bloodstream infection (CoNS CLABSI) developed during treatment

for leukemia at our institution between January 1, 2010, and March 28, 2016. The deidentified isolates

were stored in glycerol at -80 ºC and identified as CoNS via sequencing by the clinical microbiology

laboratory before being tested for association with clinical outcomes. The prototype heteroresistant

Staphylococcus aureus strain Mu3 was obtained from ATCC (Manassas, VA) and used as a positive

control. Conventional testing (vitek 2, bioMérieux, Inc., Marcy-l'Étoile, France; MIC range <0.5-2

$\mu g/mL$) classified all isolates as being susceptible to vancomycin (catalog number SBR00001, Sigma-

Aldrich, St. Louis, MO). Tryptic Soy Broth (TSB) media (catalog number 211825, BD Sciences) was

used to grow each clinical isolate overnight at 37ºC, and Brain Heart Infusion (BHI, Bacto Brain Heart

Infusion and Bacto agar, BD Biosciences) media was used for agar plates. To identify a clinical isolate as

Staphylococcus aureus or other Staphylococcal species, each one was inoculated onto a mannitol salt agar

plate, which served as a differential growth medium.

Population analysis profiles (PAPs)

117   Population analysis profiling was the basis for all heteroresistance-related detection in clinical isolates.

118   PAP was performed according to standard preparation techniques for staphylococci [11, 24]. Each isolate

119   was grown in 10 mL TSB overnight at 37ºC. Overnight stationary-phase cultures were then serially

120   diluted 1:100 in 1×PBS, ranging from 10^{-2} to 10^{-8}, and 100 μL of each dilution was streaked on BHI agar

plates containing 0, 1, 2, 3, 4, 6, or 8 $\mu g/mL$ of vancomycin. Each isolate was inoculated onto duplicate plates at each dilution at each vancomycin concentration. After incubation for 48 hours at 37ºC, colonies were counted to determine and plot the counts of bacteria in CFU/mL growing at each antibiotic concentration. The resulting PAPs were used in the analyses. The area under the colony-counts is what we used; however, the area under the log10 of the colony-counts is used for Wootton et al.'s method [11].

The isolates were analyzed in two batches, with 44 isolates (59%) that were analyzed at a later stage requiring addition of 0.2% yeast extract (catalog number 212750, BD Sciences) to the overnight culture as a growth supplement. Each batch run was accompanied by replicate runs of Mu3 strain as a positive control (3 and 7 replicas for the first and second patch, respectively, with growth supplement added to each of the 7 replicas' overnight cultures). To determine the effect of the initial inoculum on PAP-AUC, we prospectively obtained 10 more PAPs for Mu3 and 7 PAPs for 1 CoNS CLABSI isolate (CoNSB18) from 1:2 dilutions and 2x and 4x concentrations to create initial inoculum values spanning the range 10^7 to 10^9 CFU/mL for Mu3 and 10^8 to 10^9 CFU/mL for CoNSB18 for overnight cultures that otherwise were prepared and used to generate PAPs as described previously.

Etest

Each clinical isolate and Mu3 were grown in $10 mL$ of TSB overnight at 37ºC and were then diluted 1:100 in fresh 10 mL TSB. The bacterial strains were incubated at 37ºC until OD at 600 nm 0.2 (McFarland unit 2). Then 100 μL of the culture were spread on BHI before placing the vancomycin Etest strips (bio-Mérieux, Marcy-l'Étoile, France) on the plate. The plates were incubated at 37ºC for 48 hours. Each isolate was classified as heteroresistant if isolated colonies were observed in the inhibition eclipse, as determined by two independent readings.

Quantitative and statistical analysis

We created a new format for the analyses whereby each singlet PAP contributes a row, and the columns enlist the bacterial counts in CFU/mL for the PAP at 0,1,2,3,4,6, and 8 $\mu g/mL$ of vancomycin. Each PAP from the isolates and Mu3 was first checked for missing initial inoculum or for having two data points or fewer and dropped if either was true. None of the 20 Mu3 PAPs were dropped, but 14 duplicated PAPs from 7 isolates and 3 singlets were dropped from the analyses for not passing either of the two filters. The remaining isolate PAPs were then divided into two unique isolate sets: training and verification sets of 66 and 65 singlet PAPs from 67 isolates, which included 56 and 54 first-episode isolates from 56 and 54 patients, respectively.

151 The area under the counts and the area under the log10 counts were calculated by using *auc* module from

152 the Python package scikit-learn and were validated with corresponding calculations by using GraphPad

153 Prism 7.

154 The linear relationship between the area under the bacterial counts and the counts at 0 $\mu g/mL$ (the initial

155 inoculum) was firstly explored using ordinary linear square regression fit; the goodness of which was

156 judged by the p-values for the regression coefficients (slope and intercept) and the R-squared value. A

157 better fit in terms of the p-values for the coefficients was obtained using mixed effect linear regression

158 model with random intercept to adjust for PAP to PAP variation. For the cumulative survival profiling

159 (CSP) method we obtained the distribution of the cumulative survival under vancomycin gradient of the

160 reference strain using a mixed-effect linear regression analyses within Bayesian framework (further

161 details are in the Supplementary Information Appendix). Because obtaining replica PAPs is not practical

162 for the CoNS CLABSI isolates, we estimated the cumulative survival using a single PAP by dividing the

163 area under the counts by the initial inoculum. We call this estimate point-estimate-profiling (PEP).

164 We used the 2.5% quantile of the Bayesian credible interval of the distribution obtained by CSP for the

165 reference as a threshold for detection. The detection accuracy of this threshold is determined through

166 cross-validation over 50 iterations by dividing available PAPs of the reference into reference (80% of

167 PAPs) and validation (20%) sets randomly selected in each iteration and then obtaining the mean and

168 standard deviations of the percentage of validation PAPs detected by the method over the iterations. An

169 isolate is heteroresistant if its estimated cumulative survival using PEP is $\geq$ the 2.5% quantile of the

170 Bayesian credible interval of the cumulative survival estimated from the reference set.

171 The cross validation of heteroresistance detection over 20 Mu3 PAPs was compared for four methods:

172 CSP, Wootton's (AUC under the log10 counts) [11, 17, 20, 24, 25], semi-Wootton's (AUC under the

173 counts) [26–28] and altered semi-Wootton's (counts are altered to match the initial inoculum of the

174 reference before calculating AUC) [9]. The detection accuracy via cross-validation for each method was

175 estimated like the CSP method except for the detection criterion. For all methods except the CSP, a given

176 PAP from the validation set is heteroresistant if AUC is $\geq$ 0.9 the AUC of the averaged PAP of the

177 reference.

178 Following successful cross-validation, credible detection of heteroresistance in the CoNS CLABSI

179 isolates by the CSP method is deemed possible. The total number of twenty Mu3 PAPs encompassed the

180 reference set from which a distribution of the cumulative survival was obtained and the cumulative

181 survival for each isolate was estimated by PEP and compared with the 2.5% quantile of the reference

182 similar to what is described above for the Mu3 validation set.

To detect heteroresistance that may be associated with poor clinical response, we used CSP on 7 PAPs

from 1 of the CoNS isolates (CoNSB18), which is found to be both heteroresistant and associated with

poor clinical response under vancomycin treatment. As was done for Mu3, the detection accuracy was

measured first by cross-validation. Association of detections with the measured poor clinical response

was determined by using Cox proportional hazard regression.

We verified the accuracy and robustness of our CSP-based detections over replicated PAPs of isolates by

comparing heteroresistance detection (with and without association with poor clinical response) for

duplicates of 64 common isolates in the training and the verification sets.

Results

Initial inoculum strong influence on PAP-AUC

It has been reported by one study [9] that the initial inoculum has a large influence on PAP-AUC. Using

20 Mu3 PAPs, we confirmed the strong association between the area under the counts and the initial

inoculum ($R^2 = 0.90$ in the log space; Figure 1 A). A one log10 $\mu g/mL$ change in the initial inoculum of

Mu3 increases the area under the PAP counts by 1.0 log10 $CFU/mL.\mu g/mL$ (p<0.001). This strong

dependence holds true also for clinical isolates ($R^2 = 0.90$; Figure 1 B) by plating one selected CoNS

isolate (CoNSB18) using broad range of initial inoculum values. These results indicate that in PAP-based

methods that detect heteroresistance by comparing PAP-AUCs, differences in the initial inoculums of

Mu3, and an isolate such as CoNSB18 will propagate to influence their heteroresistance detection. An

alternative metric to the area under the curve that is free from the influence of the initial inoculum is

needed to adequately compare isolates to Mu3.

Figure 1: The strong dependence of the area under the counts on the initial inoculums for A) Mu3 using

20 PAPs and B) a selected CoNS isolate (CoNSB18) using 7 PAPs. In both cases, $R^2 = 0.90$ and p<0.001.

New metric: the cumulative survival under the antibiotic concentration gradient

The strong relationships in Figure 1 suggest the following model:

Bacterial counts (CFU/mL) at vancomycin concentration c = initial inoculum (CFU/mL) × survival

probability at the given concentration c ---------------------------(1)

which leads to (see Supplementary Information Appendix for the derivation and explanation)

Area under counts (CFU/mL.μg/mL) = initial inoculum (CFU/mL) × cumulative survival under the

concentration gradient (μg/mL) --------------------------(2)

That is, for a single isolate (e.g. Mu3 or CoNSB18 in Figure 1), we interpret the y-intercept in Figure 1 as

the log10 of the "cumulative survival under the antibiotic gradient". In the above model, the information

specifying the experience of the isolate under an antibiotic gradient is fully specified by the cumulative

survival under that gradient, which is the needed metric free from the influence of the initial inoculum.

CSP of Mu3: inferring the full distribution of Mu3 cumulative survival

Here we used the CSP method to infer the value of the cumulative survival of Mu3 under a vancomycin

concentration gradient up to 8 *μg/mL* utilizing 20 Mu3 PAPs.

The Bayesian linear mixed-effects regression model fits data well, as shown in Figure 2 A, and the

inferred cumulative survival for Mu3 (Figure 2 B) reveals not a single value but a wide distribution, with

a median of 1.2 *μg/mL* and a 95% Bayesian credible interval (BCI) of 0.82-1.54 *μg/mL*. To utilize this

distribution, we use the 2.5% quantile of the 95% BCI (q2.5%) to verify and detect heteroresistance of

Mu3 profiles and of clinical isolates, respectively.

Figure 2: Cumulative survival profiling (CSP) of Mu3. A) Bayesian linear mixed-effects regression fit, B)
the inferred distribution of the cumulative survival of Mu3 under a vancomycin gradient up to 8 *μg/mL*.

Cross-validation of Mu3 heteroresistance: comparison with other methods

Compared to three other PAP-based methods used in the literature, the q2.5% of the inferred distribution of cumulative-survival of Mu3 (CSP method) was superior (Table 1), cross-validating the heteroresistance of Mu3 PAPs, on average, 82% of the time. This indicates that the CSP of Mu3 would detect heteroresistance in clinical isolates at a rate higher than that of the three other methods.

Table 1: Cross-validation of 20 Mu3 PAPs using four methods. SD=standard deviation. Detection accuracy=% of Mu3 PAPs detected to be heteroresistant.

Method	AUC (mean, SD)	Detection accuracy
Wootton's[1]	(29.2,7.2)	52 +/- 23
semi-Wootton's[2]	$(7.4,8.7)\ 10^8$ $CFU/mL.\mu g/mL$	46 +/- 26
Altered semi-Wootton's[3]	(1.34,0.78)	60 +/- 26
CSP applied to Mu3	Uses cumulative survival	82 +/- 22

[1]AUC under the log10 counts [11, 17, 20, 24, 25].

[2]AUC under the counts [26–28].

[3]counts are altered to match the initial inoculum of the reference before calculating AUC [9].

Prevalence of heteroresistance in CoNS-CLABSI

Next, we used the cross-validated q2.5% threshold to detect heteroresistance in a set of 66 CoNS-CLABSI PAPs with 56 PAPs representing first-episode samples from 56 subjects (training set). Heteroresistance was detected in 37, suggesting a 66% prevalence among subjects from whom first-episode CoNS CLABSI were sampled (Table 2).

Table 2: Prevalence of heteroresistance in training and verification sets of CoNS CLABSI as revealed by q2.5% of the CSP of Mu3. The numbers and percentages of poor clinical response cases as measured by a clinical laboratory in our institution are also shown.

	Total			First episode		
CoNS CLABSI set	N	Heteroresistant, n (%)	Poor clinical response, n (%)	N	Heteroresistant, n (%)	Poor clinical response, n (%)
Training	66	46 (70)	18 (27)	56	37 (66)	15 (27)
Verification	65	48 (74)	18 (28)	54	39 (72)	15 (28)

We verified the accuracy of our detections over replicated PAPs of isolates by using another set of 65 CoNS-CLABSI PAPs (verification set) with 64 replicated PAPs of isolates in the training set and 54 PAPs representing the first episode from 54 subjects. The estimated prevalence of heteroresistance in the first-episode CoNS-CLABSI PAPs in the verification set is 72% (Table 2). The CSP method accurately classified 58 of 64 replicas by heteroresistance status, giving an accuracy of 91% over replicated PAPs.

The association of detected heteroresistance with measured poor clinical response for CoNS-CLABSI is not significant (p>0.05) regardless of the set of isolates. This is likely because, within the spectrum of the hVISA heteroresistance revealed by the distribution in Figure 2 B, only those with the highest cumulative survival might be intuitively associated with poor clinical response.

Detection of poor clinical response attributable to heteroresistant CoNS CLABSI

To establish whether poor clinical response might be attributable to detected heteroresistance, we needed to estimate the proportion of CoNS CLABSI that are *potentially* associated with poor clinical response by profiling a CoNS CLABSI isolate that is heteroresistant and associated with poor clinical response. The isolate CoNSB18, with the PAP replicas in Figure 1B, fitted these characteristics. The CSP for this isolate revealed a cumulative survival of 1.76 (95% BCI:1.15-2.81) *μg/mL,* suggesting a threshold of 1.15 *μg/mL* to detect heteroresistant CoNS CLABSI that may be associated with poor clinical response. Cross-validating the 7 PAPs for this isolate by using this q2.5% threshold gave an average detection accuracy of 80% (SD=33%) for the heteroresistance that may be associated with poor clinical response.

We then detected similar isolates among our CoNS CLABSI samples using this cross-validated threshold (Table 3). Cox proportional hazard regression analyses reveal significant association between our detections and time to measured poor clinical response or censorship (p<0.05), regardless of the set of isolates (Table 3). We estimated that more than 80% (31/37 and 32/39 from the training and verification sets, respectively) of heteroresistant CoNS CLABSI may be associated with poor clinical response. The threshold that successfully detected heteroresistance with potential association with poor clinical response is about the median for Mu3 CSP (Figure 3).

Table 3: Prevalence of heteroresistance that is potentially associated with poor clinical response in training and verification sets of CoNS CLABSI as revealed by q2.5% of the CSP of CoNSB18.

CoNS CLABSI set	Total			First episode		
	N	Heteroresistant and associated with poor outcomes, n (%)	p-value	N	Heteroresistant and associated with poor outcomes, n (%)	p-value
Training	66	40 (61)	0.018	56	31 (55)	0.019
Verification	65	41 (63)	0.030	54	32 (59)	0.039

Figure 3: Inferred cumulative survival profiles for Mu3 and CoNSB18. The cumulative survival threshold that successfully detected heteroresistant isolates with potential association with poor clinical response (hVISA + poor outcomes) is higher than the threshold that successfully detected all heteroresistant isolates regardless of association with clinical outcomes (hVISA).

Comparison with Etest

Next, we compared Etest classification results to the two heteroresistance classifications: Mu3-like and CoNSB18-like detection, taking each separately as the gold standard. Etest accurately predicted the heteroresistance status of 60% and 62% of the isolates as respectively classified by Mu3-like and CoNSB18-like detection. The corresponding sensitivities of Etest were 75% and 79% and the specificities were 26 and 33%, respectively. The results were robust across PAP replicas (Table 4).

Table 4: Comparing Etest classification to the classifications by CSP of Mu3 and of CoNSB18, taking each separately as the gold standard.

Attribute	Mu3-like (training/verification), n/n	CoNSB18-like (training/verification), n/n
Accuracy	60/61	62/63
Sensitivity	75/73	79/79
Specificity	26/29	33/38

Discussion

The initial inoculum used in PAP-based heteroresistance detections in staphylococci strongly influences the area under the PAP curve and should not be part of evaluating the experience under the antibiotic gradient. We introduced the cumulative survival under the antibiotic gradient as an alternative metric and a distribution of which must be estimated for reference strains. We introduced our method, cumulative

survival profiling (CSP), for inferring this distribution and removing the influence of the initial inoculum

by using replicated PAPs with a wide range of initial inoculum values. We applied the CSP method on

two reference isolates Mu3 (heteroresistant) and CoNSB18 (heteroresistant and associated with poor

clinical response). We estimated that about two-thirds of our first-episode CoNS CLABSI isolates are

heteroresistant to vancomycin and that 80% of the heteroresistant isolates may be associated with a poor

clinical response. These estimates raise the profile of heteroresistance to vancomycin as a risk factor for

poor clinical response in CoNS-CLABSI isolated from the pediatric leukemia population.

The inclusion of cross-validation in the CSP method is critically important, as it ensures the reliability of

reference-like detection based on cumulative survival in clinical isolates. The cross-validation of the CSP

method reveals a false-negative rate of <20%. Although this false-negative rate might never by zero for

Mu3 given that some PAPs of Mu3 inherently reflect susceptibility to vancomycin (which argues against

using single Mu3 PAPs for detecting heteroresistance), using more PAPs would decrease this rate.

Our CSP profiles are specific to the vancomycin gradient and might be specific to the experimental

preparation that was followed for PAPs and should not be directly applicable to other antibiotics, different

vancomycin gradients, or different experimental preparations. Apart from Mu3, we have not tested our

method on *S. aureus* isolates, and the samples of CoNS used might not represent other populations. Using

different methods, it has been observed that prevalence of hVISA could vary among populations [26].

The finding that a larger cumulative survival threshold (median compared to q2.5% of the CSP of Mu3)

was needed to successfully detect heteroresistance associated with a poor clinical response than that

needed to detect all heteroresistance highlights the wide spectrum of heteroresistance as revealed by the

inferred distribution from CSP of Mu3 and that only the most-resistant CoNS isolates (those with

relatively higher cumulative survival above the median CSP distribution of Mu3) may be associated with

poor clinical response (as revealed by the inferred distribution from CSP of CoNSB18). The predictability

of heteroresistance with potential association with a poor clinical response using the median of Mu3

might not be generalizable to other staphylococci strains, to other CoNS with PAPs conducted by

different experimental procedures, or to a different population from which the CoNS are sampled. This

variability suggests that an adequate reference isolate is always needed to predict/confirm association

with clinical outcomes of prospective isolates.

Therefore, we are making the method available to other investigators so that they may test it on their

isolates prospectively or retrospectively.

Although the most robust estimate for the cumulative survival would require profiling each clinical isolate

by applying CSP using multiple PAPs of the isolate, in case of multiple clinical isolates, a single PAP for

each might be enough. However, this supposition warrants further testing. We showed robustness of
classification results to replicated PAPs of isolates.

The comparison of detection results of Etest to CSP-detection of CoNSB18 indicate a relatively low false-
negative rate of 20%, suggesting that Etest might be used first on CoNS to detect hVISA-like
heteroresistance associated with poor clinical response, with an understanding of the limitation of this
false-negative rate.

Our estimate of the heteroresistance prevalence in CoNS is consistent with the 68% prevalence of
heteroresistance among MRSA CoNS [4] and with the 62.5% prevalence estimate among MRSA *S. captis*
from blood samples in neonatal intensive care units (NICUs) [7] and within the range reported among
CoNS in NICUs [6].

The CSP method offers asynchronous profiling of Mu3 and prospective isolates, whereby the same
experimental setup is used to first conduct PAPs for Mu3 and build its cumulative survival profile. The
detection accuracy is then estimated via cross-validation using the Mu3 PAPs. The same experimental
setup is used to conduct PAPs for isolates as they come forward and to detect heteroresistance in each.
This method could pave the way to standardization of the detection of heteroresistance and its association
with clinical response in staphylococci based on cumulative survival comparisons. Along these lines, the
asynchronous profiling must not mean different experimental procedures are used for the reference and
isolates' PAPs. Any differences should be investigated and reported. In this regard, the CSP method
explicitly controls for the variation in the initial inoculum and only the random experimental errors
occurring within the same experimental procedure. The addition of growth supplement to the overnight
cultures would have a direct effect on the distribution of the initial inoculum.

Acknowledgment
348 We thank Elaine Tuomanen for discussion and encouragement. We thank Cherise Guess, PhD ELS for
349 scientifically editing the manuscript.

Funding
351 EM receives funding from ALSAC. JWR is supported by 1U01AI124302 and 1RO1AI110618.

References

1. El-Halfawy OM, Valvano MA. Antimicrobial Heteroresistance: an Emerging Field in Need of Clarity.
Clin Microbiol Rev. 2015;28:191–207.

2. Hiramatsu K, Hanaki H, Ino T, Yabuta K, Oguri T, Tenover FC. Methicillin-resistant Staphylococcus
aureus clinical strain with reduced vancomycin susceptibility. J Antimicrob Chemother. 1997;40:135–6.

3. Becker K, Heilmann C, Peters G. Coagulase-Negative Staphylococci. Clinical Microbiology Reviews.
2014;27:870–926.

4. Sieradzki K, Villari P, Tomasz A. Decreased Susceptibilities to Teicoplanin and Vancomycin among
Coagulase-Negative Methicillin-Resistant Clinical Isolates of Staphylococci. Antimicrobial Agents and
Chemotherapy. 1998;42:100–7.

5. Nunes APF, Teixeira LM, Iorio NLP, Bastos CCR, de Sousa Fonseca L, Souto-Padrón T, et al.
Heterogeneous resistance to vancomycin in Staphylococcus epidermidis, Staphylococcus haemolyticus
and Staphylococcus warneri clinical strains: characterisation of glycopeptide susceptibility profiles and
cell wall thickening. Int J Antimicrob Agents. 2006;27:307–15.

6. Chong J, Caya C, Lévesque S, Quach C. Heteroresistant Vancomycin Intermediate Coagulase Negative
Staphylococcus in the NICU: A Systematic Review. PLOS ONE. 2016;11:e0164136.

7. Rasigade J-P, Raulin O, Picaud J-C, Tellini C, Bes M, Grando J, et al. Methicillin-resistant
Staphylococcus capitis with reduced vancomycin susceptibility causes late-onset sepsis in intensive care
neonates. PLoS ONE. 2012;7:e31548.

8. Van Der Zwet WC, Debets-Ossenkopp YJ, Reinders E, Kapi M, Savelkoul PHM, Van Elburg RM, et
al. Nosocomial spread of a Staphylococcus capitis strain with heteroresistance to vancomycin in a
neonatal intensive care unit. J Clin Microbiol. 2002;40:2520–5.

9. D'mello D, Daley AJ, Rahman MS, Qu Y, Garland S, Pearce C, et al. Vancomycin heteroresistance in
bloodstream isolates of Staphylococcus capitis. J Clin Microbiol. 2008;46:3124–6.

10. Antimicrobial and Antifungal Susceptibility Testing Resources. Clinical & Laboratory Standards
Institute. https://clsi.org/about/about-clsi/about-clsi-antimicrobial-and-antifungal-susceptibility-testing-
resources/. Accessed 2 Jul 2019.

11. Wootton M, Howe RA, Hillman R, Walsh TR, Bennett PM, MacGowan AP. A modified population
analysis profile (PAP) method to detect hetero-resistance to vancomycin in Staphylococcus aureus in a
UK hospital. J Antimicrob Chemother. 2001;47:399–403.

12. Satola SW, Farley MM, Anderson KF, Patel JB. Comparison of Detection Methods for
Heteroresistant Vancomycin-Intermediate Staphylococcus aureus, with the Population Analysis Profile
Method as the Reference Method. Journal of Clinical Microbiology. 2011;49:177–83.

13. Hal SJ van, Wehrhahn MC, Barbagiannakos T, Mercer J, Chen D, Paterson DL, et al. Performance of
various testing methodologies for detection of heteroresistant vancomycin intermediate Staphylococcus
aureus (hVISA) in bloodstream isolates. Journal of Clinical Microbiology. 2011.
doi:10.1128/JCM.02302-10.

14. Walsh TR, Bolmström A, Qwärnström A, Ho P, Wootton M, Howe RA, et al. Evaluation of Current
Methods for Detection of Staphylococci with Reduced Susceptibility to Glycopeptides. Journal of
Clinical Microbiology. 2001;39:2439–44.

15. Wootton M, MacGowan AP, Walsh TR, Howe RA. A Multicenter Study Evaluating the Current
Strategies for Isolating Staphylococcus aureus Strains with Reduced Susceptibility to Glycopeptides.
Journal of Clinical Microbiology. 2007;45:329–32.

16. Walsh TR, Bolmström A, Qwärnström A, Ho P, Wootton M, Howe RA, et al. Evaluation of Current
Methods for Detection of Staphylococci with Reduced Susceptibility to Glycopeptides. Journal of
Clinical Microbiology. 2001;39:2439–44.

17. Bhateja P, Purnapatre K, Dube S, Fatma T, Rattan A. Characterisation of laboratory-generated
vancomycin intermediate resistant Staphylococcus aureus strains. International Journal of Antimicrobial
Agents. 2006;27:201–11.

18. da Costa TM, Morgado PGM, Cavalcante FS, Damasco AP, Nouér SA, Dos Santos KRN. Clinical
and Microbiological Characteristics of Heteroresistant and Vancomycin-Intermediate Staphylococcus
aureus from Bloodstream Infections in a Brazilian Teaching Hospital. PLoS ONE. 2016;11:e0160506.

19. Harigaya Y, Ngo D, Lesse AJ, Huang V, Tsuji BT. Characterization of heterogeneous vancomycin-
intermediate resistance, MIC and accessory gene regulator (agr) dysfunction among clinical bloodstream
isolates of staphyloccocus aureus. BMC Infectious Diseases. 2011;11. doi:10.1186/1471-2334-11-287.

20. Khosrovaneh A, Riederer K, Saeed S, Tabriz MS, Shah AR, Hanna MM, et al. Frequency of reduced
vancomycin susceptibility and heterogeneous subpopulation in persistent or recurrent methicillin-resistant
Staphylococcus aureus bacteremia. Clin Infect Dis. 2004;38:1328–30.

21. Passalacqua KD, Satola SW, Crispell EK, Read TD. A Mutation in the PP2C Phosphatase Gene in a
Staphylococcus aureus USA300 Clinical Isolate with Reduced Susceptibility to Vancomycin and
Daptomycin. Antimicrob Agents Chemother. 2012;56:5212–23.

22. Silveira AC de O, da Cunha GR, Caierão J, de Cordova CMM, d'Azevedo PA. Molecular
epidemiology of heteroresistant vancomycin-intermediate Staphylococcus aureus in Brazil. The Brazilian
Journal of Infectious Diseases. 2015;19:466–72.

23. El-Halfawy OM, Valvano MA. Antimicrobial Heteroresistance: an Emerging Field in Need of Clarity.
Clin Microbiol Rev. 2015;28:191–207.

24. Wootton M, MacGowan AP, Walsh TR, Howe RA. A Multicenter Study Evaluating the Current
Strategies for Isolating Staphylococcus aureus Strains with Reduced Susceptibility to Glycopeptides.
Journal of Clinical Microbiology. 2007;45:329–32.

25. Satola SW, Farley MM, Anderson KF, Patel JB. Comparison of Detection Methods for
Heteroresistant Vancomycin-Intermediate Staphylococcus aureus, with the Population Analysis Profile
Method as the Reference Method. Journal of Clinical Microbiology. 2011;49:177–83.

26. Bae I-G, Federspiel JJ, Miró JM, Woods CW, Park L, Rybak MJ, et al. Heterogeneous vancomycin-
intermediate susceptibility phenotype in bloodstream methicillin-resistant Staphylococcus aureus isolates
from an international cohort of patients with infective endocarditis: prevalence, genotype, and clinical
significance. J Infect Dis. 2009;200:1355–66.

27. Rybak MJ, Leonard SN, Rossi KL, Cheung CM, Sadar HS, Jones RN. Characterization of
Vancomycin-Heteroresistant Staphylococcus aureus from the Metropolitan Area of Detroit, Michigan,
over a 22-Year Period (1986 to 2007). J Clin Microbiol. 2008;46:2950–4.

28. Hal SJ van, Jones M, Gosbell IB, Paterson DL. Vancomycin Heteroresistance Is Associated with
Reduced Mortality in ST239 Methicillin-Resistant Staphylococcus aureus Blood Stream Infections. PLOS
ONE. 2011;6:e21217.

29. Salvatier J, Wiecki TV, Fonnesbeck C. Probabilistic programming in Python using PyMC3. PeerJ
Comput Sci. 2016;2:e55.

Supplementary Information Appendix

Deriving the cumulative survival under vancomycin from PAP profile

Recasting Eq.1 in the main text using mathematical symbols for convenience:

$$B(c) = CFU_0 \times S(c) \quad \text{-------------------------------------} \text{(S1)}$$

Where (CFU_0) is the initial inoculum and does not depend on the concentration of antibiotic (c) and

$S(c)$ is the concentration-dependent survival fraction.

The cumulative survival can be derived from the PAP by mathematically integrating Eq. S1 over vancomycin gradient:

$$\int_0^{c_{max}} B(c)dc = \int_0^{c_{max}} CFU_0 \times S(c)dc = CFU_0 \times CumSurvival \quad \text{-------------------------} \text{(S2)}$$

The left-hand side of Eq.S2 is the area under the counts. Thus, the cumulative survival can be obtained from a single PAP by dividing the area under the counts by the initial inoculum and for multiple PAPs by using regression analysis.

Bayesian mixed-effects regression analysis

For inferring the cumulative survival distribution for reference strains in the CSP method, we used Bayesian linear mixed-effects model. To do that we have $1,2,3,...,N$ PAPs that span a wide range of initial inoculum. Let the areas under the counts for these PAPs in $CFU/mL.\mu g/mL$ be y_i and the initial inoculums be x_i. We wish to infer the cumulative survival ξ. From Eq.S2 we know that:

$$log10(y_i) = log10(x_i) + log10(\xi) \quad \text{-------------------------} \text{(S3)}$$

To incorporate the random effect representing PAP to PAP variation, we introduce zero-centered random deviations from the mean initial inoculum of each PAP. Let this deviation be $\epsilon_i \sim N(0,10^8)$, then the mixed-effects regression model becomes:

$$log10(y_i) = log10(x_i + \epsilon_i) + log10(\xi) \quad \text{-------------------------} \text{(S4)}$$

Equation S4 defines the mean of our model and can be fitted with zero-centered residuals:

$$log10(y_i) \sim N(log10(x_i + \epsilon_i) + log10(\xi), \sigma^2) \quad \text{-------------------------} \text{(S5)}$$

Posterior samples were obtained by using MCMC with NUTS sampler by utilizing the python package PYMC3 [29]