Current and emerging colistin resistance diagnostics: a systematic review of established and innovative detection methods

Tumisho Mmatumelo Seipei Leshaba[a], Nontombi Marylucy Mbelle[a], John Osei Sekyere[a#]

[a]Department of Medical Microbiology, School of Medicine, Faculty of Health Sciences, University of Pretoria, South Africa.

#**Address correspondence to** Dr. John Osei Sekyere, jod14139@yahoo.com; u18392467@tuks.co.za

Running head: Colistin resistance, current detection methods**.**

Tweet: "*Colistin-resistance is a global threat to bacterial infections' management. Being a last-resort antibiotic for treating of MDR Gram-negative infections, it is important to establish an efficient diagnostic method for routine use in clinical microbiology laboratories. Herein, current & emerging diagnostics for detecting colistin resistance are described to inform clinical diagnostic options.*"

Author summary/highlights/importance

- The spread of colistin resistance is rapidly increasing, threatening public and veterinary healthcare.
- As one of the last-line antibiotics, colistin has to be conserved by containing the spread of colistin resistance.
- Detecting colistin resistance relies on determining colistin MIC values by standard broth microdilution, which is labour-intensive with longer turnaround time (TAT).
- Other colistin resistance diagnostics have been developed to augment or replace the broth microdilution with faster TAT.
- Based on their respective sensitivities, specificities, TAT, skill, and cost, selected phenotypic and molecular assays are recommended for laboratories, according to their financial strengths, to enhance colistin resistance surveillance and control.

Abstract

Background. The escalation of colistin resistance, due to transferable *mcr*-genes, threatens public and animal health as there are limited therapeutic options. Given that colistin is a last-line antibiotic, there is a need to contain the spread of its resistance to conserve its efficacy. Herein, current and emerging colistin resistance diagnostics are described to inform faster clinical diagnostic choices.

Methods. A literature search in Pubmed, between 2016 and 2020, was performed. English articles evaluating colistin resistance methods/diagnostics were included.

Results. Screening resulted in the inclusion of 89 journal articles. Current colistin resistance diagnostics are either phenotypic or molecular. Broth microdilution (BMD) is currently the only golden standard for determining colistin MIC values. However, other methods have been developed to overcome the challenges associated with the BMD.

Phenotypic methods comprise agar-based methods *i.e.* ChromAgar Col-APSE, SuperPolymyxin, ChromID Colistin R *etc*., MIC-determiners *i.e*. Microscan, BD Phoenix, Sensitest colistin, Sensititre, Micronaut *etc*., MCR-detectors *i.e.* lateral flow immunoassay (LFI), chelator based assays *i.e.* EDTA- and DPA-based tests as well as biochemical colorimetric tests *i.e.* Rapid Polymyxin NP test, Rapid ResaPolymyxin NP test *etc*. Molecular methods only characterize mobile colistin resistance; they include PCR, LAMP, whole-genome sequencing (WGS) *etc*.

Conclusion. Considering the efficiency (sensitivity, specificity, turnaround time) of the Rapid ResaPolymyxin NP test, it is recommended for the initial screening of colistin-resistant isolates. This can be followed by chelator-based tests or the LFI for the rapid screening of *mcr*-genes. However molecular assays such as LAMP, PCR may be considered for use in well-equipped clinical laboratories.

Keywords: Colistin resistance; detection methods; *mcr;* diagnostics; Polymyxins

1. Introduction

The rapid dissemination of multidrug-resistant (MDR) *Acinetobacter spp.*, *Pseudomonas spp.* and carbapenemase-producing *Enterobacteriaceae* (CPE) has been of vital significance to public and veterinary health [1, 2]. Of particular concern, is infections caused by these organisms that are resistant to carbapenems [3]. Carbapenem-resistant infections caused by Gram-negative bacteria are a clinical challenge as they are associated with high mortality rates owing to

limited therapeutic options [4, 5]. The limited pipeline of new antibiotic classes has led to the

increased use of colistin alone or in combination with tigecycline or fosfomycin for the

treatment of MDR Gram-negative infections [6, 7].

Colistin is of particular value as a last-line antibiotic for treating MDR and carbapenem-

resistant Gram-negative infections as it is bactericidal unlike tigecycline, which is

bacteriostatic, is not readily available in many countries [8]. Colistin consists of hydrophilic and

lipophilic moieties that allow it to have stable polar and hydrophobic interactions with the

lipopolysaccharide (LPS) membrane of Gram-negative bacteria [9, 10]. These interactions result

in the destruction of the LPS membrane, causing the cytoplasmic content to leak out and

ultimately killing the cell [10].

However, the increased use of colistin to treat MDR Gram-negative infections has led to the

emergence of acquired colistin resistance [11]. Several mechanisms that mediate acquired

colistin resistance have been identified, the most common being chromosomal mutations and

plasmid-borne colistin resistance [10, 12, 13]. Chromosomal mutations result in modification(s) of

the LPS membrane using different mechanisms, which may include: (i) the addition of 4-

amino-4-deoxy-L-arabinose (L-Ara-4N), phosphoethanolamine (pETN) or galactosamine

moieties at the 4' or 1' position of lipid A, which reduces the overall anionic charge of the

LPS; (ii) the overexpression of efflux pumps systems; (iii) overproduction of capsule

polysaccharide that reduces the LPS membrane's permeability [8, 12].

Plasmid-borne colistin resistance involves the acquisition of a mobile colistin resistance (*mcr*)

gene that encodes a pETN transferase [7]. Since the discovery of the first plasmid-borne *mcr*-1

gene in *Escherichia coli* in China, other *mcr* variants viz., *mcr*-2, *mcr*-3, *mcr*-4, *mcr*-5, *mcr*-6,

mcr-7, *mcr*-8, *mcr*-9 and *mcr*-10, have been described worldwide [6, 14-21].

Currently, the detection of colistin resistance relies on the broth microdilution (BMD), which

only determines the minimum inhibitory concentration value of colistin [22]. Although the

BMD is the golden standard for colistin susceptibility testing, it is not suitable for routine

clinical use as it is time-consuming and is associated with methodological issues [23]. It is

imperative to establish rapid and reliable methods that are able to detect colistin resistance

regardless of the mechanism of resistance in order to effectively manage the dissemination of

colistin resistance [12, 23].

There has been an increasing interest in discovering alternative methods of detecting

resistance to colistin arising from both chromosomal mutations and plasmid-borne *mcr* genes

[1,24]. These methods can be categorized as either phenotypic or molecular methods [12,23]. This
review aims to summarize and analyse clinical diagnostic methods that are currently available
for detecting colistin resistance.

Evidence before this review

Methods used to detect colistin resistance have been reviewed [1,12,23]. Bardet and Rolain
(2018) narratively described methods used to detect colistin resistance, focusing mainly on
their ability to detect all colistin resistance mechanisms. They also analysed methods
specifically used to detect plasmid-mediated colistin resistance. Osei Sekyere (2019)
comprehensively described polymyxin resistance and *mcr*-detecting diagnostic methods,
including the composition of culture media, primers, and cycling conditions of PCR methods,
sensitivities, specificities, turnaround time, skill, relative cost, essential agreement (EA),
categorical agreement (CA), major error (ME) and very major error (VME) in a systematic
review. However, since 2018, other methods and new evaluation studies have been reported,
broadening our understanding and conclusions of the best colistin resistance
diagnostics/methods. Therefore, we herein provide a comprehensive update and expanded
review, based on broader evaluation studies, of all the current diagnostic methods designed to
detect colistin resistance.

Literature search strategy

A comprehensive literature search was performed using Pubmed. English journal articles
published from 2016 to April 2020, were retrieved and screened using the following
keywords: "colistin AND resistant" in permutation and combinations with "detection" and
"diagnostics". The search was based on articles that were evaluating methods that are
currently used for the detection of colistin resistance and *mcr*-genes. Studies that are non-
English, reviews, case reports, case studies, risk factors, epidemiology, and surveillance were
excluded. The inclusion and exclusion methods used in this review are as demonstrated in
Figure 1. The following data was extracted from the included articles and summarized in
Table 1: Diagnostic methods used, types and sample size (in numbers) of bacterial species
used in the evaluation, sensitivity, specificity, EA, CA, ME, VME, relative cost and
turnaround time.

2. Phenotypic tests

Broth Microdilution (BMD)

The CLSI and EUCAST have jointly recommended that the MIC (minimum inhibitory

concentration) testing of colistin be performed according to the ISO-20776 standard BMD

method [25, 26]. Methods such as agar dilution, disc diffusion and gradient diffusion have been

ruled out as it was shown that the large molecular size of colistin makes it poorly diffusible

through agar [1, 26, 27].

It has been recommended that BMD be used with plain polystyrene trays and colistin sulfate

salt without the addition of any surfactants [26, 28]. The CLSI had initially recommended the

addition of polysorbate-80 (P-80) to alleviate the binding of polymyxin to negatively charged

polystyrene surfaces (binding of colistin to polystyrene reduces the concentration of colistin).

However, there were concerns that the surfactant would act in a synergistic manner with

colistin, and is therefore not recommended at this time [28]. Moreover, several studies have

suggested that the loss of colistin concentration could be reduced by minimizing contact with

unexposed pipette tips and by storing colistin solution in glass tubes [28].

The CLSI and EUCAST have established a susceptible breakpoint of ≤ 2 mg/L and a resistant

breakpoint of >2 mg/L for colistin for *Acinetobacter spp.* and *Pseudomonas aeruginosa* (*P.*

aeruginosa). However, only the EUCAST has the same breakpoints for *Enterobacteriaceae*

while the CLSI has an epidemiological cut-off value of 2 mg/L that defines *Escherichia coli*,

Klebsiella pneumoniae, *Raoultella ornithiolytica*, *Enterobacter aerogenes* and *Enterobacter*

cloacae as either wild type or non-wildtype [1, 25].

BMD performed by the reference ISO-20776 is currently the only gold standard for

determining colistin MIC and evaluating CA, EA, ME and VME, yet it is laborious and rarely

performed in routine clinical microbiology laboratories [1]. Instead, diffusion methods and

automated AST systems are more commonly used [1]. Chew *et al* (2017) evaluated the

detection of *mcr*-1 positive *Enterobacteriaceae* by BMD in comparison with commercial

automated AST systems viz., Sensititre, Microscan and Vitek 2, and a gradient diffusion test

i.e., E-test. This study found that none of the commercial testing methods meet the CLSI

recommendation standard for commercial antimicrobial susceptibility testing systems: EA

$\geq 90\%$, CA $\geq 90\%$, VME $\leq 1.5\%$, and ME $\leq 3.0\%$. Even so, the Sensititre and Microscan were

shown to detect 100% of the *mcr*-1 positive isolates whereas the BMD only detected 71.4% at

a susceptible breakpoint of ≤ 2 mg/L [25]. Therefore, BMD cannot be completely reliable for the

detection of *mcr*-producing isolates, though its overall sensitivity could be improved by

reducing the susceptible breakpoint to ≤ 1 mg/L and using microtitre plates that were

manufactured to reduce adsorption [25, 28].

The methodological challenges surrounding the standard BMD have led to interest in finding
alternative means for detecting colistin resistance in Gram-negative bacteria[1, 28]. Broth
macrodilution method has been shown to be an obvious alternative, as it employs the use of
glass tubes instead of polystyrene [28]. Notwithstanding, the broth macrodilution is not a
popular alternative because it requires the same preparation and turnaround time as the
standard BMD. Hence, commercially available selective media and rapid colorimetric assays
have become popular for screening (Table 1) [28].

Diffusion and Agar dilution methods

Diffusion methods for colistin susceptibility testing are still commonly used despite being
disapproved by the CLSI and EUCAST [1, 27, 29, 30]. Two of the commercial gradient tests, E-test
and MIC Test Strip, were both found to perform poorly for colistin-resistant isolates in one
study [30]. Disc diffusion is not an MIC determiner, although it has been found to perform
similarly to gradient diffusion tests, generating high levels of VMEs (false-susceptible) while
agar dilution has a tendency of generating higher MICs (which results in high MEs) than the
reference method [1, 25, 27, 28, 30-32]. These findings further support the conclusion that
polymyxins are poorly diffused in agar and therefore corroborates the CLSI and EUCAST
recommendations to abandon diffusion and agar dilution methods [27, 30, 32]. However, disc
diffusion may be considered when detecting colistin resistant isolates with high colistin MIC
values as its evaluation in one study showed no VMEs for colistin resistant *K. pneumoniae*
and *A. baumannii* isolates that had a high level of resistance (Table 1) [28].

Manual commercial MIC-determiners: UMIC, MMS, and ComASP Colistin

UMIC (Biocentric, Bandol, France), MICRONAUT MIC-Strip (MMS) (MERLIN
Diagnostika GmbH, Bornheim, Germany), ComASP Colistin (formerly SensiTest,
Liofilchem, Italy) are non-automated broth microdilution-based tests [30]. UMIC and MMS
both consist of a plastic device with 12 wells that allow for different colistin concentrations to
be tested for a single isolate without the need for any additional equipment [26, 30, 33]. While the
ComASP (SensiTest) Colistin, consists of a compact panel for four isolates with freeze-dried
antibiotics in seven two-fold dilutions [5, 34]. Matuschek *et al* (2018) evaluated all three tests,
where ComASP had a poor performance for *Acinetobacter* spp., and UMIC was poor for
Acinetobacter spp. and *P. aeruginosa*. The overall performance of ComASP was improved
(as shown in Table 1) when certain species were removed and *K. pneumoniae* and *E. coli*
were tested, therefore suggesting that ComASP is not suitable for testing all species [5].

UMIC was generally found to lower the MIC of some colistin-resistant isolates, which may result in failure to detect colistin-resistant isolates with low MICs (≤8 mg/L) [26, 33]. More so, UMIC failed to detect four *Stenotrophomonas maltophilia* isolates with MICs ranging from 8mg/L to 32mg/L in one study, although in another study all *S. maltophilia* isolates were detected accordingly [26, 33]. The MMS had the best correlation to the BMD amongst the three tests; however, there aren't sufficient studies evaluating this test (Table 1)[30].

Automated AST systems: BD Phoenix, MICRONAUT-S, MicroScan, Sensititre and Vitek 2

Automated broth microdilution systems are of particular interest for colistin susceptibility testing due to their ease of use as compared to the reference BMD [26, 30]. Several studies have shown that some automatic BMD systems can achieve results that are relatively similar to those of the reference BMD [25, 26, 30, 32, 35]. The Sensititre recorded a high rate of agreement with the reference BMD in several studies, particularly demonstrating the highest potential for detecting *mcr*-1-producing *Enterobacteriaceae* together with MicroScan in one study [25, 26, 30]. However, MicroScan has a tendency of overestimating MICs (which may result in false-resistant isolates) of *E. cloacae*, *Salmonella spp.* and non-fermenters [5, 26]. Vitek 2 was found to have the highest rate of VMEs at 36% for *mcr*-1 producing isolates and failing to detect colistin resistance in eight *mcr*-negative *E. cloacae* complex isolates [25]. MICRONAUT-S (MERLIN Diagnostika GmbH, Bornheim, Germany) performed similarly to the Sensititre although it had two VMEs whereas the Sensititre had none [30]. The Sensititre test is unique as it can be performed manually or semi-automated unlike Microscan, MICRONAUT-S, Vitek 2and BD Phoenix that require an automated inoculation delivery system [26, 30].

The detection of colistin resistance by BD Phoenix highly correlated with that of the reference BMD method [32, 35]. Nevertheless, the BD Phoenix 100 system failed to detect colistin-resistant isolates that may have hetero-resistance and the BD Phoenix M50 had 5.7% MEs, all of which were within twofold dilutions, as compared to the reference BMD [35-37]. Notably, the BD Phoenix M50 has not been evaluated on strains that exhibit hetero-resistance; hence, its evaluation is limited as compared to BD Phoenix 100 (Table 1) [35].

Chelator-based and non-chelator-based tests

Rapid Polymyxin NP

The Rapid Polymyxin NP is a colorimetric test that is based on glucose metabolism to detect the growth of *Enterobacteriaceae* at a given concentration of a polymyxin (Colistin or polymyxin B) [11]. Resistance to polymyxins is demonstrated by a colour change (orange to

yellow) of a pH indicator, *i.e.* phenol red, as a result of acid formation associated with the

metabolism of glucose [11, 38]. Rapid polymyxin NP test is commercially available

(ELITechGroup, Puteaux, France) and can also be performed in-house with the preparation of

two solutions [11, 39]. The in-house rapid polymyxin NP test is prepared with stock solutions of

polymyxins and a rapid polymyxin NP solution, which consists of cation-adjusted Mueller-

Hinton broth powder, phenol red indicator and D(+)-glucose [11]. This test has demonstrated an

excellent detection of colistin-resistance in several studies, including detecting colistin

resistance directly from blood cultures with a sensitivity of 100% [40, 41]. However, in a study

using ComASP as the reference, the Rapid Polymyxin NP test had a lower specificity than

BD Phoenix, Vitek 2 and E-test with *K. pneumoniae* isolates (Table 1)[42].

Although the Rapid Polymyxin is limited to use on *Enterobacteriaceae*, the test is easy to

perform and the final results can be read in no more than four hours with the majority of the

results being positive in two hours [31, 32, 40, 43].

Andrade Screening Antimicrobial Test (ASAT)

Following the Rapid Polymyxin NP test, another colorimetric assay for detecting colistin

resistance in *Enterobacteriaceae* was developed [13]. The Andrade Screening Antimicrobial

Test (ASAT) was developed using an in-house broth consisting of 10g peptone, 5g sodium

chloride, 3g beef extract and 10mL Andrade indicator made with 0.5g acid fuchsin and 16mL

NaOH in 100mL water [13]. The evaluation of the ASAT was performed on 300

Enterobacteriaceae in tubes containing 175 µL of Andrade broth and colistin at a

concentration of 3.8 mg/L [13]. The test achieved a sensitivity and specificity of 90.7% and

100% respectively, where a positive result in the presence of colistin was shown by a change

in colour of the Andrade indicator (light pink to magenta) [13].

The ASAT test was further evaluated against the BD Phoenix using 1096 *Enterobacteriaceae*

clinical isolates [13]. However, this evaluation demonstrated discrepancies between the two

methods where ten *E. coli*, seven *K. pneumoniae* and one *E. cloacae* colistin resistant isolates

were accurately detected by ASAT and not by BD Phoenix [13]. Most of the isolates that were

not detected by BD Phoenix had colistin MIC values ranging between 4-8 µg/mL and carried

mcr-1 genes [13]. Although most of the colistin-resistant *K. pneumoniae* isolates had colistin

MIC values >16 µg/mL and carried bla_{KPC} genes, only three isolates that had colistin MIC

values ≤16 µg/mL showed discrepancies between the two methods [13].

Rapid Detection of non-fermenters

ElitechGroup has introduced Rapid Polymyxin tests for *Acinetobacter spp.* and *Pseudomonas*

spp. in October 2018 [44, 45]. Both tests use the same principle as the Rapid Polymyxin NP test

as they rely on the colorimetric detection of bacterial metabolism in the presence of a defined

concentration of colistin [45]. A positive result by Rapid Polymyxin *Acinetobacter* was read by

a change in colour of pH indicator, phenol red (red to orange or yellow), whereas the Rapid

Polymyxin *Pseudomonas* uses bromocresol purple pH indicator (green to violet) [45]. While the

Rapid Polymyxin *Pseudomonas* test recorded 100% correlation with the reference BMD, the

Rapid Polymyxin *Acinetobacter* had discrepancies, recording no errors in one study and eight

errors (three VMEs and five MEs) in another, respectively [44, 45].

The Rapid Resapolymyxin NP test was also developed to detect colistin resistance in all

colistin-resistant Gram-negative bacteria including *A. baumannii*, *Enterobacteriaceae* and *P.*

aeruginosa [46]. The test was carried out by inoculating a standardized bacterial suspension in

Mueller-Hinton broth containing a final concentration of 3.75 mg/L of colistin sulfate [46]. A

10% concentration of resazurin PrestoBlue was added after three hours of incubating the

medium and the results were read over a period of one hour after the addition of resazurin

PrestoBlue [46]. The detection of colistin resistance is based on the reduction of blue resazurin

to pink resorufin by metabolically active cells in the presence of a defined concentration of

colistin [46].

The evaluation of Rapid ResaPolymyxin *Acinetobacter*/*Pseudomonas* NP test showed reliable

detection of colistin resistance in non-fermenters and 100 % accuracy in *Enterobacteriaceae*

[46-48]. Although the slow growing nature of *Pseudomonas* spp. resulted in an hour delay before

a change in colour could be observed, *Enterobacteriaceae* and *Acinetobacter* spp. were

detected in less than four hours [37, 47]. The Rapid ResaPolymyxin NP test makes up for the

limitations of Rapid Polymyxin NP test in testing colistin resistance in Gram-negative bacilli

(including non-fermenters) regardless of the mechanism of resistance [46, 47]. Therefore, this test

is more suitable for general categorization of colistin resistant and colistin susceptible isolates

than the Rapid Polymyxin NP test [47].

EDTA-based

The MCR catalytic domain, phosphoethanolamine (PEtN) transferase, has been shown to be

analogous to zinc metalloproteins, where zinc deficiency reduces colistin MICs in MCR-

producing *E. coli* [49-51]. The combined disk test, colistin MIC reduction test, modified rapid

polymyxin NP test and alteration zeta potential, are four tests that are based on the inhibition

of MCR activity by EDTA and have been strategically developed to detect *mcr*-genes [49].

The combined disk test (CDT) uses 10µL of 100mM EDTA solution, which is impregnated

into one of two 10µg colistin disks [49]. The two colistin disks are placed on Mueller Hinton

agar plates swabbed with 0.5 McFarland bacterial suspension and incubated for 18 to 24h at

37□ [49]. Results are read as positive if there is an increase of ≥3mm in inhibition zone around

the colistin disk containing EDTA as compared to the colistin disk without EDTA [49]. This

method has recorded a sensitivity and specificity of 96.7% and 89.6% respectively by the

developer [49]. However, a latter study recorded less promising results, with a sensitivity and

specificity of 12% and 65.2% respectively, Hence, further evaluation studies must be

conducted on this test as the currently results seem unreliable [51].

A pre-diffusion method of the CDT was evaluated, where two colistin disks were placed and

allowed to diffuse for 2 hours on MH agar [50]. The disks were removed and the plates were left

at room temperature for 18-24h after which two disks containing 1µmol of EDTA were

strategically placed (one exactly where the colistin disk had been placed and the other, at least

30 mm away) [50]. Diameters of inhibition zones were measured after 18h of incubation and

colistin-resistant MCR-positive isolates demonstrated a ≥5mm increase in inhibition zone

around the disks [50]. The pre-diffusion method, using a cut-off value of ≥5mm, improved the

CDT to 100% accuracy for *mcr* detection [50]; it has however not been extensively evaluated to

ensure reproducibility and seems more complicated, laborious and time-consuming than the

CDT.

Colistin MIC reduction (CMR) test was performed by broth microdilution using MH broth

without cation supplementation but with 80µg/mL EDTA solution instead [49]. It was

considered that cation supplementation with calcium and magnesium would impair the

inhibitory activity of EDTA; moreover, calcium could favour the activity of putative PEtN

transferases in *E.coli* [49]. Even so, this method did not efficiently detect *mcr*-producers among

Enterobacteriaceae isolates although different concentrations of EDTA were used [49, 52].

Innovatively, EDTA has been added to the Rapid Polymyxin NP test in what is termed the

modified rapid polymyxin NP (MPNP) test to enable it to identify MCR. MPNP is the Rapid

Polymyxin NP test with the addition of two wells filled with colistin-free solution and

colistin-containing solution, both with 80µg/mL EDTA [49]. Results were read as positive for

the production of MCR-1 PEtN transferase if there was no change in colour of red phenol in

the presence of colistin and EDTA [49]. The presence of EDTA in the MPNP test successfully

detected MCR-1 positive colistin-resistant *E.coli* isolates as demonstrated by the sensitivity

and specificity of 96.7% and 100% respectively [49].

Finally, addition of EDTA results in an alteration of zeta potential of membrane charge,

which is measured to determine MCR expression [49]. Particle size and zeta potential of

colistin-susceptible and colistin-resistant bacterial cells grown in MH broth with or without

80µg/ml EDTA was measured using a ZETAPALS zeta potential analyser [49]. Colistin-

susceptible and colistin-resistant MCR-1 Enterobacteriaceae demonstrated zeta potential

values between -21.54 and -44.21 mV while colistin-resistant MCR-1 positive were ≤-20 mV

(-4.20 to -19.34 mV) [49]. In the presence of EDTA an alteration of zeta potential ranging from

-21.13 to -40.81 was observed in colistin-resistant MCR-1 positive *E.coli* isolates [49]. A zeta

potential ratio $(R_{zp} = ZP_{+EDTA} / ZP_{-EDTA})$ was calculated for all isolates and a cutoff value of

R_{zp} ≥2.5 as a criterion for the presumed detection of MCR-1 positive *E.coli* isolates was

established [49]. Alteration of zeta potential yielded a sensitivity and specificity of 95.1% and

100% respectively, however EDTA had no inhibitory effect on *mcr*-1 positive K. pneumoniae

isolates [49].

Colistin broth-disk elution (CBDE)

Simner *et al* (2019) developed the colistin broth-disk elution method (CBDE). This method

was performed on a collection of *Enterobacteriaceae*, *P. aeruginosa* and *A. baumannii*

isolates to which four tubes were each assigned [53]. The four tubes contained 10mL of CA-

MHB with colistin disks to yield final concentrations of 0,1,2 and 4 µg/mL, respectively [53].

The tubes were incubated for 30 minutes at room temperature allowing colistin to dissolve

into the broth, after which 50µL aliquot of standardized bacterial suspensions were added to

each tube [53]. Colistin MIC values were visually read after 16-20h of incubation at 35□ in

ambient air [53]. In this study, CBDE was compared to the reference BMD and Sensititre,

where the CBDE had a CA and EA of 98% and 99%, respectively as compared to both BMD

methods [53]. Three *mcr*-1-producing *E. coli* isolates resulted in a VME rate of 8% due to one

dilution difference by CBDE and BMD; however, no errors were observed when CBDE was

compared to BMAD [53].

Three studies have evaluated the modified version of the CBDE [54-56]. Bell *et al* (2019) was the

first to describe a modified CBDE method. The method was performed as previously

described, however 1mM EDTA was used by adding 20µL of 0.5M EDTA to each tube

containing CA-MHB and 10µg colistin disks [55, 56]. Fenwick *et al* (2020) added a fifth tube to
generate CBDE+EDTA with colistin concentrations of 0, 0.4, 1, 2 and 4 µg/mL, respectively.
The CBDE+EDTA method has shown overall sensitivity and specificity of 100% and 94.3-
95.8% respectively, for screening the presence of MCR in *Enterobacteriaceae* and *P.*
aeruginosa [55, 56].

Dalmolin *et al* (2019) evaluated the CBDE method using final volumes of 1mL and 200µL in
colistin broth microelution (CBM) and the microelution test (MPT), respectively. The two
methods were evaluated on 68 *Enterobacteriaceae*, nine *A. baumannii* and eight *P.*
aeruginosa isolates from human and animal samples [54]. Both CBDE methods were performed
as previously described, however the CBDE mixture was fractioned in 1mL tubes for the
CMB test and 200µL in microtiter plates for MPT test [54]. Additionally, this study evaluated
the colistin susceptibility test tube (CSTT), which was performed using one tube with 5mL
CA-MHB and 10 µg colistin disk to yield a final concentration of 2 µg/mL [54]. All three
methods presented unsatisfactory MEs and VMEs, particularly performing poorly for non-
fermenters [54].

Lateral Flow-Immunoassay

Monoclonal antibodies (MA) were used to develop lateral flow assays for the detection of
MCR-1 producing bacterial isolates [57]. In this study, 177 and 121 *Enterobacteriaceae* isolates
from human and animal samples were obtained, respectively [57]. Bacterial colonies were
isolated from agar plates, suspended in extraction buffer and dispensed on the MA-containing
cassette where they were allowed to migrate for 15 minutes [57]. All MCR-1 producing isolates
were detected accordingly as shown by a pink band on the test line and control line of the
assays. Furthermore, this test was able to detect MCR-2-producing isolates [57]. The same assay
is currently marketed as the NG-Test MCR-1 by NG Biotech in France [55]. Initially, the
evaluation of the NG-Test MCR-1 resulted in eight false-positive results that were ultimately
resolved to negatives apart from one isolate that was found to be an MCR-2 producer [55].
While the detection of the MCR-2 product by the NG Test MCR-1 further confirms the
results of the developers, cross-reactivity with MCR-2 limits the accuracy of the assay for the
MCR-1 producers. [55]. Nonetheless, the lateral flow immunoassays were found to be highly
sensitive, easy to use and cost-effective for the detection of MCR-1/-2 [55, 57].

Agar-based Medium

Superpolymyxin

Superpolymyxin is a selective medium for polymyxin-resistant Gram-negative bacteria that is

based on eosin methylene blue agar [58]. The medium was developed with the optimal colistin

concentration of 3.5 µg/mL, 10 µg/mL of daptomycin (to inhibit potential growth of Gram-

positive strains) and 5 µg/mL of amphotericin B as an antifungal. Nordmann *et al* (2016)

designed Superpolymyxin for screening intrinsic and acquired polymyxin-resistant Gram-

negative bacteria as previous screening media containing deoxycholic acids and a high

concentration of colistin inhibited the growth of strains with acquired resistance of low MIC

values (Table 1).

The use of eosin Y and methylene blue dyes helped distinguish lactose-fermenters (dark

brown to purple) from non-fermenters (colourless) [11]. This medium distinguishes lactose-

fermenting *E. coli* (metallic green sheen) from other *Enterobacteriaceae*, including non-

fermenting *E. coli* (dark brown to purple) [59, 60]. However, studies evaluating other selective

media against Superpolymyxin have shown a weaker detection of non-fermenters by

Superpolymyxin [58, 61].

This medium was able to detect colistin-resistant Gram-negative bacteria directly from

bacterial culture and clinical samples (*i.e.* rectal swabs and stool samples) with high

sensitivity and specificity [62]. Although, there is a low sensitivity for *Enterobacter* spp. which

may be due to hetero-resistant phenotypes, therefore a high inoculum was suggested [62, 63]

CHROMagar™ COL-APSE.

CHROMagar™ COL-APSE by CHROMagar (Paris, France) is the first selective medium

designed to detect and differentiate all *Acinetobacter spp.*, *Enterobacteriaceae*, *Pseudomonas*

spp. and *Stenotrophomonas spp.* [58]. The agar plates were prepared in-house using a

dehydrated CHROMagar base medium and supplements (S1 and X192) containing colistin

sulfate and oxazolidonones to enhance the growth of colistin-resistant Gram-negative bacteria

and inhibit that of Gram-positive bacteria [58]. Swarming by *Proteus spp.* was inhibited by

adding p-nitrophenyl glycerol to the medium preparation, which did not disrupt the medium's

performance. This makes CHROMagar™ COL-APSE suitable for screening mixed

specimens [58].

The accuracy in detecting and differentiating colistin-resistant Gram-negative species was

evaluated by Osei Sekyere *et al.* (2020), where the morphological appearance of the detected

strains was as described by the manufacturer. Moreover, three studies that have evaluated

CHROMagar™ COL-APSE were in agreement that the medium had a high sensitivity in

detecting isolates harbouring *mcr* genes [5, 58, 64]. However, there was significant difference in
the sensitivity and specificity recorded by Abdul Momin *et al* (2017) and Osei Sekyere *et al*
(2020) (Table 1). The poor performance in the recent study could be due to the use of cultured
bacteria instead of using serial dilutions in broth (Table 1)[5, 58].

ChromID® Colistin R

ChromID® Colistin R is a chromogenic selective medium that is primarily used for isolating
colistin-resistant *Enterobacteriaceae* from clinical stools and rectal swab samples [65]. Like
CHROMagar™ COL-APSE, the medium can differentiate between bacterial species based on
morphological appearance of bacterial colonies *i.e. E. coli* (pink to burgundy), *Klebsiella spp.,*
Enterobacter spp., Serratia (blue to green), *Salmonella spp.* (white or colourless) and
Proteeae tribe (beige-brown) (Table 1) [59, 65].

The assessment of ChromID® Colistin R and Superpolymyxin on stool and rectal swab
samples resulted in an overall better performance by ChromID® Colistin R [59]. The LOD of
this medium being at least one log lower in 69.2% of the isolates detected on both media,
whereas Superpolymyxin only had a better LOD for 7.7% isolates [59]. Nonetheless,
Superpolymyxin could be directly inoculated with stool or rectal swab samples without a 4-5
hours enrichment step required by ChromID® Colistin R; and the final sensitivity (84.9 to
86.8%) recorded for this medium was achieved after extending the turnaround time from 24
hours to 48 hours (which also allowed for the detection of an *mcr*-1 producing *E. coli* isolate)
[59].

A study by Thiry *et al* (2019) evaluated this medium against CHROMagar™ COL-APSE on
158 colistin-resistant bovine *E. coli* isolates. Half (48/96) of the isolates considered to be
intermediate to the disk diffusion test had MIC > 2 and were able to grow on both media, with
two more isolates growing on ChromID® Colistin R alone [64]. Although both media could
support the growth of (21/22) *mcr*-1-positive and (13/14) *mcr*-2-positive isolates,
CHROMagar™ COL-APSE has an advantage over ChromID® Colistin R as it is not limited
to (isolating and differentiating) *Enterobacteriaceae* (Table 1) [58, 64]

LBJMR medium

Lucie-Bardet-Jean-Marc-Rolain (LBJMR), a polyvalent medium based on Purple agar, has
been designed for the isolation of colistin-resistant Gram-negative bacteria as well as
vancomycin-resistant Gram-positive bacteria [61]. The medium was developed by adding
glucose (7.5 g/L), colistin sulfate (4 µg/mL) and vancomycin (50 µg/mL) to 31g/L Purple agar

base [61]. One hundred and forty three bacterial isolates, including colistin-resistant *Enterobacteriaceae* and non-fermentative Gram-negative bacilli were used to evaluate this medium, where the specificity and sensitivity were 100% [61]. The medium was further evaluated on 56 *mcr*-1 positive and 10 *mcr*-1-negative chicken and human stool samples as well as two clinical rectal swabs [61]. The study found that the LBJMR could detect *mcr*-1 positive isolates with high sensitivity, particularly showing a higher sensitivity for colistin-resistant non-fermenters than Superpolymyxin [61]. Furthermore, the LBJMR medium does not contain daptomycin and amphotericin B, which are used in some of the agar-based media *i.e.* Superpolymyxin to inhibit the growth of Gram-positive bacteria, including vancomycin-resistant *Enterococcus* [61]

Colistin-MAC test

The colistin-MAC test was designed to detect *mcr*-genes on the basis of colistin MIC reduction by a fixed concentration (900µg/mL) of dipicolinic acid (DPA) [66]. The test was carried out in 96-well microtitre plates using CA-MHB with DPA stock solution prepared in dimethyl sulfoxide [66]. Coppi *et al* (2018) established a cutoff value of ≥8-fold colistin MIC reduction in the presence of DPA for the presumptive identification of *mcr*-positive isolates. However, a latter study used a different cutoff value, where ≥3 twofold MIC reduction in the presence of DPA indicated a positive result [52]. The Colistin MAC test was found to perform well for *E. coli* isolates and ineffective in detecting *mcr*-genes of *K. pneumoniae* and *Salmonella* spp. [5, 52, 66]. The lack of inhibitory effect of DPA in *K. pneumoniae* isolates can be attributed to a decrease in DPA permeability or the existence of other mechanisms of colistin resistance in these strains [66]. Moreover, the addition of DPA resulted in reductions in MICs and increments of some isolates, although these adjustments did not affect the accurate classification of the isolates [5].

Luria-Bertani (LB) medium

A selective medium for detecting colistin-resistant *Enterobacteriaceae* (including those with *mcr*-1 genes) in spiked stools was evaluated [67]. The medium was developed by adding 4mg/mL colistin, 10mg/mL vancomycin, and 5mg/mL amphotericin B to agar medium made with 25g of Luria-Bertani (LB) powder [67]. Nine *Enterobacteriaceae* isolates were collected, seven of which carried *mcr*-1 genes [67]. Each of the nine isolates were spiked into donated faecal samples and serially diluted to final concentrations of 10^2 or 10^3 CFU/mL; 0.5 mL of the stool mixture were spiked in 4.5 mL of *Enterobacteriaceae* enrichment (EE) broth and

477 incubated for 24h at approximately 35▯ [67]. After which 10 µL of the spiked EE broth was
478 inoculated onto the LB medium and incubated at 35▯±2▯ for 48 hours in ambient air [67]. The
479 selective LB medium demonstrated a sensitivity of 100% (Table 1)[67].

3. Molecular tests

Amplidiag Carba-R+ MCR assay

The Amplidiag Carba-R+ MCR assay is a multiplex nucleic acid-based test developed for
detecting carbapenemase and *mcr*-1/-2 genes from rectal swabs and bacterial culture [68]. The
assay was performed retrospectively and prospectively on 215 Gram-negative bacilli and 51
Enterobacteriaceae isolates, respectively [68]. The Amplidiag Carba-R+MCR assay did not
detect one GES carbapenemase-producing *P. aeruginosa*. However, all *mcr*-1 and *mcr*-2-
producing isolates were accurately detected (Table 1)[68].

Additionally, this assay was performed on DNA extracted from 100 rectal swabs, including
40 carbapenemase-positive samples where the sensitivity was 92.5% [68]. Two NDM and one
OXA-48 producers were not detected due to a low concentration of bacteria, therefore the
samples were subjected to an overnight enrichment in brain heart infusion with 0.5 µg/mL
ertapenem [68]. The enrichment step allowed for the detection of two of the three samples that
were not previously detected, one of the samples was an OXA-48 producing *E. cloacae* with
an AcOXA (*Acinetobacter* oxacillinases with carbapenemase activity) gene [68]. The overall
performance of this assay was acceptable, demonstrating sensitivity and specificity ranging
from 92-100% and 86-100% respectively [68]. Moreover, the assay can be performed on
cultured bacteria as well as DNA extracted from rectal swabs in no more than three hours
(Table 1)[68].

AusDiagnostic MT CRE EU assay

AusDiagnostics MT CRE EU assay is a two-step nested multiplex-tandem PCR (MT-PCR)
assay by AusDiagnostics (Chesham, UK) [69]. One study evaluated the AusDiagnostics MT
CRE EU assay for detecting carbapenemase genes and colistin resistance *mcr*-1 and *mcr*-2
genes [69]. A collection of *Enterobacteriaceae*, *Pseudomonas spp.* and *Acinetobacter spp.*,
including carbapenemase or *mcr*-1/-2 producers, were used to evaluate the performance of
this assay [69]. The assay was performed by suspending two to three bacterial colonies grown
overnight on Columbia blood or cystine lactose electrolyte agar in tubes with a sample buffer
[69]. The tubes were loaded onto the AusDiagnostics MT processor platform for template
extraction and the first round of PCR [69]. Lastly, the nested RT-PCR was performed by loading

a 384-PCR plate containing the reaction mix onto the AusDiagnostic MT analyser [69]. The

results were automatically read using the AusDiagnostics MT assay software. The assay

failed to detect four out of the 22 *mcr* genes; however, the *mcr* genes were also not detected

by the reference PCR [69].

Retrospective and prospective evaluation of the assay resulted in eight and 18 false-positive

results, respectively [69]. Therefore, yielding an overall sensitivity and specificity of 95.5% and

99.8%, respectively, which were improved to yield 100% accuracy following repeats of the

assay [69]. The AusDagnostic MT CRE EU assay was able to detect *mcr*-1/-2 genes as well as

carbapenemase genes with minimal hands-on time (Table 1)[69].

Loop-mediated isothermal amplification (LAMP)

Loop-mediated isothermal amplification (LAMP) is a nucleic acid amplification method that

allows autocycling strand displacement DNA synthesis at constant temperature using Bst

DNA polymerase [70]. The use of LAMP for detecting *mcr*-1 gene was first described by two

studies [71, 72]. Zou *et al* (2017) established a LAMP assay for the detection of *mcr*-1 gene from

cultured bacteria and spiked human stools. In this study, the LAMP assay was performed in

25 µL reaction mixtures that contained 20mM Tris-HCl, 10mM KCl, 10mM $(NH4)_2 SO_4$,

8mM $MgSO_4$, 0.8 M betaine, 0.1% Tween-20, 1.4 mM of each DNTP and 8U Bst DNA

polymerase [72]. Each reaction mixture had specified amounts of forward and backward inner

primers, outer forward and backward primers, loop primers as well as the appropriate amount

of DNA template [72]. The mixture was covered with 25µL wax and incubated in a dry bath

incubator for 60 minutes at constant temperature, the amplification products were read

visually and by turbidimetry [72]. Visual detection was by colour change of a fluorescent metal

indicator, a positive reaction was demonstrated by the formation of a magnesium

pyrophosphate precipitate which changed the reaction mixture from orange to green during

amplification [72].

Imirzalioglu *et al* (2017) evaluated the eazyplex SuperBug *mcr*-1 kit (Amplex Biosystems

GmbH, Giessen, Germany) for the rapid detection of *mcr*-1 gene. Colistin-resistant isolates

were grown on LB media containing colistin sulfate to prevent the loss of *mcr*-1 encoding

plasmid, while colistin-susceptible isolates were grown in the absence of colistin [71]. The *mcr*-

1-detecting LAMP assays were found to detect *mcr*-1 genes accurately, and the in-house

LAMP assay was stated to be more sensitive than conventional PCR assays (Table 1) [71, 72].

However, *mcr*-1 detecting LAMP assays cannot detect other potential target genes [70-72].
Therefore, Zhong *et al* (2019) developed a restriction endonuclease-based multi-LAMP for
the detection of multiple *mcr*-genes. Two separate LAMP systems were established, a double-
LAMP (*mcr*-2 and *mcr*-5) and triple-LAMP system (*mcr*-1, *mcr*-3 and *mcr*-4) which were
performed in 25 µL reaction mixtures [70]. The 25 µL reaction mixture consisted of 12.5 µL
LAMP-reaction mix, 1 µL Bst 2.0 polymerase, 1.25 µL primer mix, 8.25 µL nuclease-free
water, and 2 µL DNA lysate [70]. Amplification products were detected visually by change in
colour of SYBR Green I, which changed from yellow to orange for a positive reaction [70].
Amplification products were also stained with GoldView TM and analysed by electrophoresis
on 2% agarose gel [72]. Multiplex detection of *mcr*-1 to *mcr*-5 genes was established through
restriction digestion of the LAMP products based on band numbers and fragment lengths
using Hind restriction enzyme [70].

LAMP assays have advantages over conventional PCR in that LAMP is more sensitive, has a
shorter processing time of <60 minutes, is relatively easier to run, and multiplex detections
can be conducted in the same detection system (Table 1)[70-72].

Matrix-assisted laser desorption ionization-time of flight mass spectrometer (MALDI-TOF
MS)

MALDI-TOF MS is a technique that is based on the production of mass spectra from whole
cells and their contrast to a reference spectra [73]. This method is widely used for species
identification of pathogens in clinical microbiology laboratories. Giordano and Barnini were
the first to evaluate the possibility of detecting colistin resistance using MADI-TOF MS [73]. In
this study, 139 *K. pneumoniae* isolates from clinical samples were collected, from which
protein was extracted for identification using MALDI-TOF MS (Bruker Daltonik GmbH,
Bremen, Germany)[73]. The MICs of the isolates were determined by Sensititre using a range
of antibiotics including colistin [73]. First, a training set for mass peak analysis was established
using 50/139 of the *K. pneumoniae* isolates from which 400 spectra were obtained and used
for data base entry (Main Spectrum Profile) as well as to classify algorithm models [73]. Finally,
the remaining 89/139 isolates were used to conduct the test; 712 spectra were collected from
this set [73]. However, from the 712 spectra, 158 were excluded as they constituted flat-line
spectra or outliers, demonstrating identification score below 2.3 [73]. Based on the mass signals
and intensities of the bacterial protein samples, two dimensional peak distribution classified
the training set spectra into two main groups viz., colR-KPn (colistin-resistant *K.*
pneumoniae) and colS-KPn (colistin-susceptible *K. pneumoniae*) isolates [73].

The newly created database entry consisted of using *MALDI Biotyper RTC* and *MALDI*

Biotyper v3.0 to identify *K. pneumoniae* isolates and for the automatic detection of colistin

resistance, respectively [73]. The automatic classification of the test set resulted in the correct

classification of 71% colR-KPn and 40% colS-KPn [73]. Furthermore, different algorithm

models were tested using ClinProTools v3.0 (Bruker Daltonics). The three algorithms tested

included the Genetic Algorithm (GA), Supervised Neural Network (SNN) and Quick

Classifier (QC)[73]. However the tested algorithms either had good recognition capability and

cross validation but poor classification of colistin resistance or poor recognition capability and

acceptable classification of colistin resistance [73]. Although the GA seemed more promising as

it was shown to be better suited for biological samples [73].Therefore, a reliable classification

model was created by combining the most relevant peaks detected from the GA algorithm[73].

The resulting peak combination of 4507.28/5142.84 Da from GA demonstrated a sensitivity

and specificity of 78% and 89%, respectively [73].

Three studies have evaluated MALDIxin, a MALDI-TOF-based assay, [7, 22, 74]. The

MALDIxin test was developed to detect pETN modification in lipid A directly from bacterial

colonies in less than 15 min [22, 74]. Dortet *et al* (2018) evaluated MALDIxin on *A. baumannii*

isolates, where the mass spectrum in colistin-susceptible isolates was characterised by two

sets of peaks at the centre of *m/z* 1728.1 and *m/z* 1910.3. The peaks were assigned to bis-

phosphorylated hexa-acyl and bis-phosphorylated hepta-acyl lipid A that had 12 to 14 carbons

making up the acyl chain, respectively [74]. The mass spectrum in colistin-resistant isolates was

observed by two set of peaks at the center of *m/z* 1935.3 and *m/z* 2033.3, showing m/z+25 and

m/z+123 shifts of mass unit of the bis-phosphorylated hepta-acyl lipid A at *m/z* 1910.3 [74]. The

peaks observed at *m/z* 2033.3 and *m/z* 1935.3 were assigned to pETN-modified-bis-

phosphorylated hepta-acyl and pETN-modified-mono-phosphorylated hepta-acyl lipid A,

respectively with an acyl chain of 12 carbons in length [74]. The peaks (*m/z* 2033.3 and *m/z*

1935.3) associated with pETN-modified lipid A were observed in all colistin-resistant isolates

and were not observed in any of the colistin-susceptible isolates [74].

Furniss *et al* (2019) and Dortet *et al* (2020) described the optimization of the MALDIxin test

for detecting colistin resistance in clinical *E. coli* and *K. pneumoniae* isolates, respectively.

Furniss *et al* (2019) optimized the MALDIxin test by adopting the low-resolution linear mode

used by the MALDI Biotyper Sirius system. The optimization was achieved by adding a mild-

acid hydrolysis step, which is required for analysis of clinical isolates in negative ion mode [7].

The mild-acid hydrolysis step was performed by resuspending a single bacterial colony grown

on MH agar for 18-24h in 200µL distilled water [7, 22], after which 50-100 µL of 2% acetic acid

was added to double-distilled water containing bacterial suspension and heated for 5-15 min

at 98-100□ [7, 22]. For MALDI-TOF analysis, Furniss *et al* (2019) used a MALDI Biotyper

Sirius system, whereas Dortet *et al* (2020) used a 4800 Proteonic Analyzer. The optimization

of the MALDIxin allowed for the identification of L-Ara4N- and pETN-modified lipid A in

E. coli and *K. pneumoniae* isolates. Moreover, the optimized methods were able to distinguish

between chromosome-encoded and MCR-mediated colistin resistance [7, 22].

Microarray

A commercial CT103XL microarray system that allows for the simultaneous detection of

mcr-1/-2 and β-lactamase genes was evaluated [75]. The study was conducted on 106

Enterobacteriaceae isolates including *mcr*-1 and *mcr*-2 positive strains, as well as

carbapenemase and extended-spectrum β-lactamase (ESBL)-producing strains [75]. The

CT103XL microarray, which uses a multiplex ligation detection reaction, was performed

following bacterial DNA extraction from bacterial cultures [75]. The commercial CT103XL

microarray was confirmed to simultaneously detect *mcr*-1/-2 and β-lactamase genes with

accuracy, although it failed to detect *mcr*-3, which shares 45% and 47% identity to *mcr*-1 and

mcr-2, respectively [75].

Multiple Cross Displacement Amplification coupled with Gold Nanoparticles-Based lateral

Flow Biosensor

A multiple cross displacement amplification (MCDA) method, coupled with gold

nanoparticles-based lateral flow biosensor (LFB) assay for detecting *mcr*-1 gene was

established [76]. The MCDA reaction was performed on extracted DNA from 59 bacterial

isolates, where each 25 µL reaction consisted of 12.5 µL reaction buffer, 1µL Bst DNA

polymerase 2.0, 1µL colorimetric indicator, 1.6 µM of each cross primers, 0.4 µM of each

displacement primers, 0.4 µM amplification primers and 1 µL DNA template [76]. The MCDA

reaction systems were then subjected to isothermal temperature (63□) for 40min, after which

the amplification products were analysed using 2% agarose gel electrophoresis, colorimetric

indicator and LFB [76]. For *mcr*-1 detection by LFB, 0.2 µL of the amplicons was added to the

well of the sample pad, followed by the addition of three drops of running buffer (1% Tween

20 and 0.01 mol/L phosphate-buffered saline) [76]. The results were visually read after 1-2 min;

a positive result was demonstrated by two red bands, one at the test-line and the other at the

control line [76]. The results were positive for all *mcr*-1 positive isolates and negative for all

non-*mcr*-1 isolates [76].

The MCDA-LFB assay was further applied to stool samples spiked with 100 μL dilutions of

bacterial strains [76]. The resulting detection limit was 600fg of *mcr*-1 plasmid DNA per

microliter in bacterial culture and 4.5×10^3 CFU/mL in the spiked faecal samples [76]. The

MCDA-LFB has demonstrated the same sensitivity as the *mcr*-1 LAMP, which is more

sensitive than the conventional PCR. Further, the MCDA-LFB has demonstrated a shorter

reaction time [72, 76].

Recombinase polymerase amplification (RPA)

The rapid detection of *mcr*-1, using a RPA, has been described [77]. RPA is a novel isothermal

amplification method, which can be performed in no more than 30 minutes at body

temperature without the need for thermal cycling instruments [77]. This study used basic RPA

(B-RPA) and RPA with lateral flow (LF-RPA) on 23 genomic DNA extracted from 20 *mcr*-1

positive and three *mcr*-1-negative *Enterobacteriaceae* [77]. The primers for the B-RPA assay

were designed by Clone Manager 8 and validated on three *mcr*-1 positive and one *mcr*-1

negative DNA samples [77]. The B-RPA was based on the TwistAmp Basic kit reaction system,

which was incubated at room temperature for 30 min, after which the amplicons were

extracted by phenol/chloroform solution or purified using an amplicon purification kit [77]. The

LF-RPA reaction required primers and a probe, which were labelled with biotin and

fluorescence [77]. The LF-RPA was based on TwistAmp Nfo kit reaction system, which was

incubated as described for the B-RPA [77]. The amplification products for the LF-RPA were

diluted at 1:50 with running buffer, after which a downstream operation was carried out [77].

The results for the B-RPA assay were read by agarose gel electrophoresis, whereas the results

for the LF-RPA assay were visually read using Hybridetect 2T dipsticks [77]. A positive *mcr*-1

detection by LF-RPA was demonstrated by two purple bands at the test line and the quality

control line [77]. Both assays detected the *mcr*-1 positive and *mcr*-1 negative DNA samples

accordingly; therefore, both assays are equally suitable for detecting *mcr*-1 genes [77].

Conventional & real-time PCR and Whole genome sequencing

The presence or absence of *mcr*-genes is determinable by PCR assays and whole genome

sequencing (WGS) as standard [71]. While WGS is able to characterize the mechanism of

resistance and determine the molecular evolutionary trajectory of colistin-resistant isolates,

PCR is only able to characterize resistance genes [78, 79] However, WGS technology is limited

in settings that lack adequate resources and therefore PCR assays are widely adopted for

detecting *mcr*-genes [70, 80].

Nijhuis *et al* (2016) were the first to design a real-time PCR assay for detecting *mcr*-1 from

clinical isolates using self-designed primers and probes. The assay was validated on 26 *mcr*-1

positive *E. coli* isolates, where the presence of *mcr*-1 was detected in all 26 isolates [81].

Additionally, the assay was evaluated on spiked stool samples and the efficiency of the PCR

was 102.6% and the LOD was 3-30 cfu/reaction [81]. However, *mcr*-1 genes were not detected

in other colistin-resistant strains *i.e. Klebsiella, Enterobacter, Pseudomonas, Acinetobacter*

etc.

A multiplex PCR (M-PCR) assay for the simultaneous detection of *mcr*-1 and carbapenem-

resistant genes, $bla_{KPC}, bla_{NDM}, bla_{IMP}, bla_{OXA-48-like}$, was described [82]. The assay was

validated on reference strains including *E. coli* A434-59, which contains *mcr*-1 and bla_{NDM-1}

[82]. Evaluation of the M-PCR on 127 carbapenem-resistant, eight *mcr*-1-positive and 62

carbapenem-susceptible *Enterobacteriaceae* found the assay to be 100% sensitive and

specific [82].

Additionally, three studies designed M-PCR assays to detect *mcr*-1 to *mcr*-5 genes [4, 80, 83].

The assay designed by Rebelo *et al* (2018) allowed for the simultaneous detection of *mcr*

genes and their variants in bovine and porcine isolates [80]. This study did not use internal

amplification controls as they were incompatible with DreamTaq Green PCR Master Mix [80].

The master mix contains DNA polymerase synthesized in *E. coli* and thus would produce

amplicons if 16S rRNA primers are used [80].

However, Lescat *et al* (2018) designed a more rapid (<2 hours) M-PCR assay that was

compatible with internal controls. Recently, Joussett *et al* (2019) designed and evaluated an

M-PCR assay on 50 *E. coli*, 41 *K. pneumoniae* and 12 *Salmonella enterica* isolates (from

which a total of 40 were MCR-producers), which was 100% accurate in detecting *mcr*-

positive isolates. The assay was additionally performed on 82 *Aeromonas* spp. and 10

Shewanella spp. that were previously described as potential originators of *mcr*-3 and *mcr*-4,

respectively [4]. None of the *Aeromonas* spp. were *mcr*-positive, although two *Shewanella* spp.,

S. bicestrii JAB-1 strain and *S. woody* S539 with MICs of 0.25mg/L and <0.12 mg/L

respectively, were *mcr*-4 positive [4]. However, cloning *S. bicestrii* JAB-1 genes into *E. coli*

TOP10 resulted in an *mcr*-4 positive outcome by the PCR assay with a colistin MIC of 4

mg/L [4].

Borowiak *et al* (2020) described the detection of *mcr*-1 to *mcr*-9 in colistin resistant

Salmonella enterica isolates using an M-PCR (*mcr*-1 to *mcr*-5) designed by Rebello *et al*

(2018) and a newly designed M-PCR assay (*mcr*-6 to *mcr*-9). The assay was performed on

407 colistin-resistant *S. enterica* isolates from animals, animal feed, food and the environment

[84]. Two hundred and fifty-four of the isolates had *mcr* genes. Moreover, the assay detected

mcr-9 in isolates carrying *mcr*-1 [84]. However, two separate frameshift mutations of *mcr*-9

were shown to have occurred in the respective isolates as demonstrated by WGS analysis; the

mutations are believed to have contributed to non-functional MCR-9 proteins [84].

Two studies have described methods for broth enrichment of colistin-resistant *E. coli*

followed by real-time PCR to detect *mcr*-genes [85, 86]. Chalmers *et al* (2018) were the first to

describe a SYBR Green-based real-time PCR method for *mcr*-1 and *mcr*-2 following

enrichment with *E. coli* (EC) broth containing colistin (1µg/mL). All the porcine faecal and

chicken caecal samples were screened by real-time PCR after 16h of culture in EC broth [86].

However, none of the *mcr*-1 and *mcr*-2 genes were detected by PCR in any of the samples

after 16h of enrichment [86]. As well, the method described by Chandler *et al* (2020) for

detecting *mcr*-1 included enrichment using EC broth containing colistin (1µg/mL) and

vancomycin (8µg/mL). The method was evaluated on 100 feral swine faecal samples, which

were inoculated with one of five different *mcr*-1 positive *E. coli* strains [85]. The bacteria was

inoculated at concentrations ranging between 0.1-9.99 CFU/g, 10-49.99 CFU/g, 50-99 CFU/g,

100-149 CFU/g and 200-2,200 CFU/g from which *mcr*-1 was detected with 32%, 72%, 88%,

95%, and 98% accuracy by real-time PCR, respectively [85].

Four SYBR Green-based real-time PCR assays have been developed for *mcr*-1, *mcr*-2, *mcr*-3,

mcr-4 and *mcr*-5 detection [87-90]. Bontron *et al* (2016) designed a SYBR Green-based real-time

PCR assay for detection of *mcr*-1 from cultured bacteria and stools. The assay was validated

on 20 *Enterobacteriaceae,* where it was found to accurately detect the presence or absence of

mcr-1 at a LOD of 10^2 cultured bacteria [88]. Furthermore, Dona *et al* (2017) described a

SYBR Green real-time PCR assay to also detect *mcr*-1 from human faecal samples. However,

in this study, 20µg of the stool samples were enriched overnight in 10mL LB broth containing

2µg/mL colistin and plated on four selective agar plates prior to DNA extraction [90]. The real-

time PCR accurately identified *mcr*-1 harbouring *E. coli* isolates with an LOD of 10^1 and

PCR efficiency of ca. 106% [90].

Li *et al* (2017) also designed a multiplex SYBR Green-based real-time PCR assay for *mcr*-1,

mcr-2 and *mcr*-3 detection. The assay was validated on 25 isolates including *mcr*-1 positive

and *mcr*-3 positive strains; the *mcr*-2 gene was synthesized in the study due to a lack of *mcr*-

2-positive isolates [89]. Although the *mcr* genes were detected with 100% accuracy with a LOD

of 10^2, *mcr*-2 was not validated on cultured bacteria [89]. However, in this study all three *mcr*

genes could not be simultaneously detected in one reaction unlike when the Taqman probe

was used [89]. A more recent study evaluated the SYBR Green-based real-time PCR method

proposed by Li *et al* (2017) in detecting and quantifying *mcr*-1 to *mcr*-3 as well as newly

designed assays for *mcr*-4 and *mcr*-5 [87]. The optimized *mcr*-1 to *mcr*-5 PCR assays were

validated on bacterial isolates [87]. The study found that SYBR Green real-time PCR, followed

by melting curve analysis, was more efficient in detecting and quantifying *mcr*-1 to *mcr*-5

genes in both bacterial isolates [87]. The described assays detected all five *mcr* genes with a

lower limit of 10^2. Moreover, the assays enabled screening of five individual samples in a

single reaction [87].

The parallel detection of *mcr*-1, *mcr*-2 and *mcr*-8 by real-time PCR using Taqman® probes

has been described [91-93]. Chabou *et al* (2016) designed two quantitative real-time PCR assays

with TaqMan® probes for the rapid detection of *mcr*-1 gene. Primers and probes were

designed to develop the two PCR assays, designated PE1 and PE2 [91]. The assays were

evaluated on 100 bacterial isolates (18 of which were colistin resistant) and 833 broiler faecal

samples [91]. The sensitivity and specificity of both assays were 100%, with a calibration curve

that was linear from 10^1 to 10^8. However, the PE1 assay was recommended for initial

screening of *mcr*-1 followed by PE2 assay for confirming the results [91].

Daniels *et al* (2017) developed a multiplexed real-time PCR with TaqMan® probes to detect

mcr-1 and *mcr*-2. The assay was validated on 25 bacterial isolates, some of which were *mcr*-

positive [93]. The sensitivity and specificity of the assay was 100%, being able to detect *mcr*-1

and *mcr*-2 from dilutions containing 8.5×10^3 and 7.7×10^3 cfu/mL respectively [93]. A

specific real-time PCR assay using TaqMan probes to identify *mcr*-8 was designed for the

first time by Nabti *et al* (2020). The specificity and sensitivity of the assay were evaluated on

290 bacterial isolates from clinical samples and 250 metagenomic DNA from human stools [92].

The PCR assay accurately detected *mcr*-8 from the one positive *K. pneumoniae* isolate with

an overall efficiency of 92.64% and a limit of detection of 55 CFU/mL [92].

4. Conclusion

The rapid dissemination of colistin resistance, mediated by chromosomal mutations and *mcr*-

genes, poses a threat to public and veterinary health as colistin is one of the last-line

antibiotics. It is important to establish a simple, rapid, and cost-effective diagnostic method that will not discriminate against the mechanism of colistin resistance as well as take into consideration hetero-resistant isolates. Among the available diagnostic assays, the Rapid Resapolymyxin *Acinetobacter*/*Pseudomonas* NP test is a promising initial screening method as it can be performed in-house, therefore making it relatively cheap; it is easy to perform and it is not limited to glucose-fermenting colistin-resistant Gram-negative bacteria. Other colorimetric screening methods such as ASAT, Rapid Polymyxin NP, Rapid Polymyxin *Acinetobacter,* and Rapid Polymyxin *Pseudomonas* are species specific and cannot be used for general screening in high capacity clinical laboratories.

Likewise, agar-based methods are cheap and can be used as initial screening tools in poorer settings, although most of the agar-based assays fail to detect isolates with hetero-resistance. They also have a longer turnaround time of 24 hours. Should there be a need to use agar based assays, CHROMagar COL-APSE was designed to detect and differentiate all colistin-resistant isolates, although it might be relatively expensive as compared to Superpolymyxin as well as LBJMR medium, which can be performed in-house. However, the LBJMR medium was found to detect hetero-resistance better than Superpolymyxin. The other two agar-based media can only be used to screen for *Enterobacteriaceae.*

MIC determiners could also be used for initial screening in high resource laboratories, although they relatively require a higher skill as compared to the agar-based tests and the biochemical colorimetric tests. The non-automated MIC strips *i.e*. UMIC, MMS and ComASP are cheaper than automated MIC determiners *i.e*. Microscan, Sensititre, BD Phoenix. However, automated MIC determiners could be available in most well-resourced clinical laboratories as they are generally used for Antimicrobial Susceptibility testing. Non-automated MIC strips are cheaper, require less skill, and recommendable for less-resourced laboratories.

A second screening can be performed to mainly detect *mcr*-production using chelator-based phenotypic assays, which are more suitable although most are subjected to >16h incubation. However, the Lateral Flow assay that detects MCR-1/-2 production could be used for rapid detection. Molecular methods could be considered for the detection of *mcr*-genes in well-resourced clinical microbiology laboratories. Particularly, the LAMP assays could be used as they were found to be more sensitive than PCR methods. More so, LAMP requires less equipment and has a shorter turnaround time than PCR and WGS methods.

Funding: none

Acknowledgements: none

Transparency declaration: Authors have no conflict of interest to declare.

5. References

1. Vasoo, S. Susceptibility Testing for the Polymyxins: Two Steps Back, Three Steps Forward?
Journal of clinical microbiology **55**, 2573--2582 (2017).

2. Teo, J.W., Octavia, S., Cheng, J.W. & Lin, R.T., Vol. 90 67--69 (2018).

3. Li, B. *et al.* An enzyme-free homogenous electrochemical assay for sensitive detection of the
plasmid-mediated colistin resistance gene mcr-1. *Analytical and bioanalytical chemistry* **410**,
4885--4893 (2018).

4. Jousset, A.B. *et al.* Development and validation of a multiplex polymerase chain reaction
assay for detection of the five families of plasmid-encoded colistin resistance. *Int J*
Antimicrob Agents **53**, 302-309 (2019).

5. Osei Sekyere, J., Sephofane, A.K. & Mbelle, N.M. Comparative Evaluation of CHROMagar
COL-APSE, MicroScan Walkaway, ComASPColistin, and Colistin MAC Test in Detecting
Colistin-resistant Gram-Negative Bacteria. *medRxiv*, 2019.2012.2017.19015156 (2020).

6. Liu, Y.-Y. *et al.* Emergence of plasmid-mediated colistin resistance mechanism MCR-1 in
animals and human beings in China: a microbiological and molecular biological study. *The*
Lancet infectious diseases **16**, 161-168 (2016).

7. Dortet, L. *et al.* Optimization of the MALDIxin test for the rapid identification of colistin
resistance in Klebsiella pneumoniae using MALDI-TOF MS. *J Antimicrob Chemother* **75**, 110-
116 (2020).

8. Bialvaei, A.Z. & Samadi Kafil, H. Colistin, mechanisms and prevalence of resistance. *Current*
Medical Research and Opinion **31**, 707-721 (2015).

9. Sun, J., Zhang, H., Liu, Y.-H. & Feng, Y. Towards Understanding MCR-like Colistin Resistance.
Trends in Microbiology **26**, 794-808 (2018).

10. Poirel, L., Jayol, A. & Nordmann, P. Polymyxins: Antibacterial Activity, Susceptibility Testing,
and Resistance Mechanisms Encoded by Plasmids or Chromosomes. *Clinical microbiology*
reviews **30**, 557-596 (2017).

11. Nordmann, P., Jayol, A. & Poirel, L. Rapid Detection of Polymyxin Resistance in
Enterobacteriaceae. *Emerging infectious diseases* **22**, 1038--1043 (2016).

12. Osei Sekyere, J. Mcr colistin resistance gene: a systematic review of current diagnostics and
detection methods. *Microbiologyopen* **8**, e00682 (2019).

13. Rodriguez, C.H. *et al.* In-house rapid colorimetric method for detection of colistin resistance
in Enterobacterales: A significant impact on resistance rates. *Journal of chemotherapy*
(Florence, Italy) **31**, 432--435 (2019).

14. Xavier, B.B. *et al.* Identification of a novel plasmid-mediated colistin-resistance gene, mcr-2,
in Escherichia coli, Belgium, June 2016. *Euro surveillance : bulletin Europeen sur les maladies*
transmissibles = European communicable disease bulletin **21** (2016).

15. AbuOun, M. *et al.* mcr-1 and mcr-2 variant genes identified in Moraxella species isolated
from pigs in Great Britain from 2014 to 2015. *The Journal of antimicrobial chemotherapy* **72**,
2745-2749 (2017).

16. Borowiak, M. *et al.* Identification of a novel transposon-associated phosphoethanolamine
transferase gene, mcr-5, conferring colistin resistance in d-tartrate fermenting Salmonella
enterica subsp. enterica serovar Paratyphi B. *The Journal of antimicrobial chemotherapy* **72**,
3317-3324 (2017).

17. Carattoli, A. *et al.* Novel plasmid-mediated colistin resistance mcr-4 gene in Salmonella and
Escherichia coli, Italy 2013, Spain and Belgium, 2015 to 2016. *Euro surveillance : bulletin*
Europeen sur les maladies transmissibles = European communicable disease bulletin **22**
(2017).

18. Carroll, L.M. *et al.* Identification of Novel Mobilized Colistin Resistance Gene mcr-9 in a
Multidrug-Resistant, Colistin-Susceptible Salmonella enterica Serotype Typhimurium Isolate.
mBio **10** (2019).

19. Wang, X. *et al.* Emergence of a novel mobile colistin resistance gene, mcr-8, in NDM-
producing Klebsiella pneumoniae. *Emerging microbes & infections* **7**, 122 (2018).

20. Wenjuan, Y. *et al.* in mBio, Vol. 8 (2017).

21. Yong-Qiang, Y. *et al.* Novel plasmid-mediated colistin resistance gene mcr-7.1 in Klebsiella
pneumoniae. *Journal of Antimicrobial Chemotherapy (JAC)* **73** (2018).

22. Furniss, R.C.D. *et al.* Detection of Colistin Resistance in Escherichia coli by Use of the MALDI
Biotyper Sirius Mass Spectrometry System. *Journal of clinical microbiology* **57** (2019).

23. Bardet, L. & Rolain, J.-M. Development of New Tools to Detect Colistin-Resistance among
Enterobacteriaceae Strains. *Can J Infect Dis Med Microbiol* **2018**, 3095249-3095249 (2018).

24. Furniss, R.C.D. *et al.* Detection of Colistin Resistance in <span class="named-content
genus-species" id="named-content-1">Escherichia coli by
Use of the MALDI Biotyper Sirius Mass Spectrometry System. *Journal of Clinical Microbiology*
57, e01427-01419 (2019).

25. Chew, K.L., La, M.V., Lin, R.T.P. & Teo, J.W.P. Colistin and Polymyxin B Susceptibility Testing
for Carbapenem-Resistant and mcr-Positive Enterobacteriaceae: Comparison of Sensititre,
MicroScan, Vitek 2, and Etest with Broth Microdilution. *J Clin Microbiol* **55**, 2609-2616 (2017).

26. Jayol, A., Nordmann, P., Andre, C., Poirel, L. & Dubois, V. Evaluation of three broth
microdilution systems to determine colistin susceptibility of Gram-negative bacilli. *The*
Journal of antimicrobial chemotherapy **73**, 1272--1278 (2018).

27. Simar, S., Sibley, D., Ashcraft, D. & Pankey, G. Colistin and Polymyxin B Minimal Inhibitory
Concentrations Determined by Etest Found Unreliable for Gram-Negative Bacilli. *The Ochsner*
journal **17**, 239--242 (2017).

28. Haeili, M., Kafshdouz, M., Pishnian, Z., Feizabadi, M.M. & Martinez-Martinez, L. Comparison
of susceptibility testing methods for determining the activity of colistin against Gram-
negative bacilli of clinical origin. *Journal of medical microbiology* **68**, 60--66 (2019).

29. Lutgring, J.D. *et al.* Evaluation of the MicroScan Colistin Well and Gradient Diffusion Strips for
Colistin Susceptibility Testing in Enterobacteriaceae. *Journal of Clinical*
Microbiology **57**, e01866-01818 (2019).

30. Matuschek, E., Åhman, J., Webster, C. & Kahlmeter, G. Antimicrobial susceptibility testing of
colistin–evaluation of seven commercial MIC products against standard broth microdilution

for *Escherichia coli, Klebsiella pneumoniae, Pseudomonas aeruginosa*, and *Acinetobacter*
spp. *Clinical Microbiology and Infection* **24**, 865-870 (2018).

31. Mitton, B., Kingsburgh, C., Kock, M., Mbelle, N. & Strydom, K. Evaluation of an In-House
Colistin NP Test for Use in Resource-Limited Settings. *Journal of clinical microbiology* **57**,
e00501-00519 (2019).

32. Jayol, A. *et al.* Hafnia, an enterobacterial genus naturally resistant to colistin revealed by
three susceptibility testing methods. *The Journal of antimicrobial chemotherapy* **72**, 2507--
2511 (2017).

33. Bardet, L., Okdah, L., Le Page, S., Baron, S.A. & Rolain, J.M. Comparative evaluation of the
UMIC Colistine kit to assess MIC of colistin of gram-negative rods. *BMC Microbiol* **19**, 60
(2019).

34. Carretto, E. *et al.* Clinical Validation of SensiTest Colistin, a Broth Microdilution-Based
Method To Evaluate Colistin MICs. *Journal of Clinical Microbiology* **56**, e01523-01517 (2018).

35. Hong, J.S. *et al.* Evaluation of the BD Phoenix M50 Automated Microbiology System for
Antimicrobial Susceptibility Testing with Clinical Isolates in Korea. *Microb Drug Resist* **25**,
1142-1148 (2019).

36. Jayol, A., Nordmann, P., Lehours, P., Poirel, L. & Dubois, V. Comparison of methods for
detection of plasmid-mediated and chromosomally encoded colistin resistance in
Enterobacteriaceae. *Clinical Microbiology and Infection* **24**, 175-179 (2018).

37. Rodriguez, C.H. *et al.* Discrepancies in susceptibility testing to colistin in Acinetobacter
baumannii: The influence of slow growth and heteroresistance. *International journal of*
antimicrobial agents **54**, 587--591 (2019).

38. Poirel, L. *et al.* Rapid Polymyxin NP test for the detection of polymyxin resistance mediated
by the mcr-1/mcr-2 genes. *Diagnostic microbiology and infectious disease* **90**, 7--10 (2018).

39. Jayol, A. *et al.* Evaluation of the Rapid Polymyxin NP test and its industrial version for the
detection of polymyxin-resistant Enterobacteriaceae. *Diagn Microbiol Infect Dis* **92**, 90-94
(2018).

40. Jayol, A., Dubois, V., Poirel, L. & Nordmann, P. Rapid Detection of Polymyxin-Resistant
Enterobacteriaceae from Blood Cultures. *Journal of clinical microbiology* **54**, 2273--2277
(2016).

41. Malli, E., Papagiannitsis, C.C., Xitsas, S., Tsilipounidaki, K. & Petinaki, E. Implementation of the
Rapid Polymyxin™ NP test directly to positive blood cultures bottles. *Diagnostic Microbiology*
and Infectious Disease **95**, 114889 (2019).

42. Malli, E. *et al.* Evaluation of rapid polymyxin NP test to detect colistin-resistant Klebsiella
pneumoniae isolated in a tertiary Greek hospital. *Journal of microbiological methods* **153**, 35-
-39 (2018).

43. Dalmolin, T.V. *et al.* Detection of Enterobacterales resistant to polymyxins using Rapid
Polymyxins NP test. *Braz J Microbiol* **50**, 425-428 (2019).

44. Malli, E. *et al.* Implementation of the Rapid Polymyxin Acinetobacter Test to Detect Colistin-
Resistant Acinetobacter baumannii. *Microbial drug resistance (Larchmont, N.Y.)* (2019).

45. Lescat, M., Poirel, L., Jayol, A. & Nordmann, P. Performances of the Rapid Polymyxin
Acinetobacter and Pseudomonas Tests for Colistin Susceptibility Testing. *Microb Drug Resist*
25, 520-523 (2019).

46. Lescat, M., Poirel, L., Tinguely, C. & Nordmann, P. A Resazurin Reduction-Based Assay for
Rapid Detection of Polymyxin Resistance in Acinetobacter baumannii and Pseudomonas
aeruginosa. *J Clin Microbiol* **57** (2019).

47. Jia, H. *et al.* Evaluation of resazurin-based assay for rapid detection of polymyxin-resistant
gram-negative bacteria. *BMC Microbiol* **20**, 7 (2020).

48. Germ, J. *et al.* Evaluation of resazurin-based rapid test to detect colistin resistance in
Acinetobacter baumannii isolates. *Eur J Clin Microbiol Infect Dis* **38**, 2159-2162 (2019).

49. Esposito, F. *et al.* Detection of Colistin-Resistant MCR-1-Positive Escherichia coli by Use of
Assays Based on Inhibition by EDTA and Zeta Potential. *J Clin Microbiol* **55**, 3454-3465 (2017).

50. Yauri Condor, K., Gonzales Escalante, E., Di Conza, J. & Gutkind, G. Detection of plasmid-
mediated colistin resistance by colistin pre-diffusion and inhibition with EDTA test (CPD-E) in
Enterobactereaceae. *Journal of Microbiological Methods* **167** (2019).

51. Clément, M. *et al.* The EDTA-based disk-combination tests are unreliable for the detection of
MCR-mediated colistin-resistance in Enterobacteriaceae. *Journal of Microbiological Methods*
153, 31-34 (2018).

52. Budel, T. *et al.* Evaluation of EDTA- and DPA-Based Microdilution Phenotypic Tests for the
Detection of MCR-Mediated Colistin Resistance in Enterobacteriaceae. *Microb Drug Resist*
25, 494-500 (2019).

53. Simner, P.J. *et al.* Two-Site Evaluation of the Colistin Broth Disk Elution Test To Determine
Colistin In Vitro Activity against Gram-Negative Bacilli. *Journal of*
Clinical Microbiology **57**, e01163-01118 (2019).

54. Dalmolin, T.V. *et al.* Elution methods to evaluate colistin susceptibility of Gram-negative rods.
Diagnostic Microbiology and Infectious Disease **96**, 114910 (2020).

55. Fenwick, A.J. *et al.* Evaluation of the NG-Test MCR-1 Lateral Flow Assay and EDTA-Colistin
Broth Disk Elution Methods To Detect Plasmid-Mediated Colistin Resistance Among Gram-
Negative Bacterial Isolates. *Journal of Clinical Microbiology*, JCM.01823-01819 (2020).

56. Bell, D.T. *et al.* A Novel Phenotypic Method To Screen for Plasmid-Mediated Colistin
Resistance among Enterobacteriales. *Journal of Clinical Microbiology*
57, e00040-00019 (2019).

57. Volland, H. *et al.* Development and Multicentric Validation of a Lateral Flow Immunoassay for
Rapid Detection of MCR-1-Producing Enterobacteriaceae. *Journal of clinical microbiology* **57**
(2019).

58. Abdul Momin, M.H.F. *et al.* CHROMagar COL-APSE: a selective bacterial culture medium for
the isolation and differentiation of colistin-resistant Gram-negative pathogens. *J Med*
Microbiol **66**, 1554-1561 (2017).

59. Girlich, D., Naas, T. & Dortet, L. Comparison of the Superpolymyxin and ChromID Colistin R
Screening Media for the Detection of Colistin-Resistant Enterobacteriaceae from Spiked
Rectal Swabs. *Antimicrob Agents Chemother* **63** (2019).

60. Nordmann, P., Jayol, A. & Poirel, L. A Universal Culture Medium for Screening Polymyxin-
Resistant Gram-Negative Isolates. *Journal of clinical microbiology* **54**, 1395--1399 (2016).

61. Bardet, L., Page, S.L., Leangapichart, T. & Rolain, J.-M. LBJMR medium: a new polyvalent
culture medium for isolating and selecting vancomycin and colistin-resistant bacteria. *BMC*
microbiology **17**, 220 (2017).

62. Jayol, A., Poirel, L., Andre, C., Dubois, V. & Nordmann, P. Detection of colistin-resistant Gram-
negative rods by using the SuperPolymyxin medium. *Diagn Microbiol Infect Dis* **92**, 95-101
(2018).

63. Germ, J. & Seme, K.a. Evaluation of a novel epidemiological screening approach for detection
of colistin resistant human Enterobacteriaceae isolates using a selective SuperPolymyxin
medium. *Journal of microbiological methods* **160**, 117--123 (2019).

64. Thiry, D. *et al.* Assessment of two selective agar media to isolate colistin-resistant bovine
Escherichia coli: Correlation with minimal inhibitory concentration and presence of mcr
genes. *Journal of Microbiological Methods* **159**, 174--178 (2019).

65. García-Fernández, S. *et al.* Performance of CHROMID® Colistin R agar, a new chromogenic
medium for screening of colistin-resistant Enterobacterales. *Diagnostic Microbiology and
Infectious Disease* **93**, 1-4 (2019).

66. Coppi, M. *et al.* A simple phenotypic method for screening of MCR-1-mediated colistin
resistance. *Clinical microbiology and infection : the official publication of the European
Society of Clinical Microbiology and Infectious Diseases* **24**, 1-201 (2018).

67. Turbett, S.E. *et al.* Evaluation of a Screening Method for the Detection of Colistin-Resistant
Enterobacteriaceae in Stool. *Open forum infectious diseases* **6**, ofz211 (2019).

68. Girlich, D. *et al.* Evaluation of the Amplidiag CarbaR+MCR Kit for Accurate Detection of
Carbapenemase-Producing and Colistin-Resistant Bacteria. *Journal of Clinical Microbiology*
57, e01800-01818 (2019).

69. Meunier, D., Woodford, N. & Hopkins, K.L. Evaluation of the AusDiagnostics MT CRE EU assay
for the detection of carbapenemase genes and transferable colistin resistance determinants
mcr-1/-2 in MDR Gram-negative bacteria. *The Journal of antimicrobial chemotherapy* **73**,
3355--3358 (2018).

70. Zhong, L.-L., Zhou, Q., Tan, C.-Y. & Roberts, A.P.a. Multiplex loop-mediated isothermal
amplification (multi-LAMP) assay for rapid detection of mcr-1 to mcr-5 in colistin-resistant
bacteria. *Infect Drug Resist* **12**, 1877--1887 (2019).

71. Imirzalioglu, C. *et al.*, Vol. 61 (2017).

72. Zou, D. *et al.* Sensitive and Rapid Detection of the Plasmid-Encoded Colistin-Resistance Gene
mcr-1 in Enterobacteriaceae Isolates by Loop-Mediated Isothermal Amplification. *Frontiers in
microbiology* **8**, 2356 (2017).

73. Giordano, C. & Barnini, S. Rapid detection of colistin-resistant Klebsiella pneumoniae using
MALDI-TOF MS peak-based assay. *J Microbiol Methods* **155**, 27-33 (2018).

74. Dortet, L. *et al.* Rapid detection of colistin resistance in Acinetobacter baumannii using
MALDI-TOF-based lipidomics on intact bacteria. *Scientific reports* **8**, 16910 (2018).

75. Bernasconi, O.J. *et al.*, Vol. 55 3138--3141 (2017).

76. Gong, L. *et al.* Multiple Cross Displacement Amplification Coupled With Gold Nanoparticles-
Based Lateral Flow Biosensor for Detection of the Mobilized Colistin Resistance Gene mcr-1.
Frontiers in cellular and infection microbiology **9**, 226 (2019).

77. Xu, J., Wang, X., Yang, L., Kan, B. & Lu, X. Rapid detection of mcr-1 by recombinase
polymerase amplification. *Journal of medical microbiology* **67**, 1682--1688 (2018).

78. Peter, S. *et al.* Whole-genome sequencing enabling the detection of a colistin-resistant
hypermutating Citrobacter werkmanii strain harbouring a novel metallo-beta-lactamase VIM-
48. *International journal of antimicrobial agents* **51**, 867--874 (2018).

79.   Hua, X. *et al.* Colistin Resistance in Acinetobacter baumannii MDR-ZJ06 Revealed by a
Multiomics Approach. *Frontiers in cellular and infection microbiology* **7**, 45 (2017).

80.   Rebelo, A.R. *et al.* Multiplex PCR for detection of plasmid-mediated colistin resistance
determinants, mcr-1, mcr-2, mcr-3, mcr-4 and mcr-5 for surveillance purposes. *Euro*
surveillance : bulletin Europeen sur les maladies transmissibles = European communicable
disease bulletin **23** (2018).

81.   Nijhuis, R. *et al.* Detection of the plasmid-mediated colistin-resistance gene mcr-1 in clinical
isolates and stool specimens obtained from hospitalized patients using a newly developed
real-time PCR assay. *Journal of Antimicrobial Chemotherapy* **71**, 2344-2346 (2016).

82.   Hatrongjit, R., Kerdsin, A., Akeda, Y. & Hamada, S. Detection of plasmid-mediated colistin-
resistant and carbapenem-resistant genes by multiplex PCR. *MethodsX* **5**, 532--536 (2018).

83.   Lescat, M., Poirel, L. & Nordmann, P. Rapid multiplex polymerase chain reaction for detection
of mcr-1 to mcr-5 genes. *Diagnostic Microbiology and Infectious Disease* **92**, 267-269 (2018).

84.   Borowiak, M. *et al.* Development of a Novel mcr-6 to mcr-9 Multiplex PCR and Assessment of
mcr-1 to mcr-9 Occurrence in Colistin-Resistant Salmonella enterica Isolates From
Environment, Feed, Animals and Food (2011–2018) in Germany. *Frontiers in Microbiology* **11**
(2020).

85.   Chandler, J.C. *et al.* Validation of a screening method for the detection of colistin-resistant E.
coli containing mcr-1 in feral swine feces. *Journal of microbiological methods* **172**, 105892
(2020).

86.   Chalmers, G. *et al.* A method to detect Escherichia coli carrying the colistin-resistance genes
mcr-1 and mcr-2 using a single real-time polymerase chain reaction and its application to
chicken cecal and porcine fecal samples. *Canadian journal of veterinary research = Revue*
canadienne de recherche veterinaire **82**, 312--315 (2018).

87.   Tolosi, R. *et al.* Rapid detection and quantification of plasmid-mediated colistin resistance
genes (mcr-1 to mcr-5) by real-time PCR in bacterial and environmental samples. *Journal of*
Applied Microbiology **n/a** (2020).

88.   Bontron, S.v., Poirel, L. & Nordmann, P. Real-time PCR for detection of plasmid-mediated
polymyxin resistance (mcr-1) from cultured bacteria and stools. *Journal of Antimicrobial*
Chemotherapy (JAC) **71** (2016).

89.   Li, J. *et al.* A Multiplex SYBR Green Real-Time PCR Assay for the Detection of Three Colistin
Resistance Genes from Cultured Bacteria, Feces, and Environment Samples. *Frontiers in*
microbiology **8**, 2078 (2017).

90.   Dona, V., Bernasconi, O.J., Kasraian, S., Tinguely, R. & Endimiani, A. A SYBR((R)) Green-based
real-time PCR method for improved detection of mcr-1-mediated colistin resistance in
human stool samples. *J Glob Antimicrob Resist* **9**, 57-60 (2017).

91.   Chabou, S., Leangapichart, T. & Okdah, L.a. Real-time quantitative PCR assay with
Taqman((R)) probe for rapid detection of MCR-1 plasmid-mediated colistin resistance. *New*
microbes and new infections **13**, 71--74 (2016).

92.   Nabti, L.Z. *et al.* Development of real-time PCR assay allowed describing the first clinical
Klebsiella pneumoniae isolate harboring plasmid-mediated colistin resistance mcr-8 gene in
Algeria. *Journal of global antimicrobial resistance* **20**, 266--271 (2020).

93.   Daniels, J.B. *et al.* Development and Validation of a Clinical Laboratory Improvement
Amendments-Compliant Multiplex Real-Time PCR Assay for Detection of mcr Genes. *Microb*
Drug Resist **25**, 991-996 (2019).

94.   Wang, H. *et al.* Rapid detection of colistin resistance protein MCR-1 by LC-MS/MS. *Clinical*
1057   *proteomics* **16**, 8 (2019).

1058

1059

1060   *Figure 1. Literature search strategy and sorting techniques used to obtain articles for inclusion in the systematic*
1061   *review.*

Table 1. Comparative diagnostic efficiencies of colistin resistance diagnostics

Diagnostic	Species (n)	Sensitivity (%)	Specificity (%)	Relative cost	Relative skill required	Turnaround time (hr)	CA (%)	EA (%)	ME (%)	VME (%)	LOD (CFU/ml or reaction)	References
Broth Macrodiluti on (BMAD)	Gram-negative bacilli (109)	100	100	Cheap	Low	16-20	100	100	0	0	NS	Haeili *et al.*, 2019
Diffusion and Agar dilution methods												
Disc diffusion (Rosco, Taastrup, Denmark)	Gram-negative bacilli (109)	76.19	96.59	Cheapest	Low	16-24	92.6	NA	3.4	23.8	NS	Haeili *et al.*, 2019

Disc diffusion (BBL Sensi-Disc, Becton Dickinson, Sparks, MD)	Gram-negative bacilli (109)	76.19	100	Cheapest	Low	16-24	95.4	NA	0	23.8	NS	Haeili et al., 2019
Disc diffusion	Hafnia (25)	0	100	Cheapest	Low	16-24	0	NA	0	100	NS	Jayol et al., 2017
Agar dilution	Gram-negative bacilli (109)	100	60.23	Cheap	Low	16-20	67.9	23.8	39.77	0	NS	Haeili et al., 2019
Selective agar-based media												
Superpolymyxin™	Enterobacteriaceae (94,700,231)	86.8,87.5,95.2	97.5,99.5,95.3	Cheap	Low	24-48	NA	NA	NA	NA	NS	Girlich et al., 2019; Germ et al, 2019;Jayol et al., 2018;

CHROMagar™ COL-APSE	*E.coli* (158) Gram-negative bacilli (89)	96,82.05	97,66.67	Cheap	Low	18-20	NA	NA	NA	NA	NS	Thiry et al., 2019; Osei Sekyere *et al.*, 2020
CHROMID® Colistin_R	*E. coli* 158 *Enterobacteriaceae* (94)	99, 86.8	97,100	Cheap	Low	18-24	NA	NA	NA	NA	NS	Thiry et al., 2019; Girlich *et al.*, 2019;
LB Medium	*Enterobacteriaceae* (9)	100	100	Cheap	Low	48	NA	NA	NA	NA	10^2-10^3	Turbett *et al.*, 2019
Chelator and non-chelator-based assays												
Rapid Polymyxin NP (Commercial)	*Enterobacteriaceae* (132)	100	96.7	Expensive	Low	2-3	NS	NA	3.48	0	NS	Malli *et al.*, 2019;

In-house Rapid polymyxin NP	*K. pnuemoniae* (131) *Hafnia* spp. (25) Enterobacteriaceae (200)	99,100,98, 98.21,98.1,	82,100,98, 100,98.2	Cheapest	Low	2-4	NS	NA	18.18,0,2, 0,0,1.8	1.02,0,2,1. 79,1.9	10^8	Malli *et al.*, 2018; Jayol *et al.*, 2017; Dalmolin *et al.*, 2019; Jayol *et al.*, 2016; Mitton *et al.*, 2019
ASAT	*Enterobact eriaceae* (300)	90.7	100	Cheapest	Low	3	NS	NA	0	10.17	NS	Rodriguez *et al.*, 2019a
Rapid *Acinetobact er*	*A. baumannii* (21)	100,93.3	100,86.8	Expensive	Low	3-4	100	NA	0,	0,	NS	Lescat *et al.*, 2019; Malli *et al.*, 2019b
Rapid *Pseudomo nas*	*P. aeruginosa* (17)	100	100	Expensive	Low	3-4	100	NA	0	0	NS	Lescat *et al.*, 2019

Method	Organism			Cost		Time (h)						References
Rapid Resapolymyxin NP	*A. baumannii* (82, 43,165) *Enterobacteriaceae* (20,32) *P. aeruginosa* (49) Gram-negative bacilli (253)	93.3,100,100,100,100,84	93.3,100,97,100,99,100	Cheap	Low	3-4	95.1,100,cal,99.2,cal	NA	3.7,0,4 34,0,1.01,0	1.2,0,0,0,0,16	NS	Germ *et al.*, 2019; Lescat *et al.*, 2019b; Jia *et al.*, 2020; Rodriguez *et al.*, 2019
EDTA-CDT	*Enterobacteriaceae* (48,92)	12,100	65.2,100	Cheap	Low	18-24	NS,100	NA	NS,0	NS,0	NA	Clément *et al.*, 2018; Yauri Condor *et al.*, 2019
EDTA-CMR/BMD	*Enterobacteriaceae* (92)	93.2	54.2	Cheap	Low	18-24	NS	NS	6.8	45.81	NS	Budel *et al.*, 2019

Colistin MAC test	*Enterobacteriaceae* (92) Gram-neagtive bacilli (84)	84.1, 100	100	Cheap	Low	18-24	NA	NA	0	15.9,0	NS	Budel *et al.,* 2019; Osei Sekyere *et al.,* 2020
CBDE	Gram negative bacilli (172)	100	97.76	Cheap	Low	16-20	98	99	0	8	NS	Simner *et al.,* 2019
CBDE+EDTA	*Enterobacteriaceae* (85)	100	95.8, 94.3	Cheap	Low	16-20	NS	NS	4.1, 5.75	0	NS	Bell *et al.,* 2019; Fenwick *et al.,* 2020
CBM	*Enterobacteriaceae* (68) Non-fermenters (17)	95.35,87.5	84,66.67	Cheap	Low	16-20	91.18,76.47	95.59,82.35	16,33.33	4.65,12.5	NS	Dalmolin *et al.,* 2020

MPT	_Enterobacteriaceae_ (68) Non-fermenters (17)	88.37, 75	80,77.78	Cheap	Low	16-20	85.29,76.47	98.53,76.47	20,22.22	11.63,25	NS	Dalmolin _et al._, 2020
CSTT	_Enterobacteriaceae_ (68) Non-fermenters (17)	93.02,62.5	88,100	Cheap	Low	16-20	91.18,82.35	NA	12	6.98,37.5	NS	Dalmolin _et al._, 2020
Commercial BMD-based assays												
UMIC	Gram-negative bacilli (235)	100	100	Expensive	High	18-24	100	94	0	0	NS	Bardet _et al._, 2019
ComASP Colistin (Sensitest)	Gram-negative bacilli (84)	87.18	73.33	Expensive	High	16-24	78.57	78.57	26.67	12.82	NS	Osei Sekyere _et al._, 2020

BD Phoenix 100	*Hafnia* spp. (25) *A. baumannii* (165) *K. pnuemoniae* (131)	100,48, 68, 95	100,100,100,94	Very Expensive	High	16-24	100,	100,	0,0,0,**6.06**	0,52,32,**5.1**	NS	Jayol *et al.*,2017; Rodriguez *et al.*,2019; Malli *et al.*, 2018
BD Phoenix M50	Gram-Negative Bacilli (131)	100	94.29	Very expensive	High	16-24	98.5	98.5	5.7	0	NS	Hong *et al.*, 2019
MicroScan	Enterobacteriaceae (110) Gram-negative bacilli	89.29,85.71	93.90,76.19	Very expensive	High	16-24	92.7,80.95	NS	6.1,23.81	10.7,14.29	NS	Lutgring *et al.*, 2019; Osei Sekyere *et al.*, 2020

Vitek 2	*K.pnuemoniae* (131)	98	88	Very expensive	High	18-24	NS	NS	12.12	2.04	NS	Malli *et al.*, 2018;
Molecular assays												
Amplidiag® Carba-R+MCR assay	Gram negative bacilli	92-100	86-100	Expensive	High	<3	NA	NA	NA	NA	NS	Girlich *et al.*, 2019a
AusDiagnostic MT CRE EU assay	Gram-negative bacilli (393)	95.5-100	99.8	Expensive	High	<4	NA	NA	NA	NA	NS	Meunier *et al.*, 2018

Method	Bacteria (n)											Reference
Lateral Flow immunoassay (NG-TEST MCR-1)	*Enterobacteriaceae* (298, 138) *A. baumannii* (50) *P.aeruginosa* (50)	100	98,99	Expensive	Low-High	<0.25	NA	NA	0	0.34	2.10^8	Volland *et al.,* 2019; Fenwick *et al.,* 2020
LAMP-*mcr*-1	Gram-negative bacilli (22)	100	100	Expensive	High	<1	NA	NA	NA	NA	10^8	Zou *et al.,* 2017
Multiplex-LAMP	*Enterobacteriaceae* (58)	100	100	Expensive	Very high	<1	NA	NA	NA	NA	10^4-10^5	Zhong *et al.,* 2019
MALDI-TOF MS	*K. pneumonia (139)*	78	89	Expensive	Very high	?	NA	NA	NA	NA	NS	Giordano and Barnini, 2018

MALDIxin	*A.baumannii* (17) *E. coli* (40) *K.pneumoniae* (81)	100	100	Expensive	Very high	0.25-0.5	NA	NA	NA	NA	NS	Furniss *et al.*, 2019; Dortet *et al.*, 2020; Dortet *et al.*, 2018
MCDA-LFB	*mcr*-1 positive bacteria (16) non-*mcr*-1 bacteria (43)	100	100	Expensive	High	<1	NA	NA	NA	NA	600 fg	Gong *et al.*, 2019

Multiplex PCR	E. coli (50) K. pneumoniae (41) S. enterica (12, 407) Aeromonas spp. (82) Shewenella spp. (10)	100	100	Very expensive	High	<2-3	NA	NA	NA	NA	NS	Jousset et al., 2019; Borowiak et al.,2020
Real-time PCR	E. coli (1110)	32-98	100	Very expensive	High	<3	NA	NA	NA	NA	0.1-2200 cfu/ml	Chandler et al., 2020
Real-time PCR (SYBR Green)	E. coli (332, 1062)	100	100	Very expensive	High	<3	NA	NA	NA	NA	54 cfu/ml	Chalmers et al., 2018; Tolosi et al., 2020;
Real-time PCR (Taqman)	Gram-negative bacilli (23)	100	100	Very expensive	High	<3	NA	NA	NA	NA	10^3	Daniels et al., 2019;

| Recombinase polymerase amplification | *Enterobacteriaceae* (23) | 100 | 100 | Expensive | High | <0.5 | NA | NA | NA | NA | 100fg | Xu *et al.*, 2018 |
| Whole genome sequence | *Citrobacter werkmanii* (5) | 100 | 100 | Most expensive | Highest | <48 | NA | NA | NA | NA | NS | Peter *et al.*, 2018 |