Title: Performance characteristics of the VIDAS® SARS-COV-2 IgM and IgG serological assays

Running title: VIDAS® SARS-COV-2 IgM and IgG assay performance

Nathalie Renard[a#], Soizic Daniel[a], Nadège Cayet[a], Matthieu Pecquet[b], Frédérique Raymond[a], Sylvie Pons[c], Julien Lupo[d], Carole Tourneur[a], Catherine Pretis[a], Guillaume Gerez[a], Patrick Blasco[a], Maxime Combe[a], Imen Canova[a], Mylène Lesénéchal[a#], Franck Berthier[a]

[a]R&D bioMérieux, 376 Chemin de l'orme, 69280 Marcy l'Etoile, FRANCE

[b]Laboratoire de Biologie, Centre Hospitalier Saint Joseph Saint Luc, 20 Quai Claude Bernard, 69007 Lyon, FRANCE

[c]Laboratoire commun de recherche Hospices Civils de Lyon - bioMérieux, Centre Hospitalier Lyon-Sud, Bâtiment 3F, Chemin du Grand Revoyet, 69310 Pierre-Bénite, FRANCE

[d]Institut de Biologie Structurale (IBS), CEA, CNRS, Université Grenoble-Alpes; Laboratoire de Virologie, Centre hospitalier Universitaire Grenoble-Alpes, CS10217, 38043 Grenoble cedex 09, FRANCE

[#]Nathalie Renard : Nathalie.RENARD@biomerieux.com

[#]Mylène Lesénéchal : mylene.lesenechal@biomerieux.com

ABSTRACT

The COVID-19 pandemic, caused by the new severe acute respiratory syndrome coronavirus 2

(SARS-CoV-2), continues to spread worldwide. Serological testing for SARS-CoV-2-specific

antibodies plays an important role in understanding and controlling the pandemics, notably

through epidemiological surveillance. Well validated and highly specific SARS-CoV-2 serological

assays are urgently needed. We describe here the analytical and clinical performance of VIDAS®

SARS-CoV-2 IgM and VIDAS® SARS-CoV-2 IgG, two CE-marked, EUA-authorized, automated,

qualitative assays for the detection of SARS-CoV-2-specific IgM and IgG, respectively. Both

assays showed high within-run and within-laboratory precision (coefficients of variation <

11.0%) and very low cross-reactivity towards sera of patients with a past common coronavirus

or respiratory virus infection. Clinical specificity determined on up to 989 pre-pandemic healthy

donors was ≥ 99% with a narrow 95% confidence interval for both IgM and IgG assays. Clinical

sensitivity was determined on up to 232 samples from 130 RT-PCR-confirmed SARS-CoV-2

patients. The positive percent agreement (PPA) with SARS-CoV-2 PCR reached 100% at ≥ 16 days

(VIDAS® SARS-CoV-2 IgM) and ≥ 32 days (VIDAS® SARS-CoV-2 IgG) of symptom onset. Combined

IgM/IgG test results improved the PPA compared to each test alone. SARS-CoV-2 IgG

seroconversion followed closely that of SARS-CoV-2 IgM and remained stable over time, while

SARS-CoV-2 IgM levels rapidly declined. Interestingly, SARS-CoV-2-specific IgM and IgG

responses were significantly higher in COVID-19 hospitalized vs. non-hospitalized patients.

Altogether, the VIDAS® SARS-CoV-2 IgM and IgG assays are highly specific and sensitive

serological tests suitable for the reliable detection of past acute SARS-CoV-2 infections.

INTRODUCTION

Coronavirus disease 19 (COVID-19) is an infectious disease caused by the newly discovered severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) (1, 2). Within three months of its emergence in China in December 2019, COVID-19 has been declared a global pandemic by the World Health Organisation (WHO). As of October 29, 2020, nearly 45 million COVID-19 cases and 1.2 million deaths have been reported worldwide (3–5). Accurate diagnosis is essential in managing the pandemic, not only to identify, isolate and treat affected patients, but also to characterize the epidemiology of virus transmission and develop national and international surveillance programs. WHO recommends molecular testing of SARS-CoV-2 nucleic acids for acute-phase diagnosis of suspected cases (6–8). Several nucleic acid amplification tests (NAATs), mostly based on quantitative reverse transcriptase polymerase chain reaction (RT-qPCR), have received the Conformité Européenne (CE) mark and have been approved by the United States Food and Drug Administration (FDA) under emergency use authorization (EUA) (9–11). On the other hand, serological testing for SARS-CoV-2-specific antibodies, especially immunoglobulin M (IgM) and immunoglobulin G (IgG), is not recommended as the primary method for the diagnosis of acute cases. It plays however an essential role in the diagnosis of past SARS-CoV-2 infection and in ongoing immunological and epidemiological surveillance. Serological testing might also complement molecular testing to confirm suspected cases not detected by molecular assays, either due to late (> 7 days after infection) or improper sample collection. Finally, serology screening may allow the identification of convalescent plasma donors for use as potential therapy against COVID-19 (9–18).

SARS-CoV-2 serological testing is facing several challenges. Among them, sensitivity and

specificity should be well defined for the target population and validated at different post-

infection time windows. Specificity is particularly critical in the current pandemic phase, as

seroprevalence in the population is still low. In such low-incidence settings, a specificity > 99%

and a narrow 95% confidence interval (95% CI) are required to ensure a high positive predictive

value (PPV) (11, 19, 20). Accordingly, the antigens used to design serology tests should be

properly selected and cross-reactivity with antibodies directed against other antigens, including

from other coronaviruses, should be verified. A huge number of serology assays have been

developed and marketed in the last few months, of which 56 received FDA's EUA (as of October

29, 2020) (21, 22). Clinical performance data of commercial tests are still scant, and examples of

poorly performing tests have even been reported (9). Therefore, there is an urgent need for well

validated and performant serology tests, notably demonstrating very high specificity.

We describe here the analytical and clinical performance of VIDAS® SARS-CoV-2 IgM and VIDAS®

SARS-CoV-2 IgG, two CE-marked and EUA-authorized automated qualitative assays for the

detection of SARS-CoV-2-specific IgM and IgG, respectively, in serum or plasma. Kinetics of

SARS-CoV-2-specific IgM and IgG seroconversion using the VIDAS® SARS-CoV-2 IgM and VIDAS®

SARS-CoV-2 IgG assays were also compared in hospitalized and non-hospitalized COVID-19

patients.

MATERIALS AND METHODS

Patients and samples.

SARS-CoV-2-positive samples were collected after approval by the Ethics Committee RCB 2020-

A00932-37. Informed consent was obtained in accordance with the local regulations. Pre-

pandemic samples (from healthy subjects and from donors with other medical conditions) were

collected in accordance with the Declaration of Helsinki, as revised in 2013. Collected sera and

plasma were stored frozen (< -20°C) until further testing.

Serum from up to 989 healthy pre-pandemic adult donors collected before September 2019 at

two geographical sites (Etablissement Français du Sang (EFS), France; Clinilabs, Inc., United

States) were used to determine the assay specificity of the VIDAS® SARS-CoV-2 IgM, IgG and

combined IgM/IgG test (defined as negative if both VIDAS® SARS-CoV-2 IgM and IgG assays are

negative).

For the evaluation of the positive percent agreement (PPA), 405 serum or plasma samples from

142 symptomatic patients (60 hospitalized, 61 non-hospitalized, 21 of unknown hospitalization

status) diagnosed with COVID-19 and confirmed positive for SARS-CoV-2 by molecular testing

(cobas® SARS-CoV-2, Roche 09175431190 or Real-time RT-PCR assays for the detection of SARS-

CoV-2 Institut Pasteur, Paris (23), performed at the collection site) (Fig. 1) were collected at

three local hospitals (Centre Hospitalier Saint Joseph Saint Luc, Lyon, France; Centre de

Ressources Biologiques [CRB] des Hospices Civils de Lyon, CRB Nord and CRB Sud, Lyon, France)

between March 31 and June 2, 2020. Samples were tested with the VIDAS® SARS-CoV-2 IgM and

IgG assays, and paired measurements were considered for the combined IgM/IgG test results

(defined as positive if at least one of the VIDAS® SARS-CoV-2 IgM and/or IgG assays is positive)

(Fig. 1). The PPA was evaluated according to weekly time frames (0-7, 8-15, 16-23, 24-31, ≥ 32

days) relative to the time from RT-PCR positive result and from symptom onset, when

documented (Fig. 1).

For the evaluation of serum cross-reactivity, up to 276 frozen pre-pandemic sera (i.e. negative

for SARS-CoV-2) collected from patients with other potentially interfering infections or medical

conditions (bioMérieux, Centre Hospitalier Universitaire Grenoble-Alpes and St Joseph St Luc

Lyon collections) were tested with the VIDAS® SARS-CoV-2 IgM (276 sera from 33 medical

conditions; one to 30 sera per condition) and the VIDAS® SARS-CoV-2 IgG (261 sera from 33

medical conditions; two to 30 sera per condition) assays.

Serological assays.

VIDAS® SARS-CoV-2 IgM (423833) and VIDAS® SARS-CoV-2 IgG (423834) (bioMérieux, France)

are automated qualitative CE-IVD assays developed for the VIDAS® family of instruments and

based on a two-step enzyme immunoassay combined with an enzyme-linked fluorescent assay

(ELFA) detection technique. The VIDAS® SARS-CoV-2 IgM and VIDAS® SARS-CoV-2 IgG assays are

intended for use as an aid to determine if individuals may have been exposed and infected by

SARS-CoV-2 and if they have mounted a specific anti-SARS-CoV-2 IgM and IgG immune

response. These assays allow the detection of SARS-CoV-2-specific IgM and IgG, respectively,

from 100 μl serum or plasma (lithium heparin). The VIDAS® SARS-CoV-2 IgM and VIDAS® SARS-

CoV-2 IgG serological assays were conducted according to the manufacturer's instructions.

Briefly, a solid-phase receptacle coated with the antigen (recombinant SARS-CoV-2 receptor-

binding domain [RBD] of the viral Spike protein) serves as both solid phase and pipetting device.

After the sample dilution step, SARS-CoV-2-specific IgM and IgG are captured on the coated

antigen and unbound components are washed out. In a second step, human IgM (VIDAS® SARS-

CoV-2 IgM) or IgG (VIDAS® SARS-CoV-2 IgG) are specifically detected by mouse monoclonal

antibodies conjugated to alkaline phosphatase and directed against human IgM or IgG,

respectively. Unbound components are eliminated by washing and detection is performed by

incubation with the substrate (4-Methyl-umbelliferyl phosphate) followed by measurement of

the fluorescent product (4-Methyl-umbelliferone) at 450 nm. A relative fluorescence value (RFV)

is generated (background reading subtracted from the final fluorescence reading). The assay is

conducted with a standard (S1) and a positive control (C1) that contains humanized

recombinant anti-SARS-CoV-2 antibody, either IgM or IgG depending on the assay. A negative

control (C2) is also supplied. The results are automatically calculated by the instrument,

according to the S1 standard, and an index value (i) is obtained (where $i=RFV_{sample}/RFV_{S1}$). The

test is interpreted as negative when i < 1.00 and positive when i ≥ 1.00. The positivity cut-off

values for the IgM and IgG tests were determined from a healthy pre-pandemic cohort (259

[IgM test] and 120 [IgG test] samples collected prior to August 2019), using non-parametric 99[th]

percentile because of normality rejection for the IgM positivity cut-off and using the (99,99)

tolerance intervals approach after Box-Cox transformation (24) for the IgG positivity cut-off (99[th]

percentile index values at a 99% confidence level) (data not shown).

Statistical analysis.

Assay precision was evaluated according to the Clinical & Laboratory Standards Institute (CLSI) EP05-A3 guideline (25) by the variance component method and restricted maximum likelihood (REML) using the SAS Enterprise Guide 7.13 HF8 software.

Specificity and sensitivity (PPA) estimates were evaluated according to the CLSI EP12-A2 guideline (26). The 95% confidence intervals (95% CI) were computed (either as score confidence interval if the specificity or sensitivity (PPA) belonged to] 5 ; 95[%, or as exact confidence interval otherwise) using the SAS Enterprise Guide 7.13 HF8 software.

PPA was evaluated per time windows (in days) relative to the day of RT-PCR positive result and of symptom onset (when documented). To avoid a statistical bias, only one patient's measurement per time period was included in the analysis. In case of multiple patient's measurements in one period, the first available measurement was considered for the calculation. Therefore, depending on the total number of longitudinal tests performed, each patient contributed with one to five test results in the five time windows considered (0-7, 8-15, 16-23, 24-31, ≥ 32 days).

The positive predictive value (PPV) and the negative predictive value (NPV) were calculated assuming a prevalence of 5%, as recommended by FDA for the EUA application (22), and the respective 95% CI were computed according to Mercaldo et al. (27) using the SAS Enterprise Guide 7.13 HF8 software.

VIDAS® SARS-CoV-2 index values were displayed per time frame as Tukey box plots. Two-group

comparisons of index values per time frame between hospitalized and non-hospitalized patients

were performed using the nonparametric two-tailed Mann-Whitney U-test (MWU-test) with

normal approximation. In case of multiple group comparisons, the Bonferroni method was

applied for controlling the 5% overall probability of a false-significant result. Accordingly, for the

three-group comparison shown in Figure 3, P-values < 0.017 were considered statistically

significant.

RESULTS

Analytical performance of VIDAS® SARS-CoV-2 IgM and IgG assays

Within-run and within-laboratory precisions of the ELFA-based tests were determined using

three samples (one negative and two positive for SARS-CoV-2 IgM and IgG). Samples were run in

triplicate on one VIDAS® instrument, twice a day over 10 days (with an instrument calibration

every second day), using one assay lot, thus generating 60 measurement values per sample. The

coefficient of variation (%CV) for repeatability (within-run precision) did not exceed 9.3% and

5.9% for the VIDAS® SARS-CoV-2 IgM and IgG assays, respectively. The %CV for within-laboratory

precision was also low, reaching a maximum of 10.7% and 6.9% for the VIDAS® SARS-CoV-2 IgM

and IgG assays, respectively (Table S1).

Analytical specificity and sensitivity of the VIDAS® SARS-CoV-2 IgM and IgG assays were verified

through various experiments. First, we ruled out a possible cross-reactivity of the anti-human-

IgM (VIDAS® SARS-CoV-2 IgM) or of the anti-human-IgG (VIDAS® SARS-CoV-2 IgG) with human

IgG or IgM, respectively, which might produce false-positive results. Spike-in experiments in

negative samples using either human recombinant monoclonal anti-SARS-CoV-2 IgG (10 µg/ml)

in the VIDAS® SARS-CoV-2 IgM assay or human recombinant monoclonal anti-SARS-CoV-2 IgM (3

µg/ml) in the VIDAS® SARS-CoV-2 IgG assay demonstrated neither reactivity of the alkaline-

phosphatase-conjugated anti-human-IgM toward human IgG nor reactivity of the anti-human-

IgG toward human IgM (n=10; data not shown). Second, we ruled out a possible competition

between anti-SARS-CoV-2 IgM and IgG for binding to the coated SARS-CoV-2 antigen, which

might interfere with the respective assays and generate false-negative results. Spike-in

experiments in positive samples (with index values ranging from 1.8 to 20.6) using an excess of

human recombinant monoclonal anti-SARS-CoV-2 IgG (10 µg/ml) in the VIDAS® SARS-CoV-2 IgM

assay or of human recombinant monoclonal anti-SARS-CoV-2 IgM (3 µg/ml) in the VIDAS® SARS-

CoV-2 IgG assay did not impact the qualitative test results and resulted in a maximum deviation

from the respective control of 17.6% (n=12; data not shown). Third, we evaluated the impact of

serum inactivation (56°C for 30 minutes), which might be applied by diagnostics laboratories to

inactivate potentially infectious samples (28), on test results of 10 negative and 10 positive

samples. Heat inactivation did not impact the qualitative test results of the VIDAS® SARS-CoV-2

IgM and IgG assays (20/20 = 100% concordance) and the maximum deviation to the (non-

heated) control among 18/20 samples was 17%. However, one negative sample (VIDAS® SARS-

CoV-2 IgG) and one positive sample (VIDAS® SARS-CoV-2 IgM) yielded significantly divergent

index values (data not shown). Therefore heat inactivation of sera prior to VIDAS® SARS-CoV-2

testing should be preferably avoided. Finally, we evaluated the possible cross-reactivity of

components of the assay (SARS-CoV-2 antigen RBD or immunoglobulins) with human sera from

patients with other infections (including other coronaviruses) or medical conditions (e.g.

rheumatoid factor) (29) that might interfere with the assay and yield false-positive results. Up to

276 (VIDAS® SARS-CoV-2 IgM) and 261 (VIDAS® SARS-CoV-2 IgG) sera of SARS-CoV-2-negative

patients with other infections or conditions were tested and the number of positive test results

was evaluated (Table 1). None of the 18 sera of patients with an history of infection with the

human coronaviruses CoV-NL63, CoV-229E, CoV-HKU1 or CoV-OC43 (genera *Alphacoronavirus*

and *Betacoronavirus*) were positive in the VIDAS® SARS-CoV-2 IgG assay, while the serum of one

CoV-NL63-positive patient was positive in the VIDAS® SARS-CoV-2 IgM assay. Only two out of

261 (0.8%) tested sera were positive in the VIDAS® SARS-CoV-2 IgG assay. They belonged to a

HIV-positive and a respiratory syncytial virus (RSV A)-positive patient, respectively. On the other

hand, ten out of 276 (3.6%) tested sera were positive in the VIDAS® SARS-CoV-2 IgM assay. Apart

from the one CoV-NL63-positive sample mentioned above, six sera were from patients

presenting autoantibodies (antinuclear antibody, rheumatoid factor), two were from patients

with an history of parasite infection (*Plasmodium falciparum*, *Trypanosoma cruzi*), and one from

a past Rhinovirus/Enterovirus infection. The index values associated with these 12 crossreactive

sera was low, with a median (interquartile range) of 2.1 (1.4-3.8). None of the sera from

patients infected with other respiratory viruses, including influenza virus, parainfluenza virus,

metapneumovirus or adenovirus were reactive.

Clinical performance of VIDAS® SARS-COV-2 IgM and IgG assays

The clinical specificity of the VIDAS® SARS-CoV-2 IgM and IgG assays were assessed using sera

from up to 989 pre-pandemic healthy volunteers collected in France and in the United States

before September 2019. A total of 308 sera were tested with the VIDAS® SARS-CoV-2 IgM assay

and 989 sera were tested with the VIDAS® SARS-CoV-2 IgG assay. The combined IgM/IgG assay

specificity (defined as both VIDAS® SARS-CoV-2 IgM and IgG assays being negative) was

determined on 308 paired VIDAS® SARS-CoV-2 IgM and IgG tests (Tables S2 and S3). 306/308

(VIDAS® SARS-CoV-2 IgM) and 988/989 (VIDAS® SARS-CoV-2 IgG) SARS-CoV-2-negative sera were

negative, corresponding to a specificity (95% CI) of 99.4% (97.7-99.9%) and 99.9% (99.4-100%)

for the VIDAS® SARS-CoV-2 IgM and IgG assays, respectively (Table S3). The specificity (95% CI)

of the VIDAS® SARS-CoV-2 IgG test on the common cohort (N=308) was 100.0% (98.8-100.0%).

The specificity (95% CI) of the combined IgM and IgG serology tests (306/308 tests negative in

both assays) was 99.4% (97.7-99.9%) (Table S3).

The clinical sensitivity of the VIDAS® SARS-CoV-2 IgM and IgG assays was assessed using 405

samples collected from 142 patients confirmed positive for SARS-CoV-2 by RT-PCR. The positive

percent agreement (PPA) of the serology tests with the RT-PCR test results was calculated per

weekly time windows (0-7, 8-15, 16-23, 24-31 and ≥ 32 days) relative to the time from the PCR-

positive result and to the time from symptom onset. No more than one patient sample per time

window was included in the calculation (Fig. 1 and Table 2). The PPA calculated on all available

samples for the VIDAS® SARS-CoV-2 IgM and IgG assays are shown in Tables S4 and S5,

respectively. For the sake of comparability, the PPA (95% CI) of the IgM, IgG and combined

IgM/IgG test results was also calculated on samples evaluated with both tests (paired testing;

Tables 3, 4 and S6).

The PPA from the time of RT-PCR-positive test results for the VIDAS® SARS-CoV-2 IgM assay

raised from 44.5% at 0-7 days to 100.0% at 24-31 days, before decreasing to 81.8% at ≥ 32 days

(Table 3). The PPA for the VIDAS® SARS-CoV-2 IgG assay raised from 45.5% at 0-7 days to 100.0%

from day 24 onwards(Table 3). The PPA of the combined IgM and IgG serology tests (positive in

at least one of the IgM and/or IgG assays) from the time of RT-PCR-positive test results,

increased from 53.6% at 0-7 days to 100.0% from day 16 onwards (Tables 3 and S6). The PPA

evaluated relative to the time of symptom onset raised from 31.8% at 0-7 days (VIDAS® SARS-

CoV-2 IgM and IgG) to 100.0% from day 16 (VIDAS® SARS-CoV-2 IgM) or day 24 (VIDAS® SARS-

CoV-2 IgG) (Table 4). The PPA of the combined IgM and IgG serology test results from the time

of symptom onset, increased from 36.4% at 0-7 days to 100.0% from day 16 onwards (Tables 4

and S6). It should be noted that the PPA relative to the date of symptom onset is mainly based

on hospitalized patients (Table 2).

Based on the specificity and the PPA determined on paired IgM and IgG testing (N=308 for

specificity; N=105 for PPA), the negative predictive value (NPV) and the positive predictive value

(PPV) were calculated at 5% prevalence (22), according to the time after symptom onset (Table

5). The NPV was high for both the VIDAS® SARS-CoV-2 IgM and IgG assays, whether considered

alone or in combination; NPV was ≥ 96.5% (lower 95% confidence limits ≥ 95.4%) at 0-7 days

post symptom onset and NPV increased from 99.1% to 100.0% (lower 95% confidence limits ≥

98.0%) from day 8 onwards following symptom onset. The PPV of the VIDAS® SARS-CoV-2 IgG

assay was 100% at all time frames considered. Of note, the PPV calculated using the full data set

for the VIDAS® SARS-CoV-2 IgG specificity determination (N=989) was slightly lower, increasing

from 94.3% at 0-7 days to 98.1% at ≥ 32 days (Table S7). The PPV of the VIDAS® SARS-CoV-2 IgM

assay was lower, ranging from 72.1% at 0-7 days to 89.0% from day 16 onwards following

symptom onset (Table 5). The combination of IgM and IgG test results slightly improved the PPV

and NPV of the SARS-CoV-2 IgM assay, and the NPV of the SARS-CoV-2 IgG assay. The SARS-CoV-

2 IgG assay performed best alone in terms of PPV (Table 5).

Longitudinal study of IgM and IgG seroconversion in hospitalized COVID-19 patients

The global distribution of VIDAS® SARS-CoV-2 IgM and IgG index values post symptom onset was

compared among the 105 SARS-CoV-2-positive patients described in Table 4 (i.e. mainly

including hospitalized COVID-19 patients; Table 2) (Fig. 2). IgM index values increased from the

second week of symptom onset (8-15 days) and peaked during the third week (16-23 days)

before decreasing. In comparison, the IgG index values strongly increased from the second week

of symptom onset and seemed to reach a plateau ≥ 32 days post symptom onset (Fig. 2).

The VIDAS® SARS-CoV-2 IgM and IgG index values were also compared between hospitalized

and non-hospitalized patients. Since the date of symptom onset was not documented in non-

hospitalized patients (Table 2), the index values in 54 hospitalized and 61 non-hospitalized

patients were compared relative to the time of the RT-PCR-positive results (Fig. 3). Interestingly,

the distribution of index values between hospitalized and non-hospitalized patients differ

statistically from each other for both the VIDAS® SARS-CoV-2 IgM and IgG assays at the three

compared time frames (0-7, 8-15 and 16-23 days post PCR-positive test; MWU-test p-values <

0.017). Median index values were higher in hospitalized versus non-hospitalized patients (Fig.

3).

Finally, IgM and IgG seroconversion was further investigated in four selected hospitalized

patients with either early and/or repeated measurements over an extended period of time (up

to 74 days post symptom onset; Fig. 4). IgG seroconversion closely followed IgM seroconversion

in the second week of symptom onset (Fig. 4A-B), in line with the global profile shown in Figure

2. SARS-CoV-2 IgM index rapidly decreased concomitantly with the increase of SARS-CoV-2 IgG

index (Fig. 4B-D). In the three patients shown in Fig. 4B-D, SARS-CoV-2 IgM index decreased

below the positivity cutoff (index = 1.00) 46 days after symptom onset. In contrast, SARS-CoV-2

IgG index values remained high and stable from approximately day 20 onward after symptom

onset, at least up to 74 days.

DISCUSSION

We describe here the analytical and clinical performance of the VIDAS® SARS-CoV-2 IgM and IgG assays. We demonstrate that both assays show high precision, and excellent analytical and clinical performances.

The rate of cross-reactivity with non-specific antibodies, including those of patients infected with other coronaviruses, was very low in both VIDAS® SARS-CoV-2 IgM and IgG assays. This weak cross-reactivity with other coronaviruses antibodies is likely due, at least in part, to the choice of the receptor-binding domain (RBD) of the viral Spike protein as SARS-CoV-2-specific antigen. The RBD shows a high sensitivity in ELISA, higher than that of the SARS-CoV-2 spike S1 or nucleocapsid (NC) antigens (30–33). It also presents a weaker homology and significant structural divergences with the RBD of other coronaviruses (34–36). Another advantage of using the RBD is that the viral antigen generates neutralizing antibodies likely to provide protective immunity (35–42), as previously demonstrated for SARS-CoV (43, 44). That the VIDAS® SARS-CoV-2 assays have the ability to detect SARS-CoV-2 neutralizing antibodies was recently demonstrated in mild COVID-19 patients, with an almost perfect concordance (Cohen's Kappa coefficient of 0.9) between the VIDAS® SARS-CoV-2 IgG assay and a virus neutralization test (42). Beside its strong immunogenicity and antigenicity, the RBD of SARS-CoV has been shown to elicit antibody responses that persisted many years after infection (38, 44), raising the possibility that it might also be the case for the RBD of SARS-CoV-2. Recent studies in COVID-19 patients, notably in convalescent donors (45–47), on the anti-RBD antibody dynamics post SARS-CoV-2 infection (39, 40), or demonstrating the persistence and expansion of SARS-CoV-2-specific

memory lymphocytes (48), as well as the stability of the IgG response detected with the VIDAS®

SARS-CoV-2 IgG assay up to 74 days post-symptom onset in the present study, strongly support

this proposition. Hence, a serology test such as the VIDAS® SARS-CoV-2 IgG assay is likely to be

suitable for the detection of protective immunity and the evaluation of the efficacy of future

vaccines, which are mainly based on the RBD-containing Spike protein (49, 50).

The low cross-reactivity rate with non-specific sera probably explains the very high specificity (≥

99%) and narrow 95% CI of both VIDAS® SARS-CoV-2 IgM and IgG assays. The VIDAS® SARS-CoV-2

IgG assay alone had a specificity close to 100%, slightly higher than that of the VIDAS® SARS-CoV-

2 IgM assay.

The clinical sensitivity of the VIDAS® SARS-CoV-2 assays was evaluated in SARS-CoV-2-confirmed

symptomatic cases and was determined as positive percent agreement (PPA) with the RT-PCR

assay, at successive time frames post positive PCR and, alternatively, post symptom onset. The

PPA reached 100% at 16-23 days (VIDAS® SARS-CoV-2 IgM) and at ≥ 32 days (VIDAS® SARS-CoV-2

IgG) post-symptom onset. The combined VIDAS® SARS-CoV-2 IgM/IgG test evaluation improved

the PPA of the respective IgM and IgG tests by 3.5 to 6.9 percent points during the first two

weeks (0-7 and 8-15 days) of symptom onset. Such improved sensitivity of the combined

IgM/IgG tests early after symptom onset might also be useful for the diagnosis of suspected

COVID-19 cases with negative PCR (13–15, 17, 18).

Overall, the clinical performance of the VIDAS® SARS-CoV-2 assays was excellent. It was in a

range comparable to that reported for existing EUA serological assays (22, 33, 42, 51–57).

Moreover, in several side-by-side comparisons of six to nine commercial serological assays

(specific for SARS-CoV-2 IgA, IgM, IgG or total antibodies), VIDAS® SARS-CoV-2 IgG outperformed

some of the IgG-specific competitor assays in terms of specificity and/or PPA with PCR positivity

(42, 56, 57). The high specificity of the VIDAS® SARS-CoV-2 IgG assay alone should be well suited

for epidemiological surveillance.

The kinetics of SARS-CoV-2 IgM and IgG seroconversion was also evaluated by monitoring

VIDAS® index values over time. VIDAS® SARS-CoV-2 IgM and IgG index values increased in the

second week after symptom onset. IgG index values strongly increased and remained high, as

IgM index values rapidly declined. These profiles are in agreement with those described in

recent publications (13, 30, 39, 40, 58–62). Interestingly, the magnitude of the antibody

response (index values) correlated with disease severity, as it was significantly higher in

hospitalized vs. non-hospitalized COVID-19 patients at the time frames investigated (0-7, 18-15

and 16-23 days after a PCR-positive test). This observation is in agreement with published

reports (11, 40, 62–65). It should be noted that despite the significantly lower response of sera

from mild (non-hospitalized) COVID-19 patients in the VIDAS® SARS-CoV-2 assays (Fig. 3), the

good performance of the assays as well as their strong concordance with seroneutralization was

recently demonstrated in a cohort of mild COVID-19 patients (42).

This study presents several limitations. First, assay sensitivity was evaluated on confirmed but

not on suspected SARS-CoV-2 cases (i.e. patients with symptoms but negative by PCR). It would

be interesting to evaluate and confirm the benefit of SARS-CoV-2 IgM and IgG serology to

complement PCR testing (13, 14, 17, 18). On the other hand, recent reports suggested that the

identification rate of false-negative PCR results using serology testing might be marginal,

between ~1% (20) and ~4% (62). Second, assay sensitivity was determined on symptomatic

(hospitalized and non-hospitalized) COVID-19 patients. The sensitivity of the VIDAS® SARS-CoV-2

IgM and IgG assays in asymptomatic SARS-CoV-2-infected individuals, who may represent most

of the infected patients, remains to be evaluated.

In conclusion, VIDAS® SARS-CoV-2 IgM and IgG are highly sensitive and specific assays for the

reliable screening of patients after acute SARS-CoV-2 infections (and likely after vaccination,

when available). Moreover, the VIDAS® SARS-CoV-2 IgG assay fulfils the specificity requirement

for its use in seroepidemiology studies and is well suited for the detection of past SARS-CoV-2

infections. Further studies are necessary to confirm its suitability for the detection of SARS-CoV-

2 neutralizing antibodies and to define correlates of immune protection.

ACKNOWLEDGMENTS

This work was supported by bioMérieux. We thank the Etablissement Français du Sang, notably Yves Mérieux, for providing samples. We thank all the members of R&D Immunoassay bioMérieux who took part in this work, i.e. the bioMolecule Engineering team, Prototyping, Development and Verification teams. We warmly thank Nadia Piga and the biobank team for collecting the samples in a short time frame. We are grateful to the bioMérieux Data Sciences team for the calculations and statistical evaluation of the data, and to Victor Bondanese, Laurence Bridon and the Clinical Affairs team of bioMérieux for their help and technical support. We thank Karen Brengel for her useful advices as well as Alice Banz for critically reading the manuscript. The authors thank Dr. Anne Rascle of AR Medical Writing (Regensburg, Germany) for providing medical writing support, which was funded by bioMérieux (Marcy L'Etoile, France) in accordance with Good Publication Practice (GPP3) guidelines (http://www.ismpp.org/gpp3).

Conflict of interests' statement

JL declares received research funding from bioMérieux for this study.

MP declares a consulting contract with bioMérieux.

Contributions

NR, SD, ML, CT, FB conceived the study; CP, GG, PB, MC conceived the study and performed the serological assays. FR, SP, MP, JL collected patients' samples; NC, IC, NR, CT, ML, interpreted

401   the serological assays; all authors contributed to data acquisition, data analysis, and/or data

402   interpretation, and all authors reviewed and approved the final manuscript.

REFERENCES

1. Zhou P, Yang X-L, Wang X-G, Hu B, Zhang L, Zhang W, Si H-R, Zhu Y, Li B, Huang C-L, Chen H-D, Chen J, Luo Y, Guo H, Jiang R-D, Liu M-Q, Chen Y, Shen X-R, Wang X, Zheng X-S, Zhao K, Chen Q-J, Deng F, Liu L-L, Yan B, Zhan F-X, Wang Y-Y, Xiao G-F, Shi Z-L. 2020. A pneumonia outbreak associated with a new coronavirus of probable bat origin. Nature 579:270–273.

2. Coronaviridae Study Group of the International Committee on Taxonomy of Viruses. 2020. The species Severe acute respiratory syndrome-related coronavirus: classifying 2019-nCoV and naming it SARS-CoV-2. Nat Microbiol 5:536–544.

3. World Health Organization. 2020. Coronavirus Disease (COVID-19) Situation Reports. https://www.who.int/emergencies/diseases/novel-coronavirus-2019/situation-reports. Accessed: October 29, 2020.

4. Dong E, Du H, Gardner L. 2020. An interactive web-based dashboard to track COVID-19 in real time. Lancet Infect Dis 20:533–534.

5. Johns Hopkins University. 2020. COVID-19 Resource Center. https://coronavirus.jhu.edu/map.html. Accessed: October 29, 2020.

6. World Health Organization. 2020. Laboratory testing for 2019 novel coronavirus (2019-nCoV) in suspected human cases. https://www.who.int/publications-detail-redirect/10665-331501. Accessed: October 29, 2020.

7. World Health Organization. 2020. Global surveillance for COVID-19 caused by human infection with COVID-19 virus: interim guidance. https://www.who.int/publications-detail-

redirect/global-surveillance-for-covid-19-caused-by-human-infection-with-covid-19-virus-

interim-guidance. Accessed: October 29, 2020.

8. World Health Organization. 2020. Laboratory testing strategy recommendations for COVID-

19: interim guidance. https://www.who.int/publications-detail-redirect/laboratory-testing-

strategy-recommendations-for-covid-19-interim-guidance. Accessed: October 29, 2020.

9. Vashist SK. 2020. In Vitro Diagnostic Assays for COVID-19: Recent Advances and Emerging

Trends. Diagnostics 10:202.

10. Venter M, Richter K. 2020. Towards effective diagnostic assays for COVID-19: a review. J Clin

Pathol jclinpath-2020-206685.

11. Bohn MK, Lippi G, Horvath A, Sethi S, Koch D, Ferrari M, Wang C-B, Mancini N, Steele S,

Adeli K. 2020. Molecular, serological, and biochemical diagnosis and monitoring of COVID-

19: IFCC taskforce evaluation of the latest evidence. Clinical Chemistry and Laboratory

Medicine (CCLM) 58:1037–1052.

12. Winter AK, Hegde ST. 2020. The important role of serology for COVID-19 control. The Lancet

Infectious Diseases S1473309920303224.

13. Guo L, Ren L, Yang S, Xiao M, Chang D, Yang F, Dela Cruz CS, Wang Y, Wu C, Xiao Y, Zhang L,

Han L, Dang S, Xu Y, Yang Q, Xu S, Zhu H, Xu Y, Jin Q, Sharma L, Wang L, Wang J. 2020.

Profiling Early Humoral Response to Diagnose Novel Coronavirus Disease (COVID-19). Clin

Infect Dis https://doi.org/10.1093/cid/ciaa310.

14. Zhao R, Li M, Song H, Chen J, Ren W, Feng Y, Gao GF, Song J, Peng Y, Su B, Guo X, Wang Y,

Chen J, Li J, Sun H, Bai Z, Cao W, Zhu J, Zhang Q, Sun Y, Sun S, Mao X, Su J, Chen X, He A, Gao

W, Jin R, Jiang Y, Sun L. 2020. Early detection of SARS-CoV-2 antibodies in COVID-19 patients

as a serologic marker of infection. Clinical Infectious Diseases ciaa523.

15. Caruana G, Croxatto A, Coste AT, Opota O, Lamoth F, Jaton K, Greub G. 2020. Diagnostic

strategies for SARS-CoV-2 infection and interpretation of microbiological results. Clinical

Microbiology and Infection 26:1178–1182.

16. Pancrazzi A, Magliocca P, Lorubbio M, Vaggelli G, Galano A, Mafucci M, Duranti D, Cortesi

M, Mazzeschi E, Fabbroni S, Viti G, Tartaglia Polcini A, Tripodo E, Sanchini P, Gervino S,

Tacconi D, Dei S, Mazzierli M, D'Urso A, Ognibene A. 2020. Comparison of serologic and

molecular SARS-CoV 2 results in a large cohort in Southern Tuscany demonstrates a role for

serologic testing to increase diagnostic sensitivity. Clinical Biochemistry

S0009912020307918.

17. Zhang J, Zhang X, Liu J, Ban Y, Li N, Wu Y, Liu Y, Ye R, Liu J, Li X, Li L, Qin X, Zheng R. 2020.

Serological detection of 2019-nCoV respond to the epidemic: A useful complement to

nucleic acid testing. Int Immunopharmacol 88:106861.

18. Liu R, Liu X, Yuan L, Han H, Shereen MA, Zhen J, Niu Z, Li D, Liu F, Wu K, Luo Z, Zhu C. 2020.

Analysis of adjunctive serological detection to nucleic acid test for severe acute respiratory

syndrome coronavirus 2 (SARS-CoV-2) infection diagnosis. Int Immunopharmacol

86:106746.

19. Bryant JE, Azman AS, Ferrari MJ, Arnold BF, Boni MF, Boum Yap, Hayford K, Luquero FJ, Mina

MJ, Rodriguez-Barraquer I, Wu JT, Wade D, Vernet G, Leung DT. 2020. Serology for SARS-

CoV-2: Apprehensions, opportunities, and the path forward. Sci Immunol 5:eabc6347.

20. Farnsworth CW, Anderson NW. 2020. SARS-CoV-2 Serology: Much Hype, Little Data. Clinical

Chemistry hvaa107.

21. U.S. Food and Drug Administration. 2020. In Vitro Diagnostics EUAs.

https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-

authorizations-medical-devices/vitro-diagnostics-euas. Accessed: October 29, 2020.

22. U.S. Food and Drug Administration. 2020. EUA Authorized Serology Test Performance.

https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-

authorizations-medical-devices/eua-authorized-serology-test-performance. Accessed:

October 29, 2020.

23. World Health Organization. Real-time RT-PCR assays for the detection of SARS-CoV-2 Institut

Pasteur, Paris. https://www.who.int/nepal/activities/supporting-elimination-of-kala-azar-as-

a-public-health-problem/docs/default-source/coronaviruse/real-time-rt-pcr-assays-for-the-

detection-of-sars-cov-2-institut-pasteur-paris. Accessed: October 29, 2020.

24. Dong X, Tsong Y, Shen M. 2015. Statistical considerations in setting product specifications. J

Biopharm Stat 25:280–294.

25. Clinical & Laboratory Standards Institute. EP05-A3: Evaluating Quantitative Measurement

Precision,   3rd   Edition.   https://clsi.org/standards/products/method-

evaluation/documents/ep05/. Accessed: October 29, 2020.

26. Clinical & Laboratory Standards Institute. EP12-A2: User Protocol for Evaluation of

Qualitative Test Performance, 2nd Edition. https://clsi.org/standards/products/method-

evaluation/documents/ep12/. Accessed: October 29, 2020.

27. Mercaldo ND, Lau KF, Zhou XH. 2007. Confidence intervals for predictive values with an

emphasis to case-control studies. Stat Med 26:2170–2183.

28. Pastorino B, Touret F, Gilles M, de Lamballerie X, Charrel RN. 2020. Heat Inactivation of

Different Types of SARS-CoV-2 Samples: What Protocols for Biosafety, Molecular Detection

and Serological Diagnostics? Viruses 12:735.

29. Tate J, Ward G. 2004. Interferences in immunoassay. Clin Biochem Rev 25:105–120.

30. Ma H, Zeng W, He H, Zhao D, Yang Y, Jiang D, Zhou P, Qi Y, He W, Zhao C, Yi R, Wang X,

Wang B, Xu Y, Yang Y, Kombe AJK, Ding C, Xie J, Gao Y, Cheng L, Li Y, Ma X, Jin T. 2020.

COVID-19 diagnosis and study of serum SARS-CoV-2 specific IgA, IgM and IgG by

chemiluminescence immunoanalysis. medRxiv 2020.04.17.20064907.

31. Okba NMA, Müller MA, Li W, Wang C, GeurtsvanKessel CH, Corman VM, Lamers MM,

Sikkema RS, de Bruin E, Chandler FD, Yazdanpanah Y, Le Hingrat Q, Descamps D, Houhou-

Fidouh N, Reusken CBEM, Bosch B-J, Drosten C, Koopmans MPG, Haagmans BL. 2020.

Severe Acute Respiratory Syndrome Coronavirus 2–Specific Antibody Responses in

Coronavirus Disease 2019 Patients. Emerg Infect Dis 26.

32. Liu W, Liu L, Kou G, Zheng Y, Ding Y, Ni W, Wang Q, Tan L, Wu W, Tang S, Xiong Z, Zheng S.

2020. Evaluation of Nucleocapsid and Spike Protein-Based Enzyme-Linked Immunosorbent

Assays for Detecting Antibodies against SARS-CoV-2. Journal of Clinical Microbiology

58:e00461-20.

33. Lassaunière R, Frische A, Harboe ZB, Nielsen AC, Fomsgaard A, Krogfelt KA, Jørgensen CS.

2020. Evaluation of nine commercial SARS-CoV-2 immunoassays. medRxiv

2020.04.09.20056325.

34. Chen Y, Guo Y, Pan Y, Zhao ZJ. 2020. Structure analysis of the receptor binding of 2019-

nCoV. Biochemical and Biophysical Research Communications 525:135–140.

35. Ju B, Zhang Q, Ge J, Wang R, Sun J, Ge X, Yu J, Shan S, Zhou B, Song S, Tang X, Yu J, Lan J,

Yuan J, Wang H, Zhao J, Zhang S, Wang Y, Shi X, Liu L, Zhao J, Wang X, Zhang Z, Zhang L.

2020. Human neutralizing antibodies elicited by SARS-CoV-2 infection. Nature

https://doi.org/10.1038/s41586-020-2380-z.

36. Premkumar L, Segovia-Chumbez B, Jadi R, Martinez DR, Raut R, Markmann A, Cornaby C,

Bartelt L, Weiss S, Park Y, Edwards CE, Weimer E, Scherer EM, Rouphael N, Edupuganti S,

Weiskopf D, Tse LV, Hou YJ, Margolis D, Sette A, Collins MH, Schmitz J, Baric RS, de Silva AM.

2020. The receptor binding domain of the viral spike protein is an immunodominant and

highly specific target of antibodies in SARS-CoV-2 patients. Sci Immunol 5:eabc8413.

37. Shi R, Shan C, Duan X, Chen Z, Liu P, Song J, Song T, Bi X, Han C, Wu L, Gao G, Hu X, Zhang Y,

Tong Z, Huang W, Liu WJ, Wu G, Zhang B, Wang L, Qi J, Feng H, Wang F, Wang Q, Gao GF,

Yuan Z, Yan J. 2020. A human neutralizing antibody targets the receptor binding site of

SARS-CoV-2. Nature 584:120–124.

38. Chia WN, Tan CW, Foo R, Kang AEZ, Peng Y, Sivalingam V, Tiu C, Ong XM, Zhu F, Young BE,

Chen MI-C, Tan Y-J, Lye DC, Anderson DE, Wang L-F. 2020. Serological differentiation

between COVID-19 and SARS infections. Emerging Microbes & Infections 9:1497–1505.

39. Grandjean L, Saso A, Ortiz A, Lam T, Hatcher J, Thistlethwaite R, Harris M, Best T, Johnson M,

Wagstaffe H, Ralph E, Mai A, Colijn C, Breuer J, Buckland M, Gilmour K, Goldblatt D, Team

TC-SS. 2020. Humoral Response Dynamics Following Infection with SARS-CoV-2. medRxiv

2020.07.16.20155663.

40. Crawford KHD, Dingens AS, Eguia R, Wolf CR, Wilcox N, Logue JK, Shuey K, Casto AM, Fiala B,

Wrenn S, Pettie D, King NP, Greninger AL, Chu HY, Bloom JD. 2020. Dynamics of neutralizing

antibody titers in the months after SARS-CoV-2 infection. J Infect Dis jiaa618.

41. McAndrews KM, Dowlatshahi DP, Dai J, Becker LM, Hensel J, Snowden LM, Leveille JM,

Brunner MR, Holden K, Hopkins NS, Harris A, Kumpati JJ, Whitt MA, Lee JJ, Ostrosky-

Zeichner L, Papanna R, LeBleu V, Allison J, Kalluri R. 2020. Heterogeneous antibodies against

SARS-CoV-2 spike receptor binding domain and nucleocapsid with implications on COVID-19

immunity. JCI Insight https://doi.org/10.1172/jci.insight.142386.

42. Bal A, Pozzetto B, Trabaud M-A, Escuret V, Rabilloud M, Langlois-jacques C, Paul A, Guibert

N, D'aubarde C, Massardier A, Boibieux A, Morfin F, Pitiot V, Gueyffier F, Lina B, Fassier JB,

Assant S. 2020. Evaluation of high-throughput SARS-CoV-2 serological assays in a

longitudinal cohort of mild COVID-19 patients: sensitivity, specificity and association with

virus neutralization test. medRxiv 2020.09.30.20194290.

43. He Y, Zhu Q, Liu S, Zhou Y, Yang B, Li J, Jiang S. 2005. Identification of a critical neutralization

determinant of severe acute respiratory syndrome (SARS)-associated coronavirus:

importance for designing SARS vaccines. Virology 334:74–82.

44. Cao Z, Liu L, Du L, Zhang C, Jiang S, Li T, He Y. 2010. Potent and persistent antibody

responses against the receptor-binding domain of SARS-CoV spike protein in recovered

patients. Virol J 7:299.

45. Li L, Tong X, Chen H, He R, Lv Q, Yang R, Zhao L, Wang J, Xu H, Liu C, Chen G, Chen S, Li C,

Qiao J, Yang J, Wu Y, Liu Z. 2020. Characteristics and serological patterns of COVID-19

convalescent plasma donors: optimal donors and timing of donation. Transfusion

https://doi.org/10.1111/trf.15918.

46. Salazar E, Kuchipudi SV, Christensen PA, Eagar TN, Yi X, Zhao P, Jin Z, Long SW, Olsen RJ,

Chen J, Castillo B, Leveque C, Towers DM, Lavinder J, Gollihar JD, Cardona J, Ippolito GC,

Nissly RH, Bird IM, Greenawalt D, Rossi RM, Gontu A, Srinivasan S, Poojary IB, Cattadori IM,

Hudson PJ, Joselyn N, Prugar L, Huie K, Herbert A, Bernard DW, Dye J, Kapur V, Musser JM.

2020. Relationship between Anti-Spike Protein Antibody Titers and SARS-CoV-2 In Vitro

559 Virus Neutralization in Convalescent Plasma. bioRxiv

560 https://doi.org/10.1101/2020.06.08.138990.

561 47. Yao X-Y, Liu W, Li Z-Y, Xiong H-L, Su Y-Y, Li T, Zhang S-Y, Zhang X-J, Bi Z-F, Deng C-X, Li C-Y,

562 Yuan Q, Zhang J, Zhang T-Y, Wang Z-X, Ge S, Ningshao X. 2020. Neutralizing and binding

563 antibody kinetics of COVID-19 patients during hospital and convalescent phases. medRxiv

564 2020.07.18.20156810.

565 48. Pepper M, Rodda L, Netland J, Shehata L, Pruner K, Morawski P, Thouvenel C, Takahara K,

566 Eggenberger J, Hemann E, Waterman H, Fahning M, Chen Y, Rathe J, Stokes C, Wrenn S,

567 Fiala B, Carter L, Hamerman J, King N, Gale M, Campbell D, Rawlings D. 2020. Functional

568 SARS-CoV-2-specific immune memory persists after mild COVID-19. Res Sq

569 https://doi.org/10.21203/rs.3.rs-57112/v1.

570 49. Dagotto G, Yu J, Barouch DH. 2020. Approaches and Challenges in SARS-CoV-2 Vaccine

571 Development. Cell Host Microbe https://doi.org/10.1016/j.chom.2020.08.002.

572 50. Amanat F, Krammer F. 2020. SARS-CoV-2 Vaccines: Status Report. Immunity 52:583–589.

573 51. Bryan A, Pepper G, Wener MH, Fink SL, Morishima C, Chaudhary A, Jerome KR, Mathias PC,

574 Greninger AL. 2020. Performance Characteristics of the Abbott Architect SARS-CoV-2 IgG

575 Assay and Seroprevalence in Boise, Idaho. J Clin Microbiol

576 https://doi.org/10.1128/JCM.00941-20.

577 52. Tang MS, Hock KG, Logsdon NM, Hayes JE, Gronowski AM, Anderson NW, Farnsworth CW.

578 2020. Clinical Performance of Two SARS-CoV-2 Serologic Assays. Clinical Chemistry hvaa120.

53. Perkmann T, Perkmann-Nagele N, Breyer M-K, Breyer-Kohansal R, Burghuber OC, Hartl S,

Aletaha D, Sieghart D, Quehenberger P, Marculescu R, Mucher P, Strassl R, Wagner OF,

Binder CJ, Haslacher H. 2020. Side by side comparison of three fully automated SARS-CoV-2

antibody assays with a focus on specificity. Clin Chem

https://doi.org/10.1093/clinchem/hvaa198.

54. Herroelen PH, Martens GA, De Smet D, Swaerts K, Decavele A-S. 2020. Humoral Immune

Response to SARS-CoV-2: Comparative Clinical Performance of Seven Commercial Serology

Tests. American Journal of Clinical Pathology aqaa140.

55. Pflüger LS, Bannasch JH, Brehm TT, Pfefferle S, Hoffmann A, Nörz D, van der Meirschen M,

Kluge S, Haddad M, Pischke S, Hiller J, Addo MM, Lohse AW, Wiesch JS zur, Peine S,

Aepfelbacher M, Lütgehetmann M. 2020. Clinical evaluation of five different automated

SARS-CoV-2 serology assays in a cohort of hospitalized COVID-19 patients. Journal of Clinical

Virology 130:104549.

56. Wolff F, Dahma H, Duterme C, Van den Wijngaert S, Vandenberg O, Cotton F, Montesinos I.

2020. Monitoring antibody response following SARS-CoV-2 infection: diagnostic efficiency of

4 automated immunoassays. Diagn Microbiol Infect Dis 98:115140.

57. Trabaud M-A, Icard V, Milon M-P, Bal A, Lina B, Escuret V. 2020. Comparison of eight

commercial, high-throughput, automated or ELISA assays detecting SARS-CoV-2 IgG or total

antibody. J Clin Virol 132:104613.

58. Xiao AT, Gao C, Zhang S. 2020. Profile of specific antibodies to SARS-CoV-2: The first report. J

Infect 81:147–178.

59. Padoan A, Cosma C, Sciacovelli L, Faggian D, Plebani M. 2020. Analytical performances of a

chemiluminescence immunoassay for SARS-CoV-2 IgM/IgG and antibody kinetics. Clinical

Chemistry and Laboratory Medicine (CCLM) 1.

60. Suhandynata RT, Hoffman MA, Kelner MJ, McLawhon RW, Reed SL, Fitzgerald RL. 2020.

Longitudinal Monitoring of SARS-CoV-2 IgM and IgG Seropositivity to Detect COVID-19. The

Journal of Applied Laboratory Medicine jfaa079.

61. Sun B, Feng Y, Mo X, Zheng P, Wang Q, Li P, Peng P, Liu X, Chen Z, Huang H, Zhang F, Luo W,

Niu X, Hu P, Wang L, Peng H, Huang Z, Feng L, Li F, Zhang F, Li F, Zhong N, Chen L. 2020.

Kinetics of SARS-CoV-2 specific IgM and IgG responses in COVID-19 patients. Emerg

Microbes Infect 9:940–948.

62. Long Q-X, Liu B-Z, Deng H-J, Wu G-C, Deng K, Chen Y-K, Liao P, Qiu J-F, Lin Y, Cai X-F, Wang

D-Q, Hu Y, Ren J-H, Tang N, Xu Y-Y, Yu L-H, Mo Z, Gong F, Zhang X-L, Tian W-G, Hu L, Zhang

X-X, Xiang J-L, Du H-X, Liu H-W, Lang C-H, Luo X-H, Wu S-B, Cui X-P, Zhou Z, Zhu M-M, Wang

J, Xue C-J, Li X-F, Wang L, Li Z-J, Wang K, Niu C-C, Yang Q-J, Tang X-J, Zhang Y, Liu X-M, Li J-J,

Zhang D-C, Zhang F, Liu P, Yuan J, Li Q, Hu J-L, Chen J, Huang A-L. 2020. Antibody responses

to SARS-CoV-2 in patients with COVID-19. Nat Med 26:845–848.

63. Yongchen Z, Shen H, Wang X, Shi X, Li Y, Yan J, Chen Y, Gu B. 2020. Different longitudinal

patterns of nucleic acid and serology testing results based on disease severity of COVID-19

patients. Emerging Microbes & Infections 9:833–836.

64. Wang X, Guo X, Xin Q, Pan Y, Hu Y, Li J, Chu Y, Feng Y, Wang Q. 2020. Neutralizing Antibodies

Responses to SARS-CoV-2 in COVID-19 Inpatients and Convalescent Patients. Clin Infect Dis

https://doi.org/10.1093/cid/ciaa721.

65. Hansen CB, Jarlhelt I, Pérez-Alós L, Hummelshøj Landsy L, Loftager M, Rosbjerg A,

Helgstrand C, Bjelke JR, Egebjerg T, Jardine JG, Værke Jørgensen C, Iversen K, Bayarri-Olmos

R, Garred P, Skjoedt M-O. 2020. SARS-CoV-2 antibody responses determine disease severity

in COVID-19 infected individuals. medRxiv 2020.07.27.20162321.

 FIGURE LEGENDS

Figure 1. Study flow diagram. Description of SARS-CoV-2-positive samples used to determine

the positive percent agreement relative to the time of RT-PCR-positive test result and to the

time of symptom onset. The number of and reason for sample exclusion are indicated in the

white boxes. Included samples are indicated in greyed boxes. Altogether, out of the 405

collected samples (from 142 SARS-CoV-2-positive patients), 232 samples from 130 patients with

a documented date for the SARS-CoV-2-specific PCR-positive test, and 105 samples from 63

patients with a documented date of symptom onset were tested with both VIDAS® SARS-CoV-2

IgM and IgG assays (paired tests).

Figure 2. Distribution of IgM (A) and IgG (B) index values obtained using the VIDAS® SARS-CoV-

2 IgM and IgG assays, respectively, in patients confirmed positive for SARS-CoV-2, according to

the time from symptom onset. VIDAS® SARS-CoV-2 IgM and IgG index values of 105 paired tests

from 63 SARS-CoV-2-positive patients (Table 4, Table 2 and Fig. 1) are displayed as Tukey box

plots according to the time from symptom onset. No more than one patient's sample is included

per time frame. The number of tested samples (N), and the median and interquartile range

(IQR) of index values are indicated below each graph. The dashed line shows the positivity cut-

off of both assays (i = 1.00).

Figure 3. Distribution of IgM (A) and IgG (B) index values obtained using the VIDAS® SARS-CoV-

2 IgM and IgG assays, respectively, in hospitalized vs. non-hospitalized patients confirmed

positive for SARS-CoV-2, according to the time from PCR-positive result. Out of the 232 paired

tests (Table 3), 15 were from patients with an unknown hospitalization status (Table 2) and

were thus excluded from the analysis. VIDAS® SARS-CoV-2 IgM and IgG index values of 217

paired tests from 115 SARS-CoV-2-positive patients (100 samples from 54 hospitalized patients

and 117 samples from 61 non-hospitalized patients) are depicted as Tukey box plots according

to the time from RT-PCR-positive test result. No more than one patient's sample is included per

time frame. The number of tested samples (N) and the median index are indicated below each

graph. The dashed line shows the positivity cut-off of both assays (i = 1.00). Differences between

the groups of hospitalized and non-hospitalized patients were tested at each time frame (0-7, 8-

15, 16-23 days post positive PCR) using a two-sided MWU-test; the respective p-values are

displayed above each graph. No statistical testing was performed at 24-31 and ≥ 32 days due to

the too low N values, notably in the group of non-hospitalized patients.

Figure 4. Kinetics of IgM and IgG seroconversion in four selected hospitalized patients. VIDAS®

SARS-CoV-2 IgM and IgG index values of four patients (A to D, respectively) measured over time

after symptom onset are presented. The dashed line indicates the positivity cut-off of both

assays (i = 1.00). Further patients' information is as follows: (**A**) the 78-year-old male patient

was in the intensive-care unit (ICU) at all investigated time points, except at the first (day 7) and

last (day 33) measurement time points; (**B**) the 77-year-old male patient was in ICU at all

investigated time points; (**C**) the 43-year-old male patient was in ICU at all investigated times,

except at the last two measurement time points (day 71 and 74); (**D**) the 67-year-old male

patient was in ICU at all investigated time points.

Table 1. Cross-reactivity of human sera from patients with other infections or medical conditions potentially interfering with the VIDAS® SARS-CoV-2 IgM and IgG assays

Sample category	VIDAS® SARS-CoV-2 IgM		VIDAS® SARS-CoV-2 IgG	
	No. samples tested	No. positive tests	No. samples tested	No. positive tests
Pregnant women	5	0	5	0
Antinuclear antibody (ANA)[a]	47	2	47	0
Rheumatoid factor	19	4	19	0
Human anti-mouse antibody (HAMA)	5	0	5	0
Borrelia burgdorferi [b]	10	0	6	0
Haemophilus Influenzae B	5	0	5	0
Plasmodium falciparum	3	1	3	0
Toxoplasma gondii [b]	10	0	6	0
Treponema pallidum	3	0	3	0
Trypanosoma cruzi	5	1	5	0
Hepatitis A virus (HAV)	3	0	3	0
Hepatitis B virus (HBV)	5	0	5	0
Hepatitis C virus (HCV)	5	0	5	0
Hepatitis E virus (HEV)[b]	7	0	6	0
Herpes simplex virus (HSV)[b]	6	0	6	0
Human immunodeficiency virus (HIV)	5	0	5	1
Cytomegalovirus (CMV)	4	0	3	0
Measles virus (MV)	4	0	3	0
Mumps virus (MuV)	1	0	3	0
Rubella virus (RuV)[b]	10	0	6	0
Dengue virus (DENV)	3	0	3	0
West Nile virus (WNV)	4	0	3	0
Yellow fever virus (YFV)	4	0	3	0
Zika virus (ZIKV)[b]	5	0	5	0
Adenovirus (AdV)	2	0	2	0
Metapneumovirus (MPV)	4	0	4	0
Rhinovirus/Enterovirus (RV/EnteroV)[c]	20	1	20	0
Influenza A and B virus (IAV/IBV)	30	0	30	0
Parainfluenza viruses 1/2/3 (PIV-1/2/3)	11	0	11	0
Respiratory syncytial virus A or B (RSV A or B)	13	0	13	1
Coronavirus NL63/HKU1 (CoV-NL63/HKU1)[d]	9	1	9	0
Coronavirus 229E (CoV-229E)	7	0	7	0
Coronavirus OC43 (CoV-OC43)	2	0	2	0
Total	**276**	**10**	**261**	**2**

[a]Include anti-DNA, anti-SSA, anti-SSB and anti-Sm/RNP antibodies; [b]The proportion of acute infection (i.e. IgM-positive for the respective infectious agent, when IgM levels were characterized) among samples tested with the VIDAS® SARS-CoV-2 IgM assay was: Borrelia burgdorferi: 5/10, Toxoplasma gondii: 5/10, HEV: 3/7, HSV: 3/6, RuV: 5/10, ZIKV: 3/5; [c]One out of 20 sera was from a patient with a bocavirus (BoV) co-infection and was negative in both VIDAS® SARS-CoV-2 IgM and IgG assays; [d]Six out of the nine sera tested were from patients positive for CoV-NL63 and three were from patients positive for CoV-NL63 and/or CoV-HKU1.

Table 2. Demographics of French patients positive for SARS-CoV-2 used for the determination of VIDAS® SARS-CoV-2 IgM and IgG sensitivity (positive percent agreement with RT-PCR positivity)

VIDAS® SARS-CoV-2 serology testing	IgM	IgG	Combined IgM/IgG
Patients with documented date of RT-PCR-positive test			
Number of samples tested	234	253	232
Study population, N (%)	132 (100%)	139 (100%)	130 (100%)
Age in years, median (range)			
Non-hospitalized (N=61)	missing	missing	missing
Other (N=71/78/69)	71 (27-96)	70.5 (27-96)	70 (27-96)
Gender, N (%)			
Male	47 (35.6%)	53 (38.1%)	47 (36.2%)
Female	24 (18.2%)	25 (18.0%)	22 (16.9%)
Missing	61 (46.2%)	61 (43.9%)	61 (46.9%)
Disease severity, N (%)			
Hospitalized	56 (42.4%)	57 (41.0%)	54 (41.55%)
Non-hospitalized	61 (46.2%)	61 (43.9%)	61 (46.9%)
Missing	15 (11.4%)	21 (15.1%)	15 (11.55%)
Patients with documented date of symptom onset			
Number of samples tested	108	113	105
Study population, N (%)	65 (100%)	71 (100%)	63 (100%)
Age in years, median (range)	70 (27-96)	70 (27-96)	70 (27-96)
Gender, N (%)			
Male	45 (69.2%)	51 (71.8%)	45 (71.4%)
Female	20 (30.8%)	20 (28.2%)	18 (28.6%)
Disease severity, N (%)			
Hospitalized	50 (76.9%)	50 (70.4%)	48 (76.2%)
Non-hospitalized[a]	0 (0.00%)	0 (0.0%)	0 (0.0%)
Missing	15 (23.1%)	21 (29.6%)	15 (23.8%)

[a]Date of symptom onset not documented in non-hospitalized COVID-19 patients

685   **Table 3. Positive percent agreement (PPA) of the VIDAS® SARS-CoV-2 IgM, IgG and combined**

686   **IgM/IgG test results of SARS-CoV-2-positive samples, according to the time from RT-PCR-**

687   **positive result**

VIDAS® SARS-CoV-2 serology testing	Time from RT-PCR-positive result	Median (range) time in days	No. samples[a]	No. positive results	PPA	95% CI
IgM (N=232)	0-7 days	2 (0-7)	110	49	44.5%	35.6-53.9%
	8-15 days	14 (8-15)	60	49	81.7%	70.1-89.4%
	16-23 days	20 (16-23)	38	31	81.6%	66.6-90.8%
	24-31 days	26 (24-28)	13	13	100.0%	75.3-100.0%
	≥ 32 days	33 (32-65)	11	9	81.8%	52.3-94.9%
IgG (N=232)	0-7 days	2 (0-7)	110	50	45.5%	36.5-54.8%
	8-15 days	14 (8-15)	60	53	88.3%	77.8-94.2%
	16-23 days	20 (16-23)	38	36	94.7%	82.7-98.5%
	24-31 days	26 (24-28)	13	13	100.0%	75.3-100.0%
	≥ 32 days	33 (32-65)	11	11	100.0%	71.5-100.0%
Combined IgM/IgG[b] (N=232)	0-7 days	2 (0-7)	110	59	53.6%	44.4-62.7%
	8-15 days	14 (8-15)	60	57	95.0%	86.1-99.0%
	16-23 days	20 (16-23)	38	38	100.0%	90.7-100.0%
	24-31 days	26 (24-28)	13	13	100.0%	75.3-100.0%
	≥ 32 days	33 (32-65)	11	11	100.0%	71.5-100.0%

688   [a]no more than one test result per patient per time period; [b]combined test is positive when at least one of IgM and/or IgG tests are
689   positive

691   **Table 4. Positive percent agreement (PPA) of the VIDAS® SARS-CoV-2 IgM, IgG and combined**

692   **IgM/IgG test results of SARS-CoV-2-positive samples, according to the time from symptom**

693   **onset**

VIDAS® SARS-CoV-2 serology testing	Time from symptom onset	Median (range) time in days	No. samples[a]	No. positive results	PPA	95% CI
IgM (N=105)	0-7 days	5.5 (1-7)	22	7	31.8%	16.4-52.7%
	8-15 days	12 (8-15)	29	24	82.8%	65.5-92.4%
	16-23 days	18 (16-23)	26	26	100.0%	86.8-100.0%
	24-31 days	26 (24-30)	18	18	100.0%	81.5-100.0%
	≥ 32 days	35 (32-65)	10	10	100.0%	69.2-100.0%
IgG (N=105)	0-7 days	5.5 (1-7)	22	7	31.8%	16.4-52.7%
	8-15 days	12 (8-15)	29	25	86.2%	69.4-94.5%
	16-23 days	18 (16-23)	26	25	96.2%	80.4-99.9%
	24-31 days	26 (24-30)	18	17	94.4%	74.2-99.0%
	≥ 32 days	35 (32-65)	10	10	100.0%	69.2-100.0%
Combined IgM/IgG[b] (N=105)	0-7 days	5.5 (1-7)	22	8	36.4%	19.7-57.0%
	8-15 days	12 (8-15)	29	26	89.7%	73.6-96.4%
	16-23 days	18 (16-23)	26	26	100.0%	86.8-100.0%
	24-31 days	26 (24-30)	18	18	100.0%	81.5-100.0%
	≥ 32 days	35 (32-65)	10	10	100.0%	69.2-100.0%

694   [a]no more than one test result per patient per time period; [b]combined test is positive when at least one of IgM and/or IgG tests are
695   positive

697 **Table 5. Positive and negative predictive values (PPV/NPV) at 5% prevalence of the VIDAS**[®]

698 **SARS-CoV-2 IgM, IgG and combined IgM/IgG test results, according to the time from symptom**

699 **onset**

VIDAS® SARS-CoV-2 serology testing	Time from symptom onset	PPV[a] (95% CI)	NPV[a] (95% CI)
IgM	0-7 days	72.1% (36.3-92.1%)	96.5% (95.4-97.4%)
	8-15 days	87.0% (62.5-96.4%)	99.1% (98.0-99.6%)
	16-23 days	89.0% (67.1-97.0%)	100.0% (N/A)[b]
	24-31 days	89.0% (67.1-97.0%)	100.0% (N/A)[b]
	≥ 32 days	89.0% (67.1-97.0%)	100.0% (N/A)[b]
IgG	0-7 days	100.0% (N/A)[b]	96.5% (95.4-97.4%)
	8-15 days	100.0% (N/A)[b]	99.3% (98.2-99.7%)
	16-23 days	100.0% (N/A)[b]	99.8% (98.6-100.0%)
	24-31 days	100.0% (N/A)[b]	99.7% (98.1-100.0%)
	≥ 32 days	100.0% (N/A)[b]	100.0% (N/A)[b]
Combined IgM/IgG	0-7 days	74.7% (40-92.9%)	96.7% (95.6-97.6%)
	8-15 days	87.9% (64.5-96.7%)	99.5% (98.4-99.8%)
	16-23 days	89.0% (67.1-97.0%)	100.0% (N/A)[b]
	24-31 days	89.0% (67.1-97.0%)	100.0% (N/A)[b]
	≥ 32 days	89.0% (67.1-97.0%)	100.0% (N/A)[b]

700 [a]PPV and NPV were calculated at 5% prevalence and using values of specificity and sensitivity (PPA) determined on paired
701 VIDAS® SARS-CoV-2 IgM and IgG test results (N=308 for specificity, Table S3; N=105 for PPA, Table 4); [b]95% CI not calculable
702 (division by zero)

Figure 1

Figure 2

Figure 3

Figure 4