

## 1 **A haemagglutination test for rapid detection of antibodies to SARS-CoV-2**

2

3 Alain Townsend<sup>1,2\*</sup>, Pramila Rijal<sup>1,2</sup>, Julie Xiao<sup>1</sup>, Tiong Kit Tan<sup>1</sup>, Kuan-Ying A Huang<sup>3</sup>, Lisa  
4 Schimanski<sup>1,2</sup>, Jiangdong Huo<sup>4</sup>, Nimesh Gupta<sup>5</sup>, Rolle Rahikainen<sup>6</sup>, Philippa C Matthews<sup>7,8</sup>,  
5 Derrick Crook<sup>7,9,10</sup>, Sarah Hoosdally<sup>8,10</sup>, Teresa Street<sup>9</sup>, Justine Rudkin<sup>11</sup>, Nicole Stoesser<sup>7,9</sup>,  
6 Fredrik Karpe<sup>12</sup>, Matthew Neville<sup>12</sup>, Rutger Ploeg<sup>13</sup>, Marta Oliveira<sup>13</sup>, David J Roberts<sup>14,15</sup>,  
7 Abigail A Lamikanra<sup>14</sup>, Hoi Pat Tsang<sup>14</sup>, Abbie Bown<sup>16</sup>, Richard Vipond<sup>16</sup>, Alexander J  
8 Mentzer<sup>17</sup>, Julian C Knight<sup>17</sup>, Andrew Kwok<sup>17</sup>, Gavin Screaton<sup>17,18</sup>, Juthathip  
9 Mongkolsapaya<sup>2,18,19</sup>, Wanwisa Dejnirattisai<sup>17</sup>, Piyada Supasa<sup>17</sup>, Paul Klenerman<sup>8</sup>, Christina  
10 Dold<sup>20,21</sup>, Kenneth Baillie<sup>22</sup>, Shona C Moore<sup>23</sup>, Peter JM Openshaw<sup>24</sup>, Malcolm G Semple<sup>23</sup>,  
11 Lance CW Turtle<sup>25</sup>, Mark Ainsworth<sup>26</sup>, Alice Allcock<sup>17</sup>, Sally Beer<sup>26</sup>, Sagida Bibi<sup>20</sup>, Elizabeth  
12 Clutterbuck<sup>20</sup>, Alexis Espinosa<sup>26</sup>, Maria Mendoza<sup>26</sup>, Dominique Georgiou<sup>26</sup>, Teresa Lockett<sup>26</sup>,  
13 Jose Martinez<sup>26</sup>, Elena Perez<sup>26</sup>, Veronica Sanchez<sup>26</sup>, Giuseppe Scozzafava<sup>26</sup>, Alberto  
14 Sobrinodiaz<sup>26</sup>, Hannah Thraves<sup>26</sup>, Etienne Joly<sup>27\*</sup>

15

16 <sup>1</sup>MRC Human Immunology Unit, MRC Weatherall Institute, John Radcliffe Hospital, Oxford  
17 OX3 9DS, UK

18 <sup>2</sup>Chinese Academy of Medical Science (CAMS) Oxford Institute (COI), University of Oxford,  
19 Oxford, UK

20 <sup>3</sup>Research Center for Emerging Viral Infections, College of Medicine, Chang Gung University  
21 and Division of Pediatric Infectious Diseases, Department of Pediatrics, Chang Gung Memorial  
22 Hospital, Taoyuan, Taiwan

23 <sup>4</sup>Structural Biology, The Rosalind Franklin Institute, Harwell Science & Innovation Campus,  
24 OX11 0FA, UK

25 <sup>5</sup>Vaccine Immunology Laboratory, National Institute of Immunology, Aruna Asaf Ali Marg,  
26 New Delhi – 110067, India

27 <sup>6</sup>Department of Biochemistry, University of Oxford, South Park Road OX1 3QU, UK

28 <sup>7</sup>Department of Microbiology and Infectious Diseases, John Radcliffe Hospital, Headley Way,  
29 Oxford OX3 9DU, UK

30 <sup>8</sup>Nuffield Department of Medicine, University of Oxford, Medawar Building for Pathogen  
31 Research, South Parks Road, Oxford OX1 3SY, UK

NOTE: This preprint reports new research that has not been certified by peer review and should not be used to guide clinical practice.

32 <sup>9</sup>Nuffield Department of Medicine, University of Oxford, John Radcliffe Hospital, Headley  
33 Way, Oxford OX3 9DU, UK

34 <sup>10</sup>Oxford NIHR Biomedical Research Centre, John Radcliffe Hospital, Headley Way, Oxford OX3  
35 9DU, UK

36 <sup>11</sup>Nuffield Department of Population Health, Big Data Institute, University of Oxford, Oxford,  
37 OX3 7LF

38 <sup>12</sup>Oxford Centre for Diabetes, Endocrinology and Metabolism, Radcliffe Department of  
39 Medicine, University of Oxford, Oxford, UK

40 <sup>13</sup>Nuffield Department of Surgical Sciences, University of Oxford, John Radcliffe Hospital,  
41 Headington, Oxford OX3 9DU, UK

42 <sup>14</sup>NHS Blood and Transplant, John Radcliffe Hospital, Headington, Oxford OX3 9BQ, UK

43 <sup>15</sup>BRC Haematology Theme and Radcliffe Department of Medicine, University of Oxford, John  
44 Radcliffe Hospital, Oxford, OX3 9DU

45 <sup>16</sup>Porton Down, Public Health England, Manor Farm Road, Porton Down, Salisbury, SP4 0JG,  
46 UK

47 <sup>17</sup>Wellcome Centre for Human Genetics, University of Oxford, Oxford OX3 7BN, UK

48 <sup>18</sup>Division of Medical Sciences, John Radcliffe Hospital, University of Oxford, Oxford, UK

49 <sup>19</sup>Dengue Hemorrhagic Fever Research Unit, Office for Research and Development, Faculty of  
50 Medicine, Siriraj Hospital, Mahidol University, Bangkok, Thailand

51 <sup>20</sup>Oxford Vaccine Group, Department of Paediatrics, University of Oxford, Oxford OX3 7LE, UK

52 <sup>21</sup>NIHR Oxford Biomedical Research Centre, Centre for Clinical Vaccinology and Tropical  
53 Medicine, University of Oxford, Oxford, UK

54 <sup>22</sup>Anaesthesia, Critical Care and Pain Medicine Division of Health Sciences, University of  
55 Edinburgh, Edinburgh, UK

56 <sup>23</sup>NIHR Health Protection Research Unit in Emerging and Zoonotic Infections, Institute of  
57 Infection, Veterinary and Ecological Sciences, University of Liverpool

58 <sup>24</sup>National Heart and Lung Institute, Faculty of Medicine, Imperial College London, London,  
59 UK

60 <sup>25</sup>Tropical and Infectious Diseases Unit, Liverpool University Hospitals NHS Foundation Trust,  
61 Liverpool, UK

62 <sup>26</sup>Oxford University Hospitals NHS Foundation Trust, Oxford, UK

63 <sup>27</sup>Institute of Pharmacology and Structural Biology (IPBS), University of Toulouse, CNRS,  
64 Toulouse; France

65

66 \*Corresponding Authors

67 Correspondence to Alain Townsend ([alain.townsend@imm.ox.ac.uk](mailto:alain.townsend@imm.ox.ac.uk)) and Etienne Joly  
68 ([atnjoly@mac.com](mailto:atnjoly@mac.com))

69

## 70 **ABSTRACT**

71 Serological detection of antibodies to SARS-CoV-2 is essential for establishing rates of  
72 seroconversion in populations, detection of seroconversion after vaccination, and for seeking  
73 evidence for a level of antibody that may be protective against COVID-19 disease. Several  
74 high-performance commercial tests have been described, but these require centralised  
75 laboratory facilities that are comparatively expensive, and therefore not available universally.  
76 Red cell agglutination tests have a long history in blood typing, and general serology through  
77 linkage of reporter molecules to the red cell surface. They do not require special equipment,  
78 are read by eye, have short development times, low cost and can be applied as a Point of Care  
79 Test (POCT). We describe a red cell agglutination test for the detection of antibodies to the  
80 SARS-CoV-2 receptor binding domain (RBD). We show that the Haemagglutination Test  
81 (“HAT”) has a sensitivity of 90% and specificity of 99% for detection of antibodies after a PCR  
82 diagnosed infection. The HAT can be titrated, detects rising titres in the first five days of  
83 hospital admission, correlates well with a commercial test that detects antibodies to the RBD,  
84 and can be applied as a point of care test. The developing reagent is composed of a previously  
85 described nanobody to a conserved glycoporphin A epitope on red cells, linked to the RBD from  
86 SARS-CoV-2. It can be lyophilised for ease of shipping. We have scaled up production of this  
87 reagent to one gram, which is sufficient for ten million tests, at a cost of ~0.27 UK pence per  
88 test well. Aliquots of this reagent are ready to be supplied to qualified groups anywhere in  
89 the world that need to detect antibodies to SARS-CoV-2, but do not have the facilities for high  
90 throughput commercial tests.

91

## 92 **INTRODUCTION**

93 Red cell agglutination tests have a distinguished history. Since Landsteiner’s classic  
94 observations in 1901 (Landsteiner, 1961) (English translation), they have been used for the

95 determination of blood groups (Schwarz & Dorner, 2003) detection of influenza viruses (Hirst,  
96 1941) and in a wide variety of applications championed by Prof. Robin Coombs for the  
97 detection of specific antibodies or antigens (Coombs, Mourant, & Race, 1945) (reviewed by  
98 (Pamphilon & Scott, 2007)). They have the great advantage of being simple, inexpensive, can  
99 be read by eye, and do not require sophisticated technology for their application. In the  
100 recent era the linkage of an antigen to the red cell surface has become easier with the  
101 possibility of fusing a protein antigen sequence with that of a single domain antibody or  
102 nanobody specific for a molecule on the red cell surface (discussed in (Habib et al., 2013)).


103

104 We have applied this concept to provide a simple Haemagglutination Test (“HAT”) for the  
105 detection of antibodies to the Receptor Binding Domain (RBD) of the SARS-CoV-2 spike  
106 protein. The RBD is a motile subdomain at the tip of the SARS-CoV-2 spike protein that is  
107 responsible for binding the virus to its ACE2 receptor. The RBD of betacoronaviruses folds  
108 independently of the rest of the spike protein (Lan et al., 2020; Li, Li, Farzan, & Harrison, 2005;  
109 Wang et al., 2020; Yan et al., 2020). This useful property provides an Achilles’ heel for the  
110 virus and allows many potential applications in vaccine design (Dai et al., 2020; Mulligan et  
111 al., 2020; Tan et al., 2020; Walls et al., 2020; Yang et al., 2020; Zha et al., 2020), and serology  
112 (Amanat et al., 2020; Piccoli et al., 2020; The National SARS-CoV-2 Serology Assay Evaluation  
113 Group, 2020), see also [www.gov.uk/government/publications/COVID-19-laboratory-](http://www.gov.uk/government/publications/COVID-19-laboratory-evaluations-of-serological-assays)  
114 [evaluations-of-serological-assays](http://www.gov.uk/government/publications/COVID-19-laboratory-evaluations-of-serological-assays)). The majority of neutralising antibodies bind to the RBD  
115 (Barnes et al., 2020; Piccoli et al., 2020), and the level of antibody to the RBD detected in  
116 ELISA correlates with that of neutralising antibodies (Amanat et al., 2020; Piccoli et al., 2020;  
117 Robbiani et al., 2020). We reasoned therefore that a widely applicable and inexpensive test  
118 for antibodies to the RBD would be useful for research in settings where high throughput  
119 assays were not available.

120

121 In order to link the SARS-CoV-2 RBD to red cells we selected the single domain antibody  
122 (nanobody) IH4 (Habib et al., 2013), specific for a conserved epitope on glycophorin A.  
123 Glycophorin A is expressed at up to  $10^6$  copies per red cell. The IH4 nanobody has previously  
124 been linked to HIV p24 to provide a monomeric reagent that bound p24 to the red cell surface.  
125 Antibodies to p24 present in serum cross-linked the p24 and agglutinated the red cells (Habib  
126 et al., 2013). We have adapted this approach to detection of antibodies to SARS-CoV-2 by

127 linking the RBD of the SARS-CoV-2 spike protein to IH4 via a short (GSG)<sub>2</sub> linker to produce  
 128 the fusion protein IH4-RBD-6H (Figure 1). Since we embarked on this project, another group  
 129 has described preliminary results with an approach similar to ours, but using a fusion of the  
 130 RBD to an ScFV against the H antigen to coat red blood cells with the SARS-2 RBD (Kruse et  
 131 al., 2020).  
 132  
 133



134  
 135  
 136 **Figure 1. Haemagglutination Test (HAT) for detection of antibodies to SARS-CoV-2 Receptor**  
 137 **Binding Domain.**

138 A) Concept of the HAT  
 139 B) Sequence of VHH(IH4)-RBD fusion protein. Residues underlined are encoded by cloning  
 140 sites AgeI (TG) and Sall (AST). The codon optimised cDNA sequence is shown in supplementary  
 141 Information  
 142 C) SDS-PAGE gel of purified VHH(IH4)-RBD proteins. Three micrograms of protein were run on  
 143 4-12% Bolt Bis-Tris under reducing conditions. 1: IH4-RBD produced in house in Expi293F cells,  
 144 2: IH4-RBD produced by Absolute Antibody, Oxford in HEK293 cells.

145

## 146 **RESULTS**

147

### 148 **Production of the IH4-RBD Reagent**

149 The IH4-RBD sequence (Figure 1B) was codon optimised and expressed in Expi293F cells in a  
150 standard expression vector (available on request). One advantage of this mode of production  
151 compared to bacterially produced protein as used by Habib et al, is that the reagent will carry  
152 the glycosylation moieties found in humans, which may play a role in the antigenicity of the  
153 RBD (Pinto et al., 2020). The protein (with a 6xHis tag at the C-terminus for purification) was  
154 purified by Ni-NTA chromatography which yielded ~160 mg/L. We later had one gram of the  
155 protein synthesised commercially by Absolute Antibody, Oxford. The IH4-RBD protein ran as  
156 single band at ~40 kDa on SDS PAGE (Figure 1C).

157

### 158 **Establishment of the Haemagglutination Test (HAT) with monoclonal antibodies to the RBD**

159 One purpose envisaged for the HAT is for use as an inexpensive Point of Care Test for  
160 detection of antibodies in capillary blood samples obtained by a “finger-prick”. We therefore  
161 wished to employ human red cells as indicators without the need for cell separation or  
162 washes, to mimic this setting. The use of V-bottom microtiter plates to perform simplified  
163 hemagglutination tests was first described over 50 years ago (Wegmann & Smithies, 1966).  
164 In preliminary tests, we observed that 50  $\mu$ L of whole blood (K2EDTA sample) diluted 1:40 in  
165 phosphate-buffered saline (PBS), placed in V-bottomed wells of a standard 96-well plate,  
166 settled in one hour to form a button of red cells at the bottom of the well. The normal  
167 haematocrit of blood is ~40% vol/vol, so this dilution provides ~1% red cells. If the plate was  
168 then tilted, the red cell button flowed to form a “teardrop” in ~30 seconds (for example Figure  
169 2A Row 8).


170

171 If serum or plasma samples are to be tested, a standard collection of 10 mL of Type O Rh-  
172 negative (O-ve) blood into a K2EDTA tube will thus provide sufficient red cells for 8,000 test  
173 wells.

174

175 A well characterised monoclonal antibody to the RBD, CR3022 (ter Meulen et al., 2006) added  
176 to the red cells at between 0.5-32 ng/well in 50  $\mu$ L, did not agglutinate the cells on its own

177 (Figure 2A Row 8). The addition of the IH4-RBD reagent at between 12.5-800 ng/well (in 50  
 178  $\mu$ L PBS) induced a concentration dependent agglutination of the red cells, detected by the  
 179 formation of a visible mat or plug of agglutinated cells, and the loss of teardrop formation on  
 180 tilting the plate (Figure 2A). From repeated trials of this experiment we established that a  
 181 standard addition of 100 ng/well of the IH4-RBD developer (50  $\mu$ L of a stock solution of 2  
 182  $\mu$ g/mL in PBS) induced agglutination of 50  $\mu$ L of 1:40 human red cells in the presence of as  
 183 little as 2 ng/well of the CR3022 monoclonal antibody. The standardised protocol used for the  
 184 subsequent tests were thus performed in 100 or 150  $\mu$ L final volume, containing 100 ng of  
 185 the IH4-RBD developer, and 50  $\mu$ L 1:40 whole blood ( $\sim$ 1% v/v red cells  $\sim$  0.5  $\mu$ L packed red  
 186 cells per reaction). After 60 minutes incubation at room temperature, we routinely  
 187 photographed the plates after the 30 s tilt for examination and reading.  
 188  
 189 The requirement for 100 ng of the IH4-RBD developer per test well means that the gram of  
 190 IH4-RBD protein we have had synthesised is sufficient for 10 million test wells at a cost of  
 191 approximately 0.27 UK pence per test.



192  
 193 **Figure 2. Haemagglutination with human monoclonal antibodies or nanobodies to the**  
 194 **SARS-CoV-2 RBD**

195 A) Titration of IH4-RBD and monoclonal Antibody CR3022 to RBD. Doubling dilutions of  
 196 CR3022 and IH4-RBD were prepared in separate plates. 50  $\mu$ L red cells (O-ve whole blood  
 197 diluted 1:40 in PBS) were added to the CR3022 plate, followed by transfer of 50  $\mu$ L titrated  
 198 IH4-RBD. From this titration, 100 ng/well of IH4-RBD was chosen for detection.

199 B) Detection of other anti-RBD monoclonal antibodies and ACE2-Fc. Monoclonal antibodies  
 200 were prepared in doubling dilutions in 50  $\mu$ L PBS from left to right, 50  $\mu$ L of 1:40 O-ve red cells  
 201 were added, followed by 50  $\mu$ L of IH4-RBD (2  $\mu$ g/mL in PBS). The end point was defined as the

202 last dilution without tear drop formation on tilting the plate for ~ 30 s. The binding sites for  
203 CR3022, EY6A and ACE2 on RBD have been defined (Huo, Zhao, et al., 2020; Lan et al., 2020;  
204 Yan et al., 2020; Zhou et al., 2020). EW-9B and EW-9C are monoclonal antibodies against non-  
205 RBD epitopes on the spike protein (Huang et al., 2020). ACE2-Fc has been described (Huang  
206 et al., 2020).

207  
208 Having established a standard addition of 100 ng/well of the IH4-RBD reagent, we screened a  
209 set of twelve human monoclonal antibodies, two divalent nanobodies, and divalent ACE2-Fc,  
210 that are known to bind to the RBD (Huang et al., 2020; Huo, Le Bas, et al., 2020; Pinto et al.,  
211 2020; Wrapp et al., 2020; Zhou et al., 2020). These reagents bind to at least three independent  
212 sites on the RBD, and some are strongly neutralising and capable of profound ACE2 blockade  
213 (Huang et al., 2020; Huo, Le Bas, et al., 2020). Twelve of the 15 divalent molecules  
214 agglutinated red cells and titrated in the HAT to an end point between 2-125 ng/well, after  
215 addition of 100 ng IH4-RBD (Figure 2B and Supplementary Table 1). Two monoclonal  
216 antibodies, FD-5D and EZ-7A (Huang et al., 2020) and one divalent nanobody VHH72-Fc  
217 (Wrapp et al., 2020), failed to agglutinate red cells in the presence of the IH4-RBD reagent.  
218 However, if a monoclonal antibody to human IgG was added to the reaction (50 µL of mouse  
219 anti-human IgG, Sigma Clone GG5 1:100), these molecules specifically agglutinated the red  
220 cells (Supplementary Figure 1A, B). This result, analogous to the “indirect” Coombs Test  
221 (Coombs et al., 1945; Pamphilon & Scott, 2007), suggested that these three molecules had  
222 bound to the RBD associated with the red cells but failed to crosslink to RBDs on neighbouring  
223 red cells. However, these could be crosslinked by the anti IgG reagent. Monoclonal antibodies  
224 to other regions of the spike protein (EW-9B, EW-9C and FJ-1C) failed to agglutinate red cells  
225 (Figure 2B, Supplementary Table 1). Finally we looked at the effect of a divalent ACE2-Fc  
226 molecule constructed by fusing the peptidase domain of ACE2 (amino acids 19-615) to the  
227 hinge and Fc region of human IgG1 (described in (Huo, Le Bas, et al., 2020). ACE2-Fc  
228 agglutinated red cells strongly in the presence of 100 ng/well of the IH4-RBD developer,  
229 titrating to ~ 4 ng/well (Figure 2B rows 7,8).

230  
231 In summary, these results showed that all of the known epitopes bound by characterised  
232 monoclonal antibodies were displayed by the IH4-RBD reagent, as well as the ACE2 binding  
233 site, and could mediate agglutination by specific antibodies, divalent nanobodies, or ACE2-Fc.


234

### 235 **Agglutination by plasma from donors convalescing from COVID-19**

236 These experiments established the conditions for detection of haemagglutination by  
237 monoclonal antibodies to the RBD, in particular the optimum concentration of IH4-RBD of  
238 100 ng/well. We then proceeded to look for haemagglutination by characterised plasma from  
239 COVID-19 convalescent donors. In the first trial we tested eighteen plasma samples from  
240 patients with mostly mild illness, that had been characterised with a quantitative ELISA to  
241 detect antibodies to the RBD (Peng et al., 2020). For these experiments we used fresh O-ve  
242 blood (K2EDTA sample) diluted to 1:40 as a source of red cells to avoid agglutination by  
243 natural agglutinins in the plasma. Plasmas were titrated by doubling dilution from 1:20 in 50  
244  $\mu$ L, then 50  $\mu$ L of 1:40 O-ve red cells were added, followed by addition of 100 ng of the IH4-  
245 RBD in 50  $\mu$ L PBS. After one-hour incubation, plates were tilted for ~30 seconds,  
246 photographed and read. The titre of agglutination was assessed by complete loss of teardrop  
247 formation by the red cells, any formation of a teardrop was regarded as negative. Figure 3A  
248 shows that the HAT titre matched the RBD ELISA results. Four samples were scored as  
249 negative in both assays. The remaining results showed that in general the HAT titre increased  
250 with the ELISA end point titre. One sample gave a positive titre of 1:320 in the HAT but was  
251 negative in ELISA (indicated with an arrow). We investigated this sample with a developer  
252 composed of the IH4 nanobody without the RBD component, which revealed that  
253 agglutination was RBD dependent (not shown). This sample was also positive at 1:1123 in an  
254 ELISA for full length spike protein (not shown), which suggests that the antibodies contained  
255 in this serum recognised epitope(s) present on the RBD exposed in the HAT, but not on the  
256 RBD in the RBD-ELISA reference test (Peng et al., 2020). The highest titre detected in these  
257 samples by the HAT was 1:1280 (Figure 3B).

258

259 These preliminary results showed that the HAT could detect antibodies to the RBD in plasma  
260 samples from convalescent patients in a similar manner to an ELISA test, but were not  
261 sufficient to establish the sensitivity and specificity of the HAT.



262

263 **Figure 3. Titration of stored plasma in the agglutination assay.**

264 A) Eighteen plasma samples from mild cases were compared for titration in the HAT with 1:40  
265 O-ve whole blood from a seronegative donor, and endpoint titre in an RBD ELISA (Peng et al.,  
266 2020)). Four samples were negative in both assays. The data point marked with an arrow on  
267 the graph (plasma 2 on the plate, Fig 3B) was checked with a reagent composed of IH4 without  
268 RBD and shown to be dependent on antibodies to the RBD. This sample did score positive for  
269 antibodies to full length spike in an ELISA (EPT 1:1123).

270 B) An example of titration: positive agglutination endpoints (loss of teardrop) are marked with  
271 a black solid-line circle, partial teardrops are marked with a dotted-line circle.

272

### 273 Sensitivity and specificity of the Haemagglutination Test

274 To formally assess the sensitivity and specificity of the HAT we collected a set of 98 “positive”  
275 plasma samples from donors diagnosed with COVID-19 by RT-PCR at least 28 days prior to  
276 sample collection (NHS Blood and Transplant), and 199 “negative” serum samples from  
277 healthy donors from the pre-COVID-19 era (Oxford Biobank). The samples were randomised  
278 before plating. The test wells were arranged in duplicate to contain serum/plasma at 1:40  
279 dilution, and 1:40 O-ve red cells in 50  $\mu$ L. 100 ng of IH4-RBD in 50  $\mu$ L PBS was added to one  
280 well of the pair, 50  $\mu$ L of PBS to the other (as a negative control). The negative control is  
281 important because in rare cases, particularly in donors who may have received blood  
282 transfusions, the sample may contain antibodies to non-ABO or Rhesus D antigens. After  
283 development, the plates were photographed, and read by two independent masked  
284 observers. Complete loss of teardrop was scored as positive, any flow in the teardrop as

285 negative. These rules were established before setting up the tests. Disagreements (7%  
286 overall) were resolved by accepting the weaker interpretation – i.e. if one observer scored  
287 positive but the other negative, the well was scored as negative. Having completed the  
288 scoring the columns of samples were re-randomised and the test and scoring repeated.


289  
290 Examples of test wells and scoring are shown in Figure 4. The red cells in the negative control  
291 (PBS) wells formed a clear teardrop. Red cells in positive wells (indicated with a solid ring)  
292 settled either into a mat or a button that failed to form any teardrop on tilting for 30 seconds.  
293 Occasional wells (15 of 297) formed a “partial” teardrop (shown by a dashed ring). These were  
294 scored as negative by prior agreement.

295  
296 With these rules in place we obtained in the first run sensitivity 88%, specificity 99%, and in  
297 the second run sensitivity 93%, specificity 99% (Figure 4). The Siemens Atellica  
298 Chemiluminescence assay for detection of IgG antibodies to the RBD was run in parallel on  
299 293 of these 297 samples and gave sensitivity 100%, specificity 100% for this sample set.

300  
301 We decided prior to this formal assessment to score wells with partial teardrop formation as  
302 negative, as these wells tended to give rise to disagreements between scorers and were not  
303 very helpful. Fifteen of 297 wells gave a partial teardrop. Six of these fifteen were from PCR-  
304 ve donors and scored negative on the Siemens assay, 9/16 were from PCR+ve donors and  
305 were Siemens positive. If partial teardrops were scored as positive, the sensitivity of the two  
306 assays increased to 97% and 99% (from 88% and 93%), but specificity was reduced to 96%  
307 and 98% (from 99%). This small loss of specificity would be unacceptable in sero-surveys  
308 where the expected prevalence of previous SARS-CoV-2 infection was low.

309

**A** Examples from screen of 297 test samples



**B** Operating Characteristics of the Haemagglutination Test

98 PCR +ve Samples  $\geq$ 28 days post diagnosis  
199 –ve samples from The Oxford Biobank Pre Covid-19

11 June			
Sensitivity and Specificity Calculator			
+VE = No Teardrop			
	HAT +VE	HAT -VE	
PCR +VE	86	12	98
PCR -VE	1	198	199
	87	210	297
Sensitivity %	88		
Specificity %	99		
Likelihood Ratio	174.6		
13 June REPEAT			
Sensitivity and Specificity Calculator			
+VE = No Teardrop			
	HAT +VE	HAT -VE	
PCR +VE	91	7	98
PCR -VE	1	198	199
	92	205	297
Sensitivity %	93		
Specificity %	99		
Likelihood Ratio	184.8		

310

311 **Figure 4. Operating characteristics of the HAT.**

312 A) The test set of 297 randomised plasma samples were diluted 1:40 mixed with 1:40 O-ve  
313 blood in two columns. IH4-RBD (100 ng in 50  $\mu$ L) was added to the test samples, and PBS to  
314 negative control wells. The plates were incubated at room temperature for one hour to allow  
315 the red cell pellet to form, then tilted for ~30 seconds to allow a teardrop to form. Complete  
316 loss of teardrop was scored as positive agglutination (marked with a black solid-line circle).  
317 Full teardrop or partial teardrop (marked with a dotted-line circle) were scored negative. The  
318 samples in columns were re-randomised and tested for a second time two days later.

319 B) Contingency table showing the operating characteristics of the HAT.

320

321 **HAT in the Hospital Setting**


322 We next assessed the HAT in the setting of patients recently admitted to hospital (the first  
323 five days) through access to the COMBAT collection of samples (see methods). This set  
324 comprised 153 plasma samples from donors diagnosed with COVID-19 by PCR, with clinical  
325 syndromes classified as « Critical », « Severe », « Mild », and « PCR positive Health Care  
326 Workers ». Seventy-nine control plasma samples donated in the pre-COVID-19 era were

327 obtained either from patients with bacterial sepsis (54 samples), or healthy volunteers (25  
328 samples). Samples were titrated in 11 doubling dilutions of 50  $\mu$ L from 1:40 – 1:40,096  
329 (columns 1-11). Column 12 contained 50  $\mu$ L PBS as a negative control. 50  $\mu$ L of 1:40 O-ve  
330 whole blood was added, followed by 50  $\mu$ L of 2  $\mu$ g/mL IH4-RBD (100 ng/well). In parallel, all  
331 of the 153 samples from PCR positive donors were assessed by the Siemens Atellica  
332 Chemiluminescence test for antibodies to the RBD of the spike protein. The HAT scores (as  
333 the number of doubling dilutions of the sample required to reach the endpoint of complete  
334 loss of teardrop), and representative agglutination results are shown in figure 5A. In Figure  
335 5B the HAT scores are plotted with their related Siemens test scores.

336


337 None of the seventy-nine negative control samples scored as positive in the HAT at a dilution  
338 of 1:40, thus providing 100% specificity in this set of samples. The HAT detected 131/153 (86%  
339 sensitivity) of the samples from PCR-diagnosed donors within the first five days of hospital  
340 admission, whereas the Siemens test detected 113/153 (74%). On day 5 the HAT detected  
341 41/45 (91% sensitivity). Two samples had an endpoint greater than 11 doubling dilutions in  
342 the HAT, and required a repeat measurement spanning two plates. These two samples  
343 titrated to 13 doubling dilutions (1: 163,840). Unmasking the samples revealed that both were  
344 acquired from an elderly lady with mild disease on days 3 and 5 of her admission. The range  
345 of positive titres detected by HAT was broad: 1 to 13 doubling dilutions (1:40 – 1:163,840). A  
346 correlation coefficient with the Siemens test could not be calculated as the latter has a ceiling  
347 score of 10 (Figure 5B). A comparison of the two tests in a contingency table with cut-off of  
348 1:40 (first doubling dilution) for HAT, and a score  $\geq$  1 for the Siemens test (as defined by the  
349 manufacturer), showed a strong correlation between the two tests for detection of antibodies  
350 to the RBD ( $P < 0.0001$ ; two-tailed Fisher's exact test, Figure 5B). Fifty-two of the 153 samples  
351 were from 24 donors with COVID-19 from whom repeated samples were taken on days 1, 3,  
352 or 5 of admission. The HAT detected a rise in agglutination titre over the first five days of  
353 admission in 16/24 (67%) of these patients (Table 1). Reductions in titre were not detected.

**A Titration in HAT by doubling dilution**



Donors were grouped according to clinical criteria; HCW = Front Line Health Care Workers; Sepsis = Samples with patients with Sepsis prior to Covid-19 pandemic; Healthy = Healthy volunteers prior to Covid-19; DD 1 = 1:40. Actual Titre =  $20 \times 2^{DD}$ .

**B**



354

355 **Figure 5. Titration of the set of 232 samples in the HAT**

356 A) The collection included 32 samples from 24 Critical patients, 62 samples from 48 Severe,  
 357 39 samples from 32 Mild, 20 single samples from health care workers (HCW), 54 samples from  
 358 43 patients with unrelated sepsis in the pre-COVID-19 era, 25 samples from healthy  
 359 unexposed controls. Median is indicated by a red line. DD: doubling dilutions.

360 B) Comparison to Siemens Result (anti RBD) with HAT titre by doubling dilution for 153  
 361 samples from Critical, Severe, Mild and HCW SARS-CoV-2 PCR positive donors.

362

363 These results showed that in the setting of hospital admission in the UK for suspected COVID-  
 364 19 disease, the HAT has an overall sensitivity of 86% and specificity of 100% by day five, and  
 365 frequently (67%) detected a rise in HAT titre during the first five days of admission. In this  
 366 context the HAT performed at least as well as the commercially available Siemens Atellica  
 367 Chemiluminescence assay (74%) for the detection of antibodies to the RBD of SARS-CoV-2  
 368 spike protein. Twenty samples were negative in both tests, but nine of these were taken on

369 day 1 of admission, which suggests that both tests have lower detection levels early in the  
370 course of hospital admission, before the antibody response has fully developed.

371

372 The O-ve blood used as indicator for this experiment was collected into a heparin tube, and  
373 then transferred to a K2EDTA tube. In order to be sure that the presence of heparin in the red  
374 cells had not altered the behaviour of the test, and to confirm the robustness of the results,  
375 we repeated the titrations on all of the 232 samples 34 days later, with fresh O-ve red cells  
376 from a different donor collected as usual into a K2EDTA tube. The results are shown in  
377 supplementary figures 2A-C. Specificity of the HAT remained at 100% (none of the 79 control  
378 samples were detected as positive at 1:40). The correlation with the previous assay was  
379 strong ( $R^2 = 0.975$ ), and 99% of the 232 titrations were within one doubling dilution of the  
380 matched earlier measurement. The slope of the correlation was 0.94 (95% CI 0.92-0.96),  
381 significantly less than 1. This was due to a proportion of results titrating to one doubling  
382 dilution lower titre. However, this had only a small impact on sensitivity (81% from 86%),  
383 which was still an improvement on the Siemens test (74%) in this context of the first five days  
384 of hospital admission. A rise in HAT titre in 16/24 during the first five days of admission was  
385 confirmed.

386

387

388

389

390

391

392

393

394

395

396

397

398

399

400

401 **Table 1. Fifty-two samples from 24 donors who were sampled repeatedly during the first**  
 402 **five days in hospital.**


		Day				
No	Case	1	2	3	4	5
1	C2			40		160
2	C6			40		80
3	C7	0				0
4	C9	40				2560
5	C10	320		320		
6	C23	0		160		640
7	S5	80		160		
8	S6	1280		1280		2560
9	S7	80				2560
10	S12	640		1280		
11	S13	0		40		320
12	S14			640		640
13	S17			2560		2560
14	S20	160		320		
15	S31	40				1280
16	S41	640		1280		
17	S43			40		160
18	S48			1280		2560
19	M9			0		160
20	M11	0		0		0
21	M13			0		0
22	M14			163840		163840
23	M25	40				640
24	M30			320		320

403  
 404 C: Critical, S: Severe, M: Mild

405  
 406 **HAT as a Point of Care Test on Capillary Samples**

407 The HAT is designed to detect antibodies to the RBD starting at a serum dilution of 1:40, and  
 408 we have found that that 50 µL of 1:40 dilution of whole blood provides an optimal  
 409 concentration of red blood cells for detection by agglutination in V-bottomed 96 well plates.  
 410 We have not completed an extensive analysis of the HAT as a Point of Care Test. However,  
 411 we have preliminary evidence that lyophilised IH4-RBD sent to the National Institute of  
 412 Immunology, New Delhi, functions as a Point of Care Test on capillary blood obtained by

413 finger-prick. In Figure 6, three positive (donors 1, 2 and 3) and three negative (donors 4, 5 and  
414 6) HAT results are compared to a standard ELISA for detection of antibodies to the RBD.  
415



416  
417 **Figure 6. HAT as a point of care test.**

418 Capillary blood samples were obtained by lancet. Antibodies to the RBD were detected by  
419 HAT on autologous red cells in the sample in “Test” wells (plasma at 1:40) after addition of  
420 100 ng/well IH4-RBD (see methods). NC, Negative Control (PBS replaces IH4-RBD); PC Postive  
421 Control (20 ng/well CR3022, an anti-RBD monoclonal antibody added). In parallel, after  
422 removal of red cells, the plasma was tested in a standard ELISA for detection of antibodies to  
423 the RBD. Low levels of antibody detected in the ELISA were sufficient to give a positive result  
424 in the HAT.

425  
426 Further work is needed to establish the operating characteristics of the HAT as a Point of Care  
427 Test on capillary samples. We provide a suggested operating procedure for capillary samples  
428 in methods.

#### 429 430 **Distribution of the IH4-RBD as lyophilised protein**

431 In order to ship the IH4-RBD reagent efficiently we have examined the effects of lyophilisation  
432 and reconstitution with water. IH4-RBD synthesised for the purpose of distribution was  
433 provided at 5 mg/mL in PBS by Absolute Antibody Ltd, Oxford. Two hundred microlitre  
434 aliquots (1 mg, enough for 10,000 test wells) were lyophilised overnight and stored at -20 °C.  
435 Aliquots were thawed, reconstituted with 1 ml double distilled water and titrated against the  
436 pre-lyophilisation material. No change in the titration occurred. We have synthesised one

437 gram of IH4-RBD, sufficient for 10 million test wells. This is available free of charge for any  
438 qualified group anywhere in the world in aliquots of 1 mg (20,000 test wells).

439

#### 440 **DISCUSSION**

441 The COVID-19 pandemic has had a particularly gruelling influence on the world economy, and  
442 on most populations of the world. The appearance of such a new highly contagious virus will  
443 probably not be a unique occurrence in the decades ahead. One of the lessons learned is the  
444 importance of developing affordable serological tests for detection of immune responses to  
445 SARS-CoV-2. Commercial antibody tests are not widely available to low- and middle-income  
446 countries, and lateral flow assays, while offering early promise as a near-patient test, have  
447 failed to deliver in terms of performance metrics, are expensive, and there are concerns about  
448 significant batch-to-batch variation (Adams et al., 2020). By contrast, the advantages of the  
449 HAT are the low cost of production of its single reagent (~0.27 UK pence per test well), better  
450 performance than most lateral flow devices (Adams et al., 2020), and versatility in not  
451 requiring anything other than a source of O-ve blood (10 mL of K2EDTA blood provides  
452 enough for 8,000 tests), an adjustable pipette, PBS, and a standard 96 well V-bottomed plate.

453

454 We have demonstrated that the HAT functions as a viable test for the presence of antibodies  
455 to the RBD of the SARS-CoV-2 spike protein in stored serum/plasma samples, using O-ve red  
456 cells as indicators. In the formal assessment of sensitivity and specificity we recorded an  
457 average 90% sensitivity and 99% specificity (Likelihood Ratio ~175), compared to 100%  
458 sensitivity and 100% specificity for the Siemens Atellica Chemiluminescence test on the same  
459 set of samples. The “positive” samples were selected to have been taken at least 28 days after  
460 a positive PCR test. These conditions are optimal for serological tests by allowing time for a  
461 rise in antibodies. The sensitivity of ~90% and specificity of 99% did not reach the level of  
462 98% for both recommended by the UK MHRA (The National SARS-CoV-2 Serology Assay  
463 Evaluation Group, 2020), or 99.5% for both recommended by the Infectious Disease Society  
464 of America (Hanson et al., 2020). However, these values are still consistent with a useful test  
465 in appropriate contexts, provided that users are fully aware of the operating characteristics  
466 and interpret the results correctly.

467

468 The sensitivity of the single point HAT can be enhanced (to ~98%) if wells with partial teardrop  
469 formation are scored as positive. However, this improvement in sensitivity is gained at the  
470 expense of a reduced specificity (to ~97%). If partial teardrops are to be scored as positive in  
471 the spot test at 1:40, we recommend obtaining confirmation for these by ELISA. Improvement  
472 in the operating characteristics of the HAT may be possible by a systematic analysis of buffer  
473 composition and experimental conditions, and is being investigated. We have deliberately  
474 kept complexity to a minimum, and thus all dilutions were made in standard PBS, and for the  
475 moment we recommend scoring wells with partial teardrop formation as negative.

476

477 It is interesting that the HAT titrations actually performed a little better than the Siemens test  
478 on 153 stored plasma samples from donors during the first five days of their hospital  
479 admission (note that symptom onset may have been several days earlier), in whom it  
480 detected 86% (81% in the repeat) of samples from PCR-diagnosed donors, compared to 74%  
481 for the Siemens test, and gave 100% specificity for the sample set containing control plasma  
482 from patients with sepsis and healthy controls. We speculate that at this early period of the  
483 COVID-19 illness the immune response may be dominated by IgM that would be expected to  
484 be particularly efficient at crosslinking the IH4-RBD labelled red cells. In addition, of the  
485 twenty-four donors who were tested more than once in the first five days in hospital, the HAT  
486 detected a rise in titre in sixteen (67%). A fixed high titre was detected in a further four to  
487 provide a sensitivity of 83% in these twenty-four cases. In situations of high clinical suspicion,  
488 the HAT could potentially have a place as a helpful test to support the diagnosis of COVID-19  
489 by detecting a rising titre of antibodies to the RBD during hospital admission. In patients with  
490 a prior probability of a diagnosis of COVID-19 of ~10%, the likelihood ratio of ~175 for the HAT  
491 provides a posterior probability of ~95% for this diagnosis. However, it is essential that if  
492 clinicians use a rising titre in the HAT as a diagnostic aid, they should be aware of the relatively  
493 low sensitivity (~67%) in this context.

494

495 With a specificity of 99%, the HAT could be employed for epidemiological surveys of the  
496 seropositive rate in stored serum/plasma samples from populations with a moderate  
497 expected prevalence ~10%, which would entail a negative predictive value of ~90%. The HAT  
498 may be useful to detect seroconversion after vaccination, and for the identification of  
499 potential donors of high titre plasma for therapy, if the clinical trials that are in progress

500 demonstrate a benefit. In the absence of knowledge about the level of antibody that indicates  
501 protection, the HAT should not be used to provide personal results to individuals, as discussed  
502 by the UK Royal College of Pathologists ([https://www.rcpath.org/profession/on-the-](https://www.rcpath.org/profession/on-the-agenda/COVID-19-testing-a-national-strategy.html)  
503 [agenda/COVID-19-testing-a-national-strategy.html](https://www.rcpath.org/profession/on-the-agenda/COVID-19-testing-a-national-strategy.html)). Finally, we show that the lyophilised  
504 IH4-RBD reagent sent to New Delhi functioned as expected in preliminary point of care testing  
505 on capillary samples obtained by finger-prick. However, additional evidence is needed to  
506 show that the sensitivity and specificity of the HAT, applied as a Point of Care Test in this way,  
507 are comparable to the tests on stored plasma samples, as stressed by the IDSA guideline on  
508 serological testing (Hanson et al., 2020). This will need to be done in field conditions, which is  
509 planned.

510

511 The technique required for applying the HAT can be learned in a day by a trained laboratory  
512 technician, paramedic, nurse or doctor. We have produced one gram of the developing IH4-  
513 RBD reagent (enough for ten million test wells) and offer to ship lyophilised aliquots of this  
514 material (sufficient for 10,000 tests) anywhere in the world, free of charge, for use as a  
515 research reagent for serological studies of COVID-19.

516

## 517 **METHODS**

518

### 519 **Sample Collection and Ethics**

520 **Figures 1,2:** Control whole blood (K2EDTA) as a source of red cells was collected from a  
521 healthy donor after informed consent.

522

523 **Figure 3:** Pre-pandemic negative controls: these samples were collected from healthy adults  
524 in the Oxfordshire region of the UK between 2014 and 2016, ethics approval: Oxfordshire  
525 Clinical Research Ethics Committee 08/H0606/107+5. Positive sample set: these were  
526 convalescent plasma donors recruited by NHS Blood and Transplant (NHSBT), ethics approval  
527 (NHSBT; RECOVERY [Cambridge East REC (ref: 20/EE/0101)] and REMAP-CAP [EudraCT 2015-  
528 002340-14] studies).

529

530 **Figures 4, 5 and Table 1:** Known COVID-19 positive samples were collated from three ethically  
531 approved studies: Gastro-intestinal illness in Oxford: COVID substudy [Sheffield REC,

532 reference: 16/YH/0247]ISARIC/WHO, Clinical Characterisation Protocol for Severe Emerging  
533 Infections [Oxford REC C, reference 13/SC/0149], the Sepsis Immunomics project [Oxford REC  
534 C, reference:19/SC/0296]) and by the Scotland A Research Ethics Committee (Ref:  
535 20/SS/0028). Patients were recruited from the John Radcliffe Hospital in Oxford, UK, between  
536 March and May 2020 by identification of patients hospitalised during the SARS-COV-2  
537 pandemic and recruited into the Sepsis Immunomics and ISARIC Clinical Characterisation  
538 Protocols. Time between onset of symptoms and sampling were known for all patients and if  
539 labelled as convalescent patients were sampled at least 28 days from the start of their  
540 symptoms. Written informed consent was obtained from all patients. All patients were  
541 confirmed to have a test positive for SARS-CoV-2 using reverse transcriptase polymerase  
542 chain reaction (RT-PCR) from an upper respiratory tract (nose/throat) swab tested in  
543 accredited laboratories. The degree of severity was identified as mild, severe or critical  
544 infection according to recommendations from the World Health Organisation. Severe  
545 infection was defined as COVID-19 confirmed patients with one of the following conditions:  
546 respiratory distress with RR>30/min; blood oxygen saturation<93%; arterial oxygen partial  
547 pressure (PaO<sub>2</sub>) / fraction of inspired O<sub>2</sub> (FiO<sub>2</sub>) <300 mmHg; and critical infection was defined  
548 as respiratory failure requiring mechanical ventilation or shock; or other organ failures  
549 requiring admission to ICU. Comparator samples from healthcare workers with confirmed  
550 SARS-CoV-2 infection who all had mild non-hospitalised disease were collected under the  
551 Gastro-intestinal illness in Oxford: COVID sub-study, and samples from patients with  
552 equivalently severe disease from non-COVID infection were available from the Sepsis  
553 Immunomics study where patients presenting with significantly abnormal physiological  
554 markers in the pre-pandemic timeframe had samples collected using the same methodology  
555 as that applied during the COVID pandemic. Blood samples were collected in K2EDTA  
556 vacutainers and PBMCs were separated from plasma using Sepmate isolation tubes  
557 (STEMCELL Technologies) and plasma was used in the downstream HAT assay.

558

559 **Figure 6:** Capillary samples were collected from members of institute staff with informed  
560 consent (New Delhi, India). The study is a part of the COVID-19 project 'IPA/2020/000077'.  
561 The project has been approved by the Institutional Human Ethics Committee; Ref. no. -  
562 IHEC#128/20.

563

564 **Cloning, Expression and Purification of VHH(IH4)-RBD**

565 The codon-optimised gene encoding IH4-RBD sequence (Figure 1B and supplementary for the  
566 cDNA sequence) was synthesized by Integrated DNA Technologies. The gene was cloned in  
567 to the AbVec plasmid (Genbank FJ475055) using the restriction sites AgeI and HindIII (the  
568 vector supplied the signal sequence). **This expression plasmid for IH4-RBD is available on  
569 request.** Protein was expressed in Expi293F™ cells using the manufacturer's protocol  
570 (Thermo Fisher). Protein supernatant was harvested on day 5/6 after transfection, spun and  
571 0.22 µm filtered. The protein was affinity-purified using a His-Trap HP column (Cytiva). Binding  
572 buffer consisted of 20 mM Sodium Phosphate, 150 mM NaCl and 20 mM Imidazole at pH 7.4,  
573 and the elution buffer of 500 mM Imidazole in 1 x binding buffer. Protein was concentrated  
574 using 15 ml Vivaspin 30 kDa MWCO filter and then buffer exchanged to PBS using a 10 ml 7  
575 kDa Zebaspin column (Thermo Fisher).

576

577 For large scale production, the protein was synthesized by Absolute Antibody Ltd, Oxford  
578 using the same plasmid construct in HEK293 cells.

579

580 **Lyophilisation of IH4-RBD and CR3022 monoclonal antibody**

581 For lyophilisation, 200 µL (1 mg) of IH4-RBD (5 mg/mL) and 100 µL (200 µg) CR3022 mAb (2  
582 mg/mL) in PBS buffer prepared in Protein Lo-Bind microcentrifuge tube (Fisher Scientific)  
583 were frozen at -80 °C and further cooled down to -196 °C using liquid nitrogen. Pre-cooled  
584 samples were transferred to Benchtop K freeze-dryer (VirTis) with chamber at 49 µbar and  
585 condenser pre-cooled to -72.5 °C. The samples were freeze-dried for a minimum of 24 h,  
586 wrapped in Parafilm (Merck) and stored at -20 °C. Lyophilised sample was reconstituted in  
587 the same original volume of MilliQ water.

588

589 **Indirect ELISA to detect SARS-CoV-2 specific IgG (Figure 6)**

590 A standard indirect ELISA was used to determine the SARS-CoV-2 specific IgG levels in plasma  
591 samples. A highly purified RBD protein from SARS-CoV-2 Wuhan strain (NR-52306, BEI  
592 Resources, USA), expressed in mammalian cells, was used to capture IgG in the plasma  
593 samples. Briefly, ELISA plates (Nunc, Maxisorp) were coated with 100 µL/well of RBD antigen  
594 diluted in PBS (pH 7.4) at the final concentration of 1 µg/mL and incubated overnight at 4°C.  
595 Plates were washed three times with washing buffer (0.05% Tween-20 in PBS) followed by

596 the incubation with blocking buffer (3% Skim milk and 0.05% Tween-20 in PBS). The 3-fold  
597 serially diluted heat inactivated plasma samples in dilution buffer (1% Skim milk and 0.05%  
598 Tween-20 in PBS) were added into the respective wells, followed by incubation at room  
599 temperature for 1 hour. After incubation, plates were washed, and anti-human IgG  
600 conjugated with Horseradish Peroxidase (HRP) (Southern Biotech) was added in each well.  
601 After 1 h incubation, plates were washed and developed by OPD-substrate (Sigma-Aldrich) in  
602 dark at room temperature. The reaction was stopped using 2N HCl and the optical density  
603 (OD) was measured at 492 nm. The RBD-antigen coated wells that were added with sample  
604 diluent alone were used as the blank. The OD values from sample wells were plotted after  
605 subtracting the mean of OD values obtained in the blank wells.

606

#### 607 **Equipment and Reagents for HAT**

- 608 • O-ve blood as a source of red cells collected in K2EDTA tube, diluted in **PBS to 1:20 or**  
609 **1:40 as needed**. Resuspend by inverting gently ~12 times.
- 610 • BD Contact Activated Lancet Cat. No. 366594 (2 mm x 1.5 mm)
- 611 • 100 µL, 20 µL pipettes, Multichannel pipettes
- 612 • V-bottomed 96-well plates (Greiner Bio-One, Cat. No. 651101, Microplate 96-well, PS,  
613 V-bottom, Clear, 10 pieces/bag)
- 614 • Eppendorf Tubes
- 615 • K2EDTA solution (add 5 mL PBS to 10 mL K2EDTA blood collection tube = 3.6 mg  
616 K2EDTA/mL, store at 4 °C)
- 617 • Phosphate Buffered Saline Tablets (OXOID Cat. No. BR0014G)
- 618 • IH4-RBD Reagent diluted 2 µg/mL in PBS. This remains active for at least 1-2 weeks  
619 stored at 4 °C
- 620 • V-bottomed 96-well plates, numbered, dated, timed (helps when timing many plates)
- 621 • Positive control monoclonal antibody CR3022 diluted to 2 µg/mL in PBS

622

#### 623 **Other Reagents**

624 Monoclonal antibody to human IgG (Gamma chain specific) Clone GG-5 Sigma Cat. No. I5885

625

#### 626 **1. Spot test on Stored Serum/Plasma samples (Figure 4).**

- 627 1. Plate out 50  $\mu$ L of **1:20** serum/plasma in alternate columns 1,3,5,7,9,11 (add 2.5  $\mu$ L  
628 sample to 47.5  $\mu$ L PBS).
- 629 2. Add 50  $\mu$ L **1:20** O-ve blood collected in (**so that now sample is diluted to 1:40 and red**  
630 **cells at ~1% v/v**)
- 631 3. Mix and transfer 50/100  $\mu$ L to neighbouring columns 2,4,6,8,10,12 for -ve controls.  
632 The negative control is important because in rare cases, particularly in donors who  
633 have received blood transfusions, the sample in principle may contain antibodies to  
634 non-ABO or Rhesus D antigens.
- 635 4. Add 50  $\mu$ L IH4-RBD reagent (2  $\mu$ g/mL in PBS = 100 ng/well) to Columns 1,3,5,7,9,11
- 636 5. Add 50  $\mu$ L PBS to columns 2,4,6,8,10,12.
- 637 6. Inc 1 hr RT
- 638 7. Tilt for 30 seconds
- 639 8. Photograph: with mobile phone use the zoom function to obtain a complete field
- 640 9. Read as Positive = No teardrop, Negative <1:40 = partial teardrop, Neg = complete  
641 teardrop.
- 642 10. Two readers should read the plates independently, and disagreements resolved by  
643 taking the lesser reading.
- 644 11. For each batch of samples set up positive control wells containing 20-100 ng  
645 monoclonal antibody CR3022 (as in Finger-Prick test below). This establishes that all  
646 of the reagents are working.

647

## 648 **2. Titration of Stored Serum/Plasma Samples.**

- 649 1. Dilute samples to **1:20** in 50  $\mu$ L PBS (2.5  $\mu$ L to 47.5  $\mu$ L) in V-bottomed plate in Rows A-  
650 H, column 1.
- 651 2. Prepare doubling dilutions with PBS across the plate columns 1-11 (1:40 to 1:40,960),  
652 PBS control in column 12. Eight samples can be titrated per 96-well plate.
- 653 3. Add 50  $\mu$ L 1:40 O-ve red cells (1% v/v or 1:40 fresh EDTA O-ve blood sample) to all  
654 wells
- 655 4. Add 50  $\mu$ L IH4-RBD (2  $\mu$ g/mL, = 100 ng/well). [Note: the red cells and IH4-RBD can be  
656 pre-mixed and added together in either 50  $\mu$ L or 100  $\mu$ L volume, to save a step. This  
657 variation in technique does not alter the measured titres.]
- 658 5. Allow red cells to settle for 1 hr

659 6. Tilt plate for at least 30 s and photograph. The titre is defined by the last well in which  
660 the tear drop fails to form. Partial teardrop regarded as negative.

661

### 662 3. Finger-prick test on capillary blood as a Point of Care Test

663 1. Preparation: Clean Hands, warm digit. Prepare a plate (96-well V-bottomed) labelled  
664 with Date and Time.

665 2. Prick skin on outer finger pulp with disposable, single use BD or another Lancet.

666 3. Wipe away first drop of blood with sterile towel/swab

667 4. Massage second drop

668 5. Take a minimum of 5  $\mu\text{L}$  blood with 20  $\mu\text{L}$  pipette, mix immediately into 20  $\mu\text{L}$  K2EDTA  
669 (3.6 mg/mL/PBS) in Eppendorf. If possible, take 20  $\mu\text{L}$  of blood and mix into 80  $\mu\text{L}$   
670 K2EDTA solution. Another approach is collection of blood drops into a BD Microtainer  
671 K2E EDTA lavender vials that take 250-500  $\mu\text{L}$ .

672 6. For 5  $\mu\text{L}$  sample dilute to 200  $\mu\text{L}$  with PBS (add 175  $\mu\text{L}$  PBS), for 20  $\mu\text{L}$  sample dilute to  
673 800  $\mu\text{L}$  (add 700  $\mu\text{L}$  PBS). **Sample is now at 1:40, and the red cells are at the correct**  
674 **density (~1% v/v assuming a haematocrit of 40%) to give a clear tear drop.**

675 7. Plate 50  $\mu\text{L}$  x 3 in V bottomed microtitre wells labelled T (Test), + (PC, positive Control),  
676 - (NC, negative control) – see figure 6.

677 8. Add 10  $\mu\text{L}$  of control anti RBD Mab CR3022 (2  $\mu\text{g}/\text{mL}$  stock in PBS, 20 ng/well) to “+”  
678 well

679 9. Add 50  $\mu\text{L}$  IH4-RBD (2  $\mu\text{g}/\text{mL}$  in PBS) to “T” (Test) and “+ve” wells, 50  $\mu\text{L}$  PBS to “-ve”  
680 well.

681 10. Incubate 1 hour at RT for Red Cells to form a pellet in the “-ve” well

682 11. Tilt plate against a well-lit white background for ~30 seconds to allow Tear drop to  
683 form in “-ve” well.

684 12. The presence of antibodies to RBD is shown by loss of Tear Drop formation in the “T”  
685 and “+ve” wells. Occasionally a partial tear drop forms – these wells are counted as  
686 Negative.

687 13. **Photograph the plate to record the results with the date and time.** Results can be  
688 reviewed and tabulated later. Taking picture from a distance and using the zoom  
689 function helps to take a clear picture of all wells in a 96-well plate.

- 690 14. The negative (PBS) control should be done on every sample for comparison. The  
691 Positive control induced by CR3022 is used to check that all the reagents are working,  
692 and that the glycoprotein epitope recognised by VHH(IH4) is present on the red cells.  
693 Absence of the IH4 epitope should be *very rare* (Habib et al., 2013). For setting up  
694 cohorts a positive control on every sample is therefore not necessary but should be  
695 included in every *batch* of samples.
- 696 15. If a 20 µL sample of blood was taken from the finger prick there should be 650 µL of  
697 the 1:40 diluted blood left. The red cells can be removed and a preparation of 1:40 O-  
698 ve red cells used as above to titrate the sample. In principle the autologous red cells  
699 could be washed repeatedly, resuspended in the same volume of PBS, and used as  
700 indicators for the titration, however we have not attempted to do this. The s/n is 1:40  
701 plasma that can be used in confirmatory ELISA or other tests.

702

### 703 **Author Contributions**

704 **Conceived, initiated and followed the project, documented portability and robustness of**  
705 **HAT method in a separate laboratory, recruited collaborators:** Etienne Joly.

706 **Designed and produced the IH4-RBD reagent, established conditions for lyophilisation,**  
707 **isolated and expanded human monoclonal antibodies to the RBD and ACE2-Fc, performed**  
708 **the standard HAT assays, wrote the paper:** Alain Townsend, Pramila Rijal, Julie Xiao, Tiong  
709 Kit Tan, Lisa Schimanski, Jiangdong Huo, Rolle Rahikainen, Kuan-Ying A Huang.

710 **Contributed examples of HAT as a Point of Care Test:** Nimesh Gupta.

711 **Provision of Serum/Plasma Sample Sets for Figure 4: Project management for Oxford**  
712 **serology work for sensitivity and specificity measurement, assessment of preliminary data:**  
713 Philippa Matthews, Derrick Crook, Sarah Hoosdally, Nicole Stoesser; **collection and**  
714 **processing of samples, coordination of sample banks and running Siemens assay, ethics,**  
715 **storage of pre-pandemic samples, coordination of provision of pre-pandemic samples from**  
716 **Oxford BioBank, and sero-positive donors through NHSBT :** Teresa Street, Justine Rudkin,  
717 Fredrik Karpe, Matthew Neville, Rutger Ploeg, David J Roberts, Abbie Bown, Richard Vipond,  
718 Marta Oliveira, Abigail A Lamikanra, Hoi Pat Tsang.

719 **Provision of Serum Sample Sets for Figure 5 and Table 1 (COMBAT samples):** Alexander J  
720 Mentzer, Julian C Knight, Andrew Kwok, Paul Klenerman, Christina Dold; **ISARIC4C**  
721 **Investigators:** J. Kenneth Baillie, Shona C Moore, Peter JM Openshaw, Malcolm G Semple,

722 Lance CW Turtle; **Oxford Immunology Network Covid-19 Response Clinical Sample**  
723 **Collection Consortium:** Mark Ainsworth, Alice Allcock, Sally Beer, Sagida Bibi, Elizabeth  
724 Clutterbuck, Alexis Espinosa, Maria Mendoza, Dominique Georgiou, Teresa Lockett, Jose  
725 Martinez, Elena Perez, Veronica Sanchez, Giuseppe Scozzafava, Alberto Sobrinodiaz, Hannah  
726 Thraves.

727

## 728 **ACKNOWLEDGEMENT AND FUNDING**

729 A.T. is funded by the Medical Research Council (MR/P021336/1), Townsend-Jeantet  
730 Charitable Trust (charity number 1011770) and the Chinese Academy of Medical Sciences  
731 (CAMS) Innovation Fund for Medical Science (CIFMS), China (grant no. 2018-I2M-2-002). N.G.  
732 is funded by the Science Engineering Research Board, Department of Science and Technology,  
733 India. P.C.M. is funded by the Wellcome Trust (grant ref 110110Z/15/Z). D.R. is supported by  
734 NIHR Oxford Biomedical Research Centre. National Institute for Health Research Biomedical  
735 Research Centre Funding Scheme (to G.R.S.), the Chinese Academy of Medical Sciences  
736 (CAMS) Innovation Fund for Medical Science (CIFMS), China (grant number: 2018-I2M-2-  
737 002). G.R.S. is supported as a Wellcome Trust Senior Investigator (grant 095541/A/11/Z).

738

739 The Spike Glycoprotein Receptor Binding Domain (RBD) from SARS-Related Coronavirus 2  
740 Wuhan-Hu-1 (Figure 6), Recombinant from HEK293 Cells, NR-52306 was produced under  
741 HHSN272201400008C and obtained through BEI Resources, NIAID, NIH.

742

743 We thank the healthy volunteers who kindly donated their O-ve red cells for titration of  
744 samples.

745

## 746 **DECLARATION**

747 Competing Interests none.

748 The views expressed are those of the author(s) and not necessarily those of the NHS, the  
749 NIHR, the Department of Health or Public Health England'.

750

## 751 **REFERENCES**

752

753 Adams, Emily R., Ainsworth, Mark, Anand, Rekha, Andersson, Monique I., Auckland, Kathryn, Baillie,  
754 J. Kenneth, . . . Whitehouse, James. (2020). Antibody testing for COVID-19: A report from  
755 the National COVID Scientific Advisory Panel. *Wellcome Open Research*, 5.  
756 doi:10.12688/wellcomeopenres.15927.1

757 Amanat, F., Stadlbauer, D., Strohmeier, S., Nguyen, T. H. O., Chromikova, V., McMahon, M., . . .  
758 Krammer, F. (2020). A serological assay to detect SARS-CoV-2 seroconversion in humans. *Nat*  
759 *Med*, 26(7), 1033-1036. doi:10.1038/s41591-020-0913-5

760 Barnes, C. O., West, A. P., Jr., Huey-Tubman, K. E., Hoffmann, M. A. G., Sharaf, N. G., Hoffman, P. R., .  
761 . . Bjorkman, P. J. (2020). Structures of Human Antibodies Bound to SARS-CoV-2 Spike Reveal  
762 Common Epitopes and Recurrent Features of Antibodies. *Cell*, 182(4), 828-842 e816.  
763 doi:10.1016/j.cell.2020.06.025

764 Coombs, R. R., Mourant, A. E., & Race, R. R. (1945). A new test for the detection of weak and  
765 incomplete Rh agglutinins. *Br J Exp Pathol*, 26, 255-266.

766 Dai, L., Zheng, T., Xu, K., Han, Y., Xu, L., Huang, E., . . . Gao, G. F. (2020). A Universal Design of  
767 Betacoronavirus Vaccines against COVID-19, MERS, and SARS. *Cell*, 182(3), 722-733 e711.  
768 doi:10.1016/j.cell.2020.06.035

769 Habib, I., Smolarek, D., Hattab, C., Grodecka, M., Hassanzadeh-Ghassabeh, G., Muyldermans, S., . . .  
770 Bertrand, O. (2013). V(H)H (nanobody) directed against human glycoporphin A: a tool for  
771 autologous red cell agglutination assays. *Anal Biochem*, 438(1), 82-89.  
772 doi:10.1016/j.ab.2013.03.020

773 Hanson, K. E., Caliendo, A. M., Arias, C. A., Englund, J. A., Hayden, M. K., Lee, M. J., . . . Mustafa, R. A.  
774 (2020). Infectious Diseases Society of America Guidelines on the Diagnosis of COVID-  
775 19:Serologic Testing. *Clin Infect Dis*. doi:10.1093/cid/ciaa1343

776 Hirst, G. K. (1941). The Agglutination of Red Cells by Allantoic Fluid of Chick Embryos Infected with  
777 Influenza Virus. *Science*, 94(2427), 22-23. doi:10.1126/science.94.2427.22

778 Huang, Kuan-Ying A., Tan, Tiong Kit, Chen, Ting-Hua, Huang, Chung-Guei, Harvey, Ruth, Hussain, Saira,  
779 . . . Townsend, Alain R. (2020). Plasmablast-derived antibody response to acute SARS-CoV-2  
780 infection in humans. *bioRxiv*. doi:10.1101/2020.08.28.267526

781 Huo, J., Le Bas, A., Ruza, R. R., Duyvesteyn, H. M. E., Mikolajek, H., Malinauskas, T., . . . Naismith, J. H.  
782 (2020). Neutralizing nanobodies bind SARS-CoV-2 spike RBD and block interaction with ACE2.  
783 *Nat Struct Mol Biol*, 27(9), 846-854. doi:10.1038/s41594-020-0469-6

784 Huo, J., Zhao, Y., Ren, J., Zhou, D., Duyvesteyn, H. M. E., Ginn, H. M., . . . Stuart, D. I. (2020).  
785 Neutralization of SARS-CoV-2 by Destruction of the Prefusion Spike. *Cell Host Microbe*, 28(3),  
786 497. doi:10.1016/j.chom.2020.07.002

787 Kruse, Robert L., Huang, Yuting, Smetana, Heather, Gehrie, Eric A., Amukele, Tim K., Tobian, Aaron A.  
788 R., . . . Wang, Zack Z. (2020). A rapid, point of care red blood cell agglutination assay for  
789 detecting antibodies against SARS-CoV-2. *bioRxiv*. doi:10.1101/2020.05.13.094490

790 Lan, Jun, Ge, Jiwan, Yu, Jinfang, Shan, Sisi, Zhou, Huan, Fan, Shilong, . . . Wang, Xinquan. (2020). Crystal  
791 structure of the 2019-nCoV spike receptor-binding domain bound with the ACE2 receptor.  
792 *bioRxiv*. doi:10.1101/2020.02.19.956235

793 Landsteiner, K. (1961). On agglutination of normal human blood. *Transfusion*, 1, 5-8.  
794 doi:10.1111/j.1537-2995.1961.tb00005.x

795 Li, F., Li, W., Farzan, M., & Harrison, S. C. (2005). Structure of SARS coronavirus spike receptor-binding  
796 domain complexed with receptor. *Science*, 309(5742), 1864-1868.  
797 doi:10.1126/science.1116480

798 Mulligan, M. J., Lyke, K. E., Kitchin, N., Absalon, J., Gurtman, A., Lockhart, S., . . . Jansen, K. U. (2020).  
799 Phase 1/2 study of COVID-19 RNA vaccine BNT162b1 in adults. *Nature*. doi:10.1038/s41586-  
800 020-2639-4

801 Pamphilon, D. H., & Scott, M. L. (2007). Robin Coombs: his life and contribution to haematology and  
802 transfusion medicine. *Br J Haematol*, 137(5), 401-408. doi:10.1111/j.1365-2141.2007.06531.x

803 Peng, Y., Mentzer, A. J., Liu, G., Yao, X., Yin, Z., Dong, D., . . . Dong, T. (2020). Broad and strong memory  
804 CD4(+) and CD8(+) T cells induced by SARS-CoV-2 in UK convalescent individuals following  
805 COVID-19. *Nat Immunol*. doi:10.1038/s41590-020-0782-6

806 Piccoli, Luca, Park, Young-Jun, Tortorici, M. Alejandra, Czudnochowski, Nadine, Walls, Alexandra C.,  
807 Beltramello, Martina, . . . Veesler, David. (2020). Mapping neutralizing and immunodominant  
808 sites on the SARS-CoV-2 spike receptor-binding domain by structure-guided high-resolution  
809 serology. *Cell*. doi:10.1016/j.cell.2020.09.037

810 Pinto, D., Park, Y. J., Beltramello, M., Walls, A. C., Tortorici, M. A., Bianchi, S., . . . Corti, D. (2020). Cross-  
811 neutralization of SARS-CoV-2 by a human monoclonal SARS-CoV antibody. *Nature*, 583(7815),  
812 290-295. doi:10.1038/s41586-020-2349-y

813 Robbiani, D. F., Gaebler, C., Muecksch, F., Lorenzi, J. C. C., Wang, Z., Cho, A., . . . Nussenzweig, M. C.  
814 (2020). Convergent antibody responses to SARS-CoV-2 in convalescent individuals. *Nature*,  
815 584(7821), 437-442. doi:10.1038/s41586-020-2456-9

816 Schwarz, H. P., & Dorner, F. (2003). Karl Landsteiner and his major contributions to haematology. *Br J*  
817 *Haematol*, 121(4), 556-565. doi:10.1046/j.1365-2141.2003.04295.x

818 Tan, Tiong Kit, Rijal, Pramila, Rahikainen, Rolle, Keeble, Anthony H., Schimanski, Lisa, Hussain, Saira, .  
819 . . Townsend, Alain R. (2020). COVID-19 vaccine candidate using SpyCatcher multimerization

820 of the SARS-CoV-2 spike protein receptor-binding domain induces potent neutralising  
821 antibody responses. *bioRxiv*. doi:10.1101/2020.08.31.275701

822 ter Meulen, J., van den Brink, E. N., Poon, L. L., Marissen, W. E., Leung, C. S., Cox, F., . . . Goudsmit, J.  
823 (2006). Human monoclonal antibody combination against SARS coronavirus: synergy and  
824 coverage of escape mutants. *PLoS Med*, 3(7), e237. doi:10.1371/journal.pmed.0030237

825 The National SARS-CoV-2 Serology Assay Evaluation Group, The National SARS-CoV-2 Serology Assay  
826 Evaluation Group. (2020). *Manuscript preprint: Head-to-head benchmark evaluation of the*  
827 *sensitivity and specificity of five immunoassays for SARS-CoV-2 serology on >1500 samples.*

828 Walls, A. C., Fiala, B., Schafer, A., Wrenn, S., Pham, M. N., Murphy, M., . . . King, N. P. (2020). Elicitation  
829 of potent neutralizing antibody responses by designed protein nanoparticle vaccines for SARS-  
830 CoV-2. *bioRxiv*. doi:10.1101/2020.08.11.247395

831 Wang, Q., Zhang, Y., Wu, L., Niu, S., Song, C., Zhang, Z., . . . Qi, J. (2020). Structural and Functional Basis  
832 of SARS-CoV-2 Entry by Using Human ACE2. *Cell*, 181(4), 894-904 e899.  
833 doi:10.1016/j.cell.2020.03.045

834 Wegmann, T. G., & Smithies, O. (1966). A Simple Hemagglutination System Requiring Small Amounts  
835 of Red Cells and Antibodies. *Transfusion*.

836 Wrapp, D., De Vlieger, D., Corbett, K. S., Torres, G. M., Wang, N., Van Breedam, W., . . . McLellan, J. S.  
837 (2020). Structural Basis for Potent Neutralization of Betacoronaviruses by Single-Domain  
838 Camelid Antibodies. *Cell*, 181(6), 1436-1441. doi:10.1016/j.cell.2020.05.047

839 Yan, R., Zhang, Y., Li, Y., Xia, L., Guo, Y., & Zhou, Q. (2020). Structural basis for the recognition of SARS-  
840 CoV-2 by full-length human ACE2. *Science*, 367(6485), 1444-1448.  
841 doi:10.1126/science.abb2762

842 Yang, J., Wang, W., Chen, Z., Lu, S., Yang, F., Bi, Z., . . . Wei, X. (2020). A vaccine targeting the RBD of  
843 the S protein of SARS-CoV-2 induces protective immunity. *Nature*. doi:10.1038/s41586-020-  
844 2599-8

845 Zha, Lisha, Zhao, Hongxin, Mohsen, Mona O., Hong, Liang, Zhou, Yuhang, Li, Zehua, . . . Wang, Junfeng.  
846 (2020). Development of a COVID-19 vaccine based on the receptor binding domain displayed  
847 on virus-like particles. *bioRxiv*. doi:10.1101/2020.05.06.079830

848 Zhou, D., Duyvesteyn, H. M. E., Chen, C. P., Huang, C. G., Chen, T. H., Shih, S. R., . . . Huang, K. A. (2020).  
849 Structural basis for the neutralization of SARS-CoV-2 by an antibody from a convalescent  
850 patient. *Nat Struct Mol Biol*. doi:10.1038/s41594-020-0480-y

851


852 **SUPPLEMENTARY DATA**

853

854 Supplementary Table 1. Fifteen human monoclonal antibodies or nanobodies with endpoint  
 855 titres for detection in the HA Test. One divalent nanobody VHH72-Fc and two monoclonal  
 856 antibodies FD-5D and EZ-7A specific for the RBD failed to agglutinate red cells with the IH4-  
 857 RBD reagent. Three monoclonal antibodies to other regions of the spike protein EW-8B, EW-  
 858 9C and FJ-1C acted as negative controls.

859

Mab or divalent Nanobody	HAT endpoint titre (ng/well)	Reference
CR3022	2	(Huo, Zhao, et al., 2020; ter Meulen et al., 2006)
EY-6A	16	(Zhou et al., 2020)
VHH-72-Fc	Negative	(Wrapp et al., 2020)
FI-4A	16	(Huang et al., 2020)
ACE2-Fc	4	(Huo, Le Bas, et al., 2020)
FI-3A	8	(Huang et al., 2020)
FI-1C	4	(Huang et al., 2020)
H11-H4-Fc	16	(Huo, Le Bas, et al., 2020)
FD-5D	Negative	(Huang et al., 2020)
FD-11A	31	(Huang et al., 2020)
FN-12A	31	(Huang et al., 2020)
FJ-10B	62	(Huang et al., 2020)
FM-7B	62	(Huang et al., 2020)
EZ-7A	Negative	(Huang et al., 2020)
S309	125	(Pinto et al., 2020)
EW-8B non-RBD anti Spike	Neg Control	(Huang et al., 2020)
EW-9C non-RBD anti Spike	Neg Control	(Huang et al., 2020)
FJ-1C anti NTD	Neg Control	(Huang et al., 2020)



861

862

863

864

865

866


867

868

869

870

Supplementary Figure 1 A, B) Three MAbs or Nanobody-Fc reagents of 15 tested specific for RBD failed to cross-link red cells but could be developed with an anti IgG reagent. Antibodies were titrated from 500 ng/well to 1 ng/well in 50  $\mu$ L, 1:40 red cells were added in 50  $\mu$ L, followed by 100 ng/well IH4-RBD in 50  $\mu$ L, and incubated for 1 hour. The labelled red cells were then resuspended, and either 50  $\mu$ L PBS (L hand plates), or Sigma Clone GG5 anti human IgG 1:100 in 50  $\mu$ L (R hand plates). Red cells were allowed to settle for one hour, the plate tilted for 30s and photographed.




Slope = 0.94 (95% CI 0.92 - 0.96)


Suggests second set has tendency to detect ~1 doubling dilution (DD) lower titres

230/232 (99%) were within 1 DD

**B Titration in HAT by doubling dilution (Repeat)**



Donors were grouped according to clinical criteria; HCW = Front Line Health Care Workers; Sepsis = Samples with patients with Sepsis prior to Covid-19 pandemic; Healthy = Healthy volunteers prior to Covid-19; DD 1 = 1:40. Actual Titre =  $20 \times 2^{DD}$ .



	HAT +ve	HAT -ve	Total	Fraction
Siemens +ve	108	5	113	0.96
Siemens -ve	16	24	40	0.40
Total	124	29	153	0.81
Fraction	0.87	0.83	0.74	

P value and statistical significance	
Test	Fisher's exact test
P value	<0.0001
P value summary	****
One- or two-sided	Two-sided

Sensitivity (Sn) HAT 81%, Sn Siemens 74%  
HAT still more sensitive than Siemens

Specificity 100% for HAT in these samples

871

872 Supplementary Figure 2. A) The 232 samples from figure 5 in the main paper were titrated a

873 second time 34 days after the first measurement, as described for figure 4 (Methods). The

874 values for the end point doubling dilution were compared and a correlation coefficient, and

875 slope calculated (Prism v8).

876 B) The collection included 32 samples from 24 Critical patients, 62 samples from 48 Severe,  
 877 39 samples from 32 Mild, 20 single samples from health care workers (HCW), 54 samples from  
 878 43 patients with unrelated sepsis in the pre-Covid-19 era, 25 samples from healthy unexposed  
 879 controls. Median is shown.

880 C) Comparison to Siemens Result (anti RBD) with HAT Titre by Doubling Dilution for 153  
 881 samples from Critical, Severe, Mild and HCW PCR+ve donors. The sensitivity of the HAT in this  
 882 repeat was 81%, v 86% in the first test.

883

884 Supplementary Table 2. Fifty-two samples from 24 donors who were sampled repeatedly  
 885 during the first five days in hospital; assay repeated 34 days after the measurements in Table  
 886 1. Sixteen of 24 showed a rise in titre.

		Day				
No	Case	1	2	3	4	5
1	C2			40		160
2	C6			0		80
3	C7	0				0
4	C9	0				1280
5	C10	160		320		
6	C23	0		80		320
7	S5	40		40		
8	S6	1280		2560		2560
9	S7	80				1280
10	S12	320		640		
11	S13	0		0		160
12	S14			640		1280
13	S17			1280		1280
14	S20	80		320		
15	S31	0				640
16	S41	640		1280		
17	S43			40		80
18	S48			1280		1280
19	M9			40		160
20	M11	0		0		0
21	M13			0		0
22	M14			163840		163840
23	M25	40				320
24	M30			160		160

887

888

889 **DNA sequence of codon optimised cDNA encoding IH4-RBD**

890 ATGGGATGGTCATGTATCATCCTTTTTCTAGTAGCAACTGCAACCGGTGTTTCATAGCCAGGTC  
891 CAGCTGCAAGAGTCTGGCGGAGGATCTGTTCAAGGCTGGCGGAAGCCTGAGACTGAGCTGTGT  
892 GGCCAGCGGCTACACCGATAGCACATACTGCGTCGGCTGGTTCAGACAGGCCCTGGCAAAG  
893 AGAGAGAGGGCGTCGCCAGAATCAACACCATCAGCGGCAGACCTTGGTACGCCGACTCTGTG  
894 AAGGGCAGATTCACAATCAGCCAGGACAACAGCAAGAACACCGTGTCTGCAGATGAACA  
895 GCCTGAAGCCAGAGGACACCGCCATCTACTACTGCACCCTGACCACCGCCAACAGCAGAGGC  
896 TTTTGTTCGGCGGCTACAATAAAAGGCCAGGGCACCCAAGTGACCGTGTCTAGCGCGTC  
GACCGGCTCTGGCGGCAGCGGCAACATCACCAATCTGTGCCCTTTCGGCGAGGTGTTCAACG  
CCACCAGATTTGCCAGCGTGTACGCCTGGAACCGGAAGAGAATCAGCAACTGCGTGGCCGAC  
TACAGCGTGCTGTACAATAGCGCCAGCTTCAGCACCTTCAAGTGCTACGGCGTGTCCCCTACC  
AAGCTGAACGACCTGTGCTTCACCAATGTGTACGCCGACAGCTTCGTGATCAGAGGCGACGA  
AGTTTCGGCAGATCGCTCCTGGACAGACAGGCAAGATCGCCGATTACAATAAAGCTGCCCCG  
ACGACTTCACCGGCTGCGTGATCGCCTGGAATAGCAACAACCTGGACAGCAAAGTCGGCGGC  
AACTACAATACTGTACCGGCTGTTCCGGAAGTCCAACCTGAAGCCTTTCGAGCGGGACAT  
CAGCACCGAGATCTATCAGGCCGGCAGCACCCCTTGTAATGGCGTGGAAGGCTTCAACTGCT  
ACTTCCCCTGCAGTCCTACGGCTTTCAGCCTACAACGGCGTGGGCTACCAGCCTTATAGAG  
TGGTGGTGCTGAGCTTCGAACTGCTGCATGCCCTGCTACCGTGTGCGGCCCTAAAAAACACC  
ATCACCACCACCATTGA