Impact on microbiology laboratory turnaround times following process improvements and total laboratory automation.

Carolyn Gonzalez-Ortiz, M(ASCP)CM,[1] Alanna Emrick, MLS(ASCP)CM SMCM, [1] Ying P. Tabak, PhD,[2] Latha Vankeepuram, MS,[2] Stephen Kurtz, MS,[2] David Sellers, BS,[2] Megan Wimmer, MBA,[2] Caitlin Asjes, MS,[2] Suhaireirene Suady Barake, MD,[3] Jacob Nichols, MD,[3] Fatma Levent, MD[3#]

Affiliations:

[1] University Medical Center, Lubbock, TX

[2] Becton, Dickinson and Company, Franklin Lakes, NJ

[3]Texas Tech University Health Sciences Center, Lubbock, TX

#To whom correspondence should be addressed:

Fatma Levent, MD, Clinical Associate Professor

Pediatric Infectious Diseases / Department of Pediatrics

Texas Tech Health Sciences Center

Lubbock, TX

Phone: 806-392-4976

Fax: 806-743-2314

Email: fatma.levent@ttuhsc.edu

Running title: Improved lab efficiency with automation

Abbreviations: TLA, total laboratory automation; SOP, standard operating procedure; MALDI-TOF, matrix-assisted laser desorption ionization time-of-flight mass spectrometry; TAT,

turnaround time; ID, identification; ASM, American Society of Microbiology; CLSI, Clinical

Laboratory Standards Institute; MT, medical technologist

ABSTRACT

Introduction: The impact of workflow changes and total laboratory automation (TLA) on

microbiology culture processing time was evaluated in an academic-affiliated regional hospital.

Materials and Methods: A retrospective analysis of microbiological data in a research database

was performed to compare turnaround time (TAT) for organism identification (ID) before and

after implementation of TLA (2013 versus 2016, respectively). TAT was compared using the $\chi2$

test for categorical variables and log-transformed t-test for continuous variables.

Results: A total of 9,351 predefined common and clinically important positive mono-bacterial

culture results were included in the analysis. Shorter TAT (hours) in 2016 compared to 2013

(p<0.0001) for positive result pathogen ID were observed in specimen types including blood

(51.2 vs. 70.6), urine (40.7 vs. 47.1), wound (39.6 vs. 60.2), respiratory (47.7 vs. 67), and all

specimen types combined (43.3 vs. 56.8). Although shorter TATs were not observed from all

specimen categories for negative result pathogen ID, TAT for all specimen types combined was

shorter (p≤0.001) in 2016 compared to 2013 (94 vs. 101).

Conclusions: Total laboratory automation and workflow changes—including process

standardization—facilitate shorter organism ID TAT across specimen sources.

Key Words: Laboratory automation, Microbiology, Clinical laboratory, Laboratory workflow,

Turnaround time

INTRODUCTION

Total laboratory automation (TLA) involves the integration of pre- and post-analytical processes

through incorporation of multiple, modular systems to perform different tests, on diverse

matrices, within the clinical microbiology laboratory, and typically consists of automation for

plating, incubation, reading, and processing of specimens. Through reduced manual processing

of specimens, standardization can be established to minimize the incidence of human error.[1]

Studies have demonstrated the clinical value of automating microbiological workflow including

enhanced microbial growth, better colony isolation, reduced requirements for bacterial

subculture, and reduced time to results.[2-6] Previous work suggests that TLA can facilitate rapid

reporting of microbiology culture testing, thereby informing the management of infectious

disease and leading to enhanced patient care.[6-8]

This need for laboratory automation comes from an increased demand for clinical microbiology

laboratory services, despite growing staffing challenges.[9-11] Successful automation of a

laboratory requires a thorough assessment and appropriate refinement of all laboratory workflow

practices that might impact system efficiency.[12, 13] This includes consideration of extending

service delivery to a twenty-four-hour day, seven days per week (24/7) schedule.[3, 13, 14] Such

changes facilitate the maximum benefit from laboratory automation.

Previous work demonstrated that TLA can also improve laboratory workflow, efficiency, and

reproducibility[6, 9, 15, 16] while reducing errors.[4, 8] Staffing reductions of approximately 30%

have been reported after automation despite a 27% increase in average laboratory workload per

day.[3] Thus, TLA addresses the challenges associated with a high volume workload and allows

managers to reallocate human resources to value-added activities.

In 2014, the University Medical Center (UMC), Lubbock, TX prioritized improvements to

quality and efficiency in the microbiology laboratory, including moving to TLA. After an

assessment of available TLA solutions, UMC leadership supported investment in the BD

Kiestra™ TLA system (Kiestra; Becton, Dickinson and Company, BD Life Sciences—BD

Integrated Diagnostic Systems, Sparks, MD, USA), which automates and standardizes sample

setup, incubation, and reading. UMC became the second laboratory in the United States to

implement Kiestra TLA and the first to both operate 24/7 as well as have all microbial sample

types on the system. Preparation for the transition to automation included changes in laboratory

policy and processes. The hypothesis here was that a combination of TLA and foundational

workflow improvements (made in anticipation of the new technology associated with TLA)

would result in a shorter post-TLA turnaround time (TAT). Although previous studies have

evaluated the use of TLA for single specimen sources,[5-7, 17-22] the goal here was to evaluate

TAT across a wide range of specimen sources.

MATERIALS AND METHODS

Study Design

We compared microbiological result TAT before and after implementation of TLA and process

improvements, including implementation of matrix-assisted laser desorption ionization time-of-

flight mass spectrometry (MALDI-TOF) (Bruker Daltonik, Germany).[23] The laboratory process

improvement preparation started in 2014 and the TLA system went live in May 2015. The entire

calendar years of 2013 and 2016 were used as pre- and post-periods, respectively, to evaluate

TAT before and after laboratory technology and process changes. The year 2016 was selected as

steady state to account for an implementation preparation period and a process-stabilization

period after TLA implementation in 2015. This study was approved by the Quality Improvement

Review Board at Texas Tech University Health Sciences Center. This work was conducted in a

manner consistent with the ideals set forth by the World Medical Association Declaration of

Helsinki regarding ethical conduct of research involving human subjects.

Process improvement in preparation for TLA

Review of SOPs

Process improvements and total laboratory automation at the study site went through a stepwise

and intricate process (see Figure 1). Initial steps were taken to ensure that process improvement

evaluations were generated/updated that align with the College of American Pathologists

guidelines for an accredited laboratory (https://www.cap.org/laboratory-

improvement/accreditation/laboratory-accreditation-program). In April 2014, all SOPs were

modified as needed according to procedural need; however, they remained aligned with

American Society of Microbiology (ASM)[24] and Clinical Laboratory and Standards Institute

(CLSI) guidelines,[25] as well as community practice standards, in an effort that involved

networking with physician groups.

Staff education

All medical technologists (MTs) were actively engaged in professional education, assigned

specific topics, and responsible for retraining the department members on their respective roles

and monitoring compliance with the updated workflow practices. All MTs were educated and

tested on best practice as described in the *Clinical Microbiology Procedures Handbook*.[10]

Streamlined culture order code options and standardization of agar plate setup

The UMC microbiology laboratory standardized culture setups and streamlined the number of

order codes from 49 to 31. The standardization project resolved variability by narrowing plate

setups to a single option and setting clear expectations for reading plates. Completing these

projects simultaneously in November 2014 allowed for all MTs to be trained on all sources in

preparation for the automation line, where cultures would be prepared for reading during setup.

Digital imaging offered on the TLA system enabled plate images to be taken at ideal incubation

times. This allowed reading of the plates to be optimally timed by reviewing the images rather

than the plates themselves.[14, 18, 26]

Improved communication between laboratory and physician groups

Communication between laboratory management and physician/clinician groups during the TLA

implementation period was approached here in a manner similar to that reported previously.[27-

29] A diagnostic management team was created prior to commencing the transition to TLA. The

team consisted of laboratory management and different physician/clinician groups, and included

several goals. First, the team was created to ensure more consistent, open communication

between the laboratory and clinicians. Initial meetings between physician/clinician and

laboratory management included communication on process changes for ordering, testing, and

downstream activities of reporting, charting, and consultation. Team meetings covered

adjustments by the clinician to receive laboratory results in a reduced time frame in order to

convert the time savings to a potential benefit for patients through evidence-based practices.

Feedback from physician/clinician groups was received by the laboratory regarding TLA-related

issues such as patient charting based on the previous collection day. Related to charting,

discussion items included any changes in test selection and result interpretation based on TLA

implementation. Laboratory updates were provided to physician/clinician groups at team

meetings throughout the TLA implementation process. Post-implementation meetings revealed

physician/clinician perspectives on progress based on initial goal setting for processes including

reporting and charting. Feedback was also provided to the laboratory for continued improvement.

Pre- and Post-TLA sample processing

During the 2013 study period, specimens arrived for processing throughout the day and specimen

login to the laboratory information system was considered the initial time point for TAT (this

also applies to 2016). For the conventional plating technique, urine specimens were streaked

manually using calibrated loops onto blood and MacConkey agar plates and placed into

incubators. Specimen plating was performed in batches and incubation initially occurred

overnight (minimum of 18 hours). Plate reading was performed on four source benches,

including blood/cerebral spinal fluid, wound/tissue/body fluid, urine, and stool/respiratory, each

combining all plates into one batch during the morning. However, upon review, many plates

were being read early at 16 hours per work day from 6 am to 10 pm (i.e. many plates that were

prepared very early during the day shift [~6 am] were read at the end of the day shift [~10 pm]).

Spot testing, which included oxidase catalase, pyrrolidonyl arylamidase, desoxycholate, indole,

numerous latex agglutination tests, urease, and coagulase tests for Staphylococcus aureus, were

conducted when required for identification. The BD Phoenix™ system (Becton, Dickinson and

Company, BD Integrated Diagnostic Systems, Sparks, MD), A disk, P disk, RapID™ Systems

(Thermo Fisher, Waltham, MA), or API® ID Strips (bioMérieux, Marcy-l'Étoile, France) were

utilized for identification following all spot testing.

Kiestra TLA was implemented in May 2015 (Figure 1). Using TLA, specimens were inoculated

upon reception (no batching) onto the appropriate agar plate type by the BD Kiestra InoqulA+

system (InoqulA), and plates were automatically transported via BD Kiestra ProceedA to ReadA

Compact (incubation). Digital images of the plates were automatically acquired at 18, 42, and 66

hours depending on the sample type (i.e. 18 hours for day one, 42 hours for day two, and 66

hours for day three). Positive plates were sent to the ErgonomicA module, where a technologist

performed MALDI-TOF MS for identification. Spot testing was reduced to only indole and

desoxycholate for MALDI confirmation requirements, and coagulase and latex agglutination for

a more rapid ID of S. aureus and β-hemolytic streptococci, respectively. Bruker MALDI-TOF

MS was implemented in February 2016 for gram-negative IDs; gram-positive IDs were added in

October 2016, and anaerobe and yeast IDs were added in April 2017. MALDI-TOF MS replaced

identification by conventional methods, and times to result for positive specimens, preliminary

negatives, and antimicrobial test reporting were recorded as described above. All other spot

testing and ID methods were fully replaced with MALDI-TOF MS. A similar workflow was

utilized to identify cultures as negative. Plates identified as no growth or no pathogens at 18

hours were further incubated for an additional 24 hours for a total of 42 hours before a second

image analysis was performed. Times to result for positive specimens and preliminary negatives

were as described above.

Specimen Source and Organism Identification

Specimens were classified by source as blood, urine, wound, respiratory, intra-abdominal, and

other. The top ten common and clinically relevant bacteria seen in the laboratory were

Staphylococcus aureus, Escherichia coli, Klebsiella pneumoniae, Streptococcus pneumoniae,

Streptococcus pyogenes (Group A), Streptococcus agalactiae (Group B), Pseudomonas

aeruginosa, Proteus mirabilis, Enterococcus faecalis, and Enterococcus faecium. From the

microbiology laboratory, results of 19.3% (2013) and 20.2% (2016) of the total results in each

respective year are included in this study. The analysis was restricted to positive specimen results

for mono-bacterial isolates of interest and negative culture results.[7, 18, 24, 30]

Definition of definitive results TAT

We assessed definitive positive organism identification (ID) TAT as time from laboratory login

of specimen to the earliest definitive positive result reporting time. Likewise, the TAT for

negative results was from laboratory login to the earliest definitive negative result report.

Data Source

De-identified digital microbiological data were captured in a research database (BD Insights)

created by Becton, Dickinson and Company (Franklin Lakes, NJ).[19, 20, 22] The dataset for the

current study included specimen collection time, source, laboratory login time, and culture

results reporting time.

Statistical analysis

We analyzed organism ID reporting time over a 24-hour spectrum. Definitive positive organism

ID and negative results TAT for pre- versus post-periods were compared. The χ2 test for

categorical variables and log-transformed t-test for continuous variables were utilized for

comparisons. A two-tailed p-value of <0.05 was considered statistically significant. All analyses

were conducted using SAS version 9.4 (SAS Institute, Cary, NC).

RESULTS

Overall 4,510 positive results were reported (Table 1). The mean TAT for all specimen types

was significantly shorter after TLA and process change implementation (2106) compared to pre-

TLA (2013) (all p<0.05).The overall TAT mean (standard deviation) was 43.3 (20.8) hours

versus 56.8 (24.3) hours for years 2016 and 2013, respectively. Figure 2 depicts the comparison

of cumulative distribution of TAT for positive IDs for the four most common specimen sources

(blood, urine, wound, and respiratory; all p<0.05). For blood specimens, the ID TAT was 53%

(2016) versus 17% (2013) by 48 hours from laboratory login time (Figure 2a). For samples

derived from urine, wound, and respiratory specimens, a higher percentage of results was

returned within 24 hours in 2016 compared to 2013. At least 40% of the urine, wound, and

respiratory specimens had TAT <36 hours in 2016, whereas in 2013, 40% of wound and

respiratory specimens had TAT >36 hours, and 40% of urine TAT exceeded 24 hours. For urine

specimens, the ID TAT was 51% (2016) versus 14% (2013) by 36 hours from laboratory login

time (Figure 2b). For respiratory specimens, the ID TAT was 59% (2016) versus 21% (2013) by

48 hours from laboratory login time (Figure 2c). For wound specimens, the ID TAT was 77%

(2016) versus 39% (2013) by 48 hours from laboratory login time (Figure 2d).

Overall 58,640 negative results were reported (Table 2). All sources except for respiratory and

"other" showed shortened TAT (all p<0.05). Figure 3 depicts the comparison of cumulative

distribution of negative result TAT for the four most common sources (blood, urine, wound, and

respiratory; all p<0.05). For blood specimens, the TAT was 95% (2016) versus 73% (2013) by

128 hours from laboratory login time (Figure 3a). TAT for specimens derived from urine and

wound showed the greatest reduction in 2016 compared to 2013. For urine specimens, the TAT

was 83% (2016) versus 10% (2013) by 36 hours from laboratory login time (Figure 3b). For

wound specimens, the ID TAT was 71% (2016) versus 36% (2013) by 96 hours from laboratory

login time (Figure 3c). For respiratory specimens, the negative results TAT was 79% (2016)

versus 86% (2013) by 96 hours from laboratory login time (Figure 3d). Negative results TAT

was less variable across the time frame categories for blood and respiratory between 2013 and

2016. However, the 80th percentile of TAT results for blood were returned within 126 to 128

hours in 2016, while those in 2013 were delayed to between 128 to 240 hours.

DISCUSSION

The results here suggest that TLA, in conjunction with process changes that facilitate TLA, led

to improvement in TAT for microbial ID. Although the observed shorter TAT was likely a joint

effect of intricate laboratory process improvement and implementation of the TLA, these results

are in agreement with previous studies that show a positive correlation between laboratory

automation and faster TAT. 18-22, 28 In general, the overall laboratory process improvement

interventions and TLA implementation (2016) led to shorter TAT for positive results compared

to 2013. For the highest volume specimen, urine cultures, pathogen ID TAT improved by an

average of 6.4 hours from 2013 to 2016. Since TLA processes and images the plates under

consistent incubation conditions, facilitating more robust growth,18 a 16 to 30-hour growth

period was no longer necessary, which likely facilitated earlier ID. Blood cultures showed

significant improvement in TAT to pathogen positive ID by 19 hours, potentially enabling

providers to update their course of treatment by almost an entire day for patients with

bacteremia. In addition to urine and blood cultures, wound, respiratory, intra-abdominal, and

"other" cultures showed improvements of 20.6, 19.3, 12.2, and 22.5 hours, respectively (Table 1,

Figure 2). The use of ReadA Compact smart incubators, which reduce the incubation time

needed, plus MALDI-TOF for ID of gram-negative organisms, which reduces the time to

definitive ID, could contribute to improvement in TAT.

The implementation of laboratory process improvements incorporating TLA largely led to

reductions in TAT for negative culture results. For urines, the average time to report a negative

result in 2016 improved by 14 hours compared to 2013. Prior to TLA implementation, all

negative urine cultures were read manually, which necessitated pulling plates from the incubator

racks by hand. TLA standardizes incubation duration, so cultures are available to read at the

appropriate time as opposed to times that might be dictated by staff availability or other daily

workflow considerations. Digital imaging allows negative urines to be read on-screen, reducing

the manual labor of removing each plate from the incubator, and occurred at the 18-hour mark

with TLA. These approaches are consistent with current CLSI and ASM guidelines.[24, 25] Most

likely due to TLA capabilities including digital imaging and improved incubation times, the read

time for the majority of urine cultures was reduced to one day, with the exception of suprapubic,

straight catheter, cystoscopy, and nephrostomy sources. The results of negative wound cultures

were obtained an entire day sooner. Process changes prior to TLA included the removal of

thioglycolate broths from the wound cultures and a reduction in plate incubation times.

Improvements in TAT for respiratory and "other" were not observed for negative results. The

microbiology laboratory reported that respiratory and "other" specimens grew many isolates that

were more difficult to identify at first pass. Analyses involved with this study were focused on

single isolate cultures. Respiratory and "other" specimen types often grew isolates that were

mixed, overgrown, and/or mucoid. Subcultures were often necessary, delaying ID TAT. For the

mucoid specimens, repeated IDs were sometimes necessary, causing delays.

The significant time savings reported here were likely due to several factors. The overall

laboratory process improvement interventions, including SOP update, staff education, and

hospital-wide collaborations, are likely integral parts of the improvement in the shortened ID

TAT. In addition, the TLA system aided in reducing TAT through a number of key technical

components. Because ESwab™ (Copan Diagnostics Inc., Murrieta, CA, USA) and BD

Vacutainer™ (Becton, Dickinson and Company, Franklin Lakes, NJ) system-collected urines

could be fully automated, there was more than a 75% reduction in manual plate inoculation. The

system includes a plate-streaking module, the InoqulA, which has been shown to improve colony

isolation through automated plate streaking.[2, 4, 5] BD Kiestra TLA also features a component

called the ReadA Compact, a "smart" incubator, which does not require interruptions during the

incubation process in order to improve growth conditions.[7, 15] The physical connection between

the InoqulA module, ReadA Compact, and the ProceedA (a track that automates movement of

specimens between modules) minimizes the time that plates spend in the open air at room

temperature; plates are instead transported by the ProceedA to the ReadA Compact within

minutes. Moreno-Camacho et al demonstrated that the ReadA Compact uses enhanced organism

growth cycles and improved the rate of viable organisms available for further identification and

accurate susceptibility testing.[15] Standardization of specimen preparation and detection methods

has been shown to limit the potential for variability and to reduce manual reconciliation and

interpretation by individual technologists.[4]

Previously, cultures received a non-standardized 24 hours of incubation. Incubation times varied

from 16 to 30 hours of first incubation prior to TLA and process changes. Ongoing process

review and refinement to maintain adherence to current guidelines also precipitated further

removal of redundant or obsolete steps from manual processes (such as certain broth

incubations) for specific specimen types. Interestingly, we found that laboratory process

improvement and TLA implementation led to a more even distribution of clinical results during a

24-hour time spectrum (Figure S1). However, in line with previous discussion regarding work

distribution following TLA implementation,[31] we view a more even distribution of workflow

across a 24-hour work period as an indicator of improved workflow efficiency that reflects

reductions in unnecessary or excessive work and/or idle/waiting time. Automation of negative

result identification further allows laboratory personnel to focus on other important aspects of

clinical work, for example by allowing more time to read positive plates that require technical

skill. This idea is supported by previous work, as Greub et al estimated that 50–70% of the time

spent by full-time technicians may be saved by automation of pre-analytical sample preparation

and plate inoculation alone.[32]

This study has limitations. Since this is a retrospective observational study of a microbiology

laboratory TAT improvement in a real-world setting, each intervention was not pre-designed to

last long enough to allow for assessment of independent impact. The standardization and

workflow improvement activities were integral parts of TLA implementation, in which each

intervention often overlapped with other interventions. Thus, the impact of each component

attributable to the shortened TAT could not be measured independently. This resulted in an

inability to clearly delineate the proportion of TAT improvement attributable to process

improvement versus TLA. Likewise, MALDI-TOF was not fully integrated (to include yeast

identification along with Gram negative and Gram positive bacteria) until 2017, which

admittedly affected our capability to fully assess the impact of TLA at the time of study. A goal

for future studies will be to further investigate this continued quality improvement. Another

limitation is that we did not directly assess whether improved (shorter) TAT was associated with

improved patient outcomes, which is beyond the scope of this study. Nevertheless, performance

values following TLA using Kiestra have previously been demonstrated.[17, 26, 33] Furthermore,

previous work has been able to provide more direct evidence for a benefit of specific automation

components after the implementation TLA. For example, Theparee et al [20] were able to isolate

the impact of TLA (median ID TAT decrease of ~ 2 hours compared to pre-TLA) and of

MALDI-TOF (median ID TAT decrease of ~3 hours compared to TLA alone).[20] Mutters et al

combined TLA and MALDI-TOF together versus conventional testing. However, they

completely implemented the MALDI-TOF library at one time, as opposed to the stepwise

implementation of Gram negative, Gram positive, and yeast organisms over the course of one to

two years as occurred at this study site. In that study TLA combined with MALDI-TOF led to an

antibiotic switch in 12% of cases (in addition to those cases in which a switch was based on

Gram staining) as well as shortening the first 24-hour incubation time. Ultimately, however,

randomized, controlled trials will be required to better "…evaluate the actual benefits of an

automated system in terms of morbidity, mortality, and health care costs."[6]

Conclusions

In contrast to previous studies on the impact of TLA on clinical processes involving a single

specimen type,[7, 18, 19] this study describes the significant impact of TLA and related process

changes on decreased TAT in microbial ID across a broad range of specimen types. A

combination of TLA and associated workflow improvements contributes to shorter definitive

organism ID TAT across all specimen sources and shorter TAT for negative results for most

specimen sources. Overall, there was an average of approximately 13.5 hours improvement in

TAT to organism ID across all subsets of cultures for which data were analyzed. Additionally,

the overall TAT to negative result for all cultures, even with the regression of some cultures due

to process change, improved an average of approximately 7 hours. Future studies could further

investigate the relationship of operationalizing TLA technology and optimizing workflow, and

potentially identify positive impacts on patient care by informing and thereby enabling

physicians to make antibiotic therapy choices sooner.

ACKNOWLEDGEMENTS

The authors thank Devin S. Gary, PhD (Becton, Dickinson and Company, BD Life Sciences –

Diagnostic Systems) and Andrea Gwosdow, PhD (Consultant for Becton, Dickinson and

Company, BD Life Sciences – Diagnostic Systems), for insight, discussion, and review during

the preparation of this manuscript. The individuals acknowledged here have no additional

funding or additional compensation to disclose.

DISCLOSURES

This study was approved by the Quality Improvement Review Board of the study site. The

researchers at Texas Tech University Health Sciences Center, University Medical Center,

Lubbock, TX, and Becton, Dickinson and Co., Franklin Lakes, NJ conducted the analysis as their

normal working responsibilities. YPT, LV, SK, DS, MW, and CA are current or former

employees of Becton, Dickinson, and Company, the manufacturer of the BD Kiestra TLA system

implemented by the study site. AE, CGO, SS, JN, and FL do not have any other disclosures. All

authors provided final approval of the manuscript and agree to be accountable for the accuracy

and integrity of this work.

REFERENCES

1.   Lippi G, Da Rin G. Advantages and limitations of total laboratory automation: a personal overview. *Clin Chem Lab Med.* May 27 2019;57(6):802-811.

2.   Croxatto A, Dijkstra K, Prod'hom G, Greub G. Comparison of Inoculation with the InoqulA and WASP Automated Systems with Manual Inoculation. *J Clin Microbiol.* Jul 2015;53(7):2298-2307.

3.   Croxatto A, Prod'hom G, Faverjon F, Rochais Y, Greub G. Laboratory automation in clinical bacteriology: what system to choose? *Clin Microbiol Infect.* Mar 2016;22(3):217-235.

4.   Froment P, Marchandin H, Vande Perre P, Lamy B. Automated versus manual sample inoculations in routine clinical microbiology: a performance evaluation of the fully automated InoqulA instrument. *J Clin Microbiol.* Mar 2014;52(3):796-802.

5.   Iversen J, Stendal G, Gerdes CM, Meyer CH, Andersen CO, Frimodt-Moller N. Comparative Evaluation of Inoculation of Urine Samples with the Copan WASP and BD Kiestra InoqulA Instruments. *J Clin Microbiol.* Feb 2016;54(2):328-332.

6.   Mutters NT, Hodiamont CJ, de Jong MD, Overmeijer HP, van den Boogaard M, Visser CE. Performance of Kiestra total laboratory automation combined with MS in clinical microbiology practice. *Ann Lab Med.* Mar 2014;34(2):111-117.

7.   Dauwalder O, Landrieve L, Laurent F, de Montclos M, Vandenesch F, Lina G. Does bacteriology laboratory automation reduce time to results and increase quality management? *Clin Microbiol Infect.* Mar 2016;22(3):236-243.

8.   Matthews S, Deutekom J. The future of diagnostic bacteriology. *Clin Microbiol Infect.* May 2011;17(5):651-654.

9.   Carden R, Allsbrook K, Thomas R. An examination of the supply and demand for clinical laboratory professionals in the United States. *Transfusion.* Nov 2009;49(11 Pt 2):2520-2523.

10.   Garcia E, Ali AM, Soles RM, Lewis DG. The American Society for Clinical Pathology's 2014 vacancy survey of medical laboratories in the United States. *Am J Clin Pathol.* Sep 2015;144(3):432-443.

11.   Ward-Cook K. Medical laboratory workforce trends and projections: what is past is prologue. *Clin Leadersh Manag Rev.* Nov-Dec 2002;16(6):364-369.

12.   Hawker CD. Nonanalytic Laboratory Automation: A Quarter Century of Progress. *Clin Chem.* Jun 2017;63(6):1074-1082.

13.   Murray PR. Laboratory automation: efficiency and turnaround times. *Micro Austr.* 2014;35(1).

14.   Humphrey G. MC, Gough H., Awadel-Kariem F.M. Experience with KIESTRA's Total Lab Automation solution to meet the challenge of universal MRSA screening for Lister Hospital, a large UK district general hospital. *Clinical Microbiology and Infection.* 2011;17(Supp 4).

15.   Moreno-Camacho JL, Calva-Espinosa DY, Leal-Leyva YY, Elizalde-Olivas DC, Campos-Romero A, Alcántar-Fernández J. Transformation From a Conventional Clinical Microbiology Laboratory to Full Automation. *Laboratory Medicine.* 2017;49(1):e1-e8.

16.   Thomson RB, Jr., McElvania E. Total Laboratory Automation: What Is Gained, What Is Lost, and Who Can Afford It? *Clin Lab Med.* Sep 2019;39(3):371-389.

17.   Klein S, Nurjadi D, Horner S, Heeg K, Zimmermann S, Burckhardt I. Significant increase in cultivation of Gardnerella vaginalis, Alloscardovia omnicolens, Actinotignum schaalii, and Actinomyces spp. in urine samples with total laboratory automation. *Eur J Clin Microbiol Infect Dis.* Jul 2018;37(7):1305-1311.

18.   Graham M, Tilson L, Streitberg R, Hamblin J, Korman TM. Improved standardization and potential for shortened time to results with BD Kiestra total laboratory automation of early urine cultures: A prospective comparison with manual processing. *Diagn Microbiol Infect Dis.* Sep 2016;86(1):1-4.

19.   Yarbrough ML, Lainhart W, McMullen AR, Anderson NW, Burnham CD. Impact of total laboratory automation on workflow and specimen processing time for culture of urine specimens. *Eur J Clin Microbiol Infect Dis.* Dec 2018;37(12):2405-2411.

20.   Theparee T, Das S, Thomson RB, Jr. Total Laboratory Automation and Matrix-Assisted Laser Desorption Ionization-Time of Flight Mass Spectrometry Improve Turnaround Times in the Clinical Microbiology Laboratory: a Retrospective Analysis. *J Clin Microbiol.* Jan 2018;56(1).

21.   Burckhardt I, Horner S, Burckhardt F, Zimmermann S. Detection of MRSA in nasal swabs-marked reduction of time to report for negative reports by substituting classical manual workflow with total lab automation. *Eur J Clin Microbiol Infect Dis.* Sep 2018;37(9):1745-1751.

22.   Tabak YP, Vankeepuram L, Ye G, Jeffers K, Gupta V, Murray PR. Blood Culture Turnaround Time in U.S. Acute Care Hospitals and Implications for Laboratory Process Optimization. *J Clin Microbiol.* Dec 2018;56(12).

23.   Zilberberg MD, Tabak YP, Sievert DM, et al. Using electronic health information to risk-stratify rates of Clostridium difficile infection in US hospitals. *Infect Control Hosp Epidemiol.* Jul 2011;32(7):649-655.

24.   Carroll KC, Pfaller MA, Landry ML, et al. *Manual of Clinical Microbiology, Twelfth Edition*: American Society of Microbiology; 2019.

25.   Patel JM, Cockerill FR, Eliopoulos GM, et al. CLSI. Performance Standards for Antimicrobial Susceptibility Testing: Twenty-Third Informational Supplement. CLSI document M100-S23. Wayne, PA: Clinical and Laboratory Standards Institute. 2014.

26.   Heather CS, Maley M. Automated direct screening for resistance of Gram-negative blood cultures using the BD Kiestra WorkCell. *Eur J Clin Microbiol Infect Dis.* Jan 2018;37(1):117-125.

27.   Buehler SS, Madison B, Snyder SR, et al. Effectiveness of Practices To Increase Timeliness of Providing Targeted Therapy for Inpatients with Bloodstream Infections: a Laboratory Medicine Best Practices Systematic Review and Meta-analysis. *Clin Microbiol Rev.* Jan 2016;29(1):59-103.

28.   De Socio GV, Di Donato F, Paggi R, et al. Laboratory automation reduces time to report of positive blood cultures and improves management of patients with bloodstream infection. *Eur J Clin Microbiol Infect Dis.* Dec 2018;37(12):2313-2322.

29.   Maurer FP, Christner M, Hentschke M, Rohde H. Advances in Rapid Identification and Susceptibility Testing of Bacteria in the Clinical Microbiology Laboratory: Implications for Patient Care and Antimicrobial Stewardship Programs. *Infect Dis Rep.* Mar 30 2017;9(1):6839.

30.   Garcia L. *Clinical Microbiology Procedures Handbook, 3rd Edition*; 2010.

31.   Burckhardt I. Laboratory Automation in Clinical Microbiology. *Bioengineering (Basel).* Nov 22 2018;5(4).

32.   Greub G, Prod'hom G. Automation in clinical bacteriology: what system to choose? *Clin Microbiol Infect.* May 2011;17(5):655-660.

33.   Croxatto A, Marcelpoil R, Orny C, Morel D, Prod'hom G, Greub G. Towards automated detection, semi-quantification and identification of microbial growth in clinical bacteriology: A proof of concept. *Biomed J.* Dec 2017;40(6):317-328.

Figure Legends

Figure 1: Microbiology Laboratory Quality Improvement Process Illustration. Kiestra TLA was implemented in May 2015.

Abbreviations: TLA, Total laboratory automation; MALDI-TOF, Matrix-assisted laser desorption/ionization-time of flight mass spectrometry; LIS, Laboratory information system; SOP, Standard operating proceedure.

Figure 2: Comparison of cumulative distribution of definitive positive bacterial pathogen ID turnaround time following login by most common source: a) blood, b) urine, c) respiratory, and d) wound. All values p<0.05.

Figure 3: Comparison of cumulative distribution of definitive negative culture results turnaround time following login by most common source: a) blood, b) urine, c) respiratory, and d) wound. All values p<0.05.

[a]The extended time reported for blood cultures reflects an initial growth/detection phase prior to plating.

Supplemental Figure 1: Organism ID reporting time distribution during a 24-hour spectrum.

Tables
Table 1

Table 1. Descriptive statistics for definitive positive bacterial pathogen ID turnaround time (hours)

Specimen	Year	n	Mean (SD)	Median (IQR)	p-value[a]
Blood	2013	759	70.6 (28.9)	62.8 (52.0, 74.5)	<.0001
	2016	882	51.2 (19.2)	46.6 (38.0, 59.1)	
Urine	2013	2,282	47.1 (14.9)	43.6 (38.0, 48.9)	<.0001
	2016	2,306	40.7 (18.8)	35.5 (27.7, 47.9)	
Wound	2013	1,160	60.2 (25.6)	58.6 (43.1, 69.2)	<.0001
	2016	869	39.6 (21.3)	32.8 (25.6, 47.0)	
Respiratory	2013	461	67.0 (23.5)	65.3 (49.8, 74.7)	<.0001
	2016	369	47.7 (24.9)	41.9 (28.1, 57.7)	
Intra-abdominal	2013	23	69.9 (28.8)	64.3 (53.1, 84.7)	0.0368
	2016	34	57.8 (32.1)	49.9 (25.7, 78.2)	
Other[b]	2013	56	75.1 (57.4)	64.5 (46.4, 89.8)	0.0013
	2016	50	52.6 (30.5)	46.1 (29.3, 69.7)	
Overall	2013	4,741	56.8 (24.3)	48.4 (41.1, 66.4)	<.0001
	2016	4,510	43.3 (20.8)	38.0 (28.7, 51.7)	

Abbreviations: SD, standard deviation; IQR, inter-quartile range (1st, 3rd quartile)

[a]Log-transformed t-test for continuous variables, utilized for comparisons of mean turnaround time.
[b]"Other" includes specimens from duodenal aspirate, ear, eye, quantitative tissue, quantitative bronchoscopy, environmental samples (surgical hardware, transfusion reactions), stool, cystic fibrosis cultures, dialysates, catheter tips, cerebral spinal fluid, genital, and throat.

Table 2

Table 2. Descriptive statistics for negative result turnaround time (in hours)

Specimen	Year	n	Mean (SD)	Median (IQR)	p-value[a]
Blood	2013	16,393	131 (54)	126 (122, 128)	<.0001
	2016	17,905	125 (6)	125 (123, 127)	
Urine	2013	8,115	44 (10)	45 (40, 47)	<.0001
	2016	8,880	30 (14)	26 (22, 32)	
Wound	2013	1,100	108 (40)	114 (74, 120)	<.0001
	2016	957	82 (43)	66 (50, 113)	
Respiratory	2013	1,204	62 (30)	50 (45, 64)	<.0001[b]
	2016	1,460	71 (36)	54 (47, 80)	
Intra-abdominal	2013	154	129 (26)	117 (113, 156)	<.0001
	2016	280	115 (40)	122 (103, 132)	
Other[c]	2013	613	121 (15)	118 (114, 123)	<.0001[b]
	2016	674	170 (27)	175 (164, 184)	
Overall	2013	27,579	101 (58)	121 (48, 127)	<.0001
	2016	30,156	94 (47)	123 (38, 126)	

Abbreviations: SD, standard deviation; IQR, inter-quartile range (1st, 3rd quartile)

[a]Log-transformed t-test for continuous variables, utilized for comparisons of mean turnaround time.
[b]For respiratory and "other" sources, the negative turnaround time was shorter in 2013.
[c]"Other" includes specimens from duodenal aspirate, ear, eye, quantitative tissue, quantitative bronchoscopy, environmental samples (surgical hardware, transfusion reactions), stool, cystic fibrosis cultures, dialysates, catheter tips, cerebral spinal fluid, genital, and throat.

Figures
Figure 1

Figure 1.

Implementing process improvement, TLA, MALDI-TOF, LIS, and improving system-wide communication

Improving communication between laboratory and physician groups and departments (hospital wide effort)

Implementing MALDI-TOF

Implementing TLA

Standardization of agar plate setup

Streamline culture order code options (lab Information System improvement)

Lab staff education

Review of SOPs

Pre- period

Assessment and planning for expansion to 24/7 operation

Post- period

2013   2014   2015   2016

Figure 2

Figure 2.

Figure 3

Figure 3.

Supplemental Figure 1

Figure S1.