Title

Activities of Daily Living with Bionic Arm Improved by Combination Training and Latching Filter in Prosthesis

Control Comparison

Authors

Michael D. Paskett*†[1,2], Mark R. Brinton*[3], Taylor C. Hansen[1], Jacob A. George[2,4,5], Tyler S. Davis[6], Christopher C.

Duncan[5], Gregory A. Clark[1]

[1]Department of Biomedical Engineering, University of Utah, Salt Lake City, UT 84112

[2]Center for Clinical and Translational Science, University of Utah, Salt Lake City, UT 84112

[3]School of Engineering, Math and Computer Science, Elizabethtown College, Elizabethtown, PA 17022

[4]Department of Mechanical Engineering, University of Utah, Salt Lake City, UT 84112

[5]Division of Physical Medicine and Rehabilitation, University of Utah, Salt Lake City, UT 84112

[6]Department of Neurosurgery, University of Utah, Salt Lake City, UT 84112

*contributed equally

†corresponding author

Corresponding Author: Michael D. Paskett, email: michael.paskett@utah.edu

Abstract (346/350)

Background: Advanced prostheses can restore function and improve quality of life for individuals with

amputations. Unfortunately, most commercial control strategies do not utilize the rich control information from

residual nerves and musculature. Continuous decoders can provide more intuitive prosthesis control from multi-

channel neural or electromyographic recordings. Three components influence continuous decoder performance:

the data used to train the algorithm, the algorithm, and smoothing filters on the algorithm's output. As

individual groups often focus on a single decoder, very few studies compare different decoders using otherwise

similar experimental conditions. Additionally, control improvements are generally demonstrated with offline

data or in virtual settings, which do not necessarily translate to real-world performance.

Methods: We completed a two-phase head-to-head comparison of 12 continuous decoders using activities of

daily living. In phase one, we compared two training types and a smoothing filter with three algorithms

(modified Kalman filter, multi-layer perceptron, and convolutional neural network) in a clothespin relocation

task. We compared training types that included data with only individual digit and wrist movements vs.

combination movements (e.g., simultaneous grasp and wrist flexion). We also compared raw vs. nonlinearly

smoothed algorithm outputs. In phase two, we compared the three algorithms in fragile egg, zipping, pouring,

and folding tasks using the combination training and smoothing found beneficial in phase one. In both phases,

we collected objective, performance-based (e.g., success rate) and subjective, user-focused (e.g., preference)

measures.

Results: Phase one showed that combination training improved prosthesis control accuracy and speed, and that

the nonlinear smoothing improved accuracy but generally reduced speed. In phase two, user-focused metrics

favored the convolutional neural network and modified Kalman filter, while performance-based metrics were

generally similar among all algorithms.

Conclusions: These results confirm that state-of-the-art algorithms, whether linear or nonlinear in nature,

functionally benefit from training on more complex data, and demonstrate that smoothing the output improves

control. These studies will be used to select a decoder for a long-term take-home trial with implanted

neuromyoelectric devices. Overall, clinical considerations may favor the mKF as it is similar in performance,

faster to train, and computationally less expensive than neural networks.

46   Keywords

Background

Modern prosthetic hands can now recreate the complex movements of the human hand [1]–[8].

Unfortunately, unintuitive control makes prostheses difficult to use [9] and is a major factor in prosthesis

dissatisfaction and abandonment [10]. Neural and electromyographic (EMG) signals from residual muscles

provide a rich source of movement information that could be used to provide intuitive prosthesis control. With

rare exceptions [11], commercial prosthetic control uses only two-electrode EMG setups that provide sequential

control of up to two degrees-of-freedom [1], [3], [5], [6], [12], [13], far inferior to the abilities of modern

prosthetic hands.

Using more neuromuscular inputs can provide more intuitive prosthesis control through richer, more

diverse data that can be classified into different movements based on residual neuromuscular activation

patterns. Several groups have demonstrated intuitive prosthesis control with decoders that classify distinct

movements [14]–[21]. Generally, these decoders classify EMG patterns to several pre-determined grip patterns

(e.g., close hand, open hand, pinch). More advanced classification strategies incorporate proportionality into the

classes (e.g., enable partial hand closure) [22] or allow simultaneously active classes (e.g., simultaneously rotate

wrist and close hand) [17]. Classifiers can be limited by allowing the user only a predetermined, fixed number of

movement types.

Continuous decoders provide another approach to intuitively control a prosthesis. Instead of distinct

classes of movements, continuous decoders allow independent and simultaneous movement of individual

degrees-of-freedom (e.g., finger or wrist movements), allowing the user to potentially produce any grip pattern.

Continuous decoders have been used to provide prosthesis control with linear [23]–[25] and nonlinear [26]–[28]

algorithms. Nonlinear algorithms generally take longer to train and are subject to overfitting; however, they may

capture the nonlinear nature of multi-dimensional EMG data [29], especially for complex movement patterns.

Kalman filters, and related variations, provide a linear approach to continuous control that has been

used with cortical [30] and peripheral [23], [31] recordings. Multi-layer perceptrons (MLPs), a relatively simple

artificial neural network, have been used for both classification [29], [32], [33] and continuous [26], [34]–[36]

control. Convolutional neural networks (CNNs), which, in the prosthetic domain, convolve temporally over their

input, have also been used for classification [37], [38] and continuous [27], [28] control. CNNs temporal

convolution enables them to learn features from raw EMG data [28], [37], whereas other strategies typically use

time-domain features, such as the mean absolute value of segmented EMG.

Beyond the algorithm itself, the data used to train the algorithm affects prosthesis control. Training an

algorithm in multiple arm positions can improve prosthesis control whether the position data is used to train the

algorithm [29], [39] or not [40], although incorporating positional data may not benefit linear algorithms [41].

For continuous controllers without a discrete set of classes, using training data with simultaneous movements of

more than one degree-of-freedom may expose nonlinearities in EMG data and preferentially benefit nonlinear

algorithms with higher learning capacities.

Modifying a continuous decoder's output can improve prosthesis control by improving stability.

Smoothing filters, such as a traditional low-pass filters can smooth the output and reduce unwanted jitter [42].

One study incorporated a camera system to selectively low-pass filter the output when a grasp was detected

[43]. Low-pass filters inevitably introduce delay into the control. Nonlinear smoothing filters can reduce small-

amplitude jitter without slowing larger movements [44], [45].

Prosthesis control improvements are generally demonstrated sequentially in three domains: offline,

online, and real-world. Offline comparisons measure the ability of a decoder to map EMG inputs to kinematic

outputs, without the user in the loop. Several offline comparisons have helped improve the state of the art for

both continuous [36] and classification [16], [17], [29], [38], [46]–[48] control. While a useful and necessary first

step in improving control, offline improvements do not necessarily imply online improvements [49], largely

because they do not incorporate the continuous interaction of the user with the controller. Online

improvements incorporate user feedback and have been demonstrated through Fitts' law or virtual target tasks.

Several online comparisons have demonstrated differences among control strategies [25], [27], [28], [32], [37],

[49]–[53]. The relationship of online performance to real-world performance is debated [54]–[56]. Comparing

decoders through real-world activities of daily living provides the most realistic approximation of how a decoder

would function in everyday life. Real-world comparisons have demonstrated the benefits of continuous [57],

[58] and classifier [19], [45], [59]–[62] control strategies over commercial control; however, only a couple

comparisons have been made among research-grade control strategies [45], [62] , both of which were

classifiers.

In this work, we conduct a two-phase real-world comparison of 12 continuous decoders that span the

three major components affecting control (training data, algorithms, and smoothing). In the first phase, we

compare training an algorithm with individual finger or wrist movements vs. combination movements (e.g.,

simultaneous flexion of all fingers to make a functional grasp) and compare the effect of using a nonlinear

smoothing filter [44]. We study these conditions in a two-degree-of-freedom decode with a modified Kalman

filter (mKF), an MLP, and a CNN, for a total of 12 decoders. In the second phase, we performed a more extensive

evaluation of five-degree-of-freedom decodes with the mKF, MLP, and CNN using the combination training and

smoothing filter found most beneficial in phase one. We focus on both performance-based objective and

patient-centric subjective measures. To our knowledge, this study provides the first comparison of the same

individuals using multiple continuous decoders and a physical prosthesis to complete real-world tasks. A large

part of our motivation for this study is to identify the "best" decoder among decoders we have previously

published. The results of this study provide guidance for the clinical translation of advanced prosthesis control

for a long-term take-home clinical trial with an advanced prosthesis and portable control system.

Methods

Signal Acquisition & Prosthesis Setup

Signal acquisition has been described previously [31]. In brief, surface EMG (sEMG) was collected from a

sEMG sleeve [63] with the 512-channel Grapevine System (Ripple Neuro LLC, Salt Lake City, UT). Thirty-two

single-ended channels were acquired at 1 kHz and filtered with a 6th-order high pass Butterworth filter (15 Hz),

2nd-order low-pass Butterworth filter (375 Hz), and 60, 120, and 180 Hz notch filters. After connecting the sEMG

sleeve to the acquisition device, channels were manually inspected and removed if shorted channels were

detected (generally less than two channels). The differential pairs of all monopolar channels were calculated,

and features (single-ended and differential) were created at 30 Hz using the mean absolute value of a 300-ms

buffer (i.e., 528 features from an overlapping 300-ms boxcar filter).

Non-amputee participants donned the DEKA LUKE Arm [1] with a bypass socket enabling non-amputee

use of a prosthetic arm, which has been shown to be representative of use by individuals with amputation [64].

The bypass socket, which has a freely-rotating wrist attachment, had rubber bands providing wrist rotation

resistance (0.34 Nm to 0.85 Nm at rest and maximal rotation, respectively). Baseline EMG activity was

subtracted from each feature before collecting training data by subtracting the average EMG activity during a

10-s period where the user supported the weight of the bypass socket and DEKA LUKE Arm in a neutral position

approximately waist height. Training data were recorded as participants mimicked the pre-programmed

movements of the DEKA LUKE Arm.

Decode Algorithms

The mKF was implemented as has been demonstrated previously [31]. A Gram-Schmidt

orthogonalization algorithm [65] was used to select 48 features as inputs for the mKF. The 48 EMG features and

kinematics were used to fit the parameters of a linear Kalman filter [66]. The *post-hoc* thresholds which were

used to modify the Kalman filter online [31], were optimized offline [31].

The nonlinear MLP was implemented as has been demonstrated previously [26]. The MLP consisted of

two hidden layers, each with 128 nodes, and a hyperbolic tangent (tanh) activation layer. The training and

testing data were split 80% and 20%, respectively. As with the mKF, the MLP used 48 input features as chosen by

the same stepwise Gram-Schmidt orthogonalization algorithm. The MLP trained for three dataset aggregation

iterations, each with ten epochs (found sufficient in [26]).

The nonlinear CNN was implemented as has been demonstrated previously [27]. The CNN, comprising

eight total layers, used all 528 features to fit the kinematic values. 75% of the data was used to train the

algorithm, and the remaining 25% was used as a validation set. The network trained until the loss on the

validation set was equal to or larger than the previously smallest loss for five iterations.

Smoothing Filter

A smoothing filter was implemented as has been demonstrated previously [44]. The "latching filter" we

used is a computationally inexpensive, recursive, nonlinear filter that smooths small-amplitude jitter but allows

quick changes to its output. The latching filter uses the difference between its current state and current input to

determine how much to move in the direction of the current input. For small input changes, the output is

heavily smoothed, and for large input changes, the output is nearly unsmoothed.

Study Overview

We conducted this study in two phases (Fig. 1a). In the first phase, we compared training paradigms and

the use of a smoothing filter for all three algorithms, totaling 12 decoders: three algorithms, two training

paradigms, and two output types. Due to the number of decoders and the consequential experiment duration,

we selected a simple task, the clothespin relocation task, with a low (two) degree-of-freedom control strategy,

which simplified the training and allowed the experiments to be completed in 2-3 hours.

In phase two, we increased the decoder output to five degrees-of-freedom to evaluate the three

algorithms' control in a high-degree-of-freedom setting representative of modern myoelectric prostheses. We

used the results of the first phase to inform the training paradigm and smoothing filter use. As we were

comparing only three algorithms in the second phase, we were able to perform four tasks (fragile egg, zipper,

pouring, and folding) in a 2-3-hour experimental session.

Fig. 1 (a) Overview of the two-phase functional comparison of 12 prosthesis control strategies. In phase one, we compared collecting training data with individual movements vs. combination movements and compared using raw algorithm output vs. nonlinear output smoothing with a latching filter for an mKF, MLP, and CNN in a clothespin relocation task. In phase two, we compared the mKF, MLP, and CNN with fragile egg, zipping, pouring, and folding tasks. **(b)** Example combination movement, where two degrees-of-freedom are simultaneously active during the middle portion of the trial.

Phase One: Training Type and Output Smoothing Comparison

Training Paradigms

Participants (n = 10) with no prior prosthesis experience mimicked the DEKA LUKE Arm movements in

three training periods to provide training data for the control algorithms. The two-degree-of-freedom decode

training consisted of opening and closing the hand as well as rotation of the wrist. Prior to each training period,

study administrators coached the participants through one trial of each movement, to acquaint the participants

with mimicking the movements. In the first training period, users completed 12 trials of individual movements:

close hand, open hand, pronate, supinate. Each individual movement consisted of a 0.7-s rise time, 3-s hold, and

0.7-s return to rest position (for a total of 4.4 s per movement). The second training period repeated the first

period; however, only four trials of each movement were completed.

The third period introduced combination movements, where both degrees-of-freedom were

simultaneously active for the middle portion of the trial (see Fig. 1b for a visual representation). Users

completed four trials each of eight different movements: close hand and pronate, close hand and supinate, open

hand and pronate, open hand and supinate, pronate and close hand, supinate and close hand, pronate and open

hand, and supinate and open hand. For one combination movement, the first degree-of-freedom moved to its

intended position with a 0.7-s rise time, 5.2-s hold, and 0.7-s return to resting position. The second degree-of-

freedom moved to its intended position 0.4 s after the first degree-of-freedom reached its intended position

with a 0.7-s rise time, 3.0-s hold, and 0.7-s return to its rest position. In total, one combination training trial

consisted of 6.6 s.

To compare individual and combination training paradigms, the training sets were merged into two sets

of equal duration before being used to train the decoder (so that any improved performance would not be due

to the decoder training on more data). The first and second training periods were combined to produce an

individual-movement-only set. The second and third training periods were combined to produce a set with

individual and combination movements. After merging the datasets, the mKF, CNN, and MLP were trained on

the kinematic and EMG data for proportional control.

Modified Clothespin Relocation Task

In the modified clothespin relocation task (Fig. 2), participants transferred three large clothespins from a

horizontal bar to a vertical bar on the side of the dominant hand, referred to as the "inside-out" task [67]. The

task was chosen for several reasons: its simplicity allows the user to quickly learn and master the task, mitigating

effects due to learning; its short duration allows several trials to be completed rapidly, which was necessary for

comparing 12 decoders; and, it necessarily requires simultaneously active degrees-of-freedom, typical of many

daily activities. We have used this task with the DEKA LUKE Arm previously [68].

Participants started a timer with the prosthetic hand, transferred a single clothespin, and stopped the

timer with the prosthetic hand. A trial was considered failed when a clothespin was dropped. Participants were

instructed that the success rate (proportion of successfully transferred clothespins) and the transfer time would

be recorded. There was no time limit for the task. Participants completed nine trials (three sets of three) with

each decoder. The order of the control paradigm was randomized among subjects to prevent an order effect.

After completing nine trials with a particular decoder, participants completed the NASA Task Load Index (TLX)

subjective workload survey [69].

Participants completed the task for 12 total decoders, including each possible combination of training

paradigm (individual and combination), algorithm (mKF, MLP, and CNN), and output type (raw and smoothed).

Fig. 2 Phase one: The clothespin relocation task. The clothespin relocation task tests prosthesis control with

simultaneously active grasping and wrist movements, typical of many activities of daily living.

Analysis

We compared the effect of training paradigm and smoothing filter for each algorithm with $\alpha = 0.05$. The

data were significantly non-normal for several conditions (Shapiro-Wilk test [70]), so nonparametric statistics

were used. Median completion times for successful trials were calculated for each condition for each participant

(N = 10 participants). The number of drops was calculated for each condition for each participant. For the

training paradigm comparison, the data were aggregated to compare individual and combination trainings with

Wilcoxon's signed-rank test [71] for each algorithm; each group contained both smoothed and raw output types

(N = 20; two medians for each participant). For the output type comparison, the data were aggregated to

compare raw and smoothed outputs with Wilcoxon's signed-rank test for each algorithm; each group contained

both individual and combination trainings (N = 20; two medians for each participant). Because the hypotheses

were pre-planned, no correction for multiple comparisons was applied [72].

Phase Two: Algorithm Comparison

Training Paradigm

In the second phase, we extended the degrees-of-freedom from two to five to assess the algorithms'

performance in more complex, biologically realistic conditions. Users controlled the following degrees-of-

freedom: flexion/extension of the thumb, flexion/extension of the index finger, flexion/extension of the middle,

ring, and little fingers (which are mechanically coupled on the DEKA LUKE Arm), flexion/extension of the wrist,

and pronation/supination of the wrist. Study administrators coached the participants through a single trial of

each movement type before each training period. The training for phase two consisted of four distinct training

periods.

Participants (n = 9) with some prior prosthesis experience (2-3 hours) mimicked the prosthesis through

movements of individual degrees-of-freedom. Movements consisted of a 0.7-s rise time, 5.2-s hold time, 0.7-s

fall time (for a total of 6.6 s per movement). Participants mimicked four trials on each degree-of-freedom. For

the second through fourth training periods, combination movements were introduced.

The second period consisted of grasp movements (simultaneous flexion of the thumb, index, and

middle, ring, and little degrees-of-freedom) combined with wrist movements (flexion, extension, pronation, and

supination). Participants completed the first grasp movement alone, with a 0.7-s rise time, 5.2-s hold time, and

0.7-s fall time (6.6 s total). Participants then completed grasp and wrist combination movements, where both

the grasp and wrist movement were simultaneously active in the middle portion of a trial (Fig. 1b). In these

combination trials, the grasp movement rose to its intended position in 0.7 s, held for 5.2 s, and returned to its

original position in 0.7 s. The wrist movement started 0.4 s after the first movement reached its intended

position, moved to its intended position with a 0.7-s rise time, held for 3.0 s, and returned to its original position

in 0.7 s. These combination movements were completed for each wrist movement (wrist flexion, extension,

pronation, and supination). Then, they were completed in the reverse order, where the wrist movement

253 occurred first, and the grasp movement occurred part-way through the wrist movement. We chose to complete

254 the movements in both orders due to the CNN convolving over time. Participants completed four trials of each

255 sequence, resulting in 36 total trials.

256  Periods three and four repeated the second period; however, the grasping movement was replaced with

257 opening the hand (extending the thumb, index, and middle, ring, and little digits) and a pinch movement (flexing

258 thumb and index digits), respectively. After merging the datasets from the four training periods, the mKF, CNN,

259 and MLP were trained on the kinematic and EMG data for proportional control.

260 **Activities-of-Daily-Living**

261  In the second phase, participants completed four tasks representative of activities of daily living in the

262 following order: moving a fragile egg, using a zipper, pouring, and folding a towel (Fig. 3). A time limit was

263 specified for each task. The order of the three decode algorithms (mKF, CNN, and MLP) was randomized for each

264 participant to avoid order effects, with the same random order being used within each of the four tasks. After

265 completing a task with each algorithm, participants ranked the algorithms in order of preference. Examples of

266 the tasks are included in the supplementary material (Additional file 1: Video 1). Each task is now discussed in

267 greater detail.

Fig. 3 Phase two: Activities of daily living represent realistic prosthesis use. **(a)** The fragile egg task tests precision

control and tests the proportionality of a decode algorithm. **(b)** The zipper task requires stable pinching and

simultaneous wrist rotation. **(c)** The pouring task requires stable grasping with smooth wrist rotation. **(d)** The

folding task requires two-handed object manipulation and stable grasps during positional changes.

Fragile Egg Task

In the fragile egg task, participants moved a mechanical "fragile egg" [73], [74] horizontally 15 cm from

one side of a 6.25-cm vertical barrier to the other within a 45-s time limit, beginning on the side of the dominant

hand. This task evaluated the dexterity provided by each decode algorithm in that if too much force was applied,

the mechanical egg would "break" and emit an audible click. The participants were instructed to use a two-

finger pinch to pick up the mechanical egg. The breaking force for these experiments was set at about 20 N, and

the device mass was about 615 g (0.03 N/g).

For each trial, participants started a timer with the prosthetic hand, transferred the mechanical egg, and

stopped the timer with the prosthetic hand. A trial was considered failed if the mechanical egg was broken or

dropped. Participants were instructed that both the success rate (the proportion of trials in which transfer was

successful) and the transfer time would be recorded. When introduced to the task, participants were given two

minutes of practice with each decode algorithm to decrease the learning effects in the recorded trials.

Participants attempted 15 trials with each decode algorithm. Immediately before the recorded trials

commenced, participants were given one final practice trial with the given decode algorithm. A verbal warning

alerted the participants when there were 15 s remaining. After attempting the 15 recorded trials with a

particular algorithm, participants completed a NASA-TLX survey.

Zipper Task

In the zipper task, participants closed a horizontally mounted zipper within a 30-s time limit, beginning

on the side of the dominant hand. Timed-out trials were reported as 30 s. The zipper task tests precision

grasping with combination wrist movements. For a trial to be successful, the zipper must have closed by at least

6.5 cm, as indicated with a mark. We modified the base of the zipper so that the handle would not lie less than

45 degrees from being flat because the shape of the prosthetic fingers prevented gripping the zipper when lying

flat. For each trial, participants started a timer with the prosthetic hand, closed the zipper, and stopped the

timer with the prosthetic hand. Participants were instructed that both the number of releases (times the zipper

slipped from their grasp) as well as the time would be recorded. When introduced to the task, participants were

given three practice trials with each decode algorithm in order to decrease the learning effects in the recorded

trials. Participants attempted five trials with each decode algorithm. Immediately before the recorded trials

commenced, participants were given one final practice trial with the given decode algorithm. After attempting

the five recorded trials with a particular algorithm, participants completed a NASA-TLX survey.

Pouring Task

In the pouring task, participants poured rice from an aluminum can into a cup within a 1-min time limit.

The rice-filled can was placed on the side of the dominant hand, and the empty cup was placed near the midline.

This task assessed the ability of the decode algorithms to maintain a grasp during wrist rotation. For these

experiments, the aluminum can was filled with 200 g of rice.

For each trial, participants started a timer with the prosthetic hand, poured the rice from the can (7.6

cm diameter, 11.3 cm tall) to the cup (8.6 cm diameter, 11.7 cm tall), and stopped the timer with the prosthetic

hand. Participants were instructed that the amount of rice (in grams) successfully transferred to the cup as well

as the transfer time would be recorded. The participants were not allowed to use their native hand to stabilize

the cup, and if the can was dropped the trial was considered unsuccessful. When introduced to the task,

participants were given three practice trials with each decode algorithm in order to decrease the learning effects

in the recorded trials. Participants attempted five trials with each decode algorithm. Immediately before the

recorded trials commenced, participants were given one final practice trial with the given decode algorithm.

After attempting the five recorded trials with a particular algorithm, participants completed a NASA-TLX survey.

Folding Task

In the folding task, participants folded a small hand towel (39 cm by 58 cm) twice using a pinch grasp

within a 1-min time limit. The first fold was done from short-end to short-end. The participant could then use

their non-dominant native hand to rotate the towel 90 degrees for the second fold, which was also from short-

end to short-end. For both folds, participants were instructed not to use the edge of the table to facilitate

grasping the towel. Additionally, they were not to use their native hand to guide the end of the towel into the

prosthesis. Ideally, participants simultaneously grasped the short edge of the towel with their native hand and

the prosthesis.

For each trial, participants started a timer with the prosthetic hand, folded the towel twice, and stopped

the timer with the prosthetic hand. Participants were instructed that the number of releases (times the towel

slipped from their grasp) and the time would be recorded. When introduced to the task, participants were given

three practice trials with each decode algorithm to decrease the learning effects in the recorded trials.

Participants attempted five trials with each decode algorithm. Immediately before the recorded trials

commenced, participants were given one final practice trial with the given decode algorithm. After attempting

the five recorded trials with a particular algorithm, participants completed a NASA-TLX survey.

Analysis

We compared algorithms for each metric within each task with $\alpha = 0.05$. The data were significantly

non-normal for several conditions (Shapiro-Wilk test [70]), so nonparametric statistics were used. Median

completion times for successful trials were calculated for each condition for each participant (N = 9 participants).

For the pouring task, the task was considered successful if the total amount transferred was greater than 199.9 g

(no spill). For each metric, a Kruskal-Wallis test [75] was used to determine if there were differences among the

decode algorithms. If the Kruskal-Wallis revealed differences, paired comparisons were made using Wilcoxon's

signed-rank test [71]. Because the hypotheses were pre-planned, no correction for multiple comparisons was

applied [72]. *Post-hoc*, we decided to compare aggregate user preferences. We grouped the user preferences

for each task and performed a Kruskal-Wallis test across algorithms. As this was an unplanned comparison, we

corrected the pairwise comparisons with the Dunn-Sidak correction for multiple comparisons [76].

Results

Phase one: Comparing Training Type and LF

Training Type Comparison

Training with combination movements generally improved the outcomes for the clothespin relocation

task (Fig. 4). Training with combination movements significantly reduced dropped clothespins for the CNN and

MLP ($p < 0.05$; Fig. 4a). The number of dropped clothespins for the mKF was not significantly affected by the

training type. Training with combinations significantly reduced the median time needed to transfer a clothespin

for CNN and MLP ($p < 0.05$, $p < 0.001$, respectively). The median transfer time using the mKF was not

significantly affected by the training type (Fig. 4b), though combination trainings trended toward being faster.

Combination training significantly reduced the subjective workload for the CNN (p < 0.01). Training type had no

significant effect on the subjective workload for the mKF and MLP; however, both were reduced (Fig. 4c).

Fig. 4 Phase one: Combination training improves control accuracy and speed. **(a)** Combination training

significantly reduced dropped clothespins for the MLP and CNN. **(b)** Combination training significantly reduced

transfer times for the MLP and CNN. **(c)** Combination training significantly reduced the subjective workload for

the CNN, evidenced by a lower NASA Task Load Index (TLX) score. Boxplots show median (red line), inter-quartile

range (blue box), and most extreme, non-outlier values (1.5 x inter-quartile range; black whiskers).

LF Comparison

The LF generally reduced the number of dropped clothespins but increased the clothespin transfer time

(Fig. 5). For the CNN and MLP, dropped clothespins were significantly reduced ($p < 0.05$, $p < 0.01$, respectively).

The LF significantly decreased transfer times for the MLP ($p < 0.01$), but significantly increase transfer times for

the CNN and mKF ($p < 0.05$; $p < 0.001$, respectively; Fig 5b). The LF significantly reduced subjective workload for

the MLP ($p < 0.05$) but had no significant effect on subjective workload for the CNN and mKF (Fig. 5c).

Fig. 5 Phase one: the latching filter improves prosthesis control, generally at the cost of speed. **(a)** The latching

filter significantly reduced dropped clothespins for the MLP and CNN. **(b)** The latching filter significantly

increased transfer time for the mKF and CNN, and significantly reduced transfer time for the MLP. **(c)** The

latching filter significantly reduced the subjective workload for the MLP, evidenced by a lower NASA Task Load

Index (TLX) score. Boxplots show median (red line), inter-quartile range (blue box), and most extreme, non-

outlier values (1.5 x inter-quartile range; black whiskers).

Phase Two: Comparing Algorithms

Fragile Egg Task

Preferences for the fragile egg task were significantly different for the different decode algorithms ($p <$

0.01; Fig. 6e). The CNN was significantly preferred over the MLP ($p < 0.05$). Similarly, the mKF strongly trended

toward being preferred over the MLP ($p = 0.05$). Subjective workload also significantly differed among

algorithms ($p < 0.05$; Fig. 6b). The CNN and mKF had significantly lower workload than did the MLP ($p < 0.01$, $p <$

0.05, respectively). There were no significant differences in success rate across the algorithms; however, the

mKF and CNN were generally more successful than the MLP (Fig. 6c). There were no significant differences in

completion time across algorithms, although the mKF generally performed faster than the CNN and MLP (Fig.

6a).

Zipper Task

387   Preferences for the zipper task differed significantly among algorithms ($p < 0.05$; Fig. 6e). The CNN was

388   significantly preferred over the MLP ($p < 0.05$). There were no significant differences in completion time (Fig.

389   6a), but the median completion time for the mKF was the fastest. Inadvertent releases and subjective workload

390   for the decodes did not significantly differ among algorithms (Fig. 6d, 6b, respectively).

Pouring Task

The pouring task was generally perceived as less difficult than the fragile egg and zipper tasks (Fig 6b).

There were no statistical differences for any of the recorded metrics (time, success rate, preference, and

subjective workload). The MLP was preferred by a few participants, in contrast to the fragile egg and zipper

tasks where it was never preferred (Fig. 6e). Participants had high success rates for the pouring task across

algorithms and rarely spilled rice (Fig. 6c); we report only on whether the task was completed without spills or

drops. Completion times were similar among algorithms (Fig. 6a).

Folding Task

399   The folding task was generally perceived as less difficult than the fragile egg and zipper tasks (Fig. 6b)

400   and did not result in any significant differences for the recorded metrics. More participants seem to prefer the

401   CNN for this task (Fig. 6e) even though the trend suggests they inadvertently released the towel more than with

402   the mKF or MLP. The CNN and MLP may also perform faster than the mKF (Fig. 6a).

Aggregate Preferences Across Tasks

404   Aggregated across tasks, user preferences were significantly different ($p < 0.05$; Fig. 6e). Users preferred

405   the CNN over the MLP ($p < 0.01$), and the mKF trended toward being preferred over the MLP when aggregated

406   across tasks ($p = 0.07$).

Fig. 6 Phase two: subjective user-focused metrics reveal differences that objective performance-based measures

may not. a Completion times did not significantly differ for any of the tasks. The mKF was generally faster for

tasks perceived as more difficult, and slower for tasks perceived as less difficult. b The mKF and CNN were

significantly easier to use than the MLP for the fragile egg task, as evidenced by a lower NASA Task Load Index

(TLX) score. c mKF and CNN outperform MLP in fragile egg task, although the difference was insignificant. All

algorithms had high success in the pouring task. d All algorithms performed well and did not significantly differ

in the zipping and folding tasks. e The CNN was significantly preferred over the MLP, aggregated across tasks.

The mKF trended towards significance over the MLP across all tasks. The CNN was significantly preferred over

the MLP for the fragile egg and zipper tasks. Boxplots show median (red line), inter-quartile range (colored

boxes), and most extreme, non-outlier values (1.5 x inter-quartile range; black whiskers).

Discussion

In this two-phase study, we demonstrate the importance of using diverse training sets that include

combinatorial movements; the value of smoothing the algorithm output; and the efficacy of linear and nonlinear

continuous decode algorithms for restoring functional ability to prosthesis users. To our knowledge, no study

has attempted to compare real-world functional improvements of continuous prosthesis control using multiple

algorithms, training paradigms, and output smoothing from the same participants.

In phase one, we found that training on combination movements improves functional performance and

that a latching filter [44] can reduce jitter albeit with some cost in speed. In phase two, we found that a linearly-

based decode algorithm (the mKF) can perform similarly to nonlinear deep-learning techniques (the CNN and

MLP). From a clinical perspective, the mKF may be preferable due to reduced algorithm training time and

computational resources required for real-time control. Phase two demonstrates how patient-centric outcomes

(user preferences and subjective workload) elucidated differences among algorithms when performance metrics

alone could not.

Adding combination movements to the training data benefited the participants two-fold: they

completed tasks faster and with greater control. The improved speed and control seem to preferentially benefit

nonlinear algorithms (MLP, CNN) by producing more consistent control among participants, whereas the mKF

was relatively consistent in both conditions. In terms of speed, the MLP benefited the most from the

combination movement training; however, it was also the slowest in the single-movement training paradigm

and had the most room for improvement. In terms of subjective workload, results from combination training

suggest that nonlinear approaches benefited most from combination training; however, the difference was

significant only for the CNN. Perhaps combination trainings introduce nonlinearities that preferentially benefit

nonlinear neural networks.

The LF generally enhanced control at the cost of speed. We were surprised how little the LF benefited

the mKF relative to the CNN and MLP. The LF uses a hyperparameter that adjusts the level of "latching" (and

inherent speed) of kinematic changes. The LF uses the hyperparameter to selectively dampen small changes in

the decoder estimates (which are assumed to be inadvert movements) while allowing gross movements to

occur with little or no dampening [44]. Our piloting and informal use of these algorithms in lab suggest nonlinear

approaches generally predict kinematic positions similar to the data used for training and thus move to heavily

trained positions quickly and unimpeded by the LF (e.g., rest, full flexion); in contrast, the Kalman filter predicts

kinematics better dispersed throughout the kinematic range, which may be due to a relatively low Kalman gain

[27] that places more value on past estimates than current measurements and decreases responsiveness. Thus,

gross movements could be slowed by a low Kalman gain and the LF as they are initiated. We expect that the

hyperparameter setting favored the nonlinear algorithms, as raw decoder output from the nonlinear algorithms

likely moved between extreme positions (e.g., open hand to closed hand) faster and were not slowed as much

by the LF when gross movement was intended. We expect that optimizing the hyperparameter could improve

speed and control for the mKF especially, but also the CNN and MLP; however, such an optimization is beyond

the scope of this study. The lack of improvement in dropped clothespins with the mKF might also have occurred

because its control was different than the other two decoders, and so the participants did not become as

accustomed to its responsiveness.

In phase two, we examined the decode algorithm's performance using combination training and the LF

in a high-degree-of-freedom setting (five; increased from two in phase one). With the reduced experimental

condition count in this phase, we were able to thoroughly investigate the algorithms' performance across a

broader range of functional tasks. This phase also included a larger, more diverse training dataset, which we

anticipated might preferentially benefit nonlinear algorithms.

We chose tasks representative of daily life that probe qualitatively different aspects of prosthesis

control. The fragile egg task, generally judged most difficult by participants, requires precise movements and

tests the decoder's proportionality. The zipper task requires consistent precision pinching with simultaneous

wrist rotation. The pouring task necessitates smooth wrist rotation during stable grasping. The folding task

requires two-handed object manipulation and grasp stability with positional changes which often degrade

control performance [55]. We found the variety of the tasks to be useful, but the tasks were easier for the

participants than pilot experiments suggested and, in some cases, resulted in floor and ceiling performance

effects. The tasks would need to be more difficult to better elucidate performance differences among the

algorithms in the future.

In part because of these floor and ceiling effects, user preferences provided strong evidence for

identifying the "best" algorithm. Whereas many performance metrics varied only slightly, user preferences

showed greater differences among algorithms. Aggregated across tasks, preferences indicated the CNN was

preferred over the MLP, and the mKF also trended toward being preferred over the MLP. As testing multiple

algorithms in the same experimental sessions allowed users to rank their preferred algorithms, we see

incorporating user preferences as a strength of the study. Translating research practices to the clinic requires

patient-centric, data-driven approaches. Although performance differences may be small among algorithms, our

findings show that differences in user experience exist across the participants, providing additional evidence

toward clinical translation. Subjective workload also provided useful information on user experience, showing

that the CNN and mKF were easier to use during the fragile egg task. From the user's perspective, subjective

workload may be the difference between adopting or abandoning a prosthesis; while two prosthetic systems

may be capable of similar performance, the most intuitive, easy-to-use prosthesis will certainly be preferred.

We found it surprising that the mKF performed similarly to the nonlinear CNN while the MLP

underperformed. Our informal use of the mKF suggests that the mKF performed well because it has a smoother,

more predictable response to input changes, which can improve functional control [49]. We anticipate the CNN

performed well because it used all 528 of the EMG features (versus 48 from the channel selection algorithm for

mKF and MLP) and convolves across time, thus training on more complex data and potentially learning from

temporal characteristics which could be advantageous over the MLP. We informally observed that participants

struggled to proportionally control the prosthesis for the MLP, and it would often close quickly and break the

fragile egg, which may indicate a lack of generalization from the training data.

Although common performance metrics were used, direct comparisons with other studies is difficult due

to different experimental conditions (e.g., degrees-of-freedom, physical prosthesis). In [57], where a two-

degree-of-freedom, continuous, linear regression-based control strategy was compared with a conventional

control strategy, the participant required ~10 s per clothespin for the horizontal to vertical transfer with the

regression-based approach, whereas the conventional strategy required ~14 s. In [45], the two-degree-of-

freedom "proportional simultaneous" control strategy (most similar to what we have termed continuous

control) required ~15 s to move a clothespin up and down (horizontal to vertical then vertical to horizontal),

whereas the conventional strategy required ~29 s. Our median transfer times ranged from ~7 s to ~11 s to

depending on the decoder, also two-degrees-of-freedom. Classifier approaches [19], [61], [62] on the clothespin

relocation task have ranged from ~10 s to ~30 s.

The fragile object task has been performed with various prostheses and fragile objects. Generally, the

ratio between the break force and mass determines the task difficulty; a lower ratio implies greater difficulty. In

the present study, non-amputee participants achieved a 63% median success rate. The break force in the

present experiments was about 20 N, and the device mass was 615 g, resulting in a 0.03 N/g ratio. In a previous

study from our group, an individual with an amputation achieved ~55% success with the mKF on a similar fragile

egg task without sensory feedback [24]. In [78] and [79], the success rates without sensory feedback were ~38%

and 45% with a fragile object that broke at 1.23 N and had a mass of 80 g, a ratio of 0.15 N/g, approximately five

times the ratio of our study. In [77], the success rate was ~86% without sensory feedback, also using a different

fragile object, which broke at 10.7 N and weighed 8 g, a 1.34 N/g ratio (~45 times the ratio of our study).

Transfer times, where reported, were similar among the aforementioned studies and the present study, ranging

from 10 s to 13 s per transfer.

One goal of this study was to determine the best decoders and training paradigms for a portable take-

home system. This study provides strong evidence from real-world functional tasks for using diverse training

sets as input for a decoder and modifying the decode output with an LF. Selecting between linear and nonlinear

decoders is less clear. Our findings indicate comparable performance between an mKF and CNN, where user

preferences were slightly increased for the CNN. Clinical deployment, however, may favor the mKF, which trains

faster and is less computationally expensive. Furthermore, in a take-home clinical trial where more training data

may be collected over time, aggregation of these datasets may begin to favor deep learning approaches [63] like

the CNN or MLP as technologies continue to improve. Overall, our results provide valuable information toward

the clinical implementation of advanced control strategies for prosthesis users.

Conclusions

This work compared myoelectric prosthetic control using activities of daily living with 12 continuous

decoders in a two-phase study. In the first phase, comparing training paradigms and decoder smoothing, we

found that training control algorithms with more complex movement patterns (i.e., simultaneously active digits

and/or wrist) improves control accuracy and speed. Our results showed that nonlinear smoothing of the

decoder output can improve accuracy at some cost in speed. In the second phase, comparing algorithms, our

findings show that subjective preferences favored the CNN and mKF; however, objective performance among

the mKF, MLP, and CNN were similar for many tasks.

This study was completed to select a decoder for a long-term take-home trial with implanted

neuromyoelectric devices. The results demonstrate the importance of using rich training paradigms and

nonlinear algorithm smoothing to improve continuous prosthetic control. The results also highlight the

importance of user-focused subjective metrics in comparing decoders. Where performance differences could

not effectively select a "best" decoder, user preferences added strong evidence toward selecting the mKF or

CNN. Clinical considerations may favor the mKF as it is faster to train and computationally less expensive than

the CNN. Overall, the results herein demonstrate the efficacy of continuous decoders for enabling intuitive

prosthetic control over modern bionic prostheses.

Abbreviations

539 EMG: electromyography; mKF: modified Kalman filter; MLP: multi-layer perceptron; CNN: convolutional neural

540 network; sEMG: surface electromyography; TLX: Task Load Index

Declarations

Ethics approval and consent to participate

544 All participants took part in the experiments after giving their informed consent, as approved by the University

545 of Utah IRB (ID 00098851).

Consent for publication

547 Not applicable.

Availability of data and materials

549 The datasets generated during and/or analyzed during the current study are available from the corresponding

550 author on reasonable request.

Competing interests

552 The author(s) declare(s) that they have no competing interests.

Funding

This work was sponsored by the Hand Proprioception and Touch Interfaces (HAPTIX) program administered by

the Biological Technologies Office (BTO) of the Defense Advanced Research Projects Agency (DARPA) through

the Space and Naval Warfare Systems Center, Contract No. N66001-15-C-4017. Research reported in this

publication was supported by the National Center for Advancing Translational Sciences of the National Institutes

of Health (NIH) under Award Number UL1TR002538 and TL1TR002540. The content is solely the authors'

responsibility and does not necessarily represent the official views of DARPA or the NIH.

Authors' contributions

MP designed the experiments, collected and analyzed the data, and wrote the manuscript. MB designed the

experiments and collected the data. TH designed the experiments, collected and analyzed the data, and wrote

the manuscript. JG optimized the mKF and implemented MLP and CNN algorithms. TD implemented mKF and

MLP algorithms. CD provided experimental guidance and oversaw the study. GC oversaw all aspects of the study

and writing. All authors contributed to revising the manuscript.

Acknowledgments

We thank the participants in this study for their willingness to invest their time to improve advanced prostheses

to benefit amputee populations.

Additional Files

Video 1: Example videos of phase two tasks for each algorithm

References

[1]   Mobius Bionics LLC, "LUKE Arm System," 2019. [Online]. Available: https://www.mobiusbionics.com/wp-
575   content/uploads/2019/09/Mobius-Bionics-LUKE-Product-Spec-Sheet.pdf.

576   [2]   Taska Prosthetics, "TASKA Hand," 2020. https://www.taskaprosthetics.com/the-taska (accessed Jul. 08,
577   2020).

578   [3]   Otto Bock HealthCare LP, "bebionic hand User Guide," 2018. [Online]. Available:
579   https://www.ottobock.ca/media/local-media/prosthetics/upper-limb/files/15667-bebionic-user-
580   guide.pdf.

581   [4]   Otto Bock Healthcare LP, "Michelangelo Hand Brochure," 2017. [Online]. Available:
582   http://www.ottobockus.com/media/local-media/prosthetics/upper-
583   limb/michelangelo/files/michelangelo-brochure.pdf.

[5]   Motion Control Division of Fillauer, "Utah Arm U3 and U3+ User Guide," 2020.

[6]   Open Bionics LLC, "Hero Arm - User Guide," 2020. https://openbionics.com/hero-arm-user-guide/

(accessed Jul. 08, 2020).

[7]   Össur Touch Solutions, "i-Limb Quantum Titanium," 2020.

[8]   Össur Touch Solutions, "i-Limb Access," 2020.

[9]   S. M. Engdahl, B. P. Christie, B. Kelly, A. Davis, C. A. Chestek, and D. H. Gates, "Surveying the interest of
590   individuals with upper limb loss in novel prosthetic control techniques," *J. Neuroeng. Rehabil.*, vol. 12, no.
591   1, pp. 1–11, 2015, doi: 10.1186/s12984-015-0044-2.

592   [10]   E. Biddiss and T. Chau, "Upper limb prosthesis use and abandonment: A survey of the last 25 years,"
593   *Prosthet. Orthot. Int.*, vol. 31, no. 3, pp. 236–257, 2007, doi: 10.1080/03093640600994581.

594   [11]   Coapt LLC, "Complete Control System Gen2 Handbook." 2020.

595   [12]   M. Asghari Oskoei and H. Hu, "Myoelectric control systems-A survey," *Biomed. Signal Process. Control*,
596   vol. 2, no. 4, pp. 275–294, 2007, doi: 10.1016/j.bspc.2007.07.009.

597   [13]   Touchbionics, "I-limb quantum," 2018. http://www.touchbionics.com/products/active-prostheses/i-limb-
598   quantum (accessed Mar. 07, 2019).

599   [14]   E. D'Anna *et al.*, "A closed-loop hand prosthesis with simultaneous intraneural tactile and position
600   feedback," *Sci. Robot.*, vol. 4, no. 27, 2019, doi: 10.1126/scirobotics.aau8892.

601   [15]   T. A. Kuiken *et al.*, "Targeted muscle reinnervation for real-time myoelectric control of multifunction
602   artificial arms," *JAMA - J. Am. Med. Assoc.*, vol. 301, no. 6, pp. 619–628, 2009, doi:
603   10.1001/jama.2009.116.

604   [16]   A. Waris, I. K. Niazi, M. Jamil, K. Englehart, W. Jensen, and E. N. Kamavuako, "Multiday Evaluation of
605   Techniques for EMG-Based Classification of Hand Motions," *IEEE J. Biomed. Heal. Informatics*, vol. 23, no.
606   4, pp. 1526–1534, 2019, doi: 10.1109/JBHI.2018.2864335.

607   [17]   A. J. Young, L. H. Smith, E. J. Rouse, and L. J. Hargrove, "Classification of simultaneous movements using
608   surface EMG pattern recognition," *IEEE Trans. Biomed. Eng.*, vol. 60, no. 5, pp. 1250–1258, 2013, doi:
609   10.1109/TBME.2012.2232293.

610   [18]   L. J. Hargrove, G. Li, K. B. Englehart, and B. S. Hudgins, "Principal components analysis preprocessing for
611   improved classification accuracies in pattern-recognition-based myoelectric control," *IEEE Trans. Biomed.
612   Eng.*, vol. 56, no. 5, pp. 1407–1414, 2009, doi: 10.1109/TBME.2008.2008171.

613   [19]   L. J. Hargrove, L. A. Miller, K. Turner, and T. A. Kuiken, "Myoelectric Pattern Recognition Outperforms
614   Direct Control for Transhumeral Amputees with Targeted Muscle Reinnervation: A Randomized Clinical
615   Trial," *Sci. Rep.*, vol. 7, no. 1, pp. 1–9, 2017, doi: 10.1038/s41598-017-14386-w.

616   [20]   E. D'Anna *et al.*, "A somatotopic bidirectional hand prosthesis with transcutaneous electrical nerve
617   stimulation based sensory feedback," *Sci. Rep.*, vol. 7, no. 1, pp. 1–15, Dec. 2017, doi: 10.1038/s41598-
618   017-11306-w.

619   [21]   M. Ortiz-Catalan, B. Håkansson, and R. Brånemark, "An osseointegrated human-machine gateway for
620   long term sensory feedback and control of artificial limbs," *Sci. Transl. Med.*, vol. 6, no. 257, pp. 1–9,
621   2014.

622   [22]   E. Scheme and K. Englehart, "Training strategies for mitigating the effect of proportional control on
623   classification in pattern recognition-based myoelectric control," *J. Prosthetics Orthot.*, vol. 25, no. 2, pp.
624   76–83, 2013, doi: 10.1097/JPO.0b013e318289950b.

[23]   P. P. Vu *et al.*, "A regenerative peripheral nerve interface allows real-time control of an artificial hand in

upper limb amputees," *Sci. Transl. Med.*, vol. 12, no. 533, pp. 1–12, 2020, doi:
10.1126/scitranslmed.aay2857.

[24] J. A. George *et al.*, "Biomimetic sensory feedback through peripheral nerve stimulation improves
dexterous use of a bionic hand," *Sci. Robot.*, vol. 4, no. 32, p. eaax2352, Jul. 2019, doi:
10.1126/scirobotics.aax2352.

[25] A. Ameri, E. N. Kamavuako, E. J. Scheme, K. B. Englehart, and P. A. Parker, "Support vector regression for
improved real-time, simultaneous myoelectric control," *IEEE Trans. Neural Syst. Rehabil. Eng.*, vol. 22, no.
6, pp. 1198–1209, 2014, doi: 10.1109/TNSRE.2014.2323576.

[26] H. Dantas, D. J. Warren, S. M. Wendelken, T. S. Davis, G. A. Clark, and V. J. Mathews, "Deep Learning
Movement Intent Decoders Trained with Dataset Aggregation for Prosthetic Limb Control," *IEEE Trans.*
Biomed. Eng., vol. 66, no. 11, pp. 3192–3203, 2019, doi: 10.1109/TBME.2019.2901882.

[27] J. A. George, M. R. Brinton, C. C. Duncan, D. T. Hutchinson, and G. A. Clark, "Improved Training Paradigms
and Motor-decode Algorithms: Results from Intact Individuals and a Recent Transradial Amputee with
Prior Complex Regional Pain Syndrome," in *40th International Engineering in Medicine and Biology*
Conference, 2018, vol. 2018, doi: 10.1109/EMBC.2018.8513342.

[28] A. Ameri, M. A. Akhaee, E. Scheme, and K. Englehart, "Regression convolutional neural network for
improved simultaneous EMG control," *J. Neural Eng.*, 2019, doi: 10.1088/1741-2552/ab0e2e.

[29] A. A. Adewuyi, L. J. Hargrove, and T. A. Kuiken, "Resolving the effect of wrist position on myoelectric
pattern recognition control," *J. Neuroeng. Rehabil.*, vol. 14, no. 1, 2017, doi: 10.1186/s12984-017-0246-x.

[30] V. Gilja *et al.*, "A high-performance neural prosthesis enabled by control algorithm design," *Nature*
Neuroscience, vol. 15, no. 12. pp. 1752–1757, 2012, doi: 10.1038/nn.3265.

[31] J. A. George, T. S. Davis, M. R. Brinton, and G. A. Clark, "Intuitive neuromyoelectric control of a dexterous
bionic arm using a modified Kalman filter," *J. Neurosci. Methods*, vol. 330, no. April 2019, p. 108462,
2020, doi: 10.1016/j.jneumeth.2019.108462.

[32] M. Ortiz-Catalan, B. Håkansson, and R. Brånemark, "Real-time and simultaneous control of artificial limbs
based on pattern recognition algorithms," *IEEE Trans. Neural Syst. Rehabil. Eng.*, vol. 22, no. 4, pp. 756–
764, 2014, doi: 10.1109/TNSRE.2014.2305097.

[33] L. J. Hargrove, K. Englehart, and B. Hudgins, "A comparison of surface and intramuscular myoelectric
signal classification," *IEEE Trans. Biomed. Eng.*, vol. 54, no. 5, pp. 847–853, 2007, doi:
10.1109/TBME.2006.889192.

[34] A. Ameri, E. N. Kamavuako, E. J. Scheme, K. B. Englehart, and P. A. Parker, "Real-time, simultaneous
myoelectric control using visual target-based training paradigm," *Biomed. Signal Process. Control*, vol. 13,
no. 1, pp. 8–14, 2014, doi: 10.1016/j.bspc.2014.03.006.

[35] E. N. Kamavuako, K. B. Englehart, W. Jensen, and D. Farina, "Simultaneous and proportional force
estimation in multiple degrees of freedom from intramuscular EMG," *IEEE Trans. Biomed. Eng.*, vol. 59,
no. 7, pp. 1804–1807, 2012, doi: 10.1109/TBME.2012.2197210.

[36] J. M. Hahne *et al.*, "Linear and Nonlinear Regression Techniques for Simultaneous and Proportional
Myoelectric Control," *IEEE Trans. Neural Syst. Rehabil. Eng.*, vol. 22, no. 2, pp. 269–279, Mar. 2014, doi:
10.1109/TNSRE.2014.2305520.

[37] A. Ameri, M. A. Akhaee, E. Scheme, and K. Englehart, "Real-time, simultaneous myoelectric control using
a convolutional neural network," *PLoS One*, vol. 13, no. 9, pp. 1–13, 2018, doi:

10.1371/journal.pone.0203835.

[38]   M. Zia ur Rehman *et al.*, "Multiday EMG-Based classification of hand motions with deep learning
techniques," *Sensors (Switzerland)*, vol. 18, no. 8, pp. 1–16, 2018, doi: 10.3390/s18082497.

[39]   A. Fougner, E. Scheme, A. D. C. Chan, K. Englehart, and Ø. Stavdahl, "Resolving the limb position effect in
myoelectric pattern recognition," *IEEE Trans. Neural Syst. Rehabil. Eng.*, vol. 19, no. 6, pp. 644–651, 2011,
doi: 10.1109/TNSRE.2011.2163529.

[40]   Y. Teh and L. J. Hargrove, "Understanding limb position and external load effects on real-time pattern
recognition control in amputees," *IEEE Trans. Neural Syst. Rehabil. Eng.*, vol. 28, no. 7, pp. 1–1, 2020, doi:
10.1109/tnsre.2020.2991643.

[41]   A. Radmand, E. Scheme, and K. Englehart, "On the suitability of integrating accelerometry data with
electromyography signals for resolving the effect of changes in limb position during dynamic limb
movement," *J. Prosthetics Orthot.*, vol. 26, no. 4, pp. 185–193, 2014, doi:
10.1097/JPO.0000000000000041.

[42]   D. Young *et al.*, "Closed-loop cortical control of virtual reach and posture using Cartesian and joint
velocity commands," *J. Neural Eng.*, vol. 16, no. 2, 2019, doi: 10.1088/1741-2552/aaf606.

[43]   J. E. Downey *et al.*, "Blending of brain-machine interface and vision-guided autonomous robotics
improves neuroprosthetic arm performance during grasping," *J. Neuroeng. Rehabil.*, vol. 13, no. 1, pp. 1–
12, 2016, doi: 10.1186/s12984-016-0134-9.

[44]   J. Nieveen, M. R. Brinton, D. J. Warren, and V. J. Mathews, "A Nonlinear Latching Filter to Remove Jitter
From Movement Estimates for Prostheses," *Press*.

[45]   A. L. Fougner, Ø. Stavdahl, and P. J. Kyberd, "System training and assessment in simultaneous
proportional myoelectric prosthesis control," *J. Neuroeng. Rehabil.*, vol. 11, no. 1, pp. 1–13, 2014, doi:
10.1186/1743-0003-11-75.

[46]   M. Ortiz-Catalan, F. Rouhani, R. Branemark, and B. Hakansson, "Offline accuracy: A potentially misleading
metric in myoelectric pattern recognition for prosthetic control," *Proc. Annu. Int. Conf. IEEE Eng. Med.
Biol. Soc. EMBS*, vol. 2015-Novem, pp. 1140–1143, 2015, doi: 10.1109/EMBC.2015.7318567.

[47]   E. Scheme and K. Englehart, "On the robustness of EMG features for pattern recognition based
myoelectric control; A multi-dataset comparison," *2014 36th Annu. Int. Conf. IEEE Eng. Med. Biol. Soc.
EMBC 2014*, pp. 650–653, 2014, doi: 10.1109/EMBC.2014.6943675.

[48]   M. Cracchiolo *et al.*, "Decoding of grasping tasks from intraneural recordings in trans-radial amputee," *J.
Neural Eng.*, vol. 17, no. 2, p. 026034, 2020, doi: 10.1088/1741-2552/ab8277.

[49]   N. Jiang, I. Vujaklija, H. Rehbaum, B. Graimann, and D. Farina, "Is accurate mapping of EMG signals on
kinematics needed for precise online myoelectric control?," *IEEE Trans. Neural Syst. Rehabil. Eng.*, vol. 22,
no. 3, pp. 549–558, 2014, doi: 10.1109/TNSRE.2013.2287383.

[50]   P. P. Vu *et al.*, "Closed-Loop Continuous Hand Control via Chronic Recording of Regenerative Peripheral
Nerve Interfaces," *IEEE Trans. Neural Syst. Rehabil. Eng.*, vol. 26, no. 2, pp. 515–526, Feb. 2018, doi:
10.1109/TNSRE.2017.2772961.

[51]   N. Jiang, T. Lorrain, and D. Farina, "A state-based, proportional myoelectric control method: online
validation and comparison with the clinical state-of-the-art," *J. Neuroeng. Rehabil.*, vol. 11, no. 1, p. 110,
2014, doi: 10.1186/1743-0003-11-110.

[52]   J. A. Birdwell, L. J. Hargrove, R. F. F. Weir, and T. A. Kuiken, "Extrinsic Finger and Thumb Muscles

Command a Virtual Hand to Allow Individual Finger and Grasp Control," *IEEE Trans. Biomed. Eng.*, vol. 62,
no. 1, pp. 218–226, 2014, doi: 10.1016/j.physbeh.2017.03.040.

[53] L. H. Smith, T. A. Kuiken, and L. J. Hargrove, "Real-time simultaneous and proportional myoelectric
control using intramuscular EMG," *J. Neural Eng.*, vol. 11, no. 6, 2014, doi: 10.1088/1741-
2560/11/6/066013.

[54] I. Vujaklija *et al.*, "Translating research on myoelectric control into clinics-are the performance
assessment methods adequate?," *Front. Neurorobot.*, vol. 11, no. FEB, 2017, doi:
10.3389/fnbot.2017.00007.

[55] J. L. Betthauser *et al.*, "Limb Position Tolerant Pattern Recognition for Myoelectric Prosthesis Control with
Adaptive Sparse Representations from Extreme Learning," *IEEE Trans. Biomed. Eng.*, vol. 65, no. 4, pp.
770–778, 2018, doi: 10.1109/TBME.2017.2719400.

[56] L. Hargrove, L. Miller, K. Turner, and T. Kuiken, "Control within a virtual environment is correlated to
functional outcomes when using a physical prosthesis," *J. Neuroeng. Rehabil.*, vol. 15, no. Suppl 1, 2018,
doi: 10.1186/s12984-018-0402-y.

[57] J. M. Hahne, M. A. Wilke, M. Koppe, D. Farina, and A. F. Schilling, "Longitudinal Case Study of Regression-
Based Hand Prosthesis Control in Daily Life," *Front. Neurosci.*, vol. 14, no. June, pp. 1–8, 2020, doi:
10.3389/fnins.2020.00600.

[58] J. M. Hahne, M. A. Schweisfurth, M. Koppe, and D. Farina, "Simultaneous control of multiple functions of
bionic hand prostheses: Performance and robustness in end users," *Sci. Robot.*, vol. 3, no. 19, p.
eaat3630, 2018, doi: 10.1126/scirobotics.aat3630.

[59] S. Amsuess, P. Goebel, B. Graimann, and D. Farina, "A multi-class proportional myocontrol algorithm for
upper limb prosthesis control: Validation in real-life scenarios on amputees," *IEEE Trans. Neural Syst.*
Rehabil. Eng., vol. 23, no. 5, pp. 827–836, 2015, doi: 10.1109/TNSRE.2014.2361478.

[60] T. A. Kuiken *et al.*, "Targeted reinnervation for enhanced prosthetic arm function in a woman with a
proximal amputation: a case study," *Lancet*, vol. 369, no. 9559, pp. 371–380, 2007, doi: 10.1016/S0140-
6736(07)60193-7.

[61] T. A. Kuiken, L. A. Miller, K. Turner, and L. J. Hargrove, "A Comparison of Pattern Recognition Control and
Direct Control of a Multiple Degree-of-Freedom Transradial Prosthesis," *IEEE J. Transl. Eng. Heal. Med.*,
vol. 4, 2016, doi: 10.1109/JTEHM.2016.2616123.

[62] G. K. Patel, C. Castellini, J. M. Hahne, D. Farina, and S. Dosen, "A Classification Method for Myoelectric
Control of Hand Prostheses Inspired by Muscle Coordination," *IEEE Trans. Neural Syst. Rehabil. Eng.*, vol.
26, no. 9, pp. 1745–1755, 2018, doi: 10.1109/TNSRE.2018.2861774.

[63] J. A. George, A. Neibling, M. D. Paskett, and G. A. Clark, "Inexpensive surface electromyography sleeve
with consistent electrode placement enables dexterous and stable prosthetic control through deep
learning," *41st Int. Eng. Med. Biol. Conf.*, vol. 2020, 2020, [Online]. Available:
http://arxiv.org/abs/2003.00070.

[64] M. D. Paskett *et al.*, "A Modular Transradial Bypass Socket for Surface Myoelectric Prosthetic Control in
Non-Amputees," *IEEE Trans. Neural Syst. Rehabil. Eng.*, vol. 27, no. 10, pp. 2070–2076, 2019, doi:
10.1109/TNSRE.2019.2941109.

[65] J. Nieveen, D. Warren, S. Wendelken, T. Davis, D. Kluger, and D. Page, "Channel Selection of Neural And
Electromyographic Signals for Decoding of Motor Intent," in *Myoelectric Controls Conference*, 2017, p.
720.

[66] W. Wu, Y. Gao, E. Bienenstock, J. P. Donoghue, and M. J. Black, "Bayesian population decoding of motor cortical activity using a Kalman filter," *Neural Comput.*, vol. 18, no. 1, pp. 80–118, 2006, doi: 10.1162/089976606774841585.

[67] A. Hussaini and P. Kyberd, "Refined clothespin relocation test and assessment of motion," *Prosthet. Orthot. Int.*, vol. 41, no. 3, pp. 294–302, 2017, doi: 10.1177/0309364616660250.

[68] N. R. Olsen *et al.*, "An Adaptable Prosthetic Wrist Reduces Subjective Workload," *bioRxiv*, 2019, doi: 10.1101/808634.

[69] S. G. Hart and L. E. Staveland, "Development of NASA-TLX (Task Load Index): Results of Empirical and Theoretical Research," *Adv. Psychol.*, 1988, doi: 10.1016/S0166-4115(08)62386-9.

[70] J. Peat and B. Barton, *Medical Statistics: A Guide to Data Analysis and Critical Appraisal*. 2008.

[71] J. D. Gibbons and S. Chakraborti, "Nonparametric Statistical Inference," in *International Encyclopedia of Statistical Science*, 2011.

[72] R. A. Armstrong, "When to use the Bonferroni correction," *Ophthalmic Physiol. Opt.*, vol. 34, no. 5, pp. 502–508, 2014, doi: 10.1111/opo.12131.

[73] S. G. Meek, S. C. Jacobsen, and P. P. Goulding, "Extended physiologic taction: Design and evaluation of a proportional force feedback system," *J. Rehabil. Res. Dev.*, vol. 26, no. 3, pp. 53–62, 1989.

[74] E. D. Engeberg and S. Meek, "Improved grasp force sensitivity for prosthetic hands through force-derivative feedback," *IEEE Trans. Biomed. Eng.*, vol. 55, no. 2, pp. 817–821, 2008, doi: 10.1109/TBME.2007.912675.

[75] W. H. Kruskal and W. A. Wallis, "Use of Ranks in One-Criterion Variance Analysis," *J. Am. Stat. Assoc.*, 1952, doi: 10.1080/01621459.1952.10483441.

[76] Z. Šidák, "Rectangular Confidence Regions for the Means of Multivariate Normal Distributions," *J. Am. Stat. Assoc.*, 1967, doi: 10.1080/01621459.1967.10482935.

[77] F. Clemente, M. D'Alonzo, M. Controzzi, B. B. Edin, and C. Cipriani, "Non-Invasive, Temporally Discrete Feedback of Object Contact and Release Improves Grasp Control of Closed-Loop Myoelectric Transradial Prostheses," *IEEE Trans. Neural Syst. Rehabil. Eng.*, vol. 24, no. 12, pp. 1314–1322, 2016, doi: 10.1109/TNSRE.2015.2500586.

[78] G. Valle *et al.*, "Biomimetic Intraneural Sensory Feedback Enhances Sensation Naturalness, Tactile Sensitivity, and Manual Dexterity in a Bidirectional Prosthesis," *Neuron*, vol. 100, no. 1, pp. 37-45.e7, Oct. 2018, doi: 10.1016/j.neuron.2018.08.033.

[79] F. M. Petrini *et al.*, "Six-Month Assessment of a Hand Prosthesis with Intraneural Tactile Feedback," *Ann. Neurol.*, vol. 85, no. 1, pp. 137–154, 2019, doi: 10.1002/ana.25384.