Increasing test specificity without impairing sensitivity –

lessons learned from SARS-CoV-2 serology

Thomas Perkmann MD*[1], Thomas Koller MD*[1], Nicole Perkmann-Nagele MD[1], Maria Oszvar-Kozma MSc[1], David W Eyre DPhil[2], Philippa Matthews DPhil[3], Abbie Bown MD[4], Nicole Stoesser DPhil[3], Marie-Kathrin Breyer MD PhD[5], Robab Breyer-Kohansal MD[5], Otto C Burghuber MD[6], Sylvia Hartl MD[5,6], Daniel Aletaha MD[7], Daniela Sieghart PhD[7], Peter Quehenberger MD[1], Rodrig Marculescu MD[1], Patrick Mucher MSc[1], Astrid Radakovics MSc[2], Miriam Klausberger PhD[8], Mark Duerkop PhD[9], Barbara Holzer PhD[10], Boris Hartmann PhD[10], Robert Strassl MD[1], Gerda Leitner MD[11], Florian Grebien PhD[12], Wilhelm Gerner PhD[13,14,15], Reingard Grabherr PhD[8], Oswald F Wagner MD[1], Christoph J Binder MD PhD[1], and Helmuth Haslacher MD PhD[1*]

[1]Department of Laboratory Medicine, Medical University of Vienna, Vienna, Austria

[2]Big Data Institute, Nuffield Department of Population Health, University of Oxford, Oxford, UK

[3]Nuffield Department of Medicine, University of Oxford, Oxford, UK

[4]Public Health England (PHE) Porton Down, Salisbury, UK

[5]Department of Respiratory and Critical Care Medicine and Ludwig Boltzmann Institute for Lung Health, Otto Wagner Hospital, Vienna, Austria

[6]Sigmund Freud University, Medical School and Ludwig Boltzmann Institute for Lung Health, Vienna, Austria

[7]Divison of Rheumatology, Department of Medicine III, Medical University of Vienna, Vienna, Austria

[8]Institute of Molecular Biotechnology, Department of Biotechnology, University of Natural Resources and Life Sciences (BOKU) Vienna, Vienna, Austria

[9]Institute of Bioprocess Science and Engineering, Department of Biotechnology, University of Natural Resources and Life Sciences (BOKU) Vienna, Vienna, Austria

[10]Institute for Veterinary Disease Control, Austrian Agency for Health and Food Safety (AGES), Moedling, Austria

[11]Department of Blood Group Serology and Transfusion Medicine, Medical University of Vienna, Vienna, Austria

[12]Institute for Medical Biochemistry, University of Veterinary Medicine, Vienna, Austria

[13]Institute of Immunology, University of Veterinary Medicine Vienna, Austria

[14]Christian Doppler Laboratory for an Optimized Prediction of Vaccination Success in Pigs, University of Veterinary Medicine, Vienna, Austria

[15]The Pirbright Institute, Pirbright, United Kingdom (current)

*both authors contributed equally

Address all correspondence to:

Helmuth HASLACHER, LECT MD PhD MSc BSc BA

Medical University of Vienna

Department of Laboratory Medicine

Waehringer Guertel 18-20

1090 Vienna

Austria

Tel: +43 1 40400 53190

Fax: +43 1 40495 15547

Mail: helmuth.haslacher@meduniwien.ac.at

Abstract

Background

Serological tests are widely used in various medical disciplines for diagnostic and

monitoring purposes. Unfortunately, the sensitivity and specificity of test systems is often

poor, leaving room for false positive and false negative results. However, conventional

methods used to increase specificity decrease sensitivity and vice versa. Using SARS-

CoV-2 serology as an example, we propose here a novel testing strategy: the

"Sensitivity Improved Two-Test" or "SIT²" algorithm.

Methods

SIT² involves confirmatory re-testing of samples with results falling in a predefined

retesting-zone of an initial screening test, with adjusted cut-offs to increase sensitivity.

We verified and compared the performance of SIT² to single tests and orthogonal testing

(OTA) in an Austrian cohort (1,117 negative, 64 post-COVID positive samples) and

validated the algorithm in an independent British cohort (976 negatives, 536 positives).

Results

The specificity of SIT² was superior to single tests and non-inferior to OTA. The

sensitivity was maintained or even improved using SIT² when compared to single tests

or OTA. SIT² allowed correct identification of infected individuals even when a live virus

neutralization assay could not detect antibodies. Compared to single testing or OTA,

SIT² significantly reduced total test errors to 0.46% (0.24-0.65) or 1.60% (0.94-2.38) at

both 5% or 20% seroprevalence.

Conclusion

For SARS-CoV-2 serology, SIT² proved to be the best diagnostic choice at both 5% and

20% seroprevalence in all tested scenarios. It is an easy to apply algorithm and can

potentially be helpful for the serology of other infectious diseases.

Running Head

Sensitivity-improved two-test serology

Abbreviations

Confidence interval (CI)

Coronavirus disease 2019 (COVID-19)

Enzyme-linked immunosorbent assay (ELISA)

Intensive Care Unit (ICU)

Neutralization titer (NT)

Nucleocapsid (NC)

Nucleoprotein (NP)

Orthogonal testing algorithms (OTA)

Polymerase chain reaction (PCR)

Receptor binding domain (RBD)

Sensitivity-Improved Two-Test (SIT^2)

Severe Acute Respiratory Syndrome Coronavirus 2 (SARS-CoV-2)

Spike protein (S)

Virus neutralization test (VNT)

Introduction

Serological tests are commonly used diagnostic tools in a broad medical field, spanning from infectiology (1) to autoimmunity (2), oncology (3) and transplantation medicine (4). They also play a critical role in animal disease surveillance (5). However, many serological tests come with acceptable but imperfect sensitivities and specificities. Tests with specificities slightly above 90% are considered good (6) or even highly specific (3). However, at low seroprevalence rates, every single percent counts: if the frequency of a given disease in the tested population is only 5%, a specificity of 90% would mean that - even at a sensitivity of 100% - 5 true positives would be matched by ten false positives. Thus, the probability of an individual with a positive test (positive predictive value, PPV) to be a true positive would be only 33%.

During the early phase of the SARS-CoV-2 pandemic, seroprevalences were far below 1% (7). Therefore, highly specific test systems were necessary (>99,5%) to provide good positive predictive values (8). Sensitivity and specificity can be seen as communicating vessels – the improvement of one is usually at the expense of the other (9). Consequently, test systems adjusted by the manufacturers to very high specificities (>99%) showed moderate sensitivity. This problem was particularly evident when non-hospitalized patients were included in the cohorts studied (10-12). To further increase specificity at very low seroprevalence levels, various methods have been proposed, e.g., raising the thresholds for positivity or confirming a positive result with a second test (orthogonal testing) (9, 13, 14). Unfortunately, these specificity improvement strategies inevitably lead to a further reduction of the previously moderate sensitivities.

As the pandemic progressed, the problem became more pronounced as antibodies

declined, and sophisticated statistical models were required to compensate for the

waning sensitivity (15). In the case of SARS-CoV-2, as with any evolving pandemic,

increasing seroprevalence rates worldwide have attenuated the need for higher

specificity.

However, the problem persists in non-epidemic diseases where seroprevalence remains

low. Moreover, each new pandemic begins with extremely low seroprevalence rates as

well, and in the future, we should have better diagnostic strategies in infectious serology

ready here.

In the present work, we propose for the first time an orthogonal test algorithm based on

real-life data for the SARS-CoV-2 antibody tests of Roche, Abbott, DiaSorin, and two

commercial SARS-CoV-2 ELISAs (16) intending to maximize both specificity and

sensitivity at the same time. Although our algorithm follows a general principle, it was

developed based on SARS-CoV-2 antibody tests. The SARS-CoV-2 pandemic provided

a unique opportunity in this regard, as historical samples from before the pandemic are

negative by definition (thus allowing accurate specificity testing). In addition, sufficient

PCR-confirmed positive cases were available quickly, ensuring a sophisticated and

reliable sensitivity verification. Thus, for SARS-CoV-2 - in contrast to most other

circulating microorganisms - a realistic and unusually accurate estimation of specificities

and sensitivities of serological tests was possible. We benefited from this advantage to

develop our sophisticated diagnostic algorithm.

Methods

Study design and cohorts

Sera used in this non-blinded prospective cross-sectional study were either residual

clinical specimens or samples stored in the MedUni Wien Biobank (n=1,181), a facility

specialized in the preservation and storage of human biomaterial, which operates within

a certified quality management system (ISO 9001:2015) (17).

For derivation of the SIT^2 algorithm, sample sets from individuals known to be negative

and positive were established for testing. As previously described (18), samples

collected before 01.01.2020 (i.e., assumed SARS-CoV-2 negative) were used as a

specificity cohort (n=1117): a cross-section of the Viennese population (the LEAD

study)(19), preselected for samples collected between November and April to enrich for

seasonal infections (n=494); a collection of healthy voluntary donors (n= 265); a

disease-specific collection of samples from patients with rheumatic diseases (n=358);

(see also Tables S1 and S2).

The SARS-CoV-2 positive cohort (n=64 samples from n=64 individuals) included

patients testing positive with RT-PCR during the first wave and their close, symptomatic

contacts. Of this cohort five individuals were asymptomatic, 42 had mild-moderate

symptoms, four reported severe symptoms, and 13 were admitted to the Intensive Care

Unit (ICU). The timing of symptom onset was determined by a questionnaire for

convalescent donors and by reviewing individual health records for patients and was in

median 41 [26,25-49] days. For asymptomatic donors (n=5), SARS-CoV-2 RT-PCR

confirmation time was used instead (for more details, see Tables S1 and S3). All

included participants gave written informed consent to donate their samples for research

purposes. The overall evaluation plan conformed with the Declaration of Helsinki as well

as relevant regulatory requirements. It was reviewed and approved by the ethics

committee of the Medical University of Vienna (1424/2020).

For validation of the SIT² algorithm, we used data from an independent United Kingdom

cohort (20), including 1,512 serum/plasma samples (536 PCR confirmed SARS-CoV-2

positive cases; 976 negative cases, collected earlier than 2017).

Antibody testing

For the derivation analyses, SARS-CoV-2 antibodies were either measured according to

the manufacturers' instructions on three different commercially available automated

platforms (Roche Elecsys® SARS-CoV-2 [nuclecapsid total antibody assay, further

referred to as Roche NC], Abbott SARS-CoV-2-IgG assay [nucleocapsid IgG assay,

Abbott NC], DiaSorin LIASION® SARS-CoV-2 S1/S2 assay [S1/S2 combination antigen

IgG assay, DiaSorin S1/S2]) or using 96-well enzyme-linked immunosorbent assays

(ELISAs) (Technoclone Technozym® RBD and Technozym® NP) yielding quantitative

results(16) (for details see Supplement, Supplemental Methods). The antibody assays

used in the validation cohort were Abbott NC, DiaSorin S1/S2, Roche NC, Siemens RBD

total antibody, and a novel 384-well trimeric spike protein ELISA (Oxford Immunoassay)

(20), resulting in 20 evaluable combinations. All samples from the Austrian SARS-CoV-

2-positive cohort also underwent live virus neutralization testing (VNT), and

neutralization titers (NT) were calculated, as is described in detail in the Supplemental

Methods.

Sensitivity improved two-test method (SIT²)

Our newly developed sensitivity improved two-test (SIT²) method consists of the following key components: i) sensitivity improvement by cut-off modification and ii) specificity rescue by a second, confirmatory test (Fig. 1A).

For the first component of the SIT² algorithm, positivity thresholds were optimized for sensitivity according to the first published alternative thresholds for the respective assays, calculated e.g. by ROC-analysis (21-23). Additionally, a high cut-off, above which a result can be reliably regarded as true positive without the need for further confirmation, was defined. These levels were based on in-house observations(18) and represent those values (including a safety margin) above which no more false positives were found. The highest results seen in false-positives were 1.800 COI, 2.86 Index, and 104.0 AU/mL, respectively. Hence, we defined the high cut-off for Roche and Abbott as 3.00 COI/Index and for DiaSorin as 150.0 AU/mL. The lowering of positivity thresholds improves sensitivity; the high cut-off prevents unnecessary re-testing of clearly positive samples. Moreover, the high cut-off avoids possible erroneous exclusion by the confirmatory test. The newly defined interval between the reduced threshold for positivity and the high cut-off is the re-testing zone (Fig. 1A). The initial antibody test (screening test) is then followed by a confirmatory test, whereby positive samples from the re-testing zone of the screening test are re-tested. Also, for the confirmatory test, sensitivity-adapted assay thresholds are needed (Figs.1A, 1B). As false-positive samples are usually only positive in one test system (Fig. S1), false positives can be identified, and specificity markedly restored with minimal additional testing as most samples do not fall within the re-testing zone (18, 24). A flowchart of the testing strategy

and the applied cut-off levels and their associated quality criteria are presented in Figs.

1B, 1C.

Test strategy evaluation

On the derivation cohort, we compared the overall performance of the following SARS-

CoV-2 antibody testing strategies: i) testing using single assays; ii) simple lowering of

thresholds; iii) classical orthogonal testing (OTA), and iv) our newly developed SIT^2

algorithm at assumed seroprevalences of 5% and 20%. As part of the derivation, we

then compared the performance of OTAs and SIT^2 against the results of a virus

neutralization assay. On the validation cohort, we then compared the performance of

OTAs and SIT^2. Finally, we used data from this cohort to evaluate the performance of

SIT^2 versus single tests at seroprevalences of 5%, 10%, 20%, and 50% if the Abbott and

DiaSorin assays (i.e., assays with varying degrees of discrepancies in sensitivity and

specificity) were used.

Statistical analysis

Unless otherwise indicated, categorical data are given as counts (percentages), and

continuous data are presented as median (interquartile range). Total test errors were

compared by Mann-Whitney tests or, in case they were paired, by Wilcoxon tests. 95%

confidence intervals (CI) for sensitivities and specificities were calculated according to

Wilson, 95% CI for predictive values were computed according to Mercaldo-Wald unless

otherwise indicated. Sensitivities and specificities were compared using z-scores. P

values <0.05 were considered statistically significant. All calculations were performed

using Analyse-it 5.66 (Analyse-it Software, Leeds, UK) and MedCalc 19.6 (MedCalc

bvba, Ostend, Belgium). Graphs were drawn using Microsoft Visio (Armonk, USA) and

GraphPad Prism 7.0 (La Jolla, USA).

Results

In the derivation cohort of 1,117 pre-pandemic sera and 64 sera from convalescent

COVID-19 patients (80% non-hospitalized, 20% hospitalized), the Roche NC, Abbott

NC, and DiaSorin S1/S2 antibody assays gave rise to 7/64, 10/64, and 11/64 false-

negative, as well as to 3/1,117, 9/1,117, and 20/1,117 false-positive results. Assuming a

seroprevalence of 20%, this led to 2180, 3120, and 3440 false-negative results per

100,000 tests, and 240, 650 and 1,440 false-positive results per 100,000 tests

respectively (Fig. 2A, right panel).

Effects of threshold lowering on Sensitivity and Specificity

Lowering the positivity thresholds for the Roche NC, Abbott NC, and Diasorin S1/S2 to

0.165 COI, 0.55 Index and 9 AU/mL increased the sensitivity significantly and reduced

false-negative results to 63/64, 62/64, and 57/64 (320, 620, and 2,180 per 100,000 tests

at a seroprevalence of 20%), but substantially increased false-positive results to

18/1,117, 27/1,117, and 31/1,117, respectively (1,280, 1,920 and 2,240 per 100,000

tests, an assumed seroprevalence of 20%; Table S4, Fig. 2A, right panel).

Classical OTA compared to SIT[2]

Subsequently, we evaluated 12 OTA combinations using the fully automated SARS-

CoV-2 antibody tests from Roche NC, Abbott NC, and DiaSorin S1/S2 as screening

tests, each combined with one of the other fully automated assays or a commercially

available NC or RBD-specific ELISA as a confirmation test. Combining these tests as

classical OTAs significantly increased specificity and reduced false positives to 0 (0-

1)/1,117. However, the rate of false negatives was 14 (12-16)/64 (1,095 [955-1,230] per

100,000 tests at 20% seroprevalence), and therefore considerably higher than for single

240　testing strategies. In contrast, the SIT^2 algorithm minimized false positives to 0 (0-

241　2)/1,117 (0 [0-140) per 100,000 tests at 20% seroprevalence) while also reducing false

242　negatives to 5 (3-8)/64 (1,560 [940-2420] per 100,000 tests at 20% seroprevalence, Fig.

243　2A right panel; Table S5).

Reduction of total error rates by the Sensitivity-Improved Two-Test

Of all the methods assessed, SIT^2 reached the lowest total error rates per 100,000 tests

under both 5% and 20% assumed seroprevalence (455 [235-685] and 1,600 [940-2,490]

per 100,000 tests) (Fig. 2B). At a seroprevalence of 5 %, OTA on average performed

better than individual tests, and the total error rates of the single tests were higher for

the Abbott NC and DiaSorin S1/S2 assay (OTA 1,095 [955-1,325] vs. 830 [Roche NC],

1,540 [Abbott NC] and 2,570 [DiaSorin S1/S2] per 100,000 tests). But with a

seroprevalence of 20 %, performance of OTAs, worsened compared to single tests

(OTA 4,380 [3,820-5,000] vs 1,600 [Roche], 2,540 [Abbott] and 4,420 [DiaSorin] per

100,000 tests) (Fig. 2B). Therefore, at both 5% and 20% seroprevalence, SIT^2 resulted

in the lowest overall errors. Compared to OTAs, SIT^2 yielded a similar improvement in

specificity while not suffering from the significant sensitivity reduction (Fig. S2). Since the

better overall performance of SIT^2 compared to OTAs was not due to increased

specificity but improved sensitivity, we subsequently set out to examine these

differences in more detail.

Sensitivities of single tests, OTA and SIT^2 in relation to Neutralization Testing

Next, we compared the sensitivities of the three screening tests as single tests and in

both two-test methods (OTA and SIT^2), benchmarking them using the Austrian

sensitivity cohort (n=64) simultaneously evaluated with an authentic SARS-CoV-2 virus

neutralization test (VNT). Regardless of the screening test used (Roche NC, Abbott NC,

or DiaSorin S1/S2), OTAs had lower sensitivities than single tests (80.5% [78.5-83.6],

78.1% [75.8-82.8], or 75.8% [71.5-78.9] vs. 89.1%, 84.4%, or 82.8% respectively), and

SIT^2 showed the best sensitivities of all methods (95.3% [93.0-96.5], 93.8% [92.2-96.5],

or 87.5% [85.1-88.7]) (Fig. 3). SIT^2 algorithms, including the Roche NC and Abbott NC

assays, achieved similar or even higher sensitivities than VNT (Fig. 3, VNT reference

line), made possible by the unique re-testing zone of SIT^2 (Fig. S3).

Validation of the Sensitivity-Improved Two-Test using an independent cohort

To confirm the improved sensitivity of SIT^2 compared to OTA, we analyzed the

sensitivities of OTAs and SIT^2 in an independent validation cohort of 976 pre-pandemic

samples and 536 post-COVID samples. Out of 20 combinations using the assays Roche

NC (total antibody), Abbott NC (IgG), DiaSorin S1/S2 (IgG), Siemens RBD (total

antibody), and Oxford trimeric-S (IgG), a statistically significant improvement in

sensitivities over OTAs was shown for SIT^2 in 18 combinations (Fig. 4). The

performance was comparable for the remaining two combinations (Siemens RBD with

Oxford trimeric-S and vice versa). Still, no statistically significant improvement could be

achieved due to the high pre-existing sensitivities of these assays on this particular

sample cohort.

To further illustrate the effect of SIT^2 on the outcome of SARS-CoV-2 antibody testing,

we compared single testing versus SIT^2 with the Abbott and DiaSorin assays at varying

assumed seroprevalences (5, 10, 20, and 50%), given that the Abbott NC assay is a

highly specific (99.9%), but moderately sensitive test (92.7%), and the DiaSorin S1/S2

assay has the most limited specificity (98.7%) of all evaluated assays but an acceptable

sensitivity (96.3%). Regardless of whether a lack of specificity (DiaSorin S1/S2) or

sensitivity (Abbott NC) had to be compensated for, SIT^2 improved the overall error rate

compared to the individual tests in all four combinations and at all four assumed

seroprevalence levels (Fig. 5).

Discussion

Serology is a commonly used, multi-purpose analytical method (1-4). However, not all serologic assays have appropriate sensitivities and specificities, especially in low-prevalence settings. The SARS-CoV-2 pandemic prompted the simultaneous development of several antibody tests and, which is rare otherwise, allowed to evaluate these tests with both confirmed positive and negative cases, the latter derived from biobank collections established before the virus emerged. In the case of SARS-CoV-2, false-positive samples are usually not simultaneously reactive in different test systems (14, 25). This led to the hypothesis that reducing the threshold for positivity in screening and confirmation tests would increase the specificity without impairing the sensitivity. A further improvement in sensitivity was possible by defining a high cut-off for the screening test, above which, due to the excellent reliability of high test results, no further confirmation (and, thereby, a possible false-negative result in the confirmation test) was necessary.

In the early waves of the SARS-CoV-2 pandemic, many commercially available SARS-CoV-2 antibody tests did not provide sufficient specificity to achieve acceptable positive predictive values (PPVs), for example, at a seroprevalence rate of 1-5% (13, 25). Lowering positivity thresholds might improve test sensitivity (21-23) and conventional orthogonal testing can maximize specificity (26-28). The latter might increase the positive predictive value, but PPV will only be relevantly increased at low seroprevalences. However, since seroprevalence is often neither known and varies widely from region to region, it is difficult to judge whether a less specific or less sensitive test is the lesser of two evils.

However, the reported imperfect specificities and sensitivities of SARS-CoV-2 antibody

assays are still considerably higher than those of many other serological tests used in

clinical routine. Based on actual data related to SARS-CoV-2, which are in principle

generalizable for all types of serological tests, we, therefore, we propose a new,

universally adaptable two-test system that could, in the case of SARS-CoV-2, perform

better than any other known approach regardless of the actual seroprevalence: the

sensitivity-improved Two-Test or SIT^2. Its generalizability can be inferred from the

following features: i) the adapted cut-offs used to optimize sensitivity were determined in

various independent studies and were not explicitly calculated for our cohort (21-23); ii)

SIT^2 was effective, albeit with different efficiencies, in a total of 32 different test

combinations; and iii) SIT^2 was successfully validated in an independent cohort which

was profoundly different from the derivation cohort. Therefore, SIT^2 does not require a

particular infrastructure or the availability of high-performance individual test systems but

achieves the best performance out of an available test.

Our SIT^2 strategy can rescue the specificity with minimal repeat testing required (see

Table S6). For example, when applying the Roche NC as a screening test to our cohort,

only 27 out of 1,181 samples needed confirmation testing with the Abbott NC test to

correctly identify 62/64 true positives. Simultaneously, all false-positive results were

eliminated, including those added by lowering the cut-offs (Table S4 and Fig. S1).

Additionally, it was more sensitive than virus neutralization testing, which identified only

60/64 clinical positives (Fig. 3). This result is not completely surprising as it is known that

not all patients who recovered from COVID-19 show detectable levels of neutralizing

antibodies (29). Nevertheless, it should be noted that although antibody binding assays

may have a higher sensitivity than neutralization assays, they only partially reflect the

functional activity of SARS-CoV-2-reactive antibodies(30).

Our study has both strengths and limitations. One strength is the size of the cohorts

examined, both in deriving the SIT^2 algorithm (N=1,181) and validating it (N=1,512). The

composition of our specificity cohort is also unique: it consists of three sub-cohorts with

selection criteria to further challenge analytical specificity. The lower cut-offs used to

increase sensitivity were not modeled within our datasets but were derived from ROC-

analyses data of independent studies (21-23). Furthermore, we were able to test the

performance of the two-test systems in a total of 32 combinations, 12 in the derivation

cohort and another 20 combinations in the validation cohort. As a limitation, in the

Austrian cohort, only samples ≥14 days after symptom onset were included. Therefore,

no conclusions on the sensitivity of the early seroconversion phase can be made from

these data. Furthermore, mild and asymptomatic cases were underrepresented in the

British cohort, perhaps leading to an observed higher sensitivity of the test systems.

In conclusion, we describe the novel two-test algorithm SIT^2, which makes it possible to

maintain or even significantly improve sensitivity while approaching 100% specificity.

Acknowledgments

We sincerely thank Marika Gerdov, Susanne Keim, Karin Mildner, Elisabeth Ponweiser,

Manuela Repl, Ilse Steiner, Christine Thun, and Martina Trella for excellent technical

assistance. Finally, we want to thank all the donors of the various study cohorts. The

MedUni Wien Biobank is funded to participate in the biobank consortium BBMRI.at

(www.bbmri.at) by the Austrian Federal Ministry of Science, Research and Technology.

There was no external funding received for the work presented. However, test kits for

the Technoclone ELISAs were kindly provided by the manufacturer. All authors have

read the journal's policy on disclosure of potential conflicts of interest and the following

has to be disclosed: NP-N received a travel grant from DiaSorin. DWE reports lecture

fees from Gilead outside the submitted work. OCB reports grants from GSK, grants from

Menarini, grants from Boehringer Ingelheim, grants from Astra, grants from MSD, grants

from Pfizer, and grants from Chiesi, outside the submitted work. SH does receive

unrestricted research grants (GSK, Boehringer, Menarini, Chiesi, Astra Zeneca, MSD,

Novartis, Air Liquide, Vivisol, Pfizer, TEVA) for the Ludwig Boltzmann Institute of COPD

and Respiratory Epidemiology, and is on advisory boards for G. SK, Boehringer

Ingelheim, Novartis, Menarini, Chiesi, Astra Zeneca, MSD, Roche, Abbvie, Takeda, and

TEVA for respiratory oncology and COPD. PQ is an advisory board member for Roche

Austria and reports personal fees from Takeda outside the submitted work. The Dept. of

Laboratory Medicine (Head: OWF) received compensations for advertisement on

scientific symposia from Roche, DiaSorin, and Abbott and holds a grant for evaluating

an in-vitro diagnostic device from Roche. CJB is a Board Member of Technoclone. HH

receives compensations for biobank services from Glock Health Science and Research

and BlueSky immunotherapies.

References

1.   Lass-Flörl C, Samardzic E, Knoll M. Serology anno 2021-fungal infections: from

invasive to chronic. Clin Microbiol Infect. 2021;27(9):1230-41.

2. Popp A, Kivelä L, Fuchs V, Kurppa K. Diagnosing Celiac Disease: Towards Wide-

Scale Screening and Serology-Based Criteria? Gastroenterol Res Pract.

2019;2019:2916024.

3. Hibbert J, Halec G, Baaken D, Waterboer T, Brenner N. Sensitivity and Specificity

of Human Papillomavirus (HPV) 16 Early Antigen Serology for HPV-Driven

Oropharyngeal Cancer: A Systematic Literature Review and Meta-Analysis. Cancers

(Basel). 2021;13(12).

4. Smatti MK, Al-Sadeq DW, Ali NH, Pintus G, Abou-Saleh H, Nasrallah GK.

Epstein-Barr Virus Epidemiology, Serology, and Genetic Variability of LMP-1 Oncogene

Among Healthy Population: An Update. Front Oncol. 2018;8:211.

5. Gaudino M, Moreno A, Snoeck CJ, Zohari S, Saegerman C, O'Donovan T, et al.

Emerging Influenza D virus infection in European livestock as determined in serology

studies: Are we underestimating its spread over the continent? Transbound Emerg Dis.

2021;68(3):1125-35.

6. Shibata Y, Horita N, Yamamoto M, Tsukahara T, Nagakura H, Tashiro K, et al.

Diagnostic test accuracy of anti-glycopeptidolipid-core IgA antibodies for Mycobacterium

avium complex pulmonary disease: systematic review and meta-analysis. Sci Rep.

2016;6:29325.

7. Arora RK, Joseph A, Van Wyk J, Rocco S, Atmaja A, May E, et al. SeroTracker: a

global SARS-CoV-2 seroprevalence dashboard. Lancet Infect Dis. 2020.

8. Bailey D, Konforte D, Barakauskas VE, Yip PM, Kulasingam V, Abou El Hassan

M, et al. Canadian society of clinical chemists (CSCC) interim consensus guidance for

testing and reporting of SARS-CoV-2 serology. Clin Biochem. 2020.

9.   Turbett SE, Anahtar M, Dighe AS, Garcia Beltran W, Miller T, Scott H, et al.

Evaluation of Three Commercial SARS-CoV-2 Serologic Assays and their Performance

in Two-Test Algorithms. J Clin Microbiol. 2020.

10.   Gattinger P, Borochova K, Dorofeeva Y, Henning R, Kiss R, Kratzer B, et al.

Antibodies in serum of convalescent patients following mild COVID-19 do not always

prevent virus-receptor binding. Allergy. 2020.

11.   Long QX, Tang XJ, Shi QL, Li Q, Deng HJ, Yuan J, et al. Clinical and

immunological assessment of asymptomatic SARS-CoV-2 infections. Nat Med.

2020;26(8):1200-4.

12.   Baron RC, Risch L, Weber M, Thiel S, Grossmann K, Wohlwend N, et al.

Frequency of serological non-responders and false-negative RT-PCR results in SARS-

CoV-2 testing: a population-based study. Clin Chem Lab Med. 2020.

13.   Manthei DM, Whalen JF, Schroeder LF, Sinay AM, Li SH, Valdez R, et al.

Differences in Performance Characteristics Among Four High-Throughput Assays for the

Detection of Antibodies Against SARS-CoV-2 Using a Common Set of Patient Samples.

Am J Clin Pathol. 2020.

14.   Pfluger LS, Bannasch JH, Brehm TT, Pfefferle S, Hoffmann A, Norz D, et al.

Clinical evaluation of five different automated SARS-CoV-2 serology assays in a cohort

of hospitalized COVID-19 patients. J Clin Virol. 2020;130:104549.

15.   Buss LF, Prete CA, Jr., Abrahim CMM, Mendrone A, Jr., Salomon T, de Almeida-

Neto C, et al. Three-quarters attack rate of SARS-CoV-2 in the Brazilian Amazon during

a largely unmitigated epidemic. Science. 2021;371(6526):288-92.

16.   Klausberger M, Duerkop M, Haslacher H, Wozniak-Knopp G, Cserjan-

Puschmann M, Perkmann T, et al. A comprehensive antigen production and

characterisation study for easy-to-implement, specific and quantitative SARS-CoV-2

serotests. EBioMedicine. 2021;67:103348.

17.   Eskandary F, Regele H, Baumann L, Bond G, Kozakowski N, Wahrmann M, et al.

A Randomized Trial of Bortezomib in Late Antibody-Mediated Kidney Transplant

Rejection. J Am Soc Nephrol. 2018;29(2):591-605.

18.   Perkmann T, Perkmann-Nagele N, Breyer MK, Breyer-Kohansal R, Burghuber

OC, Hartl S, et al. Side-by-Side Comparison of Three Fully Automated SARS-CoV-2

Antibody Assays with a Focus on Specificity. Clin Chem. 2020;66(11):1405-13.

19.   Breyer-Kohansal R, Hartl S, Burghuber OC, Urban M, Schrott A, Agusti A, et al.

The LEAD (Lung, Heart, Social, Body) Study: Objectives, Methodology, and External

Validity of the Population-Based Cohort Study. J Epidemiol. 2019;29(8):315-24.

20.   National S-C-SAEG. Performance characteristics of five immunoassays for

SARS-CoV-2: a head-to-head benchmark comparison. Lancet Infect Dis.

2020;20(12):1390-400.

21.   Favresse J, Eucher C, Elsen M, Tre-Hardy M, Dogne JM, Douxfils J. Clinical

Performance of the Elecsys Electrochemiluminescent Immunoassay for the Detection of

SARS-CoV-2 Total Antibodies. Clin Chem. 2020;66(8):1104-6.

22.   Lau CS, Oh HML, Hoo SP, Liang YL, Phua SK, Aw TC. Performance of an

automated chemiluminescence SARS-CoV-2 IG-G assay. Clin Chim Acta.

2020;510:760-6.

23.   Bonelli F, Sarasini A, Zierold C, Calleri M, Bonetti A, Vismara C, et al. Clinical and

Analytical Performance of an Automated Serological Test That Identifies S1/S2-

Neutralizing IgG in COVID-19 Patients Semiquantitatively. J Clin Microbiol. 2020;58(9).

24. Pfluger LS, Bannasch JH, Brehm TT, Pfefferle S, Hoffmann A, Norz D, et al.

Clinical evaluation of five different automated SARS-CoV-2 serology assays in a cohort

of hospitalized COVID-19 patients. J Clin Virol. 2020;130:104549.

25. Perkmann T, Perkmann-Nagele N, Breyer MK, Breyer-Kohansal R, Burghuber

OC, Hartl S, et al. Side by side comparison of three fully automated SARS-CoV-2

antibody assays with a focus on specificity. Clin Chem. 2020.

26. Turbett SE, Anahtar M, Dighe AS, Garcia Beltran W, Miller T, Scott H, et al.

Evaluation of Three Commercial SARS-CoV-2 Serologic Assays and Their Performance

in Two-Test Algorithms. J Clin Microbiol. 2020;59(1).

27. Bolotin S, Tran V, Osman S, Brown KA, Buchan SA, Joh E, et al. SARS-CoV-2

seroprevalence survey estimates are affected by anti-nucleocapsid antibody decline. J

Infect Dis. 2021.

28. Irsara C, Egger AE, Prokop W, Nairz M, Loacker L, Sahanic S, et al. Evaluation of

four commercial, fully automated SARS-CoV-2 antibody tests suggests a revision of the

Siemens SARS-CoV-2 IgG assay. Clin Chem Lab Med. 2021;0(0).

29. Kalkan Yazici M, Koc MM, Cetin NS, Karaaslan E, Okay G, Durdu B, et al.

Discordance between Serum Neutralizing Antibody Titers and the Recovery from

COVID-19. J Immunol. 2020;205(10):2719-25.

30. Padoan A, Bonfante F, Pagliari M, Bortolami A, Negrini D, Zuin S, et al. Analytical

and clinical performances of five immunoassays for the detection of SARS-CoV-2

antibodies in comparison with neutralization activity. EBioMedicine. 2020;62:103101.

Figure Legends

Fig. 1. A) *The Sensitivity Improved Two-Test (SIT^2) algorithm includes sensitivity improvement by adapted cut-offs and a subsequent specificity rescue by re-testing all samples within the re-testing zone of the screening test by a confirmatory test.* **B)** *Testing algorithm for SIT^2 utilizing a screening test on an automated platform (ECLIA/Roche, CMIA/Abbott, CLIA/DiaSorin) and a confirmation test, either on one of the remaining platforms or tested by means of ELISA (Technozym RBD, NP).* **C)** *All test results between a reduced cut-off suggested by the literature, and a higher cut-off, above which no more false-positives were observed, were subject to confirmation testing.* **... *results between 12.0 and 15.0, which are according to the manufacturer considered borderline, were treated as positives;* ***... *suggested as a cut-off for seroprevalence testing;* ****... *determined by in-house modeling;* [1]... *see (21);* [2]... *see (22);* [3]... *see (23).*

Fig. 2. *False-positives (FP)/false-negatives (FN) (**A**) and total error (**B**) of single tests, tests with reduced thresholds according to (21-23), orthogonal testing algorithms (OTAs) and the Sensitivity Improved Two-Test (SIT^2) algorithm at 5 and 20% estimated seroprevalence. Data in (B) were compared by Mann-Whitney tests (unpaired) or Wicoxon tests (paired).* *... *P<0.05;* **...*P<0.01;* ***...*P<0.001.*

Fig. 3. *Sensitivities of single tests, orthogonal testing algorithms (OTAs) and the Sensitivity Improved Two-Test (SIT^2) algorithm. The dotted line indicates the sensitivity of virus neutralization test (VNT).*

Fig. 4. *Differences in sensitivity and specificity (mean±95% confidence interval) between the Sensitivity Improved Two-Test (SIT^2) algorithm and standard orthogonal testing*

algorithms (OTAs) within the UK validation cohort. *… P<0.05; **...P<0.01;

…P<0.001; *...P<0.0001

***Fig. 5.** Comparing false-positives (FP), false-negatives (FN), and total error (TE) for two*

*selected test systems, **A)** Abbott, **B)** DiaSorin, between different Sensitivity Improved*

Two-Test (SIT^2) combinations and the respective single test within the UK validation

cohort for different estimated seroprevalences.

A

Step 1: Sensitivity Improvement

Step 2: Specificity Rescue

Confirmation of those between the reduced and the high cut-off in the screening test (yellow area) by a second test system

🔵 Non-infected	🔴 Infected	🟩 Test positive	🟨 Re-testing area

B

C

	Roche	Abbott	DiaSorin	Technozym RBD	Technozym NP
Sensitivity	89.1 (79.1-94.6)	84.4 (73.6-91.3)	82.8 (71.8-90.1)	85.9 (75.4-92.4)	78.1 (66.6-86.5)
Specificity	99.7 (99.2-99.9)	99.2 (98.5-99.6)	98.2 (97.3-98.8)	99.2 (98.5-99.6)	99.1 (98.4-99.5)
PPV (5%)	94.6 (84.9-98.2)	84.6 (74.0-91.4)	70.9 (60.2-79.2)	84.9 (74.4-91.6)	82.1 (71.0-89.6)
NPV (5%)	99.4 (98.9-99.7)	99.2 (98.6-99.5)	99.1 (98.4-99.5)	99.3 (98.7-99.6)	98.9 (98.2-99.3)
Cut-off (manufacturer)	1.000 COI	1.40 I	12.0 AU/mL**	8.000 U***	8.000 U***
Reduced Cut-off	>0.165[1]	≥0.55[2]	≥9.0[3]	>3.000 U****	>3.500 U****
High Cut-off	3.000	3.00	150.0	-	-

A

False positives/negatives (5% seroprevalence estimated)

False positives/negatives (20% seroprevalence estimated)

- ● FP
- ○ FN
- ⋯ FP with standard cut-off
- ⋯ FN with standard cut-off

B

Total error (5% seroprevalence estimated)

Total error (20% seroprevalence estimated)

Sensitivity with standard testing, OTA and SIT^2

A Abbott

B DiaSorin