Validation of a Computerized Cognitive Test Battery for Detection of Dementia and Mild Cognitive Impairment

Siao Ye[a], Kevin Sun[b], Huy Q. Phi[c], Brian Ko[d], Carolyn M. Parsey[d], Bin Huang[b,*], Reza Hosseini Ghomi[b,d*]

[a]Department of Biosciences, Rice University, Houston, TX, 77005, USA

[b]BrainCheck, Inc, Houston, TX, 77021, USA

[c]Clinical Research Division, Fred Hutchinson Cancer Research Center, Seattle, WA, 98109, USA

[d]Department of Neurology, University of Washington School of Medicine, Seattle, WA, 98195, USA

[*]Corresponding author(s):

E-mail: bin@braincheck.com (BH), reza@braincheck.com (RHG)

Abstract

Background

Early detection of dementia is critical for intervention and care planning, but remains difficult. Computerized cognitive testing provides an accessible and promising solution to address these current challenges. This study evaluated a computerized cognitive testing battery (BrainCheck) for its diagnostic accuracy and ability to distinguish the severity of cognitive impairment.

Methods

99 participants diagnosed with Dementia, Mild Cognitive Impairment (MCI), or Normal Cognition (NC) completed the BrainCheck battery. Statistical analyses compared participant performances on BrainCheck based on their diagnosis group.

Results

BrainCheck battery performance showed significant differences between the NC, MCI, and Dementia groups, achieving $\geq$88% sensitivity/specificity for separating NC from Dementia, and $\geq$77% sensitivity/specificity in separating the MCI group from NC/Dementia groups. Three-group-classification found true positive rates $\geq$80% for the NC and Dementia groups and $\geq$64% for the MCI group. BrainCheck was able to distinguish between diagnoses of Dementia, MCI, and NC, providing a potentially reliable tool for early detection of cognitive impairment.

Keywords: Cognitive Test, Mild Cognitive Impairment, Dementia, Cognitive Decline, Repeatable Battery, Discriminant Analysis

Background

In proportion with the increasingly aging population, the incidence of dementia is on the rise; projected to affect nearly 14 million people in the United States and upwards of 152 million people globally in the coming decades [1–3]. Current rates of undetected dementia are reported as high as 61.7% [4], and available treatments are limited to promoting quality of life rather than reversal or cure of the disease process. The ability to properly identify and treat dementia at this scope requires an active approach focused on early identification. Early detection of dementia provides access to timely interventions and knowledge to promote patient health and quality of life before symptoms become severe. Known treatment interventions help to address underlying conditions and minimize distressing symptoms, providing reasons to implement these earlier on. Early and accurate diagnosis allows for proper preparation for patients, caregivers, and their families, resulting in improved caregiver well-being and delayed nursing home placements [5, 6]. Further, it helps to characterize early-stage dementia patients for clinical trials, exploring the latest therapeutics and validating biomarkers indicative of specific pathologies. Despite the benefits, early detection is a challenge with current clinical protocols, leaving many patients undiagnosed until symptoms become noticeable in later stages of the illness.

Considered an early symptomatic stage of dementia, mild cognitive impairment (MCI) signifies a level of cognitive impairment between normal cognition (NC) and dementia [7]. While not all MCI cases progress, the conversion rate of MCI to dementia has been observed at approximately 5%-10% [8]. This stresses the importance of MCI

identification in early detection and clinical intervention for dementia, included in recommendations from the National Institute on Aging and the Alzheimer's Association [9]. Detection of MCI has been fairly successful using brief cognitive screening assessments. The widely used Montreal Cognitive Assessment (MoCA) has demonstrated 83% sensitivity and 88% specificity in distinguishing MCI from NC, and its ability to distinguish dementia from MCI has demonstrated 90% sensitivity and 63% specificity [10, 11]. The Saint Louis University Mental Status (SLUMS) exam and the Mini-Mental State Examination (MMSE) also show the ability to distinguish impaired individuals. For example, the diagnostic accuracy of the MMSE has been reported to be 62% sensitivity and 87% specificity for distinguishing MCI from NC and 80% sensitivity and 85% specificity for distinguishing dementia from MCI [12, 13]. Similar performance has been observed with the SLUMS exam, reporting 94% sensitivity and 87% specificity for diagnosing dementia and 82% sensitivity and 86% specificity for diagnosing cognitive impairment [14]. While traditional 'paper and pen' based versions of these assessments have typically been used in standard practice, digitized or computerized versions are beginning to be more available; however, they have been slow in adoption in clinical practice. Although these screening tests do well in their ability to detect MCI, they are often utilized to prompt an indication for further comprehensive testing to better specify patterns of cognitive impairment seen in neurological disorders [15].

After a diagnosis of MCI, there is often a referral for neuropsychological testing (NPT); a comprehensive evaluation of cognitive domains (e.g., attention, working memory, language, visuospatial skills, executive functioning, memory), that are used to support

clinical diagnoses and further delineate specific neurocognitive disorders. NPT can determine patterns of cognitive functioning that relate to normal aging, MCI, and dementia progression with specificity from 67% - 99% [16]. A major strength of NPT is the ability to characterize cognitive impairment, providing clues to underlying pathology, and thereby improving diagnostic accuracy to guide appropriate treatment. However, NPT comes with downsides including financial cost, long appointment times, and high levels of training and expertise required to conduct and interpret tests. To address these pitfalls and meet future demands, computerized and digitized NPT have also been developed to address issues of accessibility and efficiency [17, 18]. Prior studies have shown that some computerized cognitive tests demonstrate high accuracy in differentiating dementia patients from healthy participants, but do not have adequate psychometrics to distinguish MCI from dementia [19–22]. Ideally, computerized cognitive tests would aim to be more rapid and maximize accessibility to both patients and providers to address the expected uptick in dementia. This would also yield multiple benefits including maintaining testing standardization, alleviating time pressures of modern clinical practice, and providing a comprehensive assessment of cognitive function to strengthen a clinical diagnosis. Importantly, in the new era of practicing amid the COVID-19 pandemic [23], increasing the accessibility of remote cognitive testing for vulnerable/high-risk patients is essential.

This study evaluated BrainCheck, a computerized cognitive test battery previously validated for its diagnostic accuracy for detecting concussion [24] and dementia-related cognitive decline [25]. BrainCheck is portable, allows for remote administration, and

most insurance plans in the US reimburse it. Furthering its validation for dementia-related cognitive decline, we sought to assess BrainCheck's utility as a diagnostic aid to accurately distinguish the severity of cognitive impairment using existing clinical diagnoses of NC, MCI, and Dementia. We compared BrainCheck scores to comprehensive clinical diagnosis (subjective cognitive impairment or NC, MCI, or Dementia). Our goal was to further demonstrate the utility of BrainCheck for cognitive assessment, specifically as a diagnostic aid in cases where NPT may be unavailable, or when a comprehensive evaluation is not indicated.

Methods

Recruitment

This study was approved by the University of Washington Institutional Review Board for human subject participation. Participants were recruited using the University of Washington Alzheimer's Disease Research Center's research registry [26], a database of individuals willing to participate in dementia-related research. Many of these potential participants had been evaluated at the Memory and Brain Wellness Center in Seattle, Washington, and voluntarily signed up to be put on the registry. Those with listed addresses within a 70-mile radius of Seattle, Washington were contacted by phone or email provided within the registry. If the person was unable to physically use an iPad, too cognitively impaired to understand or follow instructions, or if the primary contact (e.g., spouse) indicated that the person was unable to participate, they were not recruited for the study. When study procedures were modified from in-person to remote administration due to the COVID-19 pandemic (approximately March 2020), participants

outside the initial geographical range were contacted. We required that these participants had access to either an iPad with iOS 10 or later, or a touchscreen computer and Wi-Fi connectivity to participate in the study.

Using the provided primary cognitive diagnosis within the registry by the memory clinic, participants were divided into one of three groups: 1) Normal Cognition (NC), indicated by subjective cognitive complaint or no diagnosis of cognitive impairment, some of which were self-reported, 2) Mild Cognitive Impairment (MCI), representing both amnestic and nonamnestic subtypes, or 3) Dementia, which included dementia due to Alzheimer's Disease (AD), Frontotemporal dementia, Vascular dementia, Lewy body dementia, Mixed dementia, or Atypical AD.

Five participants were not recruited from the registry but via snowball sampling from other participants. The recruitment of these participants was simply due to convenience, typically a spouse that was also present at the time of testing. Four of these participants were placed into the NC group after a self-report denying symptoms or a history of cognitive impairment; these four were not patients of the memory clinic. One of the participants recruited from snowball sampling was a patient from the memory clinic and placed in the AD group based on their most recent diagnosis retrieved from their medical records. All five were in the on-site administration group.

Study design and procedures

On-site administration

Data collection for on-site administration was collected from October 2019 to February 2020. A session was held either in the participant's home or in a public setting that was well-lit, quiet, and distraction-free. Consent forms were reviewed and signed by the participant, or legally authorized representative, and examiner, with both parties obtaining a copy. The study was designed for participants to complete one session with a moderator using a provided iPad (Model MR7G3LL/A) connected to Wi-Fi to complete the BrainCheck battery. Prior to testing, participants were briefed on BrainCheck, and moderator guidance was limited to questions and assistance requested by the participant during the practice portions. Participants received a gift card for participation at the conclusion of the study session.

Remote administration

Due to the COVID-19 pandemic and interest in preliminary data on remote cognitive testing, study procedures were modified to accommodate stay-at-home orders in Washington state. Data collection resumed from April 2020 to May 2020, with modified procedures using remote administration. These participants provided written and verbal consent and were administered the BrainCheck battery remotely over a video call with the moderator. Participants used their personal iPads or touchscreen computer browser to complete the BrainCheck battery. The same method for on-site administration (described above) was used for remote administration.

Measurements

A short description of each of the 5 assessments that make up the BrainCheck battery (V4.0.0) is listed in Table 2. More detailed descriptions may be found in a previous validation study [24]. After completion of the BrainCheck battery, the score of each assessment is calculated using assessment specific measurements by the BrainCheck software (Table 2). The BrainCheck Overall Score is a single, cumulative score for the BrainCheck battery that represents general cognitive functioning. This score is calculated by taking the weighted average of all completed assessment scores. If an assessment is timed-out, a penalty is applied by setting this assessment score to be zero. The normalized assessment scores and BrainCheck Overall Scores are corrected for participant age and device used (iPad versus computer) using the mean and standard deviation of the corresponding score from a normative database previously collected by BrainCheck [24,25]. The score generated follows a standard normal distribution, where a lower score indicates lower assessment performance and cognitive functioning.

Table 1. Cognitive assessments in BrainCheck

Assessment	Description	Measurement for Assessment Score	Cognitive Domain
Immediate and Delayed Recognition	First, Immediate Recognition serially displays 10 words and then asks whether a word was just seen — displays either a distractor word or a	Number of correct answers for each assessment	Memory

	target word (20 trials). At the end of the testing battery, without seeing the original list again, participants are again presented with 20 words and asked whether each word was presented before.		
Digit Symbol Substitution	Participants must match an arbitrary correspondence of symbols to digits; when presented with a new symbol, they input the corresponding digit as quickly as possible.	Median duration time of matching the digits and symbols	Executive function
Flanker	Participants are presented with a target item (in this case, a central arrow) flanked by congruent (>> > >>), or incongruent (<< > <<) arrows. Participants identify the direction of the target as quickly and accurately as possible.	Median reaction time of correct direction choice	Alertness, spatial awareness, and executive function
Stroop	Participants are instructed to find a word matching the given name of a color. There are three types of trials: NEUTRAL in which all words are presented with black font, CONGRUENT in which the word and	Median reaction time of incongruent word-color pairs	Executive function and impulsivity

font color are the same (e.g., the word RED presented in red font), and NON-CONGRUENT in which the word indicates a different color than the font (e.g., the word RED presented in green font). A time-out mechanism is triggered if there is not a completion of a trial in the assessment within 30 seconds.

Trail Making Test	Participants are instructed to connect a set of 25 dots in their correct order as rapidly as possible. Trail Making Test A uses only numbers (1 through 25), while Trail Making Test B employs alternating numbers and letters (1 – A – 2 – B – 3 – C - …). A time-out mechanism is triggered if there is not a completion of a trial in the assessment within 30 seconds	Median duration of completing the tasks for Trails A and Trails B	Visual attention and cognitive flexibility

Statistical analysis

Statistical analyses were performed using Python version 3.8.5 and R version 3.6.2 programming languages. All tests were two sided, and significance was accepted at the 5% level (alpha = 0.05). Comparison of means of groups was made by an analysis of variance (ANOVA) test for normally distributed data. The chi-square test was used to analyze differences in categorical variables.

To evaluate BrainCheck performance among participants in different diagnosis groups while adjusting for age, sex, and administration type, we used linear regression models in which the outcome variables were duration to complete BrainCheck battery, individual BrainCheck assessment scores and BrainCheck Overall Scores. P-values were corrected using Tukey's method for multiple comparisons. To assess the accuracy of the BrainCheck Overall Score in binary classification of participants in the different diagnosis groups (NC vs. Dementia, NC vs. MCI and MCI vs. Dementia), receiver operating characteristic (ROC) curves with area under the curve (AUC) calculations were generated to determine diagnostic sensitivity and specificity. In assessing BrainCheck for three group classification, we used volume under the three-class ROC surface (VUS) method from Luo et al. [27], to define optimal cutoffs for the BrainCheck Overall Score and find the maximum diagnostic accuracy.

Results

Participant characteristics and demographics

A total of 241 individuals were contacted to participate and 99 participants completed the study. Demographics details of the participants are provided in Table 2. The three groups did not differ to a significant degree in terms of education, administration type and recruitment type, but there were differences in age and sex.

Table 2. Participant demographics

Demographics	NC	MCI	Dementia	p-value

Sample (n = 99)	35	22	42	-
Age mean(SD)	67.8 (9.6)	73.5 (5.9)	71.5 (9.0)	**0.04**[a]
Sex				
Female	25	8	16	**< 0.01**[b]
Male	10	14	26	
Education level				
Some college or less	2	2	8	
BA / BS college graduate	10	6	11	0.70[b]
Post Bachelor	14	9	16	
N/A	9	5	7	
Administration Type				
On-site	29	16	29	0.37[b]
Remote	6	6	13	
Recruitment Type				
Registry	31	22	41	0.09[b]
Snowball	4	0	1	

Significant p-values are shown in bold

[a] Calculated by the ANOVA test

[b] Calculated by the chi-square test

Completion of Assessments

We found most participants in the NC group were able to complete the assessments whereas the Dementia group had a higher 'timeout' rate, with MCI falling in between the two (Fig. 1). The timeout function occurs when a participant can not complete a trial of the assessment in 30 seconds; it is embedded in the Stroop and Trails A/B assessments. Timeouts were mainly due to response delays; where participants were attempting the test but could not answer quickly enough. Overall, the Dementia group

took significantly more time to complete the BrainCheck battery (*median*=30.5 mins, *IQR*=23.4-37.1 mins) compared to the MCI (*median*=17.8 mins, *IQR*=15.4-19.6 mins) and NC group (*median*=17.8 mins, *IQR*=15.4-19.6 mins).

Fig. 1: Completion of assessments and durations to complete BrainCheck battery.

(A) Timeout rates of the Stroop and Trails A/B assessments for each diagnosis group. The BrainCheck Stroop and Trails A/B assessments timeout if participants cannot complete a trial of the assessment in 30 seconds. (B) Duration (mins) to complete the BrainCheck Battery for each diagnosis group. Letters (*a*, *b*) indicate significant differences ($p<0.05$) in linear regression models adjusted for age, gender and administration type. Any two groups sharing a letter are not significantly different.

BrainCheck performance

BrainCheck assessments were compared across the three groups using a Linear Regression Model adjusted for age and gender (Fig. 2 and Table 3). Individual scores, such as the BrainCheck Overall Score, were normalized for age and device. Overall, participants with greater cognitive impairment showed lower BrainCheck assessment

scores. All individual assessments except Trails B showed significant differences in performance between the NC and Dementia groups, while two of the seven assessments (Immediate Recognition and Digit Symbol Substitution) showed significant differences in performance between all three groups (Fig 2 and Table 3). Digit Symbol Substitution, Flanker and Trails A/B assessments showed long tails in the scores of the Dementia group because some participants in the Dementia group only completed parts of the assessments or exhibited low accuracy (SI Fig.).

Fig. 2: Pairwise comparison of participant groups based on normalized BrainCheck assessment scores of assessments.

For each assessment, any two groups sharing a letter are not significantly different. Otherwise, they are significantly different (p<0.05) in linear regression models adjusted for age, gender and administration type. The outliers identified by IQR method in each assessment are removed before the comparison.

Table 3. Linear regression analyses using each BrainCheck assessment scores and BrainCheck Overall Score as the outcome variable in separate adjusted models[1].

	NC	MCI	Dementia	Dementia vs. NC	Dementia vs. MCI	MCI vs. NC
Assessments	Estimated Marginal Means (SE)			Contrast Estimate (p)		
Immediate Recognition*	0.17 (0.44)	-1.93 (0.50)	-3.36 (0.36)	-3.54 **(<0.001)**	-1.43 **(0.04)**	-2.10 **(<0.01)**
Delayed Recognition	0.06 (0.34)	-2.16 (0.39)	-2.92 (0.28)	-2.98 **(<0.001)**	-0.76 (0.23)	-2.23 **(<0.001)**
Digit Symbol Substitution*	0.80 (0.27)	-0.21 (0.29)	-1.23 (0.29)	-2.02 **(<0.001)**	-1.01 **(0.04)**	-1.01 **(0.03)**
Flanker	0.76 (0.45)	-0.74 (0.51)	-2.64 (0.41)	-3.4 **(<0.001)**	-1.89 **(<0.01)**	-1.5 (0.06)
Stroop	-0.43 (0.12)	-0.63 (0.13)	-0.91 (0.12)	-0.49 **(0.01)**	-0.28 (0.23)	-0.21 (0.46)
Trails A	-0.01 (0.33)	-0.75 (0.36)	-1.69 (0.30)	-1.67 **(<0.01)**	-0.94 (0.11)	-0.74 (0.29)

Trails B	0.51	0.21	-0.16	-0.67	-0.37	-0.30
	(0.21)	(0.23)	(0.24)	(0.08)	(0.47)	(0.57)
Normalized BrainCheck Overall Score*	0.71	-2.15	-5.63	-6.34	-3.48	-2.86
	(0.55)	(0.62)	(0.45)	**(<0.001)**	**(<0.001)**	**(<0.01)**

Significant p-values are shown in bold. Assessments marked with * indicate significant differences across all three diagnosis groups.

[1] Adjusted for age, sex and administration type

The BrainCheck Overall Score is a composite of all individual assessments within the BrainCheck battery, representing overall performance (See details under the measurements section). Using an existing normative population database, partly compiled from controls in previous studies [24,25], the BrainCheck Overall Score was adjusted for age and the device used to generate the normalized BrainCheck Overall Scores. The normalized BrainCheck Overall Scores differed significantly among these three groups ($p<0.05$). The NC group scored significantly higher than the MCI and Dementia groups. The Dementia group demonstrated significantly lower scores than the other two groups (Fig. 3).

Fig 3: Comparison of normalized BrainCheck Overall Scores among groups.

The normalized BrainCheck Overall Score follows a standard normal distribution. Letters (*a, b, c*) indicate significant differences (p<0.05) on the linear regression model adjusted for age, gender and administration type.

BrainCheck diagnostic accuracy

Using ROC analysis, Braincheck Overall Scores achieved a sensitivity of 94% and a specificity of 88% for classifying NC and Dementia participants (AUC=0.95); a sensitivity of 83% and a specificity of 86% for classifying between NC and MCI participants (AUC=0.84); and a sensitivity of 77% and a specificity of 83% for classifying between MCI and Dementia participants (AUC=0.79) (Fig 4).

Fig 4. ROC curves for the BrainCheck Overall Score in classifying participants of different groups.

ROC curves with AUC's for the BrainCheck Overall Score in binary classification of: (A) NC vs. Dementia, (B) NC vs. MCI, (C) MCI vs. Dementia.

Using methods described by Luo and colleagues for three group classification [27], the optimal lower and upper cutoffs of the normalized BrainCheck Overall Score in maximizing diagnostic accuracy were -3.64 and -0.06, respectively. This achieved true positive rates of 83% for the NC group, 64% for the MCI group, and 83% for the Dementia group (Fig. 5).

Fig. 5: Optimal BrainCheck cutoff scores for distinguishing NC, MCI, and Dementia groups.

(A) Individual participant normalized BrainCheck Overall Scores, where the x-axis is the index of the participant, sorted by primary diagnosis (Dementia - red, MCI - green, and NC - blue). (B) Boxplots of normalized BrainCheck Overall Scores for each diagnosis group. The normalized BrainCheck Overall Score follows a standard normal distribution. The dashed lines label the optimal cutoff scores for distinguishing the diagnosis groups.

Discussion

Consistent with prior findings in concussion [24] and dementia/cognitive decline [25] samples, this study demonstrated that BrainCheck is consistent in its capability to detect cognitive impairment. Our study demonstrated that BrainCheck can detect MCI, and reliably differentiate between the severity of cognitive impairments in the diagnosis groups (NC, MCI, and Dementia). As expected, participants with more severe cognitive impairment performed worse across the individual assessments and on BrainCheck

Overall Scores. The BrainCheck Overall Scores separated participants of different diagnostic groups successfully with high sensitivity and specificity.

BrainCheck Overall Scores were more robust in distinguishing these groups where participants in the Dementia group had significantly lower scores than ones in the NC group. The BrainCheck battery was able to distinguish between NC and Dementia participants, with 94% sensitivity and 88% specificity. These findings show that the BrainCheck Overall Score demonstrates better accuracy for differentiating NC from dementia, compared to the MMSE, SLUMS, and MoCA screening measures [14, 15, 29]. People with MCI usually experience fewer cognitive deficits and preserved functioning in activities of daily living compared to those with dementia [30], and our findings of sensitivity and specificity with separating MCI from other groups were slightly lower than the NC versus the Dementia differentiations (Fig. 5). Nonetheless, the BrainCheck Overall Score showed sensitivities and specificities greater than 80% in distinguishing MCI from NC and Dementia groups, which is comparable to the MoCA, SLUMS, and MMSE [11-15]. Furthermore, a review of validated computerized cognitive tests indicated AUC's ranging from 0.803-0.970 for detecting MCI, and AUCs of 0.98 and 0.99 in detecting dementia due to AD [28], which were comparable with the results found in this study.

Although not all individual assessments in the BrainCheck battery differentiated between NC, MCI, and Dementia, we observed a general trend for each assessment showing that Dementia participants had the lowest scores whereas the NC participants

had the highest scores. Individual assessments that did show significant differences in the scores between NC and MCI groups and between Dementia and MCI groups included the Immediate Recognition and Digit Symbol Substitution. Notably, the BrainCheck Digit Symbol test showed significant differences in performance between all three diagnostic groups, whereas a previous study found that the BrainCheck Digit Symbol test did not show significant differences between cognitively healthy and cognitively impaired groups (n=18, p=0.29) [25]; likely due to this study having a larger sample size. Individual assessments with no significant differences between the MCI group and the NC/Dementia groups were the BrainCheck Stroop and Trails A/B tests (Fig 2 and Table 3). All of these tests include timeout mechanisms if participants are unable to complete the test, and time-out rates were higher in the more cognitively impaired groups (Fig 2). Therefore, when calculating the BrainCheck Overall Score, we have introduced a penalty mechanism for timed-out assessments.

In comparison to comprehensive NPT, which can typically last a few hours and sometimes require multiple visits [30], BrainCheck demonstrated shorter test duration, with median completion times of 17.8 min (*IQR*=15.4-19.6 min) for NC participants and 30.5 min (*IQR*=23.4-37.1 mins) for dementia participants (Fig. 1). Shorter test durations observed in individuals with no/less cognitive impairment suggest that computerized cognitive tests could be useful for rapid early detection in this population, prompting further evaluation, whereas those with dementia have likely already undergone a comprehensive evaluation. The wide variance in completion time for the Dementia group may have uncovered the difficulty participants with more severe cognitive

impairment may have faced in completing the BrainCheck battery, compared to the lower variance observed in the NC group.

A limitation of this study was that participants were not diagnosed by a physician at the time point of BrainCheck testing. Thus, participants were placed into the diagnosis groups based on their most recent clinical diagnosis available in their electronic health record, or for the few NC participants without medical evaluations, based on their report of no cognitive symptoms or diagnosis or cognitive impairment. The period from the most recent clinical diagnosis to the date of BrainCheck testing varied among the diagnosis groups; the Dementia group had the fewest days from their latest clinical evaluation (*median*=82.5 days; *IQR*=44.5-141.25), followed by the MCI group (*median*=244 days, *IQR*=105-346.5) and the NC group (*median*=645 days; *IQR*=225.5-1112.5). These large time intervals in a degenerative population leave room for cognition to worsen over time, potentially blurring the lines in the severity of cognitive impairment, where participants may have progressed to MCI from NC, and to Dementia from MCI during that period. This would make the distinguishing of NC from cognitive impairment more difficult, yet diagnostic accuracy among the groups still remains high. Furthermore, the median number of days since the last clinical evaluation for NC participants was as high as 645 days. This could suggest that the NC participants did not feel an inclination to seek out further cognitive evaluation during the extended time period, and may not have experienced noticeable cognitive decline. Future validity studies should ensure that a physician evaluation and diagnosis occur closer to the time of BrainCheck testing to address these limitations.

When administration type was considered in linear regression analyses, scores only showed significant differences among the three diagnosis groups instead of administration types. While remote administration was not designed into the original study, stay at home orders due to COVID-19 required modifications, and efforts were made to provide preliminary data for remote use. With preliminary outcomes indicating feasibility for remote administration, a more robust study and increased sample size will be needed to fully validate BrainCheck's cognitive assessment via it's remote feature.

Conclusions

The use of computerized cognitive tests is an opportunity to increase test accessibility for an aging population and an increasing presence of cognitive impairment. Remote test options with quick and accurate results could save on time and costs. Technological advances have allowed these tools to obtain greater sensitivity and specificity as well [28]. Our findings in this study demonstrate that BrainCheck can distinguish between three levels of cognitive impairment, NC, MCI, and Dementia. This study paves the way for a comprehensive longitudinal study, exploring BrainCheck in early detection of dementia and monitoring of cognitive symptoms over time, including comparison to 'gold standard' neuropsychological assessments. BrainCheck is automated, capable of being used remotely, and quick to administer. These qualities are critical to increasing accessibility, especially in early detection, for a population facing an expected increase of dementia in upcoming years. Increasing access may lead to growth in the normative

data needed to better characterize dementia subtypes and provide earlier detection and diagnostic utility.

Abbreviations

MCI: Mild Cognitive Impairment

NC: Normal Cognition

MoCA: Montreal Cognitive Assessment

SLUMS: Saint Louis University Mental Status

MMSE: Mini-Mental State Examination

NPT: Neuropsychological Testing

Alzheimer's Disease: AD

ANOVA: Analysis of Variance

ROC: Receiver Operating Characteristic

AUC: Area Under the Curve

Acknowledgments

Funding was provided by BrainCheck, Inc. Registry was provided by the University of Washington Alzheimer's Disease and Research Center along with the Memory and Brain Wellness Center.

Author contributions

Conceptualization: Bin Huang, Reza Hosseini Ghomi

Formal analysis: Siao Ye, Kevin Sun, Bin Huang

Funding acquisition: Reza Hosseini Ghomi

Investigation: Kevin Sun, Huy Phi, Brian Ko

Methodology: Siao Ye, Bin Huang, Reza Hosseini Ghomi

Resources: Kevin Sun, Reza Hosseini Ghomi

Supervision: Bin Huang, Reza Hosseini Ghomi

Writing – original draft: Siao Ye, Kevin Sun, Huy Phi, Brian Ko, Bin Huang

Writing – review & editing: Siao Ye, Kevin Sun, Huy Phi, Brian Ko, Carolyn M. Parsey, Bin Huang, Reza Hosseini Ghomi

References

1. 2020 Alzheimer's disease facts and figures - 2020 - Alzheimer's & Dementia - Wiley Online Library [Internet]. [cited 2020 May 22]. Available from: https://alz-journals.onlinelibrary.wiley.com/doi/10.1002/alz.12068
2. Matthews KA, Xu W, Gaglioti AH, Holt JB, Croft JB, Mack D, et al. Racial and ethnic estimates of Alzheimer's disease and related dementias in the United States (2015-2060) in adults aged ≥65 years. Alzheimers Dement J Alzheimers Assoc. 2019;15(1):17–24.
3. O'Connor D. World Alzheimer Report 2019: Attitudes to dementia. :160.
4. Lang L, Clifford A, Wei L, Zhang D, Leung D, Augustine G, et al. Prevalence and determinants of undetected dementia in the community: a systematic literature review and a meta-analysis. BMJ Open. 2017 Feb 1;7(2):e011146.
5. de Vugt ME, Verhey FRJ. The impact of early dementia diagnosis and intervention on informal caregivers. Prog Neurobiol. 2013 Nov 1;110:54–62.
6. Mittelman MS, Ferris SH, Shulman E, Steinberg G, Levin B. A family intervention to delay nursing home placement of patients with Alzheimer disease. A randomized controlled trial. JAMA. 1996 Dec 4;276(21):1725–31.
7. Morris JC, Storandt M, Miller JP, McKeel DW, Price JL, Rubin EH, et al. Mild Cognitive Impairment Represents Early-Stage Alzheimer Disease. Arch Neurol. 2001 Mar 1;58(3):397–405.

8. Mitchell AJ, Shiri-Feshki M. Rate of progression of mild cognitive impairment to dementia – meta-analysis of 41 robust inception cohort studies. Acta Psychiatr Scand. 2009 Apr 1;119(4):252–65.

9. Albert MS, DeKosky ST, Dickson D, Dubois B, Feldman HH, Fox NC, et al. The diagnosis of mild cognitive impairment due to Alzheimer's disease: Recommendations from the National Institute on Aging-Alzheimer's Association workgroups on diagnostic guidelines for Alzheimer's disease. Alzheimers Dement J Alzheimers Assoc. 2011 May;7(3):270–9.

10. A re‑examination of Montreal Cognitive Assessment (MoCA) cutoff scores - Carson - 2018 - International Journal of Geriatric Psychiatry - Wiley Online Library [Internet]. [cited 2020 Jun 26]. Available from: https://onlinelibrary.wiley.com/doi/abs/10.1002/gps.4756

11. Dautzenberg G, Lijmer J, Beekman A. Diagnostic accuracy of the Montreal Cognitive Assessment (MoCA) for cognitive screening in old age psychiatry: Determining cutoff scores in clinical practice. Avoiding spectrum bias caused by healthy controls. Int J Geriatr Psychiatry. 2020;35(3):261–9.

12. Tsoi KKF, Chan JYC, Hirai HW, Wong SYS, Kwok TCY. Cognitive Tests to Detect Dementia: A Systematic Review and Meta-analysis. JAMA Intern Med. 2015 Sep 1;175(9):1450–8.

13. Benson AD, Slavin MJ, Tran T-T, Petrella JR, Doraiswamy PM. Screening for Early Alzheimer's Disease: Is There Still a Role for the Mini-Mental State Examination? Prim Care Companion J Clin Psychiatry. 2005;7(2):62–9.

14. Szcześniak D, Rymaszewska J. The usfulness of the SLUMS test for diagnosis of mild cognitive impairment and dementia. Psychiatr Pol [Internet]. 2015 [cited 2020 Jun 22]; Available from: http://www.psychiatriapolska.pl/online-first-nr18.html

15. Roebuck-Spencer TM, Glen T, Puente AE, Denney RL, Ruff RM, Hostetter G, et al. Cognitive Screening Tests Versus Comprehensive Neuropsychological Test Batteries: A National Academy of Neuropsychology Education Paper. Arch Clin Neuropsychol. 2017 Jun 1;32(4):491–8.

16. Jacova C, Kertesz A, Blair M, Fisk JD, Feldman HH. Neuropsychological testing and assessment for dementia. Alzheimers Dement. 2007 Oct 1;3(4):299–317.

17. Wild K, Howieson D, Webbe F, Seelye A, Kaye J. Status of computerized cognitive testing in aging: A systematic review. Alzheimers Dement. 2008 Nov 1;4(6):428–37.

18. de Oliveira MO, Brucki SMD. Computerized Neurocognitive Test (CNT) in mild cognitive impairment and Alzheimer's disease. Dement Neuropsychol. 2014;8(2):112–6.

19. Lim YY, Jaeger J, Harrington K, Ashwood T, Ellis KA, Stöffler A, et al. Three-Month Stability of the CogState Brief Battery in Healthy Older Adults, Mild Cognitive Impairment, and Alzheimer's Disease: Results from the Australian Imaging, Biomarkers, and Lifestyle-Rate of Change Substudy (AIBL-ROCS). Arch Clin Neuropsychol. 2013 Jun 1;28(4):320–30.

20. Hammers D, Spurgeon E, Ryan K, Persad C, Barbas N, Heidebrink J, et al. Validity of a Brief Computerized Cognitive Screening Test in Dementia. J Geriatr Psychiatry Neurol. 2012 Jun 1;25(2):89–99.

21. Dwolatzky T, Whitehead V, Doniger GM, Simon ES, Schweiger A, Jaffe D, et al.

Validity of a novel computerized cognitive battery for mild cognitive impairment. BMC Geriatr. 2003 Nov 2;3:4.

22. Zygouris S, Tsolaki M. Computerized Cognitive Testing for Older Adults: A Review. Am J Alzheimers Dis Other Demen. 2014 Feb 13;30.

23. Cucinotta D, Vanelli M. WHO Declares COVID-19 a Pandemic. Acta Bio-Medica Atenei Parm. 2020 19;91(1):157–60.

24. Yang S, Flores B, Magal R, Harris K, Gross J, Ewbank A, et al. Diagnostic accuracy of tablet-based software for the detection of concussion. PLOS ONE. 2017 Jul 7;12(7):e0179352.

25. Groppell S, Soto-Ruiz KM, Flores B, Dawkins W, Smith I, Eagleman DM, et al. A Rapid, Mobile Neurocognitive Screening Test to Aid in Identifying Cognitive Impairment and Dementia (BrainCheck): Cohort Study. JMIR Aging. 2019 Mar 21;2(1):e12615.

26. Using ADRC and Related Resources - Alzheimer's Disease Research Center [Internet]. [cited 2020 Sep 1]. Available from: http://depts.washington.edu/mbwc/adrc/page/research-resources

27. Luo J, Xiong C. DiagTest3Grp: An R Package for Analyzing Diagnostic Tests with Three Ordinal Groups. J Stat Softw. 2012 Oct;51(3):1–24.

28. De Roeck EE, De Deyn PP, Dierckx E, Engelborghs S. Brief cognitive screening instruments for early detection of Alzheimer's disease: a systematic review. Alzheimers Res Ther. 2019 Feb 28;11(1):21.

29. Gauthier S, Reisberg B, Zaudig M, Petersen RC, Ritchie K, Broich K, et al. Mild cognitive impairment. The Lancet. 2006 Apr 15;367(9518):1262–70.

30. Neuropsychological Testing - an overview | ScienceDirect Topics [Internet]. [cited 2020 Oct 21]. Available from: https://www.sciencedirect.com/topics/medicine-and-dentistry/neuropsychological-testing