Full title: Heading for personalized rTMS in tinnitus: reliability of
individualized stimulation protocols in behavioral and electrophysiological
responses

Short title: Reliability of rTMS effects in tinnitus

Stefan Schoisswohl[1*], Berthold Langguth[1], Tobias Hebel[1], Mohamed A. Abdelnaim[1], Gregor
Volberg[2], Martin Schecklmann[1]

[1]Department of Psychiatry and Psychotherapy, University of Regensburg, Regensburg, Germany
[2]Institute of Psychology, University of Regensburg, Regensburg, Germany

* Correspondence:
Corresponding Author
Stefan Schoisswohl, University of Regensburg, Department of Psychiatry and Psychotherapy,
Universitaetsstraße 84, 93053 Regensburg, Germany. Tel.: +49-941-941-2054
E-mail: stefanschoisswohl@yahoo.de

Abstract

Background: Repetitive transcranial magnetic stimulation (rTMS) is a non-invasive brain stimulation tool potentially modulating pathological brain activity. Its clinical effectiveness is hampered by varying results and characterized by inter-individual variability in treatment responses. RTMS individualization might constitute a useful strategy to overcome this variability. A precondition for this approach would be that repeatedly applied protocols result in reliable effects. The condition tinnitus provides the advantage of immediate behavioral consequences (tinnitus loudness changes) after interventions and thus offers an excellent model to exemplify TMS personalization. **Objective**: The aim was to investigate the test-retest reliability of short rTMS stimulations in modifying tinnitus loudness and oscillatory brain activity as well as to examine the feasibility of rTMS individualization in tinnitus. **Methods:** Three short verum (1Hz, 10Hz, 20Hz; 200 pulses) and one sham (0.1Hz; 20 pulses) rTMS protocol were administered on two different days in 22 tinnitus patients. Before and after each protocol, oscillatory brain activity was recorded with electroencephalography (EEG) together with behavioral tinnitus loudness ratings. RTMS individualization was executed by dint of behavioral and electrophysiological responses. Stimulation responders were identified via consistent sham-superior increases in tinnitus loudness (behavioral responders) and alpha power increases or gamma power decreases (alpha responders/ gamma responders) in accordance to the prevalent neurophysiological models for tinnitus. **Results:** It was feasible to identify individualized rTMS protocols featuring reliable tinnitus loudness changes (55% behavioral responder), alpha increases (91% alpha responder) and gamma decreases (100% gamma responder) respectively. Alpha responses primary occurred over parieto-occipital areas, whereas gamma responses mainly appeared over frontal regions. Contrary, test-retest correlation analyses per protocol on a group-level were not significant neither for behavioral nor for electrophysiological effects. No associations between behavioral and EEG responses were given. **Conclusion:** RTMS individualization via behavioral and electrophysiological data in tinnitus can be considered as a feasible approach to overcome low reliability on group-level. The present results open the discussion favoring personalization utilizing neurophysiological markers rather than behavioral responses. These insights are not only useful for the rTMS treatment of tinnitus but also for neuromodulation interventions in other pathologies as our results suggest that the individualization of stimulation protocols is feasible despite absent group-level reliability.

Keywords: repetitive transcranial magnetic stimulation, tinnitus, neuronavigation, electroencephalography, reliability, rTMS personalization

1. Introduction

With a prevalence of about 10% of the population, tinnitus is a rather common condition. It is characterized by the perception of ringing or hissing in the absence of a corresponding external stimulus. The majority of tinnitus cases are caused by cochlear damage leading to hearing loss [1–3], whereby the absence of auditory input from the periphery provokes maladaptive neural changes which ultimately manifest in hyperactivity of auditory and non-auditory cortical regions. These pathological alterations of the central nervous system are assumed to underly the perception of tinnitus [4–8].

Recent neuromodulation techniques attempt to reverse or modulate these pathological alterations (for a recent overview see Langguth et al. (2020) [9]. One of these is repetitive transcranial magnetic stimulation (rTMS), a non-invasive brain stimulation tool with the capability to evoke neuroplastic changes by a rhythmic administration of brief electromagnetic pulses [10–12]. By means of low frequency rTMS it is possible to inhibit cortical excitability whereas high frequency protocols are assumed to increase cortical excitability. These statements should be seen as a rule of thumb as several findings show high inter- and intra-individual variability of rTMS with a lot of subjects not behaving accordingly [13–16].

On this account, temporal 1Hz rTMS was applied in order to inhibit tinnitus-associated hyperactivity of the auditory cortex and consequently suppress the tinnitus percept [17,18]. For many years various treatment trials deployed left temporal low frequency rTMS showing greatly varying findings and high inter-individual variability in responses (for an overview see Schoisswohl et al., 2019 [19]). Given this backdrop, additional studies also examined the consequences of other rTMS approaches such as multi-site stimulation protocols [20,21], prefrontal cortex stimulation [22–24] or high frequency rTMS [25–27] as well as continuous theta-burst stimulations [25,28,29] as a treatment option for tinnitus. Nevertheless, its clinical effectiveness is

still not obvious. Recent reviews and meta-analyses conclude that there is no clear indication for

an effect [9,30,31], while others report potential effectiveness for its clinical administration [32,33],

especially over the auditory cortex [34]. This lack of evidence may be explained by the fact that

there is still a certain ambiguity concerning the correct rTMS parameters (e.g., frequency or

position) for the application in tinnitus. Admittedly, the effects of non-invasive brain stimulation

methods in general are subject to rigorous intra- and inter-individual variability, conceivably

caused by a complex interplay of technical and physiological parameters [35–38].

A personalization of stimulation protocols might provide remedy and further enhance efficacy.

Individualization of magnetic stimulation might be capable of resolving the issue of effect

variability within the same type of pathology. Its implementation in neuropsychiatric research is

highly dependent on phenotypes as well as the identification of appropriate and valid clinical or

physiological outcome measures [39,40]. However, tinnitus provides the tremendous benefit of

direct reactions from the patients' side by virtue of changes in tinnitus perception and can therefore

function as a kind of role model in rTMS personalization.

Test sessions (i.e., short stimulations with recording of immediate responses) and an administration

of different protocols offer the opportunity to examine the ability of the different protocols to

suppress tinnitus loudness in an individual. In several studies it has been demonstrated that tinnitus

loudness can be suppressed by means of single rTMS sessions for some minutes [41–45], and that

suppression patterns vary among patients [46]. The concept of rTMS personalization in the clinical

context of tinnitus was first introduced by Kreuzer et al. (2017) [47]. By dint of rTMS test sessions

with different types of frequencies and positions, individualized protocols for brief tinnitus

suppression were identified and further used in the course of a daily treatment. These findings

emphasize the feasibility of rTMS individualization and imply superiority over a standard treatment

protocol by means of a higher number of treatment responders, even if no statistical differences

between the standard and the individualized treatment were detected in this relatively small pilot

study. In another preliminary work from our group using an e-field guided system we were able to

demonstrate reliable tinnitus loudness decreases for certain stimulation parameters as well as the

possibility to individualize rTMS characterized by a consistent and sham-controlled tinnitus

loudness suppression in five out of five individuals [48]. Nonetheless, behavioral responses are

governed by certain limitations such as their highly subjective matter and the difficulty to

accurately evaluate minor changes in the tinnitus percept in many cases, which might constitute an

additional reason for variability in responses.

It is widely established that tinnitus is accompanied by pathologically altered oscillatory brain

activity patterns. According to the "Thalamo-cortical Dysrhythmia Model" (TCD), the assumed

mechanism behind these changes are peripheral deafferentation causing thalamic low frequency

activity which in turn provokes a boost in neural synchrony and increased gamma activity in

auditory regions of the cortex [49–51]. An expansion of the TCD model termed "Synchronization-

by-Loss-of-Inhibition-Model" (SLIM), presumes that an ancillary suppression of neurons relevant

for inhibition additionally causes diminished activity in the alpha frequency range. Moreover

pathological gamma in tinnitus is prompted by this alpha decrease and the concomitant loss of

inhibition [52].

Several neurophysiological investigations using electro- or magnetoencephalographic

measurements revealed decreased activity in the alpha band and increased activity in the delta and

gamma frequency ranges [53–58]. In particular, gamma is suggested to be closely related to the

actual tinnitus perception [57,59,60]. These altered spontaneous brain activities might provide a

potential indicator to examine the effect of rTMS in tinnitus.

TMS-EEG investigations have already demonstrated that rTMS successfully modulates evoked

and ongoing brain activity with effects exceeding the stimulation offset (neuroplastic

consequences) [61–63]. In the clinical context, TMS-EEG approaches are already used for several

neuropsychiatric conditions such as schizophrenia, depression or Alzheimer's disease in order to

identify and investigate the relevance of neurophysiological markers as well as to provide a more

profound insight into the neuroplastic consequences of rTMS [64,65]. In case of tinnitus, a

combination of rTMS and EEG might not only help to improve our understanding of the

pathophysiology and altered ongoing brain activity patterns behind tinnitus but also shed light on

neurophysiological markers for tinnitus suppression. As of yet, only a handful of studies

investigated the consequences of single rTMS sessions on spontaneous brain activity in tinnitus.

Müller et al. (2013) [46] applied different stimulation types (1 Hz, continuous theta burst,

intermittent theta burst and individual alpha frequency rTMS) over the temporal cortex and aimed

to detect the most efficient protocol for transient tinnitus suppression per subject. By using

magnetencephalography (MEG) they could detect a significant increase of alpha power (8 – 12 Hz)

in the auditory cortex associated with tinnitus suppression. Due to the absence of clear effects on a

whole-group level, the authors emphasize the relevance of analysis on an individual subject level

to unveil the key mechanisms behind tinnitus suppression. In this context especially modulations

in the alpha frequency band might provide a potential marker for future attempts in individualizing

rTMS. Schecklmann et al. [66] compared the neuroplastic consequences of rTMS in tinnitus

patients with healthy controls and observed rTMS-specific modulations almost entirely for the

group of tinnitus patients. A left temporal 1 Hz stimulation with 200 pulses decreased power spectra

in the delta (2 – 3.5 Hz) and theta (4 – 7.5 Hz) frequency ranges as well as increased beta2 (18.5 –

21 Hz) frequency band power over frontal areas in tinnitus patients. Right prefrontal cortex 1 Hz

stimulation reduced beta3 (21.5 – 30 Hz) and gamma (30.5 – 44 Hz) band activity in tinnitus

patients in temporal areas. In contrast the control group exhibited enhanced beta3 and gamma

power. Results emphasize the ability to modulate tinnitus-related ongoing brain activity or rather

to induce neuroplastic changes in tinnitus by rTMS in accordance with the TCD model. However,

on a behavioral level reliable sham-controlled decreases in tinnitus loudness were solely present in

two out of 20 tinnitus patients (one after right prefrontal, one after left temporal rTMS).

Considering past findings, in particular the ability of certain rTMS protocols to transiently decrease

tinnitus loudness plus modulate tinnitus-associated oscillatory brain activity, up to now there is no

convincing data on test-retest reliability of the effects of single session rTMS. A test-retest design

has the potential to yield essential details about effect-consistency, more specifically if rTMS

produces reliable and valid modulations. Findings may promote decisions in choosing appropriate

rTMS protocols for daily treatments. Due to the overall high variability of non-invasive brain

stimulation effects, it is of utmost importance to identify reliable brain stimulation techniques to

cope for this variability. Hence, one of the objectives of the current experiment was to scrutinize

the test-retest reliability of different rTMS protocols in short-term tinnitus suppression and

modifying resting state EEG activity. For this purpose the study design used in Schoisswohl et al.

(2020) [48] with pre-post rTMS tinnitus loudness ratings was expanded by electrophysiological

measures before and after each rTMS. Furthermore, it was reduced by the factor stimulation

position since the majority of patients (three out of five) responded to a stimulation over temporo-

parietal regions. An e-field guided stimulation over the left and right temporo-parietal junction

allows for a more precise investigation of the parameter frequency, since past studies show

inconsistent results in this regard.

Beyond that, we wanted to undertake further research on the personalization of rTMS by means of

the model tinnitus. The aim was to assess the feasibility of identifying an rTMS protocol for

consistent and sham-superior brief tinnitus suppression per patient (behavioral response). In light

of previous work from Müller et al. (2013) [46] together with the current prevalent

neurophysiological models in tinnitus, we wanted to go one step beyond and strive for a rTMS

personalization based on electrophysiological responses. The specific objective was to ascertain a

protocol per patient, which is able to produce consistent and sham-superior increases in alpha or

decreases in the gamma frequency band (EEG response) respectively. Based on past research, we

hypothesize that for the majority of subjects it is feasible to identify a personalized rTMS protocol

by means of behavioral or EEG responses.

Additionally, we were interested if an EEG response in the alpha or gamma band is related to a

behavioral response and vice versa.

2. Material and Methods

2.1 Participants

N = 22 patients (5 female) with chronic subjective tinnitus, recruited from the Interdisciplinary

Tinnitus Centre in Regensburg, Germany, participated in this experiment. Participants had to

comply to the following inclusion criteria: age between 18 and 75 years; German-speaking; no or

stable treatment with psychoactive drugs; absence of severe somatic, neurological or psychiatric

conditions (e.g., acute psychosis, severe depression, alcohol and/ or substance abuse or known

brain tumor); no simultaneous participation in other tinnitus-related experiments or treatments.

Further, participants had to exhibit no contraindications with respect to TMS (e.g., epilepsy or state

after traumatic brain injuries; cf. [67,68]) as well as magnetic resonance imaging (MRI) (e.g., heart

pacemaker, vascular clamp, implanted insulin pump or severe claustrophobia). Each participant

received detailed clarification about study aim, procedure and applied methods as well as potential

adverse effects related to TMS and gave written informed consent before study start. The study

was approved by the ethics committee of the University of Regensburg, Germany (ethical approval

number: 17-820-101).

2.2 Test session procedure

Ahead of the rTMS tests sessions, participants' hearing thresholds for the frequency range of 125 Hz to 8 kHz were determined by the use of a clinical audiometer (Madsen Midimate, 622D; GN Otometrics, Denmark). In addition, structural MRI brain scans (T1) were undertaken with a MAGNETOM 1.5 Tesla scanner (Siemens, Germany) for neuronavigated rTMS and participants were requested to complete German versions of the Tinnitus Handicap Inventory (THI; [69,70]) the Tinnitus Sample Case History Questionnaire (TSCHQ; [71]) and the European School for Interdisciplinary Tinnitus Research Screening Questionnaire (ESIT-SQ; [72]).

The actual rTMS test sessions lasted approximately three hours and took place on two different days with a maximum inter-session interval of two days, whereby daytime was always the same ($\pm$ 1 hour). Both test session days followed the same procedure. On each day, subjects were stimulated with four different rTMS protocols over the left temporo-parietal junction (TPJ) and with the same four protocols over the right TPJ (see section repetitive transcranial magnetic stimulation). Thus, a total number of 8 different rTMS protocols was applied on each test session day. Magnetic stimulation protocols were administered in a randomized sequence on each day with the exception that the order of hemisphere was inverted (if the stimulation sequence on the first day started over the left hemisphere, it began with a stimulation over the right hemisphere on the second day). During three minutes before and after each stimulation patients were requested to focus on their tinnitus and verbally rate the current loudness of their tinnitus perception on a numeric rating scale ranging from 0% to 110% at seven different points in time with an inter-rating interval of 30 seconds (T0, T30, T60, T90, T120, T150 and T180). The numeric rating scale was anchored at 0% (absence of tinnitus) and at 110% (tinnitus loudness increase of 10%), whereby 100% corresponds to the patients usual perceived tinnitus loudness. Concurrent to these ratings, resting state oscillatory brain activity was recorded for three minutes each by means of EEG. For this purpose,

participants were instructed to sit calmly, focus on a certain point in the room and avoid

unnecessary eye blinks or muscle movements. Subsequent to each post stimulation loudness rating/

EEG recording, participants had to evaluate the level of discomfort caused by the applied rTMS

protocol on a numeric rating scale from 0 to 10, whereas 10 corresponds to utmost discomfort or

rather intolerability. **Figure 1** provides an illustration of the entire test session proceeding.

2.3. Repetitive transcranial magnetic stimulation (rTMS)

An e-field guided rTMS system (NBT System 2; Nexstim Plc., Finland) together with individual

structural MRI scans (T1) was utilized for the test sessions. Prior to the first session, participants'

resting motor threshold (RMT) was determined according to the same procedure as already

described in Schoisswohl et al. (2020) [48]. Thereby, various positions over the left primary motor

cortex were stimulated until several motor evoked potentials (MEP) derived from the thenar

muscles of the right hand with a peak-to-peak amplitude of $> 50\ \mu V$ were visible. The stimulation

position with the highest MEP amplitude was repeated via the neuronavigation system and single

pulses with intensity shifts were applied. The lowest stimulation intensity which produced MEPs

with an amplitude of at least $50\ \mu V$ in half of the administered pulses was defined as subjects'

RMT. Additionally, the RMT determination was repeated with a mounted EEG cap using the same

motor hotspot. Magnetic stimulation consisted of 1 Hz, 10 Hz and 20 Hz protocols with 200 pulses

each at 110% RMT, whereas a series of 20 pulses at 0.1 Hz was deployed as an active control

condition (sham) with no assumed neuroplastic consequences [73,74]. Each protocol was

administered over the left and right TPJ, resulting in a total number of eight different rTMS

protocols with approximately six minutes of inter-protocol intervals (3 minutes of pre and post

rTMS EEG recordings; if applicable – a waiting period until subject´s tinnitus loudness returned to

the initial level). An uncooled coil was used for the test sessions, as the cooled coil produces

additional noise and we aimed at minimizing the stimulation noise, as it could confound both

behavioral and neurophysiological effects. Stimulation positions were localized according to the

electrode positions CP5 and CP6 (10-20 System). By virtue of the application of an e-field guided

neuronavigation system and the opportunity for online visualization of strength and direction of

the TMS-induced e-field on subjects' anatomical scans, each stimulation was administered with

the induced e-field perpendicular to the sulcus of the brain area of interest. In order to ensure that

the stimulation positions respectively the induced e-field remained the same over the two test

sessions, both electrode positions were tagged with a digitization pen and marked on the individual

anatomical images. Next, the coil was placed accordingly, a single pulse with an intensity of 10%

RMT was given and the corresponding position of the coil was saved in the neuronavigation

system. With the aid of a system-intern aiming tool, the exact position of the coil could be repeated

with respect to centering, rotation and tilting. After each protocol the position was verified and

adjusted if necessary. For safety reasons all participants were wearing in-ear plugs during the whole

stimulation procedure.

2.4.Electroencephalography

2.4.1. Data acquisition and preprocessing

Participants' electrophysiological resting state brain activity was recorded for three minutes before

and after each of the eight rTMS protocols by means of EEG with a BrainAmp DC system (Brain

Products GmbH, Germany) in combination with an Easycap elastic electrode cap (Easycap GmbH,

Germany) consisting of 64 electrodes placed according to the 10-20 System (GND: AFz) and the

software Brain Vision Recorder 1.20 (Brain Products GmbH, Germany). EEG measurements were

online referenced to FCz and recorded at a sampling rate of 500 Hz. Impedances were kept lower

than 10 kΩ during the whole recording.

Preprocessing of raw EEG data was conducted with a custom-built pipeline in Matlab (Matlab

R2018b; Mathworks, USA) by employing functions implemented in the EEGLAB toolbox [75]. In

a first step, EEG data was filtered between 1 Hz and 45 Hz and segmented into 2 second epochs.

Afterwards the data was visually inspected and single epochs featuring artifacts or verbal tinnitus

loudness ratings were removed from the signal. For each dataset the first and last two epochs were

rejected as well, and noisy or deviant sensors were identified for subsequent interpolation. In order

to identify and reject vertical and horizontal eye movements, an independent component analysis

(ICA; fastICA toolbox http://research.ics.aalto.fi/ica/fastica/) was executed. In a next step, datasets

were re-referenced using an average reference and the online reference electrode FCz was

reconstructed and added as a data channel. Previously identified channels with high signal-to-noise

ratio were interpolated using the spherical spline method [76]. As a last step, preprocessed datasets

underwent a final visual inspection.

2.4.2. Power spectra

Power spectra analysis was performed with the Fieldtrip toolbox [77] in Matlab. Fast Fourier

Transformation (FFT, "mtmfft") and a hanning window with 1 Hz spectral smoothing was used for

the calculation of pre and post stimulation power spectra for each channel per rTMS protocol.

Obtained power spectra underwent a normalization through a division of the spectral power of

every single frequency by the average spectral power of the whole frequency spectrum. Grand

averages were calculated for every pre and post stimulation dataset per channel and the following

frequency bands: delta (2 – 3.5 Hz), theta (4 – 7.5 Hz), alpha (8 – 12.5 Hz), beta (13 – 32.5 Hz)

and gamma (33 – 44 Hz) [66,78]. RTMS protocol specific pre-to-post power changes in the

respective frequency bands were obtained by subtracting the pre stimulation from the post

stimulation power for each channel (Δ power = post – pre). Channels Iz, TP9 and TP10 which

usually contain noisy signal were excluded, resulting in 60 channels for further statistical analysis.

2.5.Statistical analysis

All subsequent statistical analyses were performed with the statistics software R (R version 4.0.3; R Foundation for Statistical Analysis, Austria) using the packages "psych" and "ggstatsplot". Assumptions for parametric testing were examined with Shapiro-Wilk-Tests (normal distribution) and F-Tests (sphericity) and if violated non-parametric tests were performed. Significance levels were set to $\leq .05$ for all analysis, whereby $< .10$ was considered as a statistical trend. Differences between left and right hearing level, RMT as well as induced e-field strength with and without mounted EEG cap were analyzed with paired sample t-tests and Wilcoxon signed-rank tests respectively.

2.5.1. Reliability analysis of TMS-induced tinnitus loudness changes and discomfort

Pre stimulation ratings were averaged (T0 – T180) to receive a stable tinnitus loudness baseline for each protocol per subject. TMS-induced tinnitus loudness changes were calculated via a subtraction of the mean baseline loudness level from the mean post stimulation loudness level (Δ loudness = post – baseline). Test-retest reliability of TMS-induced mean tinnitus suppression as well as discomfort evaluation over both test session days was analyzed by means of Spearman correlations for each protocol. P-values were adjusted for multiple comparisons using the Holm-Bonferroni method [79].

2.5.2. Reliability analysis of TMS-induced changes in oscillatory brain activity

Test-retest reliability of protocol-specific pre-post power spectra changes were analyzed with Spearman correlations. Therefore, the obtained changes were correlated between the first and second test session on a single sensor level per rTMS protocol and frequency band. Thus, 300 correlations were calculated per rTMS protocol (5 frequency bands x 60 sensors). Obtained p-

values of sensor correlations were adjusted according to the number of sensors separated per

frequency band and protocol using the Holm-Bonferroni method [79]. In case of significant results,

p-values were further corrected for the number of applied protocols akin to an adjustment for

multiple comparisons of behavioral results. We were exclusively interested in positive correlation

outcomes, representing reliability of TMS effects. Significant positive correlations of channel pairs

were identified via correlation matrices and projected on a 60-channel topographical scalp map

applying a color-coding for received correlation coefficients (r_s). A minimum number of two

neighboring electrodes featuring significant positive correlations (electrode cluster) was considered

as a reliable TMS-induced change in power spectra. Triangulation of 2d sensors was utilized for

neighboring sensor identification.

2.5.3. rTMS individualization via responder identification using behavioral and

electrophysiological data

Additional to test-retest reliability analysis, rTMS individualization was examined based on single

subject responses concerning loudness evaluation and electrophysiological consequences.

Behavioral response to one of the verum interventions was defined as a mean tinnitus loudness

suppression (compare section Reliability analysis of TMS-induced tinnitus loudness changes and

discomfort evaluation) superior to sham stimulation (Δ loudness verum > Δ loudness sham) on

both test session days for the same rTMS protocol and stimulated hemisphere. EEG responders

were defined for the alpha (8 – 12.5 Hz) and gamma (33 – 44 Hz) frequency bands separately. In

particular we were interested in alpha power increases and gamma power decreases from pre to

post stimulation. As indicated in the introduction, alpha increases and gamma decreases are the

most valid biomarkers for tinnitus reductions as indicated by literature review and expert

knowledge. For each subject and protocol, the top 20% of channels exhibiting alpha increases/

gamma decreases were detected and scrutinized with regards to sham-superiority per day.

If a sham-superior alpha increase respectively gamma decrease was observed in the same sensor

on both days plus the corresponding pattern occurred in a minimum number of two adjacent

channels, the subject was designated as either alpha or gamma responder. Associations of a

behavioral response to any of the eight protocols with an electrophysiological response in the alpha

or gamma frequency band and vice versa were examined with Fisher´s exact tests due to cell

frequencies less than five.

3. Results

3.1. Sample characteristics

Participants were aged from 43 to 69 years (M = 57.05, SD = 6.79) and had an average tinnitus

duration of 131.64 months (SD = 116.76). The majority reported a bilateral tinnitus perception as

well as loudness fluctuations. No difference in hearing loss between the left and the right ear was

observed, $t_{(13)}$ = -1.36, p = .198. THI severity grades ranged from mild (grade 2) to catastrophic

(grade 5) and manifested on average in severe severity (grade 4) for the whole sample (M = 58.49,

SD = 19.55). As expected, the RMTs were significantly higher with a mounted EEG cap, p < .001.

Unexpectedly, the corresponding strength of the induced e-field was similarly higher with an EEG-

cap, $t_{(21)}$ = -6.07, p < .001. Although the motor hotspot for RMT determination was identical as

well as coil centering and rotation, differences in the e-field might derive from difficulties in the

adjustment of coil tilting due to distances to the scalp as well as the concomitant stimulation at

higher intensities.

Information about participant characteristics can be seen from **table 1**. Besides anticipated side

effects such as discomfort during rTMS (in particular by rapid muscle contraction caused by high

frequency protocols) or temporary tinnitus loudness increases after stimulation, no additional side

effects could be observed in the current sample. Spearman correlations showed significant positive

correlations for discomfort evaluations of each applied rTMS protocol between the two test

sessions (cf. **figure S1** in the supplemental material).

3.2. Test-retest reliability of TMS-induced changes in tinnitus loudness and oscillatory brain

activity

No significant correlations were observed for the two test session days with regards to mean

tinnitus loudness changes for any of the protocols (cf. **figure 2**). It should be noted, that without

considering corrections for multiple comparisons, a significant test-retest reliability was found for

left hemispheric stimulation with 10 Hz and right hemispheric stimulation with 20 Hz.

On the basis of our predefined reliability criteria, reliable changes in oscillatory brain activity were

solely present for 1 Hz stimulation of the right TPJ, exclusively for the gamma frequency band. As

can be seen from **figure 3**, these reliable changes were primarily observed over parieto-occipital

regions, whereby neighboring channels PO7 and O1 constitute a significant reliable cluster. If

corrected for multiple comparisons, this effect disappears and is therefore only conditionally

interpretable.

3.3 rTMS individualization using behavioral and electrophysiological data

In 12 out of 22 (55%) participants it was possible to identify an individualized rTMS protocol for

short-term tinnitus suppression characterized by sham-superiority on both test session days. From

those 12 patients 10 (83.33%) responded to more than one stimulation type. Detailed information

about individual responses together with the induced mean tinnitus suppression (averaged over

both test session days) can be found in **table 2**. As presented in **table 3** the majority of subjects

responded to a stimulation over the left TPJ with 20 Hz (n = 7) and 10 Hz (n = 6), and over the

right TPJ with 20 Hz (n = 5).

EEG responders were defined by sham-superior alpha power increase or gamma power decrease

in minimal two neighboring electrodes on both test sessions. Using this criterion 20 alpha

responders (90.91%) were identified, showing alpha power increases for at least one of the six used

verum protocols. Left TPJ stimulation with 10 Hz (n = 10) or 1 Hz (n = 9) and right TPJ stimulation

with 10 Hz (n = 7) produced the most alpha responders (cf. **table 3**). Electrophysiological responses

in the gamma band were present in each investigated subject (100%) for at least one protocol.

Subjects responded mainly to left TPJ stimulation with 10 Hz (n = 13) or 1 Hz (n = 10) and right

TPJ stimulation with 10 Hz (n = 8) or 1 Hz (n = 8) (cf. **table 2**). Interestingly, 75% of alpha and

gamma responders showed consistent power increases/decreases for more than one protocol. **Table**

3 presents an overview of alpha increases and gamma decreases per rTMS protocol on a single

subject level. Fisher´s exact tests showed no significant association neither for behavioral

responders with alpha or gamma responders nor the other way around.

Electrodes featuring sham-superior alpha/gamma modifications on both days are presented in

figure 4 and **figure 5** by means of topographical maps displaying the quantity of electrodes within

alpha and gamma responders per rTMS protocol. It is apparent from **figure 4** that alpha responses

primary appeared in parieto-occipital regions on both hemispheres, whereas responses in the

gamma range were especially present over frontal regions bilaterally and to some extent over

parieto-occipital parts as outlined in **figure 5**. Detailed electrode information per subject and rTMS

protocol can be found in the supplementary material for alpha (**table S1**) and gamma responder

(**table S2**).

4. Discussion

The goals of this experiment were to ascertain, whether it is possible to identify treatment protocols that provide reliable effects both on the behavioral and on the electrophysiological level, as this is a precondition for individualizing treatment. We performed two test sessions at which six active and two sham protocols were administered. We evaluated reliability by comparing tinnitus suppression effects and electrophysiological effects on an individual level by responder identification. Moreover, we performed whole-group analysis in which we correlated the behavioral and electrophysiological effects of all stimulation protocols between both test sessions. The main finding of our study was, that in 55% of patients we could identify a stimulation protocol that reduced tinnitus loudness transiently in both sessions reliably more than sham stimulation. When we defined response by the effects on oscillatory brain activity, the number got even higher with 20 out of 22 patients in which a specific rTMS protocol increased alpha more than sham and with 22 out of 22 in which a specific rTMS protocol decreased gamma more than sham. These observations are in accordance with our a priori assumptions of rTMS personalization in the majority of subjects as well as findings from previous studies.

Correlation analysis for the whole group revealed reliable changes in tinnitus loudness after left TPJ stimulation with 10 Hz and right TPJ stimulation with 20 Hz. With respect to modulatory effects of rTMS on oscillatory brain activity, reliable changes could be obtained after right TPJ stimulation with 1 Hz in the gamma frequency band over parieto-occipital regions. However, these findings do not withstand a correction for multiple comparison and should therefore not be overrated and interpreted only exploratively. Nevertheless, the observed effect of rTMS on gamma activity fits well with a previous TMS-EEG study where right frontal 1 Hz rTMS decreased gamma band activity over right temporal regions in tinnitus patients, whereas it increased gamma activity in healthy controls [66]. Gamma activity is supposed to be strongly connected with tinnitus

perception [57,59,60]. However, as mentioned above, the effects of the first and the second session

were not significantly correlated when statistical corrections for multiple comparisons were

performed.

Taking our findings together we can conclude that it is possible to identify responders for specific

rTMS protocols, who demonstrate a reliable behavioral or electrophysiologic response on a specific

rTMS protocol. However, if we investigate the effects of a given protocol for the whole sample,

the variability of the effect is relatively large from session to session.

Presumably the effects depend on the brain state before or during rTMS [80] and this brain state

might vary from session to session. In the current study the experimental design with six active and

two sham protocols per session might have increased this variability, as after-effects from

preceding protocols cannot be excluded as confounding factors.

The limited test-retest reliability corroborates the claim that effects of rTMS in general [37,38,81–

83] but also treatment effects in tinnitus [9,84] are subject to high variability. Thus, it is most likely

that single subject impacts simply do not transfer to statistical analysis on a group level,

highlighting the need and importance of personalized rTMS approaches.

On the other hand, it has been demonstrated that patients, who benefit once from rTMS have a high

chance to benefit from repeated rTMS treatment [85,86].

This clinical observation fits with our finding to identify treatment responders to specific rTMS

protocols.

Our results corroborate also findings from several previous studies in which the feasibility of a

customization of rTMS for tinnitus via test sessions has been demonstrated [47,48]. The responder

rate of 55% in our study (12 out of 22) is in a similar range as in two previous studies (Kreuzer et

al. (2017) - 50% responder [47]; Schoisswohl et al. (2020) - 100% responder [48]). However, both

studies deployed magnetic stimulations over either prefrontal cortical regions or different positions

over the superior temporal gyrus, whereas the current study targeted only a single region of the

temporal cortex. The present goal was to examine the parameter frequency in more detail.

Interestingly the vast majority of participants responded to high frequency stimulations over the

left TPJ. It is assumed that high frequency protocols produce increases of cortical excitability [15].

Consequently, high frequency rTMS is most likely not able to reduce tinnitus-associated temporal

hyperactivity. But it has been suggested, that the induced effects can vary or shift across different

targets of the cortex [43] and are dependent on the intrinsic state of the brain region prior to

stimulation [80,87–89]. Based on this it could be the case that pathological auditory cortex

hyperactivity as an intrinsic state of the "tinnitus brain" might cause a reversal of supposed high

frequency effects (shift from excitatory to inhibitory). In accordance with the present results,

previous studies have demonstrated responder rates between 40% and 57% in single session

experiments with high frequency stimulation protocols [41,42,45,90]. Likewise treatment studies

with high frequency rTMS over the temporo-parietal cortex were able to show significant

reductions in tinnitus distress [26,27].

In addition to an rTMS individualization based on behavioral responses, the current investigation

has striven for a novel attempt in rTMS individualization with objective measures. By the use of

TMS-EEG combinatory measurements, we wanted to overcome the limitations of subjective

tinnitus loudness ratings and identify personalized rTMS protocols via sham-superior increases in

alpha respectively decreases in gamma frequency band power in at least two neighboring EEG

electrodes on both test session. The utilization of especially these frequency bands as an

electrophysiological response indicator is based on current neurophysiological models. In tinnitus

pathological diminished alpha band activity was observed in contrast to healthy controls [57,58],

whereby during states of brief tinnitus suppression either after sound stimulation [91] or after rTMS

[46] increases in alpha power were detected. While gamma band activity is assumed to underly the

actual tinnitus perception [57,59] by associations with tinnitus loudness [54,60]. Interestingly, we

were able to identify a substantially higher number of EEG responders than behavioral responders.

For a total number of 20 participants, it was possible to detect an alpha response for at least one

rTMS protocol, whereas the entire sample featured responses in the gamma frequency band.

Noteworthy, the quantity of alpha or gamma responders was almost twice as much as for behavioral

responders. In line with our behavioral observations, a left 10 Hz stimulation resulted in the highest

quantity of alpha and gamma responders. It has been shown that this type of stimulation can

modulate alpha oscillations [92] and that alpha and gamma oscillations show a strong interplay not

only in tinnitus [52,93].

Since neurophysiological investigations were able to demonstrate relationships of the alpha and

gamma band with tinnitus loudness [46,54,60,91], we addressed the question if a response in the

alpha or gamma band is associated with a behavioral response and vice versa. Our analysis did not

reveal any relationship of EEG and behavioral responders, hampering the interpretability of our

alpha and gamma findings as relevant for tinnitus loudness decreases in the current sample. A

potential explanation for this lack of association could derive from the possibility, that rTMS-

specific modulations/ neuroplastic consequences as measured with EEG simply do not instantly

translate into behavioral responses. Consequently, longer periods of stimulation with an

individualized rTMS protocol based on patients' EEG response (protocol with strongest alpha

increase/ gamma decrease) for example within the course of a daily treatment, might accumulate

over time into behavioral effects and could tackle the deficiency of effectiveness in current rTMS

treatment trials for tinnitus.

Future studies should strive for a systematic implementation of test session approaches and

investigate the therapeutic consequences of personalized rTMS in more detail. The present

experiment underscores the feasibility of rTMS individualization via behavioral or

electrophysiological responses. The results and methodology described in this sample of tinnitus

patients might encourage neuromodulation attempts in other pathologies to personalize rTMS in

order to account for inter-subject variability in rTMS response.

5. Conclusion

The aim of this investigation was to examine the feasibility of rTMS individualization in tinnitus.

This involved the identification of responders to specific rTMS protocols and the assessment of

test-retest reliability of rTMS effects in tinnitus suppression and ongoing brain activity. We were

able to demonstrate the feasibility of rTMS individualization by using test sessions with different

rTMS protocols. Responses to specific protocols based on electrophysiological signatures could be

identified in all patients, responses based on behavioral effects in the majority of patients. In

contrast test-retest reliability as assessed with correlation analyses for the various rTMS protocols

were rather low both for behavioral and electrophysiological effects.

Taken together these findings are highly encouraging for efforts to enhance the efficacy of rTMS

by personalizing stimulation protocols.

6. Conflict of Interest

The authors declare that they have no conflict of interest associated with this publication and there

has been no significant financial support that could have influenced the outcomes. Author TH

received a one-time travel cost coverage by Nexstim Plc. for an oral presentation.

7. Author Contributions

MS, BL and SS designed the study. MS and SS carried out the experiment. TH and MA recruited

eligible participants and conducted the magnetic resonance imaging. SS analyzed the data and

wrote the main manuscript under the supervision of MS. Authors BL, TH, MA and GV contributed

to and reviewed the manuscript.

8. Funding

The project was conducted as part of the European School for Interdisciplinary Tinnitus Research

(ESIT) and was partially funded by the European Union's Horizon 2020 Marie Sklodowska-Curie

Actions (grant agreement number 722046, ESIT project) and the European Union's Horizon 2020

Framework Programme (grant agreement number 848261, UNITI project).

9. Acknowledgments

We want to thank Nexstim Plc. for providing the electric-field guided neuronavigation rTMS

system to perform this feasibility trial.

10. References

[1]   Baguley D, McFerran D, Hall D. Tinnitus. The Lancet 2013;382:1600–7.
553   https://doi.org/10.1016/S0140-6736(13)60142-7.
554   [2]   Langguth B, Kreuzer PM, Kleinjung T, De Ridder D. Tinnitus: causes and clinical
555   management. Lancet Neurol 2013;12:920–30. https://doi.org/10.1016/S1474-
556   4422(13)70160-1.
557   [3]   McCormack A, Edmondson-Jones M, Somerset S, Hall D. A systematic review of the
558   reporting of tinnitus prevalence and severity. Hear Res 2016;337:70–9.
559   https://doi.org/10.1016/j.heares.2016.05.009.
560   [4]   Eggermont JJ. Pathophysiology of tinnitus. Prog Brain Res 2007;166:19–35.
561   https://doi.org/10.1016/S0079-6123(07)66002-6.
562   [5]   Eggermont JJ, Roberts LE. The Neuroscience of Tinnitus: Understanding Abnormal and
563   Normal Auditory Perception. Front Syst Neurosci 2012;6.
564   https://doi.org/10.3389/fnsys.2012.00053.
565   [6]   Eggermont JJ, Tass PA. Maladaptive Neural Synchrony in Tinnitus: Origin and Restoration.
566   Front Neurol 2015;6. https://doi.org/10.3389/fneur.2015.00029.
567   [7]   Henry JA, Roberts LE, Caspary DM, Theodoroff SM, Salvi RJ. Underlying Mechanisms of
568   Tinnitus: Review and Clinical Implications. J Am Acad Audiol 2014;25:5–126.
569   https://doi.org/10.3766/jaaa.25.1.2.

570	[8]	Roberts LE, Eggermont JJ, Caspary DM, Shore SE, Melcher JR, Kaltenbach JA. Ringing
571		Ears: The Neuroscience of Tinnitus. J Neurosci 2010;30:14972–9.
572		https://doi.org/10.1523/JNEUROSCI.4028-10.2010.
573	[9]	Langguth B. Non-Invasive Neuromodulation for Tinnitus. J Audiol Otol 2020;24:113–8.
574		https://doi.org/10.7874/jao.2020.00052.
575	[10]	Hallett M. Transcranial Magnetic Stimulation: A Primer. Neuron 2007;55:187–99.
576		https://doi.org/10.1016/j.neuron.2007.06.026.
577	[11]	Hoogendam JM, Ramakers GMJ, Di Lazzaro V. Physiology of repetitive transcranial
578		magnetic stimulation of the human brain. Brain Stimulat 2010;3:95–118.
579		https://doi.org/10.1016/j.brs.2009.10.005.
580	[12]	Siebner HR, Rothwell J. Transcranial magnetic stimulation: new insights into
581		representational cortical plasticity. Exp Brain Res 2003;148:1–16.
582		https://doi.org/10.1007/s00221-002-1234-2.
583	[13]	Fitzgerald P. Intensity-dependent effects of 1 Hz rTMS on human corticospinal excitability.
584		Clin Neurophysiol 2002;113:1136–41. https://doi.org/10.1016/S1388-2457(02)00145-1.
585	[14]	Fitzgerald PB, Fountain S, Daskalakis ZJ. A comprehensive review of the effects of rTMS
586		on motor cortical excitability and inhibition. Clin Neurophysiol 2006;117:2584–96.
587		https://doi.org/10.1016/j.clinph.2006.06.712.
588	[15]	Klomjai W, Katz R, Lackmy-Vallée A. Basic principles of transcranial magnetic stimulation
589		(TMS) and repetitive TMS (rTMS). Ann Phys Rehabil Med 2015;58:208–13.
590		https://doi.org/10.1016/j.rehab.2015.05.005.
591	[16]	Rossini PM, Burke D, Chen R, Cohen LG, Daskalakis Z, Di Iorio R, et al. Non-invasive
592		electrical and magnetic stimulation of the brain, spinal cord, roots and peripheral nerves:
593		Basic principles and procedures for routine clinical and research application. An updated
594		report from an I.F.C.N. Committee. Clin Neurophysiol Off J Int Fed Clin Neurophysiol
595		2015;126:1071–107. https://doi.org/10.1016/j.clinph.2015.02.001.
596	[17]	Eichhammer P, Langguth B, Marienhagen J, Kleinjung T, Hajak G. Neuronavigated
597		repetitive transcranial magnetic stimulation in patients with tinnitus: a short case series. Biol
598		Psychiatry 2003;54:862–5. https://doi.org/10.1016/S0006-3223(02)01896-6.
599	[18]	Langguth B, Eichhammer P, Wiegand R, Marienhegen J, Maenner P, Jacob P, et al.
600		Neuronavigated rTMS in a patient with chronic tinnitus. Effects of 4 weeks treatment.
601		Neuroreport 2003;14:977–80. https://doi.org/10.1097/01.wnr.0000068897.39523.41.
602	[19]	Schoisswohl S, Agrawal K, Simoes J, Neff P, Schlee W, Langguth B, et al. RTMS
603		parameters in tinnitus trials: a systematic review. Sci Rep 2019;9:1–11.
604		https://doi.org/10.1038/s41598-019-48750-9.
605	[20]	Lehner A, Schecklmann M, Greenlee MW, Rupprecht R, Langguth B. Triple-site rTMS for
606		the treatment of chronic tinnitus: a randomized controlled trial. Sci Rep 2016;6:22302.
607		https://doi.org/10.1038/srep22302.
608	[21]	Lehner A, Schecklmann M, Poeppl TB, Kreuzer PM, Vielsmeier V, Rupprecht R, et al.
609		Multisite rTMS for the treatment of chronic tinnitus: stimulation of the cortical tinnitus
610		network--a pilot study. Brain Topogr 2013;26:501–10. https://doi.org/10.1007/s10548-012-
611		0268-4.
612	[22]	De Ridder D, Song J-J, Vanneste S. Frontal cortex TMS for tinnitus. Brain Stimulat
613		2013;6:355–62. https://doi.org/10.1016/j.brs.2012.07.002.
614	[23]	Noh T-S, Kyong JS, Chang MY, Park M-K, Lee J-H, Oh S-H, et al. Comparison of
615		Treatment Outcomes Following Either Prefrontal Cortical-only or Dual-site Repetitive
616		Transcranial Magnetic Stimulation in Chronic Tinnitus Patients: A Double-blind

617   Randomized Study. Otol Neurotol Off Publ Am Otol Soc Am Neurotol Soc Eur Acad Otol
618   Neurotol 2017;38:296–303. https://doi.org/10.1097/MAO.0000000000001266.

[24]   Vanneste S, De Ridder D. Differences between a single session and repeated sessions of 1
619   Hz TMS by double-cone coil prefrontal stimulation for the improvement of tinnitus. Brain
620   Stimulat 2013;6:155–9. https://doi.org/10.1016/j.brs.2012.03.019.
621

[25]   Forogh B, Yazdi-Bahri S-M, Ahadi T, Fereshtehnejad S-M, Raissi GR. Comparison of two
622   protocols of transcranial magnetic stimulation for treatment of chronic tinnitus: a
623   randomized controlled clinical trial of burst repetitive versus high-frequency repetitive
624   Transcranial Magnetic Stimulation. Neurol Sci Off J Ital Neurol Soc Ital Soc Clin
625   Neurophysiol 2014;35:227–32. https://doi.org/10.1007/s10072-013-1487-5.
626

[26]   Khedr EM, Abo-Elfetoh N, Rothwell JC, El-Atar A, Sayed E, Khalifa H. Contralateral
627   versus ipsilateral rTMS of temporoparietal cortex for the treatment of chronic unilateral
628   tinnitus: comparative study. Eur J Neurol 2010;17:976–83. https://doi.org/10.1111/j.1468-
629   1331.2010.02965.x.
630

[27]   Khedr EM, Rothwell JC, Ahmed MA, El-Atar A. Effect of daily repetitive transcranial
631   magnetic stimulation for treatment of tinnitus: comparison of different stimulus frequencies.
632   J Neurol Neurosurg Psychiatry 2008;79:212–5. https://doi.org/10.1136/jnnp.2007.127712.
633

[28]   Plewnia C, Vonthein R, Wasserka B, Arfeller C, Naumann A, Schraven SP, et al. Treatment
634   of chronic tinnitus with theta burst stimulation: a randomized controlled trial. Neurology
635   2012;78:1628–34. https://doi.org/10.1212/WNL.0b013e3182574ef9.
636

[29]   Schecklmann M, Giani A, Tupak S, Langguth B, Raab V, Polak T, et al. Neuronavigated
637   left temporal continuous theta burst stimulation in chronic tinnitus. Restor Neurol Neurosci
638   2016;34:165–75. https://doi.org/10.3233/RNN-150518.
639

[30]   Dong C, Chen C, Wang T, Gao C, Wang Y, Guan X, et al. Low-Frequency Repetitive
640   Transcranial Magnetic Stimulation for the Treatment of Chronic Tinnitus: A Systematic
641   Review and Meta-Analysis of Randomized Controlled Trials. BioMed Res Int
642   2020;2020:3141278. https://doi.org/10.1155/2020/3141278.
643

[31]   Zenner H-P, Delb W, Kröner-Herwig B, Jäger B, Peroz I, Hesse G, et al. A
644   multidisciplinary systematic review of the treatment for chronic idiopathic tinnitus. Eur
645   Arch Otorhinolaryngol 2017;274:2079–91. https://doi.org/10.1007/s00405-016-4401-y.
646

[32]   Lefaucheur J-P, André-Obadia N, Antal A, Ayache SS, Baeken C, Benninger DH, et al.
647   Evidence-based guidelines on the therapeutic use of repetitive transcranial magnetic
648   stimulation (rTMS). Clin Neurophysiol 2014;125:2150–206.
649   https://doi.org/10.1016/j.clinph.2014.05.021.
650

[33]   Londero A, Bonfils P, Lefaucheur JP. Transcranial magnetic stimulation and subjective
651   tinnitus. A review of the literature, 2014–2016. Eur Ann Otorhinolaryngol Head Neck Dis
652   2018;135:51–8. https://doi.org/10.1016/j.anorl.2017.12.001.
653

[34]   Lefebvre-Demers M, Doyon N, Fecteau S. Non-invasive neuromodulation for tinnitus: a
654   meta-analysis and modeling studies. Brain Stimulat 2020.
655   https://doi.org/10.1016/j.brs.2020.11.014.
656

[35]   Guerra A, López-Alonso V, Cheeran B, Suppa A. Variability in non-invasive brain
657   stimulation studies: Reasons and results. Neurosci Lett 2017;719:133330.
658   https://doi.org/10.1016/j.neulet.2017.12.058.
659

[36]   López-Alonso V, Cheeran B, Río-Rodríguez D, Fernández-del-Olmo M. Inter-individual
660   Variability in Response to Non-invasive Brain Stimulation Paradigms. Brain Stimulat
661   2014;7:372–80. https://doi.org/10.1016/j.brs.2014.02.004.
662

[37] Polanía R, Nitsche MA, Ruff CC. Studying and modifying brain function with non-invasive
brain stimulation. Nat Neurosci 2018;21:174–87. https://doi.org/10.1038/s41593-017-0054-
4.
[38] Terranova C, Rizzo V, Cacciola A, Chillemi G, Calamuneri A, Milardi D, et al. Is There a
Future for Non-invasive Brain Stimulation as a Therapeutic Tool? Front Neurol
2019;9:1146. https://doi.org/10.3389/fneur.2018.01146.
[39] Cocchi L, Zalesky A. Personalized Transcranial Magnetic Stimulation in Psychiatry. Biol
Psychiatry Cogn Neurosci Neuroimaging 2018;3:731–41.
https://doi.org/10.1016/j.bpsc.2018.01.008.
[40] Garnaat SL, Fukuda AM, Yuan S, Carpenter LL. Identification of clinical features and
biomarkers that may inform a personalized approach to rTMS for depression. Pers Med
Psychiatry 2019;17–18:4–16. https://doi.org/10.1016/j.pmip.2019.09.001.
[41] Folmer RL, Carroll JR, Rahim A, Shi Y, Hal Martin W. Effects of repetitive transcranial
magnetic stimulation (rTMS) on chronic tinnitus. Acta Otolaryngol (Stockh) 2006;126:96–
101. https://doi.org/10.1080/03655230600895465.
[42] Fregni F, Marcondes R, Boggio PS, Marcolin MA, Rigonatti SP, Sanchez TG, et al.
Transient tinnitus suppression induced by repetitive transcranial magnetic stimulation and
transcranial direct current stimulation. Eur J Neurol 2006;13:996–1001.
https://doi.org/10.1111/j.1468-1331.2006.01414.x.
[43] Lorenz I, Müller N, Schlee W, Langguth B, Weisz N. Short-Term Effects of Single
Repetitive TMS Sessions on Auditory Evoked Activity in Patients With Chronic Tinnitus. J
Neurophysiol 2010;104:1497–505. https://doi.org/10.1152/jn.00370.2010.
[44] Meeus O, Blaivie C, Ost J, De Ridder D, Van de Heyning P. Influence of Tonic and Burst
Transcranial Magnetic Stimulation Characteristics on Acute Inhibition of Subjective
Tinnitus: Otol Neurotol 2009;30:697–703.
https://doi.org/10.1097/MAO.0b013e3181b05023.
[45] Plewnia C, Bartels M, Gerloff C. Transient suppression of tinnitus by transcranial magnetic
stimulation. Ann Neurol 2003;53:263–6. https://doi.org/10.1002/ana.10468.
[46] Müller N, Lorenz I, Langguth B, Weisz N. rTMS induced tinnitus relief is related to an
increase in auditory cortical alpha activity. PloS One 2013;8:e55557.
https://doi.org/10.1371/journal.pone.0055557.
[47] Kreuzer PM, Poeppl TB, Rupprecht R, Vielsmeier V, Lehner A, Langguth B, et al.
Individualized Repetitive Transcranial Magnetic Stimulation Treatment in Chronic
Tinnitus? Front Neurol 2017;8:126. https://doi.org/10.3389/fneur.2017.00126.
[48] Schoisswohl S, Langguth B, Schecklmann M. Short-Term Tinnitus Suppression With
Electric-Field Guided rTMS for Individualizing rTMS Treatment: A Technical Feasibility
Report. Front Neurol 2020;11:86. https://doi.org/10.3389/fneur.2020.00086.
[49] De Ridder D, Vanneste S, Langguth B, Llinas R. Thalamocortical Dysrhythmia: A
Theoretical Update in Tinnitus. Front Neurol 2015;6.
https://doi.org/10.3389/fneur.2015.00124.
[50] Llinás R, Urbano FJ, Leznik E, Ramírez RR, van Marle HJF. Rhythmic and dysrhythmic
thalamocortical dynamics: GABA systems and the edge effect. Trends Neurosci
2005;28:325–33. https://doi.org/10.1016/j.tins.2005.04.006.
[51] Llinás RR, Ribary U, Jeanmonod D, Kronberg E, Mitra PP. Thalamocortical dysrhythmia:
A neurological and neuropsychiatric syndrome characterized by magnetoencephalography.
Proc Natl Acad Sci U S A 1999;96:15222–7. https://doi.org/10.1073/pnas.96.26.15222.

[52] Weisz N, Dohrmann K, Elbert T. The relevance of spontaneous activity for the coding of the
tinnitus sensation. Prog Brain Res 2007:61–70. https://doi.org/10.1016/S0079-
6123(07)66006-3.
[53] Adjamian P, Sereda M, Zobay O, Hall DA, Palmer AR. Neuromagnetic Indicators of
Tinnitus and Tinnitus Masking in Patients with and without Hearing Loss. J Assoc Res
Otolaryngol 2012;13:715–31. https://doi.org/10.1007/s10162-012-0340-5.
[54] Balkenhol T, Wallhäusser-Franke E, Delb W. Psychoacoustic Tinnitus Loudness and
Tinnitus-Related Distress Show Different Associations with Oscillatory Brain Activity.
PLoS ONE 2013;8:e53180. https://doi.org/10.1371/journal.pone.0053180.
[55] Moazami-Goudarzi M, Michels L, Weisz N, Jeanmonod D. Temporo-insular enhancement
of EEG low and high frequencies in patients with chronic tinnitus. QEEG study of chronic
tinnitus patients. BMC Neurosci 2010;11. https://doi.org/10.1186/1471-2202-11-40.
[56] Schlee W, Schecklmann M, Lehner A, Kreuzer PM, Vielsmeier V, Poeppl TB, et al.
Reduced Variability of Auditory Alpha Activity in Chronic Tinnitus. Neural Plast
2014;2014:1–9. https://doi.org/10.1155/2014/436146.
[57] Weisz N, Müller S, Schlee W, Dohrmann K, Hartmann T, Elbert T. The neural code of
auditory phantom perception. J Neurosci Off J Soc Neurosci 2007;27:1479–84.
https://doi.org/10.1523/JNEUROSCI.3711-06.2007.
[58] Weisz N, Moratti S, Meinzer M, Dohrmann K, Elbert T. Tinnitus Perception and Distress Is
Related to Abnormal Spontaneous Brain Activity as Measured by
Magnetoencephalography. PLoS Med 2005;2:e153.
https://doi.org/10.1371/journal.pmed.0020153.
[59] De Ridder D, Congedo M, Vanneste S. The neural correlates of subjectively perceived and
passively matched loudness perception in auditory phantom perception. Brain Behav
2015;5. https://doi.org/10.1002/brb3.331.
[60] van der Loo E, Gais S, Congedo M, Vanneste S, Plazier M, Menovsky T, et al. Tinnitus
intensity dependent gamma oscillations of the contralateral auditory cortex. PloS One
2009;4:e7396. https://doi.org/10.1371/journal.pone.0007396.
[61] Noda Y. Toward the establishment of neurophysiological indicators for neuropsychiatric
disorders using transcranial magnetic stimulation-evoked potentials: A systematic review.
Psychiatry Clin Neurosci 2020;74:12–34. https://doi.org/10.1111/pcn.12936.
[62] Qiu S, Wang S, Yi W, Zhang C, He H. The lasting effects of 1Hz repetitive transcranial
magnetic stimulation on resting state EEG in healthy subjects. Annu Int Conf IEEE Eng
Med Biol Soc IEEE Eng Med Biol Soc Annu Int Conf 2019;2019:5918–22.
https://doi.org/10.1109/EMBC.2019.8857184.
[63] Thut G, Pascual-Leone A. A Review of Combined TMS-EEG Studies to Characterize
Lasting Effects of Repetitive TMS and Assess Their Usefulness in Cognitive and Clinical
Neuroscience. Brain Topogr 2010;22:219–32. https://doi.org/10.1007/s10548-009-0115-4.
[64] Hui J, Tremblay S, Daskalakis ZJ. The Current and Future Potential of Transcranial
Magnetic Stimulation With Electroencephalography in Psychiatry. Clin Pharmacol Ther
2019;106:734–46. https://doi.org/10.1002/cpt.1541.
[65] Tremblay S, Rogasch NC, Premoli I, Blumberger DM, Casarotto S, Chen R, et al. Clinical
utility and prospective of TMS–EEG. Clin Neurophysiol 2019;130:802–44.
https://doi.org/10.1016/j.clinph.2019.01.001.
[66] Schecklmann M, Lehner A, Gollmitzer J, Schmidt E, Schlee W, Langguth B. Repetitive
transcranial magnetic stimulation induces oscillatory power changes in chronic tinnitus.
Front Cell Neurosci 2015;9. https://doi.org/10.3389/fncel.2015.00421.

[67] Langguth B, Landgrebe M, Frank E, Schecklmann M, Sand PG, Vielsmeier V, et al.
Efficacy of different protocols of transcranial magnetic stimulation for the treatment of
tinnitus: Pooled analysis of two randomized controlled studies. World J Biol Psychiatry Off
J World Fed Soc Biol Psychiatry 2014;15:276–85.
https://doi.org/10.3109/15622975.2012.708438.
[68] Schecklmann M, Landgrebe M, Kleinjung T, Frank E, Sand PG, Rupprecht R, et al.
Changes in motor cortex excitability associated with temporal repetitive transcranial
magnetic stimulation in tinnitus: hints for cross-modal plasticity? BMC Neurosci
2014;15:71. https://doi.org/10.1186/1471-2202-15-71.
[69] Kleinjung T, Fischer B, Langguth B, Sand PG, Hajak G, Dvorakova J, et al. Validierung
einer deutschsprachigen Version des „Tinnitus Handicap Inventory". Psychiatr Prax
2007;34:S140–2. https://doi.org/10.1055/s-2006-940218.
[70] Newman CW, Jacobson GP, Spitzer JB. Development of the Tinnitus Handicap Inventory.
Arch Otolaryngol Head Neck Surg 1996;122:143–8.
https://doi.org/10.1001/archotol.1996.01890140029007.
[71] Langguth B, Goodey R, Azevedo A, Bjorne A, Cacace A, Crocetti A, et al. Consensus for
tinnitus patient assessment and treatment outcome measurement: Tinnitus Research
Initiative meeting, Regensburg, July 2006. Prog Brain Res 2007;166:525–36.
https://doi.org/10.1016/S0079-6123(07)66050-6.
[72] Genitsaridi E, Partyka M, Gallus S, Lopez-Escamez JA, Schecklmann M, Mielczarek M, et
al. Standardised profiling for tinnitus research: The European School for Interdisciplinary
Tinnitus Research Screening Questionnaire (ESIT-SQ). Hear Res 2019;377:353–9.
https://doi.org/10.1016/j.heares.2019.02.017.
[73] Chen R, Classen J, Gerloff C, Celnik P, Wassermann EM, Hallett M, et al. Depression of
motor cortex excitability by low-frequency transcranial magnetic stimulation. Neurology
1997;48:1398–403. https://doi.org/10.1212/wnl.48.5.1398.
[74] Kumru H, Albu S, Rothwell J, Leon D, Flores C, Opisso E, et al. Modulation of motor
cortex excitability by paired peripheral and transcranial magnetic stimulation. Clin
Neurophysiol 2017;128:2043–7. https://doi.org/10.1016/j.clinph.2017.06.041.
[75] Delorme A, Makeig S. EEGLAB: an open source toolbox for analysis of single-trial EEG
dynamics including independent component analysis. J Neurosci Methods 2004;134:9–21.
https://doi.org/10.1016/j.jneumeth.2003.10.009.
[76] Perrin F, Pernier J, Bertrand O, Echallier JF. Spherical splines for scalp potential and current
density mapping. Electroencephalogr Clin Neurophysiol 1989;72:184–7.
https://doi.org/10.1016/0013-4694(89)90180-6.
[77] Oostenveld R, Fries P, Maris E, Schoffelen J-M. FieldTrip: Open Source Software for
Advanced Analysis of MEG, EEG, and Invasive Electrophysiological Data. Comput Intell
Neurosci 2011;2011:1–9. https://doi.org/10.1155/2011/156869.
[78] Vanneste S, van de Heyning P, De Ridder D. The neural network of phantom sound changes
over time: a comparison between recent-onset and chronic tinnitus patients. Eur J Neurosci
2011;34:718–31. https://doi.org/10.1111/j.1460-9568.2011.07793.x.
[79] Holm S. A Simple Sequentially Rejective Multiple Test Procedure. Scand J Stat 1979;6:65–
70.
[80] Siebner HR, Lang N, Rizzo V, Nitsche MA, Paulus W, Lemon RN, et al. Preconditioning of
low-frequency repetitive transcranial magnetic stimulation with transcranial direct current
stimulation: evidence for homeostatic plasticity in the human motor cortex. J Neurosci Off J
Soc Neurosci 2004;24:3379–85. https://doi.org/10.1523/JNEUROSCI.5316-03.2004.

[81] Maeda F, Keenan JP, Tormos JM, Topka H, Pascual-Leone A. Interindividual variability of
the modulatory effects of repetitive transcranial magnetic stimulation on cortical
excitability. Exp Brain Res 2000;133:425–30. https://doi.org/10.1007/s002210000432.
[82] Nettekoven C, Volz LJ, Leimbach M, Pool E-M, Rehme AK, Eickhoff SB, et al. Inter-
individual variability in cortical excitability and motor network connectivity following
multiple blocks of rTMS. NeuroImage 2015;118:209–18.
https://doi.org/10.1016/j.neuroimage.2015.06.004.
[83] Sommer M, Wu T, Tergau F, Paulus W. Intra- and interindividual variability of motor
responses to repetitive transcranial magnetic stimulation. Clin Neurophysiol 2002;113:265–
9. https://doi.org/10.1016/S1388-2457(01)00726-X.
[84] Soleimani R, Jalali MM, Hasandokht T. Therapeutic impact of repetitive transcranial
magnetic stimulation (rTMS) on tinnitus: a systematic review and meta-analysis. Eur Arch
Otorhinolaryngol 2016;273:1663–75. https://doi.org/10.1007/s00405-015-3642-5.
[85] Lehner A, Schecklmann M, Poeppl TB, Kreuzer PM, Peytard J, Frank E, et al. Efficacy and
Safety of Repeated Courses of rTMS Treatment in Patients with Chronic Subjective
Tinnitus. BioMed Res Int 2015;2015:975808. https://doi.org/10.1155/2015/975808.
[86] Rachid F. Maintenance repetitive transcranial magnetic stimulation (rTMS) for relapse
prevention in with depression: A review. Psychiatry Res 2018;262:363–72.
https://doi.org/10.1016/j.psychres.2017.09.009.
[87] Silvanto J, Pascual-Leone A. State-Dependency of Transcranial Magnetic Stimulation.
Brain Topogr 2008;21:1–10. https://doi.org/10.1007/s10548-008-0067-0.
[88] Weisz N, Lüchinger C, Thut G, Müller N. Effects of individual alpha rTMS applied to the
auditory cortex and its implications for the treatment of chronic tinnitus. Hum Brain Mapp
2014;35:14–29. https://doi.org/10.1002/hbm.22152.
[89] Weisz N, Steidle L, Lorenz I. Formerly known as inhibitory: effects of 1-Hz rTMS on
auditory cortex are state-dependent: Formerly known as inhibitory. Eur J Neurosci
2012;36:2077–87. https://doi.org/10.1111/j.1460-9568.2012.08097.x.
[90] De Ridder D, Verstraeten E, Van der Kelen K, De Mulder G, Sunaert S, Verlooy J, et al.
Transcranial Magnetic Stimulation for Tinnitus: Influence of Tinnitus Duration on
Stimulation Parameter Choice and Maximal Tinnitus Suppression: Otol Neurotol
2005;26:616–9. https://doi.org/10.1097/01.mao.0000178146.91139.3c.
[91] King ROC, Singh Shekhawat G, King C, Chan E, Kobayashi K, Searchfield GD. The Effect
of Auditory Residual Inhibition on Tinnitus and the Electroencephalogram. Ear Hear
2021;42:130–41. https://doi.org/10.1097/AUD.0000000000000907.
[92] Thut G, Veniero D, Romei V, Miniussi C, Schyns P, Gross J. Rhythmic TMS Causes Local
Entrainment of Natural Oscillatory Signatures. Curr Biol 2011;21:1176–85.
https://doi.org/10.1016/j.cub.2011.05.049.
[93] Wagner J, Makeig S, Hoopes D, Gola M. Can Oscillatory Alpha-Gamma Phase-Amplitude
Coupling be Used to Understand and Enhance TMS Effects? Front Hum Neurosci 2019;13.
https://doi.org/10.3389/fnhum.2019.00263.
843
844
845
846

11.Tables

Table 1

Participants characteristics

N (female)	22 (5)			
Handedness (left/ right/ both) (4 missings)	0/ 14/ 4			
Tinnitus laterality (left/ right/ both/ inside head) (4 missings)	1/ 1/ 14/ 2			
Tinnitus loudness fluctuation (yes/ no) (4 missings)	14/ 4			
	M ± SD	**Md**	**Min**	**Max**
Age (years)	57.05 ± 6.79	57.50	43.00	69.00
Tinnitus duration (months) (2 missings)	131.65 ± 116.79	102.00	5.00	420.00
Hearing loss left (dB) (8 missings)	23.42 ± 9.72	22.22	7.22	41.67
Hearing loss right (dB) (8 missings)	27.87 ± 14.29	23.06	7.78	61.88
THI score (0-100)	58.49 ± 19.55	58.00	30.00	90.00
Tinnitus loudness (%) (6 missings)	70.31 ± 19.87	80.00	30.00	90.00
RMT (%)	33.95 ± 5.33	33.00	25.00	46.00
RMT (%) + EEG cap	41.91 ± 9.65	41.00	30.00	66.00
Electrical field (V/m)	54.47 ± 13.90	52.50	27.00	86.00
Electrical field (V/m) + EEG cap	66.66 ± 15.38	66.25	44.00	99.00

M = mean; SD = standard deviation; Md = Median; Min = minimum; Max = maximum; THI = Tinnitus Handicap Inventory; RMT = resting motor threshold, EEG = Electroencephalography

Table 2

Individual test session responses per protocol – behavioral and EEG responder

Subject	Behavioral responder				Alpha (8 - 12.5 Hz) / Gamma (30.5 - 44 Hz)			
		1Hz	10Hz	20Hz		1Hz	10Hz	20Hz
1	Left	12.86	-	-	Left	-	↓γ	↑α↓γ
	Right	17.15	-	-	Right	-	↓γ	-
						1Hz	10Hz	20Hz
2		-			Left	↑α↓γ	↑α↓γ	-
					Right	-	-	-
		1Hz	10Hz	20Hz		1Hz	10Hz	20Hz
3	Left	3.98	-	3.71	Left	↑α↓γ	↑α	-
	Right	-	-	-	Right	↓γ	-	-
		1Hz	10Hz	20Hz		1Hz	10Hz	20Hz
4	Left	-	10.35	-	Left	↑α↓γ	-	-
	Right	-	9.28	25.71	Right	-	-	-
		1Hz	10Hz	20Hz		1Hz	10Hz	20Hz
5	Left	-	16.42	-	Left	↑α	↓γ	↑α
	Right	10.00	16.79	21.42	Right	-	-	-

6

	1Hz	10Hz	20Hz		1Hz	10Hz	20Hz
Left	-	10.71	-	Left	↑α	-	↑α
Right	-	-	-	Right	-	↑α↓γ	-

7

					1Hz	10Hz	20Hz
		-		Left	-	↑α	-
				Right	-	↑α	↑α↓γ

8

	1Hz	10Hz	20Hz		1Hz	10Hz	20Hz
Left	-	6.07	8.93	Left	-	↑α↓γ	↓γ
Right	-	4.64	9.28	Right	-	↑α	-

9

	1Hz	10Hz	20Hz		1Hz	10Hz	20Hz
Left	-	-	3.93	Left	↑α↓γ	↑α↓γ	↑α↓γ
Right	7.55	-	-	Right	↓γ	-	-

10

	1Hz	10Hz	20Hz		1Hz	10Hz	20Hz
Left	15.36	-	5.71	Left	↓γ	↑α	↓γ
Right	-	-	-	Right	-	↓γ	↑α↓γ

11

					1Hz	10Hz	20Hz
		-		Left	↑α	-	-
				Right	↓γ	↑α↓γ	↑α

12

					1Hz	10Hz	20Hz
		-		Left	-	-	-
				Right	-	↑α	↓γ

13

					1Hz	10Hz	20Hz
		-		Left	↓γ	↑α↓γ	-
				Right	↑α↓γ	↓γ	-

14

					1Hz	10Hz	20Hz
		-		Left	↑α	↓γ	↑α↓γ
				Right	↓γ	-	↑α

15

					1Hz	10Hz	20Hz
		-		Left	↓γ	↑α	-
				Right	↓γ	-	-

16

	1Hz	10Hz	20Hz		1Hz	10Hz	20Hz
Left	-	5.00	12.85	Left	-	↑α↓γ	-
Right	-	-	10.00	Right	-	↑α	↓γ

17

					1Hz	10Hz	20Hz
		-		Left	-	↓γ	-
				Right	↑α	↓γ	↑α↓γ

18

					1Hz	10Hz	20Hz
		-		Left	↓γ	↓γ	↑α↓γ
				Right	↑α↓γ	↑α	-

19

	1Hz	10Hz	20Hz		1Hz	10Hz	20Hz
Left	-	13.57	13.57	Left	↑α	↓γ	-
Right	-	-	-	Right	↑α	↓γ	↑α

20					1Hz	10Hz	20Hz
			-	Left	↓γ	↓γ	-
				Right	-	-	-

21		1Hz	10Hz	20Hz		1Hz	10Hz	20Hz
	Left	-	-	-	Left	-	-	↓γ
	Right	7.14	-	3.93	Right	↓γ	↓γ	↓γ

22		1Hz	10Hz	20Hz		1Hz	10Hz	20Hz
	Left	-	-	7.85	Left	↓γ	↑α↓γ	-
	Right	-	-	-	Right	-	-	-

Note. Behavioral response was defined via a sham-superior tinnitus suppression on both test session days in at least one of the verum protocols. EEG response was defined as sham-superior alpha increases respectively gamma decreases in a minimum number of two neighboring electrodes on both test session days in at least one verum protocol. For behavioral responders the mean tinnitus suppression (averaged over both test session days) is presented per rTMS protocol. Alpha increases (↑α) and gamma decreases (↓γ) are equally presented per subject and protocol. Dashes indicate no response.

Table 3
Summary of behavioral and electrophysiological responder per rTMS protocol

	Behavioral responder	↑Alpha	↓Gamma
Total number of responders (N)	12	20	22
Number of responders per protocol (N)			
Left 1 Hz	3	9	10
Right 1 Hz	4	4	8
Left 10 Hz	6	10	13
Right 10 Hz	3	7	8
Left 20 Hz	7	6	7
Right 20 Hz	5	6	6

12. Figure legends

Figure 1. Test session procedure.

Test sessions were conducted on two different days according to the same procedure. Three

different verum rTMS protocols (1 Hz, 10 Hz, 20 Hz) with 200 pulses each and one sham protocol

(0.1 Hz) with 20 pulses were administered over the left and right temporo-parietal junction in a

randomized order. Thus, a total number of 8 different rTMS protocols was applied on each test

session day. Before and after each magnetic stimulation, participants were requested to rate the

subjective loudness of their tinnitus on a numeric rating scale from 0% to 110% (0% - complete

absence of tinnitus; 100% - no change; 110% - 10% loudness increase) at seven points in time (30

seconds inter-rating interval). In parallel to these ratings, resting state brain activity was recorded

with electroencephalography (EEG) for three minutes. At the end of each rTMS protocol (pre-

rating/ recording – stimulation – post-rating/ recording), patients had to rate the induced level of

discomfort on a numeric rating scale from 0 to 10 (10 – utmost discomfort).

Figure 2. Reliability of TMS-induced changes in tinnitus loudness.

Spearman correlations of rated tinnitus suppression (%) per rTMS protocol between the first and

the second test session day are illustrated. Colored ribbons represent confidence intervals (95%).

Reliable changes were observed for a (b) 10 Hz stimulation over the left temporo-parietal

junction and a (c) 20 Hz stimulation over the right temporo-parietal junction. Both correlations do

not withstand a correction for multiple comparison and should therefore not be overrated and

interpreted only exploratively.

Figure 3. Reliability of TMS-induced changes in oscillatory brain activity

Spearman correlations coefficients (r_s) of right hemispheric 1 Hz rTMS-induced changes in the

gamma frequency band (33 – 44 Hz) for the two test session days are projected on a topographical

map per electrode. Significant reliable electrodes are highlighted with asterisks. Separate reliable

changes are primary present over parieto-occipital electrodes, whereby neighboring channels PO7

and O1 form a significant reliable cluster. However, this effect disappears after a correction for

multiple comparison and is therefore only conditionally interpretable.

Figure 4. Topography of electrophysiological response – alpha responder

Topography and quantity of sensors showing sham-superior increases in the alpha frequency

band in a minimum number of two neighboring channels on both test session days are illustrated

per rTMS protocol. Responses in the alpha frequency band (8 – 12.5 Hz) to one of the verum

protocols primarily occurred over parieto-occipital electrodes.

Figure 5. Topography of electrophysiological response – gamma responder

Topography and quantity of sensors showing sham-superior decreases in the gamma frequency

band in a minimum number of two neighboring channels on both test session days are illustrated

per rTMS protocol. Responses in the gamma frequency band (33 – 44 Hz) to one of the verum

protocols mostly appeared over frontal electrodes on both hemispheres and partially over parieto-

occipital sensors.