1   Title: Neuromotor Prosthetic to Treat Stroke-Related Paresis
2

3   Authors: Mijail D. Serruya[1*], Alessandro Napoli[1], Nicholas Satterthwaite[1], Joseph Kardine, Jr[2], Joseph McCoy[2],
4   Namrata Grampurohit[3], Kiran Talekar[4], Devon Middleton[4], Feroze Mohamed[4], Michael Kogan[5], Ashwini Sharan[5],
5   Chengyuan Wu[5], Robert Rosenwasser[5]
6
[1] Department of Neurology, Thomas Jefferson University, Philadelphia, PA 19107
* Corresponding author: Mijail.Serruya@jefferson.edu
[2] Jefferson Rehabilitation (Outpatient), Thomas Jefferson University Hospital, Philadelphia, PA 19107
[3] College of Rehabilitation Science, Thomas Jefferson University, Philadelphia, PA 19107
[4] Department of Radiology, Thomas Jefferson University, Philadelphia, PA 19107
[5] Department of Neurosurgery, Thomas Jefferson University, Philadelphia, PA 19107

ABSTRACT
15
Background
Functional recovery of independent arm movement typically plateaus within six months following a stroke, leaving
chronic motor deficits. This feasibility study tested whether a wearable, powered exoskeletal orthosis, driven by a
percutaneous, implanted brain–computer interface (BCI), using the ensemble activity of individual neurons in the
precentral gyrus in the affected cortical hemisphere, could restore voluntary upper extremity function in a person
with chronic hemiparetic subsequent to a cerebral hemispheric stroke involving subcortical gray and white matter
and cortical gray matter .
23
Methods
One chronic hemiparetic stroke survivor with moderate-to-severe upper-limb motor impairment used a powered
wearable elbow-wrist-hand orthosis that opened and closed the affected hand using cortical activity, recorded from
four 64-channel microelectrode arrays implanted in the ipsilesional precentral gyrus, based on decoding of neural
spiking (discriminated single- and multi-unit) patterns and high frequency local field potentials generated by
imagined hand movements using technology and decoding methods used for people with other causes of paralysis.
The system was evaluated in a home setting daily for 12 weeks. Motor function was evaluated before, during, and
after the treatment.
32
Results
Robust single unit activity, modulating with attempted or imagined movement, was present throughout the
precentral gyrus areas implanted for the duration of the study. Although abnormal neural oscillations, that appeared
to correlate with contralateral spasticity, were present, the participant was able to acquire voluntary control over a
hand-orthosis BCI based upon ensemble single unit activity and high frequency local field potential power. The
study participant (C2) was able to achieve a maximum score of 10 points on the Action Research Arm Test (ARAT)
while using the BCI system as compared to 0 without any device, and 5 using only myoelectric control (where 0
indicates no functional abilities an upper extremity and 53 indicates full, normal healthy function). Orthosis-
powered hand-opening was faster with BCI-based control as compared to myoelectric control, on a standardized
object-movement task, possibly due to the fact that BCI-control could be deployed without triggering abnormal
spasticity inevitably reinforced by myoelectric control.
44
Conclusions
The findings demonstrate the therapeutic potential to use an implantable BCI system coupled to a limb exoskeleton
to be used to promote and assess neurally driven rehabilitation of limb function. This targeting the affected
hemisphere, to "electrically bypass" the stroke, in this n-of-1 pilot study. The participant's ability to rapidly acquire
voluntary control over otherwise paralyzed hand opening, more than 18 months after a subcortical stroke, lays the
foundation for a fully implanted movement restoration system that could restore independent arm function to stroke
survivors.

52
53   INTRODUCTION
54
Stroke is a leading cause of disability [1] with a global prevalence of over 42 million people in 2015 [2], affecting over
four million adults in the United States alone with 800,000 new cases per year [3]. Stroke leads to permanent motor

disabilities in 80% of cases [4], and half of stroke survivors require long term care. Brain computer interface (BCI)
technologies offer a potential solution to restore functional independence and improve health in people living with
its effects. In the past decade, intracortical BCI technology has continued to advance, with multiple groups
demonstrating the safety and efficacy of this approach to derive control signals [5–7] to restore communication and
control. In parallel, wearable robotic orthosis technology has evolved into FDA-cleared devices bringing concrete
benefits to patients with weakened limbs [8,9], as a therapeutic intervention, to promote recovery. This single-patient
pilot clinical trial sought to provide proof of concept that a commercially available powered arm orthosis could be
linked to thecerebral cortex in an adult with the most common form of chronic stroke. A direct path from the brain's
motor centers to the orthotic could both reanimate a paralyzed limb to enable useful hand and arm function and
potentially provide a new, neurally driven rehabilitation tool.
67
Several signal sources have been coopted to provide commands to move paralyzed limbs. Electromyographic
(EMG) control of a powered orthosis or functional electrical stimulation (FES) of muscles, has proven problematic
either because users could not generate sufficient or reliable activity to provide a good control signal , or because
voluntary activation of those recorded muscles (that were intended to generate the command) was opposed by the
stimulator's effects [10]. Contralaterally controlled electrical stimulation- where activity from the unaffected arm
triggers stimulation on the paretic arm- is a useful therapeutic intervention to improve function in the weaker limb[11]
but it is not clear how this unnatural command source could be generalized to continuously-worn devices that enable
independent arm movements. Several groups have explored scalp EEG, which is closer to the command's origins, to
derive control signals to drive robotic braces, and in one case, FES [8,12–14]. While using EEG-derived signals may be
promising for rehabilitation therapy, it would not be feasible for daily independent function. Scalp EEG requires
careful electrode positioning by a trained technician, while skin sweat and hair can cause impedances to fluctuate,
thus compromising signal quality. Daily application of even a subset of contacts to the same skin sites can lead to
skin breakdown and clinically worrisome cellulitis. Further, EEG signals are limited in the richness of commands
that can be easily and reliably derived from the available signal. By contrast, intracortical interfaces offer a rich
sources of high resolution,   multidimensional control signals, since it is the origin of such signals in healthy adults,
in non-human primates and in people with spinal or brainstem disorders [7,15,16].
84
While the vast majority of strokes involve cerebral white matter and even direct parenchymal damage. Notably,
intracortical neuromotor prosthetics have not been tested in people with strokes above the mesencephalon. It is not
known whether motor cortex remains a reliable signal source in this large population. If you used siA proof-of-
concept that a brain-computer interface, based on microelectrode arrays (MEAs) implanted in intact cortex above a
subcortical stroke, could restore behaviorally useful independent, voluntary movement, could lead to the development
of a fully implantable medical device that, in principle, could reverse the motor deficits caused by stroke, and
ultimately other lesional diseases such as tumors, multiple sclerosis, and arteriovenous malformations near or in the
cortex. This approach of using adjacent healthy cortex could also be expanded to restore non-motor functions,
including language, memory, perception and executive function.
94
95 METHODS
96
Approval for this study was granted by the US Food and Drug Administration (Investigational Device Exemption)
and the Thomas Jefferson University Institutional Review Board. The participant described in this report has provided
permission for photographs, videos and portions of his protected health information to be published for scientific and
educational purposes. After completion of informed consent, medical and surgical screening procedures, two
MultiPorts (Blackrock Microsystems, UT), each comprising two 8x8 platinum tipped microelectrode arrays tethered
to a titanium pedestal connector, were implanted into the cortex of the precentral gyrus using a pneumatic insertion
technique[1718]. Details of the human surgical procedure are in preparation for publication, but it followed procedures
used in other studies of similar technology (REFS?). Trial selection criteria are available online for this trial (see
Clinicaltrials.gov, number NCT03913286). The trial was designed with the implantation phase to last a maximum of
three months (Fig. 1). Approval for this study was granted by the US Food and Drug Administration (Investigational
Device Exemption) and the Thomas Jefferson University Institutional Review Board. The participant described in this
report has provided permission for photographs, videos and portions of his protected health information to be
published for scientific and educational purposes. After completion of informed consent, medical and surgical
screening procedures, two MultiPorts (Blackrock Microsystems, UT), each comprising two 8x8 microelectrode arrays
tethered to a titanium pedestal connector, were implanted into the cortex of the precentral gyrus using a pneumatic

insertion technique[17][18]. Details of the human surgical procedure are in preparation for publication. Selection criteria
are available online (see Clinicaltrials.gov, number NCT03913286). The trial was designed with the implantation
phase to last a maximum of three months (Fig. 1).
115

117 **Figure 1. The clinical trial timeline.**
118

Participant. The participant (ID number C2) was a right-handed male who experienced right hemispheric stroke,
manifest as acute onset dense left hemiparesis and expressive aphasia, at which time he was age between ages 35 to
40. Due to unknown time of onset and hypertension at presentation, the participant was not a candidate for
thrombolysis. CT angiogram showed occlusion of the right posterior cerebral artery and high-grade stenosis of the
left posterior cerebral artery in the proximal P2 segment. MRI of the brain showed acute infarcts in the right basal
ganglia/corona radiata and right occipital lobe. He was started on dual antiplatelet therapy for 3 weeks and then was
transitioned to aspirin 81 mg once daily, along with atorvastatin and anti-hypertensives. He had left-sided
hemiparesis, dysphagia, left homonymous hemianopsia and dense left visual neglect and was transferred to rehab
hospital for supervised therapies. He subsequently underwent two months of intensive inpatient rehabilitation and
aphasia and dysphagia resolved; over a period of three months, he acquired the ability to ambulate independently,
albeit with a persistent left foot drop. Neuroimaging showed evidence of multi-focal strokes, and prior silent strokes.
The participant had previously been in good health and did not have any known stroke risk factors such as diabetes
or smoking. There was a history of loud snoring and the participant had not been evaluated for obstructive sleep
apnea. Transthoracic and transesophageal echocardiography were normal as were serial hypercoagulability panels;
the participant was adopted and the biological family history unknown. The participant was deemed to have had
embolic strokes of unknown source (ESUS). Although serial electrocardiography since the stroke was normal, the
participant is being scheduled for a loop recorder to survey for possible paroxysmal atrial fibrillation. The
participant had learning disabilities and was presumed to have had mild cognitive impairment prior to the stroke.
Screening formal neuropsychological testing identified significant neurocognitive problems (full scale IQ 59) and
also concluded that the participant remained fully capable to provide proper informed consent and to participate in
this trial, meeting its demands and requirements. The participant provided both verbal and written informed consent,
both to participate in the trial and to share his identifying information with the public. He had been working full time
at the time of the stroke and has been unable to return to work since the stroke.
142

Pre-operative fMRI. The participant underwent MRI on a 3T Philips Ingenia MRI scanner (Amsterdam, The
Netherlands). A 1mm isotropic 3-D T2 FLAIR was obtained for structural localization. A single-shot echoplanar
gradient echo imaging (EPI) sequence with 80 volumes, repetition time (TR) = 2 s, echo time (TE) = 25 msec, voxel
size = 3 x 3 mm2, slice thickness = 3mm, axial slices = 37. _Motor Task:_ The participant was asked to visualize
movements of his paretic left hand during the MRI. Each motor trial consisted of a block design featuring a 20 s rest
block and a 20 s active block repeated. This block design was repeated between 4-6 times for a total of 160-240 s
scans. Visual stimulus comprised a 20 s video depicting two 3D modeled limbs at rest, followed by a 20 s video of
the limb(s) performing the desired motor task. Motor tasks were designated "hand" (repeated hand open/clench) and
"arm" (repeated arm extension elbow), and classified as "active" (participant performed or attempted to perform the
motion) or "passive" (physician manually moved participant's hand/arm). Hand motor was performed as bilateral-
active, paretic-active, and paretic-passive. Arm motor was performed as paretic-active and paretic-passive. In active
tasks, the participant was instructed to follow the movements in the video or concentrate on following in the case of
the paretic limb. Task choice and prioritization was based on pre-exam training and examination of participant's
capabilities and examination of BOLD activation observed during the scan. For example, a breath-hold challenge was

employed after the first couple of tasks failed to show desired activation with an attempt to improve cerebrovascular
reactivity. BOLD activation subsequently became apparent. Post processing of the BOLD data including motion
correction, smoothing, and general linear model (GLM) estimation was performed using SPM software
(www.fil.ion.ucl.ac.uk/spm) as well as Nordic brain EX software (NordicNeuroLab, Bergen, Norway). After BOLD
post-processing, generated statistical maps were overlaid on the 3D T2 FLAIR image for visualization of activation.

162
163

Cortimo system. 'Cortimo' is the designation
provided to the FDA to represent the overall
system (Fig. 2) that comprised two percutaneous
Multiports (Blackrock Microsystems, Inc), each
in turn having two multi-electrode array sensors,
the cabling, amplifiers, software and the powered
MyoPro orthosis (described subsequently). Each
sensor is an 8x8 array of silicon microelectrodes
that protrude 1.5 mm from a 3.3 x 3.3-mm
platform. At manufacture, electrodes had an
impedance ranging between 70 KOhm and 340
KOhm. The array was implanted onto the surface
of the MI arm/hand region guided by the pre-
operative fMRI; with electrodes penetrating into
the cortex to attempt to record neurons in layer V.
Recorded electrical signals pass externally
through a Ti percutaneous connector, which was
secured to the skull. Cabling attached to the
connector during recording sessions routes

Figure 2. The Cortimo system. The overall device comprises two Multiports, each with two 8x8 microelectrode arrays, patient cables linked to external amplifiers, a decoding computer and a wearable, powered arm orthosis.

signals to external amplifiers and then to a series of computers in a cart that process the signals and convert them into
different outputs, such as servo motor position of the MyoPro brace or screen position of a neural cursor that can be
viewed by the participant on a computer monitor. Currently, this system must be set up and managed by an experienced
technician.

MyoPro brace. The MyoPro (Myomo, Inc, Cambridge, MA) is an FDA-cleared myoelectric powered arm orthosis
designed to support a paretic arm [9].The rigid brace incorporates metal contacts attached to soft straps that can be
adjusted such that contacts rest on the biceps and triceps proximally, and on wrist flexors and extensors distally, on
the paretic upper extremity. The sensors continuously record the root mean square of underlying muscle activity.
Thresholds are manually set using software such that signals exceeding them will trigger one of the MyoPro motors.
Because the participant retained residual elbow flexion and extension strength, the motor at the elbow was set up such
that biceps activation triggered elbow flexion and triceps activation triggered elbow extension. For hand opening, the
MyoPro was set up to either use myoelectric control, or to use BCI-based control. Since the participant was unable to
voluntarily extend the wrist or open the fingers, the myoelectric mode was set up such that the default state was with
the hand open, and it would only be closed by activating enough wrist flexor activity.

Recording sessions. Research sessions were scheduled five days per week at a temporary residence, adjacent to the
hospital, provided to the participant. Sessions could be cancelled (for example, due to participant request or schedule
conflict) or ended early at the participant's request. Sessions would commence with neural recording and spike
discrimination. While initial sessions included filter building and structured clinical end-point (cursor control) trials,
in the final month of the trial, "training-less" algorithms were used with the participant proceeding directly to BCI-
controlled hand action as soon as the patient cables were connected. Performance of computer tasks, orthosis control
and occupational therapy exercises followed. The electrodes and neural signals selected immediately before filter
building remained constant for any given session's orthosis control trials.
Spike discrimination. Units were extracted using an automatic thresholding approach based for each electrode channel,
based on Root Mean Squared multipliers[16]. We did not perform any type of spike sorting and applied no objective
criteria to determine whether any particular unit recorded on one day was the same as that recorded on a subsequent
day.

Spike discrimination. Units were extracted using an automatic thresholding approach based on Root Mean Squared
multipliers[16]. We did not perform any type of spike sorting and applied no objective criteria to determine whether any
particular unit (or admixture of units) recorded on one day was the same as that recorded on a subsequent day. Smaller
amplitude signals likely comprised a combination of spikes from multiple neurons.

Decoder Filter Building. For each session, single and multiunit data or high frequency (100-1000 Hz) local field
potentials derived from multiple channels (20-30) were used to create a linear filter to convert these real-time
multidimensional neural features into either a one or a two-dimensional (position or velocity) output signal. During
the course of the trial, a variety of motor activity and motor imagery approaches were tested for filter building,
including imagining opening and closing the paretic hand, passively flexing and extending the elbow, passively
opening and closing the hand, and simply observing a computer cursor displayed on a monitor moving up and down
without any specific instruction. Training data for building the linear filter were collected with the participant gazing
at a screen where a target cursor was moved slowly up and down for one minute (5 seconds to go from the top to the
bottom of the screen or vice versa, at 20° visual angle). After this preliminary filter was built, a new 1-minute re-
training session was performed, this time the manually controlled target cursor was accompanied by a prediction
cursor that was neurally controlled by the participant. Using this additional training set, a second filter was built and
then tested on a simple target acquisition game in which the y-position of the predicted output was discretized into
zones such that positions on the upper part of the screen would cause an animation sprite to move up by a fixed
distance (1 cm), and positions on the lower part of the screen would cause the sprite to move down by the same fixed
distance.

BCI MyoPro use. This discrete output was then used to control the aperture of the hand via the MyoPro's hand brace
motor. The up-down mapping on the screen was translated into closed-open positions of the hand. The participant
then performed a series of functional tasks including grasping and then dropping an object, the Action Research Arm
Test[19], and a variation on Jebsen Taylor item moving test[20]. These were tested with both the participant seated and
standing.

BCI Orthosis use. The discrete output was then used to control the aperture of the hand via the MyoPro's hand brace
motor. The up-down mapping on the screen was translated into closed-open positions of the hand. The participant
then performed a series of functional tasks including grasping and then dropping an object, the Action Research Arm
Test[19], and a variation on Jebsen Taylor item moving test[20]. These were tested with both the participant seated and
standing.

"Training-less" Mapping. During certain sessions, the participant was able to effortlessly adapt the continuous cursor
control on to hand open-close control. During other sessions, however, the participant could not achieve this 'efferent
remapping,' and would attempt to overpower the orthosis motors with residual finger flexion strength. A novel
'training-less' approach was deployed in which a rolling 1-second baseline of the LFPs signals from the most
neuromodulated 20 channels was used to calculate spectral power in the high gamma band (100-500 Hz). Namely, 1-
second long LFP continuous voltages were used for computing the average spectral density estimation in the frequency
band 100-500 Hz, using non-overlapping frequency bins with a 50Hz width. Spectral density was computed using the
Matlab built-in periodogram method. Spectral values were updated every 500 ms, using 1-second-long rolling
windows with 50% overlap. Finally, these real-time spectral features derived from the 20 most neuromodulated
channels were averaged across channels to produce a single high gamma band value for each 500 ms software update.
Orthosis hand-closure would be triggered by an increase in this mean spectral power from the resting baseline ranging
between 0.5 and 3 V^2/Hz to values greater than 10 V^2/Hz, where real-time values above this threshold would make
the hand motor close.

Concomitant Occupational and Physical Therapy. Since being discharged from acute rehabilitation 60 days after the
initial stroke, the participant enrolled in outpatient physical and occupational therapy. Prior to the device implantation,
the participant completed a six-week course of occupational therapy (two to three sessions per week) intended to be a
screening phase to assess the participant's ability to wear and use the MyoPro device with myoelectric control and
follow complex instructions. Following device implantation, the participant continued occupational therapy, twice per
week, and physical therapy, once per week. Occupational therapy focused on postural training while seated and
walking, donning and doffing the MyoPro, and using the MyoPro for functional activities. Timed functional electrical
stimulation (e.g., pincer grasp programs; XCite, from Restorative Therapies, Inc, Baltimore, MD) and vibration
therapy were used for spasticity management both pre- and post-operatively. Physical therapy exercises included
scapular mobilization, progressive range of motion, weight bearing, forced use with game-related activities to
encourage left UE volitional control, aerobic endurance exercise. Broadly, the therapeutic goal was to encourage
volitional control outside of the abnormal synergistic patterns while varying postural stability.

RESULTS

The participant underwent intracortical implantation in autumn of 2020 and explantation three months later on January 2021, in accordance with the intended 3-month duration of the trial. Over the course of the study, the participant had three minor, and one serious, device-related adverse events, all of which were treated, resolved, and reported to the governing regulatory bodies. The serious adverse event was the development of a scalp infection at the left pedestal site one week prior to the device removal date, despite a regimen of topical antibiotics and regular cleaning. This infection was anticipated and was described as a potential risk in the informed consent form and consent interview. The left pedestal site had posed a challenge since the time of the initial surgery as it was not possible to exactly re-approximate the skin flap leaving the base of the pedestal exposed. This area was protected and eventually granulated and grew new skin. The participant was afebrile and asymptomatic, and the infection was detected only by close visual inspection. The participant was treated with twice daily amoxicillin/clavulanic acid for the 7 days prior to the device removal. Pedestal site skin cultures taken at device removal revealed pansensitive *staphylococcus lugdunesis* and *staphylococcus capitis*, and yeast, and appropriate antimicrobial treatment was provided. No organisms grew from cultures taken of the bone adjacent to the pedestals. The only macroscopic evidence of infection at device removal was a small area (~ 2 cm^3) of erythema and friable tissue at the skin adjacent to the right pedestal. The participant was discharged home. The participant remains in the Cortimo trial both for ongoing neurosurgical follow-up and surveillance, and to track any further performance improvements in myoelectric MyoPro use with ongoing outpatient occupational therapy.

Preoperative anatomic and functional neuroimaging. Preoperative imaging revealed the old infarct in right lentiform nucleus and adjacent white matter including corona radiata and a portion of the posterior limb of the internal capsule, along with a large old right PCA infarct, progressed since the acute stroke imaging MRI from 2019 (Fig. 3). In addition, a small region of bandlike signal abnormality involving subcortical white matter and medial aspect of hand knob region of right precentral gyrus was identified, likely reflecting retrograde neuronal degeneration, (which influenced the clinical decision for the implant placement). On DTI, there was extensive loss of fractional anisotropy in the region of right corticospinal tract from old infarct. The "imagined" left hand motor paradigm and passive elbow/forearm motor paradigm were diagnostic with good concordance. Subsequent to hypercapnia challenge, a marked BOLD signal was evident at the precentral gyrus. The brief breath-holding presumably allowed relatively increased carbon dioxide retention (relative respiratory acidosis) that in turn induced mild cerebral vasodilatation facilitating BOLD signal acquisition (in this subject with underlying vascular abnormality as a cause of his multiple

Figure 3. Neuroimaging results. Diffusion sequence from MRI scan on January 19, 2019 when the acute stroke occurred; diffusion restriction is evident in the right lentiform nucleus and adjacent white matter (a). T2-weighted MRI from January 22, 2021, shows areas of encephalomalacia and relative ventriculomegaly that have developed since the stroke (b). Functional neuroimaging revealed a hot spot of activation, indicated by a red circle, in the depth of the central sulcus along the 'hand knob' area of the precentral gyrus (c). A three-dimensional reconstruction of the participant's cortical surface derived from the MRI with the imagined left hand movement centroid of activity indicated by the red circle (d). The green shading indicates an area responsive to sensory stimulation of the left hand. The black squares indicate the location of the four microelectrode arrays.

strokes). On the "imagined" left hand motor paradigm, activation was noted in the expected location along central sulcus involving lateral aspect of the hand knob region of the precentral gyrus and the adjacent portion of postcentral gyrus (Fig. 3c). On the passive left elbow/forearm motor paradigm, activation was seen along central sulcus which shows good concordance with the "imagined" motor task as discussed with a slightly more posterior and superior extension of activation reflecting the prominent sensory component of this passive motor paradigm. A 3D brain model

320 was printed using the 3D FLAIR sequence to allow for 3D visualization of the surgical field for more accurate pre-
321 operative planning (Fig. 3d).
322
Neural recordings. Well-delineated single units were recorded from 87 of the 256
channels (Fig. 4). Neural activity correlated with actual and attempted movements
in both the paretic left arm in addition to the intact right arm. The multi-electrode
arrays that were recorded via the right pedestal connector were more lateral and
posterior along the precentral gyrus and were closer to the centroid of fMRI
imagined movement activation (Fig. 3c) and closer in absolute terms to the
subcortical stroke. Although all four arrays were closely spaces along the
precentral gyrus, cross-array discharge properties appeared to vary amongst them
(Fig. 5). The discharge rate of various units appeared to correlate with specific
residual actions, including the wrist extension that gradually developed in the
course of the three-month duration (Fig. 6). By taking the spike counts recorded
at each channel every 200 milliseconds and running them through a leaky
integrator[21], and then summing these leaky integrator outputs across all channels,
we were able to visualize the cumulative cross-array firing rate activity in
comparison to forearm electromyographic activity (Fig. 7). Of the 256 electrodes,

Figure 4. Action potential waveforms. Snapshot of action potential waveforms recorded from two of the arrays.

338 in each session, we identified 20 to 40 channels that were eventually used for neural decoding. These channels were
339 used for extraction of neural features (whether threshold crossing or spectral high-frequency power) that coded for
340 hand and elbow flexion and extension. Two main decoding approaches were used: 1) A discrete two-class (hand
341 opening/closing) classifier based on a 1-dimensional linear filter continuous output; 2) a "training-less" threshold
342 crossing approach with a rolling baseline normalization, used to control hand closing or opening.
343
Orthosis control. The left upper extremity score on the
Action Research Arm Test (ARAT) [1919] was 0 without the
orthosis on, 5 using the orthosis under myoelectric control,
and 10 using the orthosis under direct brain-control. In one
component of the Jebsen-Taylor standardized test of hand
function [2222], the goal is to pick up and move 5 cans, one at
a time, a few inches away forward on the table (normal
times are 3.23 seconds for empty soup cans in subtest 6,
and 3.30 seconds for full cans in subtest 7). Because the
design of the hand orthosis precluded the ability to grasp a
soup can (i.e., the brace only supports the thumb and next
two digits such that not all digits can be used for grasp and
the fourth and fifth digit, since they remain tightly flexed,
in fact physically block the can), the participant performed
variations on the test. It took the participant 146 seconds to
pick up, move and release 5 pill bottles using myoelectric
control, and 95 seconds to perform the identical task under
BCI control. Another task was to hold an object in the right
hand (e.g., a stress ball or a whiteboard eraser) and place it
into the paretic left hand, and then extend the left arm down
towards the floor and drop the object into a bin; this
process was then repeated 5 times in a row. Both these
tasks were performed with the participant seated. On two
trials of this pick-up-and-drop-5 in myoelectric mode, the
participant's completion times were 128 and 222 sec/trial?;
seconds; on five trials of the same task in BCI mode, times
were 81, 106, 137, 214 seconds. In addition to measuring
the total time to perform these grasp-move-release tasks,

Figure 5. Spike rasters derived from both pedestals (four arrays) in a 200-second example clip. Recordings from the left pedestal (top panel) reflect activity recorded by the two arrays located more medially and dorsally along the precentral gyrus. Recordings from the right pedestal (bottom panel), reflect activity by the two arrays located more laterally along the precentral gyrus and closer to the stroke.

372 we also quantified the time it took to release an object once the hand was in the target position. Hand release times
373 were faster under BCI control than myoelectric control (*p=0.04*, two-sample t-test).

374
Motor outcomes. Motor measures, performed
when the participant was not connected to the
BCI or wearing the MyoPro orthosis, were
tracked over time and demonstrated that the
implantation procedure did not decrease
residual strength in the paretic left arm and
left leg. In fact, muscle strength increased in
the left arm. Whereas serial neurological
exams since the time of the stroke
demonstrated an absence of voluntary wrist
extension or finger extension (0/5 on manual
muscle testing, starting two months into the
trial, the participant began to consistently
exhibit voluntary wrist extension against
gravity (3/5), and on a few occasions was able
to voluntarily extend the fingers slightly (2/5)
[23].[23]. One month prior to the device
implantation, the Fugl-Meyer upper
extremity score was 30 (out of a maximum of
66) for the left upper extremity; this increased
to a score of 36 four weeks after the two
Multiports were implanted, and a score of 38
seven weeks post-implantation [24].[24].
Although the participant did not receive
botulinum toxin injections, or receive any
type of anti-spasticity medication, during the
clinical trial, spasticity gradually decreased
with time as reflected in gradually
decreasing numbers on serial
measurements of the modified
Ashworth scale for spasticity for
passive flexion and extension
movements of the fingers, wrist, and
elbow, along with internal and
external rotation of the shoulder [25][25].
410

411 DISCUSSION
412

This pilot trial demonstrated that
ensemble single unit activity remains
active in ipsilesional cerebral cortex
overlying chronic subcortical stroke.
To our knowledge, this is the first
report of intracortical recordings in
ipsilesional cerebral cortex for a
stroke above the mesencephalon. The
trial established that single neuron,
movement related activity can be
decoded used to control a powered
orthosis that restored functionally
useful voluntary upper extremity
movement. Importantly, this brain-
computer interface system can be
used simultaneously with residual

Figure 6. Neuronal activity correlated with performed movements in the paretic limb. Over a 110-s period, the participant was asked to perform a series of left limb movements (which are described on the abscissa). Movement instruction times, delivered as a verbal command, is indicated by a small vertical hash mark. Spiking activity of two simultaneously recorded units is displayed. Rasters indicate the time of each action potential. Normalized, integrated firing rates appear beneath each raster, as derived by the 'leaky integrator' equation described in [21]; normalization is achieved by dividing by the maximum integrated firing rate from each unit's spike train over the time period displayed. Although both units tend to discharge with wrist extension and hand squeezing, the top unit (channel 61) is relatively more active for hand squeezing than wrist extension, relative to the bottom, simultaneously recorded unit (channel 62). The participant performed all movements; however, such motions required effort and he was unable to engage a consistent level of activity for each cue, and also exhibited a variable reaction time to a given cue. The participant was easily fatigued, requiring him to take a break and then adjust his posture.

Figure 7. Cumulative, integrated spike activity across channels fluctuating with joint position (top) and residual left forearm electromyographic activity (bottom). The summed spike activity across channels and run through a leaky integrator [21] appeared to fluctuate with specific residual actions in the left upper extremity. More proximal residual activity in the biceps and triceps generated a more normal-appearing pattern: as seen between 290 and 310 seconds in the bottom panel, biceps (blue trace) and triceps (red trace) activity alternate. In the distal upper extremity, however, wrist flexor (green) and wrist extensor (yellow) trace, tend to occur together in an abnormal synergy; it is also evident that wrist flexor activity is likewise abnormally synergitisic with biceps activity (an abnormal flexor synergy). The summed, integrated spiking activity across channels appears to fluctuate most consistently with wrist flexor activity.

intact movement, in particular in a limb with a gradient of intact to absent voluntary movement, as is common
following cerebral strokes. While myoelectric approaches based upon wrist flexion did enable voluntary hand opening,
this approach triggered increased muscle tone that subsequently slowed orthosis use (as the motors were opposing the
abnormal tone): the BCI control mode essentially bypassed this issue and allowed motors to operate more smoothly
and quickly. Electromyographic recordings demonstrated that while the participant did continue to engage wrist
flexors during BCI control, the amplitude was decreased from abnormally elevated levels to more normal amplitudes.

This trial was not intended to restore voluntary motor control in the hemiparetic upper extremity in the absence of any
device use, but even so, we found that strength improved, and spasticity decreased. At the very least, this suggests that
the implantation of four microelectrode arrays into ipsilesional precentral gyrus cortex did not exacerbate pre-existing
hemiparesis (i.e., it did not worsen hand or arm weakness); indeed, after the intervention hand functions improved.
One potential explanation for the participant's unexpected improvements in voluntary wrist and finger extension
include mass practice (from regular physical and occupational therapy, with daily practice reinforced by clinical trial
assistants), Although the use of an extension arm splint 8 hours each night, and the frequent use of an electric blanket
to warm the paretic forearm, undoubtedly helped reduce spasticity and stretch flexor tendons, these interventions
cannot account for the benefits as they had also been in use for months prior to enrollment when serial measurements
showed no improvements in manual muscle testing. Another, more speculative, explanation for the participant's
improved forearm function is that the daily exercise of ipsilesional cortical activity for BCI-orthosis control, promoted
a plasticity driven response to either normalize or compensate for abnormal motor synergies.

Although the limited number of trials on various tasks reduced statistical power to compare myoelectric to BCI control,
qualitatively there appeared to be a trend of faster control in the BCI mode. We speculate that this may be due to the
fact that triggering orthosis action from direct cortical recordings does not activate abnormal forearm synergies in the
same manner that myoelectric control appears to. Post-stroke spasticity is thought to represent a form of abnormal
plasticity and loss of corticoreticular facilitation of the medullary inhibition center in turn leading to decreased
inhibition from the dorsal reticulospinal tract on the spinal stretch reflex: the medial reticulospinal and vestibulospinal
tracts are unopposed leading to stretch reflex hyperexcitability [26].[26] In the myoelectric mode, where hand closing is
triggered by activation of residual wrist flexors, this hyperexcitability is inevitably triggered such that the orthosis
motors have to 'fight harder' to open the hand, slowing that process. In the BCI mode, even if residual wrist flexor
and extensor activity are engaged, it is to lesser degree such that abnormal tone is not elevated, and the orthosis motors
can more easily and rapidly achieve hand actions.

This pilot study implies that usable control signals are present in ipsilesional cerebral cortical activity. To be clinically
scalable, future devices must be fully implantable to minimize infection risk and allow mobility. With the advent of
fully implantable BCI (i.e., no percutaneous connectors, [27,28 29]), a wider range of stroke survivors could benefit.
Although a fully implantable BCI could be coupled to an external wearable orthosis, an option that may gain even
wider clinical adoption would be to couple direct cortical control to implantable functional electrical stimulation in
the paretic arm, as has been demonstrated in at least one stroke survivor[10]. The combination of a fully implantable
brain sensor with a fully implantable peripheral neuromuscular stimulation hence could be a potential solution to
"unparalyze" the arm. Direct cortically driven peripheral muscular stimulation may have both rehabilitative [30] benefits
when used intermittently and direct functional (hand use) benefits if deployed continuously in daily life. Fully
implantable brain-computer interfaces may represent a new combined medical device opportunity to help stroke
survivors break through their plateau in recovery and to achieve greater functional independence.

This pilot study implies that usable control signals are present in ipsilesional cerebral cortical activity. To be clinically
scalable, future devices must be fully implantable to minimize infection risk. With the advent of fully implantable BCI
(i.e., no percutaneous connectors, [27,28 29]), a wider range of stroke survivors could benefit. Although a fully implantable
BCI could be coupled to an external wearable orthosis, an option that may gain even wider clinical adoption would be
to couple direct cortical control to implantable functional electrical stimulation in the paretic arm, as has been
demonstrated in at least one stroke survivor [17]. The combination of a fully implantable brain sensor with a fully
implantable peripheral neuromuscular stimulation hence could be a potential solution to "unparalyze" the arm. Direct
cortically driven peripheral muscular stimulation may have both rehabilitative [30] benefits when used intermittently
and direct functional benefits if deployed continuously in daily life. Fully implantable brain-computer interfaces may

represent a medical device opportunity to help stroke survivors break through their plateau in recovery and keep going
to achieve greater functional independence.
484
485   FUNDING
486
487   This research was supported by philanthropy to the Farber Institute of Neuroscience at Thomas Jefferson University.
488
489   ETHICS COMMITTEE APPROVAL
490
491   This research was approved, and remains in approved status, by the Institutional Review Board of Thomas Jefferson
492   University, protocol number 17D.459.
493
494   CONFLICT OF INTEREST STATEMENT
495
496   Drs. Serruya and Napoli are inventors on a US provisional patent application that has been filed by Thomas Jefferson
497   University on the methods described in this paper. All authors report that they do not have any conflicts of interest
498   with the research described.
499
500   ACKNOWLEDGEMENTS
501
502   The authors would like to thank the following people for their assistance and input: Erica Jones, Shivayogi Hiremath,
503   Christopher Thompson, Carlos Vargas-Irwin, John Donoghue, Nicholas Hatsopoulos, David Weisman, Kristofer
504   Feeko, M.J. Mulcahey, Joseph Tracy, Diana Tzeng, Daniel Graves, Ashly Parekh, Joely Mass, Thomas J. Kelly, IV,
505   Stephen Valverde, Allison Weiss, Shaista Alam, Robin Dharia, Elan Miller, Lisa Bowman, Rodney Bell, Michael
506   Sperling, and the participant and his mother.
507
508   REFERENCES
509

1.   Katan, M. & Luft, A. Global Burden of Stroke. *Semin. Neurol.* (2018). doi:10.1055/s-0038-1649503
2.   Vos, T. *et al.* Global, regional, and national incidence, prevalence, and years lived with disability for 310
diseases and injuries, 1990–2015: a systematic analysis for the Global Burden of Disease Study 2015.
Lancet (2016). doi:10.1016/S0140-6736(16)31678-6
3.   Centers for Disease Control. *Prevalence of stroke--United States, 2006-2010. MMWR. Morbidity and*
mortality weekly report **61**, (2012).
4.   Kwakkel, G., Kollen, B. J., Van der Grond, J. V. & Prevo, A. J. H. Probability of regaining dexterity in the
flaccid upper limb: Impact of severity of paresis and time since onset in acute stroke. *Stroke* (2003).
doi:10.1161/01.STR.0000087172.16305.CD
5.   Milekovic T, Sarma AA, Bacher D, Simeral JD, Saab J, Pandarinath C, S. B., Blabe C, Oakley EM, Tringale
KR, Eskandar E, Cash SS, Henderson JM, S. K. & Donoghue JP, H. L. Stable long-term BCI-enabled
communication in ALS and locked-in syndrome using LFP signals. *J Neurophysiol* (2018).
doi:10.1152/jn.00493.2017
6.   Klaes, C. *et al.* A cognitive neuroprosthetic that uses cortical stimulation for somatosensory feedback. *J.*
Neural Eng. **11**, 056024 (2014).
7.   Collinger, J. L. *et al.* High-performance neuroprosthetic control by an individual with tetraplegia. *Lancet*
381, 557–564 (2013).
8.   Bhagat, N. A. *et al.* Design and optimization of an EEG-based brain machine interface (BMI) to an upper-
limb exoskeleton for stroke survivors. *Front. Neurosci.* **10**, (2016).
9.   Page, S. J., Hill, V. & White, S. Portable upper extremity robotics is as efficacious as upper extremity
rehabilitative therapy: a randomized controlled pilot trial. *Clin. Rehabil.* **27**, 494–503 (2013).
10.   Knutson, J. S. *et al.* Implanted neuroprosthesis for assisting arm and hand function after stroke: A case
study. *J. Rehabil. Res. Dev.* **49**, 1505 (2012).
11.   Knutson, J. S. *et al.* Adding contralaterally controlled electrical stimulation of the triceps to contralaterally

controlled functional electrical stimulation of the finger extensors reduces upper limb impairment and
improves reachable workspace but not dexterity: A randomized . *Am. J. Phys. Med. Rehabil.* (2020).
doi:10.1097/PHM.0000000000001363
12. Wisneski, K. J. *et al.* Unique cortical physiology associated with ipsilateral hand movements and
neuroprosthetic implications. *Stroke* **39**, 3351–3359 (2008).
13. Knutson, J. S. *et al.* Implanted neuroprosthesis for assisting arm and hand function after stroke: a case study.
J. Rehabil. Res. Dev. **49**, 1505–16 (2012).
14. Bundy, D. T. *et al.* Using ipsilateral motor signals in the unaffected cerebral hemisphere as a signal platform
for brain-computer interfaces in hemiplegic stroke survivors. *J. Neural Eng.* **9**, 036011 (2012).
15. Vargas-Irwin, C. E. *et al.* Watch, imagine, attempt: Motor cortex single-unit activity reveals context-
dependent movement encoding in humans with tetraplegia. *Front. Hum. Neurosci.* (2018).
doi:10.3389/fnhum.2018.00450
16. Hochberg, L. R. *et al.* Reach and grasp by people with tetraplegia using a neurally controlled robotic arm.
Nature **485**, 372–375 (2012).
17. Suner, S., Fellows, M. R., Vargas-Irwin, C., Nakata, G. K. & Donoghue, J. P. Reliability of signals from a
chronically implanted, silicon-based electrode array in non-human primate primary motor cortex. *IEEE*
Trans. Neural Syst. Rehabil. Eng. **13**, 524–541 (2005).
18. Rousche, P. J. & Normann, R. A. A method for pneumatically inserting an array of penetrating electrodes
into cortical tissue. *Ann. Biomed. Eng.* (1992). doi:10.1007/BF02368133
19. Yozbatiran, N., Der-Yeghiaian, L. & Cramer, S. C. A standardized approach to performing the action
research arm test. *Neurorehabil. Neural Repair* (2008). doi:10.1177/1545968307305353
20. Taylor, N., Sand, P. L. & Jebsen, R. H. Evaluation of hand function in children. *Arch. Phys. Med. Rehabil.*
(1973).
21. Fetz, E. E. & Baker, M. A. Operantly conditioned patterns on precentral unit activity and correlated
responses in adjacent cells and contralateral muscles. *J. Neurophysiol.* (1973). doi:10.1152/jn.1973.36.2.179
22. Jebsen, R. H., Taylor, N., Trieschmann, R. B., Trotter, M. J. & Howard, L. A. An objective and standardized
test of hand function. *Arch. Phys. Med. Rehabil.* (1969).
23. Martin, E. G. & Lovett, R. W. A Method of Testing Muscular Strength in Infantile Paralysis. *J. Am. Med.*
Assoc. (1915). doi:10.1001/jama.1915.02580180016006
24. Fugl-Meyer, A. R., Jääskö, L., Leyman, I., Olsson, S. & Steglind, S. The post-stroke hemiplegic patient. 1. a
method for evaluation of physical performance. *Scand. J. Rehabil. Med.* (1975). doi:10.1038/35081184
25. Ashworth, B. Preliminary trial of carisoprrodol in multiple sclerosis. *Practitioner* (1964).
26. Li, S. Spasticity, motor recovery, and neural plasticity after stroke. *Frontiers in Neurology* (2017).
doi:10.3389/fneur.2017.00120
27. Yin, M. *et al.* Wireless neurosensor for full-spectrum electrophysiology recordings during free behavior.
Neuron (2014). doi:10.1016/j.neuron.2014.11.010
28. Musk, E. An Integrated Brain-Machine Interface Platform With Thousands of Channels. *J. Med. Internet*
Res. (2019). doi:10.2196/16194
29. Kohler, F. *et al.* Closed-loop interaction with the cerebral cortex: a review of wireless implant technology§.
Brain-Computer Interfaces (2017). doi:10.1080/2326263X.2017.1338011
30. Olsen, S. *et al.* Peripheral Electrical Stimulation Paired With Movement-Related Cortical Potentials
Improves Isometric Muscle Strength and Voluntary Activation Following Stroke. *Front. Hum. Neurosci.*
(2020). doi:10.3389/fnhum.2020.00156
577