Clinical Validation of Automated and Rapid mariPOC SARS-CoV-2 Antigen Test

Juha M. Koskinen,[1,2*] Petri Antikainen,[1] Kristina Hotakainen,[3,4] Anu Haveri, [5] Niina Ikonen,[5] Carita Savolainen-Kopra,[5] Kati Sundström,[6,7] Janne O. Koskinen[a]

[1]ArcDia International Ltd, Turku, Finland

[2]Faculty of Medicine, University of Turku, Turku, Finland

[3]Mehiläinen Oy, Helsinki, Finland

[4]Department of Clinical Chemistry and Haematology, University of Helsinki, Helsinki, Finland

[5]Finnish Institute for Health and Welfare, Helsinki, Finland

[6]SataDiag, Pori, Finland

[7]Department of Clinical Chemistry, Faculty of Medicine and Health Technology, Tampere University, Tampere, Finland

*Address correspondence to Juha Koskinen, juha.koskinen@arcdia.com

JMK and PA contributed equally to the manuscript.

Abstract Novel SARS coronavirus causing COVID-19 was recognized in late 2019. Diagnostics was quickly ramped up worldwide based on the detection of viral RNA. Based on the scientific knowledge for pre-existing coronaviruses, it was expected that the RNA of this novel coronavirus will be detected from symptomatic and at significant rates also from asymptomatic

individuals due to persistence of non-infectious RNA. To increase the efficacy of diagnostics,

surveillance, screening and pandemic control, rapid methods, such as antigen tests, are needed

for decentralized testing and to assess infectiousness. The objective was to validate the analytical

and clinical performance, and usability of a novel automated mariPOC SARS-CoV-2 test, which

is based on the detection of structural viral proteins using sophisticated optical laser technology.

Clinical performance of the test was evaluated against qRT-PCR with nasopharyngeal swab

specimens collected from patients suspected of acute SARS-CoV-2 infection. Sensitivity of the

mariPOC test was 100.0% (13/13) directly from swab specimens and 84.4% (38/45) from swab

specimens in undefined transport mediums. Specificity of the test was 100.0% (201/201). The

test's limit of detection was 2.7 $TCID_{50}$/test and had no cross-reactions with the tested respiratory

microbes. Our study shows that the mariPOC can detect infectious individuals already in 20

minutes with clinical sensitivity close to qRT-PCR. The test targets conserved epitopes of SARS-

CoV-2 nucleoprotein, making it robust against strain variations. The new test is a promising and

versatile tool for syndromic testing of symptomatic cases and for high capacity infection control

screening.

Keywords COVID-19, SARS-CoV-2, antigen detection, mariPOC, PCR, infectious, screening,

infection control, pandemic

Introduction

37   Emerging pandemic coronavirus (CoV) was recognized in Wuhan, China, in late 2019. The

38   virus, isolated from patients mentioned to be pneumonic, was quickly sequenced to share 79.6%

39   full length genome similarity with the Severe Acute Respiratory Syndrome virus (SARS-CoV-1)

40   and 91.2% similarity between its nucleocapsid (N) proteins (1). The novel SARS-CoV-2,

causing COVID-19, was identified to be circulating in horseshoe bats for decades similarly to SARS-CoV-1 (2). Diagnostic nucleic acid amplification tests (NAAT), mostly quantitative real-time polymerase chain reaction (qRT-PCR), were quickly developed worldwide, based on protocol provided for World Health Organization (3). Diagnostic qRT-PCR capacities were ramped up quickly in central laboratories because such tests are fast to develop for new targets. Most often, the new qRT-PCR tests were adopted for clinical diagnostics with minimal verification and validation against other diagnostic test methods.

For the seasonal coronaviruses, the interpretation of gene positivity in clinical specimens has been challenging since the viral RNA is detected at similar rates and qRT-PCR cycle threshold (Ct) values from symptomatic and asymptomatic individuals. The viral RNA is also co-detected with genomes of other respiratory viruses (4-7). This is also the case for the SARS-CoV-2 (8, 9). Moreover, recent scientific evidence indicates that qRT-PCR positivity has poor correlation for assessment of SARS-CoV-2 infectiousness (10-16). Whereas, Pekosz et al. (2020) showed that the detection of N-protein by an antigen test correlates with SARS-CoV-2 viral culture more accurately than qRT-PCR (13). Already half a decade ago Inagaki et al. (2016) unequivocally concluded for influenza that, "PCR...is not an appropriate method for indicating infectivity" and "the antigen-detection test estimated the infectious period with comparable if not better accuracy with culture" (17). In the case of COVID-19 diagnostics, the fact that viral RNA persistence can be detected without viable virus for months, has been a known clinical challenge, as diagnostics relied in the beginning of the pandemic solely on NAAT detection (18), the efficacy of which is in ruling out positivity.

The expression of N-protein, which is the key pathogenicity factor of coronaviruses (19), is essential for the coronavirus replication and transcription of the viral RNA (20, 21). Without the

accumulation of the N-protein, the coronaviral mRNA is degraded by the nonsense-mediated

decay (NMD) pathway of eukaryotic cells (19). Alexandersen at al. (2020) concluded that the

detection of RNA is not an indicator of actively replicating SARS-CoV-2. Their data suggests

that virion and subgenomic RNAs are stable in cellular double-membrane vesicles and, therefore,

can be detected long after the acute infection (22).

Shortly after viral exposure, viral concentration is low and qRT-PCR Ct values are high. When

the virus starts replication, it happens fast. In a cell model, extensive coronavirus RNA

transcription has occurred in 6 to 8 hours after the infection (23). In addition, NAATs being

prone for reporting clinically insignificant findings (analytically the detection may be correct,

there is viral RNA in the sample) they are prone to contaminations. A study of SARS-CoV-2

primer-probe sets from four major European suppliers found a significant level of contamination

from the reagents. False positives as low as qRT-PCR Ct 17 were obtained (24). Low levels of

SARS-CoV-2 RNA contamination has also been found from surfaces and air in rooms where

mildly ill individuals were isolated without notable viable virus (25, 26). It has also been shown

that environmental contamination may yield in positive test results in PCR among individuals

sampled in the same area where intranasal influenza vaccine dosing was done (27). These data

suggests that individuals having presence near symptomatic patients can be contaminated by

RNA without being infected with viable virus. Thus, methods detecting the viral RNA by

amplification are prone for clinically insignificant positive results, especially when significant

part of the population has been infected recently. The fact that a positive NAAT result is not a

reliable biomarker of active infection or COVID-19, is a true challenge for clinicians and

decision making for quarantine. It is not only that a missed necessary quarantine has health and

epidemic costs but also that a falsely imposed quarantine has social and financial consequences

(28).

The different performance requirements of diagnostic, surveillance and screening testing have

been recently discussed by Mina and Andersen (2020). There is a need for both super sensitive

PCR based tests and rapid and appropriately sensitive antigen tests to fight the COVID-19

pandemic (29). The use of the two methodologies should supplement one another in clinical

practice and pandemic fight.

In the present study, we analytically and clinically validated the performance of a novel 2nd

generation mariPOC SARS-CoV-2 test (ArcDia International Ltd, Finland), intended for rapid

and automated detection of viral acute phase proteins when there is a clinical suspicion of acute

COVID-19. Monoclonal antibodies of the test are designed to target a conserved epitope in the

N-protein, which is the most abundant protein in coronaviruses. We have previously shown that

the presence of coronavirus OC43 N-protein in the nasopharynx correlates with the respiratory

tract infection symptoms (30). It has been shown that clinical presentations of seasonal

coronavirus OC43 infections can be similar to those of coronaviruses that are considered as

severe viruses (SARS and MERS) (31).

The mariPOC is an automated platform for the rapid multianalyte testing of acute infectious

diseases. It is based on a separation-free two-photon excitation assay technique (32, 33).

Subsequently to nasopharyngeal sampling, the mariPOC test's operational steps are: combining

the swab with the sample buffer from a bottle-top dispenser, vortexing the sample tube to release

the sample from the swab (Fig. 1), followed by automated analysis and objective fluorescent

result read out. The analysis is performed by an automated analyzer with sophisticated

autoverification functions, and the result can be transferred automatically to the laboratory

information system or as anonymized epidemiological data (34) into mariCloud[TM] service. The

hands-on time is one minute per sample, and the analyzer works in continuous-feed and walk-

away mode. The mariPOC SARS-CoV-2 test is available as a single pathogen test and as part of

syndromic multianalyte tests covering, among others, influenza viruses. The throughput of one

mariPOC analyzer can be up to 300 single analyte tests or 100 multianalyte tests in 24 hours. The

results are reported in two phases, great majority of the infectious cases in twenty minutes and

very low positive and negative cases in 55 minutes or two hours, depending on the test

configuration.

Materials and methods

Analytical sensitivity

Nasopharyngeal swab specimens from asymptomatic individuals were pooled and suspended

into mariPOC RTI sample buffer into volume corresponding to 1.3 mL per swab. This pooled

clinical sample matrix (1.3 mL) was spiked with 75 µL of gamma-irradiation inactivated SARS-

CoV-2 cell lysate (stock 2.8×10^6 $TCID_{50}$/mL, USA-WA1/2020, NR-52287, lot 70035888, BEI

Resources, Manassas, VA, USA) in culture supernatant of different viral concentrations (50%

tissue culture infective dose, $TCID_{50}$). The samples were analyzed with the mariPOC SARS-

CoV-2 test following the manufacturer's instructions. Limit of Detection (LoD) was determined

as the lowest concentration giving at least 19 positives out of 20 replicates (≥95% positivity).

Cross-reactivity

128 Analytical specificity of the mariPOC SARS-CoV-2 test was studied by challenging the test

129 against relevant microbes commonly found in the nasal cavity (Online Resource). Briefly, the

130 microbe stocks were suspended in high concentration in the mariPOC RTI sample buffer and

131 analyzed with the mariPOC test.

Clinical specificity

Validation of the specificity of the mariPOC SARS-CoV-2 test was conducted in SataDiag

laboratory unit in Pori, Finland in February 2021 by one operator following the manufacturer's

instructions. In the study, 205 freshly sampled nasopharyngeal swab specimens were analyzed.

The samples were leftover samples from routine diagnostics with the 1st generation (launched

May 2020) mariPOC SARS-CoV-2 test. Samples positive in mariPOC were also analyzed with

the Xpert Xpress SARS-CoV-2 test (ref XPRSARS-COV2-10) detecting N2 and E-genes,

Cepheid, USA.

Clinical sensitivity

Sensitivity of the mariPOC SARS-CoV-2 test was validated with 58 frozen qRT-PCR positive

nasopharyngeal samples from two specimen cohorts.

Sample cohort 1

The first cohort consisted of 13 qRT-PCR positive nasopharyngeal swab samples collected from

patients (N=211) visiting primary healthcare COVID-19 drive-in stations of Mehiläinen Oy in

Helsinki capital area of Finland from March to April 2020. The qRT-PCR negative frozen

samples were not analyzed for this study because the test specificity was studied with fresh

samples as described above. The enrollment criteria were respiratory infection symptoms and

clinician's suspicion of COVID-19, the official criteria for COVID-19-testing in Finland, and at

the clinical study sites at the time of the study. The samples were taken with a flocked swab from

the nasopharynx (8 to 12 cm deep for adults and 4 to 8 cm deep for children) by rotating the

swab in nasopharyngeal cavity for 10 seconds. Two consecutive specimens were collected from

127 patients giving oral consent to participate in the study. The specimens were collected during

an internal laboratory method validation study, which does not require external ethical review

board permission and was not linked with recruitment or treatment of patients (35). The study

was approved by the responsible chief physician of Mehiläinen Oy, and by the responsible chief

physician and area manager of SataDiag laboratory division. The specimen for standard of care

testing was collected first. These specimens were analyzed after RNA extraction with Allplex^{TM}

2019-nCoV RT-PCR assay (Seegene Inc., Republic of Korea) at Seoul Clinical laboratories

(Republic of Korea). Allplex^{TM} 2019-nCoV RT-PCR assay detects E, N and RdRP genes (36).

The second swab specimen was kept in a dry tube at +4 °C for a maximum of 8 hours and stored

frozen until analysis with the mariPOC SARS-CoV-2 test. For 84 patients, one nasopharyngeal

swab specimen was collected. These swabs were suspended into saline (0.5−1 mL) and analyzed

with Amplidiag COVID-19 qRT-PCR assay including RNA extraction (Mobidiag Ltd, Finland)

at Vita Laboratorio Ltd (Finland). Amplidiag COVID-19 qRT-PCR assay detects N and ORF1ab

genes (37). The leftover saline specimens were stored frozen until analysis with the mariPOC

SARS-CoV-2 test.

The dry swab specimens and leftover saline samples were analyzed retrospectively with the

mariPOC test by two operators following the manufacturer's instructions. Briefly, dry swabs

were suspended into 1.3 mL of the mariPOC RTI sample buffer and the leftover saline samples

(range 0.1−0.65 mL) were diluted with mariPOC RTI sample buffer to a final volume of 1.3 mL.

For this validation study, the samples were further diluted into the mariPOC RTI sample buffer

in one-to-one (1:1) ratio and analyzed with the mariPOC test. When a discrepant result between

the mariPOC and comparator RT-PCR was obtained, the swab samples taken for the mariPOC

were confirmatory tested at the Department of Clinical Microbiology, Turku University Hospital

(Finland) with an in-house reference qRT-PCR detecting E, N and RdRP genes (3).

<u>Sample cohort 2</u>

The second cohort consisted of forty five positive pseudonymized specimens with known qRT-

PCR Ct values (16 to 34) from the frozen nasopharyngeal swab specimen library of Finnish

Institute of Health and Welfare, Helsinki, Finland, in undefined transport mediums. The qRT-

PCR protocol was an in-house method based on the primers and probes by Corman et al. (2020)

(3). The cohort consisted mostly of symptomatic, but in part also of asymptomatic subjects,

while the detailed information for each subject was not available for this study. The specimens

were in either reddish (N=22) or colorless (N=23) solutions. The samples were diluted into the

mariPOC RTI sample buffer in one-to-one (1:1) ratio and analyzed with the mariPOC test. The

positivity rate of the mariPOC test was evaluated against different qRT-PCR Ct value categories

and then compared to published viral culture studies.

Results

Analytical sensitivity

LoD of the mariPOC SARS-CoV-2 test was 2.7 $TCID_{50}$/test in 20 µL reaction volume, at which

all twenty replicates gave a positive test result. Based on the certificate of analysis of the viral

preparation, the LoD equals to 1690 genome equivalents per test.

Cross-reactivity

The mariPOC test had no cross-reactions. The mariPOC test did not cross-react with seasonal

coronaviruses OC43, 229E, or NL63, but it detected the recombinant N-protein of SARS-CoV-1.

Detailed information is shown in supplemental material (Online Resource).

Clinical specificity

Specificity of the mariPOC test was 100.0% in the preliminary (200/200) and final (201/201)

result reporting phases. Three and two of the sample analyseswere rejected by the

autoverification in the preliminary and final result reporting phases giving failure rates of 1.5%

and 1.0%, respectively. Two of the samples were positive in both the mariPOC test and routine

PCR test.

Clinical sensitivity

The sensitivity of the mariPOC SARS-CoV-2 test was 100.0% (13/13) in the preliminary and

final result reporting phases in sensitivity cohort 1, where the nasopharyngeal swab specimens

were suspended directly into the mariPOC sample buffer or first into saline (Table 1). Detailed

description of the positive samples are shown in supplemental material (Online Resource).

Prevalence of SARS-CoV-2 in the first sample cohort was 6%, which is well in alignment with

the prevalence during the study time in the geographical area (5%).

The sensitivity of the test in sensitivity cohort 2 was 73.3% (33/45) and 84.4% (38/45) in the

preliminary and final result reporting phases, respectively, when the nasopharyngeal swabs were

initially suspended in undefined transport mediums and further diluted with the mariPOC sample

buffer (Table 1). Based on 95% confidence interval, both cohorts had similar statistical reliability

214   (Table 1). Overall, 38 out of 45 samples were positive with the mariPOC in sensitivity cohort 2

215   (Fig. 2). The test showed 100% (31/31) positivity rate compared to qRT-PCR for Ct values ≤28

216   (Table 2). Above the Ct value 28, the positivity rate of the mariPOC declined as typical for an

217   antigen test, reaching 91.9% (34/37) with Ct values ≤30.

Discussion

When setting up a diagnostic process or choosing a diagnostic method, one should carefully

consider, to start with, whether the disease, clinical condition and use case, require high

sensitivity for ruling out or high specificity for ruling in. There is a need for both in fighting the

COVID-19 pandemic. In general, analytically highly sensitive testing, such as PCR testing, is

good at ruling out a disease (e.g. keeping a ward clean) while highly specific testing, such as

antigen testing, is good at ruling in a disease (e.g. acute infection diagnostics and assessing

infectiousness). Because of rapidity and lesser logistic challenges compared to central lab testing,

antigen testing is particularly good in surveillance, field-testing, screening of masses, cohorting

of inpatients, acute disease diagnostics, and in assessing the infectiousness of individuals (13,

38). Especially when disease prevalence is low, clinical specificity of the screening and

diagnostic testing should be emphasized to keep unnecessary quarantines and economic damages

at minimum while still allowing sufficient enough infection control (29).

According to scientific data, to effectively prevent spread of the disease, pandemic control

should prioritize accessibility, frequency of testing, and rapid sample-to-answer time over test

sensitivity (37, 39, 40). Viral load and probability to infect others is highest just prior to onset of

symptoms and during the symptomatic phase (41).

We describe here analytical and clinical validation of mariPOC SARS-CoV-2 test sensitivity and

specificity. Determination of LoD was performed with gamma-irradiation inactivated viral

culture supernatant and showed that only less than three infectious units per test was needed for

positive test result. LoD was also determined as qRT-PCR Ct units using UV-inactivated virus

(Online Resource). The obtained Ct LoD was 33, which approaches the theoretical analytical

sensitivity of a typical PCR method with 5 µl cDNA volume and applying 95% confidence

interval. The maximum Ct values detected in clinical samples (Fig. 2 and Table 2.) were similar

to the determined Ct LoD.

Based on high identity (89.1%) between SARS-CoV-1 (Uniprot entry, P59595) and SARS-

CoV-2 (UniProt entry, P0DTC9) nucleocapsid protein sequences, and obtaining positive result

for purified SARS-CoV-1 nucleocapsid protein in cross-reactivity testing (Table S1 in the Online

Resource), it is highly likely that the mariPOC test detects also the SARS-CoV-1 virus itself.

Cross-reactions were not observed. A minor limitation of the study is that cross-reactivity for

MERS coronavirus and coronavirus HKU1 were assessed using purified protein and sequence

analysis (Online Resource) and not with clinical samples or cultured virus.

Our sensitivity validation cohort 1 showed 100% sensitivity. While the highest qRT-PCR Ct

obtained by the primary reference test in this cohort was 30.2, the cohort consisted of unselected

and consecutive samples collected from patients with clear symptoms. This might explain why

there were no samples with higher Ct values. The sensitivity cohort 2 showed 100% positivity

rate for the mariPOC below qRT-PCR Ct 28 (Table 2 and Fig. 2). This result was excellent

taking into account that the samples were unfavorable for the mariPOC platform with separation-

free fluorescent measurement. Colorful transport media are not recommended for mariPOC

testing since they elevate fluorescent signal levels (42) and unnecessarily dilute the samples,

which reduces sensitivity. The pooled sensitivity of sample cohorts 1 and 2 at qRT-PCR Ct ≤ 30

was 94%, suggesting even higher sensitivity for mariPOC compared to what has been reported in

the literature for SARS-CoV-2 viral culture against qRT-PCR, as summarized in Table 2. In

addition, our results are in line with at least two other N-protein detecting tests that were

evaluated against RT-PCR and culture (43, 44). Several studies have shown that infectivity of

SARS-CoV-2 declines rapidly in samples showing qRT-PCR Ct above 25, and viable virus is

rarely isolated after 8 days from onset of the symptoms. The detection of sole viral RNA,

especially at low levels without the detectable level of viral N-protein or culture positivity, is a

questionable marker of acute infection and infectiousness (10-14, 16, 45-47).

The results suggest that the clinical sensitivity of the mariPOC test (84.4% in unfavorable sample

matrix to 100.0% when used according to manufacturer recommendations) is similar or even

better than that of at least some rapid RT-PCR tests (93.4%), (37) when symptomatic pati ents

suspected with acute COVID-19 infection are tested within the first five days of symptoms and

prevalence among tested samples is reasonable (6% in sensitivity study 1). Recommended

sample in the mariPOC test is native nasopharyngeal swab specimen suspended into 1.3 mL of

the RTI sample buffer. Other specimen types may yield in lower apparent sensitivity. In the

sensitivity cohort 1, suspending part of the swabs first into saline prior to the addition of

mariPOC RTI sample buffer followed by a further dilution into mariPOC RTI sample buffer by a

factor of two for the testing, diluted the specimens 4 to 20 times (2 to 4.3 PCR Ct units) more

than the recommended sample pretreatment. Additional dilution lowers the sensitivity compared

to the recommended protocol, and could have led to an underestimation of the test sensitivity.

Strengths of the sensitivity validation included that the specimens were collected in the early

phase of the COVID-19 pandemic in Finland that minimized the detection of RNA persistence

with the RT-PCR among the cohort population.

Limitations of the validation study were that the patient characteristics and the number of

symptomatic days before sampling were not available for the study. Freezing and thawing of the

positive samples prior to mariPOC testing and additional dilutions to the recommended protocol

were also limitations. However, if any, these could have had a negative effect on the mariPOC

test sensitivity and, hence, the study at least did not overestimate the sensitivity of the mariPOC

SARS-CoV-2 test.

288   Conclusions

The mariPOC SARS-CoV-2 test is an automated, highly specific and clinically accurate test with

rapid sample-to-answer time for individuals with clinical suspicion and in acute phase of an

infection. The closed tube test system and the design of operational steps minimize specimen

handling and possible exposure of user to infectious material. The multianalyte syndromic tests

help to differentiate between SARS-CoV-2 and other viruses, such as influenza. The single

analyte test provides high capacity of 300 samples a day at the point-of-sampling. Objective

result read-out and LIS connectability minimize manual work and human errors. Our study

together with other scientific data suggests that the mariPOC can detect majority of the cases

already in 20 minutes with sensitivity similar to a rapid PCR while maximum sensitivity is

achieved in 55 minutes. The positivity rate of mariPOC compared to qRT-PCR Ct values in

clinical samples is very high up to Ct 28-30, and samples at least up to Ct 33.24 (Fig. 2) or 35.78

(Online Resourcetable 2) can be detected depending on the PCR method and gene target. The

detection of conserved epitope in the N-protein of SARS coronaviruses with the mariPOC likely

provides accurate information about infectiousness similarly to other antigen tests and viral

culture and suggests ability to detect also emerging virus variants. Further studies using viral

culture as comparative method and follow-up of infectiousness of patients using antigen

detection are needed in order to optimize viral respiratory tract infection management.

Supplemental material Supplemental material is available as Online Resource.

Online Resource1, PDF file, 0.4 MB.

Funding ArcDia International Ltd contributed the study with the mariPOC test system and

consumables. The study was partly supported by Business Finland, the Finnish Funding Agency

for Innovation, under the project reference 35239/31/2020.

Conflicts of interest JMK, PA and JOK are employees at ArcDia International Ltd.

Acknowledgements We thank Vita laboratories Ltd and Seoul Clinical Laboratory for

performing RT-PCR analyses and Titta Lehtonen performing the mariPOC analyses in SataDiag.

Compliance with ethical standards

Author contributions

JMK and PA: Major contributions in the development of the mariPOC test, in scientific design

and execution of the studies, result analysis, scientific analysis, and writing of the manuscript.

KH: Major contributions in providing specimens for clinical validation, in scientific analysis and

revising the manuscript.

AH: Major contributions in providing specimens for clinical validation, in scientific analysis and

revising the manuscript.

NI and CS-K: Major contributions in scientific analysis, and revision of the manuscript.

KS: Major contributions in providing specimens for clinical validation, in scientific analysis and

revising the manuscript.

JOK: Major contributions in the development of the mariPOC test and in scientific design,

execution, and analysis of the results, and revision of the manuscript.

References

1. Zhou P, Yang X, Wang X, Hu B, Zhang L, Zhang W, Si H, Zhu Y, Li B, Huang C, Chen H, Chen J,

Luo Y, Guo H, Jiang R, Liu M, Chen Y, Shen X, Wang X, Zheng X, Zhao K, Chen Q, Deng F, Liu L,

Yan B, Zhan F, Wang Y, Xiao G, Shi Z (2020) A pneumonia outbreak associated with a new coronavirus

of probable bat origin. Nature 579:270-273. https://doi.org/10.1038/s41586-020-2012-7

2. Boni MF, Lemey P, Jiang X, Lam TT, Perry BW, Castoe TA, Rambaut A, Robertson DL (2020)

Evolutionary origins of the SARS-CoV-2 sarbecovirus lineage responsible for the COVID-19 pandemic.

Nat Microbiol 5:1408-1417. https://doi.org/10.1038/s41564-020-0771-4

3. Corman VM, Landt O, Kaiser M, Molenkamp R, Meijer A, Chu DK, Bleicker T, Brünink S, Schneider

J, Schmidt ML, Mulders DG, Haagmans BL, van der Veer B, van den Brink S, Wijsman L, Goderski G,

Romette J, Ellis J, Zambon M, Peiris M, Goossens H, Reusken C, Koopmans MP, Drosten C (2020)

Detection of 2019 novel coronavirus (2019-nCoV) by real-time RT-PCR. Euro Surveill 25.

https://dx.doi.org/10.2807%2F1560-7917.ES.2020.25.3.2000045

4. Gaunt ER, Hardie A, Claas ECJ, Simmonds P, Templeton KE (2010) Epidemiology and clinical

presentations of the four human coronaviruses 229E, HKU1, NL63, and OC43 detected over 3 years

using a novel multiplex real-time PCR method. J Clin Microbiol 48:2940-2947.

https://doi.org/10.1128/jcm.00636-10

5. Prill MM, Iwane MK, Edwards KM, Williams JV, Weinberg GA, Staat MA, Willby MJ, Talbot HK,

Hall CB, Szilagyi PG, Griffin MR, Curns AT, Erdman DD (2012) Human coronavirus in young children

hospitalized for acute respiratory illness and asymptomatic controls. Pediatr Infect Dis J 31:235-240.

https://doi.org/10.1097/inf.0b013e31823e07fe

6. Rhedin S, Lindstrand A, Rotzén-Östlund M, Tolfvenstam T, Ohrmalm L, Rinder MR, Zweygberg-

Wirgart B, Ortqvist A, Henriques-Normark B, Broliden K, Naucler P (2014) Clinical utility of PCR for

common viruses in acute respiratory illness. Pediatrics 133:538. https://doi.org/10.1542/peds.2013-3042

7. Rhedin S, Lindstrand A, Hjelmgren A, Ryd-Rinder M, Öhrmalm L, Tolfvenstam T, Örtqvist Å,

Rotzén-Östlund M, Zweygberg-Wirgart B, Henriques-Normark B, Broliden K, Naucler P (2015)

Respiratory viruses associated with community-acquired pneumonia in children: matched case-control

study. Thorax 70:847-853. https://doi.org/10.1136/thoraxjnl-2015-206933

8. Chau NVV, Thanh Lam V, Thanh Dung N, Yen LM, Minh NNQ, Hung LM, Ngoc NM, Dung NT,

Man DNH, Nguyet LA, Nhat LTH, Nhu LNT, Ny NTH, Hong NTT, Kestelyn E, Dung NTP, Xuan TC,

Hien TT, Thanh Phong N, Tu TNH, Geskus RB, Thanh TT, Thanh Truong N, Binh NT, Thuong TC,

Thwaites G, Tan LV (2020) The natural history and transmission potential of asymptomatic SARS-CoV-

2 infection. Clin Infect Dis 71:2679-2687. https://doi.org/10.1093/cid/ciaa711

9. Kissler SM, Fauver JR, Mack C, Tai C, Shiue KY, Kalinich CC, Jednak S, Ott IM, Vogels CBF,

Wohlgemuth J, Weisberger J, DiFiori J, J. Anderson D, Mancell J, Ho DD, Grubaugh ND, Grad YH

(2020) Viral dynamics of SARS-CoV-2 infection and the predictive value of repeat testing. medRxiv.

https://doi.org/10.1101/2020.10.21.20217042

10. Arons MM, Hatfield KM, Reddy SC, Kimball A, James A, Jacobs JR, Taylor J, Spicer K, Bardossy

AC, Oakley LP, Tanwar S, Dyal JW, Harney J, Chisty Z, Bell JM, Methner M, Paul P, Carlson CM,

McLaughlin HP, Thornburg N, Tong S, Tamin A, Tao Y, Uehara A, Harcourt J, Clark S, Brostrom-Smith

C, Page LC, Kay M, Lewis J, Montgomery P, Stone ND, Clark TA, Honein MA, Duchin JS, Jernigan JA

(2020) Presymptomatic SARS-CoV-2 Infections and Transmission in a Skilled Nursing Facility. N Engl J

Med 382:2081-2090. https://doi.org/10.1056/nejmoa2008457

11. Bullard J, Dust K, Funk D, Strong JE, Alexander D, Garnett L, Boodman C, Bello A, Hedley A,

Schiffman Z, Doan K, Bastien N, Li Y, Van Caeseele PG, Poliquin G (2020) Predicting infectious SARS-

CoV-2 from diagnostic samples. Clin Infect Dis 71:2663–2666. https://doi.org/10.1093/cid/ciaa638

12. La Scola B, Le Bideau M, Andreani J, Hoang VT, Grimaldier C, Colson P, Gautret P, Raoult D

(2020) Viral RNA load as determined by cell culture as a management tool for discharge of SARS-CoV-2

patients from infectious disease wards. Eur J Clin Microbiol Infect Dis 39:1059-1061.

https://doi.org/10.1007/s10096-020-03913-9

13. Pekosz A, Cooper CK, Parvu V, Li M, Andrews JC, Manabe YC, Kodsi S, Leitch J, Gary DS, Roger-

Dalbert C (2020) Antigen-based testing but not real-time PCR correlates with SARS-CoV-2 virus culture.

Clin Infect Dis:ciaa1706. https://doi.org/10.1093/cid/ciaa1706

14. Singanayagam A, Patel M, Charlett A, Lopez Bernal J, Saliba V, Ellis J, Ladhani S, Zambon M,

Gopal R (2020) Duration of infectiousness and correlation with RT-PCR cycle threshold values in cases

of COVID-19, England, January to May 2020. Euro Surveill 25. https://doi.org/10.2807/1560-

7917.es.2020.25.32.2001483

15. Cao S, Gan Y, Wang C, Bachmann M, Wei S, Gong J, Huang Y, Wang T, Li L, Lu K, Jiang H, Gong

Y, Xu H, Shen X, Tian Q, Lv C, Song F, Yin X, Lu Z (2020) Post-lockdown SARS-CoV-2 nucleic acid

screening in nearly ten million residents of Wuhan, China. Nature Communications 11:1-7.

https://doi.org/10.1038/s41467-020-19802-w

16. Jaafar R, Aherfi S, Wurtz N, Grimaldier C, Van Hoang T, Colson P, Raoult D, La Scola B (2020)

Correlation Between 3790 Quantitative Polymerase Chain Reaction–Positives Samples and Positive Cell

Cultures, Including 1941 Severe Acute Respiratory Syndrome Coronavirus 2 Isolates. Clinical Infectious

Diseases. https://doi.org/10.1093/cid/ciaa1491

17. Inagaki K, Song M, Crumpton J, DeBeauchamp J, Jeevan T, Tuomanen EI, Webby RJ, Hakim H

(2016) Correlation Between the Interval of Influenza Virus Infectivity and Results of Diagnostic Assays

in a Ferret Model. J Infect Dis 213:407-410. https://doi.org/10.1093/infdis/jiv331

18. Pollock AM, Lancaster J (2020) Asymptomatic transmission of covid-19. BMJ 371:m4851.

http://dx.doi.org/10.1136/bmj.m4851

19. Wada M, Lokugamage KG, Nakagawa K, Narayanan K, Makino S (2018) Interplay between

coronavirus, a cytoplasmic RNA virus, and nonsense-mediated mRNA decay pathway. Proc Natl Acad

Sci U S A 115:E10157-E10166. https://doi.org/10.1073/pnas.1811675115

20. Almazán F, Galán C, Enjuanes L (2004) The nucleoprotein is required for efficient coronavirus

genome replication. J Virol 78:12683-12688. https://doi.org/10.1128/jvi.78.22.12683-12688.2004

21. Zúñiga S, Cruz JLG, Sola I, Mateos-Gómez PA, Palacio L, Enjuanes L (2010) Coronavirus

nucleocapsid protein facilitates template switching and is required for efficient transcription. J Virol

84:2169-2175. https://doi.org/10.1128/jvi.02011-09

22. Alexandersen S, Chamings A, Bhatta TR (2020) SARS-CoV-2 genomic and subgenomic RNAs in

diagnostic samples are not an indicator of active replication. Nat Commun 11:6059.

https://doi.org/10.1038/s41467-020-19883-7

23. Hofmann MA, Sethna PB, Brian DA (1990) Bovine coronavirus mRNA replication continues

throughout persistent infection in cell culture. J Virol 64:4108-4114.

https://doi.org/10.1128/jvi.64.9.4108-4114.1990

24. Wernike K, Keller M, Conraths FJ, Mettenleiter TC, Groschup MH, Beer M (2020) Pitfalls in SARS-

CoV-2 PCR diagnostics. Transbound Emerg Dis. https://doi.org/10.1111/tbed.13684

25. Santarpia JL, Rivera DN, Herrera VL, Morwitzer MJ, Creager HM, Santarpia GW, Crown KK, Brett-

Major DM, Schnaubelt ER, Broadhurst MJ, Lawler JV, Reid SP, Lowe JJ (2020) Aerosol and surface

contamination of SARS-CoV-2 observed in quarantine and isolation care. Sci Rep 10:12732.

https://doi.org/10.1038/s41598-020-69286-3

26. Zhou J, Otter JA, Price JR, Cimpeanu C, Garcia DM, Kinross J, Boshier PR, Mason S, Bolt F,

Holmes AH, Barclay WS (2020) Investigating SARS-CoV-2 surface and air contamination in an acute

healthcare setting during the peak of the COVID-19 pandemic in London. Clin Infect Dis.

https://doi.org/10.1093/cid/ciaa905

27. Curran T, McCaughey C, Ellis J, Mitchell SJ, Feeney SA, Watt AP, Mitchell F, Fairley D, Crawford

L, McKenna J, Coyle PV (2012) False-positive PCR results linked to administration of seasonal influenza

vaccine. J Med Microbiol 61:332-338. https://doi.org/10.1099/jmm.0.036178-0

28. Surkova E, Nikolayevskyy V, Drobniewski F (2020) False-positive COVID-19 results: hidden

problems and costs. The lancet respiratory medicine 8:1167-1168. http://dx.doi.org/10.1016/S2213-

2600(20)30453-7

29. Mina MJ, Andersen KG (2020) COVID-19 testing: One size does not fit all. Science 371:126-127.

https://doi.org/10.1126/science.abe9187

30. Bruning AHL, Aatola H, Toivola H, Ikonen N, Savolainen-Kopra C, Blomqvist S, Pajkrt D, Wolthers

KC, Koskinen JO (2018) Rapid detection and monitoring of human coronavirus infections. New

Microbes New Infect 24:52-55. https://doi.org/10.1016/j.nmni.2018.04.007

31. Vandroux D, Allou N, Jabot J, Li Pat Yuen G, Brottet E, Roquebert B, Martinet O (2018) Intensive

care admission for Coronavirus OC43 respiratory tract infections. Med Mal Infect 48:141-144.

https://doi.org/10.1016/j.medmal.2018.01.001

32. Hänninen P, Soini A, Meltola N, Soini J, Soukka J, Soini E (2000) A new microvolume technique for

bioaffinity assays using two-photon excitation. Nat Biotechnol 18:548-550. https://doi.org/10.1038/75421

33. Koskinen JO, Vainionpää R, Meltola NJ, Soukka J, Hänninen PE, Soini AE (2007) Rapid method for

detection of influenza a and B virus antigens by use of a two-photon excitation assay technique and dry-

chemistry reagents. J Clin Microbiol 45:3581-3588. https://doi.org/10.1128/JCM.00128-07

34. Gunell M, Antikainen P, Porjo N, Irjala K, Vakkila J, Hotakainen K, Kaukoranta SS, Hirvonen JJ,

Saha K, Manninen R, Forsblom B, Rantakokko-Jalava K, Peltola V, Koskinen JO, Huovinen P (2016)

Comprehensive real-time epidemiological data from respiratory infections in Finland between 2010 and

2014 obtained from an automated and multianalyte mariPOC® respiratory pathogen test. Eur J Clin

Microbiol Infect Dis 35:405-413. https://doi.org/10.1007/s10096-015-2553-0

35. Vakkila J, Koskinen JO, Brandt A, Muotiala A, Liukko V, Soittu S, Meriluoto S, Vesalainen M,

Huovinen P, Irjala K (2015) Detection of Group A Streptococcus from Pharyngeal Swab Samples by

Bacterial Culture Is Challenged by a Novel mariPOC Point-of-Care Test. Journal of Clinical

Microbiology 53:2079-2083. https://doi.org/10.1128/jcm.00018-15

36. Liotti FM, Menchinelli G, Marchetti S, Morandotti GA, Sanguinetti M, Posteraro B, Cattani P (2020)

Evaluation of three commercial assays for SARS-CoV-2 molecular detection in upper respiratory tract

samples. European Journal of Clinical Microbiology & Infectious Diseases: Official Publication of the

European Society of Clinical Microbiology 40:269-277. https://doi.org/10.1007/s10096-020-04025-0

37. Jokela P, Jääskeläinen AE, Jarva H, Holma T, Ahava MJ, Mannonen L, Lappalainen M, Kurkela S,

Loginov R (2020) SARS-CoV-2 sample-to-answer nucleic acid testing in a tertiary care emergency

department: evaluation and utility. J Clin Virol 131:104614. https://doi.org/10.1016/j.jcv.2020.104614

38. Guglielmi G (2020) Fast coronavirus tests: what they can and can't do. Nature 585:496-498.

https://doi.org/10.1038/d41586-020-02661-2

39. Larremore DB, Wilder B, Lester E, Shehata S, Burke JM, Hay JA, Tambe M, Mina MJ, Parker R

(2020) Test sensitivity is secondary to frequency and turnaround time for COVID-19 screening. Sci Adv

7:eabd5393. https://doi.org/10.1126/sciadv.abd5393

40. Mina MJ, Parker R, Larremore DB (2020) Rethinking Covid-19 Test Sensitivity - A Strategy for

Containment. N Engl J Med 383:e120. https://doi.org/10.1056/nejmp2025631

41. He X, Lau EHY, Wu P, Deng X, Wang J, Hao X, Lau YC, Wong JY, Guan Y, Tan X, Mo X, Chen Y,

Liao B, Chen W, Hu F, Zhang Q, Zhong M, Wu Y, Zhao L, Zhang F, Cowling BJ, Li F, Leung GM

(2020) Temporal dynamics in viral shedding and transmissibility of COVID-19. Nat Med 26:672-675.

https://doi.org/10.1038/s41591-020-0869-5

42. Koskinen JM, Soukka JM, Meltola NJ, Koskinen JO (2018) Microbial identification from feces and

urine in one step by two-photon excitation assay technique. Journal of Immunological Methods 460:113-

118. https://doi.org/10.1016/j.jim.2018.06.017

43. Albert E, Torres I, Bueno F, Huntley D, Molla E, Fernández-Fuentes MÁ, Martínez M, Poujois S,

Forqué L, Valdivia A, Solano de la Asunción, Carlos, Ferrer J, Colomina J, Navarro D (2020) Field

evaluation of a rapid antigen test (Panbio™ COVID-19 Ag Rapid Test Device) for COVID-19 diagnosis

in primary healthcare centres. Clin Microbiol Infect. https://doi.org/10.1016/j.cmi.2020.11.004

44. Rusanen J, Kareinen L, Szirovicza L, Uğurlu H, Levanov L, Jääskeläinen A, Ahava M, Kurkela S,

Saksela K, Hedman K, Vapalahti O, Hepojoki J (2020) A generic, scalable, and rapid TR-FRET –based

assay for SARS-CoV-2 antigen detection. medRxiv. https://doi.org/10.1101/2020.12.07.20245167

45. Wölfel R, Corman VM, Guggemos W, Seilmaier M, Zange S, Müller MA, Niemeyer D, Jones TC,

Vollmar P, Rothe C, Hoelscher M, Bleicker T, Brünink S, Schneider J, Ehmann R, Zwirglmaier K,

Drosten C, Wendtner C (2020) Virological assessment of hospitalized patients with COVID-2019. Nature

581:465-469. https://doi.org/10.1038/s41586-020-2196-x

46. Sohn Y, Jeong SJ, Chung WS, Hyun JH, Baek YJ, Cho Y, Kim JH, Ahn JY, Choi JY, Yeom J (2020)

Assessing Viral Shedding and Infectivity of Asymptomatic or Mildly Symptomatic Patients with COVID-

19 in a Later Phase. J Clin Med 9:2924. https://doi.org/10.3390/jcm9092924

47. Basile K, McPhie K, Carter I, Alderson S, Rahman H, Donovan L, Kumar S, Tran T, Ko D, Sivaruban

T, Ngo C, Toi C, O'Sullivan MV, Sintchenko V, Chen SC-, Maddocks S, Dwyer DE, Kok J (2020) Cell-

based culture of SARS-CoV-2 informs infectivity and safe de-isolation assessments during COVID-19.

Clin Infect Dis:ciaa1579. https://doi.org/10.1093/cid/ciaa1579

Table 1 Sensitivity of the mariPOC SARS-CoV-2 test when compared with the qRT-PCR methods.

Sample cohort	mariPOC result phase	No. of specimens		Sensitivity (%) (95% CI)
		TP	FN	
1	Final	13	0	100.0 (75.3– 100.0)
	Preliminary	13	0	100.0 (75.3– 100.0)
2	Final	38	7	84.4 (70.5–93.5)
	Preliminary	33	12	73.3 (58.1–85.4)

489   TP = true positive, FN = false negative, CI = confidence interval (exact Clopper-Pearson method).

490 **Table 2 Comparison of** cumulative positivity rates of viral culture (four studies) and mariPOC

491 (sensitivity sample cohort 2) to qRT-PCR below different Ct categories. (10, 12, 14, 47)

	Arons	La Scola	Singanayagam	Basile	mariPOC Preliminary	mariPOC Final
Ct < 25	87%	83%	85%	89%	100.0% (20/20)	100.0% (20/20)
Ct < 28	77%	77%	NA	86%	90.3% (28/31)	100.0% (31/31)
Ct < 30	67%	75%	74%	82%	83.8% (31/37)	91.9% (34/37)
Overall	62%	68%	30%	NA	73.3% (33/45)	84.4% (38/45)

492 NA = not applicable

493
494 **Fig. 1** Sample pretreatment in the mariPOC SARS-CoV-2 test before placing the sample tube into the

495 analyzer for automated analysis and subsequent test result read out objectively as positive or negative.

497
498 **Fig. 2** Ct values of qRT-PCR for mariPOC test positive (green dots) and negative (red dots) samples in

499 the validation sample cohort 2. Dashed line is at Ct 33.24, which was the lowest detected Ct.

1 Add the specimen   **2** Add the buffer   **3** Vortex