Practical strategies for SARS-CoV-2 RT-PCR testing in resource-constrained settings

Meredith S. Muller [1]*, Srijana Bhattarai Chhetri[1], Christopher Basham[1], Tyler Rapp[1], Feng-Chang

Lin[3], Kelly Lin[1], Daniel Westreich[3], Carla Cerami[2], Jonathan J. Juliano[1], Jessica T. Li[1]

1 Institute of Global Health and Infectious Diseases, University of North Carolina School of Medicine,

Chapel Hill, NC USA

2 MRC Unit The Gambia at the London School of Hygiene & Tropical Medicine

3 Gillings School of Global Public Health, University of North Carolina, Chapel Hill, NC USA

*Corresponding author: meredith_muller@med.unc.edu 111 Mason Farm Road, 2336 -Medical

Biomolecular Research Building, Chapel Hill, NC 27514, Telephone: 919-843-7254 Fax: 919-445-0857

Running Title: SARS-CoV-2 testing strategies

Abstract word count: 150

Manuscript word count: 3007

ABSTRACT

Alternatives to nasopharyngeal sampling are needed to increase capacity for SARS-CoV-2

testing. Among 275 participants, we piloted the collection of nasal mid-turbinate swabs

amenable to self-testing, including polyester flocked swabs as well as 3D printed plastic lattice

swabs, placed into viral transport media or an RNA stabilization agent. Flocked nasal swabs

identified 104/121 individuals who were PCR-positive for SARS-CoV-2 by nasopharyngeal

sampling (sensitivity 87%, 95% CI 79-92%), mostly missing those with low viral load ($<10^3$ viral

copies/uL). 3D-printed nasal swabs showed similar sensitivity. When nasal swabs were placed

directly into RNA preservative, the mean 1.4 log decrease in viral copies/uL compared to

nasopharyngeal samples was reduced to <1 log, even when samples were left at room

temperature for up to 7 days. We also evaluated pooling strategies that involved pooling

specimens in the lab versus pooling swabs at the point of collection, finding both successfully

detected samples $>10^2$ viral copies/uL.

KEYWORDS

SARS-CoV-2, nasal swab, qRT-PCR, pooling, viral transport medium

1. INTRODUCTION

Since the start of the SARS-CoV-2 pandemic, testing has been a cornerstone of the public health

response. The de facto standard for clinical testing is PCR from nasopharyngeal (NP) swabs.

However, nasopharyngeal sampling must be performed by trained staff using personal

protective equipment (PPE). Shortages in both, as well as NP swabs themselves, often manifest

when case counts climb. A wide array of strategies amenable to self-collection have been

piloted to expand testing capacity, including the collection of nasal swabs, oropharyngeal and

tongue swabs, saliva, and oral rinses [1-3]. The volume of tests conducted can also become

burdensome and lengthen turnaround time, spurring interest in pooled testing strategies in low

prevalence and settings [4-8]. Finally, regarding sample storage and transport, viral specimens

are typically placed in viral transport medium, and CDC recommends maintenance of cold chain

prior to processing [1], but this may not be possible in all settings.

In order to implement a household transmission study in the early phases of the epidemic in

North Carolina, when shortages of PPE and swabs were prevalent, we adopted a strategy of

self-collected nasal swabs from household members during follow-up. Here we compare this

strategy to concurrently collected nasopharyngeal swabs at enrollment in our study population.

We piloted different types of swabs stored in different media. Given interest in pooling

strategies for high throughput testing, we also used our cohort to test two different pooling

strategies: pooling swabs at the point of collection or pooling sample lysate in the lab. Our

findings provide confidence in using self-collected nasal swabs, preferably stored in an RNA

stabilizer, when nasopharyngeal sampling is not feasible.

2. METHODS

2.1 Clinical samples

Clinical samples were collected as part of a SARS-CoV-2 household transmission study

conducted in the Piedmont region of North Carolina. The study received ethical approval from

the Institutional Review Board at the University of North Carolina-Chapel Hill and is registered

at clinicaltrials.gov (NCT04445233). Participants were enrolled if they were adults that tested

positive for SARS-CoV-2 by PCR at the UNC Respiratory Diagnostic Center and shared a living

space with one or more persons who also agreed to participate. At enrollment, a standard

clinician-collected nasopharyngeal (NP) swab was performed, followed by up to two other nasal

swabs (on different sides) that were either collected by study staff or self-collected by the

participant or their guardian with guidance from study staff (**Figure S1**). For nasal sampling,

participants were instructed to insert the swab about 1-2 inches into one nostril , then swirl 5

times while slowly withdrawing the swab before placing it into the collection tube. All samples

were placed into a cooler on ice prior to processing in a BSL2+ laboratory space.

2.2 Sample collection strategies

Flocked NP swabs were collected into 3mL of Becton Dickinson's co-packaged universal viral

transport system. Two types of nasal swabs designed for mid-turbinate sampling (NMT) were

used: flocked NMT swabs (COPAN, Murrietta CA) and 3D-printed lattice NMT swabs (Resolution

Medical, Fridley MN) (**Figures S1**). Both were collected into 3mL viral transport media (VTM)

prepared using CDC SOP# DSR-052-05. Upon sample receipt in the laboratory, 1mL of the

collected sample was combined with 1mL 2X DNA/RNA Shield, a nucleic acid preservation agent

and lysis buffer (Zymo Research), and stored at -80°C until extraction. RNA was extracted from

200uL of the lysate using the Quick-RNA Viral 96 Kit (Zymo Research) and eluted in 20uL of

water. We also evaluated the effect of storage media by collecting flocked NMT swabs directly

into 3mL of 1X DNA/RNA Shield (Shield), with aliquots either frozen immediately upon return to

the lab or left at room temperature for 4 or 7 days before being stored at -80°C. RNA was

extracted from 100 uL of the lysate using the same extraction and elution protocols.

2.3 qRT-PCR viral quantification

Samples were tested using a CDC RT-qPCR protocol authorized for emergency use that consists

of three unique assays: two targeting regions of the virus' nucleocapsid gene (N1, N2) and one

targeting human RNase P gene (RP) (Catalog # 2019-nCoVEUA-01, Integrated DNA

Technologies) [9]. 5uL of extracted RNA was added to 15uL of each assay's reaction mixture

containing TaqPath 1-Step RT-qPCR Master Mix, CG (Thermofisher Scientific) and the

corresponding primer-probe set (IDT), followed by the recommended thermocycler protocol.

Plasmid DNA containing the human RPP30 gene and SARS-CoV-2 in vitro transcribed RNA

control (nCoVPC, IDT) were used as positive controls. Water was used as a negative extraction

control. Samples were designated positive if all three PCRs were positive (N1 and N2 for virus,

RP for adequate sampling). If the N1 and N2 PCRs were negative, but the RP assay had a Ct

value ≥30 or was negative, suggesting inadequate sampling, then the sample was re-extracted.

The second result was reported if the RP Ct value was <30 or if both N1 and N2 PCRs were

positive regardless of RP Ct value.

The viral load of each sample, in copies/uL, was extrapolated from standard curves generated

for each viral assay (N1 and N2) using serial dilutions of nCoVPC (2 to 100,000 viral RNA

copies/uL). The average copies/uL between the N1 and N2 assays was used as the final

quantitative viral load. Based on the sample collection and RNA extraction volumes as well as

volume of template RNA used in the RT-qPCR (5uL), this viral load represents the number of

viral RNA copies per 5 uL of VTM or Shield sample.

2.4 Pooling strategies

The efficacy of pooling NMT samples was examined through two different approaches: pooling

swabs at the point of care into the same collection vessel and pooling individual sample lysates

prior to extraction. For the first strategy, self-collected 3D-printed lattice NMT swabs from each

member of a household of three or more were collected and pooled together in 5mL of VTM.

This was done at one or more of the study visits for each household. 200uL of the sample lysate

was extracted and quantified as above. Results were compared to the self-collected individual

flocked NMT swab collected at the same visit. In the second pooling strategy, one qRT-PCR

positive sample lysate from a flocked NMT swab (pre-RNA extraction) was pooled with sample

lysate from negative individuals to construct pool sizes of 5, 10, 15, and 20. The Ct values of

twelve samples with viral copies/uL ranging from 10^1 to 10^7 were compared to the Ct values of

their corresponding pools.

2.5 Statistical analysis

A probit analysis of results from the nCoVPC plasmid control concentrations (ranging from 2 to

100,000 copies/uL as part of standard curves generated in every RT-qPCR run) by parametric

curve fitting to hit rate data was used to determine the limit of detection (LOD) of the N1 and

N2 qRT-PCR assays. Samples that were positive in both N1 and N2 assays, but with an average

viral load that fell below the LOD were categorized as indeterminate. The sensitivity and

specificity of different swab types for RT-qPCR detection of SARS-CoV-2 was calculated using

flocked NP swabs as the reference standard. Additionally, the difference in the quantitative

viral load was compared for different collection strategies. Comparisons were made on the log

scale and analyzed using Wilcoxon matched-pairs signed rank testing with a p-value<0.05

considered significant. Statistical analyses were performed using GraphPad Prism 8 and SAS 9.4

(Cary, NC).

3. RESULTS

We report data from 644 swab samples collected from 275 participants (91 households) at

enrollment, 24 pools collected at follow-up or enrollment, and 44 pools constructed from

participant samples in the lab. Participants ranged in age from 1-77 years old, with 71% >18

years of age.

3.1 Limit of detection of RT-qPCR assay

Probit analysis of nCoVPC plasmid control concentrations tested in 33 RT-qPCR runs yielded a limit of detection (LOD) for the N1 and N2 assays of 9 and 13 copies/uL, respectively (**Table S1**). The average LOD between the two assays, 11 copies/uL, was used as the cutoff for sample positivity. A sample was deemed positive if the average viral load derived from the cycle threshold (Ct) values of N1 and N2 corresponded to a concentration ≥11 copies/uL, indeterminate if <11 copies/uL, and negative if either assay failed to amplify. Altogether, 21/702 (3.0%) samples tested fell into the indeterminate category. Another 33 (4.7%) samples only amplified in one assay (N1 or N2 assay), but with a Ct value corresponding to a viral load that fell below the LOD. Only 2 samples (0.3%) were discordant between the N1 and N2 assays (positive in one but not the other).

3.2 Comparison of collection swabs and storage medium

Compared to nasopharyngeal sampling, flocked nasal mid-turbinate (NMT) swabs displayed slightly decreased sensitivity, but were well-accepted by the participants and yielded adequate sampling. Altogether, at enrollment, 275 study participants completed 226 NP swabs and 418 NMT swabs (255 flocked and 51 3D-printed in VTM, 112 flocked in Shield) (**Figure S1**). Of the 49 participants that declined to do NP swabs,

Figure 1. Concordance and comparison of SARS-CoV-2 viral loads from paired nasopharyngeal (NP) and nasal mid-turbinate (NMT) swabs. Paired NP and NMT swabs from 173 participants showed overall good concordance, with most discordances (15/16) arising from positive NP/ negative NMT samples. Quantititative viral loads derived from the average of N1 and N2 qRT-PCR assays favored NP swabs compared to NMT swabs. A y=x dashed line is drawn for reference.

46 agreed to at least one type of NMT swab. Inadequate sampling, as defined by negative N1

and N2 PCRs in concert with a negative human RP PCR or Ct ≥30, occurred in small numbers of

flocked swabs, but a substantial proportion of 3D-printed plastic lattice swabs: 1/226 (0.4%) of

NP swabs, 14/343 (4.1%) of flocked NMT swabs, and 11/51 (21.6%) of 3D-printed plastic lattice

swabs.

Using NP swabs as the reference standard, flocked NMT swabs showed excellent specificity

(98%, 95% CI 90-100%) but slightly decreased sensitivity (87%, 95% CI 79-92%) for SARS-CoV-2

detection by RT-qPCR (**Figures 1&2**). Of 173 NP-NMT swab pairs, 104 were both positive, 52

both
negative, and
10% (17/173)
were
discordant.
Three of
these
discordances
were likely
due to
inadequate
sampling (1
NP, 2 NMT
swabs with
RP Ct value
≥30), while
71% of the
rest (10/14)
occurred in
samples with
low viral
loads (<10³

A

	NP +	NP -
NMT VTM +	104	1
NMT VTM -	16	52

Sensitivity 87% (95% CI, 79 - 92%)
Specificity 98% (95% CI, 90 - 100%)
N = 185 swab pairs, (12 indeterminate swabs)

	NP +	NP -
3D VTM +	23	0
3D VTM -	1	8

Sensitivity 96% (95% CI, 79 - 100%)
Specificity 100 (95% CI, 63 - 100%)
N = 32 swab pairs (0 indeterminate swabs)

	NMT VTM +	NMT VTM -
3D VTM+	24	4
3D VTM-	4	16

Sensitivity 86% (95% CI, 67 - 96%)
Specificity 80 (95% CI, 56 - 94%)
N = 48 swab pairs (3 Indeterminate swabs)

B

	NP +	NP -
NMT D1 shield +	19	0
NMT D1 shield -	3	16

Sensitivity 86.4% (95% CI, 65 - 97%)
Specificity 100% (95% CI, 79 - 100%)
N = 40 pairs (2 indeterminate samples)

	NP +	NP -
NMT D4 shield +	29	0
NMT D4 shield -	7	18

Sensitivity 81% (95% CI, 64 - 92%)
Specificity 100% (95% CI, 82 - 100%)
N = 56 pairs (2 Indeterminate samples)

	NP +	NP -
NMT D7 shield +	40	0
NMT D7 shield -	3	23

Sensitivity 93% (95% CI, 81 - 99%)
Specificity 100% (95% CI, 85 - 100%)
N = 68 pairs (2 indeterminate samples)

Figure 2. Concordance of SARS-CoV-2 RT-qPCR detection between nasopharyngeal (NP) swabs and two different nasal mid-turbinate (NMT) swab types, stored in viral transport media (VTM) or 1x DNA/RNA shield (Shield). In (A), sensitivity and specificity of standard flocked (NMT) or 3D-printed (3D) nasal swabs collected into VTM are shown using NP swabs with co-packaged universal viral transport system as the reference standard. Concordance of flocked vs. 3D nasal swabs is also shown. In (B), flocked NMT swabs were stored in Shield, and sample aliquots were directly frozen on day 1 (D1) or kept at room temperature before being stored at -80C on day 4 (D4) or day 7 (D7). Note that samples with indeterminate viral load (<11 copies/ul) were not included in the sensitivity/specificity analyses.

viral copies detected in the NP swab). In the 104 positive swab pairs, NMT samples displayed

lower average viral loads (Spearman correlation coefficient=0.67, **Figure 1**), with a mean 1.3 log

decrease in viral copies/uL (IQR 0.6 - 2.1 log viral copies/ul) compared to NP sampling

(p<0.0001) (**Figure 3A**). This was at least partly due to a sampling difference, as NMT swabs also

showed on average 3.1 cycles higher Ct values in the human RP PCR (**Figure S2**).

Though the
3D-printed
plastic
lattice NMT
swabs were
more likely
to lead to
inadequate
sampling,
positive
samples
showed
quantitative
viral loads
similar to
flocked NMT
swabs

Figure 3. Comparison of SARS-CoV-2 viral loads between standard NP swabs and NMT or 3D NMT swabs (A) as well as NMT swabs collected into 1x DNA/RNA Shield and stored for different intervals (B). The distribution of the difference in log viral load is depicted for each comparison. Median log-fold changes are indicated by a solid line with interquartile values indicated by dotted lines. The number of sample pairs is indicated for each comparison.

(**Figure 3A**). This was true despite on average 1.2 higher Ct values for the human RP assay in the

3D vs. flocked swabs. Compared to NP sampling, 3D-printed NMT swabs displayed 95.7%

sensitivity (95% CI 78.1%-99.9%) and 100% specificity (95% CI 63.1-100%) among 48 swab pairs

(**Figure 2**).

Placing flocked NMT swabs directly into 1x DNA/RNA Shield did not improve the sensitivity of

detection, but did result in viral loads comparable to those obtained by NP sampling. Aliquots

of Shield samples were either directly stored at -80C (similar to other samples collected on day

1), or left out at room temperature for 4 or 7 days prior to freezing and processing. All NMT

Shield samples showed a specificity of 100% compared to NP swabs, while sensitivity ranged

86%, 78%, and 91% for the samples frozen at day 1, 4, and 7, respectively (**Figure 2**).

Altogether, regardless of how many days the Shield samples were left out, the overall

sensitivity was 85% (95% CI 77-92%). While sensitivity for detection was slightly diminished,

quantitative viral loads derived from NMT Shield aliquots frozen on day 1 were comparable to

NP viral loads (mean decrease of 0.5 log viral copies/uL (IQR -0.3-1.4), p=0.09) (**Figure 3B**). For

aliquots left at room temperature until day 4 and day 7, we observed a mean decrease that was

still <1 log viral copies/uL compared to NP sampling (mean 0.8 and 0.8 log viral copies/ul,

respectively (p=0.001 and p=0.0002) (**Figure 3B).**

3.3 Pooling strategies

The pooling strategies

implemented were sufficient for

detecting samples with viral

loads >10^2 copies/uL but were

not as sensitive as individual

swabs for detecting samples

with lower viral loads. Of the 24

pools of 3D-NMT swabs pooled

at the point of care, 3 were

indeterminate, and 2 (8%)

yielded discordant results

(depicted as red stars in **Figure**

4). Under the assumption that

the concurrently collected

individual flocked NMT swabs

were accurate, the two

Figure 4. Comparison of Ct values from nasal mid-turbinate (NMT) swabs pooled from households of 3-5 persons at the point of care vs. concurrently collected individual NMT swabs. Among the pools collected from 24 households (listed along the x axis in order of decreasing viral loads), 2 pools with discordant results from individual swabs are depicted as red stars. Viral loads derived from the Ct values for each sample and the corresponding pool are found in Table S2.

discordant results were false

negative pools where the

individual swab had a viral

load <100 copies/uL, close to

the limit of detection (**Table**

S2). Of the 22 concordant

pools, 8 were negative and 11

were positive, mostly with

individual swab viral loads

≥10^2 copies/uL.

Similarly, when individual

sample lysates were pooled

in the lab at varying pool

Figure 5. Ct values of increasing pool sizes constructed containing a single positive sample with varying viral loads. Viral transport media from a single positive sample with a viral load ranging from 10^1 to 10^7 were used to construct pool sizes of 5, 10, 15, 20.

sizes, none of the 2 sets of experimental pools containing a sample with a viral load of 10^1

copies/uL were positive (**Figure 5**). Of the 3 sets of pools containing a sample with a viral load of

10^2 copies/uL, 2 were positive at every pool size, while the remaining set was positive within

pools of 5 and 10 samples, but indeterminate when the pool size was increased to 15 and 20

samples. The remaining pools constructed with samples with a viral load >10^2 copies/uL were

positive across all pool sizes. The average total Ct value increase for the pools that remained

positive at a pool size of 20 samples was 5.1 cycles, close to the expected 4.3 cycle increase for

a sample diluted 1:20 using a PCR with 100% amplification efficiency.

4. DISCUSSION

In a highly exposed outpatient cohort, we found nasal swabs to be reasonably sensitive,

capturing 87% of SARS-CoV-2 infections diagnosed by nasopharyngeal sampling. This estimate

is similar to most other outpatient studies showing >85% concordance between self-collected

nasal swabs (either nasal mid-turbinate or anterior nasal swabs) and clinician-collected

nasopharyngeal sampling [2, 10-13]. Not all studies are consistent however, likely due to

heterogeneity in testing environments, and inclusion of non-acute samples collected during

follow-up [14,15].

By calculating quantitative SARS-CoV-2 viral loads, our study gives clarity on where sensitivity is

diminished [16]. For the majority of participants in which nasal sampling failed to detect virus,

the NP viral load was <1000 copies/uL, at a level that is likely non-infectious. Of these

participants, 7/11 were antibody-positive at the time of sampling (unpublished data), and for

the 8/11 participants still reporting symptoms, the average duration of reported symptoms was

6.5 days. Thus nasal samples are likely adequate for clinical diagnosis of acute infections to help

expand testing capacity, but insensitivity to low viral load infections should be taken into

consideration. On average, the decreased sensitivity of NMT swabs led to a little over a log

decrease in viral copies/uL compared to NP swabs.

Our pragmatic approach of "show one, then do one" meant that nasal swabs were both

clinician and self-collected. Also, since we often collected two nasal swabs per person, one from

each nostril, our sampling strategy may have slightly underperformed relative to other studies

that sample both nostrils with the same swab. It should be noted that we tested flocked and

3D-printed lattice swabs, but did not test dry swabs or non-flocked cotton swabs. Where high

quality swabs are not available, but other swab types are plentiful, a strategy of combining oral

and nasal samples appears promising [17].

3D-printed plastic swabs may also help address supply chain shortages [18, 19]. We first

acquired prototype NMT lattice swabs from Resolution Medical in anticipation of shortage of

supplies for our research study. In our limited testing, the prototype 3D-printed NMT lattice

swabs showed high categorical concordance with NP swabs and also yielded similar viral loads

compared to flocked NMT swabs. Similar high concordance has been demonstrated for 3D-

printed nasopharyngeal swabs [18-20]. Anecdotally, the prototype 3D-printed were observed to

be more uncomfortable for study participants compared to flocked NMT swabs, a sentiment

shared by other studies [18]. This may have contributed to the higher proportion of samples

deemed as inadequate sampling.

Labs also face VTM shortages requiring alternate transport media [21, 22]. Reagents which can

inactivate virus and also keep samples stable at ambient temperature may be particularly apt

substitutes [23]. We used 1xDNA/RNA shield (Zymogen), an RNA preservation agent that has

been widely used to inactivate SARS-CoV-2 and other respiratory viruses in various sample types and is now part of saliva and NMT Shield collection kits that have received FDA emergency use authorization [24-26]. In our hands, storage of nasal swab samples in Shield did not improve their overall diagnostic sensitivity, but positive NMT swabs stored in Shield maintained quantitative viral loads more similar to those detected in concurrently collected NP swabs.

Pooling specimens in the lab is a well-documented strategy to accelerate SARS-CoV-2 testing in high-throughput settings [4-6]. As in previous studies, we found that although Ct values do increase with pooling, the strategy can be broadly successful [27-31]. Samples with viral loads at or near the limit of detection [31], or $<10^3$ viral copies/uL in the CDC EUA assay we adopted, may go undetected as pool sizes increase. This was even more apparent when pooling swabs at the point of collection, which we piloted as unsupervised self-collection of 3D-printed swabs into the same conical tube containing 5mL of VTM.

4.1 Conclusions

Our findings add to the evidence base for nasal swabs as an adequate substitute for PCR-based clinical diagnosis of SARS-CoV-2 infection in outpatient settings where nasopharyngeal sampling is challenging. Viral recovery can be maintained even when immediate cold chain is not possible by storing swabs in an RNA preservation agent that also deactivates infectious virus. Combined with pooling specimens in the lab, these practical strategies can help expand testing in resource-constrained settings.

5. FUNDING AND ACKNOWLEDGEMENTS

This work was supported by funds from the UNC Department of Medicine, the UNC COVID-19 Response Fund/Health Foundation, and the National Center for Advancing Translational Sciences (NCATS), National Institutes of Health (UL1TR002489). We thank the Co-HOST study team, the IDEEL lab, as well as our wonderful study participants. DW was supported by the North Carolina Policy Collaboratory at UNC with funding from the North Carolina Coronavirus Relief Fund established and appropriated by the North Carolina General Assembly. Prototype 3D-printed NMT swabs were provided by Resolution Medical.

6. REFERENCES

1. CDC. 2020. Information for Laboratories about Coronavirus (COVID-19).

2. Tu Y-P, Jennings R, Hart B, Cangelosi GA, Wood RC, Wehber K, Verma P, Vojta D, Berke EM. 2020. Swabs Collected by Patients or Health Care Workers for SARS-CoV-2 Testing. N Engl J Med 383:494–496.

3. Goldfarb DM, Tilley P, Al-Rawahi GN, Srigley J, Ford G, Pedersen H, Pabbi A, Hannam-Clark S, Charles MK, Dittrick M, Gadkar VJ, Pernica JM, Hoang L. 2020. Unpublished Results. Self-collected Saline Gargle Samples as an Alternative to Healthcare Worker Collected Nasopharyngeal Swabs for COVID-19 Diagnosis in Outpatients. medRxiv 2020.09.13.20188334.

4. Lohse S, Pfuhl T, Berkó-Göttel B, Rissland J, Geißler T, Gärtner B, Becker SL, Schneitler S, Smola S. 2020. Pooling of samples for testing for SARS-CoV-2 in asymptomatic people. Lancet Infect Dis 20:1231–1232.

5. Mutesa L, Ndishimye P, Butera Y, Souopgui J, Uwineza A, Rutayisire R, Ndoricimpaye EL, Musoni E, Rujeni N, Nyatanyi T, Ntagwabira E, Semakula M, Musanabaganwa C, Nyamwasa D, Ndashimye M, Ujeneza E, Mwikarago IE, Muvunyi CM, Mazarati JB, Nsanzimana S, Turok N, Ndifon W. 2020. A pooled testing strategy for identifying SARS-CoV-2 at low prevalence. Nature https://doi.org/10.1038/s41586-020-2885-5.

6. Pilcher CD, Westreich D, Hudgens MG. 2020. Group Testing for Severe Acute Respiratory Syndrome– Coronavirus 2 to Enable Rapid Scale-up of Testing and Real-Time Surveillance of Incidence. The Journal of Infectious Diseases.

7. Majid F, Omer SB, Khwaja AI. 2020. Optimising SARS-CoV-2 pooled testing for low-resource settings. Lancet Microbe 1:e101–e102.

8. Hogan CA, Sahoo MK, Pinsky BA. 2020. Sample Pooling as a Strategy to Detect Community Transmission of SARS-CoV-2. JAMA 323:1967–1969.

9. for Disease Control C, Prevention, Others. 2020. CDC 2019-novel coronavirus (2019-nCoV) real-time RT-PCR diagnostic panel. Revis Biol Celular 3:30.

10. Hanson KE, Barker AP, Hillyard DR, Gilmore N, Barrett JW, Orlandi RR, Shakir SM. 2020.

Self-Collected Anterior Nasal and Saliva Specimens versus Health Care Worker-Collected

Nasopharyngeal Swabs for the Molecular Detection of SARS-CoV-2. J Clin Microbiol 58.

11.   Griesemer SB, Van Slyke G, Ehrbar D, Strle K, Yildirim T, Centurioni DA, Walsh AC, Chang

AK, Waxman MJ, St. George K. 2020. Evaluation of specimen types and saliva stabilization

solutions for SARS-CoV-2 testing. medRxiv 2020.06.16.20133041.

12.   Péré H, Podglajen I, Wack M, Flamarion E, Mirault T, Goudot G, Hauw-Berlemont C, Le L,

Caudron E, Carrabin S, Rodary J, Ribeyre T, Bélec L, Veyer D. 2020. Nasal Swab Sampling for

SARS-CoV-2: a Convenient Alternative in Times of Nasopharyngeal Swab Shortage. J Clin

Microbiol 58.

13.   Kojima N, Turner F, Slepnev V, Bacelar A, Deming L, Kodeboyina S, Klausner JD. 2020. Self-

Collected Oral Fluid and Nasal Swabs Demonstrate Comparable Sensitivity to Clinician

Collected Nasopharyngeal Swabs for Covid-19 Detection. medrxiv;2020.04.11.20062372v1.

Infectious Diseases (except HIV/AIDS). medRxiv.

14.   Comber L, Walsh KA, Jordan K, O'Brien KK, Clyne B, Teljeur C, Drummond L, Carty PG, De

Gascun CF, Smith SM, Harrington P, Ryan M, O'Neill M. 2020. Alternative clinical

specimens for the detection of SARS-CoV-2: A rapid review. Rev Med Virol 1:e200448.

15.   Pinninti S, Trieu C, Pati SK, Latting M, Cooper J, Seleme MC, Boppana S, Arora N, Britt WJ,

Boppana SB. 2020. Comparing Nasopharyngeal and Mid-Turbinate Nasal Swab Testing for

the Identification of SARS-CoV-2. Clin Infect Dis https://doi.org/10.1093/cid/ciaa882.

16.   McCulloch DJ, Kim AE, Wilcox NC, Logue JK, Greninger AL, Englund JA, Chu HY. 2020.

Comparison of Unsupervised Home Self-collected Midnasal Swabs With Clinician-Collected

Nasopharyngeal Swabs for Detection of SARS-CoV-2 Infection. JAMA Netw Open

3:e2016382.

17.   Shakir SM, Barker AP, Hillyard DR, Gilmore N, Barrett JW, Orlandi RR, Hanson KE. 2020.

Combined Self-collected Anterior Nasal and Oropharyngeal Specimens versus Provider-

collected Nasopharyngeal Swabs for the Detection of SARS-CoV-2. J Clin Microbiol

https://doi.org/10.1128/JCM.02291-20.

18.   Callahan CJ, Lee R, Zulauf KE, Tamburello L, Smith KP, Previtera J, Cheng A, Green A, Abdul

Azim A, Yano A, Doraiswami N, Kirby JE, Arnaout RA. 2020. Open Development and Clinical

425   Validation of Multiple 3D-Printed Nasopharyngeal Collection Swabs: Rapid Resolution of a
426   Critical COVID-19 Testing Bottleneck. J Clin Microbiol 58.

427   19.   Alghounaim M, Almazeedi S, Al Youha S, Papenburg J, Alowaish O, AbdulHussain G, Al-
428   Shemali R, Albuloushi A, Alzabin S, Al-Wogayan K, Al-Mutawa Y, Al-Sabah S. 2020. Low-
429   Cost Polyester-Tipped Three-Dimensionally Printed Nasopharyngeal Swab for the
430   Detection of Severe Acute Respiratory Syndrome-Related Coronavirus 2 (SARS-CoV-2). J
431   Clin Microbiol 58.

432   20.   Bennett I, Bulterys PL, Chang M, DeSimone JM, Fralick J, Herring M, Kabaria H, Kong C,
433   Larson B, Lu O, Maxner A, Meyer E, Patterson S, Pollack S, Rolland J, Schmidt S, Seshadri S,
434   Swarup K, Thomas C, Van Wert R. 2020. The Rapid Deployment of a 3D Printed Latticed
435   Nasopharyngeal Swab for COVID-19 Testing Made Using Digital Light Synthesis.
436   medrxiv;2020.05.25.20112201v1. Infectious Diseases (except HIV/AIDS). medRxiv.

437   21.   Rogers AA, Baumann RE, Borillo GA, Kagan RM, Batterman HJ, Galdzicka MM, Marlowe
438   EM. 2020. Evaluation of Transport Media and Specimen Transport Conditions for the
439   Detection of SARS-CoV-2 by Use of Real-Time Reverse Transcription-PCR. J Clin Microbiol
440   58.

441   22.   Rodino KG, Espy MJ, Buckwalter SP, Walchak RC, Germer JJ, Fernholz E, Boerger A, Schuetz
442   AN, Yao JD, Binnicker MJ. 2020. Evaluation of Saline, Phosphate-Buffered Saline, and
443   Minimum Essential Medium as Potential Alternatives to Viral Transport Media for SARS-
444   CoV-2 Testing. J Clin Microbiol 58.

445   23.   Welch SR, Davies KA, Buczkowski H, Hettiarachchi N, Green N, Arnold U, Jones M, Hannah
446   MJ, Evans R, Burton C, Burton JE, Guiver M, Cane PA, Woodford N, Bruce CB, Roberts ADG,
447   Killip MJ. 2020. Analysis of Inactivation of SARS-CoV-2 by Specimen Transport Media,
448   Nucleic Acid Extraction Reagents, Detergents, and Fixatives. J Clin Microbiol 58.

449   24.   Viral Sample Collection.

450   25.   IGI Testing Consortium. 2020. Blueprint for a pop-up SARS-CoV-2 testing lab. Nat
451   Biotechnol 38:791–797.

452   26.   Ranoa DRE, Holland RL, Alnaji FG, Green KJ, Wang L, Brooke CB, Burke MD, Fan TM,
453   Hergenrother PJ. 2020. Saliva-Based Molecular Testing for SARS-CoV-2 that Bypasses RNA

Extraction. Cold Spring Harbor Laboratory.

27.   Ben-Ami R, Klochendler A, Seidel M, Sido T, Gurel-Gurevich O, Yassour M, Meshorer E,

Benedek G, Fogel I, Oiknine-Djian E, Gertler A, Rotstein Z, Lavi B, Dor Y, Wolf DG, Salton M,

Drier Y, Hebrew University-Hadassah COVID-19 Diagnosis Team. 2020. Large-scale

implementation of pooled RNA extraction and RT-PCR for SARS-CoV-2 detection. Clin

Microbiol Infect https://doi.org/10.1016/j.cmi.2020.06.009.

28.   Yelin I, Aharony N, Shaer Tamar E, Argoetti A, Messer E, Berenbaum D, Shafran E, Kuzli A,

Gandali N, Shkedi O, Hashimshony T, Mandel-Gutfreund Y, Halberthal M, Geffen Y,

Szwarcwort-Cohen M, Kishony R. 2020. Evaluation of COVID-19 RT-qPCR test in multi-

sample pools. Clin Infect Dis https://doi.org/10.1093/cid/ciaa531.

29.   Perchetti GA, Sullivan K-W, Pepper G, Huang M-L, Breit N, Mathias P, Jerome KR, Greninger

AL. 2020. Pooling of SARS-CoV-2 samples to increase molecular testing throughput. J Clin

Virol 131:104570.

30.   Lim KL, Johari NA, Wong ST, Khaw LT, Tan BK, Chan KK, Wong SF, Chan WLE, Ramzi NH, Lim

PKC, Hakim SL, Voon K. 2020. A novel strategy for community screening of SARS-CoV-2

(COVID-19): Sample pooling method. PLoS One 15:e0238417.

31.   Abdalhamid B, Bilder CR, McCutchen EL, Hinrichs SH, Koepsell SA, Iwen PC. 2020.

Assessment of Specimen Pooling to Conserve SARS CoV-2 Testing Resources. Am J Clin

Pathol 153:715–718.

7. AUTHOR CONTRIBUTIONS

Meredith Muller: Writing - Original Draft, Visualization, Validation, Formal analysis,

Investigation **Srijana Chhetri**: Writing- Reviewing and Editing, Validation, Investigation

Christopher Basham: Project administration **Tyler Rapp**: Writing- Reviewing and Editing,

Investigation **Feng-Chang Lin:** Writing- Reviewing and Editing, Formal analysis **Kelly Lin:** Writing-

Reviewing and Editing **Daniel Westreich:** Writing- Reviewing and Editing **Carla Cerami:** Writing-

Reviewing and Editing, Investigation **Jonathan Juliano:** Conceptualization, Writing- Reviewing

and Editing **Jessica Lin:** Conceptualization, Supervision, Funding acquisition, Writing- Reviewing

and Editing, Visualization, Investigation

8. FIGURE LEGENDS

Figure 1. Concordance and comparison of SARS-CoV-2 viral loads from paired nasopharyngeal

(NP) and nasal mid-turbinate (NMT) swabs. Paired NP and NMT swabs from 173 participants

showed overall good concordance, with most discordances (15/16) arising from positive

NP/negative NMT samples. Quantitative viral loads derived from the average of N1 and N2 qRT-

PCR assays favored NP swabs compared to NMT swabs. A y=x dashed line is drawn for

reference.

Figure 2. Concordance of SARS-CoV-2 RT-qPCR detection between nasopharyngeal (NP) swabs

and two different nasal mid-turbinate (NMT) swab types, stored in viral transport media

(VTM) or 1x DNA/RNA shield (Shield). In (A), sensitivity and specificity of standard flocked

(NMT) or 3D-printed (3D) nasal swabs collected into VTM are shown using NP swabs with co-

packaged universal viral transport system as the reference standard. Concordance of flocked vs.

3D nasal swabs is also shown. In (B), flocked NMT swabs were stored in Shield, and sample

aliquots were directly frozen on day 1 (D1) or kept at room temperature before being stored at

-80C on day 4 (D4) or day 7 (D7). Note that samples with indeterminate viral load (<11

copies/ul) were not included in the sensitivity/specificity analyses.

Figure 3. Comparison of SARS-CoV-2 viral loads between standard NP swabs and NMT or 3D

NMT swabs (A) as well as NMT swabs collected into 1x DNA/RNA Shield and stored for

different intervals (B). The distribution of the difference in log viral load is depicted for each

comparison. Median log-fold changes are indicated by a solid line with interquartile values

indicated by dotted lines. The number of sample pairs is indicated for each comparison.

Figure 4. Comparison of Ct values from nasal mid-turbinate (NMT) swabs pooled from

households of 3-5 persons at the point of care vs. concurrently collected individual NMT

swabs. Among the pools collected from 24 households (listed along the x axis in order of

decreasing viral loads), 2 pools with discordant results from individual swabs are depicted as

red stars. Viral loads derived from the Ct values for each sample and the corresponding pool are

found in Table S2.

Figure 5. Ct values of increasing pool sizes constructed containing a single positive sample

with varying viral loads. Viral transport media from a single positive sample with a viral load

ranging from 10^1 to 10^7 were used to construct pool sizes of 5, 10, 15,

A

	NP +	NP -
NMT VTM +	104	1
NMT VTM -	16	52

Sensitivity 87% (95% CI, 79 - 92%)

Specificity 98% (95% CI, 90 - 100%)

N = 185 swab pairs, (12 indeterminate swabs)

	NP +	NP -
3D VTM +	23	0
3D VTM -	1	8

Sensitivity 96% (95% CI, 79 - 100%)

Specificity 100 (95% CI, 63 - 100%)

N = 32 swab pairs (0 indeterminate swabs)

	NMT VTM +	NMT VTM -
3D VTM+	24	4
3D VTM-	4	16

Sensitivity 86% (95% CI, 67 - 96%)

Specificity 80 (95% CI, 56 - 94%)

N = 48 swab pairs (3 Indeterminate swabs)

B

	NP +	NP -
NMT D1 shield +	19	0
NMT D1 shield -	3	16

Sensitivity 86.4% (95% CI, 65 - 97%)

Specificity 100% (95% CI, 79 - 100%)

N = 40 pairs (2 indeterminate samples)

	NP +	NP -
NMT D4 shield +	29	0
NMT D4 shield -	7	18

Sensitivity 81% (95% CI, 64 - 92%)

Specificity 100% (95% CI, 82 - 100%)

N = 56 pairs (2 Indeterminate samples)

	NP +	NP -
NMT D7 shield +	40	0
NMT D7 shield -	3	23

Sensitivity 93% (95% CI, 81 - 99%)

Specificity 100% (95% CI, 85 - 100%)

N = 68 pairs (2 indeterminate samples)