Title

Benchmarking saliency methods for chest X-ray interpretation

Authors

Adriel Saporta MS MBA[1*], Xiaotong Gui MS[1*], Ashwin Agrawal MS[1*], Anuj Pareek MD

PhD[2], Steven QH Truong MBA[3], Chanh DT Nguyen PhD[3,4], Van-Doan Ngo MD[5], Jayne

Seekins DO[6], Francis G. Blankenberg MD[6], Andrew Y. Ng PhD[1], Matthew P. Lungren MD

MPH[2†], Pranav Rajpurkar PhD [7†]

Affiliations

[1]Stanford University Department of Computer Science, USA

[2]Stanford Center for Artificial Intelligence in Medicine and Imaging, USA

[3]VinBrain, Vietnam

[4]VinUniversity, Vietnam

[5]Vinmec International Hospital, Vietnam

[6]Stanford University School of Medicine, Department of Radiology, USA

[7]Department of Biomedical Informatics, Harvard University, USA

[*]These authors contributed equally: Adriel Saporta, Xiaotong Gui, Ashwin Agrawal

†These authors contributed equally: Matthew P. Lungren, Pranav Rajpurkar

Corresponding author: Pranav Rajpurkar, PhD (pranav_rajpurkar@hms.harvard.edu)

Current word count: 3829

Abstract

Saliency methods, which "explain" deep neural networks by producing heat maps that highlight the areas of the medical image that influence model prediction, are often presented to clinicians as an aid in diagnostic decision-making. Although many saliency methods have been proposed for medical imaging interpretation, rigorous investigation of the accuracy and reliability of these strategies is necessary before they are integrated into the clinical setting. In this work, we quantitatively evaluate three saliency methods (Grad-CAM, Grad-CAM++, and Integrated Gradients) across multiple neural network architectures using two evaluation metrics. We establish the first human benchmark for chest X-ray interpretation in a multilabel classification set up, and examine under what clinical conditions saliency maps might be more prone to failure in localizing important pathologies compared to a human expert benchmark. We find that (i) while Grad-CAM generally localized pathologies better than the two other saliency methods, all three performed significantly worse compared with the human benchmark; (ii) the gap in localization performance between Grad-CAM and the human benchmark was largest for pathologies that had multiple instances, were smaller in size, and had shapes that were more complex; (iii) model confidence was positively correlated with Grad-CAM localization performance. Our work demonstrates that several important limitations of saliency methods must be addressed before we can rely on them for deep learning explainability in medical imaging.

Introduction

Deep learning has enabled automated medical imaging interpretation at the level of

practicing experts in some settings[1–3]. While the potential benefits of automated

diagnostic models are numerous, lack of model interpretability in the use of "black-box"

deep neural networks (DNNs) represents a major barrier to clinical trust and adoption[4–6].

In fact, it has been argued that the European Union's recently adopted General Data

Protection Regulation (GDPR) affirms an individual's right to an explanation in the context

of automated decision-making[7]. Although the importance of DNN interpretability is widely

acknowledged and many techniques have been proposed, little emphasis has been

placed on how best to quantitatively evaluate these explainability methods[8].

One type of DNN interpretation strategy widely used in the context of medical imaging is

based on saliency (or pixel-attribution) methods[9–12]. Saliency methods produce heat

maps highlighting the areas of the medical image that most influenced the DNN's

prediction. The heat maps help to visualize whether a DNN is concentrating on the same

regions of a medical image that a human expert would focus on, rather than concentrating

on a clinically irrelevant part of the medical image or even on confounders in the image[13–

15]. Saliency methods have been widely used for a variety of medical imaging tasks and

modalities including, but not limited to, visualizing the performance of a convolutional

neural network (CNN) in predicting (1) myocardial infarction[16] and hypoglycemia[17] from

electrocardiograms, (2) visual impairment[18], refractive error[19], and anaemia[20] from retinal

photographs, (3) long-term mortality[21] and tuberculosis[22] from chest X-ray (CXR) images,

and (4) appendicitis[23] and pulmonary embolism[24] on computed tomography scans.

However, recent work has shown that saliency methods used to validate model

predictions can be misleading in some cases and may lead to increased bias and loss of

user trust in high-stakes contexts such as healthcare[25–27]. Therefore, a rigorous

investigation of the accuracy and reliability of these strategies is necessary before they

are integrated into the clinical setting[28].

In this work, we perform a systematic evaluation of the three most common saliency

methods in medical imaging (Grad-CAM[29], Grad-CAM++[30], and Integrated Gradients[31])

using three common CNN architectures (DenseNet121[32], ResNet152[33], Inception-v4[34]).

In doing so, we establish the first human benchmark in CXR localization by collecting

radiologist segmentations for 10 pathologies using CheXpert, a large publicly available

CXR dataset[35]. To compare saliency method segmentations with expert segmentations,

we use two metrics to capture localization accuracy: (1) *mean Intersection over Union*, a

stricter metric that measures the overlap between the saliency method segmentation and

the expert segmentation, and (2) *hit rate*, a less strict metric that does not require the

saliency method to locate the full extent of a pathology. We find that (1) while Grad-CAM

generally localizes pathologies more accurately than the two other saliency methods, all

three perform significantly worse compared with a human radiologist benchmark; (2) the

gap in localization performance between Grad-CAM and the human benchmark is largest

for pathologies that have multiple instances on the same CXR, are smaller in size, and

have shapes that are more complex; (3) model confidence is positively correlated with

Grad-CAM localization performance. We publicly release a development dataset of expert

segmentations, which we call CheXplanation, to facilitate further research in DNN

explainability for medical imaging.

Results

Framework for evaluating saliency methods on multi-label classification models

Three methods were evaluated—Grad-CAM, Grad-CAM++, and Integrated Gradients—in a multi-label classification setup on the CheXpert dataset (Fig. 1a). For each of the three saliency methods, we ran experiments using three CNN architectures previously used on CheXpert: DenseNet121, ResNet152, and Inception-v4. For each combination of saliency method and model architecture, we trained and evaluated an ensemble of 30 CNNs (see Methods for ensembling details). We then passed each of the CXRs in the dataset's holdout test set into the trained ensemble model to obtain image-level predictions for the following 10 pathologies: Atelectasis, Cardiomegaly, Consolidation, Edema, Enlarged Cardiomediastinum, Lung Lesion, Lung Opacity, Pleural Effusion, Pneumothorax, and Support Devices. For each CXR, we used the saliency method to generate heat maps, one for each of the 10 pathologies, and then applied a threshold to each heat map to produce binary segmentations (top row, Fig. 1a). Thresholding is determined per pathology using Otsu's method[36], which iteratively searches for a threshold value that maximizes inter-class pixel intensity variance. We also conducted a sensitivity analysis of localization performance using different thresholds. The result shows that our evaluation of localization performance is robust to different saliency map thresholding values (see Supplementary Fig. 15). Additionally, to calculate the hit rate evaluation metric (described below), we extracted the pixel in the saliency method heat map with the largest value as the single most representative point on the CXR for that pathology.

We obtained two independent sets of pixel-level CXR segmentations on the holdout test

set: ground-truth segmentations drawn by two board-certified radiologists (middle row,

Fig. 1a) and human benchmark segmentations drawn by a separate group of three board-

certified radiologists (bottom row, Fig. 1a). The human benchmark segmentations and the

saliency method segmentations were compared with the ground-truth segmentations to

establish the human benchmark localization performance and the saliency method

localization performance, respectively. Additionally, for the hit rate evaluation metric, the

radiologists who drew the benchmark segmentations were also asked to locate a single

point on the CXR that was most representative of the pathology at hand (see

Supplementary Figs. 1 through 11 for detailed instructions given to the radiologists).

We used two evaluation metrics to compare segmentations (Fig. 1b). First, we used *mean*

Intersection over Union (mIoU), a stricter metric that measures how much, on average,

either the saliency method or benchmark segmentations overlapped with the ground-truth

segmentations. Second, we used *hit rate*, a less strict metric that does not require the

saliency method or benchmark annotators to locate the full extent of a pathology. Hit rate

is based on the pointing game setup[37], in which credit is given if the most representative

point identified by the saliency method or the benchmark annotators lies within the

ground-truth segmentation. A "hit" indicates that the correct region of the CXR was

located regardless of the exact bounds of the binary segmentations. Localization

performance is then calculated as the hit rate across the dataset[38].

a | Annotation and evaluation workflow

b | Evaluation metrics on example CXRs

IoU: 0.078 | hit rate: hit
Pleural Effusion

IoU: 0.682 | hit rate: miss
Enlarged Cardiomediastinum

Fig. 1 | Framework for evaluating saliency methods on multi-label classification models. a, Top row left: a CXR image from the holdout test set is passed into an ensemble CNN trained only on CXR images and their corresponding pathology task labels. Saliency method is used to generate 10 heat maps for the example CXR, one for each task. The pixel in the heat map with the largest value is determined to be the single most representative point on the CXR for that pathology. Top row middle: there are three pathologies present in this CXR (Airspace Opacity, Pleural Effusion, and Support Devices). Top row right: a threshold is applied to the heat maps to produce binary segmentations for each present pathology. Middle row: Two board-certified radiologists were asked to segment the pathologies that were present in the CXR as determined by the dataset's ground-truth labels. Saliency method annotations are compared to these

ground-truth annotations to evaluate how well saliency method identifies clinically-relevant areas
of the input CXR ("saliency method localization performance"). Bottom row: Two board-certified
radiologists (separate from those in middle row) were also asked to segment the pathologies that
were present in the CXR as determined by the dataset's ground-truth labels. In addition, these
radiologists were asked to locate the single point on the CXR that was most representative of
each present pathology. These benchmark annotations are compared to the ground-truth
annotations to determine a human benchmark ("human benchmark localization performance").
b, Left: CXR with ground-truth and saliency method annotations for Pleural Effusion. The
segmentations have a low overlap (IoU is 0.078), but pointing game is a "hit" since the saliency
method's most representative point is inside of the ground-truth segmentation. Right, CXR with
ground-truth and human benchmark annotations for Enlarged Cardiomediastinum. The
segmentations have a high overlap (IoU is 0.682), but pointing game is a "miss" since saliency
method's most representative point is outside of the ground-truth segmentation.
161

Evaluating localization performance of the saliency methods against the human

benchmark

In order to compare the localization performance of the saliency methods with the human

benchmark, we ran eighteen experiments, one for each combination of saliency method

(Grad-CAM, Grad-CAM++, or Integrated Gradients) and CNN architecture

(DenseNet121, ResNet152, or Inception-v4) using one of the two evaluation metrics

(mIoU or hit rate). For each evaluation metric, we chose the combination of saliency

method and architecture that demonstrated the best localization performance (Fig. 2a).

We found that Grad-CAM with DenseNet121 had the highest mIoU performance and the

highest hit rate performance. Accordingly, we compared Grad-CAM with DenseNet121

("saliency method pipeline") with the human benchmark using both mIoU and hit rate. The

localization performance for each pathology is reported on the true positive slice of the

dataset (CXRs that contain both saliency method and human benchmark segmentations

when the ground-truth label of the pathology is positive). Localization performance was

calculated this way so that saliency methods were not penalized by DNN classification

error: while the benchmark radiologists were provided with ground-truth labels when

annotating the dataset, saliency method segmentations were created based on labels

predicted by the model. (See Supplementary Fig. 16 for localization performance results

on the full dataset.)

We found that the saliency method pipeline demonstrated significantly worse localization

performance when compared with the human benchmark using both mIoU (Fig. 2b) and

hit rate (Fig. 2c) as an evaluation metric, regardless of model classification AUROC. For

each metric, we report the 95% confidence intervals using the bootstrap method with

1,000 bootstrap samples[39]. For five of the 10 pathologies, the saliency method pipeline

had a significantly lower mIoU than the human benchmark. For example, the saliency

method pipeline had one of the highest AUROC scores of the 10 pathologies for Support

Devices (0.969), but had among the worst localization performance for Support Devices

when using both mIoU (0.163 [95% CI 0.154, 0.172]) and hit rate (0.357 [95% CI 0.303,

0.408]) as evaluation metrics. On two pathologies (Atelectasis and Consolidation) the

saliency method pipeline significantly outperformed the human benchmark. On average,

across all 10 pathologies, mIoU saliency method pipeline performance was 26.6% [95%

CI 18.1%, 35.0%] worse than the human benchmark, with Lung Lesion displaying the

largest gap in performance (76.2% [95% CI 59.1%, 87.5%] worse than the human

benchmark) (Supplementary Table 4). Consolidation was the pathology on which the

mIoU saliency method pipeline performance exceeded the human benchmark the most,

by 56.1% [95% CI 42.7%, 69.4%]. For seven of the 10 pathologies, the saliency method

pipeline had a significantly lower hit rate than the human benchmark. On average, hit rate

saliency method pipeline performance was 29.4% [95% CI 15.0%, 43.2%] worse than the

human benchmark (Supplementary Table 5), with Lung Lesion again displaying the

largest gap in performance (65.9% [95% CI 35.3%, 91.7%] worse than the human

benchmark). The hit rate saliency method pipeline did not significantly outperform the

human benchmark on any of the 10 pathologies; for the remaining three of the 10

pathologies, the hit rate performance differences between the saliency method pipeline

and the human benchmark were not statistically significant. Therefore, while the saliency

method pipeline significantly underperformed the human benchmark regardless of

evaluation metric used, the average performance gap was larger when using hit rate as

an evaluation metric than when using mIoU as an evaluation metric.

a | Selection strategy for saliency method pipeline

b | mIoU localization performances

c | Hit rate localization performances

Fig. 2 | Evaluating the localization performance of the saliency methods against the human benchmark. a, The selection strategy for the mIoU saliency method pipeline and the hit rate saliency method pipeline. For each saliency method, the best model architecture is selected (highlighted in purple). Then, the best saliency method + model architecture pair, of the three, is selected. The selection strategy was performed twice, once using mIoU as the evaluation metric and once using hit rate as the evaluation metric. The best saliency method + model architecture pair was the same for both mIoU and hit rate: Grad-CAM + DenseNet121. **b**, Comparing saliency method and human benchmark localization performances under the overlap evaluation scheme (mIoU). Pathologies, along with their DenseNet121 AUROCs, are sorted on the x-axis in descending order of percentage decrease from human benchmark mIoU to saliency method pipeline mIoU for each pathology. **c**, Comparing saliency method and human benchmark localization performances under the hit rate evaluation scheme. Pathologies, along with their DenseNet121 AUROCs, are sorted on the x-axis in descending order of percentage decrease from human benchmark hit rate to saliency method pipeline hit rate for each pathology.

Characterizing the underperformance of the saliency method pipeline localization

In order to better understand the underperformance of the saliency method pipeline localization, we first conducted a qualitative analysis with a radiologist by visually inspecting both the segmentations produced by the saliency method pipeline (Grad-CAM with DenseNet121) and the human benchmark segmentations. We found that, in general, saliency method segmentations fail to capture the geometric nuances of a given pathology, and instead produce coarse, low-resolution heat maps. Specifically, our qualitative analysis found that the performance of the saliency method depended on three pathological characteristics (Fig. 3a): (1) *number of instances*: when a pathology had multiple instances on a CXR, the saliency method segmentation often highlighted one large confluent area, instead of highlighting each distinct instance of the pathology separately; (2) *size*: saliency method segmentations tended to be significantly larger than human expert segmentations, often failing to respect clear anatomical boundaries; (3) *shape complexity*: the saliency method segmentations for pathologies with complex shapes frequently included significant portions of the CXR where the pathology is not present.

Informed by our qualitative analysis and previous work in histology[40], we defined four

geometric features for our quantitative analysis (Fig. 3b): (1) *number of instances* (for

example, bilateral Pleural Effusion would have two instances, whereas there is only one

instance for Cardiomegaly), (2) *size* (pathology area with respect to the area of the whole

CXR), (3) *elongation* and (4) *irrectangularity* (the last two features measure the complexity

of the pathology shape and were calculated by fitting a rectangle of minimum area

enclosing the binary mask). See Supplementary Fig. 17 for the distribution of the four

pathological characteristics across all 10 pathologies.

For each evaluation metric, we ran 8 simple linear regressions: four with the evaluation

metric (IoU or hit rate) of the saliency method pipeline (Grad-CAM with DenseNet121) as

the dependent variable (to understand the relationship between the geometric features of

a pathology and saliency method localization performance), and four with the difference

between the evaluation metrics of the saliency method pipeline and the human

benchmark as the dependent variable (to understand the relationship between the

geometric features of a pathology and the gap in localization performance between the

saliency method pipeline and the human benchmark). Each regression used one of the

four geometric features as a single independent variable, and only the true positive slice

was included in each regression. Each feature was normalized using z-score

normalization and the regression coefficient can be interpreted as the effect of that

geometric feature on the evaluation metric at hand. See Table 1 for coefficients from the

regressions using both evaluation metrics, where we also report the 95% confidence

interval and the Bonferroni corrected p-values. For confidence intervals and p-values, we

used the standard calculation for linear models.

a | Examples from qualitative analysis

Saliency method segmentation
Ground-truth segmentation

Number of instances
Edema

Size
Airspace Opacity

Shape complexity
Pleural Effusion

b | Geometric features

Number of instances = 1
Cardiomegaly

Number of instances = 2
Atelectasis

Size = large (0.26)
Enlarged Cardiomediastinum

Size = small (0.006)
Lung Lesion

Elongation = low (1.44)
Airspace Opacity

Elongation = high (5.43)
Support Devices

Irrectangularity = low (0.24)
Enlarged Cardiomediastinum

Irrectangularity = high (0.78)
Pleural Effusion

Fig. 3 | Characterizing the underperformance of saliency method localization. a, Example
CXRs that highlight the three pathological characteristics identified by our qualitative analysis:
(1) Left, number of instances; (2) Middle, size; and (3) Right, shape complexity. **b**, Example CXRs
with the four geometric features used in our quantitative analysis: (1) Top row left, number of
instances; (2) Top row right, size = area of segmentation/area of CXR; (3) Bottom row left,
elongation; and (4) Bottom row right, irrectangularity. Elongation and irrectangularity were
calculated by fitting a rectangle of minimum area enclosing the binary mask (as indicated by the
yellow rectangles). Elongation = maxAxis/minAxis. Irrectangularity = 1 - (area of
segmentation/area of enclosing rectangle).

Our statistical analysis showed that as the area ratio of a pathology increased, mIoU

saliency method localization performance improved (0.566 [95% CI 0.526, 0.606]). We

also found that as elongation and irrectangularity increased, mIoU saliency method

localization performance worsened (elongation: -0.425 [95% CI -0.497, -0.354],

irrectangularity: -0.256 [95% CI -0.292, -0.219]). We observed that the effects of these

three geometric features were similar for hit rate saliency method localization

performance in terms of levels of statistical significance and direction of the effects.

However, there was no evidence that the number of instances of a pathology had a

significant effect on either mIoU (-0.115 [95% CI -0.220, -0.009]) or hit rate (-0.051 [95%

CI -0.364, 0.244]) saliency method localization. Therefore, regardless of evaluation

metric, saliency method localization performance suffered in the presence of pathologies

that were small in size and complex in shape.

We found that these same three pathological characteristics—larger size, and higher

elongation and irrectangularity—characterized the *gap* in mIoU localization performance

between saliency method and human benchmark. We observed that the *gap* in hit rate

localization performance was significantly characterized by all four geometric features

(number of instances, size, elongation, and irrectangularity). As the number of instances

increased, despite no significant change in hit rate localization performance itself, the *gap*

in hit rate localization performance between saliency method and the human benchmark

increased (0.470 [95% CI 0.114, 0.825]). This suggests that the saliency method performs

especially poorly in the face of a multi-instance diagnosis.

Table 1 | Coefficients from regressions on geometric features of pathologies

Geometric feature (independent variable)	Coefficient using saliency method localization (dependent variable)	Coefficient using localization difference (human benchmark - saliency method) (dependent variable)
IoU		
Number of instances	-0.115 (-0.220, -0.009)	-0.072 (-0.237, -0.094)
Size	0.566 (0.526, 0.606) ***	-0.154 (-0.231, -0.076) ***
Elongation	-0.425 (-0.497, -0.354) ***	0.476 (0.362, 0.589) ***
Irrectangularity	-0.256 (-0.292 -0.219) ***	0.307 (0.249, 0.366) ***
Hit/Miss		
Number of instances	-0.051 (-0.346, 0.244)	0.470 (0.114, 0.825) *
Size	1.269 (1.146, 1.391) ***	-0.944 (-1.104, -0.785) ***
Elongation	-0.849 (-1.053, -0.646) ***	1.110 (0.865, 1.354) ***
Irrectangularity	-0.519 (-0.624, -0.415) ***	0.689 (0.564, 0.815) ***

* p-value < 0.05, ** p-value < 0.01, *** p-value < 0.001

Effect of model confidence on saliency method localization performance

We also conducted statistical analyses to determine whether there was any correlation

between the model's confidence in its prediction and saliency method pipeline

performance (Table 2). We first ran a simple regression for each pathology using the

model's probability output as the single independent variable and using the saliency

method IoU as the dependent variable. We then performed a simple regression that uses

the same approach as above, but that includes all 10 pathologies. For each of the 11

regressions, we used the full dataset since the analysis of false positives and false

negatives was also of interest. In addition to the linear regression coefficients, we also

computed the Spearman correlation coefficients to capture any potential non-linear

associations.

We found that for all pathologies, model confidence was positively correlated with mIoU

saliency method pipeline performance. The p-values for all coefficients were below 0.001

except for the coefficients for Pneumothorax (n=11) and Lung Lesion (n=50), the two

pathologies for which we had the fewest positive examples. Of all the pathologies, model

confidence for positive predictions of Enlarged Cardiomediastinum had the largest linear

regression coefficient with mIoU saliency method pipeline performance (1.974, p-

value<0.001). Model confidence for positive predictions of Pneumothorax had the largest

Spearman correlation coefficient with mIoU saliency method pipeline performance (0.734,

p-value<0.01), followed by Pleural Effusion (0.690, p-value<0.001). Combining all

pathologies (n=2365), the linear regression coefficient was 0.109 (95% CI [0.083, 0.135]),

and the Spearman correlation coefficient was 0.285 (95% CI [0.239, 0.331]). We also

performed analogous experiments using hit rate as the dependent variable and found

comparable results (Supplementary Table 1).

Table 2 | mIoU: Coefficients from regressions on model assurance

Pathology	CXRs (n)	Linear regression coefficient	Spearman correlation coefficient
Airspace Opacity	381	0.714 (0.601, 0.826) ***	0.577 (0.542, 0.61) ***
Atelectasis	296	0.489 (0.333, 0.645) ***	0.348 (0.303, 0.391) ***
Cardiomegaly	229	0.679 (0.535, 0.823) ***	0.592 (0.559, 0.624) ***
Consolidation	120	1.155 (0.674, 1.635) ***	0.384 (0.341, 0.426) ***
Edema	124	0.642 (0.459, 0.826)***	0.548 (0.512, 0.582) ***
Enlarged Cardiomediastinum	668	1.974 (1.608, 2.340) ***	0.428 (0.386, 0.468) ***
Lung Lesion	50	0.218 (0.087, 0.349) **	0.509 (0.47, 0.545) ***
Pleural Effusion	159	0.632 (0.489, 0.776) ***	0.690 (0.663, 0.715) ***
Pneumothorax	11	0.446 (0.108, 0.783) *	0.734 (0.710, 0.756) **
Support Devices	327	0.211 (0.172, 0.25) ***	0.468 (0.428, 0.506) ***
All pathologies	2365	0.109 (0.083, 0.135) ***	0.285 (0.239, 0.331) ***

* p-value < 0.05, ** p-value < 0.01, *** p-value < 0.001

Discussion

The purpose of this work was to evaluate the performance of some of the most used

saliency methods (Grad-CAM, Grad-CAM++, Integrated Gradients) for deep learning

explainability using a variety of model architectures. We establish the first human

benchmark for CXR localization in a multilabel classification setup and demonstrate that

saliency maps are consistently worse than expert radiologists regardless of model

classification AUROC. We use qualitative and quantitative analyses to establish that

saliency method localization performance is most inferior to expert localization

performance when a pathology has multiple instances, is smaller in size, or has shapes

that are more complex, suggesting that deep learning explainability as a clinical interface

may be less reliable and less useful when used for pathologies with those characteristics.

We also show that model assurance is positively correlated with saliency method

localization performance, which could indicate that saliency methods are safer to use as

a decision aid to clinicians when the model has made a positive prediction with high

confidence.

While there are several public CXR datasets with image-level labels annotated by expert

radiologists, including MIMIC-CXR[41] and ChestX-ray8[42], and several datasets containing

segmentations for a single pathology, including SIIM-ACR Pneumothorax Segmentation[43]

and RSNA Pneumonia Detection[44], to our knowledge there are no other publicly available

CXR datasets with multilabel expert segmentations. By publicly releasing a development

dataset, CheXplanation, of 234 images with 885 expert segmentations, and a competition

with a test set of 668 images, we hope to encourage the further development of saliency

methods and other explainability techniques for medical imaging.

Our work has several potential implications for human-AI collaboration in the context of

medical decision-making. Heat maps generated using saliency methods are advocated

as clinical decision support in the hope that they not only improve clinical decision-

making, but also encourage clinicians to trust model predictions[45–47]. Many of the large

CXR vendors[48–50] use localization methods to provide pathology visualization in their

computer-aided detection (CAD) products. In addition to being used for clinical

interpretation, saliency method heat maps are also used for the evaluation of CXR

interpretation models, for quality improvement (QI) and quality assurance (QA) in clinical

practices, and for dataset annotation[51]. However, we found that saliency method

localization performance, on balance, performed worse than expert localization across

multiple analyses and across many important pathologies (our findings are consistent

with recent work focused on localizing a single pathology, Pneumothorax, in CXRs[52]). If

used in clinical practice, heat maps that incorrectly highlight medical images may

exacerbate well documented biases (chiefly, automation bias) and erode trust in model

predictions (even when model output is correct), limiting clinical translation[22].

Since IoU computes the overlap of two segmentations but pointing game hit rate better

captures diagnostic attention, we suggest using both metrics when evaluating localization

performance in the context of medical imaging. While IoU is a commonly used metric for

evaluating semantic segmentation outputs, there are inherent limitations to the metric in

the pathological context. This is indicated by our finding that even the human benchmark

segmentations had low overlap with the ground truth segmentations (the highest expert

mIoU was 0.720 for Cardiomegaly). One potential explanation for this consistent

underperformance is that pathologies can be hard to distinguish, especially without

clinical context. Furthermore, whereas many people might agree on how to segment, say,

a cat or a stop sign in traditional computer vision tasks, radiologists use a certain amount

of clinical discretion when defining the boundaries of a pathology on a CXR. There can

also be institutional and geographic differences in how radiologists are taught to

recognize pathologies, and studies have shown that there can be high interobserver

variability in the interpretation of CXRs[53–55]. We sought to address this with the hit rate

evaluation metric, which highlights when two radiologists share the same diagnostic

intention, even if it is less exact than IoU in comparing segmentations directly. The human

benchmark localization using hit rate was above 0.9 for four pathologies (Pneumothorax,

Cardiomegaly, Support Devices, and Enlarged Cardiomediastinum); these are

pathologies for which there is often little disagreement between radiologists about where

the pathologies are located, even if the expert segmentations are noisy. Further work is

needed to demonstrate which segmentation evaluation metrics, even beyond overlap and

hit rate, are more appropriate for which pathologies when evaluating saliency methods

for the clinical setting.

Our work builds upon several studies investigating the validity of saliency maps for

localization[56,57] and upon some early work on the trustworthiness of saliency methods to

explain DNNs in medical imaging[58]. However, as recent work has shown[31], evaluating

saliency methods is inherently difficult given that they are post-hoc techniques. To

illustrate this, consider the following models and saliency methods as described by some

oracle: (1) a model *M_bad* that has perfect AUROC for a given image classification task,

but that we know does *not* localize well (i.e. because the model picks up on confounders

in the image); (2) a model *M_good* that also has perfect AUROC, but that we know *does*

localize well (i.e. is looking at relevant regions of the image); (3) a saliency method *S_bad*

that does *not* properly reflect the model's attention; and (4) a saliency method *S_good*

that *does* properly reflect the model's attention. Let us say that we are evaluating the

following pipeline: we first classify an image and we then apply a saliency method post

hoc. Imagine that our evaluation reveals poor localization performance as measured by

mIoU or hit rate (as was the case in our findings). There are three possible pipelines

(combinations of model and saliency method) that would lead to this scenario: (1) *M_bad*

+ *S_good*; (2) *M_good* + *S_bad*; and (3) *M_bad* + *S_bad*. The first scenario (*M_bad* +

S_good) is the one for which saliency methods were originally intended: we have a

working saliency method that properly alerts us to models picking up on confounders. The

second scenario (*M_good* + *S_bad*) is our nightmare scenario: we have a working model

whose attention is appropriately directed, but we reject it based on a poorly localizing

saliency method. Because all three scenarios result in poor localization performance, it is

difficult—if not impossible—to know whether poor localization performance is attributable

to the model or to the saliency method (or to both). While we cannot say whether models

or saliency methods are failing in the context of medical imaging, we can say that we

should not rely on saliency methods to evaluate model localization. Future work should

explore potential techniques for localization performance attribution.

There are several limitations of our work. First, we did not investigate the impact of

pathology prevalence in the training data on saliency method localization performance.

Second, some pathologies, such as effusions and cardiomegaly, are in similar locations

across frontal view CXRs, while others, such as lesions and opacities, can vary in

locations across CXRs. Future work could investigate how the location of pathologies on

a CXR in the training/test data distribution, and the consistency of those locations, affect

saliency method localization performance. Third, while we compared saliency method-

generated pixel-level segmentations to human expert pixel-level segmentations, future

work might explore how saliency method localization performance changes when

comparing bounding-box annotations, instead of pixel-level segmentations. Finally, the

impact of saliency methods on the trust and efficacy of users is underexplored.

In conclusion, we present a rigorous evaluation of a range of saliency methods and a

human benchmark dataset, which can serve as a foundation for future work exploring

deep learning explainability techniques. This work is a reminder that care should be taken

when leveraging common saliency methods for deep learning-based workflows for

medical imaging.

Methods

Ethical and information governance approvals.

This study does not involve human subject participants.

Dataset and clinical taxonomy. *Dataset description.* The localization experiments were

performed using CheXpert, a large public dataset for chest X-ray interpretation. The

CheXpert dataset contains 224,316 chest X-rays for 65,240 patients labeled for the

presence of 14 observations (13 pathologies and an observation of "No Finding") as

positive, negative, or uncertain. The CheXpert validation set consists of 234 chest X-rays

from 200 patients randomly sampled from the full dataset and was labeled according to

the consensus of three board-certified radiologists. The test set consists of 668 chest X-

rays from 500 patients not included in the training or validation sets and was labeled

according to the consensus of five board-certified radiologists. See Supplementary Table

2 for dataset summary statistics.

Ground-truth segmentation. The chest X-rays in our validation set and test set were

manually segmented by two board-certified radiologists with 18 and 27 years of

experience, using the annotation software tool MD.ai[59] (see Supplementary Figs. 12

through 14). The radiologists were asked to contour the region of interest for all

observations in the chest X-rays for which there was a positive ground truth label in the

CheXpert dataset. For a pathology with multiple instances, all the instances were

contoured. For Support Devices, radiologists were asked to contour any implanted or

invasive devices including pacemakers, PICC/central catheters, chest tubes,

endotracheal tubes, feeding tubes and stents and ignore ECG lead wires or external

stickers visible in the chest X-ray. Finally, of the 14 observations labeled in the CheXpert

dataset, Fracture, Pleural Other, Pneumonia, and No Finding were not segmented

because they either had low prevalence and/or ill-defined boundaries unfit for

segmentation.

Evaluating the expert performance using benchmark segmentation. To evaluate the

expert performance on the test set using the IoU evaluation method, three radiologists,

certified in Vietnam with 9, 10, and 18 years of experience, were asked to segment the

regions of interest for all observations in the chest X-rays for which there was a positive

ground truth label in the CheXpert dataset. These radiologists were also provided the

same instructions for contouring as were provided to the radiologists drawing the

reference segmentations. To extract the "maximally activated" point from the benchmark

segmentations, we asked the same radiologists to locate each pathology present on each

CXR using only a single most representative point for that pathology on the CXR (see

Supplementary Figs. 1 through 11 for the detailed instructions given to the radiologists).

There was no overlap between these three radiologists and the two who drew the

reference segmentations.

Classification network architecture and training protocol. *Multi-label classification*

model. The model takes as input a single-view chest X-ray and outputs the probability for

each of the 14 observations. In case of availability of more than one view, the models

output the maximum probability of the observations across the views. Each chest X-ray

was resized to 320×320 pixels and normalized before it was fed into the network. The

model architectures (DenseNet121, ResNet152, and Inception-v4) were used. Cross-

entropy loss was used to train the model. The Adam optimizer[60] was used with default β-

parameters of β1 = 0.9 and β2 = 0.999. The learning rate was hyperparameter tuned for

the different model architectures. The best learning rate for each architecture was:

1×10−4 for DenseNet121, 1×10−5 for ResNet152, 1×10−5 for Inceptionv4. Batches were

sampled using a fixed batch size of 16 images.

Ensembling. We use an ensemble of checkpoints to create both predictions and saliency

maps to maximize model performance. In order to capture uncertainties inherent in

radiograph interpretation, we train our models using four uncertainty handling strategies

outlined in CheXpert: Ignoring, Zeroes, Ones, and 3-Class Classification. For each of the

four uncertainty handling strategies, we train our model three separate times, each time

saving the 10 checkpoints across the three epochs with the highest average AUC across

5 observations selected for their clinical importance and prevalence in the validation set:

Atelectasis, Cardiomegaly, Consolidation, Edema, and Pleural Effusion. In total, after

training, we have saved 4 x 30 = 120 checkpoints for a given model. Then, from the 120

saved checkpoints for that model, we select the top 10 performing checkpoints for each

pathology. For each CXR and each task, we compute the predictions and saliency maps

using the relevant checkpoints. We then take the mean both of the predictions and of the

saliency maps to create the final set of predictions and saliency maps for the ensemble

model. See Supplementary Table 3 for the performance of the model on each of the

pathologies.

CNN interpretation strategy. Saliency methods (Grad-CAM, Grad-CAM++, and

Integrated Gradients) were used to visualize the decision made by the classification

network. The saliency map was resized to the original image dimension using bilinear

interpolation. It was then normalized using max-min normalization and then converted

into a binary segmentation using binary thresholding (Otsu's method). We also reported

mIoU localization performance using different saliency map thresholding values. We first

applied max-min normalizations to the saliency maps so that each value gets transformed

into a decimal between 0 and 1. We then passed in a range of threshold values from 0.2

to 0.8 to create binary segmentations and calculated the mIoU score per pathology under

each threshold on the validation set. Then for the analysis with the full dataset (see

Supplementary Figure 16), we further ensure that the final binary segmentation is

consistent with model probability output by applying another layer of thresholding such

that the segmentation mask produced all zeros if the predicted probability was below a

chosen level. The probability threshold is searched on the interval of [0,0.8] with steps of

0.1. The exact value is determined per pathology by maximizing the mIoU on validation

set.

Segmentation evaluation metrics. Localization performance of each segmentation was

evaluated using Intersection over Union (IoU) score. The IoU is the ratio between the

area of overlap and the area of union between the ground truth and the predicted areas,

ranging from 0 to 1 with 0 signifying no overlap and 1 signifying perfectly overlapping

segmentation. Confidence intervals are calculated using bootstrapping with 1000

bootstrap samples. The variance in the width of CI across pathologies can be explained

by difference in sample sizes. For the percentage decrease from expert mIoU to AI mIoU,

we bootstrapped the difference between human benchmark and saliency method

localization and created the 95% confidence intervals. The confidence intervals for hit

rates were calculated in the same fashion.

Statistical analysis.

Pathology Characteristics. We used four features to characterize the pathologies. (1)

Number of instances is defined as the number of disjoint components in the

segmentation. (2) Size is the area of the pathology divided by the total image area. (3)

and (4) Elongation and irrectangularity are geometric features that measure shape

complexities. They were designed to quantify what radiologists qualitatively described as

focal or diffused. To calculate the metrics, a rectangle of minimum area enclosing the

contour is fitted to each pathology. Elongation is defined as the ratio of the rectangle's

longer side to short side. Irrectangularity = 1 - (area of segmentation/area of enclosing

rectangle), with values ranging from 0 to 1 with 1 being very irrectangular. When there

are multiple instances within one pathology, we used the characteristics of the dominant

instance (largest in perimeter).

Model Confidence. We used the probability output of the DNN architecture for model

confidence. The probabilities were normalized using max-min normalization per

pathology before aggregation.

Linear Regression. For each evaluation scheme (overlap and hit rate), we ran two groups

of simple linear regressions, with AI evaluation metrics and their differences as the

response variables. Each group has four regressions using the above four pathological

characteristics as the regressions' single attribute, respectively, and only the true positive

slice was included in each regression. All features are normalized using min-max

normalization so that they are comparable on scales of magnitudes. We report the 95%

confidence interval and Bonferroni adjusted p-value of the regression coefficients.

Acknowledgements

We would like to acknowledge MD.ai for generously providing us access to their

annotation platform. We would like to acknowledge Weights & Biases for generously

providing us access to their experiment tracking tools.

Data Availability

CheXpert data is available at https://stanfordmlgroup.github.io/competitions/chexpert/.

The validation set and corresponding benchmark radiologist annotations can be

downloaded from https://stanfordmlgroup.github.io/competitions/chexplanation/.

Code Availability

All code used to produce the results of the paper is available in the following public

repository   for   the   purpose   of   reproducing   the   study:

https://github.com/stanfordmlgroup/cheXplanation.

Competing Interests

M.L. is an advisor for and/or has research funded by GE, Philips, Carestream, Nines

Radiology, Segmed, Centaur Labs, Microsoft, and BunkerHill.

Author Contributions

(1) Conceptualization: P.R. and A.P., (2) Design: P.R., A.P., A.S., X.G. and A.A., (3) Data

analysis and interpretation: A.S., X.G., A.A., P.R., A.P., S.T., C.N., V.N., J.S., and F.B.,

(4) Drafting of the manuscript: A.S., X.G., A.A., and P.R., (5) Critical revision of the

manuscript for important intellectual content: A.P, S.T., C.N., V.N., J.S., F.B, A.N., and

M.L., (6) Supervision: A.N., M.L, and P.R.

References

1. Rajpurkar, P. *et al.* Deep learning for chest radiograph diagnosis: A retrospective

   comparison of the CheXNeXt algorithm to practicing radiologists. *PLOS Med.* **15**,

   e1002686 (2018).

2. Rajpurkar, P. *et al.* CheXNet: Radiologist-Level Pneumonia Detection on Chest X-

   Rays with Deep Learning. *ArXiv171105225 Cs Stat* (2017).

3. Bien, N. *et al.* Deep-learning-assisted diagnosis for knee magnetic resonance

   imaging: Development and retrospective validation of MRNet. *PLOS Med.* **15**,

   e1002699 (2018).

4. Baselli, G., Codari, M. & Sardanelli, F. Opening the black box of machine learning in

   radiology: can the proximity of annotated cases be a way? *Eur. Radiol. Exp.* **4**, 30

   (2020).

5. Litjens, G. *et al.* A survey on deep learning in medical image analysis. *Med. Image

   Anal.* **42**, 60–88 (2017).

6. Wang, F., Kaushal, R. & Khullar, D. Should Health Care Demand Interpretable

Artificial Intelligence or Accept "Black Box" Medicine? *Ann. Intern. Med.* **172**, 59–60

(2019).

7. Goodman, B. & Flaxman, S. European Union regulations on algorithmic decision-

making and a 'right to explanation'. *AI Mag.* **38**, 50–57 (2017).

8. Venugopal, V., Takhar, R., Gupta, S., Saboo, A. & Mahajan, V. *Clinical Explainability*

Failure (CEF) & Explainability Failure Ratio (EFR) – changing the way we

validate classification algorithms? 2020.08.12.20169607

https://www.medrxiv.org/content/10.1101/2020.08.12.20169607v1 (2020)

doi:10.1101/2020.08.12.20169607.

9. Pasa, F., Golkov, V., Pfeiffer, F., Cremers, D. & Pfeiffer, D. Efficient Deep Network

Architectures for Fast Chest X-Ray Tuberculosis Screening and Visualization. *Sci.*

Rep. **9**, 6268 (2019).

10.   Simonyan, K., Vedaldi, A. & Zisserman, A. Deep Inside Convolutional Networks:

Visualising Image Classification Models and Saliency Maps. *ArXiv13126034 Cs*

(2014).

11.   Aggarwal, M. *et al.* Towards Trainable Saliency Maps in Medical Imaging.

ArXiv201107482 Cs Eess (2020).

12.   Tjoa, E. & Guan, C. Quantifying Explainability of Saliency Methods in Deep

Neural Networks. *ArXiv200902899 Cs* (2020).

13.   Badgeley, M. A. *et al.* Deep learning predicts hip fracture using confounding

patient and healthcare variables. *Npj Digit. Med.* **2**, 1–10 (2019).

14.   Zech, J. R. *et al.* Variable generalization performance of a deep learning model

to detect pneumonia in chest radiographs: A cross-sectional study. *PLOS Med.* **15**,

e1002683 (2018).

15.   DeGrave, A. J., Janizek, J. D. & Lee, S.-I. AI for radiographic COVID-19

detection selects shortcuts over signal. *medRxiv* (2020)

doi:10.1101/2020.09.13.20193565.

16.   Makimoto, H. *et al.* Performance of a convolutional neural network derived from

an ECG database in recognizing myocardial infarction. *Sci. Rep.* **10**, 8445 (2020).

17.   Porumb, M., Stranges, S., Pescapè, A. & Pecchia, L. Precision Medicine and

Artificial Intelligence: A Pilot Study on Deep Learning for Hypoglycemic Events

Detection based on ECG. *Sci. Rep.* **10**, 1–16 (2020).

18.   Tham, Y.-C. *et al.* Referral for disease-related visual impairment using retinal

photograph-based deep learning: a proof-of-concept, model development study.

Lancet Digit. Health **3**, e29–e40 (2021).

19.   Varadarajan, A. V. *et al.* Deep Learning for Predicting Refractive Error From

Retinal Fundus Images. *Invest. Ophthalmol. Vis. Sci.* **59**, 2861–2868 (2018).

20.   Mitani, A. *et al.* Detection of anaemia from retinal fundus images via deep

learning. *Nat. Biomed. Eng.* **4**, 18–27 (2020).

21.   Deep Learning to Assess Long-term Mortality From Chest Radiographs |

Pulmonary Medicine | JAMA Network Open | JAMA Network. https://jamanetwork-

com.stanford.idm.oclc.org/journals/jamanetworkopen/fullarticle/2738349.

22.   Rajpurkar, P. *et al.* CheXaid: deep learning assistance for physician diagnosis of

tuberculosis using chest x-rays in patients with HIV. *Npj Digit. Med.* **3**, 1–8 (2020).

23.   Rajpurkar, P. *et al.* AppendiXNet: Deep Learning for Diagnosis of Appendicitis

from A Small Dataset of CT Exams Using Video Pretraining. *Sci. Rep.* **10**, 3958

(2020).

24.   Huang, S.-C. *et al.* PENet—a scalable deep-learning model for automated

diagnosis of pulmonary embolism using volumetric CT imaging. *Npj Digit. Med.* **3**, 1–

9 (2020).

25.   Rudin, C. Stop explaining black box machine learning models for high stakes

decisions and use interpretable models instead. *Nat. Mach. Intell.* **1**, 206–215 (2019).

26.   Eitel, F. & Ritter, K. Testing the robustness of attribution methods for

convolutional neural networks in MRI-based Alzheimer's disease classification.

ArXiv190908856 Cs Eess (2019).

27.   Young, K., Booth, G., Simpson, B., Dutton, R. & Shrapnel, S. Deep neural

network or dermatologist? *ArXiv190806612 Cs Eess Stat* **11797**, 48–55 (2019).

28.   Reyes, M., Meier, R., Pereira, S., Silva, C. A. & Dahlweid, F.-M. On the

Interpretability of Artificial Intelligence in Radiology: Challenges and Opportunities. **2**,

12.

29.   Selvaraju, R. R. *et al.* Grad-CAM: Visual Explanations from Deep Networks via

Gradient-based Localization. *Int. J. Comput. Vis.* **128**, 336–359 (2020).

30.   Chattopadhay, A., Sarkar, A., Howlader, P. & Balasubramanian, V. N. Grad-

CAM++: Generalized Gradient-Based Visual Explanations for Deep Convolutional

Networks. in *2018 IEEE Winter Conference on Applications of Computer Vision*

(WACV) 839–847 (2018). doi:10.1109/WACV.2018.00097.

31.   Sundararajan, M., Taly, A. & Yan, Q. Axiomatic attribution for deep networks. in

Proceedings of the 34th International Conference on Machine Learning - Volume 70

3319–3328 (JMLR.org, 2017).

32.   Huang, G., Liu, Z., Van Der Maaten, L. & Weinberger, K. Q. Densely Connected

Convolutional Networks. in *2017 IEEE Conference on Computer Vision and Pattern*

Recognition (CVPR) 2261–2269 (IEEE, 2017). doi:10.1109/CVPR.2017.243.

33.   He, K., Zhang, X., Ren, S. & Sun, J. Deep Residual Learning for Image

Recognition. in *2016 IEEE Conference on Computer Vision and Pattern Recognition*

(CVPR) 770–778 (IEEE, 2016). doi:10.1109/CVPR.2016.90.

34.   Szegedy, C. *et al.* Going deeper with convolutions. in *2015 IEEE Conference on*

Computer Vision and Pattern Recognition (CVPR) 1–9 (2015).

doi:10.1109/CVPR.2015.7298594.

35.   Irvin, J. *et al.* CheXpert: A Large Chest Radiograph Dataset with Uncertainty

Labels and Expert Comparison. *Proc. AAAI Conf. Artif. Intell.* **33**, 590–597 (2019).

36.   Otsu, N. A Threshold Selection Method from Gray-Level Histograms. *IEEE Trans*

Syst. Man Cybern. 62–66 (1979).

37.   Zhang, J., Lin, Z., Brandt, J., Shen, X. & Sclaroff, S. Top-down Neural Attention

by Excitation Backprop. *ArXiv160800507 Cs* (2016).

38.   Kim, H.-E. *et al.* Changes in cancer detection and false-positive recall in

mammography using artificial intelligence: a retrospective, multireader study. *Lancet*

Digit. Health **2**, e138–e148 (2020).

39.   Efron, B. & Tibshirani, R. J. *An Introduction to the Bootstrap*. (CRC Press, 1994).

40.   Vrabac, D. *et al.* DLBCL-Morph: Morphological features computed using deep

learning for an annotated digital DLBCL image set. *ArXiv200908123 Cs* (2020).

41.   Johnson, A. E. W. *et al.* MIMIC-CXR, a de-identified publicly available database

of chest radiographs with free-text reports. *Sci. Data* **6**, 317 (2019).

42.   Wang, X. *et al.* ChestX-ray8: Hospital-Scale Chest X-Ray Database and

Benchmarks on Weakly-Supervised Classification and Localization of Common

Thorax Diseases. in 2097–2106 (2017).

43.   SIIM-ACR Pneumothorax Segmentation. https://kaggle.com/c/siim-acr-

pneumothorax-segmentation.

44.   RSNA Pneumonia Detection Challenge. https://kaggle.com/c/rsna-pneumonia-

detection-challenge.

45.   Steiner, D. F. *et al.* Impact of Deep Learning Assistance on the Histopathologic

Review of Lymph Nodes for Metastatic Breast Cancer. *Am. J. Surg. Pathol.* **42**,

1636–1646 (2018).

46.   Uyumazturk, B. *et al.* Deep Learning for the Digital Pathologic Diagnosis of

Cholangiocarcinoma and Hepatocellular Carcinoma: Evaluating the Impact of a Web-

based Diagnostic Assistant. *ArXiv191107372 Eess* (2019).

47.   Park, A. *et al.* Deep Learning–Assisted Diagnosis of Cerebral Aneurysms Using

the HeadXNet Model. *JAMA Netw. Open* **2**, e195600 (2019).

48.   Annalise.ai - Medical imaging AI, by clinicians for clinicians. *Annalise.ai*

https://annalise.ai/.

49.   Lunit Inc. https://www.lunit.io/en.

50.   Qure.ai | Artificial Intelligence for Radiology. https://qure.ai/.

51.   Gadgil, S., Endo, M., Wen, E., Ng, A. Y. & Rajpurkar, P. CheXseg: Combining

Expert Annotations with DNN-generated Saliency Maps for X-ray Segmentation.

ArXiv210210484 Cs (2021).

52.   Crosby, J., Chen, S., Li, F., MacMahon, H. & Giger, M. Network output

visualization to uncover limitations of deep learning detection of pneumothorax. in

Medical Imaging 2020: Image Perception, Observer Performance, and Technology

Assessment vol. 11316 113160O (International Society for Optics and Photonics,

2020).

53.   Melbye, H. & Dale, K. Interobserver Variability in the Radiographic Diagnosis of

Adult Outpatient Pneumonia. *Acta Radiol.* **33**, 79–81 (1992).

54.   Herman, P. G. *et al.* Disagreements in Chest Roentgen Interpretation. *CHEST*

68, 278–282 (1975).

55.   Albaum, M. N. *et al.* Interobserver Reliability of the Chest Radiograph in

Community-Acquired Pneumonia. *CHEST* **110**, 343–350 (1996).

56.   Arun, N. T. *et al.* Assessing the validity of saliency maps for abnormality

localization in medical imaging. *ArXiv200600063 Cs* (2020).

57.   Graziani, M., Lompech, T., Müller, H. & Andrearczyk, V. *Evaluation and*

Comparison of CNN Visual Explanations for Histopathology. (2020).

58.   Arun, N. *et al.* Assessing the (Un)Trustworthiness of Saliency Maps for Localizing

Abnormalities in Medical Imaging. *ArXiv200802766 Cs* (2020).

59.   MD.ai. https://www.md.ai/.

60.   Kingma, D. P. & Ba, J. Adam: A Method for Stochastic Optimization.

ArXiv14126980 Cs (2017).