Diffusion Kurtosis Imaging of neonatal Spinal Cord in clinical routine

Rosella Trò[1*], Monica Roascio[1], Domenico Tortora[2], Mariasavina Severino[2], Andrea Rossi[2,3], Julien Cohen-Adad[4,5,6], Marco Massimo Fato[1†], Gabriele Arnulfo[1,7†]

From the [1]Department of Informatics, Bioengineering, Robotics and System Engineering (DIBRIS), University of Genoa, Genoa, Italy; [2]Neuroradiology Unit, Istituto Giannina Gaslini, Genoa, Italy; [3]Department of Health Sciences (DISSAL), University of Genoa, Genoa, Italy; [4]NeuroPoly Lab, Institute of Biomedical Engineering, Polytechnique Montreal, Montreal, QC, Canada; [5]Functional Neuroimaging Unit, CRIUGM, Université de Montréal, Montreal, QC, Canada; [6]Mila - Quebec AI Institute, Montreal, QC, Canada; [7]Neuroscience Center, Helsinki Institute of Life Science, University of Helsinki, Helsinki, Finland

*Corresponding author. Department of Informatics, Bioengineering, Robotics and System Engineering (DIBRIS), University of Genoa, Via all'Opera Pia, 13, 16145 Genoa, Italy, rosella.tro@edu.unige.it

†These authors have contributed equally to this work and share last authorship

Abstract

Diffusion Kurtosis Imaging (DKI) has undisputed advantages over more classical diffusion Magnetic Resonance Imaging (dMRI), as witnessed by a fast-increasing number of clinical applications and software packages widely adopted in brain imaging. However, in the neonatal setting, DKI is still largely underutilized in Spinal Cord (SC) imaging because of its inherently demanding technological requirements.

Due to its extreme sensitivity to non-gaussian diffusion, DKI proves particularly suitable for detecting complex, subtle, fast microstructural changes occurring in this area at this early and critical stage of development, and not identifiable with only DTI. Given the multiplicity of congenital anomalies of the spinal canal, their crucial effect on later developmental outcome, and the close interconnection between SC region and the above brain, managing to apply such a method to neonatal cohort becomes of utmost importance.

This study will (i) review the current methodological challenges associated with application of
advanced dMRI methods like DKI in early infancy, (ii) illustrate the first semi-automated pipeline
built on Spinal Cord Toolbox (SCT) for handling with DKI data of neonatal SC, from acquisition
setting to estimation of diffusion measures, through accurate adjustment of processing algorithms
customized for adult SC, and (iii) present preliminary results of its application in a pilot clinical
case study.
Results of this application agree with findings achieved in a corresponding adult survey, thus
confirming validity of adopted pipeline and diagnostic value of DKI in pediatrics.

Keywords

37 Spinal Cord, DTI, DKI, image processing pipeline, neonatal imaging, punctate white matter lesions

1. Introduction

In recent years, an increasing number of works in the field of neuroimaging are stressing the
importance to move beyond simplistic assumptions of Diffusion Tensor Imaging (DTI) model (1)
towards more advanced diffusion MRI (dMRI) methods, among which DKI (2) is one of the most
promising (3–5).
Within existing non-standard techniques, DKI has indeed turned out to be especially suitable for
imaging of SC, a structure where the assumption of Gaussian diffusion fails (6). Indeed, the
presence of Gray Matter (GM) in the central portion contains a significant amount of cell
membranes and organelles that limits diffusion to fewer directions. Taking into account
pathological processes not following a Gaussian distribution, DKI provides a better understanding
of the underlying micromolecular environment. In fact, it exhibits increased sensitivity in
microstructural assessment of both White Matter (WM) and GM (7). Hence this susceptibility
translates into an increased amount of diagnostic information, beyond that obtained with routine
diffusion metrics.
Reduced field-of-view mitigates susceptibility artifacts and cardiac/respiratory gating allows to
overcome most of the methodological challenges inherent to adult SC imaging (8). Thanks to these
strategies, DKI by now represents a promising tool for studying a plethora of spine disorders with
minor modifications to protocol parameters in use for brain imaging (9–14).

The scenario becomes definitely more complicated when attempting to translate this imaging technique to the pediatric clinical setting (15,16). Typical issues inherent to SC district include small cross-sectional area requiring high spatial resolution, interface between regions with different magnetic properties, Partial Volume Effect (PVE) of pulsating Cerebro-Spinal Fluid (CSF) with each heartbeat, and bulk physiologic motion due to proximity of heart and pulmonary parenchyma. This scenario is further complicated here from a multiplicity of factors related to the age range under analysis. Children in general have smaller anatomical structures - which in turn might result in higher risk of radiofrequency heating effects - and move more frequently (e.g. tongue sucking motion). On the other hand, artifact-reducing techniques (i.e. cardiac gating, respiratory compensation, and suppression sequences) (17) are often unfeasible since time-consuming, and sedation is typically not desirable. All aforementioned issues result in artifact-laden, low-signal images, which are often suboptimal for diagnostic evaluation.

Adopted solutions for improving image resolution and reducing artifacts comprise induction of natural sleep (by feeding the patient immediately before MR examination), the use of a vacuum fixation pillow to wrap the patient, and use of special ear muffs to protect from noise. However, the main requirement when handling with pediatric dMRI data is the choice of a proper acquisition protocol tailored for pediatric imaging, made up of low angular resolution, low b-values and few gradient directions, likewise in pediatric brain (18) in order to minimize scan time. Nevertheless, this forced time minimization clashes with specific requirements of advanced diffusion methods in terms of acquisition sequences.

Indeed, contrary to DTI, DKI and higher-order diffusion models requires multi-shell high angular resolution diffusion imaging (HARDI) sequences (19), typically involving at least three non-zero b-values distributed on hundreds of gradient directions, grouped in shells. This implies longer acquisition time, straining the feasibility of advanced dMRI methods in pediatrics. Resorting to optimized acquisition sequences (20), often combined with state-of-the-art techniques such as Parallel Imaging (21) and Multi-Band, can significantly increase acquisition speed and reduce artifacts. However, these advanced technologies are not always available in a general hospital due to high costs and technical limitations.

If extension of DTI to the pediatric SC has shown promising results in a wide range of clinical conditions, as evidenced by the increasing number of works on the topic (22–31), what immediately stands out while reviewing literature on pediatric SC is the absence of studies concerning DKI and particularly applied to the neonatal period (0-1 month).

To the best of our knowledge, the only published work on pediatric DKI (32) is limited to grown-up

children (6-16 years), whose larger anatomical structures and reduced source of movements enable

better image quality and longer scan times. Indeed, in newborns, SC dimensions themselves - 24 cm

average length and 4.4 mm diameter, which can further decrease in case of malformations (33)- are

enough to conceive amplification of aforementioned technical issues and thus justify the lack of

research towards this direction.

However, the ability of DKI to offer additional and complementary information to DTI may bring a

significant contribution in investigating such decisive and delicate stage of development, especially

if we consider the wide range of developmental anomalies of the spinal canal affecting infants at

birth (34–37).

It is on this premise that we conceived our work, whose aim is to show the feasibility of applying

DKI to neonatal SC within clinical routine with all the issues that this entails.

We thus introduce here the first complete pipeline specifically adapted to neonatal imaging acquired

for diagnostic purposes. Applicability and clinical validity of proposed method has been evaluated

analyzing a specific clinical case-study concerning a condition common to preterm birth, in

collaboration with Neuroradiology Unit of Giannina Gaslini Children's Hospital of Genova.

Specifically, we assessed effects of WM brain lesions typical of Periventricular White Matter Injury

(PWMI) on below cervical SC tracts by comparing diffusion measures between pathological

patients and healthy controls. Our findings, though preliminary, confirm ability of DKI model in

capturing subtle pathological alterations.

Since there are currently neither available protocols nor standardized methodological pipelines for

performing DKI in the infant SC, this methodological outline may at least serve as a proof-of-

concept, stressing the need for infant-specific data acquisition and processing guidelines in order to

translate DKI of neonatal SC into routine clinical practice.

2. Materials and Methods

2.1 Subjects

115 Infants whose data have been used to disclose each step of the pipeline have been enrolled since

116 August 2019 and scanned with 3.0 T MR scanner using a 32-channel head array coil (Ingenia Cx,

117 Philips, Best, the Netherlands) at the Neuroradiology Unit of Giannina Gaslini Children's Hospital

118 of Genova. Conventional MRI and DKI were performed in 17 pre-term infants (28.1 to 36.7 weeks

Gestational Age (GA); scanned at Term-Equivalent Age (TEA)). Details about subjects' demographics are reported in Table 1.

This single-center study was carried out in accordance with the recommendations of "Comitato Etico Regione Liguria, Genova, Italy" with written informed parental consent obtained for each infant prior to examination in accordance with the Declaration of Helsinki. Subjects were spontaneously breathing during examination; free-flowing oxygen was administered for all the duration of MRI session if necessary. Throughout the course of the examination, newborns were subjected to constant monitoring of the oxygen saturation and heart rate, by pulse oximeter and three-electrode electrocardiographic monitoring, respectively.

Exclusion criteria included obvious motion artifacts, oblique positioning, an incomplete imaging process or a low SNR. We opted for focusing exclusively on upper SC portion, by only acquiring the cervical district (C1-C7), affected by a significantly lower degree of physiologic motion than the dorsal and lumbar portions (38).

In consensus with a board-certified pediatric neuroradiologist, we performed QC for each of the pipeline's steps.

	Unhealthy (n=9)	Healthy (n=8)
Gender (M/F)	6/3	4/4
Mean GA (range; week)	30.3±2.6 (28.1-35.0)	31.8±3.1 (28.3-36.7)
Mean PNA (range; week)	9.2±3.9 (0.1-11.7)	8.6±3.6 (2.0-10.7)
Mean PMA (range; week)	39.4±1.6 (35.1-40.6)	40.4±1.3 (38.7-42.4)

M/F = number of male and female infants; GA = gestational age; PNA= postnatal age; PMA = postmenstrual age.

Table 1: Demographics features of infants

2.2. Full pipeline description

Our pipeline integrates MRtrix3 (v.3.0.1) (39) for setting of dMRI acquisition sequence, Spinal

Cord Toolbox (SCT, v. 5.3.0, https://github.com/neuropoly/spinalcordtoolbox) (40) for all

processing steps specific to the SC, and Diffusion Imaging in Python (Dipy, v.1.4.0) (41) for

denoising as well as computation of diffusion metrics.

Output of key processes, such as motion correction, segmentation and registration with atlas, can be

checked through an interactive SCT Quality Control (QC) module, which automatically generates

reports consisting in HTML files, containing a table of entries and allowing to show, for each entry,

animated images (background with overlay on and off) for data quality validation.

In our methodological pipeline we have opted for mainly relying on SCT, being currently the only

existing fully-comprehensive, free and open-source software dedicated to the processing and

analysis of multi-parametric MRI of the spinal cord (42–53) successfully employed in a plethora of

clinical applications concerning adult SC (54–63).

An overview of our image processing pipeline highlighting key features is shown in Figure 1. Since

SCT algorithms are validated in adult imaging, we specifically customized each processing step to

our neonatal scans. Our pipeline thus represents, to the best of our knowledge, the first semi-

automated ad-hoc procedure for imaging of neonatal spine. A fully automatic workflow is not

feasible here: acquisition time constraints, available scanner features, and subsequent image quality

require inevitable although minimal and highly reproducible manual interventions.

Figure 1. Overall Processing Pipeline: designed pipeline allows complete handling of DKI scan of neonatal Spinal Cord from acquisition setup to preprocessing, processing and postprocessing

2.3. Customized acquisition setting

In order to minimize macroscopic movement artifacts, all recommended guidelines for pediatric imaging have been adopted. So as to protect infants from acoustic disturbances caused by MR sequences, we resorted to baby earmuffs and silicone paste for hearing aids. Furthermore, we avoided most of motion by swaddling infants and by placing airbags around the baby's head. In addition, protective pads have been placed between the magnet and the patient. All these contribute to create a comfortable and warm rest environment, minimizing the chance of free movements.

MRI was performed when possible during spontaneous sleep by exploiting the administration of breast milk or formula about 30 minutes before the start of the exam. In case of spontaneous sleep failure, in order to minimize macroscopic movement artifacts, the instrumental examination was performed under mild sedation by orally administering Midazolam at 0.1-0.2 mg/kg diluted in glucose solution 33%, subject to signature of informed consent from parents and applied by expert trained nurses.

Given the lack of a specific acquisition protocol for DKI of neonatal SC, we designed the diffusion-
weighting scheme in collaboration with the neuroradiologists at Giannina Gaslini Hospital. One
constraint we had to deal with was the impossibility to perform optimized variants of Spin-Echo
Echo Planar Imaging (SE-EPI) sequence (i.e. reduced FOV or spatially selective techniques) (20)
on Philips Ingenia scanner. Therefore, minimization of scan duration was our main focus in order to
suppress motion and fast CSF pulsation artifacts typical of newborns.
We thus tested different versions of diffusion-weighted gradient scheme, adopting optimal trade-off
between Fiber Orientation Distributions profile (FODs, estimated with Mrtrix3 using Multi-Shell
Multi-Tissue Constrained Spherical Deconvolution (MS-MT CSD)), image quality and scan time.
We generated each multi-shell diffusion gradient table through Mrtrix3 script *gen_scheme*, taking
as inputs the number of phase-encoding directions to be included in the scheme (for most scanners,
including ours, typically 1), the b-value of the shell, and the number of directions to include in the
shell. This procedure ensures uniform sampling by maximizing uniformity within shells using a
bipolar electrostatic repulsion model for optimal angular coverage.
For further reducing acquisition time without significantly affecting image quality, we applied the
MultiBand slice acceleration technique (64) (https://www.usa.philips.com/healthcare/resources/
landing/compressed-sensecombined).
Moreover, spatial saturation bands have been used to suppress strong signals originating from the
subcutaneous lipids to enable volumetric region coverage.
The final version of diffusion acquisition scheme is displayed in Figure 2 as well as reported in
Table 2, and includes 6 b=0; 13 b=700 and 13 b=2100 s/mm² for a duration of 4 minutes 30
seconds. This allowed acquisition of high in-plane resolution axial diffusion weighted images,
where b=0 scans could be well discriminated from non b=0 volumes and anatomical SC features are
sharp.
This is the best possible solution that combines conflicting requirements of complex multi-shell
acquisition and minimized scan time.
201
202

203
204
205　　　**Figure 2**. **3D view of final diffusion acquisition scheme:** directions of diffusion-sensitizing
206　　gradients relative to each b-value are displayed in three different colors as reported in the legend.
207　　　Units are in s/mm^2. Markers indicates polarity: dots are the polarities in the set, asterisks their
208　　　　　　　　　　　　　　　　　　　　opposite
209　Along with dMRI, we also acquired a high-resolution structural image as anatomical reference. The
210　definitive MRI protocol thus consisted in a Turbo Spin Echo (TSE) 3D T1-weighted image
211　followed by a DKI series whose details are listed in Table 1.

	3dT1	**DKI**
TR/TE (s)	0.6/0.026337	3.378/0.128
Diffusion Scheme (s/mm²)	–	6 b=0, 13 b=700, 13 b=2100
Flip Angle (°)	90	90
Acquisition Matrix	251*512	160*160
Multi-Band Factor	–	2
Acquisition Resolution (mm)	0.50*0.38*0.38	0.8*0.8*4
Slice Thickness (mm)	0.38, 0.5 mm gap	4, without gap
Slice Orientation	–	axial
# Slices	512	24
Field of View (FOV) (mm)	195x195x126	128x93x96
Total Scan Time	4 minutes 5 s	4 minutes 30 s

212 **Table 2: Data Acquisition Details** for both structural 3D T1 and DKI image

2.4. Preprocessing

2.4.1. Denoising

SC imaging is characterized by low SNR, which can hamper accurate, repeatable, quantitative measurements. Moreover, models such as DKI are susceptible to noise and signal fluctuations often leading to degeneracies in estimation of derived parameters. SNR further lowers in case of neonates, due to relatively high overall free water content, and denoising approaches based on Principal Component Analysis (PCA) are inapplicable due to a reduced number of diffusion gradient directions.

Therefore, we adopted Patch2Self, a recently proposed self-supervised learning denoising method that outperforms existing non supervised methods. (65).

A unique advantage of Patch2Self is the lack of requirement for selecting or calibrating an explicit model either for noise or diffusion signal, so that it can be applied at any step in the pre-processing pipeline. The only assumption it relies on is randomness and uncorrelation of noise across different gradient directions. Its framework consists in holding out one volume and using patches from all other volumes to predict the center of the patches of the held-out volume using a regressor. This denoiser has already showed a significant improvement in repeatability and conspicuity of pathology in diffusion volumes and quantitative DTI metrics for adult SC (66).

Here, we chose to apply Patch2Self as first preprocessing step on raw data since it showed to offer highest SNR. The method is implemented in Dipy v.1.4.0 and applied with Ordinary Least Squares (OLS) regressor, since recommended for SC imaging (Figure 3).

Original	Denoised Output	Residuals

233
234 **Figure 3. Visual inspection of denoising:** The denoising of Patch2Self is compared against the

235 original noisy image along with their corresponding residuals. Notice that Patch2Self does not show

236 any anatomical structure in the corresponding residual plots likely neither introducing structural

237 artifacts

2.4.2. Cropping

SC scans usually include also cerebral areas such as medulla and cerebellum due to their proximity

with cervical SC. In order to exclusively focus on the area of interest excluding undesired voxels,

as a first preprocessing step, we thus recommend to apply to DKI images SCT function

sct_crop_image allowing also to fasten subsequent processing. Lower and higher bounds for

cropping along the three spatial coordinates can be specified via command line in order to select the

same area of interest (i.e. cSC) for all the cohort, considering FOV positioning is consistent across

subjects.

Specifically, in the case of our scans, FOV reduction allowed to exclude upper non-spinal areas (i.e.

cerebellum) as well as lower spinal levels whose corresponding slices are not usable due to poor

image quality (Figure 4a).

2.4.3. Motion Correction

Subjects' immobilization and anesthesia successfully minimized motion in our acquisitions.

However, since dMRI data are analyzed at a voxel level, residual intrascan and/or interslice motion

can adversely affect accuracy of the modeled results. We thus resorted to SCT complex motion

correction framework *sct_dmri_moco*, based on a combination of tools.

First of all, *SliceReg* algorithm estimates slice-by-slice translations while ensuring regularization

constraints along z axis. The latter is achieved using a polynomial function (order specified by the

user, flag *-param*). This method was shown to offer better accuracy and robustness than rigid-body

transformations and non-regularized slice-by-slice registration, respectively (40).

Moreover, motion correction in SCT includes another feature first proposed in (67) to improve the

robustness of registration in high b-value diffusion MRI data such as DKI datasets. It consists in

grouping adjacent volumes and estimating the transformation relying on these successive subsets

(typically from 3 to 5 volumes) averaged together (flag *-g*).

This robust slice- and group- wise motion correction works successfully also in case of neonatal

scans and it is hence applied here with default parameters: grouping of 3 successive dMRI volumes,

regularization with 2^{nd} order polynomial function, unitary smoothing kernel (1 mm), and final

spline interpolation (flag -x), with the exception of the metric used for registration (Figure 4b).

Indeed, Cross Correlation (CC) has been selected as similarity metric given its better performance

with respect to Mean Squares or Mutual Information (default option) at the expense of

computational time.

Since *sct_dmri_moco* works through iterative average over groups of successive slices in order to

increase the SNR of the target image, its output includes a 3D volume corresponding to the mean

from DKI slices. These motion-corrected average DKI data will serve as input for subsequent

segmentation thanks to its excellent cord contrast.

Thanks to the limited duration of our acquisition and to adopted procedures for minimizing

movement throughout the exam, amount of motion is very limited in our images. As a result,

outcome of motion correction step does not significantly differ from raw DKI image by visual

assessment. However, this represents a crucial step in case of longer scans more prone to source of

motion artefacts.

2.4.4. Segmentation

Proper segmentation of SC is decisive for subsequent steps of template registration and computation

of metrics along the cord.

Detection of SC has turned out to be a critical step, since standard SCT algorithm *propseg*, based on

multi-resolution propagation of tubular deformable models (68), is trained for adult spine.

Given the reduced size of neonatal SC and the low contrast between the spine and CSF, default

segmentation method fails in several slices even after modulating the algorithm parameters - e.g.

manual initialization of spinal cord centerline through interactive viewer (flag *-init-mask*), selection

of SC radius size (flag *-radius*) or cord rescale (flag *-rescale*).

We thus resort to a more recent and advanced method of SC extraction, based on deep learning

sct_deepseg_sc (69). This fully automatic segmentation framework was conceived for detecting SC

and intramedullary MS lesions from a variety of MRI contrasts and resolutions.

It is composed of a cascade of two convolutional neural networks (CNN), specifically designed to

deal with spinal cord morphometry: the first detects the cord centerline and reduces the space

around the spinal cord (for better class balance), and the second segments the cord.

Segmentation results outperformed *sct_propseg*, showing higher robustness to variability in both

image parameters and clinical conditions.

Thanks to its versatility, the application of this method results suitable also for neonatal imaging,
allowing robust and accurate segmentation of our scans without ever the need of additional
parameters but just specifying the kind of image contrast as _dwi_ (flag -_c_) (Figure 4c).
In case of failure of SC detection, we necessarily opt for manual correction of problematic slices on
FSL editor (_FSLeyes_) (Figure 4d,e).
This is the case of five subjects within our cohort: to validate the quality of segmentation, we
checked the QC feature on our MRI images across subjects and noticed some local segmentation
leakage - (related to the onset of artifacts at acquisition phase and not to a flaw with the algorithm) -
in a few slices and hence corrected it manually.
306
307

309   **Figure 4**. **Preprocessing:** DKI scan through preprocessing steps for one example subject: (a) FOV
310   reduction; (b) Motion Correction; (c) Segmentation: Deep Learning Segmentation algorithm
311   generally achieves satisfactory results in SC detection; (d) Example of artefactual slice due to a
312   poor fat saturation, causing the fat to alias on the spinal cord area and (e) requiring manual
313   correction of segmentation

2.5. Processing

2.5.1. Vertebral Labeling

After segmentation, labeling of vertebral levels or discs is the second mandatory step in order to

match the template to the subject's MRI (template registration).

Two vertebral levels are necessary for registering data to the template. Each of these two landmarks

consists of a voxel placed in the middle of the SC, at the level of the corresponding mid-vertebral

body, and assigned a relative number starting from 1 for C1 vertebra. However, SCT recently

introduced the possibility to alternatively use inter-vertebral disc labels with the analogous

procedure of reference numbered voxels.

We perform this step on 3D T1w images in order to achieve better accuracy given their higher

overall quality and contrast compared to DKI ones, where vertebral discs are not clearly

identifiable.

Labeling from 3D T1w anatomical image is possible as it turned out to match relatively well along

the superior-inferior (z) axis, the target direction of disc labeling, with the DKI scan (not along the

Anterior-Posterior (AP) or Right-Left (RL) direction, Figure 5).

Vertebral labeling is typically done using an automatic method *sct_label_vertebrae*, that finds C2-

C3 disc, and then locates neighbouring discs using similarity measure with the PAM50 template at

each specific level (70).

Default SCT procedure *sct_label_vertebrae* fails in automatically detecting C2-C3 vertebral disc

once again because of the small size of spines at issue and low image contrast compared to adults.

Therefore, we manually create labels with the command *sct_label_utils* through interactive viewer

option provided by SCT (flag *-create-viewer*) with little to no waste of time.

Specifically, vertebral labeling was created at the posterior tip of the top of C1 vertebra and at C3-

C4 disc, centered in the cord. Manual intervention only took a few seconds per subject (Figure 3S).

Figure 5. DKI scan overlaid on structural 3dT1w image: while both images are clearly not registered along the antero-posterior direction due to the very strong susceptibility artefact, the z-location is similar: see how the bottom tip of the cerebellum is consistent for the two scans

2.5.2. Registration to PAM50 atlas

Registration between subject's diffusion and atlas space is a very demanding task in case of neonatal imaging given the lack of a specific pediatric atlas compatible with SCT (one is currently under creation, https://github.com/neuropoly/spinalcordtoolbox/issues/2530). We thus use PAM50 atlas (71), an adult template for MRI of the full SC and brainstem in the same coordinate system as the ICBM152 (MNI) brain template, allowing to conduct simultaneous brain/spine studies. It consists of a T1w, T2w, T2*w, white and gray matter probabilistic atlas and white matter atlas of tracts as well as probabilistic labeling of spinal levels. The template has been constructed from straightened SC for facilitating registration and visualization of results.

sct_register_to_template is the main command for registering one subject to the template and vice versa, since it outputs the forward and backward warping fields. We choose subject's native diffusion space as target of registration transforms since the straightening required by opposite strategy would cause through-plane interpolation errors which would bias following extraction of diffusion measures (72).

Moreover, we suggest employing T1w atlas image for its better contrast similarity with DKI scan compared to T2w.

Application of default command does not produce satisfactory results, stressing the need to tweak all input parameters to deal with our particular contrast and resolution. Given the presence of

artifacts and some inherent features (e.g., low CSF/cord contrast) that could compromise the

registration, we use SC segmentation as input for the algorithm in order to ensure maximum

robustness.

Registration is then built through multiple steps by increasing the complexity of the transformation

performed in each step (starting with large deformation with low degree of freedom and finishing

with local adjustment). Specifically, the first step consists in vertebral alignment, that is vertebral

level matching between the subject and the template on the basis of posterior edge of the

intervertebral discs provided by previous manual vertebral labeling. Second step is slice-wise center

of mass alignment between the two images, using *centermass* algorithm instead of default

centermassrot (which also includes rotation alignment) because the cord is quasi-circular and cord

angle estimation is not reliable here. The third step is R-L scaling along x axis followed by A-P

alignment to match segmentation borders along y axis, with the ultimate aim of accommodating the

very small SC size. Finally, iterative slice-wise non-linear registration is performed through non-

linear symmetric normalization regularized with b-splines (73) using information from comparison

of Cross Correlation metric (CC) between the two images, which allows SC shape refinement. Once

the algorithm completed, one can assess the quality of registration through visual evaluation and

inspection of QC module, and thus warp the template and all its objects to each subject's DKI

image (Figure 6).

Current selection of parameters and steps successfully worked for our scans since atlas registration

algorithm robustly achieved convergence as verified through inspection of QC feature.

Figure 6. Registration with PAM50 atlas and ROI detection through atlas-based approach: (a) PAM50 atlas' cord segmentation binary mask, (b) WM and (c) GM probabilistic masks warped to subject's DKI motion-corrected mean image

2.5.3. Computation of diffusion metrics

The end point of previous preprocessing and processing steps is computation of diffusion parametric maps, from which to extract quantitative summary measures requested by the particular study in question. We estimate diffusion parametric maps through DIPY software (v. 1.4.0), (41) To avoid unnecessary calculations on the background of the image, we use a mask created by dilating spinal cord segmentation (through *sct_maths* command) because values outside the binary cord mask are important for proper account of PVE, having to be minimized in every possible way (74). Indeed, this phenomenon, because of the coarse resolution of MRI with respect to SC anatomy, may make the apparent value within a boundary voxel be a mixture between the WM and CSF compartment, thus yielding a subsequent inaccurate quantification of diffusion measures.

Since DKI model involves the estimation of a large number of parameters (75) and is more sensitive

to artefacts (76), we choose to further suppress the effects of noise and artefacts before diffusion

kurtosis fitting using a 3D Gaussian smoothing (with a Gaussian kernel with fwhm=1.25) as

suggested by pioneer DKI studies (2). This also helps in addressing the issue of implausible

negative values inherent to DKI fitting (77).

The following parametric maps can thus be generated: Mean Diffusivity (MD), Axial Diffusivity

(AD), Radial Diffusivity (RD), Fractional Anisotropy (FA) and Mean Kurtosis (MK), Axial

Kurtosis (AK), Radial Kurtosis (RK), Kurtosis Fractional Anisotropy (KFA), and Mean Signal

Kurtosis (MSK).

Given the low-angular resolution data available, to ensure robustness and reproducibility of

parameters' estimates, we opted for just computing DTI measures, whose reference tensor can be

correctly estimated from at least six independent directions, and MSK, as displayed in Figure 1S.

The latter is a robust scalar kurtosis index that can be estimated from powder averaged diffusion-

weighted data and thus independent from acquisition scheme (78,79). Indeed, fitting MSDKI is well

posed without relying on the full DK tensor, which would require a minimum of 15 non-collinear

directions per b-value. Moreover, this measure is generally more robust to low SNR situations as in

case of neonatal imaging.

This novel DKI estimator can be seen as a proxy for the MK, showing to present nearly identical

contrast to the original DK metric while improving robustness and reproducibility of the kurtosis

metrics and results in parameter maps with enhanced quality and contrast. Specifically, this measure

turns out to be less sensitive to thermal noise and imaging artifacts, and thus drastically reduces

black voxels intrinsic to DKI and challenging the visual and statistical analysis of potentially

clinically relevant biomarkers of tissue integrity. Moreover, as previously pointed (80), standard

kurtosis measures do not only depend on microstructural properties but also on mesoscopic

properties such as fiber dispersion or the intersection angle of crossing fibers. In contrary, the

kurtosis from powder-average signals has the advantage of being decoupled from confounding

effects of tissue dispersion and crossing (80,81).

427 2.6 Postprocessing

Thanks to this atlas-based analysis approach, it is possible to perform cord-specific quantification of

diffusion metrics through *sct_extract_metric* command, also restricted to specific Regions of

Interest (ROIs; labels used by default are taken from the PAM50 template, e.g. WM tracts, flag *-l*),

vertebral levels (flag *-vert*) or slice (flag *-z*), according to the specific clinical needs concerned.

Along with WM and GM probabilistic masks as a whole (Figure 6b,c), normally investigated in

medical practice, one can carry out ROI detection also in specific tracts according to the clinical

question (fifteen WM tracts and three GM regions available in total for each side).

In our example, neither DKI nor structural images ensured sufficient WM-GM-CSF contrast to

perform any manual detection of ROIs in contrast to high-contrast PSIR image of (82), whose

acquisition time would be too long for neonates. Therefore, we exploited good registration outcome

for automatic delineation of ROIs through atlas-based approach.

We opted for using lateral CSTs as ROIs for consistency with (82) - though grouping together left

and right sides in order to gain robustness by increasing volume fraction as suggested in (40) - as

well as WM and GM.

We then computed averages of diffusion measures across C1-C4 vertebral levels in order to avoid a

subject-based bias: outside of these levels the registration is inaccurate and/or MRI signal may be

corrupted. We thus checked through QC module the correctly segmented slices corresponded to the

same vertebral levels across subjects, starting from the first slice containing only SC (excluding

cerebellum, Figure 2Sc).

Moreover, estimation of DTI and MSDKI weighted average metrics was limited to those slices

where spinal cord segmentation is accurate: outside segmentation mask, metrics would indeed be

irrelevant. This was obtained by multiplying segmentation mask by specific WM, GM and CSTs

atlas labels. We quantified diffusion metrics using Weighted Average (WA) estimation to minimize

PVE avoiding bias into resulting metrics by the surrounding tissues (e.g. CSF). This is one of the

recommended methods especially in case of noisy images and small tracts as in our case. We

assessed associated voxel fraction to quantify the reliability of our diffusion measures: as

demonstrated in (40), having at least 240 voxels results in an error smaller than 1%, while having

30 voxels results in an error inferior than 2%. In this example, the metrics were computed based on

average 178.3, 50.5 and 31.5 voxels in WM, GM and CSTs respectively, thus assuring sufficient

accuracy of estimates.

2.7 Case Study

Periventricular WM Injury (PMWI) is the most frequent type of brain lesion in preterm infants, and

the spatial extent and location of WM injury correlate with distinct clinical outcomes, including

cerebral palsy and motor impairment (83).

Given the strong association of WM injury with the motor function development of preterm

neonates, we assessed the hypothesis that the presence of periventricular punctate WM lesions

identified on MRI at TEA could be associated with regionally specific alterations in cSC

microstructure.

A similar approach was already used by (82) to characterize cSC microstructural abnormalities in a

cohort of adult patients with previous unilateral ischemic stroke in the vascular territory of the

middle cerebral artery (MCA). DTI and DKI diffusion measures in cSC resulted to be a valuable

imaging marker for predicting clinical outcome. In particular, significant reduction of FA and MK

was observed in the affected lateral WM bundle of the cSC, correlating with the severity of motor

dysfunction.

Accordingly, the ultimate goal of our study was to verify whether the presence of periventricular

WM lesions affects the cSC tracts development. Specifically, we aimed to compare DTI and

MSDKI measures of cSC in two groups of preterm neonates: (i) with punctate Periventricular White

Matter lesions (PWMI), and (ii) with normal brain MRI (controls).

3. Results

3.1. Population size and classification

In order to investigate clinical differences among acquired subjects, we grouped infants as follows:

i) 9 subjects with punctate PWMI and ii) 9 subjects with normal brain MRI, used as control group.

At QC phase, in accordance with the expert neuroradiologist, we opted for excluding one control

subject due to excessively poor image quality (i.e. signal leakage at C1-C3 level, Figure 2Sa,b).

Therefore, the final number of subjects under analysis amounted to 9 and 8 infants for the two

subsets, respectively.

3.2. The role of denoising

As mentioned above, neonatal imaging is inherently affected by low SNR and sensitive to imaging

artifacts. Proper denoising of scans is therefore a crucial step in the processing pipeline. Above all,

we thus focused on quantitatively assessing the contribution of Patch2Self denoiser on subsequent

analysis.

Firstly, we computed average SNR on b=0 images, namely thermal noise due to the diffusion-

related signal attenuation, across all slices belonging to C1-C4 district of our interest.

As expected, we found a remarkable increase in mean SNR after applying Patch2Self (5.88 ± 1.41

vs 14.64 ± 4.53) as displayed in Figure 7.

Figure 7. Effects of Patch2Self denoising on thermal noise. Average Signal-to-Noise Ratio

(SNR) computed on b=0 images increases, across C1-C4 vertebral levels under analysis, when

including denoising with Patch2Self algorithm in the processing pipeline

We then inspected impact of denoising on microstructure model fitting, a critical step often leading

to degenerate parameter estimates due to the low SNR of dMRI acquisitions.

Specifically, we applied DTI and MSDKI model on raw and denoised data, resorting both to

traditional Marchenko-Pastur PCA (MP-PCA) and to Patch2Self method. We opted for comparing

our denoising procedure with MP-PCA since it represents the current state-of-the-art unsupervised

method for denoising DWI (84). In order to compare the goodness of each fit, we performed a k-

fold cross-validation (85) across an axial slice (1312 voxels) of masked data for an example subject

pulled out from our dataset. In Figure 8, we depict the improvement of the   metric by simply

subtracting the goodness-of-fit   scores of fitting noisy data from Marchenko-Pastur and

Patch2Self denoised data for both DTI and MSDKI models. Patch2Self performance is not

significantly better than MP-PCA in the goodness-of-fit for DTI model (two-sided t-test with

Bonferroni correction, p=1). On the contrary, this state-of-the-art denoiser shows a significant

improvement for MSDKI (p=1.825e-04), proving once more to be particularly suitable for DKI

model.

Figure 8. Violin plots quantifying the increase in   metric after fitting downstream DTI and

MSDKI models for one axial slice in one example subject. The   improvements in each case

are plotted by subtracting the scores of model fitting on noisy data (Raw) from   of fitting each

denoised output. Note that the consistency of microstructure model fitting on Patch2Self (P2S)

denoised data is higher than that obtained from Marchenko-Pastur (MPPCA), especially as regards

MSDKI model. ns stands for non significant, *** stands for 1.00e-04 < p <= 1.00e-03

in two-sided t-test with Bonferroni correction

3.3.   Trend of diffusion measures

For consistency with (82), we considered above all trend of cSC FA and MSK measures in the two

groups of neonates as displayed in Figure 9. For the sake of completeness, in Figure 4S we also

include remaining DTI metrics.

To an initial evaluation based on the limited sample size available, considering the variation of

average metrics between controls and patients in WM, we can notice an increase in MD, AD, RD,

parallel to an overall decrease in FA and MSK in preterm neonates with PWMI.

Furthermore, range of values of DTI measures are consistent with normative values on healthy
pediatric SC (30). MD, AD and RD values are higher while FA values are lower compared to
equivalent measures on older cohorts of patients (i.e. children/ adolescents, (22,23,25,26)). This
may be partially due to sensitivity of FA to denoising, which can imply a reduction in this metric.
Moreover, this trend is in line with the simultaneous age-related decrease in MD, AD and RD and
increase in FA metrics reflecting progressive maturation, myelination and fiber packing and
thickening within the SC, similar to that observed in the brain (86,87).
Conversely, definition of a normative variation of DKI measures across age from newborns to
adults will be feasible after further investigations from early stages of development.
542

543
Figure 9. Extraction of diffusion measures within specific ROIs: (a) White Matter (WM), (b)
Gray Matter (GM) and (c) Cortico-Spinal Tracts (CSTs) ROIs overlaid on DKI motion corrected
image; (d) Scatter plots of FA and MSK in group subjects across aforementioned ROIs: colored
spots indicate single subject's value for each metric; as reported in the legend, controls' measures
are in blue, whereas Periventricular White Matter Injury (PWMI) group's in red. Units for MSK are
in mm^2/s, while FA is dimensionless. Error bars displaying mean (diamond) and standard deviation
(bars) are overlaid on scatter plots.

4. Discussion

4.1. Research question

In the present study we tested the pipeline of SC DKI analysis in a group of neonates with PWMI, a form of mild WM injury frequently diagnosed in preterm infants. PWMI are seen at brain MRI as small, focal, multiple alterations of signal intensity (high on T1 and/or low on T2) in periventricular WM. Long-term outcome of neonates with PWMI seems to be related to the number of lesions, their pattern, and their localization. Of note, several studies have showed that greater lesion load of PWM and the involvement of frontal WM are associated with higher risk of adverse neurodevelopmental outcome, affecting both motor and cognitive functions (88). Moreover, periventricular WM lesions in preterm neonates are associated with region-specific changes in MD, FA, RD, and AD in several cerebral WM tracts that might explain the abnormal development of long-term neurological functions (89). Specifically, the involvement of pyramidal tract fibers in the periventricular WM has been demonstrated to be a relevant factor for motor dysfunction in children with PWMI (90). In our study, we found that microstructural changes can be detected by using an advanced DKI analysis also in the GM and WM of cSC of preterm neonates with PWMI studied at term-equivalent age, thus demonstrating that DKI parameters could be used as markers to unravel underlying microstructural lesions not visible on conventional structural MRI. Moreover, our preliminary findings confirm the hypothesis that in preterm neonates with PWMI WM microstructure alterations extend beyond the immediate area of periventricular injury, widening distally also in the cSC (91,92). Further analyses on a wider cohort of neonates are necessary to confirm these preliminary results, and specifically to prove if microstructural changes in cSC in preterm neonates with PWMI correlate with long-term neurological outcome.

4.2. Study significance

Here, we present the first application of DKI to neonatal SC through a pipeline able to perform complete processing on a subject within a clinically acceptable time (10 minutes average with the current setup). As regards acquisition setting, we were able to perform a time-consuming technique like DKI using a short diffusion sequence which minimizes patient's physiological motion and which likely reflects a standard clinical scenario devoid of latest technologies in terms of acquisition sequence optimization. With regard to image processing, we opted for creating this pipeline using SCT since it represents the only existing comprehensive, free and open-source software dedicated to the processing and analysis of SC multi-parametric MRI data. Adaptation of

each image processing tools already in use for adult subjects through appropriate tuning of
parameters turned out to be feasible. Indeed, it allowed to successfully overcome all the issues
mentioned in the 'Introduction' section inherent to imaging of SC and exacerbated in case of
neonatal setting, even for the most challenging steps like segmentation or registration to atlas. We
were thus able to quantify diffusion measures within specific ROIs using an atlas-based approach
which presents undisputed advantages compared to usual manual drawing of ROIs. It is automatic
and thus highly reproducible, it is not biased by the user experience and knowledge of the anatomy,
it is much faster than long and tedious manual delineation of ROIs and it allows to account for PVE.
Preliminary application of our method to a limited number of subjects has yielded physiologically
plausible findings despite suboptimal acquisition setting. In particular, differences in DTI and
MSDKI measures observed between PWMI neonates and controls confirm the hypothesis that
PWM injuries in the premature brain can be associated with microstructural alterations of both GM
and WM of neonatal cSC at term-equivalent age. These findings also highlight the importance of
complementary analysis of DTI and DKI metrics for a more accurate characterization of biological
tissues.
601

4.3   Added value of DKI

Results about feasibility of DTI and MSDKI analysis in neonatal SC with subjects collected so far
are preliminary but promising and demonstrate the clinical utility of combining DTI and DKI in the
characterization of spinal cord pathologies.
FA reduction parallel to MD increase in patients is an expected finding consistent with existing
literature and attributable to degeneration of diffusion barrier and loss of diffusion directionality.
Regarding DKI, it is interesting to note how MSK metric, although yet underused in clinical
studies, proves to be more sensitive than FA to microstructural changes in PWMI subjects by
further reducing its value in all considered ROIs.
This DKI estimator thus confirms DKI increased sensitivity in capturing alterations related to
pathology, also far from the lesion site.
Another crucial food for thought concerns observing how the presence of a WM lesion in the brain
causes subsequent alterations not only in cSC WM but also in GM, as evidence of the strong
association between brain and spine. In this respect, resorting to DKI measures becomes of utmost
importance given kurtosis sensitivity to structural changes in isotropic tissues such as GM.

Indeed, range of variation of MSDKI metric from controls to PWMI is overall higher in GM than

that of corresponding DTI measures.

This is an expected finding: DTI is extremely suitable for capturing highly directional diffusion

typical of WM bundles but shows low sensitivity to isotropic diffusion, whereas DKI allows

specific assessment of non-gaussian diffusion typical of GM areas (93).

Such findings once again stress the importance of combining DTI and DKI metrics as

complementary sensitive biomarkers in order to fully exploit the potential of dMRI compared to

conventional MRI.

However, a more comprehensive corroboration and explanation of our results is expected after

collecting an adequate number of subjects to carry out a robust statistical survey. An in-depth

interpretation of the single metrics is out of the scope of this paper. Here, we just dwell on exploring

the comparison with the work on adults which served as a starting point.

4.4   Comparison with adult study

Since MSK has proved to be an exact approximation of MK, here we assimilate it with standard

DKI metric used in (82).

Both FA and MSK trend agree with just mentioned study. Specifically, they both exhibit a

reduction in WM, GM and CSTs in case of pathology (Figure 9).

Since FA is known to be an index of structural integrity (94) and MSK a marker of tissue

microstructure complexity (2), obtained findings suggest that, in case of overlying WM brain lesion,

a loss of integrity and complexity is registered also in SC WM tracts below with a more isotropic

diffusion pattern due to disruption of WM tracts. Also MD, AD and RD follow the same trend, with

an increase in case of lesioned subjects, as in (82). This biologically plausible hypothesis already

verified for adults would thus also subsist in infants.

4.5. Study limitations

Present pipeline has been designed with current instruments at disposal and with given data whose

resolution is high in axial plane but whose contrast is very low due to short acquisition time. As a

result, any improvement in acquisition setup our pipeline will bring to even stronger and more

comprehensive results.

Major flaws of this procedure consist in basing on an adult atlas, where the exact location of tracts

may not perfectly correspond to neonatal images despite the good adjustment of registration

parameters. Current pipeline will definitely benefit from introduction of a pediatric atlas into SCT.

Moreover, image quality could be further improved: scans we acquired are extracted from routine

clinical protocol and consequently prone to noise and artifacts due to short acquisition time dictated

by clinical needs and to the lack of specific spatially-selective MR sequences.

Starting from more advanced hardware tools may significantly increase image quality and thus

accuracy of estimated metrics. Resorting to optimized acquisition sequences would also allow to

increase resolution of HARDI acquisition scheme and thus to exploit all standard DKI measures,

which can in turn increase the amount of diagnostic information.

We acknowledge the protocol in use to be on the edge for HARDI schemes required by DKI.

However, this represents an attempt to customize advanced dMRI acquisition setting within a

clinical routine protocol, already long in itself since made up of multiple MRI sequences in order to

increase diagnostic possibilities. Nevertheless, we appropriately addressed this issue at DKI tensor

and measures computation phase to ensure reliability and accuracy in their estimates.

4.6. Future developments

Validation of current pipeline can be made by testing it to a larger cohort of subjects, possibly

investigating also lower SC tracts, including thoracic and lumbar districts, and extending studies to

different clinical cases, preferably focusing on a determined pathology. For example, it would be

interesting to explore long-term correlations between DKI measures and specific clinical scores as

done in (82), where diffusion measures have been related to motor performance indexes.

A further step may be adapting this analysis pipeline to other promising higher-order diffusion

models requiring multi-shell acquisition such as NODDI (95).

5. Conclusion

676   In this work, we have showed how accurate adjustment and parameters' tuning of processing

677   algorithms customized for adult SC opens up new horizons in exploiting increased ability of

678   advanced dMRI models, also in neonatal domain, where they had never been utilized before.

Indeed, even starting from low quality data acquired for diagnostic purposes and thus suboptimal, we were able to extract from DKI information of some diagnostic use.

The case study proposed in this paper is just an example of the potential relapses of this semi-automated pipeline, which paves the wave for applying advanced dMRI models to neonatal setting in a wide range of potential clinical applications. In particular, the possibility of successfully exploiting increased sensitivity and sensibility inherent to DKI methodology also into neonatal setting would indeed be extremely useful for throwing light on complex diseases related to this critical phase of development and to deepen the knowledge about the relationship between brain and SC at birth.

Declaration of interest

Declarations of interest: none

Author contributions

Rosella Trò: Conceptualization, Methodology, Software, Writing - Original Draft, Visualization. **Monica Roascio**: Conceptualization, Supervision. **Domenico Tortora**: Conceptualization, Validation, Investigation, Resources, Supervision. **Mariasavina Severino**: Validation, Investigation, Resources, Supervision. **Andrea Rossi:** Validation, Investigation, Resources, Supervision. **Julien Cohen-Adad:** Writing- Reviewing and Editing, Supervision. **Marco Massimo Fato:** Conceptualization, Writing- Reviewing and Editing, Supervision. **Gabriele Arnulfo:** Conceptualization, Writing- Reviewing and Editing, Supervision. **These last two authors contributed equally.**

Funding sources

This research did not receive any specific grant from funding agencies in the public, commercial, or not-for-profit sectors.

Acknowledgements

707 The authors would like to thank Prof. Luca Antonio Ramenghi (Neonatal Intensive Care Unit,
708 IRCCS Istituto Giannina Gaslini, Genoa, Italy e Department of Neurosciences, Rehabilitation,
709 Ophthalmology, Genetics, Maternal and Child Health (DINOGMI), University of Genoa, Italy) and
710 the LIFT (Laboratorio di Imaging Funzionale 3 Tesla).

Data and code availability

713 Data used in this study is not able to be made openly available due to privacy restrictions of clinical
714 data imposed by the Gaslini Hospital's administration. Regarding code availability, a specific
715 pipeline compatible with SCT is integrated in an open-access GitHub repository within "SCT-
716 pipeline", a site gathering various pipelines compatible with SCT for processing MRI data
717 (https://github.com/sct-pipeline/pediatric-genova).

718
719

Supplementary material

721
722 **Figure 1S. Diffusion and kurtosis maps at the mid-C3 level for one example subject:** Units for
723 MD, AD and RD are $\mu m^2/s$, for MSK mm^2/s, while FA is dimensionless.

Figure 2S. Quality control: (a) Example of excluded DKI scan and (b) relative SC segmentation show signal loss across multiple slices as the coronal plane is not overlapping with the cord (ie: lordosis); (c) QC of C1-C4 levels: axial slices under analysis correspond to the same cervical levels for all subjects as shown in three example subjects.

Figure 3S. Vertebral labeling: Manual labeling of top of C1 vertebra and C3-C4 disc from

graphical user interface integrated in SCT

Figure 4S. Extraction of diffusion measures within specific ROIs: Scatter plots of DTI in group

subjects across aforementioned ROIs: coloured spots indicate single subject's value for each metric;

as reported in the legend, controls' measures are in blue, whereas Periventricular White Matter

Injury (PWMI) group's in red. Units for MD, AD, RD are in mm^2/s. Error bars displaying mean

(diamond) and standard deviation (bars) are overlaid on scatter plots.

References

1. Tournier JD. Diffusion MRI in the brain – Theory and concepts. *Progress in Nuclear*
Magnetic Resonance Spectroscopy (2019) doi:10.1016/j.pnmrs.2019.03.001

2. Jensen JH, Helpern JA, Ramani A, Lu H, Kaczynski K. Diffusional kurtosis imaging: The
quantification of non-Gaussian water diffusion by means of magnetic resonance imaging.
Magnetic Resonance in Medicine (2005) doi:10.1002/mrm.20508

3. Farquharson S, Tournier JD, Calamante F, Fabinyi G, Schneider-Kolsky M, Jackson GD,
Connelly A. White matter fiber tractography: Why we need to move beyond DTI. *Journal of*
Neurosurgery (2013) doi:10.3171/2013.2.JNS121294

4. Loucao R, Nunes RG, Neto-Henriques R, Correia M, Ferreira H. Human brain tractography:
A DTI vs DKI comparison analysis. in doi:10.1109/enbeng.2015.7088820

5. Mader I, Urbach H. Walk the line: From diffusion imaging to the microstructure of the brain.
Clinical Neuroradiology (2013) doi:10.1007/s00062-013-0265-3

6. Cohen Y, Anaby D, Morozov D. Diffusion MRI of the spinal cord: from structural studies to
pathology. *NMR in Biomedicine* (2017) doi:10.1002/nbm.3592

7.   Wu EX, Cheung MM. MR diffusion kurtosis imaging for neural tissue characterization. *NMR*
in Biomedicine (2010) doi:10.1002/nbm.1506

8.   Taber KH, Herrick RC, Weathers SW, Kumar A, Schomer DF, Hayman LA. Pitfalls and
Encountered Clinical Artifacts in Imaging of the Spine1. *Radiographics : a review*
publication of the Radiological Society of North America, Inc (1998)

9.   Panara V, Navarra R, Mattei PA, Piccirilli E, Cotroneo AR, Papinutto N, Henry RG, Uncini
A, Caulo M. Spinal cord microstructure integrating phase-sensitive inversion recovery and
diffusional kurtosis imaging. *Neuroradiology* (2017) doi:10.1007/s00234-017-1864-5

10. Li D, Wang X. Application value of diffusional kurtosis imaging (DKI) in evaluating
microstructural changes in the spinal cord of patients with early cervical spondylotic
myelopathy.   *Clinical   Neurology   and   Neurosurgery   (2017)*
doi:10.1016/j.clineuro.2017.03.015

11. Bester M, Sigmund E, Tabesh A, Jaggi H, Inglese M, Mitnick R. Diffusional Kurtosis
Imaging of the cervical spinal cord in multiple sclerosis patients. *Proc Intl Soc Mag Reson*
Med (2010)

12. Bester M, Sigmund E, Tabesh A, Jaggi H, Inglese M, Mitnick R. Diffusional Kurtosis
Imaging of the cervical spinal cord in multiple sclerosis patients. in

13. Raz E, Bester M, Sigmund EE, Tabesh A, Babb JS, Jaggi H, Helpern J, Mitnick RJ, Inglese
M. A better characterization of spinal cord damage in multiple sclerosis: A diffusional
kurtosis imaging study. in *American Journal of Neuroradiology* doi:10.3174/ajnr.A3512

14. Li D-W, Wang X-M. Progresses of diffusion kurtosis imaging in spinal cord injury. (2015)
31:1422–1425. doi:10.13929/j.1003-3289.2015.09.036

15. Sorantin E, Robl T, Lindbichler F, Riccabona M. MRI of the neonatal and paediatric spine
and spinal canal. *European Journal of Radiology* (2008) doi:10.1016/j.ejrad.2008.06.032

16. Thukral BB. Problems and preferences in pediatric imaging. *Indian Journal of Radiology*
and Imaging (2015) doi:10.4103/0971-3026.169466

17. Wilm BJ, Svensson J, Henning A, Pruessmann KP, Boesiger P, Kollias SS. Reduced field-of-
view MRI using outer volume suppression for spinal cord diffusion imaging. *Magnetic*
Resonance in Medicine (2007) doi:10.1002/mrm.21167

18. Toselli B, Tortora D, Severino M, Arnulfo G, Canessa A, Morana G, Rossi A, Fato MM.
Improvement in white matter tract reconstruction with constrained spherical deconvolution
and track density mapping in low angular resolution data: A pediatric study and literature
review. *Frontiers in Pediatrics* (2017) doi:10.3389/fped.2017.00182

19. Webster JG, Descoteaux M. "High Angular Resolution Diffusion Imaging (HARDI)," in
Wiley   Encyclopedia   of   Electrical   and   Electronics   Engineering
doi:10.1002/047134608x.w8258

791 20. Andre JB, Bammer R. Advanced diffusion-weighted magnetic resonance imaging techniques
792   of the human spinal cord. *Topics in Magnetic Resonance Imaging* (2010)
793   doi:10.1097/RMR.0b013e31823e65a1

794 21. Fruehwald-Pallamar J, Szomolanyi P, Fakhrai N, Lunzer A, Weber M, Thurnher MM,
795   Pallamar M, Trattnig S, Prayer D, Noebauer-Huhmann IM. Parallel imaging of the cervical
796   spine at 3T: Optimized trade-off between speed and image quality. *American Journal of*
797   *Neuroradiology* (2012) doi:10.3174/ajnr.A3101

798 22. Saksena S, Alizadeh M, Middleton DM, Conklin CJ, Krisa L, Flanders A, Mulcahey M,
799   Mohamed FB, Faro SH. Characterization of spinal cord diffusion tensor imaging metrics in
800   clinically asymptomatic pediatric subjects with incidental congenital lesions. *Spinal Cord*
801   *Series and Cases* (2018) doi:10.1038/s41394-018-0073-8

802 23. Alizadeh M, Fisher J, Saksena S, Sultan Y, Conklin CJ, Middleton DM, Krisa L,
803   Finsterbusch J, Flanders AE, Faro SH, et al. Age related diffusion and tractography changes
804   in typically developing pediatric cervical and thoracic spinal cord. *NeuroImage: Clinical*
805   (2018) doi:10.1016/j.nicl.2018.03.014

806 24. Alizadeh M, Fisher J, Saksena S, Sultan Y, Conklin CJ, Middleton DM, Finsterbusch J, Krisa
807   L, Flanders AE, Faro SH, et al. Reduced Field of View Diffusion Tensor Imaging and Fiber
808   Tractography of the Pediatric Cervical and Thoracic Spinal Cord Injury. *Journal of*
809   *Neurotrauma* (2018) doi:10.1089/neu.2017.5174

810 25. Mulcahey MJ, Samdani AF, Gaughan JP, Barakat N, Faro S, Shah P, Betz RR, Mohamed
811   FB. Diagnostic accuracy of diffusion tensor imaging for pediatric cervical spinal cord injury.
812   *Spinal Cord* (2013) doi:10.1038/sc.2013.36

813 26. Saksena S, Middleton DM, Krisa L, Shah P, Faro SH, Sinko R, Gaughan J, Finsterbusch J,
814   Mulcahey MJ, Mohamed FB. Diffusion tensor imaging of the normal cervical and thoracic
815   pediatric spinal cord. *American Journal of Neuroradiology* (2016) doi:10.3174/ajnr.A4883

816 27. Mohamed FB, Hunter LN, Barakat N, Liu CSJ, Sair H, Samdani AF, Betz RR, Faro SH,
817   Gaughan J, Mulcahey MJ. Diffusion tensor imaging of the pediatric spinal cord at 1.5T:
818   Preliminary results. *American Journal of Neuroradiology* (2011) doi:10.3174/ajnr.A2334

819 28. Antherieu P, Levy R, de Saint Denis T, Lohkamp L, Paternoster G, di Rocco F, Boddaert N,
820   Zerah M. Diffusion tensor imaging (DTI) and Tractography of the spinal cord in pediatric
821   population with spinal lipomas: preliminary study. *Child's Nervous System* (2019)
822   doi:10.1007/s00381-018-3935-2

823 29. Reynolds BB, By S, Weinberg QR, Witt AA, Newton AT, Feiler HR, Ramkorun B, Clayton
824   DB, Couture P, Martus JE, et al. Quantification of DTI in the pediatric spinal cord:
825   Application to clinical evaluation in a healthy patient population. *American Journal of*
826   *Neuroradiology* (2019) doi:10.3174/ajnr.A6104

827 30. Singhi S, Tekes A, Thurnher M, Gilson WD, Izbudak I, Thompson CB, Huisman TAGM.
828   Diffusion tensor imaging of the maturing paediatric cervical spinal cord: From the neonate to
829   the young adult. *Journal of Neuroradiology* (2012) doi:10.1016/j.neurad.2011.05.002

31. Saksena S, Mohamed FB, Middleton DM, Krisa L, Alizadeh M, Shahrampour S, Conklin CJ,
Flanders A, Finsterbusch J, Mulcahey MJ, et al. Diffusion Tensor Imaging Assessment of
Regional White Matter Changes in the Cervical and Thoracic Spinal Cord in Pediatric
Subjects. *Journal of Neurotrauma* (2018) 36:853–861. doi:10.1089/neu.2018.5826

32. Conklin CJ, Middleton DM, Alizadeh M, Finsterbusch J, Raunig DL, Faro SH, Shah P, Krisa
L, Sinko R, Delalic JZ, et al. Spatially selective 2D RF inner field of view (iFOV) diffusion
kurtosis imaging (DKI) of the pediatric spinal cord. *NeuroImage: Clinical* (2016)
doi:10.1016/j.nicl.2016.01.009

33. Singh G, True AJ, Lui CC, Prasanna P, Orleans G, Partyka L, Phatak TD. Normal anterior-
posterior diameters of the spinal cord and spinal canal in healthy term newborns on
sonography. *Pediatric Radiology* (2020) doi:10.1007/s00247-020-04879-8

34. Kaplan KM, Spivak JM, Bendo JA. Embryology of the spine and associated congenital
abnormalities. *Spine Journal* (2005) doi:10.1016/j.spinee.2004.10.044

35. Oskouian RJ, Sansur CA, Shaffrey CI. Congenital Abnormalities of the Thoracic and
Lumbar   Spine.   *Neurosurgery   Clinics   of   North   America*   (2007)
doi:10.1016/j.nec.2007.04.004

36. Rufener S, Ibrahim M, Parmar HA. Imaging of Congenital Spine and Spinal Cord
Malformations.   *Neuroimaging   Clinics   of   North   America*   (2011)
doi:10.1016/j.nic.2011.05.011

37. S Basu P, Elsebaie H, Noordeen M. Congenital spinal deformity: A comprehensive
assessment at presentation. *Spine* (2002) doi:10.1097/00007632-200210150-00014

38. Yang RK, Roth CG, Ward RJ, deJesus JO, Mitchell DG. Optimizing abdominal MR
imaging: Approaches   to   common   problems.   *Radiographics*   (2010)
doi:10.1148/rg.301095076

39. Tournier JD, Smith R, Raffelt D, Tabbara R, Dhollander T, Pietsch M, Christiaens D,
Jeurissen B, Yeh CH, Connelly A. MRtrix3: A fast, flexible and open software framework
for   medical   image   processing   and   visualisation.   *NeuroImage*   (2019)
doi:10.1016/j.neuroimage.2019.116137

40. de Leener B, Lévy S, Dupont SM, Fonov VS, Stikov N, Louis Collins D, Callot V, Cohen-
Adad J. SCT: Spinal Cord Toolbox, an open-source software for processing spinal cord MRI
data. *NeuroImage* (2017) doi:10.1016/j.neuroimage.2016.10.009

41. Garyfallidis E, Brett M, Amirbekian B, Rokem A, van der Walt S, Descoteaux M, Nimmo-
Smith I. Dipy, a library for the analysis of diffusion MRI data. *Frontiers in Neuroinformatics*
(2014) doi:10.3389/fninf.2014.00008

42. Battiston M, Grussu F, Ianus A, Schneider T, Prados F, Fairney J, Ourselin S, Alexander DC,
Cercignani M, Gandini Wheeler-Kingshott CAM, et al. An optimized framework for
quantitative magnetization transfer imaging of the cervical spinal cord in vivo. *Magnetic
Resonance in Medicine* (2018) doi:10.1002/mrm.26909

43. Brooks JCW, Büchel C, Winkler AM, Andersson JL, Tracey I. Investigating resting-state
functional connectivity in the cervical spinal cord at 3 T. *NeuroImage* (2017)
doi:10.1016/j.neuroimage.2016.12.072

44. Duval T, Lévy S, Stikov N, Cohen-Adad J, Stikov N, Campbell J, Mezer A, Witzel T, Keil
B, Smith V, et al. g-Ratio weighted imaging of the human spinal cord in vivo. *NeuroImage*
(2017) doi:10.1016/j.neuroimage.2016.09.018

45. Eippert F, Kong Y, Jenkinson M, Tracey I, Brooks JCW. Denoising spinal cord fMRI data:
Approaches to acquisition and analysis. *NeuroImage* (2017)
doi:10.1016/j.neuroimage.2016.09.065

46. Kong Y, Eippert F, Beckmann CF, Andersson J, Finsterbusch J, Büchel C, Tracey I, Brooks
JCW. Intrinsically organized resting state networks in the human spinal cord. *Proceedings of*
the National Academy of Sciences of the United States of America (2014)
doi:10.1073/pnas.1414293111

47. Samson RS, Lévy S, Schneider T, Smith AK, Smith SA, Cohen-Adad J, Wheeler-Kingshott
CAMG. ZOOM or Non-ZOOM? Assessing spinal cord diffusion tensor imaging protocols
for multi-centre studies. *PLoS ONE* (2016) doi:10.1371/journal.pone.0155557

48. Massire A, Taso M, Besson P, Guye M, Ranjeva JP, Callot V. High-resolution multi-
parametric quantitative magnetic resonance imaging of the human cervical spinal cord at 7T.
NeuroImage (2016) doi:10.1016/j.neuroimage.2016.08.055

49. Taso M, Girard OM, Duhamel G, le Troter A, Feiweier T, Guye M, Ranjeva JP, Callot V.
Tract-specific and age-related variations of the spinal cord microstructure: A multi-
parametric MRI study using diffusion tensor imaging (DTI) and inhomogeneous
magnetization transfer (ihMT). *NMR in Biomedicine* (2016) doi:10.1002/nbm.3530

50. Vahdat S, Lungu O, Cohen-Adad J, Marchand-Pauvert V, Benali H, Doyon J. Simultaneous
brain–cervical cord fMRI reveals intrinsic spinal cord plasticity during motor sequence
learning. *PLoS Biology* (2015) doi:10.1371/journal.pbio.1002186

51. Weber KA, Sentis AI, Bernadel-Huey ON, Chen Y, Wang X, Parrish TB, Mackey S.
Thermal Stimulation Alters Cervical Spinal Cord Functional Connectivity in Humans.
Neuroscience (2018) doi:10.1016/j.neuroscience.2017.10.035

52. Weber KA, Chen Y, Wang X, Kahnt T, Parrish TB. Lateralization of cervical spinal cord
activity during an isometric upper extremity motor task with functional magnetic resonance
imaging. *NeuroImage* (2016) doi:10.1016/j.neuroimage.2015.10.014

53. Ljungberg E, Vavasour I, Tam R, Yoo Y, Rauscher A, Li DKB, Traboulsee A, MacKay A,
Kolind S. Rapid myelin water imaging in human cervical spinal cord. *Magnetic Resonance in*
Medicine (2017) doi:10.1002/mrm.26551

54. Castellano A, Papinutto N, Cadioli M, Brugnara G, Iadanza A, Scigliuolo G, Pareyson D,
Uziel G, Köhler W, Aubourg P, et al. Quantitative MRI of the spinal cord and brain in

adrenomyeloneuropathy: In vivo assessment of structural changes. *Brain* (2016)
doi:10.1093/brain/aww068

55. Grabher P, Mohammadi S, Trachsler A, Friedl S, David G, Sutter R, Weiskopf N, Thompson
AJ, Curt A, Freund P. Voxel-based analysis of grey and white matter degeneration in cervical
spondylotic myelopathy. *Scientific Reports* (2016) doi:10.1038/srep24636

56. Hori M, Hagiwara A, Fukunaga I, Ueda R, Kamiya K, Suzuki Y, Liu W, Murata K,
Takamura T, Hamasaki N, et al. Application of Quantitative Microstructural MR Imaging
with Atlas-based Analysis for the Spinal Cord in Cervical Spondylotic Myelopathy.
Scientific Reports (2018) doi:10.1038/s41598-018-23527-8

57. Huber E, David G, Thompson AJ, Weiskopf N, Mohammadi S, Freund P. Dorsal and ventral
horn atrophy is associated with clinical outcome after spinal cord injury. *Neurology* (2018)
doi:10.1212/WNL.0000000000005361

58. McCoy DB, Talbott JF, Wilson M, Mamlouk MD, Cohen-Adad J, Wilson M, Narvid J. MRI
atlas-based measurement of spinal cord injury predicts outcome in acute flaccid myelitis.
American Journal of Neuroradiology (2017) doi:10.3174/ajnr.A5044

59. Martin AR, de Leener B, Cohen-Adad J, Cadotte DW, Kalsi-Ryan S, Lange SF, Tetreault L,
Nouri A, Crawley A, Mikulis DJ, et al. A novel MRI biomarker of spinal cord white matter
injury: T2∗-weighted white matter to gray matter signal intensity ratio. *American Journal of*
Neuroradiology (2017) doi:10.3174/ajnr.A5162

60. Smith AC, Weber KA, O'Dell DR, Parrish TB, Wasielewski M, Elliott JM. Lateral
Corticospinal Tract Damage Correlates With Motor Output in Incomplete Spinal Cord Injury.
Archives of Physical Medicine and Rehabilitation (2018) doi:10.1016/j.apmr.2017.10.002

61. Talbott JF, Narvid J, Chazen JL, Chin CT, Shah V. An Imaging-Based Approach to Spinal
Cord Infection. *Seminars in Ultrasound, CT and MRI* (2016) doi:10.1053/j.sult.2016.05.006

62. Ventura RE, Kister I, Chung S, Babb JS, Shepherd TM. Cervical spinal cord atrophy in
nmosd without a history of myelitis or MRI-visible lesions. *Neurology: Neuroimmunology*
and NeuroInflammation (2016) doi:10.1212/NXI.0000000000000224

63. Yiannakas MC, Mustafa AM, de Leener B, Kearney H, Tur C, Altmann DR, de Angelis F,
Plantone D, Ciccarelli O, Miller DH, et al. Fully automated segmentation of the cervical cord
from T1-weighted MRI using PropSeg: Application to multiple sclerosis. *NeuroImage:*
Clinical (2016) doi:10.1016/j.nicl.2015.11.001

64. Barth M, Breuer F, Koopmans PJ, Norris DG, Poser BA. Imaging Methodology - Review
Simultaneous Multislice (SMS) Imaging Techniques. *Magnetic Resonance in Medicine*
(2016)

65. Fadnavis S, Batson J, Garyfallidis E. Patch2Self: Denoising Diffusion MRI with Self-
Supervised Learning. (2020)1–11. Available at: http://arxiv.org/abs/2011.01355

66. Schilling K, Fadnavis S, Visagie M, Garyfallidis E, Landman B, Smith S, O'grady K.
Patch2Self denoising of diffusion MRI in the cervical spinal cord improves repeatability and
feature conspicuity. in

67. Xu J, Shimony JS, Klawiter EC, Snyder AZ, Trinkaus K, Naismith RT, Benzinger TLS,
Cross AH, Song SK. Improved in vivo diffusion tensor imaging of human cervical spinal
cord. *NeuroImage* (2013) doi:10.1016/j.neuroimage.2012.11.014

68. de Leener B, Kadoury S, Cohen-Adad J. Robust, accurate and fast automatic segmentation of
the spinal cord. *NeuroImage* (2014) doi:10.1016/j.neuroimage.2014.04.051

69. Gros C, de Leener B, Badji A, Maranzano J, Eden D, Dupont SM, Talbott J, Zhuoquiong R,
Liu Y, Granberg T, et al. Automatic segmentation of the spinal cord and intramedullary
multiple sclerosis lesions with convolutional neural networks. *NeuroImage* (2019)
doi:10.1016/j.neuroimage.2018.09.081

70. Ullmann E, Pelletier Paquette JF, Thong WE, Cohen-Adad J. Automatic Labeling of
Vertebral Levels Using a Robust Template-Based Approach. *International Journal of
Biomedical Imaging* (2014) doi:10.1155/2014/719520

71. de Leener B, Fonov VS, Collins DL, Callot V, Stikov N, Cohen-Adad J. PAM50: Unbiased
multimodal template of the brainstem and spinal cord aligned with the ICBM152 space.
NeuroImage (2018) doi:10.1016/j.neuroimage.2017.10.041

72. de Leener B, Mangeat G, Dupont S, Martin AR, Callot V, Stikov N, Fehlings MG, Cohen-
Adad J. Topologically preserving straightening of spinal cord MRI. *Journal of Magnetic
Resonance Imaging* (2017) doi:10.1002/jmri.25622

73. Tustison NJ, Avants BB. Explicit B-spline regularization in diffeomorphic image
registration. *Frontiers in Neuroinformatics* (2013) doi:10.3389/fninf.2013.00039

74. Henriques RN, Correia MM, Marrale M, Huber E, Kruper J, Koudoro S, Yeatman JD,
Garyfallidis E, Rokem A. Diffusional Kurtosis Imaging in the Diffusion Imaging in Python
Project. *Frontiers in Human Neuroscience* (2021) doi:10.3389/fnhum.2021.675433

75. Lévy S, Benhamou M, Naaman C, Rainville P, Callot V, Cohen-Adad J. White matter atlas
of the human spinal cord with estimation of partial volume effect. *NeuroImage* (2015)
doi:10.1016/j.neuroimage.2015.06.040

76. Tax CMW, Otte WM, Viergever MA, Dijkhuizen RM, Leemans A. REKINDLE: Robust
Extraction of Kurtosis INDices with Linear Estimation. *Magnetic Resonance in Medicine*
(2015) doi:10.1002/mrm.25165

77. Neto Henriques R. Diffusion kurtosis imaging of the healthy human brain. (2012)134.
Available at: http://repositorio.ul.pt/handle/10451/8511

78. Henriques RN, Jespersen SN, Jones DK, Veraart J. Toward more robust and reproducible
diffusion kurtosis imaging. *Magnetic Resonance in Medicine* (2021) doi:10.1002/mrm.28730

79.   Henriques RN. Advanced Methods for Diffusion MRI Data Analysis and their Application to
978   the Healthy Ageing Brain. (2017)

979   80.   Neto Henriques R, Correia MM, Nunes RG, Ferreira HA. Exploring the 3D geometry of the
980   diffusion kurtosis tensor-Impact on the development of robust tractography procedures and
981   novel biomarkers. *NeuroImage* (2015) doi:10.1016/j.neuroimage.2015.02.004

982   81.   Henriques RN, Jespersen SN, Shemesh N. Microscopic anisotropy misestimation in
983   spherical-mean single diffusion encoding MRI. *Magnetic Resonance in Medicine* (2019)
984   doi:10.1002/mrm.27606

985   82.   Panara V, Navarra R, Mattei PA, Piccirilli E, Bartoletti V, Uncini A, Caulo M. Correlations
986   between cervical spinal cord magnetic resonance diffusion tensor and diffusion kurtosis
987   imaging metrics and motor performance in patients with chronic ischemic brain lesions of the
988   corticospinal tract. *Neuroradiology* (2019) doi:10.1007/s00234-018-2139-5

989   83.   Volpe JJ, Inder TE, Darras BT, de Vries LS, du Plessis AJ, Neil JJ, Perlman J. *Volpe's*
990   *neurology of the newborn*. (2017). doi:10.1016/c2010-0-68825-0

991   84.   Veraart J, Novikov DS, Christiaens D, Ades-aron B, Sijbers J, Fieremans E. Denoising of
992   diffusion   MRI   using   random   matrix   theory.   *NeuroImage*   (2016)
993   doi:10.1016/j.neuroimage.2016.08.016

994   85.   Hastie T, Tibshirani R, Friedman J. *The Elements of Statistical Learning Data Mining,*
995   *Inference, and Prediction (12th printing)*. (2017).

996   86.   Pierpaoli C, Basser PJ. Toward a quantitative assessment of diffusion anisotropy. *Magnetic*
997   *Resonance in Medicine* (1996) doi:10.1002/mrm.1910360612

998   87.   Pierpaoli C, Jezzard P, Basser PJ, Barnett A, di Chiro G. Diffusion tensor MR imaging of the
999   human brain. *Radiology* (1996) doi:10.1148/radiology.201.3.8939209

88.   Parodi A, Malova M, Cardiello V, Raffa S, Re M, Calevo MG, Severino M, Tortora D,
Morana G, Rossi A, et al. Punctate white matter lesions of preterm infants: Risk factor
analysis. *European Journal of Paediatric Neurology* (2019) doi:10.1016/j.ejpn.2019.06.003

89.   Cheong JLY, Thompson DK, Wang HX, Hunt RW, Anderson PJ, Inder TE, Doyle LW.
Abnormal white matter signal on MR imaging is related to abnormal tissue microstructure.
American Journal of Neuroradiology (2009) doi:10.3174/ajnr.A1399

90.   Staudt M, Pavlova M, Böhm S, Grodd W, Krägeloh-Mann I. Pyramidal tract damage
correlates with motor dysfunction in bilateral periventricular leukomalacia (PVL).
Neuropediatrics (2003) doi:10.1055/s-2003-42206

91.   Bassi L, Chew A, Merchant N, Ball G, Ramenghi L, Boardman J, Allsop JM, Doria V,
Arichi T, Mosca F, et al. Diffusion tensor imaging in preterm infants with punctate white
matter lesions. *Pediatric Research* (2011) doi:10.1203/PDR.0b013e3182182836

92.   Tusor N, Benders MJ, Counsell SJ, Nongena P, Ederies MA, Falconer S, Chew A, Gonzalez-
Cinca N, Hajnal J v., Gangadharan S, et al. Punctate White Matter Lesions Associated with
Altered Brain Development and Adverse Motor Outcome in Preterm Infants. *Scientific*
Reports (2017) doi:10.1038/s41598-017-13753-x

93.   Paydar A, Fieremans E, Nwankwo JI, Lazar M, Sheth HD, Adisetiyo V, Helpern JA, Jensen
JH, Milla SS. Diffusional kurtosis imaging of the developing brain. *American Journal of*
Neuroradiology (2014) doi:10.3174/ajnr.A3764

94.   Hansen B. An Introduction to Kurtosis Fractional Anisotropy. *AJNR American journal of*
neuroradiology (2019) 40:1638–1641. doi:10.3174/ajnr.A6235

95.   Zhang H, Schneider T, Wheeler-Kingshott CA, Alexander DC. NODDI: Practical in vivo
neurite orientation dispersion and density imaging of the human brain. *NeuroImage* (2012)
doi:10.1016/j.neuroimage.2012.03.072

1024

1025
1026