Longitudinal assessment of diagnostic test performance over the course of acute SARS-CoV-2 infection

Rebecca L. Smith[1,2,3]*, Laura L. Gibson[4], Pamela P. Martinez[1,5,6], Ruian Ke[7], Agha Mirza[8], Madison Conte[8], Nicholas Gallagher[9], Abigail Conte[10], Leyi Wang[11], Rick Fredrickson[11], Darci C. Edmonson[1], Melinda E. Baughman[1], Karen K. Chiu[1], Hannah Choi[1], Tor W. Jensen[1,12], Kevin R. Scardina[1], Shannon Bradley[1], Stacy L. Gloss[1], Crystal Reinhart[1], Jagadeesh Yedetore[1], Alyssa N. Owens[13], John Broach[14,15], Bruce Barton[16,17], Peter Lazar[16], Darcy Henness[18], Todd Young[18], Alastair Dunnett[18], Matthew L. Robinson[8], Heba H. Mostafa[9], Andrew Pekosz[10], Yukari C. Manabe[8], William J. Heetderks[19], David D. McManus[20], Christopher B. Brooke[1,5]*

[1] Carl R. Woese Institute for Genomic Biology, University of Illinois at Urbana-Champaign, Urbana, IL
[2] Department of Pathobiology, University of Illinois at Urbana-Champaign, Urbana, IL
[3] Carle Illinois College of Medicine, University of Illinois at Urbana-Champaign, Urbana, IL
[4] Division of Infectious Diseases and Immunology, Departments of Medicine and Pediatrics, University of Massachusetts Medical School, Worcester, MA
[5] Department of Microbiology, University of Illinois at Urbana-Champaign, Urbana, IL
[6] Department of Statistics, University of Illinois at Urbana-Champaign, Urbana, IL
[7] T-6, Theoretical Biology and Biophysics, Los Alamos National Laboratory, Los Alamos, NM.
[8] Division of Infectious Diseases, Department of Medicine, Johns Hopkins School of Medicine, Baltimore, MD
[9] Division of Medical Microbiology, Department of Pathology, Johns Hopkins University School of Medicine, Baltimore, MD
[10] W. Harry Feinstone Department of Molecular Microbiology and Immunology, Johns Hopkins Bloomberg School of Public Health, Baltimore, MD
[11] Veterinary Diagnostic Laboratory, University of Illinois at Urbana-Champaign, Urbana, IL
[12] Cancer Center at Illinois, University of Illinois at Urbana-Champaign, Urbana, IL
[13] Center for Clinical and Translational Research, University of Massachusetts Medical School, Worcester, MA
[14] UMass Memorial Medical Center, Worcester, MA
[15] Department of Emergency Medicine, University of Massachusetts Medical School, Worcester, MA
[16] Division of Biostatistics and Health Services Research, University of Massachusetts Medical School, Worcester, MA
[17] Department of Population and Quantitative Health Sciences, University of Massachusetts Medical School, Worcester, MA
[18] Carle Foundation Hospital, Urbana, IL
[19] National Institute for Biomedical Imaging and Bioengineering, Bethesda, MD
[20] Division of Cardiology, University of Massachusetts Medical School, Worcester, MA

* Corresponding authors (rlsdvm@illinois.edu, cbrooke@illinois.edu)

SUMMARY:

What is already known about this topic?

Diagnostic tests and sample types for SARS-CoV-2 vary in sensitivity across the infection period.

What is added by this report?

We show that both RTqPCR (from nasal swab and saliva) and the Quidel SARS Sofia FIA rapid antigen tests peak in sensitivity during the period in which live virus can be detected in nasal swabs, but that the sensitivity of RTqPCR tests rises more rapidly in the pre-infectious period. We also use empirical data to estimate the sensitivities of RTqPCR and antigen tests as a function of testing frequency.

What are the implications for public health practice?

RTqPCR tests will be more effective than rapid antigen tests at identifying infected individuals prior to or early during the infectious period and thus for minimizing forward transmission (provided results reporting is timely). All modalities, including rapid antigen tests, showed >94% sensitivity to detect infection if used at least twice per week. Regular surveillance/screening using rapid antigen tests 2-3 times per week can be an effective strategy to achieve high sensitivity (>95%) for identifying infected individuals.

INTRODUCTION:

Frequent rapid diagnostic testing is critical for restricting community spread of SARS-CoV-2 by allowing the timely identification and isolation of infected individuals to interrupt the chain of transmission. Quantitative reverse transcription polymerase chain reaction (RTqPCR)-based detection of viral RNA within nasal swab or saliva samples represents the gold standard for sensitivity in detecting the presence of SARS-CoV-2, yet supply shortages, cost, and infrastructure limitations have made it difficult to achieve high testing frequency and volume with the rapid reporting of results needed to mitigate transmission effectively.

Recently, there has been considerable interest in the potential of rapid antigen tests to expand diagnostic testing capacity due to the ease of use, availability, cost, and rapid time-to-results[1]. However, data for their use in screening asymptomatic individuals is sparse. Enthusiasm for their widespread deployment has been further tempered by well-publicized examples of false positive results in people with low pre-test probability of infection, and by reports suggesting they lack sensitivity compared with RTqPCR, potentially making them less effective at mitigating community spread[2,3].

Here, we compare the sensitivities of nasal and saliva RTqPCR tests with the Quidel Sofia SARS Antigen Fluorescent Immunoassay (FIA) over the course of mild or asymptomatic acute SARS-CoV-2 infection through daily sampling of individuals enrolled early during infection.

METHODS:

This study was approved by the Western Institutional Review Board, and all participants consented freely.

Participants

All on-campus students and employees of the University of Illinois at Urbana-Champaign are required to submit saliva for RTqPCR testing every 2-4 days as part of the SHIELD campus surveillance testing program. Those testing positive are instructed to isolate, and were eligible to enroll in this study for a period of 24 hours following receipt of their positive test result. Close

96   contacts of individuals who test positive (particularly those co-housed with them) are instructed
97   to quarantine and were eligible to enroll for up to 5 days after their last known exposure to an
98   infected individual. All participants were also required to have received a negative saliva
99   RTqPCR result 7 days prior to enrollment.
100
101   Individuals were recruited via either a link shared in an automated text message providing
102   isolation information sent within 30 minutes of a positive test result, a call from a study recruiter,
103   or a link shared by an enrolled study participant or included in information provided to all
104   quarantining close contacts. In addition, signs were used at each testing location and a website
105   was available to inform the community about the study.
106
107   Participants were required to be at least 18 years of age, have a valid university ID, speak
108   English, have internet access, and live within 8 miles of the university campus. After enrollment
109   and consent, participants completed an initial survey to collect information on demographics and
110   health history, including suspected date of SARS-CoV-2 exposure. They were then provided
111   with sample collection supplies.
112
113   Participants who tested positive prior to enrollment or during quarantine were followed for up to
114   14 days. Quarantining participants who continued to test negative by saliva RTqPCR were
115   followed for up to 7 days after their last exposure. All participants' data and survey responses
116   were collected in the Eureka digital study platform.
117
118   ***Sample collection***
119   Each day, participants were remotely observed by study staff collecting:
120   1. 2 mL of saliva into a 50mL conical tube.
121   2. 1 nasal swab from a single nostril using a foam-tipped swab that was placed within a dry
122   collection tube.
123   3. 1 nasal swab from the other nostril using a flocked swab that was subsequently placed
124   in a collection vial containing viral transport media (VTM).
125
The order of nostrils (left vs. right) used for the two different swabs was randomized. For nasal
swabs, participants were instructed to insert the soft tip of the swab at least 1 cm into the
indicated nostril until they encountered mild resistance, rotate the swab around the nostril 5
times, leaving it in place for 10-15 seconds. After daily sample collection, participants completed
a symptom survey. A courier collected all participant samples within 1 hour of collection using a
no-contact pickup protocol designed to minimize courier exposure to infected participants.
132
Saliva RTqPCR
After collection, saliva samples were stored at room temperature and RTqPCR was run within
12 hours of initial collection. The protocol for direct saliva-to-RTqPCR assay used has been
detailed previously[4]. In brief, saliva samples were heated at 95°C for 30 minutes, followed by
the addition of 2X TBS at a 1:1 ratio (final concentration 1X TBE) and Tween-20 to a final
concentration of 0.5%. Samples were assayed using the Thermo Taqpath COVID-19 assay.
139
140   ***Quidel assay***
141   Foam-tipped nasal swabs were placed in collection tubes and stored at 4°C overnight based on
142   guidance from the manufacturer. The morning after collection, swabs were run through the Sofia
143   SARS antigen FIA on Sofia 2 devices according to the manufacturer's protocol.
144
Nasal swab RTqPCR

Collection tubes containing VTM and flocked nasal swabs were stored at -80°C after collection
and were subsequently shipped to Johns Hopkins University for RTqPCR and virus culture
testing. After thawing, VTM was aliquoted for RTqPCR and infectivity assays. One ml of VTM
from the nasal swab was assayed on the Abbott Alinity per manufacturer's instructions in a
College of American Pathologist and CLIA-certified laboratory.
151
Nasal virus culture
VeroTMPRSS2 cells were grown in complete medium (CM) consisting of DMEM with 10% fetal
bovine serum (Gibco), 1 mM glutamine (Invitrogen), 1 mM sodium pyruvate (Invitrogen), 100
U/ml of penicillin (Invitrogen), and 100 µg/ml of streptomycin (Invitrogen)[5]. Viral infectivity was
assessed on VeroTMPRSS2 cells as previously described using infection media (IM; identical to
CM except the FBS is reduced to 2.5%)[6]. When a cytopathic effect was visible in >50% of cells
in a given well, the supernatant was harvested. The presence of SARS-CoV-2 was confirmed
through RTqPCR as described previously by extracting RNA from the cell culture supernatant
using the Qiagen viral RNA isolation kit and performing RTqPCR using the N1 and N2 SARS-
CoV-2-specific primers and probes in addition to primers and probes for human RNaseP gene
using synthetic RNA target sequences to establish a standard curve[7].
163
164 *Data Analysis*
165 At the time of analysis, nasal samples from 30 participants had been analyzed by virus culture
166 and RTqPCR. Therefore, analyses that consider either nasal RTqPCR or viral culture results
167 were conducted based on a limited participant set. All confidence intervals around sensitivity
168 were calculated using binconf from the Hmisc package in R version 3.6.2.
169
170 The sensitivity of each of the tests was analyzed in three different ways:
171 First, the ability of each test (antigen, saliva RTqPCR, or nasal RTqPCR) to detect an infected
172 person on a particular day relative to the day of first positive viral culture ("daily sensitivity") was
173 calculated. Daily sensitivity was not calculated for timepoints with fewer than 5 observed
174 person-days.
175
Second, the ability of each test to detect an infected person according to their viral culture status
("status sensitivity") was calculated. Viral culture status was defined as "pre-positive" on days
prior to the first positive viral culture result, "positive" on days for which viral culture results were
positive, and "post-positive" on days with negative viral culture results that occur after the first
positive culture result. Status sensitivity was defined as the proportion of person-days with a
positive result.
182
Finally, we calculated the ability of repeated testing over a 14-day period to detect an infected
person ("protocol sensitivity") using a value-of-information approach. Seven different testing
frequencies were considered: daily, every other day, every third day, and so on, up to weekly
sampling. For each individual, the result of testing on a given schedule was calculated for each
potential starting date, with test results interpreted in parallel (all tests must be negative to be
considered negative). For instance, each person contributed two observations to the "every
other day" schedule, one starting on the first day of the study and the other starting on the
second day of the study. The proportion of "observations" with a positive result (at least one
positive test in the sampling timeframe) was considered to be the sensitivity of that testing
protocol (test and frequency combination).
193
194 All code used in analyses can be found here: https://github.com/rlsdvm/CovidDetectAnalysis
195

Results

Table 1 shows demographic information for study participants reported here. The majority of participants (21/30, 70%) were non-Hispanic white and the average age was 32.50 (SD 12.29).

Table 1: Demographic information on participants enrolled in the COVID detect study

Variable		Data
		n=30
Weight (mean (SD))		176.00 (51.17)
Height in inches (mean (SD))		67.57 (4.94)
Age (mean (SD))		32.50 (12.29)
Race (%)	Native American	0 (0.0)
	Asian	1 (3.3)
	Black	2 (6.7)
	Other	3 (10.0)
	Pacific Islander	0 (0.0)
	White	24 (80.0)
Gender (%)	Female	12 (40.0)
	Male	18 (60.0)
Ethnicity (%)	Hispanic	6 (20.0)
	Non-Hispanic	24 (80.0)

We first estimated the daily sensitivities of nasal and saliva RTqPCR and antigen tests relative to the day of first nasal swab viral culture positivity, which was used as a surrogate marker of infectious virus shedding (**Table 2, Figure 1**). We also used the viral culture data to measure the status sensitivities of each test before, during, and after viral shedding (**Figure 2**).

Prior to the first day of detectable shedding of infectious virus, both RTqPCR tests had higher daily sensitivity (0.706 for both saliva and nasal) than the antigen test (0.412). For all three tests, daily and status sensitivity peaked during days in which infectious virus shedding was detectable, as would be expected. Antigen test daily sensitivity declined precipitously after infectious virus could no longer be detected in nasal swabs, dropping below 0.5 within a week after the onset of culture positivity, while both nasal and saliva RTqPCR platforms only showed minor decreases in sensitivity, remaining at 0.792 and 0.667 after a week, respectively.

217 *Table 2: Daily sensitivity of each test platform by day relative to the day of first nasal*
218 *swab viral culture positivity.*
219

Days before (-1,-2), on (0), or after the day of first positive culture	Antigen		Saliva RTqPCR		Nasal RTqPCR		
	Daily Sensitivity	Number positive	Daily Sensitivity	Number positive	Daily Sensitivity	Number positive	Total
-2	0.333	2	0.833	5	0.667	4	6
-1	0.455	5	0.636	7	0.727	8	11
0	0.875	21	0.958	23	1.000	24	24
1	0.960	24	1.000	25	1.000	25	25
2	0.960	24	0.960	24	1.000	25	25
3	0.920	23	0.920	23	1.000	25	25
4	0.760	19	0.960	24	1.000	25	25
5	0.640	16	0.840	21	0.960	24	25
6	0.560	14	0.920	23	0.880	22	25
7	0.250	6	0.667	16	0.792	19	24
8	0.182	4	0.682	15	0.909	20	22
9	0.045	1	0.500	11	0.727	16	22
10	0	0	0.500	10	0.900	18	20
11	0.05	1	0.500	10	0.800	16	20
12	0	0	0.368	7	0.526	10	19
13	0	0	0.231	3	0.385	5	13

220
221

222 **Figure 1: Daily sensitivity of each test platform by day relative to the day of first positive**
223 **viral culture result**. *Shaded areas represent the 95% confidence interval around the observed*
224 *proportion.*

225
226
Figure 2: Status sensitivity of each test platform relative to viral culture positivity. *Bars*
indicate the 95% confidence interval around the observed proportion. Pre-positive (n=29) refers
to samples taken on days before the first viral culture-positive sample collected from each
individual. Positive (n=127) refers to samples taken on days for which viral culture results were
positive. Post-positive (n=112) refers to samples taken on days with negative viral culture
results that occur after the first positive culture result.
233

235
We next estimated the protocol sensitivities, or how the ability of each of test platform to detect
infected individuals was affected by differences in testing frequencies (**Table 3, Figure 3**).
Protocol sensitivity was defined at the schedule level, where the numerator is the number of
testing schedules resulting in at least one positive test and the denominator is the number of
testing schedules examined, where a testing schedule is defined as a set of samples from one
participant taken at a given frequency. In **Figure 3**, we calculated the effects of varying testing
frequency on sensitivity to detect infected individuals on days where nasal swabs were viral
culture positive in the top panel. In the bottom panel of **Figure 3**, we examined sensitivity to
detect infected individuals at any stage of infection.

245
246 *Table 3: **Protocol sensitivity of each test platform to detect an infected person during a**
247 **14-day testing period, relative to the frequency of testing**. "Any time" refers to detection of*
248 *the individual at any point in the 14-day testing period; "While VC+" refers to detection of the*
249 *individual before or during the time in which their viral culture was positive.*

		Nasal Antigen				Saliva RTqPCR				Nasal RTqPCR			
		Probability of Detection		Number Positive		Probability of Detection		Number Positive		Probability of Detection		Number Positive	
Testing Frequency	**N**	**Any time**	**While VC+**	**Any time**	**While VC+**	**Any time**	**While VC+**	**Any time**	**While VC+**	**Any time**	**While VC+**	**Any time**	**While VC+**
Daily	34	1.000	0.941	34	32	1.000	0.971	34	33	1.000	1.000	34	34
Every Other Day	68	0.971	0.824	66	56	0.956	0.853	65	58	0.985	0.882	67	60
Every Third Day	102	0.961	0.794	98	81	0.941	0.814	96	83	0.98	0.843	100	86
Every Fourth Day	136	0.912	0.721	124	98	0.934	0.743	127	101	0.971	0.772	132	105
Every Fifth Day	170	0.888	0.641	151	109	0.924	0.676	157	115	0.971	0.712	165	121
Every Sixth Day	204	0.833	0.569	170	116	0.907	0.608	185	124	0.956	0.632	195	129
Weekly	238	0.761	0.508	181	121	0.903	0.546	215	130	0.958	0.571	228	136

251
252

253 ***Figure 3: Protocol sensitivity of each test platform*** *to detect an infected person **(top)** before*
254 *or during days where nasal samples were viral culture positive or **(bottom)** at any time, over a*
255 *14-day testing period, relative to frequency of testing. Lines indicate 95% confidence interval*
256 *around the observed proportion.*
257

258
259
Discussion
Our data indicate that RTqPCR and antigen tests vary significantly over the course of SARS-
CoV-2 infection. Prior to the infectious period (here defined as the period during which infectious
virus could be detected in nasal swab samples), the daily sensitivities of nasal and saliva
RTqPCR tests were higher than that of the Quidel Sofia SARS Antigen FIA, suggesting that
RTqPCR tests will be more effective at identifying infected individuals before they transmit to
others.
267
Both RTqPCR and antigen tests peak in daily and status sensitivities when infectious virus is
detectable in nasal swab samples, suggesting that all three modalities can be effective at

identifying individuals during the infectious period. After this period, the daily sensitivity of
RTqPCR tests decreased gradually, with saliva RTqPCR dropping faster than nasal RTqPCR.
These dynamics are consistent with those described previously for RTqPCR[8,9]. In contrast, the
daily sensitivity of the antigen test declined very quickly, suggesting that this test will be less
effective at identifying individuals during later stages of infection. This may limit diagnosis and
contact-tracing efforts in test-limited environments.
276
277 Previous studies have suggested that frequent testing would maximize the ability of a given test
278 modality to detect infected individuals[10,11]. We found that all testing modalities showed almost
279 95% protocol sensitivity to detect infection if used at least twice per week. When applied weekly,
280 protocol sensitivity remained very high for nasal RTqPCR, declined slightly to 90% for saliva
281 RTqPCR, and dropped to only 76% for the antigen test.
282
When we compared the abilities of different testing frequencies to identify individuals while
infectious virus was detectable in nasal samples, we observed a clear reduction in protocol
sensitivity for all testing modalities when testing frequencies decreased below daily. The
reduction in protocol sensitivity was most pronounced for the antigen test, which dropped to
0.72 with testing every fourth day, however, both RTqPCR tests were only slightly better at 0.74
(saliva) and 0.77 (nasal). Altogether, these data demonstrate the importance of frequent testing
regardless of test modality for identifying individuals while they are contagious.
290
This is the first study to compare the longitudinal performance of rapid antigen and RTqPCR
tests with infectious virus shedding in a well-defined population early in SARS-CoV-2 infection.
We found that all three diagnostic tests demonstrated a high degree of daily sensitivity during
the presumed infectious period, but that the RTqPCR tests exhibited superior daily sensitivities
prior to this period. Our data suggest that RTqPCR tests can be more effective than antigen
tests at mitigating community spread of SARS-CoV-2, but only if the turnaround time for
RTqPCR results is short. Finally, these data also quantitatively demonstrate the importance of
frequent (at least twice per week) screening to maximize likelihood of detecting infected
individuals regardless of testing modality.
300

Acknowledgments
This study was funded by the NIH RADx-Tech program under 3U54HL143541-02S2. The views
expressed in this manuscript are those of the authors and do not necessarily represent the
views of the National Institute of Biomedical Imaging and Bioengineering; the National Heart,
Lung, and Blood Institute; the National Institutes of Health, or the U.S. Department of Health
and Human Services. Sofia 2 devices and associated supplies were provided to Carle
Foundation Hospital by Quidel, however Quidel played no role in the design of the study or the
interpretation or presentation of the data.
309
We also thank Shumon Ahmed, Carly Bell, Nate Bouton, Callie Brennen, Justin Brown, Coleco
Buie, Emmaline Cler, Gary Cole, Trey Coleman, Lauren Engels, Savannah Feher, Kelsey Fox,
Lexi Freeman, Yesenia Gonzalez, Montez Harris, Dan Hiser, Ayeshah Hussain, Daryl Jackson,
Michael Jenkins, Kalombo Kalonji, Syntyche Kanku, Steven Krauklis, Mary Krouse, Elmore
Leshoure, Joe Lewis, Angel Lopez, Guadalupe Lopez, Emily Luna, Chun Huai Luo, Colby
Mackey, Skyler McLain, Yared Berhanu Melesse, Madison O'Donnell, Savanna Pflugmacher,
Denver Piatt, Skyler Pierce, Jessica Quicksall, Gina Quitanilla, Ameera Samad, MacKenzie
Scroggins, Monique Settles, Macie Sinn, Pete Varney, Evette Vlach, and Raeshun Williams-
Chatman for their efforts supporting recruitment, enrollment, logistics, and sample
collection. We also thank Jeffrey Olgin, Noah Peyser, and Xochitl Butler for assistance with the

Eureka platform, Michelle Lore for assistance with REDcap, and Gillian Snyder for assistance in
development of study protocols and logistics.
322
References

1. Krüger LJ, Gaeddert M, Köppel L, Brümmer LE, Gottschalk C, Miranda IB, Schnitzler P,
Kräusslich HG, Lindner AK, Nikolai O, Mockenhaupt FP, Seybold J, Corman VM, Drosten
C, Pollock NR, Cubas-Atienzar AI, Kontogianni K, Collins A, Wright AH, Knorr B, Welker A,
de Vos M, Sacks JA, Adams ER, Denkinger CM, for the study team. Evaluation of the
accuracy, ease of use and limit of detection of novel, rapid, antigen-detecting point-of-care
diagnostics for *SARS-CoV-2* [Internet]. Infectious Diseases (except HIV/AIDS); 2020 Oct.
Available from: http://medrxiv.org/lookup/doi/10.1101/2020.10.01.20203836

2. Jääskeläinen A, Ahava M, Jokela P, Szirovicza L, Pohjala S, Vapalahti O, Lappalainen M,
Hepojoki J, Kurkela S. Evaluation of three rapid lateral flow antigen detection tests for the
diagnosis of SARS-CoV-2 infection [Internet]. Infectious Diseases (except HIV/AIDS); 2021
Jan. Available from: http://medrxiv.org/lookup/doi/10.1101/2020.12.30.20249057

3. Moreno GK, Braun KM, Pray IW, Seagaloff HE, Lim A, Poulsen K, Meiman J, Borchers J,
Westergaard RP, Moll MK, Friedrich T, O'Connor DH. SARS-CoV-2 transmission in
intercollegiate athletics not fully mitigated with daily antigen testing [Internet]. Infectious
Diseases (except HIV/AIDS); 2021 Mar. Available from:
http://medrxiv.org/lookup/doi/10.1101/2021.03.03.21252838

4. Ranoa DRE, Holland RL, Alnaji FG, Green KJ, Wang L, Brooke CB, Burke MD, Fan TM,
Hergenrother PJ. Saliva-Based Molecular Testing for SARS-CoV-2 that Bypasses RNA
Extraction [Internet]. Microbiology; 2020 Jun. Available from:
http://biorxiv.org/lookup/doi/10.1101/2020.06.18.159434

5. Matsuyama S, Nao N, Shirato K, Kawase M, Saito S, Takayama I, Nagata N, Sekizuka T,
Katoh H, Kato F, Sakata M, Tahara M, Kutsuna S, Ohmagari N, Kuroda M, Suzuki T,
Kageyama T, Takeda M. Enhanced isolation of SARS-CoV-2 by TMPRSS2-expressing
cells. Proc Natl Acad Sci U S A. 2020 Mar 31;117(13):7001–7003. PMCID: PMC7132130

6. Pekosz A, Parvu V, Li M, Andrews JC, Manabe YC, Kodsi S, Gary DS, Roger-Dalbert C,
Leitch J, Cooper CK. Antigen-Based Testing but Not Real-Time Polymerase Chain
Reaction Correlates With Severe Acute Respiratory Syndrome Coronavirus 2 Viral Culture.
Clin Infect Dis. 2021 Jan 20; PMCID: PMC7929138

7. Waggoner JJ, Stittleburg V, Pond R, Saklawi Y, Sahoo MK, Babiker A, Hussaini L, Kraft
CS, Pinsky BA, Anderson EJ, Rouphael N. Triplex Real-Time RT-PCR for Severe Acute
Respiratory Syndrome Coronavirus 2. Emerg Infect Dis. 2020 Jul;26(7):1633–1635.

8. Kissler SM, Fauver JR, Mack C, Olesen SW, Tai C, Shiue KY, Kalinich CC, Jednak S, Ott
IM, Vogels CBF, Wohlgemuth J, Weisberger J, DiFiori J, Anderson DJ, Mancell J, Ho DD,
Grubaugh ND, Grad YH. SARS-CoV-2 viral dynamics in acute infections [Internet].
Epidemiology; 2020 Oct. Available from:
http://medrxiv.org/lookup/doi/10.1101/2020.10.21.20217042

9.   Byrne AW, McEvoy D, Collins AB, Hunt K, Casey M, Barber A, Butler F, Griffin J, Lane EA,
McAloon C, O'Brien K, Wall P, Walsh KA, More SJ. Inferred duration of infectious period of
SARS-CoV-2: rapid scoping review and analysis of available evidence for asymptomatic
and symptomatic COVID-19 cases. BMJ Open. 2020 Aug;10(8):e039856.

10.   Paltiel AD, Zheng A, Walensky RP. Assessment of SARS-CoV-2 Screening Strategies to
Permit the Safe Reopening of College Campuses in the United States. JAMA Netw Open.
2020 Jul 31;3(7):e2016818.

11.   Larremore DB, Wilder B, Lester E, Shehata S, Burke JM, Hay JA, Tambe M, Mina MJ,
Parker R. Test sensitivity is secondary to frequency and turnaround time for COVID-19
screening. Sci Adv. 2021 Jan;7(1):eabd5393.