Clinical Evaluation of In House Produced 3D Printed Nasopharyngeal Swabs for

COVID-19 Testing

Simon Grandjean Lapierre[1,2], Stéphane Bedwani[1], François DeBlois[1], Audray Fortin[1], Natalia

Zamorano Cuervo[1], Karim Zerouali[1], Elise Caron[1], Philippe Morency Potvin[1,2], Simon Gagnon[1],

Nakome Nguissan[1], Pascale Arlotto[1], Isabelle Hardy[1,2], Catherine-Audrey Boutin[2], Cécile

Tremblay[1, 2], François Coutlée[1, 2], Jacques de Guise[1,4], Nathalie Grandvaux[1,3*]

1 CRCHUM - Centre Hospitalier de l'Université de Montréal, 900 rue Saint Denis, Montréal,

H2X 0A9, Québec, Canada

2 Department of Microbiology, Infectious Diseases and Immunology, Université de Montréal,

Montréal, Québec, Canada

3 Department of Biochemistry and Molecular Medicine, Faculty of Medicine, Université de

Montréal, Montréal, H3C 3J7, Québec, Canada

4 Department of System Engineering, École de technologie supérieure, Université du Québec,

Montréal, Québec, Canada.

*Corresponding author: Nathalie Grandvaux, CRCHUM, 900 rue Saint Denis, Montreal,

Quebec, H2X 0A9, nathalie.grandvaux@umontreal.ca; Tel.: (+1) 514-890-8000 ext. 35292

Keywords: SARS-CoV-2, COVID-19, 3D-printed nasopharyngeal swabs; diagnosis; PCR,

Abstract

3D printed alternatives to standard flocked swabs were rapidly developed to provide a response to the unprecedented and sudden need for an exponentially growing amounts of diagnostic tools to fight the pandemics of COVID-19. In light of the anticipated shortage, an hospital-based 3D printing platform was implemented in our institution for the production of swabs for nasopharyngeal and oropharyngeal sampling based on the freely available open-sourced design made available to the community by University of South Florida's Health Radiology and Northwell Health System teams as replacement for locally used commercial swabs. Validation of our 3D printed swabs was performed by a head-to-head diagnostic accuracy study of the 3D printed "Northwell model" with the cobas PCR Media swabs sample kit. We observed an excellent concordance (total agreement 96.8%, Kappa 0.936) in results obtained with the 3D printed and flocked swabs indicating that the in-house 3D printed swab can be used reliably in a context of shortage of flocked swabs. To our knowledge, this is the first study to report on autonomous hospital-based production and clinical validation of 3D printed swabs.

Introduction

In December 2019, the outbreak of an atypical severe respiratory syndrome, now known as coronavirus 19 disease (COVID-19) in Wuhan, China (1), rapidly led to the identification of the Severe acute respiratory syndrome (SARS)-Coronavirus 2 (SARS-CoV-2). SARS-CoV-2 is highly transmissible (2, 3) and its rapid spread throughout the world led the World Health Organization to declare the COVID-19 outbreak a global pandemic on March 11[th] 2020 (4). This triggered an unprecedented and sudden need for an exponentially growing amount of personal protection equipment and diagnostic tools. Real-time Polymerase Chain Reaction (PCR)-based detection of SARS-CoV-2 nucleic acids has rapidly become the gold standard to diagnose patients infected with SARS-CoV-2 (5). As typically performed for the detection of respiratory viruses, accurate COVID-19 PCR assay relies on the collection of samples from the upper respiratory tract including nasopharyngeal (NP) and oral mucosal surfaces (6). Flocked swabs feature perpendicular fibers that optimize specimen collection and elution in transport media and are hence considered optimal for sampling the respiratory tract mucosal surfaces. As testing rapidly became critical for the development of a COVID-19 response strategy, the world encountered shortage of PCR reagents and sampling swabs, resulting in testing backlogs, delayed diagnosis, compromised contact tracing and quarantine of patients and, potentially increased disease   transmission.

The versatility of 3D-printing as well as the possibility of rapidly developing prototypes have enabled rapid mobilization of this technology to provide a response to the interruption of supply chains (7, 8). 3D printed alternatives to standard flocked swabs were rapidly developed. Early in the pandemic, open-sourced design for 3D printed swabs, were generously made available to the

community by the teams from University of South Florida's (USF) Health Radiology and

Northwell Health System (9, 10). The design and workflow for hospital-based printing was

subsequently published (8). Local manufacturing based on 3D printing is among the strategies

that can help alleviate supply chain shortages.

In light of the anticipated shortage of sampling swabs in our hospital and in the province of

Quebec, we sought to locally manufacture and evaluate sterile 3D printed swabs based on the

freely available open source designs as replacement for locally used commercial swabs (8). Here,

we report on the fabrication process and clinical evaluation of the 3D printed swabs against the

cobas PCR Media□ swab sample kit in a prospective cohort of symptomatic healthcare workers.

To our knowledge, this is the first study to report on autonomous hospital-based production and

evaluation of 3D printed swabs. Our study showed that our locally printed swabs were a reliable

alternative to commercial swabs and confirmed the initial assumption made by the USF and

Northwell Health system groups that hospital-based production of 3D printed swabs can

constitute a rapid, on-demand local response to provide replacements for pandemic-related

disruption of the swabs supply chains.

METHODS

3D printing

3D prototypes of swabs were designed and shared by teams from the Division of 3D Clinical

Applications at USF and Northwell Health system (8). All final printed models successfully

passed a complete set of mechanical tests (11). In our hospital-based production, we used the

model referred to as "Northwell model" to which a breaking point located at 80 mm from the tip

was added to the original design.

Swabs were printed by stereolithography (SLA) using Formlabs Form 3 and 3B printers

(Formlabs, Somerville, Massachusetts, USA) with Surgical Guide resin (Formlabs, Cat#RS-F2-

SGAM-01) that is biocompatible and sterilizable. The Preform software (Formlabs) was used to

create an array of 256 swab models to be converted into printer instructions. The thickness of

each printed layer was set to 0.1 mm. Printed swabs were cleaned in 99% isopropyl alcohol

(IPA) (Sigma Aldrich, Cat#PX1835-3) for 20 minutes using the Form Wash (Formlabs) station

and allowed to air dry for at least 30 minutes before being post-cured at 70°C for 30 minutes in a

Form Cure (Formlabs) station. Visual inspection was performed to verify that printed swabs

were free of defects, otherwise they were discarded.

Sterilization

At the end of the production procedure, 3D printed swabs were immediately individually packed

in autoclavable flat pouches (4 in × 10.5 in, Stevens Company, #164-S5) previously identified

with autoclave-resistant laser labels (GA International, #AKA-13) indicating the swabs assigned

lot number and date of production. Pouches were sealed with a vacuum rotosealer machine

(Wipak Medical, #RS120). Appropriate sealing was verified visually. Sterilization was

performed by autoclaving using a pre-vacuum steam cycle set at 132°C (270°F) for 4 min

followed by a 30 min drying period.

To verify the efficacy of the sterilization procedure for each lot, the head of two swabs per

production lot were inoculated with 10 μL of the Biological Indicator *Geobacillus*

stearothermophilus spore suspension ($1.7*10^7$ CFU/0.1 mL; Steris Corporation, #NA-091) under

sterile conditions (12). Inoculated swabs were held horizontally for 30min and further let dry

vertically for 24 h under sterile conditions. Next, dried swabs were sealed individually in

sterilization pouches. One of the inoculated swabs was subjected to sterilization as described

above, together with all swabs from the same lot. The other inoculated swab was kept in the

pouch and was not sterilized. The two inoculated swabs were broken at the breakpoint and

inserted into a bacteriology tube (Sarstedt, #62515006) containing 5 mL of tryptic soy broth

culture media (Sigma Aldrich, #1463170010). Bacteria growth was monitored by incubation at

55°C with orbital agitation for 7 days. A culture media only negative control was also used.

OD600nm was measured on day 7 to assess bacterial growth.

Quality control

Swabs showing excessive warping post-sterilization were discarded after visual inspection. Basic

mechanical testing was performed using a guide formed by 3 semicircular canals with a

minimum radius of 15, 25 and 35 mm, respectively, that allowed testing the flexibility of swabs.

The test was successful if the head and neck remained intact after going through each canal. The

final test consisted in breaking the swab in half with one hand at its breaking point.

3D swabs clinical evaluation study participants' recruitment

This study was performed in our institution's COVID-19 rapid screening clinic and included symptomatic healthcare workers self-presenting for COVID-19 diagnostic testing. During an initial recruitment phase, participants were tested simultaneously with both the cobas PCR Media® swab sample kit (Roche Diagnostics, Laval, Quebec, Canada) and the 3D printed swabs. As for any COVID-19 test, all results were transmitted to the institution occupational health and safety office and public health authorities. In a subsequent recruitment phase, to increase the number of positive samples within the study, patients having previously tested positive on routine testing were contacted by clinical research personnel and offered to participate in the study. For those participants, both sampling techniques were repeated simultaneously.

Oro-nasopharyngeal swab collection and SARS-CoV-2 PCR testing

A sequential oropharyngeal and nasopharyngeal sampling with each single swab was performed. For the nasopharyngeal sampling, Cobas PCR Media swabs sample kit and 3D printed swab were used in the same nostril in a randomized order. The swabs were transferred separately to tubes containing the cobas PCR Media transport medium before proceeding with the PCR analysis. All samples were tested on the FDA emergency use authorization (EUA) approved and locally validated cobas 8800 automated RT-PCR system which simultaneously tests the *ORF1 a/b* and *E*-gene viral molecular targets together with an internal control (13).

Statistical analysis

The Lilliefors statistical test adapted from the Kolmogorov–Smirnov nonparametric test was used to assess data distribution normality. The Wilcoxon matched-pairs test for non-normal distributions was used to evaluate the difference between means of RT-PCR cycle thresholds (Ct) obtained following both sampling methods. Concordance analysis between both assays using overall, positive and negative agreement percentages was performed with calculation of the Cohen's Kappa values. By definition, Kappa values above 0.75 indicate excellent agreement, values between 0.40 and 0.75 indicate fair to good agreement, and values below 0.40 represent poor agreement beyond chance (14). Results obtained with the cobas PCR Media® swabs sample kit were considered as reference for positive and negative agreement calculation purposes.

Ethical approval

The study received ethical approval from the Comité d'éthique de la Recherche of the Centre de Recherche du Centre Hospitalier de l'Université de Montréal (CHUM, 2021-9015 / 20.079). All participants were provided written informed consent.

RESULTS

3D printed swab model

In a pilot print to start implementing a hospital production of 3D printed swabs, a biocompatible and sterilizable surgical guide resin to print two models from the designs shared by the 3D Clinical Applications Division of the USF and Northwell Health System was used. The two printed models are referred to as "USF" and "Northwell" (8). Based on the flexibility and the smaller head size, the "Northwell model" was selected by the clinical diagnostic team to pursue with the clinical tests. The cattail design of the "Northwell" model is composed of a head (18.0 mm long and 3.3 mm in diameter), a flexible neck (56.0 mm; 1.2 mm) and a handle (77.0 mm; 2.6 mm). Based on this pilot evaluation, a breaking point (1.0 mm; 1.4 mm) located at 80 mm from the tip to the original design was added to facilitate the release of the swab head in the transport tube containing the transport medium (**Figure 1A-C**). Three separate lots of 256 swabs were printed to ensure reproducibility of the production process and of the subsequent clinical evaluation. Swabs were packed individually in sterilization pouches before autoclaving (**Figure 1D**). Efficiency of the sterilization was verified for each lot through inhibition of *Geobacillus stearothermophilus spore* suspension inoculated on the head of one swab per lot (**Figure 2**). Two swabs from each printed lot were subjected to quality control check including testing of the flexibility (**Figure 1E**) and breaking of the swabs at the breaking point.

Clinical testing

A total of 63 participants were enrolled in the study. Thirty-two participants tested negative and

31 tested positive with the cobas 8800 SARS-CoV-2 PCR. Swabs from the 3 distinct printed lots

were used for sample collection (batch 1, n = 21, 11 positives, 10 negatives; batch 2, n = 21, 11

positives; 10 negatives; batch 3, n = 21, 9 positives, 12 negatives). Lilliefors statistical test and

Kolmogorov–Smirnov nonparametric test showed that cycle threshold (Ct) distributions for both

swabs were not standard (p=0.12). PCR results for the *E*-gene (p=0.27), *ORF1* gene (p=0.92) and

internal control (p=0.59) did not show significant Ct differences between flocked swabs and 3D

printed swabs (**Figure 3 and Table 1**). A full agreement table between the flocked and the 3D

printed swabs is presented in **Table 2A**. Overall, positive and negative agreements were

respectively 96.8% (61/63), 96.9% (31/32) and 96.8% (30/31) with a Kappa value of 0.936

(**Table 2B**). In 2 cases, results obtained with the 2 swabs were discordant. In one case the

flocked swab sample was positive with a Ct of 37.3 for the *E* gene only. For this same

participant, the sample obtained with the 3D printed swab led to no amplification for both genes.

In the second discordant case, the flocked swab sample was negative for both genes, while the

3D printed swab result was positive for both *ORF1* (Ct = 33.2) and *E* (Ct = 38.1) targets.

Although, in both discordant cases the detected Ct were high, several other samples included in

this study also had Ct in the high range without showing discordant results.

DISCUSSION

In this report, we described the implementation of a hospital-based 3D printing platform for the

production of swabs for nasopharyngeal and oropharyngeal sampling for COVID-19 diagnosis.

Such swabs are classified as Class I medical devices under the Canadian regulatory framework

of Health Canada. To validate our 3D printed swabs, we performed a head-to-head diagnostic

accuracy study of the 3D printed "Northwell model" swab (8) with the cobas PCR Media swabs

sample kit. We observed a high concordance (total agreement 96.8%, Kappa 0.936) in results

obtained with the 3D printed and flocked swabs indicating that the in-house 3D printed

"Northwell model" swab can be used reliably in a context of shortage of flocked swabs. Previous

clinical trial performed by the USF and Northwell Health System teams have reached similar

conclusions by comparing the 3D printed "USF model" swab with standard flocked swabs using

alternative transport medium, including the World Health Organization (WHO) approved viral

transport media, media produced in-house according to the procedure described by the Centers

for Disease Control and Prevention (CDC) or commercially available Universal Transport Media

(10). At the time of starting our clinical trial, our hospital was not facing a shortage of supply, or

planning to be out of stock in the medium term, of swabs for the diagnosis of COVID-19.

Therefore, we were able to use the transport medium provided in the cobas PCR Media swabs

sample kit with the 3D printed swab to ensure consistency in the clinical evaluation. The RT-

PCR was performed through the trial on a single local cobas 8800 automated RT-PCR system

authorized by Health Canada under an interim order for use related to COVID-19 diagnosis after

complete published validation of the assay (13, 15). Given the complexity and discomfort

associated with repeated simultaneous nasopharyngeal testing and the necessity to validate the

use of 3D printed swab in a controlled head-to-head approach, we did not include other transport

media in our validation study. We did not observe significant Ct values differences between

swabs for the *ORF1* and *E* SARS-CoV-2 genes in the Roche cobas assay. These observations are

in agreement with the data from the clinical trial performed by the USF and Northwell system

teams using the 3D printed "USF model" swab compared to the Roche cobas sampling kit

(97.03% agreement, Kappa 0.863) (10). In a study comparing swabs from four distinct 3D-

printing manufacturers to the Copan swab (501CS01) using a RT-PCR run on a Abbott m2000

RealTime system platform, high degree of concordance with Kappa values between [0.85-0.89]

were observed (16). An additional prospective clinical validation compared 3D-printed swabs

from two manufacturers with the Universal Viral Transport Kit by Becton, Dickinson &

Company using a liaison MDX RT-PCR machine (DiaSorin Molecular, LLC) and the Simplexa

COVID-19 Direct Kit. Again, an excellent concordance between sampling procedures was

observed (17).

To our knowledge, this is the first study reporting on autonomous hospital-based production and

evaluation of 3D printed swabs. The implementation of our hospital-based platform was

facilitated by the existing "Health-related 3D printing centre" in our institution, which normally

serves the needs of the Radiation Oncology department and other hospital sectors, but quickly

reoriented its efforts to provide healthcare workers with the much-needed protection equipment

and diagnostic tools. Additionally, we had access to an institutional sterilization service that

allowed us to easily perform the steam sterilization that was chosen as a rapid, nontoxic and

inexpensive technique that is microbicidal and sporicidal, and was previously shown to be

compatible with the Surgical Guide resin (18). Actual installations and staff availability allowed

us to limit the production costs of 3D printed swabs. The total cost per swab was about 0.56USD,

which included 0.26USD for consumables and 0.29USD for staff wages. Further optimization

has yet to be realized to increase the production volume and lower the production costs.

In conclusion, our study adds to the few clinical validation studies that demonstrated safety and

accuracy of 3D swabs. Our study is unique in that it tested a fully integrated hospital-based

production of 3D-printed swabs as initially suggested by the USF and Northwell Health system

groups. Our clinical trial has demonstrated that our local 3D printed swab production line offers

a reliable local alternative to commercial swabs and therefore confirms that it is a viable local

response to provide replacements in the event of pandemic supply chain disruption. In-house

production of 3D printed swabs option remains particularly relevant as the need for testing

capacity continues to increase across the world, as many countries are experiencing new waves

of infection with the emergence of variants of SARS-CoV-2 (19, 20). Our experience in the rapid

implementation of this production line could serve as an example for other institutions around

the world in the fight against COVID-19.

ACKNOWLEDGEMENT

The authors thank Floriane Point and Stephanie Matte for help with the ethics approval and

organization of patient recruitment. The authors thank all the patients who generously accepted

to participate in this study and the laboratory technicians involved in this project. This study was

funded by the TELUS Friendly Future Foundation/Fondation TELUS pour un futur meilleur and

the Fondation du Centre Hospitalier de l'Université de Montréal. The funders had no role in the

study design, data collection and interpretation, or the decision to submit the work for

publication. SGL has received funding from Roche Diagnosis unrelated to the present study to

test PCR Medium stability. FC received grants paid to the organization for research projects

unrelated to the present study from Roche Diagnostics and Merck Sharp and Dome, honorariums

for presentations from Merck Sharp and Dome and Roche diagnostics, and has participated in an

expert vaccine group by Merck Sharp and Dome.

REFERENCES

1.   Zhu N, Zhang D, Wang W, Li X, Yang B, Song J, Zhao X, Huang B, Shi W, Lu R, Niu

P, Zhan F, Ma X, Wang D, Xu W, Wu G, Gao GF, Tan W. 2020. A Novel Coronavirus

from Patients with Pneumonia in China, 2019. The New England Journal of Medicine

382:727-733.

2.   Sanche S, Lin YT, Xu C, Romero-Severson E, Hengartner N, Ke R. 2020. The Novel

Coronavirus, 2019-nCoV, is Highly Contagious and More Infectious Than Initially

Estimated. medRxiv:2020.02.07.20021154.

3.   Zheng J. 2020. SARS-CoV-2: an Emerging Coronavirus that Causes a Global Threat.

International journal of biological sciences 16:1678-1685.

4.   Organization WH. 11 March 2020 2020. WHO Director-General's opening remarks at

the media briefing on COVID-19 - 11 March 2020, *on* World Health Organization.

https://www.who.int/director-general/speeches/detail/who-director-general-s-opening-

remarks-at-the-media-briefing-on-covid-19---11-march-2020. Accessed

5.   Cheng MP, Papenburg J, Desjardins M, Kanjilal S, Quach C, Libman M, Dittrich S,

Yansouni CP. 2020. Diagnostic Testing for Severe Acute Respiratory Syndrome-Related

Coronavirus 2: A Narrative Review. Ann Intern Med 172:726-734.

6.   Wang H, Liu Q, Hu J, Zhou M, Yu MQ, Li KY, Xu D, Xiao Y, Yang JY, Lu YJ, Wang

F, Yin P, Xu SY. 2020. Nasopharyngeal Swabs Are More Sensitive Than Oropharyngeal

Swabs for COVID-19 Diagnosis and Monitoring the SARS-CoV-2 Load. Front Med

(Lausanne) 7:334.

7.   Choong YYC, Tan HW, Patel DC, Choong WTN, Chen C-H, Low HY, Tan MJ, Patel

CD, Chua CK. 2020. The global rise of 3D printing during the COVID-19 pandemic.

Nature Reviews Materials 5:637-639.

8.   Ford J, Goldstein T, Trahan S, Neuwirth A, Tatoris K, Decker S. 2020. A 3D-printed

nasopharyngeal swab for COVID-19 diagnostic testing. 3D Printing in Medicine 6:21.

9.   Rybicki FJ. 2020. 3D Printing in Medicine: COVID-19 Testing with 3D Printed

Nasopharyngeal Swabs. Clin Infect Dis ciaa1437.

10.   Decker SJ, Goldstein TA, Ford JM, Teng MN, Pugliese RS, Berry GJ, Pettengill M,

Silbert S, Hazelton TR, Wilson JW, Shine K, Wang ZX, Hutchinson M, Castagnaro J,

Bloom OE, Breining DA, Goldsmith BM, Sinnott JT, O'Donnell DG, Crawford JM,

Lockwood CJ, Kim K. 2020. 3D Printed Alternative to the Standard Synthetic Flocked

Nasopharyngeal Swabs Used for COVID-19 testing. Clin Infect Dis ciaa1366.

11.   Spadaccini C, Duoss E, Shusteff M, Tooker A, Haque R. 2020. Swab Tensile Testing

Results and Procedures. USDOE National Nuclear Security Administration (NNSA)

LLNL-TR-808538 1014512.

12.   Albert H, Davies DJ, Woodson LP, Soper CJ. 1998. Biological indicators for steam

sterilization: characterization of a rapid biological indicator utilizing Bacillus

stearothermophilus spore-associated alpha-glucosidase enzyme. J Appl Microbiol

85:865-74.

13.   Boutin CA, Grandjean-Lapierre S, Gagnon S, Labbé AC, Charest H, Roger M, Coutlée F.

2020. Comparison of SARS-CoV-2 detection from combined

nasopharyngeal/oropharyngeal swab samples by a laboratory-developed real-time RT-

PCR test and the Roche SARS-CoV-2 assay on a cobas 8800 instrument. J Clin Virol

132:104615.

14. Fleiss JL. 1981. Statistical Methods for Rates and Proportions. Second Edition. Wiley,

John and Sons, Incorporated, New York, N.Y.

15. Canada H. 18 March 2020 2020. Health Canada expedites access to COVID-19

diagnostic laboratory test kits and other medical devices, *on* Health Canada.

https://www.canada.ca/en/health-canada/news/2020/03/health-canada-expedites-access-

to-covid-19-diagnostic-laboratory-test-kits-and-other-medical-devices.html. Accessed

16. Callahan CJ, Lee R, Zulauf KE, Tamburello L, Smith KP, Previtera J, Cheng A, Green A,

Abdul Azim A, Yano A, Doraiswami N, Kirby JE, Arnaout RA. 2020. Open

Development and Clinical Validation of Multiple 3D-Printed Nasopharyngeal Collection

Swabs: Rapid Resolution of a Critical COVID-19 Testing Bottleneck. J Clin Microbiol

58:e00876-20.

17. Oland G, Garner O, de St Maurice A. 2021. Prospective clinical validation of 3D printed

nasopharyngeal swabs for diagnosis of COVID-19. Diagn Microbiol Infect Dis

99:115257.

18. Sharma N, Cao S, Msallem B, Kunz C, Brantner P, Honigmann P, Thieringer FM. 2020.

Effects of Steam Sterilization on 3D Printed Biocompatible Resin Materials for Surgical

Guides-An Accuracy Assessment Study. J Clin Med 9:1506.

19. Sabino EC, Buss LF, Carvalho MPS, Prete CA, Jr., Crispim MAE, Fraiji NA, Pereira

RHM, Parag KV, da Silva Peixoto P, Kraemer MUG, Oikawa MK, Salomon T,

Cucunuba ZM, Castro MC, de Souza Santos AA, Nascimento VH, Pereira HS, Ferguson

NM, Pybus OG, Kucharski A, Busch MP, Dye C, Faria NR. 2021. Resurgence of

COVID-19 in Manaus, Brazil, despite high seroprevalence. Lancet 397:452-455.

20. Abdool Karim SS, de Oliveira T. 2021. New SARS-CoV-2 Variants - Clinical, Public

Health, and Vaccine Implications. N Engl J Med.

FIGURE LEGENDS

Figure 1. In-house 3D printed swab model. Design of the Northwell 3D swab model with the

addition of a breakout point (**A**) used to 3D-print swab in our hospital (**B and C**). Swabs were

individually packed in autoclavable and vacuum-sealed poches (**D**) for sterilization. (**E**)

Flexibility was mechanically tested using semicircular canals (radius of 15, 25 and 35 mm).

Figure 2. Validation of 3D printed swabs sterilization. Swab heads inoculated with *G.*

stearothermophilus spore suspension before sterilization cultured in soy broth culture media.

Bacteria growth was assessed by measuring the optical density (O.D.) at 600nm. Culture media

alone was used as negative control.

Figure 3. Cycle threshold (Ct) values of reverse-transcriptase polymerase chain reaction

(RT-PCR) for the ORF1 and E genes. Participants were swabbed in the same nostril with a

flocked and a 3D printed swab, successively. RT-PCR was performed to measure the Ct values

of *ORF1* (A) and *E* (B) viral genes for each swab. An internal control (Ctrl) was also included

(C). Statistical analyses are detailed in Table 1.

377 **TABLE 1.** Means Ct values of samples collected from individuals tested with a flocked or a 3D

378 printed swab[a].

	Flocked mean Ct	3D mean Ct	Delta Ct	p-value
ORF1	26.06	26.51	0.44	0.92
E	28.03	28.30	0.26	0.27
Ctrl	35.16	35.35	0.19	0.59

389 [a]Difference between means was evaluated with the Wilcoxon matched-pairs test.

TABLE 2. Tables of agreement between flocked and 3D printed swabs showing positive and negative concordance percentages with Kappa values (**A**) and total number of positive (+), negative (-) and inconclusive (IC) PCR results (**B**)[b].

A

		3D	3D	3D	Total
		+	-	IC	
Flocked	+	30	1		31
Flocked	-	1	31	0	32
Flocked	IC	0	0	0	0
Total		31	32	0	63

B

	%	95% IC
Positive concordance	96,8	82,4 - 99,9
Negative concordance	96.9	82,9 - 99,9
Total concordance	96.8	88,5 - 99,8
Kappa value	0.936	0,738 - 0,994

[b] IC: Confidence Interval

Culture media   Sterilized   Non-sterilized

0.124   0.098   1.625

(O.D.$_{600nm}$)