A Comparison of Four Commercially Available RNA Extraction Kits for Wastewater

Surveillance of SARS-CoV-2 in a College Population

Authors: Megan O'Brien[1*], Zachary C. Rundell[2*], Michelle D. Nemec[1], Laura M. Langan[3],

Jeffrey A. Back[2], Joaquin N. Lugo[4]. [1] Molecular Biosciences Center, [2]Center for Reservoir and

Aquatic Systems, [3]Department of Environmental Sciences, [4]Department of Psychology and

Neuroscience, Baylor University, Waco, TX 76798, USA.

Keywords: wastewater-based epidemiology; WBE, sewage, COVID-19

19 Corresponding author
20

Dr. Joaquin N. Lugo

Department of Psychology and Neuroscience

One Bear Place 97334

Baylor University

Waco, TX 76798

Tel: 254-71-2389

Email: Joaquin_lugo@baylor.edu

Abstract

Localized wastewater surveillance has allowed for public health officials to gain a broader

understanding of SARS-CoV-2 viral prevalence in the community allowing public health

officials time to prepare for impending outbreaks. Given variable levels of virus in the

population through public health interventions, proper concentration and extraction of viral RNA

is a key step in ensuring accurate detections. With many commercial RNA extraction kits and

methodologies available, the performance of 4 different kits were evaluated for SARS-CoV-2

RNA detection in wastewater, specifically focusing on their applicability to lower population

densities such as those at university campus dorms. Raw wastewater samples were collected at 4

sites on a college campus over a 24 hour period as a composite sample. Included in these sites

was an isolation site that housed students that tested positive for Covid-19 via nasopharyngeal

swabs. These samples were analyzed using the following kits: Qiagen All Prep PowerViral

DNA/RNA kit, New England BioLabs Monarch RNA MiniPrep Kit, and Zymo Quick RNA-

Viral Kit, and the Zymo Quick-RNA Fecal/Soil Microbe MicroPrep Kit. All four sites were

processed according to the manufacturer's guidelines. Extractions were then quantified with RT-

qPCR one-step reactions using an N2 primer and a linearized plasmid standard. While the Zymo

Quick-RNA Fecal/Soil Microbe MicroPrep Kit (also known as the Zymo Environ Water RNA

Kit) only recovered approximately 73% (+/- 38%) SARS-CoV-2 RNA compared to the Zymo

Quick-RNA Viral kit, it was the most time efficient kit to yield comparable results. This

extraction kit had a cumulative processing time of approximately five hours compared, while the

other three kits had processing times between approximately 9 and 9.5 hours. Based on the

current research, the most effective kits for smaller population densities are pellet based and

include a homogenization, inhibitor removal, and RNA preservation step.

1. Introduction

On January 31, 2020, the WHO declared COVID-19 a Public Health Emergency of International Concern (Jee, WHO, 2020). Widespread testing of those with and without COVID-19 symptoms, in combination with the numerous vaccine roll outs is vital to curtailing the current pandemic and future outbreaks. The current estimate of infected individuals is believed to be underestimated worldwide, with numerous nations initially encouraging testing for those only with symptoms. Therefore, those who are presymptomatic or asymptomatic are often less likely to be identified, thus posing a significant potential for transmission, with studies estimating that asymptomatic or presymptomatic transmission could be responsible for up to 50% of new cases (Ghandi et. al 2020, Moghadas et. al 2020).

A variety of diagnostic testing methods are available to determine if individuals are infected with COVID-19. Current diagnostic testing involves the collection and PCR analysis of infected cells and bodily fluids for the SARS-CoV-2 virus by drawing blood or collecting samples from the nose, mouth, throat, or lungs (Ravi et. al 2020). While useful, these testing methods are hazardous, resource-intensive, and invasive. Ongoing shortages and supply chain disruptions have been a major challenge in implementing diagnostic PCR tests (Binnicker 2020). Furthermore, diagnostic PCR tests may be insensitive for the first 5-7 days following infection (Binnicker 2020). Pooled testing of samples may offer a potential solution to supply chain issues, but pooled testing increases the likelihood of false negatives in low-level positive samples and sample contamination (Binnicker 2020).

Wastewater based epidemiology offers a promising method of Covid-19 surveillance that may solve some of these pressing issues. Although a relatively new field, it has conventionally been successfully used to estimate use of legal and illegal drugs of abuse and to evaluate human

exposure to contaminants and pathogens as summarized in Lorenzo et al. (2019). Active

monitoring of SARS-CoV-2 RNA in wastewater can be a useful tool for identifying hotspots and

has been demonstrated to serve as an early warning system for new outbreaks (Xagorarki et. al

2019, Venugopal et. al 2020, Betancourt et. al 2020). This method is unique as it allows

researchers to survey large groups of people quickly with fewer resources and staff. Wastewater

based epidemiology is also less intrusive compared to nasal swabs and reduces occupational

exposure to SARS-CoV-2. Perhaps the largest benefit of wastewater surveillance is its efficiency

and ability to view community prevalence. Thus, areas with higher viral copies may be focused

on for individual testing efforts and public health interventions. Wastewater surveillance works

due to SARS-CoV-2 RNA being detectable in the feces of both symptomatic and asymptomatic

individuals (Hart et. al 2020, Mizumoto et. al 2020, Treibel et. al 2020), even after the

individuals no longer had respiratory symptoms (Zheng et. al 2020, Mesoraca et. al 2020, Jones

et. al 2020). This method of detection is exceptionally sensitive, with one modeling exercise

suggesting that wastewater surveillance could theoretically be able to detect one SARS-CoV-2

infection among 2,000,000 individuals (Hart et. al 2020). Although there is some contrasting

information, when levels go below 25 cases per 100,000 (as is likely), the methodology has to be

adjusted as outlined in D'Aoust, 2020 in which was measured and then estimated, rather than

modelled (D'Aoust et al. 2021).

Wastewater based epidemiology is a powerful tool that can provide vital information

about the spread of Covid-19 and can be useful in prioritizing diagnostic PCR testing. It has

demonstrated its ability to effectively aide in detecting affected individuals so students could be

tested or isolated to prevent further spread on a college campus (Betancourt et. al 2020).

Unfortunately, a need exists for standardized techniques in applying this method to Covid-19

surveillance (Ahmed et. al 2020). As populations contributing to the wastewater vary in their

characteristics, so does the overall matrix of the wastewater sample (Kitajima et. al 2020).

Consequently, numerous methods of sample processing exist, which can make it difficult for

researchers to choose an optimal method for accurate detection. Furthermore, sampling in small

population densities adds another layer of complexity to detection due to a potential for lower

viral loads than typical in municipalities. Monitoring in communities with low incidence has

previously demonstrated high PCR Ct values and hence variable or unquantifiable data being

collected due to very low concentrations of the viral fragment in the collected samples (D'Aoust

et al. 2021).

The selection of an appropriate RNA isolation kit is a key component of processing

samples that can have a major impact on the results yielded. This study aims to provide an

overview of the efficacy and efficiency of four common RNA isolation kits produced by Zymo,

Qiagen, and New England Biolabs when surveilling a small population for SARS-CoV-2 in

wastewater. The endpoints examined in this study include a comparison of viral detection across

all four kits as well as a qualitative description of each method.

2. Materials and Methods

2.1 Sample Collection

Wastewater was collected from 4 buildings in Waco, Texas on Baylor University's

campus. Sites A-C were collected from 3 dormitories and with a total, combined population of

approximately 850 students at time of collection. Included in these sites was a dormitory that

consistently yielded non-detectable values (site B). An additional isolation site, (Site I) was also

included, this site housed an unknown number of students who were isolated due to active

SARS-CoV-2 infections.

Composite 8.64-liters samples were collected over a 24-hr period in polypropylene

bottles from 11:00 am on 10/6/2020 to 10:45 am on 10/7/2020. ISCO model 6712 automatic

samplers were programmed to collect composite samples in 90 mL increments every 15 minutes.

The sample bottle chamber was filled with ice to keep the samples cold. Upon collection, each

composite sample was mixed by hand and an aliquot was poured into a 250 mL polypropylene

centrifuge bottle (Fisherbrand, catalog # 14-375-352) and stored on ice. Samples were processed

approximately 1 hour after collection.

2.2 RNA Concentration and Extraction

Following collection, samples were concentrated by centrifuging each 250 mL sample in

the original bottle at 4˚C for 45 minutes on a coast deceleration setting so as not to disturb the

pellet (AVANTI JXN 26, JS-7.5 rotor, 4700 RCF). Supernatant (150 mL) was collected and

aliquoted for use in three of the extraction kits. The remaining 100 mL pellet was resuspended

and used for the pellet-based extraction kit.

The resulting 150 mL filtrate was further concentrated using ultrafiltration with

AMICON 15 mL conical filtration tubes (Sigma Aldrich; UFC901024). Filtrate was loaded in

increments of 15 mL until the total collection volume had been reduced to~ 1000 µL (5000 RCF,

Eppendorf A-4-62 swing-bucket rotor with adaptors). In some cases, the filters became clogged

and a new ultrafilter was required and when necessary the sample was transferred to a new

filtration tube. Concentrate was then aliquoted (~250 µ)) and RNA exactions carried out using

extraction kits 1-3: Qiagen All Prep PowerViral DNA/RNA kit (80244), New England BioLabs

Monarch RNA MiniPrep Kit (T2010S) and Zymo Quick RNA-Viral (R1034) following

manufacturer instructions.

For the pellet, 45 out of the 100 mL of the resuspended pellet was collected into 50 mL conical

tubes. Urine conditioning buffer (3.150 mL, D3061-1-140) was added thoroughly mixed, then

centrifuged (5000 RCF x 15 min, RT; Eppendorf FA-45-6-30 fixed rotor). Following

centrifugation, the remaining filtrate was removed (~47.9 mL) leaving ~ 250 µL concentrated

pellet composite. To this, 750 µL of RNA/DNA shield (Zymo, R1100-250) was then added to

the sample and held at 4°C until extraction. All extractions took place in an RNAse, DNAse-free

hood environment, following manufacturers guidelines. A true blank site was run for primary

concentration/processing integrity. RNA extraction blanks were included with each extraction

batch and kit.

2.3 PCR Analysis

Quantification of viral load was determined via RT-qPCR (QuantStudios 6 Flex) using

New England BioLabs Enzyme and Probe Master kit (New England Biolabs, Catalog E3006X)

with IDT N2 RUO primers (IDT, catalog #10006713). Each plate contained triplicates of each

condition; whose CT's were averaged to get the mean CT for each sample for each extraction

method. The standards used contained a linearized 200,000 cp/uL N plasmid standard (IDT,

catalog #10006625) which was diluted into a 10,000 copies/µL stock. Each plate's standard

curve was conducted using an 8 series dilution, starting with 10,000 copies/µL to 2.441

copies/µL. In order to accurately assess whether inhibitors are acting upon collected samples, 1:2

dilutions were made from the RNA extract directly prior to plate analysis. Non-template controls

confirmed PCR integrity. More information on PCR analysis can be found in the Supplemental

Information.

3. Results

To benchmark the performance of the kits, the controls (viz. field blanks, extraction

blanks and PCR blanks, and no template controls) were first all confirmed to be non-detectable

for SARS-CoV-2. Thereafter, the individual sites were examined. Site B acts as a control for the

methodology, with consistency low incidence levels historically recorded. For this sampling

location, Site B resulted in a non-detectable sample across all RNA extraction kits evaluated

(Table 1), with contrastingly high viral loads identified at the isolation site. In contrast, Site A

and C both tested positive, with detectable levels consistent even between dilutions, with the

exception of the Monarch RNA Cleanup Kit, which saw a non-detected value for the original

sample, yet when diluted yielded a highly positive value. There were differences between each of

the kits' amplifications, of which the Qiagen kit performed consistently between dilutions,

potentially indicating adequate inhibitor removal, but yielded lower values compared to the other

kits.

188   **Table 1: Campus sites and SARS-CoV-2 detection in Copies/L with 4 different commercial RNA**
189   **extraction kits.**

Copies/Liter + Coefficient of Variation (CV)			
New England Biolabs Monarch RNA MiniPrep Kit	**Qiagen AllPrep PowerViral DNA/RNA Kit**	**Zymo Quick RNA-Viral Fecal/Soil Microbe Microprep Kit**	**Zymo Quick RNA-Viral with Inhibitor Removal**
Site A			
undiluted Not detected	6.2E+03 (19% CV)	1.3E+04 (20% CV)	3.9E+04 (11% CV)
1:2 1.6E+04 (8% CV)	5.3E+03 (38% CV)	1.1E+04 (12% CV)	4.0E+04 (16% CV)
Site B			
undiluted Not detected	Not detected	Not detected	Not detected
1:2 Not detected	Not detected	Not detected	Not detected
Site C			
undiluted 1.2E+04 (29% CV)	5.7E+02 (87% CV)	1.2E+04 (38% CV)	1.0E+04 (21% CV)
1:2 1.0E+04 (12% CV)	2.5E+02 (74% CV)	7.3E+03 (7% CV)	9.8E+03 (10% CV)
Site Isolation			
undiluted 4.5E+07 (5% CV)	2.1E+06 (17% CV)	1.7E+07 (26% CV)	2.3E+07 (8% CV)
1:2 2.9E+07 (21% CV)	1.2E+06 (5% CV)	1.7E+07 (4% CV)	1.6E+07 (9% CV)

Between the kits, the Zymo Quick RNA-Viral with Inhibitor Removal consistently

yielded the highest resulting viral loads. There were similarities between the Zymo Fecal Kit and

the Quick-Viral RNA kit processes, including inhibitor removal steps. On average, the Zymo

Fecal Kit values ranging from approximately 7,300 to 17,000,000 copies/L whereas the Zymo

Quick-Viral RNA kit ranged from 9,800 to 23,000,000 copies/L. The Zymo Fecal Kit recovered

approximately 73% (+/- 38%) more SARS-CoV-2 RNA than the Zymo Quick-Viral RNA kit per

site. Examining the % Coefficients of variation between each site, each method, and each

dilution factor was between 4%-87%. The Zymo Quick-Viral RNA kit had the lowest

Coefficients of variations, whereas Qiagen's kit had some of the higher values. The Zymo Fecal

Kit pellet showed dilutions may improve Coefficients of variation values.

In regards to the isolation site, the New England BioLabs kit obtained the highest viral

load at 45,000,000 copies/L in the original sample, and 28,000,000 copies/L in the diluted

sample. However, New England BioLabs kit was the most variable in terms of consistency

between dilutions. Here, the original sample carried a higher load (45.3%) than the diluted

sample. The Zymo Fecal kit had the least variation between the dilution series (0.99%), as we

saw with the site comparisons.

The Impact of PCR inhibitors was examined using the isolation location (Site I), with

little difference observed in the Zymo Fecal kit between concentrate and diluted sample in

comparison to other kits. Interestingly, cp/L at the isolation site is markedly lower in the Qiagen

kit compared to others which is surprisingly considering that this kit has been used heavily in

wastewater testing of large municipalities.

Figure 1: Investigating impacts of dilutions between the four kits for a known positive

isolation site. Impact of PCR inhibitors were examined using the isolation location (Site I), with

little difference observed in the Zymo Fecal kit between concentrate and diluted sample in

comparison to other kits. Interestingly, cp/L at the isolation site is markedly lower in the Qiagen

kit compared to others which is surprisingly considering that this kit has been used heavily in

wastewater testing of large municipalities..

3.2 *Population Normalization*

Normalizing the data by population is another key in determining hotspots for potential

outbreaks. If the viral load per resident is high relative to the surrounding sites, this could be a

potential indicator of an upcoming outbreak. Following normalizing to population size, Site A

population (539) and Site C population (153), there is consistent differences in viral loads

between extraction kits and between sites, with the Qiagen kit reporting markedly lower viral

levels then the other kits examined. (Figure 2).

Figure 2: Average Copies/L normalized by population. Average Copies/L was calculated by

dividing the number of residents for a specific site.

3.3 Qualitative Results

Previous work has found that there is a lack of qualitative information for wastewater

concentration and RNA extraction methodologies (Ahmed et. al 2020). Therefore, we have

provided qualitative information based on previous experiments. All times are approximate, and

prices are listed as current for time of publication (Table 2). Footnotes can be found in

Supplemental Information.

238   **Table 2: Qualitative descriptions for each RNA extraction kit used.**

Qualitative Table for Extraction Methods	**New England Biolabs Monarch RNA MiniPrep**	**Qiagen AllPrep PowerViral DNA/RNA Kit**	**Zymo Quick-RNA Fecal/Soil Microbe MicroPrep**	**Zymo Quick-RNA Viral with Inhibitor Removal**
Time of Kit	30 mins	1 hour	1.5 Hours	20 mins + 20 mins for Inhibitor Removal
Price (at time of publication)	$17.9/sample	$24.47/sample	$10.44/sample	$24.23/sample
Extra Equipment	Floor Centrifuge	Floor Centrifuge		Floor Centrifuge
Consumables	AMICON Ultrafilters	AMICON Ultrafilters	Nuclease Free Tubes (2 mL)	AMICON Ultrafilters; Nuclease Free Tubes (2mL)
Type of Wastewater concentrate	Filtrate	Filtrate	Pellet	Filtrate
Extra Reagents	100% Ethanol	B-Mercaptoethanol	100% Ethanol	100% Ethanol
Extra Supplies	RNAse-free Tubes	No	RNAse-free Tubes	RNAse-free Tubes
Effectiveness	Lower	Lower	Higher	Higher
Overall Turn-around Time for 4 Samples	9 hours	9.5 hours	5 hours	9 hours
Qualitative Thoughts	If using this kit, we would suggest also using the DNAse I treatment recommended by representatives from New England Biolabs	Lost a significant amount of RNA sample compared to other kits	Kit is tedious with larger sample volumes, but overall gave best results	Especially with Inhibitor Removal step, could provide a comparative alternative for Fecal Kit, but longer processing times

240   These times are total working hours and assume no breaks in between each section of

241 processing, for only four samples. There is also the assumption that the technician has conducted

242 the protocols prior to beginning the process. It should be noted that at the time of publication, the

243 Zymo Quick-RNA Fecal/Soil Microbe MicroPrep Kit has been renamed as the Zymo Environ

244 Water RNA Kit (Zymo Research, catalog #R2042), in which the only difference is that the urine

245 conditioning buffer and DNA/RNA shield are included with the kit purchase.

246   Filtrate-based methods that use ultrafiltration considerably increased processing times

247 due to long centrifuge runs and the potential for clogging filters (Table 2). The Zymo Quick-

248 RNA Fecal/Soil MicroPrep kit required the most time, but yielded results more consistently

249 across the board. This methodology also does not require the use of a floor centrifuge, meaning it

250 may be available to a wider audience looking to start local surveillance.

251 *4. Discussion*

252 The ongoing pandemic of SARS-CoV-2 has allowed for greater expansion of wastewater based

253 epidemiology (WBE) as a tool for public health surveillance. Recent testing conducted at the

254 University of Arizona demonstrated the ability for campus wastewater surveillance to be used as

255 an early detection system to pinpoint potential hot spots and isolate individuals, even before

256 symptoms show (Betancourt et. al 2021). Historically, WBE has focused mostly on enteric

257 viruses, where many of these viruses are non-enveloped. Since SARS-CoV-2 is an enveloped

258 virus, this provides additional challenges for its concentration and detection in wastewater (Ye

259 et. al 2016, Polo et. al 2020, Torii et. al 2021). Traditional methods of RNA extraction, such as

260 polyethylene glycol precipitation (PEG) are not as effective at concentrating enveloped viruses

261 since the lipid membrane more sensitive and possibly degraded by organic solvents like

262 chloroform (Polo et. al 2020) and in past publications for SARS-CoV-2 have returned lower

recovery efficiencies (Ahmed et. al 2020, Torii et. al 2021). The current study populations varied

from 140-540 persons per site, compared with upwards of 10,000 or more for a municipal

wastewater treatment plant. Since PEG precipitations require small initial volumes, low

concentration/high volume samples are at a disadvantage and rely heavily on primary

concentration methods to be effective (Lu et. al 2020). The Qiagen AllPrep PowerViral

DNA/RNA kit has been used frequently in testing municipal wastewater for SARS-CoV-2 RNA,

where a higher population density is contributing to the collection sample. When considering

smaller populations and thus smaller RNA inputs, the Qiagen kit did not perform as well in our

study. This could be due to the selectivity of combined RNA/DNA columns, or the lack of

DNA/RNA preservation (such as DNA/RNA shield or urine conditioning buffer) throughout the

long filtration steps.

Primary concentration from raw wastewater samples has been demonstrated as an

obstacle in effectively concentrating viral loads (Hamouda et. al 2021, Ahmed et. al 2020).

Pellet-based methods have gained traction recently for higher quality and higher quantity of viral

copies per sample (Kitamura et. al 2021, Perez-Cataluna et. al 2020, D'Aoust et. al 2020,

Graham et. al 2021). The filtrate based kits used in this study required larger sample volumes

compared to the pellet-based kits, and the filtrate based kits also require an additional

ultracentrifugation step. Ultrafilter clogging was an issue with ultracentrifugation, and loss of

viral load is assumed with each tube transfer.

Finding effective methods to accurately identify low SARS-CoV-2 viral loads in

wastewater are important for sectors such as hospitals, nursing homes, or schools, which have a

lower flow rates than municipal wastewater treatment plants that are normally investigated

during broad surveillances. The Qiagen All Prep PowerViral kit columns select for both DNA

and RNA, where RNA could be potentially lost. New England BioLabs suggests the use of their

DNASE 1 treatment to aide in better recovery and prevent RNA degradation. The Zymo Quick-

RNA Fecal/Soil MicroPrep Kit included two forms of RNA preservation (DNA/RNA shield and

urine conditioning buffer) which may have aided in its higher recoveries. Based on these results,

we would recommend choosing a kit that specifically targets RNA extraction, and not a

combined kit. This would allow for DNA removal as well, with which we saw increased

detection in RNA viral loads. Homogenization of samples prior to extraction can aide in

breaking viral capsids and releasing viral RNA. The Qiagen kit did not include filtrate

homogenization steps, which could have affected its lower detection values. A kit that includes

an inhibitor removal component is also vital, even in samples with low population density. The

Monarch Kit did not include an inhibitor removal step during the extraction process, whereas the

other kits involved at least one inhibitor removal step. This would indicate that inhibitor removal

was indeed necessary to see higher amplifications in CT values, since diluting inhibited samples

is known to increase the efficiency of primer binding to cDNA, where inhibitors themselves are

then diluted (Hata et. al 2015). Dilutions could be conducted in the instance that a kit with

inhibitor removal is unavailable.

A greater need for surveillance raises questions with cost and accessibility. For smaller

sampling canvases, such as hospitals or nursing homes, a kit that can quantify smaller amount of

the virus accurately is important. Further, the ability to conduct in house processing would

greatly alleviate the expenditures of sending samples out to another lab. Time is also an

important factor to consider while choosing an extraction method. Depending on the number of

samples, as well as sample composition, time can vary greatly.

Since viral shedding through feces can be seen in both symptomatic and

asymptomatic/presymptomatic patients, health officials can see larger trends in community

prevalence than through nasal swabbing alone (Hart et. al 2020, Mizumoto et. al 2020, Treibel et.

al 2020). Duration of viral shedding in symptomatic patients can vary anywhere from a matter

14-21 days (Wu et. al 2021). It is indicated that viral shedding through feces precedes symptoms

of COVID-19, and thus is imperative that precautionary measures be taken as soon as possible to

prevent widespread contagion throughout the sample population, even in low prevalence areas

(Randazzo et. al 2020).

5. Conclusion

The results of this study show that a pellet-based RNA extraction kit that includes an

inhibitor removal and RNA preservation step may yield the most consistent, timely, and accurate

results. These additional steps may be why the Zymo Quick-RNA Fecal/Soil Microbe MicroPrep

(also called the Zymo Environ Water RNA kit) was the most effective and efficient kit with the

samples we used in the study. In contrast, the least effective kit was the Qiagen All Prep

PowerViral DNA/RNA Kit. This was likely due to the absences of a preservation step and

inefficiencies stemming from the kit selecting for both DNA and RNA. For filtrate based

methods, we recommend using the Zymo Quick-RNA Viral kit as an effective and efficient

method of wastewater based epidemiological analysis in concentrated wastewater samples,

especially in smaller population densities.

Acknowledgements

We would like to acknowledge the financial support from Baylor University and for support

from Dr. Bryan Brooks and Dr. Kevin Chambliss.

CRediT roles: **Megan O'Brien:** Conceptualization, Investigation, Visualization, Writing -

original draft; Writing - review & editing. **Zachary Rundell:** Conceptualization, Investigation,

Writing - original draft; Writing - review & editing. **Michelle Nemec:** Conceptualization,

Investigation, Writing- Reviewing and Editing. **Laura Langan:** Conceptualization,

Methodology, Writing- Reviewing and Editing. **Jeffrey Back:** Conceptualization, Methodology,

Writing- Reviewing and Editing.: **Joaquin Lugo:** Conceptualization, Writing- Reviewing and

Editing, Supervision.

References

Ahmed, Warish, et al. "Surveillance of SARS-CoV-2 RNA in Wastewater: Methods

Optimization and Quality Control Are Crucial for Generating Reliable Public Health

Information." *Current Opinion in Environmental Science & Health*, vol. 17, 2020, pp. 82–

93., doi:10.1016/j.coesh.2020.09.003.

Betancourt, Walter Q., et al. "COVID-19 Containment on a College Campus via Wastewater-

Based Epidemiology, Targeted Clinical Testing and an Intervention." *Science of The Total*

Environment, vol. 779, 2021, p. 146408., doi:10.1016/j.scitotenv.2021.146408.

Binnicker, Matthew J. "Challenges and Controversies to Testing for COVID-19." *Journal of*

Clinical Microbiology, vol. 58, no. 11, 2020, doi:10.1128/jcm.01695-20.

D'Aoust, Patrick M., et al. "Quantitative Analysis of SARS-CoV-2 RNA from Wastewater

Solids in Communities with Low COVID-19 Incidence and Prevalence." 2020,

doi:10.1101/2020.08.11.20173062.

Gandhi, Monica, et al. "Asymptomatic Transmission, the Achilles' Heel of Current Strategies to

Control Covid-19." *New England Journal of Medicine*, vol. 382, no. 22, 2020, pp. 2158–

2160., doi:10.1056/nejme2009758.

Graham, Katherine E., et al. "SARS-CoV-2 in Wastewater Settled Solids Is Associated with

COVID-19 Cases in a Large Urban Sewershed." *Environmental Science and Technology*,

2020, doi:10.1101/2020.09.14.20194472.

Hamouda, Mohamed, et al. "Wastewater Surveillance for SARS-CoV-2: Lessons Learnt from

Recent Studies to Define Future Applications." *Science of The Total Environment*, vol.

759, 2021, p. 143493., doi:10.1016/j.scitotenv.2020.143493.

Hart, Olga E., and Rolf U. Halden. "Computational Analysis of SARS-CoV-2/COVID-19

Surveillance by Wastewater-Based Epidemiology Locally and Globally: Feasibility,

Economy, Opportunities and Challenges." *Science of The Total Environment*, vol. 730,

2020, p. 138875., doi:10.1016/j.scitotenv.2020.138875.

Hata, Akihiko, et al. "Organic Substances Interfere with Reverse Transcription-Quantitative

PCR-Based Virus Detection in Water Samples." *Applied and Environmental Microbiology*,

vol. 81, no. 5, 2014, pp. 1585–1593., doi:10.1128/aem.03082-14.

Jee, Youngmee. "WHO International Health Regulations Emergency Committee for the COVID-

19 Outbreak." *Epidemiology and Health*, vol. 42, 2020, doi:10.4178/epih.e2020013.

Jones, Davey, et al. "Fecal Shedding of SARS-CoV-2 and Its Potential Role in Person-To-Person

Transmission and the Environment-Based Spread of COVID-19." 2020,

doi:10.20944/preprints202007.0471.v1.

Kitajima, Masaaki, et al. "SARS-CoV-2 in Wastewater: State of the Knowledge and Research

Needs." *Science of The Total Environment*, vol. 739, 2020, p. 139076.,

doi:10.1016/j.scitotenv.2020.139076.

Kitamura, Kouichi, et al. "Efficient Detection of SARS-CoV-2 RNA in the Solid Fraction of

Wastewater." *Science of The Total Environment*, vol. 763, 2021, p. 144587.,

doi:10.1016/j.scitotenv.2020.144587.

Lorenzo, M., & Picó, Y. (2019). Wastewater-based epidemiology: Current status and future

prospects. *Current Opinion in Environmental Science & Health*, *9*, 77–84.

https://doi.org/10.1016/j.coesh.2019.05.007

Lu, Dingnan, et al. "Primary Concentration – The Critical Step in Implementing the Wastewater

Based Epidemiology for the COVID-19 Pandemic: A Mini-Review." *Science of The Total*

Environment, vol. 747, 2020, p. 141245., doi:10.1016/j.scitotenv.2020.141245.

Mesoraca, Alvaro, et al. "Evaluation of SARS-CoV-2 Viral RNA in Fecal Samples." *Virology*

Journal, vol. 17, no. 1, 2020, doi:10.1186/s12985-020-01359-1.

Mizumoto, Kenji, et al. "Estimating the Asymptomatic Proportion of Coronavirus Disease 2019

(COVID-19) Cases on Board the Diamond Princess Cruise Ship, Yokohama, Japan, 2020."

Eurosurveillance, vol. 25, no. 10, 2020, doi:10.2807/1560-7917.es.2020.25.10.2000180.

Moghadas, Seyed M., et al. "The Implications of Silent Transmission for the Control of COVID-

19 Outbreaks." *Proceedings of the National Academy of Sciences*, vol. 117, no. 30, 2020,

pp. 17513–17515., doi:10.1073/pnas.2008373117.

Polo, David, et al. "Making Waves: Wastewater-Based Epidemiology for COVID-19 –

Approaches and Challenges for Surveillance and Prediction." *Water Research*, vol. 186,

2020, p. 116404., doi:10.1016/j.watres.2020.116404.

Pérez-Cataluña, Alba, et al. "Comparing Analytical Methods to Detect SARS-CoV-2 in

Wastewater." *Science of The Total Environment*, vol. 758, 2021, p. 143870.,

doi:10.1016/j.scitotenv.2020.143870.

Randazzo, Walter, et al. "SARS-CoV-2 RNA Titers in Wastewater Anticipated COVID-19

Occurrence in a Low Prevalence Area." *Water Research*, 2020,

doi:10.1101/2020.04.22.20075200.

Ravi, Neeraja, et al. "Diagnostics for SARS-CoV-2 Detection: A Comprehensive Review of the

FDA-EUA COVID-19 Testing Landscape." *Biosensors and Bioelectronics*, vol. 165, 2020,

p. 112454., doi:10.1016/j.bios.2020.112454.

Torii, Shotaro, et al. "Applicability of Polyethylene Glycol Precipitation Followed by Acid

Guanidinium Thiocyanate-Phenol-Chloroform Extraction for the Detection of SARS-CoV-

2 RNA from Municipal Wastewater." *Science of The Total Environment*, vol. 756, 2021, p.

143067., doi:10.1016/j.scitotenv.2020.143067.

Treibel, Thomas A, et al. "COVID-19: PCR Screening of Asymptomatic Health-Care Workers at

London Hospital." *The Lancet*, vol. 395, no. 10237, 2020, pp. 1608–1610.,

doi:10.1016/s0140-6736(20)31100-4.

Venugopal, Anila, et al. "Novel Wastewater Surveillance Strategy for Early Detection of

Coronavirus Disease 2019 Hotspots." *Current Opinion in Environmental Science &*

Health, vol. 17, 2020, pp. 8–13., doi:10.1016/j.coesh.2020.05.003.

Wu, Jia-Rui. "SARS-CoV-2 and Its Infected World." *Journal of Molecular Cell Biology*, 2021,

doi:10.1093/jmcb/mjaa078.

Xagoraraki, Irene, and Evan O'Brien. "Wastewater-Based Epidemiology for Early Detection of

Viral Outbreaks." *Women in Water Quality*, 2019, pp. 75–97., doi:10.1007/978-3-030-

17819-2_5.

Ye, Yinyin, et al. "Survivability Partitioning, and Recovery of Enveloped Viruses in Untreated

Municipal Wastewater." *Environmental Science and Technology*, 2016,

doi:10.1021/acs.est.6b00876.s001.

Zheng, Shufa, et al. "Viral Load Dynamics and Disease Severity in Patients Infected with SARS-

CoV-2 in Zhejiang Province, China, January-March 2020: Retrospective Cohort Study."

BMJ, 2020, p. m1443., doi:10.1136/bmj.m1443.

Highlights

-    Samples from smaller population densities make concentrating RNA vital for detection

-    Pelleting may provide for more timely concentration and extraction of SARS-CoV-2

RNA

-    RNA shields and PCR inhibitor removal may increase detection of RNA during RT-

qPCR

1. Sample collection in intact oral cavity with/without any specimen in intact oral plaque.

2. Sucrose/Sucrose lysis. Fill tube collected cells with 400 μL Tris 10 mM, Tris 1 mM, EDTA and glucose 50 mM added. Then 20 μL of RNase to 10 μL RNase.

3. Vortex/shake. Mix tube collected cells with solution.

4. Centrifuge. Spin tube/micro centrifuge to 10 min.

5. Pour. Discard supernatant and pour solution to collect cell/filter.

6. Decanto/ Disgard. Fill tube collected cells with standard to collect cell/filter.

Created by BioRender.com