1 **TITLE:**
2 ADESSO detects SARS-CoV-2 and its variants: extensive clinical validation of an optimised
3 CRISPR-Cas13-based COVID-19 test
4
5 **AUTHORS**
6

7 Beatrice Casati[1,2], J.P. Verdi[1,3], A. Hempelmann[3], Maximilian Kittel[4,6], Andrea Gutierrez
8 Klaebisch[5,6], Sybille Welker[5,6], Sonal Asthana[1], Pavle Boskovic[7], Ka Hou Man[7], Bernhard
9 Radlwimmer[7], C. Erec Stebbins[3], Thomas Miethke*[,5,6], F. Nina Papavasiliou*[,1], Riccardo
10 Pecori*[,1]
11
12 * These corresponding authors contributed equally.
13
14 **AFFILIATIONS**
15
[1]Division of Immune Diversity, Department of Immunology and Cancer, German Cancer
Research Centre (DKFZ), 69120 Heidelberg, Germany.
[2]Faculty of Biosciences, Heidelberg University, 69120 Heidelberg, Germany.
[3]Division of Structural Biology of Infection and Immunity, Department of Immunology and
Cancer, German Cancer Research Centre (DKFZ), 69120 Heidelberg, Germany.
[4]Institute of Clinical Chemistry, Medical Faculty of Mannheim, University of Heidelberg,
Theodor-Kutzer-Ufer 1-3, 68167 Mannheim, Germany
[5]Institute of Medical Microbiology and Hygiene, Medical Faculty of Mannheim, University of
Heidelberg, Theodor-Kutzer-Ufer 1-3, 68167 Mannheim, Germany
[6]Mannheim Institute for Innate Immunoscience (MI3), Medical Faculty of Mannheim, University
of Heidelberg, Ludolf-Krehl-Str. 13–17, 68167 Mannheim, Germany
[7]Division of Molecular Genetics, German Cancer Research Centre (DKFZ), 69120 Heidelberg,
Germany.

29
30 **IMPORTANT NOTE: This protocol should not be used for clinical purposes. Although we**
31 **have validated the protocol on patient samples, this test is not officially authorized. We**
32 **hope this protocol will provide some reference points for researchers interested in**
33 **further advancing Crispr-DX diagnostic platforms. We also welcome researchers to**
34 **contact us for any assistance.**
35
36 **ABSTRACT**
37

With the coronavirus disease 19 (COVID-19) pandemic now deep into its second year,
widespread testing for the detection of the causative severe acute respiratory syndrome
coronavirus 2 (SARS-CoV-2) is fundamental. The gold standard reverse transcription
quantitative PCR (RT-qPCR) cannot keep up with the high demand alone, therefore alternative
diagnostic tests are needed. Here we present ADESSO (Accurate Detection of Evolving SARS-
CoV-2 through SHERLOCK Optimisation), an optimised version of the CRISPR-based
SHERLOCK (Specific High-sensitivity Enzymatic Reporter unLOCKing) assay. After an

extensive validation on 983 clinical samples, we demonstrated that ADESSO has a sensitivity of
96% and a specificity of 100% on extracted RNA, comparable to RT-qPCR. Its performance on
unextracted samples still allows the detection of the more infectious 75% of the COVID-19
positive population, making it suitable for point-of-care (POC) testing. Interestingly, our in
parallel comparison of 390 matching swab and gargle samples showed consistently lower viral
loads in gargle specimens. We also validated ADESSO for the detection of the B.1.1.7 variant
and demonstrated that ADESSO is adaptable to any variant of concern in less than one week, a
critical feature now that worrisome SARS-CoV-2 variants are spreading all around the world.
53
54 **INTRODUCTION**
55
Since the beginning of the global pandemic of coronavirus disease 2019 (COVID-19),
170 million confirmed cases, including 3.5 million deaths, have been reported[1]. COVID-19 is a
severe respiratory disease caused by severe acute respiratory syndrome coronavirus 2 (SARS-
CoV-2)[2,3,4]. The quick diffusion of SARS-CoV-2 is primarily attributed to the relatively long
duration of viral shedding by infected individuals, the viral load dynamics and the lengthy
incubation period[5]. Indeed, the incubation period of SARS-CoV-2 is estimated to be 5-6 days[6,7]
with a high viral load upon the onset of symptoms[8–10], suggesting that individuals with COVID-
19 begin viral shedding a few days before symptoms appear[11]. Further, a significant proportion
of infected individuals either remain entirely asymptomatic or only manifest mild symptoms[12–14].
Since these carriers are still able to transmit the virus, case identification and contact tracing
protocols alone remain inefficient[11,13,15–18], thus facilitating uncontrolled spread of the virus and
leading to the current pandemic situation.
The urgent need for a vaccine has accelerated the development of multiple effective
vaccines and more than 1.5 billion doses have already been administered[1,19–22]. However, even
in the most optimistic scenario, it will take some time before we will be able to reap the benefit of
a global vaccination campaign[23,24]. Therefore, complementary efforts to limit the spread of the
virus are still essential.
To mitigate viral spreading, many countries adopted extreme social distancing
measures, including strict lockdowns[25]. However, the socio-economic costs for such measures
are enormous and the consequences will be long lasting[26–28]. Therefore, the highest priority has
to be given to the development of strategies aimed at ensuring long-term safety through
containment and isolation of SARS-CoV-2 positive individuals and, at the same time, allowing a
safe restart of businesses and social life[29]. A recent model of viral dynamics suggests that
frequent testing for the identification of viral infections and the isolation of carriers is essential[30].
Notably, the model indicates that effective screening depends mainly on the frequency of testing
and speed of reporting, while only to some extent on test sensitivity[30].
The worldwide gold standard diagnostic test for SARS-CoV-2 infection is the reverse
transcription quantitative polymerase chain reaction (RT-qPCR). Viral RNA is isolated from
nasal swabs, throat swabs or saliva, retro-transcribed into cDNA and specific regions of the viral
genome are amplified via quantitative PCR. Multiple primer sets are utilised, allowing for the
amplification of different targets with a LoD of 10^3 viral RNA cp/ml[31].
An important limitation of RT-qPCR is the requirement for specific equipment, laboratory
infrastructures and qualified personnel. Inadequate access to such resources significantly

reduces the frequency of testing. Additionally, PCR testing facilities often require days' worth of
time to report the test outcome, resulting in a long sample-to-result turnaround time. To face
these challenges, different rapid tests have been implemented, such as rapid PCR and antigen-
based tests. However, since rapid PCR tests still require specific equipment and antigen-based
tests have lower sensitivity and specificity[32], there is a need for an alternative test that is
comparable to RT-qPCR in terms of sensitivity and specificity, but faster and independent of
complex instruments.
All these requirements are met by the so-called CRISPR diagnostic (CRISPR Dx)
technologies, which comprise multiple tools for rapid, economical, sensitive and specific nucleic
acid detection[33]. The CRISPR system is a bacterial machinery able to recognise and cleave
foreign genetic material. Among the CRISPR associated (Cas) proteins, Cas13 and Cas12 are
able to specifically bind RNA and DNA molecules, respectively, complementary to the target-
binding CRISPR RNA (crRNA). Upon target recognition, the Cas proteins cleave a reporter in
trans, which can then be detected via different readouts[34–36]. These readouts are limited by the
amount of detectable target material in the sample of interest. To circumvent this limitation,
isothermal amplification methods that do not rely on sophisticated equipment, such as loop-
mediated isothermal amplification (LAMP)[37] or recombinase polymerase amplification (RPA)[38]
have been combined with Cas-mediated nucleic acid detection[35,39,40]. CRISPR Dx technologies
were quickly adapted to offer point-of-care (POC) diagnostic tests for the detection of SARS-
CoV-2. In about an hour, test results can be read on paper-based lateral flow sticks or by
fluorescence detection with portable devices[41–48]. Despite the high potential of CRISPR Dx
technologies, only two of them have received emergency use authorisation from the Food and
Drug Administration (FDA), with use restricted to the authorised laboratories[49,50]. The analysis of
their performance on clinical samples is still not adequate enough to grant them approval for
use in routine diagnostics, therefore a more extensive study in comparison with RT-qPCR is
necessary.
Here we have optimised the Cas13a-based detection platform named "SHERLOCK"[39,40]
and developed ADESSO (Accurate Detection of Evolving SARS-CoV-2 through SHERLOCK
Optimisation) for highly sensitive detection of SARS-CoV-2 directly from patient-derived
material. The entire protocol takes one hour, does not require RNA extraction or any specific
equipment, is able to detect down to 2.5 cp/µl of SARS-CoV-2 synthetic genome and is low-cost
(less than 5€ per test). Throughout our work we extensively evaluated the real diagnostic
potential of ADESSO in direct comparison to RT-qPCR and with two different sample collection
methods (nasopharyngeal swab (NP) vs gargle of saline). Importantly, all the samples were
collected from ambulatory patients presenting minimal or mild symptoms or from people
identified as contacts of SARS-CoV-2 infected individuals, representing a part of the population
that can potentially remain undetected. Our study showed that ADESSO has a sensitivity
comparable to RT-qPCR when applied to purified RNA, while it resulted in a lower sensitivity
when performed directly on unextracted samples, yet still being more sensitive than rapid
antigen tests[32]. However, considering that the Ct values across our cohort follow a normal
distribution, we can fairly estimate that ~75% of the entire infected population would be
successfully detected by ADESSO on unextracted swabs. Importantly, the 25% portion of
infected individuals that ADESSO would miss corresponds to high RT-qPCR Ct values,
coinciding with low viral titers and minimal infectiousness[11,51]. Additionally, we also observed a

drop in sensitivity when gargling was used as sampling method for both ADESSO and RT-
qPCR. Finally, in less than one week we adapted ADESSO to specifically identify SARS-CoV-2
variants in clinical samples by modifying the primers and crRNAs used for amplification and
detection. The adapted ADESSO can identify the presence of a variant within one hour of
sample submission, thus eliminating the need for sequencing, while RT-qPCR tests would
require an additional day or two on top of the already slower turnaround. Considering the risk
posed by the spread of the new SARS-CoV-2 variants, this feature is highly relevant in the
current phase of the COVID-19 pandemic and in the near future[52].
141
142
143   **RESULTS**
144
A SHERLOCK-based assay for SARS-CoV-2 detection in clinical samples
The need for a rapid and sensitive COVID-19 POC test has been and will remain a significant
factor to contain the spread of SARS-CoV-2. CRISPR Dx technologies represent a viable option
for the development of such a test[33]. For this reason, we first aimed to reproduce and adapt a
Cas13-based molecular detection platform called SHERLOCK (Specific High Sensitivity
Enzymatic Reporter UnLOCKing)[39] for the detection of SARS-CoV-2 in clinical samples.
SHERLOCK is based on two main steps: (a) isothermal amplification of viral RNA via RT-RPA
and (b) detection of a specific RNA sequence by Cas13 protein followed by *trans*-cleavage of a
labeled reporter, which can be detected via a lateral flow-based visual readout[39] (**Figure 1A**).
During the RT-RPA, RNA is retro-transcribed and amplified to dsDNA using specific primers[38]
and a T7 promoter is added to the amplicon by including its sequence in the forward primer.
This feature is necessary for the Cas13 detection step, where a T7 RNA polymerase transcribes
the amplified dsDNA back into RNA, which can be specifically recognised by Cas13 protein in
complex with a crRNA complementary to the target. The specific binding between the Cas13-
crRNA complex and its target RNA activates Cas13 collateral activity, leading to cleavage of an
RNA reporter and generation of a detectable signal. Finally, the resulting signal can be read on
a lateral flow-based visual readout. For this readout an RNA reporter flanked by biotin and
fluorescein (FAM) is used in combination with anti-FAM antibodies labelled gold nanoparticles
used to visualise the reporter. In a negative sample, the reporter is intact and is captured by a
line of streptavidin resulting in a first band called "control band". In a positive sample, the
reporter is cut, therefore the first half of the reporter containing biotin is captured by streptavidin,
while the other half containing FAM is captured by a second line of antibodies resulting in the
appearance of the so-called "test band". The band intensity ratio between the test band and the
control band indicates the portion of reporter that has been cut, which reflects the level of Cas13
activation and thus the amount of target RNA that was detected in the sample (**Figure 1A**).
We first generated SARS-CoV-2-specific guide sequences, purified LwaCas13a protein[53]
(**Figure S1A,B**) and tested the system's detection sensitivity in the absence of a pre-
amplification step using serial dilutions of an *in-vitro*-transcribed (IVT) fragment of the SARS-
CoV-2 S gene. As previously published, a sensitivity of 10^8 aM was observed[39] (**Figure S1C**).
We then assessed the sensitivity of our test when combining the Cas13 detection with an RT-
RPA pre-amplification step on SARS-CoV-2 genes S and Orf1a, as formerly described[54]. A 2-
fold improvement of the previously published sensitivity was obtained[54]. Indeed, we detected

SARS-CoV-2 gene S at a concentration of 10aM (5 cp/µl) and gene Orf1a at a concentration of
100aM (50 cp/µl) (**Figure 1B**). This improvement is most likely due to the replacement of
ProtoScript II Reverse Transcriptase (RT) with M-MuLV RT, which retains a functional RNase H
domain that degrades DNA:RNA hybrid intermediates and thereby improves the efficiency of
RT[55] (**Figure S1D**). We then used the set of primers and crRNA for S to conduct a blind test on
30 clinical samples. These samples were NP swabs collected in saline (0.9% NaCl) and they
were previously tested for SARS-CoV-2 via RT-qPCR (Roche Cobas System) at the Medical
University Hospital Mannheim. The specimens were frozen and transported to our laboratory,
where we extracted RNA and performed SHERLOCK in duplicates. Additionally, the CDC 2019-
nCoV Real-Time RT-PCR Diagnostic Panel[56] was also performed as a control. Using
SHERLOCK, we were able to identify all 10 positive samples (**Figure 1C**, **Table S1**). Notably,
we detected sample 28, which had a very low viral titer (corresponding to a high Ct value).
These results demonstrate that SHERLOCK can be used as an alternative method to detect
SARS-CoV-2 in RNA extracted from clinical samples.
191
192
SARS-CoV-2 direct detection from clinical samples
RNA extraction is a time-consuming, labor-intensive and costly step for COVID-19 diagnosis
(**Figure 1A**) and shortage of RNA extraction kits has been a global issue throughout the
pandemic[57,58]. Different studies have already demonstrated that it is possible to omit it[41,43,59,60].
Therefore, after demonstrating the high potential of SHERLOCK as a diagnostic test for COVID-
19, we attempted to improve our protocol in order to avoid the RNA extraction step, thus making
the test faster and cheaper (**Figure 2A**). First, we compared different lysis methods by treating
one known positive sample (sample #30 in **Figure 1C** and **Supplementary File 1**) with either
QuickExtract DNA Extraction solution (Lucigen, #QE09050), Luna Cell Ready Lysis Buffer
(NEB, #E3032) or 5% Triton X 100 (Carl Roth, #3051.3) and incubating it for 5 min at 95°C. We
performed the experiment in triplicates for each lysis method and we were able to successfully
detect SARS-CoV-2 directly after lysing the sample with QuickExtract DNA Extraction solution
and Luna Cell Ready Lysis Buffer (**Figure 2B**). Lysis with 5% Triton X 100 did not allow SARS-
CoV-2 detection, although it was successfully performed in another study[59]. To improve the
sensitivity of our test, we first optimised the amount of RT units and sample input in the RT-RPA
step using dilutions of synthetic SARS-CoV-2 genome. We observed the best results with 6 U/µl
of RT and 2.5 µl of sample input per reaction (**Figure S2A**). Additionally, we compared our
standard set of RPA primers and Cas13 crRNA for S with other sets designed to target different
genes of SARS-CoV-2, namely N and Orf1a. First, we assessed their performance on serial
dilutions of a positive sample (sample #6 in **Figure 1C** and **Supplementary File 1**) (**Figure
S2B**). Then, we compared the sensitivity of the best two candidates, S and Orf1a, on dilutions of
synthetic SARS-CoV-2 genome[61]. The set of RPA primers and Cas13 crRNA for S remained the
most sensitive one (**Figure S2C**). We therefore selected it for all the following experiments. In
order to assess the sensitivity of our test on unextracted samples, we spiked in serial dilutions of
SARS-CoV-2 synthetic genome in a negative sample lysed with QuickExtract solution and we
performed SHERLOCK on the S gene. We were able to consistently detect 10 cp/µl (**Figure
2C**). We then proceeded with the evaluation of the diagnostic potential of SHERLOCK on
unextracted samples (so-called direct SHERLOCK) by performing a second blind test on 160

clinical samples. Positive samples were considered those which resulted in a band intensity
ratio (test band/control band) higher than 0.2. This threshold was defined based on the band
intensity ratio obtained in all the negative controls used in this study and the negative samples
analysed in Figures 2 and 4 (n = 282 + 467; **Figure S3A**). Direct SHERLOCK was able to
identify 73 out of 93 positive samples resulting in a sensitivity of 78% (**Figure 2D,E** and **Table**
S1). Importantly, despite an apparent LoD equivalent to Ct 27, samples with lower Ct values
and high viral loads resulted in a highly variable band intensity ratio with some being very close
to the 0.2 threshold (**Figure 2D**). For this reason, we decided to proceed with a step-by-step
optimisation of the direct SHERLOCK protocol.
230
231
ADESSO: an optimised and highly sensitive SHERLOCK assay
Considering the results of the detection of SARS-CoV-2 directly from clinical samples (**Figure**
2), we aimed at optimising the SHERLOCK protocol to develop a direct SARS-CoV-2 diagnostic
test that is as sensitive as possible. Therefore, we evaluated alternative reagents and different
reaction conditions for each one of the three main steps in SHERLOCK, namely, 1) sample
lysis, 2) RT-RPA and 3) Cas13 detection (**Figure S4A**), to increase both sensitivity and speed
of the test. At this stage, we assessed Cas13 activation also via a fluorometer to monitor the
speed of the reaction in real time. The fluorescence readout is based on the use of an RNA
reporter flanked by a fluorophore and a quencher. Upon Cas13-mediated cleavage of the
reporter, the fluorophore is cut from the quencher and its fluorescent signal can be detected by
a fluorometer (**Figure S4A**). First, we measured the RNAse activity in a swab sample collected
in saline and lysed with the method selected in Figure 2 (QuickExtract DNA Extraction solution
and incubation at 95°C for 5 min). To evaluate RNAse activity, RNaseAlert was added to the
sample following lysis and fluorescence was measured to evaluate the corresponding nuclease
activity. Notably, addition of RNase inhibitors in the lysis buffer prior to heating was sufficient to
inhibit RNAse activity almost completely (**Figure 3A**). Next, we optimised the RT-RPA step by
first comparing different RT enzymes in the presence or absence of RNase H. Once again, M-
MuLV shows the best sensitivity (5-2.5 cp/µl) in comparison to ProtoScript II or SuperScript III,
while the addition of RNase H leads to an improvement for SuperScript III only (**Figure 3B**).
Secondly, we used different final concentrations of RPA, where 1xRPA corresponds to the
standard amount of RPA described in the original SHERLOCK protocol[39,53,54] and 5xRPA
corresponds to the optimal amount according to the manufacturer's instructions. To test this, we
selected one false negative sample from our previous blind test on unextracted samples
(sample #L151, **Supplementary File 1**) and a true negative sample as negative control (sample
#L126, **Supplementary File 1**) and we repeated our assay with different concentrations of RPA.
Remarkably, while the false negative sample is still negative with 1xRPA, it results positive for
final concentrations of RPA from 2x to 5x, with a decrease in band intensity ratio when using the
4xRPA and 5xRPA concentrations (**Figure 3C, S4B**). Considering this and bearing in mind the
cost per single test, we decided to proceed using a 2xRPA concentration. To further confirm this
improvement, we compared 1xRPA and 2xRPA on 5 samples with Ct values close to the LoD
determined in Figure 2 (samples #L95, L96, L111, L122 and L123, **Supplementary File 1**). We
observed an improvement in the 2xRPA reactions with these samples as well (**Figure 3D, S4C**).
Furthermore, in order to optimise the Cas13 detection step we made a ten-fold dilution of a

positive RT-RPA reaction (50 cp/μl) and we performed Cas13 detection using the original
concentration of Cas13/crRNA (45/22.5 nM)[39,53,54], in comparison to higher amounts (**Figure**
S4D, upper panel). A concentration of Cas13/crRNA of 90 nM each leads to an improved
reaction, reaching the plateau after 15 min only, compared to 30 min for the other two
concentrations (**Figure 3E, S4D**, lower panel**).** We also confirmed that a 10-min incubation for
Cas13 detection performed in half the volume is sufficient to yield a clearly positive outcome in
the lateral flow detection assay (**Figure 3F, S4E**), which is an essential aspect for a POC test.
Moreover, a shorter Cas13 reaction allows us to extend the incubation time of the RT-RPA step
for highly sensitive reactions[53] without affecting the total time of the assay. Finally, we assessed
the sensitivity of this optimised protocol on serial dilutions of SARS-CoV-2 synthetic genome
and we observed a robustly reproducible sensitivity of 2.5 cp/μl (**Figure 3G**). We named this
new optimised diagnostic assay ADESSO (Accurate Detection of Evolving SARS-CoV-2
through SHERLOCK Optimisation) (**Figure 3H**).
278
279
Evaluation of ADESSO performance on clinical samples in direct comparison to RT-
qPCR.
We used ADESSO to test a total of 195 clinical samples in direct comparison to the RT-qPCR
protocol routinely used in the clinics. To allow a fair comparison between the methods, we first
selected 95 positive and 100 negative individuals (via COBAS RT-qPCR on NP swab). For each
of these specimens, RNA was re-extracted and analysed by RT-qPCR (Tib Molbiol) and
ADESSO. Additionally, ADESSO was also performed directly on unextracted samples. Finally,
we also obtained gargled saline from the same individuals as an alternative sampling method,
which would be ideal for POC testing. Those samples were treated exactly as the NP swabs
(**Figure 4A**). Importantly, the ADESSO results were analysed without knowing the outcome of
the COBAS RT-qPCR used as the reference.
The results of this experiment are summarised in **Table 1**. ADESSO on RNA extracted from
swabs was able to correctly identify most positive samples (91 out of 95), resulting in a
sensitivity of 96% (**Figure 4B**). Interestingly, all the false negative samples have Ct values
higher than 31, corresponding to lower viral loads (<100cp/μl) and therefore a lower probability
of spreading the virus[11,51]. RT-qPCR (Tib Molbiol) performed on the same samples was largely
in agreement with the COBAS RT-qPCR, with highly correlated Ct values (**Figure 4F**).
However, using this method we were able to identify 89 out of 95 positive samples resulting in a
sensitivity of 94% (**Table 1** and **Figure 4F**). As expected, ADESSO on unextracted samples
resulted in a lower sensitivity (77%), with all false negative samples having Ct values higher
than 29 (~100cp/μl) (**Table 1** and **Figure 4C**). The same analysis was performed on gargle
saline samples. In this case, ADESSO on extracted RNA correctly identified 74 out of 95
positive samples resulting in a sensitivity of 78%, with most false negative samples having Ct
values higher than 30 and few with Ct values between 28 and 29 (**Figure 4D**). Interestingly, this
drop in sensitivity does not seem to be related to the detection method but rather to the
sampling method. Indeed, the same decrease in sensitivity (to 79%) was observed also for RT-
qPCR (TibMolBio), with true positive samples resulting in higher Ct values (**Table 1** and **Figure**
4G). Finally, as observed for swabs, ADESSO on unextracted gargle saline samples resulted in
a lower sensitivity (65%) (**Figure 4E** and **Table 1**). In this latter case, false negative samples

have different Ct values, with some corresponding to high viral loads in swabs (analysed by
COBAS RT-qPCR) (dark red dots in **Figure 4E** and **4H**). The overall decrease in sensitivity is in
agreement with the consistently higher Ct values observed in gargle specimens compared to
their matched swab samples analysed via Tib Molbiol RT-qPCR (**Figure 4H**).
Altogether, these results validate the high potential of ADESSO as a POC test for the detection
of SARS-CoV-2 infected individuals. Notably, ADESSO on extracted RNA, either from swab or
gargle samples, performed as well as RT-qPCR (Tib Molbiol) in terms of sensitivity and
specificity. Additionally, our data also shows an important difference in the detection of SARS-
CoV-2 when gargling with saline was used as a sampling method. Even though this approach
would be better suited to a POC test, we observed a clear reduction in sensitivity, independently
of the detection method used.
320
Adaptation of ADESSO for detection of SARS-CoV-2 variants: a flexible and powerful
assay to rapidly identify specific variants or mutations.
Since the beginning of the pandemic, SARS-CoV-2 has evolved considerably. The first variants
to appear carried a D614G mutation in the spike protein[62], which is now dominant and shared
between all the existing variants. While several variants exist, here we focus our attention on
two variants of concern: SARS-CoV-2 B.1.1.7 (UK variant) and SARS-CoV-2 B.1.351 (South
Africa (SA) variant). SARS-CoV-2 B.1.1.7, also known as 501Y.V1, seems to have an enhanced
transmissibility[63] and might be more virulent[64]. It was first detected in England in late 2020 and,
after becoming the dominant variant in the UK, it has spread quickly all over Europe and
worldwide. B.1.1.7 contains eight mutations in the spike gene in addition to the mutation causing
the D614G substitution, including deletions (e.g., ΔHV69-70) (**Figure 5A**). SARS-CoV-2
B.1.351, also known as 501Y.V2, was first detected in late 2020 in Eastern Cape, South
Africa[65]. This variant quickly became dominant locally and displaced other viral lineages in
several regions, possibly as a result of increased transmissibility or immune escape[65,66]. B.1.351
contains nine mutations in the spike gene in addition to the mutation causing the D614G
substitution, including clusters of mutations (e.g., mutations leading to Δ242-244 and R246I)
(**Figure 5A**). Finally, there is growing evidence that these new variants could impair the efficacy
of current monoclonal antibody therapies and vaccines because of the several mutations
located in the spike gene[67–69]. For this reason, it is now essential to quickly identify individuals
infected by SARS-CoV-2 variants. The UK variant is the major concern in Europe and Germany,
therefore we adapted our test to detect the deletion (ΔHV69-70) specific to this strain. We called
this adapted test ADESSO-UK (**Figure 5A,B**). First, we designed two different crRNAs able to
recognise either the original Wuhan strain or the UK variant, called respectively crRNA HV69-70
and crRNA ΔHV69-70 (**Figure 5B**). Then we optimised RT-RPA primers to amplify the region of
SARS-CoV-2 genome containing HV69-70 and we selected the more sensitive set 1 for further
analysis (**Figure S5A**). Finally, we performed a blind test on positive clinical samples carrying
either UK or SA variants. We first applied ADESSO for the detection of SARS-CoV-2 and we
were able to detect all positive samples but one (sample #11, **Figure 5C**). Then, using
ADESSO-UK (crRNA ΔHV69-70 or HV69-70) we were able to correctly identify all the samples
carrying the UK variant and we could distinguish the ones bearing the SA strain (samples #1-13,
Figure 5D,E and **S5B**). Interestingly, among the three samples carrying the SA variant, only
sample #11 was not detected via ADESSO. Sequencing analysis of the viral genome in these

three samples showed that they all shared the deletion Δ242-244, but only sample #11 carried
the R246I mutation (**Figure S5C**). This mutation falls exactly within the bases recognised by the
3' end of the forward primer used in the RT-RPA step of ADESSO, thus disrupting its function
(**Figure 5F**). Notably, the assay seems to be resistant to deletions of several nucleotides
occurring in sequences that are complementary to the central region of the primer (**Figure 5F**).
Altogether, these results show how ADESSO can be readily adapted for the detection of SARS-
CoV-2 variants of concern and even specific mutations. The entire adaptation of the test took
less than one week, from the selection of a unique mutation for the UK variant to the validation
of the adapted protocol, including designing and production of the specific reagents. This
feature of our assay is a crucial aspect in the current phase of the COVID-19 pandemic, where
quick and sequencing-independent detection of variants is essential to contain their spread[52].
364
365
366   **DISCUSSION**
367
The COVID-19 pandemic has been afflicting the world for more than a year now and the
number of new weekly global cases is still hitting its highest levels, despite multiple effective
vaccines being distributed[1]. Therefore, promptly tracking infected individuals to isolate them and
prevent further spread of the virus is fundamental. The gold-standard RT-qPCR-based COVID-
19 diagnostic test alone cannot keep up with the high demand for testing and the long
turnaround time is an issue when a fast response is essential. Rapid PCR and antigen-based
tests are also available, but there are some limitations for their widespread use, such as the
requirement for sophisticated PCR equipment and the standard practice of confirming positive
antigen-based test results by RT-qPCR. Therefore, an alternative test that overcomes these
limitations is still needed. Here, we have optimised the Cas13a-based diagnostic platform called
SHERLOCK[39] and developed the improved protocol ADESSO for highly sensitive COVID-19
testing. Overall, we tested 983 samples (496 positive and 487 negative, **Supplementary File
1**), in parallel comparison with RT-qPCR. To our knowledge, it is the first time that such an
extensive study on clinical samples has been reported for CRISPR Dx technologies. ADESSO
has a sensitivity of 96% on RNA extracted from swabs and a sensitivity of 77% when performed
directly on unextracted swab samples. This drop in sensitivity is due to a decreased LoD at Ct
29, corresponding to low viral titers and minimal infectiousness[11,51]. However, skipping the RNA
extraction step considerably reduces the sample-to-result turnaround time and allows more
frequent testing, which is suggested to be essential for efficient identification of viral infections
and isolation of carriers to contain the pandemic[30]. Other advantages are the lower need for
RNA extraction kits, whose shortage has been a global issue throughout the pandemic[57,58], and
the higher test portability, which makes ADESSO appropriate for POC testing. To further
increase the POC suitability of ADESSO, we assessed its performance in comparison with RT-
qPCR on gargle samples obtained from the same individuals from whom swabs were collected.
Interestingly, we observed a loss in sensitivity independently of the diagnostic test. On RNA
extracted from gargle samples, ADESSO showed a sensitivity of 78%, comparable with a 79%
sensitivity of the RT-qPCR. In line with what we observed for swab samples, the sensitivity of
ADESSO decreased to 65% when performed directly on unextracted gargle samples (**Table 1**).
This overall loss of sensitivity can be attributed to the sampling method, which leads to a

general increase in Ct values in gargle samples compared to swabs collected from the same
individuals at the same time (**Figure 4H**). To our knowledge, this is the first study that compares
two different sampling methods in parallel on such a big cohort of patients (n = 195 swabs + 195
gargle samples), thus highlighting a consistent difference in detected viral titers depending on
the sampling method used. Despite the fact that multiple studies have shown that high SARS-
CoV-2 titers can be detected in saliva[70,71], our results show that using gargle samples instead of
NP swabs, even if more suitable for POC testing, leads to higher rates of false negative
samples, independently of the sensitivity of the downstream diagnostic test. Therefore,
alternative methods should be considered and evaluated in comparison to gargling and NP
swabbing, for example self-collection of nasal swabs, whose feasibility and reliability are already
being investigated[72].
Importantly, the cohort of positive individuals analysed in our study was selected randomly and
displays a normal distribution of Ct values, covering the full range of viral titers between Ct 17
and Ct 38 (**Figure S6A,B**). This aspect is fundamental for two reasons: first, since the samples
analysed in this study were collected from ambulatory patients presenting minimal or mild
symptoms or from people identified as contacts of SARS-CoV-2 infected individuals, it highlights
the fact that these individuals can manifest high viral loads and therefore be infectious;
moreover, it allows the inference of the test performance from the experimental cohort to the
entire SARS-CoV-2 infected population. In this way, we could confidently estimate what portion
of the population our test would detect. Mathematical models show that successful identification
and isolation of 50% of infected individuals (and tracing of their contacts) is already sufficient to
flatten the infection curve[73]. Our test exceeds this fraction in all conditions (**Table 1**), strongly
suggesting that immediate and widespread application of ADESSO would be of great help to
contain the pandemic. In particular, by applying ADESSO on swab samples without RNA
extraction, an estimated ~75% of the infected population would be successfully detected, while
~25% of the infected individuals could be missed (**Figure 4** and **S6A**). Importantly, this 25%
portion corresponds to individuals with Ct values higher than 29, associated with low viral titers
and minimum infectiousness[11,51]. Finally, our results show a disagreement between LoD on
serial dilutions of synthetic viral genome and LoD in clinical samples. Despite the "synthetic"
LoD of 2.5 cp/µl (~Ct 35; **Figure 3**), the real clinical sensitivity of ADESSO corresponds to Ct
29-31 (**Figure 4**) and the same is true for other studies although it has never been
explicated[43,46]. This aspect highlights that an extensive validation on real clinical samples
covering the full range of viral titers and following a normal distribution, as the one presented
here, is necessary to determine the real LoD of a diagnostic test. This is crucial to allow a fair
comparison between sensitivities resulting from independent studies, which can be greatly
influenced by the choice of the tested population.
The importance of testing is further highlighted by the recent emergence of SARS-CoV-2
variants, which poses a new threat for humanity, as India's recent tragic crisis has shown[52,74].
This is of utmost criticality because mutations in the viral genome might impair both molecular
and antigen-based tests, thus leading to false negative results. In these situations, being able to
promptly adapt a test is fundamental and ADESSO offers such an advantage. Here, in less than
a week we adapted the test for the detection of the B.1.1.7 variant. Based on the publicly
available SARS-CoV-2 sequences (https://www.gisaid.org), ADESSO can be adapted to any

variant of concern, thus providing an all-in-one SARS-CoV-2 detection and variant identification
tool without need for sequencing.
Finally, we calculated a cost per reaction of 2.64€ and 4.82€ for fluorometric and lateral-flow
detection, respectively (**Supplementary File 2**), which would be even lower at a production
scale. Considering the cost range for a single RT-qPCR reaction[75], three to eight tests could be
performed with ADESSO for the same price. Moreover, the cost of a thermal cycler needed to
perform RT-qPCR would also be eliminated. Lastly, the use of ADESSO for the detection of
variants would even cut the cost of sequencing. Altogether, ADESSO is cheaper than any RT-
qPCR-based COVID-19 diagnostic test and offers a more accessible option for widespread and
more frequent testing.
With the COVID-19 pandemic now deep into its second year, it has become clear that time
plays a critical role in the management of such an emergency. In order to control it, rapid
detection of new infections, quick tracing of contacts, fast vaccine distribution and prompt
reaction to emerging new variants are key-factors. Slowly but undeniably, the race against
SARS-CoV-2 has turned from a sprint into a marathon. If we want to keep up, we need to take
action faster than the virus evolves. The time is now for ADESSO to join the race.
456
457

Sampling method	Sample	Test	Test result	number			number/total number (percentage)			
				Positive samples (N=95)	Negative samples (N=100)	Total samples (N=195)	Positive predictive value	Negative predictive value	Sensitivity	Specificity
SWAB	RNA	Tib Molbiol RT-qPCR	Positive	89	0	89	89/89 (100%)		89/95 (94%)	
			Negative	6	100	106		100/106 (94%)		100/100 (100%)
		ADESSO	Positive	91	0	91	91/91 (100%)		91/95 (96%)	
			Negative	4	100	104		100/104 (96%)		100/100 (100%)
	Unextracted sample	ADESSO	Positive	73	0	73	73/73 (100%)		73/95 (77%)	
			Negative	22	100	122		100/122 (82%)		100/100 (100%)
GARGLE SAMPLE	RNA	Tib Molbiol RT-qPCR	Positive	75	0	75	75/75 (100%)		75/95 (79%)	
			Negative	20	100	120		100/120 (83%)		100/100 (100%)
		ADESSO	Positive	74	0	74	74/74 (100%)		74/95 (78%)	
			Negative	21	100	121		100/121 (83%)		100/100 (100%)
	Unextracted sample	ADESSO	Positive	62	0	62	62/62 (100%)		62/95 (65%)	
			Negative	33	100	133		100/133 (75%)		100/100 (100%)

459 **Table 1:** Positive and negative predictive values, sensitivity and specificity of ADESSO on swab
460 and gargle samples with and without RNA extraction.
461
462
463 **METHODS**
464
465 **Protocols.** The RT-RPA and Cas13 reaction protocols used for each experiment are provided
466 in **Supplementary File 4** with reference to the corresponding figures. The exact volumes are
467 given for one single reaction.
468
469 **Reagents and materials.** Detailed information about reagents and material used in this study is
470 provided in **Supplementary File 3**.
471

Cas13 purification:
Plasmid encoding LwaCas13 (pC013 - Twinstrep-SUMO-huLwCas13a was a gift from Feng Zhang (Addgene plasmid # 90097; http://n2t.net/addgene:90097; RRID:Addgene_90097)[39] was transformed into Rosetta cells and purified according to established protocols with substantial modification. Single colonies were inoculated into 25 ml Terrific Broth (TB) (100 µg/ml AMP) and grown to an OD of 0.6 at 37°C degrees while shaking at 150 rpm. The suspension was chilled for 30 min at 4°C and subsequently induced with 0.5 mM IPTG and left shaking for an additional 16h at 21°C. Cells were harvested by centrifugation at 5 k rpm for 15 min at 4°C. The pellet was resuspended in 4x (wt/vol) supplemented lysis buffer (12 cOmplete Ultra EDTA-free tablets, 600 mg of lysozyme and 6 µl of benzoase to lysis buffer (20 mM Tris pH 8.0, 500 mM NaCl, 1 mM DTT)) and lysed by sonication. Lysate was cleared by centrifugation at 10 k rpm for 1h at 4°C. Supernatant was purified using a 1 ml HIS-Trap column (Cytiva) slurry and affinity chromatography was performed using the ÄKTA pure system with lysis buffer for washing steps and an imidazole gradient for elution. After initial purification, the protein sample was incubated with SUMO protease (ThermoScientific) as per the manufacturer's instructions at 4°C overnight to remove the affinity tags. The sample was then re-applied to a 1 ml HIS-Trap column. Both the SUMO protease (which itself has a 6xHIS tag) and the cleaved affinity tag bind to the resin, while pure Cas13 eluted in the wash step. A final size-exclusion chromatography step was performed using the ÄKTA pure system using 10 mM HEPES pH 7.0, 5 mM MgCl2, 1 M NaCl and 2 mM DTT as gel filtration buffer on a Superdex 16/600 column.

Synthetic SARS-CoV-2 RNA
Fully synthetic SARS-CoV-2 RNA was purchased from Twist Biosciences (MT007544.1 or MN908947.3). In order to test SHERLOCK sensitivity, serial dilutions were prepared in water or in saline, from the initial concentration of 10^6 cp/µl to 0.01 cp/µl.

Synthetic SARS-CoV-2 S gene and Orf1ab gene RNA fragments
SARS-CoV-2 RNA, a kind gift of Prof. Bartenschlager (DKFZ, Heidelberg), was used for OneStep RT-PCR (Qiagen, #210212) as follows: 11 µl of nuclease-free water, 5 µl of 5x OneStep RT-PCR buffer, 1 µl of dNTP mix (10mM each), 1.5 µl of each primer (forward and reverse, both 10µM) and 1 µl of OneStep RT-PCR Enzyme Mix were added to 4 µl of denatured RNA. The primers used for the amplification of SARS-CoV-2 S gene and Orf1a gene are listed in **Supplementary File 3**. The RT-PCR protocol was run as follows: retrotranscription at 50°C for 30 min, denaturation at 95°C for 15 min, followed by 40 cycles of denaturation at 94°C for 30 sec, annealing at 61°C (Orf1a gene) or 62°C (S gene) for 30 sec and elongation at 72°C for 5 sec. In the end a final elongation step at 72°C was run for 10 min. PCR clean-up was performed on the RT-PCR products according to the manufacturer's instructions (Macherey-Nagel, #740609.250). The purified DNA was in-vitro-transcribed into RNA with the HiScribe T7 Quick High Yield RNA Synthesis Kit (NEB, #E2050S) following the suggested protocol for short transcripts. The IVT products were then treated with DNase I (HiScribe T7 Quick High Yield RNA Synthesis Kit, NEB, #E2050S) and purified with Monarch RNA Cleanup Kit (NEB, #T2050). The concentration of the purified products was determined by Nanodrop and Qubit. In order to test SHERLOCK sensitivity, serial dilutions were made in water from a concentration of 1µM to 1aM.

Human clinical specimen collection

Clinical specimens were collected at the Medical University Mannheim, Germany. NP swabs and gargle samples were collected from ambulatory patients presenting minimal to mild symptoms or sent by the German Health Department after having contact with a SARS-CoV-2 positive person. After verbal and visual instruction gargling was performed with 8 ml of sterile 0,9% saline (Fa. Fresenius Kabi, Bad Homburg, Germany). Samples were collected in sterile containers without additives and stored at 4°C until testing with PCR within 36 h. NP specimens were collected with flocked swabs (Improswab, Fa. Improve Medical Instruments, Guanzhou/China) and washed out with 2 ml 0,9% saline within 12 h of collection. For sample inclusion in the validation study and side-by-side comparison of ADESSO and RT-qPCR, initial PCR was performed on NP swab samples as part of routine clinical care using the cobas 6800 system (Roche, Penzberg, Germany) according to the manufacturer's instructions. Based on the results of the initial PCR, 95 positive and 100 negative samples were selected.

RNA extraction

For the first blind test (Figure 1), RNA was extracted from the clinical samples with the QIAamp® Viral RNA Mini kit (Qiagen, #52904) following the manufacturer's instructions (140µl of swab were extracted and eluted in 60µl). For the validation study (Figure 4), RNA was extracted from 200 µl of the selected gargle and NP specimens with the MagnaPure Compact System (Roche, Penzberg, Germany) using the Nucleic Acid isolation Kit I (Roche) resulting in 100 µl of eluate. Residual volume of gargle and NP specimens was stored at 4°C and sent to the DKFZ for further analysis.

RT-qPCR

CDC taqman RT-qPCR initially (Figure 1) was performed in technical triplicates according to published protocols[76], which we adapted to a 384-well plate format and a reduced reaction volume of 12.5 µl. The reaction was performed using the Superscript III One-Step RT-PCR kit with Platinum Taq Polymerase. Magnesium sulphate and BSA were added to the reaction to a final concentration of 0.8 mM and 0.04 µg/µl, respectively. Primers and probes for the viral N1 and N2 and the human RNase P genes were added as ready-made mix (1 µl; Integrated DNA Technologies Belgium; CatNo. 10006713). The E-gene probes and primers (GATC, Germany) were used at final concentrations of 500 nM for each primer and 125 nM for the probe. ROX was added to a final concentration of 50 nM. PCR was performed in a QuantStudito 5 thermocycler, with cycling conditions 55°C for 10 min, 95°C for 3 min, followed by 45 cycles of 95°C for 15 s and 58°C for 30 s.

For the validation study (Figure 4), real-time PCR of 10 µl RNA-eluate was performed on a BioRad CX96 cycler using the Sarbeco E-Gen-Kit (Fa. Tib Molbiol, Berlin, Germany) following the manufacturer's instructions. Residual volume of extracted RNA from gargle and NP specimens was stored at -20°C and sent to the DKFZ for further analysis.

Lysis of clinical samples for direct SARS-CoV-2 detection

Clinical samples were lysed for direct SHERLOCK or ADESSO assay (Figures 2) as follows: after vortexing, 10µl of sample were mixed with 10µl of QuickExtract DNA Extraction solution

(Lucigen, #QE09050). In the optimised protocol (Figure 4), QuickExtract DNA Extraction
solution is enriched with Rnase Inhibitor, Murine (NEB, #M0314) at a final concentration of
4U/µl. Samples were then incubated at 95°C for 5 min. After incubation, samples were mixed by
vortexing and spun down for 15 seconds at 10.000g. Finally, 5.6 µl of sample (for RT-RPA 2X)
were collected from the upper liquid phase, carefully avoiding to aspirate any precipitate, and
used in the RT-RPA step.
565
crRNA synthesis and purification
CRISPR-RNAs (crRNAs) were either designed in our lab or synthesised by Integrated DNA
Technologies (IDT). All crRNAs used in this study are listed in **Supplementary File 3**. To
produce the crRNAs in our lab we followed a previously published protocol[53]. In short, the
templates for the crRNAs were ordered as DNA oligonucleotides from Sigma-Aldrich with an
appended T7 promoter sequence. These oligos were annealed with a T7-3G oligonucleotide,
and used in an in vitro transcription (IVT) reaction (HiScribe T7 Quick High Yield RNA Synthesis
Kit, NEB, #E2050S). The crRNAs were then purified using Agencourt RNAClean XP Kit
(Beckman Coulter, #A63987). The correct size of the crRNAs was confirmed on a UREA gel
and the concentration evaluated by nanodrop. Aliquots of 10ng/µl of each crRNA were
produced to avoid repeated freeze and thaw cycles and stored at -80°C.
577
Reverse Transcriptase Recombinase polymerase amplification (RT-RPA)
RT-RPA reactions were carried out with TwistAmp Basic (TwistDx, #TABAS03KIT) with the
addition of M-MuLV Reverse Transcriptase (NEB, #M0253) and RNase Inhibitor, Murine (NEB,
#M0314). Reactions were run at 42°C for 45 minutes in a heat block. Here are the details for the
optimised reaction (so called RT-RPA 2X): two lyophilized pellets TwistAmp Basic are used to
prepare the following master mix for 5 reactions: 59 µl of Rehydration Buffer (RB) are mixed
with 2,5 µl of each primer (forward and reverse) at a concentration of 20µM, 1.5 µl of M-MuLV
RetroTranscriptase (200U/µl - NEB, #M0253) and 1,5 µl of Rnase Inhibitor, Murine (40U/µl -
NEB, #M0314). The RB-primer-enzyme mix is used to rehydrate two pellets and finally 5µl of
MgOAc are added. The complete mix is aliquoted (14.4µl) on top of 5,6 µl of each sample. The
RT-RPA protocol was optimised throughout the study. To avoid any confusion, we provide an
additional file (**Supplementary File 4**) with detailed protocols for each experiment presented in
this work. All RPA primers used in this study are listed in **Supplementary File 3** and were
designed following the provided guidelines[53].
592
Cas13 cleavage reaction for lateral flow readout
The reaction mix for Cas13 activity was prepared by combining 4.3 µl of nuclease-free water, 1
µl of cleavage buffer (400mM Tris pH 7.4), 1 µl of LwaCas13a protein diluted in Storage Buffer
(SB)[53] to a concentration of 126.6 µg/ml, 0.5 µl of crRNA (40 ng/µl), 0.5 µl of lateral flow reporter
(IDT, diluted in water to 20 µM), 0.5 µl of SUPERase-In RNase inhibitor (ThermoFisher
Scientific, #AM2694), 0.4 µl of rNTP solution mix (25mM each, NEB, #N0466), 0.3 µl of NxGen
T7 RNA Polymerase (Lucigen, #30223-2) and 0.5 µl of $MgCl_2$ (120mM). 1 µl of the RT-RPA-
amplified product was then added to the mix and, after vortexing and spinning down, the mixture
was incubated for 10 minutes at 37°C in a heat block. The Cas13 protocol was optimised

throughout the study. To avoid any confusion, we provide an additional file (**Supplementary**
File 4) with detailed protocols for each experiment presented in this work.
604
Lateral flow readout
Lateral flow detection was performed using commercially available detection strips (Milenia
HybriDetect 1, TwistDx, Gießen, #MILENIA01). The 10µl-LwaCas13a reactions were transferred
to a tube already containing 80 µl of HybriDetect Assay buffer. After vortexing and spinning
down the reaction mix, a lateral flow dipstick was added to the reaction tube. The result was
clearly readable after one minute. Once the whole reaction volume was absorbed, the dipstick
was removed and photographed with a smartphone camera for band intensity quantification
performed with the freely available ImageJ image processing program[77]. The results are shown
as intensity ratio (test band/control band) and test were considered positive for value of intensity
ratio above 0.2 based on the results shown in Figure S3.
615
Cas13 cleavage reaction for fluorescence readout
The reaction mix for Cas13 activity was prepared by combining 8.6 µl of nuclease-free water, 2
µl of cleavage buffer (400mM Tris pH 7.4), 2 µl of LwaCas13a protein diluted in Storage Buffer
(SB) to a concentration of 126.6 µg/ml, 1 µl of crRNA (40ng/µl), 1 µl of fluorescent reporter (IDT,
diluted in water to a final concentration of 4 µM), 1 µl of RNase inhibitor, Murine (NEB, #M0314),
0.8 µl of rNTP solution mix (25mM each, NEB, #N0466), 0.6 µl of NxGen T7 RNA Polymerase
(Lucigen, #30223-2) and 1 µl of MgCl2 (120mM). 2 µl of the RT-RPA-amplified product was
then added to the mix. The 20µl-LwaCas13a reactions were transferred in 5µl-replicates (4
wells each sample) to a 384-well, round, black-well, clear-bottom plate (Corning, #3544). The
plate was briefly spun down at 500g for 15 sec to remove potential bubbles and placed into a
pre-heated GloMax® Explorer plate reader (Promega) at 37°C.
627
Fluorescence readout
629 Fluorescence was measured every 5 min for 3 h. Data analysis, if not otherwise stated, was
630 performed at the 30-min time-point.
631
RNAse activity detection assay
In order to check for RNase activity in clinical samples, 10 µl of a negative swab and gargle
water sample (Figure 2) were mixed with 10 µl of QuickExtract DNA Extraction Solution with or
without RNase Inhibitor, Murine (NEB, M0314) at a final concentration of 4U/µl. The samples
were then incubated at 95°C for 5 min. After incubation, RNaseAlert substrate v2 (RNaseAlert
Lab Test Kit v2, #4479768, Thermo Fisher Scientific) was added at a final concentration of
200nM. The samples were mixed by vortexing, spun down and incubated at RT for 30 min in the
dark. After incubation, the samples were transferred to a 384-well, round, black-well, clear-
bottom plate (Corning, #3544) in 5µl-replicates (4 wells each sample). The plate was briefly
spun down at 500g for 15 sec to remove potential bubbles and placed into a GloMax Explorer
plate reader (Promega). RNaseAlert substrate fluorescence was measured every 5 min for 30
min. Data analysis, if not differently stated, was performed at the 5-min time-point.
644
ACKNOWLEDGEMENTS

646
We thank the transport service staff, German Cancer Research Center (DKFZ), for providing excellent transportation services of all the clinical samples from the Medical University Mannheim to the DKFZ. The work was supported by a grant from the Ministry of Science, Research and the Arts of Baden-Württemberg for COVID-19 research (grant agreement no. Kap. 1499 TG 93 to Dr. Riccardo Pecori, Prof. Dr. Nina Papavasiliou (DKFZ) and Prof. Dr. Thomas Miethke (Medical Faculty of Mannheim)). Several schematics presented here were created with Biorender.com.

AUTHOR CONTRIBUTIONS

BC, RP, and FNP designed the experiments. JPV and AH produced Cas13 protein. BC and RP performed all the experiments using SHERLOCK/ADESSO. PB and BR performed confirmative RT-qPCR on clinical samples using CDC protocol. MK collected the specimens and AGK performed the RT-qPCR on clinical samples using Tib Molbiol, under the supervision of SW. SA quantified the bands of the lateral flow strips. BC and RP analyzed the data and wrote the manuscript. RP, TM and FNP conceived the study and supervised the research. All authors have read and approved the manuscript.

COMPETING INTERESTS

The DKFZ has filed patent applications regarding this diagnostic methodology (EP 20 173 912.5). RP, FNP and BC are inventors on the above-mentioned patent applications.

REFERENCES

1. WHO Coronavirus (COVID-19) Dashboard. https://covid19.who.int.

2. Zhu, N. *et al.* A Novel Coronavirus from Patients with Pneumonia in China, 2019. *N. Engl. J. Med.* **382**, 727–733 (2020).

3. Huang, C. *et al.* Clinical features of patients infected with 2019 novel coronavirus in Wuhan, China. *The Lancet* **395**, 497–506 (2020).

4. Gorbalenya, A. E. *et al.* The species Severe acute respiratory syndrome-related coronavirus: classifying 2019-nCoV and naming it SARS-CoV-2. *Nat. Microbiol.* **5**, 536–544 (2020).

5. Cevik, M. *et al.* SARS-CoV-2, SARS-CoV, and MERS-CoV viral load dynamics, duration of viral shedding, and infectiousness: a systematic review and meta-analysis. *Lancet Microbe* **2**, e13–e22 (2021).

6. Backer, J. A., Klinkenberg, D. & Wallinga, J. Incubation period of 2019 novel coronavirus

(2019-nCoV) infections among travellers from Wuhan, China, 20–28 January 2020.

Eurosurveillance **25**, (2020).

7. Lauer, S. A. *et al.* The Incubation Period of Coronavirus Disease 2019 (COVID-19) From

Publicly Reported Confirmed Cases: Estimation and Application. *Ann. Intern. Med.* **172**, 577–

582 (2020).

8. Wölfel, R. *et al.* Virological assessment of hospitalized patients with COVID-2019. *Nature*

581, 465–469 (2020).

9. To, K. K.-W. *et al.* Temporal profiles of viral load in posterior oropharyngeal saliva samples

and serum antibody responses during infection by SARS-CoV-2: an observational cohort

study. *Lancet Infect. Dis.* **20**, 565–574 (2020).

10.   Salvatore, P. P. *et al.* Epidemiological Correlates of Polymerase Chain Reaction Cycle

Threshold Values in the Detection of Severe Acute Respiratory Syndrome Coronavirus 2

(SARS-CoV-2). *Clin. Infect. Dis.* ciaa1469 (2020) doi:10.1093/cid/ciaa1469.

11.   Singanayagam, A. *et al.* Duration of infectiousness and correlation with RT-PCR cycle

threshold values in cases of COVID-19, England, January to May 2020. *Eurosurveillance* **25**,

(2020).

12.   Ferretti, L. *et al.* Quantifying SARS-CoV-2 transmission suggests epidemic control with

digital contact tracing. *Science* **368**, eabb6936 (2020).

13.   Lavezzo, E. *et al.* Suppression of a SARS-CoV-2 outbreak in the Italian municipality of

Vo'. *Nature* **584**, 425–429 (2020).

14.   Gudbjartsson, D. F. *et al.* Spread of SARS-CoV-2 in the Icelandic Population. *N. Engl. J.*

Med. **382**, 2302–2315 (2020).

15.   Bai, Y. *et al.* Presumed Asymptomatic Carrier Transmission of COVID-19. (2020).

16.   Rothe, C. *et al.* Transmission of 2019-nCoV Infection from an Asymptomatic Contact in

Germany. *N. Engl. J. Med.* **382**, 970–971 (2020).

17.   Chan, J. F.-W. *et al.* A familial cluster of pneumonia associated with the 2019 novel

coronavirus indicating person-to-person transmission: a study of a family cluster. *The Lancet*

395, 514–523 (2020).

18.   Day, M. Covid-19: four fifths of cases are asymptomatic, China figures indicate. *BMJ*

m1375 (2020) doi:10.1136/bmj.m1375.

19.   Sahin, U. *et al.* COVID-19 vaccine BNT162b1 elicits human antibody and TH1 T cell

responses. *Nature* **586**, 594–599 (2020).

20.   Sahin, U. *et al. BNT162b2 induces SARS-CoV-2-neutralising antibodies and T cells in*

humans. http://medrxiv.org/lookup/doi/10.1101/2020.12.09.20245175 (2020)

doi:10.1101/2020.12.09.20245175.

21.   Anderson, E. J. *et al.* Safety and Immunogenicity of SARS-CoV-2 mRNA-1273 Vaccine

in Older Adults. *N. Engl. J. Med.* **383**, 2427–2438 (2020).

22.   Voysey, M. *et al.* Safety and efficacy of the ChAdOx1 nCoV-19 vaccine (AZD1222)

against SARS-CoV-2: an interim analysis of four randomised controlled trials in Brazil, South

Africa, and the UK. *The Lancet* **397**, 99–111 (2021).

23.   COVID-19 vaccines. *European Vaccination Information Portal* https://vaccination-

info.eu/en/covid-19/covid-19-vaccines.

24.   About covid19-projections.com. *COVID-19 Projections Using Machine Learning*

https://covid19-projections.com/about/.

25.   Vokó, Z. & Pitter, J. G. The effect of social distance measures on COVID-19 epidemics

in Europe: an interrupted time series analysis. *GeroScience* **42**, 1075–1082 (2020).

26.   Roberts, L. How COVID hurt the fight against other dangerous diseases. *Nature* **592**,

502–504 (2021).

27.   Nicola, M. *et al.* The socio-economic implications of the coronavirus pandemic (COVID-

19): A review. *Int. J. Surg.* **78**, 185–193 (2020).

28.   Maringe, C. *et al.* The impact of the COVID-19 pandemic on cancer deaths due to

delays in diagnosis in England, UK: a national, population-based, modelling study. *Lancet*

Oncol. **21**, 1023–1034 (2020).

29.   Donthu, N. & Gustafsson, A. Effects of COVID-19 on business and research. *J. Bus.*

Res. **117**, 284–289 (2020).

30.   Larremore, D. B. *et al.* Test sensitivity is secondary to frequency and turnaround time for

COVID-19 screening. *Sci. Adv.* **7**, eabd5393 (2021).

31.   Vogels, C. B. F. *et al.* Analytical sensitivity and efficiency comparisons of SARS-CoV-2

RT–qPCR primer–probe sets. *Nat. Microbiol.* **5**, 1299–1305 (2020).

32.   Dinnes, J. *et al.* Rapid, point-of-care antigen and molecular-based tests for diagnosis of

SARS-CoV-2 infection. *Cochrane Database Syst. Rev.* (2021)

doi:10.1002/14651858.CD013705.pub2.

33.   Abudayyeh, O. O. & Gootenberg, J. S. CRISPR diagnostics. *Science* **372**, 914–915

(2021).

34.   Abudayyeh, O. O. *et al.* C2c2 is a single-component programmable RNA-guided RNA-

targeting CRISPR effector. *Science* **353**, aaf5573 (2016).

35.   Chen, Y. *et al. N* 1-Methyladenosine detection with CRISPR-Cas13a/C2c2. *Chem. Sci.*

10, 2975–2979 (2019).

36.   East-Seletsky, A., O'Connell, M. R., Burstein, D., Knott, G. J. & Doudna, J. A. RNA

Targeting by Functionally Orthogonal Type VI-A CRISPR-Cas Enzymes. *Mol. Cell* **66**, 373-

383.e3 (2017).

37.   Notomi, T. Loop-mediated isothermal amplification of DNA. *Nucleic Acids Res.* **28**, 63e–

663 (2000).

38.   Piepenburg, O., Williams, C. H., Stemple, D. L. & Armes, N. A. DNA Detection Using

Recombination Proteins. *PLoS Biol.* **4**, e204 (2006).

39.   Gootenberg, J. S. *et al.* Nucleic acid detection with CRISPR-Cas13a/C2c2. *Science* **356**,

438–442 (2017).

40.   Gootenberg, J. S. *et al.* Multiplexed and portable nucleic acid detection platform with

Cas13, Cas12a, and Csm6. *Science* **360**, 439–444 (2018).

41.   Joung, J. *et al. Point-of-care testing for COVID-19 using SHERLOCK diagnostics.*

http://medrxiv.org/lookup/doi/10.1101/2020.05.04.20091231 (2020)

doi:10.1101/2020.05.04.20091231.

42.   Broughton, J. P. *et al.* CRISPR–Cas12-based detection of SARS-CoV-2. *Nat.*

Biotechnol. **38**, 870–874 (2020).

43.   Arizti-Sanz, J. *et al.* Streamlined inactivation, amplification, and Cas13-based detection

of SARS-CoV-2. *Nat. Commun.* **11**, 5921 (2020).

44.   Brandsma, E. *et al.* Rapid, Sensitive, and Specific Severe Acute Respiratory Syndrome

Coronavirus 2 Detection: A Multicenter Comparison Between Standard Quantitative Reverse-

Transcriptase Polymerase Chain Reaction and CRISPR-Based DETECTR. *J. Infect. Dis.*

223, 206–213 (2021).

45.   Huang, Z. *et al.* Ultra-sensitive and high-throughput CRISPR-p owered COVID-19

diagnosis. *Biosens. Bioelectron.* **164**, 112316 (2020).

46.   Patchsung, M. *et al.* Clinical validation of a Cas13-based assay for the detection of

SARS-CoV-2 RNA. *Nat. Biomed. Eng.* **4**, 1140–1149 (2020).

47.   Rauch, J. N. *et al.* A Scalable, Easy-to-Deploy Protocol for Cas13-Based Detection of

SARS-CoV-2 Genetic Material. *J. Clin. Microbiol.* **59**, 8 (2021).

48.   Fozouni, P. *et al.* Amplification-free detection of SARS-CoV-2 with CRISPR-Cas13a and

mobile phone microscopy. *Cell* **184**, 323-333.e9 (2021).

49.   FDA. LetterofAuthorizationforSherlockCRISPRSARS-CoV-2Kit. (2020).

50.   FDA. Letter of Authorization for SARS-CoV-2 RNA DETECTR Assay. (2020).

51.   Bullard, J. *et al.* Predicting Infectious Severe Acute Respiratory Syndrome Coronavirus 2

From Diagnostic Samples. *Clin. Infect. Dis.* **71**, 2663–2666 (2020).

52.   Agrawal, A. India's COVID crisis flags need to forecast variants. *Nature* **594**, 9–9 (2021).

53.   Kellner, M. J., Koob, J. G., Gootenberg, J. S., Abudayyeh, O. O. & Zhang, F.

789   SHERLOCK: nucleic acid detection with CRISPR nucleases. *Nat. Protoc.* **14**, 2986–3012

790   (2019).

791   54.   Zhang, F., Abudayyeh, O. O. & Gootenberg, J. S. A protocol for detection of COVID-19

792   using CRISPR diagnostics. (2020).

793   55.   Qian, J. *et al.* An enhanced isothermal amplification assay for viral detection. *Nat.*

794   *Commun.* **11**, 5920 (2020).

795   56.   CDC. CDC 2019-nCoV Real-Time RT-PCR Diagnostic Panel Instructions for Use.

796   (2020).

797   57.   Fomsgaard, A. S. & Rosenstierne, M. W. An alternative workflow for molecular detection

798   of SARS-CoV-2 – escape from the NA extraction kit-shortage, Copenhagen, Denmark, March

799   2020. *Eurosurveillance* **25**, (2020).

800   58.   Akst, J. RNA Extraction Kits for COVID-19 Tests Are in Short Supply in US. *The*

801   *Scientist Magazine®* https://www.the-scientist.com/news-opinion/rna-extraction-kits-for-covid-

802   19-tests-are-in-short-supply-in-us-67250 (2020).

803   59.   Smyrlaki, I. *et al.* Massive and rapid COVID-19 testing is feasible by extraction-free

804   SARS-CoV-2 RT-PCR. *Nat. Commun.* **11**, 4812 (2020).

805   60.   Bruce, E. A. *et al.* Direct RT-qPCR detection of SARS-CoV-2 RNA from patient

806   nasopharyngeal swabs without an RNA extraction step. *PLOS Biol.* **18**, e3000896 (2020).

807   61.   Metsky, H. C., Freije, C. A., Kosoko-Thoroddsen, T.-S. F., Sabeti, P. C. & Myhrvold, C.

808   *CRISPR-based surveillance for COVID-19 using genomically-comprehensive machine*

809   *learning design.* http://biorxiv.org/lookup/doi/10.1101/2020.02.26.967026 (2020)

810   doi:10.1101/2020.02.26.967026.

811   62.   Korber, B. *et al.* Tracking Changes in SARS-CoV-2 Spike: Evidence that D614G

812   Increases Infectivity of the COVID-19 Virus. *Cell* **182**, 812-827.e19 (2020).

63.   Davies, N. G. *et al.* Estimated transmissibility and impact of SARS-CoV-2 lineage

B.1.1.7 in England. *Science* **372**, eabg3055 (2021).

64.   Davies, N. G. *et al.* Increased mortality in community-tested cases of SARS-CoV-2

lineage B.1.1.7. *Nature* **593**, 270–274 (2021).

65.   Tegally, H. *et al.* Detection of a SARS-CoV-2 variant of concern in South Africa. *Nature*

592, 438–443 (2021).

66.   Pearson, C. A. B. *et al.* Estimates of severity and transmissibility of novel South Africa

SARS-CoV-2 variant 501Y.V2. (2021).

67.   Wang, P. *et al.* Antibody resistance of SARS-CoV-2 variants B.1.351 and B.1.1.7.

Nature **593**, 130–135 (2021).

68.   Madhi, S. A. *et al.* Efficacy of the ChAdOx1 nCoV-19 Covid-19 Vaccine against the

B.1.351 Variant. *N. Engl. J. Med.* **384**, 1885–1898 (2021).

69.   Abu-Raddad, L. J., Chemaitelly, H. & Butt, A. A. Effectiveness of the BNT162b2 Covid-

19 Vaccine against the B.1.1.7 and B.1.351 Variants. *N. Engl. J. Med.* NEJMc2104974

(2021) doi:10.1056/NEJMc2104974.

70.   Byrne, R. L. *et al.* Saliva Alternative to Upper Respiratory Swabs for SARS-CoV-2

Diagnosis. *Emerg. Infect. Dis.* **26**, 2769–2770 (2020).

71.   Ott, I. M. *et al. Simply saliva: stability of SARS-CoV-2 detection negates the need for*

expensive collection devices. http://medrxiv.org/lookup/doi/10.1101/2020.08.03.20165233

(2020) doi:10.1101/2020.08.03.20165233.

72.   Lindner, A. K. *et al.* SARS-CoV-2 patient self-testing with an antigen-detecting rapid test:

a head-to-head comparison with professional testing. (2021)

doi:10.1101/2021.01.06.20249009.

73.   Aleta, A. *et al.* Modelling the impact of testing, contact tracing and household quarantine

on second waves of COVID-19. *Nat. Hum. Behav.* **4**, 964–971 (2020).

74.   Gettleman, J., Yasir, S., Kumar, H., Raj, S. & Loke, A. As Covid-19 Devastates India,

Deaths Go Undercounted. *The New York Times* (2021).

75. Reagent calculator for portal. https://www.who.int/publications/m/item/reagent-calculator-

for-portal.

76. Corman, V. M. *et al.* Detection of 2019 novel coronavirus (2019-nCoV) by real-time RT-

PCR. *Eurosurveillance* **25**, (2020).

77. ImageJ. https://imagej.net/Welcome.

845
846
847
848
849 **FIGURE LEGENDS**
850

Figure 1: A SHERLOCK-based assay for SARS-CoV-2 detection in clinical samples. A.
Graphic of SHERLOCK experimental workflow to detect SARS-CoV-2 in RNA extracted from
clinical samples with lateral flow readout. **B**. SHERLOCK sensitivity on serial dilutions of an IVT
fragment of SARS-CoV-2 S and Orf1a genes. **C**. Comparison of SARS-CoV-2 detection on RNA
extracted from 30 clinical samples via SHERLOCK and RT-qPCR (Medical University Hospital
Mannheim (RT-qPCR hospital) or CDC 2019-nCoV Real-Time RT-PCR Diagnostic Panel).
SHERLOCK was performed on SARS-CoV-2 S gene; RT-qPCR at the Medical University
Hospital Mannheim was performed on SARS-CoV-2 E and Orf1a genes; CDC RT-qPCR was
performed on SARS-CoV-2 N1, N2 and E genes (CDC N1, N2, E) and human RNase P (CDC
Rp) as RNA quality control. T = test band; C = control band; nd = not detected; NTC = non
template control.

Figure 2: SARS-CoV-2 direct detection from clinical samples. A. Graphic of SHERLOCK
experimental workflow to detect SARS-CoV-2 in unextracted clinical samples with lateral flow
readout. **B**. Comparison of three lysis methods for direct detection of SARS-CoV-2 in a COVID-
19 positive clinical sample (sample #30 in Figure 1C) via SHERLOCK with lateral flow readout.
Each lysis method was performed in triplicates. **C**. Determination of SHERLOCK sensitivity with
lateral flow readout on serial dilutions of SARS-CoV-2 synthetic genome spiked in a negative
sample lysed with QuickExtract DNA Extraction Solution. For **B**-**C** band intensity ratios are
shown in the bar plots on the right. T = test band; C = control band; NTC = non template control.
D. SHERLOCK performance on 160 unextracted clinical samples with lateral flow readout. Only
the band intensity ratios of the COVID-19 positive samples (n = 93) are shown in the bar plot.
LoD = Limit of Detection. **E**. Concordance between SHERLOCK (on unextracted samples) and
RT-qPCR (on extracted RNA) for 160 clinical samples (93 positive and 67 negative).

Figure 3: ADESSO: an optimised and highly sensitive SHERLOCK assay. A. Measurement
of RNase activity in a swab sample lysed at 95°C for 5 minutes with QuickExtract DNA
Extraction Solution enriched or not with RNase inhibitor, Murine, at a final concentration of 4
U/µl. **B**. Comparison of SHERLOCK sensitivity on serial dilutions of SARS-CoV-2 genome with

different reverse transcriptases in presence or absence of RNase H. **C**. Optimisation of
SHERLOCK sensitivity with lateral flow readout by increasing the RPA reagents to detect a
false negative sample (#L151, **Supplementary File 1**). A true negative sample (#L126,
Supplementary File 1) is used as negative control. The lateral flow strips whose band intensity
ratios are plotted here are shown in Figure S4B. 1xRPA corresponds to the standard amount of
RPA described in the original SHERLOCK protocol[53] and 5xRPA corresponds to the optimal
amount recommended by the manufacturer. **D**. Confirmation of the improved SHERLOCK
sensitivity with 2xRPA compared to 1xRPA on clinical samples with Ct values close to our LoD
based on Figure 2 (**Supplementary File 1**). The lateral flow strips whose band intensity ratios
are plotted here are shown in Figure S4C. **E**. Optimisation of the Cas13 reaction kinetics by
increasing the amount of Cas13 protein and crRNA in the reaction with fluorescence readout.
The reaction kinetics is evaluated by measuring the fluorescence signal at different time-points.
The complete 3-hour analysis is shown in Figure S4D. **F**. Time-point analysis of the optimised
Cas13 reaction in half the volume to determine the shortest incubation time required to detect a
positive signal with lateral flow readout. The lateral flow strips whose band intensity ratios are
plotted here are shown in Figure S4E. **G**. Sensitivity of the improved SHERLOCK protocol with
lateral flow readout on serial dilutions of SARS-CoV-2 synthetic genome upon integration of all
the above-described optimisations. Intensity ratios are shown in the bar plot on the right. T =
test band; C = control band. **H**. Graphic of the experimental workflow of ADESSO to detect
SARS-CoV-2 in unextracted clinical samples with lateral flow or fluorescence readout.
900
Figure 4: Evaluation of ADESSO performance on clinical samples in direct comparison to
RT-qPCR. A. Schematic of the validation study to assess ADESSO performance for SARS-
CoV-2 detection in clinical samples in comparison with RT-qPCR (Tib Molbiol). The COVID-19
status of the samples included in the study was initially determined by RT-qPCR (COBAS).
ADESSO was performed on both extracted RNA and unextracted samples with lateral flow
readout. The results interpretation for ADESSO was performed without knowledge of the
outcome of RT-qPCR. **B**. ADESSO performance on RNA extracted from swab specimens in
comparison with COBAS RT-qPCR. Negative samples by Tib Molbiol RT-qPCR are represented
in orange. **C**. ADESSO performance on unextracted swab specimens in comparison with
COBAS RT-qPCR (performed on RNA extracted from swabs). **D**. ADESSO performance on
RNA extracted from gargle (G) samples in comparison with COBAS RT-qPCR (performed on
RNA extracted from swabs). Negative samples by Tib Molbiol RT-qPCR are represented in
pink. **E**. ADESSO performance on unextracted G samples in comparison with COBAS RT-
qPCR (performed on RNA extracted from swabs). Samples missed by ADESSO with low Ct
values (<28) according to COBAS RT-qPCR on RNA extracted from swabs but high Ct values
according to Tib Molbiol RT-qPCR on RNA extracted from G samples are represented in dark
red. **F**. Correlation analysis of Ct values obtained with Tib Molbiol RT-qPCR (y axis) or COBAS
RT-qPCR (x axis) on RNA extracted from swab specimens. Negative samples by Tib Molbiol
RT-qPCR are represented in orange and are excluded in the calculation of the correlation (R).
G. Correlation analysis of Ct values obtained with Tib Molbiol RT-qPCR (y axis) and COBAS
RT-qPCR (x axis) on RNA extracted from G samples. Negative samples by Tib Molbiol RT-
qPCR are represented in pink and are excluded in the calculation of the correlation (R). **H**.
Correlation analysis of Ct values obtained after Tib Molbiol RT-qPCR on RNA extracted from G

(y axis) and swab (x axis) samples. Negative swab samples are represented in orange; negative
G samples are represented in pink; negative samples both as G and swab are represented in
red. All the negative samples are excluded in the calculation of the correlation (R). Samples
missed by ADESSO with low Ct values (<28) according to COBAS RT-qPCR on RNA extracted
from swabs but high Ct values according to Tib Molbiol RT-qPCR on RNA extracted from G
samples are represented in dark red. For panels **B**, **C**, **D**, **E** only the band intensity ratios of the
positive samples are shown (n = 95). Values higher than 1 are plotted as equal to 1 for better
visualisation. LoD = Limit of detection.
932
Figure 5: Adaptation of ADESSO for detection of SARS-CoV-2 variants: a flexible and
powerful assay to rapidly identify specific variants or mutations. A. Schematic of SARS-
CoV-2 S gene with annotation of the reported mutations for SARS-CoV-2 B.1.1.7 (top) and
B.1.351 (bottom) lineages. The regions of the S gene targeted by ADESSO and ADESSO-UK
are indicated in purple and orange, respectively. **B**. Schematic of the S gene region containing
the ΔHV69-70 deletion (highlighted in pink) specific for the B.1.1.7 variant in comparison with
the original SARS-CoV-2 sequence from Wuhan and illustration of the binding of the specific
crRNAs targeting the mutated (crRNA ΔHV69-70) or Wuhan (crRNA HV69-70) sequence. The
grey sequence in the crRNAs is called direct repeat (DR) and its stem-loop structure is needed
for the recruitment of Cas13. **C**. SARS-CoV-2 detection by ADESSO in 13 clinical samples
carrying either the UK (B.1.1.7) or SA (B.1.351) SARS-CoV-2 variant. The band intensity ratios
are shown in the bar plot on the right. **D**. SARS-CoV-2 B.1.1.7 variant detection by ADESSO-UK
with ΔHV69-70 crRNA and (**E**) confirmation of the presence of SARS-CoV-2 B.1.351 variant by
ADESSO-UK with HV69-70 crRNA in the same samples. For **C**, **D**, **E**, T = test band; C = control
band; NTC = non template control. **F**. Schematic of the binding of the forward RPA primer used
in ADESSO to the complementary region in the original SARS-CoV-2 sequence from Wuhan
(top), in clinical samples #12 and #13 carrying the SA variant with the deletion Δ242-244
(middle) and in sample #11 carrying the SA variant with an additional mutation (R246I) that
disrupts the primer binding (bottom). The positions of the Δ242-244 deletion and the R246I
mutation are highlighted in grey. The point mutation causing the R246I substitution is marked in
red.
954
955
956 **SUPPLEMENTARY FIGURE LEGENDS**
957
Figure S1: Generation of LwaCas13a and first attempt of SHERLOCK. A. LwaCas13a
protein purification. The LwaCas13 fusion construct also encodes multiple affinity tags and a
protease recognition site at the N terminus of the polypeptide. We have utilized the 6xHIS tag as
the basis for our relatively inexpensive purification, while others have developed an alternative
protocol based on the Strep-tags[53]. After expression in Rosetta cells (inducible via the Lac
operon), the cells are lysed by sonication and the nucleic acid contained within the lysate is
digested. The fusion protein is then purified by nickel-affinity chromatography. The purified
fusion protein is digested with SUMO protease, which cleaves the tags and majority of the
SUMO site off of the mature protein. The SUMO protease and in-tact affinity tags are then
removed from the sample by re-applying the sample to the nickel column, leaving >98% pure

Cas13. We also employ a size exclusion chromatography step to remove any aggregated
Cas13 protein (not pictured). **B**. Serially diluted amounts of pure Cas13 were analyzed by
coomassie staining after conventional SDS-PAGE, revealing a prominent band at the
appropriate molecular weight and only minor contaminants. A serial dilution of BSA was also run
as an estimate of protein concentration by densitometry (which was also validated by BCA
assay). **C**. Sensitivity of home-made Cas13 on serial dilutions of an in-vitro-transcribed (IVT)
fragment of SARS-CoV-2 S gene in the absence of pre-amplification step. Comparison between
fully purified fresh Cas13, partially purified fresh SUMO-Cas13 and fully purified Cas13 stored
overnight (o.n.) at 4°C. Both the band intensity ratios (top) and the corresponding lateral flow
strips (bottom) are shown. **D**. Comparison of SHERLOCK sensitivity on the same IVT fragment
as in panel C when using either ProtoScript II Retro-Transcriptase (as previously published[54]) or
M-MuLV Retro-Transcriptase in the RT-RPA step.

Figure S2: SHERLOCK optimisation: input amount and test of different sets of primers-
crRNA. A. Determination of SHERLOCK sensitivity on serial dilution of SARS-CoV-2 synthetic
genome upon optimisation of RT units and RNA input in the RT-RPA reaction with lateral flow
readout. **B**. Comparison of SHERLOCK performance on different genes in SARS-CoV-2
genome by using alternative sets of primers-crRNA targeting N2 (version 1 and version 2),
Orf1a[61] and S genes on dilutions of a COVID-19 positive sample (sample #6 in Figure 1C). **C**.
Determination of SHERLOCK sensitivity on serial dilutions of SARS-CoV-2 synthetic genome by
using the two most sensitive sets of primers-crRNA selected in panel B targeting Orf1a[61] and S
genes. For panels **A**, **B**, **C**, T = test band; C = control band. For panels **A**, **C**, NTC = non
template control.

Figure S3: SARS-CoV-2 clinical samples. Definition of threshold between positive and
negative results. A. The bar plot shows the band intensity ratios of all the negative controls
utilised in this study (282, in blue) together with the negative (67(Fig. 2) + 100*4(Fig. 4) = 467, in
green) and positive (93(Fig. 2) + 95*4(Fig. 4) = 473, in pink) clinical samples analysed in
Figures 2 and 4. These data were used to define a threshold band intensity ratio of 0.2 to
distinguish between positive and negative samples.

Figure S4: ADESSO: an optimised and highly sensitive SHERLOCK assay. A. Graphic of
SHERLOCK experimental workflow to detect SARS-CoV-2 in unextracted clinical samples with
both lateral flow and fluorescence readout. **B**. Optimisation of SHERLOCK sensitivity with lateral
flow readout by increasing the RPA reagents to detect a false negative sample (#L151,
Supplementary File 1). The band intensity ratios of the lateral flow strips shown here are
plotted in Figure 3C. 1xRPA corresponds to the standard amount of RPA described in the
original SHERLOCK protocol[53] and 5xRPA corresponds to the optimal amount recommended
by the manufacturer. **C**. Confirmation of the improved SHERLOCK sensitivity with 2xRPA
compared to 1xRPA on clinical samples with Ct values close to the LoD from Figure 2
(**Supplementary File 1**). The band intensity ratios of the lateral flow strips shown here are
plotted in Figure 3D. **D**. Scheme of the experiment and complete measurement of the
fluorescence whose results are shown in Figure 3E. **E**. Time-point analysis of the Cas13
reaction to determine the shortest incubation time required to detect a positive signal with lateral

flow readout. The band intensity ratios of the lateral flow strips shown here are plotted in Figure
3F.
1014
Figure S5: Adaptation of ADESSO for detection of SARS-CoV-2 variants. A. RPA primers
optimisation to amplify the region of SARS-CoV-2 S gene surrounding the B.1.1.7 variant-
specific deletion causing ΔHV69-70. Two combinations of the same forward primer with two
alternative reverse primers were tested (set 1 and set 2) on serial dilutions of SARS-CoV-2
synthetic genome (Wuhan sequence). Cas13 detection was performed using crRNA HV69-70.
Band intensity ratios are shown on the right side. T = test band; C = control band. **B**. Band
intensity ratios of 13 clinical samples carrying either the UK (B.1.1.7) or SA (B.1.351) SARS-
CoV-2 variant tested by ADESSO-UK. The corresponding lateral flow strips are shown in Figure
5D and E. The bar plot on the left illustrates the results of the SARS-CoV-2 B.1.1.7 variant
detection by ADESSO-UK with crRNA ΔHV69-70. The bar plot on the right illustrates the results
of the confirmation of the presence of SARS-CoV-2 B.1.351 variant by ADESSO-UK with crRNA
HV69-70. NTC = non template control. **C**. Schematic of SARS-CoV-2 S gene with annotation of
the mutations identified in three patients (clinical samples #11, #12 and #13) carrying the SA
variant. The regions of the S gene targeted by ADESSO and ADESSO-UK are indicated in
purple and orange, respectively. The presence of the mutation R246I in sample #11, here
highlighted in red, disrupts the binding of the RPA forward primer used in ADESSO, thus
impeding the amplification of this region and leading to a false negative result.
1032
1033   **Figure S6: SARS-CoV-2 clinical samples. Frequency distribution of Ct values across the**
1034   **infected patients included in the study. A** and **B**. Frequency distribution and cumulative
1035   frequency distribution, respectively, of the Ct values of all the positive swab samples analysed in
1036   this work (n = 211, **Supplementary File 1**). For both distributions, the bin width is equal to 2
1037   and the R-squared (R^2) was calculated for a gaussian distribution.
1038
1039
1040   **SUPPLEMENTARY MATERIALS**
1041

| SHERLOCK on RNA | number | | | number/total number (percentage) | | | |
	Positive samples (N=10)	Negative samples (N=20)	Total samples (N=30)	Positive predictive value	Negative predictive value	Sensitivity	Specificity
Positive	10	0	10	10/10 (100%)		10/10 (100%)	
Negative	0	20	20		20/20 (100%)		20/20 (100%)

| SHERLOCK on unextracted samples | number | | | number/total number (percentage) | | | |
	Positive samples (N=93)	Negative samples (N=67)	Total samples (N=160)	Positive predictive value	Negative predictive value	Sensitivity	Specificity
Positive	73	0	73	73/73 (100%)		73/93 (78%)	
Negative	20	67	87		67/87 (77%)		67/67 (100%)

1042

1043   **Table S1**: Positive and negative predictive values, sensitivity and specificity of SHERLOCK on
1044   swab samples with (top) and without (bottom) RNA extraction.

Figure 1

Figure 2

Figure 3

Accurate Detection of Evolving SARS-CoV-2 through SHERLOCK Optimisation

Figure 4

Figure 5

A

B

C

D

E

F

Figure S1

Figure S2

A

2.5 µl input, 6 U/µl M-MuLV RT

T ▶

C ▶

cp/µl 10^4 10^3 10^2 10^1 1   0.1   0.01   NTC

1.8 µl input, 20 U/µl M-MuLV RT

T ▶

C ▶

cp/µl 10^4 10^3 10^2 10^1 1   0.1   0.01   NTC

1 µl input, 36 U/µl M-MuLV RT

T ▶

C ▶

cp/µl 10^4 10^3 10^2 10^1 1   0.1   0.01   NTC

B

Target   N2_v1   N2_v2   Orf1a Metsky et al   S

T ▶

C ▶

Dilution of sample #6
(Figure 1C)

   1:10   1:20   1:40   1:10   1:20   1:40   1:10   1:20   1:40   1:10   1:20   1:40

C

Sensitivity on synthetic SARS-CoV2 genome

T ▶

C ▶
                                               Orf1a
Metsky et al

T ▶

C ▶
                                               S

   10^4 10^3 10^2 10^1 1   0.1   0.01   NTC

Synthetic RNA (cp/µl)

Figure S3

Figure S4

Figure S5

Figure S6

A

Frequency distribution Ct values

B

Cumulative Frequency distribution Ct values