Accuracy verification of low-cost CO_2 concentration measuring devices for general use as a countermeasure against COVID-19

Yo Ishigaki[1], Koji Enoki[1], Shinji Yokogawa[1,2]

Affiliations:
1) Graduate School of Informatics and Engineering, The University of Electro-communications, Chofu, Tokyo, Japan.
2) Info-powered Energy System Research Center (iPERC), The University of Electro-communications, Chofu, Tokyo, Japan.

Corresponding author:
Yo Ishigaki, ishigaki@uec.ac.jp, 1-5-1 Chofugaoka, Chofu, Tokyo 182-8585, Japan.

Declaration of competing financial interests (CFI):
The authors indicate no conflicts of interest.

Abstract:

Within the context of the COVID-19 pandemic, CO_2 sensors that measure ventilation conditions and thereby reduce the risk of airborne infection, are gaining increasing attention. We investigated and verified the accuracy of 12 relatively low-cost sensor models that retail for less than \$45 and are advertised as infection control measures on a major e-commerce site. Our results indicate that 25% of the tested sensors can be used to identify trends in CO_2 concentration, if correctly calibrated. However, 67% of sensors did not respond to the presence of CO_2, which suggests that a type of pseudo-technique is used to display the CO_2 concentration. We recommend that these sensors are not suitable for infection prevention purposes. Furthermore, 58% of the investigated sensors showed significant responses to the presence of alcohol. Owing to the widespread use of alcohol in preventing the spread of infectious diseases,

sensors that react to alcohol can display inaccurate values, resulting in

inappropriate ventilation behavior. Therefore, we strongly recommended that

these sensors not be used. Based on our results, we offer practical

recommendations to the average consumer, who does not have special

measuring equipment, on how to identify inaccurate CO_2 sensors.

Introduction

In enclosed spaces with inadequate ventilation and air treatment, the

airborne spread of SARS-CoV-2 can occur even at a distance of 6 feet or more

from an infected person, especially as concentrations of very fine droplets and

aerosol particles drift through the air [1]. Within this context, airborne transmission

may have accounted for more than 50% of the disease transmission on the

Diamond Princess, the cruise ship associated with the SARS-CoV-2 outbreak in

Japan [2].

The concentration of indoor carbon dioxide (CO_2) has attracted attention as

a proxy for infection risk [3]. While outdoor fresh air has a constant CO_2

concentration of approximately 400 ppm, human breath contains a much larger

amount of CO_2 at 30,000–40,000 ppm. Therefore, by measuring indoor CO_2

concentration, it is possible to understand how much human breath is retained in

the air.

The American Society of Health Engineers (ASHE) states that there is a

correlation between ventilation capacity and the risk of airborne infection from

tuberculosis, measles, chicken pox, influenza, smallpox, and SARS [4]. The

Wells-Riley equation (Equation 1) can be used to quantitatively evaluate the risk

of airborne infection by bacteria and viruses [5-6].

$$P = 1 - \exp\left[-\frac{Iqpt}{Q}\left\{1 - \frac{V}{Q\theta}\left[1 - \exp\left(-\frac{Q\theta}{V}\right)\right]\right\}\right] \quad (1)$$

In this equation, P is the probability that an infected person will be infected,

I is the number of infected persons in the enclosed space, p is the respiratory

volume of one person (m^3/h), q is the rate of infectious droplets (/h), t is the time

spent by the infected person in the space (h), θ is the time spent by the infected

person in the space (h), V is the volume of the room (m^3), and Q is the ventilation

rate (m^3/h). Thus, the Wells-Riley equation models the probability of exposure to

infectious droplets emitted by an infected person and inhaled by an uninfected

person   through respiration as the risk quantity. As infectious droplets are

accompanied by a high concentration of CO_2, both of which are airborne, it is

beneficial to use CO_2 concentration as an indicator of the airborne infection risk.

A number of researchers have suggested that the threshold ventilation rate

for maintaining the effective reproduction rate of tuberculosis (the ratio of new

secondary infected persons to the source of infection) below 1 corresponds to a

CO_2 concentration of 1,000 ppm. Specifically, Du et al. found that by improving

the ventilation system in a poorly ventilated university building, where an outbreak

of tuberculosis occurred (27 tuberculosis patients and 1,665 contacts), and by

reducing the maximum CO_2 concentration from 3204 ± 50 to 591-603 ppm, the

secondary infection rate among new contacts could be reduced to zero (mean

follow-up, 5.9 years). Furthermore, the incidence of tuberculosis among contacts

could be reduced by 97% (95% CI: 50–99.9%) when the CO_2 concentration

remained below 1,000 ppm [7].

After comprehensively considering laws and regulations concerning CO_2

concentration, the Occupational Hygiene and Ergonomics Subcommittee of the

Japan Society for Occupational Health (JSOH) established classifications for

estimating ventilation status according to CO_2 concentrations (Table 1), and

recommended measures according to each classification [8].

Table 1 Ventilation criteria based on CO_2 concentrations, established by The

Occupational Hygiene and Ergonomics Subcommittee of the Japan Society for

Occupational Health (JSOH).

CO_2 concentration	Classification	Description
Less than 1,000 ppm or less	Good	High quality air. The status should be maintained.
1,000 to 1,500 ppm or less	Moderately good	Acceptable limits. Occasional opening of a number of windows is recommended (for a few minutes per hour).
1,500 to 2,500 ppm or less	Poor	Open windows for at least a few minutes every half hour (fully open) and do not use the room.
2,500 to 3,500 ppm or less	Very poor	Open the windows all the time (fully open) and do not use the room.

3,500 ppm or greater	Extremely poor	Do not use the room.

Like those of other countries, the Japanese Building Standard Law (revised

and implemented in 2003), requires the installation of a 24-hour ventilation

system in all living rooms to prevent the development of the sick building

syndrome, and requires an air change rate per hour (ACPH) of 0.5 times/h or

more inside houses. With an ACPH of 0.5 times/h, for example, in a 70 m^2 space

with a standard activity level, the CO_2 concentration at saturation is 800 ppm with

two people in the room, 1,000 ppm with three people in the room, and 1,200 ppm

with four people in the room.

While the Building Standard Law specifies the amount of ventilation, Japan's

Building Management Law (enacted in 1970) specifies CO_2 concentration

standards. These laws specify that in buildings used as entertainment venues,

department stores, restaurants, shops, offices, and schools, where the enclosed

area used for specific purposes is 3,000 m^2 or more, defined as "specified

buildings" (> 8,000 m^2 for schools), the CO_2 concentration should be 1,000 ppm

or less.

Against this background, the monitoring of CO_2 concentration by sensors is

currently being advocated to reduce the risk of COVID-19 in indoor environments.

In many instances, especially in the context of a pandemics, it is recommended

to maintain indoor CO_2 concentrations below 800 ppm (e.g., Minnesota

Department of Health, Federation of European Heating Ventilation and Air

Conditioning Associations, UK Scientific Advisory Group for Emergency, US

Centers for Disease Control and Prevention) [9]. In Japan, a widespread system

is used, in which local government supports the cost of installing CO_2 sensors in

restaurants to prevent the spread of COVID-19. For example, the Tokyo

Metropolitan Government, Kyoto Prefecture, and Osaka City support the

purchase of infection prevention devices, including CO_2 sensors, up to 100,000–

300,000 yen. In addition, Shibuya Ward in Tokyo distributes free CO_2 sensors to

restaurants located in the ward.

In response to this sudden increase in demand, low-cost CO_2 sensors have

rapidly become available on e-commerce sites such as Amazon, and are

advertised as being effective in preventing the spread of infectious diseases.

Many of these sensors have vague descriptions of their measurement principles,

with several consumers questioning the CO_2 measuring ability of these sensors.

In addition, there is a lack of guidelines for the use of CO_2 sensors in Japan, and

a lack of laws or regulations regarding their accuracy and reliability. In response

to this research need, we purchased several low-cost CO_2 sensors, to verify their

accuracy and investigate the related measurement principles.

Although several studies have already evaluated low-cost gas sensors [10,

11], including CO_2 sensors, the global outbreak of COVID-19 has led to a

proliferation of new products on the market for the control of infectious diseases

that have not been tested for efficacy and accuracy. Therefore, we conducted this

study to investigate the accuracy and reliability of these sensors and to offer

advice on these sensors to the general consumer, who does not have access to

specialized inspection equipment.

Methodology

In this study, we purchased and investigated 12 relatively low-cost CO_2 sensors (1–12; Table 2), within a price range of 2,900–4,999 JPY (approximately 26–45 USD, at the exchange rate in July 28, 2021) from the Amazon Japan website, and installed them according to instructions. The sensors were advertised as being effective in preventing COVID-19 infections. In addition to the 12 sensors, we purchased two relatively expensive CO_2 sensors, (A and B) that are widely used for industrial and experimental purposes as reference devices. Sensor A is manufactured by the T&D corporation TR-76Ui (54,780 JPY, approximately 500 USD), with a zero and 1500 ppm span calibration, while Sensor B is manufactured and factory calibrated by the C.H.C. System corporation NMA-PR-R (47,500 JPY, approximately 430 USD). Both sensors A and B are based on the non-dispersive infrared (NDIR) method.

All 14 sensors (1 to 12, A and B) were activated simultaneously and placed in a chamber (Fig. 1). The following protocol was used for the experiment:

1. Open the chamber lid.
2. Turn on the fan.
3. Ventilate the chamber with room air (adjusted to 25°C) for 5 min.
4. Close the chamber lid.
5. Turn off the fan.
6. <FRESH AIR> Record the displayed values of each measuring instrument. However, if a value outside the measurement range is displayed, it is not recorded.

7. Turn on the fan.

8. <INJECTION> Inject (additional) CO_2 gas, (additional) room air and/or

alcohol into the chamber to change the gas concentration in the chamber.

9. Wait for 5 min until the inner air and gas are evenly mixed in the chamber.

10. Turn off the fan.

11. <CONTROLLED ATMOSPHERE> Record the displayed values of each

measuring instrument. However, if a value outside the measurement

range is displayed, it is not recorded.

12. Return to step 7 (or terminate the experiment).

167   Table 2 List of CO_2 sensors used in the experiment.

ID	Manufacturer and/or product name indicated on the purchase page	Price (JPY)	Claimed specifications on the purchase page		
			Sensor type	Measurement range	Accuracy information
1	Kecheer	3,589	Unknown	Unknown	Unknown
2	Walfront	4,909	Unknown	0–9,999	±50 ppm ± 5% rdg (0-5000 ppm) 0.05% FSO
3	ROOMMATE Tokyo Deco	3,480	Unknown	400–5,000	Unknown
4	Sign Kingdom Xmonitor-r1	3,680	NDIR	Unknown	Unknown
5	KKmoon	4,899	Unknown	350–2,000	Unknown
6	Elikliv-10	2,900	Unknown	Unknown	Unknown
7	Atmonas	4,999	NDIR	400–5,000	Unknown

8	Seawardi	3,009	Unknown	Unknown	Error variation 3 to 30%
9	Rain Queen	4,700	Unknown	400–6,000	±50 ±5%
10	Lnicez	4,298	NDIR	400–5,000	Unknown
11	IUGGAN	4,800	Unknown	400–5,000	Unknown
12	YC22	4,599	Unknown	400–5,000	Unknown
A	T&D TR-76Ui	54,780	NDIR	0–9,999	CO_2 ± 50 ppm or ± 5% of reading at ≦ 5000 ppm
B	C.H.C.System NMA-PR-R	47,500	NDIR	0–3,000	±75 ppm or ±5% (whichever is greater)

Figure 1. Image of experimental equipment and setup used for an investigation

of the accuracy of 12 CO_2 sensors.

Results and Discussion

We compared the CO_2 values of the 12 investigated sensors to those of the

reference sensors (A and B; Fig. 2). The relatively expensive sensor B showed a

narrower measurement range than sensor A, although both had very similar

response characteristics. Sensor 10 generally displayed higher values than

sensor A, while sensors 4 and 7 displayed lower values than sensor A. Although

sensor 2 responded to CO_2, it seemed to malfunction, as it displayed very high

CO_2 values. Sensors 1, 3, 5, 6, 8, 9, and 12 did not seem to respond to CO_2 levels.

As these seven sensors are able to display total volatile organic compound

(tVOC) values, we conducted an additional experiment by applying approximately

5 ml of rubbing alcohol to the inside of the acrylic plate of the chamber.

Figure 2. Response characteristics of CO_2 sensors. Measured values of sensor

A, the reference device, are plotted on the horizontal axis, and the measured

values of sensors A, B, and 1 to 12 are plotted on the vertical axis, connected

190   by spline curves.

Figure 3. Response characteristics of 14 CO_2 sensors to alcohol, including two reference sensors (A and B). "1. Air" on the x-axis corresponds to the sixth step (FRESH AIR) in the experimental protocol, while "2. Alcohol" refers to the twelfth step of the protocol (CONTROLLED AIR), with 5 ml of alcohol added to the chamber at the step 8 (INJECTION) instead of CO_2.

During this experiment, rather than adding CO_2 gas, 5 ml of alcohol was sprayed on the acrylic wall in the chamber (Fig. 3). To eliminate the cross-recruitment and contamination effect of alcohol, the experiment was performed independently to the previous experiment.

Our results (Fig. 3) indicate that the readings of sensors A, B, 2, 4, 6, 7, and 10 were not affected by the presence of alcohol. On the other hand, sensors 1,

3, 8, 9, 11, and 12 reacted strongly to alcohol. Furthermore, these six sensors

were part of the group of seven sensors that did not respond to CO_2 (Fig. 2). The

remaining sensor 6 appeared to respond neither to CO_2 nor to alcohol.

Of the 12 relatively low-cost sensors investigated in this study, three models

(sensors 4, 7, and 10; 25% of the total) can potentially be used to determine CO_2

concentration levels, if correctly calibrated. However, the accuracy of the sensors

was not as good as that of the more expensive reference sensors. Machine

learning methods have been proposed for calibration of inexpensive sensors [12,

13]. Sensor 2 seemed to be malfunctioning (8% of the total), while eight sensor

models (67%) did not respond to CO_2, and seven of these (58% of the total)

responded strongly to alcohol, suggesting that the sensors were equipped with

different sensors to the CO_2 sensing module.

To investigate the sensing module, we purchased and disassembled an extra

model of sensor 12 (Fig. 4). Notably, the NDIR sensor was not present. Based on

the model number of the sensing module and the manufacturer's specifications

[14], we identified that the sensor was not designed for CO_2 measurements, but

rather for the measurement of "ammonia, hydrogen, alcohol, carbon monoxide,

methane, propane, Gan, styrene, propylene glycol, alkyl phenol, toluene,

ethylbenzene, xylene, formaldehyde, and other volatile organic gases, incense

smoke, wood, paper smoke burning out." This suggests that sensor 12 is

designed to display the concentration of CO_2 by a type of imitating algorithm

based on the value of the sensing module that detects miscellaneous gases,

excluding CO_2.

Globally, several sensors display a very vaguely defined measurement

called eCO_2 (equivalent CO_2). This is estimated from the tVOC value and has

been shown to differ from the actual CO_2 concentration [15].

Sensors 1, 3, 8, 9, 11, and 12, which responded strongly to alcohol (Fig. 3),

all displayed tVOC values separately from the CO_2 concentration. Therefore, we

suspected that these six sensor models were displaying eCO_2 values rather than

actual CO_2 concentrations. However, the sales pages for these six sensors did

not indicate that they used eCO_2 sensors.

Figure 4. Sensing module of Sensor 12. A cylindrical object was suspended

239   in the air by four wire bonds, and inside it was a coil heater.

Based on the results of our experiments, we developed a simple method for

consumers to verify the quality of sensors on offer, without needing specialized

equipment. Within this context, we recommend the following:

1. Ensure that the sensor displays a value of approximately 400 ppm in
fresh air outdoors (this will allow for the detection of faulty sensors such
as sensor 2, with a large deviation in CO_2 values).

2. Ensure that the CO_2 concentration rises significantly when breathing on
the sensor (this will help identify sensors that do not respond to increases
in CO_2, such as Sensors 1, 3, 5, 6, 8, 9, and 12).

3. Ensure that the sensor does not react erroneously to alcohol by holding
a hand dipped in an alcohol disinfectant close to the sensor (this
eliminates low-quality sensors that pseudo-indicate CO_2 concentration
from tVOC concentrations such as alcohol).

Conclusions

Twenty-five percent of the relatively low-cost sensors investigated in this
study have the potential to be used for identifying trends in CO_2 concentration, if
correctly calibrated, although with a poor accuracy. Notably, 67% of the sensors
did not respond to changes in CO_2 levels, which suggests that various pseudo-
masquerading techniques are being used to display CO_2 concentrations. These
sensors should not be used to prevent the spread of infectious diseases. In
addition, 58% of the sensors reacted strongly to alcohol. As alcohol is widely used
indoors to prevent the spread of COVID-19, sensors that respond to alcohol will
display inflated values, which can lead to incorrect ventilation behavior. Therefore,
we strongly recommended that these sensors not be used.

This study was conducted using a total of 14 commercially available sensor
models. However, there are many models of CO2 sensors available in the market.

We suggest that future research focus on an exhaustive analysis of the sensors

available in the market as a basis for a discussion on the guidelines and

regulations for CO_2 sensors. We are planning to use gas chromatography with

an accuracy of ±0.1 ppm to simultaneously detect the concentration of alcohol

and carbon dioxide for more accurate verification. With this method, we will be

able to measure both organic and inorganic gases, which was not possible before.

In addition, clarification of the electrical circuit, catalyst, semiconductor

composition, and solid-state properties of the sensing modules (Fig. 4) will assist

to pinpointing the purpose for which these sensors were developed. Furthermore,

we suggest that future research focus on not only conducting exposure

experiments with pure CO_2 gas, but also include accurate measurements with

human-derived CO_2 gases, including various volatile organic compounds.

Notes
The authors declare no conflicts of interest. This paper has been previously
submitted to medRxiv, a preprint server for health sciences. This study was
approved by the Ethics Committee on Experiments on Human Subjects (approval
number of 21005), The University of Electro-communications, located at
Chofugaoka 1-5-1, Chofu, Tokyo, Japan.

Acknowledgments
This work was supported by the Research Grant Program of the KDDI Foundation.
In addition, some of the data acquisition technologies in this study are based on
results obtained from a project commissioned by the New Energy and Industrial
Technology Development Organization (NEDO).

Data availability
The raw data of the graphs are available on FigShare.

https://figshare.com/articles/figure/Accuracy_verification_of_low-
cost_CO2_concentration_measuring_devices_for_general_use_as_a_counterm
easure_against_infectious_diseases/15067557
300
References

1) CDC (Centers for Disease Control and Prevention), SARS-CoV-2
Transmission, Summary of recent changes, Updated May 7, 2021
Online: https://www.cdc.gov/coronavirus/2019-ncov/science/science-
briefs/sars-cov-2-transmission.html (accessed 25 July 2021)
2) Azimi P, Keshavarz Z, Laurent JGC, Stephens B, Allen JG. Mechanistic
transmission modeling of COVID-19 on the Diamond Princess cruise ship
demonstrates the importance of aerosol transmission. Proc Natl Acad Sci
2021; 118(8) e2015482118; DOI: 10.1073/pnas.2015482118.
3) Memarzadeh F. Room Ventilation and Airborne Disease Transmission,
American Hospital Association, Editor: The American Society for Healthcare
Engineering; 2013.
4) Wells WF. *Airborne contagion and air hygiene*. Harvard: Harvard University
Press, 1955.
5) Riley EC, Murphy G, Riley RL. Airborne spread of measles in a suburban
elementary school. Am J Epidemiol 1978; 107: 421-432.
6) Buonanno G, Stabile L, Morawska L. Estimation of airborne viral emission;
Quanta emission rate of SARS CoV-2 for infection risk assessment. Environ
Int 2020; 141: 1–8.
7) Du CR, Wang SC, Yu MC, Chiu TF, Wang JY, Chuang PC, Jou R, Chan PC,
Fang CT. Effect of ventilation improvement during a tuberculosis outbreak in
underventilated university buildings, Indoor Air 2019: 30(3); 422-432.
8) The Occupational Hygiene and Ergonomics Subcommittee of the Japan
Society for Occupational Health (JSOH): Ventilation simulator for COVID-19
countermeasures (COVID-19 Taisaku-you Kanki Simulator). Online:
http://jsoh-
ohe.umin.jp/covid_simulator/ventilation_simulator_manual_ver1.0.pdf
(accessed 25 July 2021)
9) Eykelbosh, A. Indoor CO_2 Sensors for COVID-19 Risk Mitigation: Current
Guidance and Limitations. Vancouver, BC: National Collaborating Centre for
Environmental Health. 2021 May.

Online: https://ncceh.ca/documents/field-inquiry/indoor-co2-sensors-COVID-19-risk-mitigation-current-guidance-and (accessed 25 July 2021).

10) Lewis AC, Lee JD, Edwards PM, Shaw MD, Evans MJ, Moller SJ, Smith KR, Buckley JW, Ellis M, Gillot SR, Whited A. Evaluating the performance of low cost chemical sensors for air pollution research. Faraday Discuss 2016; 189 Online: https://doi.org/10.1039/C5FD00201J (accessed 25 July 2021).

11) Castell N, Dauge FR, Schneider P, Vogt M, Lerner U, Fishbain B, Broday D, Bartonova A. Can commercial low-cost sensor platforms contribute to air quality monitoring and exposure estimates? Environ Int 2017; 99: 293-302, ISSN 0160-4120, 2017.
Online: https://doi.org/10.1016/j.envint.2016.12.007 (accessed 25 July 2021)

12) Zimmerman N, Presto AA, Kumar SPN, Gu J, Hauryliuk A, Robinson ES, Robinson AL, Subramanian R. A machine learning calibration model using random forests to improve sensor performance for lower-cost air quality monitoring. Atmos Meas Tech, 2018; 11:291–313.
Online: https://doi.org/10.5194/amt-11-291-2018 (accessed 25 July 2021)

13) Zaidan MA, Motlagh NH, Fung PL, Lu D, Timonen H, Kuula J, Niemi JV, Tarkoma S, Petäjä T, Kulmala M, Hussein T. Intelligent Calibration and Virtual Sensing for Integrated Low-Cost Air Quality Sensors. IEEE Sensors Journal 2020; 20(22): 13638-13652 doi: 10.1109/JSEN.2020.3010316.
Online: https://ieeexplore.ieee.org/document/9144227 (accessed 25 July 2021)

14) TPM-300 Air Quality Module Intelligent Air Quality Monitoring Module Air Sensor
Online: https://ardubotics.eu/en/sensors/1243-tpm-300-air-quality-module-intelligent-air-quality-monitoring-module-air-sensor.html (accessed 25 July 2021)

15) Varzaru G, Zarnescu A, Ungurelu R, Secere M. Dismantling the confusion between the equivalent CO_2 and CO_2 concentration levels. 11th International Conference on Electronics, Computers and Artificial Intelligence (ECAI) 2019: pp. 1-4, doi: 10.1109/ECAI46879.2019.9042113.