Serological detection of SARS-CoV-2 IgG using commercially available immunoassays on

dried blood spots collected from patients

Gregory J Walker[1,2,3], Rebecca Davis[4], Zin Naing[3], Brad McEntee[5], Yonghui Lu[3], Tatijana

Denadija[4], William D Rawlinson[1,2,3,6]

Affiliations:

[1] Virology Research Laboratory, Serology and Virology Division, NSW Health Pathology SEALS

Randwick, NSW 2031, Australia

[2] School of Medical Sciences, Faculty of Medicine, University of New South Wales, Sydney,

NSW 2052, Australia

[3] Serology and Virology Division (SaViD), NSW Health Pathology, Prince of Wales Hospital,

Sydney, NSW 2031, Australia

[4] Department of Microbiology and Infectious Diseases, NSW Health Pathology, Royal Prince

Alfred Hospital, NSW

[5] Special Health Accommodation, Sydney Local Health District, NSW

[6] School of Biotechnology and Biomolecular Sciences (BABS) Faculty of Science, University of

New South Wales, Sydney, NSW 2052, Australia

*Corresponding author: Prof William Rawlinson, Virology Research Laboratory, Level 3

Clinical Sciences Building, Prince of Wales Hospital, Randwick 2031, Australia.

E-mail: w.rawlinson@unsw.edu.au   **Ph:** +61 2 93829188

Running title: SARS-CoV-2 IgG detection from dried blood spots

ABSTRACT

Serological testing for SARS-CoV-2 antibodies provides important research and diagnostic information relating to COVID-19 immune response and surveillance. A major challenge when addressing protection post- infection or vaccination is the difficulty of specimen collection from infants and children. Dried blood spots (DBS) collected by finger prick or heel prick are a minimally invasive sample collection alternative previously used to detect antibodies to other viruses. In this study we evaluated DBS for the detection of SARS-CoV-2 antibodies on three commercially available enzyme (EIA) and chemiluminescent (CLIA) immunoassays by analysing paired DBS and serum samples collected from 54 subjects. We demonstrate that testing of DBS samples was highly sensitive and specific, and quantitative results strongly correlated with those of paired serum. These results suggest that DBS derived blood is a viable alternative to plasma or serum for use in EIAs and CLIAs, and has particular utility as a minimally invasive collection tool for COVID-19 serological testing of infants and children.

Serological methods for SARS-CoV-2 antibody detection are routinely performed on plasma

or serum from peripherally collected venous blood. This can be inconvenient and often

unacceptable to children. Dried blood spots (DBS) collected by finger prick or heel prick are

a minimally-invasive alternative previously used to detect antibodies against Epstein-Barr,

HIV, Hepatitis, Rubella and other viruses (1). Evaluation of DBS for the detection of SARS-

CoV-2 antibodies on commercially available enzyme (EIA) and chemiluminescent (CLIA)

immunoassays was performed by analysing paired DBS and serum samples from 54 subjects

aged between 5 and 73 years.

Whole blood was collected from patients via venepuncture and spotted onto DBS cards

(Whatman 903 Protein Saver), with the remainder put into a serum tube. DBS were then

processed (Supplementary materials) and tested in parallel with matched sera for

antibodies against SARS-CoV-2 using the spike-based ARCHITECT SARS-CoV-2 IgG II

Quantitative Assay (Abbott Laboratories, Abbott Park, IL, USA), and the S1 spike- and

nucleocapsid-based Euroimmun Anti-SARS-CoV-2 ELISA (Euroimmun, Lübeck, Germany)

assays, following manufacturer's instructions. The ARCHITECT SARS-CoV-2 IgG II

Quantitative Assay records data as arbitrary units (AU), and these results were interpreted

as positive ($\geq$50 AU/mL) and negative (<50 AU/mL). For Euroimmun assays, the ratio of

sample absorbance over the calibrator (OD ratio) was used to categorise sample results as

positive ($\geq$1.1), equivocal (1.1> OD $\geq$0.8), and negative (<0.8). Samples returning an

equivocal result were excluded from sensitivity and specificity calculations in this analysis.

Sensitivity and specificity were calculated using results of parallel testing on peripherally

collected serum samples as the reference standard. Concordance between quantitative

measurements were calculated with Pearson's Correlation Coefficient.

For the spike-based ARCHITECT and Euroimmun assay there were no discordant results

between any of the 54 paired samples. On both assays DBS testing displayed 100%

sensitivity and 100% specificity with reference to matched serum. The correlation

coefficients between quantitative serum and DBS measurements were r= 0.9662 (0.9421 –

0.9803, p<0.0001) and r= 0.9831 (0.9708 - 0.9902, p< 0.0001), respectively (Figure). On a

subset of 46 matched samples tested on the Euroimmun nucleocapsid-based assay there

was one discordant result, and DBS testing exhibited 100% sensitivity. One of 5 pairs that

had a negative result on serum testing was positive on DBS testing. The correlation

coefficient between quantitative serum and DBS measurements were r= 0.9075 (0.8380 –

0.9480, p<0.0001).

Figure. Linear regression of CLIA arbitrary units per millilitre (AU/mL) or EIA optical density (OD)

ratios for the detection of anti-SARS-CoV-2 IgG in 54 paired dried blood spot (DBS) and serum

samples. Samples were run on the ARCHITECT SARS-CoV-2 IgG II Quantitative Assay **(A)**, and the

Euroimmun Anti-SARS-CoV-2 ELISA S1 spike **(B)** and nucleocapsid assays **(C)**. Dotted lines indicate

the positive/negative, negative/equivocal, and equivocal/positive cutoffs of respective assays.

The high sensitivity and specificity of DBS tests and strong quantitative relationship with

paired serum on the spike-based ARCHITECT CLIA and Euroimmun EIA demonstrates the

validity of DBS collection for the detection of SARS-CoV-2 antibodies on these assays. This is

the first study to validate the quantitative ARCHITECT assay for DBS testing, and the first to

include paediatric specimens in analysis (Supplementary materials). Our findings also

confirm those of recent studies in which adult DBS were validated using semi-quantitative

EIAs targeting the S1 spike (2-4) and nucleocapsid proteins (3-5). These data suggest DBS

derived blood is a viable alternative to plasma or serum collection. It has particular utility for

testing infants and children, and in hotel or home quarantine where sample collection is

performed away from clinical settings. It can reduce transmission risks to staff collecting

samples from infected patients, as less time needs to be spent with the child to obtain the

sample. Additionally, this non-invasive tool collection tool is well suited to COVID-19

surveillance.

References

1.	Amini F, Auma E, Hsia Y, et al. Reliability of dried blood spot (DBS) cards in antibody measurement: A systematic review. PLOS ONE. 2021;16(3):e0248218.

2.	Amendola A, Bianchi S, Gori M, et al. Dried Blood Spot as an Alternative to Plasma/Serum for SARS-CoV-2 IgG Detection, an Opportunity to Be Sized to Facilitate COVID-19 Surveillance Among Schoolchildren. The Pediatric Infectious Disease Journal. 2021;40(1):e46-e7.

3.	Weisser H, Steinhagen K, Höcker R, Borchardt-Lohölter V, Anvari Ö, Kern PM. Evaluation of dried blood spots as alternative sampling material for serological detection of anti-SARS-CoV-2 antibodies using established ELISAs. Clinical Chemistry and Laboratory Medicine (CCLM). 2021;59(5):979-85.

4.	Zava TT, Zava DT. Validation of dried blood spot sample modifications to two commercially available COVID-19 IgG antibody immunoassays. Bioanalysis. 2021;13(1):13-28.

5.	Mulchandani R, Brown B, Brooks T, et al. Use of dried blood spot samples for SARS-CoV-2 antibody detection using the Roche Elecsys ® high throughput immunoassay. Journal of clinical virology : the official publication of the Pan American Society for Clinical Virology. 2021;136:104739-.

Supplementary Material

Serological detection of SARS-CoV-2 IgG using a commercially available enzyme

immunoassays on dried blood spots collected from patients

Gregory J Walker, Rebecca Davis, Zin Naing, Brad McEntee, Yonghui Lu, Tatijana Denadija,

William D Rawlinson

Corresponding author: Prof William Rawlinson, Virology Research Laboratory, Level 3

Clinical Sciences Building, Prince of Wales Hospital, Randwick 2031, Australia.

E-mail: w.rawlinson@unsw.edu.au   Ph: +61 2 93829188

This file includes:

Supplementary methods. Processing of dried blood spot samples

Table S1. Age distribution of patients sampled

Processing of dried blood spot samples

Whole blood was collected from patients via venepuncture and spotted onto Whatman 903 Protein Saver DBS cards. DBS samples were then processed by methods applicable to each assay platform.

For the ARCHITECT SARS-CoV-2 IgG II Quantitative Assay (Abbott Laboratories, Abbott Park, IL, USA) a 6mm spot was punched using a standard office hole punch and transferred to a 2 ml round bottom centrifuge tube. After 300μl of 0.25% Tergitol 15-S-9 in PBS was added, the sample tube was incubated on an orbital shaker at 400 rpm for 60 min, followed by a centrifugation at 10,000g for 10 min. Supernatant (130μL) was then transferred to a sample cup and loaded onto ARCHITECT i2000SR analyzer for testing. As the current on-market ARCHITECT SARS-CoV-2 IgG II Quantitative Assay is not claimed for DBS use, a Research Use Only (RUO) assay file was installed to facilitate the DBS sample measurement. The DBS method for this platform may be further modified to improve agreement with quantitative serum results.

For the Euroimmun Anti-SARS-CoV-2 ELISA (Euroimmun, Lübeck, Germany) assays, a single 4.76mm disk was punched and incubated in 250μL of ELISA buffer (kit component) for one hour at 37°C. The eluted solution was then pipette mixed thoroughly and 100uL was taken forward for analysis as per the manufacturer's instructions.

Table S1. Age distribution of patients sampled

Age	n=
All	54
5-12	4
13-18	4
19-40	29
41-60	12
>60	5