Reduced serum neutralization capacity against SARS-CoV-2 variants in a multiplex

ACE2 RBD competition assay

Daniel Junker[1], Alex Dulovic[1], Matthias Becker[1], Teresa R. Wagner[1,2], Philipp D. Kaiser[1],

Bjoern Traenkle[1], Ulrich Rothbauer[1,2], Katharina Kienzle[3], Stefanie Bunk[3], Carlotta Strümper[3],

Helene Häberle[3], Kristina Schmauder[4,5], Nisar Malek[3,6] ,Karina Althaus[7], Michael Koeppen[8],

Michael Bitzer[3,6], Siri Göpel[3,5]*, Nicole Schneiderhan-Marra[1]*

Author Affiliations

[1] NMI Natural and Medical Sciences Institute at the University of Tübingen, Reutlingen,

Germany

[2] Pharmaceutical Biotechnology, Eberhard Karls University, Tübingen, Germany

[3] Department Internal Medicine I, Eberhard Karls University, Tübingen, Germany

[4] Institute for medical microbiology and hygiene, University Hospital Tübingen, Germany

[5] German Centre for Infection Research (DZIF), Partner Site Tübingen, Tübingen, Germany

[6] Center for Personalized Medicine, Eberhard-Karls University, Tübingen, Germany

[7] Institute for Clinical and Experimental Transfusion Medicine, University Hospital Tübingen,

Tübingen, Germany

[8] Department of Anaesthesiology and Intensive Care Medicine, University Hospital Tübingen,

Tübingen, Germany

* Denotes shared senior authorship and corresponding authors.

Contact Information

Nicole Schneiderhan-Marra – Phone number +49 (0)7121 51530 815. Email Address

Nicole.schneiderhan@nmi.de Postal Address – Markwiesenstrasse 55, 72770 Reutlingen,

Germany.

Siri Göpel – Phone number +49 (0)7071 29 85415. Email Address

siri.goepel@med.uni-tuebingen.de Postal Address – Otfried-Müller-Strasse 10, 72076

Tübingen, Germany.

Abstract

As global vaccination campaigns against SARS-CoV-2 proceed, there is emerging interest in the longevity of immune protection, especially with regard to increasingly infectious virus variants. Neutralizing antibodies (Nabs) targeting the receptor binding domain (RBD) of SARS-CoV-2 are promising correlates of protective immunity and have been successfully used for prevention and therapy. To assess neutralizing capacity, we developed a bead-based multiplex ACE2 RBD competition assay as a large scalable, time-, cost-, and material-saving alternative to infectious live-virus neutralization tests. By mimicking the interaction between ACE2 and RBD, this assay detects the presence of Nabs against SARS-CoV2 in serum. Using this multiplex approach allows the simultaneous analysis of Nabs against all SARS-CoV-2 variants of concern and variants of interest in a single well. Following validation, we analyzed 325 serum samples from 186 COVID-19 patients of varying severity. Neutralization capacity was reduced for all variants examined compared to wild-type, especially for those displaying the E484K mutation. The neutralizing immune response itself, while highly individualistic, positively correlates with IgG levels. Neutralization capacity also correlated with disease severity up to WHO grade 7, after which it reduced.

Introduction

Neutralizing antibodies (Nabs) prevent infection of the cell with pathogens or foreign particles by neutralizing them, eliminating a potential threat and rendering the pathogen or particle harmless (1). The longevity of a Nab response has important implications for immune protection and vaccination strategies. In SARS-CoV-2, Nabs interfere with the cell entry mechanism primarily by blocking the interaction of the receptor binding domain (RBD) with the human cell receptor angiotensin converting enzyme 2 (ACE2) (2, 3). The RBD of SARS-CoV-2 is the target of approximately 90 % of the neutralizing activity present in immune sera (4), with a lack of Nabs correlating with risk of fatal outcome (5, 6). Passive transfer of Nabs has been shown to provide protection from infection (7-9), with several Nabs drugs granted emergency use authorization by the U.S. Food and Drug Administration (10-13).

Since the first documented infections in Wuhan China (14), SARS-CoV-2 has continually evolved, with the emergence of global variants of concern (VOC) being of particular importance. As of this moment, the WHO currently lists the alpha (B.1.1.7)(15), beta (B.1.351)(16), gamma (P.1)(17) and delta (B.1.617.2)(18) strains as VOCs (19), in addition to further variants of interest (VOI) such as lambda (C.37)(20). The emergence and disappearance of variants and continual mutation of SARS-CoV-2 is of particular relevance for vaccine development, as all currently licensed vaccines (21-24) only elicit an immune response against the original B1 isolate (hereon referred to as "wild-type") spike protein (25). Several studies have already found that both convalescent and post-vaccinated sera have lower neutralization capacities against beta and gamma VOCs (26-28). Of particular concern are mutations on amino acid residue (aa) 484 (e.g. E484K), which seem to confer escape from vaccine control, with an additional mutation on aa 501 (e.g. N501Y) increasing this effect (29).

In order to lead development of new vaccines and safely lift social restrictions, definitive correlates of protective immunity are necessary (30). The gold standard for Nabs assessment are virus neutralization tests (VNT), however these require live infectious virions which must be handled in biosafety level 3 (BSL3) laboratories, as well as access to different variants of

SARS-CoV-2. In this study we developed and applied a multiplex ACE2 RBD competition

assay based upon the antibody-mediated inhibition of RBD-ACE2 binding. This automatable

assay enables simultaneous screening of serum samples for the presence of Nabs against an

unlimited number of VOCs/VOIs in a single well, making it a time-, material- and cost-effective

alternative to live VNTs or classical ELISAs. Following in-depth validation of the assay, we

analyzed the IgG antibody response and neutralizing capacity of 325 serum samples from 186

COVID-19 patients with mild to severe disease progression towards eleven different SARS-

CoV-2 variant RBDs including the alpha, beta, gamma and delta VOCs.

Results

ACE2 RBD competition assay validation

To investigate neutralizing capacity of serum from COVID-19 convalescent individuals, we

developed and established a high-throughput bead-based multiplex ACE2 RBD competition

assay. This assay mimics the ACE2 RBD interaction, detecting the presence of Nabs against

SARS-CoV-2. For this, we expressed RBDs of SARS-CoV-2 wild-type and 11 different variants

(alpha, beta, gamma, epsilon, eta, theta, kappa, delta, lambda, Cluster 5 and A.23.1) and

coupled them on spectrally distinct beads (see methods for details). All beads were then pooled

to produce a bead mix, which was incubated with patient serum diluted in assay buffer

containing biotinylated ACE2. RBD-bound ACE2 was then detected using phycoerythrin-

labeled streptavidin and MFI values were normalized to a control consisting of biotinylated

ACE2 without added serum.

Initially, we validated the ACE2 RBD competition assay to standard FDA bioanalytical

guidelines. To ensure reagent stability and confirm high assay precision, we measured a

sample set of five sera from vaccinated individuals across five independent experiments,

identifying that the results of the assay were highly reproducible with minimal variation (as

determined by coefficients of variation (%CVs)) (**Figure 1, Supplementary Table 3**). Intra-

assay precision was determined using 12 replicates of four samples on a single plate (**Figure**

1A), with %CVs not exceeding 5% for all RBDs (mean range 1.76 - 2.72%). Inter-assay

precision (**Figure 1B**) was tested with five samples measured in triplicates in four independent

experiments and did not exceed 7 % for any RBD (mean range 2.62 – 4.72%). Short-term

temperature stability of the ACE2 buffer was confirmed by incubation for three different

durations (2h, 4h, 24h) at two temperatures (4°C and 21 °C) (**Figure 1C**). As a control, freshly

prepared buffer was also analyzed. 98 % of %CVs were under 10 % indicating short-term

stability of the ready-to-use ACE2 buffer. We examined the stability of the biotinylated human

ACE2 stock solution in a freeze-thaw stability experiment (**Figure 1D**), finding no adverse

results (all %CVs under 20). Lastly, to investigate the effect of changes in ACE2 concentration

on the assay readout, we measured the sample set with a range of ACE2 concentrations (150 - 350 ng/mL, **Figure 1E**), with no change in neutralization capacity between different concentrations (mean range 3.54 – 6.49%).

To evaluate assay performance against commercially available products, we compared the ACE2 RBD competition assay with the FDA-approved and CE-certified NeutraLISA (Euroimmun) (**Figure 2**). Results between the assays were highly comparable (Pearson's correlation coefficient of 0.7954) (**Figure 2A**). When classifying samples as being either positive or negative, samples with a percentage reduction under 20% are considered negative for the NeutraLISA (31). As both assays detect bound ACE2, we implemented a similar cut-off for the ACE2 RBD competition assay. Overall, 31.7% of samples (59/186) were considered negative in both assays, while a further 54.3% (101/186) were considered positive in both (**Figure 2B**). Of the remaining samples, six (3.2 %) exceeded 20 % neutralization percentage only in the ACE2 RBD competition assay, while 20 (10.8 %) exceeded 20 % neutralization percentage in the NeutraLISA only.

Neutralization capacities are reduced for all investigated variant RBDs

Following validation, we analyzed neutralization capacities of 325 serum samples from 186 COVID-19 patients, including longitudinal samples from 44 donors against RBD wild-type and 11 variants (hereafter referred to as "RBD mutants") of SARS-CoV-2.

Variant-specific decreases in neutralization capacity compared to wild-type were seen for all RBD mutants (1.1-fold (A.23.1) to 19.7-fold (beta), **Figure 3**). Median neutralization percentage (57 %) was highest in wild-type and lowest (2.9 %) in beta. Among VOC RBD mutants, alpha had the lowest reduction in neutralization capacity (1.25-fold), followed by delta (1.82-fold) and gamma (7.64-fold). Beta had the largest reduction in neutralizing capacity of all variants analyzed (19.71-fold). For the VOIs, lambda had the lowest reduction (3.5 fold), followed by kappa (5.0 fold), eta (6.3 fold) and theta (10.0 fold). The former VOIs Cluster 5 (1.3 fold) and epsilon (1.7 fold) had minimal reductions.

Neutralization capacity and IgG antibody responses were positively correlated (all Pearsons

correlation coefficients above 0.70, **Figure 4**) with variant-specific differences still present. The

percentage of samples considered to be negative (<20% neutralization), followed the same

variant-specific order as **Figure 3**. For all RBD mutants, the increase in neutralization capacity

most commonly occurred once IgG RBD MFI levels exceeded 10,000. Notably, there was

individual variation among the samples, with some having high neutralizing capacity but

relatively low IgG responses. This was especially true for the RBD wild type and similar

performing RBD mutants such as A.23.1, alpha and Cluster 5. For RBD mutants eta, gamma,

theta and beta more than 75 % of all samples were considered negative, compared to 42.5 %

for wild-type (**Figure 4**).

Neutralization capacity decreases over time

We then examined longitudinal changes on neutralizing capacity and IgG responses using

samples from 44 study participants, ranging from 1 to 290 days after the first positive SARS-

CoV-2 PCR. For a subset of six individuals with similar sample collection points, both IgG

antibody response and neutralization capacity against wild-type (**Figure 5A, B**) remained

initially low following infection, with an increase no later than 10 days after the first positive

PCR test result, although one individual (Donor3) peaked 6 days after the first positive PCR

test and then declined. Neutralization capacity for most individuals (Donors1, 4, 5 and 6)

declined no later than 19 days post PCR. Generally, neutralization capacity and IgG levels

peaked at similar time points (Donors 1, 2 and 6), although for some individuals (Donors 4 and

5) neutralization capacity peaked before IgG titers. Among all longitudinal samples, while

neutralization and antibody titers followed a similar pattern (**Figure 5C, D**), responses were

highly individualistic with large variances at all time points.

As delta represents the current dominant global strain, we then examined how neutralization

capacity and antibody binding compared for this variant to wild-type. Neutralizing capacity and

IgG binding against delta was reduced for all samples (**Figure 5E and F**). Overall, neutralizing

capacity and IgG response followed the same pattern for all samples as for wild-type. When

examining all RBD mutants, mean neutralization capacity and IgG antibody responses (**Figure**

5G, H) clustered as before. Beta, theta and gamma had the largest decrease in neutralizing

capacity with eta, kappa and lambda occupying the gap between another cluster of delta,

epsilon, Cluster5 and alpha. As expected, all RBD mutants showed reduced neutralization

capacity compared to wild-type. Mean neutralization capacity was highly individualistic,

peaking for all RBD mutants 22 days post-positive PCR before decreasing below 20 % at

approximately 92 days post PCR.

Neutralization correlates with disease severity

We then examined correlations between neutralization capacity and COVID-19 disease

severity. Neutralizing capacities for wild-type and RBD delta within 7-49 days post-positive

PCR increased steadily with disease severity up to grade 7 (WHO grading scale, hospitalized

patients needing intubation and mechanical ventilation), before decreasing for patients of

grade 8 (fatal disease course) (**Figure 6A and B**). Mean neutralization capacities at WHO

grade 7 were 72.0 % for wild-type and 58.7 % for delta. From 49 days post positive PCR,

neutralization capacities correlated with disease severity (**Figure 6C and D**), however there

was an overall reduction for grades 4 to 7 compared to the first timeframe for both wild-type

and delta (**Figure 6C and D**). Within this timeframe, mean neutralization capacity for both WT

(43.56%) and delta (30.08%) decreased. Anti-RBD IgG levels also correlated with disease

severity 7 – 49 days post PCR (**Figure 6E, F**) as well as later than 49 days post PCR (**Figure**

6G, H) for both wild-type and delta. Peak IgG levels were at grade 7 severity for both wild-type

and delta 7 – 49 days post PCR. Post 49 days, mean IgG levels peaked for patients with grade

6 severity.

Neutralization correlated with both age and BMI from 49 days post-PCR onwards

Finally, to check for confounding variables, samples taken either 7 to 49 days or more than 49

days post-positive PCR were analyzed by their age, gender and BMI (**Supplementary Figure**

1). Whereas no correlation was observed between neutralization capacity and gender for both

timeframes, we found correlations between neutralization capacity and donor age (p = 0.001)

along with BMI (p = 0.0179) for samples drawn more than 49 days after the first positive PCR

test (**Supplementary Figure 1D, F**).

Discussion

With vaccination campaigns beginning to take effect and result in reduced case numbers in certain countries, interest in the serological response to SARS-CoV-2 is increasingly changing from quantitative to qualitative in nature. Whereas in the early-phase of the pandemic SARS-CoV-2 antibody assays were helpful in determining seroprevalence and support vaccine development, now a reliable correlate of immune protection is needed to securely lift social restrictions and guide future vaccine developments.

While cell-culture based virus neutralization tests (VNTs) (e.g. plaque reduction neutralization test) are the gold standard for neutralization assays, they have many disadvantages over conventional protein-based surrogate assays. Such assays require rapid access to continually changing virus variants and as such special biosafety level 3 laboratories are necessary. Additionally, VNTs are cell-culture based and therefore it takes multiple days to conduct an experiment with reproducibility potentially affected by either the cells or their long culture conditions. Consequently, highly reproducible assays under substantially faster and safer working conditions (e.g. BSL 1) would be highly beneficial. Our newly developed protein-based ACE2 RBD competition assay is finished in under 4 hours and only requires 5 µL of patient sample to measure neutralizing capacities simultaneously against multiple SARS-CoV-2 VOCs and VOIs. As a protein-based assay, it does not require enhanced safety protocols to be followed and can be completed safely in a BSL1 laboratory. Due to the bead-based nature and plate format, it is automatable, suitable for high-throughput, standardized and highly reproducible. The protein-based nature also allows for the rapid inclusion of emerging variants or single mutations.

As shown in our technical validation, assay components and as such the results are highly stable. Neutralization percentages correlate with FDA-approved in vitro diagnostic tests such as NeutraLISA (Euroimmun). The lower than 20% reduction cut-off used in the NeutraLISA to identify negative samples is also suitable for the ACE2 RBD competition assay as both measure the steric inhibition of ACE2. Direct comparison between the two assays shows that

the ACE2 RBD competition assay has a resolution range more suited to highly neutralizing

samples, while the NeutraLISA has increased resolution for samples with weak neutralization

at the cost of early saturation of neutralization percentage. Samples which had different

classifications between the two assays had a median MFI of 9056, which explains the low

neutralization capacities in the ACE2 RBD competition assay, as high neutralization capacities

are generally achieved at an MFI > 10,000.

Similarly to other authors (32, 33), we identified a significant positive correlation between anti-

RBD IgG levels and neutralization capacity. This suggests that neutralizing antibodies

represent a consistent portion of all antibodies produced. Overall, our results also show there

is a high degree of individualism in responses, with some samples having low titers yet high

neutralization capacity for specific RBD mutants. Neutralization itself followed a typical pattern

for acute infections with an initial rapid increase which peaks and then declines (34). Notably,

neutralizing capacity increased and decreased in parallel to IgG responses, which could have

important implications on the immune protection induced by prior infection or vaccination, and

should be considered when organizing booster vaccination campaigns. Like other authors (6,

35) we identified a correlation between disease severity and neutralizing capacity, however

the decrease in IgG levels and neutralizing capacity of patients with WHO disease grade 8

(death) has not to our knowledge been reported before. This decrease requires further

investigation to determine its cause given its likely role in patient mortality.

As expected, neutralizing capacity towards VOCs was highly variable. The strongest

reductions in neutralizing capacities were all from variants (eta, gamma, theta and beta) with

a E484K mutation. This specific mutation has been reported in multiple studies as an escape

mutation that enhances the RBD-ACE2 affinity (36). Neutralizing capacity was further reduced

among these variants for those which additionally had a N501Y mutation (gamma, theta and

beta), which is known to further enhance RBD-ACE2 binding (37). These results are in-line

with previous findings that have reported significant reductions in neutralization for gamma and

beta (38-40). The gamma and beta RBDs in our assays are only separated by a single K417N

mutation, which is known to significantly reduce both the RBD-ACE2 binding affinity as well as

the binding affinity to monoclonal therapeutic antibodies or other human antibodies (41). This

reduction in binding affinity could explain why beta neutralizing capacity was lower than

gamma. Among recently emerged variants (delta, kappa and lambda), neutralizing capacity

compared to wild-type was reduced for all. The reduction in neutralizing capacity seen for

kappa and delta are comparable to recent findings (42), although we could not confirm the

reduction seen by other authors for Lambda (43). This is likely due to the 7-amino acid deletion

in the N-terminal domain of lambda's spike protein, which is not present in the RBD and is

thought to contribute to its immune evading properties (44). Overall the decrease in

neutralization capacity against RBDs of all analyzed variants compared to wild-type has

important implications for the design of second generation vaccines.

Our ACE2 RBD competition assay has limitations similar to other protein-based in vitro

neutralization assays, such as only accounting for the Nabs that block the RBD-ACE2

interaction site through steric hindrance, and not for Nabs that interfere with cell entry

mechanisms as would be analyzed in a VNT. Furthermore, the binding assay is also more

prone to non-specific binding events. However, a major advantage of the multiplex ACE2 RBD

competition assay over VNTs is the speed of response toward viral evolution such as emerging

variants of concern. The bead-based format of the assay is also highly reproducible and not

susceptible to changes in experimental conditions, as is the case for cell culture based VNTs.

The plate format of the assay also enables automation and high-throughput screening. Our

ACE2 RBD competition assay only requires recombinant expressed RBD proteins which can

be quickly and easily produced. Additionally, this assay has the possibility of introducing

artificial mutants to screen for possible escape variants that could arise in the future. Similar

to other neutralization studies, our study is limited by the availability of appropriate control

samples. Additionally, as the majority of this study population were admitted to the intensive

care unit, the more serious grades of COVID-19 infection are heavily overrepresented in our

population. Our samples are also highly variable in their longitudinal nature, with no consistent

time points. However, this large variation is also beneficial as it clearly demonstrates the

individual variability in neutralizing capacity.

In conclusion, we have developed and validated an ACE2 RBD competition assay that

analyzes neutralizing antibodies in serum for all current variants of concern and interest of

SARS-CoV-2. Using this assay, we have identified differing responses in serum from infected

individuals, indicating that neutralizing antibody production is directly proportional to total

antibody production. Neutralization capacity was highly variable among all variants examined,

with the 484 residue appearing to be critical in reducing capacity. Overall, the speed and ease

of incorporating new variants into the assay makes it ideal for screening how neutralization

capacity changes in emerging variants.

Methods

Sample collection and ethics statement

325 serum samples were collected from 186 patients hospitalized at the University Hospital Tübingen, Germany between April 17, 2020 and May 12, 2021. Longitudinal samples were available from 44 of the 186 patients ranging from 2 to 13 samples per patient. All individuals were tested positive by SARS-CoV-2 PCR. Sample collection and execution of this study was approved by the Ethics committee of the Eberhard Karls University Tübingen and the University Hospital Tübingen under the ethical approval numbers 188/2020A and 764/2020BO2 to Prof. Dr. Michael Bitzer. All participants signed the broad consent of the Medical Faculty Tübingen for sample collection.

For serum collection, blood was extracted by venipuncture, with the serum blood collection tube rotated 180° two to three times to extract possible air bubbles in the sample. After a minimum coagulation time of 30 minutes at room temperature, serum was extracted by centrifugation for 15 minutes at 2000 x g (RT) and then stored at -80 °C until analysis. Time between blood sampling and centrifugation did not exceed 2 hours.

For assay validation, negative pre-pandemic serum samples were purchased from Central BioHub and four previously collected vaccinated samples from healthcare workers vaccinated with the Pfizer BNT-162b2 vaccine (28) were used. The sample collection was approved by the Ethics committee of the Eberhard Karls University Tübingen and the University Hospital Tübingen under the ethical approval number 222/2020BO2 to Dr. Karina Althaus. For all assay validation samples, written informed consent was obtained. Sample characteristics are summarized in **Supplementary Table 2**.

Expression and Purification of SARS-CoV-2 RBD mutants

The expression plasmid pCAGGS, encoding the receptor-binding domain (RBD) of SARS-CoV-2 spike protein (amino acids 319-541), was kindly provided by F. Krammer (45).

Expression and purification of VOCs alpha, beta and epsilon was carried out as previously

described (28). RBDs of SARS-CoV-2 VOCs gamma, delta, eta, theta, kappa and VOI A.23.1

were generated by PCR amplification of fragments from wild type or cognate DNA templates

and subsequent fusion PCR by overlap extension to introduce described mutations. Based on

RBD   wild   type   sequence,   primer   pairs   RBDfor

(ATATCTAGAGCCACCATGTTCGTGTTTCTGG),   E484Krev

(GCAGTTGAAGCCTTTCACGCCGTTACAAGGGGT)   and   E484Kfor,   RBDrev

(AAGATCTGCTAGCTCGAGTCGC)   for   VOC   eta   and   RBDfor,   V367Frev

(CGGAGTTGTACAGGAAGGAGTAGTCGGCCACGCA)   and   V367Ffor,   RBDrev   for   VOI

A.23.1 were used. VOC lambda was generated based on RBD wild type sequence using

primer   pairs   L452Q-for   (GGCAACTACAATTACCAGTACCGGCTGTTCCGGAAG)   and

L452Q-rev   (CGGTACTGGTAATTGTAGTTGCCGCCG),   and   F490S-for

(TCAACTGCTACTCCCCACTGCAGTCCTACGGC)   and   Lambda-F490S-rev

(CTGCAGTGGGGAGTAGCAGTTGAAGCCTTCCAC). VOC delta was generated based on

VOC   epsilon   using   primer   pairs   RBDfor,   T478Krev

(CGTTACAAGGctTGCTGCCGGCCTGATAGA)   and   T478Kfor

(CCGGCAGCAAGCCTTGTAACGGCGTGGAAG), RBDrev. Based on VOC alpha sequence,

VOC theta was generated using primer pairs RBDfor, E484Krev and E484Kfor, RBDrev. VOC

kappa was generated based on VOC eta sequence using primer pairs RBDfor, L452Rrev

(CGGTACCGGTAATTGTAGTTGCCGCCG)   and   L452Rfor

(GGCAACTACAATTACCGGTACCGGCTGTTCCGGAAG),   RBDrev.   VOC   gamma   was

generated based on VOC theta sequence using primer pairs RBDfor, K417Trev

(GTTGTAGTCGGCGATGGTGCCTGTCTGTCCAGGGG)   and   K417Tfor

(GACAGACAGGCACCATCGCCGACTACAACTACAAG),   RBDrev.   Amplificates   were

inserted into the pCDNA3.4 expression vector using XbaI and NotI restriction sites. The

integrity of all expression constructs was confirmed by standard sequencing analysis.

Confirmed constructs were expressed in Expi293 cells (28, 46). Briefly, cells were cultivated

(37 °C, 125 rpm, 8% (v/v) CO_2) to a density of 5.5×10^6 cells/mL, diluted with Expi293F

expression medium and transfection of the corresponding plasmids (1 µg/mL) with

expifectamine was performed as per the manufacturer's instructions. 20 h post transfection,

enhancers were added as per the manufacturer's instructions. Cell suspensions were then

cultivated for 2–5 days (37 °C, 125 rpm, 8 % (v/v) CO_2) and then centrifuged (4 °C, 23,900×g,

20 min) to clarify the supernatant. Supernatants were then filtered with a 0.22-µm membrane

(Millipore, Darmstadt, Germany) and supplemented with His-A buffer stock solution (20 mM

Na2HPO4, 300 mM NaCl, 20 mM imidazole, pH 7.4). The solution was then applied to a

HisTrap FF crude column on an Äkta pure system (GE Healthcare, Freiburg, Germany),

extensively washed with His-A buffer, and eluted with an imidazole gradient (50–400 mM).

Buffer exchange to PBS and concentration of eluted proteins were carried out using Amicon

10K centrifugal filter units (Millipore, Darmstadt, Germany).

Bead coupling

The in-house expressed RBD mutants were immobilized on magnetic MagPlex beads

(Luminex) using the AMG Activation Kit for Multiplex Microspheres (# A-LMPAKMM-400,

Anteo Technologies). In brief, 1 mL of spectrally distinct MagPlex beads (1.25 *10^7 beads)

were activated in 1 mL of AnteoBind Activation Reagent for 1 hour at room temperature. The

beads were washed twice with 1 mL of Conjugation buffer using a magnetic separator, before

being resuspended in 1 mL of antigen solution diluted to 25 µg/mL in Conjugation buffer. After

1 h incubation at room temperature the beads were washed twice with 1 mL Conjugation buffer

and incubated for 1 h in 0.1 % (w/v) BSA in Conjugation buffer for blocking. Following this, the

beads were washed twice with 1 mL storage buffer. Finally, the beads were resuspended in 1

mL storage buffer and stored at 4°C until further use.

ACE2 RBD competition assay

MULTICOV-AB Assay buffer (1:4 Low Cross Buffer (Candor Bioscience GmbH) in CBS (1x

PBS + 1% BSA) + 0.05 % Tween20) was supplemented with biotinylated human ACE2 (Sino

Biological, # 10108-H08H-B) to a final concentration of 571.4 ng/mL (ACE2 buffer). Working

inside a sterile laminar flow cabinet, serum samples were thawed and diluted 1:25 in assay

buffer before being further diluted 1:8 in ACE2 buffer resulting in a final concentration of 500

ng/mL ACE2 in all 1:200 diluted samples. Spectrally distinct populations of MagPlex beads

(Luminex) coupled with RBD proteins of SARS-CoV-2 wild type and variants alpha, beta,

gamma, epsilon, eta, theta, kappa, delta, lambda, Cluster 5 and A.23.1 were pooled in assay

buffer (1:4 Low Cross Buffer (Candor Bioscience GmbH) in CBS (1x PBS + 1% BSA) + 0.05%

Tween20) to create a 1x bead mix (40 beads/µL per bead population). 25 µL of diluted serum

was added to 25 µL of bead mix in each well of a 96-well plate (Corning, #3642). As

normalization control, 500 ng/mL ACE2 was used and triplicates of one quality control sample

were analyzed on every plate. Samples were incubated for 2h at 21°C while shaking at 750

rpm on a thermomixer. Following incubation, the beads were washed three times with 100 µL

wash buffer (1x PBS + 0.05 Tween20) using a microplate washer (Biotek 405TS, Biotek

Instruments GmbH). For detection of bound biotinylated ACE2, 30 µL of 2 µg/mL RPE-

Streptavidin was added to each well and incubated for 45 min at 21 °C while shaking at 750

rpm on a thermomixer. Afterwards, the beads were washed again thrice with 100 µL wash

buffer. The 96-well plate was placed for 3 min on the thermomixer at 1000 rpm to resuspend

the beads before analysis using a FLEXMAP 3D instrument with the following settings: 80 µL

(no timeout), 50 events, Gate: 7,500 – 15,000, Reporter Gain: Standard PMT. MFI values of

each sample were divided by the mean of the ACE2 normalization control. The normalized

values were converted into percent and subtracted from 100 resulting in the percentage of

neutralization. Negative neutralization percentages were set to zero.

MULTICOV-AB

MULTICOV-AB (46), an in-house produced SARS-CoV-2 antibody assay, was performed with

all serum samples to measure RBD-specific IgG levels. The antigen panel was expanded to

include RBD proteins from 11 different SARS-CoV-2 variants from which all, except the Cluster

5 variant from Sino Biological (# 40592-V08H80), were produced in-house. Briefly, spectrally

distinct populations of MagPlex beads (Luminex) coupled with RBD proteins of SARS-CoV-2

wild type and variants were pooled in assay buffer (1:4 Low Cross Buffer (Candor Bioscience

GmbH) in CBS (1x PBS + 1% BSA) + 0.05% Tween20) to create a bead mix (40 beads/µL per

bead population). Samples were diluted 1:200 in assay buffer, working inside a sterile laminar

flow cabinet. 25 µL of diluted sample were added to 25 µL of bead mix in a 96-well plate

(Corning, #3642) resulting in a total dilution of 1:400. Samples incubated for 2 h at 21°C while

shaking at 750 rpm on a thermomixer. After incubation, the beads were washed three times

with 100 µL wash buffer (1x PBS + 0.05 % Tween20) using a microplate washer (Biotek 405TS,

Biotek Instruments GmbH). Detection of bound antibodies was performed by incubating the

beads with 3 µg/mL RPE-huIgG (Dianova, # 109-116-098) for 45 min at 21 °C and 750 rpm on

a thermomixer. Following three more washing cycles with 100 µL wash buffer, the beads were

resuspended for 3 min at 21°C and 1000 rpm. For readout, a FLEXMAP 3D instrument

(Luminex) was used with the following settings: 80 µL (no timeout), 50 events, Gate: 7,500 –

15,000, Reporter Gain: Standard PMT.

408   Assay validation

409   Intra- and inter-assay precision

410   To determine the intra-assay precision of the ACE2 RBD competition assay, 12 replicates of

411   four serum samples (Vac1 – Vac4) were measured on a 96-well plate (Corning, #3642).

412   Additionally, 15 replicates of the 500 ng/mL ACE2 control and 12 replicates of a blank control

413   containing only assay buffer without sample or ACE2 were measured. For analysis, the mean,

414   standard deviation and coefficient of variation in percent of all replicates were calculated.

For inter-assay precision, four serum samples (Vac1 – Vac4) were measured in triplicates in four independent experiments. Additionally, the quality control, the ACE2 normalization control and blank were also processed in triplicates in the same four experiments. For analysis, the mean, standard deviation and coefficient of variation in percent of all replicates were calculated.

Freeze-thaw stability

Freeze-thaw stability of the biotinylated ACE2 stocks was determined by analyzing five serum samples (Vac1 – Vac4 (vaccinated) and pre-pandemic) in triplicates, with ACE2 stocks undergoing 1 to 6 freeze-thaw cycles. In addition to that, every sample was also processed with ACE2 not re-frozen once thawed (fresh). The MFI values of every sample were normalized to the values of the respective ACE2 control. For analysis, the mean, standard deviation and coefficient of variation in percent of all replicates were calculated.

Short-term stability

Short-term stability was determined by storing ACE2 buffer under six different conditions before proceeding with the assay protocol. The prepared ACE2 buffer was stored 2 h, 4 h and 24 h at both 4°C and room temperature and compared to ACE2 buffer without storage (fresh). Replicate MFI values of every sample (Vac1 – Vac4 (vaccinated) and pre-pandemic) were normalized to the values of the respective ACE2 control. For analysis, the mean, standard deviation and coefficient of variation in percent of all replicates were calculated.

Parallelism

To investigate the stability of the ACE2 RBD competition assay against variations of the used ACE2 concentration, five samples (Vac1 – Vac4 (vaccinated) and pre-pandemic) were analyzed with ACE2 concentrations ranging from 150 ng/mL to 350 ng/mL. Replicate MFI values of every sample were normalized to the values of the respective ACE2 control. For analysis, the mean, standard deviation and coefficient of variation in percent of all replicates were calculated.

Commercial assay formats

For further validation, one sample from each individual (n=186) were analyzed with the

commercially available in-vitro diagnostic test SARS-CoV-2 NeutraLISA (Euroimmun). The

assay was performed according to the manufacturer's instructions. For longitudinal donors with

more than one sample available, the sample closest to 20 days after positive PCR diagnosis

was picked. Negative neutralization percentages were set to zero.

Data analysis

Data collection and assignment to metadata was performed with Microsoft Excel 2016. Data

analysis and visualization was performed with Graphpad Prism (version 9.1.2). Longitudinal

curves were fitted using a one-site total binding equation. Correlations were analyzed using

Pearson's correlation coefficient. Figures were edited with Inkscape (version 0.92.4). Data

generated for this manuscript is available from the authors upon request.

References

1.   Forthal DN. Functions of Antibodies. *Microbiology spectrum.* 2014;2(4):Aid-0019-
2014.
2.   Jiang S, et al. Neutralizing antibodies for the treatment of COVID-19. *Nature*
biomedical engineering. 2020;4(12):1134-9.
3.   Shi R, et al. A human neutralizing antibody targets the receptor-binding site of SARS-
CoV-2. *Nature.* 2020;584(7819):120-4.
4.   Piccoli L, et al. Mapping Neutralizing and Immunodominant Sites on the SARS-CoV-2
Spike Receptor-Binding Domain by Structure-Guided High-Resolution Serology. *Cell.*
2020;183(4):1024-42.e21.
5.   Dispinseri S, et al. Neutralizing antibody responses to SARS-CoV-2 in symptomatic
COVID-19 is persistent and critical for survival. *Nature communications.*
2021;12(1):2670.
6.   Garcia-Beltran WF, et al. COVID-19-neutralizing antibodies predict disease severity
and survival. *Cell.* 2021;184(2):476-88.e11.
7.   McMahan K, et al. Correlates of protection against SARS-CoV-2 in rhesus macaques.
Nature. 2021;590(7847):630-4.
8.   Kim YI, et al. Critical role of neutralizing antibody for SARS-CoV-2 reinfection and
transmission. *Emerging microbes & infections.* 2021;10(1):152-60.
9.   Rogers TF, et al. Isolation of potent SARS-CoV-2 neutralizing antibodies and
protection from disease in a small animal model. *Science (New York, NY).*
2020;369(6506):956-63.
10. Weinreich DM, et al. REGN-COV2, a Neutralizing Antibody Cocktail, in Outpatients
with Covid-19. *The New England journal of medicine.* 2021;384(3):238-51.
11. (FDA) FaDA. Coronavirus (COVID-19) Update: FDA Authorizes Monoclonal
Antibodies for Treatment of COVID-19. https://www.fda.gov/news-events/press-
announcements/coronavirus-covid-19-update-fda-authorizes-monoclonal-antibodies-
treatment-covid-19. Accessed 23.07.2021.
12. Gottlieb RL, et al. Effect of Bamlanivimab as Monotherapy or in Combination With
Etesevimab on Viral Load in Patients With Mild to Moderate COVID-19: A
Randomized Clinical Trial. *Jama.* 2021;325(7):632-44.
13. (FDA) FaDA. Coronavirus (COVID-19) Update: FDA Authorizes Monoclonal
Antibodies for Treatment of COVID-19. https://www.fda.gov/news-events/press-
announcements/coronavirus-covid-19-update-fda-authorizes-monoclonal-antibodies-
treatment-covid-19-0. Accessed 23.07.2021.
14. Zhou P, et al. A pneumonia outbreak associated with a new coronavirus of probable
bat origin. *Nature.* 2020:1-4.

15.   Graham MS, et al. Changes in symptomatology, reinfection, and transmissibility
associated with the SARS-CoV-2 variant B.1.1.7: an ecological study. *The Lancet*
Public health. 2021;6(5):e335-e45.
16.   Tegally H, et al. Detection of a SARS-CoV-2 variant of concern in South Africa.
Nature. 2021;592(7854):438-43.
17.   Sabino EC, et al. Resurgence of COVID-19 in Manaus, Brazil, despite high
seroprevalence. *Lancet (London, England).* 2021;397(10273):452-5.
18.   Control ECfDPa. Threat Assessment Brief: Emergence of SARS-CoV-2 B.1.617
variants in India and situation in the EU/EEA.
https://www.ecdc.europa.eu/en/publications-data/threat-assessment-emergence-sars-
cov-2-b1617-variants.
19.   (WHO) WHO. Tracking SARS-CoV-2 variants.
https://www.who.int/en/activities/tracking-SARS-CoV-2-variants/. Updated 06.07.2021
Accessed 11.07.2021.
20.   O'Toole A, and Hill V. Lineage C.37. https://cov-
lineages.org/lineage.html?lineage=C.37.
21.   Skowronski DM, and De Serres G. Safety and Efficacy of the BNT162b2 mRNA
Covid-19 Vaccine. *The New England journal of medicine.* 2021;384(16):1576-7.
22.   Baden LR, et al. Efficacy and Safety of the mRNA-1273 SARS-CoV-2 Vaccine. *The*
New England journal of medicine. 2021;384(5):403-16.
23.   Voysey M, et al. Safety and efficacy of the ChAdOx1 nCoV-19 vaccine (AZD1222)
against SARS-CoV-2: an interim analysis of four randomised controlled trials in Brazil,
South Africa, and the UK. *Lancet (London, England).* 2021;397(10269):99-111.
24.   Sadoff J, et al. Safety and Efficacy of Single-Dose Ad26.COV2.S Vaccine against
Covid-19. *The New England journal of medicine.* 2021;384(23):2187-201.
25.   Krammer F. SARS-CoV-2 vaccines in development. *Nature.* 2020;586(7830):516-27.
26.   Jalkanen P, et al. COVID-19 mRNA vaccine induced antibody responses against
three SARS-CoV-2 variants. *Nature communications.* 2021;12(1):3991.
27.   Shen X, et al. Neutralization of SARS-CoV-2 Variants B.1.429 and B.1.351. *The New*
England journal of medicine. 2021;384(24):2352-4.
28.   Becker M, et al. Immune response to SARS-CoV-2 variants of concern in vaccinated
individuals. *Nature communications.* 2021;12(1):3109.
29.   Hoffmann M, et al. SARS-CoV-2 variants B.1.351 and P.1 escape from neutralizing
antibodies. *Cell.* 2021;184(9):2384-93.e12.
30.   Krammer F. A correlate of protection for SARS-CoV-2 vaccines is urgently needed.
Nature medicine. 2021.

31.   Euroimmun. SARS-CoV-2 NeutraLISA: Produkt-Datenblatt. https://www.coronavirus-
diagnostik.de/documents/Indications/Infections/Coronavirus/EI_2606_D_DE_F.pdf.
Accessed 19.08.2021.

32.   Mendrone-Junior A, et al. Correlation between SARS-COV-2 antibody screening by
immunoassay and neutralizing antibody testing. *Transfusion.* 2021;61(4):1181-90.

33.   Grenache DG, et al. Correlation of SARS-CoV-2 Neutralizing Antibodies to an
Automated Chemiluminescent Serological Immunoassay. *The journal of applied*
laboratory medicine. 2021;6(2):491-5.

34.   Seow J, et al. Longitudinal observation and decline of neutralizing antibody
responses in the three months following SARS-CoV-2 infection in humans. *Nature*
microbiology. 2020;5(12):1598-607.

35.   Legros V, et al. A longitudinal study of SARS-CoV-2-infected patients reveals a high
correlation between neutralizing antibodies and COVID-19 severity. *Cellular &*
molecular immunology. 2021;18(2):318-27.

36.   Nelson G, et al. Molecular dynamic simulation reveals E484K mutation enhances
spike RBD-ACE2 affinity and the combination of E484K, K417N and N501Y
mutations (501Y.V2 variant) induces conformational change greater than N501Y
mutant alone, potentially resulting in an escape mutant. *bioRxiv.*
2021:2021.01.13.426558.

37.   Luan B, et al. Enhanced binding of the N501Y-mutated SARS-CoV-2 spike protein to
the human ACE2 receptor: insights from molecular dynamics simulations. *FEBS*
letters. 2021;595(10):1454-61.

38.   Shen X, et al. SARS-CoV-2 variant B.1.1.7 is susceptible to neutralizing antibodies
elicited by ancestral spike vaccines. *Cell host & microbe.* 2021;29(4):529-39.e3.

39.   Wall EC, et al. Neutralising antibody activity against SARS-CoV-2 VOCs B.1.617.2
and B.1.351 by BNT162b2 vaccination. *Lancet (London, England).*
2021;397(10292):2331-3.

40.   Wang P, et al. Antibody resistance of SARS-CoV-2 variants B.1.351 and B.1.1.7.
Nature. 2021;593(7857):130-5.

41.   Luan B, and Huynh T. Insights into SARS-CoV-2's Mutations for Evading Human
Antibodies: Sacrifice and Survival. *Journal of medicinal chemistry.* 2021.

42.   Edara VV, et al. Infection and Vaccine-Induced Neutralizing-Antibody Responses to
the SARS-CoV-2 B.1.617 Variants. *The New England journal of medicine.*
2021;385(7):664-6.

43.   Kimura I, et al. SARS-CoV-2 Lambda variant exhibits higher infectivity and immune
resistance. *bioRxiv.* 2021:2021.07.28.454085.

44. Acevedo ML, et al. Infectivity and immune escape of the new SARS-CoV-2 variant of interest Lambda. *medRxiv.* 2021:2021.06.28.21259673.

45. Amanat F, et al. A serological assay to detect SARS-CoV-2 seroconversion in humans. *Nature medicine.* 2020;26(7):1033-6.

46. Becker M, et al. Exploring beyond clinical routine SARS-CoV-2 serology using MultiCoV-Ab to evaluate endemic coronavirus cross-reactivity. *Nature communications.* 2021;12(1):1152.

Competing interests

NSM was a speaker at Luminex user meetings in the past. The Natural and Medical Sciences Institute at the University of Tübingen is involved in applied research projects as a fee for services with the Luminex Corporation. The other authors declare no competing interest.

Funding

This work was financially supported by the State Ministry of Baden-Württemberg for Economic Affairs, Labour and Housing Construction (grant number FKZ 3-4332.62-NMI-68).

Acknowledgements

We thank Johanna Griesbaum, Jennifer Jüngling and Christine Geisler for excellent technical assistance.

Author Contributions

DJ and AD conceived the study and designed the experiments. NSM supervised the study. DJ and MaB performed the experiments. DJ, AD, MaB produced components for the multiplex ACE2 RBD competition assay. KK, SB, CS, HH, KS, NM, KA, MK, MB and SG

589   collected samples or organized their collection and provided clinical metadata. PDK, BT, TW

590   and UR produced and designed recombinant assay proteins. MB, SG and NSM procured

591   funding. DJ performed data analysis and generated the figures. DJ and AD wrote the

592   manuscript. All authors critically reviewed and approved the final manuscript.

Figures

Figure 1 | ACE2 RBD competition assay: Technical validation results. Results of intra-assay precision **(A)**, inter-assay precision **(B)**, short-term stability **(C)**, freeze-thaw stability **(D)** and parallelism **(E)** experiments analyzing neutralization capacity (displayed as %) using wild-type (WT) RBD. Four samples from donors vaccinated with Pfizer BNT-162b2 (n=4, black) and one pre-pandemic sample (n=1, grey) were analyzed. Data points of each sample are illustrated by different shapes according to the figure key. Percent coefficients of variation (%CV) for all included RBD mutants are summarized in **Supplementary table 3**.

A

NeutraLISA (% Inhibition)	ACE2 RBD Competition Assay (% Neutralization)		
		≤ 20 %	> 20%
	≤ 20 %	59 (31.7 %)	6 (3.2 %)
	> 20 %	20 (10.8 %)	101 (54.3 %)
Pearson r [95% CI]	0.7954 [0.7357; 0.8428]		

Figure 2 | Correlation between SARS-CoV-2 NeutraLISA and the multiplex ACE2 RBD competition assay. One sample from each individual (n=186) was measured using both assays and analyzed by linear regression. Samples were classified as being negative (non-neutralizing) if they had a value below 20% (red lines). **(A)** Descriptive statistics of the **(B)** linear regression, showing the distribution of samples as positive or negative for both assays. Correlation analysis was performed after Pearson and the correlation coefficient r is shown in the table with the 95 % confidence interval (CI).

Figure 3 | Neutralization capacity varies between RBD mutants. Violin plots showing neutralization capacities (%) of individual serum samples from 7 to 49 days post PCR (n=59, depicted as dots) against RBD mutants. Black horizontal lines represent medians. Fold-decrease of neutralization capacity in comparison to wild-type corresponds to the ratio between the medians of wild-type and the respective RBD mutant. RBDs of VOC are shown in blue. An overview of the mutations present in all RBD mutants is shown in **Supplementary Table 1**.

Figure 4 | Correlation between anti-RBD IgG MFI signals and neutralization capacities of serum samples from COVID-19 patients for wild-type and 11 RBD mutants. Regression analysis comparing neutralizing capacity (%) and IgG responses (MFI) for wild-type and all RBD mutants included in the study. Each circle represents one sample (n=186). For longitudinal donors with more than one sample available, the sample closest to 20 days post positive PCR diagnosis was selected. The percentage next to the bracket indicates the proportion of samples with neutralization capacity ≤ 20% (in orange). Pearson's correlation coefficient (r) is specified for every correlation.

Figure 5 | Longitudinal analysis of anti-RBD IgG levels and neutralization capacities in Covid-19 patients. Neutralization capacity (%) and IgG responses (MFI) against time post positive PCR test for longitudinal samples of selected donors (n=6) for wild-type **(A, B)** and RBD delta **(E, F)**. The same analysis is then shown for all individual samples from 1 to 92 days post PCR (n=168) for wild-type **(C, D)**. Black dots indicate mean responses with standard deviation indicated by the error bars. Neutralization capacity (%) and IgG responses (MFI) 1 to 92 days post PCR for all RBD mutants included in the study **(G, H)**. Each variant is illustrated by a different color according to the figure key.

Figure 6 | Correlation of anti-RBD IgG levels and neutralization capacities with SARS-CoV-2 disease severity. Bar charts showing mean neutralization capacities (%) against wild-type and RBD delta are correlated with WHO grades for disease severity for samples 7-49 days post PCR **(A, B)** and later than 49 days post PCR **(C, D)**. Mean anti-RBD WT IgG and anti-RBD delta IgG levels are shown for samples 7-49 days post PCR **(E, F)** and later than 49 days post PCR **(G, H)**. Individual samples are displayed as colored dots, bars indicate the mean of the dataset with error bars representing standard deviation. Number of samples is given below the columns (n). If no samples for a group were available, the column is labeled with "n/a". WHO grade 1 - ambulatory / no limitations of activities, 2 - ambulatory /

limitation of activities, 3 - hospitalized, mild disease / no oxygen therapy, 4 - hospitalized, mild disease / mask or nasal prongs, 6 - hospitalized, severe disease / intubation + mechanical ventilation, 7 - hospitalized, severe disease / ventilation + additional organ support (pressors, RRT, ECMO), 8 – Death. The study did not contain samples of WHO grade 5.

Supplementary Figures and Tables

Supplementary Figure 1 | Relation between neutralization capacity and gender, donor age and Body-mass-index (BMI). Correlation between wild-type neutralization capacities (%) and gender (A, B), age (C, D) and BMI (E, F) for samples 7-49 days post PCR (A,C,E) and later than 49 days post PCR (B,D,F). P-values, when significant, are shown for all panels. Pearson's r was used to determine correlations.

Supplementary Table 1| Mutations of SARS-CoV-2 RBD proteins of 11 variants used in this study.

Variant	RBD mutations							
A.23.1	V367F							
Alpha								N501Y
Cluster 5				Y453F				
Epsilon			L452R					
Delta			L452R		T478K			
Lambda			L452Q			F490S		
Kappa			L452R				E484Q	
Eta							E484K	
Gamma		K417T					E484K	N501Y
Theta							E484K	N501Y
Beta		K417N					E484K	N501Y

Supplementary Table 2|Characteristics of the analyzed serum sample set.

Characteristic	
Number of donors	186
Number of samples	325
Median age (IQR) -years	62 (22)
Female sex (%)	84 (45.7)
Median ΔT post first positive PCR test (range)	85 (1-348)
Median BMI (range)	26.5 (18.4 – 50.2)

Supplementary Table 3 | ACE2 RBD competition assay validation results. Percentage coefficients of variation (%CV) of normalized MFI values for every bead region and RBD studied for every analyzed sample (n=5) and the mean of all samples.

	Samples	RBD WT	alpha	beta	gamma	epsilon	Cluster 5	A.23.1	eta	theta	kappa
Intra-assay precision (%CV)	Vac1	1.4	1.5	2.6	2.5	1.3	1.5	1.7	1.4	2.1	1.1
	Vac2	2.2	2.1	1.1	1.6	1.7	1.2	1.5	1.9	1.6	2.0
	Vac3	2.8	4.0	3.8	2.1	4.2	3.0	3.5	2.1	1.6	4.3
	Vac4	2.2	2.1	3.4	1.9	2.0	2.0	2.3	1.7	2.1	2.4
	Mean	**2.1**	**2.4**	**2.7**	**2.0**	**2.3**	**1.9**	**2.3**	**1.8**	**1.8**	**2.5**
Inter-assay precision (%CV)	Vac1	1.9	2.7	3.0	2.9	2.8	2.3	3.5	2.7	2.9	2.9
	Vac2	2.6	3.0	3.9	2.8	3.5	2.4	2.7	3.0	2.1	3.0
	Vac3	5.7	6.9	6.9	3.5	6.4	4.6	5.5	4.9	3.6	5.2
	Vac4	4.4	3.5	4.6	2.9	4.3	2.5	1.9	3.0	2.7	3.3
	Mean	**3.3**	**3.7**	**4.3**	**3.0**	**4.0**	**2.9**	**3.0**	**3.0**	**2.6**	**3.4**
Freeze-thaw stability (%CV)	Vac1	2.7	3.8	4.5	3.1	4.0	2.7	3.3	2.1	2.9	3.2
	Vac2	4.0	3.2	2.5	2.5	3.8	2.8	2.9	2.6	3.2	2.9
	Vac3	9.3	10.2	4.7	5.8	12.1	8.9	7.3	9.8	5.5	10.6
	Vac4	7.6	6.2	6.1	5.3	7.2	5.8	6.6	6.3	4.9	6.3
	Pre-pandemic	1.8	1.8	3.2	3.7	1.8	1.4	2.1	2.3	2.3	2.0
	Mean	**4.8**	**4.7**	**4.2**	**3.9**	**5.3**	**4.1**	**4.1**	**4.3**	**3.5**	**4.7**
Short-term stability (%CV)	Vac1	2.5	3.9	3.2	2.3	2.5	2.3	2.6	2.1	2.8	3.2
	Vac2	2.3	3.1	4.5	2.3	2.3	2.8	2.6	2.0	2.4	2.6
	Vac3	8.6	9.3	6.4	3.0	6.9	4.4	4.9	7.1	3.2	7.3
	Vac4	11.6	4.9	3.4	3.2	3.0	2.4	3.1	3.2	2.3	3.3
	Pre-pandemic	1.6	2.3	3.7	1.9	1.9	1.8	2.2	1.8	2.4	2.1
	Mean	**4.8**	**4.7**	**4.2**	**3.9**	**5.3**	**4.1**	**4.1**	**4.3**	**3.5**	**4.7**
Parallelism (%CV)	Vac1	4.1	3.9	5.0	4.0	4.5	4.2	4.2	3.2	2.6	3.9
	Vac2	4.2	2.5	4.0	3.4	2.7	2.5	4.2	2.6	2.8	4.2
	Vac3	12.0	11.7	4.0	5.3	11.9	11.8	9.3	13.6	4.5	12.7
	Vac4	11.3	9.6	6.5	6.9	10.0	8.7	8.6	9.9	4.4	9.3
	Pre-pandemic	2.1	1.2	3.6	2.8	2.3	2.1	2.4	1.9	2.7	1.4
	Mean	**4.8**	**4.7**	**4.2**	**3.9**	**5.3**	**4.1**	**4.1**	**4.3**	**3.5**	**4.7**