medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

Validation of a Novel IoT and AI based Point-of-Care Testing Laboratory: Analytical Accuracy and Clinical Agreement

Lucca Centa Malucelli[a*], Gabriele Luise Neves Alves[a], Anita Leme da Rocha Saldanha[], Tereza Bellincanta Fakhouri[], Carolina Melchioretto dos Santos[a], Matheus Gonçalves Severo[a], Victor Henrique Alves Ribeiro[a], Bernardo Montesanti Machado de Almeida[a], Caio Corsi Klosovski[a], Tania Leme da Rocha Martinez[], Marileia Scartezini[], Tereza Luiza Bellincanta Fakhouri[], Marcus Vinícius Mazega Figueredo[a]

[a] Hilab Campus, José A. Possebom, 800, Curitiba - PR (Brazil). 81270-185.
[] Nephrology Department, BP - A Beneficência Portuguesa de São Paulo, Brazil

*** Corresponding author**: lucca.malucelli@hilab.com.br

ABSTRACT

Point-of-care testing (POCT) offers several advantages over traditional laboratory testing. Offering less invasive testing with a faster turnaround time is not enough if not associated with an acceptable level of accuracy. Here, we show the analytical validation behind the multi-analyte POCT immunochromatography analyser, Hilab Flow (HiF). Analyses from 4,518 clinical samples were compared to College of American Pathologists accredited laboratories for ten quantitative and thirteen qualitative exams. Compatibility between methods was evaluated in terms of association/correlation and clinical agreement. Strong correlation/ concordance was observed between quantitative (CHOL, HDL-c, TG, HbA1c, Glycemia, 25-Hydroxy Vitamin D, TSH, Uric Acid, Creatinine, Urea) and qualitative methods (COVID-19 IgG/ IgM, Beta-hCG, Syphilis, Anti-HBsAg, Zika IgG/ IgM, Influenza A/B, HIV, HCV, HBsAg, Dengue NS1, COVID-19 Ag, Dengue IgG/ IgM, PSA). Approval criteria was obtaining a kappa agreement > 0.8 or a Pearson correlation > 0.9 depending on the exam. Overall percentage agreement was greater than 95% for all exams, indicating a good clinical agreement to gold-standard laboratory-based tests. Results indicate all exams are suitable for POCT and present a reliable performance. Data support the analyser is a useful tool to aid decision-making at the clinical setting, with potential to contribute with healthcare solutions in diagnostic medicine worldwide.

Keywords: Hilab, point-of-care, POCT, performance

1. INTRODUCTION

Point-of-care testing (POCT) employs remote clinical diagnostics with faster turnaround time, next to the patient [1-2]. It is a tool evolving as fast as innovation allows, with a potentially transformative impact on health care [3]. Devices for POCT diagnostics are now equipped with embedded technology and its application in medicine is steadily growing [4], since it may optimize clinical decisions, patient outcomes and even provide financial benefits [5-6].

POCT offers an alternative to traditional laboratory testing, especially in locations with limited infrastructure, but it is also used at wealthy settings in developed countries [7] . Advantages of the testing model includes the use of portable equipment, specimen collection at the test

NOTE: This preprint reports new research that has not been certified by peer review and should not be used to guide clinical practice.

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

site, immediate identification of biological samples, small sample volume needed and fast results, features that can minimize pre-analytical errors and risks related to the transport and identification of biological samples [1-2,8].

One of the commercially available POCT devices is the Hilab Flow (HiF), a patented analyser [9-13] dedicated to read immunochromatography results from lateral flow assays and colorimetric results from vertical flow assays. The small handheld POCT device combines artificial intelligence, machine learning and deep learning techniques to enable a shorter turnaround time, but without compromisings the exam's performance, precisely the main drawback in current POCT [14]. The HiF analysers are connected to a central laboratory for real-time monitoring and management of POCT performed remotely. Another important feature of this analyser is that it is able to perform a virtually unlimited number of exams, due to its system's flexibility.

The main concern regarding POCT is to guarantee reliable results, equivalent to those acquired with traditional standard laboratory-based methods, and according to international clinical guidelines [15-17]. Rapid test suppliers usually offer limited information about test performance thus performing a robust performance analysis is crucial to offer a quality service which meets consumer expectations. Therefore, this study aims to provide data from analytical validation to assess the system performance for POCT use and its potential to aid medical decision-making.

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

2. MATERIAL AND METHODS

2.1 Sample and Data

Internal data from the routine analysis of clinical laboratory service were used retrospectively for this in-house validation study. Evaluation of respiratory viruses exams evaluation, such as COVID-19 Ag, were conducted with nasopharyngeal samples, while other exams were evaluated from whole blood, serum and/or plasma specimens. The study was approved by the Beneficência Portuguesa Research Ethics Committee: CAEE 33490420.9.0000.5483.

The clinical samples were evaluated by lateral or vertical flow assays, depending on the type of the exam. Lateral flow assays are membrane-based tests which combine colloidal gold labeled particles to detect the analyte of interest from nasopharyngeal secretion, whole blood, serum or plasma specimens. Analyte's molecules react and form a complex with labeled antibodies or antigens and, as this conjugate moves through capillarity across the membrane, the anti-analyte antibody or antigen immobilized in the membrane binds to the complex, revealing a colored line of varying intensity, that can be measured by its optical density.

For qualitative exams, a colored line in the test region, accompanied by a line in the control region, indicates a positive result (reagent). For quantitative exams, the intensity of the coloured line in the test region relates to analytes' concentration. Other analytes were evaluated by vertical flow assays. The presence of the analyte of interest promotes a measurable color change on the test membrane which is proportional to the analyte concentration and can be measured similarly to lateral flow assays.

A list of all validated exams and the respective employed tests are shown in Table 1.

Table 1 - Specifications on Analyte, Sample Type and Comparison Methods for Evaluating Clinical Correlation.

Analyte	Specimen	Gold-standard assay	Equipment	HiF
COVID-19 IgG/IgM	Serum	Chemiluminescence	*Alinity, Abbott / Sprinter, Euroimunn*	Lateral Flow
COVID-19 Ag	Nasopharynx swab	Real Time PCR	QuantStudio 5, ThermoFisher / CFX 96, Bio-Rad	Lateral Flow
Influenza A/B Virus Ag	Nasopharynx swab	Real Time PCR	QuantStudio 5, ThermoFisher / CFX 96, Bio-Rad	Lateral Flow
Syphilis Ab	Serum	Chemiluminescence	*Alinity, Abbott / Sprinter, Euroimunn*	Lateral Flow
HIV Ab	Serum	Chemiluminescence	*Alinity, Abbott*	Lateral Flow
HCV Ab	Serum	Chemiluminescence	*Alinity, Abbott*	Lateral Flow
HBsAg	Serum	Chemiluminescence	*Alinity, Abbott*	Lateral Flow
Anti-HBsAg Ab	Serum	Chemiluminescence	*Alinity, Abbott*	Lateral Flow
Dengue Virus NS1 Ag	Whole blood, plasma and serum	Immunochromatography	*Wama*	Lateral Flow

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

Dengue Virus IgG/IgM	Serum	Enzyme immunoassay	*Sprinter, Euroimunn*	Lateral Flow
Zika Virus IgG/IgM	Serum	Enzyme immunoassay	*Sprinter, Euroimunn*	Lateral Flow
PSA	Serum	Chemiluminescence	DXI 800, Beckman	Lateral Flow
TSH	Serum	Chemiluminescence	DXI 800, Beckman	Lateral Flow
Beta-hCG	Serum	Chemiluminescence	DXI800, Beckman	Lateral Flow
25-Hydroxy Vitamin D	Serum	Chemiluminescence	DXI 800, Beckman	Lateral Flow
HbA1c	Whole blood	High Performance Liquid Chromatography (HPLC)	Variant, Beckman / Cobas C513	Lateral Flow
Glycemia	Serum	Colorimetric	AU5800, Beckman	Vertical Flow
Lipid Panel*	Serum	Colorimetric	AU5800, Beckman	Vertical Flow
Renal Function**	Serum	Colorimetric	*Beckman Coulter AU-DXC 5800*	Vertical Flow

*****Lipid Panel: cholesterol, HDL-cholesterol and triglycerides. ** Renal Function: uric acid, creatinine and urea.

2.3 HiF Analyser

The HiF reader (Hilab, Curitiba-PR) is a laboratory analyser platform for professional POCT used for detection and/or quantification of various analytes. The system processes immunochromatography results from lateral flow assays and colorimetric results from vertical flow assays by measuring optical density or color model values, respectively. The equipment is composed of two main parts, a portable handheld analyser (12 cm x 12 cm x 13 cm, 450 g), which incorporates a camera-equipped light detector, and sample integrated capsules/cartridges. Such a device communicates with the laboratory's server, where artificial intelligence tools and biomedical/biochemical specialists analyze the results. The system applies artificial intelligence techniques, such as computer vision and image processing tools to find regions of interest, improve image quality and detect objects. Moreover, machine learning and deep learning techniques perform classification and regression tasks to assist the analysis of quantitative and qualitative exams, respectively. Finally, human specialists also analyze the exams. If there is divergence between the automatic and human results, a senior specialist also analyzes the exam to achieve the final result, improving the overall quality of exams. For this study, HiF analysers were used for the analytical validation of 27 different analytes.

2.4 Operation

Sample (5 µL to 80 µL) is introduced into the capsule by using a capillary tube, pipette or a medicine dropper device, with or without an adjuvant buffer (depending on the exam). The capsule, identified by a unique QR code, is then inserted into the analyser, which measures the compatible signal of the lateral/vertical flow test within the capsule. Internet Of Things (IoT) technology is used to recognize the unique QR code from each test sample and sends the reaction information via cloud to a main laboratory. There, an artificial intelligence (AI) software analyses the reaction, and a licensed health professional confirms and signs the clinical diagnosis report. The test result is released through the cloud system to the health care service location where the test was performed and via email and/or text message to the patient. Data management and protection is ensured by a robust system, enabling the tracking of processed samples based on the investigated analyte [12]. The full process occurs within 30 min.

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

2.5 Clinical Agreement Evaluation

To ensure a thorough analysis, each exam was evaluated in terms of clinical agreement and compatibility between methods.

Agreement between quantitative methods was assessed through regression analysis of the plotted curves. Interpretation of Pearson's correlation was only secondary in our analysis especially because testing in a wider range might show strong correlation, but not necessarily indicate that both methods are compatible.

2.6 Performance Evaluation

For the exam's performance evaluation, measurements from 10 quantitatives and 13 qualitatives exams were compared to those from College of American Pathologists' (CAP) accredited laboratories. Some exams are multi-analyte (e.g. renal function and lipid profile) which explains a total of 27 evaluated analytes. Analytes, specimens, gold-standard assay and equipment used for comparison and validation of the POCT assays are described in the Supplementary Data.

Quantitative clinical correlation analyses were performed for the following analytes: blood glucose (glycemia), glycosylated hemoglobin (HbA1c), total cholesterol (CHOL), high-density lipoprotein cholesterol (HDL-c), triglycerides (TG), 25-Hydroxy (25-OH) Vitamin D, thyroid-stimulating hormone (TSH), uric acid (UA), creatinine (C), and urea (UR). Other analytes from the lipid panel are indirectly estimated through the difference between CHOL and HDL-c (NHDL-c), Martin's equation (LDL-c), Friedewald equation (VLDL), but not reported here because any interpretation on correlation and analytical accuracy would be derived from the relationship between CHOl, HDL-c and TG.

Qualitative clinical correlation analyses were performed for: COVID-19 IgG / IgM, COVID-19 antigen (Ag), Influenza A / B virus Ag, Syphilis antibody (Ab), human immunodeficiency virus (HIV) Ab, hepatitis C virus (HCV), surface antigen (HBsAg), hepatitis B virus surface antibody (Anti-HBsAg), Dengue virus NS1 Ag, Dengue virus IgG / IgM, Zika virus IgG / IgM, Beta-human chorionic gonadotropin (Beta-hCG), and prostate-specific antigen (PSA).

2.7 Statistical Analysis

The adjusted Wald Interval method (Agresti–Coull interval) was applied to estimate the confidence intervals for the clinical agreement parameters of the tests: positive percentage agreement (PPA), negative percentage agreement (NPA), overall percentage agreement (OPA), and overall clinical agreement (OCA). Other approaches might render inaccurate results due to showing proportions (PPA, NPA) close to 0, thus the adjusted Wald interval estimate was chosen to prevent such behavior [18].

OCA and OPA differed in exams with more than two reference ranges, which is the case for most quantitative assays. While qualitative assays comprise binary outcomes (e.g. normal VS disease), several quantitative assays might possess multiple reference ranges depending on analyte concentration and reported clinical impact (e.g. diabetes). Thus, OCA was based

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

on the hit rate for all reference ranges of the analyte. Conversely, OPA was calculated only for samples matching the reference ranges present in the analyte individual contingency tables. The reference ranges used for plotting the contingency tables and calculating each exam's performance are shown in Table 2.

Table 2 - Interpretation and Reference Ranges for the Exams's Clinical Agreement.

Exam / Analyte	Interpretation / Reference Ranges*			
COVID-19 IgG/ IgM	Reactive	Non-reactive	-	-
Syphilis	Reactive	Non-reactive	-	-
Anti-HBsAg	Reactive	Non-reactive	-	-
Zika IgG / IgM	Reactive	Non-reactive	-	-
Influenza A / B	Reactive	Non-reactive	-	-
HIV	Reactive	Non-reactive	-	-
HCV	Reactive	Non-reactive	-	-
HBsAg	Reactive	Non-reactive	-	-
Dengue NS1	Reactive	Non-reactive	-	-
COVID-19 Ag	Reactive	Non-reactive	-	-
Dengue IgG/ IgM	Reactive	Non-reactive	-	-
PSA	Reactive	Non-reactive	-	-
Beta-hCG	Pregnancy Indicator	Non-Pregnant	-	-
CHOL (mg/dL)	High (> 190 mg/dL)	Normal (< 190)	-	-
HDL-c (mg/dL)	Low (> 40)	Normal (> 40)	-	-
TG (mg/dL)	High (> 150)	Normal (< 150)	-	-
HbA1c (%)	Diabetes (> 6.5)	Pre-Diabetes (5.7-6.4)**	Normal (< 5.7)	-
Glycemia (mg/dL)	Diabetes (> 126 mg/dL)	Pre-Diabetes (100-126)**	Normal (70-100)	Hypoglycemia (< 70)**
25-OH Vitamin D (mg/dL)	Deficiency (< 20)	Normal (20-60)	-	-
TSH (µUI/mL)	Primary Hypothyroidism (> 10)	Mild Hypothyroidism (4.5-10)**	Normal (< 4.5)	-
Uric Acid (mg/dL)	High (male > 7 female > 5.7)	Normal (male < 7 female < 5.7)	-	-
Creatinine (mg/dL)	High (male > 1.3 female > 1.1)	Normal (male < 1.3 female < 1.1)	-	-
Urea (mg/dL)	High (> 20)	Normal (< 20)	-	-

All reference ranges were used for calculating each assay overall clinical agreement (OCA).

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

* Reference ranges are based on international, national societies and/or organizations responsible for establishing guidelines for the evaluation of each analyte.
** Reference ranges excluded from calculation of the exam's positive percentage agreement (PPA), negative percentage agreement (NPA), and overall percentage agreement (OPA).

Compatibility between methods was evaluated in terms of association (Pearson's correlation coefficient for quantitative or Cohen's Kappa agreement for qualitative assays) and performance parameters (PPA and NPA). For quantitative assays, the systematic difference was also evaluated by applying a compatible paired test (t-test for parametric distribution or Mann-Whitney for non-parametric) considering the null hypothesis as 0 for the true mean difference. Finally, the regression analysis was employed for estimating the slope (and its respective confidence intervals) and correlation coefficient [19]. Samples with concentrations outside of the exam's measuring range were excluded from the linear regression analysis to avoid misinterpretation of the data. The statistical significance for all analyzes was $p < 0.05$.

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

3. RESULTS

Due to lack of consensus in literature, we followed IUPAC and CLSI definition for “analytical sensitivity” which refers to the assay sensitivity to slight changes in analyte concentration [20]. That is, the closer to 1 is the slope of the regression curve, the more precise is the method. Several guidelines on method validation highlight the importance of calculating regression parameters (e.g. correlation coefficient, y-intercept) for a comprehensive method validation [21-22], but no clear criteria is established for the acceptable range of estimated slope, especially for POCT [23]. Thus, our acceptance criteria for assay precision was obtaining a slope anywhere between 0.85 - 1.15, which is a common rule-of-thumb for most rapid test strip suppliers (Figure 1).

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .



Analyte	Slope	r
CHOL	1.080 (0.984 - 1.175)	0.920
HDL-c	0.972 (0.899 - 1.046)	0.954
Triglycerides	1.076 (0.998 - 1.155)	0.959
HbA1c	0.983 (0.947 - 1.020)	0.938
Glycemia	1.028 (1.002 - 1.054)	0.970
25-Hydroxy Vitamin D	0.908 (0.828 - 0.987)	0.906
TSH	0.935 (0.904 - 0.966)	0.953
Uric Acid	0.959 (0.879 - 1.039)	0.917
Creatinine	0.956 (0.907 - 1.005)	0.934
Urea	1.016 (0.918 - 1.113)	0.924

Figure 1 - Regression Parameters and Correlation between reference and proposed method. Correlation between methods was extremely high (r > 0.9) for all analytes.

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

All exams matched our criteria, since estimated slopes (with confidence intervals) ranged from 0.85 - 1.15. The closer to 1 is the slope the higher is the precision of an exam in relation to changes in analyte content. For the remaining exams, the variation in slope did not impact the clinical agreement of the exams even at decision limits, as indicated by the obtained PPA and NPA.

PPA, NPA and OCA were greater than 90% for all analytes, suggesting a good agreement for the proposed methods (Tables 3 and 4). Clinical differences between measurements (proposed method VS POCT) were considered acceptable for any result within a defined % of a category boundary. Ferreira et al. (2015) [24] employed a similar approach to avoid overestimation of clinical differences. However, we considered analyte-specific variation – reported by the CLIA [25] – as a standard other than an arbitrary value (e.g. 5%) for a more accurate estimate of exam's performance.

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

Table 3 - Clinical Agreement for Quantitative Assays.

Exams	PPA	NPA	OPA	OCA	n
CHOL	>99.0% (CI: 94.0% - 100.0%)	>99.0% (CI: 94.0% - 100.0%)	>99.0%	>99.0%	148
HDL-c	96.3% (CI: 80.0% - 100.0%)	>99.0% (CI: 96.0% - 100.0%)	>99.0%	>99.0%	131
TG	>99.0% (CI: 90.0% - 100.0%)	98.2% (CI: 93.3% - 100.0%)	98.7%	98.7%	157
HbA1c	90.5% (CI: 80.0% - 96.0%)	99.3% (CI: 97.0% - 100.0%)	97.7%	93.8%	3.87
Glycemia	>99.0% (CI: 89.0% - 100.0%)	>99.0% (CI: 97.0% - 100.0%)	>99.0%	91.3%	243
25-Hydroxy Vitamin D	>99.0% (CI: 90.0% - 100.0%)	>99.0% (CI: 94.0% - 100.0%)	>99.0%	>99%	115
TSH	>99.0% (CI: 87.0% - 100.0%)	>99.0% (CI: 98.0% - 100.0%)	>99.0%	98.0%	342
Uric Acid	91.7% (CI: 73.0% - 98.0%)	>99.0% (CI: 93.0% - 100.0%)	97.6%	97.6%	84
Creatinine	92.6% (CI: 82.0% - 97.0%)	>99.0% (CI: 94.0% - 100.0%)	97.0%	97.0%	133
Urea	97.8% (CI: 87.0% - 100.0%)	>99.0% (CI: 94.0% - 100.0%)	98.7%	98.7%	157
					1897

Table 4 - Clinical Agreement for Qualitative Assays.

Exams	PPA	NPA	OPA	Kappa (k)	n
COVID-19 IgG / IgM	>99.0% (CI: 98.0% - 100.0%)	>99.0% (CI: 99.0% - 100.0%)	>99.0%	0.99	1008
Beta-hCG	>99.0% (CI: 94.0% - 100.0%)	>99.0% (CI: 98.0% - 100.0%)	>99.0%	1.00	276
Syphilis	>99.0% (CI: 87.0% - 100.0%)	>99.0% (CI: 90.9% - 100.0%)	>99.0%	1.00	80
Anti-HBsAg	97.3% (CI: 90.1% - 99.8%)	>99.0% (CI: 92.8% - 100.0%)	98.5%	0.97	137
Zika IgG / IgM	>99.0% (CI: 84.0% - 100.0%)	>99.0% (CI: 93.0% - 100.0%)	>99.0%	1.00	92
Influenza A / B	>99.0% (CI: 90.0% - 100.0%)	>99.0% (CI: 89.0% - 100.0%)	>99.0%	1.00	80
HIV	>99.0% (CI: 80.6% - 100.0%)	>99.0% (CI: 87.6% - 100.0%)	>99.0%	1.00	54
HCV	>99.0% (CI: 79.8% - 100.0%)	>99.0% (CI: 90.2% - 100.0%)	>99.0%	1.00	63
HBsAg	>99.0% (CI: 76.6% - 100.0%)	>99.0% (CI: 92.0% - 100.0%)	>99.0%	1.00	69
Dengue NS1	>99.0% (CI: 94.8% - 100.0%)	>99.0% (CI: 97.8% - 100.0%)	>99.0%	1.00	300
COVID-19-Ag	>99.0% (CI: 94.0% - 100.0%)	>99.0% (CI: 95.7% - 100.0%)	>99.0%	1.00	181
Dengue IgG / IgM	>99.0% (CI: 91.0% - 100.0%)	>99.0% (CI: 97.0% - 100.0%)	>99.0%	1.00	209
PSA	>99.0% (CI: 87.0% - 100.0%)	>99.0% (CI: 89.3% - 100.0%)	>99.0%	1.00	72
					2621

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

4. DISCUSSION

Universal health care availability depends on decentralized diagnostic POCT systems in order to solve healthcare bottlenecks both in wealthy and developing countries [26-27]. Technology has incremented POCT for the last two decades, so modern devices are now smaller, smarter, easier to use, less prone to errors and go without additional testing for clinical purposes [4]. POCT is a patient-oriented technology, but health professional expertise is still required for sensitive diagnostic and clinical decision-making. Importantly, patients that used POCT reported greater levels of confidence in their health professional and motivation to manage their health condition [28]. For minor conditions and highly prevalent diseases, POCT results may be sufficient for health decisions in the clinical setting [8], without additional confirmation from a traditional laboratory. However, this scenario requires that POCT is reliable and robust regarding clinical agreement with an established laboratory method [29-30]. Combining a robust analytical validation and a reliable quality control in the clinical laboratory is the key for improving healthcare access, even in remote regions [14].

In the point of care industry, equipments are usually designed to analyse a small set of analytes which limits the spread of POCT and, consequently, offers a service at competitive prices to the consumer. The HiF analyser goes further by enabling a multi-analyte evaluation on a single device of quantitative and qualitative exams. Also, employing a dual verification of analyzed samples (by the AI and the biomedical specialist) helps minimize common analytical errors from decentralized laboratory settings. Previous works from our group show AI can improve the analytical precision of a method, thus enabling epidemiological conclusions on a populational level [13,31-32].

The integrated laboratory management model, especially developed for POCT with the HiF, is a distinguishing feature that enables a continuous enhancement of the test accuracy. These features are aligned with the understanding that POCT test performance should be compatible with any traditional tests, as required by international and regional standards [33]. POCT can be used to improve patient flow within the clinical setting and may even provide indirect cost effectiveness advantages compared to centralized laboratory testing, considering increased resource management and more efficient patient care [3,34]. Shortening the turnaround time is another asset of the HiF system, specially when delays compromises the patient's health: on an intensive care unit setting, the time of analysis shows great influence over a physician's conduct [35], thus the patient can greatly benefit from a faster, yet precise result.

Modern POCT health systems may also create a valuable database for precision public health management using Big Data. Precision public health uses data from traditional and emerging sources to deliver assertive interventions to populations, focusing on social and environmental health aspects [36-38]. Big Data analytics aims to implement evidence-based decision-making in clinical and population settings [38]. This scenario depends on available population-level data on disease surveillance [40], and requires a solid collaboration among health care systems, clinicians, governments and communities [41]. Precision public health can provide doctors and public health practitioners with new insights, with the potential of revolutionizing health care, through the use of new technologies.

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint
(which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity.
It is made available under a CC-BY-NC-ND 4.0 International license .

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

5. CONCLUSION

This study presents data from 4,518 clinical samples used to evaluate the analytical accuracy and clinical agreement between a CAP certified laboratory and a POCT based laboratory (HiF analyser). Combining effective IoT and AI tools is essential for offering POCT alternatives with a faster turnaround time and a reliable performance to benefit both health professionals and patients. Our results show that the evaluated POCT system is a useful tool to aid medical decision-making at the clinical setting, with potential to contribute with healthcare solutions in diagnostic medicine.

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

CRediT author statement

Lucca Centa Malucelli: Project Administration, Writing, Editing, and Review (Original and Final draft).
Gabriele Luise Neves Alves: Visualization, Writing, Editing, and Review (Original and Final draft).
Carolina Melchioretto dos Santos: Supervision.
Matheus Gonçalves Severo: Editing and Review (Original and Final draft).
Victor Henrique Alves Ribeiro: Writing, Editing, and Review (Final draft).
Bernardo Montesanti Machado de Almeida: Writing, Editing, and Review (Original and Final draft), Supervision.
Caio Corsi Klosovski: Editing and Review (Original and Final draft).
Tania Leme da Rocha Martinez: Editing and Review (Original and Final draft), Supervision.
Marileia Scartezini: Editing and Review (Original and Final draft), Supervision.
Tereza Luiza Bellincanta Fakhouri: Editing and Review (Original and Final draft).
Marcus Vinícius Mazega Figueredo: Funding Acquisition and Supervision.

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

ACKNOWLEDGMENT

We acknowledge funding for the analytical method validation of the exams reported here.

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

REFERENCES

[1] Larsson, A., Greig-Pylypczuk, R., Huisman, A. The state of point-of-care testing: a european perspective, Upsala Journal of Medical Sciences. 2015, 120(1): 1-10.

[2] Luppa, PB., Müller, C., Schlichtiger, A., Schlebusch, H. Point-of-care testing (POCT): Current techniques and future perspectives. Trends in Analytical Chemistry. 2011, 30(6): 887-898.

[3] Jani, IV., Peter, TF. How point-of-care testing could drive innovation in global health. New England Journal of Medicine. 2013, 368(24): 2319-24.

[4] St John, A., Price, CP. Existing and Emerging Technologies for Point-of-Care Testing. Clin Biochem Rev. 2014, 35(3): 155-67.

[5] St John, A., Price, CP. Economic evidence and point-of care testing. Clin Biochem Rev. 2013, 34:61-74.

[6] Laurence, CO., Moss, JR., Briggs, NE., Beilby, JJ., Po, CT. PoCT Trial Management Group. The cost-effectiveness of point of care testing in a general practice setting: results from a randomised controlled trial. BMC Health Serv Res. 2010,10: 165.

[7] Kozel, TR., Burnham-Marusich, AR. Point-of-Care Testing for Infectious Diseases: Past, Present, and Future. J Clin Microbiol. 2017, 55(8): 2313-2320.

[8] Park, HD. Current Status of Clinical Application of Point-of-Care Testing. Arch Pathol Lab Med. 2021, 145(2): 168-175.

[9] Figueredo, M. V. M., Junior, S. R. R., Cossetin, M. J., Cemin, R. S. C., Pinto, R. N., Santos, A. R., Andrade, G. H. B., Camati, M. E. M. (2018). Leitor de Amostras Biológicas em Testes Bioquímicos Rápidos. BR 102017008428-0 A2. Instituto Nacional da Propriedade Industrial.

[10] Figueredo, M. V. M., Junior, S. R. R., Cossetin, M. J., Cemin, R. S. C., Pinto, R. N., Santos, A. R., Andrade, G. H. B., Camati, M. E. M. (2018). Sistema de Coleta, Exame e Leitura de Amostras Biológicas em Testes Bioquímicos Rápidos e Gerenciamento de Dados. Instituto Nacional da Propriedade Industrial.

[11] Figueredo, M. V. M., Junior, S. R. R., Cossetin, M. J., Cemin, R. S. C., Pinto, R. N., Santos, A. R., Andrade, G. H. B., Camati, M. E. M. (2018). Reader Device for Biological Samples in Rapid Biochemical Tests. WO 2018/195616 Al. Intellectual Property Office.

[12] Figueredo, M. V. M., Junior, S. R. R., Cossetin, M. J., Cemin, R. S. C., Pinto, R. N., Santos, A. R., Andrade, G. H. B., Camati, M. E. M. (2020). Biological Sample Collecting, Examining and Reading System for Fast Biochemical Assays and Data Management. System for biological sample collection, examination and reading in quick biochemical tests and data management. WO 2018/195617 Al. Intellectual Property Office.

[13] Cardoso, C. Q., Almeida. B. M. M., Saldanha, A. L. R., Klosovski, C. C., Margeotto A. P. P., Gasparoto, A. L. V., Scartezini, M., Martinez, T. L. R. (2021). Brazilian Lipid Cardiovascular Risk Pre and During the Covid 19 Pandemic in Asymptomatic and Severely Affected Groups. Online Journal of Cardiology Research & Reports. DOI: 10.33552/OJCR.2021.05.000606.

[14] Florkowski, C., Don-Wauchope, A., Gimenez, N., Rodriguez-Capote, K., Wils, J., & Zemlin, A. (2017). Point-of-care testing (POCT) and evidence-based laboratory medicine (EBLM)–does it leverage any advantage in clinical decision making?. Critical reviews in clinical laboratory sciences, 54(7-8), 471-494.

[15] Pecoraro, V., Banfi, G., Germagnoli, L., Trenti, T. A systematic evaluation of immunoassay point-of-care testing to define impact on patients' outcomes. Ann Clin Biochem. 2017, 54(4): 420-431.

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

[16] International Organization for Standardization (ISO) 22870:2006 Point-of-care testing care (POCT) – Requirements for quality and competence. Geneva: ISO, 2006.

[17] CLSI. Quality Management: Approaches to Reducing Errors at the Point of Care. 1st Edition. CLSI guideline POCT07. Wayne, PA: Clinical and Laboratory Standards Institute, 2010.

[18] Agresti, A., Coull, BA. Approximate is better than “exact” for interval estimation of binomial proportions. The American Statistician. 1998, 52(2): 119-126.

[19] Watson, PF., Petrie, A. Method agreement analysis: a review of correct methodology. Theriogenology. 2010, 73(9): 1167-1179.

[20] NCCLS. Protocols for Determination of Limits of Detection and Limits of Quantitation; Approved Guideline. NCCLS document EP17-A (ISBN 1-56238-551-8). NCCLS, 940 West Valley Road, Suite 1400, Wayne, Pennsylvania 19087-1898 USA, 2004.

[21] ICH. International Conference on Harmonization. Guidance for industry: Q2B validation of analytical procedures: methodology. International Council on Harmonization of Technical Requirements for Registration of Pharmaceuticals for Human Use, 1996.

[22] VICH. Validation of analytical procedures: methodology (VICH GL2). 1999.

[23] Krause, Stephan O. Good analytical method validation practice: Deriving acceptance criteria for the AMV protocol: Part II. Journal of validation technology. 2003, 9(2):. 162-179.

[24] Ferreira, CE., França, CN., Correr, CJ., Zucker, ML., Andriolo, A., Scartezini, M. Clinical correlation between a point-of-care testing system and laboratory automation for lipid profile. Clinica Chimica Acta. 2015, 446: 263-266.

[25] US Department of Health and Human Services (HHS). Clinical laboratory improvement amendments of 1988 (CLIA) proficiency testing regulations related to analytes and acceptable performance, 2019.

[26] Urdea, M., Penny, LA., Olmsted, SS., Giovanni, MY., Kaspar, P., Shepherd, A., Wilson, P., Dahl, CA., Buchsbaum, S., Moeller, G., Hay Burgess, DC. Requirements for high impact diagnostics in the developing world. Nature. 2006, 444(1): S73–S79.

[27] Pai, NP., Vadnais, C., Denkinger, C., Engel, N., Pai, M. Point-of care testing for infectious diseases: diversity, complexity, and barriers in low- and middle-income countries. PLoS Med. 2012, 9: e1001306.

[28] Laurence, CO., Gialamas, A., Bubner, T., Yelland, L., Willson, K., Ryan, P., et al. Patient satisfaction with point-of-care testing in general practice. Br J Gen Pract. 2010, 60: e98-104.

[29] Price, C., St. John, A., Kricka, L. Point-of-Care Testing. Needs, opportunities and innovation. 3rd Edition. Washington, USA: AACC Press, 2010.

[30] Hsieh, YH., Gaydos, CA., Hogan, MT., Uy, OM., Jackman, J., Jett-Goheen, M, et al. What qualities are most important to making a point of care test desirable for clinicians and others offering sexually transmitted infection testing? PLoS One. 2011, 6: e19263.

[31] Martinez, T. L. R., Almeida, B. M. M., Cardoso, C. Q., Saldanha, A. L. R., Scartezini, M., Klosovski, C. C., Pereira, A., Santos Filho, R. D. (2021). The importance of the ratio triglycerides/HDL C in RHE Brazilian population residual risk. Atherosclerosis, 331, e204.

[32] Ribeiro, V. H. A., Steinhaus, G., Severo, E. B., Junior, J. R. F., Barbosa, L. J. L., Cossetin, M., & Figueiredo, M. V. M. (2021). A System for Enhancing Human-level Performance in COVID-19 Antibody Detection. In Anais do XXI Simpósio Brasileiro de Computação Aplicada à Saúde (pp. 224-233). SBC.

medRxiv preprint doi: https://doi.org/10.1101/2021.10.29.21264864; this version posted November 1, 2021. The copyright holder for this preprint **(which was not certified by peer review)** is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

[33] Sociedade Brasileira de Patologia Clínica/Medicina Laboratorial (SBPC/ML). Diretrizes para a gestão e garantia da qualidade de Testes Laboratoriais Remotos (TLR).2nd edition. 2016, Barueri, SP: Editora Manole, 564 p.

[34] Weber, CF., Zacharowski, K. Perioperative point of care coagulation testing. Dtsch Arztebl Int. 2012, 109: 369-375.

[35] Ferreira, C. E., Guerra, J. C., Slhessarenko, N., Scartezini, M., Franca, C. N., Colombini, M. P., Berlitz, F., Machado, A. M. O., Campana, G. A., Faulhaber, A. C. L., Galoro, C. A., Dias, C. M., Shcolnik, W., Martino, M. D. V., César, K. R., Sumita, N. M., Mendes, M. E., Faulhaber, M. H. W., Pinho, J. R. R., Barbosa, I. V., Batista, M. C., Khawali, C., Pariz, V. Mi., Andriolo, A. (2018). Point-of-Care Testing: General Aspects. Clinical laboratory, 64(1), 1-9.

[36] Dowell, SF., Blazes, D., Desmond-Hellmann, S. Four steps to precision public health. Nature. 2016, 540 (7632): 189–191.

[37] Horton, R. Offline: in defence of precision public health. Lancet. 2018, 392 (10157): 1504.

[38] Rasmussen, SA., Khoury, MJ., del Rio, C. Precision Public Health as a Key Tool in the COVID-19 Response. JAMA. 2020, 324 (10): 933–934.

[39] Dolley, S. Big data's role in precision public health. Front Public Health. 2018,6: 68.

[40] Velmovitsky, PE., Bevilacqua, T., Alencar, P., Cow, D., Morita, PP. Convergence of Precision Medicine and Public Health Into Precision Public Health: Toward a Big Data Perspective. Frontiers in Public Health. 2021, 9: 561873.

[41] Khoury, MJ., Engelgau, M., Chambers, DA., Mensah, GA: Beyond Public Health Genomics: Can Big Data and Predictive Analytics Deliver Precision Public Health? Public Health Genomics. 2018, 21: 244-250.