COVID-19 Variant Detection with a High-Fidelity CRISPR-Cas12 Enzyme

Clare L. Fasching[1]*, Venice Servellita[2,3]*, Bridget McKay[1], Vaishnavi Nagesh[1], James

P. Broughton[1], Noah Brazer[2,3], Baolin Wang[2,3], Alicia Sotomayor-Gonzalez[2,3], Kevin

Reyes[2,3], Jessica Streithorst[2,3], Rachel N. Deraney[1], Emma Stanfield[1], Carley G.

Hendriks[1], Steve Miller[2,3], Jesus Ching[1], Janice S. Chen[1†], Charles Y. Chiu[2,3,4†]

[1] Mammoth Biosciences, Inc., Brisbane, CA, USA.

[2] Department of Laboratory Medicine, University of California San Francisco, San

Francisco, CA, USA.

[3] UCSF-Abbott Viral Diagnostics and Discovery Center, San Francisco, CA, USA.

[4] Department of Medicine, Division of Infectious Diseases, University of California San

Francisco, San Francisco, CA, USA.

*These authors contributed equally

[†]Co-corresponding authors

Abstract

Laboratory tests for the accurate and rapid identification of SARS-CoV-2 variants have

the potential to guide the treatment of COVID-19 patients and inform infection control

and public health surveillance efforts. Here we present the development and validation

of a COVID-19 variant DETECTR® assay incorporating loop-mediated isothermal

amplification (LAMP) followed by CRISPR-Cas12 based identification of single

nucleotide polymorphism (SNP) mutations in the SARS-CoV-2 spike (S) gene. This

assay targets the L452R, E484K, and N501Y mutations associated with nearly all

circulating viral lineages. In a comparison of three different Cas12 enzymes, only the

newly identified enzyme CasDx1 was able to accurately identify all three targeted SNP

mutations. We developed a data analysis pipeline for CRISPR-based SNP identification

using the assay from 91 clinical samples (Ct < 30), yielding an overall SNP concordance

and agreement with SARS-CoV-2 lineage classification of 100% compared to viral

whole-genome sequencing. These findings highlight the potential utility of CRISPR-

based mutation detection for clinical and public health diagnostics.

Introduction

The emergence of new SARS-CoV-2 variants threatens to substantially prolong the

COVID-19 pandemic. SARS-CoV-2 variants, especially Variants of Concern (VOCs) (*1,*

2), have caused resurgent COVID-19 outbreaks in the United States (*2-5*) and

worldwide (*1, 6, 7*), even in populations with a high proportion of vaccinated individuals

(*8-11*). Mutations in the spike protein, which binds to the human ACE2 receptor, can

render the virus more infectious and/or more resistant to antibody neutralization,

resulting in increased transmissibility (*12*), and/or escape from immunity, whether

vaccine-mediated or naturally acquired immunity (*13, 14*). Variant identification can also

be clinically significant, as some mutations substantially reduce the effectiveness of

available monoclonal antibody therapies for the disease (*15*).

Tracking the evolution and spread of SARS-CoV-2 variants in the community is critical

to inform public policy regarding testing and vaccination, as well as guide contact

tracing and containment effects during local outbreaks (*16, 17*). Virus whole-genome

sequencing (WGS) and single nucleotide polymorphism (SNP) genotyping are

commonly used to identify variants Harper, 2021 #100;Oude Munnink, 2021 #99}, but

can be limited by long turnaround times and/or the requirement for bulky and expensive

laboratory instrumentation. Diagnostic assays based on clustered interspaced short

palindromic repeats (CRISPR) (*18*) have been developed for rapid detection of SARS-

CoV-2 in clinical samples (*13, 19-22*),, and a few have obtained Emergency Use

Authorization (EUA) by the US Food and Drug Administration (FDA) (*23, 24*). Some

advantages of these assays for use in laboratory and point of care settings include low

cost, minimal instrumentation, and a sample-to-answer turnaround time of under 2

hours (*19, 22, 25, 26*).

Here we present the development of a CRISPR-based COVID-19 variant DETECTR[®]

assay (henceforth abbreviated as DETECTR[®] assay) for the detection of SARS-CoV-2

mutations and evaluate its performance on 91 clinical samples using WGS as a

comparator method (**Fig. 1A**). The assay combines RT-LAMP pre-amplification followed

by fluorescent detection using a CRISPR-Cas12 enzyme. We perform a comparative

evaluation of multiple candidate Cas12 enzymes and demonstrate that robust assay

performance depends on the specificity of the newly identified CRISPR-Cas12 enzyme

called CasDx1 in identifying three key SNP mutations of functional relevance in the

spike protein, N501Y, L452R, and E484K (*27*).

Results

Identifying the optimal CRISPR-Cas12 enzyme for SNP detection

To determine the optimal Cas12 enzyme for SNP detection, we evaluated three

different CRISPR-Cas effectors with trans-cutting activity: LbCas12a, AsCas12a, and a

novel Cas12 enzyme called CasDx1. We initially screened guide RNAs (gRNAs) with

CasDx1 and LbCas12a for activity on synthetic gene fragments encoding regions of the

SARS-CoV-2 S-gene with either wild-type (WT) or mutant (MUT) sequences at amino

acid positions 452, 484, and 501 (**Fig. 1B-C**). From this initial activity screen, we

identified the top-performing gRNAs for each S-gene variant encoding either L452R,

E484K or N501Y **(Fig. 1D)**. Further evaluation of these guides using CasDx1,

LbCas12a and AsCas12a with their cognate gRNAs on synthetic gene fragments

revealed differences in SNP differentiation capabilities, with CasDx1 showing the

clearest SNP differentiation between wild-type (WT) and mutant (MUT) sequences for

all three S-gene variants **(Fig. 1D and Fig. S1A)**. In comparison, LbCas12a was

capable of differentiating SNPs at positions 452 and 484, but not 501, whereas

AsCas12a was only able to differentiate the SNP at position 452 **(Fig. 1D and Fig.**

S1A).

We next tested SNP differentiation capabilities on heat-inactivated viral cultures using

the full DETECTR® assay, consisting of RNA extraction, multiplexed RT-LAMP

amplification **(Fig. 1C)**, and CRISPR-Cas12 detection with guide RNAs targeting part of

the spike receptor-binding domain (RBD) **(Fig. 1B).** The LAMP primer design

incorporated two sets of six primers each, with both sets generating overlapping spike

RBD amplicons that spanned the L452R, E484K, and N501Y mutations. We chose to

adopt a redundant LAMP design for two reasons: first, this approach was shown to

improve detection sensitivity in initial experiments; second, we sought to increase assay

robustness given the continual emergence of escape mutations in the spike RBD

throughout the course of the pandemic (*13*). The tested viral cultures included an

ancestral SARS-CoV-2 lineage (WA-1) containing the wild-type spike protein (D614)

targeted by the approved mRNA (Pfizer and Moderna) (*28, 29*) and DNA adenovirus

vector (Johnson and Johnson) (*30*) vaccines, variants being monitored (VBMs) that

were previously classified as VOCs or variants of interest (VOIs), including Alpha

(B.1.1.7), Beta (B.1.351), Gamma (P.1), Epsilon (B.1.427 and B.1.429), Kappa

(B.1.617.1), and Zeta (P.2) lineages, and the current VOC Delta (B.1.617.2) lineage

(*31*). Heat-inactivated viral culture samples representing the seven SARS-CoV-2

lineages were quantified by digital droplet PCR across a 4-log dynamic range and used

to evaluate the analytical sensitivity of the pre-amplification step. RT-LAMP amplification

was evaluated using six replicates from each viral culture. We observed consistent

amplification for all seven SARS-CoV-2 lineages with 10,000 copies of target input per

reaction (200,000 copies/mL) **(Fig. 1E)**, which is comparable to the target input of

>200,000 copies/mL viruses (<30 Ct value) required for sequencing workflows used in

SARS-CoV-2 variant surveillance (*32, 33*).

To evaluate the specificity of the different Cas12 enzymes, amplified material from each

viral culture was pooled and the SNPs resulting in the L452R, E484K and N501Y

mutations were detected using CasDx1, LbCa12a and AsCas12a. Similar to the results

found using gene fragments, CasDx1 correctly identified the wild-type (WT) and

mutational (MUT) targets at positions 452, 484 and 501 in each LAMP-amplified, heat-

inactivated viral culture **(Fig. 1F and Fig. S1B)**. In comparison, LbCas12a was capable

of differentiating WT from MUT at position 501 on LAMP-amplified viral cultures but

showed much higher background for the WT target at position 452 and higher

background for both WT and MUT targets at position 484 for **(Fig. 1F and Fig. S1B)**.

Additionally, AsCas12a was able to differentiate WT from MUT targets at position 452

albeit with substantial background but was unable to differentiate WT from MUT targets

at positions 484 and 501 **(Fig. 1F and Fig. S1B)**. From these data, we concluded that

CasDx1 would provide more consistent and accurate calls for the L452R, E484K and

N501Y mutations. We thus proceeded to further develop the assay using only the high-

fidelity CasDx1 enzyme.

Fig. 1. Design and Workflow for the DETECTR® assay. (**A**) Workflow comparison

between the DETECTR® assay and SARS-CoV-2 whole-genome sequencing (WGS).

(**B**) Schematic of CRISPR-Cas gRNA design for SARS-CoV-2 S gene mutations. (**C**)

Schematic of multiplexed RT-LAMP primer design showing the SARS-CoV-2 S gene

mutations and gRNA positions. (**D**) Heat map comparison of three different Cas12

enzymes tested using 10 nM PCR-amplified synthetic gene fragments. (**E**) Dot plot

showing the number (n=6) of positive replicates across a 4-log dynamic range of the

RT-LAMP products (**F**) Heat map comparison of end-point fluorescence of three

different Cas12 enzymes tested against heat-inactivated viral cultures. Replicates (n=6)

generated using RT-LAMP were pooled and CRISPR-Cas12 reactions were then run in

triplicate (n=3).

Development of a data analysis pipeline for calling COVID-19 variant SNPs with

the DETECTR® assay

To develop a data analysis pipeline for calling SARS-CoV-2 SNP mutations and assign

lineage classifications with the DETECTR® assay (**Fig. 2A-B**), we first used data

collected from SNP synthetic gene fragment controls (n=279) that included all

mutational combinations of 452, 484 and 501 (see Methods). Based on the control

sample data, we generated allele discrimination plots (*34, 35*) to define boundaries that

separate the WT and MUT signals **(Fig. S4A)**. Clear differentiation between WT and

MUT signals was observed when plotting the ratio against the average of the WT and

MUT transformed values on a mean average (MA) plot (*34, 35*) **(Fig. S4B)**, with 100%

concordance for SNP identity at positions 452, 484, and 501 for the control samples.

Performance evaluation of the DETECTR® assay using clinical samples

Next, we assembled a blinded dataset consisting of 93 COVID-19 positive clinical

samples (previously analyzed by viral WGS) and the SNP controls run in parallel. These

samples were extracted, amplified in triplicate RT-LAMP reactions **(Fig. S2)**, and

processed further as triplicate CasDx1 reactions for each LAMP replicate **(Fig. S3)**. A

total of nine replicates were thus generated for each sample to detect WT or MUT SNPs

at positions 452, 484, and 501. The DETECTR® data analysis pipeline was then applied

to each sample to provide a final lineage categorization **(Fig. 2A-C).** For a biological

RT-LAMP replicate to be designated as either WT or MUT, the same call needed to be

made from all three technical CasDx1 replicates **(Fig. S5A).** A final SNP mutation call

was made based on ≥ 1 of the same calls from the three biological replicates, with

replicates that were designated as a No Call ignored **(Fig. S5A-C).** After excluding two

samples that were considered invalid because the fluorescence intensity from RT-LAMP

amplification did not reach a pre-established threshold determined using receiver-

operator characteristic (ROC) curve analysis **(Fig. S2 and Fig. S6)**, we evaluated a total

of 807 CasDx1 signals from the 91 remaining clinical samples, generating up to 9

replicates for each clinical sample **(Fig. S5B)**. Differentiation of WT and MUT signals

according to the allele discrimination plots was more pronounced at positions 484 and

501 than position 452 **(Fig. S4)**, whereas the MA plots showed clear separation of WT

and MUT calls for all three positions **(Fig. 3A and Fig. S4).** The variant calls made on

each sample were consistent with the difference in median values of the log-

transformed signals as determined using the data analysis pipeline **(Fig. S7)**.

Fig. 2. DETECTR® data analysis pipeline for SARS-CoV-2 SNP mutation calling.

(**A**) Interpretation table summarizing the SARS-CoV-2 mutations in this study associated with the corresponding lineage classification. (**B**) Schematic of data analysis pipeline describing the RT-LAMP QC and subsequent CasDx1 signal scaling. The scaled signals were compared across SNPs and the calls were made for each RT-LAMP replicate. The combined replicate calls defined the mutation call, which informed the final lineage classification. (**C**) Three representative clinical samples of different SARS-CoV-2 lineages depict the workflow of the DETECTR® assay. Raw fluorescence curves of each sample run in RT-LAMP amplification and subsequent triplicate DETECTR® reactions targeting both WT and MUT SNPs for L452(R), E484(K), and N501(Y). Box plot visualization of the end point fluorescence in DETECTR® across each SNP for the three representative clinical samples. Calls were made for each SNP by evaluating the median values of the DETECTR® calls and overall calls through the LAMP replicates, and given a designation of WT, MUT, or NoCall. Final calls are made on the lineage determined by each SNP. Blue represents WT and green represents MUT, with RT-LAMP replicates (n = 3), CasDx1 replicates (n = 3 per LAMP replicate) and shading around kinetic curves indicates ±1.0SD.

We then unblinded the viral WGS results to evaluate the accuracy of the DETECTR® assay for SNP calls and lineage classification. There were 14 discordant SNP calls out of 272 (94.9% SNP concordance) distributed among 11 clinical samples out of 91 **(Fig. S8A-C)**. Among the 11 discordant samples, one sample (COVID-31) was designated a 'no call' at position 452 by viral WGS and thus lacked a comparator, two samples were

designated a 'no call' due to flat WT and MUT curves (COVID-41 and COVID-73), four

samples had similar WT and MUT curve amplitudes, suggesting a mixed population

(COVID-03, COVID-56, COVID-61 and COVID-81) **(Fig. S8A)**, and four samples had

SNP assignments discordant with those from viral WGS (COVID-12, COVID-13,

COVID-20 and COVID-63) **(Fig. S8A).**

Given that the comparison data had been collected over an extended time period, we

surmised that sample stability issues arising from aliquoting and multiple freeze-thaw

cycles may have accounted for the observed discrepancies. To further investigate this

possibility, the 11 discordant clinical samples were re-extracted from original respiratory

swab matrix and re-analyzed by running both viral WGS and the DETECTR® assay in

parallel. Re-testing of the samples resulted in nearly complete agreement between the

two methods, with the exception of two SNPs that were identified as E484Q in two

samples by WGS but were incorrectly called E484 (WT) by the DETECTR® assay **(Fig.**

3B-C and Fig. S8D). Thus, based on discrepancy testing, the positive predictive

agreement (PPA) between the DETECTR® assay and viral WGS at all three WT and

MUT SNP positions was 100% (272 of 272, p<2.2e-16 by Fisher's Exact Test) **(Fig.**

3D). The corresponding negative predictive agreement (NPA) was 91.4% as the E484Q

mutation for two SNPs was incorrectly classified as WT. Nevertheless, the final viral

lineage classification for the 91 samples after discrepancy testing showed 100%

agreement with viral WGS **(Fig. 3D and Table S1)**.

Fig. 3. Evaluation of the DETECTR® assay compared to SARS-CoV-2 Whole-

Genome Sequencing. (A) MA-plots showing CasDx1 SNP detection replicates (n =

807) for each SARS-CoV-2 mutation across 91 clinical samples. WT is denoted by blue

dots, MUT is denoted by green dots, NoCall is denoted by orange dots and NTC is

denoted by grey dots. (**B**) Alignment of final mutation calls comparing the DETECTR®

and SARS-CoV-2 WGS assay results across 91 clinical samples after discordant

samples (indicated by red asterisk) were resolved. (**C**) Final lineage classification on

each clinical sample by the DETECTR® assay compared to the SARS-CoV-2 lineage

determined by viral WGS. (**D**) Final Positive Predictive Agreement (PPA), Negative

Predictive Agreement (NPA) and concordance values for each WT and MUT SNP from

the evaluation of the DETECTR® assay against the SARS-CoV-2 WGS comparator

assay after discordant samples were resolved.

Discussion

In this study, we developed a CRISPR-based DETECTR® assay for the detection of

SARS-CoV-2 variants. We evaluated three CRISPR-Cas12 enzymes, two commercially

available (LbCas12a from NEB and AsCas12a from IDT) and one proprietary (CasDx1

from Mammoth Biosciences). Based on a head-to-head comparison of these enzymes,

we observed clear differences in performance, with CasDx1 demonstrating the highest

fidelity as the only enzyme able to reliably detect all three of the targeted SNPs. A data

analysis pipeline was developed to differentiate between WT and MUT signals with the

DETECTR® assay, yielding an overall SNP concordance of 100% (272/272 total SNP

calls) and 100% agreement with lineage classification compared to viral WGS. Taken

together, these findings show robust agreement between the DETECTR® assay and

viral WGS for identification of SNP mutations and variant categorization. Thus, the

DETECTR® assay provides a faster and simpler alternative to sequencing-based

methods for COVID-19 variant surveillance.

Our results show that the choice of Cas enzyme is important to maximize the accuracy

of CRISPR-based diagnostic assays and may need to be tailored to the site that is

being targeted. As currently configured with only three SNP targets, the DETECTR®

assay cannot resolve individual major variants, except for Alpha. However, given the

rapid emergence and dynamic shifts in the distribution of variants over time (*13*), it is

likely that tracking of key mutations, several of which are suspected to arise by

convergent evolution (*36*), rather than tracking of variants, will be critical for surveillance

as the pandemic continues. Furthermore, we also develop a data analysis pipeline for

CRISPR-based SNP calling that can readily incorporate additional targets and offers a

blueprint for automated interpretation of fluorescent signals that will become more

complex as the degree of multiplexing increases.

Although CRISPR-based diagnostic assays have been previously demonstrated for the

detection of SARS-CoV-2 variants, these studies have limitations in coverage of

circulating lineages and in the extent of clinical sample evaluation. For example, the

miSHERLOCK variant assay uses LbCas12a (NEB) to detect N501Y, E484K and

Y144Del covering eight lineages (WA-1, Alpha, Beta, Gamma, Eta, Iota, Mu and Zeta)

and was tested only on contrived samples (RNA spiked into human saliva) (*20*).

Additionally, the SHINEv2 assay uses LwaCas13a to detect 69/70Del, K417N/T, L452R

and 156/157Del + R158G covering eight lineages (WA-1, Alpha, Beta, Gamma, Delta,

Epsilon, Kappa and Mu) and was tested with only the 69/70Del gRNAs on 20 Alpha-

positive NP clinical samples (*37*). In comparison, the DETECTR® assay presented here

uses CasDx1 to detect N501Y, E484K and L452R covering 11 lineages (WA-1, Alpha,

Beta, Gamma, Delta, Epsilon, Eta, Iota, Kappa, Mu and Zeta) and 91 clinical samples

representing seven out of the 11 lineages were tested with successful detection of all 7.

Some limitations of our study are as follows. First, as previously mentioned, the

DETECTR® assay currently detects only three SNPs, which may not provide enough

resolution to identify a specific lineage. Second, we observed variable performance of

the assay in SNP discrimination, with more potential overlap in the calls between WT

and MUT for the 452 position than for the other two sites, increasing the risk of

misidentification. Third, our 484 gRNA was unable to differentiate E484Q from E484

(WT) in clinical samples, which could impact the accuracy of lineage classification (we

note, however, that the E484Q mutation corresponds to a different nucleotide position in

the affected 484 codon than the E484K mutation). These first 3 limitations could

potentially be addressed by the incorporation of additional gRNAs to the assay to

provide specific and redundant coverage and to improve identification of specific

lineages. Fourth, due to the multiplexed S-gene LAMP primer design, the limit of

detection of the DETECTR® assay is higher than our previously published SARS-CoV-2

DETECTR® assay (*19*), and thus only clinical samples with a Ct < 30 were tested in our

study. Incorporation of the N-gene target as a separate reaction in the assay may be

necessary for covering the dynamic range of COVID-19 positive samples if

simultaneous detection and SNP/variant identification is desired. Finally, the current

study focuses on the development and validation of a variant DETECTR® assay using

conventional laboratory equipment. Future work will involve implementation onto

automated, portable systems for use in point of care settings.

In the near term, we suggest the use of the DETECTR® assay as an initial screen for

the presence of a rare or novel variant (e.g., carrying both L452R and E484K or carrying

all three SNPs) that could be reflexed to viral WGS. As the sequencing capacity for

most clinical and public health laboratories is limited, the DETECTR® assay would thus

enable rapid identification of variants circulating in the community to support outbreak

investigation and public health containment efforts. Furthermore, identification of

specific mutations associated with neutralizing antibody evasion, such as E484K (*14*),

could inform patient care with regards to the use of monoclonal antibodies that remain

effective in treating the infection (*15*). As the virus continues to mutate and evolve, the

DETECTR® assay can be readily reconfigured by validating new gRNAs and pre-

amplification LAMP primers and gRNAs that target emerging mutations with clinical and

epidemiological significance. For example, we postulate that the newly emerging

Omicron variant, containing at least 30 mutations in the spike protein and 11 mutations

in the spike RGD region targeted by the assay, could be detected by increasing

degeneracy in the LAMP primers and adding at least one gRNA to be able to distinguish

this variant from the others. Over the longer term, a validated CRISPR assay that

combines SARS-CoV-2 detection with variant identification would be useful as a tool for

simultaneous COVID-19 diagnosis in individual patients and surveillance for infection

control and public health purposes.

Materials and Methods

Human Sample Collection and Ethics Statement

Remnant nasopharyngeal and/or oropharyngeal (NP/OP) samples and plasma

samples from laboratory confirmed SARS-CoV-2 positive patients were retrieved from

the UCSF Clinical Laboratories and stored in a biorepository until processed. Remnant

sample biobanking was performed with a waiver of consent and according to no-subject

contact study protocols approved by the UCSF Institutional Review Board (protocol

numbers 10-01116 and 11-05519).

Synthetic Gene Fragments

Wild-type (WT) and mutant (MUT) synthetic gene fragments (Twist) were PCR amplified

using NEB 2x Phusion Master Mix following the manufacturer's protocol. The amplified

product was cleaned using AMPure XP beads following manufacturers protocol at a

0.7x concentration. The product was eluted in nuclease-free water and normalized to 10

nM. All nucleic acids used in this study are summarized in **Table S2**.

Clinical sample acquisition and extraction

De-identified residual SARS-CoV-2 RT-PCR positive nasopharyngeal and/or

oropharyngeal (NP/OP) swab samples in universal transport media (UTM) or viral

transport media (VTM) were obtained from the UCSF Clinical Microbiology Laboratory.

All samples were stored in a biorepository according to protocols approved by the

UCSF Institutional Review Board (protocol number 10-01116, 11-05519) until

processed.

All NP/OP swab samples obtained from the UCSF Clinical Microbiology Laboratory

were pretreated with DNA/RNA Shield (Zymo Research, # R1100-250) at a 1:1 ratio.

The Mag-Bind Viral DNA/RNA 96 kit (Omega Bio-Tek, # M6246-03) on the KingFisher

Flex (Thermo Fisher Scientific, # 5400630) was used for viral RNA extraction using an

input volume of 200 µl of diluted NP/OP swab sample and an elution volume of 100 µl.

The Taqpath™ COVID-19 RT-PCR kit (Thermo Fisher Scientific) was used to determine

the N gene cycle threshold values.

Heat-inactivated culture acquisition and extraction

Heat-inactivated cultures of SARS-CoV-2 Variants Being Monitored (VBM), Variants of

Concern (VOC) or Variants of Interest (VOI) were provided by the California Department

of Public Health (CDPH).

RNA from heat-inactivated SARS-CoV-2 VBM/VOC/VOI isolates were extracted using

the EZ1 Virus Mini Kit v2.0 (Qiagen, # 955134) on the EZ1 Advanced XL (Qiagen, #

9001875) according to the manufacturer's instructions. For each culture, six replicate

LAMP reactions were pooled into a single sample. DETECTR® was performed on a

1:10 dilution of the 10,000 cp/rxn LAMP amplification products.

COVID-19 variant DETECTR® assay

Two LAMP primer sets, each containing 6 primers, were designed to target the L452R,

E484K and N501Y mutations in the SARS-CoV-2 Spike (S) protein (Supplemental

Table). Sets of LAMP primers were designed from a 350 bp target sequence spanning

the 3 mutations using Primer Explorer V5 (https://primerexplorer.jp/e/). Candidate

primers were manually evaluated for inclusion using the OligoCalc online

oligonucleotide properties calculator (*38*) while ensuring that there was no overlap with

either primers from the other set or guide RNA target regions that included the L452R,

E484K, and N501Y mutations.

Multiplexed RT-LAMP was performed using a final reaction volume of 50 µl, which

consisted of 8 µl RNA template, 5 µl of L452R primer set (Eurofins Genomics), 5 µl of

E484K/N501Y primer set, 17 µl of nuclease-free water, 1 µl of SYTO-9 dye

(ThermoFisher Scientific), and 14 µl of LAMP mastermix. Each of the primer sets

consisted of 1.6 µM each of inner primers FIP and BIP, 0.2 µM each of outer primers F3

and B3, and 0.8 µM each of loop primers LF and LB). The LAMP mastermix contained 6

mM of $MgSO_4$, isothermal amplification buffer at 1X final concentration, 1.5 mM of dNTP

mix (NEB), 8 units of Bst 2.0 WarmStart DNA Polymerase (NEB), and 0.5 ul of

WarmStart RTx Reverse Transcriptase (NEB). Plates were incubated at 65°C for 40

minutes in a real-time Quantstudio™ 5 PCR instrument. Fluorescent signals were

collected every 60 seconds

40nM CasDx1 (Mammoth Biosciences), LbCas12a (EnGen® Lba Cas12a, NEB) or

AsCas12a (Alt-R® A.s. Cas12a, IDT) protein targeting the WT or MUT SNP at L452(R),

E484(K) or N501(Y) was incubated with 40nM gRNA in 1X buffer (MBuffer3 for CasDx1,

NEBuffer r2.1 for LbCas12a and AsCas12a) for 30 min at 37^{o}C. Dx1 gRNAs were used

with both CasDx1 and LbCas12a, whereas AsCas12a gRNAs were used with

AsCas12a (**Table S2**). 100nM ssDNA reporter (/5Alex594N/TTATTATT/3IAbRQSp/,

IDT) was added to the RNA-protein complex. 18µL of this DETECTR$^®$ master mix was

combined with 2 µL target amplicon. The DETECTR$^®$ assays were monitored for 30 min

at 37^{o}C in a plate reader (Tecan).

Digital PCR

Samples were evaluated at 3 dilutions (1:100; 1:1,000; and 1:10,000) using the ApexBio

Covid-19 Multiplex Digital PCR Detection Kit (Stilla Technologies) according to the

manufacturer's protocol. The controls (positive and negative provided by UCSF, the Kit

Controls, and an internal control) were run with the samples in duplicate. The dilutions

were used to determine the most accurate concentration which was determined from

the N gene concentration.

Sequencing methods

Complementary DNA (cDNA) synthesis from RNA via reverse transcription and tiling

multiplexed amplicon PCR were performed using SARS-CoV-2 primers version 3

according to a published protocol (*39*). Libraries were constructed by ligating adapters

to the amplicon products using NEBNext Ultra II DNA Library Prep Kit for Illumina (New

England Biolabs, # E7645L), barcoding using NEBNext Multiplex Oligos for Illumina

(New England Biolabs, # E6440L), and purification with AMPure XP (Beckman-Coulter,

63880). Final pooled libraries were sequenced on either Illumina NextSeq 550 or

Novaseq 6000 as 1x300 single-end reads (300 cycles).

SARS-CoV-2 viral genome assembly and variant analyses were performed using an in-

house bioinformatics pipeline. Briefly, sequencing reads generated by Illumina

sequencers (NextSeq 550 or NovaSeq 6000) were demultiplexed and converted to

FASTQ files using bcl2fastq (v2.20.0.422). Raw FASTQ files were first screened for

SARS-CoV-2 sequences using BLASTn (BLAST+ package 2.9.0) alignment against the

Wuhan-Hu-1 SARS-CoV-2 viral reference genome (NC_045512). Reads containing

adapters, the ARTIC primer sequences, and low-quality reads were filtered using

BBDuk (version 38.87) and then mapped to the NC_045512 reference genome using

BBMap (version 38.87). Variants were called with CallVariants and a depth cutoff of 5

was used to generate the final assembly. Pangolin software (version 3.0.6) (*40, 41*) was

used to identify the lineage. Using a custom in-house script, consensus FASTA files

generated by the genome assembly pipeline were scanned to confirm L452R, E484K,

and N501Y mutations.

Discordant sample retesting

Eleven samples were re-extracted as described above for the NP/OP swab samples

and evaluated by viral WGS as described above. The samples were then thawed (XXX

freeze/thaws) and amplified using the LAMP protocol described above and evaluated

using the DETECTR® assay as described above.

DETECTR® data analysis pipeline

<u>Quality Control Metric for the LAMP Reaction</u>

Prior to processing DETECTR® data from the clinical samples, we collected data

indicating the success or failure of the samples to amplify in the LAMP reaction. The

absolute truth was based on visual inspection of LAMP curves This absolute truth was

used to develop thresholds for the LAMP reactions. The positive and negative controls

from the LAMP reactions were used to derive the thresholds to qualify the samples. Two

sets of thresholds were used: time threshold and fluorescence rate threshold. The

positive LAMP controls were assumed to represent an ideal sample and displayed a

classic sigmoidal rise of fluorescence over time and the NTC represented the

background fluorescence. It was hypothesized that a sample will ideally have positive

control like fluorescence kinetics. However due to the presence of high background in

some samples, a mean value between controls for each plate was chosen as threshold.

After this, the fluorescence values at a time threshold of 18 minutes were collected. The

time point is of importance here to rule out those samples that would amplify closer to

the endpoint, signifying the LAMP intermediates to be the majority contributors of the

rise in the signal and not the actual sample itself. A score was assigned for each sample

which was calculated as a ratio of rate of fluorescence rate threshold to the rate of

fluorescence value at 18 minutes for each sample. The hypothesis is that if this ratio of

rate of fluorescence between controls and samples is less than 1, then samples have

failed to reach the minimum fluorescence required to be called out as amplified and if

the ratio is greater than or equal to 1, then samples have amplified sufficiently. To

identify the exact score value for a qualitative QC metric, an ROC analysis was done on

scores and the absolute truth (**Fig. S6).**

<u>Data Analysis for CRISPR-based SNP calling</u>

Each well has a guide specific to the mutant or the wild-type SNP. The comparison is

important to assign a genotypic call to the sample. The DETECTR® reactions across the

plate are not comparable to each other. For this purpose, the endpoint fluorescence

intensities are normalized in each well to its own minimum intensity. This term is called

fluorescence yield. The fluorescence yield can be compared across wells in a plate

under the assumption that each well will have a similar minimum fluorescence starting

point. Irrespective of the highest levels of the fluorescence intensities observed across

samples, the yield for a given target must ideally remain the same assuming that similar

concentrations of samples/target are being compared. This aids in normalizing the

signal and comparing replicates across the wells in the same plate.

$$F_y = max(F)/min(F)$$

The wildtype and mutant target guides on NTC must ideally not show any change in

intensity over time. The fluorescence yield for NTC must remain constant across

replicates, plates and close to 1.

$$F_y(NTC) = 1$$

On the contrary, if a sample has a fluorescence yield of 1, then it qualifies for a No Call.

General rules for variant calling

1. NTC was assigned NTC

2. If the Contrast of the sample for a SNP was between minimum and maximum

contrast for the plate, then the sample is assigned a NoCall.

3. If the Size of the sample is lower than the Size of the NTC on the plate, then the

sample is assigned a NoCall.

$C_{min}(NTC\text{-}snp) <= C(sample\text{-}snp) <= C_{max}(NTC\text{-}snp) \rightarrow NoCall$

$S_{min}(NTC\text{-}snp) <= S(sample\text{-}snp) <= S_{max}(NTC\text{-}snp) \rightarrow NoCall$

$log2(Fy(WT)) > log2(Fy(M)) \rightarrow Wild\ Type$

$log2(Fy(WT)) < log2(Fy(M)) \rightarrow Mutant$

<u>SNP Calls</u>

We used the following procedure to evaluate the concordance between sequencing and

DETECTR® technologies for genotypic classification of the clinical cohort dataset.

First, we considered all samples and SNPs for which both sequencing and DETECTR®

data was present in the distributed files by matching the SNP IDs and sample names.

This included cleaning and curing the dataset which had failed LAMP reactions and

identifying WT and MUT based on the spacer fluorescent. This yielded a preliminary

data set containing 279 calls across three SNPs against 93 samples. After eliminating

samples that had failed to amplify in the LAMP reaction but were assigned a genotype,

the resulting final analysis data consisted of 272 calls (WT, MUT and NoCall) spread

across three SNPs and 91 samples. For each of the three SNPs in the analysis data

set, we identified and recorded both sequencing and DETECTR® genotypes (including

NoCalls and LAMP Fails) for each of the 93 patients. The 91 patients include the

individuals for whom actual sequencing data was available.

Statistical analysis

SNP Calls

For each SNP in the analysis, we computed a variety of statistics evaluating the

concordance between genotype calls on the two different technologies. The concordant

and discordant genotypes were visualized through contingency tables. For each SNP,

there are three possible genotypes (WT, MUT and No Call). The concordance rates

were calculated without the samples that failed the LAMP reaction **(Fig. 3B and Table**

S1). The 2x2 cross tables classify all three SNPs across all the samples between

sequencing and DETECTR® technologies **(Fig. 3B and Table S1)**. The data

transformation and statistical analysis was done in R (*42*).

Acknowledgments

We thank the UCSF Center for Advanced Technology core facility (Delsy

Martinez and Tyler Miyasaki) for their efforts in high-throughput sequencing of viral

cDNA libraries using the Illumina NovaSeq 6000 instrument, and Mary Kate Morris from

the California Department of Public Health for providing the heat-inactivated viral

cultures. We also thank Lucas Harrington and Teresa Peterson for their critical review of

this manuscript.

This work has been funded by the Innovative Genomics Institute (IGI) at UC

Berkeley and UC San Francisco (C.Y.C.), the Sandler Program for Breakthrough

Biomedical Research (C.Y.C.), US Centers for Disease Control and Prevention contract

75D30121C10991 (C.Y.C.), and Mammoth Biosciences.

Author contributions

C.L.F., J.S.C., and C.Y.C. conceived and designed the study. C.Y.C, and V.S.

coordinated the SARS-CoV-2 whole-genome sequencing efforts and RT-LAMP primer

design and testing. C.L.F., J.P.B., J.C., and J.S.C. designed guide RNAs for CRISPR-

Cas12 testing. B.M., J.P.B., R.N.D., E.S., and C.G.H. tested guide RNAs and ran

DETECTR® experiments. V.S., N.B., B.W., A.S.-G., K.R., J.S., S.M., and C.Y.C.

collected samples. C.L.F., V.S., B.M., V.N., J.P.B., J.C., J.S.C., and C.Y.C. analyzed

data. C.L.F., V.S., V.N., B.M., J.P.B., J.C., J.S.C., and C.Y.C. wrote the manuscript. All

authors read the manuscript and agree to its contents.

Competing interests

C.Y.C. is the director of the UCSF-Abbott Viral Diagnostics and Discovery Center and

receives research support from Abbott Laboratories, Inc. C.L.F., B.M., V.N., J.P.B.,

R.N.D., E.S., C.G.H., J.C., and J.S.C. are employees of Mammoth Biosciences. C.Y.C.

is a member of the scientific advisory board for Mammoth Biosciences. The other

authors declare no competing interests.

Data and Materials Availability

All data needed to evaluate the conclusions in the paper are present in the paper and/or

the Supplementary Materials. The CasDx1 protein can be provided by Mammoth

Biosciences to the extent feasible, pending scientific review and a completed material

543   transfer agreement. Requests for the CasDx1 protein should be submitted to Janice

544   Chen at janice@mammothbiosci.com.

Supplementary Figures

Fig. S1. DETECTR® curves from gene fragments and heat-inactivated viral

cultures. (**A**) Raw fluorescence curves from three Cas12 enzymes (CasDx1,

LbCas12a, AsCas12a) complexed with WT and MUT SNP gRNAs run on PCR-

amplified gene fragments representing WT and MUT SNP targets. (**B**) Raw

fluorescence curves from three Cas12 enzymes (CasDx1, LbCas12a, AsCas12a) on

eight heat-inactivated viral culture samples from various SARS-CoV-2 lineages, a no

target control (RT-LAMP) and CasDx1 detection controls (WT, MUT and NTC). CasDx1

replicates (n = 3), ±1.0SD

562
Fig. S2. Raw fluorescence RT-LAMP curves for each clinical sample. The raw

fluorescence RT-LAMP amplification curves for each of the clinical samples analyzed (n

= 3 replicates). Each line is representative of the median ±1.0SD of the three RT-LAMP

replicates for each sample. RT-LAMP replicates that passed QC are represented in

navy blue and failed LAMP replicates are shown in orange. Only valid RT-LAMP

replicates were used in subsequent data analysis.

Fig. S3. Raw fluorescence CasDx1 curves for each clinical sample amplified by

RT-LAMP. Each clinical sample was amplified with RT-LAMP in triplicate, and the

resulting amplicons were detected by CasDx1 in triplicate. The raw fluorescence curves

show WT detection in blue and MUT detection in green. Each line is representative of

the median ±1.0SD of the CasDx1 replicates (n = 3) for each WT and MUT guide for

each of the RT-LAMP replicates (n = 3), represented by different patterned lines.

Fig. S4. Evaluation of the DETECTR® data analysis pipeline and making final calls.

(**A**) Allele discrimination plot visualizing the scaled signals from the COVID Variant

DETECTR® assay on gene fragments. The allele discrimination plots represent scatter

plots of scaled WT and MUT fluorescence values plotted against each other. (**B**)

Contrast-Size plots of the COVID Variant DETECTR® assay data on gene fragments to

decrease ambiguity of the scaled signals, a ratio of the WT and MUT transformed

values are plotted against the average of the WT and MUT transformed values on the

MA plot. (**C**) Allele discrimination plot visualizing the scaled signals from the COVID

Variant DETECTR® assay on clinical sample. (**D**) MA plots of the COVID Variant

DETECTR® assay on the gene fragments (n = 30 WT; n = 30 MUT for each SNP) and

no template controls (n = 33 WT; n = 33 MUT for each SNP) used to test the data

analysis.

Fig. S5. **Highly specific detection by CasDx1 for each SNP on RT-LAMP replicates**

from clinical samples. (A) The DETECTR® assay workflow from LAMP amplification to

SNP identification. (**B**) Schematic showing the relationship between clinical samples,

598 LAMP replicates and CasDx1 replicates that culminate in a final SNP call. (**C**) Heat map

599 showing CasDx1 signal (n = 3) per every LAMP replicate (n = 3) for each SNP on every

600 clinical sample reflecting samples prior to discordance testing.

606 **Fig. S6. Determination of RT-LAMP threshold with a ROC curve.**

607 Thresholds for LAMP quality analysis were derived to determine which samples had

608 amplified sufficiently. The exact score value for this qualitative QC metric was

609 determined using a ROC analysis.

Fig. S7. Visualization of SNP calls by the DETECTR® data analysis pipeline. Box

plots of all the clinical samples illustrate the spread of the scaled signals for each of the

samples across the replicates in the experiment. SNP calls were made on each sample

agreement with the median values depicted on the box plot of the sample, which also

provided an analytical confirmation of the DETECTR® results.

Fig. S8. Clinical evaluation results with clinical samples of uncertain integrity. (**A**) Raw fluorescence CasDx1 curves for the clinical samples with discordant DETECTR® and WGS results. WT detection is represented by blue lines and MUT detection is represented by green lines. Each line is representative of the median ±1.0SD of the CasDx1 replicates (n = 3) for each guide for each of the LAMP replicates (n = 3), and each RT-LAMP replicate is represented by different patterned lines. (**B**) Visualization of the COVID Variant DETECTR® and SARS-CoV-2 WGS assays showing the alignment of final calls. Across all of the clinical samples in this cohort, 80 out of the 91 clinical sample COVID Variant DETECTR® assay calls were consistent with the SARS-CoV-2 WGS calls. (**C**) Summary of re-testing of discordant samples from the original clinical sample shows nearly all SNP discrepancies are resolved.

Patient sample	DETECTR (R)				Whole Genome Sequencing					qPCR
Name	SNP 452	SNP 484	SNP 501	Generic Lineage Classification	SNP_452	SNP_484	SNP_501	Pango Lineage	Lineage/Variant Call	Ct
COVID-01	WT	WT	WT	SARS-CoV-2	WT	WT	WT	B.1.1.519	N/A	16.2
COVID-02	WT	WT	WT	SARS-CoV-2	WT	WT	WT	B.1.1.222	N/A	17.2
*COVID-03	WT	WT	WT	SARS-CoV-2	WT	WT	WT	B.1.243	N/A	13.9
COVID-04	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.427	Epsilon	15.9
COVID-05	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.427	Epsilon	22.6
COVID-06	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.429	Epsilon	15.7
COVID-07	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.427	Epsilon	16.1
COVID-08	WT	WT	WT	SARS-CoV-2	WT	WT	WT	B.1.309	N/A	19.6
COVID-09	WT	WT	WT	SARS-CoV-2	WT	WT	WT	B.1.1.348	N/A	17.2
COVID-10	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.429	Epsilon	13.1
COVID-11	WT	WT	M	Alpha	WT	WT	M	B.1.1.7	Alpha	19.3
*COVID-12	WT	M	WT	Eta/Iota/Zeta	WT	M	WT	B.1.526	Iota	21.0
*COVID-13	M	WT	WT	Delta/Epsilon/Kappa	M	Q	WT	B.1.617.1	Kappa	17.3
COVID-14	WT	WT	M	Alpha	WT	WT	M	B.1.1.7	Alpha	14.9
COVID-15	WT	WT	WT	SARS-CoV-2	WT	WT	WT	B.1.575	N/A	18.4
COVID-16	WT	WT	M	Alpha	WT	WT	M	B.1.1.7	Alpha	21.6
COVID-17	WT	WT	M	Alpha	WT	WT	M	B.1.1.7	Alpha	13.2
COVID-18	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.280	N/A	13.9
COVID-19	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	18.9
*COVID-20	WT	WT	M	Alpha	WT	WT	M	B.1.1.7	Alpha	23.5
COVID-21	WT	WT	M	Alpha	WT	WT	M	B.1.1.7	Alpha	25.9
COVID-22	WT	WT	M	Alpha	WT	WT	M	B.1.1.7	Alpha	22.2
COVID-23	WT	M	M	Beta/Gamma/Mu	WT	M	M	P.1	Gamma	17.1
COVID-24	WT	WT	M	Alpha	WT	WT	M	B.1.1.7	Alpha	23.6
COVID-25	WT	M	M	Beta/Gamma/Mu	WT	M	M	P.1	Gamma	20.4
COVID-26	WT	M	M	Beta/Gamma/Mu	WT	M	M	P.1	Gamma	21.4
COVID-27	WT	M	M	Beta/Gamma/Mu	WT	M	M	P.1	Gamma	25.0
COVID-28	WT	M	M	Beta/Gamma/Mu	WT	M	M	P.1	Gamma	15.2
COVID-29	WT	WT	M	Alpha	WT	WT	M	B.1.1.7	Alpha	16.6
COVID-30	WT	M	M	Beta/Gamma/Mu	WT	M	M	P.1	Gamma	17.5
*COVID-31	WT	WT	M	Alpha	NoCall	WT	M	B.1.1.7	Alpha	23.6
COVID-32	WT	M	WT	Eta/Iota/Zeta	WT	M	WT	B.1.1.318	N/A	15.6
COVID-33	WT	WT	M	Alpha	WT	WT	M	B.1.1.7	Alpha	22.0
COVID-34	WT	M	M	Beta/Gamma/Mu	WT	M	M	P.1	Gamma	17.2
COVID-35	WT	M	M	Beta/Gamma/Mu	WT	M	M	P.1	Gamma	18.9
COVID-36	WT	M	M	Beta/Gamma/Mu	WT	M	M	P.1	Gamma	18.9
COVID-37	WT	M	M	Beta/Gamma/Mu	WT	M	M	P.1	Gamma	18.5
COVID-38	WT	M	M	Beta/Gamma/Mu	WT	M	M	P.1	Gamma	15.4
COVID-39	WT	M	M	Beta/Gamma/Mu	WT	M	M	P.1	Gamma	16.4
COVID-40	WT	M	M	Beta/Gamma/Mu	WT	M	M	P.1	Gamma	16.7
*COVID-41	WT	M	M	Beta/Gamma/Mu	WT	M	M	P.1	Gamma	24.2
COVID-42	WT	M	M	Beta/Gamma/Mu	WT	M	M	P.1	Gamma	17.7
COVID-43	WT	WT	M	Alpha	WT	WT	M	B.1.1.7	Alpha	15.3
COVID-44	WT	WT	M	Alpha	WT	WT	M	B.1.1.7	Alpha	17.9
COVID-45	WT	M	M	Beta/Gamma/Mu	WT	M	M	P.1	Gamma	16.6
COVID-46	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	23.2
COVID-47	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.526.1	N/A	16.6
COVID-48	WT	WT	M	Alpha	WT	WT	M	B.1.1.7	Alpha	13.7
COVID-49	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	17.9
COVID-50	WT	WT	M	Alpha	WT	WT	M	B.1.1.7	Alpha	15.8
COVID-51	WT	WT	M	Alpha	WT	WT	M	B.1.1.7	Alpha	15.8
COVID-52	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	AY.1	Delta	18.1
COVID-53	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	AY.1	Delta	22.9
COVID-54	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	AY.1	Delta	22.0
COVID-55	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	19.4
*COVID-56	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	21.1
COVID-57	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.526.1	N/A	26.6
COVID-58	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	17.8
COVID-59	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	AY.1	Delta	18.5
COVID-60	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	AY.1	Delta	21.5
*COVID-61	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	16.9
COVID-62	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	14.1
*COVID-63	M	WT	WT	Delta/Epsilon/Kappa	M	Q	WT	B.1.617.2	Delta †	12.2
COVID-64	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	22.0
COVID-65	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	21.9
COVID-66	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	14.8
COVID-67	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	18.7
COVID-68	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	16.6
COVID-69	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	19.0
COVID-70	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	21.1
COVID-71	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	14.0
COVID-72	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	25.6
*COVID-73	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	21.7
COVID-74	WT	M	M	Beta/Gamma/Mu	WT	M	M	B.1.623	N/A	23.7
COVID-75	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	17.7
COVID-76	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	AY.1	Delta	20.5
COVID-77	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	25.2
COVID-78	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	18.7
COVID-79	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	19.3
COVID-80	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	18.4
*COVID-81	WT	M	M	Beta/Gamma/Mu	WT	M	M	B.1.621	Mu	19.2
COVID-82	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	AY.1	Delta	18.3
COVID-83	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	A.2.5	Delta	17.6
COVID-84	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	AY.1	Delta	12.7
COVID-85	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	21.0
COVID-86	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	16.8
COVID-87	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	17.7
COVID-88	WT	M	WT	Eta/Iota/Zeta	WT	M	WT	B.1.526	Iota	14.8
COVID-89	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	20.5
COVID-90	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	B.1.617.2	Delta	21.6
COVID-91	M	WT	WT	Delta/Epsilon/Kappa	M	WT	WT	A.2.5	Delta	15.6

Table S1. Overall results summary of final SNP calls by the DETECTR® assay and viral WGS. A summary table of the final SNP calls from the DETECTR® assay and the SARS-CoV-2 whole genome sequencing assay after discordant testing. The table includes the lineage classification from DETECTR® calls as well as the PANGO lineage and WHO labels assigned to the WGS calls. Ct values from running an FDA EUA authorized SARS-CoV-2 RT-PCR assay, the Taqpath™ COVID-19 RT-PCR kit, are shown. Discordant samples were reflexed back for reprocessing (*); COVID-63 was classified as a Delta variant by WGS despite its Q484 SNP call. (†).

Type	Name	Sequence	Source
Guide RNA	Dx1_L452	UAAUUUCUACUAAGUGUAGAUuaaacaaucuauacagguaa	Dharmacon or Synthego
	Dx1_R452	UAAUUUCUACUAAGUGUAGAUccGguauagauuguuuagga	Dharmacon or Synthego
	Dx1_E484	UAAUUUCUACUAAGUGUAGAUaaaccuucaacaccauuaca	Dharmacon or Synthego
	Dx1_K484	UAAUUUCUACUAAGUGUAGAUaaaccuuUaacaccauuaca	Dharmacon or Synthego
	Dx1_N501	UAAUUUCUACUAAGUGUAGAUcaacccacuaauggguguugg	Dharmacon or Synthego
	Dx1_Y501	UAAUUUCUACUAAGUGUAGAUaacccUcuUaugguguuggu	Dharmacon or Synthego
	AsCas12a_L452	UAAUUUCUACUCUUGUAGAUuaaacaaucuauacagguaa	Dharmacon or Synthego
	AsCas12a_R452	UAAUUUCUACUCUUGUAGAUccGguauagauuguuuagga	Dharmacon or Synthego
	AsCas12a_E484	UAAUUUCUACUCUUGUAGAUaaaccuucaacaccauuaca	Dharmacon or Synthego
	AsCas12a_K484	UAAUUUCUACUCUUGUAGAUaaaccuuUaacaccauuaca	Dharmacon or Synthego
	AsCas12a_N501	UAAUUUCUACUCUUGUAGAUcaacccacuaauggguguugg	Dharmacon or Synthego
	AsCas12a_Y501	UAAUUUCUACUCUUGUAGAUaacccUcuUaugguguuggu	Dharmacon or Synthego
Reporter	ssDNA reporter	/5Alex594N/TTATTATT/3IAbRQSp/	IDT
Gene fragment	Mutant	tctgctttactaatgtctatgcagattcatttgtaattagaggtgatgaagtcagacaaatcgctccag ggcaaactggaaatattgctgattataattatataaattaccagatgatttacaggctgcgttatagctt ggaattctaacaatcttgattctaaggttggtggtaattataattaccggtatagattgtttaggaagtct aatctcaaacctttgagagagatatttcaactgaaatctatcaggccggtagcacaccttgtaatg gtgttaaaggttttaattgttactttcctttacaatcatatggtttccaacccacttatggtgttggttacc aaccatacagagtagtagtactttcttttgaacttctacatgca	Twist Biosciences
	Wild Type	tctgctttactaatgtctatgcagattcatttgtaattagaggtgatgaagtcagacaaatcgctccag ggcaaactggaaagattgctgattataattatataaattaccagatgattttacaggctgcgttatagctt ggaattctaacaatcttgattctaaggttggtggtaattataattacctgtatagattgtttaggaagtct aatctcaaacctttgagagagatatttcaactgaaatctatcaggccggtagcacaccttgtaatg gtgttgaaggttttaattgttactttcctttacaatcatatggtttccaacccactaatggtgttggttacc aaccatacagagtagtagtactttcttttgaacttctacatgca	Twist Biosciences
LAMP Primer	set1-F3	CAAACTGGAAAGATTGCTGA	Eurofins Genomics
	set1-B3	TACTACTACTCTGTATGGTTG	Eurofins Genomics
	set1-BIP	CCACCAACCTTAGAATCAAGATTGTTAAATTACCAGATGATTTTACAGGC	Eurofins Genomics
	set1-FIP	TCAACTGAAATCTATCAGGCCGGGGAAACCATATGATTGTAAAGGAA	Eurofins Genomics
	set1-LF	TAGAATTCCAAGCTATAACGCA	Eurofins Genomics
	set1-LB	TAGCACACCTTGTAATGGTG	Eurofins Genomics
	set2-F3	TTCTAAGGTTGGTGGTAATTATAATTA	Eurofins Genomics
	set2-B3	CATTGAAGTTGAAATTGACACAT	Eurofins Genomics
	set2-BIP	TGGGTTGGAAACCATATGATTGTAAATCTATCAGGCCGGTAGC	Eurofins Genomics
	set2-FIP	TGTTGGTTACCAACCATACAGAGTAAGACTTTTTAGGTCCACAAACA	Eurofins Genomics
	set2-LF	GTAAAGGAAAGTAACAATTAAAACCT	Eurofins Genomics
	set2-LB	AACTTCTACATGCACCAGCAA	Eurofins Genomics

650 **Table S2. Nucleic acid sequences used in this study.** A list of guide RNAs, reporter

651 molecules, LAMP primers and synthetic gene fragment targets with their respective

652 suppliers.

REFERENCES

1. S. P. Otto *et al.*, The origins and potential future of SARS-CoV-2 variants of concern in the evolving COVID-19 pandemic. *Curr Biol* **31**, R918-R929 (2021).

2. R. P. Walensky, H. T. Walke, A. S. Fauci, SARS-CoV-2 Variants of Concern in the United States-Challenges and Opportunities. *JAMA* **325**, 1037-1038 (2021).

3. X. Deng *et al.*, Genomic surveillance reveals multiple introductions of SARS-CoV-2 into Northern California. *Science* **369**, 582-587 (2020).

4. S. Truelove *et al.*, Projected resurgence of COVID-19 in the United States in July-December 2021 resulting from the increased transmissibility of the Delta variant and faltering vaccination. *medRxiv*, (2021).

5. N. L. Washington *et al.*, Emergence and rapid transmission of SARS-CoV-2 B.1.1.7 in the United States. *Cell* **184**, 2587-2594 e2587 (2021).

6. M. Okereke, Spread of the delta coronavirus variant: Africa must be on watch. *Public Health Pract (Oxf)* **2**, 100209 (2021).

7. E. C. Sabino *et al.*, Resurgence of COVID-19 in Manaus, Brazil, despite high seroprevalence. *Lancet* **397**, 452-455 (2021).

8. V. Servellita *et al.*, Predominance of antibody-resistant SARS-CoV-2 variants in vaccine breakthrough cases from the San Francisco Bay Area, California. *medRxiv*, (2021).

9. T. Kustin *et al.*, Evidence for increased breakthrough rates of SARS-CoV-2 variants of concern in BNT162b2-mRNA-vaccinated individuals. *Nat Med*, (2021).

10. J. Jung, H. Sung, S. H. Kim, Covid-19 Breakthrough Infections in Vaccinated Health Care Workers. *N Engl J Med* **385**, 1629-1630 (2021).

11. C. M. Brown *et al.*, Outbreak of SARS-CoV-2 Infections, Including COVID-19 Vaccine Breakthrough Infections, Associated with Large Public Gatherings - Barnstable County, Massachusetts, July 2021. *MMWR Morb Mortal Wkly Rep* **70**, 1059-1062 (2021).

12. F. Campbell *et al.*, Increased transmissibility and global spread of SARS-CoV-2 variants of concern as at June 2021. *Euro Surveill* **26**, (2021).

13. W. T. Harvey *et al.*, SARS-CoV-2 variants, spike mutations and immune escape. *Nat Rev Microbiol* **19**, 409-424 (2021).

14. W. F. Garcia-Beltran *et al.*, Multiple SARS-CoV-2 variants escape neutralization by vaccine-induced humoral immunity. *Cell* **184**, 2523 (2021).

15. D. Focosi, M. Tuccori, A. Baj, F. Maggi, SARS-CoV-2 Variants: A Synopsis of In Vitro Efficacy Data of Convalescent Plasma, Currently Marketed Vaccines, and Monoclonal Antibodies. *Viruses* **13**, (2021).

16. B. B. Oude Munnink *et al.*, The next phase of SARS-CoV-2 surveillance: real-
time molecular epidemiology. *Nat Med* **27**, 1518-1524 (2021).

17. RockefellerFoundation, "Implementation Framework: Toward a National Genomic
Surveillance Network," (Rockefeller Foundation, New York, 2021).

18. M. S. Verosloff *et al.*, CRISPR-Cas enzymes: The toolkit revolutionizing
diagnostics. *Biotechnol J*, e2100304 (2021).

19. J. P. Broughton *et al.*, CRISPR-Cas12-based detection of SARS-CoV-2. *Nat
Biotechnol* **38**, 870-874 (2020).

20. H. de Puig *et al.*, Minimally instrumented SHERLOCK (miSHERLOCK) for
CRISPR-based point-of-care diagnosis of SARS-CoV-2 and emerging variants.
Sci Adv **7**, (2021).

21. M. Patchsung *et al.*, Clinical validation of a Cas13-based assay for the detection
of SARS-CoV-2 RNA. *Nat Biomed Eng* **4**, 1140-1149 (2020).

22. P. Fozouni *et al.*, Amplification-free detection of SARS-CoV-2 with CRISPR-
Cas13a and mobile phone microscopy. *Cell* **184**, 323-333 e329 (2021).

23. FDA. SARS-CoV-2 RNA DETECTR Assay, (2020),
https://www.fda.gov/media/139937/download, date accessed: November 14,
2021.

24. FDA. Sherlock™ CRISPR SARS-CoV-2 kit, (2020),
https://www.fda.gov/media/137748/download, date accessed: November 14,
2021.

25. J. S. Gootenberg *et al.*, Multiplexed and portable nucleic acid detection platform
with Cas13, Cas12a, and Csm6. *Science* **360**, 439-444 (2018).

26. C. Chiu, Cutting-Edge Infectious Disease Diagnostics with CRISPR. *Cell Host
Microbe* **23**, 702-704 (2018).

27. L. Zhang *et al.*, Ten emerging SARS-CoV-2 spike variants exhibit variable
infectivity, animal tropism, and antibody neutralization. *Commun Biol* **4**, 1196
(2021).

28. K. S. Corbett *et al.*, SARS-CoV-2 mRNA vaccine design enabled by prototype
pathogen preparedness. *Nature* **586**, 567-571 (2020).

29. F. P. Polack *et al.*, Safety and Efficacy of the BNT162b2 mRNA Covid-19
Vaccine. *N Engl J Med* **383**, 2603-2615 (2020).

30. R. Bos *et al.*, Ad26 vector-based COVID-19 vaccine encoding a prefusion-
stabilized SARS-CoV-2 Spike immunogen induces potent humoral and cellular
immune responses. *NPJ Vaccines* **5**, 91 (2020).

31. F. Konings *et al.*, SARS-CoV-2 Variants of Interest and Concern naming scheme
conducive for global discourse. *Nat Microbiol* **6**, 821-823 (2021).

32. F. de Mello Malta *et al.*, Mass molecular testing for COVID19 using NGS-based
technology and a highly scalable workflow. *Sci Rep* **11**, 7122 (2021).

33. Z. Igloi *et al.*, Clinical Evaluation of Roche SD Biosensor Rapid Antigen Test for
SARS-CoV-2 in Municipal Health Service Testing Site, the Netherlands. *Emerg*
Infect Dis **27**, 1323-1329 (2021).

34. C. Broccanello *et al.*, Comparison of three PCR-based assays for SNP
genotyping in plants. *Plant Methods* **14**, 28 (2018).

35. F. E. McGuigan, S. H. Ralston, Single nucleotide polymorphism detection: allelic
discrimination using TaqMan. *Psychiatr Genet* **12**, 133-136 (2002).

36. J. Singh, P. Pandit, A. G. McArthur, A. Banerjee, K. Mossman, Evolutionary
trajectory of SARS-CoV-2 and emerging variants. *Virol J* **18**, 166 (2021).

37. J. Arizti-Sanz *et al.*, Equipment-free detection of SARS-CoV-2 and Variants of
Concern using Cas13. *medRxiv*, (2021).

38. W. A. Kibbe, OligoCalc: an online oligonucleotide properties calculator. *Nucleic*
Acids Res **35**, W43-46 (2007).

39. J. Quick *et al.*, Multiplex PCR method for MinION and Illumina sequencing of
Zika and other virus genomes directly from clinical samples. *Nat Protoc* **12**, 1261-
1276 (2017).

40. A. O'Toole *et al.*, Assignment of epidemiological lineages in an emerging
pandemic using the pangolin tool. *Virus Evol* **7**, veab064 (2021).

41. A. Rambaut *et al.*, A dynamic nomenclature proposal for SARS-CoV-2 lineages
to assist genomic epidemiology. *Nat Microbiol* **5**, 1403-1407 (2020).

42. R_Core_Team. R: A Language and Environmental for Statistical Computing,
(Vienna, Austria, 2018), https://www.R-project.org, date accessed: 11/26/21.