1 # Comparison of Saliva and Mid-Turbinate Swabs for Detection of COVID-19
2

3 Jianyu Lai, B.Med., M.P.H.[a,b], Jennifer German, Ph.D.[b], Filbert Hong, Ph.D.[b], S.-H. Sheldon Tai,
4 Ph.D.[b], Kathleen M. McPhaul, Ph.D., M.P.H.[b], Donald K. Milton, M.D., Dr.P.H.[b] for the
5 University of Maryland StopCOVID Research Group[*]
6
7 [a] Department of Epidemiology and Biostatistics, University of Maryland School of Public
8   Health, College Park, Maryland, USA
9 [b] Public Health Aerobiology and Biomarker Laboratory, Institute for Applied Environmental
10   Health, University of Maryland School of Public Health, College Park, Maryland, USA
11
12
13 [*] Dr. Donald K. Milton, University of Maryland School of Public Health, 4200 Valley Drive,
14 College Park, MD 20742 email: dmilton@umd.edu; phone: 301-405-0389
15 (Alternate: Dr. Filbert Hong, at the same address, email fhong@umd.edu, phone: 301-405-4081)
16

Abstract

Background: Saliva is an attractive sample for detecting SARS-CoV-2. However, contradictory reports exist concerning the sensitivity of saliva versus nasal swabs.

Methods: We followed close contacts of COVID-19 cases for up to 14 days from last exposure and collected self-reported symptoms, mid-turbinate swabs (MTS), and saliva every two or three days. Ct values, viral load, and frequency of viral detection by MTS and saliva were compared.

Results: 58 contacts provided 200 saliva-MTS pairs; 14 contacts (13 with symptoms) had one or more positive samples. Saliva and MTS had similar rates of viral detection (p=0.78) and substantial agreement (κ=0.83). However, sensitivity varied significantly with time since symptom onset. Early on (days -3 to 2), saliva had 12 times (95%CI: 1.2, 130) greater likelihood of viral detection and 3.2 times (95% CI: 2.8, 3.8) higher RNA copy numbers compared to MTS. After day 2 post-symptoms, there was a non-significant trend toward greater sensitivity using MTS.

Conclusion: Saliva and MTS demonstrated high agreement making saliva a suitable alternative to MTS for COVID-19 detection. Saliva was more sensitive early in the infection when transmission is most likely to occur, suggesting that it may be a superior and cost-effective screening tool for COVID-19.

Introduction

The U.S. Centers for Disease Control and Prevention (CDC) recommend the use of upper respiratory specimens, including but not limited to nasopharyngeal, mid-turbinate nasal, anterior nasal and saliva specimens for the initial diagnosis of COVID-19 [1]. Although nasopharyngeal swabs (NPS) are considered to be the standard for the detection of COVID-19 by most researchers, collection requires the use of trained professionals, can cause discomfort to the patients, and may pose greater risks to healthcare workers during sample collection [1–4]. Mid-turbinate swabs (MTS) are sometimes used as an alternative to NPS in an effort to reduce patient discomfort and occupational exposures to healthcare workers [4–6]. Compared to swab-based collection, saliva is even less invasive, more affordable, and can be self-collected with minimal or no supervision [1,7,8].

Existing studies focusing on the sensitivity of NPS compared to MTS, and NPS compared to saliva have produced contradictory results [2,4,9–11]. Few studies directly compare saliva and MTS specimens. Furthermore, there is strong evidence that pre-symptomatic transmission results in higher secondary attack rates for both symptomatic and asymptomatic transmission [12,13]. However, most of the existing studies only looked at detection of symptomatic cases after symptom onset [2,4,9,10] and few looked at detection sensitivity starting with the pre-symptomatic period. Therefore, research that conducts a direct comparison of MTS and saliva, including an assessment of sensitivity over time (starting during the pre-symptomatic period) is critical to identifying optimally sensitive methods for early detection and effective control of SARS-CoV-2 transmission.

The purpose of this study was to compare the sensitivity of MTS and saliva specimens for detecting SARS-CoV-2 by actively following close contacts of COVID-19 cases and collecting MTS and saliva samples for real-time reverse transcription polymerase chain reaction (RT-PCR) during their post-exposure quarantine period.

Results

We enrolled 58 individuals with known close contact with an active COVID-19 case. Contacts provided a total of 200 saliva and MTS pairs. The number of days of sample collection per participant ranged from one to seven. Among the contacts, 14 (24%) had at least one positive sample including 11 with both positive saliva and MTS samples over the course of follow-up. One contact had only positive saliva on 3 out of 3 samples (on days -3, 0, and 1 post symptom onset) and 2 had only positive MTS samples; one was positive on 2 of 2 swabs (on days 7 and 10) and one on 1 of 5 swabs (day 21, negative on days 14, 17, 19, and 24). Most of the positive participants (92.9%) were symptomatic, whereas only one (2.3%) participant from the test negative group reported symptoms. Symptomatic participants were enrolled -3 to 14 days since symptom onset and gave samples for up to 24 days from onset of symptoms. Symptoms were mild across the follow-up period. One participant had an oral temperature ≥ 38 °C at the time of sampling; three in total had temperatures ≥ 37.8 °C, and six in total had temperatures ≥ 37.5 °C; all were in the positive group. No other significant differences were identified between the positive and negative groups (Table 1).

Viral RNA detection in and agreement between saliva and MTS
Among 200 pairs of saliva and MTS samples we detected viral RNA in 32 (16%) of the saliva
and 29 (14.5%) of the MTS samples. The frequency of detection was similar for both sample
types (p=0.781) (see Supplementary Table S1 online). Cohen's Kappa demonstrated substantial
agreement (κ=0.83) with 26 (14%) positive and 165 (82.5%) negative sample pairs (Table 2).
The 14 participants who became positive by either sample type during the follow-up period
provided 41 saliva-MTS sample pairs, among which 71% of MTS and 78% of saliva samples
were positive (see Supplementary Table S2 online), without respect to time since symptom
onset. When focusing on positive participants, however, the agreement was weak (κ=0.43 for all
and κ=0.42 for those who were symptomatic) (see Supplementary Table S3a and Table S3b
online).
94
Comparison of Ct (cycle threshold) values between saliva and MTS
Each RT-PCR reaction contained 10 µL of heat-treated saliva sample or RNA extracted from
MTS. Assuming no loss in the process, each reaction represented 7.78 µL of saliva sample or 40
µL of the MTS eluate. The Ct values for paired samples were highly correlated (rho = 0.84, r2 =
0.74, Figure 1a). The Ct values for saliva were on average slightly but significantly greater than
for MTS samples (mean difference = 0.64, p=0.01) among all 58 participants (see Figure 1b and
Supplementary Fig. S1 online), partially reflecting the difference in their input amounts.
102
Relationship between days since symptom onset, probability of detection, and viral RNA copy
numbers
The Ct values among the positive symptomatic participants increased over time (days -3 through
24), along with decreasing viral RNA copy numbers. Saliva tended to have lower Ct values and
higher viral RNA copy numbers compared to MTS from days -3 to 1.5, whereas MTS samples
had lower Ct values and higher viral load thereafter (Figure 2A and Figure 2B).
109
Among symptomatic participants who had one or more positive saliva or MTS samples, the
probability (sensitivity) of detecting viral RNA in saliva samples was 91% (10/11) from day -3
to day 2 (Table 3), was 89% (16/18) from day 3 through 8, and declined significantly thereafter
(see Figure 2C and Supplementary Fig. S2 online). The probability of detecting virus in MTS
samples from day -3 through day 2 was 45% (5/11), was 94% (17/18) from day 3 through 8, and
then declined.
116
117 Early in the course of infection (days -3 through 2) saliva had 12 times the odds of being positive
118 (95% CI: 1.2, 130) and 3.2 times higher viral RNA copy numbers (95% CI: 2.8, 3.8) compared
119 to MTS. There was a trend toward greater sensitivity and higher viral RNA copy numbers in
120 MTS than saliva samples after day 2 post onset of symptoms (Table 3).
121
122 *Asymptomatic case*
123 Only one participant from our study population was an asymptomatic case. They provided one
124 pair of saliva and MTS samples, both of which were positive, with an average Ct value of 25.8
125 for MTS and 34.7 for saliva (see Supplementary Table S4 online).
126
Discussion

Early in the course of infection, saliva was significantly more sensitive than mid-turbinate nasal
swabs (MTS). We found that the optimal performance of saliva was in the pre-symptomatic
period and was more sensitive than MTS before symptom onset. Several studies have shown that
pre-symptomatic transmission plays a more important role than symptomatic and asymptomatic
transmission in the spread of SARS-CoV-2 [12,13]. Furthermore, saliva tended to have lower Ct
values and higher viral load compared to MTS from the pre-symptomatic period through the first
days post symptom onset. Together, these findings suggest that saliva may be the most effective
method for detecting SARS-CoV-2 early during the course of infection.
136
The CDC and the Infectious Disease Society of American recommendations for COVID-19
testing allow MTS, NPS, oral swabs, anterior nasal swabs, and saliva swabs as well as saliva.[1,14]
Some studies have shown differences in the sensitivity between NPS and MTS. In older, more
acutely ill populations, NPS appears to be more sensitive than MTS, especially later in the course
of illness (greater than 7 days) [4,9]. In a study of ambulatory and symptomatic participants whose
ages were more evenly distributed, NPS and MTS swabs were highly correlated with a mean of 7
days since onset of symptoms [15]. Congrave-Wilson et al., in agreement with the current study,
found that saliva had the highest sensitivity in the first seven days post COVID-19 onset when
using NPS as the reference [2]. Similarly, Savela et al. noticed that although the peak viral load of
SARS-CoV-2 in nasal swabs were higher than saliva, the latter was more likely to be positive in
the first six days since the participants' first positive sample was detected[16]. Becker et al.
compared the sensitivity of saliva and NPS for detecting COVID-19 in a convalescent cohort 8-
56 days since first symptom and found that NPS performed better [17]. They also showed that
saliva was about 30% less sensitive than NPS in a separate diagnostic cohort, however, days
since symptom onset were not reported, so we cannot make direct comparison with our findings.
Finally, a systematic review by Bastos et al. found that saliva had similar sensitivity to NPS and
costs less [11].
154
Our findings have implications for improving public acceptance of COVID-19 testing, reducing
the cost of mass COVID-19 screening, and improving the safety of healthcare workers who
conduct testing. These findings are extremely important when considering large-scale screening
of COVID-19 in schools and workplaces. In addition to its higher sensitivity in the early stage of
the disease as demonstrated in our data, saliva has quite a few other advantages that make it an
appealing screening tool. Saliva collection is less invasive and more acceptable to the general
population [7,18]. One of the barriers hindering COVID-19 testing is people's fear of nasal swabs
due to misinformation [19]. Also, the discomfort brought by nasal swabs may also reduce people's
willingness to get tested regularly, especially among children [20,21]. With the use of saliva,
screening large groups with increased frequency may be more practicable. Saliva is cheaper than
swab-based methods, especially if pooled samples are used [11,22]. Bastos et al. estimated that
when sampling 100,000 individuals, using saliva saved more than $600,000 in comparison to
using NPS [11], using a method that is more expensive than the SalivaDirect method used here.
These cost savings are especially important in the context of low resource settings.
169
Saliva collection is also safer for healthcare workers (HCWs). Amid the pandemic, one of the
key concerns among HCWs is the occupational exposure to SARS-CoV-2 aerosols during some
medical procedures [23]. The collection of nasal swabs introduces such exposure via the close
interaction between patients and HCWs and by patients' coughing and sneezing as a result of the

procedures [24]. In contrast, saliva is the only upper respiratory specimen suggested by the CDC
that can be self-collected without supervision [1] and hence protects HCWs from directly
contacting the patients when the samples are being collected. Given all these advantages of
saliva compared to NPS, our findings further support the use of saliva for large-scale screening,
especially of pre-symptomatic patients.
179

The current study has several limitations. The contacts enrolled in this study who eventually
tested positive for SARS-CoV-2 developed mild, and in some cases, transient infections. Some
of these mild (low viral titer) infections would not have been detected by less frequent testing
protocols and may not have posed a risk for onward transmission. The sample size of those who
tested positive is relatively small. Only one asymptomatic case was identified in our study so we
could not compare the sensitivity of the two types of samples among asymptomatic COVID-19
cases. The evidence for using saliva to detect asymptomatic cases was mixed in previous studies
[2,10] and further studies are needed to clarify this issue.
188

Future research should focus on the development of rapid saliva tests with high sensitivity and
specificity. Tng et al. proposed an amplified parallel antigen rapid test (AP-ART) using saliva to
test SARS-CoV-2 with a turnaround time of only 30 minutes[25]. This test was reported to have a
sensitivity as high as 97%. However, the researchers did not compare this AP-ART with saliva-
based RT-PCR but instead NPS-based RT-PCR and estimated the specificity of this test to be
only 90%. As our study showed, this is likely an underestimate of specificity because saliva can
be more sensitive than nasal swabs early in the course of the infection. Hence, further studies are
needed to evaluate saliva rapid tests with a reference method that is also based on saliva.
197

198   In conclusion, the use of saliva is preferable for testing pre-symptomatic populations. It is more
199   acceptable to people, which reduces barriers to testing. It is also more cost effective for
200   individuals to collect their own saliva rather than using highly trained professionals to collect
201   NPS and/or MTS. Finally, self-collected saliva samples eliminate the exposure to aerosols
202   produced by sneezing, coughing and gagging of patients undergoing NPS/MTS.
203

204   Methods
205   *Study population*
206   We analyzed MTS and saliva sample data from individuals who reported close contact with
207   confirmed COVID-19 cases as part of the University of Maryland StopCOVID study [26] from
208   May 2020 to April 2021.
209

Questionnaire and sample collection
Participants were followed every two or three days for up to 14 days from their last exposure or
until SARS-CoV-2 was detected in their samples. If one or more of their screening samples
became positive, results were confirmed by an appropriate clinical diagnostic test and they were
recruited to participate in the exhaled breath aspect of the study that also involved the collection
of saliva and MTS [26]. On each day of sample collection, participants answered an online
questionnaire to update their current symptoms and medications. For those who reported having
any symptom, they also reported their symptom onset date (i.e., "When did you begin to feel
sick?").
219

The symptoms checklist in the baseline and follow-up questionnaires, as previously described [26],
included runny nose, stuffy nose, sneezing, sore throat, earache, malaise, headache, muscle
and/or joint ache, sweat/feverish/chills, nausea, loss of appetite, vomiting, abdominal pain or
diarrhea, chest tightness, shortness of breath, and cough. Participants self-reported for each of
these 16 symptoms on a scale of 0 to 3 (0 = "no symptoms," 1 = "just noticeable," 2 = "clearly
bothersome from time to time, but didn't stop me from participating in activities," 3 = "quite
bothersome most or all of the time and stopped me from participating in activities").
227
228   For saliva collection, participants were instructed to not eat or drink 30 minutes prior to the visit
229   and then collect approximately 0.5-1 mL of saliva drooled into a plastic collection tube. For
230   MTS collection, trained clinical staff inserted a mid-turbinate swab approximately 1.5-2 inches
231   into one of the participants' nostrils, rotated once, and then withdrawn. This procedure was
232   repeated in the other nostril for a total of two MTS per participant per visit.
233

Laboratory analyses
Saliva samples were processed using the SalivaDirect method [8] as previously described [26].
Briefly, 50 µL of individual saliva samples were treated with Proteinase K (New England
Biolabs), heated at 95°C for 5 minutes, and kept at 4˚C. MTS from both nostrils were combined
and processed as previously described [26]. Briefly, total nucleic acid was extracted from 200 µL
of MTS with MagMax Pathogen RNA/DNA Kit (Applied Biosystems) on KingFisher Duo Prime
(Thermo Fisher Scientific), following the manufacturers' protocols. The sample was eluted in 50
µL of Elution Buffer and kept at 4 °C. MS2 phage was spiked in each heat-treated saliva sample
and extraction to control for extraction and PCR failure. RT-PCR was set up on the same day;
each reaction consisted of 1X TaqPath 1-Step Master Mix, No ROX, 1X TaqPath COVID-19
Real Time PCR Assay Multiplex (both from Thermo Fisher Scientific), and 10 µL of heat-
treated saliva or eluted nucleic acids. Each PCR plate contained a positive control provided in the
TaqPath COVID-19 Combo Kit (Thermo Fisher Scientific) and a no template control. Viral
loads in saliva and MTS were quantified as previously described[26]. RNA copy numbers were
reported per mL for saliva and per sample for MTS. The limit of detection was 75 copies per
sample and the limit of quantification was 250 copies per sample. A positive sample was defined
as having Ct values < 40 for at least two out of three SARS-CoV-2 targets (ORF1ab, N gene, and
S gene)[27]. The average Ct values of all positive targets were used in the following analyses.
252

Statistical analyses
We analyzed only paired same-day saliva and MTS samples to ensure the comparability of the
two samples. Group comparisons were made between participants having a positive result for
either sample and those with both samples being negative. Continuous variables (age and BMI)
were compared using t-test, and categorical variables were compared using Chi-square test (sex
and chronic respiratory illness) and Fisher's exact test (age group, race, and ever smoker).
259

260   To compare the Ct values from saliva and MTS, we conducted paired t-test and Bland-Altman
261   analysis, and calculated the coefficient of determination (i.e., R squared from linear regression)
262   and Pearson correlation coefficient. The Chi-square test was used to explore the relationship
263   between detection and sample types. Cohen's Kappa was calculated to demonstrate the degree of
264   agreement between the two sample types.
265

For participants with a positive saliva or MTS sample, we used a generalized additive logistic
model [28] to estimate and plot the probability of having a positive result by days since symptom
onset for the two sample types. We also created a plot using the LOESS (locally weighted
smoothing) method with 95% confidence interval for the change of Ct values and viral RNA
copy numbers by days since symptom onset for the two sample types. Logistic regression was
used to estimate the relative odds of detection of SARS-CoV-2 in saliva over specified intervals
since symptom onset.
273
274   For the estimate of geometric means of viral RNA copy numbers and the ratio of RNA copy
275   numbers of saliva to MTS, we applied linear mixed-effect models with censored responses[29,30] to
276   handle censored observations below the limit of detection and control for random effects of
277   subjects and sample nested within subjects.
278
279   All the analyses were carried out using RStudio and R (version 4.0.4) [31].
280

281   *Ethics statement*
282   This study was approved by the University of Maryland Institutional Review Board and the
283   Human Research Protection Office of the Department of the Navy. Electronically signed
284   informed consent was obtained from all participants and questionnaire data were collected and
285   stored with REDCap [32].
286

287   Data Availability
288   The datasets generated during and/or analyzed during the current study were deposited at Open
289   Science Framework (OSF) repository: DOI 10.17605/OSF.IO/9YP3Z (https://osf.io/9yp3z/)
290

References

1. Centers for Disease Control and Prevention. Interim Guidelines for Collecting and Handling of Clinical Specimens for COVID-19 Testing. *Centers for Disease Control and Prevention* https://www.cdc.gov/coronavirus/2019-ncov/lab/guidelines-clinical-specimens.html (2020).

2. Congrave-Wilson, Z. *et al.* Change in Saliva RT-PCR Sensitivity Over the Course of SARS-CoV-2 Infection. *JAMA* (2021) doi:10.1001/jama.2021.13967.

3. Lee, R. A., Herigon, J. C., Benedetti, A., Pollock, N. R. & Denkinger, C. M. Performance of Saliva, Oropharyngeal Swabs, and Nasal Swabs for SARS-CoV-2 Molecular Detection: a Systematic Review and Meta-analysis. *Journal of Clinical Microbiology* **59**, e02881-20.

4. Jamal, A. J. *et al.* Sensitivity of midturbinate versus nasopharyngeal swabs for the detection of severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2). *Infect Control Hosp Epidemiol* 1–3 doi:10.1017/ice.2020.1326.

5. Chu, C. Y. *et al.* Performance of saliva and mid-turbinate swabs for detection of the beta variant in South Africa. *Lancet Infect Dis* **21**, 1354 (2021).

6. Yp, T. *et al.* Swabs Collected by Patients or Health Care Workers for SARS-CoV-2 Testing. *The New England journal of medicine* **383**, (2020).

7. Tan, S. H., Allicock, O., Armstrong-Hough, M. & Wyllie, A. L. Saliva as a gold-standard sample for SARS-CoV-2 detection. *The Lancet Respiratory Medicine* **9**, 562–564 (2021).

8. Vogels, C. B. F. *et al.* SalivaDirect: A simplified and flexible platform to enhance SARS-CoV-2 testing capacity. *Med* **2**, 263-280.e6 (2021).

9. Pinninti, S. *et al.* Comparing Nasopharyngeal and Mid-Turbinate Nasal Swab Testing for the Identification of SARS-CoV-2. *Clin Infect Dis* (2020) doi:10.1093/cid/ciaa882.

10. Teo, A. K. J. *et al.* Saliva is more sensitive than nasopharyngeal or nasal swabs for diagnosis

of asymptomatic and mild COVID-19 infection. *Scientific Reports* **11**, 3134 (2021).

11. Bastos, M. L., Perlman-Arrow, S., Menzies, D. & Campbell, J. R. The Sensitivity and Costs

of Testing for SARS-CoV-2 Infection With Saliva Versus Nasopharyngeal Swabs : A

Systematic Review and Meta-analysis. *Ann Intern Med* **174**, 501–510 (2021).

12. Bender, J. K., Brandl, M., Höhle, M., Buchholz, U. & Zeitlmann, N. Analysis of

Asymptomatic and Presymptomatic Transmission in SARS-CoV-2 Outbreak, Germany,

2020. *Emerg. Infect. Dis.* **27**, (2021).

13. Qiu, X. *et al.* The role of asymptomatic and pre-symptomatic infection in SARS-CoV-2

transmission—a living systematic review. *Clinical Microbiology and Infection* **27**, 511–519

(2021).

14. Hanson, K. E. *et al.* The Infectious Diseases Society of America Guidelines on the Diagnosis

of COVID-19: Molecular Diagnostic Testing. *Clinical Infectious Diseases* ciab048 (2021)

doi:10.1093/cid/ciab048.

15. Tu, Y.-P. *et al.* Swabs Collected by Patients or Health Care Workers for SARS-CoV-2

Testing. *N Engl J Med* **383**, 494–496 (2020).

16. Savela, E. S. *et al.* Quantitative SARS-CoV-2 viral-load curves in paired saliva and nasal

swabs inform appropriate respiratory sampling site and analytical test sensitivity required for

earliest viral detection. *medRxiv* 2021.04.02.21254771 (2021)

doi:10.1101/2021.04.02.21254771.

17. Becker, D. *et al.* Saliva is less sensitive than nasopharyngeal swabs for COVID-19 detection

in the community setting. *medRxiv* 2020.05.11.20092338 (2020)

doi:10.1101/2020.05.11.20092338.

18. Byrne, R. L. *et al.* Saliva Alternative to Upper Respiratory Swabs for SARS-CoV-2

Diagnosis - Volume 26, Number 11—November 2020 - Emerging Infectious Diseases

journal - CDC. doi:10.3201/eid2611.203283.

19. Yangchen, S., Ha, S., Assan, A. & Tobgay, T. *Factors Influencing COVID-19 Testing: A*

Qualitative Study in Bhutan. https://www.researchsquare.com/article/rs-885659/v1 (2021)

doi:10.21203/rs.3.rs-885659/v1.

20. Clifford, V. & Curtis, N. Saliva testing for severe acute respiratory syndrome coronavirus 2

in children. *Clinical Microbiology and Infection* **27**, 1199–1201 (2021).

21. Leach, A. J., Stubbs, E., Hare, K., Beissbarth, J. & Morris, P. S. Comparison of Nasal Swabs

with Nose Blowing for Community-Based Pneumococcal Surveillance of Healthy Children.

Journal of Clinical Microbiology **46**, 2081–2082 (2008).

22. Watkins, A. E. *et al.* Increased SARS-CoV-2 Testing Capacity with Pooled Saliva Samples.

Emerg Infect Dis **27**, (2021).

23. Nicholson, P. J. & Sen, D. Healthcare workers and protection against inhalable SARS-CoV-2

aerosols. *Occupational Medicine* **71**, 118–120 (2021).

24. Qian, Y. *et al.* Safety management of nasopharyngeal specimen collection from suspected

cases of coronavirus disease 2019. *Int J Nurs Sci* **7**, 153–156 (2020).

25. Tng, D. J. H. *et al.* Amplified parallel antigen rapid test for point-of-care salivary detection

of SARS-CoV-2 with improved sensitivity. *Microchim Acta* **189**, 1–12 (2022).

26. Adenaiye, O. O. *et al.* Infectious SARS-CoV-2 in Exhaled Aerosols and Efficacy of Masks

During Early Mild Infection. *Clin Infect Dis* ciab797 (2021) doi:10.1093/cid/ciab797.

27. ThermoFisher Scientific. TaqPath^{TM} COVID-19 Combo Kit and TaqPath^{TM} COVID-19

Combo Kit Advanced* Instructions for Use. https://www.fda.gov/media/136112/download.

28. Hastie, T. & Tibshirani, R. Generalized additive models for medical research. *Stat Methods*

Med Res **4**, 187–196 (1995).

29. Vaida, F., Fitzgerald, A. P. & DeGruttola, V. Efficient Hybrid EM for Linear and Nonlinear

Mixed Effects Models with Censored Response. *Comput Stat Data Anal* **51**, 5718–5730

(2007).

30. Vaida, F. & Liu, L. Fast Implementation for Normal Mixed Effects Models With Censored

Response. *J Comput Graph Stat* **18**, 797–817 (2009).

31. R Core Team. *R: A Language and Environment for Statistical Computing*. (R Foundation for

Statistical Computing, 2021).

32. Harris, P. A. *et al.* The REDCap consortium: Building an international community of

software platform partners. *Journal of Biomedical Informatics* **95**, 103208 (2019).

Acknowledgements

We thank all the other members of the University of Maryland StopCOVID Research Group for their efforts in recruiting participants and sample collection and processing: Oluwasanmi Oladapo Adenaiye, Barbara Albert, P. Jacob Bueno de Mesquita, Yi Esparza, Aaron Kassman, Michael Lutchenkov, Dewansh Rastogi, Maria Schanz, Isabel Sierra Maldonado, Aditya Srikakulapu, Delwin Suraj, Faith Touré, Rhonda Washington-Lewis, Somayeh Youssefi, Stuart Weston, Matthew Frieman, Mara Cai, Ashok Agrawala. We also thank Dr. Jamal Fadul and his clinic in College Park, Maryland, for assistance in recruiting study participants.

Author Contributions

Conceptualization, D.M., J.G., F.H., S.T., J.L.; Data curation, F.H.; Formal analysis, J.L.; Investigation, S.T., J.G., J.L., K.M.; Methodology, D.M., J.G., S.T., J.L.; Project administration, D.M., F.H.; Writing-original draft, J. L., J.G.; Writing-review & editing, all authors; Supervision, D.M.

Competing Interests

This work was supported by Prometheus-UMD, sponsored by the Defense Advanced Research Projects Agency (DARPA) BTO under the auspices of Col. Matthew Hepburn through agreement N66001-18-2-4015. This work was also supported by the National Institute of Allergy and Infectious Diseases Centers of Excellence for Influenza Research and Surveillance (CEIRS) Contract Number HHSN272201400008C, and the Centers for Disease Control and Prevention Contract Number 200-2020-09528. The findings and conclusions in this report are those of the authors and do not necessarily represent the official position or policy of these funding agencies and no official endorsement should be inferred.

This work was also supported by a grant from the Bill & Melinda Gates Foundation, and a generous gift from The Flu Lab (https://theflulab.org). The funders had no role in study design, data collection, analysis, decision to publish, or preparation of the manuscript.

None of the authors have a potential conflicting interest or funding source.

403 Tables
404
405 Table 1. Characteristics of the study population

		Never positive	Positive for MTS or saliva	All participants
Number of participants		44	14	58
Number of sample pairs		159	41	200
Number of days of sample collection per participant, median (range)		4 (1, 7)	3 (1, 6)	3 (1, 7)
Female, N (%)		20 (46)	8 (57)	28 (48)
Age, mean ± SD		26.5 ± 15.5	27.3 ± 13.8	26.7 ± 15
Age group, N (%)				
<18		2 (4)	1 (7)	3 (5)
18-45		38 (86)	11 (79)	49 (84)
>45		4 (9)	2 (14)	6 (10)
White, N (%)		30 (68)	11 (79)	41 (71)
BMI, mean ± SD		25.6 ± 4.9	25.2 ± 4.4	25.5 ± 4.7
Chronic respiratory illness [a], N (%)		17 (39)	5 (36)	22 (38)
Ever smoker, N (%)		1 (2)	1 (7)	2 (3)
Ever symptomatic [b], N (%)		1 (2)	13 (93)	14 (24)
Symptomatic participants	Days since symptom onset at enrollment, median (range)	2 (-)	3 (-3, 14)	2.5 (-3, 14)
	Overall days since symptom onset of sample collection, median (range)	6.5 (2, 12)	5 (-3, 24)	5 (-3, 24)
	Loss of taste/smell, N (%)	0	2 (15)	2 (14)
	Median upper [c] respiratory symptoms (IQR)	3 (1.2, 6.2)	2 (0, 3)	2 (0.2, 3)
	Median lower respiratory symptoms (IQR)	0 (0, 0)	0 (0, 1)	0 (0, 1)
	Median systemic symptoms (IQR)	0.5 (0, 1.8)	0 (0, 1.2)	0 (0, 1.8)
	Median gastrointestinal symptoms (IQR)	0 (0, 0)	0 (0, 0)	0 (0, 0)
	Temperature, mean Celsius ± SD	37 ± 0.3	37.2 ± 0.5	37.2 ± 0.4

[a] Chronic respiratory illness = volunteers with any Chronic obstructive pulmonary disease,
Asthma, Other lung diseases.
[b] Group comparison, p<0.05.

[c] Symptoms at the time of each sample collection visit. Sixteen individual symptoms were rated
from 0 to 3. Systemic (max score of 12) = malaise + headache + muscle/joint ache +
sweats/fever/chills; Gastrointestinal (max score of 12) = loss of appetite + nausea + vomit +
diarrhea; Lower Respiratory (max score of 9) = chest tightness + shortness of breath + cough;
Upper Respiratory (max score of 15) = runny nose + stuffy nose + sneeze + earache + sore
throat.
415

416   Table 2. Viral RNA detection in paired saliva and MTS samples from all participants (N=58) [a]

| Saliva Positive | MTS Positive | | Total |
	No	Yes	
No	165	3	168
Yes	6	26	32
Total	171	29	200

417   [a] Cohen's Kappa between the two sample types was calculated as κ=0.83

419 Table 3. Sensitivity of saliva and MTS and relative odds of detection and ratio of viral RNA
420 copy numbers of SARS-CoV-2 in saliva by day since symptom onset among 13 mildly
421 symptomatic contacts of known cases

Days since symptom onset [a]	Saliva Positive/N (Sensitivity %)	MTS Positive/N (Sensitivity %)	Odds Ratio[b] Saliva:MTS	Estimates[c] Saliva:MTS
All samples [d]	31/40(78)	28/40(70)	1.5 (0.54, 4)	0.083 (0.069, 0.099)
-3 through 2	10/11 (91)	5/11 (45)	12 (1.2, 130)	3.2 (2.8, 3.8)
3 through 8	16/18 (89)	17/18 (94)	0.47 (0.037, 6)	0.03 (0.026, 0.036)
9 through 24	5/11 (45)	6/11 (55)	0.7 (0.13, 3.8)	0.065 (0.057, 0.073)

[a.] Days since symptom onset inclusive of the start and end day

[b.] Odds ratios and their 95% confidence intervals were estimated using logistic regression.

[c.] Effect estimates and their 95% confidence intervals are shown as the ratio of RNA copy
numbers of saliva to MTS. Analyses were controlled for random effects of subjects and sample
nested within subjects and for censoring by the limit of detection using a linear mixed-effects
model for censored responses (R Project lmec-package).

[d.] All samples from the 13 mildly symptomatic contacts of known cases with days since symptom
onset from day -3 through day 24.

Figures

Figure 1. Association between Ct values of saliva and MTS samples. Data were from 58
participants and 400 samples.
A. Scatter plot of Ct values of saliva and MTS
B. Bland-Altman plot for comparison of saliva and MTS
438

439
440
441
Figure 2. The change of Ct values and probability of testing positive by days since symptom
onset. Data were for MTS and saliva samples from 13 participants who provided a total of 40
pairs of samples and had one or more samples positive for SARS-CoV-2 RNA.
A. Change of Ct values by days since symptom onset,
B. Change of viral RNA copy numbers (natural log scale) by days since symptom onset,
C. Probability of being tested positive by days since symptom onset estimated from a generalized
additive logistic model.
449