medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Evaluation of the Roche SARS-CoV-2 Rapid Antibody Test in samples from vaccinated individuals

Johannes Hayer[1]

Eva Urlaub[1]

[1]Roche Diagnostics GmbH, Mannheim, Germany

Corresponding author: Eva Urlaub, Roche Diagnostics GmbH, Mannheim, Germany, eva.urlaub@roche.com

Running title

SARS-CoV-2 rapid POC antibody test in a vaccinated cohort

NOTE: This preprint reports new research that has not been certified by peer review and should not be used to guide clinical practice.

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Abstract

Objective

The study aimed to establish the performance of the SARS-CoV-2 Rapid Antibody Test (IgG and IgM) and the Elecsys® Anti-SARS-CoV-2 S assay in vaccinated individuals.

Methods

A panel of serum samples from Boca Biolistics was utilized to assess antibodies following vaccination, consisting of samples drawn prior to vaccination, after the first dose, or at least 14 days after the second dose of Moderna mRNA-1273 or Pfizer-BioNTech BNT162b2 COVID-19 vaccines. Agreement between the two methods was measured and stratified by test evaluator and assay lot.

Results

Agreement between the SARS-CoV-2 Rapid Antibody Test (IgG) and Elecsys Anti-SARS-CoV-2 S assay qualitative measurements at the different assessment points for both mRNA-1273 and BNT162b2 ranged between 97.06% (95% confidence interval [CI] 84.67, 99.93) to 100% (95% CI 82.35, 100). Agreement of the SARS-CoV-2 Rapid Antibody Test (IgG) with the Elecsys Anti-SARS-CoV-2 S assay was not highly influenced by either lot or evaluator. There was a medium-to-strong correlation between the semi-quantitative SARS-CoV-2 Rapid Antibody Test (IgG) result and quantitative Elecsys Anti-SARS-CoV-2 S assay in samples taken after both doses of the vaccines, with higher intensity bands being associated with higher total anti-S antibody titer (mRNA-1273, $p=0.0019$; BNT162b2, $p<0.0001$).

Conclusion

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Semi-quantitative SARS-CoV-2 Rapid Antibody Test (IgG) and quantitative Elecsys Anti-SARS-CoV-2 S assay correlated well, suggesting that the SARS-CoV-2 Rapid Antibody Test (IgG) is helpful in understanding the immune response post-vaccination. The current data support the use of the SARS-CoV-2 Rapid Antibody Test (IgG) in the vaccinated population.

Importance

Serologic assays are an essential tool for seroprevalence surveys, for quality control of vaccines, and to determine the response to vaccination. Although a correlate of immunity has not yet been established for COVID-19 vaccines, antibody titers after natural infection and vaccination have been associated with protection from symptomatic SARS-CoV-2 infection. Rapid point-of-care assays can be of use in this context with advantages over centralized testing, such as speed and ease of use. The point-of-care SARS-CoV-2 Rapid Antibody Test (IgG) compared favorably to the Elecsys Anti-SARS-CoV-2 S assay with agreement rates above 97.06%, after one or two doses of Moderna mRNA-1273 or Pfizer-BioNTech BNT162b2. Semi-quantitative SARS-CoV-2 Rapid Antibody Test (IgG) and quantitative Elecsys Anti-SARS-CoV-2 S assay results correlated well, suggesting that SARS-CoV-2 Rapid Antibody Test (IgG) is helpful in understanding the immune response post-vaccination. The current data support the use of the SARS-CoV-2 Rapid Antibody Test (IgG) in the vaccinated population.

Keywords

antibodies, neutralizing; COVID-19 vaccines; immunity; immunoassay; nucleocapsid protein; point-of-care systems; SARS-CoV-2; spike protein

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Introduction

Serologic immunoassays are vital in the global management of the COVID-19 pandemic. Measurement of antibodies to SARS-CoV-2 is important for improving disease management, and accurate tests, with validated sensitivity and specificity, are essential for obtaining reliable results when monitoring the pandemic via seroprevalence surveys (1). Serologic assays can provide evidence of recent or past infection (2), and identify those individuals still at risk of infection (3).

Infection with SARS-CoV-2 elicits a strong neutralizing antibody response (4), and antibodies that bind via the spike protein act as neutralizing antibodies (5, 6). Several studies have shown good correlation between anti-S antibodies and functional virus neutralization (7–11). Neutralization assays utilizing live virus require biosafety level 3 containment, which can be labor intensive and low throughput (12–14); therefore anti-S immunoassays, as a surrogate for neutralization, have an important role to play (13, 15). On an individual patient level, anti-S immunoassays can confirm vaccination status or inform upon the need for booster doses (16), which may be helpful for both social and medical reasons.

Moderna mRNA-1273 (spikevax™) and Pfizer-BioNTech BNT162b2 (COMIRNATY®) COVID-19 vaccines are mRNA vaccines, approved under the U.S. Food and Drug Administration's emergency use authorization, each employing a two-dose regimen (17, 18). The immune response involves B cells that produce binding and neutralizing antibodies against the SARS-CoV-2 S protein (19, 20) and both vaccines encode the SARS-CoV-2 spike protein to generate an immune response. Serologic assays are an essential tool in vaccine quality control, being used to determine the response to vaccination and providing insight into the antibody response from the individual patient perspective (21, 22).

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

The Elecsys® Anti-SARS-CoV-2 S assay (Roche Diagnostics International Ltd, Rotkreuz, Switzerland) is an electrochemiluminescence immunoassay (ECLIA), which has been developed for quantitative *in vitro* detection of antibodies in human serum and plasma (23). The assay detects total antibodies, including immunoglobulin G (IgG), against the receptor binding domain (RBD) of the SARS-CoV-2 spike protein, and is intended for use on the fully automated and high-throughput cobas e analyzers (23). The assay has been shown to detect anti-S antibodies across a variety of different populations (2, 24), and has been utilized as a surrogate of neutralization activity (11, 25).

The SARS-CoV-2 Rapid Antibody Test (Roche Diagnostics International Ltd, Rotkreuz, Switzerland) is CE-marked rapid chromatographic immunoassay intended for the qualitative *in vitro* detection of both IgM and IgG antibodies to SARS-CoV-2 spike or nucleocapsid proteins in human serum, plasma or whole blood, for professional use in the laboratory and at point-of-care (26). A study utilizing real-world clinical samples in unvaccinated individuals demonstrated that the SARS-CoV-2 Rapid Antibody Test had comparable performance with the Elecsys Anti-SARS-CoV-2 assay that targets the nucleocapsid protein (27).

The objective of this study was to establish the performance of the SARS-CoV-2 Rapid Antibody Test and the Elecsys Anti-SARS-CoV-2 S assay in serum samples obtained from individuals who had received a COVID-19 vaccination.

Materials and Methods

Study design

The study was a retrospective performance study conducted at Roche Diagnostics GmbH (Mannheim, Germany). All samples were analyzed with the SARS-CoV-2

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Rapid Antibody Test and compared with the Elecsys Anti-SARS-CoV-2 S assay. Tests were conducted as per manufacturer's instructions (23).

Samples

To assess antibodies following vaccination, a post-vaccination panel was prepared using 74 serum samples obtained from Boca Biolistics (Florida, USA), collected between February and April 2021, and stored at -20°C or colder (all available samples at the time of request). Serum samples were from individuals from the USA sampled prior to vaccination, after the first dose, or at least 14 days after the second dose of Moderna mRNA-1273 or Pfizer-BioNTech BNT162b2 (Supplemental Table 1).

Samples were separated into aliquots of 50 µL for the SARS-CoV-2 Rapid Antibody Test measurement and 200 µL for the Elecsys Anti-SARS-CoV-2 S reference measurement on the cobas® 6000 and stored at -80°C until use.

An in-house panel of 15 negative serum samples, previously determined negative by Elecsys Anti-SARS-CoV-2 S assay, was utilized to assess the specificity of the SARS-CoV-2 Rapid Antibody Test. Samples were collected in June 2021 as part of an in-house blood donation service at Roche Diagnostics (Mannheim, Germany), anonymized, and stored at -80°C until use. No information about vaccination status was available. Evaluators were blinded as to the origin of the samples. Cross-reactivity to other pathogens was not analyzed in this study as this has been evaluated elsewhere (26, 27).

SARS-CoV-2 Rapid Antibody Test

The SARS-CoV-2 Rapid Antibody Test detects the presence of SARS-CoV-2 IgM and IgG antibodies, and the intensity of the test line is dependent upon the level of

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

SARS-CoV-2 antibodies present in the sample (26). In unvaccinated individuals, the reported sensitivity is 92.59% (7–14 days post-symptom onset [POS]) and 99.03% (>14 days POS), with specificity of 98.65% (26).

Each sample was tested with two different lots of the SARS-CoV-2 Rapid Antibody Test. The readout was visually assessed at 10 minutes by two different independent evaluators; these evaluators were not the operators who had performed the test. In accordance with manufacturer's instructions, the sample was considered positive if the control line and at least one colored band/test line at G or M became apparent (qualitative determination). As an exploratory analysis, results were further defined by the intensity of the band to provide semi-quantitative data (Supplemental Table 2), with the signal intensity being rated according to a color scale. Data are presented separated by lot and by evaluator. In the event of an invalid result, the test would be repeated.

Elecsys Anti-SARS-CoV-2 S assay

The Elecsys Anti-SARS-CoV-2 S assay results are reported automatically as the analyte concentration of each sample in U/mL, with <0.80 U/mL interpreted as negative and ≥0.80 U/mL interpreted as positive for anti-SARS-CoV-2 S antibodies (23). The reported sensitivity and specificity of this test 14 days post-PCR is 97.92% and 99.95%, respectively (24). The quantification range is between 0.4 and 250 U/mL and the reported limit of detection is 0.35 U/mL (23). If the sample concentration is >250 U/mL, the sample can be diluted up to 1:100; in these cases, the final result is determined as >25,000 U/mL (23).

In the event of an invalid result, the test would be repeated. All samples were measured in duplicate, and the result calculated as the arithmetic mean of both

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

replicates. Assessors were blinded to the SARS-CoV-2 Rapid Antibody Test results and blinded as to the origin of the samples.

Statistical analysis

The agreement rates comparing the SARS-CoV-2 Rapid Antibody Test with Elecsys Anti-SARS-CoV-2 S assay per vaccine type were calculated per lot. Two-sided 95% Clopper–Pearson confidence intervals (CI) for the agreement rates are given. Correlation between the semi-quantitative SARS-CoV-2 Rapid Antibody Test data and quantitative Elecsys Anti-SARS-CoV-2 S assay data was assessed using the Kendall's rank correlation . Accuracy of the SARS-CoV-2 Rapid Antibody Test result — how often do the qualitative results agree with one another — was assessed by lot and evaluator. Statistical analyses were conducted using R (version 3.6.3) and JMP (version 15).

Results

Overall, the study cohort consisted of 56 predominantly white/Caucasian individuals (median age 70 years; 55.36% female) with a variety of pre-existing medical conditions (Table 1). No invalid results were generated using either test.

Analyses of the SARS-CoV-2 Rapid Antibody Test (qualitative) with the Elecsys Anti-SARS-CoV-2 S assay (quantitative and qualitative)

Eight samples were obtained from patients who were sampled prior to vaccination. Whilst only two patients were reported to have been previously diagnosed with COVID-19, only one patient was IgG and IgM negative (by both the SARS-CoV-2 Rapid Antibody Test and Elecsys Anti-SARS-CoV-2 S assay). For the seven positive samples, Elecsys Anti-SARS-CoV-2 S total antibody titers ranged from 254.0 U/mL

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

to >25,000 U/mL and all were identified as positive by the SARS-CoV-2 Rapid Antibody Test.

The post-vaccination antibody analysis included 25 samples from individuals vaccinated with mRNA-1273 and 42 vaccinated with BNT162b2. The dosing interval between first and second vaccine doses for BNT162b2 was a median of 21 days (minimum 19 days, maximum 23 days) and a median of 28 days (minimum 21 days, maximum 32 days) for mRNA-1273. There was good agreement between the SARS-CoV-2 Rapid Antibody Test (IgG) and Elecsys Anti-SARS-CoV-2 S qualitative measurements at the different assessment points for both the Moderna and Pfizer-BioNTech vaccines ranging from 97.06% (95% CI 84.67, 99.93) to 100% (95% CI 89.72, 100) (Table 2). There was one false-negative sample with evaluator 2 (after second dose of BNT162b2, lot 1 and lot 2), however this sample was classified as positive by evaluator 1.

There was poor agreement between the SARS-CoV-2 Rapid Antibody Test (IgM) and Elecsys Anti-SARS-CoV-2 S assay qualitative measurements at the different assessment points for both BNT162b2 and mRNA-1273 ranging from 14.71% (95% CI 4.95, 31.06) to 80.00% (95% CI 28.36, 99.49) with many false-negatives (Table 2).

Using the in-house negative panel, there were no false-positive results for either lot or evaluator (**Error! Reference source not found.**).

Analyses of the SARS-CoV-2 Rapid Antibody Test (semi-quantitative) with the Elecsys Anti-SARS-CoV-2 S assay (quantitative)

There was a medium-to-strong correlation between the semi-quantitative SARS-CoV-2 Rapid Antibody Test (IgG) and quantitative Elecsys Anti-SARS-CoV-2 S

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

assay in samples taken after both doses of the vaccines (Figure 1), with higher intensity bands being associated with higher total anti-S antibody titer (with Kendall's up to 0.599 for IgG mRNA-1273 [p=0.0019] and 0.6833 [p<0.0001] for IgG BNT162b2) (**Error! Reference source not found.**). There was no correlation between the quantitative Elecsys Anti-SARS-CoV-2 S assay result and the intensity of the SARS-CoV-2 Rapid Antibody Test (IgM) (**Error! Reference source not found.**).

Agreement between the semi-quantitative SARS-CoV-2 Rapid Antibody Test (IgM) and quantitative Elecsys Anti-SARS-CoV-2 S assay was poor, with many samples designated as negative on the SARS-CoV-2 Rapid Antibody Test (IgM) even in the presence of high anti-S antibody titers (Results from the SARS-CoV-2 Rapid Antibody Test were classified as negative or one of the five levels of increasing positivity based upon line intensity (((+)), (+), +, ++ or +++). Maximum value attainable by the Elecsys Anti-SARS-CoV-2 assay was 25,000 U/mL. Higher values are indicated as >25,000 U/mL. Blue = Lot 1; red = Lot 2; ○ = Evaluator 1; + = Evaluator 2.

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Figure 2).

Accuracy of the SARS-CoV-2 Rapid Antibody Test

Comparison of the SARS-CoV-2 Rapid Antibody Test (IgG) results by lot or evaluator revealed that these factors did not influence the results, with a minimum accuracy estimate of 97.06% (**Error! Reference source not found.**).

SARS-CoV-2 Rapid Antibody Test (IgM) positivity appeared to be influenced by evaluator, with a minimum accuracy estimate of 40% and considerable variability in accuracy estimates across the different time points and vaccines (**Error! Reference source not found.**).

Evaluation of Elecsys Anti-SARS-CoV-2 S assay titer post-vaccination

Antibody titers by days following final vaccination are shown in Figure 3. Following complete vaccination with BNT162b2 or mRNA-1273 and up to approximately 20 days, there was significant variability with both high (maximum >25,000 U/mL) and low (minimum 238.2 U/mL for BNT162b2 and 340.2 U/mL for mRNA-1273) anti-S titers. After 30 days, anti-S titers for both vaccines commonly declined to below 5,000 U/mL (overall minimum 13.53 U/mL 25 days after vaccination with BNT162b2).

The median time between final vaccination and final measurement for those vaccinated with BNT162b2 was 18 days (minimum 14 days, maximum 44 days) and 46 days (minimum 14 days, maximum 65 days) for mRNA-1273. Anti-S antibody titers were numerically higher in participants vaccinated mRNA-1273 (median 2,307 U/mL; interquartile range [IQR] 872–6,845) than in those who received BNT162b2 (median 1,601 U/mL; IQR 467–3,131).

Longitudinal analysis

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

A small number of individuals had samples available from all three time points (BNT162b2 cohort), allowing longitudinal evaluation of anti-S antibody titer levels; all except one showed high levels of anti-S antibody titers at the first time point (prior to vaccination). There was considerable variance in the effect of vaccination upon anti-S antibody titers, however the greatest change was for the individual who was infection-naïve (<0.4 U/mL to 1,645 U/mL following the second dose). For five out of eight individuals, including those with previously diagnosed COVID-19, anti-S antibody titers decreased over time despite vaccination with BNT162b2 (**Error! Reference source not found.**).

Discussion

Serologic assays are an essential tool for quality control of vaccines and to determine the response to vaccination. Although a correlate of immunity has not yet been established for COVID-19 vaccines, antibody titers after natural infection and vaccination have been associated with protection from symptomatic SARS-CoV-2 infection (28); and rapid point-of-care assays can potentially be of use in this setting with particular advantages over centralized testing, such as speed and ease of use.

We evaluated the use of the SARS-CoV-2 Rapid Antibody Test in vaccinated individuals in comparison with the Elecsys Anti-SARS-CoV-2 S assay.

Our results indicate that the Elecsys Anti-SARS-CoV-2 S assay detects anti-S antibodies after the first dose of both BNT162b2 and mRNA-1273, in agreement with previous studies of the adaptive humoral response to vaccination in various populations (21, 22, 29–33).

Whilst the performance of rapid antibody tests has been reported in unvaccinated individuals (34–36), no studies have yet reported upon their use in vaccinated

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

individuals. The SARS-CoV-2 Rapid Antibody Test (IgG) compared favorably to the Elecsys Anti-SARS-CoV-2 S assay with agreement rates above 97.06%, after one dose and two doses, similar to what has been observed in unvaccinated individuals (27). Assessment of the influence of lot or evaluator revealed that agreement of the SARS-CoV-2 Rapid Antibody Test (IgG) with the Elecsys Anti-SARS-CoV-2 S assay was not highly influenced by either factor, suggesting that the SARS-CoV-2 Rapid Antibody Test (IgG) is robust under normal usage conditions.

There was low agreement of the SARS-CoV-2 Rapid Antibody Test (IgM) with the Elecsys Anti-SARS-CoV-2 S assay because, in general, there is a lower sensitivity for IgM compared with IgG for point-of-care lateral flow assays, as exemplified in a recent analysis of healthcare workers vaccinated with the first dose of the Pfizer vaccine and sampled 21–24 days thereafter (37). In addition, the Elecsys Anti-SARS-CoV-2 S assay preferentially detects high-affinity IgG antibodies that compete with IgM antibodies (23), contributing to the lower agreement rates with IgM results. In agreement with previous research on the use of lateral flow assays in unvaccinated individuals (38, 39), we also report lower evaluator concordance when reading IgM compared with IgG bands. Jones and colleagues found that IgM bands are often weak positives (39), consistent with our data in vaccinated individuals, potentially resulting in an increased likelihood of a miscall and contributing to the low reader agreement.

The vaccines were administered in accordance with the relevant primary dosing schedules, with approximately 3 weeks for BNT162b2 and 1 month for mRNA-1273 between doses (17, 18). Whilst the data reported herein suggest some variation in the level of antibody titers afforded by the vaccines, two doses of either mRNA-1273 or BNT162b2 induced high levels of anti-S antibodies. Anti-S total antibody titers

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

were higher in participants vaccinated with mRNA-1273 than BNT162b2, in accordance with previous publications (40). Whilst the clinical significance of quantitative differences in antibody titers following COVID-19 vaccination have not yet been established, understanding the differences in antibody titer can support individual choices and assist with policy decisions.

The longitudinal analysis revealed that five individuals in the BNT162b2 cohort appeared to have previously undiagnosed SARS-CoV-2 infection prior to vaccination, and the effect of vaccination upon anti-S antibody titers in this group at the individual level was highly variable (relative decrease or increase after the first dose). Overall, after primary vaccination with two doses, all except one of those with previous natural infection had higher anti-S antibody titers than the infection-naïve participant, in agreement with previous reports which support higher titers following vaccination in previously infected individuals (40–42).

The findings of this study are subject to several limitations. The performance of the SARS-CoV-2 Rapid Antibody Test in vaccinated individuals was based on the evaluation of a relatively small panel of samples and utilized only two vaccines. However, the cohort tested was generally representative of an aged US population, with hypertension and diabetes mellitus in line with age-related expectations of prevalence (43, 44). Whilst the median anti-S antibody titers induced by vaccination differed between those vaccinated with BNT162b2 or mRNA-1273, the majority of the post-vaccination BNT162b2 samples were collected earlier than samples collected for mRNA-1273, there were more data points for BNT162b2, and the data for mRNA-1273 samples spanned a longer time frame (last sample was post 60 days for mRNA-1273 compared with <45 days for BNT162b2). Furthermore, most of

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

the samples used in the longitudinal analysis were from positive pre-vaccinated individuals, which limits our findings.

The clinical relevance of semi-quantitative results is currently unknown and, at present, cannot be interpreted as an indication of varying levels of immunity. In addition, semi-quantitative evaluation of the intensity of the bands is determined by the reader and therefore subjective, even with the use of a color scale guide and training. The performance of the SARS-CoV-2 Rapid Antibody Test seen herein can also not be extrapolated to immunocompromised individuals whose immune response post vaccination may be more variable (45). However, semi-quantitative SARS-CoV-2 Rapid Antibody Test (IgG) and quantitative Elecsys Anti-SARS-CoV-2 S assay results correlated well, suggesting that SARS-CoV-2 Rapid Antibody Test (IgG) is helpful in understanding the immune response post vaccination, particularly in the absence of centralized and fully automated testing systems. The current data support the use of the SARS-CoV-2 Rapid Antibody Test (IgG) in the vaccinated population.

Conflicts of interest

All authors have completed the ICMJE uniform disclosure form at www.icmje.org/coi_disclosure.pdf and declare: that JH and EU are employees of Roche Diagnostics; EU participates in Roche Connect; no other relationships or activities that could appear to have influenced the submitted work.

Ethical approval and consent

Anonymized samples were provided by the commercial supplier Boca Biolistics and collected under their ethical approval system. Ethical approval for the in-house panel was provided by Ethik-Kommission II der Universität Heidelberg, Medizinische

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Fakultät, Mannheim, Germany. Samples were anonymized, and collected and tested in accordance with applicable regulations, including relevant European Union directives and regulations, and the principles of the Declaration of Helsinki.

Data availability

Data available upon request. For more information on the study and data sharing, qualified researchers may contact the corresponding author, Eva Urlaub eva.urlaub@roche.com.

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Funding

Roche Diagnostics GmbH (Mannheim, Germany) provided financial support for the study and for preparation of the article. Roche Diagnostics GmbH was involved in the study design; in the collection, analysis and interpretation of data; in the writing of the report; and in the decision to submit the article for publication. Medical writing support was funded by Roche Diagnostics GmbH.

Acknowledgements

Medical writing support was provided by Corrinne Segal, Elements Communications Ltd, Westerham, UK. The authors would like to thank Rabea Held, Regina Draude, Maureen Adamietz, Shervan Horo, and Jlenia Marconi of Roche Diagnostics for their contribution to the study. COBAS, COBAS E, and ELECSYS are trademarks of Roche. All other product names and trademarks are the property of their respective owners'. The specific application (use in vaccinated individuals) is currently outside of the intended use of the SARS-CoV-2 Rapid Antibody Test. The Elecsys Anti-SARS-CoV-2 S assay is approved under an Emergency Use Authorization in the US.

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

References

1. Krammer F, Simon V. 2020. Serology assays to manage COVID-19. Science 368:1060–1061.

2. Riester E, Majchrzak M, Mühlbacher A, Tinguely C, Findeisen P, Hegel JK, Laimighofer M, Rank CM, Schönfeld K, Langen F, Laengin T, Niederhauser C. 2021. Multicentre performance evaluation of the Elecsys Anti-SARS-CoV-2 immunoassay as an aid in determining previous exposure to SARS-CoV-2. Infect Dis Ther doi:10.1007/s40121-021-00504-9:1–17.

3. Alter G, Seder R. 2020. The power of antibody-based surveillance. N Engl J Med 383:1782–1784.

4. Lau EHY, Tsang OTY, Hui DSC, Kwan MYW, Chan WH, Chiu SS, Ko RLW, Chan KH, Cheng SMS, Perera R, Cowling BJ, Poon LLM, Peiris M. 2021. Neutralizing antibody titres in SARS-CoV-2 infections. Nat Commun 12:63.

5. Yang J, Petitjean SJL, Koehler M, Zhang Q, Dumitru AC, Chen W, Derclaye S, Vincent SP, Soumillion P, Alsteens D. 2020. Molecular interaction and inhibition of SARS-CoV-2 binding to the ACE2 receptor. Nat Commun 11:4541.

6. Yan R, Zhang Y, Li Y, Xia L, Guo Y, Zhou Q. 2020. Structural basis for the recognition of SARS-CoV-2 by full-length human ACE2. Science 367:1444–1448.

7. Bal A, Pozzetto B, Trabaud MA, Escuret V, Rabilloud M, Langlois-Jacques C, Paul A, Guibert N, D'Aubarède-Frieh C, Massardier-Pilonchery A, Fabien N, Goncalves D, Boibieux A, Morfin-Sherpa F, Pitiot V, Gueyffier F, Lina B, Fassier JB, Trouillet-Assant S. 2021. Evaluation of high-throughput SARS-CoV-2

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

serological assays in a longitudinal cohort of patients with mild COVID-19: clinical sensitivity, specificity, and association with virus neutralization test. Clin Chem 67:742–752.

8. Bonelli F, Sarasini A, Zierold C, Calleri M, Bonetti A, Vismara C, Blocki FA, Pallavicini L, Chinali A, Campisi D, Percivalle E, DiNapoli AP, Perno CF, Baldanti F. 2020. Clinical and analytical performance of an automated serological test that identifies S1/S2-neutralizing IgG in COVID-19 patients semiquantitatively. J Clin Microbiol 58:e01224–20.

9. Muecksch F, Wise H, Batchelor B, Squires M, Semple E, Richardson C, McGuire J, Clearly S, Furrie E, Greig N, Hay G, Templeton K, Lorenzi JCC, Hatziioannou T, Jenks S, Bieniasz PD. 2021. Longitudinal serological analysis and neutralizing antibody levels in coronavirus disease 2019 convalescent patients. J Infect Dis 223:389–398.

10. Weidner L, Gänsdorfer S, Unterweger S, Weseslindtner L, Drexler C, Farcet M, Witt V, Schistal E, Schlenke P, Kreil TR, Jungbauer C. 2020. Quantification of SARS-CoV-2 antibodies with eight commercially available immunoassays. J Clin Virol 129:104540.

11. Rubio-Acero R, Castelletti N, Fingerle V, Olbrich L, Bakuli A, Wölfel R, Girl P, Müller K, Jochum S, Strobl M, Hoelscher M, Wieser A. 2021. In search of the SARS-CoV-2 protection correlate: Head-to-head comparison of two quantitative S1 assays in pre-characterized oligo-/asymptomatic patients. Infect Dis Ther 10:1505–1518.

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

12. Nie J, Li Q, Wu J, Zhao C, Hao H, Liu H, Zhang L, Nie L, Qin H, Wang M, Lu Q, Li X, Sun Q, Liu J, Fan C, Huang W, Xu M, Wang Y. 2020. Establishment and validation of a pseudovirus neutralization assay for SARS-CoV-2. Emerg Microbes Infect 9:680–686.

13. Oguntuyo KY, Stevens CS, Hung CT, Ikegame S, Acklin JA, Kowdle SS, Carmichael JC, Chiu HP, Azarm KD, Haas GD, Amanat F, Klingler J, Baine I, Arinsburg S, Bandres JC, Siddiquey MNA, Schilke RM, Woolard MD, Zhang H, Duty AJ, Kraus TA, Moran TM, Tortorella D, Lim JK, Gamarnik AV, Hioe CE, Zolla-Pazner S, Ivanov SS, Kamil JP, Krammer F, Lee B. 2021. Quantifying absolute neutralization titers against SARS-CoV-2 by standardized virus neutralization assay allows for cross-cohort comparisons of COVID-19 sera. mBio 12:e02492–20.

14. Vanderheiden A, Edara VV, Floyd K, Kauffman RC, Mantus G, Anderson E, Rouphael N, Edupuganti S, Shi PY, Menachery VD, Wrammert J, Suthar MS. 2020. Development of a rapid focus reduction neutralization test assay for measuring SARS-CoV-2 neutralizing antibodies. Curr Protoc Immunol 131:e116.

15. Jochum S, Kirste I, Hortsch S, Grunert VP, Legault H, Eichenlaub U, Kashlan B, Pajon R. 2021. Clinical utility of Elecsys Anti-SARS-CoV-2 S assay in COVID-19 vaccination: An exploratory analysis of the mRNA-1273 phase 1 trial. medRxiv doi:10.1101/2021.10.04.21264521:2021.10.04.21264521.

16. Baldanti F, Ganguly N, Wang G, Möckel M, O'Neill L, Renz H, Eduardo dos Santos Ferreira C, Tateda K, Van Der Pol B. 2021. Deployment of SARS-CoV-2

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

diagnostic testing: challenges and key considerations for the future. Clin Rev Lab Sci [submitted].

17. Pfizer-BioNTech. 2021. Fact sheet for healthcare providers administering vaccine (vaccination providers): emergency use authorization (EUA) of the Pfizer-BioNTech COVID-19 vaccine to prevent coronavirus disease 2019 (COVID-19). https://www.fda.gov/media/153713/download. Accessed 26 Nov 2021.

18. Moderna TX. 2021. Fact sheet for healthcare providers administering vaccine (vaccination providers): emergency use authorization (EUA) of the moderna COVID-19 vaccine to prevent coronavirus disease 2019 (COVID-19). https://www.fda.gov/media/144637/download. Accessed 26 Nov 2021.

19. Widge AT, Rouphael NG, Jackson LA, Anderson EJ, Roberts PC, Makhene M, Chappell JD, Denison MR, Stevens LJ, Pruijssers AJ, McDermott AB, Flach B, Lin BC, Doria-Rose NA, O'Dell S, Schmidt SD, Neuzil KM, Bennett H, Leav B, Makowski M, Albert J, Cross K, Edara VV, Floyd K, Suthar MS, Buchanan W, Luke CJ, Ledgerwood JE, Mascola JR, Graham BS, Beigel JH. 2021. Durability of responses after SARS-CoV-2 mRNA-1273 vaccination. N Engl J Med 384:80–82.

20. Mrak D, Tobudic S, Koblischke M, Graninger M, Radner H, Sieghart D, Hofer P, Perkmann T, Haslacher H, Thalhammer R, Winkler S, Blüml S, Stiasny K, Aberle JH, Smolen JS, Heinz LX, Aletaha D, Bonelli M. 2021. SARS-CoV-2 vaccination in rituximab-treated patients: B cells promote humoral immune responses in the presence of T-cell-mediated immunity. Ann Rheum Dis 80:1345–1350.

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

21. Lukaszuk K, Kiewisz J, Rozanska K, Dabrowska M, Podolak A, Jakiel G, Woclawek-Potocka I, Lukaszuk A, Rabalski L. 2021. Usefulness of IVD kits for the assessment of SARS-CoV-2 antibodies to evaluate the humoral response to vaccination. Vaccines 9:840.

22. Salvagno GL, Henry BM, di Piazza G, Pighi L, De Nitto S, Bragantini D, Gianfilippi GL, Lippi G. 2021. Anti-SARS-CoV-2 receptor-binding domain total antibodies response in seropositive and seronegative healthcare workers undergoing COVID-19 mRNA BNT162b2 Vaccination. Diagnostics 11:832.

23. Roche Diagnostics. 2021. Elecsys® Anti-SARS-CoV-2 S, instructions for use (09289275190, V3.0, Dec 2021).

24. Riester E, Findeisen P, Hegel JK, Kabesch M, Ambrosch A, Rank CM, Pessl F, Laengin T, Niederhauser C. 2021. Performance evaluation of the Roche Elecsys Anti-SARS-CoV-2 S immunoassay. J Virol Methods 297:114271.

25. Resman Rus K, Korva M, Knap N, Avšič Županc T, Poljak M. 2021. Performance of the rapid high-throughput automated electrochemiluminescence immunoassay targeting total antibodies to the SARS-CoV-2 spike protein receptor binding domain in comparison to the neutralization assay. J Clin Virol 139:104820.

26. Roche Diagnostics/SD BioSensor. 2020. SARS-CoV-2 Rapid Antibody Test, instructions for use (9901-NCOV-02C, V2.0, Sept 2020).

27. Findeisen P, Stiegler H, Lopez-Calle E, Schneider T, Urlaub E, Hayer J, Zemmrich C. 2021. Clinical performance evaluation of a SARS-CoV-2 Rapid

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Antibody Test for determining past exposure to SARS-CoV-2. Int J Infect Dis 103:636–641.

28. Khoury DS, Cromer D, Reynaldi A, Schlub TE, Wheatley AK, Juno JA, Subbarao K, Kent SJ, Triccas JA, Davenport MP. 2021. Neutralizing antibody levels are highly predictive of immune protection from symptomatic SARS-CoV-2 infection. Nat Med 27:1205–1211.

29. Kennedy NA, Lin S, Goodhand JR, Chanchlani N, Hamilton B, Bewshea C, Nice R, Chee D, Cummings JF, Fraser A, Irving PM, Kamperidis N, Kok KB, Lamb CA, Macdonald J, Mehta S, Pollok RC, Raine T, Smith PJ, Verma AM, Jochum S, McDonald TJ, Sebastian S, Lees CW, Powell N, Ahmad T. 2021. Infliximab is associated with attenuated immunogenicity to BNT162b2 and ChAdOx1 nCoV-19 SARS-CoV-2 vaccines in patients with IBD. Gut 70:1884–1893.

30. Callegaro A, Borleri D, Farina C, Napolitano G, Valenti D, Rizzi M, Maggiolo F. 2021. Antibody response to SARS-CoV-2 vaccination is extremely vivacious in subjects with previous SARS-CoV-2 infection. J Med Virol 93:4612–4615.

31. Cavalcanti E, Isgrò MA, Rea D, Di Capua L, Trillò G, Russo L, Botti G, Miscio L, Buonaguro FM, Bianchi AAM. 2021. Vaccination strategy and anti - SARS-CoV-2□S titers in healthcare workers of the INT - IRCCS "Fondazione Pascale" Cancer Center (Naples, Italy). Infect Agent Cancer 16:32.

32. Herishanu Y, Avivi I, Aharon A, Shefer G, Levi S, Bronstein Y, Morales M, Ziv T, Shorer Arbel Y, Scarfò L, Joffe E, Perry C, Ghia P. 2021. Efficacy of the

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

BNT162b2 mRNA COVID-19 vaccine in patients with chronic lymphocytic leukemia. Blood 137:3165–3173.

33. Seyahi E, Bakhdiyarli G, Oztas M, Kuskucu MA, Tok Y, Sut N, Ozcifci G, Ozcaglayan A, Balkan, II, Saltoglu N, Tabak F, Hamuryudan V. 2021. Antibody response to inactivated COVID-19 vaccine (CoronaVac) in immune-mediated diseases: a controlled study among hospital workers and elderly. Rheumatol Int 41:1429–1440.

34. Batra R, Olivieri LG, Rubin D, Vallari A, Pearce S, Olivo A, Prostko J, Nebbia G, Douthwaite S, Rodgers M, Cloherty G. 2020. A comparative evaluation between the Abbott Panbio™ COVID-19 IgG/IgM rapid test device and Abbott Architect™ SARS CoV-2 IgG assay. J Clin Virol 132:104645.

35. Naranbhai V, Chang CC, Beltran WFG, Miller TE, Astudillo MG, Villalba JA, Yang D, Gelfand J, Bernstein BE, Feldman J, Hauser BM, Caradonna TM, Alter G, Murali MR, Jasrasaria R, Quinlan J, Xerras DC, Betancourt JR, Louis DN, Schmidt AG, Lennerz J, Poznansky MC, Iafrate AJ. 2020. High seroprevalence of anti-SARS-CoV-2 antibodies in Chelsea, Massachusetts. J Infect Dis 222:1955–1959.

36. Pallett SJC, Rayment M, Patel A, Fitzgerald-Smith SAM, Denny SJ, Charani E, Mai AL, Gilmour KC, Hatcher J, Scott C, Randell P, Mughal N, Jones R, Moore LSP, Davies GW. 2020. Point-of-care serological assays for delayed SARS-CoV-2 case identification among health-care workers in the UK: A prospective multicentre cohort study. Lancet Respir Med 8:885–894.

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

37. Saluzzo F, Mantegani P, Poletti De Chaurand V, Quaresima V, Cugnata F, Di Serio C, Macé A, De Vos M, Sacks JA, Cirillo DM. 2021. SARS-CoV-2 Antibody Rapid Tests: Valuable Epidemiological Tools in Challenging Settings. Microbiol Spectr 9:e0025021.

38. Whitman JD, Hiatt J, Mowery CT, Shy BR, Yu R, Yamamoto TN, Rathore U, Goldgof GM, Whitty C, Woo JM, Gallman AE, Miller TE, Levine AG, Nguyen DN, Bapat S P, Balcerek J, Bylsma SA, Lyons AM, Li S, Wong AW, Gillis-Buck EM, Steinhart ZB, Lee Y, Apathy R, Lipke MJ, Smith JA, Zheng T, Boothby IC, Isaza E, Chan J, Acenas DD 2nd, Lee J, Macrae TA, Kyaw TS, Wu D, Ng DL, Gu W, York VA, Eskandarian HA, Callaway PC, Warrier L, Moreno ME, Levan J, Torres L, Farrington LA, Loudermilk RP, Koshal K, Zorn KC, Garcia-Beltran WF, Yang D, Astudillo MG, Bernstein BE, Gelfand JA, Ryan ET, Charles RC, Iafrate AJ, Lennerz JK, Miller S, Chiu CY, Stramer SL, Wilson MR, Manglik A, Ye CJ, Krogan NJ, Anderson MS, Cyster JG, Ernst JD, Wu AHB, Lynch KL, Bern C, Hsu PD, Marson A. 2020. Evaluation of SARS-CoV-2 serology assays reveals a range of test performance. Nat Biotechnol 38: 1174–1183.

39. Jones HE, Mulchandani R, Taylor-Phillips S, Ades AE, Shute J, Perry KR, Chandra NL, Brooks T, Charlett A, Hickman M, Oliver I, Kaptoge S, Danesh J, Di Angelantonio E, Wyllie D. Accuracy of four lateral flow immunoassays for anti SARS-CoV-2 antibodies: a head-to-head comparative study. 2021. EBioMedicine 68: 103414.

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

40. Steensels D, Pierlet N, Penders J, Mesotten D, Heylen L. 2021. Comparison of SARS-CoV-2 antibody response following vaccination with BNT162b2 and mRNA-1273. JAMA 326:1533–1535.

41. Di Meo A, Miller JJ, Fabros A, Brinc D, Hall V, Pinzon N, Ierullo M, Ku T, Ferreira VH, Kumar D, Pasic MD, Kulasingam V. 2021. Evaluation of three anti-SARS-CoV-2 serologic immunoassays for post-vaccine response. J Appl Lab Med Aug 3:jfab087.

42. Terpos E, Trougakos IP, Karalis V, Ntanasis-Stathopoulos I, Gumeni S, Apostolakou F, Sklirou AD, Gavriatopoulou M, Skourti S, Kastritis E, Korompoki E, Papassotiriou I, Dimopoulos MA. 2021. Kinetics of anti-SARS-CoV-2 antibody responses 3 months post complete vaccination with BNT162b2; A prospective study in 283 health workers. Cells 10:1942.

43. Centers for Disease Control and Prevention. 2020. National diabetes statistic report, 2020. https://www.cdc.gov/diabetes/pdfs/data/statistics/national-diabetes-statistics-report.pdf. Accessed 24 Nov 2021.

44. Ostchega R, Fyar C, Nwanko T, Nguyen D. 2020. Hypertension prevalence among adults aged 18 and over: United States, 2017–2018. NCHS Data Brief 364:1–8.

45. Lee A, Wong SY, Chai LYA, Lee SC, Lee MX, Muthiah MD, Tay SH, Teo CB, Tan BKJ, Chan YH, Sundar R, Soon YY. 2022. Efficacy of covid-19 vaccines in immunocompromised patients: systematic review and meta-analysis. BMJ 376:e068632.

Tables and figures for ms

Table 1: Demographics of the donors from the Boca Biolistics panel

Characteristics	**All** **N=56**	**Pfizer** **n=34**	**Moderna** **n=22**
Male/female, %	44.64/55.36	44.12/55.88	45.45/54.55
Median age, IQR	70 (61.75–72)	70 (63–71.75)	71 (57.25–73)
Ethnicity, %			
Asian; Not Hispanic/Latino,	1.79	2.94	0
Black/African American; Not Hispanic/Latino	3.57	5.88	0
White/Caucasian; Not Hispanic/Latino	64.29	61.76	68.18
Hispanic/Latino; Black/African American	7.14	8.82	4.55
Hispanic/Latino; White/Caucasian	23.21	20.59	27.27
Pre-existing medical conditions, %			
Yes/No	51.79/48.21	50.00/50.00	54.55/45.45
Hypertension, Yes/No	60.71/39.29	35.29/64.71	54.55/45.45

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Diabetes mellitus, Yes/No	7.14/92.86	11.76/88.24	0/100

IQR, interquartile range

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Table 2: Agreement between SARS-CoV-2 Rapid Antibody Test (IgG and IgM) and Elecsys Anti-SARS-CoV-2 S assay (reference test) qualitative measurements as measured by lot and evaluator using the Boca Biolistics donor panel, and after one dose or at least 14 days after two doses of Moderna mRNA-1273 or Pfizer-BioNTech BNT162b2

		mRNA-1273								**BNT162b2**							
		1 dose				**2 doses**				**1 dose**				**2 doses**			
		Evaluator 1		**Evaluator 2**		**Evaluator 1**		**Evaluator 2**		**Evaluator 1**		**Evaluator 2**		**Evaluator 1**		**Evaluator 2**	
		Lot 1	**Lot 2**	**Lot 1**	**Lot 2**	**Lot 1**	**Lot 2**	**Lot 1**	**Lot 2**	**Lot 1**	**Lot 2**	**Lot 1**	**Lot 2**	**Lot 1**	**Lot 2**	**Lot 1**	**Lot 2**
Ig G	**N**	5	5	5	5	19	19	19	19	8	8	8	8	34	34	34	34
	N +	5	5	5	5	19	19	19	19	7	7	7	7	34	34	34	34
	N -	-	-	-	-	-	-	-	-	1*	1*	1*	1*	-	-	-	-
	TP	5	5	5	5	19	19	19	19	7	7	7	7	34	34	33	33
	FN	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1
	PPA, %	*100*	*100*	*100*	*100*	*100*	*100*	*100*	*100*	*100*	*100*	*100*	*100*	*100*	*100*	*97.06*	*97.06*
	lower	*47.82*	*47.82*	*47.82*	*47.82*	*82.35*	*82.35*	*82.35*	*82.35*	*59.04*	*59.04*	*59.04*	*59.04*	*89.72*	*89.72*	*84.67*	*84.67*

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

	CI																
	upper CI	*100*	*100*	*100*	*100*	*100*	*100*	*100*	*100*	*100*	*100*	*100*	*100*	*100*	*100*	*99.93*	*99.93*
IgM	**N**	5	5	5	5	19	19	19	19	8	8	8	8	34	34	34	34
	N +	5	5	5	5	19	19	19	19	7	7	7	7	34	34	34	34
	N -	-	-	-	-	-	-	-	-	1*	1*	1*	1*	-	-	-	-
	TP	1	1	4	3	4	4	6	4	2	2	3	2	6	6	5	5
	FN	4	4	1	2	15	15	13	15	5	5	4	5	28	28	29	29
	PPA, %	*20.00*	*20.00*	*80.00*	*60.00*	*21.05*	*21.05*	*31.58*	*21.05*	*28.57*	*28.57*	*42.86*	*28.57*	*17.65*	*17.65*	*14.71*	*14.71*
	lower CI	*0.51*	*0.51*	*28.36*	*14.66*	*6.05*	*6.05*	*12.58*	*6.05*	*3.67*	*3.67*	*9.9*	*3.67*	*6.76*	*6.76*	*4.95*	*4.95*
	upper CI	*71.64*	*71.64*	*99.49*	*94.73*	*45.57*	*45.57*	*56.55*	*45.57*	*70.96*	*70.96*	*81.59*	*70.96*	*34.53*	*34.53*	*31.06*	*31.06*

*1 sample was detected as antibody-negative by the Elecsys Anti-SARS-CoV-2 S and the SARS-CoV-2 Rapid Antibody Test.

+, positive by the reference test (Elecsys Anti-SARS-CoV-2 S assay); -, negative by the reference test (Elecsys Anti-SARS-CoV-2 S assay); CI, 95% confidence intervals; TP, true-positive; FN, false-negative; PPA, positive percent agreement

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Figure 1: SARS-CoV-2 Rapid Antibody Test semi-quantitative IgG vs Elecsys Anti-SARS-CoV-2 S assay total antibody titer following vaccination with Moderna mRNA-1273 (left panel) or Pfizer-BioNTech BNT162b2 (right panel) at least 14 days following second dose



Results from the SARS-CoV-2 Rapid Antibody Test were classified as negative or one of the five levels of increasing positivity based upon line intensity (((+)), (+), +, ++ or +++). Maximum value attainable by the Elecsys Anti-SARS-CoV-2 assay was 25,000 U/mL. Higher values are indicated as >25,000 U/mL. Blue = Lot 1; red = Lot 2; ○ = Evaluator 1; + = Evaluator 2.

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Figure 2: SARS-CoV-2 Rapid Antibody Test semi-quantitative IgM vs Elecsys Anti-SARS-CoV-2 S assay total antibody titer following vaccination with Moderna mRNA-1273 (left panel) or Pfizer-BioNTech BNT162b2 (right panel) at least 14 days following second dose



Results from the SARS-CoV-2 Rapid Antibody Test were classified as negative or one of the five levels of increasing positivity based upon line intensity (((+)), (+), +, ++ or +++). Maximum value attainable by the Elecsys Anti-SARS-CoV-2 assay was 25,000 U/mL. Higher values are indicated as >25,000 U/mL. Blue = Lot 1; red = Lot 2; ○ = Evaluator 1; + = Evaluator 2.

medRxiv preprint doi: https://doi.org/10.1101/2021.12.17.21267927; this version posted April 6, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Figure 3: Elecsys Anti-SARS-CoV-2 S assay total antibody titer over time following vaccination with Moderna mRNA-1273 or Pfizer-BioNTech BNT162b2 (second dose)



Maximum value attainable by the Elecsys Anti-SARS-CoV-2 assay was 25,000 U/mL. Higher values are indicated as >25,000 U/mL. Blue = mRNA-1273; red = BNT162b2