

1 **Title:** Novel Application of a Force Sensor during Sit-to-Stands to Measure Dynamic Cerebral
2 Autoregulation Onset

3 **Running Title:** Sit-to-Stand Force Sensor to Measure dCA

4 Alicen A. Whitaker¹, Eric D. Vidoni^{2,3}, Stacey E. Aaron¹, Adam G. Rouse^{4,5,6}, Sandra A.
5 Billinger^{1,2,3,4,7}

6 ¹Department of Physical Therapy, Rehabilitation Science, and Athletic Training, University of
7 Kansas Medical Center, Kansas City, KS

8 ²University of Kansas Alzheimer's Disease Research Center, Fairway, KS

9 ³Department of Neurology, University of Kansas Medical Center, Kansas City, KS

10 ⁴Department of Molecular and Integrative Physiology, University of Kansas Medical Center,
11 Kansas City, KS

12 ⁵Department of Neurosurgery, University of Kansas Medical Center, Kansas City, KS

13 ⁶Department of Electrical Engineering and Computer Science, University of Kansas, Lawrence,
14 KS

15 ⁷Department of Physical Medicine and Rehabilitation, University of Kansas Medical Center,
16 Kansas City, KS

17 **Corresponding Author:** Sandra A. Billinger, PT, PhD; Email: sbillinger@kumc.edu; Phone:
18 (913) 945-6685; Address: 3901 Rainbow Blvd, Mail Stop 2002, Kansas City, KS 66160

19 **Author Contributions:** AW, EV, SB conceived and designed research; AW and SA performed
20 experiments; AW, EV, SA, SB analyzed data; AW, EV, SA, AR, SB interpreted results of
21 experiments; AW, EV, SA, AR, SB drafted manuscript; AW, EV, SA, AR, SB edited and
22 revised manuscript; AW, EV, SA, AR, SB approved final version of manuscript; AW prepared
23 figures.

24 **Abstract**

25 **Purpose:** Current sit-to-stand methods measuring dynamic cerebral autoregulation (dCA) do not
26 capture the precise onset of the time delay (TD) response. Reduced sit-to-stand reactions in older
27 adults and individuals post-stroke could inadvertently introduce variability, error, and imprecise
28 timing. We applied a force sensor during a sit-to-stand task to more accurately determine how
29 TD before dCA onset may be altered.

30 **Methods:** Middle cerebral artery blood velocity (MCAv) and mean arterial pressure (MAP) were
31 measured during two sit-to-stands separated by 15 minutes. Recordings started with participants
32 sitting on a force-sensitive resistor for 60 seconds, then asked to stand for two minutes. Upon
33 standing, the force sensor voltage immediately dropped and marked the exact moment of arise-
34 and-off (AO). Time from AO until an increase in cerebrovascular conductance (CVC =
35 MCAv/MAP) was calculated as TD.

36 **Results:** We tested the sensor in 4 healthy young adults, 2 older adults, and 2 individuals post-
37 stroke. Healthy young adults stood quickly and the force sensor detected a small change in TD
38 compared to classically estimated AO, from verbal command to stand. When compared to the
39 estimated AO, older adults had a delayed measured AO and TD decreased up to ~53% while
40 individuals post-stroke had an early AO and TD increased up to ~14%.

41 **Conclusion:** The transition reaction speed during the sit to stand has the potential to influence
42 dCA metrics. As observed in the older adults and participants with stroke, this response may
43 drastically vary and influence TD.

44

45 **Key Words:** dCA, Cerebral Blood Velocity, Middle Cerebral Artery Blood Velocity , Cerebral
46 Blood Flow

47 **New & Noteworthy**

48 We developed a force sensor and are the first to apply it during sit-to-stands measuring the
49 dynamic cerebral autoregulation (dCA) response. Current methodologies estimate the transition
50 reaction speed of the sit-to-stand, which influences the dCA metrics. Compared to estimating the
51 kinetic moment of arise-and-off (AO), our force sensor can detect early or delayed AO in
52 individuals who may present with lower extremity weakness such as older adults and clinical
53 populations such as stroke.

54 **Introduction**

55 The brain requires constant cerebral blood flow, even during perturbations to the systemic
56 circulation.(1) Dynamic cerebral autoregulation (dCA) is the ability of the brain to maintain
57 blood flow by altering cerebrovascular conductance (CVC) during perturbations to circulation,
58 such as changes in peripheral blood pressure.(2-4) Recording the dCA response has become
59 prevalent with the use of transcranial Doppler ultrasound to measure middle cerebral artery
60 blood velocity (MCAv) and finger plethysmography to measure mean arterial pressure
61 (MAP).(4, 5) During a transient drop in peripheral blood pressure, for example in the transition
62 from a seated position to standing, the time delay (TD) before the onset of the dCA response is
63 measured as the elapsed time before an increase in CVC.(6, 7)

64
65 A sit-to-stand transition is a common experimental perturbation utilized to measure the temporal
66 dCA response.(8) When performing a sit-to-stand, peripheral blood pressure transiently drops
67 due to gravitational venous pooling and vasodilation.(9) To counteract the drop in blood
68 pressure, dCA responds by increasing the conductance of MCAv to the cerebrovascular system
69 before MAP increases in the peripheral vascular system.(6)

70
71 While previous studies in healthy young and older adults have performed the sit-to-stand
72 technique to measure the dCA response,(6, 8, 10-15) the exact methods of characterizing dCA
73 have varied between studies. While some report having the individual stand quickly (within 0-3
74 seconds) and mark the initiation of standing as when the verbal command was given,(6, 10-12)
75 others report having the individual place their feet on a stool in front of them (legs at 90 degrees)
76 and mark the initiation of standing from when their feet touch the floor.(8, 13-15) However, the

77 moment of standing from a seated position is not objectively marked.(16) The start of the
78 postural change, the stimulus which drops peripheral blood pressure, is critical for measuring the
79 timing of the dCA response and has been indeterminate in prior protocols. Considering the sit-to-
80 stand transition influences hemodynamics, not measuring the exact moment could inadvertently
81 introduce variability, error, and imprecise timing of the dCA metrics. Measuring the exact
82 moment of stance is especially important in older adults and clinical populations, such as stroke
83 who may present with reduced lower extremity strength, resulting in a delayed sit-to-stand
84 reaction.(17-19) Current methodologies are unable to account for the potential delay between the
85 initiation and the stance phase of a sit-to-stand.(16)

86

87 To address this gap in the literature, we developed a force sensor to determine the kinetic
88 moment of arise-and-off (AO).(16) Compared to estimating AO from a verbal command similar
89 to previous work, (6, 10-12) we hypothesized that measuring the exact moment of AO,
90 measured with a force sensor, would alter the calculation of the TD before the onset of the dCA
91 response. We conducted the procedure twice separated by 15 minutes (T1= first sit to stand and
92 T2 = second sit to stand).

93

94 **Methods**

95 This is an ongoing study, “Blood Flow Response and Acute Interval Exercise (BRAIN), focused
96 on the characterization of the cerebrovascular hemodynamic response to a single bout of high
97 intensity interval exercise and the sit-to-stand dCA response before and after exercise
98 (NCT04673994). We developed the force sensor prior to study initiation and present limited data
99 to establish “proof of principle,” not as interim analyses of the a priori stated primary outcome of

100 the clinical trial. We recruited and enrolled healthy young adults, older adults, and individuals
101 post-stroke. Healthy young adults were included if they were between the ages of 18-30 years
102 old and classified as low cardiovascular risk using the American College of Sports Medicine
103 (ACSM) Cardiovascular Risk Screen.(20) Older adults and individuals post-stroke (6 months to
104 5 years post-stroke) were included if they were: 1) between the ages of 40-80 years old, 2)
105 performed less than 150 minutes of moderate intensity exercise per week, and 3) able to answer
106 consenting questions and follow a 2-step command. Individuals were excluded if they 1) were
107 unable to stand from a sitting position without physical assistance, 2) had a previous history of
108 another neurological disease (i.e. multiple sclerosis, Alzheimer's disease, or Parkinson's
109 disease), and 3) did not have a MCAv signal on the right or left temporal window using
110 transcranial Doppler ultrasound (TCD). The study was approved by the University of Kansas
111 Medical Center Institutional Review Board Human Subjects Committee.

112
113 Prior to study procedures, all participants provided written informed consent to participate and
114 we collected demographic information. Participants also completed a standardized 5 time sit-to-
115 stand assessment, to determine leg strength and balance by measuring how quickly an individual
116 could stand up from a chair and sit down 5 times without the use of their arms.(17, 21, 22)

117 118 Set Up

119 Consistent with previously published protocols, our laboratory room was dimly lit and kept at a
120 controlled temperature of 22°C to 24°C.(23) Prior to the visit, participants were asked to abstain
121 from caffeine for 8 hours,(24-26) vigorous exercise for 24 hours,(27) and alcohol for 24
122 hours.(28) Participants were seated on a recumbent stepper (T5XR NuStep, Inc. Ann Arbor, MI)

123 with the chair rotated 90 degrees perpendicular. The participant sat with feet flat on the ground,
124 legs at 90 degrees, and an upright trunk posture. The participants rested for 20 minutes while the
125 equipment was donned: 1) a 5-lead electrocardiogram (ECG; Cardiocard, Nasiff Associates,
126 Central Square, New York) to measure heart rate, 2) a left middle finger plethysmograph
127 (Finometer, Finapres Medical Systems, Amsterdam, the Netherlands) to measure beat-to-beat
128 MAP, and 3) bilateral TCD probes (2-MHz, Multigon Industries Inc, Yonkers, New York) to
129 measure MCAv. The TCD probes were secured with headgear to maintain the optimal position
130 and angle. Consistent with previously published methodology measuring beat-to-beat MAP
131 during a sit-to-stand, participants were instructed to place their left hand with the finger
132 plethysmograph flat on the upper chest at heart level.(6, 8, 13) Due to neuromuscular fatigue, an
133 arm sling was used to help hold the Finometer in place during the sit-to-stand recording.(13) For
134 individuals with upper extremity spasticity, the Finometer was placed on the non-affected upper
135 extremity.

136


137 Force-Sensor

138 Our custom force sensor was comprised of a force-sensitive resistor (SEN-09376, Mouser
139 Electronics Inc., Tx) connected to an operational amplifier (Canduino LM393, Universal-
140 Solder, Ontario, Canada) and custom circuit housed on a breadboard (Inland 400 tie-point),
141 shown in **Figure 1**. The force-sensitive resistor was placed on the seat under the visualized right
142 ischial tuberosity and connected to the custom circuit by 6 feet of stranded wire. For individuals
143 with stroke, the force-sensitive resistor was placed underneath the non-affected lower extremity
144 due to asymmetrical weight bearing during a sit-to-stand.(29) When sitting, the force-sensitive
145 resistor converted the weight of the participant into a voltage, amplified at the breadboard, and

146 digitized via an analog-to-digital unit (NI-USB-6212, National Instruments). Once participants
147 stood from the chair the force-sensitive resistor no longer detected the participant's weight and
148 the voltage signal quickly decreased to 0 volts for the remainder of the recording. The voltage of
149 our force sensor was captured simultaneously with MCAv and MAP during the sit-to-stand using
150 custom written software. All measures were recorded at 500 Hz using a custom written software
151 within MATLAB implementing the Data Acquisition Toolbox (R2019a, TheMathworks Inc,
152 Natick, Massachusetts).

153

154 **Figure 1.**



155

156

157 Figure 1. Drawing of the Custom Force Sensor Measuring AO during Sit-to-Stands. For the
158 operational amplifier (Op Amp), the pins correspond to: 8-Vcc, 3-Input, 1-Output, 4-Ground.
159 The force-sensitive resistor has a resistance that decreases with force. As the subject stands, force
160 decreases, the resistance increases, and the voltage drop increases at the force-sensitive resistor.
161 This results in a decreased Op Amp input voltage due to decreased voltage drop across the 47
162 Ohm resistor.

163

164 Sit-to-Stand

165 Participants were familiarized with the sit-to-stand protocol prior to the recording. After a minute
166 of seated rest, participants were given a three-second countdown and asked to stand at 60
167 seconds into the recording. Participants remained standing for two minutes to allow
168 hemodynamic stabilization.(6, 10) The sit-to-stand recording was performed twice separated by
169 15 minutes (T1, T2).

170

171 Raw data were processed offline using the QRS complex within the ECG to calculate beat-to-
172 beat MCAv and MAP.(12) AO was defined as the minimum of the second derivative of the
173 sensor voltage, corresponding with the acceleration of the decreasing voltage velocity during the
174 sit-to-stand.

175

176 Data Analysis

177 Based on previously published methodology, the TD before the onset of the dCA regulatory
178 response was measured as seconds between performing the sit-to-stand and a continuous increase
179 in CVC without transient reduction ($CVC = MCAv/MAP$). (6, 10) Rather than estimating AO in
180 the recording as 60 seconds, our force sensor marks the exact AO time. We calculated the
181 difference in the TD response when using the force sensor compared to estimating AO at 60
182 seconds, or when the participant was asked to stand.

183

184 **Results**

185 A total of 8 individuals completed the sit-to-stand transition with the force sensor to measure the
 186 TD before the onset of dCA. We included healthy young adults (n=4), older adults (n=2) and
 187 individuals post-stroke (n=2). Participant characteristics are shown in **Table 1**. The sit-to-stand
 188 AO time for all participants is shown in **Table 2**.

189

190 **Table 1. Participant Characteristics**

Subject	Age Range (years)	Sex	5 Time Sit to Stand (seconds)
Young Adults			
1	21 - 25	F	5.05
2	21 - 25	F	7.42
3	21 - 25	M	6.42
4	26 - 30	M	5.39
Older Adults			
5	66 - 70	M	6.54
6	71 - 75	M	10.85
Individuals Post-Stroke			
7	41 - 45	M	10.07
8	71 - 75	M	26.35

191

192 **Table 2. Sit-to-Stand Outcomes**

Subject	T1 AO	T1 TD	TD Difference with Force Sensor	T2 AO	T2 TD	TD Difference with Force Sensor
Young Adults						
1	60.03	1.75	-0.03	59.81	1.26	+0.19
2	60.58	1.11	-0.58	60.26	2.13	-0.26
3	60.36	1.46	-0.36	60.21	6.70	-0.21
4	60.14	2.23	-0.14	59.73	1.54	+0.27
Older Adults						
5	61.43	6.79	-1.43	59.93	2.89	+0.07
6	60.32	2.33	-0.32	62.52	2.23	-2.52
Individuals Post-Stroke						
7	59.91	0.79	+0.09	60.00	1.94	0
8	59.83	1.93	+0.17	59.60	3.16	+0.40

Time in seconds. T1 = First sit-to-stand. T2 = Second sit-to-stand (separated by 15 minutes).
AO = arise-and-off. TD = time delay before the onset of the dCA response.

193

194

195 Healthy Young Adults

196 At T1, the AO time for the healthy young adults was 60.28 ± 0.24 seconds. The force sensor was
197 able to account for a shorter TD by 0.28 ± 0.24 seconds compared to estimating AO from the
198 verbal command to stand. At T2, the AO time was on average at 60.00 ± 0.27 seconds in the
199 recording. However, because 2 individuals had an AO time before being asked to stand, the force
200 sensor accounted for a change in the TD response of 0.003 ± 0.27 seconds.

201

202 Older Adults

203 At T1, participant #5's AO time was 61.43 seconds. Compared to estimating AO from the verbal
204 command to stand, the force sensor was able to account for a decrease in the TD response by
205 1.43 seconds (~17%). At T2, the same participant #5 had a 1.50 second faster AO time compared
206 to T1 and had an AO time before being asked to stand. In contrast, while participant #6 had a
207 small change in the TD at T1 of 0.32 seconds, they were 2.20 seconds slower at T2. The force
208 sensor was therefore able to account for a decrease in the TD response by 2.52 seconds (~53%)
209 at T2, compared to estimating AO from the verbal command to stand (shown in **Figure 2**).

210


211 Individuals Post-Stroke

212 At T1, participant #7's AO time was 0.09 seconds before the verbal command to stand, which
213 accounted for an increase (~13%) in the TD response, compared to estimating AO from the
214 verbal command to stand. At T2, the same participant #7's AO time was at exactly 60 seconds.

215 In contrast, participant #8's AO was before the verbal command to stand, which accounted for an
216 increase in the TD response at T1 (~10%) and T2 (~14%), respectively.

217

218 **Figure 2.**



219

220 Figure 2. Novel Application of a Force Sensor during a Sit-to-Stand in a Single Healthy Young

221 Adult (n = 1), Older Adult (n = 1), and Individual Post-Stroke (n = 1). CVC = Cerebrovascular

222 Conductance. Horizontal Dotted Line = Individuals were asked to perform the sit-to-stand at 60
223 seconds. AO = Arise-and-off when participants stood from the chair. TD = Time delay before the
224 onset of dynamic cerebral autoregulation. Force sensor values are in arbitrary units.

225

226 **Discussion**

227 Previously published sit-to-stand methodology measuring the TD before the onset of the dCA
228 response does not report the true moment of AO from the chair, and therefore does not measure
229 the exact start of the stimulus that drops peripheral blood pressure. While force sensors have
230 been previously used in the kinetic characterization of a sit-to-stand,(30) we are the first to apply
231 a force sensor during a hemodynamic recording of dCA onset. The force sensor allowed for the
232 simultaneous measurement of the sit-to-stand reaction behavior with the physiologic response.

233 We hypothesized that measuring AO precisely with a force sensor would identify previously
234 unobserved variability in standing dynamics compared to estimating AO from a verbal command
235 or other methods.(6, 8, 10, 13) We have shown that while healthy young adults may demonstrate
236 a faster transition reaction speed during a sit-to-stand dCA recording,(4, 8) older adults and
237 individuals post-stroke may differ in their AO response. We have shown that the TD before the
238 onset of dCA, which typically occurs within 10 seconds,(6, 8, 10, 14) could potentially be
239 skewed in some individuals by more than 2.5 seconds, which would directly affect dCA metrics.

240

241 Previous reviews have also reported delayed sit-to-stand reaction time in individuals with
242 stroke,(18, 31, 32) which was supported by our results (see Table 1) reporting the 5 time sit-to-
243 stand scoring 2 standard deviations away from the age-normative values.(19) While this would
244 suggest an individual post-stroke would have a delay in AO, our results show that the individuals

245 with stroke began to stand prematurely during the 3 second countdown, shown in **Figure 2**. A
246 possible explanation for an early AO could be that the individuals post-stroke anticipated longer
247 sit-to-stand transition times.(17)

248
249 We show that the healthy young adult participants performed AO at approximately the same time
250 between T1 and T2. In fact, 2 healthy young adults stood early at T2, which could have been due
251 to anticipation and learning affect. One older adult participant could have also had an increase in
252 performance learning and stood faster at T2. However, the other older adult took ~2 seconds
253 longer to perform AO at T2, which may have been due to lower extremity weakness. The
254 individuals post-stroke may have recognized their lower extremity weakness and anticipated
255 needing longer to stand.

256
257 Imprecise temporal measures of dCA could potentially impact our interpretation of data
258 especially when comparing healthy individuals to clinical populations. A previous study
259 performed repeated sit-to-stands every 10 seconds in healthy older adults compared to
260 individuals with dementia and mild cognitive impairment.(33) Contrary to their hypothesis, they
261 reported individuals with mild cognitive impairment and dementia had a better autoregulation
262 response.(33) However, the individuals with cognitive impairment could have performed the sit-
263 to-stands early and, therefore, could have skewed the response. Without measuring the exact start
264 of the sit-to-stand stimulus, the wrong conclusions could unintentionally be drawn on the
265 temporal dCA response. Another example is a study comparing the sit-to-stand dCA response
266 after a 14-week exercise intervention in pregnant women that reported no significant change in
267 the TD before the onset of dCA.(34) However, there could have been a large difference in the

268 sit-to-stand response between the baseline measures taken during the 2nd trimester of pregnancy
269 and measures taken after the exercise intervention during the 3rd trimester of pregnancy. Pregnant
270 women within the 3rd trimester could have had a slower sit-to-stand time that was not accounted
271 for and could have inadvertently skewed the calculation of TD. Imprecise timing of the dCA
272 response, supported by our results showing ~10% to 53% change in TD, could systematically
273 impact study findings.

274

275 **Conclusions**

276 The implementation of a force sensor to measure AO during a hemodynamic sit-to-stand
277 recording could improve the exact calculation of the TD before the onset of the dCA response.
278 Applying the force sensor into the sit-to-stand dCA methodology has the potential to improve
279 scientific rigor and reproducibility, especially in older adults and clinical populations such as
280 stroke.

281

282 **Acknowledgements**

283 We would like to thank Preston Judd for his contributions to editing and proof reading.

284

285 **Funding**

286 AW was supported by the Eunice Kennedy Shriver National Institute of Child Health & Human
287 Development of the National Institutes of Health (T32HD057850). SA was supported by a
288 CTSA grant from NCATS awarded to the University of Kansas for Frontiers: University of
289 Kansas Clinical and Translational Science Institute (TL1TR002368). SB and EV were supported
290 in part by the National Institute on Aging for the KU Alzheimer's Disease Research Center

291 (P30AG072973). The content is solely the responsibility of the authors and does not necessarily
292 represent the official views of the National Institutes of Health. REDCap at University of Kansas
293 Medical Center is supported by Clinical and Translational Science Awards (CTSA) Award #
294 ULTR000001 from National Center for Research Resources (NCRR). The Georgia Holland
295 Research in Exercise and Cardiovascular Health (REACH) laboratory space was supported by
296 the Georgia Holland Endowment Fund.

297

298 **Conflict of Interest Disclosure**

299 The author(s) report no conflict of interest.

300 References

- 301 1. **Willie CK, Tzeng YC, Fisher JA, and Ainslie PN.** Integrative regulation of human brain blood flow.
302 *J Physiol* 592: 841-859, 2014.
- 303 2. **Roy CS, and Sherrington CS.** On the Regulation of the Blood-supply of the Brain. *J Physiol* 11: 85-
304 158 117, 1890.
- 305 3. **Brassard P, Labrecque L, Smirl JD, Tymko MM, Caldwell HG, Hoiland RL, Lucas SJE, Denault AY,**
306 **Couture EJ, and Ainslie PN.** Losing the dogmatic view of cerebral autoregulation. *Physiol Rep* 9: e14982,
307 2021.
- 308 4. **van Beek AH, Claassen JA, Rikkert MG, and Jansen RW.** Cerebral autoregulation: an overview of
309 current concepts and methodology with special focus on the elderly. *J Cereb Blood Flow Metab* 28:
310 1071-1085, 2008.
- 311 5. **Panerai RB.** Assessment of cerebral pressure autoregulation in humans--a review of
312 measurement methods. *Physiol Meas* 19: 305-338, 1998.
- 313 6. **Labrecque L, Rahimaly K, Imhoff S, Paquette M, Le Blanc O, Malenfant S, Lucas SJE, Bailey DM,**
314 **Smirl JD, and Brassard P.** Diminished dynamic cerebral autoregulatory capacity with forced oscillations
315 in mean arterial pressure with elevated cardiorespiratory fitness. *Physiol Rep* 5: 2017.
- 316 7. **Lind-Holst M, Cotter JD, Helge JW, Boushel R, Augustesen H, Van Lieshout JJ, and Pott FC.**
317 Cerebral autoregulation dynamics in endurance-trained individuals. *J Appl Physiol (1985)* 110: 1327-
318 1333, 2011.
- 319 8. **Sorond FA, Serrador JM, Jones RN, Shaffer ML, and Lipsitz LA.** The sit-to-stand technique for
320 the measurement of dynamic cerebral autoregulation. *Ultrasound Med Biol* 35: 21-29, 2009.
- 321 9. **Olufsen MS, Ottesen JT, Tran HT, Ellwein LM, Lipsitz LA, and Novak V.** Blood pressure and
322 blood flow variation during postural change from sitting to standing: model development and validation.
323 *J Appl Physiol (1985)* 99: 1523-1537, 2005.
- 324 10. **Labrecque L, Rahimaly K, Imhoff S, Paquette M, Le Blanc O, Malenfant S, Drapeau A, Smirl JD,**
325 **Bailey DM, and Brassard P.** Dynamic cerebral autoregulation is attenuated in young fit women. *Physiol*
326 *Rep* 7: e13984, 2019.
- 327 11. **Tzeng YC, Willie CK, Atkinson G, Lucas SJ, Wong A, and Ainslie PN.** Cerebrovascular regulation
328 during transient hypotension and hypertension in humans. *Hypertension* 56: 268-273, 2010.
- 329 12. **Moir ME, Klassen SA, Zamir M, and Shoemaker JK.** Rapid changes in vascular compliance
330 contribute to cerebrovascular adjustments during transient reductions in blood pressure in young,
331 healthy adults. *J Appl Physiol (1985)* 129: 27-35, 2020.
- 332 13. **Lipsitz LA, Mukai S, Hamner J, Gagnon M, and Babikian V.** Dynamic regulation of middle
333 cerebral artery blood flow velocity in aging and hypertension. *Stroke* 31: 1897-1903, 2000.
- 334 14. **Serrador JM, Sorond FA, Vyas M, Gagnon M, Iloputaife ID, and Lipsitz LA.** Cerebral pressure-
335 flow relations in hypertensive elderly humans: transfer gain in different frequency domains. *J Appl*
336 *Physiol (1985)* 98: 151-159, 2005.
- 337 15. **Deegan BM, Sorond FA, Lipsitz LA, O'Laughlin G, and Serrador JM.** Gender related differences in
338 cerebral autoregulation in older healthy subjects. *Annu Int Conf IEEE Eng Med Biol Soc* 2009: 2859-2862,
339 2009.
- 340 16. **Stevermer CA, and Gillette JC.** Kinematic and Kinetic Indicators of Sit-to-Stand. *J Appl Biomech*
341 32: 7-15, 2016.
- 342 17. **Mong Y, Teo TW, and Ng SS.** 5-repetition sit-to-stand test in subjects with chronic stroke:
343 reliability and validity. *Arch Phys Med Rehabil* 91: 407-413, 2010.
- 344 18. **Mentiplay BF, Clark RA, Bower KJ, Williams G, and Pua YH.** Five times sit-to-stand following
345 stroke: Relationship with strength and balance. *Gait Posture* 78: 35-39, 2020.

- 346 19. **Bohannon RW.** Reference values for the five-repetition sit-to-stand test: a descriptive meta-
347 analysis of data from elders. *Percept Mot Skills* 103: 215-222, 2006.
- 348 20. **Thompson PD, Arena R, Riebe D, Pescatello LS, and American College of Sports M.** ACSM's new
349 preparticipation health screening recommendations from ACSM's guidelines for exercise testing and
350 prescription, ninth edition. *Curr Sports Med Rep* 12: 215-217, 2013.
- 351 21. **Tiedemann A, Shimada H, Sherrington C, Murray S, and Lord S.** The comparative ability of eight
352 functional mobility tests for predicting falls in community-dwelling older people. *Age Ageing* 37: 430-
353 435, 2008.
- 354 22. **Ng S.** Balance ability, not muscle strength and exercise endurance, determines the performance
355 of hemiparetic subjects on the timed-sit-to-stand test. *Am J Phys Med Rehabil* 89: 497-504, 2010.
- 356 23. **Billinger SA, Craig JC, Kwapiszeski SJ, Sisante JV, Vidoni ED, Maletsky R, and Poole DC.**
357 Dynamics of middle cerebral artery blood flow velocity during moderate-intensity exercise. *J Appl*
358 *Physiol* (1985) 122: 1125-1133, 2017.
- 359 24. **Perod AL, Roberts AE, and McKinney WM.** Caffeine can affect velocity in the middle cerebral
360 artery during hyperventilation, hypoventilation, and thinking: a transcranial Doppler study. *J*
361 *Neuroimaging* 10: 33-38, 2000.
- 362 25. In: *Caffeine for the Sustainment of Mental Task Performance: Formulations for Military*
363 *Operations*. Washington (DC): 2001.
- 364 26. **Addicott MA, Yang LL, Peiffer AM, Burnett LR, Burdette JH, Chen MY, Hayasaka S, Kraft RA,**
365 **Maldjian JA, and Laurienti PJ.** The effect of daily caffeine use on cerebral blood flow: How much
366 caffeine can we tolerate? *Hum Brain Mapp* 30: 3102-3114, 2009.
- 367 27. **Burma JS, Copeland P, Macaulay A, Khatra O, Wright AD, and Smirl JD.** Dynamic cerebral
368 autoregulation across the cardiac cycle during 8 hr of recovery from acute exercise. *Physiol Rep* 8:
369 e14367, 2020.
- 370 28. **Mathew RJ, and Wilson WH.** Regional cerebral blood flow changes associated with ethanol
371 intoxication. *Stroke* 17: 1156-1159, 1986.
- 372 29. **Brunt D, Greenberg B, Wankadia S, Trimble MA, and Shechtman O.** The effect of foot
373 placement on sit to stand in healthy young subjects and patients with hemiplegia. *Arch Phys Med*
374 *Rehabil* 83: 924-929, 2002.
- 375 30. **Galli M, Cimolin V, Crivellini M, and Campanini I.** Quantitative analysis of sit to stand
376 movement: Experimental set-up definition and application to healthy and hemiplegic adults. *Gait*
377 *Posture* 28: 80-85, 2008.
- 378 31. **Boukadida A, Piotte F, Dehail P, and Nadeau S.** Determinants of sit-to-stand tasks in individuals
379 with hemiparesis post stroke: A review. *Ann Phys Rehabil Med* 58: 167-172, 2015.
- 380 32. **Lecours J, Nadeau S, Gravel D, and Teixeira-Salmela L.** Interactions between foot placement,
381 trunk frontal position, weight-bearing and knee moment asymmetry at seat-off during rising from a
382 chair in healthy controls and persons with hemiparesis. *J Rehabil Med* 40: 200-207, 2008.
- 383 33. **de Heus RAA, de Jong DLK, Sanders ML, van Spijker GJ, Oudegeest-Sander MH, Hopman MT,**
384 **Lawlor BA, Olde Rikkert MGM, and Claassen J.** Dynamic Regulation of Cerebral Blood Flow in Patients
385 With Alzheimer Disease. *Hypertension* 72: 139-150, 2018.
- 386 34. **Skow RJ, Labrecque L, Rosenberger JA, Brassard P, Steinback CD, and Davenport MH.** Prenatal
387 exercise and cardiovascular health (PEACH) study: impact of acute and chronic exercise on
388 cerebrovascular hemodynamics and dynamic cerebral autoregulation. *J Appl Physiol* (1985) 2021.