medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

Evaluation of a Rapid and Accessible RT-qPCR Approach for SARS-CoV-2 Variant of Concern Identification

Priscilla S.-W. Yeung, MD, PhD[a]*; Hannah Wang, MD[a]*; Mamdouh Sibai, BS[a]; Daniel Solis, BS[a]; Fumiko Yamamoto, MS[a]; Naomi Iwai, CLS[b]; Becky Jiang, CLS[b]; Nathan Hammond, PhD[c]; Bernadette Truong, CLS[b]; Selamawit Bihon, CLS[b]; Suzette Santos, CLS[b]; Marilyn Mar, CLS[b]; Claire Mai, BS[c]; Kenji O. Mfuh, PhD[b]; Jacob A. Miller, MD[d]; ChunHong Huang, MD[a]; Malaya K. Sahoo, PhD[a]; James L. Zehnder, MD[a]; Benjamin A. Pinsky, MD, PhD[a,b,e]#

[a] Department of Pathology, Stanford University School of Medicine, Stanford, CA, USA

[b] Clinical Virology Laboratory, Stanford Health Care, Stanford, CA, USA

[c] Clinical Genomics Laboratory, Stanford Health Care, Stanford, CA, USA

[d] Department of Radiation Oncology, Stanford University School of Medicine, Stanford, CA, USA

[e] Division of Infectious Diseases and Geographic Medicine, Department of Medicine, Stanford University School of Medicine, Stanford, CA, USA

Running Head: SARS-CoV-2 Variant of Concern Genotyping

#Address correspondence to Benjamin A. Pinsky, MD, PhD, bpinsky@stanford.edu

*Authors Priscilla S.-W. Yeung and Hannah Wang contributed equally to this work. Author order was determined reverse-alphabetically.

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

ABSTRACT

The ability to distinguish between SARS-CoV-2 variants of concern (VOCs) is of ongoing interest due to differences in transmissibility, response to vaccination, clinical prognosis, and therapy. Although detailed genetic characterization requires whole-genome sequencing (WGS), targeted nucleic acid amplification tests can serve a complementary role in clinical settings, as they are more rapid and accessible than sequencing in most laboratories.

We designed and analytically validated a two-reaction multiplex reverse transcription quantitative PCR (RT-qPCR) assay targeting spike protein mutations L452R, E484K, and N501Y in Reaction 1, and del69-70, K417N, and T478K in Reaction 2. This assay had 95-100% agreement with WGS in 502 upper respiratory swabs collected between April 26 and August 1, 2021, consisting of 43 Alpha, 2 Beta, 20 Gamma, 378 Delta, and 59 non-VOC infections. Validation in a separate group of 230 WGS-confirmed Omicron variant samples collected in December 2021 and January 2022 demonstrated 100% agreement.

This RT-qPCR-based approach can be implemented in clinical laboratories already performing SARS-CoV-2 nucleic acid amplification tests to assist in local epidemiological surveillance and clinical decision-making.

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

INTRODUCTION

Since the original strain of SARS-CoV-2 virus was first discovered in late 2019, numerous new variants have been identified, including variants of concern (VOCs) Alpha (B.1.1.7 and Q.*), Beta (B.1.351), Gamma (P.1 and sublineages), Delta (B.1.617.2 and AY.*) and Omicron (B.1.1.529 and BA.*). Importantly, these VOCs differ in their clinical prognosis, transmissibility, antibody susceptibility, and response to vaccination (1-21). Whole-genome sequencing (WGS) has played a critical role in identifying the emergence of these new variants (22-24), and millions of distinct sequences have been deposited into public repositories such as the Global Initiative on Sharing Avian Influenza Data consortium's database (GISAID). However, WGS has a relatively long turnaround time, is labor-intensive, and requires instruments, bioinformatics support, and specially-trained staff that may not be widely available to many clinical laboratories. Therefore, the development of reverse transcription quantitative PCR (RT-qPCR) assays to detect and differentiate SARS-CoV-2 variants may be an important real-time complement to WGS epidemiologic surveillance, and may directly impact the clinical care of individual patients by informing selection of expensive and potentially difficult-to-source monoclonal antibody therapies (1, 6, 12-16, 19, 20, 25).

In this study, we report the design of a multiplex RT-qPCR assay that detects the del69-70, K417N, and T478K mutations in SARS-CoV-2 spike protein and targets the wild-type 69-70 sequence as an internal control. We further evaluate the performance of this assay in combination with our previously described RT-qPCR assay for the detection of L452R, E484K, and N501Y (26), and demonstrate the utility of this targeted mutational analysis to accurately distinguish among VOCs.

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

MATERIALS AND METHODS

Assay Design

The spike protein mutations associated with each variant that are interrogated by the RT-qPCR assays are summarized in Figure 1. In the first reaction (Reaction 1), we utilized our previously described RT-qPCR assay to detect L452R, E484K, and N501Y mutations in spike Receptor Binding Domain (RBD) (26). The present study describes the combination of this assay with a second, newly designed reaction (Reaction 2), which detects the deletion of amino acids 69-70 in the spike N-Terminal Domain (del69-70), as well as K417N and T478K mutations in the RBD. We use allele-specific RT-qPCR with probe sequences designed to maximize the difference in annealing temperature between mutant and wild-type sequences, allowing for differential binding and amplification. The primer/probe sequences for each mutation site are summarized in Table 1. Additional details are provided in the Supplemental Methods, ssDNA sequences for analytical experiments (Supplemental Table 1), guidance for interpretation and reporting (Supplemental Table 2), analytical validation data (Supplemental Table 3), and in-silico analysis of primer and probe sequences (Supplemental Figure 1).

Clinical Specimens

The samples included in the initial phase of this study were upper respiratory swab specimens collected from patients as part of routine clinical care from April 26, 2021 to August 1, 2021. Testing was performed at Stanford Clinical Virology Laboratory, which provides virologic testing for all Stanford-affiliated hospitals and outpatient centers in the San Francisco Bay Area. These initial SARS-CoV-2 nucleic acid amplification tests (NAATs) tests prior to genotyping were conducted according to manufacturer and emergency authorization instructions as previously described (26), and in the Supplemental Methods. All samples that tested positive for SARS-CoV-2 RNA were reflexed to genotyping. We then excluded samples that were initially tested by laboratory-based methods with cycle threshold values (Ct) ≥35 or relative light units (RLU) ≤1100. We included all available samples initially tested at or near the point of care as Ct data was not readily available for real-time specimen triage for these samples. We also excluded follow-up specimens to eliminate patient-level duplicates. Subsequent validation of this assay

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

for Omicron variant detection was conducted using a convenience set of 230 Omicron variant samples with available WGS data collected between December 2, 2021 and January 5, 2022. This study was conducted with Stanford institutional review board approval (protocol 57519), and individual consent was waived.

Whole-Genome Sequencing

To validate the genotyping RT-qPCR reactions, we tested their performance against WGS in a subset of the samples in the initial April 26, 2021 to August 1, 2021 cohort with Ct <30. Samples with non-dominant variant typing by RT-qPCR were prioritized for sequencing, with the remaining isolates chosen randomly to fill a sequencing run. WGS was conducted as described previously, using a lab-developed pipeline consisting of long-range PCR, followed by fragmentation, then single-end 150-cycle sequencing using MiSeq reagent kit V3 (Illumina, San Diego, CA) (26). Genomes were assembled via a custom assembly and bioinformatics pipeline using NCBI NC_045512.2 as reference. Whole-genome sequences with ≥75% genome coverage to a depth of at least 10 reads were accepted for interpretation. Median number of aligned reads was 485,870 (interquartile range [IQR] 289,363-655,481), while median genome coverage to a depth of at least 10 reads was 99.3% (IQR 97.1-99.3%). Mutation calling required a depth of at least 12 reads with a minimum variant frequency of 20%. PANGO lineage assignment was performed using https://pangolin.cog-uk.io/ running pangolin version 3.1.17, while Nextclade Web v1.13.1 and auspice.us 0.8.0. were used to perform phylogenetic placement (2, 27, 28). Both lineage and clade assignments were performed on February 1, 2022. WGS data was deposited in GISAID (Supplemental Table 4).

Statistical Analysis

Positive percent agreement (PPA) and negative percent agreement (NPA) were reported with Clopper-Pearson score 95% binomial confidence intervals (CI) using WGS as the reference method. Analyses were conducted using the R statistical software package. This study was reported in accordance with Standards for the Reporting of Diagnostic Accuracy Studies (STARD) guidelines.

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

RESULTS

During the initial study period of April 26, 2021 to August 1, 2021, the Stanford Clinical Virology Laboratory received 102,158 specimens from 70,544 unique individuals. A total of 1,657 samples from unique individuals tested positive for SARS-CoV-2, of which 1,093 (66%) had genotyping RT-qPCR Reaction 1 and Reaction 2 performed, and 502 (30.3%) had successful WGS performed (Supplemental Figure 2). Of note, Reaction 1 was performed in near real-time, while Reaction 2 was performed retrospectively. Overall, this subset of sequenced samples, had patient and testing characteristics that closely resembled those of the larger cohorts (Supplemental Table 5).

The assay resulted as “Unable to Genotype” in 152 of 1,093 samples (14%) due to lack of amplification of any target in either or both reactions. Assay failure occurred predominantly in samples originally tested at or near the point of care (119/341, 35%), where all positive samples were triaged for genotyping without any filter. In contrast, assay failure occurred much less frequently in samples originally tested in the moderate-to-high complexity virology lab (33/752, 4%), where samples with lower viral loads were not triaged for genotyping.

For the combination of Reactions 1 and 2, the PPAs for del69-70, L452R, T478K, E484K, and N501Y were 100% (Table 2). Across all six loci, only K417N had a false negative, resulting in a PPA of 96% (27/28). In this sample, WGS showed a synonymous T to C mutation at position 1254 of the spike gene corresponding to amino acid position 418, changing the codon from ATT to ATC. This single base pair substitution likely decreased the annealing temperature, causing probe dropout and a false negative result.

The NPAs for del69-70, K417N, T478K, and N501Y were 100% (Table 2). L452R had an NPA of 95% (94/99) and E484K had an NPA of 99% (464/467). At the L452 locus, there were five samples positive for L452R mutation by RT-qPCR that were negative by WGS. Manual review of the WGS data showed that these were likely false negative WGS results due to insufficient (<12 reads) coverage at this codon. There

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

were 3-9 reads containing the L452R mutation identified in the WGS primary data in each of these five samples. These five samples were all in the Delta lineage based on mutations found at other positions by sequencing.

For the E484K target, there were three samples that tested positive for the E to K mutation but in fact had a E484Q mutation determined by WGS. In both the E to K mutation (GAA to AAA) as well as the E to Q mutation (GAA to CAA), there was a single base substitution at the first position of the codon resulting in nonspecific probe binding. These three samples had a distinct blunted amplification curve with high Ct values associated with E484Q, as previously described (29).

Of note, there was a subset of variant AY.2, involving four specimens in our cohort, that had a V70F mutation causing both del69-70 and wt69-70 probes not to bind. However, because this variant would have T478K and K417N detected, the wt69-70 signal was not needed as an amplification control. This scenario has been reflected in the clinical interpretation table (Supplemental Table 2).

SARS-CoV-2 positive specimens collected starting December 2, 2021 began to show an unusual combination of mutations: presence of K417N and del69-70 only in Reaction 2, with all targets including internal control N501 not detected in Reaction 1. Based on in-silico analysis, we determined that these cases likely represented Omicron variant. While most Omicron variant strains possess del69-70, K417N, T478K, and N501Y mutations, they also have mutations at A67V, S477N, and Q498R, which would be predicted to interfere with binding of the del69-70/wt69-70, T478K, and N501Y/N501 probes, respectively. The del69-70 probe likely was able to retain some degree of binding due to the wider melting temperature differential of a 6-nucleotide deletion compared to a point mutation. As such, we validated this assay for Omicron detection using a set of 230 SARS-CoV-2 positive samples confirmed to be Omicron by WGS. We found that the unique pattern of K417N and del69-70 in Reaction 2, along with failure to amplify any target including internal control in Reaction 1, was present in 230/230 (100%, 95% CI 98-100%) Omicron samples tested. This pattern was not seen in any of the 1,093 non-Omicron samples previously genotyped.

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

We next predicted the WHO variant designation of samples using RT-qPCR and correlated them with the PANGO lineage assignments based on WGS data (Table 3). Mapping the genotyping results of the cohort based on RT-qPCR mutation analysis onto the Nextclade phylogenetic tree demonstrated close correlation with their WHO variant designations (Figure 2). Among the 732 clinical samples that were tested by both RT-qPCR and WGS, 43 samples (5.9%) were Alpha (B.1.1.7 or Q.3), 2 samples (0.3%) were Beta (B.1.351), 20 samples (2.7%) were Gamma (P.1 and sublineages), 378 samples (51.6%) were Delta (B.1.617.2 or AY.*), and 230 samples (31.4%) were Omicron (B.1.1.529 or BA.*). There were no RT-qPCR false negatives in assigning samples to these lineages. In addition, there were 59 samples (8.1%) tested by WGS that did not correspond to a WHO VOC as of February 2, 2022. Within this subset, there were 4 samples that were erroneously assigned as Gamma and 1 that was assigned as Beta by RT-qPCR. By WGS, these samples were variant of interest (VOI) Mu (B.1.621 or BB.2). This variant shares mutations E484K and N501Y with both the Beta and Gamma variants. A subset of Mu also includes the K417N mutation which is seen in the Beta variant. Thus, our PCR assay could not distinguish VOI Mu from VOCs Beta and Gamma. Our interpretation table included in the supplementary information reflects this limitation (Supplemental Table 2). The remaining 54 samples did not contain mutation patterns associated with VOCs.

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

DISCUSSION

The ability to distinguish between SARS-CoV-2 VOCs is important for epidemiologic surveillance, and in certain circumstances, the care of individual COVID-19 patients. In this study, we describe a two-reaction, multiplex RT-qPCR genotyping approach that examines the spike mutations del69-70, K417N, L452R, T478K, E484K, and N501Y. This targeted mutational analysis can be used to differentiate between the WHO VOCs Alpha (B.1.1.7 and Q.*), Beta (B.1.351), Gamma (P.1 and sublineages), Delta (B.1.617.2 and AY.*), and Omicron (B.1.1.519 and BA.*), as well as identify samples which cannot be categorized into a known VOC or VOI. Because the first part of this approach, Reaction 1, has been previously described, this current study focuses on Reaction 2 and the integrated results of the two-reaction test (26). Overall, these reactions showed high concordance with WGS, demonstrating over 95% PPA and NPA for all targeted mutations.

Several groups have previously described similar approaches to SARS-CoV-2 variant determination by RT-qPCR and digital droplet RT-PCR, particularly for the spike del69-70, E484K, and N501Y positions (30-37). Some of these assays included additional mutation sites that were not in our study, such as spike del144 or ORF1a Δ3675–3677 (30, 36). These earlier assays, published prior to the rise of Delta, primarily targeted VOCs Alpha, Beta, and Gamma. This was then followed by a surge of reports on the detection of the Delta variant. Garson et al. utilized double-mismatch allele-specific RT-PCR at L452R and T478K to differentiate Delta variant from other VOCs in 42 UK patient samples (38). Aoki et al. described an approach that combines nested PCR along with high-resolution melting analysis at those same mutations, which was validated in a small Japanese patient cohort (39). Barua et al. used a slightly different approach, taking advantage of the difference in melting temperature of a probe targeted to Delta mutation in spike T478K compared to other variants for a Delta-specific RT-FRET-PCR assay (40). Another defining feature of VOC Delta is spike del156-157, which was the target of a Delta variant PCR test developed by Hamill et al. (41). To our knowledge, the two-reaction multiplex RT-qPCR approach outlined in this study examining six different mutation sites is the most comprehensive variant genotyping test described that can identify Alpha, Beta, Gamma, Delta, and Omicron variants.

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

Multiplex RT-qPCR SARS-CoV-2 genotyping takes advantage of a commonly-used molecular technique that can be implemented by laboratories using existing equipment, materials, and personnel. Because this assay is more accessible and has a more rapid turnaround time than WGS, we envision it serving a complimentary role to sequencing. The genotyping RT-qPCR can provide more detailed and up-to-date epidemiological information by increasing the sample size of categorized variants in each geographic region, and can be essential in tracking local outbreaks in areas without direct access to WGS. For individual patients, the turnaround time of several hours also allows it to directly impact clinical care. For example, VOCs show differential susceptibility to monoclonal antibody treatments, and variant reporting could include this information (Supplemental Table 2) (1). Furthermore, current ongoing trials for small molecule drugs and other treatments may yield more information about variant-specific treatment strategies.

Importantly, RT-qPCR genotyping can help prioritize samples for sequencing. Although sequencing is needed for identification of novel variants and characterization of viral evolution, pre-screening by RT-qPCR can enrich for samples with atypical mutation patterns, lead to more efficient use of sequencing resources, and potentially more rapid identification of new variants.

This RT-qPCR approach has several limitations as evidenced by its assay failure rate of 14% across all tested samples in our initial cohort. Because multiplex RT-qPCRs involve a mixture of multiple sets of primers and probes, they are inherently less sensitive than single-target assays. For samples with RNA concentrations near the lower limits of detection, freeze-thaw cycles could impact RNA stability, and may not yield consistent results due to stochastic variation. This issue could be alleviated by implementing a Ct/RLU filter to only genotype samples most likely to yield interpretable results. Within our 1,093 sample cohort, the lower assay failure rate in samples tested in our clinical virology laboratory (4%) compared to near-care settings (35%) is likely attributable to genotyping only specimens with higher viral RNA levels. Note, however, that even with such filtering, mutation analysis by RT-qPCR remains more sensitive than WGS. The other limitation to this approach is the rapidly changing variant landscape which may render

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

such an assay obsolete in the matter of weeks. However, the inclusion of multiple targets in key residues that influence viral fitness helps guard against this possibility, as evidenced by our ability to detect the emergence of Omicron variant in our population. Still, flexibility and vigilance are required to re-design and re-validate these types of assays as novel variants emerge.

In summary, we developed and validated a two-reaction multiplex RT-qPCR genotyping strategy that interrogates six clinically relevant mutations within the SARS-CoV-2 spike: del69-70, K417N, L452R, T478K, E484K, and N501Y. This approach allows for identification of WHO VOCs Alpha, Beta, Gamma, Delta and Omicron with excellent concordance to WGS. Overall, this method complements WGS, and is suitable for clinical decision-making, near real-time variant surveillance, and the triage of samples for sequencing.

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

ACKNOWLEDGEMENTS

We would like to thank the Stanford Clinical Virology Laboratory Staff for their excellence and perseverance in caring for our community while facing the unique challenges of this pandemic. We additionally are grateful to the Stanford Protein and Nucleic Acid Facility for their synthesis of reagents on short notice.

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

REFERENCES

1. Tao K, Tzou PL, Nouhin J, Gupta RK, de Oliveira T, Kosakovsky Pond SL, Fera D, Shafer RW. 2021. The biological and clinical significance of emerging SARS-CoV-2 variants. Nat Rev Genet 22:757-773.
2. Hadfield J, Megill C, Bell SM, Huddleston J, Potter B, Callender C, Sagulenko P, Bedford T, Neher RA. 2018. Nextstrain: real-time tracking of pathogen evolution. Bioinformatics 34:4121-4123.
3. Day T, Gandon S, Lion S, Otto SP. 2020. On the evolutionary epidemiology of SARS-CoV-2. Curr Biol 30:R849-R857.
4. Davies NG, Abbott S, Barnard RC, Jarvis CI, Kucharski AJ, Munday JD, Pearson CAB, Russell TW, Tully DC, Washburne AD, Wenseleers T, Gimma A, Waites W, Wong KLM, van Zandvoort K, Silverman JD, Group CC-W, Consortium C-GU, Diaz-Ordaz K, Keogh R, Eggo RM, Funk S, Jit M, Atkins KE, Edmunds WJ. 2021. Estimated transmissibility and impact of SARS-CoV-2 lineage B.1.1.7 in England. Science 372.
5. Muik A, Wallisch AK, Sanger B, Swanson KA, Muhl J, Chen W, Cai H, Maurus D, Sarkar R, Tureci O, Dormitzer PR, Sahin U. 2021. Neutralization of SARS-CoV-2 lineage B.1.1.7 pseudovirus by BNT162b2 vaccine-elicited human sera. Science 371:1152-1153.
6. Wang P, Casner RG, Nair MS, Wang M, Yu J, Cerutti G, Liu L, Kwong PD, Huang Y, Shapiro L, Ho DD. 2021. Increased resistance of SARS-CoV-2 variant P.1 to antibody neutralization. Cell Host Microbe 29:747-751 e4.
7. Collier DA, De Marco A, Ferreira I, Meng B, Datir RP, Walls AC, Kemp SA, Bassi J, Pinto D, Silacci-Fregni C, Bianchi S, Tortorici MA, Bowen J, Culap K, Jaconi S, Cameroni E, Snell G, Pizzuto MS, Pellanda AF, Garzoni C, Riva A, Collaboration C-NBC-, Elmer A, Kingston N, Graves B, McCoy LE, Smith KGC, Bradley JR, Temperton N, Ceron-Gutierrez L, Barcenas-Morales G, Consortium C-GU, Harvey W, Virgin HW, Lanzavecchia A, Piccoli L, Doffinger R, Wills M, Veesler D, Corti D, Gupta RK. 2021. Sensitivity of SARS-CoV-2 B.1.1.7 to mRNA vaccine-elicited antibodies. Nature 593:136-141.

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

8. Shen X, Tang H, McDanal C, Wagh K, Fischer W, Theiler J, Yoon H, Li D, Haynes BF, Sanders KO, Gnanakaran S, Hengartner N, Pajon R, Smith G, Dubovsky F, Glenn GM, Korber B, Montefiori DC. 2021. SARS-CoV-2 variant B.1.1.7 is susceptible to neutralizing antibodies elicited by ancestral Spike vaccines. bioRxiv doi:10.1101/2021.01.27.428516.

9. Karim SSA. 2021. Vaccines and SARS-CoV-2 variants: the urgent need for a correlate of protection. Lancet 397:1263-1264.

10. Bolze A, Cirulli ET, Luo S, White S, Wyman D, Rossi AD, Machado H, Cassens T, Jacobs S, Schiabor Barrett KM, Tsan K, Nguyen J, Ramirez JM, Sandoval E, Wang X, Wong D, Becker D, Laurent M, Lu JT, Isaksson M, Washington NL, Lee W. 2021. SARS-CoV-2 variant Delta rapidly displaced variant Alpha in the United States and led to higher viral loads. medRxiv doi:10.1101/2021.06.20.21259195:2021.06.20.21259195.

11. Washington NL, Gangavarapu K, Zeller M, Bolze A, Cirulli ET, Schiabor Barrett KM, Larsen BB, Anderson C, White S, Cassens T, Jacobs S, Levan G, Nguyen J, Ramirez JM, Rivera-Garcia C, Sandoval E, Wang X, Wong D, Spencer E, Robles-Sikisaka R, Kurzban E, Hughes LD, Deng X, Wang C, Servellita V, Valentine H, De Hoff P, Seaver P, Sathe S, Gietzen K, Sickler B, Antico J, Hoon K, Liu J, Harding A, Bakhtar O, Basler T, Austin B, Isaksson M, Febbo P, Becker D, Laurent M, McDonald E, Yeo GW, Knight R, Laurent LC, de Feo E, Worobey M, Chiu C, Suchard MA, et al. 2021. Genomic epidemiology identifies emergence and rapid transmission of SARS-CoV-2 B.1.1.7 in the United States. medRxiv doi:10.1101/2021.02.06.21251159.

12. Chen X, Chen Z, Azman AS, Sun R, Lu W, Zheng N, Zhou J, Wu Q, Deng X, Zhao Z, Chen X, Ge S, Yang J, Leung DT, Yu H. 2021. Neutralizing antibodies against SARS-CoV-2 variants induced by natural infection or vaccination: a systematic review and pooled meta-analysis. Clin Infect Dis doi:10.1093/cid/ciab646.

13. Supasa P, Zhou D, Dejnirattisai W, Liu C, Mentzer AJ, Ginn HM, Zhao Y, Duyvesteyn HME, Nutalai R, Tuekprakhon A, Wang B, Paesen GC, Slon-Campos J, Lopez-Camacho C, Hallis B, Coombes N, Bewley KR, Charlton S, Walter TS, Barnes E, Dunachie SJ, Skelly D, Lumley SF, Baker N, Shaik I, Humphries HE, Godwin K, Gent N, Sienkiewicz A, Dold C, Levin R, Dong T, Pollard AJ, Knight JC, Klenerman P, Crook D, Lambe T, Clutterbuck E, Bibi S, Flaxman A, Bittaye

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

M, Belij-Rammerstorfer S, Gilbert S, Hall DR, Williams MA, Paterson NG, James W, Carroll MW, Fry EE, Mongkolsapaya J, et al. 2021. Reduced neutralization of SARS-CoV-2 B.1.1.7 variant by convalescent and vaccine sera. Cell 184:2201-2211 e7.

14. Hoffmann M, Arora P, Gross R, Seidel A, Hornich BF, Hahn AS, Kruger N, Graichen L, Hofmann-Winkler H, Kempf A, Winkler MS, Schulz S, Jack HM, Jahrsdorfer B, Schrezenmeier H, Muller M, Kleger A, Munch J, Pohlmann S. 2021. SARS-CoV-2 variants B.1.351 and P.1 escape from neutralizing antibodies. Cell 184:2384-2393 e12.

15. Liu C, Ginn HM, Dejnirattisai W, Supasa P, Wang B, Tuekprakhon A, Nutalai R, Zhou D, Mentzer AJ, Zhao Y, Duyvesteyn HME, Lopez-Camacho C, Slon-Campos J, Walter TS, Skelly D, Johnson SA, Ritter TG, Mason C, Costa Clemens SA, Gomes Naveca F, Nascimento V, Nascimento F, Fernandes da Costa C, Resende PC, Pauvolid-Correa A, Siqueira MM, Dold C, Temperton N, Dong T, Pollard AJ, Knight JC, Crook D, Lambe T, Clutterbuck E, Bibi S, Flaxman A, Bittaye M, Belij-Rammerstorfer S, Gilbert SC, Malik T, Carroll MW, Klenerman P, Barnes E, Dunachie SJ, Baillie V, Serafin N, Ditse Z, Da Silva K, Paterson NG, Williams MA, et al. 2021. Reduced neutralization of SARS-CoV-2 B.1.617 by vaccine and convalescent serum. Cell 184:4220-4236 e13.

16. Planas D, Veyer D, Baidaliuk A, Staropoli I, Guivel-Benhassine F, Rajah MM, Planchais C, Porrot F, Robillard N, Puech J, Prot M, Gallais F, Gantner P, Velay A, Le Guen J, Kassis-Chikhani N, Edriss D, Belec L, Seve A, Courtellemont L, Pere H, Hocqueloux L, Fafi-Kremer S, Prazuck T, Mouquet H, Bruel T, Simon-Loriere E, Rey FA, Schwartz O. 2021. Reduced sensitivity of SARS-CoV-2 variant Delta to antibody neutralization. Nature 596:276-280.

17. Liu J, Liu Y, Xia H, Zou J, Weaver SC, Swanson KA, Cai H, Cutler M, Cooper D, Muik A, Jansen KU, Sahin U, Xie X, Dormitzer PR, Shi PY. 2021. BNT162b2-elicited neutralization of B.1.617 and other SARS-CoV-2 variants. Nature 596:273-275.

18. Pulliam JRC, van Schalkwyk C, Govender N, von Gottberg A, Cohen C, Groome MJ, Dushoff J, Mlisana K, Moultrie H. 2021. Increased risk of SARS-CoV-2 reinfection associated with emergence of the Omicron variant in South Africa. medRxiv doi:10.1101/2021.11.11.21266068:2021.11.11.21266068.

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

19. Wilhelm A, Widera M, Grikscheit K, Toptan T, Schenk B, Pallas C, Metzler M, Kohmer N, Hoehl S, Helfritz FA, Wolf T, Goetsch U, Ciesek S. 2021. Reduced Neutralization of SARS-CoV-2 Omicron Variant by Vaccine Sera and monoclonal antibodies. medRxiv doi:10.1101/2021.12.07.21267432:2021.12.07.21267432.

20. Planas D, Saunders N, Maes P, Guivel-Benhassine F, Planchais C, Buchrieser J, Bolland WH, Porrot F, Staropoli I, Lemoine F, Pere H, Veyer D, Puech J, Rodary J, Baele G, Dellicour S, Raymenants J, Gorissen S, Geenen C, Vanmechelen B, Wawina-Bokalanga T, Marti-Carreras J, Cuypers L, Seve A, Hocqueloux L, Prazuck T, Rey F, Simon-Loriere E, Bruel T, Mouquet H, Andre E, Schwartz O. 2021. Considerable escape of SARS-CoV-2 Omicron to antibody neutralization. Nature doi:10.1038/s41586-021-04389-z.

21. Salvatore M, Bhattacharyya R, Purkayastha S, Zimmermann L, Ray D, Hazra A, Kleinsasser M, Mellan T, Whittaker C, Flaxman S, Bhatt S, Mishra S, Mukherjee B. 2021. Resurgence of SARS-CoV-2 in India: Potential role of the B.1.617.2 (Delta) variant and delayed interventions. medRxiv doi:10.1101/2021.06.23.21259405:2021.06.23.21259405.

22. Wang L, Cheng G. 2021. Sequence analysis of the emerging SARS-CoV-2 variant Omicron in South Africa. J Med Virol doi:10.1002/jmv.27516.

23. Zhang W, Davis BD, Chen SS, Sincuir Martinez JM, Plummer JT, Vail E. 2021. Emergence of a Novel SARS-CoV-2 Variant in Southern California. JAMA 325:1324-1326.

24. Frampton D, Rampling T, Cross A, Bailey H, Heaney J, Byott M, Scott R, Sconza R, Price J, Margaritis M, Bergstrom M, Spyer MJ, Miralhes PB, Grant P, Kirk S, Valerio C, Mangera Z, Prabhahar T, Moreno-Cuesta J, Arulkumaran N, Singer M, Shin GY, Sanchez E, Paraskevopoulou SM, Pillay D, McKendry RA, Mirfenderesky M, Houlihan CF, Nastouli E. 2021. Genomic characteristics and clinical effect of the emergent SARS-CoV-2 B.1.1.7 lineage in London, UK: a whole-genome sequencing and hospital-based cohort study. Lancet Infect Dis 21:1246-1256.

25. Mlcochova P, Kemp SA, Dhar MS, Papa G, Meng B, Ferreira IATM, Datir R, Collier DA, Albecka A, Singh S, Pandey R, Brown J, Zhou J, Goonawardane N, Mishra S, Whittaker C, Mellan T, Marwal R, Datta M, Sengupta S, Ponnusamy K, Radhakrishnan VS, Abdullahi A, Charles O,

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

Chattopadhyay P, Devi P, Caputo D, Peacock T, Wattal C, Goel N, Satwik A, Vaishya R, Agarwal M, Chauhan H, Dikid T, Gogia H, Lall H, Verma K, Dhar MS, Singh MK, Soni N, Meena N, Madan P, Singh P, Sharma R, Sharma R, Kabra S, Kumar S, Kumari S, Sharma U, et al. 2021. SARS-CoV-2 B.1.617.2 Delta variant replication and immune evasion. Nature 599:114-119.

26. Wang H, Miller JA, Verghese M, Sibai M, Solis D, Mfuh KO, Jiang B, Iwai N, Mar M, Huang C, Yamamoto F, Sahoo MK, Zehnder J, Pinsky BA. 2021. Multiplex SARS-CoV-2 Genotyping Reverse Transcriptase PCR for Population-Level Variant Screening and Epidemiologic Surveillance. J Clin Microbiol 59:e0085921.

27. Sagulenko P, Puller V, Neher RA. 2018. TreeTime: Maximum-likelihood phylodynamic analysis. Virus Evol 4:vex042.

28. Bossuyt PM, Reitsma JB, Bruns DE, Gatsonis CA, Glasziou PP, Irwig L, Lijmer JG, Moher D, Rennie D, de Vet HC, Kressel HY, Rifai N, Golub RM, Altman DG, Hooft L, Korevaar DA, Cohen JF, Group S. 2015. STARD 2015: an updated list of essential items for reporting diagnostic accuracy studies. BMJ 351:h5527.

29. Verghese M, Jiang B, Iwai N, Mar M, Sahoo MK, Yamamoto F, Mfuh KO, Miller J, Wang H, Zehnder J, Pinsky BA. 2021. A SARS-CoV-2 Variant with L452R and E484Q Neutralization Resistance Mutations. J Clin Microbiol 59:e0074121.

30. Borsova K, Paul ED, Kovacova V, Radvanszka M, Hajdu R, Cabanova V, Slavikova M, Lickova M, Lukacikova L, Belak A, Roussier L, Kosticova M, Liskova A, Madarova L, Stefkovicova M, Reizigova L, Novakova E, Sabaka P, Koscalova A, Brejova B, Staronova E, Misik M, Vinar T, Nosek J, Cekan P, Klempa B. 2021. Surveillance of SARS-CoV-2 lineage B.1.1.7 in Slovakia using a novel, multiplexed RT-qPCR assay. Sci Rep 11:20494.

31. Korukluoglu G, Kolukirik M, Bayrakdar F, Ozgumus GG, Altas AB, Cosgun Y, Ketre Kolukirik CZ. 2021. 40 minutes RT-qPCR Assay for Screening Spike N501Y and HV69-70del Mutations. bioRxiv doi:10.1101/2021.01.26.428302:2021.01.26.428302.

32. Perchetti GA, Zhu H, Mills MG, Shrestha L, Wagner C, Bakhash SM, Lin MJ, Xie H, Huang ML, Mathias P, Bedford T, Jerome KR, Greninger AL, Roychoudhury P. 2021. Specific allelic

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

discrimination of N501Y and other SARS-CoV-2 mutations by ddPCR detects B.1.1.7 lineage in Washington State. J Med Virol 93:5931-5941.

33. Banada P, Green R, Banik S, Chopoorian A, Streck D, Jones R, Chakravorty S, Alland D. 2021. A Simple Reverse Transcriptase PCR Melting-Temperature Assay To Rapidly Screen for Widely Circulating SARS-CoV-2 Variants. J Clin Microbiol 59:e0084521.

34. Bal A, Destras G, Gaymard A, Stefic K, Marlet J, Eymieux S, Regue H, Semanas Q, d'Aubarede C, Billaud G, Laurent F, Gonzalez C, Mekki Y, Valette M, Bouscambert M, Gaudy-Graffin C, Lina B, Morfin F, Josset L, Group CO-DHS. 2021. Two-step strategy for the identification of SARS-CoV-2 variant of concern 202012/01 and other variants with spike deletion H69-V70, France, August to December 2020. Euro Surveill 26.

35. Neopane P, Nypaver J, Shrestha R, Beqaj SS. 2021. SARS-CoV-2 Variants Detection Using TaqMan SARS-CoV-2 Mutation Panel Molecular Genotyping Assays. Infect Drug Resist 14:4471-4479.

36. Vogels CBF, Breban MI, Ott IM, Alpert T, Petrone ME, Watkins AE, Kalinich CC, Earnest R, Rothman JE, Goes de Jesus J, Morales Claro I, Magalhaes Ferreira G, Crispim MAE, Brazil UKCGN, Singh L, Tegally H, Anyaneji UJ, Network for Genomic Surveillance in South A, Hodcroft EB, Mason CE, Khullar G, Metti J, Dudley JT, MacKay MJ, Nash M, Wang J, Liu C, Hui P, Murphy S, Neal C, Laszlo E, Landry ML, Muyombwe A, Downing R, Razeq J, de Oliveira T, Faria NR, Sabino EC, Neher RA, Fauver JR, Grubaugh ND. 2021. Multiplex qPCR discriminates variants of concern to enhance global surveillance of SARS-CoV-2. PLoS Biol 19:e3001236.

37. Hale R, Crowley P, Dervisevic S, Coupland L, Cliff PR, Ebie S, Snell LB, Paul J, Williams C, Randell P, Pond M, Stanley K. 2021. Development of a Multiplex Tandem PCR (MT-PCR) Assay for the Detection of Emerging SARS-CoV-2 Variants. Viruses 13.

38. Garson JA, Badru S, Parker E, Tedder RS, McClure MO. 2022. Highly sensitive and specific detection of the SARS-CoV-2 Delta variant by double-mismatch allele-specific real time reverse transcription PCR. J Clin Virol 146:105049.

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

39. Aoki A, Adachi H, Mori Y, Ito M, Sato K, Okuda K, Sakakibara T, Okamoto Y, Jinno H. 2021. A rapid screening assay for L452R and T478K spike mutations in SARS-CoV-2 Delta variant using high-resolution melting analysis. J Toxicol Sci 46:471-476.

40. Barua S, Hoque M, Kelly PJ, Bai J, Hanzlicek G, Noll L, Walz H, Johnson C, Kyriakis C, Wang C. 2021. High-resolution melting curve FRET-PCR rapidly identifies SARS-CoV-2 mutations. J Med Virol 93:5588-5593.

41. Hamill V, Noll L, Lu N, Tsui WNT, Porter EP, Gray M, Sebhatu T, Goerl K, Brown S, Palinski R, Thomason S, Almes K, Retallick J, Bai J. 2021. Molecular detection of SARS-CoV-2 strains and differentiation of Delta variant strains. Transbound Emerg Dis doi:10.1111/tbed.14443.

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

FIGURE LEGENDS

Figure 1. Summary of current World Health Organization (WHO)-designated variants of concern (VOC) along with their expected spike mutations at sites targeted by this two-reaction multiplex SARS-CoV-2 RT-qPCR genotyping approach. These reactions are designed to detect the following mutations: del69-70, K417N, L452R, T478K, E484K, and N501Y. Shading indicates predicted versus empiric performance of this assay for the detection and differentiation of these VOCs. While the del69-70, T478K, and N501Y mutations were all predicted to be not-detected by this assay in samples from Omicron-infected individuals due to known adjacent mutations in the probe binding site (A67V, S477N, Q498R), del69-70 was detected empirically with diminished efficiency. An asterisk denotes a known limitation of the assay in differentiating VOCs Beta and Gamma from the variant of interest Mu.

Figure 2. Nextclade phylogenetic tree of 3,097 SARS-CoV-2 genomes, including all 732 of the sequenced genomes from this study, and 2,365 genomes from the Nextstrain global reference tree as of February 2, 2022. The 732 included genomes are colored by RT-qPCR genotyping predicted variant type, with each circle representing a sequenced genome. Branch length corresponds to nucleotide divergence. Sequenced genomes span the breadth of the reference tree. Annotation to the right of the tree demonstrates the variant type based on whole-genome sequencing (WGS). Variant determination by RT-qPCR matched WGS except for 1 sequence typed as Beta, and 4 sequences typed as Gamma by RT-qPCR which clustered with variant of interest Mu by WGS.

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

TABLES

Table 1. Reaction 2 Primer and Probe Oligonucleotide Sequences

	Name	Sequence (5' → 3')	Final Concentration
Primers	del69-70_FWD	CATTAAATGGTAGGACAGGGTTA	300 nM
	del69-70_REV	ACATTCAACTCAGGACTTGTT	300 nM
	K417N_FWD	GCAGCCTGTAAAATCATCTG	300 nM
	K417N_REV	CATTTGTAATTAGAGGTGATGAAGTC	300 nM
	T478K_FWD	AAAGGAAAGTAACAATTAAAACCT	300 nM
	T478K_REV	AGGAAGTCTAATCTCAAACCT	300 nM
Probes	del69-70_MT_HEX	HEX-CCTAAACAATCTATACCGGTAATT-BHQ1	50 nM
	wt69-70_WT_CY3.5[a]	CY3.5-GGTCCCAGAGACATGTATAG-BHQ2	50 nM
	K417N_MT_CY5	CY5-TAATCAGCAATATTTCCAGT-BHQ2	50 nM
	T478K_MT_FAM	FAM-ACCATTACAAGGTTTGCTAC-BHQ1	50 nM

FWD, forward; REV, reverse; WT, wild-type; MT, mutant; HEX, hexachlorofluorescein; CY3.5, cyanine 3.5; CY5, cyanine 5; FAM, 5(6)-carboxyfluorescein; BHQ, Black hole quencher

[a] Included as an internal amplification control for samples without the del69-70 mutation

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

Table 2. Comparison of RT-qPCR and WGS Results for SARS-CoV-2 Spike Gene Mutation Detection in the Initial Cohort (n=502)

Spike Mutations		WGS pos	WGS neg	PPA (95% CI)	NPA (95% CI)
Del69-70	**RT-qPCR pos**	43	0	100% (92-100%)	100% (99-100%)
	RT-qPCR neg	0	459		
K417N	**RT-qPCR pos**	27	0	96% (82-100%)	100% (99-100%)
	RT-qPCR neg	1[a]	474		
L452R	**RT-qPCR pos**	403	5[b]	100% (99-100%)	95% (89-98%)
	RT-qPCR neg	0	94		
T478K	**RT-qPCR pos**	379	0	100% (99-100%)	100% (97-100%)
	RT-qPCR neg	0	123		
E484K	**RT-qPCR pos**	35	3[c]	100% (90-100%)	99% (98-100%)
	RT-qPCR neg	0	464		
N501Y	**RT-qPCR pos**	70	0	100% (95-100%)	100% (99-100%)
	RT-qPCR neg	0	432		

RT-qPCR, reverse transcription quantitative polymerase chain reaction; WGS, whole-genome sequencing; PPA, positive percent agreement; NPA, negative percent agreement; CI, confidence interval

[a] False negative RT-qPCR result due to synonymous mutation in spike gene amino acid position 418 (codon ATT -> ATC) causing probe dropout.

[b] False negative WGS results due to insufficient read count (<12) at this codon. Manual review of sequences revealed 3-9 mutant reads in each sample.

[c] These three samples were found on WGS to be positive for E484Q. While positive for the E484K target on RT-qPCR, these samples had a distinct blunted amplification curve associated with E484Q as previously described (29).

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

Table 3. Comparison of RT-qPCR and WGS for SARS-CoV-2 Variant of Concern Detection (n=732)

WGS		RT-qPCR						
WHO VOC	**PANGO lineage**	**Alpha**	**Beta**	**Gamma**	**Delta**	**Omicron**	**Not a VOC**	**All**
Alpha	All Alpha	43	-	-	-	-	-	43
	B.1.1.7	37	-	-	-	-	-	37
	Q.3	6	-	-	-	-	-	6
Beta	Beta	-	2	-	-	-	-	2
	B.1.351	-	2	-	-	-	-	2
Gamma	All Gamma	-	-	20	-	-	-	20
	P.1	-	-	13	-	-	-	13
	P.1.10	-	-	5	-	-	-	5
	P.1.17	-	-	2	-	-	-	2
Delta	All Delta	-	-	-	378	-	-	378
	B.1.617.2	-	-	-	29	-	-	29
	AY.1	-	-	-	20	-	-	20
	AY.2	-	-	-	5	-	-	5
	AY.3	-	-	-	5	-	-	5
	AY.4	-	-	-	1	-	-	1
	AY.13	-	-	-	32	-	-	32
	AY.14	-	-	-	59	-	-	59
	AY.19	-	-	-	1	-	-	1
	AY.20	-	-	-	5	-	-	5
	AY.23	-	-	-	1	-	-	1
	AY.25	-	-	-	7	-	-	7
	AY.25.1	-	-	-	25	-	-	25
	AY.26	-	-	-	15	-	-	15
	AY.35	-	-	-	2	-	-	2
	AY.43	-	-	-	2	-	-	2
	AY.44	-	-	-	77	-	-	77
	AY.46.2	-	-	-	1	-	-	1
	AY.47	-	-	-	8	-	-	8
	AY.48	-	-	-	1	-	-	1
	AY.52	-	-	-	1	-	-	1
	AY.54	-	-	-	3	-	-	3
	AY.59	-	-	-	1	-	-	1
	AY.62	-	-	-	1	-	-	1
	AY.67	-	-	-	3	-	-	3
	AY.74	-	-	-	1	-	-	1
	AY.75	-	-	-	10	-	-	10
	AY.98.1	-	-	-	1	-	-	1
	AY.100	-	-	-	3	-	-	3
	AY.103	-	-	-	26	-	-	26
	AY.110	-	-	-	9	-	-	9
	AY.114	-	-	-	1	-	-	1
	AY.116.1	-	-	-	2	-	-	2
	AY.118	-	-	-	5	-	-	5
	AY.119	-	-	-	4	-	-	4
	AY.120.1	-	-	-	1	-	-	1
	AY.121	-	-	-	3	-	-	3
	AY.122	-	-	-	5	-	-	5
	AY.126	-	-	-	2	-	-	2
Omicron	All Omicron	-	-	-	-	230	-	230
	BA.1	-	-	-	-	123	-	123
	BA.1.1	-	-	-	-	107	-	107
Not a VOC	All Non- VOC	-	1	4	-	-	54	59

medRxiv preprint doi: https://doi.org/10.1101/2022.02.07.22270629; this version posted February 8, 2022. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY-NC-ND 4.0 International license .

	A.2.5	-	-	-	-	-	6	6
	B.1	-	-	-	-	-	3	3
	B.1.1.318	-	-	-	-	-	1	1
	B.1.1.519	-	-	-	-	-	1	1
	B.1.311	-	-	-	-	-	1	1
	B.1.427	-	-	-	-	-	3	3
	B.1.429	-	-	-	-	-	8	8
	B.1.526	-	-	-	-	-	10	10
	B.1.621[a]	-	1	2	-	-	-	3
	BB.2[a]	-	-	2	-	-	-	2
	B.1.627	-	-	-	-	-	1	1
	B.1.637	-	-	-	-	-	11	11
	XB	-	-	-	-	-	9	9
All	All Variants	43	3	24	378	230	54	732

WGS, whole-genome sequencing; RT-qPCR, reverse transcription quantitative polymerase chain reaction; WHO, World Health Organization; VOC, variant of concern

[a] Variant of interest Mu with E484K and N501Y mutations, and a subset with K417N, which overlaps with VOCs Beta and Gamma

477

	Spike Protein Amino Acid Position					
WHO VOC	**69-70**	**K417**	**L452**	**T478**	**E484**	**N501**
Alpha	deletion	K	L	T	E	Y
Beta*	WT	N	L	T	K	Y
Gamma*	WT	T	L	T	K	Y
Delta	WT	K	R	K	E	N
Omicron	deletion	N	L	K	A	Y

- Mutation is predicted to be detected & is empirically detected in this variant
- Mutation is predicted to be not detected & is empirically not detected in this variant
- Mutation is predicted to be not detected but can be empirically detected in this variant


Variant RT-qPCR
Alpha
Beta
Gamma
Delta
Omicron
Not a VOC
Delta
Alpha
Gamma
Omicron
Beta
Mu
0
10
20
30
40
50
60
70
Mutations