Accuracy of tongue swab testing using Xpert MTB-RIF Ultra for tuberculosis diagnosis

Andama A[1,2β], Whitman GR[3β], Crowder R[4,5], Reza TF[4,5], Jaganath D[5,6], Mulondo J[2], Nalugwa TK[4], Semitala C. F[1,2], Worodria W[1,7], Cook C[5], Wood RC[3], Weigel KM[3], Olson AM[3], Lohmiller Shaw J[3], Denkinger C[8], Nahid P[4,5], Cangelosi G[3], and Cattamanchi A[4,5].*

[1]Department of Medicine, College of Health Sciences, Makerere University, Kampala, Uganda.

[2]Infectious Diseases Research Collaboration, Kampala, Uganda.

[3]Department of Environmental and Occupational Health Sciences, School of Public Health, University of Washington, Seattle, WA USA.

[4]Uganda Tuberculosis Implementation Research consortium, Kampala, Uganda

[5]Division of Pulmonary and Critical Care Medicine and Center for Tuberculosis, San Francisco General Hospital, University of California, San Francisco, CA USA.

[6]Division of Pediatric Infectious Diseases, University of California, San Francisco, CA USA

[7]Mulago National Referral Hospital, Kampala, Uganda.

[8]University Hospital of Heidelberg, Germany

β These authors contributed equally to this work.

*Corresponding author

Abstract

Tongue dorsum swabs have shown promise as alternatives to sputum for detecting *Mycobacterium tuberculosis* (MTB) in patients with pulmonary tuberculosis (TB). Some of the most encouraging results have come from studies that used manual quantitative PCR (qPCR) to analyze swabs. Studies using the automated Cepheid Xpert® MTB/RIF Ultra qPCR test (Xpert Ultra) have yielded less encouraging results with tongue swabs, possibly because Xpert Ultra is optimized for testing sputum, not tongue swab samples. Using two new sample processing methods that demonstrated good sensitivity in preliminary experiments, we assessed diagnostic accuracy and semi-quantitative signals of 183 tongue swab samples using Xpert Ultra in a clinical setting. Relative to a sputum Xpert Ultra reference standard, the sensitivity of tongue swab Xpert Ultra was 68.5% (95% CI 54.4-80.5) and specificity was 100.0% (95% CI 97.2-100.0). When compared to a microbiological reference standard (MRS) incorporating both sputum Xpert Ultra and sputum culture, sensitivity was 64.9 (95% CI 51.1-77.1) and specificity remained the same. Higher sensitivity was observed (77.8% CI 64.4-88.0) when "trace" Xpert results were included among positive swabs. Semi-quantitative Xpert Ultra results were generally lower with swabs than with sputum. None of the eight sputum Xpert Ultra "trace" or "very low" results were detected using tongue swabs. Tongue swabs should be considered when sputum cannot be collected for Xpert Ultra testing, or in certain mass-screening settings. Further optimization of tongue swab analysis is needed to achieve parity with sputum-based molecular testing for TB.

Abstract word count: 240

1. Introduction

Each year, an estimated 10 million people develop tuberculosis (TB) and over one million die of TB despite the availability of effective treatment for most forms of the disease [1]. A significant barrier to current TB care and control efforts is the difficulty of rapidly screening large numbers of people for active disease. Historically, TB diagnosis has relied on passive case finding, in which people with symptoms self-report to a healthcare facility for further evaluation. Pulmonary TB diagnosis most often relies on microbiological or molecular analysis of sputum for the presence of *Mycobacterium tuberculosis* (MTB) cells or DNA to confirm diagnosis [2]. The production of sputum can be burdensome for those providing the specimen and hazardous to health workers and others when adequate safety measures are not in place. Moreover, children and people living with HIV (PLHIV) are often unable to produce adequate sputum [3, 4]. These limitations have spurred efforts to identify alternative sample types for TB detection that are easier, less invasive, and safer to collect.

Molecular analysis of sputum, such as with Xpert MTB/RIF Ultra (Xpert, Cepheid, USA) is now the first-line test recommended for diagnosis of pulmonary TB [5-7]. Sputum sample processing for Xpert testing involves the addition of Cepheid's proprietary Sample Reagent (SR). This reagent liquefies and decontaminates the sample without lysing bacilli. Once the treated sample is loaded into the cartridge and the run started, the sample is drawn through a membrane filter, trapping intact bacilli. After wash steps, a sonic horn lyses the bacilli on the membrane and releases their DNA. The DNA is then subjected to a semi-nested quantitative polymerase chain-reaction (qPCR), targeting the rifampicin resistance determining region (RRDR) of the

rpoB gene. A mixed two-insertion sequence (IS1081 and IS6110) probe is also included to enhance detection of low bacillary load samples [5, 8].

Tongue swabbing, also known as oral swab analysis (OSA), has emerged as a potential alternative to sputum-based molecular testing [9]. Compared to sputum collection, tongue swabbing is faster, easier, and safer. Because tongue swabs are easy to collect from any person in any setting, OSA may be especially useful for non-clinical and community-based screening. We have previously optimized and clinically validated OSA procedures for manual DNA extraction and IS6110-targeted qPCR [10-14]. Sensitivity in adults relative to positive sputum Xpert or positive sputum culture has ranged from 88% to 93%, with specificity ranging from 79% to 92% [11, 13].

Given the widespread use of Xpert for TB diagnosis in high burden countries, it is important to harmonize protocols for Xpert testing using new sampling methods. To date, only a few studies have evaluated OSA in conjunction with Xpert testing. Success has been limited, with reported sensitivities ranging from ~45% relative to sputum culture [15] to ~51% relative to sputum Xpert [16]. Although these studies utilized different swabbing and sample handling methods, the results underscore the need to optimize and standardize Xpert protocols for detecting MTB in tongue swabs.

In this study, we sought to identify and characterize improved methods and provide updated guidance and protocols for use of Xpert to detect MTB in tongue swabs using tongue swab samples spiked with standardized cultured MTB cells. We then evaluated the diagnostic accuracy of tongue swab Xpert when using the updated protocols among adults being evaluated for pulmonary TB at two outpatient clinics in Kampala, Uganda.

2. Materials and Methods

2.1. Method optimization experiments.

Contrived samples consisting of tongue swabs collected from healthy volunteers spiked with cultured MTB were used for optimization of tongue swab processing methods (Supplemental information). Study participants for these experiments were adults (>18 years) recruited from the University of Washington, School of Public Health, Department of Environmental and Occupational Health Sciences. Participants provided samples between November 2019 and February 2021. Participants were assumed to be TB-negative based on low risk of exposure and lack of symptoms. Inclusion criteria and screening protocols have been described previously by Wood et al, 2015 [12]. A total of 15 participants were enrolled and sampled. Participants were contacted for repeated sampling as needed. In order to protect the safety of participants and study personnel, swabs were self-collected with study staff leaving the room while collection took place. Procedures for collection of swab samples from human participants were approved of by the University of Washington Human Subjects Division (STUDY00001840).

2.2. Clinical evaluation.

For the clinical evaluation of Xpert performed on tongue swabs, we conducted a cross-sectional study between April and September 2021 at outpatient clinics at Mulago National Referral Hospital and Kisenyi Health Center IV in Kampala, Uganda. We screened consecutive adults (>18 years) presenting to the health centers for any reason for cough lasting more than two weeks. Among those with cough lasting more than two weeks, we excluded patients who 1) had completed latent or active TB treatment within

the past 12 months; 2) had taken any medication with anti-mycobacterial activity

(including fluoroquinolones) for any reason, within 2 weeks of study entry; 3) resided

>20km from the study site or were unwilling to return for follow-up visits; or 4) were

unwilling or unable to provide informed consent. The Makerere University School of

Medicine Research and Ethics Committee, the Uganda National Council for Science

and Technology and the University of California San Francisco Committee on Human

Research reviewed and approved the study. The study was conducted and reported in

accordance with the Standards for Reporting of Diagnostic Accuracy Studies (STARD)

guidelines [2].

2.3. Method development and limit of detection (LoD) quantification.

Using Copan FLOQSwab® tongue swabs collected from TB-negative volunteers

and spiked with cultured MTB cells, we evaluated three strategies for improving the

sensitivity of Xpert performed on tongue swabs. The LoD of each method was

compared to that of manual methods reported previously [11, 13]. Methods are

presented in detail in the Supplemental Information and summarized here. In Method 1,

a single FLOQSwab was processed with Cepheid Sample Reagent (SR) using a

protocol similar to that recommended by the manufacturer for sputum. In Method 2, two

FLOQSwabs were collected and combined into a single tube and processed as in

Method 1. In Method 3, a single FLOQSwab was processed by boiling, incubation, and

mixing (without using SR). Methods 1 and 2, which used SR, were assessed because

they closely resemble sputum processing protocols currently recommended by Cepheid

for Xpert Ultra. Method 2 (SR, two swabs) tested the hypothesis that the processing

steps applied to one FLOQSwab in Method 1 could be applied to two FLOQSwabs to

improve the sensitivity of TB detection. A recent study demonstrated that a single

flocked swab collects ≤10% of the tongue dorsum biomass that is available for TB

testing [13]. Therefore, if two flocked swabs are collected in succession, suspended in

buffer at the same volume as a single swab, and tested as a single sample, then the

collection of MTB bacilli from TB patients may be doubled on average relative to single

swabbing. Method 3 (boil method) was selected for this study because it resembled the

first steps of our previous manual methods [11-14].

Preliminary evaluations of these methods used tongue swab samples collected

from healthy US volunteers as follows. Samples were collected using COPAN

FLOQSwabs® Regular Flocked Swab with an 80 mm or 30 mm breakpoint (Copan Italia

SpA, Brescia Italy). These swabs were shown to maximize tongue dorsum biomass

collection, relative to other swab products [13]. Participants self-swabbed along the

breadth of the mid-tongue dorsum, firmly pressing and rolling the swab head for

approximately 10 seconds. Swab heads were then immediately spiked with 10 µL of

serially-diluted cultured MTB strain H37Ra stored in 1x phosphate buffered saline with

15% glycerol and 0.05% Tween® 80 (PBSGT), or with blank PBSGT. Swab heads were

then broken off into 5 mL polypropylene transport tubes (Corning) containing 800µL

sterile 1x Tris-EDTA (TE) buffer (10 mM Tris-HCl containing 1 mM EDTA•Na$_2$, pH 8.0)

(Corning) for Methods 1-3.

LoD's of the experimental methods were quantified using serial dilutions as

described in the Supplemental Information. LoDs are reported on a colony forming unit

(CFU) per swab basis (which is the same as the CFU per sample amount for Xpert

Methods 1 and 3, but is half of the CFU per swab basis for Method 2 which included 2

swabs each spiked with the same number of bacilli as a single swab). Raw dose-

response data are shown in **Table S1**. Exploratory runs at some dilutions (*e.g.*, 100

cfu/swab) increased the number of runs at those dilutions, relative to other dilutions.

Results of all experiments were used in LoD calculations. This minimized bias that could

have resulted from the designation of specific runs as exploratory.

2.4. Clinical evaluation procedures.

Upon enrollment at the outpatient clinics, we obtained detailed demographic and

medical history using a standardized form, collected blood for HIV testing and CD4

count (if HIV-positive) and collected up to three expectorated sputum specimens. One

sputum sample was used to perform Xpert testing for all patients (an additional sample

was collected and tested if the initial Xpert result was trace-positive). If the initial or

repeat sputum Xpert was negative or invalid, the remaining two sputum samples were

cultured in liquid media. Trained laboratory technologists conducted all TB testing

following standard protocols for Xpert testing [17] and mycobacterial culture [18].

For culture, sputum samples were digested and decontaminated using the

sodium hydroxide/N-acetyl-L-cysteine method with a final concentration of 1.5%,

neutralized with sterile phosphate-buffered solution, centrifuged, and re-suspended in

phosphate-buffered solution. One Mycobacterial Growth Indicator Tube (MGIT) was

inoculated with each 0.5mL decontaminated sample. When MGIT tubes were not

available, solid 7h10 media was used for culture, and in the event of a negative 7h10

culture result, leftover sputum was preserved for culture in liquid media (MGIT) when

available. During a period where MGIT tubes were not available, 32 patients had

mycobacterial cultures performed using 7H10 media (of whom 28 subsequently had

MGIT cultures done using frozen sputum sediment).

Prior to inoculation a cocktail of antibiotics containing PANTA (polymyxin,

amphotericin B, nalidixic acid, trimethoprim, azlocillin) mixed with OADC (oleic acid,

bovine serum albumin, dextrose, and catalase) were added to MGIT tubes. MGIT tubes

were incubated in a BACTEC MGIT 960 instrument (BD; Franklin Lakes, USA) for up to

42 days. Blood agar plates were prepared to check for contamination of positive MGIT

tubes, and a stained Ziehl Neelsen (ZN) smear was examined to check for the presence

of acid fast bacilli (AFB). Speciation was then performed to confirm the presence of *M.*

tuberculosis complex using SD Bioline strips (SD MPT64TB Ag kit, South Korea).

Prior to sputum collection, patients were also asked to self-collect up to 3 tongue

swabs (COPAN FLOQswab) using the same procedure described for healthy volunteers

in Washington. Two tongue swabs were placed in a single tube with 800μL of TE buffer

and processed within one hour for Xpert testing using Method 2. In a subset of patients,

an additional single swab stored at -80 degrees Celsius in a tube with 800μL of TE

buffer was processed for Xpert testing using Method 3.

2.5. Definitions of reference standard and index tests for the clinical evaluation.

We considered patients to have TB if they had a positive result on sputum Xpert

Ultra (grade very low, low, medium, high), two trace-positive sputum Xpert Ultra results

or a culture result positive for MTB. We considered patients not to have TB if they had

no positive result (sputum Xpert Ultra and culture), and two negative cultures. We

considered patients to have indeterminate TB status if they had no positive result

(sputum Xpert Ultra and culture) or <2 trace-positive sputum Xpert Ultra results but <2

negative MGIT cultures due to contamination.

Staff performing tongue swab Xpert Ultra (the index test) were blinded to

reference standard TB test results. The primary clinical evaluation analysis for Methods

2 and 3 followed the WHO recommendations for interpreting Xpert results, with trace-

positive results counted as positive for HIV+ participants only. Secondary analyses

considered trace-positive results positive for all participants.

2.6. Statistical approach of the clinical evaluation.

We summarized demographic and clinical data using appropriate descriptive

statistics. We calculated sensitivity and specificity of tongue swab Xpert in reference to

sputum Xpert results alone, and in reference to sputum Xpert and culture results (with

TB status defined as described above). The difference in sensitivity between Method 2

and Method 3 was assessed using McNemar's test. We used STATA 15 (StataCorp,

College Station, TX, USA) for all analyses.

3. Results

3.1 Limits of detection of tongue swab processing methods for Xpert testing.

Figure S1 plots the limits of detection (LoD) curves of each of the described

methods, and LoDs are summarized in Table 1. Method 1 (SR, single swab), with 69

spiked samples, was relatively insensitive, with an LoD of 101.7 CFU/swab (95% CI

64.5 – 144.0). With 64 spiked samples, Method 2 (SR, double swab) improved

sensitivity with a LoD of 76.5 CFU/swab (95% CI 54.2 – 104.1). Method 3 (boil method)

demonstrated the highest sensitivity, equal to or better than the manual methods used

previously (Table 1). With a total of 38 spiked samples represented in the dose-
response data, Method 3 exhibited an LoD of 22.3 CFU/swab (95% CI 15.3 – 34.3). For
purposes of comparison, with 40 spiked samples, the LoD of our clinically-validated
manual qPCR method was 53.5 CFU/swab (95% CI 36.9 – 73.0). This value fell
between those of Methods 1 and 2 (**Table 1**), and therefore Method 2 was selected for
the main clinical evaluation. Of note, Methods 1-3 had no false positives among
negative control tongue swab samples (n = 27). Additionally, there were no false-
positive rifampin resistance determinations (n = 61).

3.2. Study population.

Between April 14 and September 9, 2021, 184 eligible patients were enrolled and
underwent tongue swab Xpert testing using Method 2 (SR, two swabs). One patient had
an indeterminate tongue swab Xpert result and was excluded from this analysis. Among
the remaining 183 patients, median age was 33 years; 76 (41.5%) were female, 58
(31.7%) were HIV-positive, and 22 (12.0) had previously been diagnosed with TB
(**Table 2**). TB symptoms were common, with 142 (77.6%) experiencing weight loss, 137
(74.9%) fever, 122 (66.7%) night sweats, 115 (62.8%) decrease in appetite, 26 (14.2%)
bumps in neck, armpit, or groin, and 25 (13.7%) had hemoptysis (**Table 2**).

3.3. Diagnostic accuracy.

Relative to sputum Xpert, the sensitivity of tongue swab Xpert was 68.5% (95%
CI 54.4, 80.5) and specificity was 100.0% (95% CI 97.2-100.0). Sensitivity reduced to
64.9% (95% CI 51.1, 77.1) against the microbiological reference standard and
specificity remained 100% (95% CI 96.8, 100). When all trace tongue swab Xpert

results were considered positive, sensitivity improved to 77.8% (95% CI 64.4, 88.0) with

no impact on specificity (**Table 3**). Sensitivity of tongue swab Xpert was consistently

higher among patients with a higher mycobacterial load, as assessed using sputum

Xpert Ultra (**Table 4**). Semi-quantitative Xpert results were generally lower with tongue

swabs than with sputum samples, with none of the eight sputum Xpert "trace" or "very

low" results being detected by tongue swabs (**Table 5**).

In a subset of 37 patients, 35 (94.6%) with sputum Xpert Ultra positive

(semiquantitative result of very low or higher) and 2 with sputum Xpert Ultra trace

positive results, frozen tongue swab samples were also processed and tested using

Method 3 (boil method). The two participants with sputum Xpert Ultra trace positive

results were not detected by either tongue swab method. Among the remaining 35

participants with very low or higher semi-quantitative sputum Xpert Ultra positive results,

31 were identified by both swab methods, 2 were identified by Method 3 and missed by

Method 2, and 2 were identified by Method 2 but missed by Method 3, resulting in no

difference in sensitivity. When all sputum Xpert Ultra trace positive results were

considered to be true positive, Method 3 identified 3 positives missed by Method 2,

while Method 2 identified 1 positive that was missed by Method 3 (sensitivity difference

5.7% [95% CI -19.6 to 8.2, p-value=0.6]).

4. Discussion

OSA is a simpler alternative to sputum collection for TB testing, but standardized

methods for pairing tongue swabs with widely available molecular testing platforms are

needed. In laboratory spiking experiments, we showed that LoD is comparable to

manual DNA extraction and qPCR when two tongue swabs are processed for Xpert

testing using a protocol similar to that used for sputum-based testing. In clinical

evaluation, this method resulted in three-quarters of patients with TB being detected by

Xpert testing of tongue swabs while retaining high specificity (100%, 95% CI 95.8-100).

These data highlight that OSA should be considered an alternative when sputum cannot

be collected, and the need for continued optimization of OSA-based TB diagnosis to

detect pauci-bacillary disease.

Compared with previous studies of OSA-based Xpert testing, we show a marked

improvement in sensitivity while retaining high specificity. Sensitivity was 45% in a study

from Peru that collected a single cheek swab using one of three swab types and storage

buffers, and processed swabs for Xpert testing in a manger similar to Methods 1 and 2

used in our study. Based on our previous studies, more bacterial biomass is present on

the tongue dorsum than on cheeks, and FLOQswabs pick up more biomass than the

swab types used in the Peru study. In a study of mass TB screening among prisoners in

Brazil, sensitivity was 51% overall when testing two tongue swabs sequentially and

ranged from 38% to 90% in patients with very low/low to high semi-quantitative sputum

Xpert results. Sensitivity was modestly higher in our study, especially among patients

with higher mycobacterial load, supporting the use of our protocol involving the use of

FLOQswabs and simultaneous testing of two swabs processed with Xpert SR.

However, our results showed the same pattern of inability to detect paucibacillary

disease.

Although our data support the use of tongue swabs when sputum cannot be

collected, there remains a need to further optimize OSA-based molecular testing for TB.

Our laboratory experiments showed that boiling a single swab results in lower LoD

compared with processing one or two swabs with Xpert SR. However, benefits were

marginal in our small clinical sub-study of the boil method. We reported recently that

flocked tongue swabs collect only a small percentage of the biomass that is present on

tongue dorsa [13]. Therefore, efforts to optimize sensitivity are focusing on the use of

alternative swab products that collect more biomass.

Key strengths of the study include the use of laboratory- and clinic-based studies

to validate protocols for Xpert testing using tongue swabs and clinical validation among

a consecutive sample of patients with cough identified through clinic-based active TB

case finding. However, there are also some limitations to this study. Our clinical study

population is reflective of patients seeking healthcare in Uganda. OSA-based Xpert

testing may have lower sensitivity if applied for TB screening among high-risk but

asymptomatic individuals. Our clinical sub-study of the boil method was limited to a

small subset of patients. A larger evaluation may confirm some benefits, but unless the

difference is meaningful, the additional laboratory work required is unlikely to be taken

up in routine practice. Last, we did not confirm positive sputum Xpert Ultra results with

culture. However, Xpert has been shown to have very high specificity (>98%) for TB

[20].

In summary, we demonstrate that Xpert testing of tongue swabs can be a viable

alternative when sputum collection is not feasible and should be done using

standardized protocols. With further improvements in sample collection and processing

and leveraging advances in molecular testing of swabs for SARS-CoV-2, OSA-based

molecular testing has the potential to open the door to expanded TB case-finding

efforts. This non-invasive, fast, and safer diagnostic specimen and processing method

may provide a useful new tool in the global fight against TB.

Acknowledgements

This research was supported by NIH (NIAID) grant #s U01AI152087 (AC) and R01AI139254 (GC), and Global Health Labs (AC). We are grateful to Cepheid, Inc., for providing research-use-only Xpert Ultra test kits for LoD experiments, to Copan Italia for providing swabs, and to the patients and staff at Mulago National Referral Hospital and Kisenyi Health Center IV.

Table 1. Comparison of Methods 1-3 and manual qPCR method LoDs

Method	Description	H37Ra LoDs in cfu/swab (95% CI)[1]
Method 1	1 FLOQswab, SR[2], Xpert Ultra ("single swab SR")	101.7 (64.5 - 144.0)
Method 2	2 FLOQswabs, SR, Xpert Ultra ("double swab SR")	76.5 (54.2 - 104.9)
Method 3	1 FLOQswab, boil w/o SR[2], Xpert Ultra ("boil method")	22.3 (15.3 - 34.3)
Manual (Reference)[3]	1 FLOQswab, Qiagen extraction and EtOH precipitation, manual IS6110 qPCR	47.6 (32.1 – 65.8)

[1]LoDs, limits of detection. Contrived samples were tongue swabs from healthy volunteers spiked with dilution series of cultured MTB H37Ra.

[2]SR, GeneXpert Sample Reagent

[3]Method used in Luabeya et al (2019) and Wood et al (2021)

Table 2. Demographic characteristics by Method 2 tongue swab Xpert Ultra result

	Total (N=183)	Tongue swab Ultra* Negative (n=146)	Tongue swab Ultra* Positive (n=37)
Female	76 (41.5)	65 (44.5)	11 (29.7)
Age, median (IRQ)	33 (26-43)	35 (28-45)	26 (23-35)
HIV-positive	58 (31.7)	53 (36.3)	5 (13.5)
CD4 count<100	10 (17.2)	9 (17.0)	1 (20.0)
Prior TB	22 (12.0)	17 (11.6)	5 (13.5)
Smoked tobacco in last 7 days	16 (8.7)	11 (7.5)	5 (13.5)
BCG	158 (86.3)	128 (87.7)	30 (81.1)
BMI, median (IQR)	21.4 (19.3-23.9)	21.7 (19.8-24.6)	19.6 (17.9-20.9)
Underweight (BMI<18.5)	36 (19.7)	23 (15.8)	13 (19.7)
Symptoms			
Cough >2 weeks	183 (100)	146 (100)	37 (100)
Fever	137 (74.9)	103 (70.6)	34 (91.9)
Hemoptysis	25 (13.7)	17 (11.6)	8 (21.6)
Night sweats	122 (66.7)	97 (66.4)	25 (67.6)
Weight loss	142 (77.6)	108 (74.0)	34 (91.9)
Decrease in appetite	115 (62.8)	86 (58.9)	29 (78.4)
Bumps in neck, armpit, groin	26 (14.2)	20 (13.7)	6 (16.2)
Proteinuria 2+	21 (11.5)	19 (13.0)	2 (5.4)

TB: tuberculosis

* Tongue swab Ultra conducted using Method 2 (SR, two swabs). Trace results considered positive for HIV+ participants only.

346 **Table 3. Diagnostic accuracy of tongue swab Xpert Ultra**

Tongue swab Ultra, Method 2 (N=183) trace+ for HIV+ only		
	Sputum Xpert	**Microbiologic reference standard***
Sensitivity	68.5 (54.4-80.5)	64.9 (51.1-77.1)
Specificity	100 (97.2-100)	100 (95.8-100)
PPV	100 (90.5-100)	100 (90.5-100)
NPV	88.4 (82.0-93.1)	81.3 (72.6-88.2)
Tongue swab Ultra, Method 2 (N=183) trace+ for all		
	Sputum Xpert	**Microbiologic reference standard***
Sensitivity	77.8 (64.4-88.0)	73.4 (59.1-83.3)
Specificity	100 (97.2-100)	100 (96.9-100)
PPV	100 (91.6-100)	100 (91.6-100)
NPV	91.5 (85.6-95.5)	88.1 (81.5-93.1)
Tongue swab Ultra, Method 3 (N=37) trace+for HIV+ only		
	Sputum Xpert	**Microbiologic reference standard***
Sensitivity	62.9 (44.9-78.5)	62.9 (44.9-78.5)
Specificity	100 (15.8-100)	100 (15.8-100)
PPV	100 (84.6-100)	100 (84.6-100)
NPV	13.3 (1.7-40.5)	13.3 (1.7-40.5)
Tongue swab Ultra, Method 3 (N=37) trace+ for all		
	Sputum Xpert	**Microbiologic reference standard***
Sensitivity	62.9 (44.9-78.5)	77.1 (59.9-89.6)
Specificity	100 (15.8-100)	100 (15.8-100)
PPV	100 (84.6-100)	100 (87.2-100)
NPV	20.0 (2.5-55.6)	20.0 (2.5-55.6)

348 * excluding 6 indeterminate MRS results (<2 negative cultures due to contamination)

Table 4. Sensitivity of tongue swab Xpert Ultra compared by sputum Xpert Ultra semiquantitative grace

Sputum Xpert Ultra Semiquantitative grade	Detected by Method 2		Detected by Method 3	
	trace+ for HIV+ only	trace+ for all	trace+ for HIV+ only	trace+ for all
Trace	0/2 (0%)	0/2 (0%)	0/2 (0%)	0/2 (0%)
Very Low	0/6 (0%)	0/6 (0%)	3/9 (33.3%)	1/5 (20.0%)
Low	4/12 (33.3%)	6/12 (50.0%)	3/9 (33.3%)	5/9 (55.6%)
Medium	13/15 (86.7%)	15/15 (100%)	9/11 (81.8%)	11/11 (100%)
High	20/21 (95.2%)	21/21 (100%)	10/10 (100%)	10/10 (100%)
All	37/56 (66.1%)	42/56 (75.0%)	22/37 (59.5%)	27/37 (73.0%)

Table 5. Semiquantitative results

Tongue swab Xpert Ultra Method 2

		Negative	Trace	Very low	Low	Medium	Total
	Negative	127	0	0	0	0	127
	Trace	2	0	0	0	0	2
Sputum Xpert Ultra*	Very low	6	0	0	0	0	6
	Low	6	3	0	3	0	12
	Medium	0	3	5	7	0	15
	High	0	1	4	14	2	21
	Total	141	7	9	24	2	183

*Highest semiquantitative grade from sputum Xpert Ultra; test repeated if the first result = trace or indeterminate

References

1.   Organization WH. Global Tuberculosis Report 20202020 2020.

2.   Cohen JF, Korevaar DA, Altman DG, Bruns DE, Gatsonis CA, Hooft L, et al. STARD 2015 guidelines for reporting diagnostic accuracy studies: explanation and elaboration. BMJ Open. 2016;6(11):e012799. doi: 10.1136/bmjopen-2016-012799.

3.   World Health Organization. The use of lateral flow urine lipoarabinomannan assay (LF-LAM) for the diagnosis and screening of active tuberculosis in people living with HIV: Policy guidance. Available at: http://www.who.int/tb/areas-of-work/laboratory/policy_statement_lam_web.pdf. Last accessed 4 Aug 2019.2015 14 Dec 2018.

4.   Denkinger CM, Kik SV, Cirillo DM, Casenghi M, Shinnick T, Weyer K, et al. Defining the Needs for Next Generation Assays for Tuberculosis. Journal of Infectious Diseases. 2015;211(suppl 2):S29-S38.

5.   Chakravorty S, Simmons AM, Rowneki M, Parmar H, Cao Y, Ryan J, et al. The New Xpert MTB/RIF Ultra: Improving Detection of Mycobacterium tuberculosis and Resistance to Rifampin in an Assay Suitable for Point-of-Care Testing. mBio. 2017;8(4).

6.   Boehme CC, Nabeta P, Hillemann D, Nicol MP, Shenai S, Krapp F, et al. Rapid Molecular Detection of Tuberculosis and Rifampin Resistance. New England Journal of Medicine. 2010;363(11):1005-15. doi: doi: 10.1056/NEJMoa0907847.

7.   Helb D, Jones M, Story E, Boehme C, Wallace E, Ho K, et al. Rapid Detection of Mycobacterium tuberculosis and Rifampin Resistance by Use of On-Demand, Near-Patient Technology. Journal of Clinical Microbiology. 2010;48(1):229-37.

8.   World Health Organization. WHO meeting report of a technical expert consultation: Non-inferiority analysis of Xpert MTB/RIF Ultra compared to Xpert MTB/RIF. Available at: https://www.who.int/tb/publications/2017/XpertUltra/en/. Last accessed 4 Aug 2019.2017.

9.   Valinetz ED, Cangelosi GA. A Look Inside: Oral Sampling for Detection of Non-Oral Infectious Diseases. Journal of Clinical Microbiology. 2021:JCM.02360-20. doi: 10.1128/jcm.02360-20.

10.   Deviaene M, Weigel KM, Wood RC, Luabeya AKK, Jones-Engel L, Hatherill M, et al. Sample adequacy controls for infectious disease diagnosis by oral swabbing. PloS one. 2020;15(10):e0241542.

11.   Luabeya AK, Wood RC, Shenje J, Filander E, Ontong C, Mabwe S, et al. Noninvasive Detection of Tuberculosis by Oral Swab Analysis. Journal of Clinical Microbiology. 2019;57(3):e01847-18.

12.   Wood RC, Luabeya AK, Weigel KM, Wilbur AK, Jones-Engel L, Hatherill M, et al. Detection of Mycobacterium tuberculosis DNA on the oral mucosa of tuberculosis patients. Sci Rep. 2015;5. doi: Article.

13.   Wood RC, Andama A, Hermansky G, Burkot S, Asege L, Job M, et al. Characterization of oral swab samples for diagnosis of pulmonary tuberculosis. PLOS ONE. 2021;16(5):e0251422. doi: 10.1371/journal.pone.0251422.

14.   Nicol MP, Wood RC, Workman L, Prins M, Whitman C, Ghebrekristos Y, et al. Microbiological diagnosis of pulmonary tuberculosis in children by oral swab polymerase chain reaction. Scientific Reports. 2019;9(1):10789.

15.   Mesman AW, Calderon R, Soto M, Coit J, Aliaga J, Mendoza M, et al. Mycobacterium tuberculosis detection from oral swabs with Xpert MTB/RIF ULTRA: a pilot study. BMC Res Notes. 2019;12(1):349-.

16.   Lima F, Santos AS, Oliveira RD, Silva CCR, Gon+ºalves CCM, Andrews JR, et al. Oral swab testing by Xpert-« MTB/RIF Ultra for mass tuberculosis screening in prisons. Journal of Clinical Tuberculosis and Other Mycobacterial Diseases. 2020;19:100148.

17.   Chakravorty S, Simmons AM, Rowneki M, Parmar H, Cao Y, Ryan J, et al. The New Xpert MTB/RIF Ultra: Improving Detection of Mycobacterium tuberculosis and Resistance to Rifampin in an Assay Suitable for Point-of-Care Testing. MBio. 2017;8(4):e00812-17. Epub 2017/08/31. doi: 10.1128/mBio.00812-17. PubMed PMID: 28851844; PubMed Central PMCID: PMCPMC5574709.

18.   Stop TB Partnership Global Laboratory Initiative. Mycobacteriology Laboratory Manual.   Available   at: http://www.stoptb.org/wg/gli/assets/documents/gli_mycobacteriology_lab_manual_web.pdf. Last accessed 4 Aug 2019.2014.