<u>Title:</u> Digitizing Non-Invasive Neuromodulation Trials: Scoping Review, Process Mapping, and Recommendations from a Delphi Panel

Authors: Andre R Brunoni^{1,2,3*}, Hamed Ekhtiari⁴, Andrea Antal⁵, Paradee Auvichayapat⁶, Chris Baeken^{7,8,9,10}, Isabela M. Benseñor¹¹, Marom Bikson¹², Paulo Boggio¹³, Barbara Borroni¹⁴, Filippo Brighina¹⁵, Jerome Brunelin^{16,17}, Sandra Carvalho¹⁸, Wolnei Caumo^{19,20,21,22}, Patrick Ciechanski²³, Leigh Charvet²⁴, Vincent P. Clark²⁵, Roi Cohen Kadosh²⁶, Maria Cotelli²⁷, Abhishek Datta²⁸, Zhi-De Deng²⁹, Rudi De Raedt³⁰, Dirk De Ridder³¹, Paul B. Fitzgerald³², Agnes Floel^{33,34}, Flavio Frohlich^{35,36,37,38,39,40}, Mark S. George^{41,42}, Peyman Ghobadi-Azbari^{43,44}, Stephan Goerigk^{45,46,47}, Roy H. Hamilton⁴⁸, Shapour J. Jaberzadeh⁴⁹, Kate Hoy³², Dawson J. Kidgell⁵⁰, Arash Khojasteh Zonozi^{43,51}, Adam Kirton⁵², Steven Laureys⁵³, Michal Lavidor⁵⁴, Kiwon Lee⁵⁵, Jorge Leite⁵⁶, Sarah H. Lisanby⁵⁷, Colleen Loo^{58,59}, Donel M. Martin^{58,59}, Carlo Miniussi⁶⁰, Marine Mondino^{15,16}, Katia Monte-Silva^{61,62}, Leon Morales-Quezada⁶³, Michael A. Nitsche^{64,65}, Alexandre H. Okano^{62,66,67}, Claudia S. Oliveira^{68,69}, Balder Onarheim⁷⁰, Kevin Pacheco-Barrios^{71,72}, Frank Padberg⁷³, Ester M Nakamura-Palacios⁷⁴, Ulrich Palm^{75,76}, Walter Paulus⁷⁷, Christian Plewnia⁷⁸, Alberto Priori⁷⁹, Tarek K. Rajji^{80,81,82}, Lais B. Razza⁸³, Erik M. Rehn⁸⁴, Giulio Ruffini⁸⁵, Klaus Schellhorn⁸⁶, Mehran Zare-Bidoky^{43,87}, Marcel Simis⁸⁸, Pawel Skorupinski⁸⁹, Paulo Suen⁸³, Aurore Thibaut⁹⁰, Leandro C. L. Valiengo³, Marie-Anne Vanderhasselt^{8,9}, Sven Vanneste⁹¹, Ganesan Venkatasubramanian⁹², Ines R. Violante⁹³, Anna Wexler⁹⁴, Adam J. Woods^{95,96,97}, Felipe Fregni⁷¹

 Department and Institute of Psychiatry, Faculdade de Medicina da Universidade de São Paulo, São Paulo, Brazil
 Department of Internal Medicine, Faculdade de Medicina da Universidade de São Paulo & Hospital Universitário, Universidade de São Paulo, São Paulo, Brazil

3. Laboratory of Neurosciences (LIM-27), Instituto Nacional de Biomarcadores em Neuropsiquiatria (INBioN), Service of Interdisciplinary Neuromodulation (SIN), Department and Institute of Psychiatry, Faculdade de Medicina da Universidade de São Paulo, São Paulo, Brazil

4. Laureate Institute for Brain Research (LIBR), Tulsa, OK, USA

5. Department of Neurology, University Medical Center Göttingen, Göttingen, Germany

6. Department of Physiology, Faculty of Medicine, Khon Kaen University, Khon Kaen, Thailand

7. Vrije Universiteit Brussel (VUB): Department of Psychiatry University Hospital (UZBrussel), Brussels, Belgium

8. Department of Head and Skin, Ghent University Hospital, Ghent University, Ghent, Belgium

9. Ghent Experimental Psychiatry (GHEP) Lab, Ghent, Belgium

10. Eindhoven University of Technology, Department of Electrical Engineering, the Netherlands

11. Center for Clinical and Epidemiological Research, University of São Paulo, São Paulo, Brazil

12. The Department of Biomedical Engineering, The City College of New York, The City University of New York, New York, USA

13. Social and Cognitive Neuroscience Laboratory, Center for Biological Science and Health, Mackenzie Presbyterian University, São Paulo, Brazil

14. Centre for Neurodegenerative Disorders, Neurology Unit, Department of Clinical and Experimental Sciences, University of Brescia, Italy

15. Department of Biomedicine, Neuroscience and Advanced Diagnostics (Bi.N.D.), University of Palermo, Palermo, Italy

16. Centre Hospitalier le Vinatier, Bron, France

17. INSERM U1028, CNRS UMR 5292, PSYR2 Team, Centre de recherche en Neurosciences de Lyon (CRNL), Université Lyon 1, Lyon, France

1

18. Translational Neuropsychology Lab, Department of Education and Psychology and William James Center for Research (WJCR), University of Aveiro, Campus Universitário de Santiago, Aveiro, Portugal

19. Post-Graduate Program in Medical Sciences, School of Medicine, Universidade Federal do Rio Grande do Sul (UFRGS), Brazil

20. Laboratory of Pain and Neuromodulation at Hospital de Clínicas de Porto Alegre (HCPA), Porto Alegre, Brazil

21. Pain and Palliative Care Service at HCPA, Brazil

22. Department of Surgery, School of Medicine, UFRGS, Brazil

23. Faculty of Medicine and Dentistry, University of Alberta, 1-002 Katz Group Centre for Pharmacy and Health Research, Edmonton, Alberta, Canada

24. Department of Neurology, NYU Grossman School of Medicine, New York, NY, USA

25. Psychology Clinical Neuroscience Center, Department of Psychology, The University of New Mexico, Albuquerque, NM, USA

26. School of Psychology, Faculty of Health and Medical Sciences, University of Surrey, Guildford, UK

27. Neuropsychology Unit, IRCCS Istituto Centro San Giovanni di Dio Fatebenefratelli, Brescia, Italy

28. Research and Development, Soterix Medical Inc., New York, USA

29. Noninvasive Neuromodulation Unit, Experimental Therapeutics & Pathophysiology Branch, National Institute of Mental Health, Bethesda, MD, USA

30. Department of Experimental Clinical and Health Psychology, Ghent University, Belgium

31. Section of Neurosurgery, Department of Surgical Sciences, Dunedin School of Medicine, University of Otago, Dunedin, New Zealand

32. Epworth Centre for Innovation in Mental Health, Epworth Healthcare and Monash University Department of Psychiatry, Camberwell, Victoria, Australia

33. Department of Neurology, University Medicine Greifswald, Greifswald, Germany

34. German Center for Neurodegenerative Diseases (DZNE), Rostock/Greifswald, Germany

35. Department of Psychiatry, University of North Carolina, Chapel Hill, NC, USA

36. Carolina Center for Neurostimulation, University of North Carolina, Chapel Hill, NC, USA

37. Neuroscience Center, University of North Carolina, Chapel Hill, NC, USA

38. Department of Cell Biology and Physiology, University of North Carolina, Chapel Hill, NC, USA

39. Department of Biomedical Engineering, University of North Carolina, Chapel Hill, NC, USA

40. Department of Neurology, University of North Carolina, Chapel Hill, NC, USA

41. Department of Psychiatry, Medical University of South Carolina, Charleston, SC, USA

42. Ralph H. Johnson VA Medical Center, Charleston, SC, USA

43. Iranian National Center for Addiction Studies, Tehran University of Medical Sciences, Tehran, Iran

44. Department of Biomedical Engineering, Shahed University, Tehran, Iran

45. Department of Psychiatry and Psychotherapy, LMU Hospital, Munich, Germany

46. Department of Psychological Methodology and Assessment, LMU, Munich, Germany

47. Hochschule Fresenius, University of Applied Sciences, Munich, Germany

48. Department of Neurology, University of Pennsylvania, Philadelphia, PA, USA

49. Department of Physiotherapy, Faculty of Medicine, Nursing and Health Sciences, Monash University, Melbourne, Australia

50. Department of Physiotherapy, School of Primary and Allied Health Care, Faculty of Medicine, Nursing and Health Science, Monash University, Melbourne, Australia

51. Faculty of Medicine, Mashhad University of Medical Sciences, Mashhad, Iran

52. Department of Clinical Neurosciences and Department of Pediatrics, University of Calgary, Calgary, Alberta, Canada

53. Coma Science Group, GIGA-Consciousness, GIGA Institute, University of Liège, Liege, Belgium

54. Bar Ilan University, Department of Psychology, and the Gonda Brain Research Center, Israel

55. Ybrain Corporation, Gyeonggi-do, Republic of Korea

56. INPP, Portucalense University, Porto, Portugal

57. Noninvasive Neuromodulation Unit, Experimental Therapeutics & Pathophysiology Branch, National Institute of Mental Health, Bethesda, MD, USA

58. School of Psychiatry, University of New South Wales, Sydney, NSW, Australia

59. Black Dog Institute, Sydney, NSW, Australia

60. Center for Mind/Brain Sciences - CIMeC, University of Trento, Rovereto, Italy

61. Applied Neuroscience Laboratory, Department of Physical Therapy, Universidade Federal de Pernambuco, UFPE, Recife, PE, Brazil

62. NAPeN Network (Núcleo de Assistência e Pesquisa em Neuromodulação), Brazil

63. Department of Physical Medicine and Rehabilitation, Spaulding Rehabilitation Hospital, Harvard Medical School, Boston, MA 02215, USA

64. Department of Psychology and Neurosciences, Leibniz Research Centre for Working Environment and Human Factors, Dortmund, Germany

65. Department of Neurology, University Medical Hospital Bergmannsheil, Bochum, Germany

66. Center for Mathematics, Computation, and Cognition, Universidade Federal do ABC, São Bernardo do Campo, Brazil

67. Brazilian Institute of Neuroscience and Neurotechnology (BRAINN/ CEPID-FAPESP), University of Campinas, Campinas, São Paulo, Brazil

68. Master's and Doctoral Program in Health Sciences, Faculty of Medical Sciences, Santa Casa de São Paulo, São Paulo, Brazil

69. Master's and Doctoral Program in Human Movement and Rehabilitation, Evangelical University of Goiás, Anápolis, Brazil

70. PlatoScience ApS, Copenhagen, Denmark

71. Neuromodulation Center and Center for Clinical Research Learning, Spaulding Rehabilitation Hospital and Massachusetts General Hospital, Harvard Medical School, Boston, MA, USA

72. Universidad San Ignacio de Loyola, Vicerrectorado de Investigación, Unidad de Investigación para la Generación y Síntesis de Evidencias en Salud, Lima, Peru

73. Department of Psychiatry and Psychotherapy, University Hospital, LMU Munich, Munich, Germany

74. Laboratory of Cognitive Sciences and Neuropsychopharmacology, Program of Post-Graduation in Physiological Sciences, Health Sciences Center, Federal University of Espirito Santo, Vitória, ES, Brazil

75. Department of Psychiatry and Psychotherapy, Klinikum der Universität München, Munich, Germany

76. Medical Park Chiemseeblick, Rasthausstr. 25, 83233 Bernau-Felden, Germany

77. Department of Neurology. Ludwig Maximilians University Munich, Klinikum Großhadern, Marchioninistr, München, Germany

78. Department of Psychiatry and Psychotherapy, Tübingen Center for Mental Health (TüCMH), Neurophysiology and Interventional Neuropsychiatry, University of Tübingen, Tübingen, Germany

79. Aldo Ravelli Research Center for Neurotechnology and Experimental Neurotherapeutics, Department of Health Sciences, University of Milan, Milan, Italy

80. Centre for Addiction and Mental Health, Toronto, Canada

81. Temerty Faculty of Medicine, University of Toronto, Toronto, Canada

82. Toronto Dementia Research Alliance, Toronto, Canada

83. Service of Interdisciplinary Neuromodulation (SIN), Department and Institute of Psychiatry, Faculdade de Medicina da Universidade de São Paulo, São Paulo, Brazil

84. Flow Neuroscience, Malmö, Skane Lan, Sweden

85. Neuroelectrics Corporation, Cambridge, Massachusetts, USA

86. neuroConn GmbH, Ilmenau, Germany

87. School of Medicine, Shahid-Sadoughi University of Medical Sciences, Yazd, Iran

88. Physical and Rehabilitation Medicine Institute, General Hospital, Medical School of the University of Sao Paulo, São Paulo, Brazil

89. Neuroelectrics Corporation, Cambridge, Massachusetts, USA

90. Coma Science Group, GIGA-Consciousness & Centre du Cerveau, University and University Hospital of Liège, Liège, Belgium

91. Lab for Clinical & Integrative Neuroscience, Trinity College of Neuroscience, Trinity College Dublin, Ireland

92. Department of Psychiatry, National Institute of Mental Health and Neuro Sciences (NIMHANS), Bengaluru, India

93. School of Psychology, Faculty of Health and Medical Sciences, University of Surrey, Guildford, UK

94. Department of Medical Ethics and Health Policy, University of Pennsylvania, Philadelphia, PA, USA

95. Center for Cognitive Aging and Memory, McKnight Brain Institute, University of Florida, Gainesville, FL, USA

96. Department of Clinical and Health Psychology, University of Florida, Gainesville, FL, USA

97. Department of Neuroscience, University of Florida, Gainesville, FL, USA

*Corresponding Author:

André Russowsky Brunoni, MD, Ph.D.,


Service of Interdisciplinary Neuromodulation, R Dr Ovídio Pires de Campos 785, 20 andar Ala Sul, Instituto de Psiquiatria, CEP 05403-000, São Paulo, Brazil email: brunoni@usp.br

Abstract

Although relatively costly and non-scalable, non-invasive neuromodulation interventions are treatment alternatives for neuropsychiatric disorders. The recent developments of highly-deployable transcranial electric stimulation (tES) systems, combined with mobile-Health technologies, could be incorporated in digital trials to overcome methodological barriers and increase equity of access. We convened 61 highly-productive specialists and contacted 8 tES companies to assess 71 issues related to tES digitalization readiness, and processes, barriers, advantages, and opportunities for implementing tES digital trials. Delphi-based recommendations (>60% agreement) were provided. Device appraisal showed moderate digitalization readiness, with high safety and the possibility of trial implementation, but low connectivity. Panelists recognized the potential of tES for scalability, generalizability, and leverage of digital trials processes; although they reached no consensus about aspects regarding methodological biases. We further propose and discuss a conceptual framework for exploiting shared aspects between mobile-Health tES technologies with digital trials methodology to drive future efforts for digitizing tES trials.

Keywords: Mobile Health; non-invasive neuromodulation; digital health, systematic review; Delphi panel

Graphical Abstract. Consensus Roadmap.



(A) Recruitment process. The study procedure started with defining the components of the research problem by the core research team. After defining the problems, two different sets of participants (the steering committee (SC) including key leaders of the field identified by the core team and the expert panel (EP) as a more diverse group of experts identified based on the number of publications based on a systematic review) were identified and were invited to participate in a Delphi study. The study facilitators (first and last authors) led the communications with the SC to design the initial questionnaire through an iterative approach. (B) *Evidence synthesis:* To collect the available evidence, companies producing portable tES (ptES) devices were contacted, based on the companies suggested by the SC and EP to provide details about the available devices. For mapping methodological processes of digitizing tES trials, two distinct strategies were performed and embedded into the questionnaire, namely SIPOC (Suppliers, Inputs, Process, Outputs, and Customer) and SWOT (Strengths, Weaknesses, Opportunities, and Threats) assessment were performed and embedded into the questionnaire was validated and finalized via collecting and summarizing opinions. Afterward, the SC and EP responded to the final questionnaire, and results were analyzed providing a list of recommendations for running tES digital trials based on a pre-registered consensus threshold.

1. Introduction

Transcranial electric stimulation (tES) is a non-invasive neuromodulation intervention that uses electric currents applied over the scalp to modify cortical activity and treat neuropsychiatric disorders and has high safety and tolerability (Fregni et al., 2021). Notwithstanding, due to its low-to-moderate efficacy for several conditions, the consensus of its readiness for clinical use across indications varies (Ekhtiari et al., 2019; Fregni et al., 2021), and regulatory approvals across regions are mixed (Bikson et al., 2018b; Fregni et al., 2015), warranting larger-scale clinical trials (Brunoni et al., 2012). However, these trials are hampered by the need for daily visits to the research center to deliver the necessary number of tES sessions, limiting recruitment, harming adherence, increasing costs (5), and restricting diversity (Bikson et al., 2018a; Charvet et al., 2020).

Unlike other non-invasive neuromodulation modalities, tES devices, by virtue of being affordable and battery-powered (Woods et al., 2016), are portable, intervening an appealing brain stimulation modality for patients who do not tolerate pharmacotherapy (Brunoni et al., 2019) or have difficulty attending treatment at a clinical center several companies have been designing highly-deployable tES devices that could be used to address issues of scale, access, and patients' burden in the context of digital trials - i.e., trials that leverage aspects such as recruitment, assessment, and data analyses through the implementation of digital technologies (Inan et al., 2020). These approaches could be further integrated with mobile Health devices, apps, and wearables, allowing for several new implementations, such as simultaneous combination with cognitive and psychological interventions, ecological momentary assessment of behaviors, passive data collection, and digital phenotyping (Insel, 2018; Torous et al., 2017). However, since protocols and standards for digital trials using mobile tES are still emerging, the challenges and opportunities of their implementation processes have not yet been systematically examined. Moreover, issues on rigor and reproducibility - for instance, best practices to perform randomization, allocation concealment, and sham stimulation - and generalizability - how to fully explore their potential for scalability while ensuring adherence and representativeness - have only been discussed in non-digital contexts (Bikson et al., 2018a; Charvet et al., 2020, 2015). Finally, implementation challenges are different in low-/middle- income countries, which present lower digital literacy, fluency in non-native languages, and wireless connectivity (Silva-Filho et al., 2022); conversely, scalable mobile Health interventions have even higher impact potential in such countries (Torous et al., 2021).

Considering these challenges, we systematically identified non-invasive neuromodulation specialists to elaborate and discuss issues related to the extent, processes, and methodological characteristics of digitizing non-invasive neuromodulation trials. These findings, supplemented by a systematic scoping review of tES clinical research articles and an assessment of the digitalization degree of commercially available tES devices, provided a key summary of Delphi-based recommendations for enhancing the implementation of digital tES trials.

2. Methods

Our protocol was pre-registered at the Open Science Framework (OSF) (Razza et al., 2021) and is depicted in Fig. 1. Minor deviations occurred (Sup. Material - Appendix 1).




Fig. 1. Study workflow. First, the steering committee (SC) was formed, including prominent researchers in the field. Then, supplemented by the results of a systematic review conducted on tES trials, the SC developed the questionnaire, which was sent to all participants of the study (SC and expert panel (EP)) to answer it. Simultaneously, companies producing tES devices were contacted, based on the companies known by the SC and EP, to provide details about the available devices. Finally, the SC analyzed the data received from the questionnaire and all participants took part in reporting the results.

2.1. Specialist panel

We used a Delphi technique, in which comments and feedback are iteratively discussed (Hsu and Sandford, 2019), for addressing challenges and proposing recommendations for digitizing tES trials. Following recent papers (Ekhtiari et al., 2022a, 2022b), we initially convened a steering committee (SC) group, formed based on the collaborative network of the leading authors, to develop structured questionnaires with items using Five-point Likert scales or open questions (Supplementary Material - Appendix 2). The SC also provided qualitative feedback on several topics that were analyzed by the leading authors and qualitatively described here. Afterward, the SC interacted with a larger expert panel (EP) to

rate each item. The EP members were selected among the most productive authors in the field, identified through a systematic review of recent tES clinical trials in 10 years (Sup. Material - Appendix 3). Several interactions were performed between the EP and SC until a final manuscript version was assembled. The consensus was achieved by a >60% agreement of all panelists. Electronic questionnaires were used in all steps of this process. All members of the SC and the EP consented to have their names listed and identified in the manuscript.

2.2. Systematic scoping review

A systematic scoping review (Levac et al., 2010) was performed to identify tES reviews, consensus papers, and guidelines to select characteristics for composing the questionnaires used in the rating phase (Sup. Material - Appendix 4).

2.3. tES digitalization readiness

Companies producing tES devices were identified through specialist referrals and web search and surveyed using structured questionnaires to assess their digitalization readiness, according to connectivity, readiness for digital trials, parameter space flexibility, ecological footprint, front-end interface, and data security (Sup. Material - Appendix 5).

2.4. Process mapping and methodological assessment

We used SWOT (Strengths, Weaknesses, Opportunities, and Threats) and SIPOC (Suppliers, Inputs, Process, Outputs, Customers) approaches to respectively identify external and internal negative and positive aspects for digitizing tES trials and map and compare processes related to standard and digital tES trials. The methodological assessment was based on perceived challenges and advantages, identified through questionnaires, of conducting such trials (Sup. Material - Appendix 2).

2.5. Role of the funding source

This work received no specific funding from any source.

3. Results

3.1. Specialist panel

For the SC, 34 authors were invited and all agreed to participate. For the EP, out of 43 authors who were identified, 14 did not reply to our contacts, and 2 declined to participate. Finally, 27 participants constituted the EP (Sup. Material - Appendix 6). Most panelists were men (70%), experienced (78% with > 10 years of experience in the field), and between 40 to 49 years old (44% and 33% of the SC and the EP). They resided in the US (SC n=11, EP n=3), Brazil (SC n=6, EP n=5), Germany (SC n=5, EP n=4), and 13 other countries (Sup. Material - Appendix 7). Only 15% and 18% of the SC and EP members, respectively, were not conducting at least one tES trial with digital features; most were principal investigators (83%) of such trials.

3.2. Systematic scoping review

Our initial search yielded 443 references, and 34 articles met the inclusion criteria of our scoping review, including 9 recommendation articles (Bikson et al., 2018a; Deer et al., 2014; Fregni et al., 2015; Fried et

al., 2021; Kim et al., 2020; Sandars et al., 2016; Sierawska et al., 2019; Thibaut et al., 2017; Zhang et al., 2019), 10 guidelines (Antal et al., 2017; Bikson et al., 2020; Charvet et al., 2020, 2015; Cruccu et al., 2016; Fregni et al., 2021; Gillick et al., 2018; Lefaucheur et al., 2017; Legatt et al., 2016; Parikh et al., 2016), 10 critical reviews (ALHarbi et al., 2017; Cappon et al., 2016; Godeiro et al., 2021; Lucchiari et al., 2018; Maatoug et al., 2021; McClintock et al., 2019; Sanches et al., 2021; Santos et al., 2021; Shiozawa et al., 2017; Workman et al., 2020), and 5 expert consensus articles (Baptista et al., 2019; Buch et al., 2017; Ekhtiari et al., 2019; Grimaldi et al., 2014; Martelletti et al., 2013), which were used for elaborating the study questionnaires (Sup. Material - Appendix 8).

3.3. tES digitalization readiness

Eight of 13 companies contacted provided feedback. Digitization readiness varied according to their wireless connectivity, readiness for digital trials, the flexibility of parameter space, ecological footprint, front-end interface, and data security. Markedly, current systems have limited wireless connectivity, which is a barrier for device-to-device communication with wearables and third-party apps that could enhance portability potential (e.g., apps collecting biological data, and mobile mental health apps). Conversely, most systems currently present good data security protocols (reported HIPAA or GDPR compliance), the flexibility of tES parameter space, and readiness for digital trials (Fig. 2).




Fig. 2. Digitalization readiness of tES devices. Based on the feedback of 8 out of 13 major transcranial electric current stimulation (tES) companies, we evaluated the readiness of these systems for digital trials, considering their connectivity (capability of communicating with other devices and the Web, due to the presence of Bluetooth, Wi-Fi, 3G/4G/5G, and communication with third-party apps), methodological aspects (randomization, sham, blinding, built-in data collection, optional data collection, and optional research dashboard), parameter space flexibility (current intensity, session duration, number of sessions, electrode positioning), ecological footprint (rechargeability and replaceability of batteries, recyclability and reusability of sponges, and recyclability of devices), front-end interface (smartphone app, touch screen, device itself, no interaction), and data security (compliance to laws such as GDPR, mention of encryption and anonymization procedures, and option for not collecting sensitive data). The questionnaires and rating systems are described in the Sup. Material-Appendix 5.

3.4. SWOT assessment

The identified characteristics and quantitative agreement rating composing the SWOT assessment are displayed in Fig. 3. Qualitative aspects are briefly discussed here and detailed in the Sup. Material - Appendix 9.




Fig. 3. SWOT (Strengths, Weaknesses, Opportunities, and Threats) assessment for digital tES trials. This figure depicts the ratings of 55 participants (24 from the steering committee and 31 from the expert panel) for the ptES clinical trials strengths and opportunities (A) and threats and weaknesses (B). Each item was rated from strongly disagree to strongly agree. In ratings for the potential strengths and opportunities of tES clinical trials (A), all of the items have reached the 50% threshold of agreement (rated as either agree or strongly agree by more than 50% of the respondents). These items have also reached a more stringent threshold of 60%. However, in ratings for the potential threats and weaknesses of ptES clinical trials (B), two of the items (2.4. Skin burns as a relevant side effect, and 2.6. Higher rates of serious adverse events) did not reach the 50% threshold. Items are represented by their summary in the figure. Full text of the items is provided in Supplementary Tables 1 and 2. tDCS=transcranial direct current stimulation. ptES=portable transcranial electrical stimulation. DIY=do-it-yourself.

Regarding strengths, the panelists agreed on four features: (a) high safety, considering previous evidence from non-digital trials and studies in humans (Antal et al., 2017; Aparício et al., 2016; Bikson et al., 2016; Moffa et al., 2017); (b) feasibility of self-application, owing to recent developments of devices in which electrode placement is fixed, methods for easy strap positioning, and friendly end-user interface of mobile tES device (Charvet et al., 2020); (c) being a non-pharmacological intervention; and (d) affordability, as tES devices are simple to be built in terms of electric engineering (Woods et al., 2016), costs of high-end features (e.g., microprocessors, Bluetooth and wireless connectivity, miniaturization) are decreasing over time, and self-application saves human resources.

Regarding weaknesses, panelists agreed on two aspects: (a) difficulties in remote supervision, raising concerns regarding patients themselves manipulating tES devices, which could lead to misuse, diversion of the device, or its use outside of medical contexts, further impacting on the reproducibility of findings; (b) and difficulties in obtaining accurate placement of electrodes, as deviations in electrode positioning and orientation might affect outcomes (Woods et al., 2016). Therefore, companies should develop and test new methods for assuring the correct placement of electrodes (Fried et al., 2021). Other potential weaknesses did not reach consensus, such as concerns related to bioethics, particularly regarding equity

to access; increased (compared to on-site tES trials) risks of common and serious adverse events (Lefaucheur et al., 2017); and relatively low evidence of clinical efficacy for most conditions (Fregni et al., 2021).

Regarding opportunities, six aspects reached agreement: (a) scalability, as, compared to on-site tES trials that need physical space, staff to apply sessions and commute of participants, digital trials using mobile tES devices do not have such constraints, allowing research assistants to monitor several participants at once, at any distance from the study centers; (b) telemedicine, which has been widely adopted during the COVID-19 pandemic, facilitating its adoption in digital trials; (c) employment of combined mobile Health technologies, permitting digital phenotyping (Torous et al., 2017) and combination with digital interventions when using paired wearables and smartphone applications; (d) automatization of procedures (see SIPOC below); (e) 5G / Internet of Things, which can boost connectivity and data processing, leveraging data collection (Torous et al., 2017) and eventually allowing the development of mobile closed-loop tES systems (Sanches et al., 2021); and (f) use of design thinking approaches, i.e., customizing mobile tES devices around the patients' perspectives (Polhemus et al., 2020), for instance, to accommodate those with physical or cognitive impairments.

Finally, two threats reached a consensus: (a) recreation and do-it-yourself misuse, which could lead to unexpected adverse events and safety issues (Sierawska et al., 2019); and (b) regulatory status, as medical devices require formal regulatory approval in the US (Darrow et al., 2021) and Europe (Antich-Isern et al., 2021), although some tES devices are marketed as wellness devices, have regulatory device exemptions (Bikson et al., 2018b), or can be approved by similarity (Bikson et al., 2018b). Further, mobile tES devices could have additional regulations, if framed as mobile Health systems (Onodera and Sengoku, 2018). Additionally, two potential threats were identified by most of the panel, but did not reach the 60% consensus threshold: (a) risks related to hacking and cyber-security, as observed in mobile Health devices (Aljedaani and Babar, 2021), and (b) risks related to confidentiality and anonymity.

3.5. SIPOC

We identified 4 main processes (recruitment, pre-screening, screening, and participation) in which digitization and automatization procedures can leverage mobile tES trials (Table 1). The panelists noted that trials might not be purely digital or analog, and different degrees of digital features can occur at each process. For instance, participants can be recruited through social media but screened onsite. Moreover, digital processes can provide enhanced metrics to adjust processes, recruit faster, and follow participants for longer periods. Finally, digitization processes provide scalability due to the use of digital assessments and telemedicine. (Table 1).

Digital features	Suppliers	Inputs	Outputs	Clients		
Process: Recruitment						
Absent	Traditional Media	Telephone call or email	Pool of volunteers constrained due to geographical barriers	Pre-screening / screening onsite teams		

 Table 1. Comparison of SIPOC processes of trials in which digital features are present and absent.

Present	Targeted Social Media and Google Ads, public segmentation	Electronic forms, ChatBots	Larger pool of volunteers	Online pre-screening / screening teams			
Process: Pre-screening							
Absent	Volunteers	Onsite checklist of eligibility criteria	Potential Participants	Onsite screening team			
Present	Volunteers	Online assessments, AI techniques can increase the likelihood of inclusion	Potential Participants	Online screening team			
Process: Screening							
Absent	Potential participants	Onsite clinical interview; written consent	Participants	Onsite clinical team			
Present	Potential participants	Clinical interviews aided by digital assessments; digital consent	Participants	Online clinical team			
Process: Participation in the trial							
Absent	Participants	Staff delivers tES sessions and assessments	Completers and patients who dropped out	Trial complete; possibly follow-up studies			
Present	Participants	tES devices shipped to participants; online videos and telesupport to guide self-applied tES; digital assessments via online interviews and mHealth technologies	Completers and patients who dropped out	Trial complete; possibly follow-up studies			

The table illustrates how processes of clinical trials can be leveraged using digital approaches. In (a) recruitment, volunteers, unconstrained by geographical barriers, can fill out electronic forms and be automatically excluded according to eligibility criteria; (b) pre-screening, potential participants can be scheduled and contacted automatically for online screening (a step that can be enriched by using machine learning algorithms); (c) screening, enrolled participants can provide digital consent, have transcranial electric stimulation (tES) devices shipped to their homes, and being instructed how to use them via videos and/or augmented reality techniques; and (d) participation in the trial, interaction with the staff can be mediated by psychiatric chatbots for filtering between simple questions, reporting of adverse events and need of medical care.

3.6. Methodological aspects

The panelists examined 24 methodological aspects of digital trials using mobile tES devices, reaching consensus in 12 of them (Fig. 4). They are briefly discussed here and detailed in the Sup. Material - Appendix 10.




Fig. 4. Ratings for advantages and challenges of conducting tES digital trials. This figure depicts the ratings of 55 raters (24 from the steering committee and 31 from the expert panel) for methodological aspects of tES. Each item was rated from strongly disagree to strongly agree. In ratings for the advantages of transcranial electric stimulation (tES) digital trials from a methodological perspective, only one item (1.8. Greater internal validity) did not reach the agreement threshold of 50% (rated as either agree or strongly agree by more than 50% of the respondents). All of the other items in this category also reached a more stringent threshold of 60%. However, in ratings for challenges of tES digital trials from a methodological perspective, 9 items (2.3. Increased risk for suicidality, 2.4. Increased risk for the manic switch, 2.5. Increased risk for cognitive effects, 2.6. Increased risk for SAEs, 2.7. ptES will substitute on-site, 2.8. More dropouts, 2.9. More selection biases, 2.10. More randomization biases, and 2.13. More assessment biases) did not reach the 50% threshold of agreement. With a more stringent threshold of 60%, two additional items (2.11. More blinding biases and 2.12. More sham biases) dropped out from the agreement. Items are represented by their summary in the figure. Full texts of the items are provided in Supplementary Table 3. ptES=portable transcranial electrical stimulation. SAE=serious adverse event.

Of the 12 aspects that reached consensus, 10 were perceived as advantages of digital trials, which included (a) the adoption of different study designs, including (b) adaptive designs, as adaptation rules can be performed automatically and remotely whether data are collected and analyzed by mobile tES systems. Panelists also considered that (c) tES devices are already sufficiently developed to be used

remotely, which allows for (d) longer follow-up periods and (e) higher recruitment rates, being (f) faster and more efficient than on-site trials. Also, (g) greater external validity compared to on-site trials were perceived. Finally, other advantages were the potential for (h) collecting more data than on-site trials, (i) combination with other therapies and (j) scalability. The 2 disadvantages/challenges that reached consensus were: (a) the necessity of validating new tES parameters, methodologies, and indications first in on-site studies, and (b) the need of developing better remote assessment methods, such as behavioral clinical scales properly designed and validated to be employed remotely.

The items that did not reach consensus mostly pertain to internal validity issues. Interestingly, a significant portion of panelists was undecided on these issues (Fig. 5). Of note, interesting remarks (detailed in Sup. Material - Appendix 10) were made for: (a) randomization-allocation procedures that are done using either specific devices or apps/software, but can be vulnerable to contamination biases due to hacking and exposure of the randomization list; (b) study blinding, as blinding breaking can occur if devices are manipulated; (c) sham stimulation, which can also be revealed due to device manipulation.

3.7. Recommendations

The panelists recommended that teams performing tES digital trials should have members specialized in (a) digital marketing strategies, to enhance online recruitment; (b) data science and visualization, for data collection and analysis; (c) front-end interfaces, to enhance user experience; (d) back-end programming; (e) issues related to data security, integrity, anonymity, and replicability. Also, they suggested that (f) a team member should be always ("24/7") available (Fig. 5).




Fig. 5. Ratings of the tES clinical trials team features and general recommendations on digitizing tES trials. This figure depicts the ratings of 55 participants (24 from the steering committee and 31 from the expert panel) for the ptES clinical trials team features (A) and general recommendations for digitizing ptES clinical trials (B). Each item was rated from strongly disagree to strongly agree. In ratings for the ptES clinical trials team features (A), all of the items have reached the 50% threshold of agreement (rated as either agree or strongly agree by more than 50% of the respondents). These items have also reached a more stringent threshold of 60%. Similarly, in ratings for the general recommendations for digitizing ptES clinical trials (B), all of the items have reached the 50% threshold of agreement. However, with a more stringent threshold of 60%, one item (8. Develop methods to assess AEs) dropped out from the agreement. Items are represented by their summary in the figure. Full text of the items is provided in Supplementary Tables 4 and 5. ptES=portable transcranial electrical stimulation. AE=adverse event. SAE=serious adverse event.

Regarding further research, most panelists recommended that better methods for (a) randomization/allocation, and (b) sham should be explored. Also, further research was recommended to develop or refine methods to enhance (c) adherence and (d) external validity of the trials. Also, more research should be devoted to aspects such as (e) combination of interventions, (f) biobehavioral data collection, (g) enhanced data security, and methods to assess (h) serious and (i) unexpected adverse events (details in Sup. Material - Appendix 11).

4. Discussion

By convening a diverse group of 61 worldwide specialists in the field of non-invasive neuromodulation, we performed the first systematic assessment and Delphi-based validation of perceived challenges, opportunities, methodological issues, and recommendations on digitizing non-invasive neuromodulation trials. We used several strategies to organize and describe these assessments, such as processing mapping strategies, a systematic scoping review of the literature, iterative rating and validation of structured questionnaires by specialists, and assessment of the digital readiness of commercial tES devices. Taken together, our findings show that performing digital trials using mobile tES devices has complementary advantages and can overcome major on-site tES trial challenges, namely the intensive treatment schedules (Thibaut et al., 2017), transportation costs, accessibility, and scalability (Charvet et al., 2020). By performing trials remotely, dislocation burdens are decreased, as well as the need for space at the research center, and of trained staff for delivering tES sessions, aspects that increase the trial duration (Brunoni et al., 2012; Parikh et al., 2016, p. 4). Additionally, tES devices are highly scalable, as a single team member can monitor several people at once, provided that tES devices are easy to manipulate, handle, and can be self-delivered. Such scalability gains could be leveraged in faster trials with larger sample sizes, longer follow-up periods, and employing digital recruitment strategies. Considering our results, we propose and further discuss a conceptual framework for digitizing neuromodulation, combining concepts of digital clinical trials with mobile tES (Fig. 6).




Fig. 6. A conceptual framework for digitizing neuromodulation. As depicted on the left side of the figure, transcranial electric stimulation devices had been relatively simple, essentially using batteries connected to electrodes to deliver constant currents,

and contain few (micro) electronic components. Although portable and safe, they had not been specifically designed for use outside hospital or academic center settings. New and future generations of tES will be mobile Health tES systems using digital technologies for improving health outcomes. They are and will be smaller and lighter than previous generations, possessing wireless connectivity. Such devices are already used at home and are self-delivered, usually with some degree of remote supervision. Their use will be supported by proprietary or third-party apps and wearables. Resulting together with the aforementioned concept as digitizing neuromodulation, the right side of the figure shows digital trials as clinical trials that use digital features to enhance recruitment, assessment, and data analysis and could unleash the full potential of tES regarding scalability and equity of access. There are many similarities between the assumptions of digital trials and the capabilities of mtES, which are discussed in this work. EMA=Ecological momentary assessment. mtES=mobile transcranial electrical stimulation.

4.1. Mobile tES devices

To the best of our knowledge, the term "mobile tES" has not been used yet to describe the combination of highly-deployable neuromodulation devices paired with other wearables or apps. This terminology frames tES in the context of mobile Health devices (Onodera and Sengoku, 2018), encouraging the exploitation of contact points between these two growing fields. However, even though several issues for deployable and remote use have already been addressed (e.g., decreasing prices, rechargeable batteries, tailored sponges, sham stimulation, easiness of electrode positioning, and programming session stimulation parameters), our findings showed that no commercial tES devices have been fully digitized yet, presenting different degrees of online, wired, or wireless connectivity. Also, especially for offline devices, methods for restricting the number of sessions allowed per day were not identified. Additionally, most systems neither collect active or passive data, nor present friendly-user interfaces.

The panelists agreed on several opportunities; however, most are distant from immediate implementation. For instance, device-to-device communication ("Internet of Things") would need pairing with third-party apps or wearables, which is not yet available. This limits other perceived opportunities such as digital phenotyping, combination with psychological or cognitive app interventions, and seamless automatization with other platforms and digital processes.

4.2. Methodological challenges, advantages, and processes of digital trials

The impact of digitizing neuromodulation trials on external validity seems mostly positive, considering that subjects who would not be enrolled in on-site trials are reached. On-site trial samples are likely to be composed of those with free time and/or living near the clinical center to receive the sessions. Notwithstanding, it is still possible that those younger, richer, more educated, with higher digital literacy, and living in urban areas are preferentially enrolled in digital trials. Also, if recruitment strategies are performed solely using social media, the trial results would have restricted generalizability for people that do not use such media. This could be overcome by using segmented digital marketing campaigns. Likewise, attrition rates might not necessarily be lower in digital trials - although not needing to return daily to the clinical center could decrease the burden and minimize dropouts, samples from digital trials might face less engagement and more difficulties in self-delivering the sessions. The lack of daily contact with the clinical staff could also decrease motivation and increase dropouts. In addition, direct social contact, social support, and social connectedness outside the digital environment can influence attrition rates of clinical trials. Therefore, telemonitoring and proper interaction with participants, managing their expectations and credibility of the team, and reinforcing the need to abide by the study protocol, could avoid dropouts.

Panelists also emphasized that methodological issues that have not been completely addressed in on-site trials can be magnified in digital trials (Charvet et al., 2015; Fried et al., 2021). For instance, if the process of randomization - allocation concealment - is hacked from the server and publicly exposed, the entire trial can be lost (or, at least, suspended until a new list is generated). Moreover, blinding and sham stimulation issues are not completely resolved issues in on-site trials (Fonteneau et al., 2019), and biases arising from these steps are more likely to occur (e.g., sham stimulation can be unconcealed by measuring the voltage between electrodes (Woods et al., 2016)) and harm the entire study (for example, by exposing methodological vulnerabilities on the Web). Additionally, issues that would be minor in on-site trials might be more relevant in online trials. For instance, stimulation sessions can be performed in on-site trials appropriately and guarantee adherence (Woods et al., 2016), but, in digital trials, some degree of remote monitoring would be necessary for ensuring these aspects.

Finally, there are unique new challenges for digitizing neuromodulation. Even though cyber-hacking is not usually discussed in the environment of clinical trials, government and big company systems are being increasingly hacked. Data anonymity and confidentiality are additional aspects of vulnerability more relevant than in on-site trials, if, for instance, information is also recorded in the devices, smartphones, or is transmitted remotely. Data collection using standard behavioral scales (for instance, scales for depression) and adverse events need to be further validated to be used online and remotely to avoid instrumental biases. Finally, even open pre-publication of study protocols, which enhances transparency and reproducibility, cannot be fully detailed in digital trials, as a complete description of the groups, blinding methods, and sham stimulation of the protocols could be harmful to the internal validity of digital trials (Charvet et al., 2020, 2015).

4.3. Limitations

Although we systematically reviewed the literature for selecting the most productive authors in neuromodulation, experts publishing in non-English databases might not have been selected. In addition, most of the panelists are from high-income countries, limiting the experience, feedback, and the number of votes of panelists from low-/middle- income countries, where 85% of the world population lives, and with probably additional issues on digitalization, including availability. Moreover, our search might not have identified emerging young experts as we have established a threshold based on the number of publications. Although we considered different methods for composing the EP, such as "snowball sampling" based on recommendations from the SC members, and search of other databases such as clinicaltrials.gov, preprints, and conference publications, these processes would be non-systematic or involve gray areas in the literature. We also did not assess other stakeholders besides people from industry and academia that could have been relevant for our work, such as patients, governmental and non-governmental organizations. Moreover, only 8 of 13 companies replied to our contacts, despite several emails that were sent to reach them, and even offering the possibility of online meetings to discuss this work. Although we could have extracted company information based on public information, we opted not to do that as some information could have been inaccurate. Finally, no large tES digital trials have been finished and published yet; therefore, the processes and challenges described here are mostly theoretical and should be iteratively updated as the field develops. Interestingly, the lack of consensus on issues related to its disadvantages, risks, and biases, with many specialists remaining undecided, indicates that the field is still in its infancy.

4.4. Future directions

The recommendations for teams conducting digital neuromodulation trials markedly diverge from on-site trials that are centered on clinical specialists and staff trained in performing biomedical procedures. The feasibility of these recommendations should be further debated, as they would require more resources. Most recommendations fit with companies and for pivotal studies, and not necessarily for teams running pilot studies that would not have all the capabilities recommended above. For instance, third-party services could be contracted to do specific tasks related to software and hardware development, or such aspects could be developed together by researchers and companies. Moreover, recommendations such as a support team being always available for medical urgencies, although optimal, might be unrealistic even with large resources. Such issues would need to be carefully discussed with internal review boards and ethics committees to guarantee patient safety without harming trial feasibility.

The recommendations for further research in some aspects specifically related to internal validity, and also external validity, were largely convergent, reaching 70-80% agreement rates. Taken together, these recommendations call for new standards and best practices of fundamental pillars of tES clinical research, such as methods for sham stimulation, randomization, allocation, and assessment of adverse events. These methodologies have been steadily built over the last decade (Bikson et al., 2018a; Brunoni et al., 2019, 2012) and, although challenged in certain aspects such as sham and blinding (Fonteneau et al., 2019; Turner et al., 2021), they have been largely used in clinical trials (Fried et al., 2021). Although several approaches could be used, in a first step relatively simple randomized studies involving healthy participants could use parameters such as blinding efficacy and rate of adverse events as outcome measures, comparing whether they are different in those receiving on-site vs. online tES. Pilot studies using mobile tES in clinical samples are also encouraged to report their methodological approaches and challenges (Alonzo et al., 2019; Eilam-Stock et al., 2021).

5. Conclusion

In this first Delphi Panel evaluating opportunities, risks, and methodological issues regarding digitizing tES trials, we provided a landscape of this new approach and reached a consensus on several recommendations that should be evaluated in further studies. The panel of specialists agreed on the advantages associated with the implementation of tES trials; however, considering the fast-growing digitalization in Medicine and Biotechnology, there is a pressing need to better understand how to adapt tES trials to be performed remotely, with a clearer knowledge regarding its positive and negative aspects.

Declaration of Competing Interest

AD, RCK, ER, F. Frohlich, KH, KS, MB, PF: have equity in a company. F. Frohlich, FP, MG, MB, MN, PF, RCK, S. Laureys, WP: are members of the Scientific Advisory Board of a company. AA, F. Frohlich, FP, KMS, MB, MN, RCK, S. Laureys, UP: received honoraria of a company. CB, CL, DD, FP, KMS, S. Laureys, TR, WP: received in-kind support of a company; MC, received financial support by the Italian Ministry of Health (Ricerca Corrente); AD, RCK, ER, F. Frohlich, KS, CP: has a chief role in a company. EMNP: Scientific consulting for a company. AD, KS, MB, PF, RCK, WP: have patents related to a company. ARB receives scholarships and support from FAPESP, the Brazilian National Council of Scientific Development (CNPq-1B), University of São Paulo Medical School (FMUSP), the UK Academy of Medical Sciences (Newton

Advanced Fellowship), and the International Health Cohort Consortium (IHCC). HE is supported by the Laureate Institute for Brain Research (LIBR), Warren K. Family Foundation, Oklahoma Center for Advancement of Science and Technologies (OCAST, #HR18-139), and Brain and Behavior Foundation (NARSAD Young Investigator Award #27305). AA is supported by the VW Foundation German 2018 - Israeli Cooperation in Biological and Life Sciences, Medicine (A128416) and by the BMBF (Stimcode, 01FP2124B). TKR has received research support from Brain Canada, Brain and Behavior Research Foundation, BrightFocus Foundation, Canada Foundation for Innovation, Canada Research Chair, Canadian Institutes of Health Research, Centre for Aging and Brain Health Innovation, National Institutes of Health, Ontario Ministry of Health and Long-Term Care, Ontario Ministry of Research and Innovation, and the Weston Brain Institute. TKR also received for an investigator-initiated study in-kind equipment support from Newronika, and in-kind research online accounts from Scientific Brain Training Pro, and participated in 2021 in an advisory board for Biogen Canada Inc. TKR is also an inventor on the United States Provisional Patent No. 17/396,030 that describes cell-based assays and kits for assessing serum cholinergic receptor activity. JL receives scholarship and support from the Portuguese Foundation for Science and Technology (FCT), co-funded through COMPETE 2020 – PO Competitividade e Internacionalização/Portugal 2020/European Union, FEDER (Fundos Europeus Estruturais e de Investimento – FEEI) under the number: PTDC/PSI-ESP/30280/2017. SC receives scholarship and support from the Portuguese Foundation for Science and Technology (FCT), co-funded through COMPETE 2020 – PO Competitividade e Internacionalização/Portugal 2020/European Union, FEDER (Fundos Europeus Estruturais e de Investimento – FEEI) under the number: PTDC/PSI-ESP/29701/2017. CL has received research funding from the Australian National Health and Medical Research Council and Stanley Medical Research Foundation. GV (Ganesan Venkatasubramanian) is supported by the Department of Biotechnology - Wellcome Trust India Alliance (IA/CRC/19/1/610005) and the Department of Biotechnology, Government of India (BT/HRD-NBA-NWB/38/2019-20(6)). LBR has received support from FAPESP (grant 2019/07256-7). MB has received support from NIH (grant 1R01NS112996 and 1R01NS101362). IRV received funding from the BBSRC (Ref: BB/S008314/1). AHO has received scholarships (Process 13/10187-0 and 14/10134-7) and support from FAPESP (CEPID/BRAINN - The Brazilian Institute of Neuroscience and Neurotechnology Process: 13/07559–3), and the Brazilian National Council of Scientific Development (CNPg 479000/2012-3; 487361/2013-0; PQ-2). LC has received support from the NIH grants NIH 1R01NS112996-01A, R21NS101712-0), US Department of Defense (grant W81XWH-17-1-0320), VA Healthcare (GRANT13010404), and National MS Society (RG-1803-30492). KPB has received support from the Spaulding Research Institute (Spaulding Research Accelerator Program).

References

- ALHarbi, M.F., Armijo-Olivo, S., Kim, E.S., 2017. Transcranial direct current stimulation (tDCS) to improve naming ability in post-stroke aphasia: A critical review. Behav Brain Res 332, 7–15. https://doi.org/10.1016/j.bbr.2017.05.050
- Aljedaani, B., Babar, M.A., 2021. Challenges With Developing Secure Mobile Health Applications: Systematic Review. JMIR Mhealth Uhealth 9, e15654. https://doi.org/10.2196/15654
- Alonzo, A., Fong, J., Ball, N., Martin, D., Chand, N., Loo, C., 2019. Pilot trial of home-administered transcranial direct current stimulation for the treatment of depression. J Affect Disord 252, 475– 483. https://doi.org/10.1016/j.jad.2019.04.041
- Antal, A., Alekseichuk, I., Bikson, M., Brockmöller, J., Brunoni, A.R., Chen, R., Cohen, L.G., Dowthwaite, G., Ellrich, J., Flöel, A., Fregni, F., George, M.S., Hamilton, R., Haueisen, J., Herrmann, C.S., Hummel, F.C., Lefaucheur, J.P., Liebetanz, D., Loo, C.K., McCaig, C.D., Miniussi, C., Miranda, P.C., Moliadze, V., Nitsche, M.A., Nowak, R., Padberg, F., Pascual-Leone, A., Poppendieck, W., Priori, A., Rossi, S., Rossini, P.M., Rothwell, J., Rueger, M.A., Ruffini, G., Schellhorn, K., Siebner, H.R., Ugawa, Y., Wexler, A., Ziemann, U., Hallett, M., Paulus, W., 2017. Low intensity transcranial electric stimulation: Safety, ethical, legal regulatory and application guidelines. Clin Neurophysiol 128, 1774–1809. https://doi.org/10.1016/j.clinph.2017.06.001
- Antich-Isern, P., Caro-Barri, J., Aparicio-Blanco, J., 2021. The combination of medical devices and medicinal products revisited from the new European legal framework. Int J Pharm 607, 120992. https://doi.org/10.1016/j.ijpharm.2021.120992
- Aparício, L.V.M., Guarienti, F., Razza, L.B., Carvalho, A.F., Fregni, F., Brunoni, A.R., 2016. A Systematic Review on the Acceptability and Tolerability of Transcranial Direct Current Stimulation Treatment in Neuropsychiatry Trials. Brain Stimul 9, 671–681. https://doi.org/10.1016/j.brs.2016.05.004
- Baptista, A.F., Fernandes, A.M.B.L., Sá, K.N., Okano, A.H., Brunoni, A.R., Lara-Solares, A., Jreige Iskandar, A., Guerrero, C., Amescua-García, C., Kraychete, D.C., Caparelli-Daguer, E., Atencio, E., Piedimonte, F., Colimon, F., Hazime, F.A., Garcia, J.B.S., Hernández-Castro, J.J., Cantisani, J.A.F., Karina do Monte-Silva, K., Lemos Correia, L.C., Gallegos, M.S., Marcolin, M.A., Ricco, M.A., Cook, M.B., Bonilla, P., Schestatsky, P., Galhardoni, R., Silva, V., Delgado Barrera, W., Caumo, W., Bouhassira, D., Chipchase, L.S., Lefaucheur, J.-P., Teixeira, M.J., de Andrade, D.C., 2019. Latin American and Caribbean consensus on noninvasive central nervous system neuromodulation for chronic pain management (LAC2-NIN-CP). Pain Rep e692. 4, https://doi.org/10.1097/PR9.0000000000000692
- Bikson, M., Brunoni, A.R., Charvet, L.E., Clark, V.P., Cohen, L.G., Deng, Z.-D., Dmochowski, J., Edwards, D.J., Frohlich, F., Kappenman, E.S., Lim, K.O., Loo, C., Mantovani, A., McMullen, D.P., Parra, L.C., Pearson, M., Richardson, J.D., Rumsey, J.M., Sehatpour, P., Sommers, D., Unal, G., Wassermann, E.M., Woods, A.J., Lisanby, S.H., 2018a. Rigor and reproducibility in research with transcranial electrical stimulation: An NIMH-sponsored workshop. Brain Stimulation: Basic, Translational, and Clinical Research in Neuromodulation 11, 465–480. https://doi.org/10.1016/j.brs.2017.12.008
- Bikson, M., Grossman, P., Thomas, C., Zannou, A.L., Jiang, J., Adnan, T., Mourdoukoutas, A.P., Kronberg, G., Truong, D., Boggio, P., Brunoni, A.R., Charvet, L., Fregni, F., Fritsch, B., Gillick, B., Hamilton, R.H., Hampstead, B.M., Jankord, R., Kirton, A., Knotkova, H., Liebetanz, D., Liu, A., Loo, C., Nitsche, M.A., Reis, J., Richardson, J.D., Rotenberg, A., Turkeltaub, P.E., Woods, A.J., 2016. Safety of Transcranial Direct Current Stimulation: Evidence Based Update 2016. Brain Stimul 9, 641–661. https://doi.org/10.1016/j.brs.2016.06.004
- Bikson, M., Hanlon, C.A., Woods, A.J., Gillick, B.T., Charvet, L., Lamm, C., Madeo, G., Holczer, A., Almeida, J., Antal, A., Ay, M.R., Baeken, C., Blumberger, D.M., Campanella, S., Camprodon, J.A., Christiansen, L., Loo, C., Crinion, J.T., Fitzgerald, P., Gallimberti, L., Ghobadi-Azbari, P., Ghodratitoostani, I., Grabner, R.H., Hartwigsen, G., Hirata, A., Kirton, A., Knotkova, H., Krupitsky,

E., Marangolo, P., Nakamura-Palacios, E.M., Potok, W., Praharaj, S.K., Ruff, C.C., Schlaug, G., Siebner, H.R., Stagg, C.J., Thielscher, A., Wenderoth, N., Yuan, T.-F., Zhang, X., Ekhtiari, H., 2020. Guidelines for TMS/tES clinical services and research through the COVID-19 pandemic. Brain Stimul 13, 1124–1149. https://doi.org/10.1016/j.brs.2020.05.010

- Bikson, M., Paneri, B., Mourdoukoutas, A., Esmaeilpour, Z., Badran, B.W., Azzam, R., Adair, D., Datta, A., Fang, X.H., Wingeier, B., Chao, D., Alonso-Alonso, M., Lee, K., Knotkova, H., Woods, A.J., Hagedorn, D., Jeffery, D., Giordano, J., Tyler, W.J., 2018b. Limited output transcranial electrical stimulation (LOTES-2017): Engineering principles, regulatory statutes, and industry standards for wellness, over-the-counter, or prescription devices with low risk. Brain Stimulation: Basic, Translational, and Clinical Research in Neuromodulation 11, 134–157. https://doi.org/10.1016/j.brs.2017.10.012
- Brunoni, A.R., Nitsche, M.A., Bolognini, N., Bikson, M., Wagner, T., Merabet, L., Edwards, D.J., Valero-Cabre, A., Rotenberg, A., Pascual-Leone, A., Ferrucci, R., Priori, A., Boggio, P.S., Fregni, F., 2012. Clinical research with transcranial direct current stimulation (tDCS): Challenges and future directions. Brain Stimulation: Basic, Translational, and Clinical Research in Neuromodulation 5, 175–195. https://doi.org/10.1016/j.brs.2011.03.002
- Brunoni, A.R., Sampaio-Junior, B., Moffa, A.H., Aparício, L.V., Gordon, P., Klein, I., Rios, R.M., Razza, L.B., Loo, C., Padberg, F., Valiengo, L., 2019. Noninvasive brain stimulation in psychiatric disorders: a primer. Braz J Psychiatry 41, 70–81. https://doi.org/10.1590/1516-4446-2017-0018
- Buch, E.R., Santarnecchi, E., Antal, A., Born, J., Celnik, P.A., Classen, J., Gerloff, C., Hallett, M., Hummel, F.C., Nitsche, M.A., Pascual-Leone, A., Paulus, W.J., Reis, J., Robertson, E.M., Rothwell, J.C., Sandrini, M., Schambra, H.M., Wassermann, E.M., Ziemann, U., Cohen, L.G., 2017. Effects of tDCS on motor learning and memory formation: A consensus and critical position paper. Clin Neurophysiol 128, 589–603. https://doi.org/10.1016/j.clinph.2017.01.004
- Cappon, D., Jahanshahi, M., Bisiacchi, P., 2016. Value and Efficacy of Transcranial Direct Current Stimulation in the Cognitive Rehabilitation: A Critical Review Since 2000. Front Neurosci 10, 157. https://doi.org/10.3389/fnins.2016.00157
- Charvet, L.E., Kasschau, M., Datta, A., Knotkova, H., Stevens, M.C., Alonzo, A., Loo, C., Krull, K.R., Bikson, M., 2015. Remotely-supervised transcranial direct current stimulation (tDCS) for clinical trials: guidelines for technology and protocols. Frontiers in Systems Neuroscience 9.
- Charvet, L.E., Shaw, M.T., Bikson, M., Woods, A.J., Knotkova, H., 2020. Supervised transcranial direct current stimulation (tDCS) at home: A guide for clinical research and practice. Brain Stimulation: Basic, Translational, and Clinical Research in Neuromodulation 13, 686–693. https://doi.org/10.1016/j.brs.2020.02.011
- Cruccu, G., Garcia-Larrea, L., Hansson, P., Keindl, M., Lefaucheur, J.-P., Paulus, W., Taylor, R., Tronnier, V., Truini, A., Attal, N., 2016. EAN guidelines on central neurostimulation therapy in chronic pain conditions. Eur J Neurol 23, 1489–1499. https://doi.org/10.1111/ene.13103
- Darrow, J.J., Avorn, J., Kesselheim, A.S., 2021. FDA Regulation and Approval of Medical Devices: 1976-2020. JAMA 326, 420–432. https://doi.org/10.1001/jama.2021.11171
- Deer, T.R., Mekhail, N., Provenzano, D., Pope, J., Krames, E., Leong, M., Levy, R.M., Abejon, D., Buchser, E., Burton, A., Buvanendran, A., Candido, K., Caraway, D., Cousins, M., DeJongste, M., Diwan, S., Eldabe, S., Gatzinsky, K., Foreman, R.D., Hayek, S., Kim, P., Kinfe, T., Kloth, D., Kumar, K., Rizvi, S., Lad, S.P., Liem, L., Linderoth, B., Mackey, S., McDowell, G., McRoberts, P., Poree, L., Prager, J., Raso, L., Rauck, R., Russo, M., Simpson, B., Slavin, K., Staats, P., Stanton-Hicks, M., Verrills, P., Wellington, J., Williams, K., North, R., 2014. The Appropriate Use of Neurostimulation of the Spinal Cord and Peripheral Nervous System for the Treatment of Chronic Pain and Ischemic Diseases: The Neuromodulation Appropriateness Consensus Committee. Neuromodulation 17, 515–550. https://doi.org/10.1111/ner.12208

- Eilam-Stock, T., George, A., Lustberg, M., Wolintz, R., Krupp, L.B., Charvet, L.E., 2021. Telehealth transcranial direct current stimulation for recovery from Post-Acute Sequelae of SARS-CoV-2 (PASC). Brain Stimul 14, 1520–1522. https://doi.org/10.1016/j.brs.2021.10.381
- Ekhtiari, H., Ghobadi-Azbari, P., Thielscher, A., Antal, A., Li, L.M., Shereen, A.D., Cabral-Calderin, Y., Keeser, D., Bergmann, T.O., Jamil, A., Violante, I.R., Almeida, J., Meinzer, M., Siebner, H.R., Woods, A.J., Stagg, C.J., Abend, R., Antonenko, D., Auer, T., Bächinger, M., Baeken, C., Barron, H.C., Chase, H.W., Crinion, J., Datta, A., Davis, M.H., Ebrahimi, M., Esmaeilpour, Z., Falcone, B., Fiori, V., Ghodratitoostani, I., Gilam, G., Grabner, R.H., Greenspan, J.D., Groen, G., Hartwigsen, G., Hauser, T.U., Herrmann, C.S., Juan, C.-H., Krekelberg, B., Lefebvre, S., Liew, S.-L., Madsen, K.H., Mahdavifar-Khayati, R., Malmir, N., Marangolo, P., Martin, A.K., Meeker, T.J., Ardabili, H.M., Moisa, M., Momi, D., Mulyana, B., Opitz, A., Orlov, N., Ragert, P., Ruff, C.C., Ruffini, G., Ruttorf, M., Sangchooli, A., Schellhorn, K., Schlaug, G., Sehm, B., Soleimani, G., Tavakoli, H., Thompson, B., Timmann, D., Tsuchiyagaito, A., Ulrich, M., Vosskuhl, J., Weinrich, C.A., Zare-Bidoky, M., Zhang, X., Zoefel, B., Nitsche, M.A., Bikson, M., 2022a. A checklist for assessing the methodological quality of concurrent tES-fMRI studies (ContES checklist): a consensus study and statement. Nat Protoc 17, 596–617. https://doi.org/10.1038/s41596-021-00664-5
- Ekhtiari, H., Tavakoli, H., Addolorato, G., Baeken, C., Bonci, A., Campanella, S., Castelo-Branco, L., Challet-Bouju, G., Clark, V.P., Claus, E., Dannon, P.N., Del Felice, A., den Uyl, T., Diana, M., di Giannantonio, M., Fedota, J.R., Fitzgerald, P., Gallimberti, L., Grall-Bronnec, M., Herremans, S.C., Herrmann, M.J., Jamil, A., Khedr, E., Kouimtsidis, C., Kozak, K., Krupitsky, E., Lamm, C., Lechner, W.V., Madeo, G., Malmir, N., Martinotti, G., McDonald, W.M., Montemitro, C., Nakamura-Palacios, E.M., Nasehi, M., Noël, X., Nosratabadi, M., Paulus, M., Pettorruso, M., Pradhan, B., Praharaj, S.K., Rafferty, H., Sahlem, G., Salmeron, B. jo, Sauvaget, A., Schluter, R.S., Sergiou, C., Shahbabaie, A., Sheffer, C., Spagnolo, P.A., Steele, V.R., Yuan, T., van Dongen, J.D.M., Van Waes, V., Venkatasubramanian, G., Verdejo-García, A., Verveer, I., Welsh, J.W., Wesley, M.J., Witkiewitz, K., Yavari, F., Zarrindast, M.-R., Zawertailo, L., Zhang, X., Cha, Y.-H., George, T.P., Frohlich, F., Goudriaan, A.E., Fecteau, S., Daughters, S.B., Stein, E.A., Fregni, F., Nitsche, M.A., Zangen, A., Bikson, M., Hanlon, C.A., 2019. Transcranial electrical and magnetic stimulation (tES and TMS) for addiction medicine: A consensus paper on the present state of the science and the road ahead. Neuroscience & Biobehavioral Reviews 104, 118–140. https://doi.org/10.1016/j.neubiorev.2019.06.007
- Ekhtiari, H., Zare-Bidoky, M., Sangchooli, A., Janes, A.C., Kaufman, M.J., Oliver, J.A., Prisciandaro, J.J., Wüstenberg, T., Anton, R.F., Bach, P., Baldacchino, A., Beck, A., Bjork, J.M., Brewer, J., Childress, A.R., Claus, E.D., Courtney, K.E., Ebrahimi, M., Filbey, F.M., Ghahremani, D.G., Azbari, P.G., Goldstein, R.Z., Goudriaan, A.E., Grodin, E.N., Hamilton, J.P., Hanlon, C.A., Hassani-Abharian, P., Heinz, A., Joseph, J.E., Kiefer, F., Zonoozi, A.K., Kober, H., Kuplicki, R., Li, Q., London, E.D., McClernon, J., Noori, H.R., Owens, M.M., Paulus, M.P., Perini, I., Potenza, M., Potvin, S., Ray, L., Schacht, J.P., Seo, D., Sinha, R., Smolka, M.N., Spanagel, R., Steele, V.R., Stein, E.A., Steins-Loeber, S., Tapert, S.F., Verdejo-Garcia, A., Vollstädt-Klein, S., Wetherill, R.R., Wilson, S.J., Witkiewitz, K., Yuan, K., Zhang, X., Zilverstand, A., 2022b. A methodological checklist for fMRI drug cue reactivity studies: development and expert consensus. Nat Protoc 17, 567-595. https://doi.org/10.1038/s41596-021-00649-4
- Fonteneau, C., Mondino, M., Arns, M., Baeken, C., Bikson, M., Brunoni, A.R., Burke, M.J., Neuvonen, T., Padberg, F., Pascual-Leone, A., Poulet, E., Ruffini, G., Santarnecchi, E., Sauvaget, A., Schellhorn, K., Suaud-Chagny, M.-F., Palm, U., Brunelin, J., 2019. Sham tDCS: A hidden source of variability? Reflections for further blinded, controlled trials. Brain Stimul 12, 668–673. https://doi.org/10.1016/j.brs.2018.12.977
- Fregni, F., El-Hagrassy, M.M., Pacheco-Barrios, K., Carvalho, S., Leite, J., Simis, M., Brunelin, J., Nakamura-Palacios, E.M., Marangolo, P., Venkatasubramanian, G., San-Juan, D., Caumo, W., Bikson, M.,

Brunoni, A.R., Neuromodulation Center Working Group, 2021. Evidence-Based Guidelines and Secondary Meta-Analysis for the Use of Transcranial Direct Current Stimulation in Neurological and Psychiatric Disorders. Int J Neuropsychopharmacol 24, 256–313. https://doi.org/10.1093/ijnp/pyaa051

- Fregni, F., Nitsche, M., Loo, C.K., Brunoni, A., Marangolo, P., Leite, J., Carvalho, S., Bolognini, N., Caumo, W., Paik, N., Simis, M., Ueda, K., Ekhitari, H., Luu, P., Tucker, D., Tyler, W., Brunelin, J., Datta, A., Juan, C., Venkatasubramanian, G., Boggio, P., Bikson, M., 2015. Regulatory Considerations for the Clinical and Research Use of Transcranial Direct Current Stimulation (tDCS): review and recommendations from an expert panel. Clin Res Regul Aff 22-35. 32, https://doi.org/10.3109/10601333.2015.980944
- Fried, P.J., Santarnecchi, E., Antal, A., Bartres-Faz, D., Bestmann, S., Carpenter, L.L., Celnik, P., Edwards, D., Farzan, F., Fecteau, S., George, M.S., He, B., Kim, Y.-H., Leocani, L., Lisanby, S.H., Loo, C., Luber, B., Nitsche, M.A., Paulus, W., Rossi, S., Rossini, P.M., Rothwell, J., Sack, A.T., Thut, G., Ugawa, Y., Ziemann, U., Hallett, M., Pascual-Leone, A., 2021. Training in the practice of noninvasive brain stimulation: Recommendations from an IFCN committee. Clinical Neurophysiology 132, 819–837. https://doi.org/10.1016/j.clinph.2020.11.018
- Gillick, B.T., Gordon, A.M., Feyma, T., Krach, L.E., Carmel, J., Rich, T.L., Bleyenheuft, Y., Friel, K., 2018. Non-Invasive Brain Stimulation in Children With Unilateral Cerebral Palsy: A Protocol and Risk Mitigation Guide. Front Pediatr 6, 56. https://doi.org/10.3389/fped.2018.00056
- Godeiro, C., França, C., Carra, R.B., Saba, F., Saba, R., Maia, D., Brandão, P., Allam, N., Rieder, C.R.M., Freitas, F.C., Capato, T., Spitz, M., Faria, D.D. de, Cordellini, M., Veiga, B.A.A.G., Rocha, M.S.G., Maciel, R., Melo, L.B.D., Möller, P.D.S., R R Júnior, M., Fornari, L.H.T., Mantese, C.E., Barbosa, E.R., Munhoz, R.P., Coletta, M.V.D., Cury, R.G., 2021. Use of non-invasive stimulation in movement disorders: a critical review. Arq Neuropsiquiatr 79, 630–646. https://doi.org/10.1590/0004-282X-ANP-2020-0381
- Grimaldi, G., Argyropoulos, G.P., Boehringer, A., Celnik, P., Edwards, M.J., Ferrucci, R., Galea, J.M., Groiss,
 S.J., Hiraoka, K., Kassavetis, P., Lesage, E., Manto, M., Miall, R.C., Priori, A., Sadnicka, A., Ugawa,
 Y., Ziemann, U., 2014. Non-invasive cerebellar stimulation--a consensus paper. Cerebellum 13, 121–138. https://doi.org/10.1007/s12311-013-0514-7
- Hsu, C.-C., Sandford, B., 2019. The Delphi Technique: Making Sense of Consensus. Practical Assessment, Research, and Evaluation 12. https://doi.org/10.7275/pdz9-th90
- Inan, O.T., Tenaerts, P., Prindiville, S.A., Reynolds, H.R., Dizon, D.S., Cooper-Arnold, K., Turakhia, M., Pletcher, M.J., Preston, K.L., Krumholz, H.M., Marlin, B.M., Mandl, K.D., Klasnja, P., Spring, B., Iturriaga, E., Campo, R., Desvigne-Nickens, P., Rosenberg, Y., Steinhubl, S.R., Califf, R.M., 2020. Digitizing clinical trials. npj Digit. Med. 3, 1–7. https://doi.org/10.1038/s41746-020-0302-y
- Insel, T.R., 2018. Digital phenotyping: a global tool for psychiatry. World Psychiatry 17, 276–277. https://doi.org/10.1002/wps.20550
- Kim, D.-J., Moussa-Tooks, A.B., Bolbecker, A.R., Apthorp, D., Newman, S.D., O'Donnell, B.F., Hetrick, W.P., 2020. Cerebellar-cortical dysconnectivity in resting-state associated with sensorimotor tasks in schizophrenia. Hum Brain Mapp 41, 3119–3132. https://doi.org/10.1002/hbm.25002
- Lefaucheur, J.-P., Antal, A., Ayache, S.S., Benninger, D.H., Brunelin, J., Cogiamanian, F., Cotelli, M., De Ridder, D., Ferrucci, R., Langguth, B., Marangolo, P., Mylius, V., Nitsche, M.A., Padberg, F., Palm, U., Poulet, E., Priori, A., Rossi, S., Schecklmann, M., Vanneste, S., Ziemann, U., Garcia-Larrea, L., Paulus, W., 2017. Evidence-based guidelines on the therapeutic use of transcranial direct current stimulation (tDCS). Clin Neurophysiol 128, 56–92. https://doi.org/10.1016/j.clinph.2016.10.087
- Legatt, A.D., Emerson, R.G., Epstein, C.M., MacDonald, D.B., Deletis, V., Bravo, R.J., López, J.R., 2016. ACNS Guideline: Transcranial Electrical Stimulation Motor Evoked Potential Monitoring. J Clin Neurophysiol 33, 42–50. https://doi.org/10.1097/WNP.00000000000253

- Levac, D., Colquhoun, H., O'Brien, K.K., 2010. Scoping studies: advancing the methodology. Implement Sci 5, 69. https://doi.org/10.1186/1748-5908-5-69
- Lucchiari, C., Sala, P.M., Vanutelli, M.E., 2018. Promoting Creativity Through Transcranial Direct Current Stimulation (tDCS). A Critical Review. Front Behav Neurosci 12, 167. https://doi.org/10.3389/fnbeh.2018.00167
- Maatoug, R., Bihan, K., Duriez, P., Podevin, P., Silveira-Reis-Brito, L., Benyamina, A., Valero-Cabré, A., Millet, B., 2021. Non-invasive and invasive brain stimulation in alcohol use disorders: A critical review of selected human evidence and methodological considerations to guide future research. Comprehensive Psychiatry 109, 152257. https://doi.org/10.1016/j.comppsych.2021.152257
- Martelletti, P., Jensen, R.H., Antal, A., Arcioni, R., Brighina, F., de Tommaso, M., Franzini, A., Fontaine, D., Heiland, M., Jürgens, T.P., Leone, M., Magis, D., Paemeleire, K., Palmisani, S., Paulus, W., May, A., European Headache Federation, 2013. Neuromodulation of chronic headaches: position statement from the European Headache Federation. J Headache Pain 14, 86. https://doi.org/10.1186/1129-2377-14-86
- McClintock, S.M., Kallioniemi, E., Martin, D.M., Kim, J.U., Weisenbach, S.L., Abbott, C.C., 2019. A Critical Review and Synthesis of Clinical and Neurocognitive Effects of Noninvasive Neuromodulation Antidepressant Therapies. Focus (Am Psychiatr Publ) 17, 18–29. https://doi.org/10.1176/appi.focus.20180031
- Moffa, A.H., Brunoni, A.R., Fregni, F., Palm, U., Padberg, F., Blumberger, D.M., Daskalakis, Z.J., Bennabi, D., Haffen, E., Alonzo, A., Loo, C.K., 2017. Safety and acceptability of transcranial direct current stimulation for the acute treatment of major depressive episodes: Analysis of individual patient data. J Affect Disord 221, 1–5. https://doi.org/10.1016/j.jad.2017.06.021
- Onodera, R., Sengoku, S., 2018. Innovation process of mHealth: An overview of FDA-approved mobile medical applications. Int J Med Inform 118, 65–71. https://doi.org/10.1016/j.ijmedinf.2018.07.004
- Parikh, S.V., Quilty, L.C., Ravitz, P., Rosenbluth, M., Pavlova, B., Grigoriadis, S., Velyvis, V., Kennedy, S.H., Lam, R.W., MacQueen, G.M., Milev, R.V., Ravindran, A.V., Uher, R., CANMAT Depression Work Group, 2016. Canadian Network for Mood and Anxiety Treatments (CANMAT) 2016 Clinical Guidelines for the Management of Adults with Major Depressive Disorder: Section 2. Psychological Treatments. Can J Psychiatry 61, 524–539. https://doi.org/10.1177/0706743716659418
- Polhemus, A.M., Novák, J., Ferrao, J., Simblett, S., Radaelli, M., Locatelli, P., Matcham, F., Kerz, M., Weyer, J., Burke, P., Huang, V., Dockendorf, M.F., Temesi, G., Wykes, T., Comi, G., Myin-Germeys, I., Folarin, A., Dobson, R., Manyakov, N.V., Narayan, V.A., Hotopf, M., 2020. Human-Centered Design Strategies for Device Selection in mHealth Programs: Development of a Novel Framework and Case Study. JMIR Mhealth Uhealth 8, e16043. https://doi.org/10.2196/16043
- Razza, L., Brunoni, A., Suen, P.J.C., Ekhtiari, H., Zare-Bidoky, M., Ghobadi-Azbari, P., Zonoozi, A.K., 2021.
 Digitizing Portable Transcranial Electric Stimulation clinical trials: Scoping Review, Process
 Mapping, and recommendations from a Delphi Panel. https://doi.org/10.17605/OSF.IO/K83VP
- Sanches, C., Stengel, C., Godard, J., Mertz, J., Teichmann, M., Migliaccio, R., Valero-Cabré, A., 2021. Past, Present, and Future of Non-invasive Brain Stimulation Approaches to Treat Cognitive Impairment in Neurodegenerative Diseases: Time for a Comprehensive Critical Review. Frontiers in Aging Neuroscience 12.
- Sandars, M., Cloutman, L., Woollams, A.M., 2016. Taking Sides: An Integrative Review of the Impact of Laterality and Polarity on Efficacy of Therapeutic Transcranial Direct Current Stimulation for Anomia in Chronic Poststroke Aphasia. Neural Plast 2016, 8428256. https://doi.org/10.1155/2016/8428256

- Santos, F.H., Mosbacher, J.A., Menghini, D., Rubia, K., Grabner, R.H., Cohen Kadosh, R., 2021. Effects of transcranial stimulation in developmental neurocognitive disorders: A critical appraisal. Prog Brain Res 264, 1–40. https://doi.org/10.1016/bs.pbr.2021.01.012
- Shiozawa, P., Cordeiro, Q., Cho, H.J., Trevizol, A.P., Brietzke, E., 2017. A critical review of trials of transcranial direct current stimulation and trigeminal nerve stimulation for depression: the issue of treatment-emergent mania. Trends Psychiatry Psychother 39, 48–53. https://doi.org/10.1590/2237-6089-2016-0027
- Sierawska, A., Prehn-Kristensen, A., Moliadze, V., Krauel, K., Nowak, R., Freitag, C.M., Siniatchkin, M., Buyx, A., 2019. Unmet Needs in Children With Attention Deficit Hyperactivity Disorder-Can Transcranial Direct Current Stimulation Fill the Gap? Promises and Ethical Challenges. Front Psychiatry 10, 334. https://doi.org/10.3389/fpsyt.2019.00334
- Silva-Filho, E., Pilloni, G., Charvet, L.E., Fregni, F., Brunoni, A.R., Bikson, M., 2022. Factors supporting availability of home-based Neuromodulation using remote supervision in middle-income countries; Brazil experience. Brain Stimulation: Basic, Translational, and Clinical Research in Neuromodulation 15, 385–387. https://doi.org/10.1016/j.brs.2022.02.005
- Thibaut, A., O'Brien, A.T., Fregni, F., 2017. Strategies for replacing non-invasive brain stimulation sessions: recommendations for designing neurostimulation clinical trials. Expert Rev Med Devices 14, 633– 649. https://doi.org/10.1080/17434440.2017.1352470
- Torous, J., Bucci, S., Bell, I.H., Kessing, L.V., Faurholt-Jepsen, M., Whelan, P., Carvalho, A.F., Keshavan, M., Linardon, J., Firth, J., 2021. The growing field of digital psychiatry: current evidence and the future of apps, social media, chatbots, and virtual reality. World Psychiatry 20, 318–335. https://doi.org/10.1002/wps.20883
- Torous, J., Onnela, J.-P., Keshavan, M., 2017. New dimensions and new tools to realize the potential of RDoC: digital phenotyping via smartphones and connected devices. Transl Psychiatry 7, e1053. https://doi.org/10.1038/tp.2017.25
- Turner, C., Jackson, C., Learmonth, G., 2021. Is the "end-of-study guess" a valid measure of sham blinding during transcranial direct current stimulation? Eur J Neurosci 53, 1592–1604. https://doi.org/10.1111/ejn.15018
- Woods, A.J., Antal, A., Bikson, M., Boggio, P.S., Brunoni, A.R., Celnik, P., Cohen, L.G., Fregni, F., Herrmann, C.S., Kappenman, E.S., Knotkova, H., Liebetanz, D., Miniussi, C., Miranda, P.C., Paulus, W., Priori, A., Reato, D., Stagg, C., Wenderoth, N., Nitsche, M.A., 2016. A technical guide to tDCS, and related non-invasive brain stimulation tools. Clinical Neurophysiology 127, 1031–1048. https://doi.org/10.1016/j.clinph.2015.11.012
- Workman, C.D., Fietsam, A.C., Rudroff, T., 2020. Tolerability and Blinding of Transcranial Direct Current Stimulation in People with Parkinson's Disease: A Critical Review. Brain Sci 10, E467. https://doi.org/10.3390/brainsci10070467
- Zhang, Y., Song, H., Chen, Y., Zuo, L., Xia, X., Zhang, X., 2019. Thinking on Transcranial Direct Current Stimulation (tDCS) in Reading Interventions: Recommendations for Future Research Directions. Front Hum Neurosci 13, 157. https://doi.org/10.3389/fnhum.2019.00157