Computer-assisted analysis of polysomnographic recordings improves inter-scorer associated agreement and scoring times

Diego Alvarez-Estevez[1], Roselyne M. Rijsman[2]

[1] Universidade da Coruña, Center for Information and Communications Technology Research (CITIC), A Coruña, Spain

[2] Sleep center and clinical neurophysiology department, Haaglanden Medisch Centrum, The Hague, The Netherlands

Corresponding author:

Diego Alvarez-Estevez, PhD

Center for Information and Communications Technology Research (CITIC)

University of A Coruña

Edif. CITIC, planta 2, Campus Elviña s/n

15071 A Coruña, Spain

Tel: +34 881 01 5531

e-mail: diego.alvareze@udc.es

ABSTRACT

Study Objectives: To evaluate the current levels of human-expert performance and to assess the possible benefits of computer-assisted analysis for the scoring of polysomnographic sleep recordings (PSGs).

Methods: A group of 12 expert scorers independently reviewed 5 PSGs for each of the following tasks: (i) sleep stating, (ii) detection of EEG arousals, (iii) analysis of the respiratory activity, and (iv) identification of leg movements, thus amounting to a total of 20 recordings. For each one, rescoring was repeated separately, using the classical visually-based manual approach and a computer-assisted semi-automatic procedure. Resulting inter-scorer agreement and scoring times were examined and compared among the two methods.

Results: Computer-assisted sleep scoring showed a consistent and statistically relevant effect toward less time required for the completion of each of the PSG scoring tasks. Gain factors ranged from 1.26 (EEG arousals) to 2.41 (limb movements). Inter-scorer kappa agreement was also consistently increased with the use of supervised semi-automatic scoring. Specifically, from K=0.76 to K=0.80 (sleep stages), K=0.72 to K=0.91 (limb movements), K=0.55 to K=0.66 (respiratory activity), and K = 0.58 to K=0.65 (EEG arousals). Inter-scorer agreement on the examined set of diagnostic indices did also show a trend toward higher Interclass Correlation Coefficient scores when using the semi-automatic scoring approach.

Conclusions: Computer-assisted analysis can improve inter-scorer agreement and scoring times associated with the review of PSG studies resulting in higher efficiency and overall quality in the diagnosis sleep disorders.

Keywords

Polysomnographic recordings; Inter-scorer variability; Sleep staging; Limb movements; Respiratory activity; EEG arousals; Manual scoring; Automatic analysis;

STATEMENT OF SIGNIFICANCE

This study shows that, in comparison to manual scoring, the use of semi-automatic PSG analysis has benefits in the form of faster scoring and higher associated inter-scorer agreement. A number of firsts are included, as to our knowledge, this is the first work to systematically address the comparison between manual and semi-automatic scoring through the use of quantifiable metrics of scoring time and inter-scorer reliability. Existing literature regarding manual or semi-automatic scoring time analysis of limb movements or EEG arousals is scarce, and for sleep staging or apneic event detection, directly inexistent. Likewise, this is the first work to provide metrics of inter-scorer reliability for derived semi-automatic event markings as well as the resulting indices of diagnose.

1. INTRODUCTION

The analysis of polysomnographic sleep recordings (PSGs) constitutes one of the most time-consuming tasks in the daily work of a Sleep Center. A typical PSG examination contains somewhere between eight up to twenty-four hours of continuous neurophysiological activity recording. Common PSG data include, among others, different traces of electroencephalographic (EEG), electrooculographic (EOG), electrocardiographic (ECG), electromyographic (EMG), and respiratory activity [1]. Likewise, analysis of the PSG can be organized into different subtasks, for instance, analysis of the macro and micro structure of sleep, characterization of the respiratory function, or identification and scoring of limb movement activity.

Clinical findings over the last years have uncovered the negative consequences that Sleep Disorders exert over health, contributing to the general public awareness. The former has led to a steady increase in the demand for PSG investigations. This represents a challenge for the already congested sleep centers. Clinician's time is expensive and scant. In addition, the large amount and the complexity of the associated data, makes of PSG analysis a task prone to errors and to subjective interpretations [2]. Indeed, despite homogenization procedures promoted by development and usage of clinical standard guidelines [1] [3], different grades of intra- and inter-expert variability have been reported in the literature, affecting the resulting PSG outcomes, which vary among the specific references or tasks subject to evaluation [4] [5] [6] [7] [8]. In this context, the use of automatic scoring algorithms to support the clinician in the analysis of the PSG presents potential advantages. On one hand, in terms of great savings in scoring time, and thus in human resources, helping reducing the overall costs associated with the diagnosis. Literature, in fact, is rich on examples that focus on the development and validation of automatic analysis methods in different areas related to the scoring of sleep studies [5] [9] [10] [11] [12] [13] [14] [15] [16] [17]. However, despite the promising man-machine agreement results reported in some of these works, practical acceptance of these systems among the clinical community remains low. In fact, there are still open questions on whether "are we there yet", in terms of acceptable performance and enough generalization capabilities of these algorithms as compared to well-trained human clinicians [2].

Automatic scoring, on the other hand, has another interesting property, which is the ability to provide

deterministic (repeatable) diagnostic outcomes. Hence, it is conceivable that automatic scoring could

play a role to help reduce the levels of inter-scorer variability, thus contributing to standardization and

overall quality improvement in the diagnostic process. That we know of, however, no previous

attempts have been made to formally address this hypothesis.

The main goal of this study is to evaluate the current levels of human-expert performance and to

assess the possible benefits of using computer-assisted analysis in the scoring of PSGs. For this

purpose, we are making use of objective (quantifiable) quality indicators regarding the scoring time

(time needed to score a full PSG) and the repeatability of the scoring outcomes. More specifically, the

following four (usually, the most important) PSG analysis subtasks are considered: sleep staging,

analysis of EEG micro-arousals, evaluation of the respiratory function, and scoring of limb movement

activity. For each of these tasks, the standard manual and computer-assisted scoring approaches are

evaluated and compared. Under the manual approach the clinician performs analysis of PSG data

using the commonly accepted 'gold standard' approach based on visual scoring. In the computer-

assisted so-called "semi-automatic" approach, an automatic algorithm performs a preliminary analysis

pass, and the clinician then reviews the results by adding, removing or editing miss-scorings from the

automatic algorithm.

Our work is novel as no previous studies have attempted to examine the hypothesis on whether semi-

automatic scoring can contribute to reduction of inter-scorer variability. To our knowledge this is also

the first study to systematically address possible scoring time differences between the manual and

semi-automatic approaches. Furthermore, there is a general lack of comprehensive studies that

confront human and (semi)automatic scoring. In effect, the available literature usually targets the

validation of the automatic algorithm by assessing its performance in comparison to one or more

reference scorers. However, contextualization of the results with respect to the baseline levels of

human performance in the corresponding task is often omitted. This is a problem as this is necessary

to set a valid target reference for the automatic algorithm and to discard possible overfitting to specific

scorers or bias effects. Extrapolation of human baseline levels from independent external studies, on

the other hand, is in general not possible. As stated before, inter-scorer and database variability plays

a role, for which ideally the respective human and computer performances should be derived, and

compared, using a common patient dataset.

To have objective (measurable) references of the levels of human performance is as well of

fundamental importance to allow implementation of quality control mechanisms in the patient care. Our

work contributes to this aspect as well by adding to the existing literature. In this respect it should be

noted that evolution of the scoring methods and reference clinical guidelines motivates reassessment

of the existing references, for which some of them are outdated. Furthermore, for some of the

examined scoring tasks literature references of related quality factors have never been reported.

2. METHODS

2.1. Study database

PSG data for this study has been gathered by retrospective inspection of the Haaglanden Medisch

Centrum (HMC, The Hague, The Netherlands) Sleep Center clinical patient database. All data were

acquired in the course of the common clinical practice. No patient was therefore subjected to any

additional behavior in relation to this study, nor was prescribed any additional treatment outside of the

regular clinical workflow.

Recordings were de-identified and subrogate study numbers were assigned to each patient, avoiding

any possibility of individual patient identification. Under these conditions the study obtained approval

of the local Medical Ethics Committee (Medisch Ethische Toestsingscomissie Zuidwest Holland) under

code MTEC-19-065, who considered that the protocol did not fall under the scope of the Medical

Scientific Research Involving Human Subjects Act (WMO) and that no explicit informed consent was

required by participants. Study has as well obtained written permission from the database owner for

publication.

PSG data consisted of raw biomedical signals following standard acquisition procedures according to

the AASM guidelines [1]. SOMNOscreen^{TM} plus devices (SOMNOmedics, Germany) were used as the

acquisition hardware. Event annotations resulting from the scorers' reviews during regular clinical

workflow accompanied each recording. Clinical scorings were carried out by HMC sleep technicians

including the scoring of sleep stages, EEG arousals, and respiratory activity, following the AASM

standards [1], and the scoring of the limb movement activity following the WASM2016 guidelines [3].

Both the raw signal's data and the resulting clinical scoring annotations were digitally stored using the

EDF+ format [18].

2.2. Rescoring task

In the present study a group of sleep technicians were prompted to review 5 PSG recordings,

independently for each of the following scoring subtasks: (i) sleep stating, (ii) detection of EEG

arousals, (iii) analysis of the respiratory activity, and (iv) identification of leg movements. Thus, in total,

20 different PSG recordings were included in the study. All sleep scorers belonged to HMC staff and

participated regularly on clinical workflow in an autonomous mode. Sleep technicians undergoing

training or supervision were excluded. In total 12 scorers participated in this task.

Each of the participant scorers were tasked to review the exact same recordings. In all cases, scorers

were blinded to both the patient identity (by using de-identified recordings) and possible results of

previous scorings (e.g. that took place during regular clinical workflow, from other scorers, or during a

previous self-rescoring subtask).

Rescoring was repeated, separately for each recording, using a purely manual and a semi-automatic

approach. To avoid learning effects, at least 4 months of separation were scheduled between these

two moments. For reference, the number of PSG recordings scored by each sleep technician due to

the normal working routine in that time approximately averages to 70 patients. Scorers were also not

informed of the fact that manual and semi-automatic scorings would involve the same recordings.

All scorings took place using the Polyman software [19]. For each task a timer was automatically set in

the background by the program (not available to the human scorer). The tick counting was

automatically paused if no mouse or keyboard interaction was detected during more than a minute,

and the offline time was subtracted from the total scoring time. The resulting active scoring time was

saved separately in a file for later analysis.

Scoring took place between time in bed periods only (between lights off – on markers, which were

provided as pre-filled annotations). For scoring of EEG arousals, respiratory events and limb

movements, the pre-filled clinical hypnogram was also provided as additional source for contextual

interpretation. Regardless, scorers were instructed to stick to the scoring of the relevant events in the

context of the specific target task, not being allowed to change any pre-filled contextual information

provided.

During the semi-automatic scoring process, the annotations that resulted from the output of the

corresponding automatic analysis algorithms were provided, in addition, at the start of the scoring.

Scorers were instructed to review these scorings by adding, deleting, or editing the event's onset and

offset times, where corresponds, and according to their own expertise. Details regarding the

development and validation of the automatic scoring algorithms that were used for this purpose have

been reported in past works. The reader is referred to check the specific references regarding the

automatic scoring of sleep stages [2] [20], leg movement activity [21] [22], respiratory events [23] [24],

and EEG arousals [25] [26].

2.3. Selection of PSGs

For each of the scoring subtasks targeted in this study (i.e. sleep staging, EEG arousals, respiration,

and limb movements) 5 PSG recordings were selected from an initial pre-sample dataset of 2801

recordings, corresponding to the most recent one-year data from the HMC database at the time

(2019). Sampling size per task was determined by availability of resources with regard to expert

scoring time. As result, 20 PSG recordings were selected that integrate the final scoring sample for the

study.

The selection process was implemented with the objective to minimize the chance of selection bias

and deal with the inherent inter-subject variability among the PSGs' initial pre-sample distribution. No

specific exclusion criteria were applied during the selection process to filter out recordings, for

instance, due to specific patient conditions, or poor signal quality. A sufficient condition was that the

recording had been accepted for manual scoring during the regular clinical workflow, a condition that,

by definition, was already satisfied by all recordings included in the pre-sample dataset. The

underlying motivation was that of reproducing, as close as possible, the same conditions as in real

practice, by considering a representative sample of the general patient phenotype.

Hence, for each of the four scoring subtasks, the following selection procedure was scheduled:

(i)   Taking as reference the complete pre-sample dataset (2801 PSGs) a full automatic

analysis (no human intervention at all) of the recording was performed. This analysis led

142    to a list of automatically scored events $L_a(i)$, for each recording i, related to the

143    corresponding target task on each case.

145  (ii)  Using the list of automatically-generated events, $L_a(i)$, each PSG was compared with the

146    corresponding list of events that resulted during routine clinical examination, $L_c(i)$.

147    Confronting $L_a(i)$ with $L_c(i)$, a preliminary metric of performance agreement between the

148    two scoring outputs, $K_{ac}(i)$, was obtained. Specifically, K_{ac} was calculated using the

149    Cohen´s Kappa statistic [27]. Details on the implementation of K_{ac} for each of the four

150    target subtasks are described in the "analysis methods" section.

152  (iii)  By repeating this operation through all 2801 PSGs available in the initial pre-sample

153    dataset, a distribution DK_{ac} of $K_{ac}(i)$ values was obtained.

155  (iv)  Using DK_{ac} as reference, uniform sampling was performed to select the target number

156    $N=5$ of recordings to be included in each subtasks' final study dataset. Specifically, the 5

157    recordings whose associated $K_{ac}(i)$ performance metrics represent the middle of each

158    inter-quartile range, plus the median, were selected as representatives of their respective

159    whole populations. In other words, the recordings with performance scores representing

160    the 12.5th, 37.5th, 50th, 62.5th and 87.5th percentiles of each DK_{ac} distribution were

161    selected for the final study dataset.

163 Effectively, the above described procedure is preferable over random resampling as it avoids potential

164 selection of outliers by chance (i.e. extreme favorable or unfavorable cases for the automatic

165 algorithm) that might bias the resulting sample. Instead, the procedure seeks an even distribution of

166 the relative recording scoring difficulty throughout the final study dataset. In this respect, it is

167 hypothesized that automatic scoring difficulty is correlated with the manual scoring difficulty: the lower

168 (higher) the agreement between the automatic and the clinical reference would be, the lower (higher)

169 the resulting inter-rater agreement is expected to be. Similar selection procedures were scheduled

170 during the validation of different automatic scoring algorithms that were reported in the past [26] [20].

171 Correlation analyses are scheduled to analyze validity of this hypothesis whose results are provided in

Supplementary Table D1 and discussed in the corresponding Section D of the Supplementary

Materials.

Table 1 summarizes the general demographics and PSG descriptors in the resulting patient study

sample. Data are presented stratified among the corresponding task-specific subgroups.

2.4. Analysis methods

Analysis of inter-scorer agreement is carried out in the first place by discretizing the recording time into

non-overlapping analysis mini-epochs. Each analysis mini-epoch is assigned the corresponding

scorer's output in the context of the specific target subtask. Duration of the mini-epochs are task-

related as well. In the case of the sleep scoring, analysis epochs have the standard duration of 30s

and take possible values according to the AASM clinical guidelines, that is, either W, N1, N2, N3, or R

[1]. In the context of the EEG arousals, respiratory events, and limb movement's scoring subtasks,

each mini-epoch takes a binary value noting the presence or absence of event, respectively, if

overlapping or not with the events marked by the scorer. Analysis mini-epoch duration is set to 0.5s for

all the three subtasks.

Time discretization in the above terms leads to the construction of k-dimensional contingency tables

(k=5 for sleep staging, k=2 otherwise) from which standard metrics of agreement for categorical data

can be derived. Within each task, agreement between each score's pair combination is calculated.

More specifically, the Cohen's kappa statistic is used for this purpose. The use of the Cohen's kappa

is motivated given its widespread use in the field, as well as its robustness in the case of imbalanced

class distributions as it corrects for agreement due to chance [27].

Inter-scorer agreement is also evaluated among the diagnostic indices resulting from the respective

scorings. Following the list of recommended parameters to be reported in PSG studies [1] [3], a

representative subset for each of the subtasks targeted in this study is selected. In particular, sleep

quality-related parameters of Sleep Efficiency (SE), Sleep Onset Latency (SOL), and Wake After

Sleep Onset (WASO) [28], in relation to the sleep scoring task; Apnea-Hypopnea Index (AHI), Apnea

Index (AI), Hypopnea Index (HI) and Oxygen Desaturation Index (ODI), in relation to the scoring of

respiratory events; Arousal Index (ArI), in relation to the scoring of EEG arousal events; and Leg

Movement Index (LMI) and the Periodic Leg Movement Index (PLMI), in relation to the limb movement

scoring task. LMI and PLMI indices are calculated according to the WASM2016 scoring guidelines, the

former as the number of limb movements ≥ 0.5s after bilateral combinations per hour of sleep, and the

latter including respiratory-related LMs as well [3]. Inter-scorer agreement among the resulting indices

within each method is evaluated using the Intraclass Correlation Coefficient (ICC) [29]. Specifically, a

two-way absolute single-measures variant of the statistic, ICC(A,1), is used [30]. A Matlab

implementation for calculation of this coefficient has been used whose source code is available at [31].

Hypothesis testing is carried out to check for significant differences between the manual and semi-

automatic scoring approaches. For this purpose, the reference level for statistical significance is set to

$\alpha = 0.05$. Kappa agreement differences between the two approaches are examined using the paired

version of the Wilcoxon signed rank test among all the matched scorer pair combinations. Analogous

analysis is performed for checking out differences in the respective scoring times among the matched

individual scorers. For each test the corresponding effect size is reported using the Cohen's D statistic.

Statistical significance on inter-scorer ICC agreement differences among diagnostic indices is also

evaluated. For this purpose, the a priori expected agreement (*r0*) for the semi-automatic approach is

set to the effective ICC levels achieved with manual scoring.

Results of the above-mentioned analyses are presented in the subsequent section by aggregating the

respective scorings among the five recordings involved within each scoring task. In order to keep the

main text extension attainable, individualized per-recording results are provided as Supplementary

Material. Manual vs. semi-automatic differences in diagnostic indices in this case are examined using

paired analyses, and comparison of the respective variance distributions is examined using the Brown-

Forsythe (unpaired) test. For the latter, i.e. comparison of distribution's variance, the corresponding

manual and semi-automatic indices are first mean normalized within their respective distributions to

avoid possible bias due to differences in the respective population means. Supplementary materials

do also provide correlation analyses aiming to test the PSG selection hypothesis, namely that

automatic scoring difficulty is correlated with manual scoring variability.

3. RESULTS

3.1. Analysis of scoring time

Figure 1 shows the median scoring time associated with the completion of the different analysis tasks according to the followed approach, i.e. manual or semi-automatic. Values on the bar plot are shown in minutes and aggregate the results among the five recordings involved on each case.

Table 2 expands the results of Figure 1 and shows the results of the associated statistical analyses involving the two scoring approaches. Data in Table 2 unveil a consistent and statistically relevant effect toward less time required for the completion of each task when using the semi-automatic scoring approach. Gain factors vary per task, with the largest time savings relating to the scoring of limb movements, followed by the analysis of the respiratory activity, and a less pronounced effect associated with the sleep staging task and the scoring of EEG arousals. The associated effect sizes on each case support these interpretations. In this regard, notice that a positive sign on the corresponding index indicates that the overall effect (in this case scoring time) is bigger in the manual scoring scenario, with the associated absolute value being an indicative of how much bigger the effect is.

When comparing absolute time values among the different tasks, our results show that identification of limb movements is the most time consuming task when using manual analysis. Scoring of respiratory events is relatively the quickest. The trend changes a bit when using the semi-automatic approach, resulting in sleep staging being the slowest, with analysis of respiratory activity remains as the fastest task.

Individualized per-recording and per-scorer analyses for each task can be found, respectively, in Supplementary Tables A1-A4 and Supplementary Figures A1-A4.

3.2. Analysis of kappa agreement

Figure 2 shows the global kappa agreement results per scoring task when comparing manual and semi-automatic scoring approaches. Values on the bar plot represent the median expert paired agreements among the five recordings within the corresponding task.

Table 3 shows results of the statistical analyses between the corresponding manual and semi-

automatic scoring differences. Moreover, results are subcategorized for some of the tasks into

different contexts of clinical interest. In particular, differences between wake and sleep periods are

reported for limb movements, as well as results involving different types of events within the

respiratory analysis task. For the analysis of the limb movement activity, the individual kappa scores

for each leg channel (left / right) were averaged together before statistical analysis was executed.

Results from Table 3 show that statistically significant differences between manual and semi-

automatic scoring are reached regardless of the specific task or the event subtype. A consistent trend

toward higher inter-scorer agreement associated with the use of semi-automatic scoring is shown.

Notice the associated effect sizes overall show a negative sign, being indicative of the general smaller

agreement achieved in the manual scoring scenario. The highest absolute effect in this context is

associated with the analysis of the limb movement activity task.

When comparing among the different tasks, the highest (either manual or semi-automatic) agreements

are achieved in the case of the sleep staging and limb movements' analysis tasks. For the latter,

higher agreement is obtained during sleep periods than during wakefulness. With regard to the

analysis of respiratory activity, and attending to the different event subtypes, higher agreement is

achieved for the scoring of apneas than of hypopneas. Rather low agreement is achieved associated

with manual scoring of oxygen desaturation events (median K = 0.40), which is much improved when

using the semi-automatic approach (median K = 0.82). Finally, reliability associated with the scoring of

EEG arousal events reaches agreement levels similar to those obtained for the identification of

respiratory events in general (i.e. apneas, hypopneas and RERAs altogether).

Individualized per-recording analyses for each of the tasks are supplied in Supplementary Tables B1-

B9.

3.3. Analysis of derived diagnostic indices

Table 4 examines inter-scorer agreement among the selected list of diagnostic parameters for the

manual and semi-automatic scoring approaches. Agreement is evaluated using the Interclass

Correlation Coefficient (ICC).

When comparing absolute ICC values among the different tasks, a trend can be seen toward higher

inter-scorer agreement when using the semi-automatic scoring approach, with the only exception of

SOL. Regardless of the scoring approach, the highest general agreement is achieved in the case of

SE, AI, and WASO (ICC > 0.99 in all cases). Agreement associated with the scoring of apneas

probably contributes to the relative high scores achieved for the AHI too. Detection of hypopneas as

reflected by HI, on the other hand, shows relative lower levels of ICC agreement. HI is, in fact, is the

index where the lowest overall agreement is achieved, followed by ArI. All the obtained ICC values,

also regardless of the scoring approach, achieve statistical significance when the null hypothesis

assumes no a priori agreement ($r0 = 0$)

For examining statistical significance of the observed differences between the manual and semi-

automatic approaches, the null hypothesis is set, in addition, to assume baseline ICC levels

corresponding to manual scoring. In this case, significant differences are obtained for the indices of

WASO, LMI, PLMI, AI and ODI. In the case of SE, AHI, HI and ArI, the trend remains consistent

toward higher ICC values when using the semi-automatic scoring approach, albeit analyses do not

reach statistical relevance. And finally, only for SOL, higher ICC values are obtained using manual

scoring, but again differences do not reach the level of statistical significance.

Individualized per-recording analyses are supplied in Supplementary Tables C1-C10. In this case

manual vs. semi-automatic differences are examined both using paired Wilcoxon sign-rank and

unpaired Brown-Forsythe tests, as described in the methods section.

4. DISCUSSION

The main goal of this study was to evaluate the current levels of human-expert performance in the

scoring of PSGs, and to assess the possible benefits of using automatic analysis in supporting this

task. For this purpose, we have considered four of the most common subtasks involved in the analysis

of PSGs: sleep staging, identification of EEG arousals, evaluation of the respiratory function, and

scoring of limb movement activity. On each case, quantifiable metrics of performance regarding the

scoring time, and inter-scorer agreement, have been examined and compared among the manual and

semi-automatic scoring procedures. To our knowledge, this is the first study to systematically address

the differences between these two approaches.

Our experimentation has shown that the use of semi-automatic analysis has benefits in the form of

faster scoring and higher inter-scorer agreement. Faster scoring means reduced associated costs, and

the possibility to reduce the waiting lists by the consequent increase in the scoring production. Higher

inter-scorer agreement means better consistency and reliability of the PSG outcomes, and therefore

improved quality of the diagnosis. The trend is consistent though all of the four examined tasks.

Differences between the two approaches have achieved statistical significance both for the scoring

time and the expert agreement measured in terms of Cohen's K. The impact of these differences on a

subset of derived diagnostic indices, analyzed in terms of ICC agreement, has shown a more

heterogeneous pattern. While statistical significant differences have been observed for indices of

WASO, LMI, LMI, AI and ODI, this was not the case for indices of SE, SOL, AHI, HI and ArI.

Regardless, for all the examined indices with the exception of SOL, the trend was consistent toward

higher ICC values when using the semi-automatic scoring approach.

Our study provides a number of firsts. To the authors knowledge, for example, this is the first study

reporting and comparing the time associated with the scoring of sleep stages and respiratory events

(both for the manual and semi-automatic approaches). Our data shows an average gain factor in

scoring time of 1.33 and 1.63, respectively, when using semi-automatic scoring. With regard to the

detection of limb movements, at least one earlier work has also examined the hypothesis of scoring

time differences between manual and semi-automatic scoring. Specifically, in Roessen et al. [32], the

manual review of 30 PSGs took a total of 28 man-hours, whereas the semi-automatic procedure

required 10 man-hours, hence resulting in a 2.8 gain factor favoring the latter. This result is consistent

with our experiments in which we obtained a 2.41 gain factor for the limb movement detection task.

More specifically, in our case the associated median scoring time per recording was of 44.53 minutes

for the manual approach, and 18.50 minutes when using the semi-automatic procedure. With regard to

the scoring of EEG arousals, in a past work we have already reported an estimated 20-25% speed

improvement associated with the use of semi-automatic scoring [26]. In both studies the same

automatic EEG arousal scoring algorithm was used, with the estimation reported back then not

differing much from the 26% improvement obtained in the present study. In [26] only the mean manual

scoring time per individual was reported, averaging to 25 minutes (18 – 50 minute range), which is

also not far from the median 27.50 minutes, but with higher variability (12.91 – 92.15 range), obtained

in the present work. In addition, the respective values when using semi-automatic scoring are also

reported in this study, decreasing to a median of 21.78 minutes per recording for the same task.

Our study is also the first to investigate inter-scorer reliability in the context of semi-automatic scoring.

In the case of manual scoring, the availability of past references for contextualization of our results

depends on the specific task. For example, several studies have previously addressed inter-scorer

reliability associated with manual scoring of sleep stages. Detailed discussion can be found in another

recent publication by the authors [20], resulting in kappa coefficients ranging widely between 0.46 and

0.89 depending on the consulted reference. This range is compatible with the median agreement

achieved in this study (K = 0. 76 for manual, K = 0.80 for semi-auto). On the other hand, only one past

reference was found examining inter-scorer kappa agreement in the detection of limb movements. In

the study of Pittman et al. [5] K = 0.77 was obtained between two scoring experts on a dataset of 31

PSGs. Notice, however, that agreement reported in Pittman et al. refers only to the scoring of PLMS,

not LMs, and that the scoring reference was based on older standards (ASDA1993 [33]). Moreover,

analysis was constrained to sleep periods only, and its associated resolution was 30s. Our results

using the recent WASM2016 scoring standards resulted in a global K = 0.72 for manual, and K = 0.91

for semi-automatic scoring, when examining LMs during TIB using a 0.5s analysis step. Agreement

falls respectively to K = 0.67 and K = 0.89 during wake periods, and improves to K = 0.75 and K =

0.92 during TST.

Pitman et al. have reported as well a K = 0.82 for the manual scoring of apneas and hypopneas using

the 2001 AASM Medicare scoring definitions on a 30s analysis epoch [5]. With our settings, we have

achieved rather lower agreement resulting in median K = 0.55 (improving to K = 0.66 with semi-

automatic scoring). We have obtained higher agreement for the scoring of apneas (median K = 0.74

for manual, K = 0.88 for semi-automatic) than for the case of hypopneas (respectively K = 0.46 and K

= 0.61). This is an expected result, however no study that we know of had attempted to quantify this

difference in terms of kappa agreement so far. Perhaps more surprising, rather low agreement has

been obtained for the manual scoring of desaturation events (median K = 0.40), which however

improved significantly with the use of semi-automatic scoring (K = 0.82). The underlying reason that

explains this difference has yet to be determined.

As for the EEG arousal task, some studies can be found reporting kappa values for manual scoring in

the 0.47 – 0.59 range [26] [34] [35]. Once again, some of these studies use older scoring guidelines

(ASDA1992) besides other sources of variability, and therefore direct comparison has to be carefully

considered. Regardless, the reported range is consistent with our experimental results in the case of

manual scoring (K = 0.58). Our study shows, in addition, that better inter-scorer agreement can be

achieved if using semi-automatic scoring (up to K = 0.65 in our dataset).

In this study we have also examined possible differences among the resulting derived indices for

clinical diagnosis. This is, to our knowledge, the first study to address inter-scorer agreement of

diagnostic indices obtained during semi-automatic scoring. In the context of manual analysis, on the

other hand, one can find several other past studies reporting on related ICC agreement scores [5] [36]

[7] [37] [38] [39] [6] [40] [41]. In this regard, the specific agreement references vary per study. Danker-

Hopfe et al. [6] and Kuna et al. [41], for example, have reported ICC values for SE of 0.91 and 0.77

respectively, which is below the agreement obtained in this study (ICC = 0.99). Reliability on PLMS

has been reported by Pittman et al. [5] (ICC = 0.93) and Bliwise et al. [36] (ICC = 0.91 – 0.99),

however, using older definitions of the index [42] [43]. This is relevant as recent studies [44] [45] [46]

have pointed out to significant differences in the resulting PLM index calculations when using as

reference the latest clinical scoring guidelines. The agreement results are nevertheless comparable to

the levels obtained in our work (ICC = 0.94), which use the recent WASM2016 standards [3]. Under

this reference, our study is in fact the first one to set a reference for the agreement associated with

manual derivation of the LM and PLM indices (ICC = 0.92 and 0.94, respectively). With regard to

respiratory-derived indices, possibly the most widely reported is the AHI, with reliability scores for

manual scoring nevertheless ranging widely between ICC 0.54-0.99 depending on the consulted study

[5] [7] [41] [47]. Most likely, these differences are to a great extent driven by the specific rule used for

the scoring of hypopneas. As stated before, it is widely accepted that agreement regarding scoring of

hypopneas is lower as compared to that of apneas. This can also be observed by comparing ICC

agreement values associated with AI and HI indices reported the literature. This is also the case in our

study, with ICC agreement for the manual derivation of respiratory related indices showing a relatively

high score for AHI (ICC = 0.98), following the expected trend of higher agreement for AI (ICC = 0.99)

in comparison to HI (ICC = 0.60). In the case of the associated oxygen desaturation index, and ICC of

0.84 was obtained. Finally, and with regard to reliability of ArI indices resulting from manual scoring,

literature shows high variability as well (ICC = 0.52-0.96 [39] [37] [7] [38] [5]). Our results for manual

scoring approximately fit in the middle of that range (ICC = 0.68).

A general warning when comparing results from different works in the literature, is that one has always

to bear in mind that relevant differences might exist between the respective population samples, the

analysis methods, or the clinical scoring references valid at the time of the study.

That said, some limitations of our study have to be mentioned as well. First, it is important to remark

that absolute values of the investigated quality scores are associated with performance of expert

scorers from one specific sleep lab. This study does not involve analysis of inter-scorer variability

across multiple centers, and thus results might not generalize to other centers. In such scenario, the

respective values of scoring agreement are expected to be lower in comparison due to the greater

amount of variability involved [2]. This study neither has attempted to quantify the corresponding levels

of intra-scorer variability. It cannot completely be ruled-out that some of the differences between

manual and semi-automatic approaches could be influenced by a component of intra-scorer variability

effects, at the individual scorer case at least. Nevertheless, the relative high number of involved

experts (12 in our study) should contribute to limit its the impact on the global results.

It should also be remarked that quality indicators derived from the semi-automatic scoring procedure

are likely modulated by the reliability and performance of the specific automatic analysis algorithms

used in the first instance. One might speculate with the idea that the better the algorithm, the higher

the improvement on expert agreement with respect to the manual approach. However, there is no

actual evidence allowing the support of this hypothesis. The usage of alternative automatic scoring

methods might lead to different results. Regardless, our results support the hypothesis that semi-

automatic algorithm can improve scoring quality in terms of both speed and resulting inter-scorer

agreement. Also interesting, inter-scorer reliability studies available through literature, and this is no

exception, implicitly assume that the outcome of all human scorers is equally valid. This might be a

risky assumption, although there is no clear formula to discern who (out of a set of human experts)

represents the best reference, and who does not. This propounds an interesting line of future research

linked to another non-less interesting debate: can automatic scoring perform better than human

experts? Of course, in terms of its capacity to correctly identify the relevant events associated with the

physiological activity's ground truth, as automatic analysis can obviously outperform manual

counterpart in terms of scoring speed. If, like in this case, the standard reference is subject to

variability associated with human decision, it does not seem very plausible that any automatic

algorithm could perform beyond the limit set by the average human agreement. After all, as stated

before, deviations from such a reference do not necessarily correlate with the quality of the associated

scorings. This is a subject that deserves more study.

Last but not least, one another possible limitation of this study relates to the number of PSGs involved

in the evaluation of each analysis subtask. The relative high number of sleep experts involved partially

counteracts this fact, and indeed, the number has proven enough to reach statistical significance

among many of the reported hypothesis tests. However, a higher number of PSGs per task would in

general be desirable. More specifically, for those cases in which the reported trends did not achieve

significant effects, the question remains open on whether this could be attributed to the relative small

PSG sample size. Notice, on the other hand, that post-hoc power analyses were consciously omitted

because no useful conclusions are expected from them [48]. A higher sample size would also

contribute to spread the bias risk due to demographic and physiological subject variability.

Unfortunately, the chosen sample size was imposed by the available resources; thus this was not a

design parameter we were able to tune. As noticed, scoring of PSG data is complex and time-

consuming, and expert's time is expensive and scant.

In conclusion, our results provide an updated reference for inter-scorer agreement levels and scoring

times associated with both manual and semi-automatic scoring of PSG studies. We have

systematically analyzed and compared the resulting differences, showing that the use of semi-

automatic scoring can improve both speed and consistency of the PSG analysis outcomes. Benefits

include reduction of the associated diagnostic costs, shortening of patient's waiting lists, and overall

improved productivity. In addition, enhancement of inter-scorer agreement leads to higher repeatability

and quality of the diagnosis. More work has to be done to investigate generalization of these results by

increasing the subject sample and its heterogeneity. Future work should also assess the effects of

inter-center and intra-expert scoring variability, and goodness of fully automatic scoring in comparison

to manual and semi-automatic approaches.

DISCLOSURE STATEMENT

Financial Disclosure: none

Non-financial Disclosure: none

ACKNOWLEDGMENT

This study has been initiated at Haaglanden Medisch Centrum under project number 2019-073. The

study has also been partially funded under project ED431H 2020/10 of Xunta de Galicia. Authors wish

to acknowledge the support received from the Centro de Investigación de Galicia "CITIC", funded by

Xunta de Galicia and the European Union (European Regional Development Fund-Galicia 2014-2020

Program), by grant ED431G 2019/01.

REFERENCES

[1] R. Berry, R. Brooks, C. Gamaldo, S. Harding, R. Lloyd, S. Quan, M. Troester and B. Vaughn, The AASM manual for the scoring of sleep and associated events: rules, terminology and technical specifications, Version 2.4, Darien, IL: American Academy of Sleep Medicine, 2017.

[2] D. Alvarez-Estevez and I. Fernández-Varela, "Addressing database variability in learning from medical data: An ensemble-based approach using convolutional neural networks and a case of study applied to automatic sleep scoring," *Computers in Biology and Medicine,* vol. 119, no. 103697, 2020.

[3] R. Ferri, S. Fulda, R. Allen, M. Zucconi, O. Bruni, S. Chokroverty, L. Ferini-Strambi, B. Frauscher, D. Garcia-Borreguero, M. Hirshkowitz, B. H:ogl, Y. Inoue, A. Jahangir, M. Manconi, C. Marcus, D. Picchietti, G. Plazzi, J. Winkelman and R. Zak, "World Association of Sleep Medicine (WASM) 2016 standards for recroding and socring leg movements in polysomnograms developed by a joint task force from the International and the European Restless Legs Syndrome Study Group (IRLSSG and EURLSSG)," *Sleep Medicine,* vol. 26, pp. 86-95, 2016.

[4] R. Norman, I. Pal, C. Stewart, J. Walsleben and D. Rapoport, "Interobserver agreement among sleep scorers from different centers in a large dataset," *Sleep,* vol. 23, no. 7, pp. 901-908, 2000.

[5] S. Pittman, M. MacDonald, R. Fogel, A. Malhotra, K. Todros, B. Levy, A. Geva and D. White, "Assessment of automated scoring of polysomnographic recordings in a population with suspected sleep-disordered breathing," *Sleep,* vol. 27, no. 7, pp. 1394-1403, 2004.

[6] H. Danker-Hopfe, P. Anderer, J. Zeitlhofer, M. Boeck, H. Dorn, G. Gruber, E. Heller, E. Loretz, D. Moser, S. Parapatics, B. Saletu, A. Schmidt and G. Dorffner, "Interrater reliability for sleep scoring according to the Rechtschaffen & Kales and the new AASM standard," *Journal of Sleep Research,* vol. 18, pp. 74-84, 2009.

[7] C. Whitney, D. Gottlieb, S. Redline, R. Norman, R. Dodge, E. Shahar, S. Surovec en F. Nieto, „Reliability of scoring respiratory disturbance indices and sleep staging," *Sleep,* vol. 21, nr. 7, pp. 749-757, 1998.

[8] R. Rosenberg and S. van Hout, "The American Academy of Sleep Medicine inter-scorer reliability program: sleep stage scoring," *Journal of Clinical Sleep Medicine,* vol. 9, no. 1, pp. 81-87, 2013.

[9] D. Alvarez-Estevez and V. Moret-Bonillo, "Fuzzy reasoning used to detect apneic events in the sleep apnea-hypopnea syndrome," *Expert Systems with Applications,* vol. 36, no. 4, pp. 7778-7785, 2009.

[10] D. Alvarez-Estevez en V. Moret-Bonillo, „Computer-Assisted Diagnosis of the Sleep Apnea-Hypopnea Syndrome: A review," *Sleep Disorders,* vol. 2015, nr. 237878, 2015.

[11] D. Alvarez-Estevez, I. van Velzen, T. Ottolini-Capellen and B. Kemp, "Derivation and modeling of two new features for the characterization of rapid and slow eye movements in electrooculographic sleep recordings," *Biomedical Signal Processing and Control,* vol. 35, pp. 87-99, 2017.

[12] J. Stephansen, A. Olesen, M. Olsen, A. Ambati, E. Leary, H. Moore, O. Carrillo, L. Lin, F. Han, H. Yan, Y. Sun, Y. Dauvilliers, S. Scholz, L. Barateau, B. Hogl, A. Stefani, S. Hong, T. Kim, F. Pizza, G. Plazzi, S. Vandi, E. Antelmi, D. Perrin, S. Kuna and P. Schweitzer, "Neural network analysis of sleep stages enables efficient diagnosis of narcolepsy," *Nature Communications,* vol. 9, no. 5229, pp. 1-15, 2018.

[13] M. Cesari, A. Stefani, T. Penzel, A. Ibrahim, H. Hackner, A. Heidbreder, A. Szentkiralyi, B. Stubbe, H. Völzke, K. Berger and B. Högl, "Inter-rater sleep stage scoring reliability between manual scoring from two European sleep centers and automatic scoring performed by the artificial

intelligence-based Stanford-STAGES algorithm," *Journal of Clinical Sleep Medicine,* vol. 17, no. 6, pp. 1237-1247, 2021.

[14] L. Carvelli, A. Olesen, A. Brink-Kjaer, E. Leary, P. Peppard, E. Mignot, H. Sorensen and P. Jennum, "Design of a deep learning model for automatic scoring of periodic and non-periodic leg movements during sleep validated against multiple human experts," *Sleep Medicine,* vol. 69, pp. 109-119, 2020.

[15] A. Brink-Kjaer, A. Olesen, P. Peppard, K. Stone, P. Jennum, E. Mignot en H. Sorensen, „Automatic detection of cortical arousals in sleep and their contribution to daytime sleepiness," *Clinical Neurophysiology,* vol. 131, pp. 1187-1203, 2020.

[16] M. Cesari, M. Christensen, B. Mollenhauer, F. Sixel-Döring, H. Sorensen, C. Trenkwalder en P. Jennum, „A data-driven system to identify REM sleep behavior disorder and to predict its progression from the prodromal stage in Parkinson's disease," *Sleep Medicine,* vol. 77, pp. 238-248, 2021.

[17] S. Biswal, H. Sun, B. Goparaju, M. Westover, J. Sun and M. Bianchi, "Expert-level sleep scoring with deep neural networks," *Journal of the American Medical Informatics Association,* vol. 25, no. 12, pp. 1643-1650, 2018.

[18] B. Kemp and J. Olivan, "European data format 'plus' (EDF+), an EDF alike standard format for the exchange of physiological data," *Clinical Neurophysiology,* vol. 114, pp. 1755-1761, 2003.

[19] B. Kemp and M. Roessen, "Polyman: a free(ing) viewer for standard EDF(+) recordings and scorings," in *Sleep and Wake Research in The Netherlands*, Dutch Society for Wake-Sleep Research, 2007.

[20] D. Alvarez-Estevez and R. Rijsman, "Inter-database validation of a deep learning approach for automatic sleep scoring," *PLoS ONE,* vol. 16, no. 8, p. e0256111, 2021.

[21] D. Alvarez-Estevez, D. Wahid and R. Rijsman, "Validation of an automatic scoring algorithm for the analysis of periodic limb movements according to the WASM2016 guidelines," *Sleep Medicine,* vol. 40, pp. e13-e14, 2017.

[22] D. Alvarez-Estevez, "A new automatic method for the detection of limb movements and the analysis of their periodicity," *Biomedical Signal Processing and Control,* vol. 26, pp. 117-125, 2016.

[23] D. Alvarez-Estevez, D. Wahid and R. Rijsman, "Objective validation of an automatic desaturation detector on a clinical polygraphic database," in *Sleep-Wake Research in The Netherlands*, vol. 29, Dutch Society for Sleep-Wake Research, 2018.

[24] V. Moret-Bonillo, D. Alvarez-Estevez, A. Fernández-Leal and E. Hernández-Pereira, "Intelligent approach for analysis of respiratory signals and oxygen saturation in the Sleep Apnea/Hypopnea Syndrome," *The Open Medical Informatics Journal,* vol. 8, 2014.

[25] I. Fernández-Varela, D. Alvarez-Estevez, E. Hernández-Pereira and V. Moret-Bonillo, "A simple and robust method for the automatic scoring of EEG arousals in polysomnographic recordings," *Computers in Biology and Medicine,* vol. 87, pp. 77-86, 2017.

[26] D. Alvarez-Estevez and I. Fernández-Varela, "Large-scale validation of an automatic EEG arousal detection algorithm using different heterogeneous databases," *Sleep Medicine,* vol. 57, pp. 6-14, 2019.

[27] J. Cohen, "A coefficient of agreement for nominal scales," *Educational and Psychological Measurement,* vol. 20, pp. 37-46, 1960.

[28] M. Ohayon, E. Wickwire, M. Hirshkowitz, S. Albert, A. Avidan, F. Daly, Y. Dauvilliers, R. Ferri, C. Fung, D. Gozal, N. Hazen, A. Krystal, K. Lichstein, M. Mallampalli, G. Plazzi, R. Rawding, F. Scheer, V. Somers and M. Vitiello, "National Sleep Foundation's sleep quality recommendations: first report," *Sleep Health,* vol. 3, pp. 6-19, 2017.

[29] P. Shrout and J. Fleiss, "Intraclass correlations: uses in assessing rater reliability," *Psychological Bulletin,* vol. 86, no. 2, pp. 420-428, 1979.

[30] K. McGraw en S. Wong, „Forming inferences about some intraclass correlation coefficients," *Psychological Methods,* vol. 1, nr. 1, pp. 30-46, 1996.

[31] A. Salarian, „MATLAB Central File Exchange. Intraclass Correlation Coefficient (ICC)," [Online]. Available: https://www.mathworks.com/matlabcentral/fileexchange/22099-intraclass-correlation-coefficient-icc. [Geopend 25 February 2021].

[32] M. Roessen, M. Thijssen and B. Kemp, "Semi-automatic detection of leg movements: program features and scoring results," in *Sleep-Wake Research in The Netherlands*, Dutch Association for Wake-Sleep Research, 1998, pp. 101-105.

[33] M. Bonnet, D. Carley, M. Carskadon, P. Easton, C. Guilleminault, R. Harper, B. Hayes, M. Hirshkowitz, P. Ktonas, S. Keenan, M. Pressman, T. Roehrs, J. Smith, J. Walsh, S. Weber and P. Westbrook, "The Atlas Task Force: Recording and Scoring Leg Movements," *Sleep,* vol. 16, no. 8, pp. 748-759, 1993.

[34] W. Ruehland, T. Churchward, L. Schachter, T. Lakey, N. Tarquinio, F. O'Donoghue, M. Barnes and P. Rochford, "Polysomnography using abbreviated signal montages: impact on sleep and cortical arousal scoring," *Sleep Medicine,* vol. 16, pp. 173-180, 2015.

[35] M. Drinnan, A. Murray, C. Griffiths and J. Gibson, "Interobserver variability in recognizing arousal in respiratory sleep disorders," *Am J Respir Crit Care Med,* vol. 158, no. 2, pp. 358-362, 1998.

[36] D. Bliwise, S. Keenan, D. Burnburg, C. Mattice, P. Minkley, A. Pursley and J. Zelonis, "Inter-Rater Reliability for Scoring Periodic Leg Movements in Sleep," *Sleep,* vol. 14, no. 3, pp. 249-251, 1991.

[37] M. Smurra, M. Dury, G. Aubert, D. Rodenstein and G. Liistro, "Sleep fragmentation: comparison of two definitions of short arousals during sleep in OSAS patients," *European Respiratory Journal,* vol. 17, pp. 723-727, 2001.

[38] T. Wong, P. Galster, T. Lau, J. Lutz and C. Marcus, "Reliability of Scoring Arousals in Normal Children and Children with Obstructive Sleep Apnea Syndrome," *Sleep,* vol. 27, no. 6, pp. 1139-11145, 2004.

[39] J. Loredo, J. Clausen, S. Ancoli-Israel and J. Dimsdale, "Night-to-night arousal variability and interscorer reliability of arousal measurements," *Sleep,* vol. 22, no. 7, pp. 916-920, 1999.

[40] U. Magalang, N. Chen, P. Cistulli, A. Fedson, T. Gislason, D. Hillman, T. Penzel, R. Tamisier, S. Tufik, G. Philips and A. Pack, "Agreement in the scoring of respiratory events and sleep among international sleep centers," *Sleep,* vol. 36, no. 4, pp. 591-596, 2013.

[41] S. Kuna, R. Benca, C. Kushida, J. Walsh, M. Younes, B. Staley, A. Hanlon, A. Pack, G. Pien and A. Malhotra, "Agreement in computer-assisted manual scoring of polysomnograms across sleep centers," *Sleep,* vol. 4, no. 583-589, p. 36, 2013.

[42] R. Coleman, D. Bliwise, N. Sajben, A. Boomkamp, L. deBruyn and W. Dement, "Daytime sleepiness in patients with periodic movements in sleep," *Sleep,* vol. 5, pp. S191-202, 1982.

[43] The Atlas Task Force of the American Sleep Disorders Association, "EEG arousals: scoring rules

and examples: a preliminary report from the Sleep Disorders Atlas Task Force of the American Sleep Disorders Association," *Sleep,* vol. 15, pp. 173-184, 1992.

[44] M. Schipper, D. Alvarez-Estevez, K. Jellema, J. Verbraecken, S. Fulda and R. Rijsman, "Sleep-related leg movements in obstructive sleep apnea: definitions, determinants, and clinical consequences," *Sleep Medicine,* vol. 75, pp. 131-140, 2020.

[45] S. Lee, Y. Lee, C. Cho, H. Wang en K. Im, „Different scoring rules for respiratory event-related leg movements: effects on the prevalence of peridic limb movements during sleep and their association with depressive symptoms in patients with obstructive sleep apnea," *Sleep Medicine,* vol. 85, pp. 204-210, 2021.

[46] S. Fulda, "Periodic Leg Movements During Sleep," *Sleep Medicine Clinics,* vol. 16, no. 2, pp. 289-303, 2021.

[47] A. Malhotra, M. Younes, S. Kuna, R. Benca, C. Kushida, J. Walsh, A. Hanlon, B. Staley, A. Pack and G. Pien, "Performance of an automated polysomnography scoring system versus computer-assisted manual scoring," *Sleep,* vol. 36, no. 4, pp. 573-582, 2013.

[48] J. Hoenig and D. Heisey, "The abuse of power: The pervasive fallacy of power calculations for data analysis," *The American Statistician,* vol. 55, no. 1, 2001.

FIGURE CAPTIONS

Figure 1. Differences in scoring time between manual and semi-automatic scoring approaches.

Median scoring time per task is shown in minutes.

Figure 2. Agreement comparison between manual and semi-automatic scoring approaches. Median

Kappa agreement for each task is shown by aggregating all the expert pair combinations throughout

the five corresponding recordings.

TABLES

Table 1. Summary of general demographics and PSG descriptors in the study dataset. PSG descriptors correspond to values resulting from retrospective examination in the clinical database, i.e. prior to the multi-expert rescoring procedures carried out in this study. Distributions are characterized using the median and the corresponding interquartile ranges

Parameter	Group values				
	Subset A	Subset B	Subset C	Subset D	Total
n	5	5	5	5	20
Age (years)	52.0 [47.0, 57.0]	57.0 [51.0, 68.0]	59.0 [57.0, 61.0]	55.0 [52.0, 63.0]	57.0 [51.8, 61.5]
Male (n, %)	5 (100%)	1 (20%)	3 (60%)	3 (60%)	12 (60%)
Time In Bed (TIB, hours)	7.5 [7.4, 8.0]	7.3 [7.0, 7.3]	6.5 [6.4, 7.4]	8.1 [7.2, 8.2]	7.3 [7.0, 8.0]
Total Sleep Time (TST, hours)	5.9 [5.9, 7.1]	6.7 [5.9, 6.7]	6.0 [5.8, 6.1]	7.3 [5.5, 7.4]	6.0 [5.7, 7.0]
Sleep Latency (SL, min)	4.6 [3.0, 5.4]	1.8 [1.5, 3.3]	9.7 [2.6, 14.5]	1.9 [1.0, 21.5]	3.1 [1.6, 16.2]
Stage R latency (min)	69.5 [40.0, 174.0]	104 [64.5, 124.0]	59.0 [52.0, 138.0]	83.5 [80.0, 85.0]	81.8 [58.0, 127.0]
Wake After Sleep Onset (WASO, min)	72.8 [24.7, 124.0]	53.0 [39.9, 68.2]	38.9 [28.1, 52.2]	89.6 [46.7, 100.4]	52.6 [27.3, 106.3]
Sleep Efficiency (SE, %)	91.1 [74.2, 94.5]	88.6 [83.8, 90.9]	89.9 [84.7, 92.8]	83.0 [76.6, 90.5]	89.2 [76.0, 93.2]
Arousal Index (ArI, n/TST)	4.7 [1.8, 17.6]	14.3 [13.4, 16.0]	19.8 [9.6, 23.4]	9.6 [7.9, 13.7]	13.5 [7.5, 20.7]
Apnea-Hypopnea Index (AHI, n/TST)	9.1 [6.0, 9.4]	5.1 [0.6, 20.7]	6.2 [5.8, 13.1]	7.2 [3.0, 20.6]	6.7 [5.0, 20.6]
Oxygen Desaturation Index (ODI, n/TST)	10.7 [10.5, 10.8]	1.4 [0.6, 24.8]	9.9 [5.5, 15.4]	9.9 [2.8, 15.7]	10.2 [4.4, 18.0]
Leg Movement Index (n/TST)	52.3 [9.0, 56.5]	32.0 [21.0, 57.7]	14.2 [8.6, 39.8]	30.0 [19.3, 43.2]	31.0 [12.9, 53.4]
Periodic Leg Movement Index (PLMI, n/TST)	13.0 [0.3, 45.7]	13.8 [13.4, 49.0]	0.8 [0.4, 26.7]	22.4 [4.1, 28.4]	13.6 [0.7, 32.8]

Subset-scoring task correspondences: A - sleep stages; B - Limb Movements; C - Respiratory Events; D - EEG arousals

Table 2. Analysis of scoring time differences per task between manual and semi-automatic approaches. Distributions are characterized using the corresponding median and interquartile ranges in minutes. Gain factors are calculated on each case as the ratio between the corresponding median scoring times

Scoring task	Manual	Semi-Auto	Gain factor	*p*-value	Effect size
Sleep staging	32.62 [21.74, 48.64]	24.54 [16.25, 39.81]	1.33	0.0005*	0.4297
Limb movements	44.53 [31.00, 65.30]	18.50 [12.63, 26.73]	2.41	< 0.0001*	1.3475
Respiration	23.81 [17.62, 46.72]	14.58 [10.46, 20.68]	1.63	< 0.0001*	0.9474
EEG arousals	27.50 [21.22, 37.65]	21.78 [15.96, 28.59]	1.26	0.0011*	0.4233
Altogether	134.92 [113.08, 187.65]	80.59 [66.75, 107.89]	1.67	< 0.0001*	1.7527

n = 60 resulting from all scoring experts (n = 12) and analyzed recordings (n = 5) for each of the corresponding tasks;

*Statistically significant result

Table 3. Overall kappa inter-scorer agreement per scoring task and comparison between manual and semi-automatic approaches. Distributions are characterized using the corresponding median and interquartile ranges. TIB = Time in Bed

Scoring task	Context	Manual	Semi-auto	*p*-value	Effect size
Sleep staging	TIB	0.76 [0.69, 0.80]	0.80 [0.76, 0.83]	< 0.0001*	-0.7146
Limb Movements	TIB	0.72 [0.64, 0.79]	0.91 [0.86, 0.95]	< 0.0001*	-2.0223
	Wake	0.67 [0.57, 0.77]	0.89 [0.82, 0.94]	< 0.0001*	-1.7121
	Sleep	0.75 [0.65, 0.81]	0.92 [0.86, 0.95]	< 0.0001*	-1.6704
Respiratory activity	Apnea, Hypopnea, RERA (TIB)	0.55 [0.43, 0.78]	0.66 [0.53, 0.89]	< 0.0001*	-0.8315
	Apneas (TIB)	0.74 [0.35, 0.88]	0.88 [0.57, 0.98]	< 0.0001*	-0.3783
	Hypopneas (TIB)	0.46 [0.36, 0.53]	0.61 [0.51, 0.68]	< 0.0001*	-0.9569
	Desaturations (TIB)	0.40 [0.24, 0.55]	0.82 [0.71, 0.92]	< 0.0001*	-1.8525
EEG Arousals	TIB	0.58 [0.48, 0.65]	0.65 [0.56, 0.71]	< 0.0001*	-0.6166

*n = 330 resulting from all distinct combinations of expert scorer pairs (n = 66) on each of the five recordings related to the corresponding scoring task; *Statistically significant result*

Table 4. Comparison of inter-scoring agreement among diagnostic indices between manual and semi-automatic approaches. Agreement is characterized in terms of Interclass Correlation Coefficient (ICC), and corresponding index distributions using the respective median and interquartile ranges. r0 = null hypothesis on ICC score; SE = Sleep Efficiency; SOL = Sleep Onset Latency; WASO = Wake After Sleep Onset; (P)LMI = (Periodic) Leg Movement Index; AHI = Apnea-Hypopnea Index; AI = Apnea Index; HI = Hypopnea Index; ODI = Oxygen Desaturation Index; ArI = Arousal Index.

Index (TST)	Summary of index distributions		ICC		r0	ICC *p*-value	
	Manual	Semi-auto	Manual	Semi-auto		Manual	Semi-auto
SE (%)	89.11 [68.05, 94.87]	90.31 [70.59, 94.86]	0.9938	0.9974	0	< 0.0001*	< 0.0001*
					0.9938	---	0.0665
SOL (min)	5.27 2.50, 107.41]	4.11 [2.31, 104.81]	0.8720	0.8367	0	< 0.0001*	< 0.0001*
					0.8720	---	0.5589
WASO (min)	86.20 [22.27, 151.77]	79.18 [23.02, 139.51]	0.9924	0.9970	0	< 0.0001*	< 0.0001*
					0.9924	---	0.0481*
LMI	25.27 [19.78, 70.19]	31.73 [20.89, 61.82]	0.9227	0.9757	0	< 0.0001*	< 0.0001*
					0.9227	---	0.0166*
PLMI	13.91 [9.20, 61.70]	14.14 [12.79, 53.70]	0.9351	0.9771	0	< 0.0001*	< 0.0001*
					0.9351	---	0.0282*
AHI	5.86 [3.71, 12.47]	6.54 [4.70, 14.61]	0.9878	0.9924	0	< 0.0001*	< 0.0001*
					0.9878	---	0.1901
AI	1.26 [0.31, 3.46]	1.57 [0.00, 3.54]	0.9962	0.9986	0	< 0.0001*	< 0.0001*
					0.9962	---	0.0441*
HI	3.10 [2.38, 4.87]	4.55 [3.13, 6.68]	0.6010	0.7540	0	< 0.0001*	< 0.0001*
					0.6010	---	0.1160
ODI	6.48 [4.24, 14.38]	12.04 [6.01, 15.89]	0.8370	0.9843	0	< 0.0001*	< 0.0001*
					0.8370	---	< 0.0001*
ArI	17.57 [12.98, 25.64]	18.38 [13.82, 25.41]	0.6824	0.7673	0	< 0.0001*	< 0.0001*
					0.6824	---	0.2290

n = 60 resulting from all scoring experts (n = 12) and analyzed recordings (n = 5) for each of the corresponding

tasks;

*Statistically significant result