Validation of Reduced S-gene Target Performance and Failure for Rapid Surveillance of SARS-CoV-2 Variants

Cyndi Clark, PhD[1,3]; Joshua Schrecker, PharmD[1,3]; Matthew Hardison, PhD[1]; Michael S. Taitel, PhD[2]

[1]Aegis Sciences Corporation, Nashville, TN

[2]Walgreens, Deerfield, IL

[3]These authors contributed equally to this manuscript.

Corresponding Author: Michael S. Taitel, PhD Michael.taitel@walgreens.com

Funding Statement: This research received no specific grant from any funding agency in the public, commercial, or not-for-profit sectors.

Conflict of Interest Statement: The authors declare no conflict of interest in preparing this article.

Author Approval: all authors have seen and approved the manuscript.

Data Availability Statement: All data referred to in the manuscript will be made available in a deidentified dataset.

Abstract

SARS-CoV-2, the virus that causes COVID-19, has many variants capable of rapid transmission causing serious illness. Timely surveillance of new variants is essential for an effective public health response. Ensuring availability and access to diagnostic and molecular testing is key to this type of surveillance. This study utilized reverse transcription polymerase chain reaction (RT-PCR) and whole genome sequencing results from COVID-19-positive patient samples obtained through a collaboration between Aegis Sciences Corporation and Walgreens Pharmacy that has conducted more than 8.5 million COVID-19 tests at ~5,200 locations across the United States and Puerto Rico.

Viral evolution of SARS-CoV-2 can lead to mutations in the S-gene that cause reduced or failed S-gene amplification in diagnostic PCR tests. These anomalies, labeled reduced S-gene target performance (rSGTP) and S-gene target failure (SGTF), are characteristic of Alpha and Omicron (B.1.1.529, BA.1, and BA.1.1) lineages. This observation has been validated by whole genome sequencing and can provide presumptive lineage data following completion of diagnostic PCR testing in 24-48 hours from collection. Active surveillance of trends in PCR and sequencing results is key to the identification of changes in viral transmission and emerging variants. This study shows that rSGTP and SGTF can be utilized for near real-time tracking and surveillance of SARS-CoV-2 variants.

Introduction

Throughout the duration of the COVID-19 pandemic, mutations in the SARS-CoV-2 genome have led to the emergence of numerous variants. As of March 2022, over 1,900 unique lineages of SARS-CoV-2 have been identified via whole genome sequencing (1), a methodology that has proven to be of great importance to public health surveillance during the pandemic (2). Most new variants retain the same properties as their parent lineage, whereas others are classified as variants being monitored (VBM), variants of interest (VOI), variants of concern (VOC), or variants of high consequence (VOHC) based on an increased risk to global public health (3). Since December 2020, there have been six VOCs: Alpha, Beta, Gamma, Delta, Epsilon, and Omicron based on data showing a reduced response to diagnostics, treatment, or vaccines, evidence of increased transmissibility, and/or demonstration of increased disease severity (4). Subsequently, Alpha, Beta, Gamma, and Epsilon were reclassified as VBM amid the Delta surge in September of 2021. Omicron emerged in late November and has garnered much attention due to its significant increase in transmissibility and its expedient replacement of Delta as the predominant variant (5).

Surveillance data of circulating variants are typically available 2 weeks after the date of specimen collection due to the length of time required to complete whole genome sequencing. Efforts have been made to optimize sequencing methods and validate other methods to track the spread of viral lineages more efficiently. Some variants, such as Alpha and Omicron (B.1.1.529, BA.1, BA.1.1), have demonstrated unique patterns during real-time reverse transcriptase polymerase chain reaction (RT-PCR) testing, allowing for rapid assessment of their presumptive presence in a patient specimen. When using the ThermoFisher TaqPath™ COVID-19 Combo Kit RT-PCR assay to detect N, ORF1ab, and S-genes, Alpha's and Omicron's mutated S-gene prevents target amplification. This anomaly is labeled S-gene target failure (SGTF) (6) and is caused by deletions at positions 69 and 70 (del69-70) in the S protein (7). Alpha, the dominant variant of concern in early 2021, shares the characteristic S-gene mutation (del69-70) with Omicron lineages B.1.1.529, BA.1, and BA.1.1. Surveillance of positive PCR results exhibiting SGTF allows for the identification of variants, such as Omicron, simultaneously with a positive PCR result within 24-48 hours; whereas strain identification to confirm lineage via SARS-CoV-2 whole genome sequencing can take up to 2 weeks after initial diagnostic testing.

Although the use of SGTF for rapid surveillance of viral spread has been reported in peer-reviewed literature, the rates of transmission are likely underreported based on analysis of SGTF alone (8). This manuscript describes the validation of reduced S-gene target performance (rSGTP) analysis in addition to SGTF for early surveillance of variants with characteristic S-gene mutations, like del69-70. The use of additional measures for rapid surveillance of variants with rSGTP and SGTF allow for a real-time assessment of transmission and assist with identification of emerging variants during periods of transition.

Herein we show the use of a novel algorithm for rSGTP and SGTF based on the Ct values of the RT-PCR results as a proxy to rapidly assess the spread of certain SARS-CoV-2 lineages and validation of the algorithm with sequencing results in a randomized subset of specimens.

With the recent emergence of BA.2, the Omicron lineage that does not share the del69-70 mutation or exhibit rSGTP or SGTF, we were also able to track the trajectory of B.1.1.529/BA.1/BA.1.1 versus BA.2 in near real-time using the algorithm and confirmed our analyses with sequencing data, demonstrating the utility of this tool during viral evolution and how this methodology may inform future surveillance measures.

Methods

This study describes SARS-CoV-2 RT-PCR and whole genome sequencing results obtained via a collaboration between Aegis Sciences Corporation and Walgreens Pharmacy that has administered ~8.5 million COVID-19 tests at ~5,200 locations across the United States and Puerto Rico (9). RT-PCR testing utilizes primers and probes that bind to regions along the viral genome and release fluorescent compounds through repeated cycles of gene target amplification. Samples tested via the ThermoFisher TaqPath™ COVID-19 Combo Kit are positive if the fluorescent signal is detected above threshold in two out of three gene targets (10). SARS-CoV-2 lineages with characteristic S-gene mutations have altered amplification of the TaqPath™ S-gene target. This study analyzed rSGTP and SGTF for genome sequencing results meeting the following criteria:

- Reduced S-gene Target Performance (rSGTP):
    - Amplification of N and ORF1ab, but reduced amplification of S-gene.
    - S-gene Ct value $\geq$ 4 + Average (N-gene Ct value: ORF1ab-gene Ct value)
- S-Gene Target Failure (SGTF):
    - Amplification of N and ORF1ab genes, but no amplification of the S-gene.
    - S-gene Ct value > 37 or null

These data were generated in a subset of COVID-19 positive specimens with an average N-gene and ORF1ab-gene Ct value $\leq$ 30 due to Ct fluctuations and poor sequencing results in samples with high Ct values and lower viral titers. Presumed Alpha and Omicron

B.1.1.529/BA.1/BA.1.1 prevalence is calculated by taking the sum of rSGTP and SGTF RT-PCR positive results divided by the total number of RT-PCR positives for each collection date or a range of collection dates.

- Presumed Alpha or Omicron B.1.1.529/BA.1/BA.1.1 % = [(rSGTP + SGTF) / TOTAL POSITIVES] * 100.

Presumed Omicron BA.2 prevalence is calculated by subtracting the sum of rSGTP and SGTF RT-PCR positive results from the total number of RT-PCR positives divided by the total number of RT-PCR positives for each collection date or a range of collection dates.

- Presumed Omicron BA.2 % = [[TOTAL POSITIVES - (rSGTP + SGTF)] / TOTAL POSITIVES] * 100.

Next-generation sequencing (NGS) was performed using the Illumina® COVIDSeq Test and the NovaSeq 6000 sequencer. The Illumina® DRAGEN COVID Lineage App uses the FASTQ files to align reads to the SARS-CoV-2 reference genome (NC_045512.2) and reports coverage of targeted regions. Indeterminate (N) bases are assigned in regions covered with less than 20 unique reads or where the allele frequencies of A, T, G, or C are equal or too low to confidently make a call. Sequences with greater than 10% N bases were not included in the analysis. Variant calling, consensus sequence generation, and lineage/clade analysis is completed via Pangolin and NextClade pipelines.

Sensitivity, Specificity, Positive Predictive Value (PPV), and Negative Predictive Value (NPV) was determined for Presumed Alpha, Presumed Omicron B.1.1.529/BA.1/BA.1.1, and Presumed Omicron BA.2 calculations.

Results

A randomized subset of COVID-19 positive samples collected between 1/1/2021 - 3/23/2022 with an average N-gene and ORF1ab-gene Ct value $\leq$ 30 (N = 374,469) were sequenced. Lineages identified in these samples were: Alpha – 8.40% (n = 31,442), Delta – 60.15%

(n = 225,226), Omicron B.1.1.529/BA.1/BA.1.1 – 21.32% (n = 79,826), Omicron BA.2 – 1.00% (n = 3,737) and Other – 9.14% (n = 34,238) (Table 1 and Figure 1). S-gene target failure (SGTF) was dominant in Alpha (79.35%) and Omicron B.1.1.529/BA.1/BA.1.1 (92.18%) but was not exhibited in all cases (Table 1 and Figure 1). Analysis of the non-SGTF samples that were classified as Alpha and Omicron B.1.1.529/BA.1/BA.1.1 showed lower amplification of the S-gene target when compared to the N-gene and ORF1ab-gene Ct values within the same sample.

Reduced S-gene target performance (rSGTP) was validated by comparing the frequency that the S-gene amplified and the distribution of the Ct value differences between the S-gene and the average of N and ORF1ab-genes for each sample where the lineage was confirmed as Alpha, Delta, Omicron, or Other (Table 2). The Ct value differences amongst samples confirmed as Delta (Mean: 0.36, Median: 0.28), Omicron BA.2 (Mean: 0.49, Median: 0.42), or Other (Mean: 0.23, Median: 0.19) were negligible and the S-gene amplified in greater than 95% of cases. Conversely, S-gene amplification was observed in only 17.07% and 7.47% of Alpha and Omicron B.1.1.529/BA.1/BA.1.1 samples, respectively. Moreover, the Ct value of the S-gene was 3-4 cycles greater than the average Ct value of N and ORF1ab-genes in more than 90% of Alpha and Omicron B.1.1.529/BA.1/BA.1.1 cases (Figure 2). The S-gene amplified on average 5 cycles higher in Alpha and 6 cycles higher in Omicron B.1.1.529/BA.1/BA.1.1 samples (Figure 2). Using the algorithms for rSGTP: S-gene Ct value $\geq$ 4 + Average Ct value(N: ORF1ab) and SGTF: S-gene Ct value > 37, or null rSGTP + SGTF was observed in 96.42% of Alpha and 99.65% Omicron B.1.1.529/BA.1/BA.1.1 samples (Table 2).

Sensitivity and specificity of using rSGTP + SGTF or SGTF alone was calculated for Alpha, Omicron B.1.1.529/BA.1/BA.1.1, and Omicron BA.2 lineages that were confirmed by sequencing within the timeframe of first and last identification based on date of specimen collection (Table 3). Sensitivity improved by utilizing rSGTP + SGTF as a proxy for the identification of Alpha (96.4% vs. 79.3%) and Omicron B.1.1.529/BA.1/BA.1.1 (99.6% vs.

92.2%) in comparison to SGTF alone. Sensitivity did not change for the inverse of rSGTP + SGTF compared to the inverse of only SGTF in Omicron BA.2 (99.7%). Specificity remained consistent for both Alpha (99.5% vs. 99.6%) and Omicron B.1.1.529/BA.1/BA.1.1 (97.6% vs 97.8%) but increased for Omicron BA.2 (99.1 vs. 94.1%) when comparing rSGTP + SGTF to SGTF alone. The increase in sensitivity and negative predictive value (NPV) when rSGTP + SGTF is used as a proxy for surveillance of Alpha and Omicron B.1.1.529/BA.1/BA.1.1 lineages create a more accurate presumptive dataset with fewer false negatives. Furthermore, the utility of the inverse rSGTP + SGTF calculation is demonstrated by higher specificity and positive predictive value, or less false positives, in the presumed Omicron BA.2 dataset.

The prevalence of rSGTP + SGTF samples was graphed over time and compared to the prevalence of NGS-confirmed Alpha, Delta, Omicron B.1.1.529/BA.1/BA.1.1, and Omicron BA.2 (Figure 3). These data demonstrate the strong correlation between rSGTP + SGTF positives and the emergence and spread of both Alpha and Omicron B.1.1.529/BA.1/BA.1.1 VOC within the population and, for the first time, supports the use of reduced S-gene amplification (rSGTP) in samples as a proxy for early detection and surveillance of SARS-CoV-2 variants.

Discussion

SARS-CoV-2 genomes accumulate ~2 mutations per month, and over time, viral evolution gives rise to new lineages that are partitioned into phylogenetic trees and classified as variant strains (11). Mutations in variants of concern (VOC), such as Omicron and previously Alpha, pose a challenge for some PCR primers to anneal to the S-gene. The phenomenon defined as reduced S-gene target performance (rSGTP) and failure (SGTF) is characteristic of both Alpha and Omicron B.1.1.529/BA.1/BA.1.1 variants and can be used as a proxy to track emergence and spread of these and, potentially, future variants in real-time, enabling rapid dissemination of COVID-19 surveillance patterns. Additionally, the absence of rSGTP and SGTF has recently allowed for early assessment of transmission patterns of BA.2. Using the algorithm proposed for

rSGTP and SGTF as a proxy for surveillance of Omicron and other variants of concern with these characteristic PCR result patterns, significantly accelerates the ability to identify changes in viral spread versus an additional week or more necessary for viral genome sequencing.

Several previous publications have reported similar mechanisms for utilizing SGTF to increase efficiency of surveillance of certain variants of concern. To date, the publications have primarily focused on early detection of the Alpha and Omicron B.1.1.529/BA.1/BA.1.1. Investigators utilized similar methods for identification of SGTF as what is presented in this paper. This included identification of PCR results with amplification of the N and ORF1ab-genes in the absence of amplification for the S-gene. Similarly, other groups targeted PCR results with cycle threshold values ≤ 30 for the N and ORF1ab-genes to reduce the incidence of identifying a sample as exhibiting SGTF in cases where viral load was low, and amplification was near the limit of the detection of the assay. These publications typically focused on smaller sample sizes but did confirm SGTF as a proxy for identification of the Alpha and B.1.1.529/BA.1/BA.1.1 variants through subsequent genome sequencing (12-17). Similar to our results, the vast majority of samples with SGTF, which were expected to have been caused by commonly circulating variants at the time of the study, were confirmed as Alpha or Omicron B.1.1.529/BA.1/BA.1.1 via sequencing. Additionally, we determined sensitivity (92.2%) and specificity (97.8%) was similar to previously published calculations utilizing SGTF to detect Omicron (B.1.1.529, BA.1, BA.1.1 lineages) (18). Our analyses also revealed a significant proportion of Alpha and Omicron B.1.1.529/BA.1/BA.1.1 samples with S-gene amplification (Table 1). For this reason, we believe that utilization of SGTF alone for early surveillance of variants with del69-70 would underestimate the actual changes in transmission of these variants.

Migueres et al. reported a similar phenomenon that we identified in both Alpha and Omicron B.1.1.529/BA.1/BA.1.1 cases, which we refer to as rSGTP, though the analysis was performed

only during a time at which Alpha was in circulation (19). These investigators labeled samples

with reduced S-gene amplification in comparison to that of N and ORF1ab-genes as "S-gene

target late detection," or SGTL. More specifically, these samples were identified when S-gene

Ct values were ≥ 5 cycles higher than N or ORF1ab-genes and were reported as totaling less

than 1% of all Alpha cases. Our analysis defined samples as having rSGTP when the S-gene

cycle threshold value was at least 4 cycles higher than the average Ct values of N and ORF1ab,

and nearly 17% of Alpha and 8% of Omicron B.1.1.529/BA.1/BA.1.1 cases exhibited rSGTP.

As we endure an ever-changing pandemic, the development of novel tools to both diagnose

acute infections and to expedite identification of emerging variants will be critical during the

transition into an endemic phase. Collaborative efforts between laboratories and healthcare

providers to identify unique trends within large, shared datasets can provide support for public

health decision making. Those engaged in protecting public health during the COVID-19

pandemic are urged to adopt new methods to identify and track this virus as it continues to

mutate.

Table 1: Rates of rSGTP and SGTF in NGS confirmed lineages

Lineage Category	Total Samples Analyzed by Lineage	% of Total Samples Analyzed for Study	Total Samples with SGTF by Lineage	% of Samples with SGTF by Lineage	Total Samples with rSGTP by Lineage	% of Samples with rSGTP by Lineage	% of Samples with SGTF or rSGTP by Lineage
Alpha	31,442	8.40%	24,949	79.35%	5,367	17.07%	96.42%
Delta	225,226	60.15%	283	0.13%	151	0.07%	0.19%
Omicron B.1.1.529/ BA.1/ BA.1.1	79,829	23.32%	73,583	92.18%	5,960	7.47%	99.65%
Omicron BA.2	3,737	1.00%	11	0.29%	2	0.05%	0.35%
Other	34,238	9.14%	1,135	3.32%	123	0.36%	3.67%

Table 2: Ct value differences in NGS confirmed lineages with S-gene amplification

Lineage Category	Total Samples with S-gene Present	Mean Ct Value Difference (S-gene Ct value – AVG Ct value (N: ORF1ab-genes)	Median Ct Value Difference (S-gene Ct value – AVG Ct value (N: ORF1ab-genes)
Alpha	6,493	5.08	5.31
Delta	224,943	0.36	0.28
Omicron B.1.1.529/BA.1/BA.1.1	6,243	6.01	6.07
Omicron BA.2	3,726	0.49	0.42
Other	33,103	0.23	0.19

Figure 2: Distribution of Ct value Differences between the S-gene and the average of N and ORF1ab-genes in Alpha and Omicron B.1.1.529/BA.1/BA.1.1 variants

Alpha	B.1.1.529/BA.1/BA.1.1

Quantiles			Quantiles		
100.0%	maximum	18.57431425	100.0%	maximum	22.818301
99.5%		9.637467605	99.5%		11.96101157
97.5%		7.6214769	97.5%		8.3299015
90.0%		6.750757	90.0%		7.2518295
75.0%	quartile	6.0895925	75.0%	quartile	6.6525205
50.0%	median	5.3118435	50.0%	median	6.070979
25.0%	quartile	4.425855	25.0%	quartile	5.4155845
10.0%		3.1902082	10.0%		4.6954773
2.5%		0.1716099125	2.5%		3.22237485
0.5%		-0.773169885	0.5%		0.15341977
0.0%	minimum	-9.345805	0.0%	minimum	-15.025473

Summary Statistics			Summary Statistics	
Mean		5.0755514	Mean	6.0076117
Std Dev		1.7566684	Std Dev	1.430093
Std Err Mean		0.0218005	Std Err Mean	0.0180995
Upper 95% Mean		5.1182876	Upper 95% Mean	6.0430931
Lower 95% Mean		5.0328152	Lower 95% Mean	5.9721304
N		6493	N	6243

Table 3: Sensitivity, Specificity, PPV, and NPV of Proxy Calculations by Lineage

Presumed Alpha				
Alpha (Samples Collected: 1/1/2021-12/03/2021; n = 273,318)	Sensitivity	Specificity	PPV	NPV
SGTF	79.3%	99.6%	96.5%	97.4%
rSGTP + SGTF	96.4%	99.5%	96.4%	99.5%
Presumed Omicron B.1.1.529/BA.1/BA.1.1				
Omicron - B.1.1.529/BA.1/BA.1.1 (Samples Collected: 11/24/2021-3/23/2022; n = 105,646)	Sensitivity	Specificity	PPV	NPV
SGTF	92.2%	97.8%	99.2%	80.2%
rSGTP + SGTF	99.6%	97.6%	99.2%	98.9%
Presumed Omicron BA.2				
Omicron - BA.2 (Samples Collected: 1/3/2022-3/23/2022; n =: 59,920)	Sensitivity	Specificity	PPV	NPV
Inverse SGTF	99.7%	94.1%	53.1%	100.0%
Inverse rSGTP + SGTF	99.7%	99.1%	88.0%	100.0%

Figure 3: Surveillance of SARS-CoV-2 variants via rSGTP + SGTF tracking versus NGS-lineage confirmation

Date range truncated at March 2021 due to low sample volume for sequenced samples prior to this date

References:

1. SARS-CoV-2 lineages. Pango Lineages Website. https://cov-lineages.org/lineage_list.html. Accessed April 8, 2022.

2. Anaclerio F, Ferrante R, Mandatori D, Antonucci I, Capanna M, Damiani V, Di Tomo P, Ferrante R, Ranaudo M, De Laurenzi V, Stuppia L, De Fabritiis S. Different Strategies for the Identification of SARS-CoV-2 Variants in the Laboratory Practice. Genes (Basel). 2021; 12(9). doi: 10.3390/genes12091428

3. Tracking SARS-CoV-2 Variants. World Health Organization Website. https://www.who.int/en/activities/tracking-SARS-CoV-2-variants. Accessed April 8, 2022.

4. SARS-CoV-2 Variant Classifications and Definitions. Centers for Disease Control and Prevention Website. https://www.cdc.gov/coronavirus/2019-ncov/variants/variant-classifications.html Accessed April 8, 2022.

5. Centers for Disease Control and Prevention. Science Brief: Omicron (B.1.1.529) Variant. Updated: December 1, 2021 [cited 2021 Dec 26]. Available from: https://www.cdc.gov/coronavirus/2019-ncov/science/science-briefs/scientific-brief-omicron-variant.html.

6. Li A, Maier A, Carter M, Hugh Guan T. Omicron and S-gene target failure cases in the highest COVID-19 case rate region in Canada - December 2021 [published online ahead of print, 2021 Dec 29]. J Med Virol. 2021;10.1002/jmv.27562. doi:10.1002/jmv.27562.

7. Galloway S, Paul P, MacCannell D, Johansson M, Brooks J, MacNeil A, Slayton R, Tong S, Silk B, Armstrong G, Biggerstaff M, Dugan V. Emergence of SARS-CoV-2 B.1.1.7 Lineage – United States, December 29, 2020-January 12, 2021. Morbidity and Mortality Weekly Report (MMWR). 2021 Jan 22; 70(3).

8. Blomquist P, Bridgen J, Bray N, O'Connell AM, West D, Groves N, Gallagher E, Utsi L, Jarvis C, Hardstaff J, Byers C, Metelmann S, Simons D, Zaidi A, Twohig K, Savagar Bm

Lochen A, Ryan C, Wrenn K, Saavedra-Campos M, Abedin Z, Florence I, Cleary P, Elson R, Vivancos R, Lake I. Enhancing epidemiological surveillance of the emergence of the SARS-CoV-2 Omicron variant using spike gene target failure data, England, 15 November to 31 December 2021. Euro Surveill. 2022; 27(11);

https://doi.org/10.2807/1560-7917.ES.2022.27.11.2200143

9. Clark C, Singh T, Schrecker J, Feng L, Galanter J, Shah A, Smith-Ray R, Hardison M, Taitel M. Surveillance and epidemiology of SARS-CoV-2 viral variants in the United States, May-October 2021. JMIR Preprints. 25/11/2021:35060.

10. ThermoFisher Scientific. TaqPath™ COVID-19 Combo Kit and TaqPath™ COVID-19 Combo Kit Advanced INSTRUCTIONS FOR USE: Multiplex real-time RT-PCR test intended for the qualitative detection of nucleic acid from SARS-CoV-2. Updated: October 4, 2021 [Cited 2021 Dec 26]. Available from:

https://www.fda.gov/media/136112/download.

11. Harvey WT, Carabelli AM, Jackson, B, Gupta RK, Thomson EC, Harrison EM, Ludden C, Reeve R, Rambaut A, COVID-19 Geonomics UK (COG-UK) Consortium, Peacock SJ, Robertson DL. SARS-CoV-2 variants, spike mutations and immune escape. Nature Reviews Microbiology. 2021; 19, 409-424.

12. Blomquist PB, Bridgen J, Bray N, et al. Enhancing epidemiological surveillance of the emergence of the SARS-CoV-2 Omicron variant using spike gene target failure data, England, 15 November to 31 December 2021. *Euro Surveill.* 2022;27(11):10.2807/1560-7917.ES.2022.27.11.2200143. doi:10.2807/1560-7917.ES.2022.27.11.2200143.

13. Caputo V, Calvino G, Strafella C, et al. Tracking the Initial Diffusion of SARS-CoV-2 Omicron Variant in Italy by RT-PCR and Comparison with Alpha and Delta Variants Spreading. *Diagnostics (Basel)*. 2022;12(2):467. Published 2022 Feb 11. doi:10.3390/diagnostics12020467.

14. Eggink D, Andeweg SP, Vennema H, et al. Increased risk of infection with SARS-CoV-2 Omicron BA.1 compared with Delta in vaccinated and previously infected individuals, the

Netherlands, 22 November 2021 to 19 January 2022. *Euro Surveill.*
2022;27(4):2101196. doi:10.2807/1560-7917.ES.2022.27.4.2101196.

15. Wolter N, Jassat W, Walaza S, et al. Early assessment of the clinical severity of the SARS-CoV-2 omicron variant in South Africa: a data linkage study. *Lancet.* 2022;399(10323):437-446. doi:10.1016/S0140-6736(22)00017-4.

16. Li A, Maier A, Carter M, Guan TH. Omicron and S-gene target failure cases in the highest COVID-19 case rate region in Canada-December 2021. *J Med Virol.* 2022;94(5):1784-1786. doi:10.1002/jmv.27562.

17. Washington NL, Gangavarapu K, Zeller M, et al. Emergence and rapid transmission of SARS-CoV-2 B.1.1.7 in the United States. Cell. 2021;184(10):2587-2594.e7. doi:10.1016/j.cell.2021.03.052.

18. Accorsi EK, Britton A, Fleming-Dutra KE, et al. Association Between 3 Doses of mRNA COVID-19 Vaccine and Symptomatic Infection Caused by the SARS-CoV-2 Omicron and Delta Variants. JAMA. 2022;327(7):639-651. doi:10.1001/jama.2022.0470

19. Migueres M, Lhomme S, Trémeaux P, et al. Evaluation of two RT-PCR screening assays for identifying SARS-CoV-2 variants. J Clin Virol. 2021;143:104969. doi:10.1016/j.jcv.2021.104969